Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: CHAVARRIA-SORTO

DATE/TIME OF TEST: 02/01/2018 00:40:43

Diagnostics Check		OK	00:34
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.216	00:36
Breath Volume		2.210	00:36
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.215	00:39
Breath Volume		2.742	00:39
Air Blank		0.000	00:39
Control Test		0.080	00:40
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 02/02/2018 14:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PORCELLI,ANDREW,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PORCELLI,ANDREW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: MATHIAS

DATE/TIME OF TEST: 02/01/2018 16:59:28

Diagnostics Check		OK	16:48	
Air Blank		0.000	16:48	
Control Test		0.081	16:48	
Air Blank		0.000	16:49	
Subject Sample	# 1		16:52	No Sample Provided
Breath Volume		0.000	16:52	
Air Blank		0.000	16:52	
Air Blank		0.000	16:54	
Subject Sample	# 2		16:58	No Sample Provided
Breath Volume		0.000	16:58	
Air Blank		0.000	16:58	
Control Test		0.081	16:58	
Air Blank		0.000	16:59	
Diagnostics Check		OK	16:59	

Data Download Date/Time: 02/02/2018 14:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 01/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HAUGHTON

DATE/TIME OF TEST: 02/02/2018 02:19:36

Diagnostics Check		OK	02:11
Air Blank		0.000	02:12
Control Test		0.081	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.173	02:13
Breath Volume		1.539	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.167	02:18
Breath Volume		1.894	02:18
Air Blank		0.000	02:18
Control Test		0.080	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 02/02/2018 14:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: LOPEZ RODRIGUEZ

DATE/TIME OF TEST: 02/02/2018 19:10:03

Diagnostics Check		OK	19:02
Air Blank		0.000	19:03
Control Test		0.077	19:03
Air Blank		0.000	19:04
Subject Sample	# 1	0.000	19:05
Breath Volume		1.902	19:05
Air Blank		0.000	19:05
Air Blank		0.000	19:07
Subject Sample	# 2	0.000	19:08
Breath Volume		2.550	19:08
Air Blank		0.000	19:09
Control Test		0.078	19:09
Air Blank		0.000	19:09
Diagnostics Check		OK	19:10

Data Download Date/Time: 03/15/2018 14:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PECKO,MATTHEW
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: BUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: TELLAM

DATE/TIME OF TEST: 02/02/2018 20:18:28

Diagnostics Check		OK	20:15	
Air Blank		0.000	20:15	
Control Test		0.078	20:15	
Air Blank		0.000	20:16	
Subject Sample	# 1		20:17	Subject Test Refused
Breath Volume		0.000	20:17	
Air Blank		0.000	20:17	
Control Test		0.076	20:17	
Air Blank		0.000	20:18	
Diagnostics Check		OK	20:18	

Data Download Date/Time: 03/15/2018 14:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROSE,JORDAN,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: GROSE,JORDAN
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SUGRIM

DATE/TIME OF TEST: 02/03/2018 03:44:40

Diagnostics Check		OK	03:34
Air Blank		0.000	03:35
Control Test		0.080	03:35
Air Blank		0.000	03:36
Subject Sample	# 1	0.269	03:36
Breath Volume		1.832	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.246	03:40
Breath Volume		1.812	03:40
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 3	0.266	03:43
Breath Volume		2.480	03:43
Air Blank		0.000	03:43
Control Test		0.076	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 03/15/2018 14:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PORCELLI,ANDREW,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PORCELLI,ANDREW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:12

SUBJECT'S LAST NAME: NAIDOO

DATE/TIME OF TEST: 02/03/2018 12:46:15

Diagnostics Check		OK	12:34	
Air Blank		0.000	12:35	
Control Test		0.079	12:35	
Air Blank		0.000	12:35	
Subject Sample	# 1	0.156	12:39	Volume Not Met
Breath Volume		0.171	12:39	
Air Blank		0.000	12:39	
Air Blank		0.000	12:41	
Subject Sample	# 2	0.154	12:44	Volume Not Met
Breath Volume		0.082	12:44	
Air Blank		0.000	12:45	
Control Test		0.077	12:45	
Air Blank		0.000	12:46	
Diagnostics Check		OK	12:46	

Data Download Date/Time: 03/15/2018 14:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEEDHAM,K,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: LEEDHAM,K
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:46

SUBJECT'S LAST NAME: ZAYES

DATE/TIME OF TEST: 02/04/2018 16:12:40

Diagnostics Check		OK	16:06
Air Blank		0.000	16:06
Control Test		0.078	16:07
Air Blank		0.000	16:07
Subject Sample	# 1	0.254	16:08
Breath Volume		2.238	16:08
Air Blank		0.000	16:08
Air Blank		0.000	16:10
Subject Sample	# 2	0.247	16:11
Breath Volume		1.492	16:11
Air Blank		0.000	16:11
Control Test		0.077	16:12
Air Blank		0.000	16:12
Diagnostics Check		OK	16:12

Data Download Date/Time: 03/15/2018 14:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRA,CORP,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SERRA,CORP
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: OBRIEN

DATE/TIME OF TEST: 02/04/2018 21:32:28

Diagnostics Check		OK	21:25
Air Blank		0.000	21:25
Control Test		0.077	21:25
Air Blank		0.000	21:26
Subject Sample	# 1	0.159	21:27
Breath Volume		3.179	21:27
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.157	21:30
Breath Volume		2.898	21:30
Air Blank		0.000	21:31
Control Test		0.077	21:31
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 03/15/2018 14:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: ROGINSKI

DATE/TIME OF TEST: 02/07/2018 20:32:39

Diagnostics Check		OK	20:26
Air Blank		0.000	20:26
Control Test		0.077	20:27
Air Blank		0.000	20:27
Subject Sample	# 1	0.252	20:28
Breath Volume		1.503	20:28
Air Blank		0.000	20:28
Air Blank		0.000	20:30
Subject Sample	# 2	0.260	20:31
Breath Volume		1.609	20:31
Air Blank		0.000	20:31
Control Test		0.077	20:32
Air Blank		0.000	20:32
Diagnostics Check		OK	20:32

Data Download Date/Time: 03/15/2018 14:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTY,J,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CUTY,J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: LIQOURI

DATE/TIME OF TEST: 02/08/2018 11:10:15

Diagnostics Check		OK	11:04
Air Blank		0.000	11:04
Control Test		0.080	11:05
Air Blank		0.000	11:05
Subject Sample	# 1	0.000	11:06
Breath Volume		4.367	11:06
Air Blank		0.000	11:06
Air Blank		0.000	11:08
Subject Sample	# 2	0.000	11:08
Breath Volume		3.031	11:08
Air Blank		0.000	11:09
Control Test		0.076	11:09
Air Blank		0.000	11:10
Diagnostics Check		OK	11:10

Data Download Date/Time: 03/15/2018 14:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: WARTHEN

DATE/TIME OF TEST: 02/08/2018 18:11:07

Diagnostics Check		OK	18:04
Air Blank		0.000	18:05
Control Test		0.077	18:05
Air Blank		0.000	18:06
Subject Sample	# 1	0.173	18:06
Breath Volume		1.496	18:06
Air Blank		0.000	18:07
Air Blank		0.000	18:09
Subject Sample	# 2	0.170	18:09
Breath Volume		1.496	18:09
Air Blank		0.000	18:10
Control Test		0.077	18:10
Air Blank		0.000	18:11
Diagnostics Check		OK	18:11

Data Download Date/Time: 03/15/2018 14:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTY,J,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CUTY,J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:46

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 02/09/2018 16:16:30

Diagnostics Check		OK	16:07
Air Blank		0.000	16:07
Control Test		0.077	16:08
Air Blank		0.000	16:08
Subject Sample	# 1	0.215	16:09
Breath Volume		1.656	16:09
Air Blank		0.000	16:09
Air Blank		0.000	16:11
Subject Sample	# 2	0.192	16:12
Breath Volume		1.910	16:12
Air Blank		0.000	16:12
Air Blank		0.000	16:14
Subject Sample	# 3	0.196	16:14
Breath Volume		1.980	16:14
Air Blank		0.000	16:15
Control Test		0.077	16:15
Air Blank		0.000	16:16
Diagnostics Check		OK	16:16

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 02/10/2018 02:57:01

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.078	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.161	02:52
Breath Volume		1.707	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:54
Subject Sample	# 2	0.162	02:55
Breath Volume		1.554	02:55
Air Blank		0.000	02:56
Control Test		0.076	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:57

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: SOMMER

DATE/TIME OF TEST: 02/10/2018 18:54:10

Diagnostics Check		OK	18:46
Air Blank		0.000	18:47
Control Test		0.078	18:47
Air Blank		0.000	18:47
Subject Sample	# 1	0.137	18:49
Breath Volume		3.652	18:49
Air Blank		0.000	18:49
Air Blank		0.000	18:51
Subject Sample	# 2	0.133	18:52
Breath Volume		3.113	18:52
Air Blank		0.000	18:53
Control Test		0.077	18:53
Air Blank		0.000	18:54
Diagnostics Check		OK	18:54

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WYATT,PAMELA
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 02/10/2018 22:36:58

Diagnostics Check		OK	22:30
Air Blank		0.000	22:31
Control Test		0.078	22:31
Air Blank		0.000	22:32
Subject Sample	# 1	0.215	22:32
Breath Volume		3.542	22:32
Air Blank		0.000	22:33
Air Blank		0.000	22:35
Subject Sample	# 2	0.213	22:35
Breath Volume		3.453	22:35
Air Blank		0.000	22:36
Control Test		0.077	22:36
Air Blank		0.000	22:36
Diagnostics Check		OK	22:36

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 02/11/2018 02:06:42

Diagnostics Check		OK	01:59
Air Blank		0.000	02:00
Control Test		0.077	02:00
Air Blank		0.000	02:00
Subject Sample	# 1	0.268	02:01
Breath Volume		4.132	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:03
Subject Sample	# 2	0.268	02:05
Breath Volume		4.109	02:05
Air Blank		0.000	02:05
Control Test		0.077	02:06
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: ENGLISH

DATE/TIME OF TEST: 02/11/2018 21:14:03

Diagnostics Check		OK	21:07
Air Blank		0.000	21:07
Control Test		0.078	21:07
Air Blank		0.000	21:08
Subject Sample	# 1	0.175	21:09
Breath Volume		2.296	21:09
Air Blank		0.000	21:10
Air Blank		0.000	21:11
Subject Sample	# 2	0.165	21:12
Breath Volume		1.199	21:12
Air Blank		0.000	21:13
Control Test		0.076	21:13
Air Blank		0.000	21:13
Diagnostics Check		OK	21:14

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PORCELLI,ANDREW,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PORCELLI,ANDREW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 02/13/2018 13:28:58

Diagnostics Check		OK	13:23
Air Blank		0.000	13:23
Control Test		0.077	13:23
Air Blank		0.000	13:24
Subject Sample	# 1	0.000	13:24
Breath Volume		1.531	13:24
Air Blank		0.000	13:25
Air Blank		0.000	13:27
Subject Sample	# 2	0.000	13:27
Breath Volume		1.457	13:27
Air Blank		0.000	13:28
Control Test		0.078	13:28
Air Blank		0.000	13:28
Diagnostics Check		OK	13:28

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PORCELLI,ANDREW,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PORCELLI,ANDREW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: HULL

DATE/TIME OF TEST: 02/13/2018 17:24:42

Diagnostics Check		OK	17:18
Air Blank		0.000	17:18
Control Test		0.077	17:19
Air Blank		0.000	17:19
Subject Sample	# 1	0.102	17:20
Breath Volume		2.667	17:20
Air Blank		0.000	17:20
Air Blank		0.000	17:22
Subject Sample	# 2	0.107	17:23
Breath Volume		2.890	17:23
Air Blank		0.000	17:23
Control Test		0.077	17:24
Air Blank		0.000	17:24
Diagnostics Check		OK	17:24

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRA,CORP,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SERRA,CORP
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 02/13/2018 22:56:05

Diagnostics Check		OK	22:41	
Air Blank		0.000	22:42	
Control Test		0.078	22:42	
Air Blank		0.000	22:43	
Subject Sample	# 1	0.151	22:46	Volume Not Met
Breath Volume		0.000	22:46	
Air Blank		0.000	22:47	
Air Blank		0.000	22:48	
Subject Sample	# 2	0.184	22:50	
Breath Volume		2.183	22:50	
Air Blank		0.000	22:51	
Air Blank		0.000	22:53	
Subject Sample	# 3	0.194	22:54	
Breath Volume		1.445	22:54	
Air Blank		0.000	22:55	
Control Test		0.073	22:55	Control Outside Tolerance
Air Blank		0.000	22:56	

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRA,CORP,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SERRA,CORP
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 02/13/2018 23:14:42

Diagnostics Check		OK	23:03
Air Blank		0.000	23:04
Control Test		0.076	23:04
Air Blank		0.000	23:05
Subject Sample	# 1	0.183	23:05
Breath Volume		1.574	23:05
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.205	23:09
Breath Volume		1.980	23:09
Air Blank		0.000	23:09
Air Blank		0.000	23:11
Subject Sample	# 3	0.205	23:13
Breath Volume		2.527	23:13
Air Blank		0.000	23:13
Control Test		0.078	23:14
Air Blank		0.000	23:14
Diagnostics Check		OK	23:14

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: CROWE

DATE/TIME OF TEST: 02/15/2018 20:34:49

Diagnostics Check		OK	20:28
Air Blank		0.000	20:28
Control Test		0.078	20:29
Air Blank		0.000	20:29
Subject Sample	# 1	0.065	20:30
Breath Volume		1.710	20:30
Air Blank		0.000	20:30
Air Blank		0.000	20:32
Subject Sample	# 2	0.063	20:33
Breath Volume		1.765	20:33
Air Blank		0.000	20:33
Control Test		0.077	20:34
Air Blank		0.000	20:34
Diagnostics Check		OK	20:34

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROSE,JORDAN,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: GROSE,JORDAN
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: OLSON-LANZETTA

DATE/TIME OF TEST: 02/16/2018 22:43:40

Diagnostics Check		OK	22:36
Air Blank		0.000	22:36
Control Test		0.078	22:36
Air Blank		0.000	22:37
Subject Sample	# 1	0.214	22:38
Breath Volume		1.238	22:38
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.214	22:41
Breath Volume		2.070	22:41
Air Blank		0.000	22:42
Control Test		0.078	22:43
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SCHEMENAUER

DATE/TIME OF TEST: 02/18/2018 02:27:03

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.077	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.110	02:22
Breath Volume		1.761	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.109	02:25
Breath Volume		1.820	02:25
Air Blank		0.000	02:26
Control Test		0.077	02:26
Air Blank		0.000	02:26
Diagnostics Check		OK	02:27

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: AGUILAR CHUN

DATE/TIME OF TEST: 02/21/2018 00:44:43

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.078	00:36
Air Blank		0.000	00:37
Subject Sample	# 1	0.220	00:39
Breath Volume		1.726	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.221	00:43
Breath Volume		2.214	00:43
Air Blank		0.000	00:43
Control Test		0.077	00:44
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SINDICIC

DATE/TIME OF TEST: 02/21/2018 03:40:14

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.078	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.236	03:35
Breath Volume		3.023	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.240	03:38
Breath Volume		3.066	03:38
Air Blank		0.000	03:39
Control Test		0.076	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATEL,MITHIL,V
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PATEL,MITHIL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: CORBIN

DATE/TIME OF TEST: 02/23/2018 17:10:12

Diagnostics Check		OK	17:02
Air Blank		0.000	17:02
Control Test		0.078	17:03
Air Blank		0.000	17:03
Subject Sample	# 1	0.184	17:04
Breath Volume		1.804	17:04
Air Blank		0.000	17:04
Air Blank		0.000	17:06
Subject Sample	# 2	0.192	17:08
Breath Volume		1.179	17:08
Air Blank		0.000	17:09
Control Test		0.079	17:09
Air Blank		0.000	17:10
Diagnostics Check		OK	17:10

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: FREITAS

DATE/TIME OF TEST: 02/24/2018 01:53:38

Diagnostics Check		OK	01:47
Air Blank		0.000	01:47
Control Test		0.079	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.141	01:49
Breath Volume		1.839	01:49
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.137	01:52
Breath Volume		1.628	01:52
Air Blank		0.000	01:52
Control Test		0.078	01:53
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/24/2018 02:36:24

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.077	02:31
Air Blank		0.000	02:31
Subject Sample	# 1	0.197	02:31
Breath Volume		3.152	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.198	02:34
Breath Volume		2.980	02:34
Air Blank		0.000	02:35
Control Test		0.077	02:35
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/26/2018 22:56:28

Diagnostics Check		OK	22:49
Air Blank		0.000	22:49
Control Test		0.079	22:50
Air Blank		0.000	22:50
Subject Sample	# 1	0.000	22:52
Breath Volume		1.320	22:52
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.000	22:55
Breath Volume		1.335	22:55
Air Blank		0.000	22:55
Control Test		0.079	22:55
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: CONCEPCION

DATE/TIME OF TEST: 02/28/2018 19:58:07

Diagnostics Check		OK	19:51
Air Blank		0.000	19:52
Control Test		0.080	19:52
Air Blank		0.000	19:53
Subject Sample	# 1	0.131	19:53
Breath Volume		2.449	19:53
Air Blank		0.000	19:54
Air Blank		0.000	19:56
Subject Sample	# 2	0.128	19:56
Breath Volume		2.347	19:56
Air Blank		0.000	19:57
Control Test		0.080	19:57
Air Blank		0.000	19:58
Diagnostics Check		OK	19:58

Data Download Date/Time: 03/15/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: THAKKAR,AMIT
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: DE ZAYAS

DATE/TIME OF TEST: 02/03/2018 22:20:11

Diagnostics Check		OK	22:10	
Air Blank		0.000	22:11	
Control Test		0.080	22:11	
Air Blank		0.000	22:12	
Subject Sample	# 1		22:15	No Sample Provided
Breath Volume		0.000	22:15	
Air Blank		0.000	22:15	
Air Blank		0.000	22:17	
Subject Sample	# 2		22:18	Subject Test Refused
Breath Volume		0.000	22:18	
Air Blank		0.000	22:19	
Control Test		0.082	22:19	
Air Blank		0.000	22:20	
Diagnostics Check		OK	22:20	

Data Download Date/Time: 02/27/2018 14:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: WILSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: BELKIN

DATE/TIME OF TEST: 02/09/2018 05:09:00

Diagnostics Check		OK	05:01
Air Blank		0.000	05:02
Control Test		0.081	05:02
Air Blank		0.000	05:03
Subject Sample	# 1	0.162	05:04
Breath Volume		2.453	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:06
Subject Sample	# 2	0.155	05:07
Breath Volume		2.421	05:07
Air Blank		0.000	05:08
Control Test		0.081	05:08
Air Blank		0.000	05:08
Diagnostics Check		OK	05:09

Data Download Date/Time: 02/27/2018 14:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELO,FRANCISCO,J
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: AGUIRRE,RICARDO
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: HERNANDEZ RAYMUNDO

DATE/TIME OF TEST: 02/09/2018 22:31:05

Diagnostics Check		OK	22:22	
Air Blank		0.000	22:22	
Control Test		0.081	22:22	
Air Blank		0.000	22:23	
Subject Sample	# 1		22:26	No Sample Provided
Breath Volume		0.000	22:26	
Air Blank		0.000	22:26	
Air Blank		0.000	22:28	
Subject Sample	# 2		22:29	Subject Test Refused
Breath Volume		0.000	22:29	
Air Blank		0.000	22:30	
Control Test		0.080	22:30	
Air Blank		0.000	22:31	
Diagnostics Check		OK	22:31	

Data Download Date/Time: 02/27/2018 14:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/15/2018 20:18:02

Diagnostics Check		OK	20:10
Air Blank		0.000	20:11
Control Test		0.081	20:11
Air Blank		0.000	20:12
Subject Sample	# 1	0.155	20:12
Breath Volume		1.691	20:12
Air Blank		0.000	20:13
Air Blank		0.000	20:15
Subject Sample	# 2	0.158	20:16
Breath Volume		3.113	20:16
Air Blank		0.000	20:17
Control Test		0.078	20:17
Air Blank		0.000	20:17
Diagnostics Check		OK	20:18

Data Download Date/Time: 02/27/2018 14:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ZAMORA CRIBAS

DATE/TIME OF TEST: 02/17/2018 01:47:29

Diagnostics Check		OK	01:40	
Air Blank		0.000	01:40	
Control Test		0.081	01:41	
Air Blank		0.000	01:41	
Subject Sample	# 1	0.254	01:42	
Breath Volume		2.304	01:42	
Air Blank		0.000	01:43	
Air Blank		0.000	01:45	
Subject Sample	# 2	0.243	01:45	
Breath Volume		1.703	01:45	
Air Blank		0.000	01:46	
Control Test		0.080	01:46	
Air Blank			01:47	RFI Detect
Air Blank		0.000	01:47	

Data Download Date/Time: 02/27/2018 14:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ZAMORA CRIBAS

DATE/TIME OF TEST: 02/17/2018 01:56:34

Diagnostics Check		OK	01:49
Air Blank		0.000	01:50
Control Test		0.079	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.241	01:51
Breath Volume		1.960	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:53
Subject Sample	# 2	0.243	01:54
Breath Volume		2.089	01:54
Air Blank		0.000	01:55
Control Test		0.080	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 02/27/2018 14:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/17/2018 21:00:25

Diagnostics Check		OK	20:53
Air Blank		0.000	20:53
Control Test		0.081	20:53
Air Blank		0.000	20:54
Subject Sample	# 1	0.183	20:55
Breath Volume		2.312	20:55
Air Blank		0.000	20:55
Air Blank		0.000	20:57
Subject Sample	# 2	0.185	20:58
Breath Volume		1.390	20:58
Air Blank		0.000	20:59
Control Test		0.079	20:59
Air Blank		0.000	21:00
Diagnostics Check		OK	21:00

Data Download Date/Time: 02/27/2018 14:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,GEORGE,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEARCE,AARON
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 13:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:47

SUBJECT'S LAST NAME: CRISP

DATE/TIME OF TEST: 02/18/2018 14:19:53

Diagnostics Check		OK	14:13
Air Blank		0.000	14:13
Control Test		0.081	14:13
Air Blank		0.000	14:14
Subject Sample	# 1	0.119	14:15
Breath Volume		1.531	14:15
Air Blank		0.000	14:15
Air Blank		0.000	14:17
Subject Sample	# 2	0.134	14:18
Breath Volume		2.695	14:18
Air Blank		0.000	14:18
Control Test		0.081	14:19
Air Blank		0.000	14:19
Diagnostics Check		OK	14:19

Data Download Date/Time: 02/27/2018 14:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEWTON,JOSHUA,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: NEWTON,JOSHUA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: SUTHERLAND

DATE/TIME OF TEST: 02/20/2018 03:55:40

Diagnostics Check		OK	03:47
Air Blank		0.000	03:48
Control Test		0.081	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.175	03:50
Breath Volume		1.964	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.171	03:53
Breath Volume		1.886	03:53
Air Blank		0.000	03:54
Control Test		0.077	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 02/27/2018 14:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA,MAYKEL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACOSTA,MAYKEL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: ALVAREZ RUIZ

DATE/TIME OF TEST: 02/22/2018 02:40:19

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.080	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.191	02:34
Breath Volume		1.628	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.197	02:38
Breath Volume		2.242	02:38
Air Blank		0.000	02:39
Control Test		0.080	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 02/27/2018 14:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS,HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: KIRK

DATE/TIME OF TEST: 02/07/2018 21:40:53

Diagnostics Check		OK	21:34
Air Blank		0.000	21:34
Control Test		0.078	21:34
Air Blank		0.000	21:35
Subject Sample	# 1	0.058	21:36
Breath Volume		2.332	21:36
Air Blank		0.000	21:36
Air Blank		0.000	21:38
Subject Sample	# 2	0.058	21:39
Breath Volume		2.335	21:39
Air Blank		0.000	21:39
Control Test		0.078	21:40
Air Blank		0.000	21:40
Diagnostics Check		OK	21:40

Data Download Date/Time: 03/05/2018 09:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON,JOSH,G
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: EICHNER,AMANDA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:09

SUBJECT'S LAST NAME: SOUTHARD

DATE/TIME OF TEST: 02/14/2018 15:36:41

Diagnostics Check		OK	15:30
Air Blank		0.000	15:30
Control Test		0.078	15:31
Air Blank		0.000	15:31
Subject Sample	# 1	0.000	15:32
Breath Volume		1.679	15:32
Air Blank		0.000	15:32
Air Blank		0.000	15:34
Subject Sample	# 2	0.000	15:35
Breath Volume		3.312	15:35
Air Blank		0.000	15:35
Control Test		0.076	15:36
Air Blank		0.000	15:36
Diagnostics Check		OK	15:36

Data Download Date/Time: 03/05/2018 09:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: YEARTY,WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 02/22/2018 04:02:13

Diagnostics Check		OK	03:55
Air Blank		0.000	03:55
Control Test		0.078	03:56
Air Blank		0.000	03:56
Subject Sample	# 1	0.070	03:57
Breath Volume		2.218	03:57
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.069	04:00
Breath Volume		2.312	04:00
Air Blank		0.000	04:01
Control Test		0.078	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 03/05/2018 09:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: OHARROW

DATE/TIME OF TEST: 02/25/2018 19:22:59

Diagnostics Check		OK	19:08
Air Blank		0.000	19:09
Control Test		0.078	19:09
Air Blank		0.000	19:10
Subject Sample	# 1	0.203	19:12
Breath Volume		1.332	19:12
Air Blank		0.000	19:13
Air Blank		0.000	19:15
Subject Sample	# 2	0.176	19:15
Breath Volume		1.621	19:15
Air Blank		0.000	19:16
Air Blank		0.000	19:18
Subject Sample	# 3	0.184	19:21
Breath Volume		1.371	19:21
Air Blank		0.000	19:22
Control Test		0.076	19:22
Air Blank		0.000	19:22
Diagnostics Check		OK	19:22

Data Download Date/Time: 03/05/2018 09:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: KRASINSKAS,RICHARD
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 01/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: HOLM

DATE/TIME OF TEST: 02/10/2018 20:29:08

Diagnostics Check		OK	20:20	
Air Blank		0.000	20:21	
Control Test		0.081	20:21	
Air Blank		0.000	20:21	
Subject Sample	# 1		20:25	No Sample Provided
Breath Volume		0.000	20:25	
Air Blank		0.000	20:25	
Air Blank		0.000	20:27	
Subject Sample	# 2		20:27	Subject Test Refused
Breath Volume		0.000	20:27	
Air Blank		0.000	20:28	
Control Test		0.081	20:28	
Air Blank		0.000	20:29	
Diagnostics Check		OK	20:29	

Data Download Date/Time: 02/23/2018 20:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,MICHAEL,K
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: REED,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: PALLADINO

DATE/TIME OF TEST: 02/21/2018 23:21:31

Diagnostics Check		OK	23:13
Air Blank		0.000	23:14
Control Test		0.081	23:14
Air Blank		0.000	23:14
Subject Sample	# 1	0.000	23:16
Breath Volume		1.933	23:16
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.000	23:20
Breath Volume		2.890	23:20
Air Blank		0.000	23:20
Control Test		0.081	23:20
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 02/23/2018 20:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CHALKER,S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 01/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: MCCLURE

DATE/TIME OF TEST: 02/05/2018 03:33:09

Diagnostics Check		OK	03:19	
Air Blank		0.000	03:19	
Control Test		0.081	03:20	
Air Blank		0.000	03:20	
Subject Sample	# 1	0.195	03:23	Volume Not Met
Breath Volume		0.457	03:23	
Air Blank		0.000	03:24	
Air Blank		0.000	03:26	
Subject Sample	# 2	0.188	03:27	
Breath Volume		1.515	03:27	
Air Blank		0.000	03:28	
Air Blank		0.000	03:29	
Subject Sample	# 3	0.199	03:31	
Breath Volume		1.949	03:31	
Air Blank		0.000	03:32	
Control Test		0.079	03:32	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:33	

Data Download Date/Time: 02/19/2018 16:14

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROOKS,E,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JETT,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:51

SUBJECT'S LAST NAME: MARKHAM

DATE/TIME OF TEST: 02/14/2018 07:19:48

Diagnostics Check		OK	07:12
Air Blank		0.000	07:13
Control Test		0.080	07:13
Air Blank		0.000	07:14
Subject Sample	# 1	0.089	07:14
Breath Volume		3.007	07:14
Air Blank		0.000	07:15
Air Blank		0.000	07:17
Subject Sample	# 2	0.086	07:18
Breath Volume		3.062	07:18
Air Blank		0.000	07:18
Control Test		0.079	07:19
Air Blank		0.000	07:19
Diagnostics Check		OK	07:19

Data Download Date/Time: 02/19/2018 16:14

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,C,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CHALKER,S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 01/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BADEA

DATE/TIME OF TEST: 02/17/2018 04:10:10

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:04
Subject Sample	# 1	0.175	04:04
Breath Volume		1.183	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.188	04:08
Breath Volume		1.554	04:08
Air Blank		0.000	04:09
Control Test		0.080	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 02/19/2018 16:14

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAUNDRS,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SYDOROWICZ,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:09

SUBJECT'S LAST NAME: LYVERS

DATE/TIME OF TEST: 02/18/2018 14:32:46

Diagnostics Check		OK	14:29	
Air Blank		0.000	14:29	
Control Test		0.079	14:29	
Air Blank		0.000	14:30	
Subject Sample	# 1		14:31	RFI Detect
Breath Volume		0.015	14:31	
Air Blank		0.000	14:31	
Control Test		0.079	14:32	
Air Blank		0.000	14:32	
Diagnostics Check		OK	14:32	

Data Download Date/Time: 02/19/2018 16:14

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAUNDERS,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SYDOROWICZ,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:09

SUBJECT'S LAST NAME: LYVERS

DATE/TIME OF TEST: 02/18/2018 14:42:56

Diagnostics Check		OK	14:34	
Air Blank		0.000	14:35	
Control Test		0.079	14:35	
Air Blank		0.000	14:35	
Subject Sample	# 1	0.220	14:37	
Breath Volume		1.191	14:37	
Air Blank		0.000	14:37	
Air Blank		0.000	14:39	
Subject Sample	# 2		14:41	RFI Detect
Breath Volume		0.898	14:41	
Air Blank		0.000	14:42	
Control Test		0.078	14:42	
Air Blank		0.000	14:42	
Diagnostics Check		OK	14:42	

Data Download Date/Time: 02/19/2018 16:14

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAUNDERS,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SYDOROWICZ,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:09

SUBJECT'S LAST NAME: LYVERS

DATE/TIME OF TEST: 02/18/2018 14:53:37

Diagnostics Check		OK	14:44
Air Blank		0.000	14:45
Control Test		0.077	14:45
Air Blank		0.000	14:46
Subject Sample	# 1	0.214	14:46
Breath Volume		1.316	14:46
Air Blank		0.000	14:47
Air Blank		0.000	14:49
Subject Sample	# 2	0.214	14:51
Breath Volume		1.945	14:51
Air Blank		0.000	14:52
Control Test		0.078	14:52
Air Blank		0.000	14:53
Diagnostics Check		OK	14:53

Data Download Date/Time: 02/19/2018 16:14

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 02/28/2018 17:03:12

Diagnostics Check		OK	16:48	
Air Blank		0.000	16:49	
Control Test		0.081	16:49	
Air Blank		0.000	16:49	
Subject Sample	# 1	0.000	16:50	
Breath Volume		1.250	16:50	
Air Blank		0.000	16:50	
Air Blank		0.000	16:52	
Subject Sample	# 2	0.000	16:56	Volume Not Met
Breath Volume		0.574	16:56	
Air Blank		0.000	16:56	
Air Blank		0.000	16:58	
Subject Sample	# 3	0.000	17:01	Volume Not Met
Breath Volume		0.085	17:01	
Air Blank		0.000	17:02	
Control Test		0.079	17:02	
Air Blank		0.000	17:03	
Diagnostics Check		OK	17:03	

Data Download Date/Time: 03/24/2018 07:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGOG,DEPUTY,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 02/28/2018 17:16:53

Diagnostics Check		OK	17:10
Air Blank		0.000	17:10
Control Test		0.079	17:11
Air Blank		0.000	17:11
Subject Sample	# 1	0.000	17:12
Breath Volume		2.195	17:12
Air Blank		0.000	17:12
Air Blank		0.000	17:14
Subject Sample	# 2	0.000	17:15
Breath Volume		1.753	17:15
Air Blank		0.000	17:15
Control Test		0.080	17:16
Air Blank		0.000	17:16
Diagnostics Check		OK	17:16

Data Download Date/Time: 03/24/2018 07:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 02/03/2018 02:35:30

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.079	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.180	02:30
Breath Volume		1.398	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.170	02:33
Breath Volume		1.843	02:33
Air Blank		0.000	02:34
Control Test		0.077	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 03/09/2018 12:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 02/03/2018 03:12:36

Diagnostics Check		OK	03:00	
Air Blank		0.000	03:01	
Control Test		0.078	03:01	
Air Blank		0.000	03:02	
Subject Sample	# 1	0.127	03:05	Volume Not Met
Breath Volume		0.636	03:05	
Air Blank		0.000	03:06	
Air Blank		0.000	03:07	
Subject Sample	# 2	0.100	03:10	Volume Not Met
Breath Volume		0.683	03:10	
Air Blank		0.000	03:11	
Control Test		0.077	03:11	
Air Blank		0.000	03:12	
Diagnostics Check		OK	03:12	

Data Download Date/Time: 03/09/2018 12:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: APPLEMAN

DATE/TIME OF TEST: 02/03/2018 05:12:36

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.079	05:04
Air Blank		0.000	05:04
Subject Sample	# 1	0.164	05:07
Breath Volume		1.246	05:07
Air Blank		0.000	05:08
Air Blank		0.000	05:10
Subject Sample	# 2	0.153	05:11
Breath Volume		1.222	05:11
Air Blank		0.000	05:11
Control Test		0.076	05:12
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 03/09/2018 12:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 02/04/2018 01:42:30

Diagnostics Check		OK	01:34
Air Blank		0.000	01:34
Control Test		0.079	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.251	01:37
Breath Volume		1.769	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:39
Subject Sample	# 2	0.249	01:40
Breath Volume		2.335	01:40
Air Blank		0.000	01:41
Control Test		0.077	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 03/09/2018 12:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SIFUENTES OVALLE

DATE/TIME OF TEST: 02/04/2018 04:01:56

Diagnostics Check		OK	03:54
Air Blank		0.000	03:54
Control Test		0.079	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.220	03:57
Breath Volume		2.343	03:57
Air Blank		0.000	03:58
Air Blank		0.000	03:59
Subject Sample	# 2	0.216	04:00
Breath Volume		1.511	04:00
Air Blank		0.000	04:01
Control Test		0.077	04:01
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 03/09/2018 12:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: MONTANO

DATE/TIME OF TEST: 02/04/2018 04:19:03

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.078	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.180	04:14
Breath Volume		1.765	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:16
Subject Sample	# 2	0.177	04:17
Breath Volume		2.132	04:17
Air Blank		0.000	04:18
Control Test		0.076	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:19

Data Download Date/Time: 03/09/2018 12:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WRIGHT,DELSHAWNTA
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: CARDENAS

DATE/TIME OF TEST: 02/04/2018 07:46:05

Diagnostics Check		OK	07:38
Air Blank		0.000	07:38
Control Test		0.079	07:39
Air Blank		0.000	07:39
Subject Sample	# 1	0.144	07:41
Breath Volume		1.375	07:41
Air Blank		0.000	07:42
Air Blank		0.000	07:44
Subject Sample	# 2	0.146	07:44
Breath Volume		1.445	07:44
Air Blank		0.000	07:45
Control Test		0.077	07:45
Air Blank		0.000	07:46
Diagnostics Check		OK	07:46

Data Download Date/Time: 03/09/2018 12:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MEBANE

DATE/TIME OF TEST: 02/04/2018 22:36:34

Diagnostics Check		OK	22:30
Air Blank		0.000	22:31
Control Test		0.079	22:31
Air Blank		0.000	22:31
Subject Sample	# 1	0.167	22:32
Breath Volume		1.574	22:32
Air Blank		0.000	22:32
Air Blank		0.000	22:34
Subject Sample	# 2	0.167	22:35
Breath Volume		1.539	22:35
Air Blank		0.000	22:35
Control Test		0.078	22:35
Air Blank		0.000	22:36
Diagnostics Check		OK	22:36

Data Download Date/Time: 03/09/2018 12:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: LONGEST

DATE/TIME OF TEST: 02/04/2018 23:29:10

Diagnostics Check		OK	23:17	
Air Blank		0.000	23:17	
Control Test		0.079	23:18	
Air Blank		0.000	23:18	
Subject Sample	# 1	0.116	23:21	Volume Not Met
Breath Volume		0.148	23:21	
Air Blank		0.000	23:22	
Air Blank		0.000	23:24	
Subject Sample	# 2	0.113	23:27	Volume Not Met
Breath Volume		0.164	23:27	
Air Blank		0.000	23:28	
Control Test		0.078	23:28	
Air Blank		0.000	23:29	
Diagnostics Check		OK	23:29	

Data Download Date/Time: 03/09/2018 12:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: RIVAS GUEVARA

DATE/TIME OF TEST: 02/05/2018 00:36:17

Diagnostics Check		OK	00:29
Air Blank		0.000	00:29
Control Test		0.078	00:30
Air Blank		0.000	00:30
Subject Sample	# 1	0.246	00:31
Breath Volume		2.367	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.249	00:34
Breath Volume		2.496	00:34
Air Blank		0.000	00:35
Control Test		0.078	00:35
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 03/09/2018 12:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: AMAYA

DATE/TIME OF TEST: 02/05/2018 01:33:18

Diagnostics Check		OK	01:26
Air Blank		0.000	01:26
Control Test		0.078	01:27
Air Blank		0.000	01:27
Subject Sample	# 1	0.112	01:28
Breath Volume		1.480	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.117	01:31
Breath Volume		1.343	01:31
Air Blank		0.000	01:32
Control Test		0.079	01:32
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 03/09/2018 12:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: AMIN

DATE/TIME OF TEST: 02/05/2018 04:45:45

Diagnostics Check		OK	04:39
Air Blank		0.000	04:39
Control Test		0.079	04:40
Air Blank		0.000	04:40
Subject Sample	# 1	0.148	04:41
Breath Volume		1.906	04:41
Air Blank		0.000	04:41
Air Blank		0.000	04:43
Subject Sample	# 2	0.149	04:44
Breath Volume		1.816	04:44
Air Blank		0.000	04:44
Control Test		0.079	04:45
Air Blank		0.000	04:45
Diagnostics Check		OK	04:45

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MORRIS JR

DATE/TIME OF TEST: 02/06/2018 01:43:37

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.079	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.341	01:39
Breath Volume		2.417	01:39
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.341	01:42
Breath Volume		2.953	01:42
Air Blank		0.000	01:42
Control Test		0.079	01:43
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,DAYRON
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: CERMENO GONZALEZ

DATE/TIME OF TEST: 02/08/2018 04:01:34

Diagnostics Check		OK	03:54
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:55
Subject Sample	# 1	0.145	03:56
Breath Volume		1.101	03:56
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.145	03:59
Breath Volume		1.503	03:59
Air Blank		0.000	04:00
Control Test		0.079	04:00
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEHENY,JACOB
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: AMAYA BONILLA

DATE/TIME OF TEST: 02/10/2018 01:24:05

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.080	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.193	01:15
Breath Volume		1.453	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:17
Subject Sample	# 2	0.217	01:19
Breath Volume		3.062	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:21
Subject Sample	# 3	0.210	01:22
Breath Volume		2.570	01:22
Air Blank		0.000	01:23
Control Test		0.077	01:23
Air Blank		0.000	01:23
Diagnostics Check		OK	01:24

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,DAYRON
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 02/10/2018 02:32:56

Diagnostics Check		OK	02:26
Air Blank		0.000	02:27
Control Test		0.079	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.000	02:28
Breath Volume		1.277	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.000	02:31
Breath Volume		1.921	02:31
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BALDWIN,TIMOTHY
ARRESTING OFFICER AGENCY: Tequesta PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: IANNOTTI

DATE/TIME OF TEST: 02/11/2018 01:47:10

Diagnostics Check		OK	01:37	
Air Blank		0.000	01:38	
Control Test		0.079	01:38	
Air Blank		0.000	01:38	
Subject Sample	# 1		01:42	No Sample Provided
Breath Volume		0.000	01:42	
Air Blank		0.000	01:42	
Air Blank		0.000	01:44	
Subject Sample	# 2		01:45	Subject Test Refused
Breath Volume		0.000	01:45	
Air Blank		0.000	01:46	
Control Test		0.079	01:46	
Air Blank		0.000	01:47	
Diagnostics Check		OK	01:47	

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KRIGGER,ROHALIO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: CORBINO

DATE/TIME OF TEST: 02/11/2018 17:39:29

Diagnostics Check		OK	17:27	
Air Blank		0.000	17:28	
Control Test		0.080	17:28	
Air Blank		0.000	17:29	
Subject Sample	# 1	0.225	17:32	Volume Not Met
Breath Volume		0.085	17:32	
Air Blank		0.000	17:32	
Air Blank		0.000	17:34	
Subject Sample	# 2	0.234	17:37	Volume Not Met
Breath Volume		0.015	17:37	
Air Blank		0.000	17:38	
Control Test		0.077	17:38	
Air Blank		0.000	17:39	
Diagnostics Check		OK	17:39	

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCCOY,JOSHUA
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: SONTAY PELICO

DATE/TIME OF TEST: 02/11/2018 19:32:46

Diagnostics Check		OK	19:25
Air Blank		0.000	19:26
Control Test		0.080	19:26
Air Blank		0.000	19:26
Subject Sample	# 1	0.219	19:28
Breath Volume		1.746	19:28
Air Blank		0.000	19:28
Air Blank		0.000	19:30
Subject Sample	# 2	0.232	19:31
Breath Volume		1.972	19:31
Air Blank		0.000	19:31
Control Test		0.079	19:32
Air Blank		0.000	19:32
Diagnostics Check		OK	19:32

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROCHA,LUIS
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: LEON JESUS

DATE/TIME OF TEST: 02/11/2018 20:48:48

Diagnostics Check		OK	20:40
Air Blank		0.000	20:40
Control Test		0.079	20:40
Air Blank		0.000	20:41
Subject Sample	# 1	0.223	20:44
Breath Volume		1.566	20:44
Air Blank		0.000	20:44
Air Blank		0.000	20:46
Subject Sample	# 2	0.225	20:47
Breath Volume		1.441	20:47
Air Blank		0.000	20:47
Control Test		0.078	20:48
Air Blank		0.000	20:48
Diagnostics Check		OK	20:48

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LONEY,MICHAEL
ARRESTING OFFICER AGENCY: Tequesta PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: LEONE

DATE/TIME OF TEST: 02/11/2018 23:59:00

Diagnostics Check		OK	23:52
Air Blank		0.000	23:52
Control Test		0.080	23:53
Air Blank		0.000	23:53
Subject Sample	# 1	0.107	23:54
Breath Volume		3.265	23:54
Air Blank		0.000	23:55
Air Blank		0.000	23:56
Subject Sample	# 2	0.112	23:57
Breath Volume		2.851	23:57
Air Blank		0.000	23:58
Control Test		0.080	23:58
Air Blank		0.000	23:58
Diagnostics Check		OK	23:59

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CROWE,DOUGLAS
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: BUCKINGHAM

DATE/TIME OF TEST: 02/12/2018 20:20:12

Diagnostics Check		OK	20:13
Air Blank		0.000	20:14
Control Test		0.079	20:14
Air Blank		0.000	20:14
Subject Sample	# 1	0.000	20:15
Breath Volume		1.515	20:15
Air Blank		0.000	20:16
Air Blank		0.000	20:18
Subject Sample	# 2	0.000	20:18
Breath Volume		2.070	20:18
Air Blank		0.000	20:19
Control Test		0.079	20:19
Air Blank		0.000	20:20
Diagnostics Check		OK	20:20

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 02/13/2018 00:57:23

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.078	00:52
Air Blank		0.000	00:52
Subject Sample	# 1	0.149	00:52
Breath Volume		3.093	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.151	00:55
Breath Volume		2.593	00:55
Air Blank		0.000	00:56
Control Test		0.078	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MEREDITH,JORDAN
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:05

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 02/13/2018 12:02:15

Diagnostics Check		OK	11:55
Air Blank		0.000	11:56
Control Test		0.080	11:56
Air Blank		0.000	11:57
Subject Sample	# 1	0.231	11:57
Breath Volume		1.574	11:57
Air Blank		0.000	11:58
Air Blank		0.000	12:00
Subject Sample	# 2	0.223	12:00
Breath Volume		1.523	12:00
Air Blank		0.000	12:01
Control Test		0.079	12:01
Air Blank		0.000	12:02
Diagnostics Check		OK	12:02

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARDENAS,FLORANGEL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: AYRES

DATE/TIME OF TEST: 02/14/2018 19:08:55

Diagnostics Check		OK	19:02
Air Blank		0.000	19:02
Control Test		0.080	19:03
Air Blank		0.000	19:03
Subject Sample	# 1	0.213	19:04
Breath Volume		1.671	19:04
Air Blank		0.000	19:05
Air Blank		0.000	19:06
Subject Sample	# 2	0.213	19:07
Breath Volume		1.394	19:07
Air Blank		0.000	19:08
Control Test		0.080	19:08
Air Blank		0.000	19:08
Diagnostics Check		OK	19:08

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GASSANT,GAEL
ARRESTING OFFICER AGENCY: FL Atlantic UPD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: VENEZIO

DATE/TIME OF TEST: 02/14/2018 23:26:52

Diagnostics Check		OK	23:19
Air Blank		0.000	23:19
Control Test		0.080	23:19
Air Blank		0.000	23:20
Subject Sample	# 1	0.000	23:22
Breath Volume		1.730	23:22
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.000	23:25
Breath Volume		1.191	23:25
Air Blank		0.000	23:25
Control Test		0.079	23:26
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLOWERS,GARY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: OSEGUEDA AGUDELO

DATE/TIME OF TEST: 02/15/2018 01:44:09

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.080	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.123	01:39
Breath Volume		3.035	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.115	01:42
Breath Volume		1.214	01:42
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: BRITTON

DATE/TIME OF TEST: 02/16/2018 03:06:03

Diagnostics Check		OK	02:58
Air Blank		0.000	02:59
Control Test		0.080	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.046	03:01
Breath Volume		1.980	03:01
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.045	03:04
Breath Volume		2.292	03:04
Air Blank		0.000	03:05
Control Test		0.080	03:05
Air Blank		0.000	03:05
Diagnostics Check		OK	03:06

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KUEBLER,COLE
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 02/17/2018 00:56:16

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.080	00:50
Air Blank		0.000	00:51
Subject Sample	# 1	0.154	00:51
Breath Volume		2.601	00:51
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.159	00:54
Breath Volume		2.789	00:54
Air Blank		0.000	00:55
Control Test		0.079	00:55
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,DAYRON
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 02/17/2018 01:39:53

Diagnostics Check		OK	01:32
Air Blank		0.000	01:33
Control Test		0.080	01:33
Air Blank		0.000	01:33
Subject Sample	# 1	0.153	01:34
Breath Volume		1.281	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.152	01:38
Breath Volume		2.085	01:38
Air Blank		0.000	01:38
Control Test		0.078	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 03/09/2018 12:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: DIAZ MENDEZ

DATE/TIME OF TEST: 02/17/2018 23:53:05

Diagnostics Check		OK	23:44
Air Blank		0.000	23:45
Control Test		0.080	23:45
Air Blank		0.000	23:46
Subject Sample	# 1	0.180	23:48
Breath Volume		1.523	23:48
Air Blank		0.000	23:49
Air Blank		0.000	23:50
Subject Sample	# 2	0.170	23:51
Breath Volume		1.261	23:51
Air Blank		0.000	23:52
Control Test		0.078	23:52
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROCHA,LUIS
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: MINKER

DATE/TIME OF TEST: 02/18/2018 05:12:12

Diagnostics Check		OK	05:05
Air Blank		0.000	05:06
Control Test		0.080	05:06
Air Blank		0.000	05:06
Subject Sample	# 1	0.145	05:07
Breath Volume		1.144	05:07
Air Blank		0.000	05:08
Air Blank		0.000	05:10
Subject Sample	# 2	0.152	05:10
Breath Volume		1.312	05:10
Air Blank		0.000	05:11
Control Test		0.079	05:11
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALONSO,NICHOLAS
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: VILLATORO LEON

DATE/TIME OF TEST: 02/18/2018 17:27:17

Diagnostics Check		OK	17:18
Air Blank		0.000	17:19
Control Test		0.081	17:19
Air Blank		0.000	17:20
Subject Sample	# 1	0.258	17:22
Breath Volume		1.386	17:22
Air Blank		0.000	17:23
Air Blank		0.000	17:25
Subject Sample	# 2	0.264	17:25
Breath Volume		1.734	17:25
Air Blank		0.000	17:26
Control Test		0.078	17:26
Air Blank		0.000	17:27
Diagnostics Check		OK	17:27

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,DAVID
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: ADRIANI

DATE/TIME OF TEST: 02/18/2018 21:54:12

Diagnostics Check		OK	21:47
Air Blank		0.000	21:47
Control Test		0.080	21:47
Air Blank		0.000	21:48
Subject Sample	# 1	0.100	21:49
Breath Volume		1.898	21:49
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.100	21:52
Breath Volume		1.730	21:52
Air Blank		0.000	21:53
Control Test		0.079	21:53
Air Blank		0.000	21:54
Diagnostics Check		OK	21:54

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FEQUIERE,RALPH
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 02/19/2018 01:27:42

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:14	
Control Test		0.079	01:14	
Air Blank		0.000	01:14	
Subject Sample	# 1	0.192	01:18	Volume Not Met
Breath Volume		0.195	01:18	
Air Blank		0.000	01:18	
Air Blank		0.000	01:20	
Subject Sample	# 2	0.190	01:21	
Breath Volume		1.859	01:21	
Air Blank		0.000	01:21	
Air Blank		0.000	01:23	
Subject Sample	# 3	0.182	01:26	
Breath Volume		1.246	01:26	
Air Blank		0.000	01:26	
Control Test		0.078	01:27	
Air Blank		0.000	01:27	
Diagnostics Check		OK	01:27	

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,EMMA
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: SHALDONE

DATE/TIME OF TEST: 02/19/2018 11:36:25

Diagnostics Check		OK	11:30
Air Blank		0.000	11:30
Control Test		0.080	11:31
Air Blank		0.000	11:31
Subject Sample	# 1	0.000	11:32
Breath Volume		1.710	11:32
Air Blank		0.000	11:32
Air Blank		0.000	11:34
Subject Sample	# 2	0.000	11:35
Breath Volume		1.820	11:35
Air Blank		0.000	11:35
Control Test		0.080	11:35
Air Blank		0.000	11:36
Diagnostics Check		OK	11:36

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: KLOBA

DATE/TIME OF TEST: 02/19/2018 20:07:33

Diagnostics Check		OK	20:00
Air Blank		0.000	20:01
Control Test		0.080	20:01
Air Blank		0.000	20:02
Subject Sample	# 1	0.211	20:03
Breath Volume		1.679	20:03
Air Blank		0.000	20:03
Air Blank		0.000	20:05
Subject Sample	# 2	0.206	20:05
Breath Volume		1.847	20:05
Air Blank		0.000	20:06
Control Test		0.080	20:06
Air Blank		0.000	20:07
Diagnostics Check		OK	20:07

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCCOY,JOSHUA
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: LAKIN

DATE/TIME OF TEST: 02/20/2018 17:43:29

Diagnostics Check		OK	17:36
Air Blank		0.000	17:37
Control Test		0.080	17:37
Air Blank		0.000	17:37
Subject Sample	# 1	0.157	17:38
Breath Volume		1.984	17:38
Air Blank		0.000	17:39
Air Blank		0.000	17:41
Subject Sample	# 2	0.152	17:41
Breath Volume		1.929	17:41
Air Blank		0.000	17:42
Control Test		0.079	17:42
Air Blank		0.000	17:43
Diagnostics Check		OK	17:43

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: FRANKA

DATE/TIME OF TEST: 02/21/2018 17:55:08

Diagnostics Check		OK	17:45
Air Blank		0.000	17:45
Control Test		0.080	17:46
Air Blank		0.000	17:46
Subject Sample	# 1	0.264	17:47
Breath Volume		1.953	17:47
Air Blank		0.000	17:48
Air Blank		0.000	17:49
Subject Sample	# 2	0.233	17:50
Breath Volume		1.792	17:50
Air Blank		0.000	17:51
Air Blank		0.000	17:53
Subject Sample	# 3	0.256	17:53
Breath Volume		2.722	17:53
Air Blank		0.000	17:54
Control Test		0.079	17:54
Air Blank		0.000	17:55
Diagnostics Check		OK	17:55

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCCLUSKEY,RYAN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: TORRES FIGUEROA

DATE/TIME OF TEST: 02/21/2018 22:19:58

Diagnostics Check		OK	22:13
Air Blank		0.000	22:13
Control Test		0.080	22:14
Air Blank		0.000	22:14
Subject Sample	# 1	0.235	22:15
Breath Volume		1.468	22:15
Air Blank		0.000	22:16
Air Blank		0.000	22:17
Subject Sample	# 2	0.236	22:18
Breath Volume		1.691	22:18
Air Blank		0.000	22:19
Control Test		0.079	22:19
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 02/21/2018 22:50:02

Diagnostics Check		OK	22:43
Air Blank		0.000	22:43
Control Test		0.080	22:44
Air Blank		0.000	22:44
Subject Sample	# 1	0.115	22:45
Breath Volume		2.140	22:45
Air Blank		0.000	22:46
Air Blank		0.000	22:48
Subject Sample	# 2	0.116	22:48
Breath Volume		2.085	22:48
Air Blank		0.000	22:49
Control Test		0.080	22:49
Air Blank		0.000	22:49
Diagnostics Check		OK	22:50

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 02/22/2018 23:41:08

Diagnostics Check		OK	23:36	
Air Blank		0.000	23:37	
Control Test		0.079	23:37	
Air Blank		0.000	23:38	
Subject Sample	# 1		23:39	Subject Test Refused
Breath Volume		0.000	23:39	
Air Blank		0.000	23:40	
Control Test		0.080	23:40	
Air Blank		0.000	23:41	
Diagnostics Check		OK	23:41	

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITTAKER,JAMES
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: AKERBLOM

DATE/TIME OF TEST: 02/23/2018 12:46:02

Diagnostics Check		OK	12:39
Air Blank		0.000	12:39
Control Test		0.080	12:40
Air Blank		0.000	12:40
Subject Sample	# 1	0.000	12:41
Breath Volume		1.828	12:41
Air Blank		0.000	12:42
Air Blank		0.000	12:44
Subject Sample	# 2	0.000	12:44
Breath Volume		1.968	12:44
Air Blank		0.000	12:45
Control Test		0.079	12:45
Air Blank		0.000	12:45
Diagnostics Check		OK	12:46

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ALDOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GONZALEZ ESQUIVEL

DATE/TIME OF TEST: 02/24/2018 04:25:42

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.161	04:20
Breath Volume		1.164	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.158	04:24
Breath Volume		3.640	04:24
Air Blank		0.000	04:24
Control Test		0.078	04:25
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: NEVILLE

DATE/TIME OF TEST: 02/24/2018 05:02:20

Diagnostics Check		OK	04:55
Air Blank		0.000	04:56
Control Test		0.080	04:56
Air Blank		0.000	04:57
Subject Sample	# 1	0.142	04:57
Breath Volume		1.960	04:57
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.123	05:00
Breath Volume		1.148	05:00
Air Blank		0.000	05:01
Control Test		0.080	05:01
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BEDWARD

DATE/TIME OF TEST: 02/25/2018 00:17:14

Diagnostics Check		OK	00:08	
Air Blank		0.000	00:09	
Control Test		0.080	00:09	
Air Blank		0.000	00:10	
Subject Sample	# 1	0.133	00:13	Volume Not Met
Breath Volume		0.277	00:13	
Air Blank		0.000	00:13	
Air Blank		0.000	00:15	
Subject Sample	# 2		00:15	Subject Test Refused
Breath Volume		0.000	00:15	
Air Blank		0.000	00:16	
Control Test		0.079	00:16	
Air Blank		0.000	00:17	
Diagnostics Check		OK	00:17	

Data Download Date/Time: 03/09/2018 12:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 02/25/2018 01:18:33

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.080	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.000	01:14
Breath Volume		1.785	01:14
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.000	01:17
Breath Volume		2.582	01:17
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEHENY,JACOB
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: ETIENNE

DATE/TIME OF TEST: 02/25/2018 04:04:12

Diagnostics Check		OK	03:57
Air Blank		0.000	03:58
Control Test		0.080	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.117	03:59
Breath Volume		3.085	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.121	04:02
Breath Volume		2.996	04:02
Air Blank		0.000	04:03
Control Test		0.081	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BATISTA,DANIEL
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: SYSYN

DATE/TIME OF TEST: 02/25/2018 08:29:58

Diagnostics Check		OK	08:23
Air Blank		0.000	08:24
Control Test		0.081	08:24
Air Blank		0.000	08:24
Subject Sample	# 1	0.105	08:25
Breath Volume		3.273	08:25
Air Blank		0.000	08:26
Air Blank		0.000	08:28
Subject Sample	# 2	0.087	08:28
Breath Volume		1.136	08:28
Air Blank		0.000	08:29
Control Test		0.080	08:29
Air Blank		0.000	08:29
Diagnostics Check		OK	08:29

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MACKEY,JAMES
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:42

SUBJECT'S LAST NAME: MARRERO HERNANDEZ

DATE/TIME OF TEST: 02/25/2018 09:21:51

Diagnostics Check		OK	09:08	
Air Blank		0.000	09:09	
Control Test		0.080	09:09	
Air Blank		0.000	09:09	
Subject Sample	# 1	0.185	09:12	Volume Not Met
Breath Volume		0.816	09:12	
Air Blank		0.000	09:13	
Air Blank		0.000	09:15	
Subject Sample	# 2	0.211	09:17	
Breath Volume		2.007	09:17	
Air Blank		0.000	09:18	
Air Blank		0.000	09:19	
Subject Sample	# 3	0.216	09:20	
Breath Volume		2.804	09:20	
Air Blank		0.000	09:20	
Control Test		0.079	09:21	
Air Blank		0.000	09:21	
Diagnostics Check		OK	09:21	

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BIDO,GISELLE
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: GORMAN

DATE/TIME OF TEST: 02/25/2018 21:01:59

Diagnostics Check		OK	20:55
Air Blank		0.000	20:55
Control Test		0.080	20:56
Air Blank		0.000	20:56
Subject Sample	# 1	0.072	20:57
Breath Volume		1.312	20:57
Air Blank		0.000	20:58
Air Blank		0.000	20:59
Subject Sample	# 2	0.076	21:00
Breath Volume		1.746	21:00
Air Blank		0.000	21:01
Control Test		0.080	21:01
Air Blank		0.000	21:01
Diagnostics Check		OK	21:01

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: LUCAS ROSA

DATE/TIME OF TEST: 02/26/2018 02:44:09

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.080	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.182	02:38
Breath Volume		2.246	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:40
Subject Sample	# 2	0.170	02:42
Breath Volume		1.675	02:42
Air Blank		0.000	02:43
Control Test		0.079	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VALERIO,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: FORLENZA

DATE/TIME OF TEST: 02/27/2018 18:17:03

Diagnostics Check		OK	18:10
Air Blank		0.000	18:10
Control Test		0.080	18:10
Air Blank		0.000	18:11
Subject Sample	# 1	0.064	18:12
Breath Volume		2.277	18:12
Air Blank		0.000	18:12
Air Blank		0.000	18:14
Subject Sample	# 2	0.069	18:15
Breath Volume		1.832	18:15
Air Blank		0.000	18:16
Control Test		0.080	18:16
Air Blank		0.000	18:16
Diagnostics Check		OK	18:17

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/28/2018 03:05:30

Diagnostics Check		OK	02:59
Air Blank		0.000	03:00
Control Test		0.080	03:00
Air Blank		0.000	03:00
Subject Sample	# 1	0.000	03:01
Breath Volume		1.503	03:01
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.000	03:04
Breath Volume		1.726	03:04
Air Blank		0.000	03:04
Control Test		0.080	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 03/09/2018 12:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ANGELONE,JOHN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: KNOBBE

DATE/TIME OF TEST: 02/28/2018 08:23:45

Diagnostics Check		OK	08:17
Air Blank		0.000	08:17
Control Test		0.080	08:17
Air Blank		0.000	08:18
Subject Sample	# 1	0.317	08:19
Breath Volume		3.722	08:19
Air Blank		0.000	08:19
Air Blank		0.000	08:21
Subject Sample	# 2	0.314	08:22
Breath Volume		2.789	08:22
Air Blank		0.000	08:22
Control Test		0.079	08:23
Air Blank		0.000	08:23
Diagnostics Check		OK	08:23

Data Download Date/Time: 03/09/2018 12:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: LANNON

DATE/TIME OF TEST: 02/02/2018 03:29:21

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.080	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.197	03:24
Breath Volume		3.128	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.180	03:27
Breath Volume		2.140	03:27
Air Blank		0.000	03:28
Control Test		0.078	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 02/02/2018 15:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALKINS,BRIAN,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WEST,JOHN
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 01/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: CORTES COLON

DATE/TIME OF TEST: 02/23/2018 19:33:03

Diagnostics Check		OK	19:22	
Air Blank		0.000	19:22	
Control Test		0.079	19:23	
Air Blank		0.000	19:23	
Subject Sample	# 1	0.142	19:26	Volume Not Met
Breath Volume		0.449	19:26	
Air Blank		0.000	19:27	
Air Blank		0.000	19:29	
Subject Sample	# 2		19:31	Subject Test Refused
Breath Volume		0.058	19:31	
Air Blank		0.000	19:32	
Control Test		0.077	19:32	
Air Blank		0.000	19:32	
Diagnostics Check		OK	19:33	

Data Download Date/Time: 02/26/2018 11:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: MCCLOUD

DATE/TIME OF TEST: 02/02/2018 04:12:18

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.081	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.000	04:08
Breath Volume		3.062	04:08
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.000	04:10
Breath Volume		2.906	04:10
Air Blank		0.000	04:11
Control Test		0.081	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 02/14/2018 20:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: ZERO TOL
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 02/04/2018 02:35:13

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.081	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.086	02:30
Breath Volume		1.281	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.085	02:33
Breath Volume		2.152	02:33
Air Blank		0.000	02:34
Control Test		0.081	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 02/14/2018 20:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: YORK

DATE/TIME OF TEST: 02/10/2018 03:45:07

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.080	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.241	03:40
Breath Volume		2.593	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.242	03:43
Breath Volume		2.628	03:43
Air Blank		0.000	03:44
Control Test		0.079	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 02/14/2018 20:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWARDS,JACOB,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: L200179811790

DATE/TIME OF TEST: 02/11/2018 02:40:44

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:31	
Control Test		0.080	02:32	
Air Blank		0.000	02:32	
Subject Sample	# 1	0.285	02:34	
Breath Volume		1.820	02:34	
Air Blank		0.000	02:35	
Air Blank		0.000	02:36	
Subject Sample	# 2	0.257	02:39	
Breath Volume		1.718	02:39	
Air Blank			02:40	Purge Fail
Air Blank		0.000	02:40	

Data Download Date/Time: 02/14/2018 20:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWARDS,JACOB,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 02/11/2018 02:58:23

Diagnostics Check		OK	02:43	
Air Blank		0.000	02:44	
Control Test		0.079	02:44	
Air Blank		0.000	02:45	
Subject Sample	# 1	0.282	02:45	
Breath Volume		1.562	02:45	
Air Blank		0.000	02:46	
Air Blank		0.000	02:48	
Subject Sample	# 2	0.280	02:51	Volume Not Met
Breath Volume		0.859	02:51	
Air Blank		0.000	02:52	
Air Blank		0.000	02:53	
Subject Sample	# 3		02:56	No Sample Provided
Breath Volume		0.000	02:56	
Air Blank		0.000	02:57	
Control Test		0.075	02:57	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 02/14/2018 20:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 02/11/2018 04:26:56

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.080	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.116	04:22
Breath Volume		3.585	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:24
Subject Sample	# 2	0.116	04:25
Breath Volume		3.039	04:25
Air Blank		0.000	04:25
Control Test		0.079	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 02/14/2018 20:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDES,LYNNE,S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: DYKES,HUDSON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: COMPTE

DATE/TIME OF TEST: 02/11/2018 05:15:46

Diagnostics Check		OK	05:08
Air Blank		0.000	05:09
Control Test		0.080	05:09
Air Blank		0.000	05:10
Subject Sample	# 1	0.179	05:10
Breath Volume		1.234	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.185	05:14
Breath Volume		2.070	05:14
Air Blank		0.000	05:14
Control Test		0.080	05:15
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 02/14/2018 20:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSANO,DESIREE,B
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BOSTICK,JESSE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: BEASLEY

DATE/TIME OF TEST: 02/12/2018 23:52:46

Diagnostics Check		OK	23:41	
Air Blank		0.000	23:41	
Control Test		0.080	23:42	
Air Blank		0.000	23:42	
Subject Sample	# 1		23:45	No Sample Provided
Breath Volume		0.000	23:45	
Air Blank		0.000	23:46	
Air Blank		0.000	23:48	
Subject Sample	# 2		23:51	No Sample Provided
Breath Volume		0.000	23:51	
Air Blank		0.000	23:51	
Control Test		0.080	23:52	
Air Blank		0.000	23:52	
Diagnostics Check		OK	23:52	

Data Download Date/Time: 02/14/2018 20:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KANKAM,DENZEL
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 02/14/2018 02:04:43

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.079	01:59
Air Blank		0.000	01:59
Subject Sample	# 1	0.226	02:00
Breath Volume		2.148	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.227	02:02
Breath Volume		2.570	02:02
Air Blank		0.000	02:03
Control Test		0.078	02:04
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 02/14/2018 20:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: ANDERSON PIOT

DATE/TIME OF TEST: 02/17/2018 02:04:10

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:53	
Control Test		0.080	01:53	
Air Blank		0.000	01:54	
Subject Sample	# 1		01:57	No Sample Provided
Breath Volume		0.000	01:57	
Air Blank		0.000	01:57	
Air Blank		0.000	01:59	
Subject Sample	# 2		02:02	No Sample Provided
Breath Volume		0.000	02:02	
Air Blank		0.000	02:03	
Control Test		0.081	02:03	
Air Blank		0.000	02:04	
Diagnostics Check		OK	02:04	

Data Download Date/Time: 03/14/2018 21:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WELCH,CHAD
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 02/18/2018 21:47:15

Diagnostics Check		OK	21:35	
Air Blank		0.000	21:36	
Control Test		0.081	21:36	
Air Blank		0.000	21:37	
Subject Sample	# 1		21:40	No Sample Provided
Breath Volume		0.000	21:40	
Air Blank		0.000	21:40	
Air Blank		0.000	21:42	
Subject Sample	# 2		21:45	No Sample Provided
Breath Volume		0.000	21:45	
Air Blank		0.000	21:46	
Control Test		0.080	21:46	
Air Blank		0.000	21:47	
Diagnostics Check		OK	21:47	

Data Download Date/Time: 03/14/2018 21:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANKAM,DENZEL,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KANKAM,DENZEL
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 02/21/2018 00:31:04

Diagnostics Check		OK	00:24
Air Blank		0.000	00:24
Control Test		0.081	00:25
Air Blank		0.000	00:25
Subject Sample	# 1	0.000	00:26
Breath Volume		1.406	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.000	00:29
Breath Volume		3.230	00:29
Air Blank		0.000	00:30
Control Test		0.078	00:30
Air Blank		0.000	00:30
Diagnostics Check		OK	00:31

Data Download Date/Time: 03/14/2018 21:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANKAM,DENZEL,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BEVILLE,TRAVIS
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 02/23/2018 03:06:38

Diagnostics Check		OK	02:59
Air Blank		0.000	03:00
Control Test		0.081	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.206	03:02
Breath Volume		3.398	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.193	03:04
Breath Volume		2.218	03:04
Air Blank		0.000	03:05
Control Test		0.078	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 03/14/2018 21:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWARDS,JACOB,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 02/25/2018 03:13:41

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.080	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.136	03:09
Breath Volume		1.351	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.137	03:12
Breath Volume		1.226	03:12
Air Blank		0.000	03:12
Control Test		0.078	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 03/14/2018 21:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WENDT,HENDRIK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: VINSON

DATE/TIME OF TEST: 02/28/2018 23:37:04

Diagnostics Check		OK	23:25	
Air Blank		0.000	23:26	
Control Test		0.080	23:26	
Air Blank		0.000	23:26	
Subject Sample	# 1		23:30	No Sample Provided
Breath Volume		0.000	23:30	
Air Blank		0.000	23:30	
Air Blank		0.000	23:32	
Subject Sample	# 2		23:35	No Sample Provided
Breath Volume		0.000	23:35	
Air Blank		0.000	23:36	
Control Test		0.080	23:36	
Air Blank		0.000	23:36	
Diagnostics Check		OK	23:37	

Data Download Date/Time: 03/14/2018 21:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: URDANETA DIAZ

DATE/TIME OF TEST: 02/01/2018 02:04:09

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:57	
Control Test		0.081	01:57	
Air Blank		0.000	01:57	
Subject Sample	# 1		02:01	No Sample Provided
Breath Volume		0.000	02:01	
Air Blank		0.000	02:01	
Air Blank		0.000	02:03	
Subject Sample	# 2		02:03	Sequence Aborted
Breath Volume		0.000	02:03	
Air Blank		0.000	02:04	

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: URDANETA DIAZ

DATE/TIME OF TEST: 02/01/2018 02:17:00

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:14	
Control Test		0.081	02:14	
Air Blank		0.000	02:14	
Subject Sample	# 1		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:16	
Control Test		0.082	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: CARPENA

DATE/TIME OF TEST: 02/02/2018 02:23:06

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.081	02:14
Air Blank		0.000	02:15
Subject Sample	# 1	0.213	02:16
Breath Volume		1.445	02:16
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.210	02:21
Breath Volume		1.320	02:21
Air Blank		0.000	02:22
Control Test		0.080	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/02/2018 04:31:01

Diagnostics Check		OK	04:23
Air Blank		0.000	04:24
Control Test		0.080	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.110	04:25
Breath Volume		1.617	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.113	04:29
Breath Volume		2.156	04:29
Air Blank		0.000	04:30
Control Test		0.080	04:30
Air Blank		0.000	04:30
Diagnostics Check		OK	04:31

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: BAGGETT

DATE/TIME OF TEST: 02/03/2018 01:04:19

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.080	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.154	00:59
Breath Volume		2.699	00:59
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.166	01:02
Breath Volume		3.867	01:02
Air Blank		0.000	01:03
Control Test		0.080	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BOLANOS,H
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ENLOW

DATE/TIME OF TEST: 02/03/2018 01:48:42

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.080	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.000	01:44
Breath Volume		1.953	01:44
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.000	01:47
Breath Volume		1.484	01:47
Air Blank		0.000	01:47
Control Test		0.080	01:48
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 02/03/2018 02:32:29

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.080	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.170	02:27
Breath Volume		1.250	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.162	02:30
Breath Volume		1.218	02:30
Air Blank		0.000	02:31
Control Test		0.079	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 02/03/2018 03:10:24

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.080	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.099	03:05
Breath Volume		2.082	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:07
Subject Sample	# 2	0.097	03:08
Breath Volume		1.546	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GORDON,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:17

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 02/03/2018 15:46:48

Diagnostics Check		OK	15:39
Air Blank		0.000	15:40
Control Test		0.081	15:40
Air Blank		0.000	15:40
Subject Sample	# 1	0.192	15:41
Breath Volume		1.500	15:41
Air Blank		0.000	15:42
Air Blank		0.000	15:44
Subject Sample	# 2	0.192	15:45
Breath Volume		1.960	15:45
Air Blank		0.000	15:45
Control Test		0.079	15:46
Air Blank		0.000	15:46
Diagnostics Check		OK	15:46

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: PASCUAL-OSORIO

DATE/TIME OF TEST: 02/03/2018 19:49:09

Diagnostics Check		OK	19:42
Air Blank		0.000	19:43
Control Test		0.081	19:43
Air Blank		0.000	19:43
Subject Sample	# 1	0.229	19:44
Breath Volume		1.191	19:44
Air Blank		0.000	19:45
Air Blank		0.000	19:47
Subject Sample	# 2	0.216	19:47
Breath Volume		2.585	19:47
Air Blank		0.000	19:48
Control Test		0.079	19:48
Air Blank		0.000	19:49
Diagnostics Check		OK	19:49

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: COURTNEY

DATE/TIME OF TEST: 02/03/2018 22:11:50

Diagnostics Check		OK	21:58	
Air Blank		0.000	21:58	
Control Test		0.080	21:58	
Air Blank		0.000	21:59	
Subject Sample	# 1	0.239	22:02	Volume Not Met
Breath Volume		0.429	22:02	
Air Blank		0.000	22:03	
Air Blank		0.000	22:05	
Subject Sample	# 2	0.240	22:06	
Breath Volume		1.816	22:06	
Air Blank		0.000	22:06	
Air Blank		0.000	22:08	
Subject Sample	# 3	0.244	22:10	
Breath Volume		1.253	22:10	
Air Blank		0.000	22:10	
Control Test		0.078	22:11	
Air Blank		0.000	22:11	
Diagnostics Check		OK	22:11	

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: GOMEZ-GOMEZ

DATE/TIME OF TEST: 02/04/2018 03:24:33

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.125	03:19
Breath Volume		2.226	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.140	03:22
Breath Volume		3.722	03:22
Air Blank		0.000	03:23
Control Test		0.079	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: GONZALEZ-PEREZ

DATE/TIME OF TEST: 02/04/2018 03:54:56

Diagnostics Check		OK	03:45
Air Blank		0.000	03:45
Control Test		0.080	03:45
Air Blank		0.000	03:46
Subject Sample	# 1	0.136	03:48
Breath Volume		1.535	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.129	03:53
Breath Volume		1.949	03:53
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/04/2018 05:09:52

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.080	05:04
Air Blank		0.000	05:04
Subject Sample	# 1	0.146	05:05
Breath Volume		2.917	05:05
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.149	05:08
Breath Volume		2.898	05:08
Air Blank		0.000	05:08
Control Test		0.079	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: SCARPA

DATE/TIME OF TEST: 02/04/2018 06:12:37

Diagnostics Check		OK	05:59	
Air Blank		0.000	05:59	
Control Test		0.080	06:00	
Air Blank		0.000	06:00	
Subject Sample	# 1	0.165	06:03	Volume Not Met
Breath Volume		0.167	06:03	
Air Blank		0.000	06:04	
Air Blank		0.000	06:06	
Subject Sample	# 2	0.151	06:07	
Breath Volume		1.632	06:07	
Air Blank		0.000	06:08	
Air Blank		0.000	06:10	
Subject Sample	# 3	0.171	06:10	
Breath Volume		1.445	06:10	
Air Blank		0.000	06:11	
Control Test		0.078	06:11	
Air Blank		0.000	06:12	
Diagnostics Check		OK	06:12	

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: DOWDEN

DATE/TIME OF TEST: 02/05/2018 00:10:40

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.079	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.000	00:06
Breath Volume		1.125	00:06
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.000	00:09
Breath Volume		1.734	00:09
Air Blank		0.000	00:09
Control Test		0.080	00:10
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: VACA

DATE/TIME OF TEST: 02/05/2018 03:02:36

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.101	02:58
Breath Volume		1.648	02:58
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.102	03:00
Breath Volume		3.375	03:00
Air Blank		0.000	03:01
Control Test		0.080	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: JOSE-PAZ

DATE/TIME OF TEST: 02/05/2018 23:19:51

Diagnostics Check		OK	23:12
Air Blank		0.000	23:12
Control Test		0.079	23:13
Air Blank		0.000	23:13
Subject Sample	# 1	0.178	23:14
Breath Volume		1.441	23:14
Air Blank		0.000	23:15
Air Blank		0.000	23:17
Subject Sample	# 2	0.177	23:18
Breath Volume		1.578	23:18
Air Blank		0.000	23:18
Control Test		0.079	23:19
Air Blank		0.000	23:19
Diagnostics Check		OK	23:19

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PALFY,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/07/2018 00:04:57

Diagnostics Check		OK	23:56	
Air Blank		0.000	23:56	
Control Test		0.080	23:57	
Air Blank		0.000	23:57	
Subject Sample	# 1		00:00	No Sample Provided
Breath Volume		0.000	00:00	
Air Blank		0.000	00:01	
Air Blank		0.000	00:03	
Subject Sample	# 2		00:03	Subject Test Refused
Breath Volume		0.000	00:03	
Air Blank		0.000	00:03	
Control Test		0.080	00:04	
Air Blank		0.000	00:04	
Diagnostics Check		OK	00:04	

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: LOZANO

DATE/TIME OF TEST: 02/08/2018 01:01:11

Diagnostics Check		OK	00:54
Air Blank		0.000	00:55
Control Test		0.081	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.132	00:56
Breath Volume		2.191	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.128	00:59
Breath Volume		1.921	00:59
Air Blank		0.000	01:00
Control Test		0.080	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: PERMAUL

DATE/TIME OF TEST: 02/08/2018 22:04:07

Diagnostics Check		OK	21:55
Air Blank		0.000	21:56
Control Test		0.079	21:56
Air Blank		0.000	21:57
Subject Sample	# 1	0.194	21:59
Breath Volume		1.339	21:59
Air Blank		0.000	22:00
Air Blank		0.000	22:01
Subject Sample	# 2	0.195	22:02
Breath Volume		1.281	22:02
Air Blank		0.000	22:03
Control Test		0.078	22:03
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: RENDON

DATE/TIME OF TEST: 02/09/2018 03:22:58

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.080	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.143	03:18
Breath Volume		1.929	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:20
Subject Sample	# 2	0.147	03:21
Breath Volume		1.605	03:21
Air Blank		0.000	03:22
Control Test		0.081	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEINBERG,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: PRUETT

DATE/TIME OF TEST: 02/09/2018 05:48:33

Diagnostics Check		OK	05:40
Air Blank		0.000	05:40
Control Test		0.081	05:40
Air Blank		0.000	05:41
Subject Sample	# 1	0.000	05:44
Breath Volume		1.261	05:44
Air Blank		0.000	05:44
Air Blank		0.000	05:46
Subject Sample	# 2	0.000	05:47
Breath Volume		1.179	05:47
Air Blank		0.000	05:47
Control Test		0.081	05:47
Air Blank		0.000	05:48
Diagnostics Check		OK	05:48

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:59

SUBJECT'S LAST NAME: RAHMAN

DATE/TIME OF TEST: 02/09/2018 07:26:16

Diagnostics Check		OK	07:20
Air Blank		0.000	07:20
Control Test		0.079	07:20
Air Blank		0.000	07:21
Subject Sample	# 1	0.101	07:21
Breath Volume		3.269	07:21
Air Blank		0.000	07:22
Air Blank		0.000	07:24
Subject Sample	# 2	0.101	07:24
Breath Volume		2.996	07:24
Air Blank		0.000	07:25
Control Test		0.080	07:25
Air Blank		0.000	07:26
Diagnostics Check		OK	07:26

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: DIAZ OLIVA

DATE/TIME OF TEST: 02/10/2018 03:11:45

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.081	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.177	03:07
Breath Volume		1.753	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:09
Subject Sample	# 2	0.173	03:10
Breath Volume		1.484	03:10
Air Blank		0.000	03:10
Control Test		0.077	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 02/10/2018 04:25:28

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.080	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.150	04:20
Breath Volume		1.277	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.147	04:23
Breath Volume		1.214	04:23
Air Blank		0.000	04:24
Control Test		0.079	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERD,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:44

SUBJECT'S LAST NAME: SANCHEZ RODRIGUEZ

DATE/TIME OF TEST: 02/10/2018 17:11:18

Diagnostics Check		OK	17:04
Air Blank		0.000	17:05
Control Test		0.080	17:05
Air Blank		0.000	17:06
Subject Sample	# 1	0.000	17:06
Breath Volume		1.796	17:06
Air Blank		0.000	17:07
Air Blank		0.000	17:09
Subject Sample	# 2	0.000	17:09
Breath Volume		1.453	17:09
Air Blank		0.000	17:10
Control Test		0.080	17:10
Air Blank		0.000	17:11
Diagnostics Check		OK	17:11

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: RIVERA COLON

DATE/TIME OF TEST: 02/11/2018 00:48:45

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.081	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.084	00:44
Breath Volume		2.824	00:44
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.085	00:47
Breath Volume		2.789	00:47
Air Blank		0.000	00:47
Control Test		0.080	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: WOMBLE

DATE/TIME OF TEST: 02/12/2018 23:07:24

Diagnostics Check		OK	23:00
Air Blank		0.000	23:00
Control Test		0.080	23:01
Air Blank		0.000	23:01
Subject Sample	# 1	0.081	23:02
Breath Volume		1.199	23:02
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.083	23:05
Breath Volume		1.609	23:05
Air Blank		0.000	23:06
Control Test		0.079	23:06
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BISHOP-THOMPSON

DATE/TIME OF TEST: 02/13/2018 03:17:10

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.151	03:12
Breath Volume		1.476	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:14
Subject Sample	# 2	0.159	03:15
Breath Volume		2.218	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 02/13/2018 17:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: LANDIS

DATE/TIME OF TEST: 02/15/2018 01:40:45

Diagnostics Check		OK	01:29
Air Blank		0.000	01:30
Control Test		0.080	01:30
Air Blank		0.000	01:31
Subject Sample	# 1	0.202	01:34
Breath Volume		1.207	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:36
Subject Sample	# 2	0.188	01:39
Breath Volume		1.605	01:39
Air Blank		0.000	01:39
Control Test		0.077	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 03/05/2018 17:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: LAPORTE ORTIZ

DATE/TIME OF TEST: 02/16/2018 19:10:59

Diagnostics Check		OK	19:04
Air Blank		0.000	19:04
Control Test		0.080	19:04
Air Blank		0.000	19:05
Subject Sample	# 1	0.130	19:06
Breath Volume		1.886	19:06
Air Blank		0.000	19:07
Air Blank		0.000	19:08
Subject Sample	# 2	0.133	19:09
Breath Volume		2.179	19:09
Air Blank		0.000	19:10
Control Test		0.080	19:10
Air Blank		0.000	19:10
Diagnostics Check		OK	19:10

Data Download Date/Time: 03/05/2018 17:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 02/17/2018 02:23:59

Diagnostics Check		OK	02:16
Air Blank		0.000	02:17
Control Test		0.081	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.150	02:19
Breath Volume		1.140	02:19
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.169	02:22
Breath Volume		2.382	02:22
Air Blank		0.000	02:23
Control Test		0.079	02:23
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 03/05/2018 17:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HILES,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: AJCUC

DATE/TIME OF TEST: 02/18/2018 00:25:22

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.081	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.199	00:20
Breath Volume		1.816	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.207	00:23
Breath Volume		1.566	00:23
Air Blank		0.000	00:24
Control Test		0.080	00:24
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 03/05/2018 17:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: LAGARES

DATE/TIME OF TEST: 02/19/2018 01:47:12

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.081	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.132	01:42
Breath Volume		1.277	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:44
Subject Sample	# 2	0.134	01:45
Breath Volume		1.285	01:45
Air Blank		0.000	01:46
Control Test		0.079	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 03/05/2018 17:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: PEREZ BERMUDEZ

DATE/TIME OF TEST: 02/19/2018 05:31:42

Diagnostics Check		OK	05:23
Air Blank		0.000	05:23
Control Test		0.080	05:24
Air Blank		0.000	05:24
Subject Sample	# 1	0.167	05:26
Breath Volume		1.906	05:26
Air Blank		0.000	05:27
Air Blank		0.000	05:29
Subject Sample	# 2	0.164	05:30
Breath Volume		2.207	05:30
Air Blank		0.000	05:30
Control Test		0.078	05:31
Air Blank		0.000	05:31
Diagnostics Check		OK	05:31

Data Download Date/Time: 03/05/2018 17:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: TAVAREZ

DATE/TIME OF TEST: 02/21/2018 04:31:56

Diagnostics Check		OK	04:24
Air Blank		0.000	04:25
Control Test		0.080	04:25
Air Blank		0.000	04:26
Subject Sample	# 1	0.048	04:26
Breath Volume		1.460	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.055	04:30
Breath Volume		1.316	04:30
Air Blank		0.000	04:30
Control Test		0.080	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 03/05/2018 17:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: MCCONNELL

DATE/TIME OF TEST: 02/21/2018 05:53:00

Diagnostics Check		OK	05:45
Air Blank		0.000	05:45
Control Test		0.080	05:46
Air Blank		0.000	05:46
Subject Sample	# 1	0.237	05:47
Breath Volume		2.636	05:47
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.231	05:51
Breath Volume		2.785	05:51
Air Blank		0.000	05:52
Control Test		0.079	05:52
Air Blank		0.000	05:52
Diagnostics Check		OK	05:53

Data Download Date/Time: 03/05/2018 17:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 02/23/2018 01:40:24

Diagnostics Check		OK	01:33
Air Blank		0.000	01:34
Control Test		0.080	01:34
Air Blank		0.000	01:35
Subject Sample	# 1	0.000	01:36
Breath Volume		2.312	01:36
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.000	01:39
Breath Volume		1.250	01:39
Air Blank		0.000	01:39
Control Test		0.080	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 03/05/2018 17:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/24/2018 01:39:58

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.080	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.136	01:35
Breath Volume		2.839	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:37
Subject Sample	# 2	0.121	01:38
Breath Volume		1.410	01:38
Air Blank		0.000	01:39
Control Test		0.079	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 03/05/2018 17:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARBOSA,F
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 02/24/2018 03:53:27

Diagnostics Check		OK	03:46
Air Blank		0.000	03:47
Control Test		0.080	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.145	03:48
Breath Volume		2.558	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.153	03:51
Breath Volume		2.535	03:51
Air Blank		0.000	03:52
Control Test		0.078	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 03/05/2018 17:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: UNO,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: NEIDZWIECKI

DATE/TIME OF TEST: 02/25/2018 19:59:08

Diagnostics Check		OK	19:51
Air Blank		0.000	19:52
Control Test		0.080	19:52
Air Blank		0.000	19:52
Subject Sample	# 1	0.154	19:54
Breath Volume		1.628	19:54
Air Blank		0.000	19:55
Air Blank		0.000	19:56
Subject Sample	# 2	0.162	19:57
Breath Volume		2.457	19:57
Air Blank		0.000	19:58
Control Test		0.078	19:58
Air Blank		0.000	19:59
Diagnostics Check		OK	19:59

Data Download Date/Time: 03/05/2018 17:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: HERNANDEZ-SALAZAR

DATE/TIME OF TEST: 02/26/2018 19:56:25

Diagnostics Check		OK	19:44	
Air Blank		0.000	19:45	
Control Test		0.080	19:45	
Air Blank		0.000	19:45	
Subject Sample	# 1	0.208	19:49	Volume Not Met
Breath Volume		0.000	19:49	
Air Blank		0.000	19:49	
Air Blank		0.000	19:51	
Subject Sample	# 2	0.206	19:54	Volume Not Met
Breath Volume		0.000	19:54	
Air Blank		0.000	19:55	
Control Test		0.079	19:55	
Air Blank		0.000	19:56	
Diagnostics Check		OK	19:56	

Data Download Date/Time: 03/05/2018 17:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 02/24/2018 02:21:53

Diagnostics Check		OK	02:08	
Air Blank		0.000	02:08	
Control Test		0.081	02:08	
Air Blank		0.000	02:09	
Subject Sample	# 1	0.178	02:12	Volume Not Met
Breath Volume		0.113	02:12	
Air Blank		0.000	02:13	
Air Blank		0.000	02:14	
Subject Sample	# 2	0.171	02:16	
Breath Volume		1.863	02:16	
Air Blank		0.000	02:17	
Air Blank		0.000	02:18	
Subject Sample	# 3	0.164	02:20	
Breath Volume		1.917	02:20	
Air Blank		0.000	02:20	
Control Test		0.079	02:21	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 03/29/2018 09:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,TROY,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LAND,DUSTIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: STEEVES

DATE/TIME OF TEST: 02/25/2018 06:48:20

Diagnostics Check		OK	06:42
Air Blank		0.000	06:42
Control Test		0.080	06:43
Air Blank		0.000	06:43
Subject Sample	# 1	0.229	06:43
Breath Volume		3.375	06:43
Air Blank		0.000	06:44
Air Blank		0.000	06:46
Subject Sample	# 2	0.240	06:46
Breath Volume		3.179	06:46
Air Blank		0.000	06:47
Control Test		0.079	06:47
Air Blank		0.000	06:48
Diagnostics Check		OK	06:48

Data Download Date/Time: 03/29/2018 09:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: OSBORNE,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: COE

DATE/TIME OF TEST: 02/01/2018 21:37:40

Diagnostics Check		OK	21:30
Air Blank		0.000	21:31
Control Test		0.079	21:31
Air Blank		0.000	21:32
Subject Sample	# 1	0.000	21:33
Breath Volume		2.058	21:33
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 2	0.000	21:36
Breath Volume		3.871	21:36
Air Blank		0.000	21:36
Control Test		0.078	21:37
Air Blank		0.000	21:37
Diagnostics Check		OK	21:37

Data Download Date/Time: 02/20/2018 09:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORMLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GORMLEY,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: NONNENMOCHER

DATE/TIME OF TEST: 02/02/2018 17:47:21

Diagnostics Check		OK	17:41
Air Blank		0.000	17:41
Control Test		0.079	17:41
Air Blank		0.000	17:42
Subject Sample	# 1	0.275	17:42
Breath Volume		4.125	17:42
Air Blank		0.000	17:43
Air Blank		0.000	17:45
Subject Sample	# 2	0.272	17:45
Breath Volume		2.738	17:45
Air Blank		0.000	17:46
Control Test		0.079	17:46
Air Blank		0.000	17:47
Diagnostics Check		OK	17:47

Data Download Date/Time: 02/20/2018 09:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,BRANDT,J
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: KESSINGER,BRANDT
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 02/03/2018 01:33:26

Diagnostics Check		OK	01:29	
Air Blank		0.000	01:29	
Control Test		0.081	01:30	
Air Blank		0.000	01:30	
Subject Sample	# 1		01:32	Sequence Aborted
Breath Volume		0.000	01:32	
Air Blank		0.000	01:33	

Data Download Date/Time: 02/20/2018 09:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY,WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/04/2018 21:08:09

Diagnostics Check		OK	20:58
Air Blank		0.000	20:58
Control Test		0.079	20:58
Air Blank		0.000	20:59
Subject Sample	# 1	0.215	21:02
Breath Volume		3.378	21:02
Air Blank		0.000	21:03
Air Blank		0.000	21:04
Subject Sample	# 2	0.212	21:06
Breath Volume		3.519	21:06
Air Blank		0.000	21:07
Control Test		0.078	21:07
Air Blank		0.000	21:08
Diagnostics Check		OK	21:08

Data Download Date/Time: 02/20/2018 09:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: OSBORNE,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: GRIESER

DATE/TIME OF TEST: 02/06/2018 04:33:37

Diagnostics Check		OK	04:27
Air Blank		0.000	04:27
Control Test		0.079	04:27
Air Blank		0.000	04:28
Subject Sample	# 1	0.199	04:28
Breath Volume		3.347	04:28
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.196	04:31
Breath Volume		2.917	04:31
Air Blank		0.000	04:32
Control Test		0.079	04:32
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 02/20/2018 09:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEITE,STEVEN,G
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEITE,STEVEN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/07/2018 06:18:39

Diagnostics Check		OK	06:13	
Air Blank		0.000	06:13	
Control Test		0.078	06:14	
Air Blank		0.000	06:14	
Subject Sample	# 1		06:17	Subject Test Refused
Breath Volume		0.000	06:17	
Air Blank		0.000	06:17	
Control Test		0.078	06:18	
Air Blank		0.000	06:18	
Diagnostics Check		OK	06:18	

Data Download Date/Time: 02/20/2018 09:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 02/08/2018 00:37:46

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.079	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.071	00:33
Breath Volume		2.105	00:33
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.062	00:36
Breath Volume		2.429	00:36
Air Blank		0.000	00:36
Control Test		0.075	00:37
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 02/20/2018 09:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEITE,STEVEN,G
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEITE,STEVEN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: SASSER

DATE/TIME OF TEST: 02/08/2018 23:46:22

Diagnostics Check		OK	23:34	
Air Blank		0.000	23:35	
Control Test		0.079	23:35	
Air Blank		0.000	23:36	
Subject Sample	# 1		23:39	No Sample Provided
Breath Volume		0.000	23:39	
Air Blank		0.000	23:40	
Air Blank		0.000	23:41	
Subject Sample	# 2		23:44	No Sample Provided
Breath Volume		0.000	23:44	
Air Blank		0.000	23:45	
Control Test		0.079	23:45	
Air Blank		0.000	23:46	
Diagnostics Check		OK	23:46	

Data Download Date/Time: 02/20/2018 09:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAMUEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 02/09/2018 03:56:04

Diagnostics Check		OK	03:48
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.143	03:51
Breath Volume		4.289	03:51
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.137	03:54
Breath Volume		4.808	03:54
Air Blank		0.000	03:55
Control Test		0.079	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:56

Data Download Date/Time: 02/20/2018 09:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TASCHETTA,JOSEPH,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BOATWRIGHT,J
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 02/10/2018 00:54:36

Diagnostics Check		OK	00:48
Air Blank		0.000	00:48
Control Test		0.078	00:49
Air Blank		0.000	00:49
Subject Sample	# 1	0.160	00:50
Breath Volume		1.421	00:50
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.158	00:53
Breath Volume		2.808	00:53
Air Blank		0.000	00:53
Control Test		0.078	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 02/20/2018 09:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENE,ROBERT,R
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: GREENE,ROBERT
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: MONISH

DATE/TIME OF TEST: 02/12/2018 17:33:58

Diagnostics Check		OK	17:30	
Air Blank		0.000	17:31	
Control Test		0.078	17:31	
Air Blank		0.000	17:32	
Subject Sample	# 1		17:32	Subject Test Refused
Breath Volume		0.000	17:32	
Air Blank		0.000	17:33	
Control Test		0.078	17:33	
Air Blank		0.000	17:33	
Diagnostics Check		OK	17:33	

Data Download Date/Time: 02/20/2018 09:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANGELISTA,CURTIS,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LACKEY,NICHOLAS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/17/2018 03:48:28

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.078	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.223	03:43
Breath Volume		2.488	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.232	03:46
Breath Volume		2.855	03:46
Air Blank		0.000	03:47
Control Test		0.076	03:47
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 02/20/2018 09:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,MARK,A
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: TURNER,MARK
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 02/18/2018 08:13:51

Diagnostics Check		OK	08:07
Air Blank		0.000	08:08
Control Test		0.079	08:08
Air Blank		0.000	08:08
Subject Sample	# 1	0.142	08:09
Breath Volume		3.125	08:09
Air Blank		0.000	08:10
Air Blank		0.000	08:11
Subject Sample	# 2	0.136	08:12
Breath Volume		3.101	08:12
Air Blank		0.000	08:12
Control Test		0.079	08:13
Air Blank		0.000	08:13
Diagnostics Check		OK	08:13

Data Download Date/Time: 02/20/2018 09:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLIS,MARCO,A
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: POWERS,BRITTANY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: SOLIS

DATE/TIME OF TEST: 02/18/2018 09:48:10

Diagnostics Check		OK	09:36	
Air Blank		0.000	09:36	
Control Test		0.079	09:36	
Air Blank		0.000	09:37	
Subject Sample	# 1		09:40	No Sample Provided
Breath Volume		0.000	09:40	
Air Blank		0.000	09:41	
Air Blank		0.000	09:43	
Subject Sample	# 2	0.152	09:43	
Breath Volume		2.222	09:43	
Air Blank		0.000	09:44	
Air Blank		0.000	09:46	
Subject Sample	# 3	0.147	09:46	
Breath Volume		2.011	09:46	
Air Blank		0.000	09:47	
Control Test		0.077	09:47	
Air Blank		0.000	09:48	
Diagnostics Check		OK	09:48	

Data Download Date/Time: 02/20/2018 09:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA,CAROLINA,M
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ACOSTA,CAROLINA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 02/01/2018 21:37:05

Diagnostics Check		OK	21:31	
Air Blank		0.000	21:32	
Control Test		0.081	21:32	
Air Blank		0.000	21:33	
Subject Sample	# 1		21:35	Subject Test Refused
Breath Volume		0.000	21:35	
Air Blank		0.000	21:36	
Control Test		0.080	21:36	
Air Blank		0.000	21:37	
Diagnostics Check		OK	21:37	

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENA,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MENA,ALEXANDER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 05:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: LEO RAMIREZ

DATE/TIME OF TEST: 02/04/2018 06:28:28

Diagnostics Check	OK	06:27	
Air Blank	0.000	06:27	
Control Test	0.080	06:27	
Air Blank		06:27	RFI Detect
Air Blank	0.000	06:28	

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENA,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MENA,ALEXANDER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 05:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: LEO RAMIREZ

DATE/TIME OF TEST: 02/04/2018 06:37:25

Diagnostics Check		OK	06:30
Air Blank		0.000	06:30
Control Test		0.080	06:30
Air Blank		0.000	06:31
Subject Sample	# 1	0.178	06:32
Breath Volume		1.800	06:32
Air Blank		0.000	06:33
Air Blank		0.000	06:35
Subject Sample	# 2	0.177	06:35
Breath Volume		1.632	06:35
Air Blank		0.000	06:36
Control Test		0.080	06:36
Air Blank		0.000	06:37
Diagnostics Check		OK	06:37

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ROY,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUEZ,ROY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: CUROTTO

DATE/TIME OF TEST: 02/04/2018 17:25:03

Diagnostics Check		OK	17:14	
Air Blank		0.000	17:15	
Control Test		0.080	17:15	
Air Blank		0.000	17:16	
Subject Sample	# 1	0.228	17:18	
Breath Volume		2.105	17:18	
Air Blank		0.000	17:19	
Air Blank		0.000	17:21	
Subject Sample	# 2	0.222	17:23	
Breath Volume		2.207	17:23	
Air Blank			17:24	Purge Fail
Air Blank		0.000	17:25	

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,JASIEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ZALDIVAR

DATE/TIME OF TEST: 02/08/2018 02:24:44

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:16	
Control Test		0.080	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 1		02:20	No Sample Provided
Breath Volume		0.000	02:20	
Air Blank		0.000	02:21	
Air Blank		0.000	02:23	
Subject Sample	# 2		02:23	Subject Test Refused
Breath Volume		0.000	02:23	
Air Blank		0.000	02:23	
Control Test		0.081	02:24	
Air Blank		0.000	02:24	
Diagnostics Check		OK	02:24	

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,AQUILES,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SANCHEZ,AQUILES
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/11/2018 01:34:12

Diagnostics Check	OK	01:33	
Air Blank		01:33	RFI Detect
Air Blank	0.000	01:34	

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,AQUILES,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SANCHEZ,AQUILES
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/11/2018 01:49:23

Diagnostics Check		OK	01:45	
Air Blank		0.000	01:46	
Control Test		0.080	01:46	
Air Blank		0.000	01:47	
Subject Sample	# 1		01:48	Subject Test Refused
Breath Volume		0.000	01:48	
Air Blank		0.000	01:48	
Control Test		0.080	01:48	
Air Blank		0.000	01:49	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,KEITH,C
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: HERNANDEZ,KEITH
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: BUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 02/11/2018 19:52:26

Diagnostics Check		OK	19:46	
Air Blank		0.000	19:47	
Control Test		0.080	19:47	
Air Blank		0.000	19:48	
Subject Sample	# 1	0.193	19:51	Volume Not Met
Breath Volume		1.011	19:51	
Air Blank			19:51	Purge Fail
Air Blank		0.000	19:52	

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,KEITH,C
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: HERNANDEZ,KEITH
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: BUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 02/11/2018 20:01:52

Diagnostics Check		OK	19:58	
Air Blank		0.000	19:58	
Control Test		0.076	19:59	
Air Blank		0.000	19:59	
Subject Sample	# 1	0.148	20:00	
Breath Volume		3.316	20:00	
Air Blank			20:01	Purge Fail
Air Blank		0.000	20:01	

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,KEITH,C
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: HERNANDEZ,KEITH
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: BUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 02/11/2018 20:23:29

Diagnostics Check		OK	20:09
Air Blank		0.000	20:10
Control Test		0.076	20:10
Air Blank		0.000	20:10
Subject Sample	# 1	0.175	20:12
Breath Volume		2.515	20:12
Air Blank		0.000	20:13
Air Blank		0.000	20:14
Subject Sample	# 2	0.154	20:16
Breath Volume		2.902	20:16
Air Blank		0.000	20:17
Air Blank		0.000	20:18
Subject Sample	# 3	0.168	20:21
Breath Volume		2.589	20:21
Air Blank		0.000	20:22
Control Test		0.078	20:22
Air Blank		0.000	20:23
Diagnostics Check		OK	20:23

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ADRIAN,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: MOSCHIANO,DAVID
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: BUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 02/15/2018 20:24:27

Diagnostics Check		OK	20:21	
Air Blank		0.000	20:21	
Control Test		0.080	20:21	
Air Blank		0.000	20:22	
Subject Sample	# 1		20:23	Subject Test Refused
Breath Volume		0.000	20:23	
Air Blank		0.000	20:23	
Control Test		0.080	20:23	
Air Blank		0.000	20:24	
Diagnostics Check		OK	20:24	

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,DIYANGO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VILLEGAS,A
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 02/16/2018 00:05:26

Diagnostics Check		OK	23:58
Air Blank		0.000	23:59
Control Test		0.081	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.150	00:00
Breath Volume		2.746	00:00
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.150	00:03
Breath Volume		2.398	00:03
Air Blank		0.000	00:04
Control Test		0.078	00:04
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,RAYMON,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: WASHINGTON,JAYSON
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 02/18/2018 01:26:25

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.081	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.198	01:21
Breath Volume		1.125	01:21
Air Blank		0.000	01:22
Air Blank		0.000	01:23
Subject Sample	# 2	0.198	01:24
Breath Volume		1.546	01:24
Air Blank		0.000	01:25
Control Test		0.079	01:25
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENA,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HARMON,FABIANA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 05:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: BALDO CASTELLANOS

DATE/TIME OF TEST: 02/18/2018 06:18:04

Diagnostics Check		OK	06:11
Air Blank		0.000	06:11
Control Test		0.080	06:12
Air Blank		0.000	06:12
Subject Sample	# 1	0.132	06:13
Breath Volume		1.222	06:13
Air Blank		0.000	06:14
Air Blank		0.000	06:16
Subject Sample	# 2	0.135	06:16
Breath Volume		1.511	06:16
Air Blank		0.000	06:17
Control Test		0.079	06:17
Air Blank		0.000	06:17
Diagnostics Check		OK	06:18

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUAREZ,BERNICE,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SUAREZ,BERNICE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 08:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: ORELLANA MUNOZ

DATE/TIME OF TEST: 02/18/2018 09:01:34

Diagnostics Check		OK	08:52
Air Blank		0.000	08:53
Control Test		0.080	08:53
Air Blank		0.000	08:53
Subject Sample	# 1	0.209	08:56
Breath Volume		1.710	08:56
Air Blank		0.000	08:57
Air Blank		0.000	08:58
Subject Sample	# 2	0.203	08:59
Breath Volume		2.558	08:59
Air Blank		0.000	09:00
Control Test		0.075	09:00
Air Blank		0.000	09:01
Diagnostics Check		OK	09:01

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA,CAROLINA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TARGIA,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 02/18/2018 23:31:34

Diagnostics Check		OK	23:28	
Air Blank		0.000	23:29	
Control Test		0.080	23:29	
Air Blank		0.000	23:30	
Subject Sample	# 1		23:30	Subject Test Refused
Breath Volume		0.000	23:30	
Air Blank		0.000	23:30	
Control Test		0.080	23:30	
Air Blank		0.000	23:31	
Diagnostics Check		OK	23:31	

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,DIYANGO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: FERNANDEZ,FERNANDEZ
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: DAKOTA

DATE/TIME OF TEST: 02/19/2018 04:45:22

Diagnostics Check		OK	04:41	
Air Blank		0.000	04:42	
Control Test		0.080	04:42	
Air Blank		0.000	04:42	
Subject Sample	# 1		04:44	Subject Test Refused
Breath Volume		0.000	04:44	
Air Blank		0.000	04:44	
Control Test		0.080	04:44	
Air Blank		0.000	04:45	
Diagnostics Check		OK	04:45	

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYA,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: BECKER,ALEXA
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 02/23/2018 04:20:57

Diagnostics Check		OK	04:18	
Air Blank		0.000	04:18	
Control Test		0.080	04:18	
Air Blank		0.000	04:19	
Subject Sample	# 1		04:19	Subject Test Refused
Breath Volume		0.000	04:19	
Air Blank		0.000	04:20	
Control Test		0.080	04:20	
Air Blank		0.000	04:20	
Diagnostics Check		OK	04:20	

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,AQUILES,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SANCHEZ,AQUILES
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: SARAGOVI

DATE/TIME OF TEST: 02/24/2018 01:54:38

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:54	
Control Test	0.000	01:54	RFI Detect
Air Blank		01:54	RFI Detect

Data Download Date/Time: 02/27/2018 13:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,AQUILES,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SANCHEZ,AQUILES
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: SARAGOVI

DATE/TIME OF TEST: 02/24/2018 02:03:39

Diagnostics Check		OK	01:57
Air Blank		0.000	01:57
Control Test		0.080	01:57
Air Blank		0.000	01:58
Subject Sample	# 1	0.095	01:59
Breath Volume		2.167	01:59
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.091	02:02
Breath Volume		2.144	02:02
Air Blank		0.000	02:02
Control Test		0.079	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,JOHN,V
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,JOHN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: CRUZGALIANO

DATE/TIME OF TEST: 02/25/2018 03:38:25

Diagnostics Check		OK	03:29	
Air Blank		0.000	03:29	
Control Test		0.081	03:30	
Air Blank		0.000	03:30	
Subject Sample	# 1	0.202	03:33	Volume Not Met
Breath Volume		0.015	03:33	
Air Blank		0.000	03:34	
Air Blank		0.000	03:36	
Subject Sample	# 2		03:37	Subject Test Refused
Breath Volume		0.000	03:37	
Air Blank		0.000	03:37	
Control Test		0.076	03:37	
Air Blank		0.000	03:38	
Diagnostics Check		OK	03:38	

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,CHIRLEDY,M
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GARCIA,CHIRLEDY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: GARBUZOVA

DATE/TIME OF TEST: 02/26/2018 00:33:42

Diagnostics Check		OK	00:30	
Air Blank		0.000	00:31	
Control Test		0.081	00:31	
Air Blank		0.000	00:32	
Subject Sample	# 1		00:32	Subject Test Refused
Breath Volume		0.000	00:32	
Air Blank		0.000	00:32	
Control Test		0.080	00:33	
Air Blank		0.000	00:33	
Diagnostics Check		OK	00:33	

Data Download Date/Time: 02/27/2018 13:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANCO,JUAN,S
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PEREZ,ADRIAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: BUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: AVILES

DATE/TIME OF TEST: 02/01/2018 22:06:35

Diagnostics Check		OK	22:00
Air Blank		0.000	22:00
Control Test		0.080	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.000	22:02
Breath Volume		2.847	22:02
Air Blank		0.000	22:02
Air Blank		0.000	22:04
Subject Sample	# 2	0.000	22:05
Breath Volume		4.281	22:05
Air Blank		0.000	22:05
Control Test		0.079	22:05
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 02/27/2018 12:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,GUILLERMO,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GARCIA,BRYAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 07:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 02/12/2018 10:07:31

Diagnostics Check		OK	10:04	
Air Blank		0.000	10:05	
Control Test		0.081	10:05	
Air Blank		0.000	10:05	
Subject Sample	# 1		10:06	Subject Test Refused
Breath Volume		0.000	10:06	
Air Blank		0.000	10:06	
Control Test		0.083	10:06	
Air Blank		0.000	10:07	
Diagnostics Check		OK	10:07	

Data Download Date/Time: 02/27/2018 12:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARGIA,FRANK,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,JASIEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: MADDIX

DATE/TIME OF TEST: 02/16/2018 00:58:44

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.081	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.213	00:54
Breath Volume		1.675	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.214	00:57
Breath Volume		1.152	00:57
Air Blank		0.000	00:57
Control Test		0.081	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 02/27/2018 12:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARGIA,FRANK,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TARGIA,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 02/24/2018 02:36:16

Diagnostics Check		OK	02:33	
Air Blank		0.000	02:33	
Control Test		0.080	02:34	
Air Blank		0.000	02:34	
Subject Sample	# 1		02:34	Subject Test Refused
Breath Volume		0.000	02:34	
Air Blank		0.000	02:35	
Control Test		0.080	02:35	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:36	

Data Download Date/Time: 02/27/2018 12:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TALANIA,ADRIAN,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUEZ,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 02/24/2018 04:42:48

Diagnostics Check		OK	04:39	
Air Blank		0.000	04:39	
Control Test		0.080	04:39	
Air Blank		0.000	04:40	
Subject Sample	# 1		04:41	Subject Test Refused
Breath Volume		0.000	04:41	
Air Blank		0.000	04:41	
Control Test		0.080	04:42	
Air Blank		0.000	04:42	
Diagnostics Check		OK	04:42	

Data Download Date/Time: 02/27/2018 12:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,TIMOTHY,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BROWN,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: PIRACOW

DATE/TIME OF TEST: 02/08/2018 21:44:46

Diagnostics Check		OK	21:41	
Air Blank		0.000	21:42	
Control Test		0.082	21:42	
Air Blank		0.000	21:42	
Subject Sample	# 1		21:43	Subject Test Refused
Breath Volume		0.000	21:43	
Air Blank		0.000	21:43	
Control Test		0.082	21:44	
Air Blank		0.000	21:44	
Diagnostics Check		OK	21:44	

Data Download Date/Time: 02/28/2018 17:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEON,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: ROBLES

DATE/TIME OF TEST: 02/15/2018 19:50:11

Diagnostics Check		OK	19:43
Air Blank		0.000	19:43
Control Test		0.080	19:44
Air Blank		0.000	19:44
Subject Sample	# 1	0.238	19:45
Breath Volume		2.121	19:45
Air Blank		0.000	19:46
Air Blank		0.000	19:47
Subject Sample	# 2	0.231	19:48
Breath Volume		2.351	19:48
Air Blank		0.000	19:49
Control Test		0.077	19:49
Air Blank		0.000	19:50
Diagnostics Check		OK	19:50

Data Download Date/Time: 02/28/2018 17:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,TIMOTHY,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BROWN,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: DRAKE

DATE/TIME OF TEST: 02/18/2018 00:33:56

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.080	00:28
Air Blank		0.000	00:28
Subject Sample	# 1	0.280	00:29
Breath Volume		1.785	00:29
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.288	00:32
Breath Volume		2.335	00:32
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 02/28/2018 17:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,TIMOTHY,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BROWN,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: GUTU

DATE/TIME OF TEST: 02/26/2018 03:25:03

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.081	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.175	03:20
Breath Volume		1.968	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:22
Subject Sample	# 2	0.173	03:23
Breath Volume		1.171	03:23
Air Blank		0.000	03:24
Control Test		0.079	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:25

Data Download Date/Time: 02/28/2018 17:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRER,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: CARBALLEIRA,JEANETTE
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 01/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/03/2018 20:44:00

Diagnostics Check		OK	20:40	
Air Blank		0.000	20:41	
Control Test		0.078	20:41	
Air Blank		0.000	20:42	
Subject Sample	# 1		20:42	Subject Test Refused
Breath Volume		0.000	20:42	
Air Blank		0.000	20:43	
Control Test		0.078	20:43	
Air Blank		0.000	20:43	
Diagnostics Check		OK	20:44	

Data Download Date/Time: 02/28/2018 09:14

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,EDISON,B
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: CRUZ,EDISON
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 01/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: BRADSHAW

DATE/TIME OF TEST: 02/05/2018 06:28:19

Diagnostics Check		OK	06:20
Air Blank		0.000	06:20
Control Test		0.077	06:20
Air Blank		0.000	06:21
Subject Sample	# 1	0.054	06:23
Breath Volume		2.671	06:23
Air Blank		0.000	06:24
Air Blank		0.000	06:25
Subject Sample	# 2	0.053	06:26
Breath Volume		2.148	06:26
Air Blank		0.000	06:27
Control Test		0.076	06:27
Air Blank		0.000	06:28
Diagnostics Check		OK	06:28

Data Download Date/Time: 02/28/2018 09:14

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAMBRANA,ERICK,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: FERRER,RAFAEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 01/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ZAMBRANA

DATE/TIME OF TEST: 02/08/2018 01:00:49

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.077	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.136	00:55
Breath Volume		2.703	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:57
Subject Sample	# 2	0.133	00:59
Breath Volume		2.382	00:59
Air Blank		0.000	00:59
Control Test		0.078	01:00
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 02/28/2018 09:14

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: DI DOMIZO

DATE/TIME OF TEST: 02/06/2018 02:02:25

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.076	01:55
Air Blank		0.000	01:55
Subject Sample	# 1	0.164	01:56
Breath Volume		1.425	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:58
Subject Sample	# 2	0.170	02:00
Breath Volume		1.703	02:00
Air Blank		0.000	02:01
Control Test		0.077	02:01
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 02/07/2018 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/06/2018 02:59:03

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.077	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.000	02:54
Breath Volume		2.910	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.000	02:57
Breath Volume		2.585	02:57
Air Blank		0.000	02:58
Control Test		0.077	02:58
Air Blank		0.000	02:58
Diagnostics Check		OK	02:59

Data Download Date/Time: 02/07/2018 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 02/11/2018 23:56:29

Diagnostics Check		OK	23:47
Air Blank		0.000	23:48
Control Test		0.077	23:48
Air Blank		0.000	23:49
Subject Sample	# 1	0.093	23:49
Breath Volume		1.214	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.095	23:54
Breath Volume		1.394	23:54
Air Blank		0.000	23:55
Control Test		0.079	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 03/01/2018 13:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: TERZADO

DATE/TIME OF TEST: 02/13/2018 03:03:56

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.076	02:58
Air Blank		0.000	02:58
Subject Sample	# 1	0.162	02:59
Breath Volume		2.308	02:59
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.162	03:02
Breath Volume		3.316	03:02
Air Blank		0.000	03:03
Control Test		0.077	03:03
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 03/01/2018 13:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MITRA

DATE/TIME OF TEST: 02/21/2018 00:23:57

Diagnostics Check		OK	00:16
Air Blank		0.000	00:16
Control Test		0.077	00:16
Air Blank		0.000	00:17
Subject Sample	# 1	0.133	00:18
Breath Volume		2.894	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.138	00:22
Breath Volume		2.156	00:22
Air Blank		0.000	00:23
Control Test		0.076	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 03/01/2018 13:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/21/2018 22:55:27

Diagnostics Check		OK	22:49
Air Blank		0.000	22:49
Control Test		0.076	22:49
Air Blank		0.000	22:50
Subject Sample	# 1	0.204	22:50
Breath Volume		2.042	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.208	22:53
Breath Volume		2.191	22:53
Air Blank		0.000	22:54
Control Test		0.077	22:54
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 03/01/2018 13:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: JEAN-BAPTISTE

DATE/TIME OF TEST: 02/25/2018 21:46:47

Diagnostics Check		OK	21:41	
Air Blank		0.000	21:42	
Control Test		0.077	21:42	
Air Blank		0.000	21:43	
Subject Sample	# 1		21:45	Subject Test Refused
Breath Volume		0.000	21:45	
Air Blank		0.000	21:45	
Control Test		0.078	21:46	
Air Blank		0.000	21:46	
Diagnostics Check		OK	21:46	

Data Download Date/Time: 03/01/2018 13:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MENGHI

DATE/TIME OF TEST: 02/26/2018 03:34:59

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.077	03:28
Air Blank		0.000	03:29
Subject Sample	# 1	0.162	03:29
Breath Volume		3.007	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.166	03:33
Breath Volume		3.234	03:33
Air Blank		0.000	03:34
Control Test		0.077	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 03/01/2018 13:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: HILLAN

DATE/TIME OF TEST: 02/27/2018 00:44:21

Diagnostics Check		OK	00:36
Air Blank		0.000	00:37
Control Test		0.077	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.212	00:39
Breath Volume		1.851	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.210	00:42
Breath Volume		1.222	00:42
Air Blank		0.000	00:43
Control Test		0.075	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 03/01/2018 13:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: PARRIS

DATE/TIME OF TEST: 02/03/2018 02:17:18

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.080	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.129	02:12
Breath Volume		1.511	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.135	02:15
Breath Volume		2.546	02:15
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: MORALES MARQUEZ

DATE/TIME OF TEST: 02/03/2018 02:55:55

Diagnostics Check		OK	02:49
Air Blank		0.000	02:49
Control Test		0.079	02:49
Air Blank		0.000	02:50
Subject Sample	# 1	0.213	02:51
Breath Volume		2.480	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:53
Subject Sample	# 2	0.199	02:54
Breath Volume		1.339	02:54
Air Blank		0.000	02:54
Control Test		0.079	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 02/03/2018 04:24:58

Diagnostics Check		OK	04:18
Air Blank		0.000	04:18
Control Test		0.080	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.168	04:20
Breath Volume		2.136	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:22
Subject Sample	# 2	0.170	04:23
Breath Volume		2.988	04:23
Air Blank		0.000	04:24
Control Test		0.079	04:24
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: RODRIGUEZ ORTIZ

DATE/TIME OF TEST: 02/03/2018 21:19:01

Diagnostics Check		OK	21:10
Air Blank		0.000	21:11
Control Test		0.080	21:11
Air Blank		0.000	21:12
Subject Sample	# 1	0.186	21:14
Breath Volume		1.945	21:14
Air Blank		0.000	21:14
Air Blank		0.000	21:16
Subject Sample	# 2	0.187	21:17
Breath Volume		2.226	21:17
Air Blank		0.000	21:18
Control Test		0.079	21:18
Air Blank		0.000	21:18
Diagnostics Check		OK	21:19

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 02/03/2018 21:49:47

Diagnostics Check		OK	21:42
Air Blank		0.000	21:43
Control Test		0.079	21:43
Air Blank		0.000	21:44
Subject Sample	# 1	0.105	21:44
Breath Volume		2.726	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.101	21:48
Breath Volume		2.957	21:48
Air Blank		0.000	21:48
Control Test		0.079	21:49
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MENDEZ PUAC

DATE/TIME OF TEST: 02/04/2018 00:49:45

Diagnostics Check		OK	00:36	
Air Blank		0.000	00:36	
Control Test		0.079	00:36	
Air Blank		0.000	00:37	
Subject Sample	# 1	0.281	00:40	Slope Not Level
Breath Volume		1.363	00:40	
Air Blank		0.000	00:41	
Air Blank		0.000	00:43	
Subject Sample	# 2	0.306	00:44	
Breath Volume		2.460	00:44	
Air Blank		0.000	00:45	
Air Blank		0.000	00:47	
Subject Sample	# 3	0.296	00:48	
Breath Volume		2.847	00:48	
Air Blank		0.000	00:48	
Control Test		0.077	00:49	
Air Blank		0.000	00:49	
Diagnostics Check		OK	00:49	

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: MEJIA MORALES

DATE/TIME OF TEST: 02/04/2018 01:54:22

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.079	01:49
Air Blank		0.000	01:49
Subject Sample	# 1	0.000	01:50
Breath Volume		1.328	01:50
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.000	01:52
Breath Volume		1.804	01:52
Air Blank		0.000	01:53
Control Test		0.079	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 02/05/2018 01:29:27

Diagnostics Check		OK	01:23
Air Blank		0.000	01:23
Control Test		0.079	01:23
Air Blank		0.000	01:24
Subject Sample	# 1	0.082	01:24
Breath Volume		1.218	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.082	01:27
Breath Volume		1.675	01:27
Air Blank		0.000	01:28
Control Test		0.080	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KELLY,MAX
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 02/05/2018 02:14:44

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.080	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.208	02:10
Breath Volume		3.460	02:10
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.209	02:13
Breath Volume		2.753	02:13
Air Blank		0.000	02:13
Control Test		0.079	02:14
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: AMY

DATE/TIME OF TEST: 02/07/2018 20:34:11

Diagnostics Check		OK	20:26
Air Blank		0.000	20:27
Control Test		0.080	20:27
Air Blank		0.000	20:27
Subject Sample	# 1	0.154	20:28
Breath Volume		1.175	20:28
Air Blank		0.000	20:29
Air Blank		0.000	20:31
Subject Sample	# 2	0.147	20:32
Breath Volume		1.207	20:32
Air Blank		0.000	20:33
Control Test		0.079	20:33
Air Blank		0.000	20:34
Diagnostics Check		OK	20:34

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GRAHAM,DALE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: SPAGNO

DATE/TIME OF TEST: 02/10/2018 00:26:44

Diagnostics Check		OK	00:19
Air Blank		0.000	00:20
Control Test		0.081	00:20
Air Blank		0.000	00:20
Subject Sample	# 1	0.202	00:21
Breath Volume		2.300	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.200	00:25
Breath Volume		2.253	00:25
Air Blank		0.000	00:25
Control Test		0.079	00:26
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 03/09/2018 14:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: OCHOA-LOPEZ

DATE/TIME OF TEST: 02/10/2018 03:28:29

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.080	03:22
Air Blank		0.000	03:22
Subject Sample	# 1	0.097	03:23
Breath Volume		1.539	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.101	03:26
Breath Volume		2.457	03:26
Air Blank		0.000	03:27
Control Test		0.079	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: VROMAN

DATE/TIME OF TEST: 02/10/2018 04:51:42

Diagnostics Check		OK	04:44
Air Blank		0.000	04:45
Control Test		0.079	04:45
Air Blank		0.000	04:45
Subject Sample	# 1	0.143	04:46
Breath Volume		1.570	04:46
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.146	04:50
Breath Volume		2.527	04:50
Air Blank		0.000	04:50
Control Test		0.078	04:51
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HERNANDEZ,MARK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:03

SUBJECT'S LAST NAME: SCITES

DATE/TIME OF TEST: 02/10/2018 14:55:18

Diagnostics Check		OK	14:48
Air Blank		0.000	14:49
Control Test		0.080	14:49
Air Blank		0.000	14:50
Subject Sample	# 1	0.000	14:50
Breath Volume		2.234	14:50
Air Blank		0.000	14:51
Air Blank		0.000	14:53
Subject Sample	# 2	0.000	14:53
Breath Volume		2.804	14:53
Air Blank		0.000	14:54
Control Test		0.081	14:54
Air Blank		0.000	14:55
Diagnostics Check		OK	14:55

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TRABAL,MILTON
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: GARCIA SANCHEZ

DATE/TIME OF TEST: 02/10/2018 22:20:23

Diagnostics Check		OK	22:13
Air Blank		0.000	22:14
Control Test		0.080	22:14
Air Blank		0.000	22:14
Subject Sample	# 1	0.116	22:15
Breath Volume		3.636	22:15
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.115	22:18
Breath Volume		3.941	22:18
Air Blank		0.000	22:19
Control Test		0.080	22:19
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: HERNANDEZ VASQUEZ

DATE/TIME OF TEST: 02/10/2018 23:56:03

Diagnostics Check		OK	23:47
Air Blank		0.000	23:48
Control Test		0.080	23:48
Air Blank		0.000	23:49
Subject Sample	# 1	0.215	23:50
Breath Volume		1.730	23:50
Air Blank		0.000	23:51
Air Blank		0.000	23:52
Subject Sample	# 2	0.222	23:54
Breath Volume		2.582	23:54
Air Blank		0.000	23:55
Control Test		0.079	23:55
Air Blank		0.000	23:55
Diagnostics Check		OK	23:56

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: LOZANO

DATE/TIME OF TEST: 02/11/2018 23:20:39

Diagnostics Check		OK	23:14
Air Blank		0.000	23:14
Control Test		0.080	23:15
Air Blank		0.000	23:15
Subject Sample	# 1	0.215	23:16
Breath Volume		4.285	23:16
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.214	23:18
Breath Volume		4.097	23:18
Air Blank		0.000	23:19
Control Test		0.080	23:20
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STREIGOLD,DAVID
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 02/13/2018 01:09:50

Diagnostics Check		OK	01:03
Air Blank		0.000	01:03
Control Test		0.080	01:04
Air Blank		0.000	01:04
Subject Sample	# 1	0.194	01:05
Breath Volume		2.339	01:05
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.195	01:08
Breath Volume		2.132	01:08
Air Blank		0.000	01:08
Control Test		0.079	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: RUBIESCOBAR

DATE/TIME OF TEST: 02/13/2018 22:46:11

Diagnostics Check		OK	22:38
Air Blank		0.000	22:39
Control Test		0.080	22:39
Air Blank		0.000	22:40
Subject Sample	# 1	0.179	22:41
Breath Volume		1.863	22:41
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.181	22:44
Breath Volume		1.773	22:44
Air Blank		0.000	22:45
Control Test		0.079	22:45
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHANK=,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SALM,PHILIP
ARRESTING OFFICER AGENCY: Ocean Ridge PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PASCAL

DATE/TIME OF TEST: 02/14/2018 02:52:20

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.080	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.146	02:47
Breath Volume		1.582	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:49
Subject Sample	# 2	0.144	02:50
Breath Volume		2.460	02:50
Air Blank		0.000	02:51
Control Test		0.077	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAFER,NANCY,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FRANQUI,JASON
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:07

SUBJECT'S LAST NAME: ROQUE

DATE/TIME OF TEST: 02/15/2018 12:50:02

Diagnostics Check		OK	12:43
Air Blank		0.000	12:43
Control Test		0.080	12:43
Air Blank		0.000	12:44
Subject Sample	# 1	0.272	12:45
Breath Volume		3.054	12:45
Air Blank		0.000	12:46
Air Blank		0.000	12:47
Subject Sample	# 2	0.265	12:48
Breath Volume		2.347	12:48
Air Blank		0.000	12:49
Control Test		0.079	12:49
Air Blank		0.000	12:49
Diagnostics Check		OK	12:50

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JENKINS II,HOWARD
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: DOERR

DATE/TIME OF TEST: 02/15/2018 23:54:12

Diagnostics Check		OK	23:46
Air Blank		0.000	23:47
Control Test		0.080	23:47
Air Blank		0.000	23:48
Subject Sample	# 1	0.218	23:49
Breath Volume		1.546	23:49
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.221	23:52
Breath Volume		1.480	23:52
Air Blank		0.000	23:53
Control Test		0.078	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: THEVENIN

DATE/TIME OF TEST: 02/16/2018 03:50:31

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.080	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.148	03:45
Breath Volume		1.410	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:47
Subject Sample	# 2	0.150	03:48
Breath Volume		2.019	03:48
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DOMINGUEZ,JAYMES
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: HEINZ

DATE/TIME OF TEST: 02/16/2018 15:39:33

Diagnostics Check		OK	15:33
Air Blank		0.000	15:34
Control Test		0.080	15:34
Air Blank		0.000	15:34
Subject Sample	# 1	0.000	15:35
Breath Volume		2.406	15:35
Air Blank		0.000	15:35
Air Blank		0.000	15:37
Subject Sample	# 2	0.000	15:38
Breath Volume		2.558	15:38
Air Blank		0.000	15:38
Control Test		0.079	15:38
Air Blank		0.000	15:39
Diagnostics Check		OK	15:39

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 02/17/2018 01:09:41

Diagnostics Check		OK	01:03
Air Blank		0.000	01:03
Control Test		0.080	01:04
Air Blank		0.000	01:04
Subject Sample	# 1	0.183	01:05
Breath Volume		1.902	01:05
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.182	01:08
Breath Volume		1.820	01:08
Air Blank		0.000	01:08
Control Test		0.079	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: DEBRIERE

DATE/TIME OF TEST: 02/17/2018 01:46:36

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.080	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.096	01:41
Breath Volume		2.410	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.095	01:44
Breath Volume		2.128	01:44
Air Blank		0.000	01:45
Control Test		0.079	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TRENTMAN,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: YLIPELKONEN

DATE/TIME OF TEST: 02/18/2018 20:03:04

Diagnostics Check		OK	19:56
Air Blank		0.000	19:56
Control Test		0.080	19:57
Air Blank		0.000	19:57
Subject Sample	# 1	0.030	19:58
Breath Volume		1.777	19:58
Air Blank		0.000	19:59
Air Blank		0.000	20:01
Subject Sample	# 2	0.026	20:01
Breath Volume		1.128	20:01
Air Blank		0.000	20:02
Control Test		0.080	20:02
Air Blank		0.000	20:02
Diagnostics Check		OK	20:03

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: MOULATRE

DATE/TIME OF TEST: 02/18/2018 22:12:47

Diagnostics Check		OK	22:04
Air Blank		0.000	22:04
Control Test		0.080	22:04
Air Blank		0.000	22:05
Subject Sample	# 1	0.216	22:08
Breath Volume		3.046	22:08
Air Blank		0.000	22:09
Air Blank		0.000	22:10
Subject Sample	# 2	0.204	22:11
Breath Volume		2.468	22:11
Air Blank		0.000	22:11
Control Test		0.076	22:12
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SHACKELFORD,JAMES
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: MEZA SANCHEZ

DATE/TIME OF TEST: 02/19/2018 04:16:08

Diagnostics Check		OK	04:09
Air Blank		0.000	04:10
Control Test		0.080	04:10
Air Blank		0.000	04:10
Subject Sample	# 1	0.206	04:11
Breath Volume		1.437	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.213	04:14
Breath Volume		1.812	04:14
Air Blank		0.000	04:15
Control Test		0.079	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWE,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GRUBER,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: CARY

DATE/TIME OF TEST: 02/19/2018 21:46:40

Diagnostics Check		OK	21:35
Air Blank		0.000	21:35
Control Test		0.080	21:35
Air Blank		0.000	21:36
Subject Sample	# 1	0.257	21:37
Breath Volume		2.226	21:37
Air Blank		0.000	21:38
Air Blank		0.000	21:39
Subject Sample	# 2	0.225	21:41
Breath Volume		1.109	21:41
Air Blank		0.000	21:42
Air Blank		0.000	21:44
Subject Sample	# 3	0.245	21:44
Breath Volume		2.718	21:44
Air Blank		0.000	21:45
Control Test		0.077	21:46
Air Blank		0.000	21:46
Diagnostics Check		OK	21:46

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: STRICKLER

DATE/TIME OF TEST: 02/20/2018 00:42:10

Diagnostics Check		OK	00:35
Air Blank		0.000	00:35
Control Test		0.080	00:36
Air Blank		0.000	00:36
Subject Sample	# 1	0.165	00:37
Breath Volume		1.386	00:37
Air Blank		0.000	00:38
Air Blank		0.000	00:39
Subject Sample	# 2	0.165	00:40
Breath Volume		1.421	00:40
Air Blank		0.000	00:41
Control Test		0.080	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 02/21/2018 22:29:19

Diagnostics Check		OK	22:22
Air Blank		0.000	22:22
Control Test		0.080	22:22
Air Blank		0.000	22:23
Subject Sample	# 1	0.064	22:24
Breath Volume		1.460	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:26
Subject Sample	# 2	0.063	22:27
Breath Volume		1.652	22:27
Air Blank		0.000	22:28
Control Test		0.078	22:28
Air Blank		0.000	22:29
Diagnostics Check		OK	22:29

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: BRIANT

DATE/TIME OF TEST: 02/23/2018 20:01:32

Diagnostics Check		OK	19:53
Air Blank		0.000	19:53
Control Test		0.081	19:53
Air Blank		0.000	19:54
Subject Sample	# 1	0.090	19:56
Breath Volume		4.003	19:56
Air Blank		0.000	19:57
Air Blank		0.000	19:58
Subject Sample	# 2	0.087	19:59
Breath Volume		3.070	19:59
Air Blank		0.000	20:00
Control Test		0.079	20:00
Air Blank		0.000	20:01
Diagnostics Check		OK	20:01

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: SANCHEZ MUX

DATE/TIME OF TEST: 02/23/2018 23:39:49

Diagnostics Check		OK	23:32
Air Blank		0.000	23:33
Control Test		0.080	23:33
Air Blank		0.000	23:33
Subject Sample	# 1	0.240	23:35
Breath Volume		1.390	23:35
Air Blank		0.000	23:35
Air Blank		0.000	23:37
Subject Sample	# 2	0.255	23:38
Breath Volume		1.320	23:38
Air Blank		0.000	23:38
Control Test		0.079	23:39
Air Blank		0.000	23:39
Diagnostics Check		OK	23:39

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROTHENBURG,JONATHAN
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: FRANCK

DATE/TIME OF TEST: 02/24/2018 03:16:46

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.080	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.140	03:09
Breath Volume		1.562	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.142	03:14
Breath Volume		1.628	03:14
Air Blank		0.000	03:15
Control Test		0.077	03:16
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AIME,JEAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: CLEMENS

DATE/TIME OF TEST: 02/24/2018 04:30:43

Diagnostics Check		OK	04:24
Air Blank		0.000	04:24
Control Test		0.080	04:25
Air Blank		0.000	04:25
Subject Sample	# 1	0.153	04:26
Breath Volume		1.273	04:26
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.156	04:29
Breath Volume		1.320	04:29
Air Blank		0.000	04:29
Control Test		0.080	04:30
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: GUERRA HERNANDEZ

DATE/TIME OF TEST: 02/24/2018 05:15:51

Diagnostics Check		OK	05:09
Air Blank		0.000	05:09
Control Test		0.079	05:10
Air Blank		0.000	05:10
Subject Sample	# 1	0.139	05:11
Breath Volume		2.808	05:11
Air Blank		0.000	05:12
Air Blank		0.000	05:13
Subject Sample	# 2	0.138	05:14
Breath Volume		2.464	05:14
Air Blank		0.000	05:14
Control Test		0.080	05:15
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LOIACONO,LEONARD
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 02/24/2018 07:17:38

Diagnostics Check		OK	07:11
Air Blank		0.000	07:12
Control Test		0.080	07:12
Air Blank		0.000	07:12
Subject Sample	# 1	0.185	07:13
Breath Volume		1.773	07:13
Air Blank		0.000	07:13
Air Blank		0.000	07:15
Subject Sample	# 2	0.183	07:15
Breath Volume		2.062	07:15
Air Blank		0.000	07:16
Control Test		0.080	07:16
Air Blank		0.000	07:17
Diagnostics Check		OK	07:17

Data Download Date/Time: 03/09/2018 14:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VALERIO,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: BOUCHARD

DATE/TIME OF TEST: 02/24/2018 19:43:45

Diagnostics Check		OK	19:36
Air Blank		0.000	19:36
Control Test		0.080	19:37
Air Blank		0.000	19:37
Subject Sample	# 1	0.181	19:38
Breath Volume		2.476	19:38
Air Blank		0.000	19:39
Air Blank		0.000	19:41
Subject Sample	# 2	0.165	19:42
Breath Volume		1.363	19:42
Air Blank		0.000	19:42
Control Test		0.079	19:43
Air Blank		0.000	19:43
Diagnostics Check		OK	19:43

Data Download Date/Time: 03/09/2018 14:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: ANDINO IZAGUIRRE

DATE/TIME OF TEST: 02/24/2018 23:56:26

Diagnostics Check		OK	23:49
Air Blank		0.000	23:50
Control Test		0.080	23:50
Air Blank		0.000	23:50
Subject Sample	# 1	0.207	23:51
Breath Volume		2.277	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.207	23:54
Breath Volume		1.480	23:54
Air Blank		0.000	23:55
Control Test		0.079	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 03/09/2018 14:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLOWERS,GARY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: CLEARY

DATE/TIME OF TEST: 02/25/2018 02:55:39

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.000	02:51
Breath Volume		1.359	02:51
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.000	02:54
Breath Volume		1.828	02:54
Air Blank		0.000	02:54
Control Test		0.081	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 03/09/2018 14:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HORNE,ASHTON
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/27/2018 03:11:15

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.080	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.000	03:07
Breath Volume		2.066	03:07
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.000	03:09
Breath Volume		2.828	03:09
Air Blank		0.000	03:10
Control Test		0.081	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 03/09/2018 14:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: KELAHER

DATE/TIME OF TEST: 02/27/2018 23:19:50

Diagnostics Check		OK	23:16	
Air Blank		0.000	23:16	
Control Test		0.081	23:16	
Air Blank		0.000	23:17	
Subject Sample	# 1		23:18	Subject Test Refused
Breath Volume		0.000	23:18	
Air Blank		0.000	23:18	
Control Test		0.081	23:19	
Air Blank		0.000	23:19	
Diagnostics Check		OK	23:19	

Data Download Date/Time: 03/09/2018 14:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,J,G
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: OCAMPO,OMAR
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: O'NEAL

DATE/TIME OF TEST: 02/02/2018 02:42:52

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:38	
Control Test		0.080	02:38	
Air Blank		0.000	02:39	
Subject Sample	# 1		02:41	Subject Test Refused
Breath Volume		0.000	02:41	
Air Blank		0.000	02:41	
Control Test		0.080	02:42	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:42	

Data Download Date/Time: 02/28/2018 18:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRAY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WRAY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: CANCEL

DATE/TIME OF TEST: 02/04/2018 00:27:20

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.080	00:21
Air Blank		0.000	00:22
Subject Sample	# 1	0.266	00:22
Breath Volume		2.917	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.270	00:25
Breath Volume		2.812	00:25
Air Blank		0.000	00:26
Control Test		0.079	00:26
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 02/28/2018 18:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GADSON,LAUREN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: TOWNS,R
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 02/04/2018 23:40:53

Diagnostics Check		OK	23:37	
Air Blank		0.000	23:37	
Control Test		0.080	23:38	
Air Blank		0.000	23:38	
Subject Sample	# 1		23:39	Subject Test Refused
Breath Volume		0.000	23:39	
Air Blank		0.000	23:39	
Control Test		0.080	23:40	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 02/28/2018 18:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JENNIFER,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PREVATT,JENNIFER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: MILLMAN

DATE/TIME OF TEST: 02/05/2018 23:21:34

Diagnostics Check		OK	23:18	
Air Blank		0.000	23:19	
Control Test		0.080	23:19	
Air Blank		0.000	23:19	
Subject Sample	# 1		23:20	Subject Test Refused
Breath Volume		0.000	23:20	
Air Blank		0.000	23:20	
Control Test		0.079	23:20	
Air Blank		0.000	23:21	
Diagnostics Check		OK	23:21	

Data Download Date/Time: 02/28/2018 18:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,KYLE,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SYLVESTER,T
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MCNITT

DATE/TIME OF TEST: 02/10/2018 03:32:15

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:28	
Control Test		0.081	03:29	
Air Blank		0.000	03:29	
Subject Sample	# 1		03:30	Subject Test Refused
Breath Volume		0.000	03:30	
Air Blank		0.000	03:31	
Control Test		0.080	03:31	
Air Blank		0.000	03:32	
Diagnostics Check		OK	03:32	

Data Download Date/Time: 02/28/2018 18:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,KYLE,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: ST JOHN,JORDAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: IVERSON

DATE/TIME OF TEST: 02/10/2018 04:28:16

Diagnostics Check		OK	04:18
Air Blank		0.000	04:18
Control Test		0.080	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.135	04:21
Breath Volume		1.210	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.135	04:26
Breath Volume		1.414	04:26
Air Blank		0.000	04:27
Control Test		0.079	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 02/28/2018 18:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHEWNING,AUSTIN,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CHEWNING,AUSTIN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: CHAPEL

DATE/TIME OF TEST: 02/10/2018 22:28:51

Diagnostics Check		OK	22:22
Air Blank		0.000	22:22
Control Test		0.080	22:22
Air Blank		0.000	22:23
Subject Sample	# 1	0.186	22:24
Breath Volume		2.597	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:26
Subject Sample	# 2	0.186	22:27
Breath Volume		2.304	22:27
Air Blank		0.000	22:27
Control Test		0.079	22:28
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 02/28/2018 18:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,KYLE,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: ST JOHN,JORDAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: RIVERA DE JESUS

DATE/TIME OF TEST: 02/11/2018 22:34:26

Diagnostics Check		OK	22:27
Air Blank		0.000	22:28
Control Test		0.080	22:28
Air Blank		0.000	22:29
Subject Sample	# 1	0.198	22:29
Breath Volume		1.285	22:29
Air Blank		0.000	22:30
Air Blank		0.000	22:32
Subject Sample	# 2	0.205	22:32
Breath Volume		3.332	22:32
Air Blank		0.000	22:33
Control Test		0.078	22:33
Air Blank		0.000	22:34
Diagnostics Check		OK	22:34

Data Download Date/Time: 02/28/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GADSON,LAUREN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CHAMBERS,J
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: BREHM

DATE/TIME OF TEST: 02/13/2018 00:51:20

Diagnostics Check		OK	00:48	
Air Blank		0.000	00:48	
Control Test		0.080	00:49	
Air Blank		0.000	00:49	
Subject Sample	# 1		00:49	Subject Test Refused
Breath Volume		0.000	00:49	
Air Blank		0.000	00:50	
Control Test		0.079	00:50	
Air Blank		0.000	00:51	
Diagnostics Check		OK	00:51	

Data Download Date/Time: 02/28/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,MITA,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: ST. JOHN,JORDAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: CARLIN

DATE/TIME OF TEST: 02/20/2018 04:02:48

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.080	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.151	03:58
Breath Volume		2.039	03:58
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.151	04:01
Breath Volume		1.765	04:01
Air Blank		0.000	04:01
Control Test		0.079	04:02
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 02/28/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEMP,F,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GADDIE,K
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BONNER

DATE/TIME OF TEST: 02/25/2018 02:28:29

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.189	02:23
Breath Volume		1.750	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.197	02:26
Breath Volume		2.070	02:26
Air Blank		0.000	02:27
Control Test		0.076	02:27
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 02/28/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEMP,F,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BARKOSKIE,A
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 02/25/2018 03:16:40

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.080	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.065	03:12
Breath Volume		1.757	03:12
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.061	03:15
Breath Volume		2.933	03:15
Air Blank		0.000	03:15
Control Test		0.078	03:16
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 02/28/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGLYNN,JOHN,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MCGLYNN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: GARVEY

DATE/TIME OF TEST: 02/03/2018 02:36:58

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:34	
Control Test		0.080	02:34	
Air Blank		0.000	02:35	
Subject Sample	# 1		02:35	Subject Test Refused
Breath Volume		0.000	02:35	
Air Blank		0.000	02:36	
Control Test		0.079	02:36	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:36	

Data Download Date/Time: 02/28/2018 09:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGLYNN,JOHN,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MCGLYNN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 02/04/2018 02:40:32

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:37	
Control Test		0.079	02:38	
Air Blank		0.000	02:38	
Subject Sample	# 1		02:39	Subject Test Refused
Breath Volume		0.000	02:39	
Air Blank		0.000	02:39	
Control Test		0.083	02:39	
Air Blank			02:40	RFI Detect
Air Blank		0.000	02:40	

Data Download Date/Time: 02/28/2018 09:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGLYNN,JOHN,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MCGLYNN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 02/04/2018 02:46:36

Diagnostics Check		OK	02:43	
Air Blank		0.000	02:44	
Control Test		0.079	02:44	
Air Blank		0.000	02:45	
Subject Sample	# 1		02:45	Subject Test Refused
Breath Volume		0.000	02:45	
Air Blank		0.000	02:45	
Control Test		0.079	02:46	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:46	

Data Download Date/Time: 02/28/2018 09:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGLYNN,JOHN,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MCGLYNN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 02/17/2018 03:40:08

Diagnostics Check		OK	03:35	
Air Blank		0.000	03:36	
Control Test		0.079	03:36	
Air Blank		0.000	03:37	
Subject Sample	# 1		03:38	Subject Test Refused
Breath Volume		0.000	03:38	
Air Blank		0.000	03:39	
Control Test		0.080	03:39	
Air Blank		0.000	03:40	
Diagnostics Check		OK	03:40	

Data Download Date/Time: 02/28/2018 09:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGLYNN,JOHN,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MCGLYNN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/26/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 16080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 02/20/2018 23:26:31

Diagnostics Check		OK	23:21	
Air Blank		0.000	23:21	
Control Test		0.078	23:21	
Air Blank		0.000	23:22	
Subject Sample	# 1		23:25	RFI Detect
Breath Volume		0.000	23:25	
Air Blank		0.000	23:25	
Control Test		0.079	23:25	
Air Blank		0.000	23:26	
Diagnostics Check		OK	23:26	

Data Download Date/Time: 02/28/2018 09:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGLYNN,JOHN,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MCGLYNN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 02/20/2018 23:36:20

Diagnostics Check		OK	23:29
Air Blank		0.000	23:30
Control Test		0.078	23:30
Air Blank		0.000	23:30
Subject Sample	# 1	0.194	23:31
Breath Volume		1.855	23:31
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.184	23:34
Breath Volume		1.667	23:34
Air Blank		0.000	23:35
Control Test		0.079	23:35
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 02/28/2018 09:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: MARES AZPEITIA

DATE/TIME OF TEST: 02/02/2018 21:34:35

Diagnostics Check		OK	21:28
Air Blank		0.000	21:28
Control Test		0.079	21:28
Air Blank		0.000	21:29
Subject Sample	# 1	0.146	21:30
Breath Volume		1.511	21:30
Air Blank		0.000	21:30
Air Blank		0.000	21:32
Subject Sample	# 2	0.156	21:32
Breath Volume		2.730	21:32
Air Blank		0.000	21:33
Control Test		0.080	21:33
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 02/26/2018 11:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: CIURCA

DATE/TIME OF TEST: 02/16/2018 23:41:45

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.079	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.127	23:37
Breath Volume		1.757	23:37
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.140	23:40
Breath Volume		1.707	23:40
Air Blank		0.000	23:40
Control Test		0.078	23:41
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 02/26/2018 11:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 02/01/2018 02:28:55

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.082	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.182	02:24
Breath Volume		1.632	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:26
Subject Sample	# 2	0.176	02:27
Breath Volume		1.410	02:27
Air Blank		0.000	02:27
Control Test		0.081	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 02/27/2018 22:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: STRONG

DATE/TIME OF TEST: 02/06/2018 01:58:26

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.082	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.173	01:53
Breath Volume		1.859	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.164	01:56
Breath Volume		1.296	01:56
Air Blank		0.000	01:57
Control Test		0.082	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 02/27/2018 22:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: JIMON RODRIGUEZ

DATE/TIME OF TEST: 02/08/2018 21:10:24

Diagnostics Check		OK	21:02
Air Blank		0.000	21:02
Control Test		0.082	21:03
Air Blank		0.000	21:03
Subject Sample	# 1	0.216	21:04
Breath Volume		1.820	21:04
Air Blank		0.000	21:05
Air Blank		0.000	21:07
Subject Sample	# 2	0.217	21:08
Breath Volume		1.898	21:08
Air Blank		0.000	21:09
Control Test		0.080	21:09
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 02/27/2018 22:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 02/10/2018 02:18:32

Diagnostics Check		OK	02:12
Air Blank		0.000	02:12
Control Test		0.083	02:13
Air Blank		0.000	02:13
Subject Sample	# 1	0.139	02:14
Breath Volume		1.441	02:14
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.138	02:16
Breath Volume		2.363	02:16
Air Blank		0.000	02:17
Control Test		0.082	02:17
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 02/27/2018 22:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 02/11/2018 00:19:11

Diagnostics Check		OK	00:12
Air Blank		0.000	00:13
Control Test		0.081	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.171	00:14
Breath Volume		2.105	00:14
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.164	00:17
Breath Volume		2.238	00:17
Air Blank		0.000	00:18
Control Test		0.081	00:18
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 02/27/2018 22:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: GORHAM

DATE/TIME OF TEST: 02/20/2018 03:21:01

Diagnostics Check		OK	03:14
Air Blank		0.000	03:14
Control Test		0.081	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.143	03:16
Breath Volume		1.359	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:18
Subject Sample	# 2	0.137	03:19
Breath Volume		1.316	03:19
Air Blank		0.000	03:20
Control Test		0.080	03:20
Air Blank		0.000	03:20
Diagnostics Check		OK	03:21

Data Download Date/Time: 02/27/2018 22:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: DEJESUS

DATE/TIME OF TEST: 02/24/2018 21:13:07

Diagnostics Check		OK	21:04
Air Blank		0.000	21:04
Control Test		0.080	21:05
Air Blank		0.000	21:05
Subject Sample	# 1	0.136	21:08
Breath Volume		1.503	21:08
Air Blank		0.000	21:09
Air Blank		0.000	21:10
Subject Sample	# 2	0.120	21:11
Breath Volume		1.238	21:11
Air Blank		0.000	21:12
Control Test		0.078	21:12
Air Blank		0.000	21:13
Diagnostics Check		OK	21:13

Data Download Date/Time: 02/27/2018 22:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMBRIATI,NICHOLAS,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: AMBRIATI,NICHOLAS
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 14:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 02/06/2018 15:52:33

Diagnostics Check		OK	15:46
Air Blank		0.000	15:46
Control Test		0.080	15:46
Air Blank		0.000	15:47
Subject Sample	# 1	0.000	15:47
Breath Volume		1.578	15:47
Air Blank		0.000	15:48
Air Blank		0.000	15:50
Subject Sample	# 2	0.000	15:51
Breath Volume		4.492	15:51
Air Blank		0.000	15:51
Control Test		0.078	15:51
Air Blank		0.000	15:52
Diagnostics Check		OK	15:52

Data Download Date/Time: 02/26/2018 11:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,BARBARA,B
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,BARBARA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 12:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 02/08/2018 16:01:54

Diagnostics Check		OK	15:55
Air Blank		0.000	15:56
Control Test		0.080	15:56
Air Blank		0.000	15:56
Subject Sample	# 1	0.000	15:57
Breath Volume		4.203	15:57
Air Blank		0.000	15:58
Air Blank		0.000	16:00
Subject Sample	# 2	0.000	16:00
Breath Volume		3.937	16:00
Air Blank		0.000	16:00
Control Test		0.080	16:01
Air Blank		0.000	16:01
Diagnostics Check		OK	16:01

Data Download Date/Time: 02/26/2018 11:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: QUILES RAMOS

DATE/TIME OF TEST: 02/11/2018 03:34:27

Diagnostics Check		OK	03:27
Air Blank		0.000	03:28
Control Test		0.080	03:28
Air Blank		0.000	03:29
Subject Sample	# 1	0.157	03:29
Breath Volume		1.886	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.145	03:32
Breath Volume		1.277	03:32
Air Blank		0.000	03:33
Control Test		0.080	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 02/26/2018 11:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: CASTRO ZAVALA

DATE/TIME OF TEST: 02/12/2018 01:45:46

Diagnostics Check		OK	01:38
Air Blank		0.000	01:39
Control Test		0.080	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.086	01:40
Breath Volume		2.761	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.079	01:44
Breath Volume		2.199	01:44
Air Blank		0.000	01:44
Control Test		0.077	01:45
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 02/26/2018 11:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMBRIATI,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: RIVARD,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 10:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:12

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 02/13/2018 11:38:19

Diagnostics Check		OK	11:35	
Air Blank		0.000	11:35	
Control Test		0.080	11:36	
Air Blank		0.000	11:36	
Subject Sample	# 1		11:36	Subject Test Refused
Breath Volume		0.000	11:36	
Air Blank		0.000	11:37	
Control Test		0.080	11:37	
Air Blank		0.000	11:38	
Diagnostics Check		OK	11:38	

Data Download Date/Time: 02/26/2018 11:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GWODZ,OFC
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 02/14/2018 05:00:27

Diagnostics Check		OK	04:55	
Air Blank		0.000	04:56	
Control Test		0.079	04:56	
Air Blank		0.000	04:57	
Subject Sample	# 1		04:59	Subject Test Refused
Breath Volume		0.000	04:59	
Air Blank		0.000	04:59	
Control Test		0.080	04:59	
Air Blank		0.000	05:00	
Diagnostics Check		OK	05:00	

Data Download Date/Time: 02/26/2018 11:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,RONALD,E
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: PETERSON,RONALD
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: OLITSKY

DATE/TIME OF TEST: 02/15/2018 02:54:48

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.079	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.173	02:49
Breath Volume		1.449	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.168	02:53
Breath Volume		2.359	02:53
Air Blank		0.000	02:53
Control Test		0.079	02:54
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 02/26/2018 11:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: DAVIS,D
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: DUDEMAINE

DATE/TIME OF TEST: 02/17/2018 04:27:56

Diagnostics Check		OK	04:21
Air Blank		0.000	04:22
Control Test		0.080	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.079	04:23
Breath Volume		2.007	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:25
Subject Sample	# 2	0.078	04:26
Breath Volume		1.699	04:26
Air Blank		0.000	04:27
Control Test		0.080	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 02/26/2018 11:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GWODZ,OFC
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 02/17/2018 05:06:48

Diagnostics Check		OK	04:57
Air Blank		0.000	04:57
Control Test		0.079	04:57
Air Blank		0.000	04:58
Subject Sample	# 1	0.140	05:01
Breath Volume		1.867	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.135	05:05
Breath Volume		1.664	05:05
Air Blank		0.000	05:05
Control Test		0.076	05:06
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 02/26/2018 11:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: DAVIS,DONALD
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: LUNSFORD

DATE/TIME OF TEST: 02/21/2018 21:59:56

Diagnostics Check		OK	21:51
Air Blank		0.000	21:51
Control Test		0.080	21:51
Air Blank		0.000	21:52
Subject Sample	# 1	0.214	21:55
Breath Volume		1.175	21:55
Air Blank		0.000	21:55
Air Blank		0.000	21:57
Subject Sample	# 2	0.210	21:58
Breath Volume		2.601	21:58
Air Blank		0.000	21:59
Control Test		0.078	21:59
Air Blank		0.000	21:59
Diagnostics Check		OK	21:59

Data Download Date/Time: 02/26/2018 11:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 02/24/2018 00:47:36

Diagnostics Check		OK	00:44	
Air Blank		0.000	00:44	
Control Test		0.079	00:45	
Air Blank		0.000	00:45	
Subject Sample	# 1		00:46	Subject Test Refused
Breath Volume		0.000	00:46	
Air Blank		0.000	00:46	
Control Test		0.080	00:47	
Air Blank		0.000	00:47	
Diagnostics Check		OK	00:47	

Data Download Date/Time: 02/26/2018 11:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,AARON,K
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: GONZALES,AARON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: BUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BOOTH

DATE/TIME OF TEST: 02/25/2018 21:46:40

Diagnostics Check		OK	21:40
Air Blank		0.000	21:40
Control Test		0.079	21:40
Air Blank		0.000	21:41
Subject Sample	# 1	0.135	21:42
Breath Volume		2.644	21:42
Air Blank		0.000	21:42
Air Blank		0.000	21:44
Subject Sample	# 2	0.135	21:45
Breath Volume		3.460	21:45
Air Blank		0.000	21:45
Control Test		0.079	21:46
Air Blank		0.000	21:46
Diagnostics Check		OK	21:46

Data Download Date/Time: 02/26/2018 11:57

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GWODZ,OFC
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 02/27/2018 01:24:56

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.080	01:19
Air Blank		0.000	01:19
Subject Sample	# 1	0.220	01:20
Breath Volume		2.437	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:22
Subject Sample	# 2	0.216	01:23
Breath Volume		2.234	01:23
Air Blank		0.000	01:24
Control Test		0.078	01:24
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 03/26/2018 11:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,FABIO,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: MEJIAS ROSADO

DATE/TIME OF TEST: 02/16/2018 01:48:42

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.078	01:43
Air Blank		0.000	01:43
Subject Sample	# 1	0.108	01:44
Breath Volume		2.261	01:44
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.092	01:47
Breath Volume		1.128	01:47
Air Blank		0.000	01:47
Control Test		0.076	01:48
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 03/01/2018 22:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ ORTIZ,JESUS,E
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: ORTIZ ORTIZ

DATE/TIME OF TEST: 02/18/2018 05:06:21

Diagnostics Check		OK	04:59
Air Blank		0.000	05:00
Control Test		0.079	05:00
Air Blank		0.000	05:01
Subject Sample	# 1	0.144	05:01
Breath Volume		3.066	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.138	05:04
Breath Volume		2.007	05:04
Air Blank		0.000	05:05
Control Test		0.076	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 03/01/2018 22:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,TABARIE,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,TABARIE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12346
DRY GAS EXPIRY DATE: 11/11/2017
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: TRACY

DATE/TIME OF TEST: 02/01/2018 20:11:05

Diagnostics Check		OK	19:57	
Air Blank		0.000	19:58	
Control Test		0.080	19:58	
Air Blank		0.000	19:59	
Subject Sample	# 1	0.008	20:00	
Breath Volume		1.167	20:00	
Air Blank		0.000	20:01	
Air Blank		0.000	20:03	
Subject Sample	# 2	0.007	20:06	Volume Not Met
Breath Volume		1.097	20:06	
Air Blank		0.000	20:06	
Air Blank		0.000	20:08	
Subject Sample	# 3	0.000	20:09	
Breath Volume		1.402	20:09	
Air Blank		0.000	20:10	
Control Test		0.081	20:10	
Air Blank		0.000	20:11	
Diagnostics Check		OK	20:11	

Data Download Date/Time: 02/27/2018 23:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,TABARIE,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,TABARIE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12346
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: ALONZO GARCIA

DATE/TIME OF TEST: 02/10/2018 02:07:24

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.080	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.195	02:02
Breath Volume		1.621	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:04
Subject Sample	# 2	0.198	02:05
Breath Volume		2.035	02:05
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 02/27/2018 23:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,TABARIE,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,TABARIE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12346
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: ZAGAL BENITEZ

DATE/TIME OF TEST: 02/12/2018 03:37:21

Diagnostics Check		OK	03:24	
Air Blank		0.000	03:25	
Control Test		0.079	03:25	
Air Blank		0.000	03:26	
Subject Sample	# 1	0.141	03:26	
Breath Volume		1.156	03:26	
Air Blank		0.000	03:27	
Air Blank		0.000	03:28	
Subject Sample	# 2	0.137	03:32	Volume Not Met
Breath Volume		0.089	03:32	
Air Blank		0.000	03:32	
Air Blank		0.000	03:34	
Subject Sample	# 3	0.132	03:35	
Breath Volume		1.183	03:35	
Air Blank		0.000	03:36	
Control Test		0.078	03:36	
Air Blank		0.000	03:37	
Diagnostics Check		OK	03:37	

Data Download Date/Time: 02/27/2018 23:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,TABARIE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 02/16/2018 02:11:20

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.079	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.171	02:06
Breath Volume		1.339	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:08
Subject Sample	# 2	0.181	02:09
Breath Volume		1.843	02:09
Air Blank		0.000	02:10
Control Test		0.077	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 02/27/2018 23:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,TABARIE,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,TABARIE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: ELMHURST

DATE/TIME OF TEST: 02/20/2018 18:57:48

Diagnostics Check		OK	18:50
Air Blank		0.000	18:50
Control Test		0.079	18:51
Air Blank		0.000	18:51
Subject Sample	# 1	0.000	18:53
Breath Volume		1.636	18:53
Air Blank		0.000	18:53
Air Blank		0.000	18:55
Subject Sample	# 2	0.000	18:56
Breath Volume		1.140	18:56
Air Blank		0.000	18:56
Control Test		0.080	18:57
Air Blank		0.000	18:57
Diagnostics Check		OK	18:57

Data Download Date/Time: 02/27/2018 23:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: ALAMILLO

DATE/TIME OF TEST: 02/21/2018 22:20:27

Diagnostics Check		OK	22:14
Air Blank		0.000	22:14
Control Test		0.078	22:15
Air Blank		0.000	22:15
Subject Sample	# 1	0.163	22:16
Breath Volume		2.441	22:16
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.158	22:18
Breath Volume		1.304	22:18
Air Blank		0.000	22:19
Control Test		0.077	22:19
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 02/27/2018 22:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: RICHTER

DATE/TIME OF TEST: 02/22/2018 00:19:11

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:16	
Control Test		0.078	00:16	
Air Blank		0.000	00:17	
Subject Sample	# 1		00:17	Subject Test Refused
Breath Volume		0.000	00:17	
Air Blank		0.000	00:18	
Control Test		0.078	00:18	
Air Blank		0.000	00:19	
Diagnostics Check		OK	00:19	

Data Download Date/Time: 02/27/2018 22:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: THOMPSON JR

DATE/TIME OF TEST: 02/24/2018 01:01:09

Diagnostics Check		OK	00:54
Air Blank		0.000	00:55
Control Test		0.079	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.185	00:56
Breath Volume		3.250	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.181	00:59
Breath Volume		3.527	00:59
Air Blank		0.000	01:00
Control Test		0.078	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 02/27/2018 22:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:38

SUBJECT'S LAST NAME: DA ROSA

DATE/TIME OF TEST: 02/24/2018 17:03:52

Diagnostics Check		OK	16:58	
Air Blank		0.000	16:59	
Control Test		0.078	16:59	
Air Blank		0.000	16:59	
Subject Sample	# 1		17:02	Subject Test Refused
Breath Volume		0.000	17:02	
Air Blank		0.000	17:02	
Control Test		0.077	17:03	
Air Blank		0.000	17:03	
Diagnostics Check		OK	17:03	

Data Download Date/Time: 02/27/2018 22:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTSON,CHRIS,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: MATTSON,CHRIS
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: RIVERA HERNANDEZ

DATE/TIME OF TEST: 02/03/2018 23:07:33

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:55	
Control Test		0.080	22:55	
Air Blank		0.000	22:56	
Subject Sample	# 1	0.157	22:59	Volume Not Met
Breath Volume		0.097	22:59	
Air Blank		0.000	23:00	
Air Blank		0.000	23:02	
Subject Sample	# 2	0.156	23:02	
Breath Volume		2.128	23:02	
Air Blank		0.000	23:03	
Air Blank		0.000	23:05	
Subject Sample	# 3	0.158	23:05	
Breath Volume		1.171	23:05	
Air Blank		0.000	23:06	
Control Test		0.078	23:06	
Air Blank		0.000	23:07	
Diagnostics Check		OK	23:07	

Data Download Date/Time: 03/01/2018 16:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 16:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:56

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 02/07/2018 17:24:33

Diagnostics Check		OK	17:16
Air Blank		0.000	17:17
Control Test		0.080	17:17
Air Blank		0.000	17:18
Subject Sample	# 1	0.027	17:20
Breath Volume		2.203	17:20
Air Blank		0.000	17:20
Air Blank		0.000	17:22
Subject Sample	# 2	0.026	17:23
Breath Volume		1.316	17:23
Air Blank		0.000	17:23
Control Test		0.079	17:23
Air Blank		0.000	17:24
Diagnostics Check		OK	17:24

Data Download Date/Time: 03/01/2018 16:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'KEEFE,PATRICIA,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: PALETTI

DATE/TIME OF TEST: 02/07/2018 23:32:07

Diagnostics Check		OK	23:15	
Air Blank		0.000	23:16	
Control Test		0.080	23:16	
Air Blank		0.000	23:17	
Subject Sample	# 1		23:20	No Sample Provided
Breath Volume		0.000	23:20	
Air Blank		0.000	23:20	
Air Blank		0.000	23:22	
Subject Sample	# 2	0.339	23:24	
Breath Volume		2.304	23:24	
Air Blank		0.000	23:25	
Air Blank		0.000	23:27	
Subject Sample	# 3	0.273	23:30	Volume Not Met
Breath Volume		0.300	23:30	
Air Blank		0.000	23:31	
Control Test		0.080	23:31	
Air Blank		0.000	23:32	
Diagnostics Check		OK	23:32	

Data Download Date/Time: 03/01/2018 16:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'KEEFE,PATRICIA,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: PALETTI

DATE/TIME OF TEST: 02/07/2018 23:44:22

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:35	
Control Test		0.080	23:36	
Air Blank		0.000	23:36	
Subject Sample	# 1	0.306	23:39	
Breath Volume		1.500	23:39	
Air Blank		0.000	23:40	
Air Blank		0.000	23:41	
Subject Sample	# 2		23:43	Subject Test Refused
Breath Volume		0.000	23:43	
Air Blank		0.000	23:43	
Control Test		0.077	23:43	
Air Blank		0.000	23:44	
Diagnostics Check		OK	23:44	

Data Download Date/Time: 03/01/2018 16:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ALVAREZ,ORLANDO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: ROCKCASTLE

DATE/TIME OF TEST: 02/09/2018 20:21:04

Diagnostics Check		OK	20:17	
Air Blank		0.000	20:18	
Control Test		0.080	20:18	
Air Blank		0.000	20:19	
Subject Sample	# 1		20:19	Subject Test Refused
Breath Volume		0.000	20:19	
Air Blank		0.000	20:20	
Control Test		0.079	20:20	
Air Blank		0.000	20:20	
Diagnostics Check		OK	20:21	

Data Download Date/Time: 03/01/2018 16:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERNS,CODY,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MILLER,JAMIE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: KILMER

DATE/TIME OF TEST: 02/09/2018 21:26:09

Diagnostics Check		OK	21:23	
Air Blank		0.000	21:23	
Control Test		0.080	21:24	
Air Blank		0.000	21:24	
Subject Sample	# 1		21:24	Subject Test Refused
Breath Volume		0.000	21:24	
Air Blank		0.000	21:25	
Control Test		0.079	21:25	
Air Blank		0.000	21:26	
Diagnostics Check		OK	21:26	

Data Download Date/Time: 03/01/2018 16:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORELL,JONAS,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ROCHA,PAUL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: FERRIE

DATE/TIME OF TEST: 02/10/2018 01:49:16

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:46	
Control Test		0.080	01:47	
Air Blank		0.000	01:47	
Subject Sample	# 1		01:47	Subject Test Refused
Breath Volume		0.000	01:47	
Air Blank		0.000	01:48	
Control Test		0.079	01:48	
Air Blank		0.000	01:49	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 03/01/2018 16:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERNS,CODY,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MILLER,JAMIE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 02/10/2018 04:20:57

Diagnostics Check		OK	04:17	
Air Blank		0.000	04:18	
Control Test		0.080	04:18	
Air Blank		0.000	04:19	
Subject Sample	# 1		04:19	Subject Test Refused
Breath Volume		0.000	04:19	
Air Blank		0.000	04:20	
Control Test		0.080	04:20	
Air Blank		0.000	04:20	
Diagnostics Check		OK	04:20	

Data Download Date/Time: 03/01/2018 16:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORELL,JONAS,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/18/2018 04:45:13

Diagnostics Check		OK	04:35
Air Blank		0.000	04:35
Control Test		0.079	04:36
Air Blank		0.000	04:36
Subject Sample	# 1	0.202	04:37
Breath Volume		2.101	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.225	04:40
Breath Volume		2.460	04:40
Air Blank		0.000	04:41
Air Blank		0.000	04:43
Subject Sample	# 3	0.213	04:43
Breath Volume		1.933	04:43
Air Blank		0.000	04:44
Control Test		0.076	04:44
Air Blank		0.000	04:45
Diagnostics Check		OK	04:45

Data Download Date/Time: 03/01/2018 16:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: MOLINA BONILLA

DATE/TIME OF TEST: 02/01/2018 00:15:56

Diagnostics Check		OK	00:06	
Air Blank		0.000	00:06	
Control Test		0.081	00:06	
Air Blank		0.000	00:07	
Subject Sample	# 1	0.104	00:10	Volume Not Met
Breath Volume		0.214	00:10	
Air Blank		0.000	00:11	
Air Blank		0.000	00:13	
Subject Sample	# 2		00:14	Subject Test Refused
Breath Volume		0.085	00:14	
Air Blank		0.000	00:15	
Control Test		0.080	00:15	
Air Blank		0.000	00:15	
Diagnostics Check		OK	00:15	

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SOMMERS

DATE/TIME OF TEST: 02/01/2018 01:39:56

Diagnostics Check		OK	01:37	
Air Blank		0.000	01:37	
Control Test		0.081	01:37	
Air Blank		0.000	01:38	
Subject Sample	# 1		01:38	Subject Test Refused
Breath Volume		0.000	01:38	
Air Blank		0.000	01:39	
Control Test		0.082	01:39	
Air Blank		0.000	01:39	
Diagnostics Check		OK	01:39	

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: VISZMEG

DATE/TIME OF TEST: 02/02/2018 01:24:06

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:21	
Control Test		0.081	01:22	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:22	Subject Test Refused
Breath Volume		0.000	01:22	
Air Blank		0.000	01:23	
Control Test		0.081	01:23	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,STEVEN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMPSON,STEVEN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: BONNETT

DATE/TIME OF TEST: 02/02/2018 22:55:29

Diagnostics Check		OK	22:48
Air Blank		0.000	22:49
Control Test		0.081	22:49
Air Blank		0.000	22:50
Subject Sample	# 1	0.119	22:50
Breath Volume		1.257	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.119	22:53
Breath Volume		1.539	22:53
Air Blank		0.000	22:54
Control Test		0.080	22:54
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORES,ALAN,P
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: BORES,ALAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 02/03/2018 04:28:01

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.081	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.178	04:23
Breath Volume		4.347	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:25
Subject Sample	# 2	0.171	04:26
Breath Volume		2.949	04:26
Air Blank		0.000	04:27
Control Test		0.081	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:28

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: RODRIGUEZ EUTIMIO

DATE/TIME OF TEST: 02/03/2018 23:08:23

Diagnostics Check		OK	23:03	
Air Blank		0.000	23:04	
Control Test		0.081	23:04	
Air Blank		0.000	23:05	
Subject Sample	# 1		23:07	Subject Test Refused
Breath Volume		0.000	23:07	
Air Blank		0.000	23:07	
Control Test		0.081	23:07	
Air Blank		0.000	23:08	
Diagnostics Check		OK	23:08	

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: RODRIGUEZ EUTIMIO

DATE/TIME OF TEST: 02/03/2018 23:44:20

Diagnostics Check		OK	23:37
Air Blank		0.000	23:37
Control Test		0.081	23:38
Air Blank		0.000	23:38
Subject Sample	# 1	0.195	23:39
Breath Volume		1.625	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.199	23:42
Breath Volume		1.824	23:42
Air Blank		0.000	23:43
Control Test		0.080	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MERRILL,JACOB
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 02/05/2018 02:26:25

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:15	
Control Test		0.081	02:15	
Air Blank		0.000	02:15	
Subject Sample	# 1	0.187	02:19	Volume Not Met
Breath Volume		0.187	02:19	
Air Blank		0.000	02:20	
Air Blank		0.000	02:21	
Subject Sample	# 2	0.187	02:24	Volume Not Met
Breath Volume		0.285	02:24	
Air Blank		0.000	02:25	
Control Test		0.078	02:25	
Air Blank		0.000	02:26	
Diagnostics Check		OK	02:26	

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MERRILL,JACOB
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 02/05/2018 02:35:52

Diagnostics Check		OK	02:27	
Air Blank		0.000	02:27	
Control Test		0.078	02:28	
Air Blank		0.000	02:28	
Subject Sample	# 1	0.184	02:31	Volume Not Met
Breath Volume		0.441	02:31	
Air Blank		0.000	02:32	
Air Blank		0.000	02:34	
Subject Sample	# 2		02:34	Subject Test Refused
Breath Volume		0.000	02:34	
Air Blank		0.000	02:34	
Control Test		0.079	02:35	
Air Blank		0.000	02:35	
Diagnostics Check		OK	02:35	

Data Download Date/Time: 02/26/2018 11:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 16:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 02/05/2018 17:35:49

Diagnostics Check		OK	17:28
Air Blank		0.000	17:29
Control Test		0.080	17:29
Air Blank		0.000	17:30
Subject Sample	# 1	0.000	17:31
Breath Volume		1.855	17:31
Air Blank		0.000	17:31
Air Blank		0.000	17:33
Subject Sample	# 2	0.000	17:34
Breath Volume		2.871	17:34
Air Blank		0.000	17:34
Control Test		0.078	17:35
Air Blank		0.000	17:35
Diagnostics Check		OK	17:35

Data Download Date/Time: 02/26/2018 11:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: 2130,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: COSSEY

DATE/TIME OF TEST: 02/05/2018 22:39:11

Diagnostics Check		OK	22:26	
Air Blank		0.000	22:26	
Control Test		0.081	22:27	
Air Blank		0.000	22:27	
Subject Sample	# 1	0.158	22:28	
Breath Volume		1.601	22:28	
Air Blank		0.000	22:29	
Air Blank		0.000	22:31	
Subject Sample	# 2	0.171	22:34	Volume Not Met
Breath Volume		0.511	22:34	
Air Blank		0.000	22:35	
Air Blank		0.000	22:37	
Subject Sample	# 3	0.160	22:38	
Breath Volume		1.964	22:38	
Air Blank			22:38	Improper Sample
Air Blank			22:39	Improper Sample

Data Download Date/Time: 02/26/2018 11:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: ALARCON

DATE/TIME OF TEST: 02/05/2018 22:52:50

Diagnostics Check	OK	22:51	
Air Blank		22:52	Ambient Fail
Air Blank	0.000	22:52	

Data Download Date/Time: 02/26/2018 11:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: ALARCON

DATE/TIME OF TEST: 02/05/2018 22:58:50

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:56	
Control Test		0.079	22:56	
Air Blank		0.000	22:57	
Subject Sample	# 1		22:57	Subject Test Refused
Breath Volume		0.000	22:57	
Air Blank		0.000	22:57	
Control Test		0.079	22:58	
Air Blank		0.000	22:58	
Diagnostics Check		OK	22:58	

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: BUTTON

DATE/TIME OF TEST: 02/07/2018 01:15:42

Diagnostics Check		OK	01:02	
Air Blank		0.000	01:03	
Control Test		0.081	01:03	
Air Blank		0.000	01:03	
Subject Sample	# 1	0.167	01:07	Volume Not Met
Breath Volume		0.617	01:07	
Air Blank		0.000	01:08	
Air Blank		0.000	01:09	
Subject Sample	# 2	0.162	01:10	
Breath Volume		1.972	01:10	
Air Blank		0.000	01:11	
Air Blank		0.000	01:13	
Subject Sample	# 3	0.162	01:14	
Breath Volume		1.515	01:14	
Air Blank		0.000	01:14	
Control Test		0.076	01:15	
Air Blank		0.000	01:15	
Diagnostics Check		OK	01:15	

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 02/10/2018 00:59:08

Diagnostics Check		OK	00:56	
Air Blank		0.000	00:56	
Control Test		0.080	00:56	
Air Blank		0.000	00:57	
Subject Sample	# 1		00:57	Subject Test Refused
Breath Volume		0.000	00:57	
Air Blank		0.000	00:58	
Control Test		0.081	00:58	
Air Blank		0.000	00:59	
Diagnostics Check		OK	00:59	

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/10/2018 18:47:38

Diagnostics Check		OK	18:44	
Air Blank		0.000	18:45	
Control Test		0.081	18:45	
Air Blank		0.000	18:46	
Subject Sample	# 1		18:46	Subject Test Refused
Breath Volume		0.000	18:46	
Air Blank		0.000	18:46	
Control Test		0.081	18:47	
Air Blank		0.000	18:47	
Diagnostics Check		OK	18:47	

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: ST JOHN

DATE/TIME OF TEST: 02/10/2018 22:31:10

Diagnostics Check		OK	22:24
Air Blank		0.000	22:25
Control Test		0.080	22:25
Air Blank		0.000	22:26
Subject Sample	# 1	0.235	22:26
Breath Volume		2.468	22:26
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.225	22:29
Breath Volume		2.421	22:29
Air Blank		0.000	22:30
Control Test		0.079	22:30
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: MAYE

DATE/TIME OF TEST: 02/11/2018 00:18:18

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:15	
Control Test		0.081	00:16	
Air Blank		0.000	00:16	
Subject Sample	# 1		00:16	Subject Test Refused
Breath Volume		0.000	00:16	
Air Blank		0.000	00:17	
Control Test		0.081	00:17	
Air Blank		0.000	00:18	
Diagnostics Check		OK	00:18	

Data Download Date/Time: 02/26/2018 11:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: TERZO

DATE/TIME OF TEST: 02/11/2018 02:24:48

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.080	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.205	02:20
Breath Volume		1.519	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:22
Subject Sample	# 2	0.202	02:23
Breath Volume		1.312	02:23
Air Blank		0.000	02:23
Control Test		0.078	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: CALIXTRO

DATE/TIME OF TEST: 02/11/2018 22:01:06

Diagnostics Check		OK	21:58	
Air Blank		0.000	21:58	
Control Test		0.080	21:59	
Air Blank		0.000	21:59	
Subject Sample	# 1		21:59	Subject Test Refused
Breath Volume		0.000	21:59	
Air Blank		0.000	22:00	
Control Test		0.081	22:00	
Air Blank		0.000	22:01	
Diagnostics Check		OK	22:01	

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ
BURROUGHS,DOMINIC,J
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MAROSE,CHARLES
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 02/17/2018 07:30:45

Diagnostics Check		OK	07:24
Air Blank		0.000	07:24
Control Test		0.078	07:24
Air Blank		0.000	07:25
Subject Sample	# 1	0.125	07:25
Breath Volume		1.160	07:25
Air Blank		0.000	07:26
Air Blank		0.000	07:28
Subject Sample	# 2	0.137	07:29
Breath Volume		3.125	07:29
Air Blank		0.000	07:29
Control Test		0.076	07:30
Air Blank		0.000	07:30
Diagnostics Check		OK	07:30

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ROBLERO PEREZ

DATE/TIME OF TEST: 02/18/2018 02:15:47

Diagnostics Check		OK	02:09
Air Blank		0.000	02:10
Control Test		0.081	02:10
Air Blank		0.000	02:11
Subject Sample	# 1	0.194	02:11
Breath Volume		1.945	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:13
Subject Sample	# 2	0.201	02:14
Breath Volume		1.945	02:14
Air Blank		0.000	02:14
Control Test		0.080	02:15
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,STEVEN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMPSON,STEVEN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 02/19/2018 00:57:49

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.081	00:52
Air Blank		0.000	00:52
Subject Sample	# 1	0.237	00:53
Breath Volume		1.386	00:53
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.230	00:56
Breath Volume		2.039	00:56
Air Blank		0.000	00:56
Control Test		0.078	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: VERDUGO DIAZ

DATE/TIME OF TEST: 02/19/2018 01:51:20

Diagnostics Check		OK	01:42	
Air Blank		0.000	01:43	
Control Test		0.081	01:43	
Air Blank		0.000	01:44	
Subject Sample	# 1		01:47	No Sample Provided
Breath Volume		0.000	01:47	
Air Blank		0.000	01:47	
Air Blank		0.000	01:49	
Subject Sample	# 2		01:49	Subject Test Refused
Breath Volume		0.000	01:49	
Air Blank		0.000	01:50	
Control Test		0.081	01:50	
Air Blank		0.000	01:51	
Diagnostics Check		OK	01:51	

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPURLOCK,ROBERT,V
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPURLOCK,ROBEERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: K300432740840

DATE/TIME OF TEST: 02/20/2018 21:44:38

Diagnostics Check		OK	21:38
Air Blank		0.000	21:38
Control Test		0.080	21:38
Air Blank		0.000	21:39
Subject Sample	# 1	0.140	21:40
Breath Volume		2.535	21:40
Air Blank		0.000	21:40
Air Blank		0.000	21:42
Subject Sample	# 2	0.133	21:43
Breath Volume		3.101	21:43
Air Blank		0.000	21:43
Control Test		0.079	21:44
Air Blank		0.000	21:44
Diagnostics Check		OK	21:44

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: WOODWARD

DATE/TIME OF TEST: 02/21/2018 23:53:44

Diagnostics Check		OK	23:47
Air Blank		0.000	23:47
Control Test		0.081	23:48
Air Blank		0.000	23:48
Subject Sample	# 1	0.190	23:49
Breath Volume		2.507	23:49
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.181	23:52
Breath Volume		1.761	23:52
Air Blank		0.000	23:52
Control Test		0.080	23:53
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,STEVEN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMPSON,STEVEN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: FALLON

DATE/TIME OF TEST: 02/22/2018 01:46:10

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.080	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.224	01:41
Breath Volume		1.285	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.229	01:44
Breath Volume		1.414	01:44
Air Blank		0.000	01:45
Control Test		0.080	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,DANIEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WATSON,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 11:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: STROUP

DATE/TIME OF TEST: 02/24/2018 00:27:38

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:25	
Control Test		0.081	00:25	
Air Blank		0.000	00:26	
Subject Sample	# 1		00:26	Subject Test Refused
Breath Volume		0.000	00:26	
Air Blank		0.000	00:26	
Control Test		0.081	00:27	
Air Blank		0.000	00:27	
Diagnostics Check		OK	00:27	

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SOUTH

DATE/TIME OF TEST: 02/24/2018 01:33:01

Diagnostics Check		OK	01:30	
Air Blank		0.000	01:30	
Control Test		0.081	01:30	
Air Blank		0.000	01:31	
Subject Sample	# 1		01:31	Subject Test Refused
Breath Volume		0.000	01:31	
Air Blank		0.000	01:32	
Control Test		0.082	01:32	
Air Blank		0.000	01:32	
Diagnostics Check		OK	01:33	

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: YAGUA GARCIA

DATE/TIME OF TEST: 02/24/2018 02:23:07

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.080	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.162	02:18
Breath Volume		2.585	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.162	02:21
Breath Volume		2.878	02:21
Air Blank		0.000	02:22
Control Test		0.080	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: SCHROEDER

DATE/TIME OF TEST: 02/24/2018 19:59:37

Diagnostics Check		OK	19:53
Air Blank		0.000	19:53
Control Test		0.081	19:53
Air Blank		0.000	19:54
Subject Sample	# 1	0.236	19:54
Breath Volume		2.707	19:54
Air Blank		0.000	19:55
Air Blank		0.000	19:57
Subject Sample	# 2	0.244	19:58
Breath Volume		2.757	19:58
Air Blank		0.000	19:58
Control Test		0.080	19:59
Air Blank		0.000	19:59
Diagnostics Check		OK	19:59

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: CRUZ MURILLO

DATE/TIME OF TEST: 02/24/2018 22:02:24

Diagnostics Check		OK	21:57	
Air Blank		0.000	21:58	
Control Test		0.080	21:58	
Air Blank		0.000	21:58	
Subject Sample	# 1		22:01	Subject Test Refused
Breath Volume		0.000	22:01	
Air Blank		0.000	22:01	
Control Test		0.081	22:01	
Air Blank		0.000	22:02	
Diagnostics Check		OK	22:02	

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: TAMBURINO

DATE/TIME OF TEST: 02/24/2018 23:27:03

Diagnostics Check		OK	23:24	
Air Blank		0.000	23:24	
Control Test		0.081	23:24	
Air Blank		0.000	23:25	
Subject Sample	# 1		23:25	Subject Test Refused
Breath Volume		0.000	23:25	
Air Blank		0.000	23:26	
Control Test		0.081	23:26	
Air Blank		0.000	23:26	
Diagnostics Check		OK	23:27	

Data Download Date/Time: 02/26/2018 11:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: VERVILLE

DATE/TIME OF TEST: 02/25/2018 23:17:05

Diagnostics Check		OK	23:10
Air Blank		0.000	23:10
Control Test		0.081	23:11
Air Blank		0.000	23:11
Subject Sample	# 1	0.192	23:12
Breath Volume		1.968	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:14
Subject Sample	# 2	0.184	23:15
Breath Volume		1.808	23:15
Air Blank		0.000	23:16
Control Test		0.079	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 02/26/2018 02:16:44

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:13	
Control Test		0.080	02:14	
Air Blank		0.000	02:14	
Subject Sample	# 1		02:15	Slope Not Met
Breath Volume		0.945	02:15	
Air Blank		0.000	02:15	
Control Test		0.080	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 02/26/2018 02:25:59

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.080	02:19
Air Blank		0.000	02:20
Subject Sample	# 1	0.255	02:21
Breath Volume		1.285	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.249	02:24
Breath Volume		1.820	02:24
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 02/26/2018 11:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: DONAHUE

DATE/TIME OF TEST: 02/28/2018 23:05:41

Diagnostics Check		OK	23:02	
Air Blank		0.000	23:03	
Control Test		0.080	23:03	
Air Blank		0.000	23:04	
Subject Sample	# 1		23:04	Subject Test Refused
Breath Volume		0.000	23:04	
Air Blank		0.000	23:04	
Control Test		0.081	23:05	
Air Blank		0.000	23:05	
Diagnostics Check		OK	23:05	

Data Download Date/Time: 03/26/2018 11:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 123456789
DRY GAS EXPIRY DATE: 11/11/2018
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: VALENTINE

DATE/TIME OF TEST: 02/15/2018 21:33:09

Diagnostics Check		OK	21:25
Air Blank		0.000	21:25
Control Test		0.082	21:25
Air Blank		0.000	21:26
Subject Sample	# 1	0.136	21:26
Breath Volume		1.218	21:26
Air Blank		0.000	21:27
Air Blank		0.000	21:29
Subject Sample	# 2	0.136	21:31
Breath Volume		1.238	21:31
Air Blank		0.000	21:32
Control Test		0.080	21:32
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 02/26/2018 09:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: VARNADORE

DATE/TIME OF TEST: 02/01/2018 01:08:41

Diagnostics Check		OK	01:01
Air Blank		0.000	01:02
Control Test		0.080	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.304	01:03
Breath Volume		1.718	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:05
Subject Sample	# 2	0.312	01:06
Breath Volume		1.785	01:06
Air Blank		0.000	01:07
Control Test		0.081	01:08
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 02/15/2018 21:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BRIGGS,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: BENDER

DATE/TIME OF TEST: 02/05/2018 21:33:01

Diagnostics Check		OK	21:25
Air Blank		0.000	21:26
Control Test		0.080	21:26
Air Blank		0.000	21:27
Subject Sample	# 1	0.240	21:27
Breath Volume		3.386	21:27
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.226	21:31
Breath Volume		3.875	21:31
Air Blank		0.000	21:32
Control Test		0.079	21:32
Air Blank		0.000	21:32
Diagnostics Check		OK	21:33

Data Download Date/Time: 02/15/2018 21:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 02/10/2018 00:46:35

Diagnostics Check		OK	00:40
Air Blank		0.000	00:41
Control Test		0.079	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.000	00:42
Breath Volume		4.093	00:42
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.000	00:45
Breath Volume		2.757	00:45
Air Blank		0.000	00:45
Control Test		0.078	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 02/15/2018 21:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: GUYER

DATE/TIME OF TEST: 02/11/2018 19:29:49

Diagnostics Check		OK	19:23
Air Blank		0.000	19:23
Control Test		0.079	19:24
Air Blank		0.000	19:24
Subject Sample	# 1	0.153	19:25
Breath Volume		1.664	19:25
Air Blank		0.000	19:25
Air Blank		0.000	19:27
Subject Sample	# 2	0.161	19:28
Breath Volume		3.207	19:28
Air Blank		0.000	19:28
Control Test		0.080	19:29
Air Blank		0.000	19:29
Diagnostics Check		OK	19:29

Data Download Date/Time: 02/15/2018 21:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MINTZ

DATE/TIME OF TEST: 02/15/2018 00:18:10

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.079	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.188	00:13
Breath Volume		2.460	00:13
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.178	00:16
Breath Volume		2.269	00:16
Air Blank		0.000	00:17
Control Test		0.079	00:17
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 02/15/2018 21:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONIE,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: NEGRONI PEREZ

DATE/TIME OF TEST: 02/24/2018 00:04:52

Diagnostics Check		OK	23:57
Air Blank		0.000	23:58
Control Test		0.081	23:58
Air Blank		0.000	23:59
Subject Sample	# 2	0.191	23:59
Breath Volume		1.820	23:59
Air Blank		0.000	00:00
Air Blank		0.000	00:02
Subject Sample	# 1	0.189	00:03
Breath Volume		1.542	00:03
Air Blank		0.000	00:03
Control Test		0.078	00:04
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 02/28/2018 10:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SCHUKAY

DATE/TIME OF TEST: 02/24/2018 01:31:59

Diagnostics Check		OK	01:25
Air Blank		0.000	01:26
Control Test		0.080	01:26
Air Blank		0.000	01:27
Subject Sample	# 1	0.007	01:27
Breath Volume		3.234	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.000	01:30
Breath Volume		3.335	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 02/28/2018 10:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MANGOLD,K
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 02/24/2018 22:34:02

Diagnostics Check		OK	22:25
Air Blank		0.000	22:25
Control Test		0.080	22:26
Air Blank		0.000	22:26
Subject Sample	# 1	0.251	22:28
Breath Volume		2.066	22:28
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.243	22:32
Breath Volume		2.140	22:32
Air Blank		0.000	22:33
Control Test		0.076	22:33
Air Blank		0.000	22:33
Diagnostics Check		OK	22:34

Data Download Date/Time: 02/28/2018 10:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: RODRIGUEZ MORALES

DATE/TIME OF TEST: 02/01/2018 22:59:00

Diagnostics Check		OK	22:52
Air Blank		0.000	22:53
Control Test		0.080	22:53
Air Blank		0.000	22:54
Subject Sample	# 1	0.122	22:54
Breath Volume		1.386	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:56
Subject Sample	# 2	0.117	22:57
Breath Volume		1.417	22:57
Air Blank		0.000	22:58
Control Test		0.082	22:58
Air Blank		0.000	22:58
Diagnostics Check		OK	22:59

Data Download Date/Time: 02/15/2018 21:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 02/03/2018 01:12:00

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.082	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.110	01:07
Breath Volume		1.355	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:09
Subject Sample	# 2	0.123	01:10
Breath Volume		1.535	01:10
Air Blank		0.000	01:11
Control Test		0.082	01:11
Air Blank		0.000	01:11
Diagnostics Check		OK	01:12

Data Download Date/Time: 02/15/2018 21:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 02/09/2018 00:49:06

Diagnostics Check		OK	00:42
Air Blank		0.000	00:43
Control Test		0.080	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.128	00:44
Breath Volume		1.605	00:44
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.120	00:47
Breath Volume		1.828	00:47
Air Blank		0.000	00:48
Control Test		0.081	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 02/15/2018 21:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: MOUGHTON

DATE/TIME OF TEST: 02/10/2018 01:38:40

Diagnostics Check		OK	01:32
Air Blank		0.000	01:33
Control Test		0.081	01:33
Air Blank		0.000	01:33
Subject Sample	# 1	0.116	01:34
Breath Volume		2.003	01:34
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.116	01:37
Breath Volume		2.781	01:37
Air Blank		0.000	01:37
Control Test		0.081	01:38
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 02/15/2018 21:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: PRITCHARD

DATE/TIME OF TEST: 02/10/2018 03:08:16

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.081	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.000	03:04
Breath Volume		3.582	03:04
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.000	03:06
Breath Volume		3.781	03:06
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 02/15/2018 21:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 02/11/2018 00:33:03

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.081	00:27
Air Blank		0.000	00:27
Subject Sample	# 1	0.143	00:28
Breath Volume		1.976	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.145	00:31
Breath Volume		1.843	00:31
Air Blank		0.000	00:32
Control Test		0.081	00:32
Air Blank		0.000	00:32
Diagnostics Check		OK	00:33

Data Download Date/Time: 02/15/2018 21:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 02/21/2018 00:37:27

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.081	00:31
Air Blank		0.000	00:32
Subject Sample	# 1	0.273	00:32
Breath Volume		1.300	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.281	00:35
Breath Volume		1.652	00:35
Air Blank		0.000	00:36
Control Test		0.079	00:36
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 03/05/2018 21:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 02/03/2018 00:51:02

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.121	00:46
Breath Volume		3.496	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:48
Subject Sample	# 2	0.134	00:49
Breath Volume		4.531	00:49
Air Blank		0.000	00:50
Control Test		0.078	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:51

Data Download Date/Time: 03/02/2018 09:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: OLDHAM

DATE/TIME OF TEST: 02/03/2018 01:34:52

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.079	01:29
Air Blank		0.000	01:29
Subject Sample	# 1	0.298	01:30
Breath Volume		2.347	01:30
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.306	01:33
Breath Volume		2.554	01:33
Air Blank		0.000	01:33
Control Test		0.079	01:34
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 03/02/2018 09:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GUEVARA,ERICK
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CAMPOS-BRENES

DATE/TIME OF TEST: 02/03/2018 01:53:14

Diagnostics Check		OK	01:50	
Air Blank		0.000	01:50	
Control Test		0.079	01:51	
Air Blank		0.000	01:51	
Subject Sample	# 1		01:51	Subject Test Refused
Breath Volume		0.000	01:51	
Air Blank		0.000	01:52	
Control Test		0.078	01:52	
Air Blank		0.000	01:53	
Diagnostics Check		OK	01:53	

Data Download Date/Time: 03/02/2018 09:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PHILPOT,CEDRIC
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: GIVANS

DATE/TIME OF TEST: 02/03/2018 02:17:50

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.079	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.154	02:13
Breath Volume		2.253	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:15
Subject Sample	# 2	0.151	02:16
Breath Volume		1.425	02:16
Air Blank		0.000	02:16
Control Test		0.078	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 03/02/2018 09:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,CARLOS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: ARCE

DATE/TIME OF TEST: 02/03/2018 02:27:53

Diagnostics Check		OK	02:21
Air Blank		0.000	02:22
Control Test		0.079	02:22
Air Blank		0.000	02:22
Subject Sample	# 1	0.130	02:23
Breath Volume		1.886	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:25
Subject Sample	# 2	0.135	02:26
Breath Volume		1.800	02:26
Air Blank		0.000	02:26
Control Test		0.079	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 03/02/2018 09:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ACOSTA,ALBERTO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: OYARZUN

DATE/TIME OF TEST: 02/08/2018 21:39:36

Diagnostics Check		OK	21:25	
Air Blank		0.000	21:25	
Control Test		0.080	21:26	
Air Blank		0.000	21:26	
Subject Sample	# 1	0.160	21:29	Volume Not Met
Breath Volume		0.015	21:29	
Air Blank		0.000	21:30	
Air Blank		0.000	21:32	
Subject Sample	# 2	0.174	21:34	
Breath Volume		1.269	21:34	
Air Blank		0.000	21:35	
Air Blank		0.000	21:36	
Subject Sample	# 3	0.164	21:37	
Breath Volume		2.109	21:37	
Air Blank		0.000	21:38	
Control Test		0.078	21:38	
Air Blank		0.000	21:39	
Diagnostics Check		OK	21:39	

Data Download Date/Time: 03/02/2018 09:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RUTLEDGE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MALLEA

DATE/TIME OF TEST: 02/09/2018 00:48:07

Diagnostics Check		OK	00:41
Air Blank		0.000	00:42
Control Test		0.080	00:42
Air Blank		0.000	00:42
Subject Sample	# 1	0.000	00:43
Breath Volume		2.167	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.000	00:46
Breath Volume		1.699	00:46
Air Blank		0.000	00:47
Control Test		0.079	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 03/02/2018 09:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GARCIA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BARREAU

DATE/TIME OF TEST: 02/09/2018 01:00:24

Diagnostics Check		OK	00:52
Air Blank		0.000	00:53
Control Test		0.079	00:53
Air Blank		0.000	00:54
Subject Sample	# 1	0.053	00:54
Breath Volume		1.203	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.059	00:58
Breath Volume		2.632	00:58
Air Blank		0.000	00:59
Control Test		0.081	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 03/02/2018 09:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PHILPOT,CEDRIC
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/23/2018 01:19:15

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.078	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.110	01:14
Breath Volume		2.703	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.115	01:17
Breath Volume		2.835	01:17
Air Blank		0.000	01:18
Control Test		0.079	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 03/02/2018 09:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GUEVARA,ERICK
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 02/25/2018 00:24:41

Diagnostics Check		OK	00:21	
Air Blank		0.000	00:22	
Control Test		0.079	00:22	
Air Blank		0.000	00:23	
Subject Sample	# 1		00:23	Subject Test Refused
Breath Volume		0.000	00:23	
Air Blank		0.000	00:23	
Control Test		0.079	00:24	
Air Blank		0.000	00:24	
Diagnostics Check		OK	00:24	

Data Download Date/Time: 03/02/2018 09:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: POSADAS MARTINEZ

DATE/TIME OF TEST: 02/25/2018 01:08:46

Diagnostics Check		OK	01:05	
Air Blank		0.000	01:06	
Control Test		0.080	01:06	
Air Blank		0.000	01:07	
Subject Sample	# 1		01:07	Subject Test Refused
Breath Volume		0.000	01:07	
Air Blank		0.000	01:07	
Control Test		0.079	01:08	
Air Blank		0.000	01:08	
Diagnostics Check		OK	01:08	

Data Download Date/Time: 03/02/2018 09:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RUTLEDGE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SHEPHERD

DATE/TIME OF TEST: 02/25/2018 01:33:55

Diagnostics Check		OK	01:27
Air Blank		0.000	01:27
Control Test		0.079	01:28
Air Blank		0.000	01:28
Subject Sample	# 1	0.202	01:29
Breath Volume		1.324	01:29
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.198	01:32
Breath Volume		2.292	01:32
Air Blank		0.000	01:32
Control Test		0.077	01:33
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 03/02/2018 09:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VALDEZ,MICHAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 02/25/2018 01:50:10

Diagnostics Check		OK	01:43
Air Blank		0.000	01:44
Control Test		0.079	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.000	01:45
Breath Volume		1.359	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.000	01:48
Breath Volume		1.746	01:48
Air Blank		0.000	01:49
Control Test		0.079	01:49
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 03/02/2018 09:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PHILPOT,CEDRIC
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: ESPINOSA CEPERO

DATE/TIME OF TEST: 02/25/2018 02:08:58

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.077	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.232	02:02
Breath Volume		1.246	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.234	02:07
Breath Volume		2.250	02:07
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 03/02/2018 09:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MARTOS,JULIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: DEL RIO

DATE/TIME OF TEST: 02/25/2018 02:26:11

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.078	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.186	02:20
Breath Volume		1.378	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.175	02:24
Breath Volume		1.519	02:24
Air Blank		0.000	02:25
Control Test		0.077	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 03/02/2018 09:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 02/25/2018 02:36:58

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:34	
Control Test		0.079	02:34	
Air Blank		0.000	02:35	
Subject Sample	# 1		02:35	Subject Test Refused
Breath Volume		0.000	02:35	
Air Blank		0.000	02:36	
Control Test		0.078	02:36	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:36	

Data Download Date/Time: 03/02/2018 09:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HARRISON,JORDAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: JEANCHARLES

DATE/TIME OF TEST: 02/08/2018 02:11:33

Diagnostics Check		OK	01:59	
Air Blank		0.000	01:59	
Control Test		0.081	02:00	
Air Blank		0.000	02:00	
Subject Sample	# 1	0.125	02:03	Volume Not Met
Breath Volume		0.050	02:03	
Air Blank		0.000	02:04	
Air Blank		0.000	02:06	
Subject Sample	# 2		02:09	No Sample Provided
Breath Volume		0.000	02:09	
Air Blank		0.000	02:10	
Control Test		0.083	02:10	
Air Blank		0.000	02:11	
Diagnostics Check		OK	02:11	

Data Download Date/Time: 03/01/2018 12:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: DURAN

DATE/TIME OF TEST: 02/08/2018 03:34:31

Diagnostics Check		OK	03:27
Air Blank		0.000	03:28
Control Test		0.081	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.151	03:29
Breath Volume		1.679	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:31
Subject Sample	# 2	0.159	03:32
Breath Volume		1.277	03:32
Air Blank		0.000	03:33
Control Test		0.081	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TONNA,GREG
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:53

SUBJECT'S LAST NAME: MONTEAGUDO

DATE/TIME OF TEST: 02/08/2018 10:28:22

Diagnostics Check		OK	10:14	
Air Blank		0.000	10:14	
Control Test		0.080	10:15	
Air Blank		0.000	10:15	
Subject Sample	# 1	0.000	10:17	
Breath Volume		1.468	10:17	
Air Blank		0.000	10:18	
Air Blank		0.000	10:20	
Subject Sample	# 2	0.000	10:23	Volume Not Met
Breath Volume		0.246	10:23	
Air Blank		0.000	10:23	
Air Blank		0.000	10:25	
Subject Sample	# 3	0.000	10:26	
Breath Volume		1.160	10:26	
Air Blank		0.000	10:27	
Control Test		0.080	10:27	
Air Blank		0.000	10:28	
Diagnostics Check		OK	10:28	

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,YVETTE,D
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEMP,MARC
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: KEYES

DATE/TIME OF TEST: 02/08/2018 19:09:51

Diagnostics Check		OK	19:00
Air Blank		0.000	19:00
Control Test		0.081	19:01
Air Blank		0.000	19:01
Subject Sample	# 1	0.179	19:03
Breath Volume		1.175	19:03
Air Blank		0.000	19:04
Air Blank		0.000	19:05
Subject Sample	# 2	0.186	19:08
Breath Volume		1.531	19:08
Air Blank		0.000	19:08
Control Test		0.080	19:09
Air Blank		0.000	19:09
Diagnostics Check		OK	19:09

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MENDOZA,THOMAS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PYRAM

DATE/TIME OF TEST: 02/09/2018 01:33:38

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.082	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.170	01:28
Breath Volume		1.382	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:30
Subject Sample	# 2	0.172	01:31
Breath Volume		1.191	01:31
Air Blank		0.000	01:32
Control Test		0.080	01:33
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: DEANS

DATE/TIME OF TEST: 02/09/2018 02:22:32

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.081	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.155	02:18
Breath Volume		1.359	02:18
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.157	02:20
Breath Volume		1.464	02:20
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: REYNOLDS,JOSHUA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 02/09/2018 02:57:02

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.080	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.172	02:52
Breath Volume		2.292	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.190	02:55
Breath Volume		2.781	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:57

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DALTON,RYAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: GARCIA NARANJO

DATE/TIME OF TEST: 02/09/2018 03:45:32

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.080	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.122	03:40
Breath Volume		1.125	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.123	03:43
Breath Volume		1.144	03:43
Air Blank		0.000	03:44
Control Test		0.077	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GONZALEZ,CHRISTIAN
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ANYIKWA

DATE/TIME OF TEST: 02/09/2018 04:37:31

Diagnostics Check		OK	04:25	
Air Blank		0.000	04:26	
Control Test		0.080	04:26	
Air Blank		0.000	04:27	
Subject Sample	# 1	0.199	04:30	Volume Not Met
Breath Volume		0.203	04:30	
Air Blank		0.000	04:31	
Air Blank		0.000	04:32	
Subject Sample	# 2	0.178	04:35	Volume Not Met
Breath Volume		0.058	04:35	
Air Blank		0.000	04:36	
Control Test		0.078	04:36	
Air Blank		0.000	04:37	
Diagnostics Check		OK	04:37	

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GONZALEZ,CHRISTIAN
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ANYIKWA

DATE/TIME OF TEST: 02/09/2018 04:47:30

Diagnostics Check		OK	04:39
Air Blank		0.000	04:40
Control Test		0.078	04:40
Air Blank		0.000	04:41
Subject Sample	# 1	0.197	04:42
Breath Volume		1.445	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:44
Subject Sample	# 2	0.191	04:45
Breath Volume		1.656	04:45
Air Blank		0.000	04:46
Control Test		0.079	04:46
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SEPOT,MICHAEL
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: BOCHAN

DATE/TIME OF TEST: 02/09/2018 17:22:47

Diagnostics Check		OK	17:15
Air Blank		0.000	17:15
Control Test		0.080	17:16
Air Blank		0.000	17:16
Subject Sample	# 1	0.176	17:18
Breath Volume		2.386	17:18
Air Blank		0.000	17:18
Air Blank		0.000	17:20
Subject Sample	# 2	0.175	17:21
Breath Volume		1.734	17:21
Air Blank		0.000	17:21
Control Test		0.080	17:22
Air Blank		0.000	17:22
Diagnostics Check		OK	17:22

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: ORLANDO

DATE/TIME OF TEST: 02/09/2018 22:17:27

Diagnostics Check		OK	22:11
Air Blank		0.000	22:11
Control Test		0.080	22:11
Air Blank		0.000	22:12
Subject Sample	# 1	0.202	22:13
Breath Volume		2.257	22:13
Air Blank		0.000	22:13
Air Blank		0.000	22:15
Subject Sample	# 2	0.209	22:15
Breath Volume		1.968	22:15
Air Blank		0.000	22:16
Control Test		0.080	22:16
Air Blank		0.000	22:17
Diagnostics Check		OK	22:17

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: MOBLEY

DATE/TIME OF TEST: 02/10/2018 03:55:19

Diagnostics Check		OK	03:48
Air Blank		0.000	03:49
Control Test		0.081	03:49
Air Blank		0.000	03:49
Subject Sample	# 1	0.133	03:50
Breath Volume		1.648	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.133	03:53
Breath Volume		1.417	03:53
Air Blank		0.000	03:54
Control Test		0.080	03:54
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: FEDER

DATE/TIME OF TEST: 02/10/2018 07:07:08

Diagnostics Check		OK	07:00
Air Blank		0.000	07:01
Control Test		0.081	07:01
Air Blank		0.000	07:02
Subject Sample	# 1	0.172	07:02
Breath Volume		3.144	07:02
Air Blank		0.000	07:03
Air Blank		0.000	07:05
Subject Sample	# 2	0.172	07:05
Breath Volume		2.773	07:05
Air Blank		0.000	07:06
Control Test		0.082	07:06
Air Blank		0.000	07:07
Diagnostics Check		OK	07:07

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SCHEMTOB,ASHLEY
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: JEFF

DATE/TIME OF TEST: 02/10/2018 19:53:35

Diagnostics Check		OK	19:50	
Air Blank		0.000	19:50	
Control Test		0.081	19:51	
Air Blank		0.000	19:51	
Subject Sample	# 1		19:52	Subject Test Refused
Breath Volume		0.000	19:52	
Air Blank		0.000	19:52	
Control Test		0.081	19:52	
Air Blank		0.000	19:53	
Diagnostics Check		OK	19:53	

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: AKERS,DAVE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:11

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 02/11/2018 07:37:55

Diagnostics Check		OK	07:31
Air Blank		0.000	07:31
Control Test		0.081	07:32
Air Blank		0.000	07:32
Subject Sample	# 1	0.131	07:33
Breath Volume		1.335	07:33
Air Blank		0.000	07:34
Air Blank		0.000	07:35
Subject Sample	# 2	0.133	07:36
Breath Volume		1.492	07:36
Air Blank		0.000	07:36
Control Test		0.081	07:37
Air Blank		0.000	07:37
Diagnostics Check		OK	07:37

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: PENNINGTON

DATE/TIME OF TEST: 02/12/2018 01:47:27

Diagnostics Check		OK	01:35	
Air Blank		0.000	01:36	
Control Test		0.081	01:36	
Air Blank		0.000	01:37	
Subject Sample	# 1	0.120	01:40	Volume Not Met
Breath Volume		0.484	01:40	
Air Blank		0.000	01:40	
Air Blank		0.000	01:42	
Subject Sample	# 2	0.083	01:45	Volume Not Met
Breath Volume		0.078	01:45	
Air Blank		0.000	01:46	
Control Test		0.079	01:46	
Air Blank		0.000	01:47	
Diagnostics Check		OK	01:47	

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HARRISON,JORDAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ERBY

DATE/TIME OF TEST: 02/12/2018 02:48:05

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.081	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.063	02:42
Breath Volume		1.152	02:42
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.057	02:46
Breath Volume		1.347	02:46
Air Blank		0.000	02:47
Control Test		0.082	02:47
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LATA,CHRISTIAN
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/12/2018 04:15:17

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.082	04:08
Air Blank		0.000	04:09
Subject Sample	# 1	0.000	04:10
Breath Volume		1.312	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.000	04:13
Breath Volume		1.816	04:13
Air Blank		0.000	04:14
Control Test		0.081	04:14
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MCGILLICUDY,STEVEN
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: DRAGOI

DATE/TIME OF TEST: 02/12/2018 05:30:53

Diagnostics Check		OK	05:22
Air Blank		0.000	05:22
Control Test		0.081	05:23
Air Blank		0.000	05:23
Subject Sample	# 1	0.142	05:26
Breath Volume		1.171	05:26
Air Blank		0.000	05:26
Air Blank		0.000	05:28
Subject Sample	# 2	0.147	05:29
Breath Volume		1.406	05:29
Air Blank		0.000	05:29
Control Test		0.079	05:30
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BERMAN,MARC
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: CROCKETT

DATE/TIME OF TEST: 02/13/2018 23:37:02

Diagnostics Check		OK	23:30
Air Blank		0.000	23:31
Control Test		0.080	23:31
Air Blank		0.000	23:32
Subject Sample	# 1	0.115	23:32
Breath Volume		2.066	23:32
Air Blank		0.000	23:33
Air Blank		0.000	23:34
Subject Sample	# 2	0.119	23:35
Breath Volume		2.191	23:35
Air Blank		0.000	23:36
Control Test		0.081	23:36
Air Blank		0.000	23:36
Diagnostics Check		OK	23:37

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VALDES,ELIO
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 02/14/2018 02:43:39

Diagnostics Check		OK	02:32	
Air Blank		0.000	02:32	
Control Test		0.081	02:32	
Air Blank		0.000	02:33	
Subject Sample	# 1	0.000	02:36	Volume Not Met
Breath Volume		0.015	02:36	
Air Blank		0.000	02:37	
Air Blank		0.000	02:39	
Subject Sample	# 2		02:42	No Sample Provided
Breath Volume		0.000	02:42	
Air Blank		0.000	02:42	
Control Test		0.080	02:43	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: BILOTTI

DATE/TIME OF TEST: 02/16/2018 01:41:25

Diagnostics Check		OK	01:33
Air Blank		0.000	01:34
Control Test		0.082	01:34
Air Blank		0.000	01:35
Subject Sample	# 1	0.153	01:36
Breath Volume		1.460	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.154	01:39
Breath Volume		2.140	01:39
Air Blank		0.000	01:40
Control Test		0.081	01:40
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PEERANI

DATE/TIME OF TEST: 02/17/2018 04:17:03

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.081	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.083	04:12
Breath Volume		1.628	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.085	04:15
Breath Volume		2.000	04:15
Air Blank		0.000	04:16
Control Test		0.081	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:17

Data Download Date/Time: 03/01/2018 12:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JASON
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: GADOMSKI

DATE/TIME OF TEST: 02/17/2018 05:27:58

Diagnostics Check		OK	05:10	
Air Blank		0.000	05:11	
Control Test		0.082	05:11	
Air Blank		0.000	05:12	
Subject Sample	# 1	0.164	05:15	
Breath Volume		1.238	05:15	
Air Blank		0.000	05:15	
Air Blank		0.000	05:17	
Subject Sample	# 2	0.169	05:20	Volume Not Met
Breath Volume		0.542	05:20	
Air Blank		0.000	05:21	
Air Blank		0.000	05:23	
Subject Sample	# 3	0.158	05:26	Volume Not Met
Breath Volume		0.179	05:26	
Air Blank		0.000	05:27	
Control Test		0.078	05:27	
Air Blank		0.000	05:27	
Diagnostics Check		OK	05:27	

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JASON
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: GADOMSKI

DATE/TIME OF TEST: 02/17/2018 05:38:54

Diagnostics Check		OK	05:31
Air Blank		0.000	05:31
Control Test		0.080	05:32
Air Blank		0.000	05:32
Subject Sample	# 1	0.159	05:33
Breath Volume		1.390	05:33
Air Blank		0.000	05:33
Air Blank		0.000	05:35
Subject Sample	# 2	0.164	05:37
Breath Volume		1.765	05:37
Air Blank		0.000	05:37
Control Test		0.080	05:38
Air Blank		0.000	05:38
Diagnostics Check		OK	05:38

Data Download Date/Time: 03/01/2018 12:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: LYKENS,W
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: BERNARD

DATE/TIME OF TEST: 02/11/2018 02:58:31

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:52	
Control Test		0.078	02:52	
Air Blank		0.000	02:52	
Subject Sample	# 1	0.171	02:53	
Breath Volume		1.140	02:53	
Air Blank		0.000	02:54	
Air Blank		0.000	02:56	
Subject Sample	# 2		02:56	Interferent Detect
Breath Volume		1.472	02:56	
Air Blank		0.000	02:57	
Control Test		0.076	02:57	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 02/28/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: LYKENS,W
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: BERNARD

DATE/TIME OF TEST: 02/11/2018 03:06:58

Diagnostics Check		OK	03:00	
Air Blank		0.000	03:00	
Control Test		0.077	03:01	
Air Blank		0.000	03:01	
Subject Sample	# 1	0.180	03:02	
Breath Volume		1.128	03:02	
Air Blank		0.000	03:02	
Air Blank		0.000	03:04	
Subject Sample	# 2		03:05	Interferent Detect
Breath Volume		1.371	03:05	
Air Blank		0.000	03:06	
Control Test		0.078	03:06	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:06	

Data Download Date/Time: 02/28/2018 14:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,HECTOR,J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: VEGA,HECTOR
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: URIOSTE

DATE/TIME OF TEST: 02/02/2018 23:47:33

Diagnostics Check	OK	23:46	
Air Blank		23:47	Improper Sample
Air Blank		23:47	Improper Sample

Data Download Date/Time: 02/21/2018 14:53

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,HECTOR,J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: VEGA,HECTOR
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: URIOSTE

DATE/TIME OF TEST: 02/02/2018 23:52:43

Diagnostics Check	OK	23:51	
Air Blank		23:52	Ambient Fail
Air Blank	0.000	23:52	

Data Download Date/Time: 02/21/2018 14:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,HECTOR,J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: VEGA,HECTOR
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: URIOSTE

DATE/TIME OF TEST: 02/03/2018 00:02:33

Diagnostics Check		OK	23:56
Air Blank		0.000	23:56
Control Test		0.078	23:57
Air Blank		0.000	23:57
Subject Sample	# 2	0.201	23:58
Breath Volume		3.156	23:58
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.193	00:00
Breath Volume		2.871	00:00
Air Blank		0.000	00:01
Control Test		0.076	00:01
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 02/21/2018 14:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAILOR,TRAVIS,C
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: VEGA,HECTOR
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:31

SUBJECT'S LAST NAME: QUEZADA

DATE/TIME OF TEST: 02/04/2018 07:05:01

Diagnostics Check		OK	07:02	
Air Blank		0.000	07:02	
Control Test		0.079	07:02	
Air Blank		0.000	07:03	
Subject Sample	# 1		07:03	Subject Test Refused
Breath Volume		0.000	07:03	
Air Blank		0.000	07:04	
Control Test		0.079	07:04	
Air Blank		0.000	07:04	
Diagnostics Check		OK	07:05	

Data Download Date/Time: 02/21/2018 14:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENBURG,DUSTIN,L
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MONTGOMERY,MARCUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 02/18/2018 05:39:40

Diagnostics Check		OK	05:32
Air Blank		0.000	05:33
Control Test		0.078	05:33
Air Blank		0.000	05:34
Subject Sample	# 1	0.082	05:35
Breath Volume		1.644	05:35
Air Blank		0.000	05:35
Air Blank		0.000	05:37
Subject Sample	# 2	0.089	05:38
Breath Volume		2.292	05:38
Air Blank		0.000	05:38
Control Test		0.077	05:39
Air Blank		0.000	05:39
Diagnostics Check		OK	05:39

Data Download Date/Time: 02/21/2018 14:53

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CRAWFORD,BENJAMIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: ATTEA

DATE/TIME OF TEST: 02/25/2018 21:16:24

Diagnostics Check		OK	21:13	
Air Blank		0.000	21:14	
Control Test		0.079	21:14	
Air Blank		0.000	21:14	
Subject Sample	# 1		21:15	Subject Test Refused
Breath Volume		0.000	21:15	
Air Blank		0.000	21:15	
Control Test		0.078	21:15	
Air Blank		0.000	21:16	
Diagnostics Check		OK	21:16	

Data Download Date/Time: 04/04/2018 13:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMSEY,JOHNATHAN,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: TODD,CALEB
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MCELRATH

DATE/TIME OF TEST: 02/01/2018 01:12:12

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.081	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.166	01:07
Breath Volume		1.523	01:07
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.167	01:10
Breath Volume		1.121	01:10
Air Blank		0.000	01:11
Control Test		0.081	01:11
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 02/06/2018 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: 015823080

DATE/TIME OF TEST: 02/03/2018 00:11:13

Diagnostics Check		OK	00:00
Air Blank		0.000	00:01
Control Test		0.081	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.157	00:05
Breath Volume		1.531	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:07
Subject Sample	# 2	0.159	00:09
Breath Volume		1.628	00:09
Air Blank		0.000	00:10
Control Test		0.076	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 02/06/2018 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JONES,JAMES
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 80-006937
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: HALE

DATE/TIME OF TEST: 02/03/2018 04:02:15

Diagnostics Check	OK	04:01	
Air Blank		04:01	Ambient Fail
Air Blank	0.000	04:02	

Data Download Date/Time: 02/06/2018 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JONES,JAMES
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 80-006937
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: HALE'

DATE/TIME OF TEST: 02/03/2018 04:10:11

Diagnostics Check		OK	04:03
Air Blank		0.000	04:04
Control Test		0.079	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.229	04:05
Breath Volume		2.839	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.227	04:08
Breath Volume		2.765	04:08
Air Blank		0.000	04:09
Control Test		0.078	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 02/06/2018 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: YOUNG,ALEX
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: JOSETT

DATE/TIME OF TEST: 02/04/2018 00:34:58

Diagnostics Check		OK	00:25
Air Blank		0.000	00:26
Control Test		0.080	00:26
Air Blank		0.000	00:26
Subject Sample	# 1	0.158	00:29
Breath Volume		1.335	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.152	00:33
Breath Volume		1.914	00:33
Air Blank		0.000	00:34
Control Test		0.076	00:34
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

Data Download Date/Time: 02/06/2018 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 80-006937
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 02/04/2018 01:59:18

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.079	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.229	01:51
Breath Volume		1.421	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:53
Subject Sample	# 2	0.206	01:54
Breath Volume		1.539	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:56
Subject Sample	# 3	0.212	01:57
Breath Volume		1.804	01:57
Air Blank		0.000	01:58
Control Test		0.077	01:58
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 02/06/2018 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,BLAKE,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: LANCASTER,JONATHAN
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/04/2018 02:55:58

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.175	02:51
Breath Volume		2.753	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:53
Subject Sample	# 2	0.170	02:54
Breath Volume		3.363	02:54
Air Blank		0.000	02:55
Control Test		0.077	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 02/06/2018 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.