Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNOR,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: O'CONNOR,STEPHEN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 03/01/2018 04:38:21

Diagnostics Check		OK	04:31
Air Blank		0.000	04:32
Control Test		0.080	04:32
Air Blank		0.000	04:33
Subject Sample	# 1	0.132	04:33
Breath Volume		3.609	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.132	04:36
Breath Volume		3.832	04:36
Air Blank		0.000	04:37
Control Test		0.079	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 03/14/2018 10:57

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVERETT,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: EVERETT,MATTHEW
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 03/03/2018 01:32:21

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:21	
Control Test		0.081	01:21	
Air Blank		0.000	01:21	
Subject Sample	# 1		01:24	No Sample Provided
Breath Volume		0.000	01:24	
Air Blank		0.000	01:25	
Air Blank		0.000	01:27	
Subject Sample	# 2	0.230	01:27	
Breath Volume		3.078	01:27	
Air Blank		0.000	01:28	
Air Blank		0.000	01:30	
Subject Sample	# 3	0.232	01:30	
Breath Volume		3.843	01:30	
Air Blank		0.000	01:31	
Control Test		0.081	01:31	
Air Blank		0.000	01:32	
Diagnostics Check		OK	01:32	

Data Download Date/Time: 03/14/2018 10:57

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULLOCK,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BULLOCK,JOSHUA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: PRESTON

DATE/TIME OF TEST: 03/03/2018 02:39:55

Diagnostics Check		OK	02:31
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.196	02:34
Breath Volume		2.351	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:36
Subject Sample	# 2	0.197	02:38
Breath Volume		1.886	02:38
Air Blank		0.000	02:39
Control Test		0.081	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 03/14/2018 10:57

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAXTON,D,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BLEVINS,A
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: HAMES

DATE/TIME OF TEST: 03/03/2018 06:46:37

Diagnostics Check		OK	06:39
Air Blank		0.000	06:39
Control Test		0.081	06:40
Air Blank		0.000	06:40
Subject Sample	# 1	0.105	06:42
Breath Volume		1.406	06:42
Air Blank		0.000	06:42
Air Blank		0.000	06:44
Subject Sample	# 2	0.102	06:45
Breath Volume		1.359	06:45
Air Blank		0.000	06:45
Control Test		0.080	06:46
Air Blank		0.000	06:46
Diagnostics Check		OK	06:46

Data Download Date/Time: 03/14/2018 10:57

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 03/04/2018 00:54:28

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.081	00:48
Air Blank		0.000	00:49
Subject Sample	# 1	0.258	00:49
Breath Volume		2.664	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.252	00:52
Breath Volume		2.414	00:52
Air Blank		0.000	00:53
Control Test		0.080	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 03/14/2018 10:57

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARNER,TYLER,J
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: MIMS,MITCHELL
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: FAGAN

DATE/TIME OF TEST: 03/10/2018 00:31:42

Diagnostics Check		OK	00:24
Air Blank		0.000	00:25
Control Test		0.080	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.145	00:26
Breath Volume		1.328	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.150	00:30
Breath Volume		2.183	00:30
Air Blank		0.000	00:30
Control Test		0.080	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 03/14/2018 10:57

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: PEARSON,DAMIEN
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 03/10/2018 02:00:26

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:53	
Control Test		0.080	01:53	
Air Blank		0.000	01:54	
Subject Sample	# 1	0.136	01:55	
Breath Volume		2.183	01:55	
Air Blank		0.000	01:56	
Air Blank		0.000	01:58	
Subject Sample	# 2	0.138	01:58	
Breath Volume		3.078	01:58	
Air Blank		0.000	01:59	
Control Test		0.073	01:59	Control Outside Tolerance
Air Blank		0.000	02:00	

Data Download Date/Time: 03/14/2018 10:57

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARTHY,THEODORE,G
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: MCCARTHY,THEODORE
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:43

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/11/2018 08:23:04

Diagnostics Check		OK	08:16
Air Blank		0.000	08:17
Control Test		0.080	08:17
Air Blank		0.000	08:18
Subject Sample	# 1	0.115	08:18
Breath Volume		2.980	08:18
Air Blank		0.000	08:19
Air Blank		0.000	08:21
Subject Sample	# 2	0.107	08:21
Breath Volume		2.695	08:21
Air Blank		0.000	08:22
Control Test		0.080	08:22
Air Blank		0.000	08:22
Diagnostics Check		OK	08:23

Data Download Date/Time: 03/14/2018 10:57

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVERETT,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: VANN,TOM
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: HINTON

DATE/TIME OF TEST: 03/12/2018 23:17:31

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.080	23:12
Air Blank		0.000	23:12
Subject Sample	# 1	0.177	23:13
Breath Volume		1.574	23:13
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.180	23:16
Breath Volume		1.894	23:16
Air Blank		0.000	23:16
Control Test		0.080	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 03/14/2018 10:57

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COHEN,VALARIE,L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MARTINA,RYAN
ARRESTING OFFICER AGENCY: FHP Headquarters
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: KEVER

DATE/TIME OF TEST: 03/14/2018 22:50:17

Diagnostics Check		OK	22:42
Air Blank		0.000	22:42
Control Test		0.081	22:43
Air Blank		0.000	22:43
Subject Sample	# 1	0.211	22:46
Breath Volume		1.929	22:46
Air Blank		0.000	22:46
Air Blank		0.000	22:48
Subject Sample	# 2	0.206	22:48
Breath Volume		1.718	22:48
Air Blank		0.000	22:49
Control Test		0.080	22:49
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 04/25/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVERETT,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: EVERETT,MATTHEW
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:09

SUBJECT'S LAST NAME: APLIN

DATE/TIME OF TEST: 03/15/2018 17:44:43

Diagnostics Check		OK	17:37
Air Blank		0.000	17:37
Control Test		0.080	17:38
Air Blank		0.000	17:38
Subject Sample	# 1	0.303	17:39
Breath Volume		1.527	17:39
Air Blank		0.000	17:40
Air Blank		0.000	17:42
Subject Sample	# 2	0.308	17:43
Breath Volume		1.496	17:43
Air Blank		0.000	17:43
Control Test		0.079	17:44
Air Blank		0.000	17:44
Diagnostics Check		OK	17:44

Data Download Date/Time: 04/25/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVERETT,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARVEY,RHETT
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ELLINGTON

DATE/TIME OF TEST: 03/16/2018 00:18:21

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:06	
Control Test		0.081	00:06	
Air Blank		0.000	00:06	
Subject Sample	# 1		00:10	No Sample Provided
Breath Volume		0.000	00:10	
Air Blank		0.000	00:10	
Air Blank		0.000	00:12	
Subject Sample	# 2	0.139	00:13	
Breath Volume		2.101	00:13	
Air Blank		0.000	00:13	
Air Blank		0.000	00:15	
Subject Sample	# 3	0.130	00:16	
Breath Volume		1.742	00:16	
Air Blank		0.000	00:17	
Control Test		0.080	00:17	
Air Blank		0.000	00:18	
Diagnostics Check		OK	00:18	

Data Download Date/Time: 04/25/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BLEVINS,A.
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: BOYNTON

DATE/TIME OF TEST: 03/16/2018 05:13:50

Diagnostics Check		OK	05:01	
Air Blank		0.000	05:01	
Control Test		0.080	05:02	
Air Blank		0.000	05:02	
Subject Sample	# 1	0.224	05:03	
Breath Volume		1.898	05:03	
Air Blank		0.000	05:04	
Air Blank		0.000	05:06	
Subject Sample	# 2	0.186	05:09	Volume Not Met
Breath Volume		0.394	05:09	
Air Blank		0.000	05:09	
Air Blank		0.000	05:11	
Subject Sample	# 3	0.223	05:12	
Breath Volume		1.925	05:12	
Air Blank		0.000	05:12	
Control Test		0.080	05:13	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 04/25/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEIDEL,JEFFERY,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HEIDEL,JEFFERY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: EL-SINNARY

DATE/TIME OF TEST: 03/17/2018 02:59:48

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.079	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.033	02:55
Breath Volume		2.601	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.032	02:58
Breath Volume		3.304	02:58
Air Blank		0.000	02:58
Control Test		0.079	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 04/25/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAS,BRANDYN,M
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: DEAS,BRANDYN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BORDEN

DATE/TIME OF TEST: 03/17/2018 03:50:50

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.077	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.183	03:46
Breath Volume		2.878	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.191	03:49
Breath Volume		2.437	03:49
Air Blank		0.000	03:49
Control Test		0.078	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 04/25/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: COLLINS,C
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 03/18/2018 02:10:06

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.080	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.195	02:05
Breath Volume		1.535	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.190	02:08
Breath Volume		1.343	02:08
Air Blank		0.000	02:09
Control Test		0.079	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 04/25/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAXTON,D,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BRAXTON,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:31

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/20/2018 08:06:12

Diagnostics Check		OK	07:54
Air Blank		0.000	07:54
Control Test		0.078	07:55
Air Blank		0.000	07:55
Subject Sample	# 1	0.221	07:58
Breath Volume		2.527	07:58
Air Blank		0.000	07:59
Air Blank		0.000	08:00
Subject Sample	# 2	0.256	08:01
Breath Volume		4.328	08:01
Air Blank		0.000	08:02
Air Blank		0.000	08:04
Subject Sample	# 3	0.250	08:04
Breath Volume		4.253	08:04
Air Blank		0.000	08:05
Control Test		0.075	08:05
Air Blank		0.000	08:06
Diagnostics Check		OK	08:06

Data Download Date/Time: 04/25/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAXTON,D,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BRAXTON,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: SHELTON

DATE/TIME OF TEST: 03/20/2018 23:26:49

Diagnostics Check		OK	23:20
Air Blank		0.000	23:21
Control Test		0.080	23:21
Air Blank		0.000	23:22
Subject Sample	# 1	0.135	23:22
Breath Volume		1.949	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:24
Subject Sample	# 2	0.133	23:25
Breath Volume		1.562	23:25
Air Blank		0.000	23:25
Control Test		0.080	23:26
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 04/25/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 03/24/2018 23:05:57

Diagnostics Check		OK	22:59
Air Blank		0.000	22:59
Control Test		0.079	23:00
Air Blank		0.000	23:00
Subject Sample	# 1	0.285	23:01
Breath Volume		2.035	23:01
Air Blank		0.000	23:02
Air Blank		0.000	23:03
Subject Sample	# 2	0.270	23:04
Breath Volume		1.902	23:04
Air Blank		0.000	23:05
Control Test		0.077	23:05
Air Blank		0.000	23:05
Diagnostics Check		OK	23:05

Data Download Date/Time: 04/25/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: COLLINS,CAMERON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/25/2018 00:14:08

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.079	00:05
Air Blank		0.000	00:05
Subject Sample	# 1	0.150	00:06
Breath Volume		1.183	00:06
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.126	00:09
Breath Volume		1.371	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 3	0.136	00:12
Breath Volume		1.156	00:12
Air Blank		0.000	00:13
Control Test		0.079	00:13
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 04/25/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: STEIN

DATE/TIME OF TEST: 03/25/2018 02:20:10

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.079	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.255	02:11
Breath Volume		1.757	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.230	02:15
Breath Volume		1.820	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 3	0.232	02:18
Breath Volume		2.796	02:18
Air Blank		0.000	02:19
Control Test		0.077	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 04/25/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIMBA,MIKE,J
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: ZIMBA,MIKE
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: OQUERDO

DATE/TIME OF TEST: 03/26/2018 03:06:32

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.080	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.180	03:01
Breath Volume		2.320	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.188	03:05
Breath Volume		2.457	03:05
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 04/25/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,PAUL,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: AZRAK,PAUL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/28/2018 00:06:31

Diagnostics Check		OK	00:00
Air Blank		0.000	00:01
Control Test		0.080	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.195	00:02
Breath Volume		2.710	00:02
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.192	00:05
Breath Volume		2.265	00:05
Air Blank		0.000	00:05
Control Test		0.080	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 04/25/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: POSCHARSKY

DATE/TIME OF TEST: 03/29/2018 02:20:39

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.080	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.134	02:16
Breath Volume		1.406	02:16
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.139	02:19
Breath Volume		1.605	02:19
Air Blank		0.000	02:19
Control Test		0.079	02:20
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 04/25/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,R,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: STEELE,R
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 03/31/2018 03:51:14

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.080	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.197	03:46
Breath Volume		1.648	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.192	03:49
Breath Volume		2.386	03:49
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 04/25/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/03/2018 01:44:54

Diagnostics Check		OK	01:38
Air Blank		0.000	01:39
Control Test		0.080	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.104	01:40
Breath Volume		2.601	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.097	01:43
Breath Volume		2.257	01:43
Air Blank		0.000	01:43
Control Test		0.083	01:44
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 03/05/2018 21:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: GRAF

DATE/TIME OF TEST: 03/03/2018 21:16:33

Diagnostics Check	OK	21:15	
Air Blank		21:15	Ambient Fail
Air Blank	0.000	21:16	

Data Download Date/Time: 03/05/2018 21:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: GRAF

DATE/TIME OF TEST: 03/03/2018 21:19:09

Diagnostics Check	OK	21:18	
Air Blank		21:18	Ambient Fail
Air Blank	0.000	21:19	

Data Download Date/Time: 03/05/2018 21:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/04/2018 02:17:03

Diagnostics Check	OK	02:16	
Air Blank		02:16	Ambient Fail
Air Blank	0.000	02:17	

Data Download Date/Time: 03/05/2018 21:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/04/2018 02:24:06

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.078	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.172	02:19
Breath Volume		1.859	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.164	02:22
Breath Volume		1.636	02:22
Air Blank		0.000	02:23
Control Test		0.082	02:23
Air Blank		0.000	02:23
Diagnostics Check		OK	02:24

Data Download Date/Time: 03/05/2018 21:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 03/06/2018 02:06:36

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.077	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.202	02:02
Breath Volume		2.480	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:04
Subject Sample	# 2	0.206	02:05
Breath Volume		2.945	02:05
Air Blank		0.000	02:05
Control Test		0.082	02:06
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 04/03/2018 19:29

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SHECKELLS

DATE/TIME OF TEST: 03/10/2018 00:17:18

Diagnostics Check		OK	00:13	
Air Blank		0.000	00:13	
Control Test		0.077	00:13	
Air Blank		0.000	00:14	
Subject Sample	# 1	0.196	00:15	
Breath Volume		2.742	00:15	
Air Blank			00:16	Purge Fail
Air Blank			00:17	Purge Fail

Data Download Date/Time: 04/03/2018 19:29

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SHECKELLS

DATE/TIME OF TEST: 03/10/2018 00:21:00

Diagnostics Check	OK	00:20	
Air Blank		00:20	Ambient Fail
Air Blank	0.000	00:21	

Data Download Date/Time: 04/03/2018 19:29

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HALPIN

DATE/TIME OF TEST: 03/23/2018 02:06:18

Diagnostics Check	OK	02:05	
Air Blank		02:05	Ambient Fail
Air Blank	0.000	02:06	

Data Download Date/Time: 04/03/2018 19:29

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HALPIN

DATE/TIME OF TEST: 03/23/2018 02:08:47

Diagnostics Check	OK	02:07	
Air Blank		02:08	Ambient Fail
Air Blank		02:08	Purge Fail

Data Download Date/Time: 04/03/2018 19:29

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HALPIN

DATE/TIME OF TEST: 03/23/2018 02:19:12

Diagnostics Check	OK	02:18	
Air Blank		02:18	Ambient Fail
Air Blank	0.000	02:19	

Data Download Date/Time: 04/03/2018 19:29

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HALPIN

DATE/TIME OF TEST: 03/23/2018 02:21:13

Diagnostics Check	OK	02:20	
Air Blank		02:20	Ambient Fail
Air Blank	0.000	02:21	

Data Download Date/Time: 04/03/2018 19:29

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: USRY,STEPHEN,K
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: TRAN,DAVID
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 03/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: REVELL

DATE/TIME OF TEST: 03/23/2018 05:06:25

Diagnostics Check		OK	04:59
Air Blank		0.000	04:59
Control Test		0.081	05:00
Air Blank		0.000	05:00
Subject Sample	# 1	0.248	05:01
Breath Volume		3.601	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.251	05:04
Breath Volume		3.769	05:04
Air Blank		0.000	05:05
Control Test		0.078	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 04/28/2018 16:08

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHRE,PETER,M
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: HERRERA,MARK
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 03/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 03/26/2018 00:33:20

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:28
Subject Sample	# 1	0.000	00:29
Breath Volume		1.710	00:29
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.000	00:31
Breath Volume		2.894	00:31
Air Blank		0.000	00:32
Control Test		0.078	00:32
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 04/28/2018 16:08

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LECOCQ,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: KOWAL

DATE/TIME OF TEST: 03/03/2018 01:17:05

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.082	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.103	01:12
Breath Volume		3.550	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.099	01:15
Breath Volume		3.121	01:15
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 04/20/2018 10:10

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORTON,JAMES,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LEON,ERICK
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 03/06/2018 00:47:33

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.081	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.222	00:42
Breath Volume		2.238	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.229	00:45
Breath Volume		2.574	00:45
Air Blank		0.000	00:46
Control Test		0.080	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILHELM,TONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: DUCHARME

DATE/TIME OF TEST: 03/09/2018 03:03:58

Diagnostics Check		OK	03:01	
Air Blank		0.000	03:01	
Control Test		0.082	03:01	
Air Blank		0.000	03:02	
Subject Sample	# 1		03:02	Subject Test Refused
Breath Volume		0.000	03:02	
Air Blank		0.000	03:03	
Control Test		0.083	03:03	
Air Blank		0.000	03:03	
Diagnostics Check		OK	03:03	

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNHAM,J,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURNHAM,J
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 03/10/2018 01:05:20

Diagnostics Check		OK	00:55	
Air Blank		0.000	00:55	
Control Test		0.081	00:55	
Air Blank		0.000	00:56	
Subject Sample	# 1	0.166	00:59	Volume Not Met
Breath Volume		0.199	00:59	
Air Blank		0.000	01:00	
Air Blank		0.000	01:01	
Subject Sample	# 2		01:03	Subject Test Refused
Breath Volume		0.820	01:03	
Air Blank		0.000	01:04	
Control Test		0.079	01:04	
Air Blank		0.000	01:05	
Diagnostics Check		OK	01:05	

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLACKMON,DARRELL,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,LANCE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: BLACKMON

DATE/TIME OF TEST: 03/11/2018 21:33:28

Diagnostics Check		OK	21:26
Air Blank		0.000	21:27
Control Test		0.080	21:27
Air Blank		0.000	21:28
Subject Sample	# 1	0.216	21:28
Breath Volume		3.273	21:28
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 2	0.205	21:31
Breath Volume		2.582	21:31
Air Blank		0.000	21:32
Control Test		0.076	21:32
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILHELM,TONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: GAFFORD

DATE/TIME OF TEST: 03/12/2018 19:52:12

Diagnostics Check		OK	19:45
Air Blank		0.000	19:46
Control Test		0.081	19:46
Air Blank		0.000	19:47
Subject Sample	# 1	0.221	19:47
Breath Volume		3.468	19:47
Air Blank		0.000	19:48
Air Blank		0.000	19:50
Subject Sample	# 2	0.228	19:50
Breath Volume		2.742	19:50
Air Blank		0.000	19:51
Control Test		0.079	19:51
Air Blank		0.000	19:52
Diagnostics Check		OK	19:52

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILHELM,TONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: ORCUTT

DATE/TIME OF TEST: 03/13/2018 20:53:22

Diagnostics Check		OK	20:50	
Air Blank		0.000	20:50	
Control Test		0.081	20:51	
Air Blank		0.000	20:51	
Subject Sample	# 1		20:51	Subject Test Refused
Breath Volume		0.000	20:51	
Air Blank		0.000	20:52	
Control Test		0.081	20:52	
Air Blank		0.000	20:53	
Diagnostics Check		OK	20:53	

Data Download Date/Time: 04/20/2018 10:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILDE,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: SHARITT

DATE/TIME OF TEST: 03/13/2018 22:18:43

Diagnostics Check		OK	22:15	
Air Blank		0.000	22:16	
Control Test		0.082	22:16	
Air Blank		0.000	22:16	
Subject Sample	# 1		22:17	Subject Test Refused
Breath Volume		0.000	22:17	
Air Blank		0.000	22:17	
Control Test		0.082	22:18	
Air Blank		0.000	22:18	
Diagnostics Check		OK	22:18	

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUH,CASEY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DUH,CASEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: SHUBERT

DATE/TIME OF TEST: 03/14/2018 00:35:01

Diagnostics Check		OK	00:27
Air Blank		0.000	00:28
Control Test		0.081	00:28
Air Blank		0.000	00:29
Subject Sample	# 1	0.187	00:30
Breath Volume		3.304	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.180	00:33
Breath Volume		3.355	00:33
Air Blank		0.000	00:34
Control Test		0.077	00:34
Air Blank		0.000	00:34
Diagnostics Check		OK	00:35

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,ROBERT,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NAYLOR,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 03/15/2018 01:52:54

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.082	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.090	01:48
Breath Volume		3.835	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:50
Subject Sample	# 2	0.088	01:51
Breath Volume		3.722	01:51
Air Blank		0.000	01:51
Control Test		0.081	01:52
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORTON,JAMES,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MONARCH,BRIAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: MCNUTT

DATE/TIME OF TEST: 03/15/2018 05:26:33

Diagnostics Check		OK	05:15	
Air Blank		0.000	05:15	
Control Test		0.082	05:15	
Air Blank		0.000	05:16	
Subject Sample	# 1	0.000	05:19	Volume Not Met
Breath Volume		0.152	05:19	
Air Blank		0.000	05:20	
Air Blank		0.000	05:22	
Subject Sample	# 2		05:25	No Sample Provided
Breath Volume		0.000	05:25	
Air Blank		0.000	05:25	
Control Test		0.082	05:25	
Air Blank		0.000	05:26	
Diagnostics Check		OK	05:26	

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RACER,JACOB,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RACER,JACOB
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 14:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: PYLANT

DATE/TIME OF TEST: 03/15/2018 16:44:01

Diagnostics Check		OK	16:38	
Air Blank		0.000	16:39	
Control Test		0.081	16:39	
Air Blank		0.000	16:40	
Subject Sample	# 1		16:42	Subject Test Refused
Breath Volume		0.000	16:42	
Air Blank		0.000	16:43	
Control Test		0.082	16:43	
Air Blank		0.000	16:43	
Diagnostics Check		OK	16:44	

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORTON,JAMES,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ORDOQUI,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: RONDON

DATE/TIME OF TEST: 03/16/2018 05:13:33

Diagnostics Check		OK	05:02	
Air Blank		0.000	05:02	
Control Test		0.081	05:02	
Air Blank		0.000	05:03	
Subject Sample	# 1		05:06	No Sample Provided
Breath Volume		0.000	05:06	
Air Blank		0.000	05:07	
Air Blank		0.000	05:09	
Subject Sample	# 2		05:12	No Sample Provided
Breath Volume		0.000	05:12	
Air Blank		0.000	05:12	
Control Test		0.082	05:12	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUH,CASEY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DUH,CASEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: HALE

DATE/TIME OF TEST: 03/16/2018 22:42:17

Diagnostics Check		OK	22:38	
Air Blank		0.000	22:39	
Control Test		0.081	22:39	
Air Blank		0.000	22:40	
Subject Sample	# 1		22:40	Subject Test Refused
Breath Volume		0.000	22:40	
Air Blank		0.000	22:41	
Control Test		0.081	22:41	
Air Blank		0.000	22:42	
Diagnostics Check		OK	22:42	

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NABULSI,MOOTAZ
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: CHASTAIN

DATE/TIME OF TEST: 03/18/2018 05:41:40

Diagnostics Check		OK	05:38	
Air Blank		0.000	05:39	
Control Test		0.081	05:39	
Air Blank		0.000	05:40	
Subject Sample	# 1		05:40	Subject Test Refused
Breath Volume		0.000	05:40	
Air Blank		0.000	05:40	
Control Test		0.081	05:41	
Air Blank		0.000	05:41	
Diagnostics Check		OK	05:41	

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAWYER,DARRYL,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAWYER,DARRYL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: MADDOX

DATE/TIME OF TEST: 03/18/2018 20:53:38

Diagnostics Check		OK	20:47	
Air Blank		0.000	20:47	
Control Test		0.080	20:47	
Air Blank		0.000	20:48	
Subject Sample	# 1	0.225	20:50	
Breath Volume		2.742	20:50	
Air Blank		0.000	20:51	
Air Blank			20:53	Ambient Fail
Air Blank		0.000	20:53	

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAWYER,DARRYL,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAWYER,DARRYL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: MADDOX

DATE/TIME OF TEST: 03/18/2018 21:04:17

Diagnostics Check		OK	20:57	
Air Blank		0.000	20:58	
Control Test		0.080	20:58	
Air Blank		0.000	20:59	
Subject Sample	# 1	0.232	20:59	
Breath Volume		2.406	20:59	
Air Blank		0.000	21:00	
Air Blank		0.000	21:02	
Subject Sample	# 2	0.215	21:02	
Breath Volume		2.656	21:02	
Air Blank		0.000	21:03	
Control Test		0.074	21:03	Control Outside Tolerance
Air Blank		0.000	21:04	

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LECOCQ,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: DURAND

DATE/TIME OF TEST: 03/18/2018 23:01:48

Diagnostics Check		OK	22:55
Air Blank		0.000	22:56
Control Test		0.081	22:56
Air Blank		0.000	22:57
Subject Sample	# 1	0.000	22:57
Breath Volume		1.539	22:57
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.000	23:00
Breath Volume		1.355	23:00
Air Blank		0.000	23:00
Control Test		0.077	23:01
Air Blank		0.000	23:01
Diagnostics Check		OK	23:01

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMSON,TYLER,H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMSON,TYLER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 12:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/23/2018 01:16:32

Diagnostics Check		OK	01:09
Air Blank		0.000	01:10
Control Test		0.082	01:10
Air Blank		0.000	01:10
Subject Sample	# 1	0.157	01:12
Breath Volume		2.308	01:12
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.155	01:14
Breath Volume		2.074	01:14
Air Blank		0.000	01:15
Control Test		0.081	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORTON,JAMES,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ALEXANDER,C
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 03/25/2018 00:41:41

Diagnostics Check		OK	00:34
Air Blank		0.000	00:35
Control Test		0.081	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.162	00:37
Breath Volume		1.949	00:37
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.156	00:40
Breath Volume		2.171	00:40
Air Blank		0.000	00:40
Control Test		0.080	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,ROBERT,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ORDOQUI,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: ROMANS

DATE/TIME OF TEST: 03/25/2018 04:36:51

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.080	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.065	04:32
Breath Volume		1.558	04:32
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.074	04:35
Breath Volume		2.277	04:35
Air Blank		0.000	04:35
Control Test		0.081	04:36
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 04/20/2018 10:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMSON,TYLER,H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STEVENS,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: RICKS

DATE/TIME OF TEST: 03/27/2018 21:02:10

Diagnostics Check		OK	20:55
Air Blank		0.000	20:55
Control Test		0.081	20:55
Air Blank		0.000	20:56
Subject Sample	# 1	0.219	20:57
Breath Volume		2.925	20:57
Air Blank		0.000	20:58
Air Blank		0.000	20:59
Subject Sample	# 2	0.204	21:00
Breath Volume		1.875	21:00
Air Blank		0.000	21:01
Control Test		0.080	21:01
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 04/20/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHRE,PETER,M
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: SOWELL,JONATHAN
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MOMMAERTS

DATE/TIME OF TEST: 03/03/2018 05:49:53

Diagnostics Check		OK	05:43
Air Blank		0.000	05:43
Control Test		0.081	05:44
Air Blank		0.000	05:44
Subject Sample	# 1	0.231	05:45
Breath Volume		2.437	05:45
Air Blank		0.000	05:45
Air Blank		0.000	05:47
Subject Sample	# 2	0.230	05:48
Breath Volume		2.699	05:48
Air Blank		0.000	05:48
Control Test		0.080	05:49
Air Blank		0.000	05:49
Diagnostics Check		OK	05:49

Data Download Date/Time: 03/12/2018 14:02

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACBETH,MATTHEW,R
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: MACBETH,MATTHEW
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BALDREE

DATE/TIME OF TEST: 03/06/2018 01:11:40

Diagnostics Check		OK	01:08	
Air Blank		0.000	01:09	
Control Test		0.081	01:09	
Air Blank		0.000	01:10	
Subject Sample	# 1		01:10	Subject Test Refused
Breath Volume		0.000	01:10	
Air Blank		0.000	01:10	
Control Test		0.081	01:11	
Air Blank		0.000	01:11	
Diagnostics Check		OK	01:11	

Data Download Date/Time: 03/12/2018 14:03

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,ALLISON,T
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: KING,JONATHAN
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: ADMIN
ARREST TIME: 08:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: SANDOVAL

DATE/TIME OF TEST: 03/08/2018 09:56:43

Diagnostics Check		OK	09:50
Air Blank		0.000	09:51
Control Test		0.080	09:51
Air Blank		0.000	09:51
Subject Sample	# 1	0.068	09:52
Breath Volume		2.187	09:52
Air Blank		0.000	09:52
Air Blank		0.000	09:54
Subject Sample	# 2	0.067	09:55
Breath Volume		2.746	09:55
Air Blank		0.000	09:55
Control Test		0.082	09:56
Air Blank		0.000	09:56
Diagnostics Check		OK	09:56

Data Download Date/Time: 03/12/2018 14:03

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY,WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/02/2018 23:51:57

Diagnostics Check		OK	23:44
Air Blank		0.000	23:45
Control Test		0.080	23:45
Air Blank		0.000	23:46
Subject Sample	# 1	0.153	23:47
Breath Volume		2.277	23:47
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.155	23:50
Breath Volume		1.296	23:50
Air Blank		0.000	23:51
Control Test		0.080	23:51
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: IVANS,KATRINA
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 03/03/2018 19:18:52

Diagnostics Check		OK	19:12
Air Blank		0.000	19:12
Control Test		0.081	19:12
Air Blank		0.000	19:13
Subject Sample	# 1	0.270	19:14
Breath Volume		2.843	19:14
Air Blank		0.000	19:14
Air Blank		0.000	19:16
Subject Sample	# 2	0.254	19:17
Breath Volume		3.296	19:17
Air Blank		0.000	19:17
Control Test		0.081	19:18
Air Blank		0.000	19:18
Diagnostics Check		OK	19:18

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY,WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: RITTER

DATE/TIME OF TEST: 03/04/2018 00:15:49

Diagnostics Check		OK	00:08
Air Blank		0.000	00:08
Control Test		0.080	00:09
Air Blank		0.000	00:09
Subject Sample	# 1	0.164	00:11
Breath Volume		3.300	00:11
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.159	00:14
Breath Volume		2.992	00:14
Air Blank		0.000	00:14
Control Test		0.079	00:15
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: GREEN,HARRY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: NETZER

DATE/TIME OF TEST: 03/04/2018 04:06:55

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.080	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.153	04:02
Breath Volume		2.792	04:02
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.146	04:05
Breath Volume		3.347	04:05
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARACLE,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GAFKJEN,THOMAS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: BROYLES

DATE/TIME OF TEST: 03/08/2018 17:05:48

Diagnostics Check		OK	17:02	
Air Blank		0.000	17:03	
Control Test		0.080	17:03	
Air Blank		0.000	17:03	
Subject Sample	# 1		17:04	Slope Not Met
Breath Volume		1.218	17:04	
Air Blank		0.000	17:04	
Control Test		0.081	17:05	
Air Blank		0.000	17:05	
Diagnostics Check		OK	17:05	

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARACLE,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GAFKJEN,THOMAS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: BROYLES

DATE/TIME OF TEST: 03/08/2018 17:13:51

Diagnostics Check		OK	17:07
Air Blank		0.000	17:08
Control Test		0.081	17:08
Air Blank		0.000	17:08
Subject Sample	# 1	0.281	17:09
Breath Volume		3.023	17:09
Air Blank		0.000	17:10
Air Blank		0.000	17:11
Subject Sample	# 2	0.269	17:12
Breath Volume		2.914	17:12
Air Blank		0.000	17:12
Control Test		0.079	17:13
Air Blank		0.000	17:13
Diagnostics Check		OK	17:13

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VO,SEBASTIAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/08/2018 20:23:40

Diagnostics Check		OK	20:13
Air Blank		0.000	20:13
Control Test		0.080	20:14
Air Blank		0.000	20:14
Subject Sample	# 1	0.201	20:17
Breath Volume		1.843	20:17
Air Blank		0.000	20:18
Air Blank		0.000	20:20
Subject Sample	# 2	0.182	20:21
Breath Volume		1.664	20:21
Air Blank		0.000	20:22
Control Test		0.076	20:23
Air Blank		0.000	20:23
Diagnostics Check		OK	20:23

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY,WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: GRANO

DATE/TIME OF TEST: 03/10/2018 04:16:44

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.079	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.136	04:12
Breath Volume		1.609	04:12
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.144	04:15
Breath Volume		1.671	04:15
Air Blank		0.000	04:15
Control Test		0.079	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POOL,HAROLD,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ROBINSON,ANDREW
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/10/2018 20:32:18

Diagnostics Check		OK	20:26
Air Blank		0.000	20:26
Control Test		0.080	20:27
Air Blank		0.000	20:27
Subject Sample	# 1	0.000	20:28
Breath Volume		2.867	20:28
Air Blank		0.000	20:28
Air Blank		0.000	20:30
Subject Sample	# 2	0.000	20:30
Breath Volume		2.789	20:30
Air Blank		0.000	20:31
Control Test		0.080	20:31
Air Blank		0.000	20:32
Diagnostics Check		OK	20:32

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY,WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: CHAP

DATE/TIME OF TEST: 03/11/2018 19:36:13

Diagnostics Check		OK	19:28
Air Blank		0.000	19:29
Control Test		0.079	19:29
Air Blank		0.000	19:29
Subject Sample	# 1	0.241	19:30
Breath Volume		1.714	19:30
Air Blank		0.000	19:31
Air Blank		0.000	19:33
Subject Sample	# 2	0.235	19:34
Breath Volume		2.410	19:34
Air Blank		0.000	19:35
Control Test		0.079	19:35
Air Blank		0.000	19:36
Diagnostics Check		OK	19:36

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SOSA,SANTIAGO
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: ESTRADA

DATE/TIME OF TEST: 03/11/2018 21:31:30

Diagnostics Check		OK	21:20
Air Blank		0.000	21:20
Control Test		0.079	21:21
Air Blank		0.000	21:21
Subject Sample	# 1	0.232	21:24
Breath Volume		1.386	21:24
Air Blank		0.000	21:24
Air Blank		0.000	21:26
Subject Sample	# 2	0.219	21:29
Breath Volume		1.593	21:29
Air Blank		0.000	21:30
Control Test		0.079	21:30
Air Blank		0.000	21:31
Diagnostics Check		OK	21:31

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARY,ADAM,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BOATWRIGHT,JAMIE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 03/11/2018 23:07:52

Diagnostics Check		OK	23:00
Air Blank		0.000	23:01
Control Test		0.078	23:01
Air Blank		0.000	23:02
Subject Sample	# 1	0.168	23:03
Breath Volume		1.210	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:05
Subject Sample	# 2	0.157	23:06
Breath Volume		1.757	23:06
Air Blank		0.000	23:06
Control Test		0.078	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY,WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: SCHULER

DATE/TIME OF TEST: 03/14/2018 01:50:05

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.081	01:44
Air Blank		0.000	01:44
Subject Sample	# 1	0.184	01:45
Breath Volume		2.406	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:47
Subject Sample	# 2	0.187	01:48
Breath Volume		2.679	01:48
Air Blank		0.000	01:49
Control Test		0.081	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:50

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEITE,STEVEN,G
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SECOR,BROC
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: DONAGHY

DATE/TIME OF TEST: 03/15/2018 23:25:59

Diagnostics Check		OK	23:18
Air Blank		0.000	23:18
Control Test		0.080	23:19
Air Blank		0.000	23:19
Subject Sample	# 1	0.238	23:20
Breath Volume		1.296	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:22
Subject Sample	# 2	0.224	23:24
Breath Volume		1.128	23:24
Air Blank		0.000	23:25
Control Test		0.079	23:25
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDRIX,TIMOTHY,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENDRIX,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 03/16/2018 21:09:10

Diagnostics Check		OK	20:57	
Air Blank		0.000	20:58	
Control Test		0.080	20:58	
Air Blank		0.000	20:58	
Subject Sample	# 1	0.114	21:01	Volume Not Met
Breath Volume		0.253	21:01	
Air Blank		0.000	21:02	
Air Blank		0.000	21:04	
Subject Sample	# 2	0.102	21:07	Volume Not Met
Breath Volume		0.031	21:07	
Air Blank		0.000	21:08	
Control Test		0.074	21:08	Control Outside Tolerance
Air Blank		0.000	21:09	

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDRIX,TIMOTHY,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENDRIX,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 03/16/2018 21:23:25

Diagnostics Check		OK	21:16	
Air Blank		0.000	21:17	
Control Test		0.078	21:17	
Air Blank		0.000	21:18	
Subject Sample	# 1	0.126	21:18	
Breath Volume		1.187	21:18	
Air Blank		0.000	21:19	
Air Blank		0.000	21:21	
Subject Sample	# 2		21:22	Subject Test Refused
Breath Volume		0.000	21:22	
Air Blank		0.000	21:22	
Control Test		0.080	21:22	
Air Blank		0.000	21:23	
Diagnostics Check		OK	21:23	

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,HARRY,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: GREEN,HARRY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: GULLEY

DATE/TIME OF TEST: 03/18/2018 01:01:42

Diagnostics Check		OK	00:54
Air Blank		0.000	00:55
Control Test		0.078	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.117	00:56
Breath Volume		1.324	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.127	01:00
Breath Volume		1.820	01:00
Air Blank		0.000	01:00
Control Test		0.080	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BYRD,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: BUCK

DATE/TIME OF TEST: 03/20/2018 04:34:37

Diagnostics Check		OK	04:31	
Air Blank		0.000	04:32	
Control Test		0.078	04:32	
Air Blank		0.000	04:33	
Subject Sample	# 1		04:33	Subject Test Refused
Breath Volume		0.000	04:33	
Air Blank		0.000	04:33	
Control Test		0.079	04:34	
Air Blank		0.000	04:34	
Diagnostics Check		OK	04:34	

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 03/22/2018 00:06:59

Diagnostics Check		OK	00:00
Air Blank		0.000	00:01
Control Test		0.080	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.101	00:02
Breath Volume		1.554	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.107	00:05
Breath Volume		1.394	00:05
Air Blank		0.000	00:06
Control Test		0.079	00:06
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEITE,STEVEN,G
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEITE,STEVEN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/22/2018 01:19:50

Diagnostics Check		OK	01:13
Air Blank		0.000	01:14
Control Test		0.080	01:14
Air Blank		0.000	01:15
Subject Sample	# 1	0.127	01:15
Breath Volume		2.914	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:17
Subject Sample	# 2	0.125	01:18
Breath Volume		2.292	01:18
Air Blank		0.000	01:18
Control Test		0.079	01:19
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAMUEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: CHEATHAM

DATE/TIME OF TEST: 03/22/2018 23:23:45

Diagnostics Check		OK	23:17
Air Blank		0.000	23:17
Control Test		0.080	23:17
Air Blank		0.000	23:18
Subject Sample	# 1	0.184	23:19
Breath Volume		2.039	23:19
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.183	23:22
Breath Volume		2.542	23:22
Air Blank		0.000	23:22
Control Test		0.079	23:23
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 03/24/2018 03:02:17

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.176	02:57
Breath Volume		2.015	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.163	03:00
Breath Volume		1.410	03:00
Air Blank		0.000	03:01
Control Test		0.080	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SMALLWOOD,ELLIS
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: LACEY

DATE/TIME OF TEST: 03/25/2018 19:13:52

Diagnostics Check		OK	19:07
Air Blank		0.000	19:08
Control Test		0.078	19:08
Air Blank		0.000	19:09
Subject Sample	# 1	0.129	19:09
Breath Volume		3.664	19:09
Air Blank		0.000	19:10
Air Blank		0.000	19:12
Subject Sample	# 2	0.123	19:12
Breath Volume		2.933	19:12
Air Blank		0.000	19:12
Control Test		0.078	19:13
Air Blank		0.000	19:13
Diagnostics Check		OK	19:13

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 03/25/2018 22:19:03

Diagnostics Check		OK	22:12
Air Blank		0.000	22:13
Control Test		0.079	22:13
Air Blank		0.000	22:14
Subject Sample	# 1	0.229	22:14
Breath Volume		2.230	22:14
Air Blank		0.000	22:15
Air Blank		0.000	22:17
Subject Sample	# 2	0.233	22:17
Breath Volume		1.421	22:17
Air Blank		0.000	22:18
Control Test		0.079	22:18
Air Blank		0.000	22:18
Diagnostics Check		OK	22:19

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: COATES

DATE/TIME OF TEST: 03/26/2018 01:51:57

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.080	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.000	01:47
Breath Volume		2.152	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.000	01:50
Breath Volume		1.449	01:50
Air Blank		0.000	01:51
Control Test		0.079	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESILVA,JARRETT,B
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: DESILVA,JARRETT
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 03/27/2018 02:12:07

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:00	
Control Test		0.079	02:01	
Air Blank		0.000	02:01	
Subject Sample	# 1	0.079	02:04	Volume Not Met
Breath Volume		0.015	02:04	
Air Blank		0.000	02:05	
Air Blank		0.000	02:07	
Subject Sample	# 2	0.177	02:10	Volume Not Met
Breath Volume		0.238	02:10	
Air Blank		0.000	02:11	
Control Test		0.078	02:11	
Air Blank		0.000	02:12	
Diagnostics Check		OK	02:12	

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BYRD,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: STARNES

DATE/TIME OF TEST: 03/29/2018 02:34:16

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.079	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.102	02:29
Breath Volume		1.632	02:29
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.109	02:32
Breath Volume		1.292	02:32
Air Blank		0.000	02:33
Control Test		0.079	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 03/29/2018 09:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNDON,DAVID,E
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: HERNDON,DAVID
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: VICKERY

DATE/TIME OF TEST: 03/06/2018 15:47:40

Diagnostics Check		OK	15:41
Air Blank		0.000	15:41
Control Test		0.079	15:42
Air Blank		0.000	15:42
Subject Sample	# 1	0.190	15:42
Breath Volume		1.816	15:42
Air Blank		0.000	15:43
Air Blank		0.000	15:45
Subject Sample	# 2	0.184	15:46
Breath Volume		2.027	15:46
Air Blank		0.000	15:46
Control Test		0.078	15:47
Air Blank		0.000	15:47
Diagnostics Check		OK	15:47

Data Download Date/Time: 04/05/2018 20:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYLE JR,HARRY,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HUYKE,ANGELTHON
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/11/2018 03:01:40

Diagnostics Check		OK	01:53
Air Blank		0.000	01:54
Control Test		0.079	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.224	01:56
Breath Volume		1.652	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.218	03:00
Breath Volume		1.492	03:00
Air Blank		0.000	03:00
Control Test		0.078	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 04/05/2018 20:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COWAN,RICHARD,T
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: BURRELL,CODY
ARRESTING OFFICER AGENCY: Graceville PD
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 03/19/2018 02:47:29

Diagnostics Check		OK	02:41
Air Blank		0.000	02:42
Control Test		0.079	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.049	02:43
Breath Volume		2.667	02:43
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.045	02:45
Breath Volume		2.660	02:45
Air Blank		0.000	02:46
Control Test		0.078	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 04/05/2018 20:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COWAN,RICHARD,T
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: THOMAS,JON
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MOUNT

DATE/TIME OF TEST: 03/19/2018 04:52:29

Diagnostics Check		OK	04:45
Air Blank		0.000	04:46
Control Test		0.079	04:46
Air Blank		0.000	04:47
Subject Sample	# 1	0.102	04:47
Breath Volume		1.339	04:47
Air Blank		0.000	04:48
Air Blank		0.000	04:50
Subject Sample	# 2	0.099	04:50
Breath Volume		1.246	04:50
Air Blank		0.000	04:51
Control Test		0.077	04:51
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 04/05/2018 20:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 03/10/2018 20:23:16

Diagnostics Check		OK	20:17
Air Blank		0.000	20:17
Control Test		0.079	20:17
Air Blank		0.000	20:18
Subject Sample	# 1	0.293	20:18
Breath Volume		2.738	20:18
Air Blank		0.000	20:19
Air Blank		0.000	20:21
Subject Sample	# 2	0.292	20:21
Breath Volume		2.460	20:21
Air Blank		0.000	20:22
Control Test		0.079	20:22
Air Blank		0.000	20:23
Diagnostics Check		OK	20:23

Data Download Date/Time: 03/27/2018 09:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,MIGUEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ROBLES,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: ADDISON

DATE/TIME OF TEST: 03/10/2018 23:53:24

Diagnostics Check		OK	23:46
Air Blank		0.000	23:47
Control Test		0.078	23:47
Air Blank		0.000	23:48
Subject Sample	# 1	0.125	23:48
Breath Volume		2.355	23:48
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.117	23:51
Breath Volume		1.183	23:51
Air Blank		0.000	23:52
Control Test		0.080	23:52
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 03/27/2018 09:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LAKE,JUSTIN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: SHEPARD

DATE/TIME OF TEST: 03/11/2018 05:24:51

Diagnostics Check		OK	05:18
Air Blank		0.000	05:18
Control Test		0.080	05:18
Air Blank		0.000	05:19
Subject Sample	# 1	0.146	05:19
Breath Volume		1.480	05:19
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.153	05:23
Breath Volume		2.542	05:23
Air Blank		0.000	05:23
Control Test		0.078	05:24
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 03/27/2018 09:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 03/16/2018 01:01:25

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.079	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.100	00:57
Breath Volume		2.824	00:57
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.103	00:59
Breath Volume		2.714	00:59
Air Blank		0.000	01:00
Control Test		0.079	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 03/27/2018 09:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHARDSON,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: BEVILLE,TRAVIS
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: NOBLES JR

DATE/TIME OF TEST: 03/17/2018 04:45:37

Diagnostics Check		OK	04:39
Air Blank		0.000	04:40
Control Test		0.079	04:40
Air Blank		0.000	04:40
Subject Sample	# 1	0.072	04:41
Breath Volume		3.222	04:41
Air Blank		0.000	04:41
Air Blank		0.000	04:43
Subject Sample	# 2	0.071	04:44
Breath Volume		3.214	04:44
Air Blank		0.000	04:44
Control Test		0.078	04:44
Air Blank		0.000	04:45
Diagnostics Check		OK	04:45

Data Download Date/Time: 03/27/2018 09:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 03/17/2018 23:39:36

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:38	
Control Test	0.079	23:39	
Air Blank		23:39	RFI Detect
Air Blank		23:39	RFI Detect

Data Download Date/Time: 03/27/2018 09:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 03/17/2018 23:47:39

Diagnostics Check		OK	23:41
Air Blank		0.000	23:41
Control Test		0.080	23:42
Air Blank		0.000	23:42
Subject Sample	# 1	0.197	23:43
Breath Volume		1.605	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:45
Subject Sample	# 2	0.198	23:46
Breath Volume		1.898	23:46
Air Blank		0.000	23:46
Control Test		0.079	23:47
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 03/27/2018 09:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANKAM,DENZEL,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WENDT,HENDRIK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: STABILE

DATE/TIME OF TEST: 03/18/2018 03:02:51

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.080	02:57
Air Blank		0.000	02:57
Subject Sample	# 1	0.097	02:58
Breath Volume		3.585	02:58
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.097	03:01
Breath Volume		3.679	03:01
Air Blank		0.000	03:01
Control Test		0.078	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 03/27/2018 09:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,JAVON,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MURPHY,JAVON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: POWERS

DATE/TIME OF TEST: 03/22/2018 00:12:26

Diagnostics Check		OK	00:06
Air Blank		0.000	00:06
Control Test		0.080	00:06
Air Blank		0.000	00:07
Subject Sample	# 1	0.031	00:07
Breath Volume		1.957	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.030	00:10
Breath Volume		1.890	00:10
Air Blank		0.000	00:11
Control Test		0.080	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 03/27/2018 09:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: ARGENTINO

DATE/TIME OF TEST: 03/23/2018 22:30:43

Diagnostics Check		OK	22:22	
Air Blank		0.000	22:22	
Control Test		0.080	22:22	
Air Blank		0.000	22:23	
Subject Sample	# 1	0.000	22:26	Volume Not Met
Breath Volume		0.628	22:26	
Air Blank		0.000	22:26	
Air Blank		0.000	22:28	
Subject Sample	# 2		22:29	Subject Test Refused
Breath Volume		0.000	22:29	
Air Blank		0.000	22:29	
Control Test		0.079	22:30	
Air Blank		0.000	22:30	
Diagnostics Check		OK	22:30	

Data Download Date/Time: 03/27/2018 09:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 03/25/2018 02:19:19

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.080	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.195	02:14
Breath Volume		1.648	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.193	02:17
Breath Volume		2.093	02:17
Air Blank		0.000	02:18
Control Test		0.078	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 03/27/2018 09:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHARDSON,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: BEVILLE,TRAVIS
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: JOHN

DATE/TIME OF TEST: 03/28/2018 02:01:36

Diagnostics Check		OK	01:54
Air Blank		0.000	01:55
Control Test		0.079	01:55
Air Blank		0.000	01:56
Subject Sample	# 1	0.252	01:56
Breath Volume		2.093	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.237	01:59
Breath Volume		1.378	01:59
Air Blank		0.000	02:00
Control Test		0.078	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 04/23/2018 09:41

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,VICTOR,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACKSON,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 07:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:02

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 03/08/2018 09:28:26

Diagnostics Check		OK	09:22
Air Blank		0.000	09:22
Control Test		0.082	09:23
Air Blank		0.000	09:23
Subject Sample	# 1	0.000	09:24
Breath Volume		2.847	09:24
Air Blank		0.000	09:24
Air Blank		0.000	09:26
Subject Sample	# 2	0.000	09:27
Breath Volume		3.945	09:27
Air Blank		0.000	09:27
Control Test		0.081	09:27
Air Blank		0.000	09:28
Diagnostics Check		OK	09:28

Data Download Date/Time: 05/04/2018 18:39

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,ROBERT,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELLEY,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: CULLY

DATE/TIME OF TEST: 03/09/2018 02:04:38

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:57	
Control Test		0.079	01:58	
Air Blank		0.000	01:58	
Subject Sample	# 1	0.189	01:59	
Breath Volume		2.351	01:59	
Air Blank		0.000	02:00	
Air Blank		0.000	02:02	
Subject Sample	# 2		02:03	Subject Test Refused
Breath Volume		0.000	02:03	
Air Blank		0.000	02:03	
Control Test		0.079	02:04	
Air Blank		0.000	02:04	
Diagnostics Check		OK	02:04	

Data Download Date/Time: 05/04/2018 18:39

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONDY,GEOFFREY,C
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: CLARK,ETHAN
ARRESTING OFFICER AGENCY: Lafayette County SO
LAST AGENCY INSPECTION: 03/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:19

SUBJECT'S LAST NAME: LAND

DATE/TIME OF TEST: 03/27/2018 14:49:02

Diagnostics Check		OK	14:40
Air Blank		0.000	14:41
Control Test		0.079	14:41
Air Blank		0.000	14:41
Subject Sample	# 1	0.012	14:43
Breath Volume		1.703	14:43
Air Blank		0.000	14:43
Air Blank		0.000	14:45
Subject Sample	# 2	0.009	14:47
Breath Volume		2.738	14:47
Air Blank		0.000	14:48
Control Test		0.077	14:48
Air Blank		0.000	14:48
Diagnostics Check		OK	14:49

Data Download Date/Time: 05/24/2018 12:01

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY,MATHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY,MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: DORTCH

DATE/TIME OF TEST: 03/04/2018 17:56:04

Diagnostics Check		OK	17:48
Air Blank		0.000	17:49
Control Test		0.078	17:49
Air Blank		0.000	17:49
Subject Sample	# 1	0.189	17:50
Breath Volume		1.355	17:50
Air Blank		0.000	17:50
Air Blank		0.000	17:52
Subject Sample	# 2	0.200	17:54
Breath Volume		1.371	17:54
Air Blank		0.000	17:55
Control Test		0.078	17:55
Air Blank		0.000	17:55
Diagnostics Check		OK	17:56

Data Download Date/Time: 03/31/2018 16:52

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATES,BRIAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BATES,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: TOMBLIN

DATE/TIME OF TEST: 03/06/2018 22:19:31

Diagnostics Check		OK	22:13	
Air Blank		0.000	22:13	
Control Test		0.077	22:14	
Air Blank		0.000	22:14	
Subject Sample	# 1	0.104	22:15	
Breath Volume		2.855	22:15	
Air Blank		0.000	22:15	
Air Blank		0.000	22:17	
Subject Sample	# 2	0.103	22:18	
Breath Volume		3.058	22:18	
Air Blank		0.000	22:18	
Control Test		0.000	22:18	RFI Detect
Air Blank		0.000	22:19	

Data Download Date/Time: 03/31/2018 16:52

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATES,BRIAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BATES,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: TOMBLIN

DATE/TIME OF TEST: 03/06/2018 22:29:05

Diagnostics Check		OK	22:22	
Air Blank		0.000	22:23	
Control Test		0.075	22:23	
Air Blank		0.000	22:24	
Subject Sample	# 1	0.102	22:24	
Breath Volume		1.390	22:24	
Air Blank		0.000	22:25	
Air Blank		0.000	22:27	
Subject Sample	# 2	0.096	22:27	
Breath Volume		2.703	22:27	
Air Blank		0.000	22:28	
Control Test		0.074	22:28	Control Outside Tolerance
Air Blank		0.000	22:29	

Data Download Date/Time: 03/31/2018 16:52

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATES,BRIAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BATES,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: TOMBLIN

DATE/TIME OF TEST: 03/06/2018 22:45:13

Diagnostics Check		OK	22:38	
Air Blank		0.000	22:38	
Control Test		0.076	22:39	
Air Blank		0.000	22:39	
Subject Sample	# 1	0.097	22:40	
Breath Volume		2.585	22:40	
Air Blank		0.000	22:41	
Air Blank		0.000	22:42	
Subject Sample	# 2	0.095	22:43	
Breath Volume		3.015	22:43	
Air Blank		0.000	22:44	
Control Test		0.074	22:44	Control Outside Tolerance
Air Blank		0.000	22:45	

Data Download Date/Time: 03/31/2018 16:52

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLIE,JOHN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TILLIE,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 03/07/2018 06:07:23

Diagnostics Check		OK	06:01
Air Blank		0.000	06:01
Control Test		0.077	06:02
Air Blank		0.000	06:02
Subject Sample	# 1	0.103	06:02
Breath Volume		2.582	06:02
Air Blank		0.000	06:03
Air Blank		0.000	06:05
Subject Sample	# 2	0.103	06:05
Breath Volume		2.457	06:05
Air Blank		0.000	06:06
Control Test		0.076	06:06
Air Blank		0.000	06:07
Diagnostics Check		OK	06:07

Data Download Date/Time: 03/31/2018 16:52

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: USECHE,IVAN,D
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: BASS,ERICA
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/08/2018 01:41:39

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:39	
Control Test		0.078	01:39	
Air Blank		0.000	01:39	
Subject Sample	# 1		01:40	Subject Test Refused
Breath Volume		0.000	01:40	
Air Blank		0.000	01:40	
Control Test		0.078	01:41	
Air Blank		0.000	01:41	
Diagnostics Check		OK	01:41	

Data Download Date/Time: 03/31/2018 16:52

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: COLBERT,BRANDON
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: REGAR

DATE/TIME OF TEST: 03/10/2018 00:39:03

Diagnostics Check		OK	00:30
Air Blank		0.000	00:31
Control Test		0.078	00:31
Air Blank		0.000	00:32
Subject Sample	# 1	0.158	00:33
Breath Volume		1.316	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:35
Subject Sample	# 2	0.141	00:37
Breath Volume		2.089	00:37
Air Blank		0.000	00:38
Control Test		0.075	00:38
Air Blank		0.000	00:38
Diagnostics Check		OK	00:39

Data Download Date/Time: 03/31/2018 16:52

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: MOSLEY,K
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 03/11/2018 03:43:23

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.077	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.085	03:39
Breath Volume		3.886	03:39
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.082	03:41
Breath Volume		2.984	03:41
Air Blank		0.000	03:42
Control Test		0.076	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 03/31/2018 16:52

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIEGEL,TMOTHY,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RIEGEL,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: JANSEN

DATE/TIME OF TEST: 03/14/2018 02:39:05

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.078	02:33
Air Blank		0.000	02:34
Subject Sample	# 1	0.204	02:34
Breath Volume		1.738	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.201	02:37
Breath Volume		1.777	02:37
Air Blank		0.000	02:38
Control Test		0.076	02:38
Air Blank		0.000	02:38
Diagnostics Check		OK	02:39

Data Download Date/Time: 03/31/2018 16:52

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSGROVE,DUSTIN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MUSGROVE,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: CHILDS

DATE/TIME OF TEST: 03/19/2018 23:21:16

Diagnostics Check		OK	23:18	
Air Blank		0.000	23:18	
Control Test		0.076	23:19	
Air Blank		0.000	23:19	
Subject Sample	# 1		23:20	Sequence Aborted
Breath Volume		0.000	23:20	
Air Blank		0.000	23:21	

Data Download Date/Time: 03/31/2018 16:52

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: CREWS,MICHAEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: ZERO TOL
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/30/2018 18:58:08

Diagnostics Check		OK	18:51
Air Blank		0.000	18:52
Control Test		0.077	18:52
Air Blank		0.000	18:52
Subject Sample	# 1	0.129	18:53
Breath Volume		1.386	18:53
Air Blank		0.000	18:53
Air Blank		0.000	18:55
Subject Sample	# 2	0.132	18:56
Breath Volume		2.394	18:56
Air Blank		0.000	18:57
Control Test		0.076	18:57
Air Blank		0.000	18:58
Diagnostics Check		OK	18:58

Data Download Date/Time: 03/31/2018 16:52

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: ROBINSON,STUART
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 03/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: HOLLIE

DATE/TIME OF TEST: 03/31/2018 20:00:45

Diagnostics Check		OK	19:57	
Air Blank		0.000	19:57	
Control Test		0.077	19:58	
Air Blank		0.000	19:58	
Subject Sample	# 1		19:59	Subject Test Refused
Breath Volume		0.000	19:59	
Air Blank		0.000	19:59	
Control Test		0.076	20:00	
Air Blank		0.000	20:00	
Diagnostics Check		OK	20:00	

Data Download Date/Time: 04/30/2018 01:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

INSTRUMENT REGISTERED TO: NAS JACKSONVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUSINO,LUIS,H
BREATH TEST OPERATOR AGENCY: NAS Jacksonville PD
ARRESTING OFFICER: MEDINA,MERCEDES
ARRESTING OFFICER AGENCY: NAS Jacksonville PD
LAST AGENCY INSPECTION: 02/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ELLISON

DATE/TIME OF TEST: 03/05/2018 00:22:40

Diagnostics Check		OK	00:07	
Air Blank		0.000	00:08	
Control Test		0.079	00:08	
Air Blank		0.000	00:09	
Subject Sample	# 1	0.188	00:12	Volume Not Met
Breath Volume		0.117	00:12	
Air Blank		0.000	00:13	
Air Blank		0.000	00:14	
Subject Sample	# 2	0.175	00:16	
Breath Volume		2.101	00:16	
Air Blank		0.000	00:17	
Air Blank		0.000	00:19	
Subject Sample	# 3	0.190	00:20	
Breath Volume		1.167	00:20	
Air Blank		0.000	00:21	
Control Test		0.075	00:21	
Air Blank		0.000	00:22	
Diagnostics Check		OK	00:22	

Data Download Date/Time: 03/29/2018 03:09

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADFORD,CHARLES,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WADFORD,CHARLES
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HOLDEN

DATE/TIME OF TEST: 03/02/2018 02:46:57

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.082	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.107	02:42
Breath Volume		1.593	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:44
Subject Sample	# 2	0.116	02:45
Breath Volume		2.644	02:45
Air Blank		0.000	02:46
Control Test		0.083	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 03/07/2018 14:57

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENSON,DAVID,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WATSON,CLAYTON
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: MINCEY

DATE/TIME OF TEST: 03/04/2018 00:36:58

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.083	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.036	00:32
Breath Volume		4.324	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:34
Subject Sample	# 2	0.032	00:35
Breath Volume		2.242	00:35
Air Blank		0.000	00:36
Control Test		0.084	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 03/07/2018 14:57

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRS,RUSSELL,B
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: BROOKS,JAKE
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/04/2018 02:49:45

Diagnostics Check		OK	02:46	
Air Blank		0.000	02:46	
Control Test		0.082	02:47	
Air Blank		0.000	02:47	
Subject Sample	# 1		02:48	Subject Test Refused
Breath Volume		0.000	02:48	
Air Blank		0.000	02:48	
Control Test		0.083	02:49	
Air Blank		0.000	02:49	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 03/07/2018 14:57

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRS,RUSSELL,B
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TOMPKINS,CHARLES
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 03/04/2018 04:28:23

Diagnostics Check		OK	04:21
Air Blank		0.000	04:22
Control Test		0.083	04:22
Air Blank		0.000	04:23
Subject Sample	# 1	0.103	04:23
Breath Volume		1.972	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.098	04:26
Breath Volume		2.859	04:26
Air Blank		0.000	04:27
Control Test		0.082	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 03/07/2018 14:57

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIEGEL,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RIEGEL,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SHEPPARD

DATE/TIME OF TEST: 03/17/2018 23:20:40

Diagnostics Check		OK	23:13	
Air Blank		0.000	23:14	
Control Test		0.083	23:14	
Air Blank		0.000	23:15	
Subject Sample	# 1	0.136	23:15	
Breath Volume		2.246	23:15	
Air Blank		0.000	23:16	
Air Blank		0.000	23:17	
Subject Sample	# 2	0.127	23:19	
Breath Volume		2.609	23:19	
Air Blank			23:19	Purge Fail
Air Blank		0.000	23:20	

Data Download Date/Time: 04/30/2018 23:29

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALEXIS,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HERNANDEZ,ALEXIS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 03/18/2018 03:58:08

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.079	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.127	03:53
Breath Volume		2.707	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.130	03:56
Breath Volume		2.234	03:56
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 04/30/2018 23:29

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: LEE,WILLIAM
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: GOLENIEWSKI

DATE/TIME OF TEST: 03/24/2018 21:32:16

Diagnostics Check		OK	21:29	
Air Blank		0.000	21:29	
Control Test		0.084	21:30	
Air Blank		0.000	21:30	
Subject Sample	# 1		21:30	Subject Test Refused
Breath Volume		0.000	21:30	
Air Blank		0.000	21:31	
Control Test		0.084	21:31	
Air Blank		0.000	21:32	
Diagnostics Check		OK	21:32	

Data Download Date/Time: 04/30/2018 23:29

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COWAN,RICHARD,T
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: ARROYO,BRANDON
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 03/20/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 03/27/2018 00:51:38

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.077	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.238	00:47
Breath Volume		2.132	00:47
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.225	00:50
Breath Volume		1.718	00:50
Air Blank		0.000	00:50
Control Test		0.081	00:50
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 04/05/2018 20:16

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: STEDGE

DATE/TIME OF TEST: 03/18/2018 20:57:24

Diagnostics Check		OK	20:54	
Air Blank		0.000	20:54	
Control Test		0.081	20:55	
Air Blank		0.000	20:55	
Subject Sample	# 1		20:56	Subject Test Refused
Breath Volume		0.000	20:56	
Air Blank		0.000	20:56	
Control Test		0.081	20:56	
Air Blank		0.000	20:57	
Diagnostics Check		OK	20:57	

Data Download Date/Time: 04/05/2018 20:06

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COWAN,RICHARD,T
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: COWAN,RICHARD
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: BISS

DATE/TIME OF TEST: 03/03/2018 22:16:47

Diagnostics Check		OK	22:10
Air Blank		0.000	22:10
Control Test		0.082	22:10
Air Blank		0.000	22:11
Subject Sample	# 1	0.138	22:11
Breath Volume		1.218	22:11
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.131	22:14
Breath Volume		2.164	22:14
Air Blank		0.000	22:15
Control Test		0.079	22:16
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 04/02/2018 23:07

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD,MATTHEW,E
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: MAYFIELD,MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: MCCAIN

DATE/TIME OF TEST: 03/10/2018 21:13:03

Diagnostics Check		OK	21:06
Air Blank		0.000	21:06
Control Test		0.080	21:07
Air Blank		0.000	21:07
Subject Sample	# 1	0.307	21:08
Breath Volume		1.867	21:08
Air Blank		0.000	21:09
Air Blank		0.000	21:10
Subject Sample	# 2	0.299	21:11
Breath Volume		1.652	21:11
Air Blank		0.000	21:12
Control Test		0.077	21:12
Air Blank		0.000	21:12
Diagnostics Check		OK	21:13

Data Download Date/Time: 04/02/2018 23:07

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,BLAKE,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BROOKS,BLAKE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: MCDONNELL

DATE/TIME OF TEST: 03/11/2018 03:06:08

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.082	01:59
Air Blank		0.000	01:59
Subject Sample	# 1	0.102	03:01
Breath Volume		1.144	03:01
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.098	03:04
Breath Volume		1.644	03:04
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 04/02/2018 23:07

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: CUZNER

DATE/TIME OF TEST: 03/13/2018 00:03:09

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:00	
Control Test		0.083	00:00	
Air Blank		0.000	00:01	
Subject Sample	# 1		00:01	RFI Detect
Breath Volume		0.000	00:01	
Air Blank		0.000	00:02	
Control Test		0.083	00:02	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 04/02/2018 23:07

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: CUZNER

DATE/TIME OF TEST: 03/13/2018 00:09:44

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:06	
Control Test		0.083	00:06	
Air Blank		0.000	00:06	
Subject Sample	# 1	0.106	00:08	
Breath Volume		1.398	00:08	
Air Blank			00:09	Purge Fail
Air Blank		0.000	00:09	

Data Download Date/Time: 04/02/2018 23:07

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: CUZNER

DATE/TIME OF TEST: 03/13/2018 00:15:04

Diagnostics Check		OK	00:12	
Air Blank		0.000	00:12	
Control Test		0.077	00:12	
Air Blank		0.000	00:13	
Subject Sample	# 1		00:13	RFI Detect
Breath Volume		0.000	00:13	
Air Blank		0.000	00:14	
Control Test		0.080	00:14	
Air Blank		0.000	00:14	
Diagnostics Check		OK	00:15	

Data Download Date/Time: 04/02/2018 23:07

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: CUZNER

DATE/TIME OF TEST: 03/13/2018 00:24:42

Diagnostics Check		OK	00:17
Air Blank		0.000	00:18
Control Test		0.080	00:18
Air Blank		0.000	00:18
Subject Sample	# 1	0.104	00:19
Breath Volume		1.785	00:19
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.092	00:22
Breath Volume		1.445	00:22
Air Blank		0.000	00:23
Control Test		0.077	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 04/02/2018 23:07

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD,MATTHEW,E
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: MAYFIELD,MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: ALMUTAIRY

DATE/TIME OF TEST: 03/14/2018 20:11:18

Diagnostics Check		OK	19:55	
Air Blank		0.000	19:55	
Control Test		0.082	19:56	
Air Blank		0.000	19:56	
Subject Sample	# 1	0.132	19:59	Volume Not Met
Breath Volume		0.222	19:59	
Air Blank		0.000	20:00	
Air Blank		0.000	20:02	
Subject Sample	# 2	0.168	20:03	
Breath Volume		1.414	20:03	
Air Blank		0.000	20:04	
Air Blank		0.000	20:06	
Subject Sample	# 3	0.156	20:09	Volume Not Met
Breath Volume		0.054	20:09	
Air Blank		0.000	20:10	
Control Test		0.080	20:10	
Air Blank		0.000	20:11	
Diagnostics Check		OK	20:11	

Data Download Date/Time: 04/02/2018 23:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,BLAKE,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BROOKS,BLAKE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: SOWDER

DATE/TIME OF TEST: 03/15/2018 00:27:48

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.082	00:21
Air Blank		0.000	00:22
Subject Sample	# 1	0.163	00:23
Breath Volume		2.214	00:23
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.156	00:25
Breath Volume		1.828	00:25
Air Blank		0.000	00:26
Control Test		0.079	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 04/02/2018 23:07

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: DEMO

DATE/TIME OF TEST: 03/17/2018 02:35:25

Diagnostics Check		OK	02:32	
Air Blank		0.000	02:32	
Control Test		0.082	02:33	
Air Blank		0.000	02:33	
Subject Sample	# 1		02:34	RFI Detect
Breath Volume		0.000	02:34	
Air Blank		0.000	02:34	
Control Test		0.082	02:34	
Air Blank		0.000	02:35	
Diagnostics Check		OK	02:35	

Data Download Date/Time: 04/02/2018 23:07

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: DEMO

DATE/TIME OF TEST: 03/17/2018 02:44:55

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.082	02:38
Air Blank		0.000	02:38
Subject Sample	# 1	0.085	02:39
Breath Volume		1.191	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.087	02:43
Breath Volume		1.566	02:43
Air Blank		0.000	02:43
Control Test		0.081	02:44
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 04/02/2018 23:07

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,BLAKE,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BROOKS,BLAKE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: DAVEY

DATE/TIME OF TEST: 03/17/2018 23:31:13

Diagnostics Check		OK	23:26	
Air Blank		0.000	23:26	
Control Test		0.082	23:26	
Air Blank		0.000	23:27	
Subject Sample	# 1	0.249	23:29	
Breath Volume		1.593	23:29	
Air Blank			23:30	Purge Fail
Air Blank		0.000	23:31	

Data Download Date/Time: 04/02/2018 23:07

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,BLAKE,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BROOKS,BLAKE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: DAVEY

DATE/TIME OF TEST: 03/17/2018 23:39:44

Diagnostics Check		OK	23:32
Air Blank		0.000	23:32
Control Test		0.076	23:33
Air Blank		0.000	23:33
Subject Sample	# 1	0.243	23:34
Breath Volume		1.425	23:34
Air Blank		0.000	23:35
Air Blank		0.000	23:36
Subject Sample	# 2	0.249	23:37
Breath Volume		2.136	23:37
Air Blank		0.000	23:38
Control Test		0.077	23:39
Air Blank		0.000	23:39
Diagnostics Check		OK	23:39

Data Download Date/Time: 04/02/2018 23:07

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: MOERKE

DATE/TIME OF TEST: 03/18/2018 01:38:11

Diagnostics Check		OK	01:29
Air Blank		0.000	01:29
Control Test		0.080	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.133	01:31
Breath Volume		1.269	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:33
Subject Sample	# 2	0.137	01:36
Breath Volume		1.628	01:36
Air Blank		0.000	01:37
Control Test		0.080	01:37
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 04/02/2018 23:07

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DARYL,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MILLER,DARYL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: LAURO

DATE/TIME OF TEST: 03/22/2018 00:53:16

Diagnostics Check		OK	00:46
Air Blank		0.000	00:47
Control Test		0.084	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.239	00:48
Breath Volume		1.437	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.252	00:51
Breath Volume		1.562	00:51
Air Blank		0.000	00:52
Control Test		0.082	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 04/02/2018 23:07

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: CLYBURN

DATE/TIME OF TEST: 03/27/2018 02:53:09

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:50	
Control Test		0.082	02:51	
Air Blank		0.000	02:51	
Subject Sample	# 1		02:51	Subject Test Refused
Breath Volume		0.000	02:51	
Air Blank		0.000	02:52	
Control Test		0.083	02:52	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 04/02/2018 23:07

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 461951
DRY GAS EXPIRY DATE: 06/15/2019
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 03/27/2018 20:54:14

Diagnostics Check		OK	20:45
Air Blank		0.000	20:46
Control Test		0.083	20:46
Air Blank		0.000	20:47
Subject Sample	# 1	0.162	20:49
Breath Volume		1.781	20:49
Air Blank		0.000	20:50
Air Blank		0.000	20:51
Subject Sample	# 2	0.157	20:52
Breath Volume		1.718	20:52
Air Blank		0.000	20:53
Control Test		0.077	20:53
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 04/02/2018 23:07

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: JOYNER BARNES

DATE/TIME OF TEST: 03/04/2018 03:39:39

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.080	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.160	03:34
Breath Volume		3.898	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.163	03:38
Breath Volume		3.914	03:38
Air Blank		0.000	03:38
Control Test		0.081	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 03/10/2018 11:29

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,JOHN,H
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: HUNTER,JOHN
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: PEOPLES

DATE/TIME OF TEST: 03/04/2018 19:09:16

Diagnostics Check		OK	19:03
Air Blank		0.000	19:03
Control Test		0.080	19:03
Air Blank		0.000	19:04
Subject Sample	# 1	0.184	19:04
Breath Volume		2.007	19:04
Air Blank		0.000	19:05
Air Blank		0.000	19:07
Subject Sample	# 2	0.180	19:07
Breath Volume		1.820	19:07
Air Blank		0.000	19:08
Control Test		0.081	19:08
Air Blank		0.000	19:09
Diagnostics Check		OK	19:09

Data Download Date/Time: 03/10/2018 11:29

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEFF,STEPHEN,
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: NEFF,STEPHEN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: BUNSEE

DATE/TIME OF TEST: 03/07/2018 20:38:28

Diagnostics Check		OK	20:32	
Air Blank		0.000	20:32	
Control Test		0.081	20:33	
Air Blank		0.000	20:33	
Subject Sample	# 1	0.168	20:34	
Breath Volume		1.289	20:34	
Air Blank		0.000	20:34	
Air Blank		0.000	20:36	
Subject Sample	# 2	0.168	20:37	
Breath Volume		2.566	20:37	
Air Blank		0.000	20:37	
			20:37	RFI Detect
Air Blank		0.000	20:38	

Data Download Date/Time: 03/10/2018 11:29

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEFF,STEPHEN,
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: BUNSEE

DATE/TIME OF TEST: 03/07/2018 20:46:23

Diagnostics Check		OK	20:40
Air Blank		0.000	20:40
Control Test		0.080	20:40
Air Blank		0.000	20:41
Subject Sample	# 1	0.168	20:42
Breath Volume		2.390	20:42
Air Blank		0.000	20:42
Air Blank		0.000	20:44
Subject Sample	# 2	0.174	20:44
Breath Volume		2.035	20:44
Air Blank		0.000	20:45
Control Test		0.081	20:45
Air Blank		0.000	20:46
Diagnostics Check		OK	20:46

Data Download Date/Time: 03/10/2018 11:29

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 03/09/2018 01:11:27

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.080	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.137	01:07
Breath Volume		2.656	01:07
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.142	01:09
Breath Volume		3.035	01:09
Air Blank		0.000	01:10
Control Test		0.081	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 03/10/2018 11:29

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: NEFF,STEFF
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 03/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 03/10/2018 21:50:11

Diagnostics Check		OK	21:47	
Air Blank		0.000	21:47	
Control Test		0.080	21:48	
Air Blank		0.000	21:48	
Subject Sample	# 1		21:48	Subject Test Refused
Breath Volume		0.000	21:48	
Air Blank		0.000	21:49	
Control Test		0.080	21:49	
Air Blank		0.000	21:50	
Diagnostics Check		OK	21:50	

Data Download Date/Time: 04/16/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 03/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MCMANUS

DATE/TIME OF TEST: 03/17/2018 03:47:17

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.078	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.184	03:42
Breath Volume		3.375	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.187	03:45
Breath Volume		3.035	03:45
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 04/16/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: CASTALDO,JOHN
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 03/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: WEEKLEY

DATE/TIME OF TEST: 03/17/2018 22:25:38

Diagnostics Check		OK	22:19
Air Blank		0.000	22:19
Control Test		0.080	22:20
Air Blank		0.000	22:20
Subject Sample	# 1	0.224	22:21
Breath Volume		2.140	22:21
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.216	22:24
Breath Volume		1.601	22:24
Air Blank		0.000	22:24
Control Test		0.079	22:25
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 04/16/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIDD,JERMEL,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WHITE,TRAVIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 03/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: HIPPARD

DATE/TIME OF TEST: 03/19/2018 20:03:29

Diagnostics Check		OK	20:00	
Air Blank		0.000	20:01	
Control Test		0.079	20:01	
Air Blank		0.000	20:01	
Subject Sample	# 1		20:02	Subject Test Refused
Breath Volume		0.000	20:02	
Air Blank		0.000	20:02	
Control Test		0.079	20:02	
Air Blank		0.000	20:03	
Diagnostics Check		OK	20:03	

Data Download Date/Time: 04/16/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIERS,BLAINE,T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WEIERS,BLAINE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 03/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: BOZEMAN

DATE/TIME OF TEST: 03/21/2018 20:05:51

Diagnostics Check		OK	19:56
Air Blank		0.000	19:56
Control Test		0.081	19:56
Air Blank		0.000	19:57
Subject Sample	# 1	0.114	19:59
Breath Volume		1.640	19:59
Air Blank		0.000	19:59
Air Blank		0.000	20:01
Subject Sample	# 2	0.119	20:04
Breath Volume		2.027	20:04
Air Blank		0.000	20:04
Control Test		0.083	20:05
Air Blank		0.000	20:05
Diagnostics Check		OK	20:05

Data Download Date/Time: 04/16/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIDD,JERMEL,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WHITE,TRAVIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 03/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: SRIPROMMA

DATE/TIME OF TEST: 03/24/2018 06:15:31

Diagnostics Check		OK	06:09
Air Blank		0.000	06:09
Control Test		0.080	06:09
Air Blank		0.000	06:10
Subject Sample	# 1	0.122	06:11
Breath Volume		2.406	06:11
Air Blank		0.000	06:11
Air Blank		0.000	06:13
Subject Sample	# 2	0.124	06:14
Breath Volume		3.437	06:14
Air Blank		0.000	06:14
Control Test		0.080	06:14
Air Blank		0.000	06:15
Diagnostics Check		OK	06:15

Data Download Date/Time: 04/16/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIDD,JERMEL,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: KIDD,JERMEL
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 03/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: KIRCHHEVEL

DATE/TIME OF TEST: 03/28/2018 21:36:01

Diagnostics Check		OK	21:30
Air Blank		0.000	21:30
Control Test		0.080	21:30
Air Blank		0.000	21:31
Subject Sample	# 1	0.000	21:31
Breath Volume		2.375	21:31
Air Blank		0.000	21:32
Air Blank		0.000	21:34
Subject Sample	# 2	0.000	21:34
Breath Volume		3.304	21:34
Air Blank		0.000	21:35
Control Test		0.081	21:35
Air Blank		0.000	21:35
Diagnostics Check		OK	21:36

Data Download Date/Time: 04/16/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 03/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: GRACE

DATE/TIME OF TEST: 03/31/2018 20:48:04

Diagnostics Check		OK	20:40
Air Blank		0.000	20:41
Control Test		0.079	20:41
Air Blank		0.000	20:42
Subject Sample	# 1	0.208	20:42
Breath Volume		1.339	20:42
Air Blank		0.000	20:43
Air Blank		0.000	20:45
Subject Sample	# 2	0.209	20:46
Breath Volume		1.109	20:46
Air Blank		0.000	20:47
Control Test		0.081	20:47
Air Blank		0.000	20:47
Diagnostics Check		OK	20:48

Data Download Date/Time: 04/16/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 03/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: STACHOWICZ

DATE/TIME OF TEST: 03/31/2018 23:46:44

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.080	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.138	23:42
Breath Volume		2.882	23:42
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.134	23:45
Breath Volume		2.640	23:45
Air Blank		0.000	23:45
Control Test		0.080	23:46
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 04/16/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIERS,BLAINE,T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WEIERS,BLAINE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 03/10/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:54

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 03/01/2018 18:21:10

Diagnostics Check		OK	18:18	
Air Blank		0.000	18:18	
Control Test		0.080	18:19	
Air Blank		0.000	18:19	
Subject Sample	# 1		18:19	Subject Test Refused
Breath Volume		0.000	18:19	
Air Blank		0.000	18:20	
Control Test		0.080	18:20	
Air Blank		0.000	18:21	
Diagnostics Check		OK	18:21	

Data Download Date/Time: 04/16/2018 10:45

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WALTENBURG,MARCUS
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 03/02/2018 16:40:11

Diagnostics Check		OK	16:28	
Air Blank		0.000	16:28	
Control Test		0.081	16:29	
Air Blank		0.000	16:29	
Subject Sample	# 1		16:32	No Sample Provided
Breath Volume		0.000	16:32	
Air Blank		0.000	16:33	
Air Blank		0.000	16:35	
Subject Sample	# 2	0.253	16:38	Volume Not Met
Breath Volume		0.066	16:38	
Air Blank		0.000	16:39	
Control Test		0.078	16:39	
Air Blank		0.000	16:40	
Diagnostics Check		OK	16:40	

Data Download Date/Time: 03/29/2018 08:23

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ANTON,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: EVOY

DATE/TIME OF TEST: 03/02/2018 23:42:35

Diagnostics Check		OK	23:34
Air Blank		0.000	23:35
Control Test		0.080	23:35
Air Blank		0.000	23:35
Subject Sample	# 1	0.228	23:37
Breath Volume		1.277	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.238	23:41
Breath Volume		1.214	23:41
Air Blank		0.000	23:41
Control Test		0.079	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 03/29/2018 08:23

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 03/07/2018 01:03:30

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.199	00:58
Breath Volume		3.160	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.196	01:01
Breath Volume		2.816	01:01
Air Blank		0.000	01:02
Control Test		0.079	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 03/29/2018 08:23

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: CURRICA

DATE/TIME OF TEST: 03/08/2018 05:09:10

Diagnostics Check		OK	05:02
Air Blank		0.000	05:03
Control Test		0.081	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.128	05:04
Breath Volume		2.484	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.129	05:07
Breath Volume		2.500	05:07
Air Blank		0.000	05:08
Control Test		0.080	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 03/29/2018 08:23

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: FRAZIER,MICHAEL
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: RULAPAUGH

DATE/TIME OF TEST: 03/11/2018 22:43:21

Diagnostics Check		OK	22:36
Air Blank		0.000	22:36
Control Test		0.080	22:37
Air Blank		0.000	22:37
Subject Sample	# 1	0.291	22:38
Breath Volume		2.429	22:38
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.286	22:41
Breath Volume		1.671	22:41
Air Blank		0.000	22:42
Control Test		0.078	22:42
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 03/29/2018 08:23

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,JAMES,B
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PAYNE,JAMES
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: BUI
ARREST TIME: 12:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: ROBY

DATE/TIME OF TEST: 03/14/2018 14:18:21

Diagnostics Check		OK	14:10
Air Blank		0.000	14:10
Control Test		0.081	14:11
Air Blank		0.000	14:11
Subject Sample	# 1	0.199	14:12
Breath Volume		1.636	14:12
Air Blank		0.000	14:13
Air Blank		0.000	14:15
Subject Sample	# 2	0.192	14:16
Breath Volume		2.277	14:16
Air Blank		0.000	14:17
Control Test		0.080	14:17
Air Blank		0.000	14:18
Diagnostics Check		OK	14:18

Data Download Date/Time: 03/29/2018 08:23

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODRUFF,MARK,D
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: CLARK,NOAH
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: HAYLING

DATE/TIME OF TEST: 03/15/2018 03:47:47

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.081	03:42
Air Blank		0.000	03:42
Subject Sample	# 1	0.080	03:43
Breath Volume		1.886	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:45
Subject Sample	# 2	0.082	03:46
Breath Volume		1.800	03:46
Air Blank		0.000	03:46
Control Test		0.081	03:47
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 03/29/2018 08:23

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WALTENBURG,MARCUS
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: BOESCH

DATE/TIME OF TEST: 03/15/2018 17:54:03

Diagnostics Check		OK	17:48
Air Blank		0.000	17:48
Control Test		0.081	17:48
Air Blank		0.000	17:49
Subject Sample	# 1	0.174	17:49
Breath Volume		3.773	17:49
Air Blank		0.000	17:50
Air Blank		0.000	17:52
Subject Sample	# 2	0.172	17:52
Breath Volume		2.464	17:52
Air Blank		0.000	17:53
Control Test		0.080	17:53
Air Blank		0.000	17:53
Diagnostics Check		OK	17:54

Data Download Date/Time: 03/29/2018 08:23

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,TANYA,L
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WRIGHT,TANYA
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/15/2018 22:00:45

Diagnostics Check		OK	21:53
Air Blank		0.000	21:54
Control Test		0.081	21:54
Air Blank		0.000	21:55
Subject Sample	# 1	0.219	21:56
Breath Volume		2.968	21:56
Air Blank		0.000	21:56
Air Blank		0.000	21:58
Subject Sample	# 2	0.215	21:59
Breath Volume		3.519	21:59
Air Blank		0.000	21:59
Control Test		0.080	22:00
Air Blank		0.000	22:00
Diagnostics Check		OK	22:00

Data Download Date/Time: 03/29/2018 08:24

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: FRAZIER,MICHAEL
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: LUNA

DATE/TIME OF TEST: 03/18/2018 21:57:03

Diagnostics Check		OK	21:50
Air Blank		0.000	21:51
Control Test		0.079	21:51
Air Blank		0.000	21:51
Subject Sample	# 1	0.138	21:52
Breath Volume		2.343	21:52
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.132	21:55
Breath Volume		1.500	21:55
Air Blank		0.000	21:56
Control Test		0.079	21:56
Air Blank		0.000	21:56
Diagnostics Check		OK	21:57

Data Download Date/Time: 03/29/2018 08:24

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: HELLER

DATE/TIME OF TEST: 03/19/2018 23:46:36

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.080	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.126	23:42
Breath Volume		2.933	23:42
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.126	23:45
Breath Volume		2.175	23:45
Air Blank		0.000	23:45
Control Test		0.080	23:46
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 03/29/2018 08:24

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODRUFF,MARK,D
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: CLARK,NOAH
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 03/20/2018 21:39:34

Diagnostics Check		OK	21:33
Air Blank		0.000	21:33
Control Test		0.080	21:34
Air Blank		0.000	21:34
Subject Sample	# 1	0.070	21:35
Breath Volume		2.144	21:35
Air Blank		0.000	21:35
Air Blank		0.000	21:37
Subject Sample	# 2	0.067	21:38
Breath Volume		1.949	21:38
Air Blank		0.000	21:38
Control Test		0.081	21:38
Air Blank		0.000	21:39
Diagnostics Check		OK	21:39

Data Download Date/Time: 03/29/2018 08:24

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAX,ERIC,C
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BURNS,KYLE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 03/23/2018 04:13:39

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.081	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.231	04:09
Breath Volume		1.289	04:09
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.233	04:12
Breath Volume		1.550	04:12
Air Blank		0.000	04:12
Control Test		0.079	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 03/29/2018 08:24

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER,MATTHEW,T
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GALLAGHER,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: HAIGHT

DATE/TIME OF TEST: 03/23/2018 21:43:52

Diagnostics Check		OK	21:35	
Air Blank		0.000	21:35	
Control Test		0.080	21:36	
Air Blank		0.000	21:36	
Subject Sample	# 1		21:39	No Sample Provided
Breath Volume		0.000	21:39	
Air Blank		0.000	21:40	
Air Blank		0.000	21:42	
Subject Sample	# 2		21:42	Subject Test Refused
Breath Volume		0.000	21:42	
Air Blank		0.000	21:42	
Control Test		0.080	21:43	
Air Blank		0.000	21:43	
Diagnostics Check		OK	21:43	

Data Download Date/Time: 03/29/2018 08:24

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,TANYA,L
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BRENNAN,JOSEPH
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 03/24/2018 23:44:01

Diagnostics Check		OK	23:37
Air Blank		0.000	23:38
Control Test		0.081	23:38
Air Blank		0.000	23:38
Subject Sample	# 1	0.125	23:39
Breath Volume		2.593	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.131	23:42
Breath Volume		2.582	23:42
Air Blank		0.000	23:43
Control Test		0.080	23:43
Air Blank		0.000	23:43
Diagnostics Check		OK	23:44

Data Download Date/Time: 03/29/2018 08:24

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGEL,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: RIDLE,E
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 03/25/2018 05:01:18

Diagnostics Check		OK	04:54
Air Blank		0.000	04:54
Control Test		0.081	04:55
Air Blank		0.000	04:55
Subject Sample	# 1	0.273	04:56
Breath Volume		2.390	04:56
Air Blank		0.000	04:57
Air Blank		0.000	04:59
Subject Sample	# 2	0.264	04:59
Breath Volume		2.425	04:59
Air Blank		0.000	05:00
Control Test		0.078	05:00
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 03/29/2018 08:24

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 03/27/2018 21:33:39

Diagnostics Check		OK	21:27
Air Blank		0.000	21:27
Control Test		0.080	21:27
Air Blank		0.000	21:28
Subject Sample	# 1	0.143	21:29
Breath Volume		3.386	21:29
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 2	0.147	21:32
Breath Volume		3.148	21:32
Air Blank		0.000	21:32
Control Test		0.080	21:33
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 03/29/2018 08:24

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CESAR

DATE/TIME OF TEST: 03/29/2018 02:52:06

Diagnostics Check		OK	02:39	
Air Blank		0.000	02:40	
Control Test		0.080	02:40	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:44	No Sample Provided
Breath Volume		0.000	02:44	
Air Blank		0.000	02:44	
Air Blank		0.000	02:46	
Subject Sample	# 2	0.149	02:47	
Breath Volume		2.632	02:47	
Air Blank		0.000	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 3	0.152	02:50	
Breath Volume		3.644	02:50	
Air Blank		0.000	02:51	
Control Test		0.080	02:51	
Air Blank		0.000	02:52	
Diagnostics Check		OK	02:52	

Data Download Date/Time: 03/29/2018 08:24

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,TANYA,L
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: RIDLE,EVAN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: LEITNAKER

DATE/TIME OF TEST: 03/29/2018 23:27:04

Diagnostics Check		OK	23:20
Air Blank		0.000	23:21
Control Test		0.081	23:21
Air Blank		0.000	23:22
Subject Sample	# 1	0.061	23:22
Breath Volume		3.972	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.056	23:25
Breath Volume		2.566	23:25
Air Blank		0.000	23:26
Control Test		0.080	23:26
Air Blank		0.000	23:26
Diagnostics Check		OK	23:27

Data Download Date/Time: 04/23/2018 07:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAX,ERIC,C
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: STUBLEY,NATHANIEL
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: DEJESUS

DATE/TIME OF TEST: 03/30/2018 22:12:34

Diagnostics Check		OK	22:05
Air Blank		0.000	22:05
Control Test		0.080	22:05
Air Blank		0.000	22:06
Subject Sample	# 1	0.255	22:06
Breath Volume		1.316	22:06
Air Blank		0.000	22:07
Air Blank		0.000	22:09
Subject Sample	# 2	0.235	22:10
Breath Volume		1.828	22:10
Air Blank		0.000	22:11
Control Test		0.079	22:11
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 04/23/2018 07:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAX,ERIC,C
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: STUBLEY,NATHANIEL
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: DEJESUS

DATE/TIME OF TEST: 03/30/2018 22:26:56

Diagnostics Check		OK	22:18
Air Blank		0.000	22:19
Control Test		0.079	22:19
Air Blank		0.000	22:20
Subject Sample	# 1	0.224	22:22
Breath Volume		2.210	22:22
Air Blank		0.000	22:22
Air Blank		0.000	22:24
Subject Sample	# 2	0.243	22:25
Breath Volume		2.117	22:25
Air Blank		0.000	22:26
Control Test		0.078	22:26
Air Blank		0.000	22:26
Diagnostics Check		OK	22:26

Data Download Date/Time: 04/23/2018 07:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BARRERA,JOSE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 03/31/2018 05:05:17

Diagnostics Check		OK	04:58
Air Blank		0.000	04:58
Control Test		0.080	04:59
Air Blank		0.000	04:59
Subject Sample	# 1	0.223	05:00
Breath Volume		1.597	05:00
Air Blank		0.000	05:01
Air Blank		0.000	05:02
Subject Sample	# 2	0.207	05:03
Breath Volume		2.589	05:03
Air Blank		0.000	05:04
Control Test		0.078	05:04
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 04/23/2018 07:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,JESUS,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: ANDRADE,JESUS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 03/31/2018 12:27:37

Diagnostics Check		OK	12:15	
Air Blank		0.000	12:16	
Control Test		0.079	12:16	
Air Blank		0.000	12:17	
Subject Sample	# 1	0.196	12:20	Volume Not Met
Breath Volume		0.402	12:20	
Air Blank		0.000	12:21	
Air Blank		0.000	12:22	
Subject Sample	# 2	0.219	12:25	Volume Not Met
Breath Volume		0.000	12:25	
Air Blank		0.000	12:26	
Control Test		0.078	12:26	
Air Blank		0.000	12:27	
Diagnostics Check		OK	12:27	

Data Download Date/Time: 04/23/2018 07:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,JESUS,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: ANDRADE,JESUS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 03/31/2018 12:32:50

Diagnostics Check		OK	12:29	
Air Blank		0.000	12:29	
Control Test		0.077	12:30	
Air Blank		0.000	12:30	
Subject Sample	# 1		12:32	Sequence Aborted
Breath Volume		0.218	12:32	
Air Blank		0.000	12:32	

Data Download Date/Time: 04/23/2018 07:33

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALEXIS,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HERNANDEZ,ALEXIS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PADGETT-ZANDERS

DATE/TIME OF TEST: 03/04/2018 03:21:06

Diagnostics Check		OK	03:14
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.220	03:16
Breath Volume		2.042	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.220	03:19
Breath Volume		1.441	03:19
Air Blank		0.000	03:20
Control Test		0.079	03:20
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 03/30/2018 01:56

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: EVERETT

DATE/TIME OF TEST: 03/02/2018 01:19:41

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.081	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.168	01:14
Breath Volume		2.156	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.153	01:18
Breath Volume		1.140	01:18
Air Blank		0.000	01:18
Control Test		0.081	01:19
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 03/07/2018 07:13

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/02/2018 04:31:43

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.081	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.181	04:27
Breath Volume		1.343	04:27
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.194	04:30
Breath Volume		1.273	04:30
Air Blank		0.000	04:30
Control Test		0.080	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 03/07/2018 07:13

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: FANNIN

DATE/TIME OF TEST: 03/02/2018 23:00:34

Diagnostics Check		OK	22:53
Air Blank		0.000	22:54
Control Test		0.081	22:54
Air Blank		0.000	22:54
Subject Sample	# 1	0.022	22:56
Breath Volume		2.332	22:56
Air Blank		0.000	22:56
Air Blank		0.000	22:58
Subject Sample	# 2	0.018	22:59
Breath Volume		1.843	22:59
Air Blank		0.000	22:59
Control Test		0.081	22:59
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 03/07/2018 07:13

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: MARINARO

DATE/TIME OF TEST: 03/03/2018 00:15:30

Diagnostics Check		OK	00:12	
Air Blank		0.000	00:12	
Control Test		0.081	00:12	
Air Blank		0.000	00:13	
Subject Sample	# 1		00:14	Subject Test Refused
Breath Volume		0.000	00:14	
Air Blank		0.000	00:14	
Control Test		0.082	00:14	
Air Blank		0.000	00:15	
Diagnostics Check		OK	00:15	

Data Download Date/Time: 03/07/2018 07:13

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: KOUBA

DATE/TIME OF TEST: 03/03/2018 03:03:09

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.082	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.192	02:58
Breath Volume		1.601	02:58
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.181	03:01
Breath Volume		1.714	03:01
Air Blank		0.000	03:02
Control Test		0.081	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 03/07/2018 07:13

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: GABLE

DATE/TIME OF TEST: 03/03/2018 05:01:18

Diagnostics Check		OK	04:55
Air Blank		0.000	04:55
Control Test		0.082	04:55
Air Blank		0.000	04:56
Subject Sample	# 1	0.128	04:56
Breath Volume		1.656	04:56
Air Blank		0.000	04:57
Air Blank		0.000	04:59
Subject Sample	# 2	0.138	04:59
Breath Volume		1.824	04:59
Air Blank		0.000	05:00
Control Test		0.082	05:00
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 03/07/2018 07:13

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: RENTERIA HERRERA

DATE/TIME OF TEST: 03/03/2018 22:05:21

Diagnostics Check		OK	21:59
Air Blank		0.000	21:59
Control Test		0.082	21:59
Air Blank		0.000	22:00
Subject Sample	# 1	0.194	22:00
Breath Volume		1.210	22:00
Air Blank		0.000	22:01
Air Blank		0.000	22:03
Subject Sample	# 2	0.189	22:03
Breath Volume		1.457	22:03
Air Blank		0.000	22:04
Control Test		0.081	22:04
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 03/07/2018 07:13

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: ADAME

DATE/TIME OF TEST: 03/04/2018 01:44:17

Diagnostics Check		OK	01:38
Air Blank		0.000	01:38
Control Test		0.082	01:38
Air Blank		0.000	01:39
Subject Sample	# 1	0.000	01:39
Breath Volume		1.613	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.000	01:42
Breath Volume		1.152	01:42
Air Blank		0.000	01:43
Control Test		0.081	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 03/07/2018 07:13

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMS,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: VERGARA

DATE/TIME OF TEST: 03/04/2018 03:22:01

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.082	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.056	03:17
Breath Volume		1.980	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.063	03:20
Breath Volume		2.195	03:20
Air Blank		0.000	03:21
Control Test		0.081	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:22

Data Download Date/Time: 03/07/2018 07:13

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 03/04/2018 04:00:05

Diagnostics Check		OK	03:53
Air Blank		0.000	03:53
Control Test		0.082	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.210	03:55
Breath Volume		1.789	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.200	03:58
Breath Volume		1.972	03:58
Air Blank		0.000	03:59
Control Test		0.081	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	04:00

Data Download Date/Time: 03/07/2018 07:13

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: LARSON

DATE/TIME OF TEST: 03/05/2018 00:40:30

Diagnostics Check		OK	00:33
Air Blank		0.000	00:34
Control Test		0.081	00:34
Air Blank		0.000	00:35
Subject Sample	# 1	0.181	00:35
Breath Volume		3.031	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.183	00:38
Breath Volume		3.453	00:38
Air Blank		0.000	00:39
Control Test		0.082	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 03/07/2018 07:13

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 03/05/2018 03:09:55

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.083	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.114	03:05
Breath Volume		1.511	03:05
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.106	03:08
Breath Volume		1.628	03:08
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 03/07/2018 07:13

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NONE,NONE
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 03/05/2018 18:53:30

Diagnostics Check		OK	18:47	
Air Blank		0.000	18:48	
Control Test		0.081	18:48	
Air Blank		0.000	18:49	
Subject Sample	# 1	0.310	18:49	
Breath Volume		2.171	18:49	
Air Blank		0.000	18:50	
Air Blank		0.000	18:51	
Subject Sample	# 2		18:52	Subject Test Refused
Breath Volume		0.000	18:52	
Air Blank		0.000	18:52	
Control Test		0.081	18:52	
Air Blank		0.000	18:53	
Diagnostics Check		OK	18:53	

Data Download Date/Time: 03/07/2018 07:13

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: POOLE

DATE/TIME OF TEST: 03/05/2018 22:44:59

Diagnostics Check		OK	22:38
Air Blank		0.000	22:39
Control Test		0.082	22:39
Air Blank		0.000	22:39
Subject Sample	# 1	0.206	22:40
Breath Volume		1.589	22:40
Air Blank		0.000	22:41
Air Blank		0.000	22:42
Subject Sample	# 2	0.200	22:43
Breath Volume		1.558	22:43
Air Blank		0.000	22:44
Control Test		0.081	22:44
Air Blank		0.000	22:44
Diagnostics Check		OK	22:44

Data Download Date/Time: 03/07/2018 07:13

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/06/2018 03:46:37

Diagnostics Check		OK	03:43	
Air Blank		0.000	03:44	
Control Test		0.081	03:44	
Air Blank		0.000	03:44	
Subject Sample	# 1		03:45	Subject Test Refused
Breath Volume		0.000	03:45	
Air Blank		0.000	03:45	
Control Test		0.081	03:46	
Air Blank		0.000	03:46	
Diagnostics Check		OK	03:46	

Data Download Date/Time: 03/07/2018 07:13

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: ROSARIO GONZALEZ

DATE/TIME OF TEST: 03/06/2018 22:06:34

Diagnostics Check		OK	22:00
Air Blank		0.000	22:01
Control Test		0.081	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.000	22:02
Breath Volume		2.746	22:02
Air Blank		0.000	22:02
Air Blank		0.000	22:04
Subject Sample	# 2	0.000	22:05
Breath Volume		1.679	22:05
Air Blank		0.000	22:05
Control Test		0.081	22:05
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 03/07/2018 07:13

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: LOGSDON V

DATE/TIME OF TEST: 03/07/2018 21:02:26

Diagnostics Check		OK	20:55
Air Blank		0.000	20:56
Control Test		0.082	20:56
Air Blank		0.000	20:57
Subject Sample	# 1	0.174	20:57
Breath Volume		1.648	20:57
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.169	21:00
Breath Volume		2.300	21:00
Air Blank		0.000	21:01
Control Test		0.081	21:01
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 04/05/2018 17:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 03/07/2018 22:23:45

Diagnostics Check		OK	22:16
Air Blank		0.000	22:16
Control Test		0.081	22:17
Air Blank		0.000	22:17
Subject Sample	# 1	0.000	22:19
Breath Volume		1.234	22:19
Air Blank		0.000	22:19
Air Blank		0.000	22:21
Subject Sample	# 2	0.000	22:22
Breath Volume		1.796	22:22
Air Blank		0.000	22:22
Control Test		0.081	22:23
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 04/05/2018 17:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: EVERETT JR

DATE/TIME OF TEST: 03/08/2018 04:05:57

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.082	04:00
Air Blank		0.000	04:00
Subject Sample	# 1	0.140	04:01
Breath Volume		2.886	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:03
Subject Sample	# 2	0.145	04:04
Breath Volume		3.003	04:04
Air Blank		0.000	04:05
Control Test		0.081	04:05
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 04/05/2018 17:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: CROTEAU

DATE/TIME OF TEST: 03/08/2018 23:06:35

Diagnostics Check		OK	22:59
Air Blank		0.000	22:59
Control Test		0.083	22:59
Air Blank		0.000	23:00
Subject Sample	# 1	0.166	23:02
Breath Volume		2.019	23:02
Air Blank		0.000	23:02
Air Blank		0.000	23:04
Subject Sample	# 2	0.178	23:05
Breath Volume		1.535	23:05
Air Blank		0.000	23:05
Control Test		0.082	23:05
Air Blank		0.000	23:06
Diagnostics Check		OK	23:06

Data Download Date/Time: 04/05/2018 17:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIERRA,P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: QUARTMAN

DATE/TIME OF TEST: 03/09/2018 04:24:50

Diagnostics Check		OK	04:17
Air Blank		0.000	04:18
Control Test		0.082	04:18
Air Blank		0.000	04:19
Subject Sample	# 1	0.125	04:20
Breath Volume		1.375	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:22
Subject Sample	# 2	0.132	04:23
Breath Volume		1.351	04:23
Air Blank		0.000	04:23
Control Test		0.081	04:24
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 04/05/2018 17:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: INSISIENGMAY

DATE/TIME OF TEST: 03/09/2018 21:34:26

Diagnostics Check		OK	21:27
Air Blank		0.000	21:28
Control Test		0.082	21:28
Air Blank		0.000	21:29
Subject Sample	# 1	0.165	21:29
Breath Volume		2.773	21:29
Air Blank		0.000	21:30
Air Blank		0.000	21:32
Subject Sample	# 2	0.168	21:32
Breath Volume		1.984	21:32
Air Blank		0.000	21:33
Control Test		0.082	21:33
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 04/05/2018 17:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: DIENG

DATE/TIME OF TEST: 03/10/2018 00:42:04

Diagnostics Check		OK	00:35
Air Blank		0.000	00:35
Control Test		0.083	00:36
Air Blank		0.000	00:36
Subject Sample	# 1	0.148	00:37
Breath Volume		1.648	00:37
Air Blank		0.000	00:38
Air Blank		0.000	00:39
Subject Sample	# 2	0.149	00:40
Breath Volume		2.472	00:40
Air Blank		0.000	00:41
Control Test		0.081	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:42

Data Download Date/Time: 04/05/2018 17:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: ESCALONA MARTINEZ

DATE/TIME OF TEST: 03/10/2018 03:24:38

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.082	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.156	03:20
Breath Volume		2.414	03:20
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.147	03:23
Breath Volume		1.707	03:23
Air Blank		0.000	03:23
Control Test		0.081	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 04/05/2018 17:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,L
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: ABDELHAQ

DATE/TIME OF TEST: 03/10/2018 04:31:49

Diagnostics Check		OK	04:17	
Air Blank		0.000	04:17	
Control Test		0.082	04:18	
Air Blank		0.000	04:18	
Subject Sample	# 1	0.103	04:19	
Breath Volume		4.246	04:19	
Air Blank		0.000	04:19	
Air Blank		0.000	04:21	
Subject Sample	# 2		04:24	No Sample Provided
Breath Volume		0.000	04:24	
Air Blank		0.000	04:25	
Air Blank		0.000	04:27	
Subject Sample	# 3		04:30	No Sample Provided
Breath Volume		0.000	04:30	
Air Blank		0.000	04:30	
Control Test		0.082	04:31	
Air Blank		0.000	04:31	
Diagnostics Check		OK	04:31	

Data Download Date/Time: 04/05/2018 17:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,L
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: ABDELHAQ

DATE/TIME OF TEST: 03/10/2018 04:50:44

Diagnostics Check		OK	04:44
Air Blank		0.000	04:44
Control Test		0.082	04:45
Air Blank		0.000	04:45
Subject Sample	# 1	0.095	04:46
Breath Volume		4.250	04:46
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.094	04:49
Breath Volume		3.226	04:49
Air Blank		0.000	04:49
Control Test		0.081	04:50
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 04/05/2018 17:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARRELL,Q
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: SPEROW

DATE/TIME OF TEST: 03/10/2018 05:47:12

Diagnostics Check		OK	05:39
Air Blank		0.000	05:40
Control Test		0.081	05:40
Air Blank		0.000	05:40
Subject Sample	# 1	0.236	05:41
Breath Volume		2.152	05:41
Air Blank		0.000	05:42
Air Blank		0.000	05:44
Subject Sample	# 2	0.241	05:45
Breath Volume		4.082	05:45
Air Blank		0.000	05:46
Control Test		0.081	05:46
Air Blank		0.000	05:47
Diagnostics Check		OK	05:47

Data Download Date/Time: 04/05/2018 17:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: REICHERT,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:23

SUBJECT'S LAST NAME: CRUZ-GONZALEZ

DATE/TIME OF TEST: 03/10/2018 15:55:54

Diagnostics Check		OK	15:44	
Air Blank		0.000	15:44	
Control Test		0.081	15:44	
Air Blank		0.000	15:45	
Subject Sample	# 1	0.256	15:48	Volume Not Met
Breath Volume		0.230	15:48	
Air Blank		0.000	15:49	
Air Blank		0.000	15:50	
Subject Sample	# 2	0.244	15:54	Volume Not Met
Breath Volume		0.273	15:54	
Air Blank		0.000	15:54	
Control Test		0.079	15:55	
Air Blank		0.000	15:55	
Diagnostics Check		OK	15:55	

Data Download Date/Time: 04/05/2018 17:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: REICHERT,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:23

SUBJECT'S LAST NAME: CRUZ-GONZALEZ

DATE/TIME OF TEST: 03/10/2018 16:13:08

Diagnostics Check		OK	15:58	
Air Blank		0.000	15:58	
Control Test		0.079	15:59	
Air Blank		0.000	15:59	
Subject Sample	# 1	0.240	16:02	Volume Not Met
Breath Volume		0.203	16:02	
Air Blank		0.000	16:03	
Air Blank		0.000	16:05	
Subject Sample	# 2	0.270	16:05	
Breath Volume		1.355	16:05	
Air Blank		0.000	16:06	
Air Blank		0.000	16:08	
Subject Sample	# 3	0.235	16:11	Volume Not Met
Breath Volume		0.269	16:11	
Air Blank		0.000	16:12	
Control Test		0.078	16:12	
Air Blank		0.000	16:13	
Diagnostics Check		OK	16:13	

Data Download Date/Time: 04/05/2018 17:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 03/10/2018 22:48:29

Diagnostics Check		OK	22:45	
Air Blank		0.000	22:45	
Control Test		0.080	22:46	
Air Blank		0.000	22:46	
Subject Sample	# 1		22:47	Subject Test Refused
Breath Volume		0.000	22:47	
Air Blank		0.000	22:47	
Control Test		0.081	22:47	
Air Blank		0.000	22:48	
Diagnostics Check		OK	22:48	

Data Download Date/Time: 04/05/2018 17:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 03/11/2018 00:35:52

Diagnostics Check		OK	00:22	
Air Blank		0.000	00:22	
Control Test		0.081	00:23	
Air Blank		0.000	00:23	
Subject Sample	# 1	0.166	00:26	Volume Not Met
Breath Volume		0.355	00:26	
Air Blank		0.000	00:27	
Air Blank		0.000	00:29	
Subject Sample	# 2	0.210	00:30	No .020 agreement
Breath Volume		2.203	00:30	
Air Blank		0.000	00:30	
Air Blank		0.000	00:32	
Subject Sample	# 3	0.187	00:34	No .020 agreement
Breath Volume		1.363	00:34	
Air Blank		0.000	00:34	
Control Test		0.079	00:35	
Air Blank		0.000	00:35	
Diagnostics Check		OK	00:35	

Data Download Date/Time: 04/05/2018 17:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 03/11/2018 01:05:58

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.079	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.196	01:01
Breath Volume		1.601	01:01
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.191	01:04
Breath Volume		1.554	01:04
Air Blank		0.000	01:05
Control Test		0.079	01:05
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 04/05/2018 17:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: VASQUEZ JR

DATE/TIME OF TEST: 03/11/2018 03:48:47

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.081	03:42
Air Blank		0.000	03:43
Subject Sample	# 1	0.000	03:44
Breath Volume		4.113	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.000	03:47
Breath Volume		4.226	03:47
Air Blank		0.000	03:47
Control Test		0.081	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 04/05/2018 17:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAULINO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: MADGETT

DATE/TIME OF TEST: 03/12/2018 02:29:14

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.081	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.182	02:24
Breath Volume		1.363	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.180	02:27
Breath Volume		1.300	02:27
Air Blank		0.000	02:28
Control Test		0.080	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 04/05/2018 17:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: DABROWSKI

DATE/TIME OF TEST: 03/13/2018 20:46:06

Diagnostics Check		OK	20:39
Air Blank		0.000	20:40
Control Test		0.082	20:40
Air Blank		0.000	20:41
Subject Sample	# 1	0.187	20:41
Breath Volume		2.027	20:41
Air Blank		0.000	20:42
Air Blank		0.000	20:44
Subject Sample	# 2	0.182	20:44
Breath Volume		1.472	20:44
Air Blank		0.000	20:45
Control Test		0.081	20:45
Air Blank		0.000	20:46
Diagnostics Check		OK	20:46

Data Download Date/Time: 04/05/2018 17:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MNTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: ANTONIO

DATE/TIME OF TEST: 03/13/2018 23:56:28

Diagnostics Check		OK	23:49
Air Blank		0.000	23:49
Control Test		0.082	23:50
Air Blank		0.000	23:50
Subject Sample	# 1	0.141	23:51
Breath Volume		1.812	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.136	23:54
Breath Volume		1.726	23:54
Air Blank		0.000	23:55
Control Test		0.081	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 04/05/2018 17:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: JOHANSEN II

DATE/TIME OF TEST: 03/15/2018 03:11:17

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.083	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.155	03:06
Breath Volume		5.058	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.156	03:09
Breath Volume		4.957	03:09
Air Blank		0.000	03:10
Control Test		0.081	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 04/05/2018 17:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 03/15/2018 03:58:58

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.083	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.138	03:54
Breath Volume		1.273	03:54
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.145	03:57
Breath Volume		3.089	03:57
Air Blank		0.000	03:58
Control Test		0.082	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 04/05/2018 17:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,L
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 03/15/2018 04:53:56

Diagnostics Check		OK	04:47
Air Blank		0.000	04:48
Control Test		0.082	04:48
Air Blank		0.000	04:49
Subject Sample	# 1	0.104	04:49
Breath Volume		1.394	04:49
Air Blank		0.000	04:50
Air Blank		0.000	04:51
Subject Sample	# 2	0.102	04:52
Breath Volume		1.378	04:52
Air Blank		0.000	04:53
Control Test		0.082	04:53
Air Blank		0.000	04:53
Diagnostics Check		OK	04:53

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TORRES,H
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: LEVINE

DATE/TIME OF TEST: 03/16/2018 01:58:21

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.082	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.130	01:53
Breath Volume		3.613	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.123	01:56
Breath Volume		3.449	01:56
Air Blank		0.000	01:57
Control Test		0.082	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: BUDHU

DATE/TIME OF TEST: 03/17/2018 01:29:53

Diagnostics Check		OK	01:23
Air Blank		0.000	01:23
Control Test		0.082	01:24
Air Blank		0.000	01:24
Subject Sample	# 1	0.079	01:25
Breath Volume		2.019	01:25
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.081	01:28
Breath Volume		2.441	01:28
Air Blank		0.000	01:28
Control Test		0.082	01:29
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: EASON

DATE/TIME OF TEST: 03/17/2018 03:40:27

Diagnostics Check		OK	03:33
Air Blank		0.000	03:34
Control Test		0.081	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.209	03:35
Breath Volume		2.843	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.221	03:38
Breath Volume		2.914	03:38
Air Blank		0.000	03:39
Control Test		0.081	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRIGHTWELL,R
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 03/17/2018 04:15:46

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.081	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.070	04:11
Breath Volume		2.050	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:13
Subject Sample	# 2	0.066	04:14
Breath Volume		1.210	04:14
Air Blank		0.000	04:14
Control Test		0.080	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 04/05/2018 17:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MOREHOUSE

DATE/TIME OF TEST: 03/17/2018 04:53:25

Diagnostics Check		OK	04:46
Air Blank		0.000	04:46
Control Test		0.082	04:46
Air Blank		0.000	04:47
Subject Sample	# 1	0.217	04:48
Breath Volume		1.167	04:48
Air Blank		0.000	04:48
Air Blank		0.000	04:50
Subject Sample	# 2	0.225	04:51
Breath Volume		1.996	04:51
Air Blank		0.000	04:52
Control Test		0.081	04:52
Air Blank		0.000	04:53
Diagnostics Check		OK	04:53

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: BURRIS

DATE/TIME OF TEST: 03/17/2018 05:13:18

Diagnostics Check		OK	05:04	
Air Blank		0.000	05:04	
Control Test		0.080	05:05	
Air Blank		0.000	05:05	
Subject Sample	# 1	0.083	05:08	Volume Not Met
Breath Volume		0.921	05:08	
Air Blank		0.000	05:09	
Air Blank		0.000	05:11	
Subject Sample	# 2		05:11	Subject Test Refused
Breath Volume		0.000	05:11	
Air Blank		0.000	05:12	
Control Test		0.080	05:12	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 04/05/2018 17:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,A
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: BUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: SCHERP

DATE/TIME OF TEST: 03/17/2018 22:16:36

Diagnostics Check		OK	22:08
Air Blank		0.000	22:08
Control Test		0.082	22:08
Air Blank		0.000	22:09
Subject Sample	# 1	0.151	22:09
Breath Volume		1.316	22:09
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.158	22:14
Breath Volume		1.562	22:14
Air Blank		0.000	22:15
Control Test		0.082	22:15
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: GRANADOS CABRERA

DATE/TIME OF TEST: 03/18/2018 01:45:21

Diagnostics Check		OK	01:37
Air Blank		0.000	01:38
Control Test		0.081	01:38
Air Blank		0.000	01:39
Subject Sample	# 1	0.210	01:40
Breath Volume		2.628	01:40
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.210	01:43
Breath Volume		1.488	01:43
Air Blank		0.000	01:44
Control Test		0.080	01:44
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/18/2018 02:47:13

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.081	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.216	02:42
Breath Volume		3.160	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.229	02:45
Breath Volume		2.218	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 04/05/2018 17:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: ESCALANTE RESINO

DATE/TIME OF TEST: 03/18/2018 03:54:14

Diagnostics Check		OK	03:47
Air Blank		0.000	03:48
Control Test		0.081	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.200	03:49
Breath Volume		2.472	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:51
Subject Sample	# 2	0.203	03:52
Breath Volume		3.539	03:52
Air Blank		0.000	03:53
Control Test		0.080	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: VARGAS IZQUIERDO

DATE/TIME OF TEST: 03/18/2018 05:05:24

Diagnostics Check		OK	04:58
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Subject Sample	# 1	0.142	05:00
Breath Volume		1.328	05:00
Air Blank		0.000	05:01
Air Blank		0.000	05:03
Subject Sample	# 2	0.149	05:03
Breath Volume		1.492	05:03
Air Blank		0.000	05:04
Control Test		0.079	05:04
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:52

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 03/18/2018 07:20:11

Diagnostics Check		OK	07:13
Air Blank		0.000	07:13
Control Test		0.081	07:13
Air Blank		0.000	07:14
Subject Sample	# 1	0.014	07:15
Breath Volume		2.140	07:15
Air Blank		0.000	07:16
Air Blank		0.000	07:18
Subject Sample	# 2	0.013	07:18
Breath Volume		2.464	07:18
Air Blank		0.000	07:19
Control Test		0.080	07:19
Air Blank		0.000	07:20
Diagnostics Check		OK	07:20

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MARTINEZ-RIVERA,A
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: BUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: GARCIA-LLORCA

DATE/TIME OF TEST: 03/18/2018 22:31:42

Diagnostics Check		OK	22:25
Air Blank		0.000	22:26
Control Test		0.082	22:26
Air Blank		0.000	22:26
Subject Sample	# 1	0.069	22:27
Breath Volume		2.449	22:27
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.064	22:30
Breath Volume		1.511	22:30
Air Blank		0.000	22:30
Control Test		0.081	22:31
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 03/18/2018 23:58:14

Diagnostics Check		OK	23:51
Air Blank		0.000	23:51
Control Test		0.082	23:52
Air Blank		0.000	23:52
Subject Sample	# 1	0.167	23:53
Breath Volume		1.636	23:53
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.158	23:56
Breath Volume		1.820	23:56
Air Blank		0.000	23:57
Control Test		0.081	23:57
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 03/19/2018 04:08:05

Diagnostics Check		OK	04:01
Air Blank		0.000	04:02
Control Test		0.081	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.217	04:03
Breath Volume		2.179	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.233	04:06
Breath Volume		2.492	04:06
Air Blank		0.000	04:07
Control Test		0.081	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: GAMBER

DATE/TIME OF TEST: 03/20/2018 00:38:27

Diagnostics Check		OK	00:28
Air Blank		0.000	00:29
Control Test		0.080	00:29
Air Blank		0.000	00:30
Subject Sample	# 1	0.229	00:30
Breath Volume		1.945	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:32
Subject Sample	# 2	0.208	00:33
Breath Volume		1.218	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 3	0.238	00:36
Breath Volume		2.792	00:36
Air Blank		0.000	00:37
Control Test		0.080	00:37
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: MEDINA VAZQUEZ

DATE/TIME OF TEST: 03/20/2018 05:44:19

Diagnostics Check		OK	05:38
Air Blank		0.000	05:38
Control Test		0.081	05:38
Air Blank		0.000	05:39
Subject Sample	# 1	0.000	05:39
Breath Volume		2.433	05:39
Air Blank		0.000	05:40
Air Blank		0.000	05:42
Subject Sample	# 2	0.000	05:42
Breath Volume		1.964	05:42
Air Blank		0.000	05:43
Control Test		0.081	05:43
Air Blank		0.000	05:44
Diagnostics Check		OK	05:44

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KRYGIER,B
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: ROSARIO JR

DATE/TIME OF TEST: 03/20/2018 15:51:32

Diagnostics Check		OK	15:45
Air Blank		0.000	15:46
Control Test		0.080	15:46
Air Blank		0.000	15:46
Subject Sample	# 1	0.000	15:47
Breath Volume		2.320	15:47
Air Blank		0.000	15:47
Air Blank		0.000	15:49
Subject Sample	# 2	0.000	15:50
Breath Volume		2.109	15:50
Air Blank		0.000	15:50
Control Test		0.080	15:50
Air Blank		0.000	15:51
Diagnostics Check		OK	15:51

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARBOSA,F
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:39

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 03/21/2018 14:08:34

Diagnostics Check		OK	14:01
Air Blank		0.000	14:01
Control Test		0.082	14:01
Air Blank		0.000	14:02
Subject Sample	# 1	0.215	14:03
Breath Volume		1.335	14:03
Air Blank		0.000	14:04
Air Blank		0.000	14:06
Subject Sample	# 2	0.218	14:06
Breath Volume		2.011	14:06
Air Blank		0.000	14:07
Control Test		0.081	14:07
Air Blank		0.000	14:08
Diagnostics Check		OK	14:08

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: HANNAH

DATE/TIME OF TEST: 03/22/2018 00:38:19

Diagnostics Check		OK	00:30
Air Blank		0.000	00:31
Control Test		0.081	00:31
Air Blank		0.000	00:32
Subject Sample	# 1	0.114	00:33
Breath Volume		1.597	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.112	00:36
Breath Volume		1.855	00:36
Air Blank		0.000	00:37
Control Test		0.081	00:37
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 04/05/2018 17:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 03/22/2018 23:13:02

Diagnostics Check		OK	23:03
Air Blank		0.000	23:03
Control Test		0.082	23:04
Air Blank		0.000	23:04
Subject Sample	# 1	0.309	23:05
Breath Volume		1.441	23:05
Air Blank		0.000	23:06
Air Blank		0.000	23:07
Subject Sample	# 2	0.278	23:08
Breath Volume		1.371	23:08
Air Blank		0.000	23:09
Air Blank		0.000	23:10
Subject Sample	# 3	0.286	23:11
Breath Volume		1.453	23:11
Air Blank		0.000	23:12
Control Test		0.081	23:12
Air Blank		0.000	23:12
Diagnostics Check		OK	23:13

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMS,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: BOESEL

DATE/TIME OF TEST: 03/23/2018 00:23:11

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.079	00:17
Air Blank		0.000	00:17
Subject Sample	# 1	0.247	00:18
Breath Volume		2.644	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.245	00:21
Breath Volume		2.628	00:21
Air Blank		0.000	00:22
Control Test		0.082	00:22
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: GIVENS

DATE/TIME OF TEST: 03/23/2018 04:59:14

Diagnostics Check		OK	04:52
Air Blank		0.000	04:52
Control Test		0.082	04:53
Air Blank		0.000	04:53
Subject Sample	# 1	0.172	04:54
Breath Volume		1.562	04:54
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.174	04:57
Breath Volume		1.578	04:57
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KENNEDY,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: ESPOLA

DATE/TIME OF TEST: 03/23/2018 22:53:26

Diagnostics Check		OK	22:46
Air Blank		0.000	22:46
Control Test		0.083	22:47
Air Blank		0.000	22:47
Subject Sample	# 1	0.149	22:48
Breath Volume		2.648	22:48
Air Blank		0.000	22:49
Air Blank		0.000	22:51
Subject Sample	# 2	0.142	22:51
Breath Volume		2.546	22:51
Air Blank		0.000	22:52
Control Test		0.082	22:52
Air Blank		0.000	22:53
Diagnostics Check		OK	22:53

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 03/24/2018 00:22:53

Diagnostics Check		OK	00:11	
Air Blank		0.000	00:11	
Control Test		0.082	00:11	
Air Blank		0.000	00:12	
Subject Sample	# 1	0.172	00:15	Volume Not Met
Breath Volume		0.601	00:15	
Air Blank		0.000	00:16	
Air Blank		0.000	00:17	
Subject Sample	# 2	0.185	00:21	Volume Not Met
Breath Volume		0.042	00:21	
Air Blank		0.000	00:21	
Control Test		0.081	00:22	
Air Blank		0.000	00:22	
Diagnostics Check		OK	00:22	

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 03/24/2018 00:44:20

Diagnostics Check		OK	00:31	
Air Blank		0.000	00:31	
Control Test		0.080	00:32	
Air Blank		0.000	00:32	
Subject Sample	# 1	0.193	00:35	Volume Not Met
Breath Volume		1.003	00:35	
Air Blank		0.000	00:36	
Air Blank		0.000	00:38	
Subject Sample	# 2	0.209	00:38	
Breath Volume		1.562	00:38	
Air Blank		0.000	00:39	
Air Blank		0.000	00:41	
Subject Sample	# 3		00:42	Subject Test Refused
Breath Volume		0.000	00:42	
Air Blank		0.000	00:43	
Control Test		0.081	00:43	
Air Blank		0.000	00:44	
Diagnostics Check		OK	00:44	

Data Download Date/Time: 04/05/2018 17:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MORANIEC,AMANDA
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 03/24/2018 17:11:05

Diagnostics Check		OK	17:02	
Air Blank		0.000	17:03	
Control Test		0.083	17:03	
Air Blank		0.000	17:03	
Subject Sample	# 1		17:07	No Sample Provided
Breath Volume		0.000	17:07	
Air Blank		0.000	17:07	
Air Blank		0.000	17:09	
Subject Sample	# 2		17:09	Subject Test Refused
Breath Volume		0.000	17:09	
Air Blank		0.000	17:10	
Control Test		0.082	17:10	
Air Blank		0.000	17:10	
Diagnostics Check		OK	17:11	

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MORANIEC,AMANDA
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 03/24/2018 17:22:11

Diagnostics Check		OK	17:12
Air Blank		0.000	17:12
Control Test		0.082	17:12
Air Blank		0.000	17:13
Subject Sample	# 1	0.193	17:14
Breath Volume		2.914	17:14
Air Blank		0.000	17:14
Air Blank		0.000	17:16
Subject Sample	# 2	0.160	17:17
Breath Volume		1.339	17:17
Air Blank		0.000	17:18
Air Blank		0.000	17:20
Subject Sample	# 3	0.182	17:20
Breath Volume		1.914	17:20
Air Blank		0.000	17:21
Control Test		0.081	17:21
Air Blank		0.000	17:22
Diagnostics Check		OK	17:22

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/25/2018 00:44:17

Diagnostics Check		OK	00:37
Air Blank		0.000	00:38
Control Test		0.081	00:38
Air Blank		0.000	00:39
Subject Sample	# 1	0.195	00:39
Breath Volume		2.750	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.192	00:42
Breath Volume		2.574	00:42
Air Blank		0.000	00:43
Control Test		0.081	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIERRA,P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 03/25/2018 03:13:04

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.082	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.122	03:07
Breath Volume		2.316	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.128	03:11
Breath Volume		1.933	03:11
Air Blank		0.000	03:12
Control Test		0.081	03:12
Air Blank		0.000	03:12
Diagnostics Check		OK	03:13

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: WILTSHIRE

DATE/TIME OF TEST: 03/25/2018 03:32:27

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.081	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.130	03:27
Breath Volume		3.085	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.133	03:30
Breath Volume		3.339	03:30
Air Blank		0.000	03:31
Control Test		0.081	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: KIESNER

DATE/TIME OF TEST: 03/25/2018 04:25:07

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.081	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.107	04:20
Breath Volume		1.875	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.107	04:23
Breath Volume		2.519	04:23
Air Blank		0.000	04:24
Control Test		0.081	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 03/25/2018 23:09:31

Diagnostics Check		OK	23:03
Air Blank		0.000	23:03
Control Test		0.082	23:04
Air Blank		0.000	23:04
Subject Sample	# 1	0.040	23:05
Breath Volume		1.656	23:05
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.038	23:07
Breath Volume		2.945	23:07
Air Blank		0.000	23:08
Control Test		0.082	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 03/28/2018 23:43:34

Diagnostics Check		OK	23:36
Air Blank		0.000	23:37
Control Test		0.082	23:37
Air Blank		0.000	23:37
Subject Sample	# 1	0.220	23:38
Breath Volume		1.449	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.222	23:41
Breath Volume		1.554	23:41
Air Blank		0.000	23:42
Control Test		0.081	23:42
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 03/29/2018 04:51:16

Diagnostics Check		OK	04:44
Air Blank		0.000	04:44
Control Test		0.082	04:45
Air Blank		0.000	04:45
Subject Sample	# 1	0.141	04:46
Breath Volume		2.750	04:46
Air Blank		0.000	04:47
Air Blank		0.000	04:48
Subject Sample	# 2	0.147	04:49
Breath Volume		2.769	04:49
Air Blank		0.000	04:50
Control Test		0.080	04:50
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEDINA,P
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: PEREZ-LOPEZ

DATE/TIME OF TEST: 03/30/2018 03:10:19

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.082	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.146	03:05
Breath Volume		1.976	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.150	03:08
Breath Volume		1.957	03:08
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 03/30/2018 05:01:07

Diagnostics Check		OK	04:53
Air Blank		0.000	04:53
Control Test		0.081	04:54
Air Blank		0.000	04:54
Subject Sample	# 1	0.212	04:56
Breath Volume		2.875	04:56
Air Blank		0.000	04:57
Air Blank		0.000	04:58
Subject Sample	# 2	0.213	04:59
Breath Volume		1.519	04:59
Air Blank		0.000	05:00
Control Test		0.080	05:00
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SOSBEE

DATE/TIME OF TEST: 03/31/2018 00:19:57

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.081	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.149	00:14
Breath Volume		1.960	00:14
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.154	00:18
Breath Volume		1.734	00:18
Air Blank		0.000	00:19
Control Test		0.081	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIERRA,P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: CUFF

DATE/TIME OF TEST: 03/31/2018 02:19:27

Diagnostics Check		OK	02:08	
Air Blank		0.000	02:08	
Control Test		0.081	02:08	
Air Blank		0.000	02:09	
Subject Sample	# 1		02:12	No Sample Provided
Breath Volume		0.000	02:12	
Air Blank		0.000	02:12	
Air Blank		0.000	02:14	
Subject Sample	# 2		02:18	No Sample Provided
Breath Volume		0.000	02:18	
Air Blank		0.000	02:18	
Control Test		0.081	02:18	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 04/05/2018 17:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: WILLIAMSON III

DATE/TIME OF TEST: 03/31/2018 03:30:00

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.081	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.165	03:25
Breath Volume		1.242	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.167	03:28
Breath Volume		1.996	03:28
Air Blank		0.000	03:29
Control Test		0.081	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:30

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 03/31/2018 04:45:08

Diagnostics Check		OK	04:38
Air Blank		0.000	04:39
Control Test		0.081	04:39
Air Blank		0.000	04:40
Subject Sample	# 1	0.056	04:40
Breath Volume		2.660	04:40
Air Blank		0.000	04:41
Air Blank		0.000	04:43
Subject Sample	# 2	0.059	04:43
Breath Volume		3.125	04:43
Air Blank		0.000	04:44
Control Test		0.081	04:44
Air Blank		0.000	04:45
Diagnostics Check		OK	04:45

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROBINSON,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:42

SUBJECT'S LAST NAME: GAISBAUER

DATE/TIME OF TEST: 03/31/2018 16:09:33

Diagnostics Check		OK	16:03
Air Blank		0.000	16:03
Control Test		0.081	16:03
Air Blank		0.000	16:04
Subject Sample	# 1	0.273	16:04
Breath Volume		1.203	16:04
Air Blank		0.000	16:05
Air Blank		0.000	16:07
Subject Sample	# 2	0.293	16:07
Breath Volume		2.281	16:07
Air Blank		0.000	16:08
Control Test		0.081	16:08
Air Blank		0.000	16:09
Diagnostics Check		OK	16:09

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HILES,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: SHREVE

DATE/TIME OF TEST: 03/31/2018 23:38:39

Diagnostics Check		OK	23:32
Air Blank		0.000	23:32
Control Test		0.082	23:33
Air Blank		0.000	23:33
Subject Sample	# 1	0.114	23:34
Breath Volume		2.000	23:34
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.110	23:37
Breath Volume		1.687	23:37
Air Blank		0.000	23:37
Control Test		0.081	23:38
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 04/05/2018 17:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTMAS,DAVID,D
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ESQUIVEL,ANTHONY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:29

SUBJECT'S LAST NAME: PAPESH

DATE/TIME OF TEST: 03/24/2018 13:02:28

Diagnostics Check		OK	12:56
Air Blank		0.000	12:56
Control Test		0.081	12:57
Air Blank		0.000	12:57
Subject Sample	# 1	0.000	12:58
Breath Volume		2.257	12:58
Air Blank		0.000	12:58
Air Blank		0.000	13:00
Subject Sample	# 2	0.000	13:01
Breath Volume		1.824	13:01
Air Blank		0.000	13:01
Control Test		0.080	13:01
Air Blank		0.000	13:02
Diagnostics Check		OK	13:02

Data Download Date/Time: 04/27/2018 03:41

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID,MICHAEL,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: REID,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 64476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: ROSAS-MEDINA

DATE/TIME OF TEST: 03/25/2018 16:46:02

Diagnostics Check		OK	16:39
Air Blank		0.000	16:39
Control Test		0.080	16:39
Air Blank		0.000	16:40
Subject Sample	# 1	0.121	16:41
Breath Volume		1.105	16:41
Air Blank		0.000	16:42
Air Blank		0.000	16:44
Subject Sample	# 2	0.122	16:44
Breath Volume		2.175	16:44
Air Blank		0.000	16:45
Control Test		0.080	16:45
Air Blank		0.000	16:45
Diagnostics Check		OK	16:46

Data Download Date/Time: 04/27/2018 03:41

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HECKMAN,SHELBY,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PESCIA,DEAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 03/27/2018 03:50:06

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.165	03:45
Breath Volume		1.753	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.165	03:48
Breath Volume		2.878	03:48
Air Blank		0.000	03:49
Control Test		0.080	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 04/27/2018 03:41

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTIGLIONE,SALVATORE,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CASTIGLIONE,SALVATORE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: MC CRADY

DATE/TIME OF TEST: 03/29/2018 03:57:30

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.080	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.142	03:53
Breath Volume		1.593	03:53
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.136	03:55
Breath Volume		1.921	03:55
Air Blank		0.000	03:56
Control Test		0.081	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 04/27/2018 03:41

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINES,RAINES,L
BREATH TEST OPERATOR AGENCY: Dunnellon PD
ARRESTING OFFICER: RAINES,ELDON
ARRESTING OFFICER AGENCY: Dunnellon PD
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: STOCKER

DATE/TIME OF TEST: 03/29/2018 05:09:20

Diagnostics Check		OK	05:06	
Air Blank		0.000	05:06	
Control Test		0.081	05:07	
Air Blank		0.000	05:07	
Subject Sample	# 1		05:07	Subject Test Refused
Breath Volume		0.000	05:07	
Air Blank		0.000	05:08	
Control Test		0.080	05:08	
Air Blank		0.000	05:09	
Diagnostics Check		OK	05:09	

Data Download Date/Time: 04/27/2018 03:41

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: MCLAREN

DATE/TIME OF TEST: 03/30/2018 02:57:50

Diagnostics Check	Fail	02:57	Temperature Regulation Test
Air Blank	0.000	02:57	

Data Download Date/Time: 04/27/2018 03:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: MCLAREN

DATE/TIME OF TEST: 03/30/2018 02:59:33

Diagnostics Check	Fail	02:59	Temperature Regulation Test
Air Blank	0.000	02:59	

Data Download Date/Time: 04/27/2018 03:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: MCLAREN

DATE/TIME OF TEST: 03/30/2018 03:11:44

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.080	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.048	03:07
Breath Volume		2.312	03:07
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.046	03:10
Breath Volume		2.289	03:10
Air Blank		0.000	03:10
Control Test		0.080	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 04/27/2018 03:41

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: DANNEMAN

DATE/TIME OF TEST: 03/30/2018 23:51:56

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.080	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.325	23:42
Breath Volume		1.136	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.348	23:46
Breath Volume		1.515	23:46
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 3	0.356	23:50
Breath Volume		1.925	23:50
Air Blank		0.000	23:50
Control Test		0.080	23:51
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 04/27/2018 03:41

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 03/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SANTOS PADRON

DATE/TIME OF TEST: 03/31/2018 03:27:31

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.080	03:22
Air Blank		0.000	03:22
Subject Sample	# 1	0.120	03:23
Breath Volume		2.109	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.113	03:25
Breath Volume		2.109	03:25
Air Blank		0.000	03:26
Control Test		0.081	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 04/27/2018 03:41

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/17/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: FALKIN

DATE/TIME OF TEST: 03/01/2018 00:06:07

Diagnostics Check		OK	00:02	
Air Blank		0.000	00:03	
Control Test		0.080	00:03	
Air Blank		0.000	00:04	
Subject Sample	# 1		00:04	Subject Test Refused
Breath Volume		0.000	00:04	
Air Blank		0.000	00:05	
Control Test		0.081	00:05	
Air Blank		0.000	00:06	
Diagnostics Check		OK	00:06	

Data Download Date/Time: 03/26/2018 03:49

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTA,ANGEL,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: HOFFLAND,R
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 03/04/2018 03:02:06

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.082	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.150	02:57
Breath Volume		2.464	02:57
Air Blank		0.000	02:58
Air Blank		0.000	02:59
Subject Sample	# 2	0.158	03:00
Breath Volume		1.855	03:00
Air Blank		0.000	03:01
Control Test		0.081	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 03/26/2018 03:49

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTA,ANGEL,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MATHIS,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: MCLELLAND

DATE/TIME OF TEST: 03/04/2018 21:54:19

Diagnostics Check		OK	21:47
Air Blank		0.000	21:48
Control Test		0.082	21:48
Air Blank		0.000	21:49
Subject Sample	# 1	0.114	21:49
Breath Volume		1.109	21:49
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.110	21:52
Breath Volume		1.562	21:52
Air Blank		0.000	21:53
Control Test		0.081	21:53
Air Blank		0.000	21:54
Diagnostics Check		OK	21:54

Data Download Date/Time: 03/26/2018 03:49

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,D,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: GREIS,D
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: COULTER

DATE/TIME OF TEST: 03/08/2018 01:35:44

Diagnostics Check		OK	01:28
Air Blank		0.000	01:29
Control Test		0.081	01:29
Air Blank		0.000	01:30
Subject Sample	# 1	0.190	01:30
Breath Volume		2.628	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:32
Subject Sample	# 2	0.189	01:34
Breath Volume		2.617	01:34
Air Blank		0.000	01:34
Control Test		0.081	01:35
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 03/26/2018 03:49

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTA,ANGEL,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MATHIS,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 03/08/2018 22:02:07

Diagnostics Check		OK	21:55
Air Blank		0.000	21:55
Control Test		0.082	21:56
Air Blank		0.000	21:56
Subject Sample	# 1	0.143	21:57
Breath Volume		2.839	21:57
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.145	22:00
Breath Volume		2.511	22:00
Air Blank		0.000	22:01
Control Test		0.081	22:01
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 03/26/2018 03:49

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: HARPER,W
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/17/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/10/2018 04:29:23

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.081	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.090	04:24
Breath Volume		1.320	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.091	04:27
Breath Volume		2.683	04:27
Air Blank		0.000	04:28
Control Test		0.081	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 03/26/2018 03:49

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COCHRANE,AARON,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,AARON
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: PIMIENTA

DATE/TIME OF TEST: 03/11/2018 01:11:25

Diagnostics Check		OK	01:06	
Air Blank		0.000	01:07	
Control Test		0.081	01:07	
Air Blank		0.000	01:07	
Subject Sample	# 1		01:10	Subject Test Refused
Breath Volume		0.000	01:10	
Air Blank		0.000	01:10	
Control Test		0.081	01:10	
Air Blank		0.000	01:11	
Diagnostics Check		OK	01:11	

Data Download Date/Time: 03/26/2018 03:49

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/17/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: PACKAN

DATE/TIME OF TEST: 03/11/2018 04:01:47

Diagnostics Check		OK	03:57	
Air Blank		0.000	03:58	
Control Test		0.080	03:58	
Air Blank		0.000	03:59	
Subject Sample	# 1		04:00	Subject Test Refused
Breath Volume		0.000	04:00	
Air Blank		0.000	04:00	
Control Test		0.081	04:01	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 03/26/2018 03:49

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COCHRANE,AARON,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,AARON
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 03/11/2018 23:57:38

Diagnostics Check		OK	23:50
Air Blank		0.000	23:51
Control Test		0.080	23:51
Air Blank		0.000	23:52
Subject Sample	# 1	0.204	23:52
Breath Volume		3.847	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.205	23:56
Breath Volume		3.656	23:56
Air Blank		0.000	23:56
Control Test		0.079	23:57
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 03/26/2018 03:49

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,CHRISTINA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RITCHIE,FRANCIS
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: LECORNU

DATE/TIME OF TEST: 03/13/2018 21:56:04

Diagnostics Check		OK	21:53	
Air Blank		0.000	21:53	
Control Test		0.082	21:53	
Air Blank		0.000	21:54	
Subject Sample	# 1		21:54	Subject Test Refused
Breath Volume		0.000	21:54	
Air Blank		0.000	21:55	
Control Test		0.081	21:55	
Air Blank		0.000	21:55	
Diagnostics Check		OK	21:56	

Data Download Date/Time: 03/26/2018 03:49

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,CHRISTINA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: HOFFLAND,RYAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/14/2018 02:39:57

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:36	
Control Test		0.081	02:36	
Air Blank		0.000	02:37	
Subject Sample	# 1		02:38	Subject Test Refused
Breath Volume		0.000	02:38	
Air Blank		0.000	02:39	
Control Test		0.081	02:39	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 03/26/2018 03:49

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: FOREHAND

DATE/TIME OF TEST: 03/17/2018 07:54:34

Diagnostics Check		OK	07:47
Air Blank		0.000	07:48
Control Test		0.081	07:48
Air Blank		0.000	07:48
Subject Sample	# 1	0.112	07:49
Breath Volume		1.269	07:49
Air Blank		0.000	07:50
Air Blank		0.000	07:52
Subject Sample	# 2	0.111	07:52
Breath Volume		2.429	07:52
Air Blank		0.000	07:53
Control Test		0.081	07:53
Air Blank		0.000	07:54
Diagnostics Check		OK	07:54

Data Download Date/Time: 03/26/2018 03:49

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: RAYBURN

DATE/TIME OF TEST: 03/17/2018 22:43:56

Diagnostics Check		OK	22:41	
Air Blank		0.000	22:41	
Control Test		0.082	22:41	
Air Blank		0.000	22:42	
Subject Sample	# 1		22:42	Subject Test Refused
Breath Volume		0.000	22:42	
Air Blank		0.000	22:43	
Control Test		0.081	22:43	
Air Blank		0.000	22:43	
Diagnostics Check		OK	22:43	

Data Download Date/Time: 03/26/2018 03:49

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 03/18/2018 03:28:00

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.081	03:22
Air Blank		0.000	03:22
Subject Sample	# 1	0.308	03:23
Breath Volume		3.265	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:25
Subject Sample	# 2	0.316	03:26
Breath Volume		3.667	03:26
Air Blank		0.000	03:27
Control Test		0.080	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:28

Data Download Date/Time: 03/26/2018 03:49

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SPENCE

DATE/TIME OF TEST: 03/27/2018 02:34:29

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.080	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.121	02:29
Breath Volume		1.390	02:29
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.128	02:32
Breath Volume		2.726	02:32
Air Blank		0.000	02:33
Control Test		0.079	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 04/27/2018 05:08

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COCHRANE,AARON,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,AARON
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 03/30/2018 01:05:05

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.081	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.202	01:00
Breath Volume		3.050	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:02
Subject Sample	# 2	0.193	01:03
Breath Volume		2.167	01:03
Air Blank		0.000	01:04
Control Test		0.079	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:05

Data Download Date/Time: 04/27/2018 05:08

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,JAMES,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: STEWART,DAVID
ARRESTING OFFICER AGENCY: Center Hill PD
LAST AGENCY INSPECTION: 03/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: H161479761910

DATE/TIME OF TEST: 03/31/2018 23:17:25

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.079	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.244	23:12
Breath Volume		3.394	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.247	23:15
Breath Volume		3.210	23:15
Air Blank		0.000	23:16
Control Test		0.079	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 04/26/2018 10:29

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NAVARRO,BENJAMIN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: BEATTIE

DATE/TIME OF TEST: 03/02/2018 00:14:20

Diagnostics Check		OK	00:07
Air Blank		0.000	00:08
Control Test		0.077	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.146	00:09
Breath Volume		2.371	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:11
Subject Sample	# 2	0.148	00:12
Breath Volume		2.484	00:12
Air Blank		0.000	00:13
Control Test		0.077	00:13
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ZINN,BRYAN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: HEIM

DATE/TIME OF TEST: 03/03/2018 01:00:33

Diagnostics Check		OK	00:57	
Air Blank		0.000	00:58	
Control Test		0.078	00:58	
Air Blank		0.000	00:58	
Subject Sample	# 1		00:59	Subject Test Refused
Breath Volume		0.000	00:59	
Air Blank		0.000	00:59	
Control Test		0.079	00:59	
Air Blank		0.000	01:00	
Diagnostics Check		OK	01:00	

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MCCLINTON

DATE/TIME OF TEST: 03/03/2018 05:13:32

Diagnostics Check		OK	05:10	
Air Blank		0.000	05:11	
Control Test		0.079	05:11	
Air Blank		0.000	05:11	
Subject Sample	# 1		05:12	Subject Test Refused
Breath Volume		0.000	05:12	
Air Blank		0.000	05:12	
Control Test		0.078	05:12	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: CAUSEY,JOHN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/05/2018 00:32:29

Diagnostics Check		OK	00:25
Air Blank		0.000	00:26
Control Test		0.078	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.148	00:27
Breath Volume		2.171	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.135	00:30
Breath Volume		2.617	00:30
Air Blank		0.000	00:31
Control Test		0.078	00:31
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAMOS,JOSEPH
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: NOELL

DATE/TIME OF TEST: 03/06/2018 22:37:31

Diagnostics Check		OK	22:34	
Air Blank		0.000	22:35	
Control Test		0.078	22:35	
Air Blank		0.000	22:35	
Subject Sample	# 1		22:36	Subject Test Refused
Breath Volume		0.000	22:36	
Air Blank		0.000	22:36	
Control Test		0.078	22:36	
Air Blank		0.000	22:37	
Diagnostics Check		OK	22:37	

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HOLLOWAY,JACOB
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: SWINDALL

DATE/TIME OF TEST: 03/07/2018 21:22:27

Diagnostics Check		OK	21:15
Air Blank		0.000	21:15
Control Test		0.079	21:16
Air Blank		0.000	21:16
Subject Sample	# 1	0.195	21:17
Breath Volume		1.835	21:17
Air Blank		0.000	21:18
Air Blank		0.000	21:19
Subject Sample	# 2	0.200	21:20
Breath Volume		1.511	21:20
Air Blank		0.000	21:21
Control Test		0.078	21:21
Air Blank		0.000	21:22
Diagnostics Check		OK	21:22

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: STRIMIKE,ALEX
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: MERCADO

DATE/TIME OF TEST: 03/08/2018 23:41:31

Diagnostics Check		OK	23:38	
Air Blank		0.000	23:38	
Control Test		0.079	23:39	
Air Blank		0.000	23:39	
Subject Sample	# 1		23:40	Subject Test Refused
Breath Volume		0.000	23:40	
Air Blank		0.000	23:40	
Control Test		0.078	23:40	
Air Blank		0.000	23:41	
Diagnostics Check		OK	23:41	

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BARD,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: CHURCH

DATE/TIME OF TEST: 03/09/2018 22:11:10

Diagnostics Check		OK	22:04
Air Blank		0.000	22:04
Control Test		0.080	22:04
Air Blank		0.000	22:05
Subject Sample	# 1	0.276	22:06
Breath Volume		3.683	22:06
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.282	22:09
Breath Volume		3.539	22:09
Air Blank		0.000	22:10
Control Test		0.078	22:10
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL,JARROD,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BARD,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 03/10/2018 17:15:57

Diagnostics Check		OK	17:13	
Air Blank		0.000	17:13	
Control Test		0.078	17:13	
Air Blank		0.000	17:14	
Subject Sample	# 1		17:14	Subject Test Refused
Breath Volume		0.000	17:14	
Air Blank		0.000	17:15	
Control Test		0.078	17:15	
Air Blank		0.000	17:15	
Diagnostics Check		OK	17:15	

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HENRY,DENNIS
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: BARROWS

DATE/TIME OF TEST: 03/12/2018 05:26:28

Diagnostics Check		OK	05:19
Air Blank		0.000	05:19
Control Test		0.077	05:19
Air Blank		0.000	05:20
Subject Sample	# 1	0.146	05:21
Breath Volume		2.765	05:21
Air Blank		0.000	05:21
Air Blank		0.000	05:23
Subject Sample	# 2	0.133	05:24
Breath Volume		1.816	05:24
Air Blank		0.000	05:25
Control Test		0.077	05:25
Air Blank		0.000	05:26
Diagnostics Check		OK	05:26

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVERS,DAWN,T
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GREGORY,TYLER
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 03/12/2018 23:48:52

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:45	
Control Test		0.078	23:45	
Air Blank		0.000	23:46	
Subject Sample	# 1		23:47	Subject Test Refused
Breath Volume		0.000	23:47	
Air Blank		0.000	23:47	
Control Test		0.079	23:48	
Air Blank		0.000	23:48	
Diagnostics Check		OK	23:48	

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVERS,DAWN,T
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ZINN,BRYAN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: SQUIRE

DATE/TIME OF TEST: 03/13/2018 23:24:33

Diagnostics Check		OK	23:17
Air Blank		0.000	23:17
Control Test		0.079	23:18
Air Blank		0.000	23:18
Subject Sample	# 1	0.163	23:19
Breath Volume		1.886	23:19
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.155	23:22
Breath Volume		1.714	23:22
Air Blank		0.000	23:23
Control Test		0.078	23:23
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NAVARRO,BENJAMIN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: NOE

DATE/TIME OF TEST: 03/14/2018 23:11:38

Diagnostics Check		OK	23:03
Air Blank		0.000	23:04
Control Test		0.079	23:04
Air Blank		0.000	23:05
Subject Sample	# 1	0.182	23:06
Breath Volume		2.324	23:06
Air Blank		0.000	23:07
Air Blank		0.000	23:09
Subject Sample	# 2	0.174	23:09
Breath Volume		1.503	23:09
Air Blank		0.000	23:10
Control Test		0.078	23:10
Air Blank		0.000	23:11
Diagnostics Check		OK	23:11

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NAVARRO,BENJAMIN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: CABRAL

DATE/TIME OF TEST: 03/20/2018 23:22:54

Diagnostics Check		OK	23:12
Air Blank		0.000	23:12
Control Test		0.077	23:13
Air Blank		0.000	23:13
Subject Sample	# 1	0.150	23:16
Breath Volume		1.449	23:16
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.147	23:21
Breath Volume		1.781	23:21
Air Blank		0.000	23:21
Control Test		0.077	23:22
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SMITH,JERRY
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: BURTON

DATE/TIME OF TEST: 03/22/2018 00:55:59

Diagnostics Check		OK	00:48
Air Blank		0.000	00:49
Control Test		0.078	00:49
Air Blank		0.000	00:49
Subject Sample	# 1	0.231	00:50
Breath Volume		1.390	00:50
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.237	00:54
Breath Volume		1.644	00:54
Air Blank		0.000	00:55
Control Test		0.077	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NAVARRO,BENJAMIN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: CARDONE

DATE/TIME OF TEST: 03/23/2018 22:55:51

Diagnostics Check		OK	22:48
Air Blank		0.000	22:49
Control Test		0.079	22:49
Air Blank		0.000	22:49
Subject Sample	# 1	0.096	22:50
Breath Volume		1.117	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.114	22:54
Breath Volume		3.187	22:54
Air Blank		0.000	22:54
Control Test		0.078	22:55
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BOUTROS,PHILLIP
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: SCHAEFER

DATE/TIME OF TEST: 03/24/2018 21:58:42

Diagnostics Check		OK	21:51
Air Blank		0.000	21:52
Control Test		0.078	21:52
Air Blank		0.000	21:53
Subject Sample	# 1	0.230	21:53
Breath Volume		2.156	21:53
Air Blank		0.000	21:54
Air Blank		0.000	21:56
Subject Sample	# 2	0.238	21:57
Breath Volume		1.574	21:57
Air Blank		0.000	21:57
Control Test		0.078	21:58
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NAVARRO,BENJAMIN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: MARTY

DATE/TIME OF TEST: 03/25/2018 00:12:54

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.079	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.152	00:07
Breath Volume		1.671	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:09
Subject Sample	# 2	0.145	00:11
Breath Volume		1.773	00:11
Air Blank		0.000	00:11
Control Test		0.078	00:12
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NAVARRO,BENJAMIN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: COLDREN

DATE/TIME OF TEST: 03/28/2018 22:51:05

Diagnostics Check		OK	22:41
Air Blank		0.000	22:41
Control Test		0.077	22:42
Air Blank		0.000	22:42
Subject Sample	# 1	0.128	22:45
Breath Volume		1.402	22:45
Air Blank		0.000	22:46
Air Blank		0.000	22:48
Subject Sample	# 2	0.130	22:49
Breath Volume		1.820	22:49
Air Blank		0.000	22:50
Control Test		0.076	22:50
Air Blank		0.000	22:50
Diagnostics Check		OK	22:51

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HENRY,DENNIS
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: KALEY

DATE/TIME OF TEST: 03/29/2018 04:14:41

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.079	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.000	04:09
Breath Volume		1.589	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.000	04:13
Breath Volume		2.035	04:13
Air Blank		0.000	04:13
Control Test		0.079	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 03/31/2018 01:47

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CHASE,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOLTZHOUSE,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: SELF

DATE/TIME OF TEST: 03/05/2018 02:16:49

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:14	
Control Test		0.081	02:14	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Control Test		0.081	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 03/21/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ORTIZ,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: KNACK

DATE/TIME OF TEST: 03/10/2018 02:59:57

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.081	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.134	02:54
Breath Volume		1.179	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.132	02:58
Breath Volume		1.675	02:58
Air Blank		0.000	02:59
Control Test		0.078	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 03/21/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FERRER,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: BROGAN

DATE/TIME OF TEST: 03/10/2018 03:59:04

Diagnostics Check		OK	03:47	
Air Blank		0.000	03:48	
Control Test		0.079	03:48	
Air Blank		0.000	03:48	
Subject Sample	# 1		03:52	No Sample Provided
Breath Volume		0.000	03:52	
Air Blank		0.000	03:52	
Air Blank		0.000	03:54	
Subject Sample	# 2		03:57	No Sample Provided
Breath Volume		0.000	03:57	
Air Blank		0.000	03:58	
Control Test		0.079	03:58	
Air Blank		0.000	03:58	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 03/21/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 03/10/2018 04:46:54

Diagnostics Check		OK	04:39
Air Blank		0.000	04:39
Control Test		0.080	04:39
Air Blank		0.000	04:40
Subject Sample	# 1	0.146	04:41
Breath Volume		1.410	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.134	04:45
Breath Volume		1.757	04:45
Air Blank		0.000	04:45
Control Test		0.080	04:46
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 03/21/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,TROY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: JONIKA

DATE/TIME OF TEST: 03/12/2018 02:22:13

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.079	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.284	02:17
Breath Volume		2.828	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.275	02:20
Breath Volume		3.437	02:20
Air Blank		0.000	02:21
Control Test		0.078	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 03/21/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLTZHOUSE,JONATHAN,D
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: HOLTZHOUSE,JONATHAN
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: CLARI

DATE/TIME OF TEST: 03/14/2018 02:07:21

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.080	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.157	02:02
Breath Volume		3.195	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.144	02:05
Breath Volume		1.781	02:05
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 03/21/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: HALES

DATE/TIME OF TEST: 03/15/2018 23:20:50

Diagnostics Check		OK	23:09	
Air Blank		0.000	23:09	
Control Test		0.081	23:10	
Air Blank		0.000	23:10	
Subject Sample	# 1		23:13	No Sample Provided
Breath Volume		0.000	23:13	
Air Blank		0.000	23:14	
Air Blank		0.000	23:16	
Subject Sample	# 2		23:19	No Sample Provided
Breath Volume		0.000	23:19	
Air Blank		0.000	23:19	
Control Test		0.080	23:20	
Air Blank		0.000	23:20	
Diagnostics Check		OK	23:20	

Data Download Date/Time: 03/21/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DOCTOR,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: EMERY

DATE/TIME OF TEST: 03/16/2018 01:34:17

Diagnostics Check	OK	01:33	
Air Blank		01:33	Sequence Aborted
Air Blank	0.000	01:34	

Data Download Date/Time: 03/21/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CHASE,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOLTZHOUSE,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/17/2018 02:14:48

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.080	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.317	02:09
Breath Volume		2.429	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.315	02:13
Breath Volume		2.324	02:13
Air Blank		0.000	02:13
Control Test		0.078	02:14
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 03/21/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOLTZHOUSE,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: HOSKINS

DATE/TIME OF TEST: 03/17/2018 22:22:38

Diagnostics Check		OK	22:19	
Air Blank		0.000	22:20	
Control Test		0.079	22:20	
Air Blank		0.000	22:20	
Subject Sample	# 1		22:21	Subject Test Refused
Breath Volume		0.000	22:21	
Air Blank		0.000	22:21	
Control Test		0.079	22:22	
Air Blank		0.000	22:22	
Diagnostics Check		OK	22:22	

Data Download Date/Time: 03/21/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FERRER,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/21/2018 04:14:20

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:00	
Control Test		0.080	04:01	
Air Blank		0.000	04:01	
Subject Sample	# 1	0.182	04:04	Slope Not Level
Breath Volume		1.386	04:04	
Air Blank		0.000	04:05	
Air Blank		0.000	04:07	
Subject Sample	# 2	0.184	04:08	
Breath Volume		1.714	04:08	
Air Blank		0.000	04:09	
Air Blank		0.000	04:11	
Subject Sample	# 3	0.196	04:12	
Breath Volume		2.085	04:12	
Air Blank		0.000	04:13	
Control Test		0.078	04:13	
Air Blank		0.000	04:14	
Diagnostics Check		OK	04:14	

Data Download Date/Time: 03/21/2018 10:11

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CHASE,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: SAAYMAN,F
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 03/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: ECK

DATE/TIME OF TEST: 03/22/2018 01:02:02

Diagnostics Check		OK	00:57	
Air Blank		0.000	00:57	
Control Test		0.080	00:58	
Air Blank		0.000	00:58	
Subject Sample	# 1		01:01	RFI Detect
Breath Volume		0.000	01:01	
Air Blank			01:01	RFI Detect
Air Blank		0.000	01:02	

Data Download Date/Time: 04/30/2018 14:06

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: KJEKSTAD,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: DALY

DATE/TIME OF TEST: 03/25/2018 22:52:08

Diagnostics Check		OK	22:44
Air Blank		0.000	22:44
Control Test		0.079	22:44
Air Blank		0.000	22:45
Subject Sample	# 1	0.275	22:46
Breath Volume		2.507	22:46
Air Blank		0.000	22:47
Air Blank		0.000	22:48
Subject Sample	# 2	0.272	22:50
Breath Volume		2.535	22:50
Air Blank		0.000	22:51
Control Test		0.082	22:51
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

Data Download Date/Time: 04/30/2018 14:06

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WARD,TRP
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: TOMAS-LOPEZ

DATE/TIME OF TEST: 03/26/2018 01:24:32

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:22	
Control Test		0.080	01:22	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:23	Subject Test Refused
Breath Volume		0.000	01:23	
Air Blank		0.000	01:23	
Control Test		0.080	01:23	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 04/30/2018 14:06

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOLTZHOUSE,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 03/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: JONIKA

DATE/TIME OF TEST: 03/26/2018 21:28:12

Diagnostics Check		OK	21:20
Air Blank		0.000	21:20
Control Test		0.079	21:21
Air Blank		0.000	21:21
Subject Sample	# 1	0.157	21:22
Breath Volume		2.085	21:22
Air Blank		0.000	21:23
Air Blank		0.000	21:25
Subject Sample	# 2	0.170	21:26
Breath Volume		3.265	21:26
Air Blank		0.000	21:27
Control Test		0.078	21:27
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 04/30/2018 14:06

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CHASE,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: KJEKSTAD,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: BRANT

DATE/TIME OF TEST: 03/27/2018 22:40:14

Diagnostics Check		OK	22:28	
Air Blank		0.000	22:29	
Control Test		0.080	22:29	
Air Blank		0.000	22:29	
Subject Sample	# 1	0.103	22:33	Volume Not Met
Breath Volume		0.097	22:33	
Air Blank		0.000	22:33	
Air Blank		0.000	22:35	
Subject Sample	# 2	0.129	22:38	Volume Not Met
Breath Volume		0.460	22:38	
Air Blank		0.000	22:39	
Control Test		0.079	22:39	
Air Blank		0.000	22:40	
Diagnostics Check		OK	22:40	

Data Download Date/Time: 04/30/2018 14:06

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TITUS,JOSH
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 03/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: TARLETON

DATE/TIME OF TEST: 03/29/2018 15:41:50

Diagnostics Check		OK	15:38	
Air Blank		0.000	15:39	
Control Test		0.079	15:39	
Air Blank		0.000	15:40	
Subject Sample	# 1		15:40	Subject Test Refused
Breath Volume		0.000	15:40	
Air Blank		0.000	15:40	
Control Test		0.079	15:41	
Air Blank		0.000	15:41	
Diagnostics Check		OK	15:41	

Data Download Date/Time: 04/30/2018 14:06

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: HILBERER,BRIAN
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: WENTINK

DATE/TIME OF TEST: 03/03/2018 22:57:06

Diagnostics Check		OK	22:49
Air Blank		0.000	22:49
Control Test		0.080	22:50
Air Blank		0.000	22:50
Subject Sample	# 1	0.207	22:51
Breath Volume		2.148	22:51
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.214	22:55
Breath Volume		2.132	22:55
Air Blank		0.000	22:56
Control Test		0.082	22:56
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 03/09/2018 18:39

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BELL,ETHAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: RAMBHAROS

DATE/TIME OF TEST: 03/03/2018 23:36:31

Diagnostics Check		OK	23:30
Air Blank		0.000	23:30
Control Test		0.080	23:31
Air Blank		0.000	23:31
Subject Sample	# 1	0.143	23:32
Breath Volume		2.023	23:32
Air Blank		0.000	23:32
Air Blank		0.000	23:34
Subject Sample	# 2	0.143	23:34
Breath Volume		2.378	23:34
Air Blank		0.000	23:35
Control Test		0.081	23:35
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 03/09/2018 18:39

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,B
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: BROUGHTON

DATE/TIME OF TEST: 03/04/2018 03:21:39

Diagnostics Check		OK	03:14
Air Blank		0.000	03:15
Control Test		0.081	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.092	03:17
Breath Volume		3.250	03:17
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.089	03:20
Breath Volume		3.472	03:20
Air Blank		0.000	03:20
Control Test		0.081	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 03/09/2018 18:39

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGDON,JOHN,K
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: VINSON,HEATHER
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: SHELL

DATE/TIME OF TEST: 03/10/2018 18:34:37

Diagnostics Check		OK	18:31	
Air Blank		0.000	18:31	
Control Test		0.080	18:32	
Air Blank		0.000	18:32	
Subject Sample	# 1		18:33	Subject Test Refused
Breath Volume		0.000	18:33	
Air Blank		0.000	18:33	
Control Test		0.079	18:34	
Air Blank		0.000	18:34	
Diagnostics Check		OK	18:34	

Data Download Date/Time: 04/07/2018 22:37

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON,AMY,B
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: DEVINE,MICHAEL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: DOBBS

DATE/TIME OF TEST: 03/12/2018 03:17:26

Diagnostics Check	OK	03:16	
Air Blank		03:16	RFI Detect
Air Blank	0.000	03:17	

Data Download Date/Time: 04/07/2018 22:37

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON,AMY,B
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: DEVINE,MICHAEL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: DOBBS

DATE/TIME OF TEST: 03/12/2018 03:26:24

Diagnostics Check		OK	03:19
Air Blank		0.000	03:19
Control Test		0.080	03:20
Air Blank		0.000	03:20
Subject Sample	# 1	0.168	03:21
Breath Volume		2.636	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.159	03:24
Breath Volume		1.925	03:24
Air Blank		0.000	03:25
Control Test		0.080	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 04/07/2018 22:37

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERDIKES,PANDELIS,C
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: HILBERER,BRIAN
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: JOYNER

DATE/TIME OF TEST: 03/17/2018 03:56:22

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.080	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.160	03:49
Breath Volume		1.714	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.155	03:54
Breath Volume		2.785	03:54
Air Blank		0.000	03:55
Control Test		0.081	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 04/07/2018 22:37

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON,AMY,B
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: DEVINE,MICHAEL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 03/21/2018 02:13:41

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.080	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.131	02:09
Breath Volume		2.605	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.134	02:12
Breath Volume		3.593	02:12
Air Blank		0.000	02:12
Control Test		0.081	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 04/07/2018 22:37

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,S,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PERDIKES,P
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: HINTON

DATE/TIME OF TEST: 03/21/2018 22:59:59

Diagnostics Check		OK	22:53
Air Blank		0.000	22:54
Control Test		0.079	22:54
Air Blank		0.000	22:54
Subject Sample	# 1	0.159	22:55
Breath Volume		3.230	22:55
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.154	22:58
Breath Volume		2.632	22:58
Air Blank		0.000	22:59
Control Test		0.080	22:59
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 04/07/2018 22:37

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GOOD

DATE/TIME OF TEST: 03/22/2018 01:57:53

Diagnostics Check		OK	01:50
Air Blank		0.000	01:51
Control Test		0.080	01:51
Air Blank		0.000	01:52
Subject Sample	# 1	0.163	01:53
Breath Volume		1.625	01:53
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.168	01:56
Breath Volume		1.554	01:56
Air Blank		0.000	01:56
Control Test		0.079	01:57
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 04/07/2018 22:37

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON,AMY,B
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: SANDBAKKEN,PAUL
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 03/24/2018 01:51:30

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:38	
Control Test		0.079	01:38	
Air Blank		0.000	01:39	
Subject Sample	# 1	0.115	01:42	Volume Not Met
Breath Volume		0.492	01:42	
Air Blank		0.000	01:43	
Air Blank		0.000	01:44	
Subject Sample	# 2	0.160	01:46	
Breath Volume		1.640	01:46	
Air Blank		0.000	01:47	
Air Blank		0.000	01:49	
Subject Sample	# 3	0.158	01:50	
Breath Volume		1.808	01:50	
Air Blank		0.000	01:50	
Control Test		0.079	01:50	
Air Blank		0.000	01:51	
Diagnostics Check		OK	01:51	

Data Download Date/Time: 04/07/2018 22:37

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON,AMY,B
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: DEVINE,MICHAEL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: POWERS

DATE/TIME OF TEST: 03/24/2018 02:23:01

Diagnostics Check		OK	02:20	
Air Blank		0.000	02:20	
Control Test		0.079	02:20	
Air Blank		0.000	02:21	
Subject Sample	# 1		02:21	Subject Test Refused
Breath Volume		0.000	02:21	
Air Blank		0.000	02:22	
Control Test		0.079	02:22	
Air Blank		0.000	02:22	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 04/07/2018 22:38

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,ALEC,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:34

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 03/24/2018 07:02:06

Diagnostics Check		OK	06:55
Air Blank		0.000	06:55
Control Test		0.080	06:55
Air Blank		0.000	06:56
Subject Sample	# 1	0.153	06:57
Breath Volume		2.410	06:57
Air Blank		0.000	06:58
Air Blank		0.000	07:00
Subject Sample	# 2	0.152	07:00
Breath Volume		1.414	07:00
Air Blank		0.000	07:01
Control Test		0.079	07:01
Air Blank		0.000	07:02
Diagnostics Check		OK	07:02

Data Download Date/Time: 04/07/2018 22:38

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON,AMY,B
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: DEVINE,MICHAEL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: CAMPOS

DATE/TIME OF TEST: 03/26/2018 02:58:32

Diagnostics Check		OK	02:55	
Air Blank		0.000	02:56	
Control Test		0.080	02:56	
Air Blank		0.000	02:56	
Subject Sample	# 1		02:57	Subject Test Refused
Breath Volume		0.000	02:57	
Air Blank		0.000	02:57	
Control Test		0.080	02:57	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 04/07/2018 22:38

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUROUMOUSIS,GEORGE,T
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: KOUROUMOUSIS,GEORGE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 03/27/2018 06:23:01

Diagnostics Check		OK	06:20	
Air Blank		0.000	06:20	
Control Test		0.079	06:20	
Air Blank		0.000	06:21	
Subject Sample	# 1		06:21	Subject Test Refused
Breath Volume		0.000	06:21	
Air Blank		0.000	06:22	
Control Test		0.080	06:22	
Air Blank		0.000	06:22	
Diagnostics Check		OK	06:23	

Data Download Date/Time: 04/07/2018 22:38

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,ALEC,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/31/2018 01:43:00

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:39	
Control Test		0.079	01:39	
Air Blank		0.000	01:39	
Subject Sample	# 1		01:41	Subject Test Refused
Breath Volume		0.000	01:41	
Air Blank		0.000	01:42	
Control Test		0.080	01:42	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:43	

Data Download Date/Time: 04/07/2018 22:38

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 03/10/2018 02:37:10

Diagnostics Check		OK	02:27
Air Blank		0.000	02:28
Control Test		0.081	02:28
Air Blank		0.000	02:28
Subject Sample	# 1	0.142	02:31
Breath Volume		1.414	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.145	02:35
Breath Volume		1.910	02:35
Air Blank		0.000	02:36
Control Test		0.081	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 04/07/2018 22:46

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: TOLER

DATE/TIME OF TEST: 03/10/2018 03:44:46

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.082	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 1		03:44	Sequence Aborted
Breath Volume		0.000	03:44	
Air Blank		0.000	03:44	

Data Download Date/Time: 04/07/2018 22:46

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,B
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: TOLER

DATE/TIME OF TEST: 03/10/2018 04:12:15

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.081	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.176	04:04
Breath Volume		1.550	04:04
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.197	04:07
Breath Volume		2.789	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 3	0.197	04:10
Breath Volume		2.886	04:10
Air Blank		0.000	04:11
Control Test		0.081	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 04/07/2018 22:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/15/2018 21:31:57

Diagnostics Check		OK	21:25
Air Blank		0.000	21:25
Control Test		0.081	21:26
Air Blank		0.000	21:26
Subject Sample	# 1	0.041	21:27
Breath Volume		1.683	21:27
Air Blank		0.000	21:27
Air Blank		0.000	21:29
Subject Sample	# 2	0.045	21:30
Breath Volume		2.085	21:30
Air Blank		0.000	21:31
Control Test		0.080	21:31
Air Blank		0.000	21:31
Diagnostics Check		OK	21:31

Data Download Date/Time: 04/07/2018 22:46

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: BARBIN

DATE/TIME OF TEST: 03/16/2018 19:46:57

Diagnostics Check		OK	19:40
Air Blank		0.000	19:41
Control Test		0.080	19:41
Air Blank		0.000	19:41
Subject Sample	# 1	0.162	19:42
Breath Volume		2.492	19:42
Air Blank		0.000	19:43
Air Blank		0.000	19:44
Subject Sample	# 2	0.160	19:45
Breath Volume		2.757	19:45
Air Blank		0.000	19:46
Control Test		0.081	19:46
Air Blank		0.000	19:46
Diagnostics Check		OK	19:46

Data Download Date/Time: 04/07/2018 22:46

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BORRELLI

DATE/TIME OF TEST: 03/16/2018 23:20:10

Diagnostics Check		OK	23:13
Air Blank		0.000	23:13
Control Test		0.082	23:14
Air Blank		0.000	23:14
Subject Sample	# 1	0.079	23:15
Breath Volume		1.781	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:17
Subject Sample	# 2	0.080	23:18
Breath Volume		2.152	23:18
Air Blank		0.000	23:19
Control Test		0.081	23:19
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 04/07/2018 22:46

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 03/24/2018 02:33:08

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.082	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.079	02:28
Breath Volume		3.054	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.077	02:31
Breath Volume		3.050	02:31
Air Blank		0.000	02:32
Control Test		0.082	02:32
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 04/07/2018 22:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: KINNEY,CARRIANNE
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SEIGLER

DATE/TIME OF TEST: 03/06/2018 22:16:17

Diagnostics Check		OK	22:09
Air Blank		0.000	22:10
Control Test		0.081	22:10
Air Blank		0.000	22:10
Subject Sample	# 1	0.110	22:11
Breath Volume		2.613	22:11
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.109	22:14
Breath Volume		1.781	22:14
Air Blank		0.000	22:15
Control Test		0.080	22:15
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 03/09/2018 03:37

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: EDWARDS,KEVIN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: MOBLEY

DATE/TIME OF TEST: 03/15/2018 01:20:50

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.081	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.154	01:15
Breath Volume		1.207	01:15
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.136	01:19
Breath Volume		1.660	01:19
Air Blank		0.000	01:19
Control Test		0.079	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 04/15/2018 11:18

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: EDWARDS,KEVIN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: SFORZA

DATE/TIME OF TEST: 03/15/2018 21:19:47

Diagnostics Check		OK	21:16	
Air Blank		0.000	21:17	
Control Test		0.082	21:17	
Air Blank		0.000	21:18	
Subject Sample	# 1		21:18	Subject Test Refused
Breath Volume		0.000	21:18	
Air Blank		0.000	21:18	
Control Test		0.081	21:19	
Air Blank		0.000	21:19	
Diagnostics Check		OK	21:19	

Data Download Date/Time: 04/15/2018 11:18

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOCH,ADAM,G
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: TAYLOR,KEITH
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 583457
DRY GAS EXPIRY DATE: 10/20/2019
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 03/10/2018 23:07:57

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.078	23:02
Air Blank		0.000	23:02
Subject Sample	# 1	0.229	23:03
Breath Volume		1.937	23:03
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.236	23:06
Breath Volume		3.375	23:06
Air Blank		0.000	23:07
Control Test		0.077	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 03/29/2018 03:30

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,A,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: TAYLOR,K
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 583457
DRY GAS EXPIRY DATE: 10/20/2019
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: CANTRELL

DATE/TIME OF TEST: 03/20/2018 23:48:06

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:45	
Control Test		0.078	23:46	
Air Blank		0.000	23:46	
Subject Sample	# 1		23:46	Subject Test Refused
Breath Volume		0.000	23:46	
Air Blank		0.000	23:47	
Control Test		0.079	23:47	
Air Blank		0.000	23:48	
Diagnostics Check		OK	23:48	

Data Download Date/Time: 03/29/2018 03:30

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,KEITH,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: KEMENY,CHRISTA
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 583457
DRY GAS EXPIRY DATE: 10/20/2019
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: HARBIN

DATE/TIME OF TEST: 03/24/2018 23:24:10

Diagnostics Check		OK	23:17
Air Blank		0.000	23:18
Control Test		0.079	23:18
Air Blank		0.000	23:19
Subject Sample	# 1	0.196	23:19
Breath Volume		2.539	23:19
Air Blank		0.000	23:20
Air Blank		0.000	23:22
Subject Sample	# 2	0.184	23:22
Breath Volume		1.445	23:22
Air Blank		0.000	23:23
Control Test		0.080	23:23
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 03/29/2018 03:30

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,BRANDON,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MEAD,R
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WOLFORD

DATE/TIME OF TEST: 03/01/2018 21:55:47

Diagnostics Check		OK	21:52	
Air Blank		0.000	21:53	
Control Test		0.081	21:53	
Air Blank		0.000	21:54	
Subject Sample	# 1		21:54	Subject Test Refused
Breath Volume		0.000	21:54	
Air Blank		0.000	21:54	
Control Test		0.080	21:55	
Air Blank		0.000	21:55	
Diagnostics Check		OK	21:55	

Data Download Date/Time: 03/22/2018 13:36

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,LINSEY,R
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MENDOZA,LINSEY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: PERALES

DATE/TIME OF TEST: 03/04/2018 23:27:33

Diagnostics Check		OK	23:21
Air Blank		0.000	23:21
Control Test		0.080	23:22
Air Blank		0.000	23:22
Subject Sample	# 1	0.019	23:23
Breath Volume		2.578	23:23
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.017	23:26
Breath Volume		2.640	23:26
Air Blank		0.000	23:26
Control Test		0.078	23:26
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 03/22/2018 13:36

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,MEGAN,K
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: SULLIVAN,MEGAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ARANCA

DATE/TIME OF TEST: 03/05/2018 01:24:32

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.080	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.042	01:18
Breath Volume		1.503	01:18
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.042	01:23
Breath Volume		2.234	01:23
Air Blank		0.000	01:23
Control Test		0.080	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 03/22/2018 13:36

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,MEGAN,K
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: SULLIVAN,MEGAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: CUNI

DATE/TIME OF TEST: 03/08/2018 02:52:03

Diagnostics Check		OK	02:45
Air Blank		0.000	02:46
Control Test		0.082	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.000	02:47
Breath Volume		2.269	02:47
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.000	02:50
Breath Volume		2.890	02:50
Air Blank		0.000	02:51
Control Test		0.082	02:51
Air Blank		0.000	02:51
Diagnostics Check		OK	02:52

Data Download Date/Time: 03/22/2018 13:36

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 03/08/2018 04:47:21

Diagnostics Check		OK	04:40
Air Blank		0.000	04:41
Control Test		0.081	04:41
Air Blank		0.000	04:42
Subject Sample	# 1	0.217	04:42
Breath Volume		2.191	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.205	04:45
Breath Volume		1.527	04:45
Air Blank		0.000	04:46
Control Test		0.081	04:46
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 03/22/2018 13:36

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORFEL,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WORFEL,KYLE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/08/2018 14:55:03

Diagnostics Check		OK	14:52	
Air Blank		0.000	14:52	
Control Test		0.082	14:52	
Air Blank		0.000	14:53	
Subject Sample	# 1		14:53	Subject Test Refused
Breath Volume		0.000	14:53	
Air Blank		0.000	14:54	
Control Test		0.081	14:54	
Air Blank		0.000	14:54	
Diagnostics Check		OK	14:55	

Data Download Date/Time: 03/22/2018 13:36

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,MEGAN,K
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: SULLIVAN,MEGAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 03/09/2018 03:32:44

Diagnostics Check		OK	03:21	
Air Blank		0.000	03:21	
Control Test		0.082	03:22	
Air Blank		0.000	03:22	
Subject Sample	# 1		03:25	No Sample Provided
Breath Volume		0.000	03:25	
Air Blank		0.000	03:26	
Air Blank		0.000	03:28	
Subject Sample	# 2		03:31	No Sample Provided
Breath Volume		0.000	03:31	
Air Blank		0.000	03:31	
Control Test		0.082	03:32	
Air Blank		0.000	03:32	
Diagnostics Check		OK	03:32	

Data Download Date/Time: 03/22/2018 13:36

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: CHANCEY,SAMANTHA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: POLLARD

DATE/TIME OF TEST: 03/11/2018 04:04:16

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:51	
Control Test		0.080	03:51	
Air Blank		0.000	03:52	
Subject Sample	# 1	0.093	03:55	Volume Not Met
Breath Volume		0.753	03:55	
Air Blank		0.000	03:56	
Air Blank		0.000	03:57	
Subject Sample	# 2	0.096	03:58	
Breath Volume		1.714	03:58	
Air Blank		0.000	03:59	
Air Blank		0.000	04:00	
Subject Sample	# 3	0.084	04:02	
Breath Volume		1.113	04:02	
Air Blank		0.000	04:03	
Control Test		0.079	04:03	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 03/22/2018 13:36

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTS,C,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BATTS,C
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: BUCKNER JR

DATE/TIME OF TEST: 03/11/2018 23:06:05

Diagnostics Check		OK	22:59
Air Blank		0.000	22:59
Control Test		0.079	22:59
Air Blank		0.000	23:00
Subject Sample	# 1	0.223	23:01
Breath Volume		2.675	23:01
Air Blank		0.000	23:02
Air Blank		0.000	23:03
Subject Sample	# 2	0.214	23:04
Breath Volume		2.015	23:04
Air Blank		0.000	23:05
Control Test		0.078	23:05
Air Blank		0.000	23:05
Diagnostics Check		OK	23:06

Data Download Date/Time: 03/22/2018 13:36

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,KYLE,W
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BROWN,KYLE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/12/2018 00:50:34

Diagnostics Check		OK	00:45	
Air Blank		0.000	00:45	
Control Test		0.080	00:46	
Air Blank		0.000	00:46	
Subject Sample	# 1		00:49	Subject Test Refused
Breath Volume		0.000	00:49	
Air Blank		0.000	00:49	
Control Test		0.079	00:49	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 03/22/2018 13:37

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: O'NEAL

DATE/TIME OF TEST: 03/16/2018 03:09:00

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:06	
Control Test		0.081	03:06	
Air Blank		0.000	03:07	
Subject Sample	# 1		03:07	Subject Test Refused
Breath Volume		0.000	03:07	
Air Blank		0.000	03:08	
Control Test		0.081	03:08	
Air Blank		0.000	03:08	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 03/22/2018 13:37

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: SAMMONS

DATE/TIME OF TEST: 03/17/2018 00:06:24

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.080	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.172	00:01
Breath Volume		3.296	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.172	00:04
Breath Volume		3.437	00:04
Air Blank		0.000	00:05
Control Test		0.081	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 03/22/2018 13:37

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,LINSEY,R
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MENDOZA,LINSEY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 03/17/2018 01:05:31

Diagnostics Check		OK	00:53	
Air Blank		0.000	00:53	
Control Test		0.080	00:54	
Air Blank		0.000	00:54	
Subject Sample	# 1	0.128	00:57	Volume Not Met
Breath Volume		0.742	00:57	
Air Blank		0.000	00:58	
Air Blank		0.000	01:00	
Subject Sample	# 2	0.154	01:00	
Breath Volume		1.160	01:00	
Air Blank		0.000	01:01	
Air Blank		0.000	01:03	
Subject Sample	# 3	0.155	01:03	
Breath Volume		1.839	01:03	
Air Blank		0.000	01:04	
Control Test		0.077	01:04	
Air Blank		0.000	01:05	
Diagnostics Check		OK	01:05	

Data Download Date/Time: 03/22/2018 13:37

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: TURGEON

DATE/TIME OF TEST: 03/17/2018 01:47:40

Diagnostics Check		OK	01:41
Air Blank		0.000	01:41
Control Test		0.080	01:42
Air Blank		0.000	01:42
Subject Sample	# 1	0.114	01:43
Breath Volume		2.933	01:43
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.113	01:46
Breath Volume		2.757	01:46
Air Blank		0.000	01:46
Control Test		0.079	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 03/22/2018 13:37

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: YORK

DATE/TIME OF TEST: 03/18/2018 20:55:00

Diagnostics Check	OK	20:54	
Air Blank		20:54	RFI Detect
Air Blank	0.000	20:55	

Data Download Date/Time: 03/22/2018 13:37

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: YORK

DATE/TIME OF TEST: 03/18/2018 21:02:29

Diagnostics Check		OK	20:56
Air Blank		0.000	20:56
Control Test		0.080	20:57
Air Blank		0.000	20:57
Subject Sample	# 1	0.000	20:58
Breath Volume		4.203	20:58
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.000	21:01
Breath Volume		3.859	21:01
Air Blank		0.000	21:01
Control Test		0.080	21:01
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 03/22/2018 13:37

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GERNERT,JOHN,D
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ESQUIVEL,ANTHONY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: AKINS

DATE/TIME OF TEST: 03/20/2018 19:04:55

Diagnostics Check		OK	18:57
Air Blank		0.000	18:57
Control Test		0.080	18:58
Air Blank		0.000	18:58
Subject Sample	# 1	0.229	18:59
Breath Volume		2.082	18:59
Air Blank		0.000	19:00
Air Blank		0.000	19:02
Subject Sample	# 2	0.222	19:03
Breath Volume		1.417	19:03
Air Blank		0.000	19:04
Control Test		0.078	19:04
Air Blank		0.000	19:04
Diagnostics Check		OK	19:04

Data Download Date/Time: 03/22/2018 13:37

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIERRA,P 51736
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:33

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 03/01/2018 07:03:10

Diagnostics Check		OK	06:53
Air Blank		0.000	06:54
Control Test		0.080	06:54
Air Blank		0.000	06:55
Subject Sample	# 1	0.166	06:57
Breath Volume		1.429	06:57
Air Blank		0.000	06:58
Air Blank		0.000	07:00
Subject Sample	# 2	0.174	07:01
Breath Volume		2.117	07:01
Air Blank		0.000	07:02
Control Test		0.077	07:02
Air Blank		0.000	07:03
Diagnostics Check		OK	07:03

Data Download Date/Time: 03/07/2018 07:17

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUTIERREZ,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: BLY

DATE/TIME OF TEST: 03/01/2018 16:42:44

Diagnostics Check		OK	16:36
Air Blank		0.000	16:37
Control Test		0.079	16:37
Air Blank		0.000	16:37
Subject Sample	# 1	0.249	16:38
Breath Volume		1.953	16:38
Air Blank		0.000	16:39
Air Blank		0.000	16:40
Subject Sample	# 2	0.247	16:41
Breath Volume		2.875	16:41
Air Blank		0.000	16:41
Control Test		0.078	16:42
Air Blank		0.000	16:42
Diagnostics Check		OK	16:42

Data Download Date/Time: 03/07/2018 07:17

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 03/01/2018 23:44:25

Diagnostics Check		OK	23:37
Air Blank		0.000	23:37
Control Test		0.080	23:37
Air Blank		0.000	23:38
Subject Sample	# 1	0.131	23:39
Breath Volume		1.882	23:39
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.139	23:42
Breath Volume		2.902	23:42
Air Blank		0.000	23:43
Control Test		0.079	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 03/07/2018 07:17

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 03/02/2018 02:45:40

Diagnostics Check		OK	02:38
Air Blank		0.000	02:39
Control Test		0.080	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.145	02:41
Breath Volume		2.839	02:41
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.142	02:44
Breath Volume		2.734	02:44
Air Blank		0.000	02:44
Control Test		0.079	02:45
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 03/07/2018 07:17

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALE,W
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: TROKANSKI

DATE/TIME OF TEST: 03/02/2018 21:45:10

Diagnostics Check		OK	21:38
Air Blank		0.000	21:38
Control Test		0.080	21:38
Air Blank		0.000	21:39
Subject Sample	# 1	0.115	21:40
Breath Volume		1.738	21:40
Air Blank		0.000	21:41
Air Blank		0.000	21:42
Subject Sample	# 2	0.118	21:43
Breath Volume		1.847	21:43
Air Blank		0.000	21:44
Control Test		0.079	21:44
Air Blank		0.000	21:45
Diagnostics Check		OK	21:45

Data Download Date/Time: 03/07/2018 07:17

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: YCAZA

DATE/TIME OF TEST: 03/02/2018 23:24:09

Diagnostics Check		OK	23:17
Air Blank		0.000	23:17
Control Test		0.079	23:18
Air Blank		0.000	23:18
Subject Sample	# 1	0.180	23:19
Breath Volume		2.144	23:19
Air Blank		0.000	23:20
Air Blank		0.000	23:22
Subject Sample	# 2	0.183	23:22
Breath Volume		2.273	23:22
Air Blank		0.000	23:23
Control Test		0.080	23:23
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 03/07/2018 07:17

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: MAYBERRY

DATE/TIME OF TEST: 03/03/2018 20:11:06

Diagnostics Check		OK	20:05
Air Blank		0.000	20:05
Control Test		0.081	20:05
Air Blank		0.000	20:06
Subject Sample	# 1	0.181	20:06
Breath Volume		2.605	20:06
Air Blank		0.000	20:07
Air Blank		0.000	20:09
Subject Sample	# 2	0.184	20:09
Breath Volume		2.457	20:09
Air Blank		0.000	20:10
Control Test		0.080	20:10
Air Blank		0.000	20:11
Diagnostics Check		OK	20:11

Data Download Date/Time: 03/07/2018 07:17

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 03/03/2018 21:07:37

Diagnostics Check		OK	21:01
Air Blank		0.000	21:01
Control Test		0.081	21:02
Air Blank		0.000	21:02
Subject Sample	# 1	0.253	21:03
Breath Volume		1.574	21:03
Air Blank		0.000	21:03
Air Blank		0.000	21:05
Subject Sample	# 2	0.254	21:06
Breath Volume		1.847	21:06
Air Blank		0.000	21:06
Control Test		0.080	21:07
Air Blank		0.000	21:07
Diagnostics Check		OK	21:07

Data Download Date/Time: 03/07/2018 07:17

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,L
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: BAUTISTA ORTIZ

DATE/TIME OF TEST: 03/04/2018 01:14:55

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.081	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.247	01:10
Breath Volume		2.226	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:12
Subject Sample	# 2	0.243	01:13
Breath Volume		1.652	01:13
Air Blank		0.000	01:14
Control Test		0.079	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 03/07/2018 07:17

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: QUINTANA CARMONA

DATE/TIME OF TEST: 03/04/2018 03:24:57

Diagnostics Check		OK	03:17
Air Blank		0.000	03:17
Control Test		0.080	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.138	03:19
Breath Volume		2.203	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.125	03:23
Breath Volume		1.960	03:23
Air Blank		0.000	03:24
Control Test		0.078	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 03/07/2018 07:17

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: BROWN JR.

DATE/TIME OF TEST: 03/04/2018 04:18:09

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.080	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.144	04:13
Breath Volume		1.246	04:13
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.154	04:16
Breath Volume		2.050	04:16
Air Blank		0.000	04:17
Control Test		0.079	04:17
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 03/07/2018 07:17

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 03/04/2018 20:25:11

Diagnostics Check		OK	20:18
Air Blank		0.000	20:18
Control Test		0.081	20:18
Air Blank		0.000	20:19
Subject Sample	# 1	0.088	20:20
Breath Volume		1.890	20:20
Air Blank		0.000	20:20
Air Blank		0.000	20:22
Subject Sample	# 2	0.083	20:23
Breath Volume		2.488	20:23
Air Blank		0.000	20:24
Control Test		0.081	20:24
Air Blank		0.000	20:25
Diagnostics Check		OK	20:25

Data Download Date/Time: 03/07/2018 07:17

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: SEIJA

DATE/TIME OF TEST: 03/05/2018 04:59:14

Diagnostics Check		OK	04:52
Air Blank		0.000	04:53
Control Test		0.080	04:53
Air Blank		0.000	04:54
Subject Sample	# 1	0.166	04:54
Breath Volume		1.902	04:54
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.161	04:57
Breath Volume		2.500	04:57
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 03/07/2018 07:18

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: CRUTCHER

DATE/TIME OF TEST: 03/08/2018 03:57:41

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.080	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.093	03:53
Breath Volume		1.792	03:53
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.098	03:56
Breath Volume		1.980	03:56
Air Blank		0.000	03:56
Control Test		0.080	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 04/05/2018 19:27

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: SCOPPETTUOLO

DATE/TIME OF TEST: 03/09/2018 03:23:29

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.080	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.186	03:18
Breath Volume		2.578	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.184	03:21
Breath Volume		1.777	03:21
Air Blank		0.000	03:22
Control Test		0.080	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 04/05/2018 19:27

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: BARNETT II

DATE/TIME OF TEST: 03/09/2018 04:24:07

Diagnostics Check		OK	04:17
Air Blank		0.000	04:18
Control Test		0.080	04:18
Air Blank		0.000	04:19
Subject Sample	# 1	0.226	04:19
Breath Volume		1.449	04:19
Air Blank		0.000	04:20
Air Blank		0.000	04:22
Subject Sample	# 2	0.208	04:22
Breath Volume		1.402	04:22
Air Blank		0.000	04:23
Control Test		0.079	04:23
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 04/05/2018 19:27

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: DEAKIN

DATE/TIME OF TEST: 03/10/2018 04:17:56

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.081	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.118	04:13
Breath Volume		1.832	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:15
Subject Sample	# 2	0.126	04:16
Breath Volume		1.460	04:16
Air Blank		0.000	04:17
Control Test		0.080	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 04/05/2018 19:27

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: HORTON

DATE/TIME OF TEST: 03/10/2018 05:16:37

Diagnostics Check		OK	05:10
Air Blank		0.000	05:10
Control Test		0.080	05:10
Air Blank		0.000	05:11
Subject Sample	# 1	0.118	05:11
Breath Volume		1.718	05:11
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.124	05:15
Breath Volume		1.253	05:15
Air Blank		0.000	05:15
Control Test		0.079	05:16
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 04/05/2018 19:27

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HARVESTER,W
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: ROSALEZ

DATE/TIME OF TEST: 03/11/2018 01:02:43

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.080	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.194	00:56
Breath Volume		1.171	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.212	01:01
Breath Volume		2.023	01:01
Air Blank		0.000	01:01
Control Test		0.079	01:02
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 04/05/2018 19:27

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 03/11/2018 01:59:57

Diagnostics Check		OK	01:53
Air Blank		0.000	01:53
Control Test		0.078	01:54
Air Blank		0.000	01:54
Subject Sample	# 1	0.094	01:55
Breath Volume		2.445	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:57
Subject Sample	# 2	0.098	01:58
Breath Volume		2.078	01:58
Air Blank		0.000	01:59
Control Test		0.078	01:59
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 04/05/2018 19:27

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,L
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: CARAWAY

DATE/TIME OF TEST: 03/11/2018 03:25:47

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.078	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.191	03:21
Breath Volume		1.785	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.184	03:24
Breath Volume		1.687	03:24
Air Blank		0.000	03:24
Control Test		0.079	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,',A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,L
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: CARAWAY

DATE/TIME OF TEST: 03/11/2018 05:30:48

Diagnostics Check		OK	05:24
Air Blank		0.000	05:24
Control Test		0.079	05:24
Air Blank		0.000	05:25
Subject Sample	# 1	0.147	05:26
Breath Volume		1.207	05:26
Air Blank		0.000	05:26
Air Blank		0.000	05:28
Subject Sample	# 2	0.150	05:29
Breath Volume		2.242	05:29
Air Blank		0.000	05:29
Control Test		0.079	05:30
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/12/2018 00:12:08

Diagnostics Check		OK	23:57	
Air Blank		0.000	23:57	
Control Test		0.080	23:58	
Air Blank		0.000	23:58	
Subject Sample	# 1	0.148	00:01	Volume Not Met
Breath Volume		0.101	00:01	
Air Blank		0.000	00:02	
Air Blank		0.000	00:04	
Subject Sample	# 2	0.183	00:06	
Breath Volume		1.851	00:06	
Air Blank		0.000	00:07	
Air Blank		0.000	00:08	
Subject Sample	# 3	0.174	00:10	
Breath Volume		1.765	00:10	
Air Blank		0.000	00:11	
Control Test		0.078	00:11	
Air Blank		0.000	00:12	
Diagnostics Check		OK	00:12	

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 03/14/2018 03:45:20

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:31	
Control Test		0.080	03:31	
Air Blank		0.000	03:32	
Subject Sample	# 1	0.126	03:35	
Breath Volume		1.253	03:35	
Air Blank		0.000	03:35	
Air Blank		0.000	03:37	
Subject Sample	# 2	0.118	03:40	Volume Not Met
Breath Volume		0.460	03:40	
Air Blank		0.000	03:41	
Air Blank		0.000	03:43	
Subject Sample	# 3	0.135	03:43	
Breath Volume		1.632	03:43	
Air Blank		0.000	03:44	
Control Test		0.080	03:44	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: MILNES

DATE/TIME OF TEST: 03/16/2018 04:04:09

Diagnostics Check		OK	03:57
Air Blank		0.000	03:58
Control Test		0.081	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.168	03:59
Breath Volume		2.574	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.179	04:02
Breath Volume		2.402	04:02
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HILES,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: REECE

DATE/TIME OF TEST: 03/17/2018 00:27:59

Diagnostics Check		OK	00:20
Air Blank		0.000	00:21
Control Test		0.080	00:21
Air Blank		0.000	00:22
Subject Sample	# 1	0.115	00:23
Breath Volume		2.167	00:23
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.113	00:26
Breath Volume		2.480	00:26
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/17/2018 03:23:29

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.080	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.129	03:18
Breath Volume		3.601	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.132	03:21
Breath Volume		3.558	03:21
Air Blank		0.000	03:22
Control Test		0.080	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: BUCKNER

DATE/TIME OF TEST: 03/17/2018 03:48:49

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.080	03:43
Air Blank		0.000	03:43
Subject Sample	# 1	0.169	03:44
Breath Volume		1.535	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.180	03:47
Breath Volume		2.199	03:47
Air Blank		0.000	03:47
Control Test		0.079	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 03/17/2018 06:02:56

Diagnostics Check		OK	05:56
Air Blank		0.000	05:57
Control Test		0.080	05:57
Air Blank		0.000	05:57
Subject Sample	# 1	0.176	05:58
Breath Volume		1.730	05:58
Air Blank		0.000	05:59
Air Blank		0.000	06:01
Subject Sample	# 2	0.175	06:01
Breath Volume		1.640	06:01
Air Blank		0.000	06:02
Control Test		0.079	06:02
Air Blank		0.000	06:02
Diagnostics Check		OK	06:02

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: VALDEZ

DATE/TIME OF TEST: 03/18/2018 01:26:15

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.080	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.220	01:20
Breath Volume		1.566	01:20
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.231	01:24
Breath Volume		2.398	01:24
Air Blank		0.000	01:25
Control Test		0.078	01:25
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: NEDD JR

DATE/TIME OF TEST: 03/18/2018 03:51:33

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.079	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.101	03:46
Breath Volume		1.644	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.107	03:50
Breath Volume		3.257	03:50
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMS,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: JUSTICE JR

DATE/TIME OF TEST: 03/19/2018 00:41:49

Diagnostics Check		OK	00:27	
Air Blank		0.000	00:28	
Control Test		0.079	00:28	
Air Blank		0.000	00:28	
Subject Sample	# 1	0.180	00:29	
Breath Volume		1.320	00:29	
Air Blank		0.000	00:30	
Air Blank		0.000	00:32	
Subject Sample	# 2	0.152	00:35	Volume Not Met
Breath Volume		0.210	00:35	
Air Blank		0.000	00:36	
Air Blank		0.000	00:38	
Subject Sample	# 3	0.175	00:40	
Breath Volume		2.511	00:40	
Air Blank		0.000	00:40	
Control Test		0.078	00:41	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ',CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: WYLIE

DATE/TIME OF TEST: 03/19/2018 01:19:51

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.079	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.179	01:15
Breath Volume		1.996	01:15
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.179	01:18
Breath Volume		1.722	01:18
Air Blank		0.000	01:18
Control Test		0.078	01:19
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAP,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: HERNANDEZ PEREZ

DATE/TIME OF TEST: 03/23/2018 03:43:33

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.081	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.188	03:38
Breath Volume		2.117	03:38
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.194	03:42
Breath Volume		2.671	03:42
Air Blank		0.000	03:42
Control Test		0.078	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/23/2018 05:06:28

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.079	05:01
Air Blank		0.000	05:01
Subject Sample	# 1	0.152	05:02
Breath Volume		2.640	05:02
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.142	05:04
Breath Volume		2.347	05:04
Air Blank		0.000	05:05
Control Test		0.080	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARRELL,Q
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 03/25/2018 02:57:17

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.080	02:51
Air Blank		0.000	02:52
Subject Sample	# 1	0.169	02:52
Breath Volume		2.550	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.164	02:55
Breath Volume		1.957	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: WEBBER

DATE/TIME OF TEST: 03/30/2018 01:49:17

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.080	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.210	01:42
Breath Volume		1.558	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.208	01:47
Breath Volume		1.500	01:47
Air Blank		0.000	01:48
Control Test		0.080	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 04/05/2018 19:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMS,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: KOPER

DATE/TIME OF TEST: 03/31/2018 03:39:48

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.080	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.183	03:35
Breath Volume		3.589	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.187	03:38
Breath Volume		3.644	03:38
Air Blank		0.000	03:38
Control Test		0.080	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: GIRON RIVERA

DATE/TIME OF TEST: 03/31/2018 05:05:17

Diagnostics Check		OK	04:52	
Air Blank		0.000	04:53	
Control Test		0.080	04:53	
Air Blank		0.000	04:53	
Subject Sample	# 1	0.147	04:54	No .020 agreement
Breath Volume		1.140	04:54	
Air Blank		0.000	04:55	
Air Blank		0.000	04:57	
Subject Sample	# 2	0.169	04:58	No .020 agreement
Breath Volume		1.953	04:58	
Air Blank		0.000	04:58	
Air Blank		0.000	05:00	
Subject Sample	# 3		05:03	No Sample Provided
Breath Volume		0.000	05:03	
Air Blank		0.000	05:04	
Control Test		0.078	05:04	
Air Blank		0.000	05:05	
Diagnostics Check		OK	05:05	

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: GIRON RIVERA

DATE/TIME OF TEST: 03/31/2018 05:34:18

Diagnostics Check		OK	05:27
Air Blank		0.000	05:28
Control Test		0.080	05:28
Air Blank		0.000	05:29
Subject Sample	# 1	0.157	05:29
Breath Volume		3.003	05:29
Air Blank		0.000	05:30
Air Blank		0.000	05:32
Subject Sample	# 2	0.155	05:32
Breath Volume		1.824	05:32
Air Blank		0.000	05:33
Control Test		0.079	05:33
Air Blank		0.000	05:34
Diagnostics Check		OK	05:34

Data Download Date/Time: 04/05/2018 19:28

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,D
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: KEENEY

DATE/TIME OF TEST: 03/02/2018 22:53:48

Diagnostics Check		OK	22:49	
Air Blank		0.000	22:49	
Control Test		0.080	22:49	
Air Blank		0.000	22:50	
Subject Sample	# 1	0.163	22:52	
Breath Volume		1.855	22:52	
Air Blank			22:53	Purge Fail
Air Blank		0.000	22:53	

Data Download Date/Time: 03/09/2018 18:53

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,D
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: KEENEY

DATE/TIME OF TEST: 03/02/2018 23:02:26

Diagnostics Check		OK	22:55
Air Blank		0.000	22:55
Control Test		0.080	22:56
Air Blank		0.000	22:56
Subject Sample	# 1	0.183	22:57
Breath Volume		1.886	22:57
Air Blank		0.000	22:57
Air Blank		0.000	22:59
Subject Sample	# 2	0.179	23:00
Breath Volume		1.812	23:00
Air Blank		0.000	23:01
Control Test		0.080	23:01
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 03/09/2018 18:53

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 03/08/2018 20:26:44

Diagnostics Check		OK	20:20
Air Blank		0.000	20:20
Control Test		0.080	20:21
Air Blank		0.000	20:21
Subject Sample	# 1	0.202	20:22
Breath Volume		2.511	20:22
Air Blank		0.000	20:22
Air Blank		0.000	20:24
Subject Sample	# 2	0.199	20:25
Breath Volume		2.828	20:25
Air Blank		0.000	20:25
Control Test		0.081	20:26
Air Blank		0.000	20:26
Diagnostics Check		OK	20:26

Data Download Date/Time: 03/09/2018 18:53

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON,AMY,B
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: CHESSHER,SANDI
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: VELAZQUEZ GONZALEZ

DATE/TIME OF TEST: 03/10/2018 05:46:47

Diagnostics Check	OK	05:45	
Air Blank		05:46	Ambient Fail
Air Blank	0.000	05:46	

Data Download Date/Time: 04/07/2018 23:00

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON,AMY,B
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: VELAZQUEZ GONZALEZ

DATE/TIME OF TEST: 03/10/2018 05:57:31

Diagnostics Check		OK	05:50
Air Blank		0.000	05:51
Control Test		0.080	05:51
Air Blank		0.000	05:51
Subject Sample	# 1	0.091	05:52
Breath Volume		3.656	05:52
Air Blank		0.000	05:53
Air Blank		0.000	05:55
Subject Sample	# 2	0.093	05:55
Breath Volume		2.925	05:55
Air Blank		0.000	05:56
Control Test		0.081	05:56
Air Blank		0.000	05:57
Diagnostics Check		OK	05:57

Data Download Date/Time: 04/07/2018 23:00

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAFA,YUSUF,A
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: ARAFA,YUSUF
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1861309
DRY GAS EXPIRY DATE: 05/19/2018
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/29/2018 20:45:37

Diagnostics Check		OK	20:38
Air Blank		0.000	20:38
Control Test		0.079	20:39
Air Blank		0.000	20:39
Subject Sample	# 1	0.212	20:40
Breath Volume		2.980	20:40
Air Blank		0.000	20:41
Air Blank		0.000	20:43
Subject Sample	# 2	0.213	20:44
Breath Volume		1.554	20:44
Air Blank		0.000	20:44
Control Test		0.079	20:45
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 05/01/2018 15:50

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUTTON,KERRY,D
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: DUTTON,KERRY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 02/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: RENFROE

DATE/TIME OF TEST: 03/10/2018 00:03:05

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:00	
Control Test		0.078	00:01	
Air Blank		0.000	00:01	
Subject Sample	# 1		00:01	Subject Test Refused
Breath Volume		0.000	00:01	
Air Blank		0.000	00:02	
Control Test		0.077	00:02	
Air Blank		0.000	00:02	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 03/31/2018 10:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUTTON,KERRY,D
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: DUTTON,KERRY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 02/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 03/11/2018 19:55:06

Diagnostics Check		OK	19:51	
Air Blank		0.000	19:51	
Control Test		0.077	19:52	
Air Blank		0.000	19:52	
Subject Sample	# 1		19:53	Subject Test Refused
Breath Volume		0.000	19:53	
Air Blank		0.000	19:54	
Control Test		0.077	19:54	
Air Blank		0.000	19:55	
Diagnostics Check		OK	19:55	

Data Download Date/Time: 03/31/2018 10:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUTTON,KERRY,D
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: DUTTON,KERRY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 02/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 03/19/2018 18:49:27

Diagnostics Check		OK	18:46	
Air Blank		0.000	18:46	
Control Test		0.076	18:47	
Air Blank		0.000	18:47	
Subject Sample	# 1		18:48	Subject Test Refused
Breath Volume		0.000	18:48	
Air Blank		0.000	18:48	
Control Test		0.077	18:48	
Air Blank		0.000	18:49	
Diagnostics Check		OK	18:49	

Data Download Date/Time: 03/31/2018 10:12

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: PINA

DATE/TIME OF TEST: 03/02/2018 01:45:13

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.080	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.112	01:40
Breath Volume		2.308	01:40
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.107	01:43
Breath Volume		1.925	01:43
Air Blank		0.000	01:44
Control Test		0.080	01:44
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 03/29/2018 08:41

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: RIENDEAU

DATE/TIME OF TEST: 03/08/2018 02:51:27

Diagnostics Check		OK	02:44
Air Blank		0.000	02:45
Control Test		0.082	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.283	02:46
Breath Volume		2.863	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.278	02:49
Breath Volume		2.660	02:49
Air Blank		0.000	02:50
Control Test		0.081	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 03/29/2018 08:41

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 03/18/2018 02:27:00

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.081	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.175	02:22
Breath Volume		1.890	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.176	02:25
Breath Volume		2.421	02:25
Air Blank		0.000	02:26
Control Test		0.081	02:26
Air Blank		0.000	02:26
Diagnostics Check		OK	02:27

Data Download Date/Time: 03/29/2018 08:41

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: O'NEIL

DATE/TIME OF TEST: 03/23/2018 18:55:20

Diagnostics Check		OK	18:48
Air Blank		0.000	18:49
Control Test		0.082	18:49
Air Blank		0.000	18:50
Subject Sample	# 1	0.127	18:50
Breath Volume		1.640	18:50
Air Blank		0.000	18:51
Air Blank		0.000	18:53
Subject Sample	# 2	0.128	18:53
Breath Volume		1.957	18:53
Air Blank		0.000	18:54
Control Test		0.082	18:54
Air Blank		0.000	18:55
Diagnostics Check		OK	18:55

Data Download Date/Time: 03/29/2018 08:41

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: CRAIN

DATE/TIME OF TEST: 03/29/2018 03:28:25

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.081	03:22
Air Blank		0.000	03:22
Subject Sample	# 1	0.231	03:23
Breath Volume		1.328	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.236	03:26
Breath Volume		2.574	03:26
Air Blank		0.000	03:27
Control Test		0.081	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 03/29/2018 08:41

Printed: 05 June 2018

COBRA ©1997-2014 CMI, Inc.