

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 03/16/2018 00:52:21

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:48	
Control Test	0.081	00:48	
Air Blank	0.000	00:49	
Subject Sample	# 1 0.221	00:50	
Breath Volume	1.105	00:50	
Air Blank		00:51	Purge Fail
Air Blank		00:52	Purge Fail

Data Download Date/Time: 04/06/2018 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 03/16/2018 01:02:51

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.080	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.230	00:58
Breath Volume	1.671	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.241	01:01
Breath Volume	1.488	01:01
Air Blank	0.000	01:01
Control Test	0.080	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 04/06/2018 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LONES,SHAUN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 03/22/2018 17:58:52

Diagnostics Check	OK	17:52	
Air Blank	0.000	17:53	
Control Test	0.080	17:53	
Air Blank	0.000	17:54	
Subject Sample	# 1 0.151	17:57	Volume Not Met
Breath Volume	0.191	17:57	
Air Blank		17:58	Purge Fail
Air Blank		17:58	Purge Fail

Data Download Date/Time: 04/06/2018 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: SCHMUCKER

DATE/TIME OF TEST: 03/24/2018 03:50:19

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:47	
Control Test	0.080	03:48	
Air Blank	0.000	03:48	
Subject Sample	# 1	03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:49	
Control Test	0.081	03:49	
Air Blank	0.000	03:50	
Diagnostics Check	OK	03:50	

Data Download Date/Time: 04/06/2018 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: STRUYF

DATE/TIME OF TEST: 03/24/2018 21:56:20

Diagnostics Check	OK	21:53	
Air Blank	0.000	21:53	
Control Test	0.080	21:54	
Air Blank	0.000	21:54	
Subject Sample	# 1	21:54	Subject Test Refused
Breath Volume	0.000	21:54	
Air Blank	0.000	21:55	
Control Test	0.080	21:55	
Air Blank	0.000	21:56	
Diagnostics Check	OK	21:56	

Data Download Date/Time: 04/06/2018 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: BRAJDIC

DATE/TIME OF TEST: 03/25/2018 00:12:42

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.082	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.178	00:07
Breath Volume	1.675	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.182	00:11
Breath Volume	1.300	00:11
Air Blank	0.000	00:11
Control Test	0.080	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 04/06/2018 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,SHANNON,N
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: O'CONNELL,SHANNON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: VICKERS

DATE/TIME OF TEST: 03/31/2018 04:28:44

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:25	
Control Test	0.080	04:25	
Air Blank	0.000	04:26	
Subject Sample	# 1	04:27	Subject Test Refused
Breath Volume	0.000	04:27	
Air Blank	0.000	04:27	
Control Test	0.081	04:28	
Air Blank	0.000	04:28	
Diagnostics Check	OK	04:28	

Data Download Date/Time: 04/06/2018 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,GIL,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GONZALEZ,GIL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 10:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:42

SUBJECT'S LAST NAME: OKESON

DATE/TIME OF TEST: 03/08/2018 12:07:38

Diagnostics Check	OK	12:04	
Air Blank	0.000	12:05	
Control Test	0.081	12:05	
Air Blank	0.000	12:05	
Subject Sample	# 1	12:06	Subject Test Refused
Breath Volume	0.000	12:06	
Air Blank	0.000	12:06	
Control Test	0.081	12:07	
Air Blank	0.000	12:07	
Diagnostics Check	OK	12:07	

Data Download Date/Time: 04/04/2018 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,RACQUEL,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: DANIELS,RACQUEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: BORRON

DATE/TIME OF TEST: 03/14/2018 23:02:11

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:59	
Control Test	0.081	22:59	
Air Blank	0.000	23:00	
Subject Sample	# 1	23:00	Subject Test Refused
Breath Volume	0.000	23:00	
Air Blank	0.000	23:01	
Control Test	0.081	23:01	
Air Blank	0.000	23:02	
Diagnostics Check	OK	23:02	

Data Download Date/Time: 04/04/2018 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERNS,CODY,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: JANZEN,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: ;6360100265074185285

DATE/TIME OF TEST: 03/24/2018 21:10:36

Diagnostics Check	OK	21:04
Air Blank	0.000	21:04
Control Test	0.082	21:05
Air Blank	0.000	21:05
Subject Sample # 1	0.120	21:06
Breath Volume	1.792	21:06
Air Blank	0.000	21:06
Air Blank	0.000	21:08
Subject Sample # 2	0.118	21:09
Breath Volume	1.996	21:09
Air Blank	0.000	21:09
Control Test	0.081	21:10
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 04/04/2018 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERNS,CODY,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: RIOS,JOEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: CUBAS RAMOS

DATE/TIME OF TEST: 03/25/2018 22:19:12

Diagnostics Check		OK	22:09	
Air Blank		0.000	22:09	
Control Test		0.082	22:10	
Air Blank		0.000	22:10	
Subject Sample	# 1	0.158	22:13	Volume Not Met
Breath Volume		0.175	22:13	
Air Blank		0.000	22:14	
Air Blank		0.000	22:16	
Subject Sample	# 2		22:17	Subject Test Refused
Breath Volume		0.039	22:17	
Air Blank		0.000	22:18	
Control Test		0.080	22:18	
Air Blank		0.000	22:19	
Diagnostics Check		OK	22:19	

Data Download Date/Time: 04/04/2018 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'KEEFE,PATRICIA,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HAGER,JAMES
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: PIEROG

DATE/TIME OF TEST: 03/27/2018 05:47:56

Diagnostics Check	OK	05:44	
Air Blank	0.000	05:45	
Control Test	0.081	05:45	
Air Blank	0.000	05:46	
Subject Sample	# 1	05:46	Subject Test Refused
Breath Volume	0.000	05:46	
Air Blank	0.000	05:47	
Control Test	0.081	05:47	
Air Blank	0.000	05:47	
Diagnostics Check	OK	05:47	

Data Download Date/Time: 04/04/2018 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: FINCH

DATE/TIME OF TEST: 03/31/2018 03:43:21

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.081	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.242	03:38
Breath Volume	2.332	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.236	03:41
Breath Volume	1.285	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 04/04/2018 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA QUINONES,IGNACIO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: RAMBERT,MICHAEL
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: PEREZ VALDES

DATE/TIME OF TEST: 03/11/2018 03:49:15

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:46	
Control Test	0.078	03:46	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Control Test	0.079	03:48	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 03/23/2018 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA QUINONES,IGNACIO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: RAMBERT,MICHAEL
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 03/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BANEGAS

DATE/TIME OF TEST: 03/24/2018 03:51:08

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:48	
Control Test	0.079	03:49	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:49	Subject Test Refused
Breath Volume	0.000	03:49	
Air Blank	0.000	03:50	
Control Test	0.079	03:50	
Air Blank	0.000	03:51	
Diagnostics Check	OK	03:51	

Data Download Date/Time: 05/29/2018 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA QUINONES,IGNACIO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: RODRIGUEZ,ERIK
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 03/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HO

DATE/TIME OF TEST: 03/26/2018 23:08:00

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.112	23:02
Breath Volume	2.148	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.105	23:06
Breath Volume	2.320	23:06
Air Blank	0.000	23:07
Control Test	0.079	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:08

Data Download Date/Time: 05/29/2018 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINCORE, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ROA, JUAN
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 16:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: MIRANDALOPEZ

DATE/TIME OF TEST: 03/03/2018 16:46:44

Diagnostics Check	OK	16:43	
Air Blank	0.000	16:44	
Control Test	0.082	16:44	
Air Blank	0.000	16:45	
Subject Sample	# 1	16:45	Subject Test Refused
Breath Volume	0.000	16:45	
Air Blank	0.000	16:45	
Control Test	0.082	16:46	
Air Blank	0.000	16:46	
Diagnostics Check	OK	16:46	

Data Download Date/Time: 03/29/2018 09:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANCO,JUAN,S
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BLANCO,JUAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 03/18/2018 15:27:38

Diagnostics Check	OK	15:20
Air Blank	0.000	15:21
Control Test	0.080	15:21
Air Blank	0.000	15:21
Subject Sample # 1	0.267	15:22
Breath Volume	3.390	15:22
Air Blank	0.000	15:23
Air Blank	0.000	15:25
Subject Sample # 2	0.257	15:26
Breath Volume	2.851	15:26
Air Blank	0.000	15:26
Control Test	0.078	15:27
Air Blank	0.000	15:27
Diagnostics Check	OK	15:27

Data Download Date/Time: 04/20/2018 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELFNER,ISAAC,S
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: HELFNER,ISAAC
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 03/22/2018 23:14:01

Diagnostics Check	OK	23:06
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.129	23:09
Breath Volume	1.800	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:11
Subject Sample # 2	0.127	23:12
Breath Volume	2.398	23:12
Air Blank	0.000	23:13
Control Test	0.079	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:14

Data Download Date/Time: 04/20/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUNCAS,ANDREW,J
BREATH TEST OPERATOR AGENCY: Bal Harbour Vil PD
ARRESTING OFFICER: KUNCAS,ANDREW
ARRESTING OFFICER AGENCY: Bal Harbour Vil PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: TOPCHIEV

DATE/TIME OF TEST: 03/24/2018 23:29:59

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:27	
Control Test	0.080	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:28	Subject Test Refused
Breath Volume	0.000	23:28	
Air Blank	0.000	23:29	
Control Test	0.077	23:29	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 04/20/2018 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNSCHY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Bay Harbor Island PD
ARRESTING OFFICER: GUADALUPE,MARCOS
ARRESTING OFFICER AGENCY: Bay Harbor Island PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: FEALEY JR

DATE/TIME OF TEST: 03/24/2018 23:46:12

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.080	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.000	23:41
Breath Volume	4.375	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.000	23:44
Breath Volume	4.199	23:44
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 04/20/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,MIGUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: COLLAZO

DATE/TIME OF TEST: 03/29/2018 22:43:41

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:41	
Control Test	0.081	22:41	
Air Blank	0.000	22:42	
Subject Sample	# 1	22:42	Subject Test Refused
Breath Volume	0.000	22:42	
Air Blank	0.000	22:42	
Control Test	0.081	22:43	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:43	

Data Download Date/Time: 04/20/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: LAGOS

DATE/TIME OF TEST: 03/16/2018 04:19:49

Diagnostics Check	OK	04:16	
Air Blank	0.000	04:17	
Control Test	0.081	04:17	
Air Blank	0.000	04:18	
Subject Sample	# 1	04:18	Subject Test Refused
Breath Volume	0.000	04:18	
Air Blank	0.000	04:18	
Control Test	0.080	04:19	
Air Blank	0.000	04:19	
Diagnostics Check	OK	04:19	

Data Download Date/Time: 04/20/2018 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATTORNO,ALEXANDER,F
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GATTORNO,ALEXANDER
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 03/17/2018 06:10:07

Diagnostics Check	OK	06:03
Air Blank	0.000	06:04
Control Test	0.080	06:04
Air Blank	0.000	06:05
Subject Sample # 1	0.123	06:05
Breath Volume	1.906	06:05
Air Blank	0.000	06:06
Air Blank	0.000	06:08
Subject Sample # 2	0.124	06:08
Breath Volume	2.089	06:08
Air Blank	0.000	06:09
Control Test	0.079	06:09
Air Blank	0.000	06:10
Diagnostics Check	OK	06:10

Data Download Date/Time: 04/20/2018 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: DINGLE

DATE/TIME OF TEST: 03/18/2018 00:44:03

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:41	
Control Test	0.080	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:42	Subject Test Refused
Breath Volume	0.000	00:42	
Air Blank	0.000	00:43	
Control Test	0.080	00:43	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 04/20/2018 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: SANAC

DATE/TIME OF TEST: 03/16/2018 05:32:29

Diagnostics Check	OK	05:26
Air Blank	0.000	05:26
Control Test	0.081	05:27
Air Blank	0.000	05:27
Subject Sample # 1	0.153	05:28
Breath Volume	2.628	05:28
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.153	05:30
Breath Volume	2.929	05:30
Air Blank	0.000	05:31
Control Test	0.080	05:31
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 04/17/2018 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,MARK
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 03/17/2018 01:48:16

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.000	01:44
Breath Volume	2.398	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.000	01:46
Breath Volume	2.789	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 04/17/2018 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 03/18/2018 02:51:31

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.080	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.160	02:47
Breath Volume	2.308	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.155	02:49
Breath Volume	2.187	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 04/17/2018 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/18/2018 04:29:28

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:27	
Control Test	0.079	04:27	
Air Blank	0.000	04:27	
Subject Sample	# 1	04:28	Subject Test Refused
Breath Volume	0.000	04:28	
Air Blank	0.000	04:28	
Control Test	0.080	04:28	
Air Blank	0.000	04:29	
Diagnostics Check	OK	04:29	

Data Download Date/Time: 04/17/2018 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 03/18/2018 06:05:46

Diagnostics Check	OK	06:02	
Air Blank	0.000	06:03	
Control Test	0.079	06:03	
Air Blank	0.000	06:04	
Subject Sample	# 1	06:04	Subject Test Refused
Breath Volume	0.000	06:04	
Air Blank	0.000	06:04	
Control Test	0.080	06:05	
Air Blank	0.000	06:05	
Diagnostics Check	OK	06:05	

Data Download Date/Time: 04/17/2018 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: AFRA

DATE/TIME OF TEST: 03/22/2018 23:53:42

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:51	
Control Test	0.081	23:51	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:52	
Control Test	0.081	23:53	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

Data Download Date/Time: 04/17/2018 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 03/29/2018 04:09:24

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.131	04:04
Breath Volume	2.640	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.133	04:07
Breath Volume	2.429	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 04/17/2018 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/31/2018 02:54:04

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.080	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.080	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 04/17/2018 13:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: DIAZ,JAVIER
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: ABERLE

DATE/TIME OF TEST: 03/01/2018 01:25:23

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:23	
Control Test	0.080	01:23	
Air Blank	0.000	01:23	
Subject Sample	# 1	01:24	Subject Test Refused
Breath Volume	0.000	01:24	
Air Blank	0.000	01:24	
Control Test	0.080	01:24	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 03/09/2018 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 03/01/2018 03:07:54

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.294	03:02
Breath Volume	2.199	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.277	03:06
Breath Volume	2.421	03:06
Air Blank	0.000	03:06
Control Test	0.078	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 03/09/2018 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: LOPEZ-LALINDE

DATE/TIME OF TEST: 03/01/2018 04:47:26

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.080	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.126	04:42
Breath Volume	2.718	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.124	04:45
Breath Volume	2.285	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 03/09/2018 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: FRANCO HERNANDEZ

DATE/TIME OF TEST: 03/01/2018 22:06:18

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.080	22:00
Air Blank	0.000	22:00
Subject Sample # 1	0.173	22:01
Breath Volume	1.722	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.174	22:04
Breath Volume	1.902	22:04
Air Blank	0.000	22:05
Control Test	0.079	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 03/09/2018 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,YOSNIEL,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: LORENZO,YOSNIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 03/02/2018 03:16:52

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.081	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:10	No Sample Provided
Breath Volume	0.000	03:10	
Air Blank	0.000	03:10	
Air Blank	0.000	03:12	
Subject Sample	# 2	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.081	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 03/09/2018 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,YOSNIEL,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: LORENZO,YOSNIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 03/02/2018 03:21:23

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.079	03:19	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:20	
Control Test	0.079	03:20	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 03/09/2018 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO, YOSNIEL,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: LORENZO, YOSNIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: PENA MEJIA

DATE/TIME OF TEST: 03/02/2018 23:42:25

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.081	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.088	23:37
Breath Volume	1.972	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.090	23:40
Breath Volume	1.546	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 03/09/2018 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 03/03/2018 05:37:48

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.081	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.130	05:33
Breath Volume	1.328	05:33
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.139	05:36
Breath Volume	1.328	05:36
Air Blank	0.000	05:36
Control Test	0.081	05:37
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 03/09/2018 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,RENE,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MARTINEZ,RENE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: ROSELI ARCANGEL

DATE/TIME OF TEST: 03/04/2018 01:22:00

Diagnostics Check	OK	01:21	
Air Blank		01:21	RFI Detect
Air Blank		01:22	RFI Detect

Data Download Date/Time: 03/09/2018 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,RENE,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MARTINEZ,RENE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: ROSELI ARCANGEL

DATE/TIME OF TEST: 03/04/2018 01:32:11

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.081	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.200	01:27
Breath Volume	2.574	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.199	01:30
Breath Volume	3.003	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 03/09/2018 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORGES,SEAN,P
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: PICALLO,RUBEN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: SAFI NASSER

DATE/TIME OF TEST: 03/04/2018 07:01:35

Diagnostics Check	OK	07:00	
Air Blank	0.000	07:01	
		07:01	RFI Detect
Air Blank	0.000	07:01	

Data Download Date/Time: 03/09/2018 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORGES,SEAN,P
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: PICALLO,RUBEN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: SAFI NASSER

DATE/TIME OF TEST: 03/04/2018 07:07:11

Diagnostics Check	OK	07:04	
Air Blank	0.000	07:04	
Control Test	0.082	07:04	
Air Blank	0.000	07:05	
Subject Sample	# 1	07:05	Subject Test Refused
Breath Volume	0.000	07:05	
Air Blank	0.000	07:06	
Control Test	0.080	07:06	
Air Blank	0.000	07:07	
Diagnostics Check	OK	07:07	

Data Download Date/Time: 03/09/2018 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,YOSNIEL,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: LORENZO,YOSNIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: MESA

DATE/TIME OF TEST: 03/07/2018 00:47:44

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:44	
Control Test	0.080	00:45	
Air Blank	0.000	00:45	
Subject Sample	# 1	00:46	Subject Test Refused
Breath Volume	0.000	00:46	
Air Blank	0.000	00:46	
Control Test	0.080	00:47	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 03/09/2018 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: OLEA ZAMBRANO

DATE/TIME OF TEST: 03/08/2018 06:54:31

Diagnostics Check	OK	06:47
Air Blank	0.000	06:47
Control Test	0.081	06:47
Air Blank	0.000	06:48
Subject Sample # 1	0.143	06:48
Breath Volume	1.437	06:48
Air Blank	0.000	06:49
Air Blank	0.000	06:51
Subject Sample # 2	0.162	06:52
Breath Volume	1.562	06:52
Air Blank	0.000	06:53
Control Test	0.081	06:53
Air Blank	0.000	06:54
Diagnostics Check	OK	06:54

Data Download Date/Time: 03/09/2018 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/03/2018 06:15:04

Diagnostics Check	OK	06:12	
Air Blank	0.000	06:12	
Control Test	0.080	06:12	
Air Blank	0.000	06:13	
Subject Sample	# 1	06:13	Subject Test Refused
Breath Volume	0.000	06:13	
Air Blank	0.000	06:14	
Control Test	0.080	06:14	
Air Blank	0.000	06:14	
Diagnostics Check	OK	06:15	

Data Download Date/Time: 03/22/2018 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JULIO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FERNANDEZ,JULIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: VILLATORO

DATE/TIME OF TEST: 03/11/2018 04:38:36

Diagnostics Check	OK	04:37	
Air Blank	0.000	04:38	
Control Test	0.079	04:38	
Air Blank		04:38	Improper Sample
Air Blank		04:38	Improper Sample

Data Download Date/Time: 03/22/2018 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JULIO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FERNANDEZ,JULIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: VILLATORO

DATE/TIME OF TEST: 03/11/2018 04:40:48

Diagnostics Check	OK	04:39	
Air Blank		04:40	Improper Sample
Air Blank	0.000	04:40	

Data Download Date/Time: 03/22/2018 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JULIO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FERNANDEZ,JULIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: VILLATORO

DATE/TIME OF TEST: 03/11/2018 04:45:16

Diagnostics Check	OK	04:44	
Air Blank	0.000	04:44	
Control Test	0.074	04:44	Control Outside Tolerance
Air Blank		04:45	RFI Detect

Data Download Date/Time: 03/22/2018 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JULIO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FERNANDEZ,JULIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: VILLATORO

DATE/TIME OF TEST: 03/11/2018 04:52:58

Diagnostics Check		OK	04:46
Air Blank		0.000	04:46
Control Test		0.079	04:47
Air Blank		0.000	04:47
Subject Sample	# 1	0.224	04:48
Breath Volume		1.300	04:48
Air Blank		0.000	04:48
Air Blank		0.000	04:50
Subject Sample	# 2	0.220	04:51
Breath Volume		1.500	04:51
Air Blank		0.000	04:52
Control Test		0.080	04:52
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 03/22/2018 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNOZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MUNOZ,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: DELEON

DATE/TIME OF TEST: 03/13/2018 04:28:09

Diagnostics Check		OK	04:21
Air Blank		0.000	04:22
Control Test		0.080	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.158	04:23
Breath Volume		3.984	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.156	04:26
Breath Volume		3.863	04:26
Air Blank		0.000	04:27
Control Test		0.080	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 03/22/2018 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 03/14/2018 05:57:23

Diagnostics Check		OK	05:49
Air Blank		0.000	05:50
Control Test		0.080	05:50
Air Blank		0.000	05:51
Subject Sample	# 1	0.191	05:52
Breath Volume		1.132	05:52
Air Blank		0.000	05:53
Air Blank		0.000	05:55
Subject Sample	# 2	0.186	05:55
Breath Volume		2.320	05:55
Air Blank		0.000	05:56
Control Test		0.079	05:56
Air Blank		0.000	05:57
Diagnostics Check		OK	05:57

Data Download Date/Time: 03/22/2018 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:16

SUBJECT'S LAST NAME: LACALLE

DATE/TIME OF TEST: 03/17/2018 08:52:37

Diagnostics Check	OK	08:45
Air Blank	0.000	08:46
Control Test	0.079	08:46
Air Blank	0.000	08:47
Subject Sample # 1	0.200	08:47
Breath Volume	1.402	08:47
Air Blank	0.000	08:48
Air Blank	0.000	08:50
Subject Sample # 2	0.197	08:50
Breath Volume	1.320	08:50
Air Blank	0.000	08:51
Control Test	0.078	08:52
Air Blank	0.000	08:52
Diagnostics Check	OK	08:52

Data Download Date/Time: 03/22/2018 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MARTOS,JULIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: PAZ

DATE/TIME OF TEST: 03/18/2018 00:18:01

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.080	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.000	00:13
Breath Volume	1.332	00:13
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.000	00:16
Breath Volume	1.417	00:16
Air Blank	0.000	00:17
Control Test	0.079	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:18

Data Download Date/Time: 03/22/2018 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,F.
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: OLEA ZAMBRANO

DATE/TIME OF TEST: 03/08/2018 05:34:37

Diagnostics Check	OK	05:31	
Air Blank	0.000	05:31	
Control Test	0.079	05:32	
Air Blank	0.000	05:32	
Subject Sample	# 1	05:33	RFI Detect
Breath Volume	0.000	05:33	
Air Blank	0.000	05:33	
Control Test	0.080	05:34	
Air Blank	0.000	05:34	
Diagnostics Check	OK	05:34	

Data Download Date/Time: 03/22/2018 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: OLEA ZAMBRANO

DATE/TIME OF TEST: 03/08/2018 05:40:43

Diagnostics Check	OK	05:38	
Air Blank	0.000	05:38	
Control Test	0.079	05:38	
Air Blank	0.000	05:39	
Subject Sample	# 1	05:39	RFI Detect
Breath Volume	0.000	05:39	
Air Blank	0.000	05:40	
		05:40	RFI Detect
Air Blank	0.000	05:40	

Data Download Date/Time: 03/22/2018 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: OLEA ZAMBRANO

DATE/TIME OF TEST: 03/08/2018 05:48:54

Diagnostics Check	OK	05:43	
Air Blank	0.000	05:44	
Control Test	0.079	05:44	
Air Blank	0.000	05:45	
Subject Sample	# 1 0.183	05:47	
Breath Volume	1.617	05:47	
Air Blank		05:47	RFI Detect
Air Blank		05:48	Purge Fail
Air Blank		05:48	Purge Fail

Data Download Date/Time: 03/22/2018 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AJO,MICHAEL,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AJO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: BERKOWITZ

DATE/TIME OF TEST: 03/07/2018 00:49:24

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:37	
Control Test	0.079	00:38	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:41	No Sample Provided
Breath Volume	0.000	00:41	
Air Blank	0.000	00:42	
Air Blank	0.000	00:44	
Subject Sample	# 2	0.188	00:44
Breath Volume	1.859	00:44	
Air Blank	0.000	00:45	
Air Blank	0.000	00:47	
Subject Sample	# 3	0.196	00:47
Breath Volume	2.519	00:47	
Air Blank	0.000	00:48	
Control Test	0.078	00:48	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 03/30/2018 00:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ZOILA,M
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: CASTILLO,CARLOS
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 03/19/2018 00:20:13

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:17	
Control Test	0.079	00:18	
Air Blank	0.000	00:18	
Subject Sample	# 1	00:18	Subject Test Refused
Breath Volume	0.000	00:18	
Air Blank	0.000	00:19	
Control Test	0.079	00:19	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

Data Download Date/Time: 03/30/2018 00:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAFLIN,DAVID,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: TEST,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 03/06/2018 19:37:47

Diagnostics Check	OK	19:31
Air Blank	0.000	19:31
Control Test	0.079	19:31
Air Blank	0.000	19:32
Subject Sample # 1	0.000	19:33
Breath Volume	2.667	19:33
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample # 2	0.000	19:36
Breath Volume	1.144	19:36
Air Blank	0.000	19:36
Control Test	0.080	19:37
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 03/07/2018 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AXELSON,B,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCCOY,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: BUCKEY

DATE/TIME OF TEST: 03/02/2018 21:29:45

Diagnostics Check	OK	21:21
Air Blank	0.000	21:22
Control Test	0.079	21:22
Air Blank	0.000	21:23
Subject Sample # 1	0.170	21:24
Breath Volume	1.132	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:26
Subject Sample # 2	0.166	21:28
Breath Volume	1.117	21:28
Air Blank	0.000	21:28
Control Test	0.077	21:29
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 03/05/2018 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: ZAREMBKA

DATE/TIME OF TEST: 03/02/2018 21:40:49

Diagnostics Check	OK	21:33
Air Blank	0.000	21:34
Control Test	0.079	21:34
Air Blank	0.000	21:35
Subject Sample # 1	0.237	21:36
Breath Volume	2.218	21:36
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.252	21:39
Breath Volume	2.371	21:39
Air Blank	0.000	21:39
Control Test	0.078	21:40
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

Data Download Date/Time: 03/05/2018 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULRY, THOMAS, B
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MULRY, THOMAS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: DE VAUL

DATE/TIME OF TEST: 03/02/2018 23:05:55

Diagnostics Check	OK	22:59
Air Blank	0.000	23:00
Control Test	0.078	23:00
Air Blank	0.000	23:01
Subject Sample # 1	0.123	23:01
Breath Volume	1.496	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:03
Subject Sample # 2	0.120	23:04
Breath Volume	1.996	23:04
Air Blank	0.000	23:05
Control Test	0.078	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 03/05/2018 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNARD,NICK,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BERNARD,NICK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: BRIGGS

DATE/TIME OF TEST: 03/02/2018 23:17:27

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.015	23:12
Breath Volume	3.074	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.012	23:15
Breath Volume	1.875	23:15
Air Blank	0.000	23:16
Control Test	0.078	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 03/05/2018 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATSON,MITCHELL,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCCOY,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: MENDOZA LOPEZ

DATE/TIME OF TEST: 03/03/2018 03:14:12

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.079	03:00	
Air Blank	0.000	03:00	
Subject Sample	# 1 0.167	03:03	
Breath Volume	1.628	03:03	
Air Blank	0.000	03:03	
Air Blank	0.000	03:05	
Subject Sample	# 2 0.132	03:08	Volume Not Met
Breath Volume	0.378	03:08	
Air Blank	0.000	03:09	
Air Blank	0.000	03:11	
Subject Sample	# 3 0.163	03:12	
Breath Volume	1.687	03:12	
Air Blank	0.000	03:13	
Control Test	0.077	03:13	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 03/05/2018 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,REID,B
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: WHITE,REID
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: BUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 03/03/2018 18:56:06

Diagnostics Check	OK	18:49
Air Blank	0.000	18:50
Control Test	0.079	18:50
Air Blank	0.000	18:50
Subject Sample # 1	0.169	18:51
Breath Volume	3.117	18:51
Air Blank	0.000	18:52
Air Blank	0.000	18:54
Subject Sample # 2	0.160	18:54
Breath Volume	2.261	18:54
Air Blank	0.000	18:55
Control Test	0.078	18:55
Air Blank	0.000	18:56
Diagnostics Check	OK	18:56

Data Download Date/Time: 03/05/2018 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CONTESSA,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 03/03/2018 20:15:29

Diagnostics Check	OK	20:09
Air Blank	0.000	20:09
Control Test	0.078	20:09
Air Blank	0.000	20:10
Subject Sample # 1	0.155	20:10
Breath Volume	2.617	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:13
Subject Sample # 2	0.155	20:13
Breath Volume	1.683	20:13
Air Blank	0.000	20:14
Control Test	0.079	20:14
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 03/05/2018 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/04/2018 00:52:34

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.078	00:48	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:51	Subject Test Refused
Breath Volume	0.000	00:51	
Air Blank	0.000	00:51	
Control Test	0.079	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 03/05/2018 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEAR,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: CAPLAN

DATE/TIME OF TEST: 03/04/2018 03:42:38

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.093	03:36
Breath Volume	1.609	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.085	03:40
Breath Volume	1.390	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 03/05/2018 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,Y,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NODARSE,Y
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/07/2018 20:17:49

Diagnostics Check	OK	20:11
Air Blank	0.000	20:12
Control Test	0.078	20:12
Air Blank	0.000	20:12
Subject Sample # 1	0.166	20:13
Breath Volume	2.304	20:13
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.176	20:16
Breath Volume	2.531	20:16
Air Blank	0.000	20:16
Control Test	0.079	20:17
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: PAUZUS JR

DATE/TIME OF TEST: 03/08/2018 02:23:28

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.078	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.174	02:18
Breath Volume	3.859	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.180	02:21
Breath Volume	3.906	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: RODARTE

DATE/TIME OF TEST: 03/09/2018 04:29:04

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.079	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.207	04:24
Breath Volume	4.332	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:26
Subject Sample # 2	0.203	04:27
Breath Volume	4.222	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:29

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,D,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: JACKSON,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 03/10/2018 03:56:19

Diagnostics Check		OK	03:44	
Air Blank		0.000	03:45	
Control Test		0.079	03:45	
Air Blank		0.000	03:46	
Subject Sample	# 1	0.148	03:49	Volume Not Met
Breath Volume		0.035	03:49	
Air Blank		0.000	03:49	
Air Blank		0.000	03:51	
Subject Sample	# 2	0.157	03:54	Volume Not Met
Breath Volume		0.070	03:54	
Air Blank		0.000	03:55	
Control Test		0.079	03:55	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOLO,J,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CAMPOLO,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: PARKS

DATE/TIME OF TEST: 03/10/2018 22:48:09

Diagnostics Check	OK	22:43	
Air Blank	0.000	22:43	
Control Test	0.079	22:44	
Air Blank	0.000	22:44	
Subject Sample	# 1	22:47	Sequence Aborted
Breath Volume	0.164	22:47	
Air Blank	0.000	22:48	

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOLO,J,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CAMPOLO,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: PARKS

DATE/TIME OF TEST: 03/10/2018 23:02:29

Diagnostics Check		OK	22:49	
Air Blank		0.000	22:49	
Control Test		0.078	22:50	
Air Blank		0.000	22:50	
Subject Sample	# 1	0.043	22:53	Volume Not Met
Breath Volume		1.074	22:53	
Air Blank		0.000	22:54	
Air Blank		0.000	22:56	
Subject Sample	# 2	0.050	22:57	
Breath Volume		2.152	22:57	
Air Blank		0.000	22:58	
Air Blank		0.000	23:00	
Subject Sample	# 3	0.044	23:01	
Breath Volume		1.492	23:01	
Air Blank		0.000	23:01	
Control Test		0.078	23:01	
Air Blank		0.000	23:02	
Diagnostics Check		OK	23:02	

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOODMAN,ROBERT,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GROSSKLAS,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 03/11/2018 05:58:55

Diagnostics Check	OK	05:52
Air Blank	0.000	05:52
Control Test	0.078	05:53
Air Blank	0.000	05:53
Subject Sample # 1	0.115	05:54
Breath Volume	1.425	05:54
Air Blank	0.000	05:54
Air Blank	0.000	05:56
Subject Sample # 2	0.105	05:57
Breath Volume	1.304	05:57
Air Blank	0.000	05:58
Control Test	0.078	05:58
Air Blank	0.000	05:58
Diagnostics Check	OK	05:58

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNORAT,ADELY,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORWIN,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: CIFUENTES GONZALEZ

DATE/TIME OF TEST: 03/11/2018 06:26:58

Diagnostics Check	OK	06:19
Air Blank	0.000	06:20
Control Test	0.078	06:20
Air Blank	0.000	06:21
Subject Sample # 1	0.168	06:21
Breath Volume	1.882	06:21
Air Blank	0.000	06:22
Air Blank	0.000	06:24
Subject Sample # 2	0.172	06:25
Breath Volume	1.902	06:25
Air Blank	0.000	06:26
Control Test	0.079	06:26
Air Blank	0.000	06:26
Diagnostics Check	OK	06:26

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRILLO,XAVIER,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MERCATO,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: ZAREMBA

DATE/TIME OF TEST: 03/14/2018 16:49:18

Diagnostics Check	OK	16:43
Air Blank	0.000	16:43
Control Test	0.079	16:43
Air Blank	0.000	16:44
Subject Sample # 1	0.043	16:44
Breath Volume	1.339	16:44
Air Blank	0.000	16:45
Air Blank	0.000	16:47
Subject Sample # 2	0.046	16:47
Breath Volume	1.437	16:47
Air Blank	0.000	16:48
Control Test	0.079	16:48
Air Blank	0.000	16:49
Diagnostics Check	OK	16:49

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GUTH,ROBERT
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: GIRALDO-OCAMPO

DATE/TIME OF TEST: 03/15/2018 23:51:23

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.079	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.186	23:47
Breath Volume	2.457	23:47
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.182	23:49
Breath Volume	1.753	23:49
Air Blank	0.000	23:50
Control Test	0.078	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON, DEMETRIS, L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: JACKSON, DEMETRIS
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: ANTELO BEJARANO

DATE/TIME OF TEST: 03/16/2018 06:17:59

Diagnostics Check		OK	06:06	
Air Blank		0.000	06:06	
Control Test		0.079	06:07	
Air Blank		0.000	06:07	
Subject Sample	# 1	0.123	06:10	Volume Not Met
Breath Volume		0.242	06:10	
Air Blank		0.000	06:11	
Air Blank		0.000	06:13	
Subject Sample	# 2	0.134	06:16	Volume Not Met
Breath Volume		0.000	06:16	
Air Blank		0.000	06:17	
Control Test		0.078	06:17	
Air Blank		0.000	06:17	
Diagnostics Check		OK	06:17	

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,DEMETRIS,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: JACKSON,DEMETRIS
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: ANTELO BEJARANO

DATE/TIME OF TEST: 03/16/2018 06:28:45

Diagnostics Check	OK	06:19	
Air Blank	0.000	06:20	
Control Test	0.077	06:20	
Air Blank	0.000	06:21	
Subject Sample	# 1 0.125	06:24	Volume Not Met
Breath Volume	0.062	06:24	
Air Blank	0.000	06:25	
Air Blank	0.000	06:26	
Subject Sample	# 2	06:28	Sequence Aborted
Breath Volume	0.113	06:28	
Air Blank	0.000	06:28	

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,U,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: COSTELLO,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: VELASQUEZ RANALETTA

DATE/TIME OF TEST: 03/17/2018 01:42:38

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.169	01:38
Breath Volume	2.410	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.162	01:41
Breath Volume	2.316	01:41
Air Blank	0.000	01:41
Control Test	0.079	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,KEVIN,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: POUNDS,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/17/2018 02:48:28

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.078	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.183	02:41
Breath Volume	1.464	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.193	02:46
Breath Volume	1.402	02:46
Air Blank	0.000	02:47
Control Test	0.078	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVAS,W,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCGILVREY,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 03/17/2018 04:09:38

Diagnostics Check		OK	03:58	
Air Blank		0.000	03:58	
Control Test		0.077	03:58	
Air Blank		0.000	03:59	
Subject Sample	# 1	0.124	04:02	Volume Not Met
Breath Volume		0.085	04:02	
Air Blank		0.000	04:03	
Air Blank		0.000	04:05	
Subject Sample	# 2	0.117	04:08	Volume Not Met
Breath Volume		0.320	04:08	
Air Blank		0.000	04:08	
Control Test		0.077	04:09	
Air Blank		0.000	04:09	
Diagnostics Check		OK	04:09	

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CARPENTER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/17/2018 22:03:20

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.079	21:58
Air Blank	0.000	21:58
Subject Sample # 1	0.198	21:58
Breath Volume	2.425	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.193	22:01
Breath Volume	2.316	22:01
Air Blank	0.000	22:02
Control Test	0.078	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ ANGEL,MARTIN,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: LOPEZ ANGEL

DATE/TIME OF TEST: 03/18/2018 03:42:57

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:31	
Control Test		0.078	03:31	
Air Blank		0.000	03:31	
Subject Sample	# 1	0.142	03:35	Volume Not Met
Breath Volume		0.957	03:35	
Air Blank		0.000	03:35	
Air Blank		0.000	03:37	
Subject Sample	# 2	0.184	03:38	No .020 agreement
Breath Volume		1.277	03:38	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 3	0.151	03:41	No .020 agreement
Breath Volume		1.761	03:41	
Air Blank		0.000	03:42	
Control Test		0.078	03:42	
Air Blank		0.000	03:42	
Diagnostics Check		OK	03:42	

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: LOPEZ ANGEL

DATE/TIME OF TEST: 03/18/2018 03:55:24

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.078	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.176	03:47
Breath Volume	1.890	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:49
Subject Sample # 2	0.151	03:50
Breath Volume	1.636	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 3	0.154	03:53
Breath Volume	2.468	03:53
Air Blank	0.000	03:54
Control Test	0.078	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,N
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DISARRO,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: ILIE

DATE/TIME OF TEST: 03/18/2018 04:09:54

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.078	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.237	04:05
Breath Volume	1.933	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.218	04:08
Breath Volume	2.390	04:08
Air Blank	0.000	04:09
Control Test	0.078	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINTFORT,PAUL,W
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KIRSCH,Z
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:41

SUBJECT'S LAST NAME: GRANTHAM

DATE/TIME OF TEST: 03/18/2018 15:12:14

Diagnostics Check	OK	15:05
Air Blank	0.000	15:06
Control Test	0.078	15:06
Air Blank	0.000	15:06
Subject Sample # 1	0.310	15:07
Breath Volume	1.789	15:07
Air Blank	0.000	15:08
Air Blank	0.000	15:10
Subject Sample # 2	0.301	15:10
Breath Volume	1.558	15:10
Air Blank	0.000	15:11
Control Test	0.077	15:11
Air Blank	0.000	15:12
Diagnostics Check	OK	15:12

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,HENRRY,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GOMEZ,HENRRY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: JEAN-BAPTISTE

DATE/TIME OF TEST: 03/20/2018 19:06:01

Diagnostics Check	OK	19:02	
Air Blank	0.000	19:03	
Control Test	0.079	19:03	
Air Blank	0.000	19:03	
Subject Sample	# 1	19:04	RFI Detect
Breath Volume	0.000	19:04	
Air Blank	0.000	19:05	
Control Test	0.078	19:05	
Air Blank	0.000	19:05	
Diagnostics Check	OK	19:06	

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNARD,NICK,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BERNARD,NICK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 03/22/2018 20:22:00

Diagnostics Check	OK	20:15
Air Blank	0.000	20:16
Control Test	0.079	20:16
Air Blank	0.000	20:17
Subject Sample # 1	0.130	20:17
Breath Volume	2.597	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.129	20:20
Breath Volume	3.046	20:20
Air Blank	0.000	20:21
Control Test	0.080	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:22

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHLINE,C,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BUTTLER,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 03/22/2018 22:24:26

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.079	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.261	22:18
Breath Volume	2.781	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.261	22:22
Breath Volume	2.589	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: FORD,L
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MEJIA MUNOZ

DATE/TIME OF TEST: 03/23/2018 22:27:35

Diagnostics Check	OK	22:20
Air Blank	0.000	22:21
Control Test	0.079	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.177	22:23
Breath Volume	2.253	22:23
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.173	22:26
Breath Volume	2.511	22:26
Air Blank	0.000	22:26
Control Test	0.078	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCCOY,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: TABORA

DATE/TIME OF TEST: 03/26/2018 21:51:38

Diagnostics Check	OK	21:41
Air Blank	0.000	21:42
Control Test	0.078	21:42
Air Blank	0.000	21:42
Subject Sample # 1	0.214	21:43
Breath Volume	1.199	21:43
Air Blank	0.000	21:44
Air Blank	0.000	21:45
Subject Sample # 2	0.181	21:46
Breath Volume	1.746	21:46
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 3	0.188	21:50
Breath Volume	1.757	21:50
Air Blank	0.000	21:50
Control Test	0.078	21:51
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO, JONATHAN, B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TROTTER, W
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: KREUZER

DATE/TIME OF TEST: 03/27/2018 22:03:30

Diagnostics Check	OK	21:56
Air Blank	0.000	21:57
Control Test	0.079	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.130	21:58
Breath Volume	1.941	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.133	22:02
Breath Volume	1.832	22:02
Air Blank	0.000	22:02
Control Test	0.079	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 03/28/2018 03:48:01

Diagnostics Check		OK	03:41
Air Blank		0.000	03:42
Control Test		0.079	03:42
Air Blank		0.000	03:42
Subject Sample	# 1	0.170	03:43
Breath Volume		1.964	03:43
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.171	03:46
Breath Volume		2.128	03:46
Air Blank		0.000	03:47
Control Test		0.079	03:47
Air Blank		0.000	03:47
Diagnostics Check		OK	03:48

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRATT,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GROENDYK,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: NAPP

DATE/TIME OF TEST: 03/28/2018 22:25:35

Diagnostics Check	OK	22:17	
Air Blank	0.000	22:17	
Control Test	0.078	22:18	
Air Blank	0.000	22:18	
Subject Sample	# 1 0.181	22:20	
Breath Volume	1.652	22:20	
Air Blank	0.000	22:21	
Air Blank	0.000	22:22	
Subject Sample	# 2	22:25	Sequence Aborted
Breath Volume	0.000	22:25	
Air Blank	0.000	22:25	

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUSCH,JJ
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DUSCH,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: PROB/PAROL
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: CRUZOLANDEZ

DATE/TIME OF TEST: 03/30/2018 18:42:56

Diagnostics Check	OK	18:36
Air Blank	0.000	18:37
Control Test	0.078	18:37
Air Blank	0.000	18:37
Subject Sample # 1	0.074	18:38
Breath Volume	2.289	18:38
Air Blank	0.000	18:38
Air Blank	0.000	18:40
Subject Sample # 2	0.070	18:41
Breath Volume	3.976	18:41
Air Blank	0.000	18:42
Control Test	0.079	18:42
Air Blank	0.000	18:42
Diagnostics Check	OK	18:42

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,Y,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LIPARI,
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: BOUSQUET

DATE/TIME OF TEST: 03/30/2018 22:38:20

Diagnostics Check	OK	22:32
Air Blank	0.000	22:32
Control Test	0.078	22:33
Air Blank	0.000	22:33
Subject Sample # 1	0.111	22:33
Breath Volume	3.468	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.103	22:36
Breath Volume	1.871	22:36
Air Blank	0.000	22:37
Control Test	0.078	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,KEVIN,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WOODS,H
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: PHILBERT

DATE/TIME OF TEST: 03/31/2018 02:14:13

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.078	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.069	02:09
Breath Volume	2.023	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.066	02:12
Breath Volume	1.738	02:12
Air Blank	0.000	02:13
Control Test	0.077	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,RASHAAD,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SMITH,RASHAAD
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/31/2018 05:09:14

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:59	
Control Test	0.078	04:59	
Air Blank	0.000	05:00	
Subject Sample	# 1 0.168	05:03	Volume Not Met
Breath Volume	0.226	05:03	
Air Blank	0.000	05:04	
Air Blank	0.000	05:05	
Subject Sample	# 2	05:07	Subject Test Refused
Breath Volume	0.277	05:07	
Air Blank	0.000	05:08	
Control Test	0.073	05:08	Control Outside Tolerance
Air Blank	0.000	05:09	

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 03/31/2018 06:17:15

Diagnostics Check	OK	06:11
Air Blank	0.000	06:11
Control Test	0.077	06:11
Air Blank	0.000	06:12
Subject Sample # 1	0.119	06:12
Breath Volume	2.027	06:12
Air Blank	0.000	06:13
Air Blank	0.000	06:15
Subject Sample # 2	0.118	06:15
Breath Volume	1.531	06:15
Air Blank	0.000	06:16
Control Test	0.077	06:16
Air Blank	0.000	06:17
Diagnostics Check	OK	06:17

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIKEMAN,HOUSTON,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TABARES,G
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/17/2018 03:33:11

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.079	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.198	03:28
Breath Volume		1.628	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:30
Subject Sample	# 2	0.187	03:31
Breath Volume		2.792	03:31
Air Blank		0.000	03:32
Control Test		0.079	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 04/02/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,KEVIN,G
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MORALES,GUALBERTO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: DILLON

DATE/TIME OF TEST: 03/29/2018 03:15:58

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.024	03:11
Breath Volume	2.281	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.021	03:14
Breath Volume	2.730	03:14
Air Blank	0.000	03:15
Control Test	0.078	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 04/02/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: SORIANOCARRENO

DATE/TIME OF TEST: 03/31/2018 20:48:13

Diagnostics Check	OK	20:41
Air Blank	0.000	20:42
Control Test	0.078	20:42
Air Blank	0.000	20:42
Subject Sample # 1	0.184	20:43
Breath Volume	2.410	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.184	20:46
Breath Volume	2.730	20:46
Air Blank	0.000	20:47
Control Test	0.078	20:47
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 04/02/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON, DEMETRIS, L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: JACKSON, DEMETRIS
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: ANTELO BEJARANO

DATE/TIME OF TEST: 03/16/2018 06:39:19

Diagnostics Check	OK	06:27	
Air Blank	0.000	06:28	
Control Test	0.079	06:28	
Air Blank	0.000	06:29	
Subject Sample	# 1 0.114	06:32	Volume Not Met
Breath Volume	0.191	06:32	
Air Blank	0.000	06:33	
Air Blank	0.000	06:34	
Subject Sample	# 2	06:37	No Sample Provided
Breath Volume	0.000	06:37	
Air Blank	0.000	06:38	
Control Test	0.077	06:38	
Air Blank	0.000	06:39	
Diagnostics Check	OK	06:39	

Data Download Date/Time: 04/02/2018 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEPARD,CHAD,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ALLEN,MARTIN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:23

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 03/03/2018 06:58:49

Diagnostics Check		OK	06:46	
Air Blank		0.000	06:47	
Control Test		0.080	06:47	
Air Blank		0.000	06:48	
Subject Sample	# 1	0.194	06:51	Slope Not Level
Breath Volume		1.519	06:51	
Air Blank		0.000	06:52	
Air Blank		0.000	06:53	
Subject Sample	# 2	0.205	06:54	
Breath Volume		2.441	06:54	
Air Blank		0.000	06:55	
Air Blank		0.000	06:56	
Subject Sample	# 3	0.199	06:57	
Breath Volume		2.175	06:57	
Air Blank		0.000	06:57	
Control Test		0.078	06:58	
Air Blank		0.000	06:58	
Diagnostics Check		OK	06:58	

Data Download Date/Time: 03/19/2018 02:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BOVEY,TAYLOR
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: AMADOR

DATE/TIME OF TEST: 03/06/2018 01:38:05

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.082	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.000	01:33
Breath Volume	3.675	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.000	01:36
Breath Volume	3.660	01:36
Air Blank	0.000	01:37
Control Test	0.081	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:38

Data Download Date/Time: 03/19/2018 02:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ROMAN,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: STAMBAUGH

DATE/TIME OF TEST: 03/06/2018 03:32:44

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.210	03:28
Breath Volume	1.714	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.225	03:31
Breath Volume	1.738	03:31
Air Blank	0.000	03:31
Control Test	0.081	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 03/19/2018 02:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLATZ,KEVIN,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WOELK,MATT
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: CINCOTTI

DATE/TIME OF TEST: 03/06/2018 21:15:34

Diagnostics Check	OK	21:07
Air Blank	0.000	21:08
Control Test	0.081	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.110	21:10
Breath Volume	2.058	21:10
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample # 2	0.112	21:14
Breath Volume	2.011	21:14
Air Blank	0.000	21:14
Control Test	0.081	21:14
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time: 03/19/2018 02:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TAYLOR,ROBERT
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: MIHM

DATE/TIME OF TEST: 03/06/2018 21:31:57

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.081	21:24
Air Blank	0.000	21:25
Subject Sample # 1	0.093	21:27
Breath Volume	1.707	21:27
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample # 2	0.089	21:30
Breath Volume	2.070	21:30
Air Blank	0.000	21:31
Control Test	0.081	21:31
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time: 03/19/2018 02:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,CELESTE,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SOLTIS,T
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: LEFEBVRE

DATE/TIME OF TEST: 03/08/2018 16:12:04

Diagnostics Check	OK	16:05
Air Blank	0.000	16:06
Control Test	0.082	16:06
Air Blank	0.000	16:07
Subject Sample # 1	0.264	16:07
Breath Volume	1.449	16:07
Air Blank	0.000	16:08
Air Blank	0.000	16:10
Subject Sample # 2	0.248	16:10
Breath Volume	1.980	16:10
Air Blank	0.000	16:11
Control Test	0.080	16:11
Air Blank	0.000	16:11
Diagnostics Check	OK	16:12

Data Download Date/Time: 03/19/2018 02:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: EGAN,MELISSA
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 13:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: LEFEBVRE

DATE/TIME OF TEST: 03/08/2018 21:53:39

Diagnostics Check	OK	21:47
Air Blank	0.000	21:48
Control Test	0.082	21:48
Air Blank	0.000	21:49
Subject Sample # 1	0.108	21:49
Breath Volume	1.621	21:49
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.106	21:52
Breath Volume	1.617	21:52
Air Blank	0.000	21:52
Control Test	0.081	21:53
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 03/19/2018 02:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: EGAN,MELISSA
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 13:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: LEFEBVRE

DATE/TIME OF TEST: 03/09/2018 00:42:04

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.082	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.068	00:37
Breath Volume	1.402	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.065	00:40
Breath Volume	1.347	00:40
Air Blank	0.000	00:41
Control Test	0.082	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:42

Data Download Date/Time: 03/19/2018 02:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: EGAN,MELISSA
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 13:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: LEFEBVRE

DATE/TIME OF TEST: 03/09/2018 02:54:16

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.082	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.037	02:49
Breath Volume	1.511	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.038	02:52
Breath Volume	1.410	02:52
Air Blank	0.000	02:53
Control Test	0.082	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 03/19/2018 02:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: OSKEY,AUSTIN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 03/10/2018 00:25:48

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:20	
Control Test	0.081	00:20	
Air Blank	0.000	00:21	
Subject Sample	# 1 0.138	00:21	
Breath Volume	1.386	00:21	
Air Blank	0.000	00:22	
Air Blank	0.000	00:24	
Subject Sample	# 2	00:24	Subject Test Refused
Breath Volume	0.000	00:24	
Air Blank	0.000	00:24	
Control Test	0.082	00:25	
Air Blank	0.000	00:25	
Diagnostics Check	OK	00:25	

Data Download Date/Time: 03/19/2018 02:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLATZ,KEVIN,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORSE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ST.JOHN

DATE/TIME OF TEST: 03/10/2018 22:59:45

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.128	22:55
Breath Volume	1.515	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.128	22:58
Breath Volume	1.859	22:58
Air Blank	0.000	22:58
Control Test	0.080	22:59
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 03/19/2018 02:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORSE,JUSTIN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: REILLY

DATE/TIME OF TEST: 03/11/2018 03:57:33

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.081	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.149	03:53
Breath Volume	2.398	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.151	03:56
Breath Volume	2.449	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 03/19/2018 02:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLATZ,KEVIN,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORSE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: ST.JOHN

DATE/TIME OF TEST: 03/11/2018 05:00:05

Diagnostics Check	OK	04:52
Air Blank	0.000	04:53
Control Test	0.081	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.046	04:55
Breath Volume	2.175	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:57
Subject Sample # 2	0.042	04:58
Breath Volume	1.402	04:58
Air Blank	0.000	04:59
Control Test	0.082	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	05:00

Data Download Date/Time: 03/19/2018 02:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,J,H
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CAMPBELL,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: ST JOHN

DATE/TIME OF TEST: 03/11/2018 07:40:03

Diagnostics Check	OK	07:33
Air Blank	0.000	07:34
Control Test	0.081	07:34
Air Blank	0.000	07:35
Subject Sample # 1	0.000	07:35
Breath Volume	1.968	07:35
Air Blank	0.000	07:36
Air Blank	0.000	07:38
Subject Sample # 2	0.000	07:38
Breath Volume	1.531	07:38
Air Blank	0.000	07:39
Control Test	0.080	07:39
Air Blank	0.000	07:39
Diagnostics Check	OK	07:40

Data Download Date/Time: 03/19/2018 02:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SLEDIASKI,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: BENTLEY

DATE/TIME OF TEST: 03/13/2018 21:54:25

Diagnostics Check	OK	21:47
Air Blank	0.000	21:48
Control Test	0.081	21:48
Air Blank	0.000	21:49
Subject Sample # 1	0.201	21:49
Breath Volume	2.808	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.201	21:52
Breath Volume	3.765	21:52
Air Blank	0.000	21:53
Control Test	0.081	21:53
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 03/19/2018 02:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: COTE,JOSH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: BEBON

DATE/TIME OF TEST: 03/15/2018 03:52:18

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.082	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.189	03:46
Breath Volume	1.890	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.197	03:50
Breath Volume	3.207	03:50
Air Blank	0.000	03:51
Control Test	0.082	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 03/19/2018 02:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: EARNER,JACOB
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: TATUM

DATE/TIME OF TEST: 03/17/2018 02:12:20

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.081	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.172	02:07
Breath Volume	1.660	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.173	02:10
Breath Volume	1.304	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 03/19/2018 02:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HENSHAW,N
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: SHERRY

DATE/TIME OF TEST: 03/18/2018 01:08:09

Diagnostics Check		OK	01:01
Air Blank		0.000	01:02
Control Test		0.081	01:02
Air Blank		0.000	01:03
Subject Sample	# 1	0.105	01:03
Breath Volume		2.476	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.105	01:06
Breath Volume		1.933	01:06
Air Blank		0.000	01:07
Control Test		0.080	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 03/19/2018 02:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: MADDEN

DATE/TIME OF TEST: 03/18/2018 03:35:34

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.128	03:30
Breath Volume	1.656	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.124	03:34
Breath Volume	2.738	03:34
Air Blank	0.000	03:34
Control Test	0.081	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 03/19/2018 02:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMS,DONNA,J
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: REED,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/05/2018 12:15:30

Diagnostics Check	OK	12:09
Air Blank	0.000	12:09
Control Test	0.080	12:10
Air Blank	0.000	12:10
Subject Sample # 1	0.000	12:11
Breath Volume	3.445	12:11
Air Blank	0.000	12:11
Air Blank	0.000	12:13
Subject Sample # 2	0.000	12:14
Breath Volume	1.792	12:14
Air Blank	0.000	12:14
Control Test	0.080	12:14
Air Blank	0.000	12:15
Diagnostics Check	OK	12:15

Data Download Date/Time: 03/30/2018 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,J,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: REED,M
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/05/2018 20:39:20

Diagnostics Check	OK	20:33
Air Blank	0.000	20:33
Control Test	0.080	20:34
Air Blank	0.000	20:34
Subject Sample # 1	0.000	20:35
Breath Volume	2.882	20:35
Air Blank	0.000	20:35
Air Blank	0.000	20:37
Subject Sample # 2	0.000	20:37
Breath Volume	1.820	20:37
Air Blank	0.000	20:38
Control Test	0.080	20:38
Air Blank	0.000	20:39
Diagnostics Check	OK	20:39

Data Download Date/Time: 03/30/2018 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONDRAGON, EDWIN, F
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: MURRAY, S
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 03/13/2018 20:58:53

Diagnostics Check		OK	20:52
Air Blank		0.000	20:52
Control Test		0.079	20:52
Air Blank		0.000	20:53
Subject Sample	# 1	0.223	20:53
Breath Volume		2.417	20:53
Air Blank		0.000	20:54
Air Blank		0.000	20:56
Subject Sample	# 2	0.210	20:57
Breath Volume		2.171	20:57
Air Blank		0.000	20:57
Control Test		0.079	20:58
Air Blank		0.000	20:58
Diagnostics Check		OK	20:58

Data Download Date/Time: 03/30/2018 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANCO,ARIEL,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: MURRAY,S
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: K400512697870

DATE/TIME OF TEST: 03/14/2018 05:07:29

Diagnostics Check	OK	05:00
Air Blank	0.000	05:00
Control Test	0.079	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.071	05:03
Breath Volume	1.824	05:03
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.068	05:05
Breath Volume	1.730	05:05
Air Blank	0.000	05:06
Control Test	0.078	05:06
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 03/30/2018 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNASTAL,SAMUEL,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SALGADO,TICO
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:44

SUBJECT'S LAST NAME: PROVOST

DATE/TIME OF TEST: 03/14/2018 14:36:02

Diagnostics Check	OK	14:24	
Air Blank	0.000	14:24	
Control Test	0.080	14:24	
Air Blank	0.000	14:25	
Subject Sample	# 1	14:28	No Sample Provided
Breath Volume	0.000	14:28	
Air Blank	0.000	14:29	
Air Blank	0.000	14:31	
Subject Sample	# 2 0.270	14:32	
Breath Volume	2.867	14:32	
Air Blank	0.000	14:33	
Air Blank		14:35	Ambient Fail
Air Blank	0.000	14:36	

Data Download Date/Time: 03/30/2018 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONDRAGON, EDWIN, F
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: JEAN, FRANTZ
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 13:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:44

SUBJECT'S LAST NAME: PROVOST

DATE/TIME OF TEST: 03/14/2018 22:10:55

Diagnostics Check	OK	22:04
Air Blank	0.000	22:05
Control Test	0.081	22:05
Air Blank	0.000	22:05
Subject Sample # 1	0.073	22:06
Breath Volume	2.242	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:08
Subject Sample # 2	0.072	22:09
Breath Volume	2.035	22:09
Air Blank	0.000	22:10
Control Test	0.079	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 03/30/2018 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANCO,ARIEL,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: NIEVES,TITO
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: HARLOW

DATE/TIME OF TEST: 03/18/2018 00:45:15

Diagnostics Check		OK	00:38
Air Blank		0.000	00:38
Control Test		0.080	00:38
Air Blank		0.000	00:39
Subject Sample	# 1	0.216	00:40
Breath Volume		1.902	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.215	00:43
Breath Volume		1.546	00:43
Air Blank		0.000	00:44
Control Test		0.079	00:44
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 03/30/2018 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDE,JOEL,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SILVA,J
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: HARLOW

DATE/TIME OF TEST: 03/18/2018 08:09:52

Diagnostics Check	OK	08:03
Air Blank	0.000	08:03
Control Test	0.080	08:03
Air Blank	0.000	08:04
Subject Sample # 1	0.087	08:05
Breath Volume	1.148	08:05
Air Blank	0.000	08:05
Air Blank	0.000	08:07
Subject Sample # 2	0.081	08:08
Breath Volume	1.363	08:08
Air Blank	0.000	08:08
Control Test	0.078	08:09
Air Blank	0.000	08:09
Diagnostics Check	OK	08:09

Data Download Date/Time: 03/30/2018 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER, DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER, DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416090A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: MCCULLERS

DATE/TIME OF TEST: 03/02/2018 00:28:37

Diagnostics Check		OK	00:17	
Air Blank		0.000	00:17	
Control Test		0.079	00:17	
Air Blank		0.000	00:18	
Subject Sample	# 1	0.175	00:21	Volume Not Met
Breath Volume		0.265	00:21	
Air Blank		0.000	00:22	
Air Blank		0.000	00:23	
Subject Sample	# 2	0.148	00:27	Volume Not Met
Breath Volume		0.097	00:27	
Air Blank		0.000	00:27	
Control Test		0.080	00:28	
Air Blank		0.000	00:28	
Diagnostics Check		OK	00:28	

Data Download Date/Time: 03/29/2018 14:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416090A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: BALE

DATE/TIME OF TEST: 03/03/2018 21:49:27

Diagnostics Check	OK	21:41
Air Blank	0.000	21:42
Control Test	0.083	21:42
Air Blank	0.000	21:42
Subject Sample # 1	0.197	21:43
Breath Volume	1.406	21:43
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample # 2	0.211	21:47
Breath Volume	1.648	21:47
Air Blank	0.000	21:48
Control Test	0.084	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 03/29/2018 14:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416090A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: WOMER

DATE/TIME OF TEST: 03/10/2018 21:46:28

Diagnostics Check	OK	21:40
Air Blank	0.000	21:40
Control Test	0.081	21:40
Air Blank	0.000	21:41
Subject Sample # 1	0.000	21:42
Breath Volume	1.910	21:42
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.000	21:45
Breath Volume	1.542	21:45
Air Blank	0.000	21:45
Control Test	0.081	21:45
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 03/29/2018 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER, DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER, DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416090A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 03/11/2018 01:44:41

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.082	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.199	01:39
Breath Volume	3.250	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.198	01:43
Breath Volume	2.949	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 03/29/2018 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER, DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER, DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416090A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 03/16/2018 02:24:53

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:17	
Control Test	0.082	02:17	
Air Blank	0.000	02:17	
Subject Sample # 1	0.251	02:18	
Breath Volume	2.191	02:18	
Air Blank	0.000	02:18	
Air Blank	0.000	02:20	
Subject Sample # 2	0.250	02:23	
Breath Volume	2.238	02:23	
Air Blank		02:24	Purge Fail
Air Blank	0.000	02:24	

Data Download Date/Time: 03/29/2018 14:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416090A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/23/2018 23:19:48

Diagnostics Check	OK	23:15	
Air Blank	0.000	23:16	
Control Test	0.079	23:16	
Air Blank	0.000	23:17	
Subject Sample # 1	0.194	23:18	
Breath Volume	1.683	23:18	
Air Blank		23:19	Purge Fail
Air Blank	0.000	23:19	

Data Download Date/Time: 03/29/2018 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTOPASSI,BARON,H
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: CORTOPASSI,BARON
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 02/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: TULLY

DATE/TIME OF TEST: 03/03/2018 17:35:34

Diagnostics Check	OK	17:28
Air Blank	0.000	17:29
Control Test	0.083	17:29
Air Blank	0.000	17:29
Subject Sample # 1	0.132	17:30
Breath Volume	1.289	17:30
Air Blank	0.000	17:31
Air Blank	0.000	17:33
Subject Sample # 2	0.126	17:34
Breath Volume	1.191	17:34
Air Blank	0.000	17:34
Control Test	0.083	17:34
Air Blank	0.000	17:35
Diagnostics Check	OK	17:35

Data Download Date/Time: 03/06/2018 10:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUVERSBURK,BRENNA,M
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: HAUVERSBURK,BRENNA
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 02/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GEARHART

DATE/TIME OF TEST: 03/04/2018 01:16:00

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:15	
Control Test	0.083	01:15	
Air Blank		01:15	RFI Detect
Air Blank	0.000	01:16	

Data Download Date/Time: 03/06/2018 10:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUVERSBURK,BRENNA,M
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: HAUVERSBURK,BRENNA
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 02/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GEARHART

DATE/TIME OF TEST: 03/04/2018 01:29:29

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.083	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1 0.151	01:21	Volume Not Met
Breath Volume	0.710	01:21	
Air Blank	0.000	01:22	
Air Blank	0.000	01:24	
Subject Sample	# 2 0.159	01:24	
Breath Volume	3.851	01:24	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample	# 3 0.154	01:27	
Breath Volume	2.914	01:27	
Air Blank	0.000	01:28	
Control Test	0.082	01:28	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 03/06/2018 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,LONI,A
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: HOLTON,AMBER
ARRESTING OFFICER AGENCY: Carrabelle PD
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 15:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:01

SUBJECT'S LAST NAME: DIESTELHORST

DATE/TIME OF TEST: 03/07/2018 16:27:02

Diagnostics Check	OK	16:21
Air Blank	0.000	16:21
Control Test	0.084	16:22
Air Blank	0.000	16:22
Subject Sample # 1	0.000	16:22
Breath Volume	1.574	16:22
Air Blank	0.000	16:23
Air Blank	0.000	16:25
Subject Sample # 2	0.000	16:25
Breath Volume	1.703	16:25
Air Blank	0.000	16:26
Control Test	0.082	16:26
Air Blank	0.000	16:26
Diagnostics Check	OK	16:27

Data Download Date/Time: 04/12/2018 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEWITT,JEFF,
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: HEWITT,JEFF
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 15:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 03/09/2018 16:36:21

Diagnostics Check	OK	16:25	
Air Blank	0.000	16:25	
Control Test	0.082	16:25	
Air Blank	0.000	16:26	
Subject Sample	# 1	16:29	No Sample Provided
Breath Volume	0.000	16:29	
Air Blank	0.000	16:30	
Air Blank	0.000	16:31	
Subject Sample	# 2	16:35	No Sample Provided
Breath Volume	0.000	16:35	
Air Blank	0.000	16:35	
Control Test	0.080	16:35	
Air Blank	0.000	16:36	
Diagnostics Check	OK	16:36	

Data Download Date/Time: 04/12/2018 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUVERSBURK,BRENNA,M
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: HAUVERSBURK,BRENNA
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: DE LA CRUZ

DATE/TIME OF TEST: 03/11/2018 03:58:24

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.083	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.128	03:53
Breath Volume	2.335	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.131	03:56
Breath Volume	2.308	03:56
Air Blank	0.000	03:57
Control Test	0.082	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 04/12/2018 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS, SR.,MICHAEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: CROSS, SR.,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MUGGRIDGE

DATE/TIME OF TEST: 03/24/2018 23:12:15

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.082	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.186	23:07
Breath Volume	1.621	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:09
Subject Sample # 2	0.174	23:10
Breath Volume	1.246	23:10
Air Blank	0.000	23:11
Control Test	0.080	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 04/12/2018 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMM,JAMES,L
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: MILLER,KRISTEN
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 03/27/2018 20:05:08

Diagnostics Check	OK	19:58
Air Blank	0.000	19:59
Control Test	0.083	19:59
Air Blank	0.000	20:00
Subject Sample # 1	0.173	20:00
Breath Volume	2.347	20:00
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample # 2	0.177	20:03
Breath Volume	2.269	20:03
Air Blank	0.000	20:04
Control Test	0.082	20:04
Air Blank	0.000	20:05
Diagnostics Check	OK	20:05

Data Download Date/Time: 04/12/2018 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUVERSBURK,BRENNA,M
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: HAUVERSBURK,BRENNA
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: CADWALLADER

DATE/TIME OF TEST: 03/28/2018 22:00:30

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:55	
Control Test	0.082	21:56	
Air Blank	0.000	21:56	
Subject Sample	# 1	21:58	Subject Test Refused
Breath Volume	0.000	21:58	
Air Blank	0.000	21:59	
Control Test	0.082	21:59	
Air Blank		22:00	RFI Detect
Air Blank	0.000	22:00	

Data Download Date/Time: 04/12/2018 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUVERSBURK,BRENNA,M
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: HAUVERSBURK,BRENNA
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: CADWALLADER

DATE/TIME OF TEST: 03/28/2018 22:05:34

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:03	
Control Test	0.082	22:03	
Air Blank	0.000	22:03	
Subject Sample	# 1	22:04	Subject Test Refused
Breath Volume	0.000	22:04	
Air Blank	0.000	22:04	
Control Test	0.082	22:04	
Air Blank	0.000	22:05	
Diagnostics Check	OK	22:05	

Data Download Date/Time: 04/12/2018 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMM,JAMES,L
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: SMITH,LARRY
ARRESTING OFFICER AGENCY: Apalachicola PD
LAST AGENCY INSPECTION: 03/06/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LINGG

DATE/TIME OF TEST: 03/31/2018 00:00:22

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.083	23:55
Air Blank	0.000	23:55
Subject Sample # 1	0.000	23:56
Breath Volume	1.562	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.000	23:59
Breath Volume	1.832	23:59
Air Blank	0.000	23:59
Control Test	0.081	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 04/12/2018 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: GUTSHALL,NICHOLAS
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: DOMINGUEZ MOLINA

DATE/TIME OF TEST: 03/04/2018 19:49:23

Diagnostics Check	OK	19:46	
Air Blank	0.000	19:46	
Control Test	0.082	19:47	
Air Blank	0.000	19:47	
Subject Sample	# 1	19:48	Subject Test Refused
Breath Volume	0.000	19:48	
Air Blank	0.000	19:48	
Control Test	0.081	19:48	
Air Blank	0.000	19:49	
Diagnostics Check	OK	19:49	

Data Download Date/Time: 03/29/2018 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: BLACKMAN,ANTHONY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: FLANNERY

DATE/TIME OF TEST: 03/20/2018 00:41:19

Diagnostics Check		OK	00:28	
Air Blank		0.000	00:29	
Control Test		0.080	00:29	
Air Blank		0.000	00:30	
Subject Sample	# 1	0.258	00:30	No .020 agreement
Breath Volume		1.433	00:30	
Air Blank		0.000	00:31	
Air Blank		0.000	00:33	
Subject Sample	# 2	0.282	00:34	No .020 agreement
Breath Volume		1.546	00:34	
Air Blank		0.000	00:34	
Air Blank		0.000	00:36	
Subject Sample	# 3	0.244	00:39	Volume Not Met
Breath Volume		0.218	00:39	
Air Blank		0.000	00:40	
Control Test		0.081	00:40	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 03/29/2018 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: BLACKMAN,ANTHONY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: FLANNERY

DATE/TIME OF TEST: 03/20/2018 01:17:05

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.079	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.245	01:12
Breath Volume		1.972	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.246	01:15
Breath Volume		1.738	01:15
Air Blank		0.000	01:16
Control Test		0.080	01:16
Air Blank		0.000	01:16
Diagnostics Check		OK	01:17

Data Download Date/Time: 03/29/2018 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HESTER,JAMES,R

BREATH TEST OPERATOR AGENCY: Clewiston PD

ARRESTING OFFICER: GANNON,DAVID

ARRESTING OFFICER AGENCY: Clewiston PD

LAST AGENCY INSPECTION: 02/27/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 644476

DRY GAS EXPIRY DATE: 12/01/2019

VIOLATION CODE: DUI

ARREST TIME: 01:46

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: GARCIA ALVEREZ

DATE/TIME OF TEST: 03/25/2018 02:30:16

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:21	
Control Test	0.081	02:21	
Air Blank	0.000	02:22	
Subject Sample	# 1 0.124	02:25	Volume Not Met
Breath Volume	0.058	02:25	
Air Blank	0.000	02:25	
Air Blank	0.000	02:27	
Subject Sample	# 2	02:28	Subject Test Refused
Breath Volume	0.050	02:28	
Air Blank	0.000	02:29	
Control Test	0.081	02:29	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 03/29/2018 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAUFFMAN, MITCHELL, P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: KAUFFMAN, MITCHELL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: BRUMBLEY

DATE/TIME OF TEST: 03/17/2018 08:32:20

Diagnostics Check	OK	08:25
Air Blank	0.000	08:26
Control Test	0.081	08:26
Air Blank	0.000	08:27
Subject Sample # 1	0.144	08:27
Breath Volume	3.394	08:27
Air Blank	0.000	08:28
Air Blank	0.000	08:30
Subject Sample # 2	0.141	08:30
Breath Volume	2.175	08:30
Air Blank	0.000	08:31
Control Test	0.079	08:31
Air Blank	0.000	08:32
Diagnostics Check	OK	08:32

Data Download Date/Time: 04/09/2018 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAUFFMAN, MITCHELL, P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HEDGES, MATTHEW
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 03/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 03/26/2018 19:44:47

Diagnostics Check	OK	19:40	
Air Blank	0.000	19:41	
Control Test	0.080	19:41	
Air Blank	0.000	19:42	
Subject Sample	# 1	19:43	Subject Test Refused
Breath Volume	1.761	19:43	
Air Blank	0.000	19:43	
Control Test	0.081	19:44	
Air Blank	0.000	19:44	
Diagnostics Check	OK	19:44	

Data Download Date/Time: 04/09/2018 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAUFFMAN, MITCHELL, P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HEDGES, MATTHEW
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 03/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 03/26/2018 19:58:22

Diagnostics Check		OK	19:46	
Air Blank		0.000	19:47	
Control Test		0.080	19:47	
Air Blank		0.000	19:48	
Subject Sample	# 1	0.187	19:51	Volume Not Met
Breath Volume		0.175	19:51	
Air Blank		0.000	19:52	
Air Blank		0.000	19:53	
Subject Sample	# 2	0.180	19:56	Volume Not Met
Breath Volume		0.015	19:56	
Air Blank		0.000	19:57	
Control Test		0.081	19:57	
Air Blank		0.000	19:58	
Diagnostics Check		OK	19:58	

Data Download Date/Time: 04/09/2018 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WONG SHUE,A
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 03/01/2018 01:33:48

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.081	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.144	01:29
Breath Volume	2.867	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.145	01:32
Breath Volume	3.273	01:32
Air Blank	0.000	01:32
Control Test	0.081	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 03/08/2018 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BILLINGS

DATE/TIME OF TEST: 03/01/2018 03:20:24

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:09	
Control Test		0.082	03:09	
Air Blank		0.000	03:10	
Subject Sample	# 1	0.134	03:13	Volume Not Met
Breath Volume		0.699	03:13	
Air Blank		0.000	03:13	
Air Blank		0.000	03:15	
Subject Sample	# 2	0.143	03:18	Volume Not Met
Breath Volume		0.812	03:18	
Air Blank		0.000	03:19	
Control Test		0.081	03:19	
Air Blank		0.000	03:20	
Diagnostics Check		OK	03:20	

Data Download Date/Time: 03/08/2018 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROBBINS,NATHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: REDMOND

DATE/TIME OF TEST: 03/02/2018 01:08:54

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:57	
Control Test	0.082	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1	01:01	No Sample Provided
Breath Volume	0.000	01:01	
Air Blank	0.000	01:02	
Air Blank	0.000	01:04	
Subject Sample	# 2	01:07	No Sample Provided
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	
Control Test	0.082	01:08	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 03/08/2018 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WARNER,MICHAEL
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 03/03/2018 01:46:07

Diagnostics Check	OK	01:39
Air Blank	0.000	01:40
Control Test	0.082	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.139	01:41
Breath Volume	3.351	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.135	01:44
Breath Volume	2.773	01:44
Air Blank	0.000	01:45
Control Test	0.081	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 03/08/2018 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATHURIN,WILBERTE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: PEREZ SOTO

DATE/TIME OF TEST: 03/03/2018 02:53:12

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.082	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.198	02:48
Breath Volume	2.453	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.180	02:51
Breath Volume	1.378	02:51
Air Blank	0.000	02:52
Control Test	0.081	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 03/08/2018 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DAVILA,S
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: CUELLAR

DATE/TIME OF TEST: 03/03/2018 04:25:38

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.082	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.097	04:21
Breath Volume	3.171	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.086	04:24
Breath Volume	2.535	04:24
Air Blank	0.000	04:24
Control Test	0.081	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 03/08/2018 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MASSICOTTE,MICHAEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: MEZA

DATE/TIME OF TEST: 03/04/2018 02:22:29

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:14	
Control Test	0.082	02:14	
Air Blank	0.000	02:15	
Subject Sample	# 1	02:18	No Sample Provided
Breath Volume	0.000	02:18	
Air Blank	0.000	02:18	
Air Blank	0.000	02:20	
Subject Sample	# 2	02:21	Subject Test Refused
Breath Volume	0.000	02:21	
Air Blank	0.000	02:21	
Control Test	0.082	02:21	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 03/08/2018 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BECK,KURTIS
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: MYRIE

DATE/TIME OF TEST: 03/04/2018 03:23:24

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:11	
Control Test	0.082	03:11	
Air Blank	0.000	03:12	
Subject Sample # 1	0.207	03:13	
Breath Volume	1.519	03:13	
Air Blank	0.000	03:13	
Air Blank	0.000	03:15	
Subject Sample # 2	0.173	03:19	Volume Not Met
Breath Volume	0.218	03:19	
Air Blank	0.000	03:19	
Air Blank	0.000	03:21	
Subject Sample # 3		03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:22	
Control Test	0.080	03:22	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 03/08/2018 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: SHAH

DATE/TIME OF TEST: 03/04/2018 04:59:26

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:45	
Control Test	0.082	04:45	
Air Blank	0.000	04:46	
Subject Sample	# 1	04:49	No Sample Provided
Breath Volume	0.000	04:49	
Air Blank	0.000	04:49	
Air Blank	0.000	04:51	
Subject Sample	# 2	0.104	04:53
Breath Volume	2.390	04:53	
Air Blank	0.000	04:54	
Air Blank	0.000	04:56	
Subject Sample	# 3	0.095	04:57
Breath Volume	1.570	04:57	
Air Blank	0.000	04:58	
Control Test	0.081	04:58	
Air Blank	0.000	04:59	
Diagnostics Check	OK	04:59	

Data Download Date/Time: 03/08/2018 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBIJ
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GUEVARA,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 03/04/2018 05:43:12

Diagnostics Check	OK	05:40	
Air Blank	0.000	05:40	
Control Test	0.082	05:41	
Air Blank	0.000	05:41	
Subject Sample	# 1	05:41	Subject Test Refused
Breath Volume	0.000	05:41	
Air Blank	0.000	05:42	
Control Test	0.082	05:42	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time: 03/08/2018 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 03/09/2018 01:30:25

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:19	
Control Test	0.082	01:19	
Air Blank	0.000	01:20	
Subject Sample	# 1	01:23	No Sample Provided
Breath Volume	0.000	01:23	
Air Blank	0.000	01:23	
Air Blank	0.000	01:25	
Subject Sample	# 2	01:29	No Sample Provided
Breath Volume	0.000	01:29	
Air Blank	0.000	01:29	
Control Test	0.082	01:29	
Air Blank	0.000	01:30	
Diagnostics Check	OK	01:30	

Data Download Date/Time: 03/27/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROBBINS,NATHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: CHURCH

DATE/TIME OF TEST: 03/09/2018 03:02:34

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.082	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.207	02:58
Breath Volume	2.175	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.213	03:01
Breath Volume	2.820	03:01
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 03/27/2018 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ,ERNESTO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: CROTTY

DATE/TIME OF TEST: 03/10/2018 03:56:21

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.081	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.177	03:51
Breath Volume	1.792	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.172	03:54
Breath Volume	1.437	03:54
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 03/27/2018 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPENCER,JOHN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/11/2018 01:31:36

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.082	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.000	01:27
Breath Volume	1.332	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.000	01:30
Breath Volume	1.296	01:30
Air Blank	0.000	01:30
Control Test	0.081	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 03/27/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HEBEL,GRIFFIN
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PERALTA

DATE/TIME OF TEST: 03/11/2018 04:21:00

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.082	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.138	04:16
Breath Volume	2.339	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:18
Subject Sample # 2	0.126	04:19
Breath Volume	1.750	04:19
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:21

Data Download Date/Time: 03/27/2018 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 03/11/2018 05:07:31

Diagnostics Check		OK	05:00
Air Blank		0.000	05:01
Control Test		0.081	05:01
Air Blank		0.000	05:01
Subject Sample	# 1	0.240	05:02
Breath Volume		1.785	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.239	05:06
Breath Volume		1.792	05:06
Air Blank		0.000	05:06
Control Test		0.080	05:06
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 03/27/2018 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MURPHY,SEAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 03/11/2018 06:05:27

Diagnostics Check		OK	05:59
Air Blank		0.000	05:59
Control Test		0.081	05:59
Air Blank		0.000	06:00
Subject Sample	# 1	0.212	06:01
Breath Volume		2.589	06:01
Air Blank		0.000	06:01
Air Blank		0.000	06:03
Subject Sample	# 2	0.204	06:03
Breath Volume		1.566	06:03
Air Blank		0.000	06:04
Control Test		0.080	06:04
Air Blank		0.000	06:05
Diagnostics Check		OK	06:05

Data Download Date/Time: 03/27/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BODE,JOSHUA
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 03/12/2018 01:24:50

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:13	
Control Test	0.081	01:14	
Air Blank	0.000	01:14	
Subject Sample	# 1	01:17	No Sample Provided
Breath Volume	0.000	01:17	
Air Blank	0.000	01:18	
Air Blank	0.000	01:20	
Subject Sample	# 2	01:23	No Sample Provided
Breath Volume	0.000	01:23	
Air Blank	0.000	01:23	
Control Test	0.081	01:24	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 03/27/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BODE,JOSHUA
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 03/12/2018 02:04:00

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:53	
Control Test	0.081	01:53	
Air Blank	0.000	01:53	
Subject Sample	# 1	01:57	No Sample Provided
Breath Volume	0.000	01:57	
Air Blank	0.000	01:57	
Air Blank	0.000	01:59	
Subject Sample	# 2	02:02	No Sample Provided
Breath Volume	0.000	02:02	
Air Blank	0.000	02:03	
Control Test	0.081	02:03	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:04	

Data Download Date/Time: 03/27/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 03/14/2018 01:55:05

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.083	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.000	01:50
Breath Volume	2.773	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.000	01:53
Breath Volume	2.925	01:53
Air Blank	0.000	01:54
Control Test	0.082	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 03/27/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: OCHOA,DAVID
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: NEGRON

DATE/TIME OF TEST: 03/17/2018 01:11:39

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.082	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.161	01:07
Breath Volume	1.156	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.173	01:10
Breath Volume	1.765	01:10
Air Blank	0.000	01:10
Control Test	0.081	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 03/27/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEAVER,WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: ADECLAT

DATE/TIME OF TEST: 03/17/2018 22:45:31

Diagnostics Check	OK	22:39
Air Blank	0.000	22:39
Control Test	0.083	22:40
Air Blank	0.000	22:40
Subject Sample # 1	0.184	22:41
Breath Volume	1.417	22:41
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.192	22:44
Breath Volume	1.605	22:44
Air Blank	0.000	22:44
Control Test	0.082	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 03/27/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,BRIAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: ARJUNE

DATE/TIME OF TEST: 03/18/2018 03:46:01

Diagnostics Check	OK	03:34		
Air Blank	0.000	03:34		
Control Test	0.083	03:35		
Air Blank	0.000	03:35		
Subject Sample	# 1	03:38	No Sample Provided	
Breath Volume	0.000	03:38		
Air Blank	0.000	03:39		
Air Blank	0.000	03:41		
Subject Sample	# 2	0.162	03:44	Volume Not Met
Breath Volume	0.492	03:44		
Air Blank	0.000	03:45		
Control Test	0.082	03:45		
Air Blank	0.000	03:45		
Diagnostics Check	OK	03:46		

Data Download Date/Time: 03/27/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,BRIAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: ARJUNE

DATE/TIME OF TEST: 03/18/2018 03:56:51

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:48	
Control Test	0.081	03:49	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:52	No Sample Provided
Breath Volume	0.000	03:52	
Air Blank	0.000	03:53	
Air Blank	0.000	03:55	
Subject Sample	# 2	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Control Test	0.081	03:56	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 03/27/2018 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGAN,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TUERO,ADRIAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: LANGAN

DATE/TIME OF TEST: 03/23/2018 19:25:12

Diagnostics Check	OK	19:18
Air Blank	0.000	19:19
Control Test	0.082	19:19
Air Blank	0.000	19:20
Subject Sample # 1	0.231	19:20
Breath Volume	1.687	19:20
Air Blank	0.000	19:21
Air Blank	0.000	19:22
Subject Sample # 2	0.231	19:23
Breath Volume	2.710	19:23
Air Blank	0.000	19:24
Control Test	0.081	19:24
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 03/27/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCWHERTER,PETER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: HEGGOOD

DATE/TIME OF TEST: 03/25/2018 20:49:42

Diagnostics Check	OK	20:43
Air Blank	0.000	20:43
Control Test	0.082	20:44
Air Blank	0.000	20:44
Subject Sample # 1	0.182	20:45
Breath Volume	3.531	20:45
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.178	20:48
Breath Volume	2.726	20:48
Air Blank	0.000	20:48
Control Test	0.081	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 03/27/2018 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALLWORTH, MICHAEL, L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: STALLWORTH, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: MCMILLAN

DATE/TIME OF TEST: 03/13/2018 19:57:18

Diagnostics Check	OK	19:54	
Air Blank	0.000	19:54	
Control Test	0.081	19:55	
Air Blank	0.000	19:55	
Subject Sample	# 1	19:55	Subject Test Refused
Breath Volume	0.000	19:55	
Air Blank	0.000	19:56	
Control Test	0.081	19:56	
Air Blank	0.000	19:57	
Diagnostics Check	OK	19:57	

Data Download Date/Time: 03/15/2018 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENING,FREDRICK,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HENING,FREDRICK
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 03/15/2018 01:55:30

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.081	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.172	01:51
Breath Volume	2.222	01:51
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.165	01:54
Breath Volume	1.433	01:54
Air Blank	0.000	01:54
Control Test	0.081	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 03/15/2018 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SANTIAGO MARTINEZ

DATE/TIME OF TEST: 03/24/2018 02:48:11

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.081	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.189	02:43
Breath Volume	1.886	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.195	02:46
Breath Volume	1.960	02:46
Air Blank	0.000	02:47
Control Test	0.081	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 04/24/2018 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN,CHARLES,E

BREATH TEST OPERATOR AGENCY: Gretna PD

ARRESTING OFFICER: COOPER,JAMES

ARRESTING OFFICER AGENCY: Gretna PD

LAST AGENCY INSPECTION: 03/15/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14816080A3

DRY GAS EXPIRY DATE: 07/05/2018

VIOLATION CODE: DUI

ARREST TIME: 17:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: WELSHANS

DATE/TIME OF TEST: 03/28/2018 18:19:46

Diagnostics Check	OK	18:16	
Air Blank	0.000	18:17	
Control Test	0.081	18:17	
Air Blank	0.000	18:18	
Subject Sample	# 1	18:18	Subject Test Refused
Breath Volume	0.000	18:18	
Air Blank	0.000	18:18	
Control Test	0.081	18:19	
Air Blank	0.000	18:19	
Diagnostics Check	OK	18:19	

Data Download Date/Time: 04/24/2018 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: STFLEUR

DATE/TIME OF TEST: 03/05/2018 04:17:18

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.082	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.104	04:12
Breath Volume	1.378	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.106	04:15
Breath Volume	1.488	04:15
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 03/26/2018 21:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: DOUVRES,MICHAEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: SALLEY

DATE/TIME OF TEST: 03/14/2018 03:46:00

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.081	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.177	03:41
Breath Volume	1.863	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:43
Subject Sample # 2	0.167	03:44
Breath Volume	1.871	03:44
Air Blank	0.000	03:45
Control Test	0.078	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:46

Data Download Date/Time: 03/26/2018 21:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN, JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: CHINTAMAN

DATE/TIME OF TEST: 03/19/2018 00:18:24

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.081	00:11
Air Blank	0.000	00:12
Subject Sample # 1	0.173	00:13
Breath Volume	1.292	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.164	00:16
Breath Volume	1.312	00:16
Air Blank	0.000	00:17
Control Test	0.080	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 03/26/2018 21:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: MENDOZA OCHOA

DATE/TIME OF TEST: 03/11/2018 06:55:08

Diagnostics Check	OK	06:47
Air Blank	0.000	06:48
Control Test	0.078	06:48
Air Blank	0.000	06:48
Subject Sample # 1	0.109	06:50
Breath Volume	1.257	06:50
Air Blank	0.000	06:51
Air Blank	0.000	06:53
Subject Sample # 2	0.120	06:53
Breath Volume	1.855	06:53
Air Blank	0.000	06:54
Control Test	0.078	06:54
Air Blank	0.000	06:55
Diagnostics Check	OK	06:55

Data Download Date/Time: 03/27/2018 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: SANCHEZ,JOSE
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:47

SUBJECT'S LAST NAME: RUSS

DATE/TIME OF TEST: 03/11/2018 08:16:53

Diagnostics Check		OK	08:10
Air Blank		0.000	08:10
Control Test		0.078	08:10
Air Blank		0.000	08:11
Subject Sample	# 1	0.206	08:11
Breath Volume		2.710	08:11
Air Blank		0.000	08:12
Air Blank		0.000	08:14
Subject Sample	# 2	0.195	08:15
Breath Volume		2.128	08:15
Air Blank		0.000	08:15
Control Test		0.077	08:16
Air Blank		0.000	08:16
Diagnostics Check		OK	08:16

Data Download Date/Time: 03/27/2018 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCHBACH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HOBUS,CAVIN
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: SWILPA

DATE/TIME OF TEST: 03/08/2018 21:07:56

Diagnostics Check	OK	21:06	
Air Blank		21:07	Ambient Fail
Air Blank	0.000	21:07	

Data Download Date/Time: 03/30/2018 12:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCHBACH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HOBUS,CAVIN
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: SWILPA

DATE/TIME OF TEST: 03/08/2018 21:12:35

Diagnostics Check	OK	21:11	
Air Blank		21:12	Ambient Fail
Air Blank	0.000	21:12	

Data Download Date/Time: 03/30/2018 12:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCHBACH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HOBUS,CAVIN
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: SWILPA

DATE/TIME OF TEST: 03/08/2018 21:23:12

Diagnostics Check	OK	21:17	
Air Blank	0.000	21:17	
Control Test	0.078	21:18	
Air Blank	0.000	21:18	
Subject Sample	# 1 0.340	21:18	
Breath Volume	3.160	21:18	
Air Blank	0.000	21:19	
Air Blank	0.000	21:21	
Subject Sample	# 2	21:21	Slope Not Met
Breath Volume	2.625	21:21	
Air Blank	0.000	21:22	
Control Test	0.080	21:22	
Air Blank	0.000	21:23	
Diagnostics Check	OK	21:23	

Data Download Date/Time: 03/30/2018 12:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCHBACH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HOBUS,CAVIN
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: SWILPA

DATE/TIME OF TEST: 03/08/2018 21:34:12

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.079	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.326	21:29
Breath Volume	2.992	21:29
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.334	21:32
Breath Volume	3.226	21:32
Air Blank	0.000	21:33
Control Test	0.080	21:33
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

Data Download Date/Time: 03/30/2018 12:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: WRIGHT,TJ
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/16/2018 22:47:08

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.081	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.113	22:42
Breath Volume	2.269	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:44
Subject Sample # 2	0.108	22:45
Breath Volume	2.375	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 04/30/2018 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: DONALDSON,LAUREN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: ROLFE

DATE/TIME OF TEST: 03/18/2018 21:10:05

Diagnostics Check	OK	21:03
Air Blank	0.000	21:04
Control Test	0.080	21:04
Air Blank	0.000	21:05
Subject Sample # 1	0.241	21:05
Breath Volume	3.156	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:08
Subject Sample # 2	0.229	21:08
Breath Volume	2.804	21:08
Air Blank	0.000	21:09
Control Test	0.081	21:09
Air Blank	0.000	21:09
Diagnostics Check	OK	21:10

Data Download Date/Time: 04/30/2018 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOYCE,KEVIN,G
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: GIBBS,JAMES
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:12

SUBJECT'S LAST NAME: LUDDEN

DATE/TIME OF TEST: 03/29/2018 14:45:01

Diagnostics Check	OK	14:38
Air Blank	0.000	14:39
Control Test	0.080	14:39
Air Blank	0.000	14:40
Subject Sample # 1	0.374	14:40
Breath Volume	1.921	14:40
Air Blank	0.000	14:41
Air Blank	0.000	14:42
Subject Sample # 2	0.372	14:43
Breath Volume	1.894	14:43
Air Blank	0.000	14:44
Control Test	0.079	14:44
Air Blank	0.000	14:44
Diagnostics Check	OK	14:45

Data Download Date/Time: 04/30/2018 15:55