

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WESTERBAND,FRANK
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: RODDENBERRY

DATE/TIME OF TEST: 04/01/2018 04:58:09

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.080	04:49
Air Blank	0.000	04:50
Subject Sample # 1	0.189	04:50
Breath Volume	2.351	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.159	04:53
Breath Volume	1.386	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 3	0.182	04:56
Breath Volume	2.515	04:56
Air Blank	0.000	04:57
Control Test	0.081	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 04/06/2018 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 04/03/2018 23:36:53

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:34	
Control Test	0.080	23:34	
Air Blank	0.000	23:35	
Subject Sample	# 1	23:35	Subject Test Refused
Breath Volume	0.000	23:35	
Air Blank	0.000	23:35	
Control Test	0.081	23:36	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 04/06/2018 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 04/07/2018 01:11:02

Diagnostics Check	OK	01:08	
Air Blank	0.000	01:08	
Control Test	0.081	01:08	
Air Blank	0.000	01:09	
Subject Sample	# 1	01:09	Subject Test Refused
Breath Volume	0.000	01:09	
Air Blank	0.000	01:10	
Control Test	0.080	01:10	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 05/06/2018 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SWOGGER, EDWARD
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 04/19/2018 23:28:57

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.081	23:23
Air Blank	0.000	23:23
Subject Sample # 1	0.210	23:24
Breath Volume	2.695	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:26
Subject Sample # 2	0.222	23:27
Breath Volume	2.429	23:27
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 05/06/2018 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:37

SUBJECT'S LAST NAME: FONT

DATE/TIME OF TEST: 04/24/2018 18:01:10

Diagnostics Check	OK	17:58	
Air Blank	0.000	17:58	
Control Test	0.080	17:59	
Air Blank	0.000	17:59	
Subject Sample	# 1	17:59	Subject Test Refused
Breath Volume	0.000	17:59	
Air Blank	0.000	18:00	
Control Test	0.080	18:00	
Air Blank	0.000	18:01	
Diagnostics Check	OK	18:01	

Data Download Date/Time: 05/06/2018 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SWOGGER, EDWARD
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: ROEILL

DATE/TIME OF TEST: 04/24/2018 22:44:14

Diagnostics Check	OK	22:41	
Air Blank	0.000	22:41	
Control Test	0.080	22:42	
Air Blank	0.000	22:42	
Subject Sample	# 1	22:42	Subject Test Refused
Breath Volume	0.000	22:42	
Air Blank	0.000	22:43	
Control Test	0.080	22:43	
Air Blank	0.000	22:44	
Diagnostics Check	OK	22:44	

Data Download Date/Time: 05/06/2018 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SWOGGER, EDWARD
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: PANGMAN

DATE/TIME OF TEST: 04/26/2018 16:41:04

Diagnostics Check	OK	16:34
Air Blank	0.000	16:35
Control Test	0.080	16:35
Air Blank	0.000	16:36
Subject Sample # 1	0.229	16:36
Breath Volume	1.882	16:36
Air Blank	0.000	16:37
Air Blank	0.000	16:39
Subject Sample # 2	0.213	16:39
Breath Volume	1.140	16:39
Air Blank	0.000	16:40
Control Test	0.080	16:40
Air Blank	0.000	16:40
Diagnostics Check	OK	16:41

Data Download Date/Time: 05/06/2018 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SWOGGER, EDWARD
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 04/28/2018 03:43:15

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.080	03:41	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:42	
Control Test	0.081	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 05/06/2018 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WESTERBAND, FRANK
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: LAGOS

DATE/TIME OF TEST: 04/29/2018 04:00:43

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:58	
Control Test	0.079	03:58	
Air Blank	0.000	03:59	
Subject Sample	# 1	03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	03:59	
Control Test	0.080	04:00	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 05/06/2018 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAGO,SHEILA,Y
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MOLINA,IGNACIO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 04/01/2018 23:59:01

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:56	
Control Test	0.081	23:56	
Air Blank	0.000	23:57	
Subject Sample	# 1	23:57	Subject Test Refused
Breath Volume	0.000	23:57	
Air Blank	0.000	23:58	
Control Test	0.081	23:58	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 04/04/2018 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KILMURRAY,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KILMURRAY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: FIGLIUOLO

DATE/TIME OF TEST: 04/09/2018 02:36:24

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:32	
Control Test	0.081	02:32	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:35	
Control Test	0.081	02:35	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 05/04/2018 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERNS,CODY,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WILLIAMS,TONY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 04/12/2018 23:38:16

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.082	23:36	
Air Blank	0.000	23:36	
Subject Sample	# 1	23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:37	
Control Test	0.082	23:37	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 05/04/2018 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KILMURRAY,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ROCHA,PAUL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: ARRAZCAETA

DATE/TIME OF TEST: 04/13/2018 01:32:44

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:30	
Control Test	0.082	01:30	
Air Blank	0.000	01:31	
Subject Sample	# 1	01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:31	
Control Test	0.082	01:32	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 05/04/2018 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MCKENNA 3

DATE/TIME OF TEST: 04/13/2018 01:50:43

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:48	
Control Test	0.081	01:48	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:49	
Control Test	0.081	01:50	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 05/04/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,BRANDON,P
BREATH TEST OPERATOR AGENCY: Everglades Nat Park
ARRESTING OFFICER: KIGER,DANIEL
ARRESTING OFFICER AGENCY: Everglades Nat Park
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: BUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: DEFERIA

DATE/TIME OF TEST: 04/14/2018 17:21:09

Diagnostics Check	OK	17:12
Air Blank	0.000	17:13
Control Test	0.081	17:13
Air Blank	0.000	17:14
Subject Sample # 1	0.216	17:16
Breath Volume	1.375	17:16
Air Blank	0.000	17:16
Air Blank	0.000	17:18
Subject Sample # 2	0.219	17:19
Breath Volume	2.742	17:19
Air Blank	0.000	17:20
Control Test	0.079	17:20
Air Blank	0.000	17:21
Diagnostics Check	OK	17:21

Data Download Date/Time: 05/04/2018 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAGO,SHEILA,Y
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SEAGO,SHEILA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: ABATE

DATE/TIME OF TEST: 04/15/2018 02:18:26

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.081	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.136	02:13
Breath Volume	1.777	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.136	02:16
Breath Volume	1.761	02:16
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 05/04/2018 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAGO,SHEILA,Y
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SEAGO,SHEILA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: ABATE

DATE/TIME OF TEST: 04/15/2018 02:57:47

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.082	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.123	02:53
Breath Volume	1.960	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.124	02:56
Breath Volume	1.898	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 05/04/2018 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ROCHA,PAUL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: LYNN

DATE/TIME OF TEST: 04/17/2018 03:00:35

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:58	
Control Test	0.082	02:58	
Air Blank	0.000	02:58	
Subject Sample	# 1	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	02:59	
Control Test	0.082	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 05/04/2018 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'KEEFE,PATRICIA,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: PARRIS

DATE/TIME OF TEST: 04/19/2018 00:07:59

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:05	
Control Test	0.082	00:05	
Air Blank	0.000	00:06	
Subject Sample	# 1	00:06	Subject Test Refused
Breath Volume	0.000	00:06	
Air Blank	0.000	00:07	
Control Test	0.082	00:07	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 05/04/2018 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 04/20/2018 19:03:56

Diagnostics Check	OK	18:56
Air Blank	0.000	18:57
Control Test	0.081	18:57
Air Blank	0.000	18:58
Subject Sample # 1	0.147	18:59
Breath Volume	3.578	18:59
Air Blank	0.000	18:59
Air Blank	0.000	19:01
Subject Sample # 2	0.142	19:02
Breath Volume	2.976	19:02
Air Blank	0.000	19:03
Control Test	0.081	19:03
Air Blank	0.000	19:03
Diagnostics Check	OK	19:03

Data Download Date/Time: 05/04/2018 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEAN,PAUL,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WILLIAMS,TONY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: HERNANDEZ ARRIETA

DATE/TIME OF TEST: 04/20/2018 22:20:15

Diagnostics Check	OK	22:13
Air Blank	0.000	22:14
Control Test	0.082	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.158	22:15
Breath Volume	1.968	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.159	22:18
Breath Volume	1.550	22:18
Air Blank	0.000	22:19
Control Test	0.082	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 05/04/2018 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAGO,SHEILA,Y
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: DUTOIT

DATE/TIME OF TEST: 04/29/2018 20:41:48

Diagnostics Check	OK	20:38	
Air Blank	0.000	20:39	
Control Test	0.082	20:39	
Air Blank	0.000	20:39	
Subject Sample	# 1	20:40	Subject Test Refused
Breath Volume	0.000	20:40	
Air Blank	0.000	20:40	
Control Test	0.081	20:41	
Air Blank	0.000	20:41	
Diagnostics Check	OK	20:41	

Data Download Date/Time: 05/04/2018 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PAGE,KYLE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 04/30/2018 21:06:28

Diagnostics Check	OK	20:59
Air Blank	0.000	21:00
Control Test	0.081	21:00
Air Blank	0.000	21:00
Subject Sample # 1	0.117	21:01
Breath Volume	2.050	21:01
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample # 2	0.114	21:04
Breath Volume	1.832	21:04
Air Blank	0.000	21:05
Control Test	0.081	21:05
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 05/04/2018 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYA,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: SMITH,RYAN
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 03/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: DIAZORTEGA

DATE/TIME OF TEST: 04/06/2018 23:32:19

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:29	
Control Test	0.078	23:30	
Air Blank	0.000	23:30	
Subject Sample	# 1	23:30	Subject Test Refused
Breath Volume	0.000	23:30	
Air Blank	0.000	23:31	
Control Test	0.078	23:31	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

Data Download Date/Time: 05/29/2018 09:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYA,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: NAYA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 03/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: JIMENEZHIDALGO

DATE/TIME OF TEST: 04/07/2018 04:06:33

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:02	
Control Test	0.079	04:02	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.253	04:04	
Air Blank	0.000	04:05	
Control Test	0.079	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 05/29/2018 09:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY, JONATHAN, P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY, JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: WORLEY

DATE/TIME OF TEST: 04/08/2018 02:59:06

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.081	02:51	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:54	No Sample Provided
Breath Volume	0.000	02:54	
Air Blank	0.000	02:55	
Air Blank	0.000	02:57	
Subject Sample	# 2	02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:58	
Control Test	0.082	02:58	
Air Blank	0.000	02:59	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 05/16/2018 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: REIS

DATE/TIME OF TEST: 04/28/2018 01:39:43

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.082	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.073	01:34
Breath Volume	2.644	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.064	01:38
Breath Volume	1.601	01:38
Air Blank	0.000	01:38
Control Test	0.081	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 04/30/2018 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,FREDERICK,A
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: DOMINGUEZ,FREDERICK
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: VELOSO

DATE/TIME OF TEST: 04/30/2018 09:56:30

Diagnostics Check	OK	09:50
Air Blank	0.000	09:51
Control Test	0.081	09:51
Air Blank	0.000	09:52
Subject Sample # 1	0.000	09:52
Breath Volume	2.609	09:52
Air Blank	0.000	09:52
Air Blank	0.000	09:54
Subject Sample # 2	0.000	09:55
Breath Volume	2.937	09:55
Air Blank	0.000	09:55
Control Test	0.081	09:55
Air Blank	0.000	09:56
Diagnostics Check	OK	09:56

Data Download Date/Time: 04/30/2018 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 14:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: SUSKEY

DATE/TIME OF TEST: 04/30/2018 15:02:29

Diagnostics Check	OK	14:55
Air Blank	0.000	14:56
Control Test	0.081	14:56
Air Blank	0.000	14:57
Subject Sample # 1	0.000	14:58
Breath Volume	1.410	14:58
Air Blank	0.000	14:58
Air Blank	0.000	15:00
Subject Sample # 2	0.000	15:01
Breath Volume	1.410	15:01
Air Blank	0.000	15:01
Control Test	0.081	15:01
Air Blank	0.000	15:02
Diagnostics Check	OK	15:02

Data Download Date/Time: 05/31/2018 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELFNER,ISAAC,S
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: HELFNER,ISAAC
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:53

SUBJECT'S LAST NAME: BEITING

DATE/TIME OF TEST: 04/10/2018 14:24:59

Diagnostics Check	OK	14:23	
Air Blank	0.000	14:24	
Control Test	0.074	14:24	Control Outside Tolerance
Air Blank	0.000	14:24	

Data Download Date/Time: 04/20/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELFNER,ISAAC,S
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: HELFNER,ISAAC
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:53

SUBJECT'S LAST NAME: BEITING

DATE/TIME OF TEST: 04/10/2018 14:26:59

Diagnostics Check	OK	14:26	
Air Blank		14:26	RFI Detect
Air Blank		14:26	RFI Detect

Data Download Date/Time: 04/20/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELFNER,ISAAC,S
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: HELFNER,ISAAC
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:53

SUBJECT'S LAST NAME: BEITING

DATE/TIME OF TEST: 04/10/2018 14:35:03

Diagnostics Check	OK	14:28
Air Blank	0.000	14:29
Control Test	0.075	14:29
Air Blank	0.000	14:30
Subject Sample # 1	0.204	14:30
Breath Volume	2.777	14:30
Air Blank	0.000	14:31
Air Blank	0.000	14:33
Subject Sample # 2	0.195	14:33
Breath Volume	2.113	14:33
Air Blank	0.000	14:34
Control Test	0.077	14:34
Air Blank	0.000	14:34
Diagnostics Check	OK	14:35

Data Download Date/Time: 04/20/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:41

SUBJECT'S LAST NAME: HILAIRE

DATE/TIME OF TEST: 04/13/2018 07:09:09

Diagnostics Check	OK	07:01
Air Blank	0.000	07:02
Control Test	0.082	07:02
Air Blank	0.000	07:03
Subject Sample # 1	0.100	07:04
Breath Volume	1.425	07:04
Air Blank	0.000	07:04
Air Blank	0.000	07:06
Subject Sample # 2	0.101	07:07
Breath Volume	2.839	07:07
Air Blank	0.000	07:08
Control Test	0.079	07:08
Air Blank	0.000	07:09
Diagnostics Check	OK	07:09

Data Download Date/Time: 04/20/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE,JASON,B
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: WAITE,JASON
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:10

SUBJECT'S LAST NAME: RAMOS-HERNANDEZ

DATE/TIME OF TEST: 04/14/2018 12:46:53

Diagnostics Check		OK	12:33	
Air Blank		0.000	12:34	
Control Test		0.080	12:34	
Air Blank		0.000	12:35	
Subject Sample	# 1	0.103	12:38	Volume Not Met
Breath Volume		0.000	12:38	
Air Blank		0.000	12:38	
Air Blank		0.000	12:40	
Subject Sample	# 2	0.120	12:42	
Breath Volume		1.289	12:42	
Air Blank		0.000	12:43	
Air Blank		0.000	12:44	
Subject Sample	# 3	0.121	12:45	
Breath Volume		1.425	12:45	
Air Blank		0.000	12:45	
Control Test		0.077	12:46	
Air Blank		0.000	12:46	
Diagnostics Check		OK	12:46	

Data Download Date/Time: 04/20/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROIG,DAWN,M
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: ROIG,DAWN
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: GARCIA BARRERA

DATE/TIME OF TEST: 04/14/2018 19:11:26

Diagnostics Check	OK	19:08	
Air Blank	0.000	19:09	
Control Test	0.078	19:09	
Air Blank	0.000	19:09	
Subject Sample	# 1	19:10	Subject Test Refused
Breath Volume	0.000	19:10	
Air Blank	0.000	19:10	
Control Test	0.079	19:10	
Air Blank	0.000	19:11	
Diagnostics Check	OK	19:11	

Data Download Date/Time: 04/20/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROIG,DAWN,M
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: ROIG,DAWN
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: GARCIA BARRERA

DATE/TIME OF TEST: 04/14/2018 19:23:25

Diagnostics Check	OK	19:14
Air Blank	0.000	19:15
Control Test	0.079	19:15
Air Blank	0.000	19:15
Subject Sample # 1	0.196	19:18
Breath Volume	1.769	19:18
Air Blank	0.000	19:19
Air Blank	0.000	19:20
Subject Sample # 2	0.203	19:21
Breath Volume	2.191	19:21
Air Blank	0.000	19:22
Control Test	0.077	19:22
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 04/20/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: ORELL

DATE/TIME OF TEST: 04/18/2018 21:43:16

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:40	
Control Test	0.080	21:41	
Air Blank	0.000	21:41	
Subject Sample	# 1	21:41	Subject Test Refused
Breath Volume	0.000	21:41	
Air Blank	0.000	21:42	
Control Test	0.080	21:42	
Air Blank	0.000	21:43	
Diagnostics Check	OK	21:43	

Data Download Date/Time: 04/20/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: NORWOOD

DATE/TIME OF TEST: 04/07/2018 02:31:26

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:29	
Control Test	0.080	02:29	
Air Blank	0.000	02:29	
Subject Sample	# 1	02:30	Subject Test Refused
Breath Volume	0.000	02:30	
Air Blank	0.000	02:30	
Control Test	0.080	02:30	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

Data Download Date/Time: 04/20/2018 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 04/15/2018 06:32:17

Diagnostics Check	OK	06:29	
Air Blank	0.000	06:29	
Control Test	0.080	06:30	
Air Blank	0.000	06:30	
Subject Sample	# 1	06:30	Subject Test Refused
Breath Volume	0.000	06:30	
Air Blank	0.000	06:31	
Control Test	0.080	06:31	
Air Blank	0.000	06:32	
Diagnostics Check	OK	06:32	

Data Download Date/Time: 04/20/2018 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 04/15/2018 06:41:39

Diagnostics Check	OK	06:38	
Air Blank	0.000	06:39	
Control Test	0.079	06:39	
Air Blank	0.000	06:40	
Subject Sample	# 1	06:40	Subject Test Refused
Breath Volume	0.000	06:40	
Air Blank	0.000	06:40	
Control Test	0.078	06:41	
Air Blank	0.000	06:41	
Diagnostics Check	OK	06:41	

Data Download Date/Time: 04/20/2018 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: VARGAS-MARTINEZ

DATE/TIME OF TEST: 04/01/2018 02:34:08

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.080	02:32	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:33	
Control Test	0.080	02:33	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 04/17/2018 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: DEL RIO CORREA

DATE/TIME OF TEST: 04/05/2018 02:07:35

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.086	02:03
Breath Volume	1.117	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.095	02:06
Breath Volume	2.433	02:06
Air Blank	0.000	02:06
Control Test	0.081	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 04/17/2018 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 04/05/2018 05:50:10

Diagnostics Check	OK	05:41
Air Blank	0.000	05:42
Control Test	0.080	05:42
Air Blank	0.000	05:43
Subject Sample # 1	0.094	05:45
Breath Volume	1.644	05:45
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.098	05:48
Breath Volume	2.414	05:48
Air Blank	0.000	05:49
Control Test	0.080	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time: 04/17/2018 13:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIESTEBAN,RICHARD,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GEBBIA,SALVATORE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 04/28/2018 20:58:44

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:47	
Control Test	0.079	20:48	
Air Blank	0.000	20:48	
Subject Sample	# 1	20:51	No Sample Provided
Breath Volume	0.000	20:51	
Air Blank	0.000	20:52	
Air Blank	0.000	20:54	
Subject Sample	# 2	20:57	Volume Not Met
Breath Volume	0.019	20:57	
Air Blank	0.000	20:57	
Control Test	0.079	20:58	
Air Blank	0.000	20:58	
Diagnostics Check	OK	20:58	

Data Download Date/Time: 05/15/2018 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIESTEBAN,RICHARD,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GEBBIA,SALVATORE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 04/28/2018 21:09:52

Diagnostics Check		OK	21:00	
Air Blank		0.000	21:01	
Control Test		0.078	21:01	
Air Blank		0.000	21:01	
Subject Sample	# 1	0.216	21:05	Volume Not Met
Breath Volume		0.000	21:05	
Air Blank		0.000	21:05	
Air Blank		0.000	21:07	
Subject Sample	# 2		21:08	Subject Test Refused
Breath Volume		0.000	21:08	
Air Blank		0.000	21:08	
Control Test		0.079	21:09	
Air Blank		0.000	21:09	
Diagnostics Check		OK	21:09	

Data Download Date/Time: 05/15/2018 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIESTEBAN,RICHARD,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GEBBIA,SALVATORE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 04/28/2018 21:26:18

Diagnostics Check		OK	21:11	
Air Blank		0.000	21:11	
Control Test		0.079	21:12	
Air Blank		0.000	21:12	
Subject Sample	# 1	0.232	21:15	Volume Not Met
Breath Volume		0.308	21:15	
Air Blank		0.000	21:16	
Air Blank		0.000	21:18	
Subject Sample	# 2	0.257	21:19	
Breath Volume		1.304	21:19	
Air Blank		0.000	21:19	
Air Blank		0.000	21:21	
Subject Sample	# 3	0.255	21:24	
Breath Volume		1.785	21:24	
Air Blank		0.000	21:25	
Control Test		0.075	21:25	
Air Blank		0.000	21:26	
Diagnostics Check		OK	21:26	

Data Download Date/Time: 05/15/2018 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: VALDIVIE

DATE/TIME OF TEST: 04/08/2018 06:48:39

Diagnostics Check	OK	06:45	
Air Blank	0.000	06:46	
Control Test	0.080	06:46	
Air Blank	0.000	06:46	
Subject Sample	# 1	06:47	Subject Test Refused
Breath Volume	0.000	06:47	
Air Blank	0.000	06:47	
Control Test	0.080	06:48	
Air Blank	0.000	06:48	
Diagnostics Check	OK	06:48	

Data Download Date/Time: 05/21/2018 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CECILI

DATE/TIME OF TEST: 04/13/2018 02:05:12

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.186	02:00
Breath Volume	2.144	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.194	02:03
Breath Volume	1.984	02:03
Air Blank	0.000	02:04
Control Test	0.079	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 05/21/2018 13:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, GUILLERMO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PONCE, ALFONSO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 04/12/2018 00:11:36

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.079	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.159	00:07
Breath Volume		3.945	00:07
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.154	00:10
Breath Volume		3.183	00:10
Air Blank		0.000	00:10
Control Test		0.079	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 05/21/2018 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON, GLADSTON,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LAWSON, GLADSTON
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: BARNARD

DATE/TIME OF TEST: 04/18/2018 02:08:23

Diagnostics Check	OK	02:07	
Air Blank		02:07	RFI Detect
Air Blank	0.000	02:08	

Data Download Date/Time: 05/21/2018 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON, GLADSTON,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LAWSON, GLADSTON
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: BARNARD

DATE/TIME OF TEST: 04/18/2018 02:15:56

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.078	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.145	02:11
Breath Volume	1.902	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.144	02:14
Breath Volume	2.160	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 05/21/2018 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NADAL,JOSE,V
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: NADAL,JOSE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: JERVIS

DATE/TIME OF TEST: 04/23/2018 03:55:50

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.181	03:50
Breath Volume	1.863	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.170	03:54
Breath Volume	1.609	03:54
Air Blank	0.000	03:54
Control Test	0.080	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 05/21/2018 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON, GLADSTON,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LAWSON, GLADSTON
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: TERRERO

DATE/TIME OF TEST: 04/24/2018 01:59:08

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.081	01:57	
Air Blank	0.000	01:57	
Subject Sample	# 1	01:57	Subject Test Refused
Breath Volume	0.000	01:57	
Air Blank	0.000	01:58	
Control Test	0.080	01:58	
Air Blank	0.000	01:59	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 05/21/2018 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,KENNETH,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: DURHAM,S
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: KOCH

DATE/TIME OF TEST: 04/04/2018 19:26:45

Diagnostics Check	OK	19:23	
Air Blank	0.000	19:23	
Control Test	0.079	19:24	
Air Blank	0.000	19:24	
Subject Sample	# 1	19:25	Subject Test Refused
Breath Volume	0.000	19:25	
Air Blank	0.000	19:25	
Control Test	0.076	19:26	
Air Blank	0.000	19:26	
Diagnostics Check	OK	19:26	

Data Download Date/Time: 04/30/2018 20:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AJO,MICHAEL,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AJO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: TRABANINO

DATE/TIME OF TEST: 04/12/2018 23:56:38

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:53	
Control Test	0.079	23:54	
Air Blank	0.000	23:54	
Subject Sample	# 1	23:55	Subject Test Refused
Breath Volume	0.000	23:55	
Air Blank	0.000	23:55	
Control Test	0.078	23:56	
Air Blank	0.000	23:56	
Diagnostics Check	OK	23:56	

Data Download Date/Time: 04/30/2018 20:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRERA, INGRID,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: DARDEN, DARDEN
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: MOLINARES

DATE/TIME OF TEST: 04/14/2018 19:42:46

Diagnostics Check	OK	19:38	
Air Blank	0.000	19:38	
Control Test	0.079	19:39	
Air Blank	0.000	19:39	
Subject Sample	# 1	19:41	Subject Test Refused
Breath Volume	0.000	19:41	
Air Blank	0.000	19:41	
Control Test	0.076	19:42	
Air Blank	0.000	19:42	
Diagnostics Check	OK	19:42	

Data Download Date/Time: 04/30/2018 20:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TANIS,MIKE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: OLIVARES,ANDRES
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: ROQUE

DATE/TIME OF TEST: 04/15/2018 00:32:32

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.079	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.216	00:27
Breath Volume	1.144	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:29
Subject Sample # 2	0.213	00:30
Breath Volume	1.179	00:30
Air Blank	0.000	00:31
Control Test	0.076	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 04/30/2018 20:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAPIRO,SABINA,M
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: JENKINS,DARREN
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 04/21/2018 19:57:00

Diagnostics Check		OK	19:50
Air Blank		0.000	19:50
Control Test		0.078	19:51
Air Blank		0.000	19:51
Subject Sample	# 1	0.254	19:52
Breath Volume		2.921	19:52
Air Blank		0.000	19:53
Air Blank		0.000	19:54
Subject Sample	# 2	0.261	19:55
Breath Volume		1.960	19:55
Air Blank		0.000	19:56
Control Test		0.076	19:56
Air Blank		0.000	19:56
Diagnostics Check		OK	19:57

Data Download Date/Time: 04/30/2018 20:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AJO,MICHAEL,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AJO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 04/25/2018 02:39:03

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:38	
Control Test	0.078	02:38	
Air Blank		02:38	RFI Detect
Air Blank	0.000	02:39	

Data Download Date/Time: 04/30/2018 20:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AJO,MICHAEL,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AJO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 04/25/2018 02:44:01

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.078	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:43	
Control Test	0.078	02:43	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 04/30/2018 20:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AJO,MICHAEL,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: MENDEZ,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: YOUTHE

DATE/TIME OF TEST: 04/25/2018 02:58:15

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.155	02:53
Breath Volume	1.863	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:55
Subject Sample # 2	0.154	02:56
Breath Volume	1.332	02:56
Air Blank	0.000	02:57
Control Test	0.077	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 04/30/2018 20:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000903

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCK,NATHAN,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BROCK,NATHAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: BUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: EVERHART

DATE/TIME OF TEST: 04/29/2018 18:20:21

Diagnostics Check	OK	18:13
Air Blank	0.000	18:14
Control Test	0.077	18:14
Air Blank	0.000	18:14
Subject Sample # 1	0.204	18:15
Breath Volume	2.402	18:15
Air Blank	0.000	18:16
Air Blank	0.000	18:18
Subject Sample # 2	0.203	18:18
Breath Volume	2.109	18:18
Air Blank	0.000	18:19
Control Test	0.076	18:19
Air Blank	0.000	18:20
Diagnostics Check	OK	18:20

Data Download Date/Time: 05/29/2018 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUSA,KYLE,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PLUSA,KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 04/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: BUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: HARBIN

DATE/TIME OF TEST: 04/29/2018 17:58:51

Diagnostics Check	OK	17:55	
Air Blank	0.000	17:56	
Control Test	0.080	17:56	
Air Blank	0.000	17:57	
Subject Sample	# 1	17:57	Subject Test Refused
Breath Volume	0.000	17:57	
Air Blank	0.000	17:57	
Control Test	0.079	17:58	
Air Blank	0.000	17:58	
Diagnostics Check	OK	17:58	

Data Download Date/Time: 05/11/2018 12:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 04/14/2018 04:18:46

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.079	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.146	04:14
Breath Volume	3.226	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.152	04:17
Breath Volume	2.921	04:17
Air Blank	0.000	04:17
Control Test	0.078	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 04/27/2018 12:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 04/19/2018 01:39:00

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.000	01:34
Breath Volume	2.984	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.000	01:37
Breath Volume	3.976	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:39

Data Download Date/Time: 04/27/2018 12:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEINER,RYAN,T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: BEINER,RYAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: SWETMAN

DATE/TIME OF TEST: 04/19/2018 15:30:16

Diagnostics Check	OK	15:14	
Air Blank	0.000	15:14	
Control Test	0.078	15:15	
Air Blank	0.000	15:15	
Subject Sample	# 1 0.282	15:17	
Breath Volume	1.707	15:17	
Air Blank	0.000	15:18	
Air Blank	0.000	15:19	
Subject Sample	# 2 0.230	15:22	Volume Not Met
Breath Volume	0.148	15:22	
Air Blank	0.000	15:23	
Air Blank	0.000	15:25	
Subject Sample	# 3 0.285	15:28	
Breath Volume	1.417	15:28	
Air Blank	0.000	15:29	
Control Test	0.076	15:29	
Air Blank	0.000	15:30	
Diagnostics Check	OK	15:30	

Data Download Date/Time: 04/27/2018 12:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 04/20/2018 19:02:04

Diagnostics Check	OK	18:55
Air Blank	0.000	18:56
Control Test	0.079	18:56
Air Blank	0.000	18:57
Subject Sample # 1	0.000	18:57
Breath Volume	2.554	18:57
Air Blank	0.000	18:58
Air Blank	0.000	19:00
Subject Sample # 2	0.000	19:00
Breath Volume	2.789	19:00
Air Blank	0.000	19:01
Control Test	0.078	19:01
Air Blank	0.000	19:01
Diagnostics Check	OK	19:02

Data Download Date/Time: 04/27/2018 12:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Lee Port Authority
ARRESTING OFFICER: FREEMAN,MONICA
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: DEXTER

DATE/TIME OF TEST: 04/21/2018 01:13:20

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.078	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.093	01:07
Breath Volume	1.421	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.101	01:11
Breath Volume	2.625	01:11
Air Blank	0.000	01:12
Control Test	0.077	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 04/27/2018 12:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALCON,IVAN,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FALCON,IVAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: DENNY

DATE/TIME OF TEST: 04/21/2018 23:47:02

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.078	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.248	23:42
Breath Volume	2.859	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:44
Subject Sample # 2	0.250	23:45
Breath Volume	3.730	23:45
Air Blank	0.000	23:46
Control Test	0.077	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:47

Data Download Date/Time: 04/27/2018 12:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALCON,IVAN,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: RYAN,MATHEW
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MULVIHILL

DATE/TIME OF TEST: 04/22/2018 00:46:43

Diagnostics Check	OK	00:40
Air Blank	0.000	00:41
Control Test	0.078	00:41
Air Blank	0.000	00:42
Subject Sample # 1	0.101	00:42
Breath Volume	1.871	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:44
Subject Sample # 2	0.097	00:45
Breath Volume	2.273	00:45
Air Blank	0.000	00:45
Control Test	0.077	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 04/27/2018 12:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,A,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BOVILL,R
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:35

SUBJECT'S LAST NAME: PARK

DATE/TIME OF TEST: 04/22/2018 14:02:08

Diagnostics Check	OK	13:55
Air Blank	0.000	13:56
Control Test	0.079	13:56
Air Blank	0.000	13:57
Subject Sample # 1	0.152	13:57
Breath Volume	3.675	13:57
Air Blank	0.000	13:58
Air Blank	0.000	14:00
Subject Sample # 2	0.149	14:00
Breath Volume	3.589	14:00
Air Blank	0.000	14:01
Control Test	0.078	14:01
Air Blank	0.000	14:02
Diagnostics Check	OK	14:02

Data Download Date/Time: 04/27/2018 12:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAEFER, KENDALL, A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: URENA, ROSEMARY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 04/26/2018 20:49:41

Diagnostics Check	OK	20:43
Air Blank	0.000	20:43
Control Test	0.080	20:44
Air Blank	0.000	20:44
Subject Sample # 1	0.166	20:45
Breath Volume	2.976	20:45
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.163	20:48
Breath Volume	2.832	20:48
Air Blank	0.000	20:48
Control Test	0.079	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 04/27/2018 12:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRANO,ROMAN,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: PAHLKE,MITCH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ANTONIO

DATE/TIME OF TEST: 04/27/2018 23:18:54

Diagnostics Check	OK	23:13	
Air Blank	0.000	23:13	
Control Test	0.079	23:14	
Air Blank	0.000	23:14	
Subject Sample	# 1 0.187	23:17	
Breath Volume	2.136	23:17	
Air Blank		23:18	Purge Fail
Air Blank	0.000	23:18	

Data Download Date/Time: 05/09/2018 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINDHOLM,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: BAEZ-PEREZ

DATE/TIME OF TEST: 04/30/2018 04:22:12

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.282	04:17
Breath Volume	3.207	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.285	04:20
Breath Volume	3.277	04:20
Air Blank	0.000	04:21
Control Test	0.078	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 05/09/2018 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRITY, THOMAS, A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GARRITY, THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: CURTISS JR

DATE/TIME OF TEST: 04/27/2018 21:47:23

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.077	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.143	21:42
Breath Volume	2.152	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.147	21:45
Breath Volume	2.644	21:45
Air Blank	0.000	21:46
Control Test	0.078	21:46
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 05/23/2018 18:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRITY, THOMAS, A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GARRITY, THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CAROUTHERS

DATE/TIME OF TEST: 04/29/2018 03:08:37

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.078	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.147	03:03
Breath Volume	2.039	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.150	03:07
Breath Volume	1.773	03:07
Air Blank	0.000	03:07
Control Test	0.079	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 05/23/2018 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ONDREA,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: FORD,L
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: GERACI

DATE/TIME OF TEST: 04/03/2018 22:34:48

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.079	22:29
Air Blank	0.000	22:29
Subject Sample # 1	0.262	22:30
Breath Volume	1.984	22:30
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.259	22:33
Breath Volume	2.269	22:33
Air Blank	0.000	22:33
Control Test	0.077	22:34
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPkins,JENNIFER,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROULEAU,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:29

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 04/04/2018 10:59:42

Diagnostics Check	OK	10:53
Air Blank	0.000	10:53
Control Test	0.078	10:54
Air Blank	0.000	10:54
Subject Sample # 1	0.161	10:55
Breath Volume	1.890	10:55
Air Blank	0.000	10:55
Air Blank	0.000	10:57
Subject Sample # 2	0.171	10:58
Breath Volume	3.226	10:58
Air Blank	0.000	10:58
Control Test	0.078	10:59
Air Blank	0.000	10:59
Diagnostics Check	OK	10:59

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 04/04/2018 21:44:40

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.078	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.173	21:39
Breath Volume	2.398	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.174	21:43
Breath Volume	1.898	21:43
Air Blank	0.000	21:43
Control Test	0.076	21:44
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: BOMHOLT

DATE/TIME OF TEST: 04/10/2018 00:27:33

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.078	00:22
Air Blank	0.000	00:22
Subject Sample # 1	0.151	00:23
Breath Volume	2.085	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.137	00:26
Breath Volume	1.390	00:26
Air Blank	0.000	00:26
Control Test	0.078	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHLINE,C,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STEVENSON,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: SPAGNUOLO

DATE/TIME OF TEST: 04/10/2018 03:13:34

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.078	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.170	03:09
Breath Volume	2.156	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.170	03:12
Breath Volume	1.417	03:12
Air Blank	0.000	03:12
Control Test	0.078	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATSON,MITCHELL,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: COSTELLO,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: ANDRADE

DATE/TIME OF TEST: 04/10/2018 05:41:35

Diagnostics Check		OK	05:25	
Air Blank		0.000	05:26	
Control Test		0.079	05:26	
Air Blank		0.000	05:27	
Subject Sample	# 1	0.195	05:30	Slope Not Level
Breath Volume		1.203	05:30	
Air Blank		0.000	05:30	
Air Blank		0.000	05:32	
Subject Sample	# 2	0.203	05:34	
Breath Volume		1.312	05:34	
Air Blank		0.000	05:34	
Air Blank		0.000	05:36	
Subject Sample	# 3	0.168	05:39	Volume Not Met
Breath Volume		0.234	05:39	
Air Blank		0.000	05:40	
Control Test		0.078	05:40	
Air Blank		0.000	05:41	
Diagnostics Check		OK	05:41	

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: GOSS

DATE/TIME OF TEST: 04/10/2018 18:32:40

Diagnostics Check	OK	18:26
Air Blank	0.000	18:26
Control Test	0.079	18:26
Air Blank	0.000	18:27
Subject Sample # 1	0.263	18:28
Breath Volume	1.382	18:28
Air Blank	0.000	18:28
Air Blank	0.000	18:30
Subject Sample # 2	0.262	18:31
Breath Volume	1.726	18:31
Air Blank	0.000	18:31
Control Test	0.078	18:32
Air Blank	0.000	18:32
Diagnostics Check	OK	18:32

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEAR,DALE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 04/11/2018 00:43:02

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.079	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.055	00:38
Breath Volume	1.820	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:40
Subject Sample # 2	0.057	00:41
Breath Volume	2.378	00:41
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:43

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: MACDONALD

DATE/TIME OF TEST: 04/14/2018 01:05:29

Diagnostics Check	OK	00:58
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	01:00
Subject Sample # 1	0.000	01:01
Breath Volume	1.480	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.000	01:04
Breath Volume	2.937	01:04
Air Blank	0.000	01:04
Control Test	0.078	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVAS,W,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GROSSKLAS,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: NOVA

DATE/TIME OF TEST: 04/14/2018 02:45:23

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.078	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.089	02:40
Breath Volume	1.628	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:42
Subject Sample # 2	0.090	02:43
Breath Volume	1.101	02:43
Air Blank	0.000	02:44
Control Test	0.077	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOEL,Y,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STEVENSON,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CRESPO

DATE/TIME OF TEST: 04/14/2018 03:38:18

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.078	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.099	03:33
Breath Volume	2.414	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.100	03:36
Breath Volume	2.320	03:36
Air Blank	0.000	03:37
Control Test	0.078	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/14/2018 04:47:55

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.078	04:42
Air Blank	0.000	04:43
Subject Sample # 1	0.099	04:43
Breath Volume	2.207	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:45
Subject Sample # 2	0.102	04:46
Breath Volume	1.917	04:46
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUSA,KYLE,A
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PLUSA,KYLE
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: BUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 04/14/2018 19:49:44

Diagnostics Check	OK	19:46	
Air Blank	0.000	19:47	
Control Test	0.079	19:47	
Air Blank	0.000	19:48	
Subject Sample	# 1	19:48	Subject Test Refused
Breath Volume	0.000	19:48	
Air Blank	0.000	19:48	
Control Test	0.078	19:49	
Air Blank	0.000	19:49	
Diagnostics Check	OK	19:49	

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: EASTERLING

DATE/TIME OF TEST: 04/15/2018 02:38:20

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:25	
Control Test	0.079	02:25	
Air Blank	0.000	02:26	
Subject Sample # 1	0.092	02:27	
Breath Volume	1.332	02:27	
Air Blank	0.000	02:28	
Air Blank	0.000	02:30	
Subject Sample # 2	0.088	02:33	Volume Not Met
Breath Volume	0.183	02:33	
Air Blank	0.000	02:33	
Air Blank	0.000	02:35	
Subject Sample # 3	0.090	02:36	
Breath Volume	1.984	02:36	
Air Blank	0.000	02:37	
Control Test	0.078	02:37	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIANSEN,DAVID,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CHRISTIANSEN,DAVID
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: MEUNIER

DATE/TIME OF TEST: 04/15/2018 03:26:27

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.078	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.122	03:21
Breath Volume	1.875	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.129	03:24
Breath Volume	1.375	03:24
Air Blank	0.000	03:25
Control Test	0.078	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: MAZURE

DATE/TIME OF TEST: 04/15/2018 03:44:46

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.078	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.120	03:40
Breath Volume	3.035	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.125	03:43
Breath Volume	2.773	03:43
Air Blank	0.000	03:43
Control Test	0.078	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,KEVIN,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCCOY,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: TURUK

DATE/TIME OF TEST: 04/16/2018 04:13:06

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.079	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.179	04:08
Breath Volume	1.863	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.160	04:11
Breath Volume	1.699	04:11
Air Blank	0.000	04:12
Control Test	0.078	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIBEY,AMANDA,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DELGADO,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ALEXIS

DATE/TIME OF TEST: 04/17/2018 00:25:12

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:17	
Control Test	0.079	00:17	
Air Blank	0.000	00:18	
Subject Sample	# 1	00:21	No Sample Provided
Breath Volume	0.000	00:21	
Air Blank	0.000	00:21	
Air Blank	0.000	00:23	
Subject Sample	# 2	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:24	
Control Test	0.079	00:24	
Air Blank	0.000	00:25	
Diagnostics Check	OK	00:25	

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: PONGSAWADI

DATE/TIME OF TEST: 04/17/2018 02:02:45

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.141	01:57
Breath Volume	1.332	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.144	02:01
Breath Volume	1.605	02:01
Air Blank	0.000	02:01
Control Test	0.079	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: HOLLER

DATE/TIME OF TEST: 04/17/2018 22:41:08

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.078	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.120	22:36
Breath Volume	2.855	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.120	22:39
Breath Volume	2.304	22:39
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLMSTEAD,J,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CRUZ,I
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: STERLING

DATE/TIME OF TEST: 04/18/2018 01:07:02

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.247	01:02
Breath Volume		2.929	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:04
Subject Sample	# 2	0.232	01:05
Breath Volume		1.957	01:05
Air Blank		0.000	01:06
Control Test		0.078	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:07

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AXELSON,B,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SILL,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/19/2018 02:36:41

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.078	02:31
Air Blank		0.000	02:31
Subject Sample	# 1	0.230	02:32
Breath Volume		2.210	02:32
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.230	02:35
Breath Volume		2.386	02:35
Air Blank		0.000	02:35
Control Test		0.078	02:36
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULRY, THOMAS, B
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MULRY, THOMAS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: SOMMA

DATE/TIME OF TEST: 04/19/2018 19:20:28

Diagnostics Check	OK	19:14
Air Blank	0.000	19:14
Control Test	0.078	19:14
Air Blank	0.000	19:15
Subject Sample # 1	0.207	19:15
Breath Volume	2.046	19:15
Air Blank	0.000	19:16
Air Blank	0.000	19:18
Subject Sample # 2	0.210	19:18
Breath Volume	3.812	19:18
Air Blank	0.000	19:19
Control Test	0.078	19:19
Air Blank	0.000	19:20
Diagnostics Check	OK	19:20

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: KUBICSEK

DATE/TIME OF TEST: 04/21/2018 00:33:52

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.079	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.176	00:29
Breath Volume	1.574	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:31
Subject Sample # 2	0.190	00:32
Breath Volume	1.171	00:32
Air Blank	0.000	00:32
Control Test	0.078	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALCON,IVAN,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FALCON,IVAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: DOMIGUEZ MARTINEZ

DATE/TIME OF TEST: 04/21/2018 02:35:54

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.078	02:33	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:34	RFI Detect
Breath Volume	0.886	02:34	
Air Blank	0.000	02:34	
Control Test	0.079	02:35	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALCON,IVAN,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FALCON,IVAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DOMINGUEZ MARTINEZ

DATE/TIME OF TEST: 04/21/2018 02:45:11

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.078	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.209	02:40
Breath Volume	2.167	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.212	02:43
Breath Volume	3.171	02:43
Air Blank	0.000	02:44
Control Test	0.076	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: GAULT

DATE/TIME OF TEST: 04/21/2018 02:56:52

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.078	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.155	02:52
Breath Volume	3.472	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.154	02:55
Breath Volume	3.507	02:55
Air Blank	0.000	02:55
Control Test	0.078	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,REID,B
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: YUREWITCH,KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: BUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: KLINE

DATE/TIME OF TEST: 04/21/2018 20:00:18

Diagnostics Check	OK	19:59	
Air Blank		19:59	RFI Detect
Air Blank	0.000	20:00	

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,REID,B
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: YUREWITCH,KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: BUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: KLINE

DATE/TIME OF TEST: 04/21/2018 20:04:15

Diagnostics Check	OK	20:01	
Air Blank	0.000	20:01	
Control Test	0.078	20:02	
Air Blank	0.000	20:02	
Subject Sample	# 1	20:02	RFI Detect
Breath Volume	0.000	20:02	
Air Blank	0.000	20:03	
Control Test	0.078	20:03	
Air Blank	0.000	20:04	
Diagnostics Check	OK	20:04	

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,REID,B
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: YUREWITCH,KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: BUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: KLINE

DATE/TIME OF TEST: 04/21/2018 20:12:33

Diagnostics Check	OK	20:05
Air Blank	0.000	20:06
Control Test	0.079	20:06
Air Blank	0.000	20:07
Subject Sample # 1	0.142	20:08
Breath Volume	1.867	20:08
Air Blank	0.000	20:08
Air Blank	0.000	20:10
Subject Sample # 2	0.135	20:11
Breath Volume	1.300	20:11
Air Blank	0.000	20:11
Control Test	0.078	20:11
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERMAN,BRENT,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GROENDYK,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: SKIVER

DATE/TIME OF TEST: 04/21/2018 23:56:13

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.078	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.194	23:51
Breath Volume	2.171	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.185	23:54
Breath Volume	1.394	23:54
Air Blank	0.000	23:55
Control Test	0.079	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: FLUDGATE

DATE/TIME OF TEST: 04/22/2018 03:46:46

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.078	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.114	03:42
Breath Volume	2.917	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.117	03:45
Breath Volume	3.343	03:45
Air Blank	0.000	03:45
Control Test	0.079	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,T,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GROENDYK,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 04/22/2018 20:08:07

Diagnostics Check	OK	20:02
Air Blank	0.000	20:02
Control Test	0.078	20:02
Air Blank	0.000	20:03
Subject Sample # 1	0.000	20:03
Breath Volume	2.800	20:03
Air Blank	0.000	20:04
Air Blank	0.000	20:06
Subject Sample # 2	0.000	20:06
Breath Volume	3.261	20:06
Air Blank	0.000	20:07
Control Test	0.078	20:07
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JOSE,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LOPEZ,JOSE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/23/2018 01:40:55

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.078	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.164	01:35
Breath Volume	1.539	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.169	01:39
Breath Volume	1.488	01:39
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,Y,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DESIMONE,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/23/2018 22:36:46

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.078	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.168	22:32
Breath Volume	2.593	22:32
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.170	22:35
Breath Volume	2.703	22:35
Air Blank	0.000	22:35
Control Test	0.078	22:36
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LIPARI,M
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: SALDANA

DATE/TIME OF TEST: 04/24/2018 20:29:59

Diagnostics Check	OK	20:23
Air Blank	0.000	20:24
Control Test	0.079	20:24
Air Blank	0.000	20:25
Subject Sample # 1	0.129	20:25
Breath Volume	2.050	20:25
Air Blank	0.000	20:26
Air Blank	0.000	20:28
Subject Sample # 2	0.123	20:28
Breath Volume	1.925	20:28
Air Blank	0.000	20:29
Control Test	0.076	20:29
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,COLTON,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SOLIS,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: HARLOW

DATE/TIME OF TEST: 04/24/2018 22:07:51

Diagnostics Check	OK	22:01
Air Blank	0.000	22:02
Control Test	0.078	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.246	22:03
Breath Volume	1.582	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:05
Subject Sample # 2	0.233	22:06
Breath Volume	1.402	22:06
Air Blank	0.000	22:06
Control Test	0.078	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,N
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STEVENSON,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: MISOVIC

DATE/TIME OF TEST: 04/25/2018 00:51:35

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.078	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.128	00:47
Breath Volume	3.378	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.125	00:50
Breath Volume	3.460	00:50
Air Blank	0.000	00:50
Control Test	0.078	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERTILUS,FREDELYNE,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LEONI,STEPHEN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: GOODHEW

DATE/TIME OF TEST: 04/26/2018 05:41:27

Diagnostics Check	OK	05:40	
Air Blank	0.000	05:40	
Control Test	0.074	05:40	Control Outside Tolerance
Air Blank	0.000	05:41	

Data Download Date/Time: 05/01/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,N
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GROSSLAS,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: CAMERO

DATE/TIME OF TEST: 04/01/2018 04:27:42

Diagnostics Check		OK	04:13	
Air Blank		0.000	04:14	
Control Test		0.078	04:14	
Air Blank		0.000	04:15	
Subject Sample	# 1	0.146	04:18	Volume Not Met
Breath Volume		0.000	04:18	
Air Blank		0.000	04:19	
Air Blank		0.000	04:20	
Subject Sample	# 2	0.151	04:22	
Breath Volume		2.578	04:22	
Air Blank		0.000	04:23	
Air Blank		0.000	04:24	
Subject Sample	# 3	0.166	04:26	
Breath Volume		1.765	04:26	
Air Blank		0.000	04:26	
Control Test		0.077	04:27	
Air Blank		0.000	04:27	
Diagnostics Check		OK	04:27	

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREISHEIMER,JOHN,V
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GREISHEIMER,JOHN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: ZIELINSKI

DATE/TIME OF TEST: 04/01/2018 05:14:32

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.077	05:08
Air Blank	0.000	05:09
Subject Sample # 1	0.181	05:09
Breath Volume	3.839	05:09
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.187	05:13
Breath Volume	3.414	05:13
Air Blank	0.000	05:13
Control Test	0.078	05:13
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 04/02/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERN,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 04/30/2018 12:12:39

Diagnostics Check	OK	12:07
Air Blank	0.000	12:07
Control Test	0.081	12:07
Air Blank	0.000	12:08
Subject Sample # 1	0.000	12:08
Breath Volume	2.289	12:08
Air Blank	0.000	12:08
Air Blank	0.000	12:11
Subject Sample # 2	0.000	12:11
Breath Volume	2.550	12:11
Air Blank	0.000	12:11
Control Test	0.081	12:12
Air Blank	0.000	12:12
Diagnostics Check	OK	12:12

Data Download Date/Time: 05/28/2018 00:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLATZ,KEVIN,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ROOT,DUANE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: VILMORNE

DATE/TIME OF TEST: 04/30/2018 21:55:31

Diagnostics Check	OK	21:48
Air Blank	0.000	21:48
Control Test	0.081	21:48
Air Blank	0.000	21:49
Subject Sample # 1	0.062	21:50
Breath Volume	2.062	21:50
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.062	21:54
Breath Volume	1.953	21:54
Air Blank	0.000	21:54
Control Test	0.081	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 05/28/2018 00:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALLADINO,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: MORGAN,D/S
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 03/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BILLIE

DATE/TIME OF TEST: 04/06/2018 00:36:07

Diagnostics Check	OK	00:29
Air Blank	0.000	00:29
Control Test	0.081	00:29
Air Blank	0.000	00:30
Subject Sample # 1	0.051	00:31
Breath Volume	3.234	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.049	00:34
Breath Volume	2.519	00:34
Air Blank	0.000	00:35
Control Test	0.078	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 06/11/2018 18:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSUE,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CLENDENING,JAMES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 03/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ALMANZA

DATE/TIME OF TEST: 04/16/2018 23:31:00

Diagnostics Check	OK	23:19	
Air Blank	0.000	23:19	
Control Test	0.080	23:19	
Air Blank	0.000	23:20	
Subject Sample	# 1	23:23	No Sample Provided
Breath Volume	0.000	23:23	
Air Blank	0.000	23:24	
Air Blank	0.000	23:26	
Subject Sample	# 2	0.214	23:26
Breath Volume	1.417	23:26	
Air Blank	0.000	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 3	0.212	23:29
Breath Volume	1.472	23:29	
Air Blank	0.000	23:30	
Control Test	0.078	23:30	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:31	

Data Download Date/Time: 06/11/2018 18:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SILVA, JUAN
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 03/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 04/21/2018 21:16:51

Diagnostics Check		OK	21:08
Air Blank		0.000	21:08
Control Test		0.080	21:08
Air Blank		0.000	21:09
Subject Sample	# 1	0.220	21:11
Breath Volume		1.269	21:11
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 2	0.217	21:15
Breath Volume		2.128	21:15
Air Blank		0.000	21:15
Control Test		0.078	21:16
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

Data Download Date/Time: 06/11/2018 18:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUVERSBURK,BRENNA,M
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: HAUVERSBURK,BRENNA
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 04/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 04/26/2018 23:34:57

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.079	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.262	23:30
Breath Volume	3.308	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:32
Subject Sample # 2	0.254	23:33
Breath Volume	2.679	23:33
Air Blank	0.000	23:34
Control Test	0.078	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 05/10/2018 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CORONILLA,DANIEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 04/06/2018 02:38:55

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.081	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:37	Subject Test Refused
Breath Volume	0.000	02:37	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 04/29/2018 14:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAMBERLAIN,STEVEN,M
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: TAYLOR,ANDREW
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 04/08/2018 02:44:40

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.081	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.095	02:40
Breath Volume	1.242	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.104	02:43
Breath Volume	1.785	02:43
Air Blank	0.000	02:43
Control Test	0.081	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 04/29/2018 14:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARROW,JESHUA,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARROW,JESHUA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: GOLKO

DATE/TIME OF TEST: 04/07/2018 21:48:54

Diagnostics Check	OK	21:47	
Air Blank	0.000	21:47	
Control Test	0.080	21:48	
Air Blank		21:48	RFI Detect
Air Blank	0.000	21:48	

Data Download Date/Time: 04/09/2018 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARROW,JESHUA,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARROW,JESHUA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: GOLKO

DATE/TIME OF TEST: 04/07/2018 23:24:31

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:22	
Control Test	0.081	23:22	
Air Blank	0.000	23:23	
Subject Sample	# 1	23:23	Subject Test Refused
Breath Volume	0.000	23:23	
Air Blank	0.000	23:23	
Control Test	0.081	23:23	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:24	

Data Download Date/Time: 04/09/2018 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAUFFMAN, MITCHELL, P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SCHILLING, JOHN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: BURTON

DATE/TIME OF TEST: 04/26/2018 22:52:31

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:50	
Control Test	0.081	22:50	
Air Blank	0.000	22:50	
Subject Sample	# 1	22:51	Subject Test Refused
Breath Volume	0.000	22:51	
Air Blank	0.000	22:51	
Control Test	0.080	22:51	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 05/08/2018 10:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPANGLER,GARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SPANGLER,GARY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 04/27/2018 06:54:21

Diagnostics Check	OK	06:48
Air Blank	0.000	06:48
Control Test	0.081	06:49
Air Blank	0.000	06:49
Subject Sample # 1	0.237	06:49
Breath Volume	2.015	06:49
Air Blank	0.000	06:50
Air Blank	0.000	06:52
Subject Sample # 2	0.241	06:52
Breath Volume	2.300	06:52
Air Blank	0.000	06:53
Control Test	0.080	06:53
Air Blank	0.000	06:54
Diagnostics Check	OK	06:54

Data Download Date/Time: 05/08/2018 10:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,VICKI,N
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: SHEFFIELD,BRANDON
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 04/28/2018 23:11:10

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:08	
Control Test	0.080	23:08	
Air Blank	0.000	23:09	
Subject Sample	# 1	23:09	Subject Test Refused
Breath Volume	0.000	23:09	
Air Blank	0.000	23:10	
Control Test	0.081	23:10	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 05/08/2018 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENING,FREDRICK,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HENING,FREDRICK
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: OCHAT

DATE/TIME OF TEST: 04/30/2018 18:31:01

Diagnostics Check	OK	18:20
Air Blank	0.000	18:21
Control Test	0.081	18:21
Air Blank	0.000	18:21
Subject Sample # 1	0.272	18:23
Breath Volume	3.773	18:23
Air Blank	0.000	18:24
Air Blank	0.000	18:25
Subject Sample # 2	0.235	18:26
Breath Volume	1.386	18:26
Air Blank	0.000	18:27
Air Blank	0.000	18:28
Subject Sample # 3	0.240	18:29
Breath Volume	1.679	18:29
Air Blank	0.000	18:30
Control Test	0.080	18:30
Air Blank	0.000	18:30
Diagnostics Check	OK	18:31

Data Download Date/Time: 05/08/2018 10:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHRITENOUR,STEVEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: F,F
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: F

DATE/TIME OF TEST: 04/07/2018 20:07:39

Diagnostics Check	OK	20:04	
Air Blank	0.000	20:05	
Control Test	0.080	20:05	
Air Blank	0.000	20:06	
Subject Sample	# 1	20:06	Subject Test Refused
Breath Volume	0.000	20:06	
Air Blank	0.000	20:06	
Control Test	0.081	20:07	
Air Blank	0.000	20:07	
Diagnostics Check	OK	20:07	

Data Download Date/Time: 04/24/2018 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: MONROE

DATE/TIME OF TEST: 04/07/2018 20:16:13

Diagnostics Check		OK	20:10
Air Blank		0.000	20:10
Control Test		0.080	20:10
Air Blank		0.000	20:11
Subject Sample	# 1	0.237	20:11
Breath Volume		1.644	20:11
Air Blank		0.000	20:12
Air Blank		0.000	20:14
Subject Sample	# 2	0.240	20:14
Breath Volume		1.734	20:14
Air Blank		0.000	20:15
Control Test		0.081	20:15
Air Blank		0.000	20:16
Diagnostics Check		OK	20:16

Data Download Date/Time: 04/24/2018 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TALLMAN,JOHN,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: TALLMAN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:36

SUBJECT'S LAST NAME: FRANKS

DATE/TIME OF TEST: 04/16/2018 12:02:33

Diagnostics Check	OK	11:56
Air Blank	0.000	11:56
Control Test	0.080	11:57
Air Blank	0.000	11:57
Subject Sample # 1	0.302	11:58
Breath Volume	2.300	11:58
Air Blank	0.000	11:58
Air Blank	0.000	12:00
Subject Sample # 2	0.304	12:00
Breath Volume	2.601	12:00
Air Blank	0.000	12:01
Control Test	0.081	12:01
Air Blank	0.000	12:02
Diagnostics Check	OK	12:02

Data Download Date/Time: 04/24/2018 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBER,HAROLD,J
BREATH TEST OPERATOR AGENCY: Quincy PD
ARRESTING OFFICER: FLINT,KILIAN
ARRESTING OFFICER AGENCY: Quincy PD
LAST AGENCY INSPECTION: 03/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 04/20/2018 23:48:22

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:45	
Control Test	0.080	23:46	
Air Blank	0.000	23:46	
Subject Sample	# 1	23:47	Subject Test Refused
Breath Volume	0.000	23:47	
Air Blank	0.000	23:47	
Control Test	0.081	23:47	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 04/24/2018 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAFFORD,STERLING,
BREATH TEST OPERATOR AGENCY: Gadsden County SO
ARRESTING OFFICER: SAFFORD,STERLING
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: EVERARDO

DATE/TIME OF TEST: 04/28/2018 00:53:59

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:51	
Control Test	0.080	00:51	
Air Blank	0.000	00:52	
Subject Sample	# 1	00:52	Subject Test Refused
Breath Volume	0.000	00:52	
Air Blank	0.000	00:53	
Control Test	0.080	00:53	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

Data Download Date/Time: 05/09/2018 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS, MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 04/01/2018 21:45:28

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:40
Subject Sample # 1	0.158	21:40
Breath Volume	1.492	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.161	21:43
Breath Volume	1.566	21:43
Air Blank	0.000	21:44
Control Test	0.080	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 04/17/2018 18:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: DOUVRES, MICHAEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 04/01/2018 23:25:00

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.080	23:20
Air Blank	0.000	23:20
Subject Sample # 1	0.000	23:20
Breath Volume	3.406	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.000	23:23
Breath Volume	2.906	23:23
Air Blank	0.000	23:24
Control Test	0.079	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:25

Data Download Date/Time: 04/17/2018 18:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: WEISS

DATE/TIME OF TEST: 04/21/2018 23:48:43

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.080	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.107	23:43
Breath Volume	1.554	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.107	23:47
Breath Volume	1.824	23:47
Air Blank	0.000	23:47
Control Test	0.080	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 05/03/2018 19:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MARTINEZ,YOMARI
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: RAMIREZ-RIVERA

DATE/TIME OF TEST: 04/22/2018 04:53:54

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.081	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.106	04:49
Breath Volume	3.386	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:51
Subject Sample # 2	0.106	04:52
Breath Volume	2.937	04:52
Air Blank	0.000	04:52
Control Test	0.079	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 05/03/2018 19:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVA,EMILIANO,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SILVA,EMILIANO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 04/27/2018 03:19:33

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:16	
Control Test	0.080	03:17	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:18	RFI Detect
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Control Test	0.081	03:18	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 05/03/2018 19:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVA,EMILIANO,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SILVA,EMILIANO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 04/27/2018 03:23:44

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.080	03:23	RFI Detect
Air Blank	0.000	03:23	

Data Download Date/Time: 05/03/2018 19:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIARRATANA,JOHN,R
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: MAYORGA,OSCAR
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: VILLANUEVA

DATE/TIME OF TEST: 04/01/2018 22:04:51

Diagnostics Check		OK	21:58
Air Blank		0.000	21:59
Control Test		0.076	21:59
Air Blank		0.000	21:59
Subject Sample	# 1	0.251	22:00
Breath Volume		2.457	22:00
Air Blank		0.000	22:01
Air Blank		0.000	22:02
Subject Sample	# 2	0.242	22:03
Breath Volume		1.988	22:03
Air Blank		0.000	22:03
Control Test		0.077	22:04
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 04/27/2018 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIARRATANA,JOHN,R
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: GIARRATANA,JOHN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: CHARNECO

DATE/TIME OF TEST: 04/15/2018 01:14:22

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.078	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:12	Interferent Detect
Breath Volume	1.375	01:12	
Air Blank	0.000	01:13	
Control Test	0.078	01:13	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 04/27/2018 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIARRATANA,JOHN,R
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: GIARRATANA,JOHN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: CHARNECO

DATE/TIME OF TEST: 04/15/2018 01:19:26

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:16	
Control Test	0.078	01:16	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:17	Interferent Detect
Breath Volume	2.582	01:17	
Air Blank	0.000	01:18	
Control Test	0.078	01:18	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 04/27/2018 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIARRATANA,JOHN,R
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: GIARRATANA,JOHN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 03/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: CHARNECO

DATE/TIME OF TEST: 04/15/2018 01:49:24

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:46	
Control Test	0.078	01:46	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:47	Interferent Detect
Breath Volume	2.351	01:47	
Air Blank	0.000	01:48	
Control Test	0.078	01:48	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 04/27/2018 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIARRATANA,JOHN,R
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: TUCK,NATHANIEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: CULLIVAN

DATE/TIME OF TEST: 04/27/2018 23:04:27

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:03	
Control Test	0.071	23:04	Control Outside Tolerance
Air Blank		23:04	RFI Detect

Data Download Date/Time: 05/29/2018 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIARRATANA,JOHN,R
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: TUCK,NATHANIEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: CULLIVAN

DATE/TIME OF TEST: 04/27/2018 23:17:12

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.076	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.183	23:12
Breath Volume	2.214	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:14
Subject Sample # 2	0.177	23:15
Breath Volume	2.328	23:15
Air Blank	0.000	23:16
Control Test	0.076	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 05/29/2018 16:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORUJO,PEDRO,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: CORUJO,PEDRO
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: BADILLO BADILLO

DATE/TIME OF TEST: 04/01/2018 03:07:12

Diagnostics Check	OK	03:04	
Air Blank	0.000	03:04	
Control Test	0.079	03:05	
Air Blank	0.000	03:05	
Subject Sample	# 1	03:05	Subject Test Refused
Breath Volume	0.000	03:05	
Air Blank	0.000	03:06	
Control Test	0.078	03:06	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 04/17/2018 17:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: CANDELARIA

DATE/TIME OF TEST: 04/01/2018 03:53:29

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.078	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.136	03:48
Breath Volume	1.570	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.127	03:51
Breath Volume	1.488	03:51
Air Blank	0.000	03:52
Control Test	0.078	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 04/17/2018 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 04/05/2018 03:46:16

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.078	03:38	
Air Blank	0.000	03:38	
Subject Sample	# 1 0.120	03:41	Volume Not Met
Breath Volume	0.000	03:41	
Air Blank	0.000	03:42	
Air Blank	0.000	03:44	
Subject Sample	# 2	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:45	
Control Test	0.077	03:45	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 04/17/2018 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: COTTO GONZALEZ

DATE/TIME OF TEST: 04/07/2018 03:37:26

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.110	03:32
Breath Volume	2.140	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.107	03:35
Breath Volume	1.921	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 04/17/2018 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,CHARLES,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCONNELL,CHARLES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:42

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 04/07/2018 06:13:16

Diagnostics Check	OK	06:03
Air Blank	0.000	06:04
Control Test	0.079	06:04
Air Blank	0.000	06:05
Subject Sample # 1	0.235	06:08
Breath Volume	1.984	06:08
Air Blank	0.000	06:08
Air Blank	0.000	06:10
Subject Sample # 2	0.234	06:11
Breath Volume	2.191	06:11
Air Blank	0.000	06:12
Control Test	0.077	06:12
Air Blank	0.000	06:13
Diagnostics Check	OK	06:13

Data Download Date/Time: 04/17/2018 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORUJO,PEDRO,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: CORUJO,PEDRO
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: ROSARIO GONZALEZ

DATE/TIME OF TEST: 04/10/2018 22:06:44

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.078	22:00
Air Blank	0.000	22:00
Subject Sample # 1	0.111	22:02
Breath Volume	4.156	22:02
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.113	22:05
Breath Volume	4.164	22:05
Air Blank	0.000	22:05
Control Test	0.079	22:06
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 04/17/2018 17:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,CHARLES,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCONNELL,CHARLES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: LASH

DATE/TIME OF TEST: 04/13/2018 04:24:38

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.079	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.007	04:20
Breath Volume	2.496	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.000	04:23
Breath Volume	3.480	04:23
Air Blank	0.000	04:23
Control Test	0.078	04:24
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 04/17/2018 17:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: RODRIGUEZ, ALBERT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 04/14/2018 19:49:44

Diagnostics Check	OK	19:40
Air Blank	0.000	19:41
Control Test	0.079	19:41
Air Blank	0.000	19:42
Subject Sample # 1	0.324	19:42
Breath Volume	1.882	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:44
Subject Sample # 2	0.301	19:45
Breath Volume	2.167	19:45
Air Blank	0.000	19:46
Air Blank	0.000	19:47
Subject Sample # 3	0.306	19:48
Breath Volume	2.160	19:48
Air Blank	0.000	19:48
Control Test	0.078	19:49
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

Data Download Date/Time: 04/17/2018 17:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RODRIGUEZ,ALBERT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: LOCHBAUM

DATE/TIME OF TEST: 04/15/2018 01:07:28

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:06	
Control Test	0.079	01:06	
Air Blank	0.000	01:06	
		01:07	RFI Detect
Air Blank	0.000	01:07	

Data Download Date/Time: 04/17/2018 17:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RODRIGUEZ,ALBERT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: LOCHBAUM

DATE/TIME OF TEST: 04/15/2018 01:19:07

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.031	01:14
Breath Volume	1.960	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.030	01:17
Breath Volume	2.410	01:17
Air Blank	0.000	01:18
Control Test	0.078	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 04/17/2018 17:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS, MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: PESANTEZ

DATE/TIME OF TEST: 04/15/2018 03:34:33

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.079	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.256	03:28
Breath Volume	1.851	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.241	03:32
Breath Volume	1.742	03:32
Air Blank	0.000	03:33
Control Test	0.077	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 04/17/2018 17:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:32

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 04/16/2018 06:00:56

Diagnostics Check	OK	05:53
Air Blank	0.000	05:53
Control Test	0.079	05:54
Air Blank	0.000	05:54
Subject Sample # 1	0.152	05:56
Breath Volume	1.511	05:56
Air Blank	0.000	05:57
Air Blank	0.000	05:58
Subject Sample # 2	0.139	05:59
Breath Volume	1.320	05:59
Air Blank	0.000	06:00
Control Test	0.078	06:00
Air Blank	0.000	06:00
Diagnostics Check	OK	06:00

Data Download Date/Time: 04/17/2018 17:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEVES, CHARLIE,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: NIEVES, CHARLIE
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 03/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MCCALLA

DATE/TIME OF TEST: 04/17/2018 04:16:41

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.115	04:11
Breath Volume	2.453	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.111	04:15
Breath Volume	2.425	04:15
Air Blank	0.000	04:15
Control Test	0.078	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 04/17/2018 17:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: DAUGHERTY, KENYATTA
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: BAERT

DATE/TIME OF TEST: 04/24/2018 18:14:27

Diagnostics Check	OK	18:08
Air Blank	0.000	18:08
Control Test	0.079	18:09
Air Blank	0.000	18:09
Subject Sample # 1	0.304	18:09
Breath Volume	2.914	18:09
Air Blank	0.000	18:10
Air Blank	0.000	18:12
Subject Sample # 2	0.286	18:12
Breath Volume	3.210	18:12
Air Blank	0.000	18:13
Control Test	0.078	18:13
Air Blank	0.000	18:14
Diagnostics Check	OK	18:14

Data Download Date/Time: 05/03/2018 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORUJO,PEDRO,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: BINDER,JARED
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: SPELLS

DATE/TIME OF TEST: 04/24/2018 21:01:32

Diagnostics Check	OK	20:56	
Air Blank	0.000	20:56	
Control Test	0.078	20:57	
Air Blank	0.000	20:57	
Subject Sample	# 1	21:00	Subject Test Refused
Breath Volume	0.000	21:00	
Air Blank	0.000	21:00	
Control Test	0.079	21:00	
Air Blank	0.000	21:01	
Diagnostics Check	OK	21:01	

Data Download Date/Time: 05/03/2018 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVA, EMILIANO, R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SILVA, EMILIANO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 04/27/2018 03:33:36

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.179	03:28
Breath Volume	2.019	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.180	03:31
Breath Volume	2.453	03:31
Air Blank	0.000	03:32
Control Test	0.078	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 05/03/2018 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: SCHEINER,JAMES
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: MCMAHON

DATE/TIME OF TEST: 04/14/2018 18:58:15

Diagnostics Check	OK	18:51
Air Blank	0.000	18:51
Control Test	0.079	18:52
Air Blank	0.000	18:52
Subject Sample # 1	0.190	18:53
Breath Volume	2.664	18:53
Air Blank	0.000	18:54
Air Blank	0.000	18:56
Subject Sample # 2	0.176	18:56
Breath Volume	1.429	18:56
Air Blank	0.000	18:57
Control Test	0.078	18:57
Air Blank	0.000	18:58
Diagnostics Check	OK	18:58

Data Download Date/Time: 05/27/2018 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: SCHEINER,JAMES
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 04/16/2018 22:29:31

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:16	
Control Test	0.078	22:16	
Air Blank	0.000	22:16	
Subject Sample	# 1 0.274	22:17	
Breath Volume	1.164	22:17	
Air Blank	0.000	22:18	
Air Blank	0.000	22:20	
Subject Sample	# 2 0.252	22:23	Volume Not Met
Breath Volume	0.050	22:23	
Air Blank	0.000	22:24	
Air Blank	0.000	22:25	
Subject Sample	# 3	22:27	Subject Test Refused
Breath Volume	0.242	22:27	
Air Blank	0.000	22:28	
Control Test	0.077	22:28	
Air Blank	0.000	22:29	
Diagnostics Check	OK	22:29	

Data Download Date/Time: 05/27/2018 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEINER,JAMES,R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: FLAMMIA,A
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BOSS

DATE/TIME OF TEST: 04/17/2018 02:17:26

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.151	02:12
Breath Volume	2.378	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.148	02:16
Breath Volume	1.226	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 05/27/2018 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUSLER,ETHAN,D
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: PAYNE,TAYLOR
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 04/21/2018 01:42:58

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.158	01:38
Breath Volume	1.914	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.163	01:41
Breath Volume	1.804	01:41
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 05/27/2018 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MARTINS,ALEX
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 04/04/2018 16:57:02

Diagnostics Check	OK	16:50
Air Blank	0.000	16:51
Control Test	0.081	16:51
Air Blank	0.000	16:52
Subject Sample # 1	0.000	16:52
Breath Volume	2.527	16:52
Air Blank	0.000	16:53
Air Blank	0.000	16:55
Subject Sample # 2	0.000	16:55
Breath Volume	2.296	16:55
Air Blank	0.000	16:56
Control Test	0.081	16:56
Air Blank	0.000	16:56
Diagnostics Check	OK	16:57

Data Download Date/Time: 04/30/2018 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,ROBERT,A
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: HUMPHREY,MATTHEW
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: FLACK II

DATE/TIME OF TEST: 04/05/2018 08:31:07

Diagnostics Check	OK	08:24
Air Blank	0.000	08:25
Control Test	0.080	08:25
Air Blank	0.000	08:26
Subject Sample # 1	0.009	08:26
Breath Volume	2.761	08:26
Air Blank	0.000	08:27
Air Blank	0.000	08:29
Subject Sample # 2	0.007	08:29
Breath Volume	2.675	08:29
Air Blank	0.000	08:30
Control Test	0.082	08:30
Air Blank	0.000	08:31
Diagnostics Check	OK	08:31

Data Download Date/Time: 04/30/2018 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOLE,GEOFFERY,E
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MCDOLE,GEOFFERY
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: VANSYCKLE

DATE/TIME OF TEST: 04/07/2018 02:32:18

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.081	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.025	02:27
Breath Volume	1.746	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.026	02:30
Breath Volume	1.718	02:30
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 04/30/2018 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MCGRORY,DIARMUID
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: BREWER

DATE/TIME OF TEST: 04/10/2018 17:27:48

Diagnostics Check	OK	17:20
Air Blank	0.000	17:21
Control Test	0.081	17:21
Air Blank	0.000	17:22
Subject Sample # 1	0.343	17:22
Breath Volume	2.242	17:22
Air Blank	0.000	17:23
Air Blank	0.000	17:25
Subject Sample # 2	0.348	17:25
Breath Volume	2.191	17:25
Air Blank	0.000	17:26
Control Test	0.080	17:27
Air Blank	0.000	17:27
Diagnostics Check	OK	17:27

Data Download Date/Time: 04/30/2018 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURDEN,SHAWNTRALE,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: DURDEN,SHAWNTRALE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: CHUBIOGLU

DATE/TIME OF TEST: 04/15/2018 20:57:30

Diagnostics Check	OK	20:51
Air Blank	0.000	20:51
Control Test	0.080	20:51
Air Blank	0.000	20:52
Subject Sample # 1	0.000	20:53
Breath Volume	1.968	20:53
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.000	20:56
Breath Volume	2.585	20:56
Air Blank	0.000	20:56
Control Test	0.081	20:56
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 04/30/2018 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: CABRANES,JORGE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: SLAVENS

DATE/TIME OF TEST: 04/28/2018 02:06:50

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:04	
Control Test	0.080	02:04	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:05	Subject Test Refused
Breath Volume	0.000	02:05	
Air Blank	0.000	02:05	
Control Test	0.080	02:06	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

Data Download Date/Time: 04/30/2018 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MCGRORY,DIARMUID
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: ROWLAND

DATE/TIME OF TEST: 04/29/2018 04:35:58

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.081	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1 0.234	04:34	
Breath Volume	1.867	04:34	
Air Blank		04:35	Purge Fail
Air Blank		04:35	Purge Fail

Data Download Date/Time: 04/30/2018 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MCGRORY,DIARMUID
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: ROWLAND

DATE/TIME OF TEST: 04/29/2018 04:45:19

Diagnostics Check		OK	04:37
Air Blank		0.000	04:38
Control Test		0.079	04:38
Air Blank		0.000	04:39
Subject Sample	# 1	0.226	04:40
Breath Volume		1.632	04:40
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.235	04:43
Breath Volume		1.453	04:43
Air Blank		0.000	04:44
Control Test		0.080	04:44
Air Blank		0.000	04:45
Diagnostics Check		OK	04:45

Data Download Date/Time: 04/30/2018 15:58