Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: STONE,JIMMIE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 13:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: KADASHAW

DATE/TIME OF TEST: 05/02/2018 14:52:35

Diagnostics Check		OK	14:46
Air Blank		0.000	14:47
Control Test		0.080	14:47
Air Blank		0.000	14:48
Subject Sample	# 1	0.000	14:48
Breath Volume		2.039	14:48
Air Blank		0.000	14:48
Air Blank		0.000	14:50
Subject Sample	# 2	0.000	14:51
Breath Volume		2.757	14:51
Air Blank		0.000	14:51
Control Test		0.080	14:51
Air Blank		0.000	14:52
Diagnostics Check		OK	14:52

Data Download Date/Time: 05/30/2018 11:35

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MYERS,MATT
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: MONTALVO

DATE/TIME OF TEST: 05/04/2018 22:19:36

Diagnostics Check		OK	22:12
Air Blank		0.000	22:13
Control Test		0.081	22:13
Air Blank		0.000	22:14
Subject Sample	# 1	0.000	22:14
Breath Volume		2.289	22:14
Air Blank		0.000	22:15
Air Blank		0.000	22:17
Subject Sample	# 2	0.000	22:18
Breath Volume		1.816	22:18
Air Blank		0.000	22:18
Control Test		0.081	22:19
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

Data Download Date/Time: 05/30/2018 11:35

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: NEALIS,JACOB
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: KINSLEY

DATE/TIME OF TEST: 05/04/2018 23:27:02

Diagnostics Check		OK	23:10	
Air Blank		0.000	23:11	
Control Test		0.080	23:11	
Air Blank		0.000	23:12	
Subject Sample	# 1	0.223	23:14	
Breath Volume		1.218	23:14	
Air Blank		0.000	23:14	
Air Blank		0.000	23:16	
Subject Sample	# 2	0.206	23:19	Volume Not Met
Breath Volume		0.214	23:19	
Air Blank		0.000	23:20	
Air Blank		0.000	23:22	
Subject Sample	# 3	0.186	23:25	Volume Not Met
Breath Volume		0.562	23:25	
Air Blank		0.000	23:26	
Control Test		0.081	23:26	
Air Blank		0.000	23:26	
Diagnostics Check		OK	23:27	

Data Download Date/Time: 05/30/2018 11:35

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: ROYCE,JAMES
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 05/10/2018 23:50:59

Diagnostics Check		OK	23:43
Air Blank		0.000	23:43
Control Test		0.080	23:44
Air Blank		0.000	23:44
Subject Sample	# 1	0.059	23:46
Breath Volume		2.011	23:46
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.053	23:49
Breath Volume		1.433	23:49
Air Blank		0.000	23:50
Control Test		0.079	23:50
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 05/30/2018 11:35

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: SANTANA,JOHN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: COSME

DATE/TIME OF TEST: 05/13/2018 21:32:09

Diagnostics Check		OK	21:26
Air Blank		0.000	21:26
Control Test		0.080	21:26
Air Blank		0.000	21:27
Subject Sample	# 1	0.175	21:27
Breath Volume		2.148	21:27
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.176	21:30
Breath Volume		2.558	21:30
Air Blank		0.000	21:31
Control Test		0.081	21:31
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 05/30/2018 11:35

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MYERS,MATTHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: COFFIE

DATE/TIME OF TEST: 05/19/2018 23:28:55

Diagnostics Check		OK	23:22
Air Blank		0.000	23:22
Control Test		0.079	23:22
Air Blank		0.000	23:23
Subject Sample	# 1	0.182	23:24
Breath Volume		1.187	23:24
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.185	23:27
Breath Volume		1.406	23:27
Air Blank		0.000	23:28
Control Test		0.079	23:28
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 05/30/2018 11:35

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CASELLAS PEREZ

DATE/TIME OF TEST: 05/26/2018 23:18:57

Diagnostics Check		OK	23:10	
Air Blank		0.000	23:10	
Control Test		0.079	23:11	
Air Blank		0.000	23:11	
Subject Sample	# 1	0.413	23:12	
Breath Volume		1.156	23:12	
Air Blank		0.000	23:12	
Air Blank		0.000	23:14	
Subject Sample	# 2	0.382	23:15	
Breath Volume		1.335	23:15	
Air Blank		0.000	23:15	
Air Blank		0.000	23:17	
Subject Sample	# 3	0.361	23:18	
Breath Volume		1.691	23:18	
Air Blank			23:18	RFI Detect
Air Blank		0.000	23:18	

Data Download Date/Time: 05/30/2018 11:35

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CASELLAS PEREZ

DATE/TIME OF TEST: 05/26/2018 23:28:53

Diagnostics Check		OK	23:22
Air Blank		0.000	23:22
Control Test		0.079	23:23
Air Blank		0.000	23:23
Subject Sample	# 1	0.395	23:24
Breath Volume		1.972	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:26
Subject Sample	# 2	0.384	23:27
Breath Volume		1.820	23:27
Air Blank		0.000	23:27
Control Test		0.079	23:28
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 05/30/2018 11:35

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YARBOROUGH,ERIC,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: YARBOROUGH,ERIC
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 05/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/04/2018 22:32:03

Diagnostics Check		OK	22:24
Air Blank		0.000	22:24
Control Test		0.080	22:25
Air Blank		0.000	22:25
Subject Sample	# 1	0.176	22:27
Breath Volume		1.269	22:27
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.180	22:30
Breath Volume		1.832	22:30
Air Blank		0.000	22:31
Control Test		0.080	22:31
Air Blank		0.000	22:31
Diagnostics Check		OK	22:32

Data Download Date/Time: 06/05/2018 12:22

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILLAN,THOMAS,C
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: CAMPBELL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: ROREGUEZ RIOS

DATE/TIME OF TEST: 05/06/2018 03:50:29

Diagnostics Check		OK	03:47	
Air Blank		0.000	03:48	
Control Test		0.077	03:48	
Air Blank		0.000	03:48	
Subject Sample	# 1		03:49	Subject Test Refused
Breath Volume		0.000	03:49	
Air Blank		0.000	03:49	
Control Test		0.079	03:49	
Air Blank		0.000	03:50	
Diagnostics Check		OK	03:50	

Data Download Date/Time: 05/30/2018 19:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORYANCE,JASON,R
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: RILEY,RACHEL
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: VILLASENOR

DATE/TIME OF TEST: 05/08/2018 22:42:47

Diagnostics Check		OK	22:36
Air Blank		0.000	22:36
Control Test		0.079	22:37
Air Blank		0.000	22:37
Subject Sample	# 1	0.203	22:38
Breath Volume		2.421	22:38
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.206	22:41
Breath Volume		2.914	22:41
Air Blank		0.000	22:41
Control Test		0.078	22:42
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 05/30/2018 19:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEMAN,EDWARD,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CARBONE,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/09/2018 18:15:23

Diagnostics Check		OK	18:08
Air Blank		0.000	18:09
Control Test		0.079	18:09
Air Blank		0.000	18:10
Subject Sample	# 1	0.207	18:10
Breath Volume		1.773	18:10
Air Blank		0.000	18:11
Air Blank		0.000	18:13
Subject Sample	# 2	0.196	18:13
Breath Volume		1.304	18:13
Air Blank		0.000	18:14
Control Test		0.077	18:14
Air Blank		0.000	18:15
Diagnostics Check		OK	18:15

Data Download Date/Time: 05/30/2018 19:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: MUINA,NICOLE
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: HOPE

DATE/TIME OF TEST: 05/17/2018 01:38:05

Diagnostics Check		OK	01:31
Air Blank		0.000	01:32
Control Test		0.078	01:32
Air Blank		0.000	01:33
Subject Sample	# 1	0.199	01:33
Breath Volume		1.968	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.196	01:36
Breath Volume		2.480	01:36
Air Blank		0.000	01:37
Control Test		0.077	01:37
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 05/30/2018 19:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOUGHRAN,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/03/2018 04:19:33

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.082	04:14
Air Blank		0.000	04:14
Subject Sample	# 1	0.138	04:15
Breath Volume		2.812	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.131	04:18
Breath Volume		1.640	04:18
Air Blank		0.000	04:18
Control Test		0.082	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 05/10/2018 12:08

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARRO,CASIMIRO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: HAMMER

DATE/TIME OF TEST: 05/03/2018 21:47:19

Diagnostics Check		OK	21:35	
Air Blank		0.000	21:36	
Control Test		0.082	21:36	
Air Blank		0.000	21:36	
Subject Sample	# 1	0.206	21:40	Volume Not Met
Breath Volume		1.070	21:40	
Air Blank		0.000	21:40	
Air Blank		0.000	21:42	
Subject Sample	# 2	0.199	21:45	Volume Not Met
Breath Volume		0.210	21:45	
Air Blank		0.000	21:46	
Control Test		0.081	21:46	
Air Blank		0.000	21:47	
Diagnostics Check		OK	21:47	

Data Download Date/Time: 05/10/2018 12:08

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARRO,CASIMIRO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: HAMMER

DATE/TIME OF TEST: 05/03/2018 22:15:29

Diagnostics Check		OK	22:10	
Air Blank		0.000	22:10	
Control Test		0.082	22:10	
Air Blank		0.000	22:11	
Subject Sample	# 1		22:13	RFI Detect
Breath Volume		0.632	22:13	
Air Blank		0.000	22:14	
Control Test		0.082	22:14	
Air Blank		0.000	22:15	
Diagnostics Check		OK	22:15	

Data Download Date/Time: 05/10/2018 12:08

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARRO,CASIMIRO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: HAMMER

DATE/TIME OF TEST: 05/03/2018 22:27:22

Diagnostics Check		OK	22:18
Air Blank		0.000	22:18
Control Test		0.081	22:19
Air Blank		0.000	22:19
Subject Sample	# 1	0.205	22:21
Breath Volume		1.121	22:21
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.190	22:25
Breath Volume		1.128	22:25
Air Blank		0.000	22:26
Control Test		0.081	22:26
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 05/10/2018 12:08

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: AKERS,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: ESTEBAN

DATE/TIME OF TEST: 05/04/2018 18:43:09

Diagnostics Check		OK	18:36
Air Blank		0.000	18:36
Control Test		0.081	18:37
Air Blank		0.000	18:37
Subject Sample	# 1	0.191	18:38
Breath Volume		2.058	18:38
Air Blank		0.000	18:39
Air Blank		0.000	18:41
Subject Sample	# 2	0.204	18:41
Breath Volume		1.777	18:41
Air Blank		0.000	18:42
Control Test		0.082	18:42
Air Blank		0.000	18:43
Diagnostics Check		OK	18:43

Data Download Date/Time: 05/10/2018 12:08

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: PESCE

DATE/TIME OF TEST: 05/04/2018 21:15:25

Diagnostics Check		OK	21:05
Air Blank		0.000	21:06
Control Test		0.081	21:06
Air Blank		0.000	21:07
Subject Sample	# 1	0.164	21:09
Breath Volume		2.085	21:09
Air Blank		0.000	21:10
Air Blank		0.000	21:12
Subject Sample	# 2	0.159	21:13
Breath Volume		2.656	21:13
Air Blank		0.000	21:14
Control Test		0.081	21:14
Air Blank		0.000	21:15
Diagnostics Check		OK	21:15

Data Download Date/Time: 05/10/2018 12:08

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: COATES

DATE/TIME OF TEST: 05/05/2018 06:05:27

Diagnostics Check		OK	05:56	
Air Blank		0.000	05:57	
Control Test		0.081	05:57	
Air Blank		0.000	05:58	
Subject Sample	# 1		06:01	No Sample Provided
Breath Volume		0.000	06:01	
Air Blank		0.000	06:01	
Air Blank		0.000	06:03	
Subject Sample	# 2		06:04	Subject Test Refused
Breath Volume		0.000	06:04	
Air Blank		0.000	06:04	
Control Test		0.082	06:04	
Air Blank		0.000	06:05	
Diagnostics Check		OK	06:05	

Data Download Date/Time: 05/10/2018 12:08

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: BROMFIELD

DATE/TIME OF TEST: 05/06/2018 04:42:31

Diagnostics Check		OK	04:36
Air Blank		0.000	04:36
Control Test		0.082	04:37
Air Blank		0.000	04:37
Subject Sample	# 1	0.053	04:38
Breath Volume		2.394	04:38
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.047	04:41
Breath Volume		1.820	04:41
Air Blank		0.000	04:41
Control Test		0.082	04:41
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 05/10/2018 12:08

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,YVETTE,D
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: AKERS,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: COHEE

DATE/TIME OF TEST: 05/07/2018 18:07:19

Diagnostics Check		OK	17:59
Air Blank		0.000	17:59
Control Test		0.082	17:59
Air Blank		0.000	18:00
Subject Sample	# 1	0.281	18:02
Breath Volume		2.703	18:02
Air Blank		0.000	18:03
Air Blank		0.000	18:05
Subject Sample	# 2	0.288	18:05
Breath Volume		2.121	18:05
Air Blank		0.000	18:06
Control Test		0.081	18:06
Air Blank		0.000	18:07
Diagnostics Check		OK	18:07

Data Download Date/Time: 05/10/2018 12:08

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STRZALKOWSKI,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: FALLMANN

DATE/TIME OF TEST: 05/07/2018 23:56:30

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.082	23:51
Air Blank		0.000	23:51
Subject Sample	# 1	0.160	23:52
Breath Volume		2.300	23:52
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.151	23:55
Breath Volume		1.945	23:55
Air Blank		0.000	23:55
Control Test		0.081	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 05/10/2018 12:08

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FOREMAN,ANTOINETTE
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: JEAN-PAUL

DATE/TIME OF TEST: 05/16/2018 03:28:58

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.082	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.089	03:24
Breath Volume		1.382	03:24
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.085	03:27
Breath Volume		1.457	03:27
Air Blank		0.000	03:28
Control Test		0.082	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 06/21/2018 11:42

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MORELAND

DATE/TIME OF TEST: 05/17/2018 23:21:31

Diagnostics Check		OK	23:13
Air Blank		0.000	23:14
Control Test		0.082	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.075	23:16
Breath Volume		2.601	23:16
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.072	23:20
Breath Volume		2.832	23:20
Air Blank		0.000	23:20
Control Test		0.081	23:20
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 06/21/2018 11:42

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/18/2018 02:06:42

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:55	
Control Test		0.082	01:56	
Air Blank		0.000	01:56	
Subject Sample	# 1		01:59	No Sample Provided
Breath Volume		0.000	01:59	
Air Blank		0.000	02:00	
Air Blank		0.000	02:02	
Subject Sample	# 2		02:05	No Sample Provided
Breath Volume		0.000	02:05	
Air Blank		0.000	02:05	
Control Test		0.081	02:06	
Air Blank		0.000	02:06	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 06/21/2018 11:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: GUERRA

DATE/TIME OF TEST: 05/18/2018 04:22:15

Diagnostics Check		OK	04:15
Air Blank		0.000	04:16
Control Test		0.083	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.207	04:17
Breath Volume		3.250	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.207	04:20
Breath Volume		2.199	04:20
Air Blank		0.000	04:21
Control Test		0.081	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 06/21/2018 11:42

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JASON
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 05/19/2018 22:53:16

Diagnostics Check		OK	22:46
Air Blank		0.000	22:47
Control Test		0.081	22:47
Air Blank		0.000	22:48
Subject Sample	# 1	0.388	22:48
Breath Volume		1.597	22:48
Air Blank		0.000	22:49
Air Blank		0.000	22:51
Subject Sample	# 2	0.369	22:51
Breath Volume		2.078	22:51
Air Blank		0.000	22:52
Control Test		0.081	22:52
Air Blank		0.000	22:53
Diagnostics Check		OK	22:53

Data Download Date/Time: 06/21/2018 11:42

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ARTURO

DATE/TIME OF TEST: 05/20/2018 01:10:16

Diagnostics Check		OK	01:02
Air Blank		0.000	01:02
Control Test		0.081	01:03
Air Blank		0.000	01:03
Subject Sample	# 1	0.169	01:05
Breath Volume		2.226	01:05
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.159	01:08
Breath Volume		1.875	01:08
Air Blank		0.000	01:09
Control Test		0.081	01:09
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 06/21/2018 11:42

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLAWAY,DAVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 05/20/2018 01:45:03

Diagnostics Check		OK	01:38
Air Blank		0.000	01:39
Control Test		0.081	01:39
Air Blank		0.000	01:39
Subject Sample	# 1	0.273	01:40
Breath Volume		2.015	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:42
Subject Sample	# 2	0.253	01:43
Breath Volume		1.238	01:43
Air Blank		0.000	01:44
Control Test		0.081	01:44
Air Blank		0.000	01:44
Diagnostics Check		OK	01:45

Data Download Date/Time: 06/21/2018 11:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HARDING,JENNIFER
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 05/20/2018 04:55:12

Diagnostics Check		OK	04:45
Air Blank		0.000	04:46
Control Test		0.081	04:46
Air Blank		0.000	04:47
Subject Sample	# 1	0.177	04:49
Breath Volume		1.578	04:49
Air Blank		0.000	04:49
Air Blank		0.000	04:51
Subject Sample	# 2	0.175	04:53
Breath Volume		1.285	04:53
Air Blank		0.000	04:54
Control Test		0.081	04:54
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 06/21/2018 11:42

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SPEZIALE,ADAM
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 05/21/2018 05:39:23

Diagnostics Check		OK	05:32
Air Blank		0.000	05:33
Control Test		0.082	05:33
Air Blank		0.000	05:34
Subject Sample	# 1	0.200	05:34
Breath Volume		2.187	05:34
Air Blank		0.000	05:35
Air Blank		0.000	05:37
Subject Sample	# 2	0.197	05:37
Breath Volume		1.710	05:37
Air Blank		0.000	05:38
Control Test		0.081	05:38
Air Blank		0.000	05:39
Diagnostics Check		OK	05:39

Data Download Date/Time: 06/21/2018 11:42

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STRZALKOWSKI,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 05/21/2018 20:39:04

Diagnostics Check		OK	20:30
Air Blank		0.000	20:30
Control Test		0.082	20:30
Air Blank		0.000	20:31
Subject Sample	# 1	0.238	20:32
Breath Volume		1.183	20:32
Air Blank		0.000	20:32
Air Blank		0.000	20:34
Subject Sample	# 2	0.230	20:37
Breath Volume		1.601	20:37
Air Blank		0.000	20:38
Control Test		0.082	20:38
Air Blank		0.000	20:38
Diagnostics Check		OK	20:39

Data Download Date/Time: 06/21/2018 11:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 05/23/2018 22:59:53

Diagnostics Check		OK	22:53
Air Blank		0.000	22:54
Control Test		0.081	22:54
Air Blank		0.000	22:55
Subject Sample	# 1	0.265	22:55
Breath Volume		1.105	22:55
Air Blank		0.000	22:56
Air Blank		0.000	22:57
Subject Sample	# 2	0.270	22:58
Breath Volume		1.214	22:58
Air Blank		0.000	22:58
Control Test		0.082	22:59
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 06/21/2018 11:42

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: GELLER

DATE/TIME OF TEST: 05/23/2018 23:50:13

Diagnostics Check		OK	23:44
Air Blank		0.000	23:44
Control Test		0.082	23:44
Air Blank		0.000	23:45
Subject Sample	# 1	0.000	23:45
Breath Volume		2.398	23:45
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.000	23:48
Breath Volume		1.355	23:48
Air Blank		0.000	23:49
Control Test		0.082	23:49
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 06/21/2018 11:43

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: FRANCISCO-LAPAIX

DATE/TIME OF TEST: 05/24/2018 02:50:10

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.082	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.117	02:45
Breath Volume		3.277	02:45
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.117	02:48
Breath Volume		3.687	02:48
Air Blank		0.000	02:49
Control Test		0.083	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 06/21/2018 11:43

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: GILL

DATE/TIME OF TEST: 05/25/2018 02:13:43

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.081	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.145	02:09
Breath Volume		2.890	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.139	02:12
Breath Volume		2.550	02:12
Air Blank		0.000	02:12
Control Test		0.081	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 06/21/2018 11:43

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JAY
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GASKIN

DATE/TIME OF TEST: 05/26/2018 03:58:15

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.082	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.153	03:52
Breath Volume		1.488	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.155	03:56
Breath Volume		1.625	03:56
Air Blank		0.000	03:57
Control Test		0.082	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 06/21/2018 11:43

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SAEHUNG

DATE/TIME OF TEST: 05/27/2018 00:47:47

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.083	00:42
Air Blank		0.000	00:42
Subject Sample	# 1	0.275	00:43
Breath Volume		1.175	00:43
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.260	00:46
Breath Volume		1.464	00:46
Air Blank		0.000	00:46
Control Test		0.081	00:47
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 06/21/2018 11:43

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JAY
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 05/28/2018 02:22:40

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.082	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.209	02:18
Breath Volume		1.898	02:18
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.200	02:21
Breath Volume		1.531	02:21
Air Blank		0.000	02:21
Control Test		0.081	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 06/21/2018 11:43

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JAY
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: CUSHING

DATE/TIME OF TEST: 05/29/2018 02:46:41

Diagnostics Check		OK	02:39
Air Blank		0.000	02:40
Control Test		0.083	02:40
Air Blank		0.000	02:40
Subject Sample	# 1	0.114	02:42
Breath Volume		1.957	02:42
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.117	02:45
Breath Volume		1.765	02:45
Air Blank		0.000	02:45
Control Test		0.082	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 06/21/2018 11:43

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/30/2018 01:39:58

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.082	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.153	01:35
Breath Volume		1.632	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:37
Subject Sample	# 2	0.150	01:38
Breath Volume		2.535	01:38
Air Blank		0.000	01:39
Control Test		0.082	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 06/21/2018 11:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BAIN,MERVIN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 05/31/2018 23:11:06

Diagnostics Check		OK	23:04
Air Blank		0.000	23:04
Control Test		0.082	23:04
Air Blank		0.000	23:05
Subject Sample	# 1	0.192	23:06
Breath Volume		1.421	23:06
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.210	23:09
Breath Volume		1.910	23:09
Air Blank		0.000	23:10
Control Test		0.082	23:10
Air Blank		0.000	23:11
Diagnostics Check		OK	23:11

Data Download Date/Time: 06/21/2018 11:43

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBLINO,NICK,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ADAMS,RON
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: DISCHINGER

DATE/TIME OF TEST: 05/15/2018 02:03:43

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:01	
Control Test		0.084	02:01	
Air Blank		0.000	02:02	
Subject Sample	# 1		02:02	Subject Test Refused
Breath Volume		0.000	02:02	
Air Blank		0.000	02:02	
Control Test		0.083	02:03	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 06/10/2018 22:13

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KALLMEYER,DANIEL,K
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: OTHER
ARREST TIME: 18:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: ARMBRUSTER

DATE/TIME OF TEST: 05/18/2018 18:37:10

Diagnostics Check		OK	18:31
Air Blank		0.000	18:31
Control Test		0.084	18:31
Air Blank		0.000	18:32
Subject Sample	# 1	0.000	18:32
Breath Volume		2.101	18:32
Air Blank		0.000	18:33
Air Blank		0.000	18:35
Subject Sample	# 2	0.000	18:35
Breath Volume		2.902	18:35
Air Blank		0.000	18:36
Control Test		0.084	18:36
Air Blank		0.000	18:37
Diagnostics Check		OK	18:37

Data Download Date/Time: 06/10/2018 22:13

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBLINO,NICK,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: LUNA,ISAEL
ARRESTING OFFICER AGENCY: Davenport PD
LAST AGENCY INSPECTION: 05/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: COLIN

DATE/TIME OF TEST: 05/27/2018 23:53:10

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:52	
Control Test	0.092	23:52	Control Outside Tolerance
Air Blank	0.000	23:53	

Data Download Date/Time: 06/10/2018 22:13

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBLINO,NICK,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: LUNA,ISAEL
ARRESTING OFFICER AGENCY: Davenport PD
LAST AGENCY INSPECTION: 05/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: COLIN

DATE/TIME OF TEST: 05/28/2018 00:07:48

Diagnostics Check		OK	23:55	
Air Blank		0.000	23:55	
Control Test		0.084	23:56	
Air Blank		0.000	23:56	
Subject Sample	# 3	0.206	23:59	Slope Not Level
Breath Volume		1.269	23:59	
Air Blank		0.000	00:00	
Air Blank		0.000	00:02	
Subject Sample	# 1	0.239	00:02	
Breath Volume		1.292	00:02	
Air Blank		0.000	00:03	
Air Blank		0.000	00:05	
Subject Sample	# 2	0.219	00:06	
Breath Volume		1.167	00:06	
Air Blank		0.000	00:06	
Control Test		0.080	00:07	
Air Blank		0.000	00:07	
Diagnostics Check		OK	00:07	

Data Download Date/Time: 06/10/2018 22:13

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBLINO,NICK,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: COSSIO GARCIA

DATE/TIME OF TEST: 05/28/2018 22:51:30

Diagnostics Check		OK	22:42	
Air Blank		0.000	22:43	
Control Test		0.084	22:43	
Air Blank		0.000	22:43	
Subject Sample	# 1	0.190	22:45	
Breath Volume		1.406	22:45	
Air Blank		0.000	22:46	
Air Blank		0.000	22:47	
Subject Sample	# 2		22:50	Subject Test Refused
Breath Volume		0.000	22:50	
Air Blank		0.000	22:50	
Control Test		0.083	22:50	
Air Blank		0.000	22:51	
Diagnostics Check		OK	22:51	

Data Download Date/Time: 06/10/2018 22:13

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,BRIAN,R
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: PAGE,BRIAN
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: VOGT

DATE/TIME OF TEST: 05/10/2018 21:08:47

Diagnostics Check	Fail	21:08	Voltage/Current Test
Air Blank	0.000	21:08	

Data Download Date/Time: 05/26/2018 09:39

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,BRIAN,R
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: PAGE,BRIAN
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: VOGT

DATE/TIME OF TEST: 05/10/2018 21:21:47

Diagnostics Check		OK	21:11	
Air Blank		0.000	21:11	
Control Test		0.078	21:11	
Air Blank		0.000	21:12	
Subject Sample	# 1	0.249	21:13	
Breath Volume		1.289	21:13	
Air Blank		0.000	21:13	
Air Blank		0.000	21:15	
Subject Sample	# 2	0.210	21:16	
Breath Volume		1.253	21:16	
Air Blank		0.000	21:17	
Air Blank		0.000	21:19	
Subject Sample	# 3	0.227	21:19	
Breath Volume		1.730	21:19	
Air Blank		0.000	21:20	
Control Test		0.078	21:20	
Air Blank		0.000	21:21	
Diagnostics Check		Fail	21:21	Voltage/Current Test
Air Blank		0.000	21:21	

Data Download Date/Time: 05/26/2018 09:39

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,CAITLIN,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: JENKINS,ANDRE
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: KUBIK

DATE/TIME OF TEST: 05/13/2018 02:52:22

Diagnostics Check		OK	02:43
Air Blank		0.000	02:44
Control Test		0.080	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.136	02:47
Breath Volume		1.871	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.132	02:50
Breath Volume		1.933	02:50
Air Blank		0.000	02:51
Control Test		0.077	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 05/26/2018 09:39

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: KAVANAUGH,ANTHONY
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/11/2018 03:24:32

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.082	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.000	03:20
Breath Volume		1.847	03:20
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.000	03:23
Breath Volume		2.144	03:23
Air Blank		0.000	03:23
Control Test		0.082	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 05/30/2018 11:23

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIGLER,RICKIE,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ZIGLER,RICKIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 05/21/2018 01:59:37

Diagnostics Check		OK	01:52
Air Blank		0.000	01:53
Control Test		0.082	01:53
Air Blank		0.000	01:54
Subject Sample	# 1	0.186	01:54
Breath Volume		1.839	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.181	01:58
Breath Volume		1.976	01:58
Air Blank		0.000	01:58
Control Test		0.081	01:59
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 05/30/2018 11:23

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,RONELL,T
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: BYLICKI,DANIEL
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 05/19/2018 22:19:33

Diagnostics Check		OK	22:11
Air Blank		0.000	22:12
Control Test		0.079	22:12
Air Blank		0.000	22:13
Subject Sample	# 1	0.186	22:14
Breath Volume		2.871	22:14
Air Blank		0.000	22:15
Air Blank		0.000	22:17
Subject Sample	# 2	0.183	22:17
Breath Volume		2.464	22:17
Air Blank		0.000	22:18
Control Test		0.077	22:18
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

Data Download Date/Time: 05/31/2018 14:42

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,RONELL,T
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: MARTYN,KYLE
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/20/2018 00:43:48

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.078	00:42	
Air Blank		00:43	RFI Detect
Air Blank	0.000	00:43	

Data Download Date/Time: 05/31/2018 14:42

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,RONELL,T
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: MARTYN,KYLE
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/20/2018 00:55:24

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.079	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.127	00:47
Breath Volume		1.414	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.148	00:50
Breath Volume		1.546	00:50
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 3	0.139	00:53
Breath Volume		2.253	00:53
Air Blank		0.000	00:54
Control Test		0.077	00:54
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 05/31/2018 14:42

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAKUBOWSKI,SCOTT,T
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: DILBERT,EZRA
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: LORANGE

DATE/TIME OF TEST: 05/26/2018 21:15:31

Diagnostics Check		OK	21:04	
Air Blank		0.000	21:04	
Control Test		0.079	21:04	
Air Blank		0.000	21:05	
Subject Sample	# 1	0.000	21:08	Volume Not Met
Breath Volume		0.277	21:08	
Air Blank		0.000	21:09	
Air Blank		0.000	21:11	
Subject Sample	# 2	0.000	21:14	Volume Not Met
Breath Volume		0.839	21:14	
Air Blank		0.000	21:14	
Control Test		0.078	21:14	
Air Blank		0.000	21:15	
Diagnostics Check		OK	21:15	

Data Download Date/Time: 05/31/2018 14:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAKUBOWSKI,SCOTT,T
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: JAKUBOWSKI,SCOTT
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: KECK

DATE/TIME OF TEST: 05/26/2018 23:58:50

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.078	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.179	23:54
Breath Volume		2.125	23:54
Air Blank		0.000	23:55
Air Blank		0.000	23:56
Subject Sample	# 2	0.170	23:57
Breath Volume		1.886	23:57
Air Blank		0.000	23:57
Control Test		0.077	23:58
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 05/31/2018 14:42

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARMINGTON,WILLIAM,G
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: STANLEY,LLOYD
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 05/01/2018 02:54:30

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.079	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.154	02:50
Breath Volume		1.765	02:50
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.156	02:52
Breath Volume		2.035	02:52
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 05/28/2018 11:31

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARMINGTON,WILLIAM,G
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: STAGE,NICHOLAS
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: SINGHAL

DATE/TIME OF TEST: 05/02/2018 04:46:58

Diagnostics Check		OK	04:44	
Air Blank		0.000	04:44	
Control Test		0.081	04:44	
Air Blank		0.000	04:45	
Subject Sample	# 1		04:45	Subject Test Refused
Breath Volume		0.000	04:45	
Air Blank		0.000	04:46	
Control Test		0.081	04:46	
Air Blank		0.000	04:46	
Diagnostics Check		OK	04:46	

Data Download Date/Time: 05/28/2018 11:31

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARMINGTON,WILLIAM,G
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: KAVANAUGH,ANTHONY
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: OHEARN

DATE/TIME OF TEST: 05/06/2018 00:59:48

Diagnostics Check		OK	00:45	
Air Blank		0.000	00:46	
Control Test		0.080	00:46	
Air Blank		0.000	00:47	
Subject Sample	# 1	0.166	00:50	Volume Not Met
Breath Volume		0.203	00:50	
Air Blank		0.000	00:50	
Air Blank		0.000	00:52	
Subject Sample	# 2	0.155	00:55	
Breath Volume		1.148	00:55	
Air Blank		0.000	00:55	
Air Blank		0.000	00:57	
Subject Sample	# 3	0.167	00:58	
Breath Volume		1.609	00:58	
Air Blank		0.000	00:58	
Control Test		0.077	00:59	
Air Blank		0.000	00:59	
Diagnostics Check		OK	00:59	

Data Download Date/Time: 05/28/2018 11:31

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARMINGTON,WILLIAM,G
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: KAVANAUGH,ANTHONY
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: OHEARN

DATE/TIME OF TEST: 05/06/2018 01:08:04

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.079	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.174	01:03
Breath Volume		1.167	01:03
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.173	01:06
Breath Volume		1.554	01:06
Air Blank		0.000	01:07
Control Test		0.079	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:08

Data Download Date/Time: 05/28/2018 11:31

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARMINGTON,WILLIAM,G
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: ROBINSON,IVAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: THOMASSON

DATE/TIME OF TEST: 05/06/2018 02:48:06

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.080	02:42
Air Blank		0.000	02:43
Subject Sample	# 1	0.114	02:43
Breath Volume		1.164	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.117	02:46
Breath Volume		1.187	02:46
Air Blank		0.000	02:47
Control Test		0.080	02:47
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 05/28/2018 11:31

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBIN,RICHARD,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: RAIHALA,J
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: USEFOF

DATE/TIME OF TEST: 05/08/2018 03:45:13

Diagnostics Check		OK	03:38
Air Blank		0.000	03:38
Control Test		0.079	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.209	03:40
Breath Volume		2.355	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.219	03:43
Breath Volume		1.214	03:43
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 05/28/2018 11:31

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBIN,RICHARD,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: ROBINSON,I
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 05/12/2018 00:53:26

Diagnostics Check		OK	00:50	
Air Blank		0.000	00:51	
Control Test		0.081	00:51	
Air Blank		0.000	00:51	
Subject Sample	# 1		00:52	Subject Test Refused
Breath Volume		0.000	00:52	
Air Blank		0.000	00:52	
Control Test		0.081	00:52	
Air Blank		0.000	00:53	
Diagnostics Check		OK	00:53	

Data Download Date/Time: 05/28/2018 11:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBIN,RICHARD,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: RAIHALA,J
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: FORTH

DATE/TIME OF TEST: 05/13/2018 04:57:18

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.081	04:50
Air Blank		0.000	04:51
Subject Sample	# 1	0.201	04:52
Breath Volume		1.718	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:54
Subject Sample	# 2	0.197	04:55
Breath Volume		1.320	04:55
Air Blank		0.000	04:56
Control Test		0.081	04:56
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 05/28/2018 11:31

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,BENITA,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: DOUGLAS,THOMAS
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: FIRESTONE

DATE/TIME OF TEST: 05/28/2018 00:30:09

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:16	
Control Test		0.080	00:16	
Air Blank		0.000	00:17	
Subject Sample	# 1	0.148	00:20	Volume Not Met
Breath Volume		0.019	00:20	
Air Blank		0.000	00:20	
Air Blank		0.000	00:22	
Subject Sample	# 2	0.153	00:23	
Breath Volume		1.476	00:23	
Air Blank		0.000	00:24	
Air Blank		0.000	00:26	
Subject Sample	# 3	0.142	00:28	
Breath Volume		1.359	00:28	
Air Blank		0.000	00:29	
Control Test		0.079	00:29	
Air Blank		0.000	00:30	
Diagnostics Check		OK	00:30	

Data Download Date/Time: 05/28/2018 11:31

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001158

INSTRUMENT REGISTERED TO: PONCE INLET PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAIG,BRADLEY,A
BREATH TEST OPERATOR AGENCY: Ponce Inlet PD
ARRESTING OFFICER: CRAIG,BRADLEY
ARRESTING OFFICER AGENCY: Ponce Inlet PD
LAST AGENCY INSPECTION: 05/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: HOFFMANN

DATE/TIME OF TEST: 05/29/2018 00:39:16

Diagnostics Check		OK	00:36	
Air Blank		0.000	00:36	
Control Test		0.082	00:36	
Air Blank		0.000	00:37	
Subject Sample	# 1		00:37	Subject Test Refused
Breath Volume		0.000	00:37	
Air Blank		0.000	00:38	
Control Test		0.081	00:38	
Air Blank		0.000	00:39	
Diagnostics Check		OK	00:39	

Data Download Date/Time: 06/25/2018 21:25

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: LADY LAKE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER,TRAVIS,J
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: THOMAS,MATT
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 04/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/10/2018 02:09:42

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.076	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.201	02:04
Breath Volume		2.097	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.186	02:08
Breath Volume		1.367	02:08
Air Blank		0.000	02:08
Control Test		0.076	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 05/15/2018 09:58

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001178

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PASTOR,JOSE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: MENA ARELLANO

DATE/TIME OF TEST: 05/26/2018 18:50:51

Diagnostics Check		OK	18:42
Air Blank		0.000	18:43
Control Test		0.079	18:43
Air Blank		0.000	18:44
Subject Sample	# 1	0.156	18:45
Breath Volume		1.785	18:45
Air Blank		0.000	18:45
Air Blank		0.000	18:47
Subject Sample	# 2	0.172	18:49
Breath Volume		2.433	18:49
Air Blank		0.000	18:49
Control Test		0.079	18:50
Air Blank		0.000	18:50
Diagnostics Check		OK	18:50

Data Download Date/Time: 06/28/2018 12:51

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: ZOLNAY

DATE/TIME OF TEST: 05/05/2018 20:49:52

Diagnostics Check		OK	20:43
Air Blank		0.000	20:44
Control Test		0.081	20:44
Air Blank		0.000	20:44
Subject Sample	# 1	0.000	20:45
Breath Volume		3.437	20:45
Air Blank		0.000	20:45
Air Blank		0.000	20:47
Subject Sample	# 2	0.000	20:48
Breath Volume		2.742	20:48
Air Blank		0.000	20:48
Control Test		0.080	20:49
Air Blank		0.000	20:49
Diagnostics Check		OK	20:49

Data Download Date/Time: 05/26/2018 22:47

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: DOMINGUEZ,JAYMES
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 05/06/2018 01:36:52

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.080	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.108	01:31
Breath Volume		1.195	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.111	01:35
Breath Volume		1.835	01:35
Air Blank		0.000	01:35
Control Test		0.081	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 05/26/2018 22:47

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/06/2018 03:23:12

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:08	
Control Test		0.080	03:09	
Air Blank		0.000	03:09	
Subject Sample	# 1	0.000	03:12	Volume Not Met
Breath Volume		0.050	03:12	
Air Blank		0.000	03:13	
Air Blank		0.000	03:15	
Subject Sample	# 2	0.205	03:16	No .020 agreement
Breath Volume		1.671	03:16	
Air Blank		0.000	03:17	
Air Blank		0.000	03:19	
Subject Sample	# 3	0.181	03:21	No .020 agreement
Breath Volume		1.511	03:21	
Air Blank		0.000	03:22	
Control Test		0.079	03:22	
Air Blank		0.000	03:23	
Diagnostics Check		OK	03:23	

Data Download Date/Time: 05/26/2018 22:47

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HERNANDEZ SANCHEZ

DATE/TIME OF TEST: 05/27/2018 01:34:01

Diagnostics Check		OK	01:26
Air Blank		0.000	01:27
Control Test		0.080	01:27
Air Blank		0.000	01:28
Subject Sample	# 1	0.186	01:29
Breath Volume		1.617	01:29
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.198	01:32
Breath Volume		1.179	01:32
Air Blank		0.000	01:33
Control Test		0.077	01:33
Air Blank		0.000	01:33
Diagnostics Check		OK	01:34

Data Download Date/Time: 06/28/2018 13:24

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAFA,YUSUF,A
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: ARAFA,YUSUF
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: STEVERSON

DATE/TIME OF TEST: 05/06/2018 18:01:57

Diagnostics Check		OK	17:55
Air Blank		0.000	17:55
Control Test		0.080	17:56
Air Blank		0.000	17:56
Subject Sample	# 1	0.255	17:57
Breath Volume		1.902	17:57
Air Blank		0.000	17:58
Air Blank		0.000	17:59
Subject Sample	# 2	0.264	18:00
Breath Volume		3.632	18:00
Air Blank		0.000	18:01
Control Test		0.080	18:01
Air Blank		0.000	18:01
Diagnostics Check		OK	18:01

Data Download Date/Time: 06/22/2018 18:33

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,JACOB,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MOORE,JACOB
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: ENTREKIN

DATE/TIME OF TEST: 05/06/2018 21:27:16

Diagnostics Check		OK	21:21
Air Blank		0.000	21:21
Control Test		0.080	21:21
Air Blank		0.000	21:22
Subject Sample	# 1	0.184	21:22
Breath Volume		1.925	21:22
Air Blank		0.000	21:23
Air Blank		0.000	21:25
Subject Sample	# 2	0.185	21:25
Breath Volume		2.632	21:25
Air Blank		0.000	21:26
Control Test		0.080	21:26
Air Blank		0.000	21:27
Diagnostics Check		OK	21:27

Data Download Date/Time: 06/22/2018 18:33

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRITT,DAKOTA,A
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MERRITT,DAKOTA
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: QUATTLEBAUM

DATE/TIME OF TEST: 05/04/2018 18:29:37

Diagnostics Check		OK	18:23
Air Blank		0.000	18:24
Control Test		0.076	18:24
Air Blank		0.000	18:24
Subject Sample	# 1	0.178	18:25
Breath Volume		2.085	18:25
Air Blank		0.000	18:25
Air Blank		0.000	18:27
Subject Sample	# 2	0.176	18:28
Breath Volume		2.355	18:28
Air Blank		0.000	18:28
Control Test		0.079	18:29
Air Blank		0.000	18:29
Diagnostics Check		OK	18:29

Data Download Date/Time: 05/28/2018 10:54

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BRENNAN,C
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 05/05/2018 22:14:30

Diagnostics Check		OK	22:08
Air Blank		0.000	22:08
Control Test		0.079	22:09
Air Blank		0.000	22:09
Subject Sample	# 1	0.155	22:10
Breath Volume		2.023	22:10
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.155	22:13
Breath Volume		2.175	22:13
Air Blank		0.000	22:13
Control Test		0.080	22:13
Air Blank		0.000	22:14
Diagnostics Check		OK	22:14

Data Download Date/Time: 05/28/2018 10:54

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,DAVID,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: RHODES,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 05/13/2018 01:16:29

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:13	
Control Test		0.079	01:14	
Air Blank		0.000	01:14	
Subject Sample	# 1		01:15	Subject Test Refused
Breath Volume		0.000	01:15	
Air Blank		0.000	01:15	
Control Test		0.079	01:15	
Air Blank		0.000	01:16	
Diagnostics Check		OK	01:16	

Data Download Date/Time: 05/28/2018 10:54

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CORONILLA,DANIEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: PERALTA

DATE/TIME OF TEST: 05/31/2018 03:21:38

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:19	
Control Test		0.081	03:19	
Air Blank		0.000	03:19	
Subject Sample	# 1		03:20	Subject Test Refused
Breath Volume		0.000	03:20	
Air Blank		0.000	03:20	
Control Test		0.081	03:21	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 06/25/2018 21:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: VANDETTI,MATTHEW
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 05/31/2018 20:24:00

Diagnostics Check		OK	20:15
Air Blank		0.000	20:15
Control Test		0.081	20:15
Air Blank		0.000	20:16
Subject Sample	# 1	0.217	20:18
Breath Volume		1.246	20:18
Air Blank		0.000	20:19
Air Blank		0.000	20:21
Subject Sample	# 2	0.219	20:22
Breath Volume		1.132	20:22
Air Blank		0.000	20:23
Control Test		0.079	20:23
Air Blank		0.000	20:23
Diagnostics Check		OK	20:24

Data Download Date/Time: 06/25/2018 21:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: JOHNSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: LACY

DATE/TIME OF TEST: 05/03/2018 01:33:34

Diagnostics Check		OK	01:27
Air Blank		0.000	01:27
Control Test		0.079	01:28
Air Blank		0.000	01:28
Subject Sample	# 1	0.230	01:28
Breath Volume		2.882	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.229	01:32
Breath Volume		2.468	01:32
Air Blank		0.000	01:32
Control Test		0.079	01:32
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 05/10/2018 15:21

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLACE,NICHOLAS,T
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: GARCIA,MICHAEL
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 04/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: HARMON

DATE/TIME OF TEST: 05/05/2018 00:42:53

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.080	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.115	00:38
Breath Volume		2.527	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.111	00:41
Breath Volume		3.011	00:41
Air Blank		0.000	00:41
Control Test		0.080	00:42
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 05/10/2018 15:21

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: HASTINGS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 05/12/2018 00:31:30

Diagnostics Check		OK	00:24
Air Blank		0.000	00:24
Control Test		0.080	00:24
Air Blank		0.000	00:25
Subject Sample	# 1	0.103	00:26
Breath Volume		2.113	00:26
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.095	00:30
Breath Volume		2.003	00:30
Air Blank		0.000	00:30
Control Test		0.079	00:30
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 06/22/2018 16:39

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: JOHNSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ARNULFO

DATE/TIME OF TEST: 05/13/2018 02:42:36

Diagnostics Check		OK	02:36
Air Blank		0.000	02:36
Control Test		0.080	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.170	02:38
Breath Volume		2.640	02:38
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.165	02:41
Breath Volume		2.128	02:41
Air Blank		0.000	02:41
Control Test		0.080	02:42
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 06/22/2018 16:39

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: RHODES

DATE/TIME OF TEST: 05/13/2018 03:28:24

Diagnostics Check		OK	03:19
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.123	03:22
Breath Volume		2.472	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:24
Subject Sample	# 2	0.115	03:26
Breath Volume		2.593	03:26
Air Blank		0.000	03:27
Control Test		0.078	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 06/22/2018 16:39

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOUZA,AARON,B
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: KELLY,ROBERT
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: OTHER
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:02

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/15/2018 15:29:16

Diagnostics Check		OK	15:23
Air Blank		0.000	15:23
Control Test		0.079	15:23
Air Blank		0.000	15:24
Subject Sample	# 1	0.284	15:24
Breath Volume		2.714	15:24
Air Blank		0.000	15:25
Air Blank		0.000	15:27
Subject Sample	# 2	0.278	15:27
Breath Volume		1.699	15:27
Air Blank		0.000	15:28
Control Test		0.079	15:28
Air Blank		0.000	15:29
Diagnostics Check		OK	15:29

Data Download Date/Time: 06/22/2018 16:39

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: STONE,RODNEYSIA
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: COYLE

DATE/TIME OF TEST: 05/16/2018 22:18:20

Diagnostics Check		OK	22:11
Air Blank		0.000	22:11
Control Test		0.081	22:12
Air Blank		0.000	22:12
Subject Sample	# 1	0.000	22:13
Breath Volume		1.527	22:13
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.000	22:16
Breath Volume		4.203	22:16
Air Blank		0.000	22:17
Control Test		0.079	22:17
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 06/22/2018 16:39

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: BAKER,MATTHEW
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: HERNANDEZ MARTINEZ

DATE/TIME OF TEST: 05/25/2018 21:29:05

Diagnostics Check		OK	21:26	
Air Blank		0.000	21:26	
Control Test		0.080	21:26	
Air Blank		0.000	21:27	
Subject Sample	# 1		21:27	Subject Test Refused
Breath Volume		0.000	21:27	
Air Blank		0.000	21:28	
Control Test		0.080	21:28	
Air Blank		0.000	21:29	
Diagnostics Check		OK	21:29	

Data Download Date/Time: 06/26/2018 03:17

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICAH,L
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: MCDONALD,GREGORY
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 152169
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 05/15/2018 21:55:11

Diagnostics Check		OK	21:46
Air Blank		0.000	21:46
Control Test		0.080	21:46
Air Blank		0.000	21:47
Subject Sample	# 1	0.255	21:47
Breath Volume		1.945	21:47
Air Blank		0.000	21:48
Air Blank		0.000	21:50
Subject Sample	# 2	0.282	21:50
Breath Volume		1.843	21:50
Air Blank		0.000	21:51
Air Blank		0.000	21:53
Subject Sample	# 3	0.280	21:53
Breath Volume		2.835	21:53
Air Blank		0.000	21:54
Control Test		0.080	21:54
Air Blank		0.000	21:55
Diagnostics Check		OK	21:55

Data Download Date/Time: 05/30/2018 08:50

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: STREIGOLD,DAVID
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 152169
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BEDFORD

DATE/TIME OF TEST: 05/19/2018 01:03:14

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:00	
Control Test		0.079	01:01	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:01	Subject Test Refused
Breath Volume		0.000	01:01	
Air Blank		0.000	01:02	
Control Test		0.080	01:02	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 05/30/2018 08:50

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICAH,L
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: THOMAS,MICAH
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 152169
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: BANGOR

DATE/TIME OF TEST: 05/22/2018 02:09:48

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:07	
Control Test		0.080	02:07	
Air Blank		0.000	02:08	
Subject Sample	# 1		02:08	Subject Test Refused
Breath Volume		0.000	02:08	
Air Blank		0.000	02:08	
Control Test		0.080	02:09	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 05/30/2018 08:50

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 05/09/2018 03:14:28

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.197	03:09
Breath Volume		3.648	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.204	03:12
Breath Volume		3.718	03:12
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 05/22/2018 16:18

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: WEIDL-PARNELL

DATE/TIME OF TEST: 05/26/2018 19:18:49

Diagnostics Check		OK	19:13	
Air Blank		0.000	19:14	
Control Test		0.080	19:14	
Air Blank		0.000	19:14	
Subject Sample	# 1		19:17	RFI Detect
Breath Volume		0.742	19:17	
Air Blank		0.000	19:17	
Control Test		0.082	19:18	
Air Blank		0.000	19:18	
Diagnostics Check		OK	19:18	

Data Download Date/Time: 06/27/2018 15:39

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: WEIDL-PARNELL

DATE/TIME OF TEST: 05/26/2018 19:32:35

Diagnostics Check		OK	19:21	
Air Blank		0.000	19:21	
Control Test		0.079	19:21	
Air Blank		0.000	19:22	
Subject Sample	# 1	0.000	19:25	Volume Not Met
Breath Volume		0.816	19:25	
Air Blank		0.000	19:26	
Air Blank		0.000	19:28	
Subject Sample	# 2	0.000	19:31	Volume Not Met
Breath Volume		0.839	19:31	
Air Blank		0.000	19:31	
Control Test		0.080	19:31	
Air Blank		0.000	19:32	
Diagnostics Check		OK	19:32	

Data Download Date/Time: 06/27/2018 15:39

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: MOTT III

DATE/TIME OF TEST: 05/26/2018 20:55:47

Diagnostics Check		OK	20:47	
Air Blank		0.000	20:47	
Control Test		0.080	20:48	
Air Blank		0.000	20:48	
Subject Sample	# 1		20:51	No Sample Provided
Breath Volume		0.000	20:51	
Air Blank		0.000	20:52	
Air Blank		0.000	20:54	
Subject Sample	# 2		20:54	Subject Test Refused
Breath Volume		0.000	20:54	
Air Blank		0.000	20:54	
Control Test		0.079	20:55	
Air Blank		0.000	20:55	
Diagnostics Check		OK	20:55	

Data Download Date/Time: 06/27/2018 15:39

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: EARLS,CHRISTIAN
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: SACA

DATE/TIME OF TEST: 05/26/2018 21:36:21

Diagnostics Check		OK	21:24	
Air Blank		0.000	21:25	
Control Test		0.080	21:25	
Air Blank		0.000	21:26	
Subject Sample	# 1	0.099	21:29	Volume Not Met
Breath Volume		0.046	21:29	
Air Blank		0.000	21:29	
Air Blank		0.000	21:31	
Subject Sample	# 2	0.085	21:34	Volume Not Met
Breath Volume		0.046	21:34	
Air Blank		0.000	21:35	
Control Test		0.080	21:35	
Air Blank		0.000	21:36	
Diagnostics Check		OK	21:36	

Data Download Date/Time: 06/27/2018 15:39

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/27/2018 20:54:49

Diagnostics Check		OK	20:48
Air Blank		0.000	20:48
Control Test		0.077	20:49
Air Blank		0.000	20:49
Subject Sample	# 1	0.198	20:50
Breath Volume		2.605	20:50
Air Blank		0.000	20:51
Air Blank		0.000	20:52
Subject Sample	# 2	0.184	20:53
Breath Volume		1.851	20:53
Air Blank		0.000	20:53
Control Test		0.078	20:54
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 06/27/2018 15:39

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CARTER,JACOB
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: KERCHER

DATE/TIME OF TEST: 05/28/2018 00:53:21

Diagnostics Check		OK	00:47
Air Blank		0.000	00:47
Control Test		0.080	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.083	00:48
Breath Volume		1.171	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.080	00:51
Breath Volume		2.558	00:51
Air Blank		0.000	00:52
Control Test		0.080	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 06/27/2018 15:39

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SNOW

DATE/TIME OF TEST: 05/28/2018 01:03:27

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:00	
Control Test		0.079	01:01	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:02	Subject Test Refused
Breath Volume		0.000	01:02	
Air Blank		0.000	01:02	
Control Test		0.079	01:02	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 06/27/2018 15:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKE,CLEVELAND,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: VINCENT

DATE/TIME OF TEST: 05/28/2018 20:18:57

Diagnostics Check		OK	20:12
Air Blank		0.000	20:13
Control Test		0.080	20:13
Air Blank		0.000	20:13
Subject Sample	# 1	0.136	20:14
Breath Volume		3.253	20:14
Air Blank		0.000	20:15
Air Blank		0.000	20:16
Subject Sample	# 2	0.128	20:17
Breath Volume		2.531	20:17
Air Blank		0.000	20:18
Control Test		0.078	20:18
Air Blank		0.000	20:18
Diagnostics Check		OK	20:18

Data Download Date/Time: 06/27/2018 15:39

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKE,CLEVELAND,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEWEESE,JEREMY
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BRANCH

DATE/TIME OF TEST: 05/29/2018 01:30:55

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:28	
Control Test		0.080	01:28	
Air Blank		0.000	01:28	
Subject Sample	# 1		01:29	Subject Test Refused
Breath Volume		0.000	01:29	
Air Blank		0.000	01:30	
Control Test		0.080	01:30	
Air Blank		0.000	01:30	
Diagnostics Check		OK	01:30	

Data Download Date/Time: 06/27/2018 15:39

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: DALBEY

DATE/TIME OF TEST: 05/12/2018 23:59:14

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.081	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.114	23:54
Breath Volume		2.664	23:54
Air Blank		0.000	23:55
Air Blank		0.000	23:56
Subject Sample	# 2	0.116	23:57
Breath Volume		2.949	23:57
Air Blank		0.000	23:58
Control Test		0.079	23:58
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 06/07/2018 23:06

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: BUCHANAN

DATE/TIME OF TEST: 05/13/2018 02:58:32

Diagnostics Check		OK	02:55	
Air Blank		0.000	02:56	
Control Test		0.079	02:56	
Air Blank		0.000	02:56	
Subject Sample	# 1		02:57	Subject Test Refused
Breath Volume		0.000	02:57	
Air Blank		0.000	02:57	
Control Test		0.077	02:57	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 06/07/2018 23:06

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: BUCHANAN

DATE/TIME OF TEST: 05/13/2018 03:07:03

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.078	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.133	03:02
Breath Volume		1.683	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:04
Subject Sample	# 2	0.141	03:05
Breath Volume		2.140	03:05
Air Blank		0.000	03:06
Control Test		0.080	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:07

Data Download Date/Time: 06/07/2018 23:06

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MATHEWS

DATE/TIME OF TEST: 05/19/2018 01:09:33

Diagnostics Check		OK	01:03
Air Blank		0.000	01:03
Control Test		0.081	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.142	01:04
Breath Volume		2.832	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.146	01:08
Breath Volume		3.062	01:08
Air Blank		0.000	01:08
Control Test		0.076	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 06/07/2018 23:06

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 05/19/2018 03:18:15

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.078	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.169	03:12
Breath Volume		2.015	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.166	03:16
Breath Volume		1.531	03:16
Air Blank		0.000	03:17
Control Test		0.078	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 06/07/2018 23:06

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MEEKS

DATE/TIME OF TEST: 05/30/2018 21:35:51

Diagnostics Check		OK	21:32	
Air Blank		0.000	21:32	
Control Test		0.079	21:33	
Air Blank		0.000	21:33	
Subject Sample	# 1		21:34	Interferent Detect
Breath Volume		2.968	21:34	
Air Blank		0.000	21:34	
Control Test		0.078	21:35	
Air Blank		0.000	21:35	
Diagnostics Check		OK	21:35	

Data Download Date/Time: 06/07/2018 23:06

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MEEKS

DATE/TIME OF TEST: 05/30/2018 21:45:05

Diagnostics Check		OK	21:38
Air Blank		0.000	21:39
Control Test		0.078	21:39
Air Blank		0.000	21:39
Subject Sample	# 1	0.105	21:40
Breath Volume		2.613	21:40
Air Blank		0.000	21:41
Air Blank		0.000	21:43
Subject Sample	# 2	0.096	21:43
Breath Volume		2.472	21:43
Air Blank		0.000	21:44
Control Test		0.079	21:44
Air Blank		0.000	21:44
Diagnostics Check		OK	21:45

Data Download Date/Time: 06/07/2018 23:06

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 05/31/2018 03:00:10

Diagnostics Check		OK	02:45	
Air Blank		0.000	02:46	
Control Test		0.078	02:46	
Air Blank		0.000	02:46	
Subject Sample	# 1	0.162	02:49	
Breath Volume		1.929	02:49	
Air Blank		0.000	02:50	
Air Blank		0.000	02:52	
Subject Sample	# 2	0.124	02:55	Volume Not Met
Breath Volume		0.703	02:55	
Air Blank		0.000	02:55	
Air Blank		0.000	02:57	
Subject Sample	# 3	0.171	02:58	
Breath Volume		2.667	02:58	
Air Blank		0.000	02:59	
Control Test		0.077	02:59	
Air Blank		0.000	03:00	
Diagnostics Check		OK	03:00	

Data Download Date/Time: 06/07/2018 23:06

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTZ,KATHRYN,N
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: GOTZ,KATHRYN
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: PALLON

DATE/TIME OF TEST: 05/04/2018 02:23:51

Diagnostics Check	OK	02:23	
Air Blank		02:23	RFI Detect
Air Blank	0.000	02:23	

Data Download Date/Time: 05/22/2018 05:24

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTZ,KATHRYN,N
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: GOTZ,KATHRYN
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: PALLON

DATE/TIME OF TEST: 05/04/2018 02:31:33

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.080	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.000	02:27
Breath Volume		1.796	02:27
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.000	02:30
Breath Volume		2.566	02:30
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 05/22/2018 05:24

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: CASTELLANO,VINCENT
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 04/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 05/18/2018 23:34:23

Diagnostics Check		OK	23:23
Air Blank		0.000	23:24
Control Test		0.079	23:24
Air Blank		0.000	23:25
Subject Sample	# 1	0.237	23:25
Breath Volume		1.777	23:25
Air Blank		0.000	23:26
Air Blank		0.000	23:28
Subject Sample	# 2	0.208	23:29
Breath Volume		3.113	23:29
Air Blank		0.000	23:30
Air Blank		0.000	23:32
Subject Sample	# 3	0.200	23:32
Breath Volume		2.714	23:32
Air Blank		0.000	23:33
Control Test		0.078	23:33
Air Blank		0.000	23:34
Diagnostics Check		OK	23:34

Data Download Date/Time: 05/22/2018 05:24

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDEN,ALTON,G
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: OGDEN,ALTON
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/26/2018 23:59:33

Diagnostics Check		OK	23:47	
Air Blank		0.000	23:48	
Control Test		0.080	23:48	
Air Blank		0.000	23:48	
Subject Sample	# 1		23:52	No Sample Provided
Breath Volume		0.000	23:52	
Air Blank		0.000	23:52	
Air Blank		0.000	23:54	
Subject Sample	# 2	0.220	23:57	Slope Not Level
Breath Volume		1.195	23:57	
Air Blank		0.000	23:58	
Control Test		0.078	23:58	
Air Blank		0.000	23:59	
Diagnostics Check		OK	23:59	

Data Download Date/Time: 06/20/2018 02:40

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DANIEL,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: OGDEN,ALTON
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/27/2018 00:08:52

Diagnostics Check		OK	00:06	
Air Blank		0.000	00:06	
Control Test		0.077	00:06	
Air Blank		0.000	00:07	
Subject Sample	# 1		00:07	Subject Test Refused
Breath Volume		0.000	00:07	
Air Blank		0.000	00:07	
Control Test		0.080	00:08	
Air Blank		0.000	00:08	
Diagnostics Check		OK	00:08	

Data Download Date/Time: 06/20/2018 02:40

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GINTZ,BRIAN,W
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: HARRIS,BILLIE
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 765462
DRY GAS EXPIRY DATE: 04/13/2020
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 05/18/2018 19:25:24

Diagnostics Check		OK	19:22	
Air Blank		0.000	19:23	
Control Test		0.078	19:23	
Air Blank		0.000	19:23	
Subject Sample	# 1		19:24	Subject Test Refused
Breath Volume		0.000	19:24	
Air Blank		0.000	19:24	
Control Test		0.078	19:24	
Air Blank		0.000	19:25	
Diagnostics Check		OK	19:25	

Data Download Date/Time: 05/30/2018 23:28

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GABOURY,SCOTT,E
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: EPITROPOULOS,SAMUAL
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 765462
DRY GAS EXPIRY DATE: 04/13/2020
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 05/27/2018 19:34:36

Diagnostics Check		OK	19:27
Air Blank		0.000	19:28
Control Test		0.079	19:28
Air Blank		0.000	19:29
Subject Sample	# 1	0.324	19:29
Breath Volume		1.988	19:29
Air Blank		0.000	19:30
Air Blank		0.000	19:32
Subject Sample	# 2	0.309	19:32
Breath Volume		2.160	19:32
Air Blank		0.000	19:33
Control Test		0.078	19:33
Air Blank		0.000	19:34
Diagnostics Check		OK	19:34

Data Download Date/Time: 05/30/2018 23:28

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANDESE,ANGELO,
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MIALKI,ANDREW
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 13:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:36

SUBJECT'S LAST NAME: BOHL

DATE/TIME OF TEST: 05/09/2018 14:08:55

Diagnostics Check		OK	14:02
Air Blank		0.000	14:03
Control Test		0.081	14:03
Air Blank		0.000	14:03
Subject Sample	# 1	0.010	14:04
Breath Volume		3.171	14:04
Air Blank		0.000	14:04
Air Blank		0.000	14:06
Subject Sample	# 2	0.009	14:07
Breath Volume		2.507	14:07
Air Blank		0.000	14:08
Control Test		0.081	14:08
Air Blank		0.000	14:08
Diagnostics Check		OK	14:08

Data Download Date/Time: 05/31/2018 15:47

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURTZ,THEODORE,F
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: ALAMOS,JAMIE
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: CAIN

DATE/TIME OF TEST: 05/16/2018 23:07:30

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.081	23:02
Air Blank		0.000	23:02
Subject Sample	# 1	0.000	23:03
Breath Volume		2.839	23:03
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.000	23:06
Breath Volume		2.140	23:06
Air Blank		0.000	23:06
Control Test		0.082	23:06
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 05/31/2018 15:47

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIRD,JEREMY,E
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MIALKI,ANDREW
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: POWLES

DATE/TIME OF TEST: 05/26/2018 01:39:47

Diagnostics Check		OK	01:33
Air Blank		0.000	01:34
Control Test		0.081	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.170	01:35
Breath Volume		4.714	01:35
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.165	01:38
Breath Volume		2.707	01:38
Air Blank		0.000	01:38
Control Test		0.080	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 05/31/2018 15:47

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIRD,JEREMY,E
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: SCHNIBBEN,BRIAN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: LOPES

DATE/TIME OF TEST: 05/26/2018 01:51:51

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.081	01:46
Air Blank		0.000	01:46
Subject Sample	# 1	0.128	01:47
Breath Volume		2.304	01:47
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.128	01:50
Breath Volume		2.410	01:50
Air Blank		0.000	01:50
Control Test		0.080	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 05/31/2018 15:47

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAIG,BRYAN,D
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: SANDERSON,TAYLOR
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 04/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 894572
DRY GAS EXPIRY DATE: 01/17/2021
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: SHELTON

DATE/TIME OF TEST: 05/01/2018 22:31:32

Diagnostics Check		OK	22:28	
Air Blank		0.000	22:29	
Control Test		0.082	22:29	
Air Blank		0.000	22:29	
Subject Sample	# 1		22:30	Subject Test Refused
Breath Volume		0.000	22:30	
Air Blank		0.000	22:30	
Control Test		0.082	22:30	
Air Blank		0.000	22:31	
Diagnostics Check		OK	22:31	

Data Download Date/Time: 05/17/2018 23:06

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: AKERS,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: DRAKE

DATE/TIME OF TEST: 05/11/2018 19:42:54

Diagnostics Check		OK	19:36
Air Blank		0.000	19:36
Control Test		0.080	19:37
Air Blank		0.000	19:37
Subject Sample	# 1	0.194	19:38
Breath Volume		1.855	19:38
Air Blank		0.000	19:39
Air Blank		0.000	19:40
Subject Sample	# 2	0.185	19:41
Breath Volume		1.789	19:41
Air Blank		0.000	19:42
Control Test		0.083	19:42
Air Blank		0.000	19:42
Diagnostics Check		OK	19:42

Data Download Date/Time: 06/21/2018 12:32

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SGT BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: LOPEZ HERNANDEZ

DATE/TIME OF TEST: 05/12/2018 02:11:39

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.080	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.237	02:07
Breath Volume		1.980	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.255	02:10
Breath Volume		3.640	02:10
Air Blank		0.000	02:10
Control Test		0.082	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 06/21/2018 12:32

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: CRISTO

DATE/TIME OF TEST: 05/12/2018 04:23:39

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:16	
Control Test		0.080	04:16	
Air Blank		0.000	04:16	
Subject Sample	# 1	0.118	04:17	
Breath Volume		1.289	04:17	
Air Blank		0.000	04:18	
Air Blank		0.000	04:19	
Subject Sample	# 2	0.112	04:22	
Breath Volume		1.484	04:22	
Air Blank		0.000	04:22	
Control Test		0.086	04:23	Control Outside Tolerance
Air Blank		0.000	04:23	

Data Download Date/Time: 06/21/2018 12:32

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: CRISTO

DATE/TIME OF TEST: 05/12/2018 04:32:31

Diagnostics Check		OK	04:25
Air Blank		0.000	04:26
Control Test		0.080	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.119	04:27
Breath Volume		2.246	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.124	04:31
Breath Volume		2.546	04:31
Air Blank		0.000	04:31
Control Test		0.083	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 06/21/2018 12:32

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: FLUITT

DATE/TIME OF TEST: 05/12/2018 05:55:52

Diagnostics Check		OK	05:49
Air Blank		0.000	05:49
Control Test		0.081	05:50
Air Blank		0.000	05:50
Subject Sample	# 1	0.134	05:51
Breath Volume		1.757	05:51
Air Blank		0.000	05:52
Air Blank		0.000	05:53
Subject Sample	# 2	0.139	05:54
Breath Volume		2.375	05:54
Air Blank		0.000	05:54
Control Test		0.080	05:55
Air Blank		0.000	05:55
Diagnostics Check		OK	05:55

Data Download Date/Time: 06/21/2018 12:32

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:11

SUBJECT'S LAST NAME: GRINNELL

DATE/TIME OF TEST: 05/12/2018 06:50:26

Diagnostics Check		OK	06:41
Air Blank		0.000	06:42
Control Test		0.080	06:42
Air Blank		0.000	06:43
Subject Sample	# 1	0.149	06:44
Breath Volume		1.664	06:44
Air Blank		0.000	06:44
Air Blank		0.000	06:46
Subject Sample	# 2	0.146	06:48
Breath Volume		1.609	06:48
Air Blank		0.000	06:49
Control Test		0.085	06:49
Air Blank		0.000	06:50
Diagnostics Check		OK	06:50

Data Download Date/Time: 06/21/2018 12:32

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 06:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 05/12/2018 07:14:52

Diagnostics Check		OK	07:05
Air Blank		0.000	07:06
Control Test		0.081	07:06
Air Blank		0.000	07:07
Subject Sample	# 1	0.115	07:08
Breath Volume		1.527	07:08
Air Blank		0.000	07:09
Air Blank		0.000	07:11
Subject Sample	# 2	0.112	07:13
Breath Volume		1.566	07:13
Air Blank		0.000	07:13
Control Test		0.084	07:14
Air Blank		0.000	07:14
Diagnostics Check		OK	07:14

Data Download Date/Time: 06/21/2018 12:32

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: DESTEFANO

DATE/TIME OF TEST: 05/13/2018 02:55:08

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:48	
Control Test		0.080	02:48	
Air Blank		0.000	02:48	
Subject Sample	# 1	0.141	02:49	
Breath Volume		1.378	02:49	
Air Blank		0.000	02:50	
Air Blank		0.000	02:52	
Subject Sample	# 2	0.141	02:53	
Breath Volume		2.562	02:53	
Air Blank		0.000	02:54	
Control Test		0.086	02:54	Control Outside Tolerance
Air Blank		0.000	02:55	

Data Download Date/Time: 06/21/2018 12:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: DESTEFANO

DATE/TIME OF TEST: 05/13/2018 03:05:59

Diagnostics Check		OK	02:57
Air Blank		0.000	02:58
Control Test		0.080	02:58
Air Blank		0.000	02:58
Subject Sample	# 1	0.139	03:00
Breath Volume		1.785	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:02
Subject Sample	# 2	0.137	03:04
Breath Volume		1.503	03:04
Air Blank		0.000	03:05
Control Test		0.085	03:05
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 06/21/2018 12:32

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRITY,THOMAS,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: OUELLETTE,SARAH
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/27/2018 03:27:02

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.081	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.107	03:22
Breath Volume		1.929	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.104	03:25
Breath Volume		1.660	03:25
Air Blank		0.000	03:26
Control Test		0.080	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:27

Data Download Date/Time: 06/13/2018 12:43

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRITY,THOMAS,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GARRITY,THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 05/31/2018 04:05:41

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:53	
Control Test		0.079	03:54	
Air Blank		0.000	03:54	
Subject Sample	# 1	0.144	03:57	Volume Not Met
Breath Volume		1.097	03:57	
Air Blank		0.000	03:58	
Air Blank		0.000	04:00	
Subject Sample	# 2	0.147	04:01	
Breath Volume		1.632	04:01	
Air Blank		0.000	04:01	
Air Blank		0.000	04:03	
Subject Sample	# 3	0.152	04:04	
Breath Volume		1.480	04:04	
Air Blank		0.000	04:04	
Control Test		0.080	04:05	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:05	

Data Download Date/Time: 06/13/2018 12:43

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROBBINS,NATHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 05/03/2018 03:00:32

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.083	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.118	02:55
Breath Volume		3.437	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.116	02:58
Breath Volume		3.832	02:58
Air Blank		0.000	02:59
Control Test		0.081	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 05/03/2018 16:39

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALMONTE,JOSE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: GARITY

DATE/TIME OF TEST: 05/05/2018 01:25:43

Diagnostics Check		OK	01:19
Air Blank		0.000	01:19
Control Test		0.083	01:20
Air Blank		0.000	01:20
Subject Sample	# 1	0.250	01:21
Breath Volume		1.750	01:21
Air Blank		0.000	01:22
Air Blank		0.000	01:23
Subject Sample	# 2	0.250	01:24
Breath Volume		1.546	01:24
Air Blank		0.000	01:24
Control Test		0.081	01:25
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 05/08/2018 10:03

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BORRAS,TAMARA
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: CASTILLOESPANA

DATE/TIME OF TEST: 05/05/2018 04:03:19

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.081	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.160	03:58
Breath Volume		1.679	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.167	04:01
Breath Volume		2.679	04:01
Air Blank		0.000	04:02
Control Test		0.081	04:02
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 05/08/2018 10:03

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WING,RYAN
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 05/05/2018 04:50:49

Diagnostics Check		OK	04:43
Air Blank		0.000	04:44
Control Test		0.082	04:44
Air Blank		0.000	04:45
Subject Sample	# 1	0.107	04:45
Breath Volume		1.234	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.121	04:49
Breath Volume		2.156	04:49
Air Blank		0.000	04:49
Control Test		0.081	04:50
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 05/08/2018 10:03

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ODEA,PATRICIA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:47

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 05/05/2018 07:26:12

Diagnostics Check		OK	07:19
Air Blank		0.000	07:19
Control Test		0.082	07:20
Air Blank		0.000	07:20
Subject Sample	# 1	0.173	07:21
Breath Volume		1.503	07:21
Air Blank		0.000	07:22
Air Blank		0.000	07:24
Subject Sample	# 2	0.177	07:24
Breath Volume		1.273	07:24
Air Blank		0.000	07:25
Control Test		0.080	07:25
Air Blank		0.000	07:26
Diagnostics Check		OK	07:26

Data Download Date/Time: 05/08/2018 10:03

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PEREZ,JUAN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/05/2018 23:49:19

Diagnostics Check		OK	23:42
Air Blank		0.000	23:43
Control Test		0.082	23:43
Air Blank		0.000	23:43
Subject Sample	# 1	0.164	23:44
Breath Volume		3.113	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.160	23:47
Breath Volume		2.843	23:47
Air Blank		0.000	23:48
Control Test		0.082	23:48
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 05/08/2018 10:03

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BAKER,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: DENOR

DATE/TIME OF TEST: 05/06/2018 02:48:04

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.081	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.224	02:43
Breath Volume		1.421	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.218	02:46
Breath Volume		1.457	02:46
Air Blank		0.000	02:47
Control Test		0.081	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:48

Data Download Date/Time: 05/08/2018 10:03

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORALES,JOSE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 05/06/2018 03:44:25

Diagnostics Check		OK	03:38
Air Blank		0.000	03:38
Control Test		0.082	03:38
Air Blank		0.000	03:39
Subject Sample	# 1	0.175	03:39
Breath Volume		1.542	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.158	03:42
Breath Volume		1.566	03:42
Air Blank		0.000	03:43
Control Test		0.080	03:43
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 05/08/2018 10:03

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PHILLIP,NOVEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: COTTY SANCHEZ

DATE/TIME OF TEST: 05/06/2018 05:10:49

Diagnostics Check		OK	04:58	
Air Blank		0.000	04:59	
Control Test		0.081	04:59	
Air Blank		0.000	05:00	
Subject Sample	# 1	0.153	05:03	Volume Not Met
Breath Volume		0.031	05:03	
Air Blank		0.000	05:03	
Air Blank		0.000	05:05	
Subject Sample	# 2	0.169	05:06	
Breath Volume		1.640	05:06	
Air Blank		0.000	05:07	
Air Blank		0.000	05:08	
Subject Sample	# 3	0.171	05:09	
Breath Volume		1.324	05:09	
Air Blank		0.000	05:09	
Control Test		0.081	05:10	
Air Blank		0.000	05:10	
Diagnostics Check		OK	05:10	

Data Download Date/Time: 05/08/2018 10:03

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 05/18/2018 03:01:13

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.081	02:55
Air Blank		0.000	02:55
Subject Sample	# 1	0.133	02:56
Breath Volume		3.957	02:56
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.127	02:59
Breath Volume		4.101	02:59
Air Blank		0.000	03:00
Control Test		0.078	03:00
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 06/07/2018 14:58

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILKES,DAVID
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 05/18/2018 05:15:04

Diagnostics Check		OK	05:07
Air Blank		0.000	05:07
Control Test		0.081	05:08
Air Blank		0.000	05:08
Subject Sample	# 1	0.164	05:09
Breath Volume		1.335	05:09
Air Blank		0.000	05:10
Air Blank		0.000	05:12
Subject Sample	# 2	0.164	05:13
Breath Volume		1.460	05:13
Air Blank		0.000	05:14
Control Test		0.082	05:14
Air Blank		0.000	05:14
Diagnostics Check		OK	05:15

Data Download Date/Time: 06/07/2018 14:58

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AUSTIN,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: SCHULTZ

DATE/TIME OF TEST: 05/20/2018 03:24:45

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.082	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.185	03:20
Breath Volume		1.679	03:20
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.188	03:23
Breath Volume		1.433	03:23
Air Blank		0.000	03:23
Control Test		0.082	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 06/07/2018 14:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN,JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 05/20/2018 06:22:38

Diagnostics Check		OK	06:16
Air Blank		0.000	06:16
Control Test		0.082	06:16
Air Blank		0.000	06:17
Subject Sample	# 1	0.000	06:18
Breath Volume		1.210	06:18
Air Blank		0.000	06:18
Air Blank		0.000	06:20
Subject Sample	# 2	0.000	06:21
Breath Volume		1.746	06:21
Air Blank		0.000	06:21
Control Test		0.080	06:22
Air Blank		0.000	06:22
Diagnostics Check		OK	06:22

Data Download Date/Time: 06/07/2018 14:58

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VORHEIS,JUSTIN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: BRAVO VELASQUEZ

DATE/TIME OF TEST: 05/20/2018 23:42:58

Diagnostics Check		OK	23:35
Air Blank		0.000	23:36
Control Test		0.082	23:36
Air Blank		0.000	23:37
Subject Sample	# 1	0.319	23:38
Breath Volume		2.546	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:40
Subject Sample	# 2	0.313	23:41
Breath Volume		2.011	23:41
Air Blank		0.000	23:42
Control Test		0.081	23:42
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 06/07/2018 14:58

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KANE,ALEXANDER
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: NIETO

DATE/TIME OF TEST: 05/21/2018 04:23:15

Diagnostics Check		OK	04:14
Air Blank		0.000	04:15
Control Test		0.083	04:15
Air Blank		0.000	04:16
Subject Sample	# 1	0.133	04:18
Breath Volume		1.156	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:20
Subject Sample	# 2	0.153	04:21
Breath Volume		1.175	04:21
Air Blank		0.000	04:22
Control Test		0.081	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 06/07/2018 14:58

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 05/24/2018 04:13:29

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.082	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.109	04:08
Breath Volume		1.675	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:10
Subject Sample	# 2	0.107	04:11
Breath Volume		1.429	04:11
Air Blank		0.000	04:12
Control Test		0.082	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 06/07/2018 14:58

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: STEPHENSON,AARON
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/27/2018 01:44:29

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.082	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.126	01:39
Breath Volume		1.425	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:41
Subject Sample	# 2	0.126	01:42
Breath Volume		1.648	01:42
Air Blank		0.000	01:43
Control Test		0.081	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 06/07/2018 14:58

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEAVER,WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 05/27/2018 05:52:52

Diagnostics Check		OK	05:46
Air Blank		0.000	05:46
Control Test		0.083	05:46
Air Blank		0.000	05:47
Subject Sample	# 1	0.169	05:48
Breath Volume		3.027	05:48
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.174	05:51
Breath Volume		3.085	05:51
Air Blank		0.000	05:51
Control Test		0.081	05:52
Air Blank		0.000	05:52
Diagnostics Check		OK	05:52

Data Download Date/Time: 06/07/2018 14:58

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ZOLY,NICHOLAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:18

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 05/27/2018 07:51:07

Diagnostics Check		OK	07:42
Air Blank		0.000	07:42
Control Test		0.083	07:43
Air Blank		0.000	07:43
Subject Sample	# 1	0.122	07:44
Breath Volume		1.355	07:44
Air Blank		0.000	07:45
Air Blank		0.000	07:47
Subject Sample	# 2	0.116	07:49
Breath Volume		1.121	07:49
Air Blank		0.000	07:50
Control Test		0.080	07:50
Air Blank		0.000	07:51
Diagnostics Check		OK	07:51

Data Download Date/Time: 06/07/2018 14:58

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARIANO,SEBASTIAN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: PEREZ GARCIA

DATE/TIME OF TEST: 05/27/2018 23:21:16

Diagnostics Check		OK	23:13
Air Blank		0.000	23:13
Control Test		0.082	23:13
Air Blank		0.000	23:14
Subject Sample	# 1	0.245	23:15
Breath Volume		2.105	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:17
Subject Sample	# 2	0.236	23:19
Breath Volume		1.773	23:19
Air Blank		0.000	23:20
Control Test		0.080	23:20
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 06/07/2018 14:58

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TRANSULE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/28/2018 05:22:30

Diagnostics Check		OK	05:14
Air Blank		0.000	05:14
Control Test		0.082	05:15
Air Blank		0.000	05:15
Subject Sample	# 1	0.269	05:17
Breath Volume		1.578	05:17
Air Blank		0.000	05:18
Air Blank		0.000	05:19
Subject Sample	# 2	0.262	05:20
Breath Volume		1.457	05:20
Air Blank		0.000	05:21
Control Test		0.078	05:21
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 06/07/2018 14:58

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LANDIN,CARLOS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: VAZQUEZ LOPEZ

DATE/TIME OF TEST: 05/29/2018 02:01:16

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.081	01:55
Air Blank		0.000	01:55
Subject Sample	# 1	0.190	01:56
Breath Volume		1.132	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.204	01:59
Breath Volume		1.253	01:59
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 06/07/2018 14:58

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORALES,JOSE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 05/30/2018 03:32:22

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.083	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.138	03:27
Breath Volume		4.261	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.141	03:30
Breath Volume		4.542	03:30
Air Blank		0.000	03:31
Control Test		0.081	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 06/07/2018 14:58

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORALES,JOSE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: GRINGERI

DATE/TIME OF TEST: 05/01/2018 02:08:15

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.080	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.178	02:03
Breath Volume		1.300	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.170	02:06
Breath Volume		1.398	02:06
Air Blank		0.000	02:07
Control Test		0.080	02:07
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 05/03/2018 16:55

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MENDOZA,ANGEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: MCCLINTON

DATE/TIME OF TEST: 05/02/2018 04:44:28

Diagnostics Check		OK	04:31	
Air Blank		0.000	04:32	
Control Test		0.080	04:32	
Air Blank		0.000	04:32	
Subject Sample	# 1	0.106	04:36	Volume Not Met
Breath Volume		0.312	04:36	
Air Blank		0.000	04:36	
Air Blank		0.000	04:38	
Subject Sample	# 2	0.126	04:39	
Breath Volume		2.058	04:39	
Air Blank		0.000	04:40	
Air Blank		0.000	04:42	
Subject Sample	# 3	0.118	04:43	
Breath Volume		1.113	04:43	
Air Blank		0.000	04:43	
Control Test		0.080	04:43	
Air Blank		0.000	04:44	
Diagnostics Check		OK	04:44	

Data Download Date/Time: 05/03/2018 16:55

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROBBINS,NATHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: KITCHELL

DATE/TIME OF TEST: 05/06/2018 02:42:44

Diagnostics Check		OK	02:36
Air Blank		0.000	02:37
Control Test		0.081	02:37
Air Blank		0.000	02:37
Subject Sample	# 1	0.249	02:38
Breath Volume		1.804	02:38
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.251	02:41
Breath Volume		1.699	02:41
Air Blank		0.000	02:41
Control Test		0.080	02:42
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 05/08/2018 12:34

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPENCER,JOHN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HAMMOND

DATE/TIME OF TEST: 05/06/2018 03:22:56

Diagnostics Check		OK	03:11	
Air Blank		0.000	03:12	
Control Test		0.078	03:12	
Air Blank		0.000	03:12	
Subject Sample	# 1		03:15	No Sample Provided
Breath Volume		0.000	03:15	
Air Blank		0.000	03:16	
Air Blank		0.000	03:18	
Subject Sample	# 2		03:21	No Sample Provided
Breath Volume		0.000	03:21	
Air Blank		0.000	03:22	
Control Test		0.080	03:22	
Air Blank		0.000	03:22	
Diagnostics Check		OK	03:22	

Data Download Date/Time: 05/08/2018 12:34

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HERNANDEZ,EDUARDO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: RAYMUNDO LOPEZ

DATE/TIME OF TEST: 05/07/2018 00:52:07

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.080	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.072	00:47
Breath Volume		1.296	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:49
Subject Sample	# 2	0.069	00:50
Breath Volume		1.925	00:50
Air Blank		0.000	00:51
Control Test		0.081	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 05/08/2018 12:34

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KREKEL,ALEC
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: SILER

DATE/TIME OF TEST: 05/07/2018 03:45:36

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.081	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.194	03:41
Breath Volume		1.398	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.203	03:44
Breath Volume		1.605	03:44
Air Blank		0.000	03:44
Control Test		0.081	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 05/08/2018 12:34

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LANDIN,CARLOS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: COSME

DATE/TIME OF TEST: 05/10/2018 01:04:24

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.080	00:59
Air Blank		0.000	00:59
Subject Sample	# 1	0.273	01:00
Breath Volume		2.800	01:00
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.272	01:02
Breath Volume		2.292	01:02
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 05/17/2018 11:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORALES,JOSE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 05/11/2018 02:58:05

Diagnostics Check		OK	02:51
Air Blank		0.000	02:52
Control Test		0.081	02:52
Air Blank		0.000	02:53
Subject Sample	# 1	0.098	02:53
Breath Volume		2.898	02:53
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.098	02:56
Breath Volume		3.226	02:56
Air Blank		0.000	02:57
Control Test		0.080	02:57
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 05/17/2018 11:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MENDOZA,ANGEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: SANTA GONZALEZ

DATE/TIME OF TEST: 05/11/2018 04:52:07

Diagnostics Check		OK	04:45
Air Blank		0.000	04:45
Control Test		0.081	04:46
Air Blank		0.000	04:46
Subject Sample	# 1	0.101	04:47
Breath Volume		1.160	04:47
Air Blank		0.000	04:48
Air Blank		0.000	04:50
Subject Sample	# 2	0.105	04:50
Breath Volume		1.519	04:50
Air Blank		0.000	04:51
Control Test		0.080	04:51
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 05/17/2018 11:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL,TALMON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: PFLANZ

DATE/TIME OF TEST: 05/12/2018 03:17:40

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.081	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.233	03:13
Breath Volume		2.910	03:13
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.237	03:16
Breath Volume		3.429	03:16
Air Blank		0.000	03:16
Control Test		0.079	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 05/17/2018 11:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,LIBORIO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 05/14/2018 22:41:46

Diagnostics Check		OK	22:35
Air Blank		0.000	22:35
Control Test		0.081	22:36
Air Blank		0.000	22:36
Subject Sample	# 1	0.186	22:37
Breath Volume		3.703	22:37
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.189	22:40
Breath Volume		3.390	22:40
Air Blank		0.000	22:40
Control Test		0.080	22:41
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 05/17/2018 11:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: FERNANDES VIEIRA PED

DATE/TIME OF TEST: 05/18/2018 03:51:47

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.081	03:46
Air Blank		0.000	03:47
Subject Sample	# 1	0.122	03:47
Breath Volume		1.328	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:49
Subject Sample	# 2	0.116	03:50
Breath Volume		1.105	03:50
Air Blank		0.000	03:50
Control Test		0.081	03:51
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 06/07/2018 14:50

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FERRARA,DAWN
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 05/19/2018 03:22:16

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.082	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.048	03:17
Breath Volume		1.886	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.044	03:20
Breath Volume		1.957	03:20
Air Blank		0.000	03:21
Control Test		0.081	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 06/07/2018 14:50

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 05/20/2018 02:19:51

Diagnostics Check		OK	02:03	
Air Blank		0.000	02:03	
Control Test		0.081	02:03	
Air Blank		0.000	02:04	
Subject Sample	# 1	0.161	02:07	Volume Not Met
Breath Volume		0.035	02:07	
Air Blank		0.000	02:08	
Air Blank		0.000	02:10	
Subject Sample	# 2	0.162	02:13	
Breath Volume		1.687	02:13	
Air Blank		0.000	02:13	
Air Blank		0.000	02:15	
Subject Sample	# 3	0.162	02:18	
Breath Volume		2.328	02:18	
Air Blank		0.000	02:18	
Control Test		0.079	02:19	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 06/07/2018 14:50

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HANSTINE,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: CHABLE GIL

DATE/TIME OF TEST: 05/20/2018 04:51:30

Diagnostics Check		OK	04:44
Air Blank		0.000	04:45
Control Test		0.080	04:45
Air Blank		0.000	04:46
Subject Sample	# 1	0.207	04:46
Breath Volume		1.214	04:46
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.227	04:50
Breath Volume		2.707	04:50
Air Blank		0.000	04:50
Control Test		0.080	04:50
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 06/07/2018 14:50

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILKES,DAVID
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 05/23/2018 04:01:26

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:47	
Control Test		0.080	03:47	
Air Blank		0.000	03:47	
Subject Sample	# 1	0.137	03:51	Volume Not Met
Breath Volume		0.402	03:51	
Air Blank		0.000	03:51	
Air Blank		0.000	03:53	
Subject Sample	# 2	0.183	03:54	
Breath Volume		1.441	03:54	
Air Blank		0.000	03:54	
Air Blank		0.000	03:56	
Subject Sample	# 3	0.158	03:59	Volume Not Met
Breath Volume		1.070	03:59	
Air Blank		0.000	04:00	
Control Test		0.080	04:00	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 06/07/2018 14:50

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILKES,DAVID
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 05/23/2018 04:09:09

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.080	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.175	04:04
Breath Volume		1.628	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.178	04:07
Breath Volume		1.640	04:07
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 06/07/2018 14:50

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: POWELL,KRISTI
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: HEPBURN

DATE/TIME OF TEST: 05/25/2018 05:39:31

Diagnostics Check		OK	05:32
Air Blank		0.000	05:33
Control Test		0.082	05:33
Air Blank		0.000	05:33
Subject Sample	# 1	0.184	05:35
Breath Volume		1.207	05:35
Air Blank		0.000	05:35
Air Blank		0.000	05:37
Subject Sample	# 2	0.196	05:38
Breath Volume		1.226	05:38
Air Blank		0.000	05:38
Control Test		0.081	05:38
Air Blank		0.000	05:39
Diagnostics Check		OK	05:39

Data Download Date/Time: 06/07/2018 14:50

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: DANISH

DATE/TIME OF TEST: 05/25/2018 22:23:24

Diagnostics Check		OK	22:17
Air Blank		0.000	22:17
Control Test		0.081	22:18
Air Blank		0.000	22:18
Subject Sample	# 1	0.000	22:19
Breath Volume		2.441	22:19
Air Blank		0.000	22:19
Air Blank		0.000	22:21
Subject Sample	# 2	0.000	22:22
Breath Volume		2.468	22:22
Air Blank		0.000	22:22
Control Test		0.081	22:22
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 06/07/2018 14:50

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALEXANDER,MICHAEL
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: GARCIA OCASIO

DATE/TIME OF TEST: 05/26/2018 02:57:50

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.080	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.168	02:52
Breath Volume		1.457	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:54
Subject Sample	# 2	0.168	02:56
Breath Volume		1.589	02:56
Air Blank		0.000	02:56
Control Test		0.080	02:57
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 06/07/2018 14:50

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 05/26/2018 03:53:09

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.081	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.114	03:48
Breath Volume		1.835	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.110	03:51
Breath Volume		1.734	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 06/07/2018 14:50

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LUCIANO PORRAS

DATE/TIME OF TEST: 05/28/2018 00:16:30

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.081	00:07
Air Blank		0.000	00:08
Subject Sample	# 1	0.170	00:10
Breath Volume		2.113	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.159	00:15
Breath Volume		1.859	00:15
Air Blank		0.000	00:15
Control Test		0.080	00:15
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 06/07/2018 14:50

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/29/2018 01:37:34

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.081	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.000	01:33
Breath Volume		1.859	01:33
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.000	01:36
Breath Volume		1.691	01:36
Air Blank		0.000	01:36
Control Test		0.081	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 06/07/2018 14:50

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,CRAIG
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: MALDONADO-RIVERA

DATE/TIME OF TEST: 05/10/2018 02:10:09

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.080	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.131	02:05
Breath Volume		1.281	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.127	02:08
Breath Volume		2.605	02:08
Air Blank		0.000	02:09
Control Test		0.079	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 05/17/2018 11:22

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ALSING

DATE/TIME OF TEST: 05/11/2018 03:49:58

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.080	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.144	03:45
Breath Volume		2.515	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:47
Subject Sample	# 2	0.148	03:48
Breath Volume		3.140	03:48
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 05/17/2018 11:22

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,CRAIG
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: ZAPATA

DATE/TIME OF TEST: 05/11/2018 04:55:33

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.080	04:49
Air Blank		0.000	04:50
Subject Sample	# 1	0.209	04:51
Breath Volume		1.523	04:51
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.202	04:54
Breath Volume		1.832	04:54
Air Blank		0.000	04:54
Control Test		0.079	04:54
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 05/17/2018 11:22

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BROWN,QUINTEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 05/12/2018 04:52:32

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:52	
Control Test	0.079	04:52	RFI Detect
Air Blank		04:52	RFI Detect

Data Download Date/Time: 05/17/2018 11:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BROWN,QUINTEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 05/12/2018 05:01:09

Diagnostics Check		OK	04:54
Air Blank		0.000	04:55
Control Test		0.080	04:55
Air Blank		0.000	04:55
Subject Sample	# 1	0.191	04:56
Breath Volume		3.250	04:56
Air Blank		0.000	04:57
Air Blank		0.000	04:58
Subject Sample	# 2	0.184	04:59
Breath Volume		2.976	04:59
Air Blank		0.000	05:00
Control Test		0.080	05:00
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 05/17/2018 11:22

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 05/13/2018 05:08:52

Diagnostics Check		OK	05:06	
Air Blank		0.000	05:06	
Control Test		0.080	05:06	
Air Blank		0.000	05:07	
Subject Sample	# 1		05:07	Subject Test Refused
Breath Volume		0.000	05:07	
Air Blank		0.000	05:07	
Control Test		0.080	05:08	
Air Blank		0.000	05:08	
Diagnostics Check		OK	05:08	

Data Download Date/Time: 05/17/2018 11:22

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PARADIS,GEORGE
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 05/15/2018 22:51:14

Diagnostics Check		OK	22:44
Air Blank		0.000	22:45
Control Test		0.080	22:45
Air Blank		0.000	22:46
Subject Sample	# 1	0.192	22:46
Breath Volume		2.039	22:46
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.201	22:49
Breath Volume		1.402	22:49
Air Blank		0.000	22:50
Control Test		0.079	22:50
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 05/17/2018 11:22

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 05/16/2018 02:24:23

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.190	02:19
Breath Volume		1.121	02:19
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.181	02:22
Breath Volume		1.332	02:22
Air Blank		0.000	02:23
Control Test		0.079	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 05/17/2018 11:22

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JONES III,PAUL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: BRUNO

DATE/TIME OF TEST: 05/17/2018 02:33:39

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:31	
Control Test		0.080	02:31	
Air Blank		0.000	02:32	
Subject Sample	# 1		02:32	Subject Test Refused
Breath Volume		0.000	02:32	
Air Blank		0.000	02:32	
Control Test		0.079	02:33	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

Data Download Date/Time: 05/17/2018 11:22

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 05/19/2018 04:05:12

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:02	
Control Test		0.080	04:02	
Air Blank		0.000	04:03	
Subject Sample	# 1		04:03	Subject Test Refused
Breath Volume		0.000	04:03	
Air Blank		0.000	04:04	
Control Test		0.080	04:04	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:05	

Data Download Date/Time: 06/07/2018 14:53

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: ROSA-VIERA-FILHO

DATE/TIME OF TEST: 05/20/2018 02:44:40

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.080	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.112	02:39
Breath Volume		3.203	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.117	02:43
Breath Volume		3.324	02:43
Air Blank		0.000	02:43
Control Test		0.080	02:44
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 06/07/2018 14:53

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 05/20/2018 05:48:18

Diagnostics Check		OK	05:35	
Air Blank		0.000	05:35	
Control Test		0.080	05:35	
Air Blank		0.000	05:36	
Subject Sample	# 1	0.082	05:39	Volume Not Met
Breath Volume		0.128	05:39	
Air Blank		0.000	05:40	
Air Blank		0.000	05:42	
Subject Sample	# 2	0.084	05:42	
Breath Volume		1.746	05:42	
Air Blank		0.000	05:43	
Air Blank		0.000	05:45	
Subject Sample	# 3	0.079	05:46	
Breath Volume		1.605	05:46	
Air Blank		0.000	05:47	
Control Test		0.080	05:47	
Air Blank		0.000	05:48	
Diagnostics Check		OK	05:48	

Data Download Date/Time: 06/07/2018 14:53

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 05/23/2018 23:30:06

Diagnostics Check		OK	23:23
Air Blank		0.000	23:24
Control Test		0.080	23:24
Air Blank		0.000	23:25
Subject Sample	# 1	0.031	23:25
Breath Volume		2.871	23:25
Air Blank		0.000	23:26
Air Blank		0.000	23:28
Subject Sample	# 2	0.031	23:28
Breath Volume		2.320	23:28
Air Blank		0.000	23:29
Control Test		0.080	23:29
Air Blank		0.000	23:30
Diagnostics Check		OK	23:30

Data Download Date/Time: 06/07/2018 14:53

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,LIBORIO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: HALL JR

DATE/TIME OF TEST: 05/24/2018 23:18:50

Diagnostics Check		OK	23:16	
Air Blank		0.000	23:16	
Control Test		0.080	23:16	
Air Blank		0.000	23:17	
Subject Sample	# 1		23:17	Subject Test Refused
Breath Volume		0.000	23:17	
Air Blank		0.000	23:17	
Control Test		0.080	23:18	
Air Blank		0.000	23:18	
Diagnostics Check		OK	23:18	

Data Download Date/Time: 06/07/2018 14:53

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRINCE,BLAKE
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: GUTIERREZ-ROSAS

DATE/TIME OF TEST: 05/25/2018 22:37:01

Diagnostics Check		OK	22:33	
Air Blank		0.000	22:34	
Control Test		0.079	22:34	
Air Blank		0.000	22:35	
Subject Sample	# 1		22:35	Subject Test Refused
Breath Volume		0.000	22:35	
Air Blank		0.000	22:36	
Control Test		0.079	22:36	
Air Blank		0.000	22:36	
Diagnostics Check		OK	22:37	

Data Download Date/Time: 06/07/2018 14:53

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: GARCIA JR

DATE/TIME OF TEST: 05/26/2018 03:17:34

Diagnostics Check		OK	03:06	
Air Blank		0.000	03:06	
Control Test		0.080	03:06	
Air Blank		0.000	03:07	
Subject Sample	# 1		03:10	No Sample Provided
Breath Volume		0.000	03:10	
Air Blank		0.000	03:11	
Air Blank		0.000	03:13	
Subject Sample	# 2		03:16	No Sample Provided
Breath Volume		0.000	03:16	
Air Blank		0.000	03:16	
Control Test		0.080	03:16	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:17	

Data Download Date/Time: 06/07/2018 14:53

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN,JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: VELASCO GOMEZ

DATE/TIME OF TEST: 05/26/2018 03:58:58

Diagnostics Check		OK	03:52
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.117	03:54
Breath Volume		1.656	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:56
Subject Sample	# 2	0.106	03:57
Breath Volume		1.761	03:57
Air Blank		0.000	03:58
Control Test		0.079	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 06/07/2018 14:53

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MENDOZA,ANGEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: PIERRE FRANCOIS

DATE/TIME OF TEST: 05/30/2018 00:34:42

Diagnostics Check		OK	00:31	
Air Blank		0.000	00:32	
Control Test		0.080	00:32	
Air Blank		0.000	00:33	
Subject Sample	# 1		00:33	Subject Test Refused
Breath Volume		0.000	00:33	
Air Blank		0.000	00:33	
Control Test		0.080	00:34	
Air Blank		0.000	00:34	
Diagnostics Check		OK	00:34	

Data Download Date/Time: 06/07/2018 14:53

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: OSARNO,MONICA
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: ?

DATE/TIME OF TEST: 05/30/2018 03:46:22

Diagnostics Check		OK	03:43	
Air Blank		0.000	03:43	
Control Test		0.080	03:44	
Air Blank		0.000	03:44	
Subject Sample	# 1		03:45	Subject Test Refused
Breath Volume		0.000	03:45	
Air Blank		0.000	03:45	
Control Test		0.079	03:45	
Air Blank		0.000	03:46	
Diagnostics Check		OK	03:46	

Data Download Date/Time: 06/07/2018 14:53

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 05/31/2018 03:43:25

Diagnostics Check		OK	03:36
Air Blank		0.000	03:37
Control Test		0.079	03:37
Air Blank		0.000	03:38
Subject Sample	# 1	0.170	03:38
Breath Volume		1.171	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.169	03:41
Breath Volume		1.890	03:41
Air Blank		0.000	03:42
Control Test		0.079	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 06/07/2018 14:53

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: HERNANDEZ,L
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: CAPECE

DATE/TIME OF TEST: 05/03/2018 22:56:12

Diagnostics Check		OK	22:53	
Air Blank		0.000	22:53	
Control Test		0.081	22:54	
Air Blank		0.000	22:54	
Subject Sample	# 1		22:54	Subject Test Refused
Breath Volume		0.000	22:54	
Air Blank		0.000	22:55	
Control Test		0.080	22:55	
Air Blank		0.000	22:56	
Diagnostics Check		OK	22:56	

Data Download Date/Time: 05/26/2018 20:33

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: HALL,NICHOLAS
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: FRIEND

DATE/TIME OF TEST: 05/07/2018 02:36:22

Diagnostics Check		OK	02:33	
Air Blank		0.000	02:33	
Control Test		0.080	02:33	
Air Blank		0.000	02:34	
Subject Sample	# 1		02:35	Subject Test Refused
Breath Volume		0.000	02:35	
Air Blank		0.000	02:35	
Control Test		0.080	02:35	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:36	

Data Download Date/Time: 05/26/2018 20:33

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON,EDWIN,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: JONES,JOSEPH
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: BAUER

DATE/TIME OF TEST: 05/05/2018 00:41:21

Diagnostics Check		OK	00:27	
Air Blank		0.000	00:28	
Control Test		0.080	00:28	
Air Blank		0.000	00:29	
Subject Sample	# 1	0.182	00:32	Volume Not Met
Breath Volume		0.984	00:32	
Air Blank		0.000	00:33	
Air Blank		0.000	00:34	
Subject Sample	# 2	0.199	00:35	
Breath Volume		1.738	00:35	
Air Blank		0.000	00:36	
Air Blank		0.000	00:38	
Subject Sample	# 3	0.199	00:39	
Breath Volume		1.585	00:39	
Air Blank		0.000	00:40	
Control Test		0.079	00:40	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 05/31/2018 08:10

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: HUNT,KELLY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: SANTASIERO

DATE/TIME OF TEST: 05/12/2018 03:19:38

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.080	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.154	03:15
Breath Volume		2.816	03:15
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.159	03:18
Breath Volume		2.246	03:18
Air Blank		0.000	03:18
Control Test		0.079	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 05/31/2018 08:10

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,WILLIAM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD,WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: TULLIS

DATE/TIME OF TEST: 05/13/2018 00:08:35

Diagnostics Check		OK	00:00
Air Blank		0.000	00:01
Control Test		0.080	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.239	00:03
Breath Volume		1.511	00:03
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.242	00:07
Breath Volume		1.867	00:07
Air Blank		0.000	00:07
Control Test		0.079	00:07
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 05/31/2018 08:10

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSION,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: HESSION,CHRISTOPHER
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: BRANDOW

DATE/TIME OF TEST: 05/18/2018 17:34:45

Diagnostics Check		OK	17:27
Air Blank		0.000	17:28
Control Test		0.080	17:28
Air Blank		0.000	17:29
Subject Sample	# 1	0.342	17:30
Breath Volume		2.125	17:30
Air Blank		0.000	17:30
Air Blank		0.000	17:32
Subject Sample	# 2	0.336	17:33
Breath Volume		2.101	17:33
Air Blank		0.000	17:33
Control Test		0.079	17:34
Air Blank		0.000	17:34
Diagnostics Check		OK	17:34

Data Download Date/Time: 05/31/2018 08:10

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANQUI,DANIEL,C
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: FRANQUI,DANIEL
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BURGER

DATE/TIME OF TEST: 05/19/2018 01:28:22

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.079	01:22
Air Blank		0.000	01:23
Subject Sample	# 1	0.199	01:23
Breath Volume		1.972	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.192	01:26
Breath Volume		2.628	01:26
Air Blank		0.000	01:27
Control Test		0.079	01:27
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 05/31/2018 08:10

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUBRANO,OLIVIA,C
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: LUBRANO,OLIVIA
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 14:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 05/22/2018 14:59:42

Diagnostics Check		OK	14:53
Air Blank		0.000	14:53
Control Test		0.079	14:53
Air Blank		0.000	14:54
Subject Sample	# 1	0.000	14:55
Breath Volume		1.718	14:55
Air Blank		0.000	14:55
Air Blank		0.000	14:57
Subject Sample	# 2	0.000	14:58
Breath Volume		1.625	14:58
Air Blank		0.000	14:58
Control Test		0.079	14:59
Air Blank		0.000	14:59
Diagnostics Check		OK	14:59

Data Download Date/Time: 05/31/2018 08:10

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINEHAN,DANIEL,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: LINEHAN,DANIEL
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 14:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:02

SUBJECT'S LAST NAME: WHITEWAY

DATE/TIME OF TEST: 05/24/2018 15:36:49

Diagnostics Check		OK	15:30
Air Blank		0.000	15:30
Control Test		0.079	15:31
Air Blank		0.000	15:31
Subject Sample	# 1	0.000	15:32
Breath Volume		2.511	15:32
Air Blank		0.000	15:32
Air Blank		0.000	15:34
Subject Sample	# 2	0.000	15:35
Breath Volume		1.183	15:35
Air Blank		0.000	15:35
Control Test		0.080	15:36
Air Blank		0.000	15:36
Diagnostics Check		OK	15:36

Data Download Date/Time: 05/31/2018 08:10

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSION,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: DUMMER,LUKE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 05/26/2018 01:31:28

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.080	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.000	01:27
Breath Volume		2.226	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.000	01:30
Breath Volume		1.972	01:30
Air Blank		0.000	01:30
Control Test		0.081	01:30
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 05/31/2018 08:10

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,JEFFERY,L
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BLAIR,JEFFERY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 04/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:49

SUBJECT'S LAST NAME: KILLMER

DATE/TIME OF TEST: 05/28/2018 17:24:42

Diagnostics Check		OK	17:18
Air Blank		0.000	17:18
Control Test		0.079	17:19
Air Blank		0.000	17:19
Subject Sample	# 1	0.000	17:20
Breath Volume		1.312	17:20
Air Blank		0.000	17:20
Air Blank		0.000	17:22
Subject Sample	# 2	0.000	17:23
Breath Volume		1.515	17:23
Air Blank		0.000	17:23
Control Test		0.080	17:24
Air Blank		0.000	17:24
Diagnostics Check		OK	17:24

Data Download Date/Time: 05/31/2018 08:10

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,WILLAIM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD,WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 05/05/2018 23:09:56

Diagnostics Check		OK	23:06	
Air Blank		0.000	23:06	
Control Test		0.080	23:07	
Air Blank		0.000	23:07	
Subject Sample	# 1	0.261	23:08	
Breath Volume		1.335	23:08	
Air Blank			23:09	Purge Fail
Air Blank			23:09	Purge Fail

Data Download Date/Time: 05/29/2018 05:53

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,WILLIAM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD,WILLAIM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 05/05/2018 23:18:55

Diagnostics Check		OK	23:11	
Air Blank		0.000	23:12	
Control Test		0.079	23:12	
Air Blank		0.000	23:13	
Subject Sample	# 1	0.256	23:14	
Breath Volume		1.742	23:14	
Air Blank		0.000	23:15	
Air Blank		0.000	23:16	
Subject Sample	# 2	0.250	23:17	
Breath Volume		1.238	23:17	
Air Blank		0.000	23:18	
Control Test		0.080	23:18	Interferent Detect
Air Blank		0.000	23:18	

Data Download Date/Time: 05/29/2018 05:53

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIMMER,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: WIMMER,MATTHEW
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: MIRABEL

DATE/TIME OF TEST: 05/28/2018 13:05:33

Diagnostics Check	OK	13:04	
Air Blank	0.000	13:04	
Control Test	0.081	13:04	Interferent Detect
Air Blank	0.000	13:05	

Data Download Date/Time: 05/29/2018 05:53

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIMMER,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: WIMMER,MATTHEW
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: MIRABEL

DATE/TIME OF TEST: 05/28/2018 13:08:41

Diagnostics Check	OK	13:07	
Air Blank	0.000	13:07	
Control Test	0.081	13:08	Interferent Detect
Air Blank	0.000	13:08	

Data Download Date/Time: 05/29/2018 05:53

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NUTTING,KAYLEE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: HORN

DATE/TIME OF TEST: 05/03/2018 00:07:17

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.081	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.095	00:02
Breath Volume		2.699	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.096	00:05
Breath Volume		2.523	00:05
Air Blank		0.000	00:06
Control Test		0.080	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 05/23/2018 18:25

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,NICHOLAS,W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FREEMAN,W
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: MCGILL

DATE/TIME OF TEST: 05/04/2018 23:44:34

Diagnostics Check		OK	23:31	
Air Blank		0.000	23:31	
Control Test		0.080	23:32	
Air Blank		0.000	23:32	
Subject Sample	# 1	0.314	23:34	
Breath Volume		2.222	23:34	
Air Blank		0.000	23:35	
Air Blank		0.000	23:36	
Subject Sample	# 2	0.282	23:39	Volume Not Met
Breath Volume		0.785	23:39	
Air Blank		0.000	23:40	
Air Blank		0.000	23:42	
Subject Sample	# 3	0.309	23:42	
Breath Volume		1.695	23:42	
Air Blank		0.000	23:43	
Control Test		0.078	23:43	
Air Blank		0.000	23:44	
Diagnostics Check		OK	23:44	

Data Download Date/Time: 05/23/2018 18:25

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,NICHOLAS,W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LEWIS,B
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 05/06/2018 04:57:19

Diagnostics Check		OK	04:49
Air Blank		0.000	04:50
Control Test		0.081	04:50
Air Blank		0.000	04:51
Subject Sample	# 1	0.205	04:52
Breath Volume		1.472	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.212	04:55
Breath Volume		2.082	04:55
Air Blank		0.000	04:56
Control Test		0.079	04:56
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 05/23/2018 18:25

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,NICHOLAS,W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: ROCCATANI

DATE/TIME OF TEST: 05/06/2018 20:49:51

Diagnostics Check		OK	20:41
Air Blank		0.000	20:41
Control Test		0.080	20:42
Air Blank		0.000	20:42
Subject Sample	# 1	0.242	20:45
Breath Volume		1.968	20:45
Air Blank		0.000	20:46
Air Blank		0.000	20:47
Subject Sample	# 2	0.222	20:48
Breath Volume		1.621	20:48
Air Blank		0.000	20:48
Control Test		0.079	20:49
Air Blank		0.000	20:49
Diagnostics Check		OK	20:49

Data Download Date/Time: 05/23/2018 18:25

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWELL,NATHANAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROWELL,NATHANAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: VIEIRA

DATE/TIME OF TEST: 05/12/2018 00:10:10

Diagnostics Check		OK	00:03
Air Blank		0.000	00:04
Control Test		0.081	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.085	00:05
Breath Volume		4.089	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.082	00:08
Breath Volume		4.222	00:08
Air Blank		0.000	00:09
Control Test		0.080	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 05/23/2018 18:25

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: JEFFERSON,ZACHERY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 09:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:38

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 05/12/2018 12:05:44

Diagnostics Check		OK	11:59
Air Blank		0.000	12:00
Control Test		0.081	12:00
Air Blank		0.000	12:01
Subject Sample	# 1	0.000	12:01
Breath Volume		2.191	12:01
Air Blank		0.000	12:01
Air Blank		0.000	12:03
Subject Sample	# 2	0.000	12:04
Breath Volume		1.878	12:04
Air Blank		0.000	12:04
Control Test		0.081	12:05
Air Blank		0.000	12:05
Diagnostics Check		OK	12:05

Data Download Date/Time: 05/23/2018 18:25

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWELL,NATHANAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NUTTING,KAYLEE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: STREVAL

DATE/TIME OF TEST: 05/12/2018 22:52:00

Diagnostics Check		OK	22:45
Air Blank		0.000	22:45
Control Test		0.081	22:46
Air Blank		0.000	22:46
Subject Sample	# 1	0.207	22:47
Breath Volume		1.417	22:47
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.209	22:50
Breath Volume		1.406	22:50
Air Blank		0.000	22:51
Control Test		0.081	22:51
Air Blank		0.000	22:51
Diagnostics Check		OK	22:52

Data Download Date/Time: 05/23/2018 18:25

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWELL,NATHANAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROWELL,NATHANAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 05/18/2018 03:56:54

Diagnostics Check		OK	03:48	
Air Blank		0.000	03:48	
Control Test		0.081	03:49	
Air Blank		0.000	03:49	
Subject Sample	# 1	0.149	03:52	Volume Not Met
Breath Volume		0.335	03:52	
Air Blank		0.000	03:53	
Air Blank		0.000	03:55	
Subject Sample	# 2		03:55	Subject Test Refused
Breath Volume		0.000	03:55	
Air Blank		0.000	03:56	
Control Test		0.077	03:56	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 05/23/2018 18:25

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MARCUS,H
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BARRETT,T
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 05/19/2018 02:33:42

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.080	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.171	02:29
Breath Volume		2.714	02:29
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.165	02:32
Breath Volume		2.820	02:32
Air Blank		0.000	02:32
Control Test		0.079	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 05/23/2018 18:25

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MARCUS,H
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HARRIS,T
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: KOMADA

DATE/TIME OF TEST: 05/23/2018 23:27:40

Diagnostics Check		OK	23:21
Air Blank		0.000	23:21
Control Test		0.082	23:22
Air Blank		0.000	23:22
Subject Sample	# 1	0.031	23:23
Breath Volume		3.355	23:23
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.028	23:26
Breath Volume		3.238	23:26
Air Blank		0.000	23:26
Control Test		0.080	23:27
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 06/13/2018 12:47

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWELL,NATHANAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROWELL,NATHANAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 05/25/2018 21:51:53

Diagnostics Check		OK	21:45
Air Blank		0.000	21:46
Control Test		0.081	21:46
Air Blank		0.000	21:47
Subject Sample	# 1	0.114	21:47
Breath Volume		2.500	21:47
Air Blank		0.000	21:48
Air Blank		0.000	21:49
Subject Sample	# 2	0.105	21:50
Breath Volume		1.761	21:50
Air Blank		0.000	21:50
Control Test		0.080	21:51
Air Blank		0.000	21:51
Diagnostics Check		OK	21:51

Data Download Date/Time: 06/13/2018 12:48

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWELL,NATHANAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROWELL,NATHANAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GAW

DATE/TIME OF TEST: 05/27/2018 23:27:25

Diagnostics Check		OK	23:21
Air Blank		0.000	23:21
Control Test		0.080	23:22
Air Blank		0.000	23:22
Subject Sample	# 1	0.180	23:22
Breath Volume		1.945	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.172	23:25
Breath Volume		1.523	23:25
Air Blank		0.000	23:26
Control Test		0.081	23:26
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 06/13/2018 12:48

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: DANISH

DATE/TIME OF TEST: 05/06/2018 00:28:11

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.081	00:22
Air Blank		0.000	00:23
Subject Sample	# 1	0.163	00:23
Breath Volume		3.097	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.158	00:26
Breath Volume		1.511	00:26
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 05/25/2018 18:38

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: SHEPHARD

DATE/TIME OF TEST: 05/15/2018 22:12:12

Diagnostics Check		OK	22:04
Air Blank		0.000	22:04
Control Test		0.083	22:05
Air Blank		0.000	22:05
Subject Sample	# 1	0.125	22:07
Breath Volume		1.414	22:07
Air Blank		0.000	22:08
Air Blank		0.000	22:10
Subject Sample	# 2	0.128	22:10
Breath Volume		1.550	22:10
Air Blank		0.000	22:11
Control Test		0.080	22:11
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 05/25/2018 18:38

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 05/16/2018 01:47:07

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.151	01:42
Breath Volume		1.753	01:42
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.152	01:45
Breath Volume		1.835	01:45
Air Blank		0.000	01:46
Control Test		0.079	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 05/25/2018 18:38

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ARTEAGA

DATE/TIME OF TEST: 05/18/2018 03:26:50

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.083	03:26	Interferent Detect
Air Blank	0.000	03:26	

Data Download Date/Time: 05/25/2018 18:38

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ARTEAGA

DATE/TIME OF TEST: 05/18/2018 03:41:49

Diagnostics Check		OK	03:29	
Air Blank		0.000	03:29	
Control Test		0.081	03:29	
Air Blank		0.000	03:30	
Subject Sample	# 1	0.103	03:33	Volume Not Met
Breath Volume		0.066	03:33	
Air Blank		0.000	03:34	
Air Blank		0.000	03:35	
Subject Sample	# 2	0.107	03:36	
Breath Volume		1.671	03:36	
Air Blank		0.000	03:37	
Air Blank		0.000	03:38	
Subject Sample	# 3	0.108	03:40	
Breath Volume		1.738	03:40	
Air Blank		0.000	03:40	
Control Test		0.081	03:41	
Air Blank		0.000	03:41	
Diagnostics Check		OK	03:41	

Data Download Date/Time: 05/25/2018 18:38

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: CHILDS

DATE/TIME OF TEST: 05/30/2018 01:19:56

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:05	
Control Test		0.081	01:05	
Air Blank		0.000	01:06	
Subject Sample	# 1	0.138	01:09	Volume Not Met
Breath Volume		0.136	01:09	
Air Blank		0.000	01:10	
Air Blank		0.000	01:11	
Subject Sample	# 2	0.130	01:12	
Breath Volume		1.183	01:12	
Air Blank		0.000	01:13	
Air Blank		0.000	01:15	
Subject Sample	# 3	0.134	01:18	
Breath Volume		1.734	01:18	
Air Blank		0.000	01:19	
Control Test		0.080	01:19	
Air Blank		0.000	01:19	
Diagnostics Check		OK	01:19	

Data Download Date/Time: 06/27/2018 18:38

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: COUCH

DATE/TIME OF TEST: 05/05/2018 23:40:29

Diagnostics Check		OK	23:34
Air Blank		0.000	23:34
Control Test		0.078	23:35
Air Blank		0.000	23:35
Subject Sample	# 1	0.116	23:35
Breath Volume		3.035	23:35
Air Blank		0.000	23:36
Air Blank		0.000	23:38
Subject Sample	# 2	0.109	23:39
Breath Volume		2.539	23:39
Air Blank		0.000	23:39
Control Test		0.079	23:39
Air Blank		0.000	23:40
Diagnostics Check		OK	23:40

Data Download Date/Time: 05/25/2018 18:42

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 05/12/2018 04:06:51

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.079	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.212	04:02
Breath Volume		2.156	04:02
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.208	04:05
Breath Volume		3.441	04:05
Air Blank		0.000	04:05
Control Test		0.079	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 05/25/2018 18:42

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: WOLFINGBARGER

DATE/TIME OF TEST: 05/16/2018 23:39:58

Diagnostics Check		OK	23:33
Air Blank		0.000	23:34
Control Test		0.081	23:34
Air Blank		0.000	23:34
Subject Sample	# 1	0.065	23:35
Breath Volume		1.722	23:35
Air Blank		0.000	23:36
Air Blank		0.000	23:38
Subject Sample	# 2	0.068	23:38
Breath Volume		2.046	23:38
Air Blank		0.000	23:39
Control Test		0.081	23:39
Air Blank		0.000	23:39
Diagnostics Check		OK	23:39

Data Download Date/Time: 05/25/2018 18:42

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: LILES

DATE/TIME OF TEST: 05/18/2018 01:28:52

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.080	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.106	01:24
Breath Volume		2.687	01:24
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.109	01:27
Breath Volume		2.621	01:27
Air Blank		0.000	01:27
Control Test		0.080	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 05/25/2018 18:42

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: MOYENO SOTO

DATE/TIME OF TEST: 05/21/2018 22:15:57

Diagnostics Check		OK	22:06
Air Blank		0.000	22:06
Control Test		0.079	22:06
Air Blank		0.000	22:07
Subject Sample	# 1	0.196	22:07
Breath Volume		2.253	22:07
Air Blank		0.000	22:08
Air Blank		0.000	22:10
Subject Sample	# 2	0.175	22:10
Breath Volume		1.296	22:10
Air Blank		0.000	22:11
Air Blank		0.000	22:13
Subject Sample	# 3	0.186	22:14
Breath Volume		2.320	22:14
Air Blank		0.000	22:15
Control Test		0.080	22:15
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 05/25/2018 18:42

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: DOODY

DATE/TIME OF TEST: 05/29/2018 03:34:37

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.081	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.116	03:30
Breath Volume		2.648	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.109	03:33
Breath Volume		2.500	03:33
Air Blank		0.000	03:33
Control Test		0.078	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 06/27/2018 18:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: HUBBS

DATE/TIME OF TEST: 05/29/2018 20:51:37

Diagnostics Check		OK	20:48	
Air Blank		0.000	20:48	
Control Test		0.079	20:49	
Air Blank		0.000	20:49	
Subject Sample	# 1		20:50	Subject Test Refused
Breath Volume		0.000	20:50	
Air Blank		0.000	20:50	
Control Test		0.079	20:51	
Air Blank		0.000	20:51	
Diagnostics Check		OK	20:51	

Data Download Date/Time: 06/27/2018 18:47

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 05/30/2018 01:14:52

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.079	01:09
Air Blank		0.000	01:09
Subject Sample	# 1	0.128	01:10
Breath Volume		1.503	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.131	01:13
Breath Volume		1.746	01:13
Air Blank		0.000	01:13
Control Test		0.078	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 06/27/2018 18:47

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: INOA,AMELIA
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: HODSDON

DATE/TIME OF TEST: 05/27/2018 03:25:38

Diagnostics Check		OK	03:19
Air Blank		0.000	03:19
Control Test		0.080	03:20
Air Blank		0.000	03:20
Subject Sample	# 1	0.283	03:21
Breath Volume		1.695	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.269	03:24
Breath Volume		1.777	03:24
Air Blank		0.000	03:24
Control Test		0.078	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 05/29/2018 00:57

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HENRY,CAITLIN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GROOMS

DATE/TIME OF TEST: 05/28/2018 03:56:14

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.000	03:51
Breath Volume		3.195	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.000	03:54
Breath Volume		2.843	03:54
Air Blank		0.000	03:55
Control Test		0.081	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 05/29/2018 00:57

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PONDER,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: PHILLIPS,R
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 05/28/2018 21:21:59

Diagnostics Check		OK	21:15
Air Blank		0.000	21:15
Control Test		0.079	21:16
Air Blank		0.000	21:16
Subject Sample	# 1	0.195	21:17
Breath Volume		1.503	21:17
Air Blank		0.000	21:17
Air Blank		0.000	21:19
Subject Sample	# 2	0.193	21:20
Breath Volume		1.472	21:20
Air Blank		0.000	21:21
Control Test		0.080	21:21
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 05/29/2018 00:57

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ROSS,JACOB
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: HUBER

DATE/TIME OF TEST: 05/31/2018 00:51:19

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.080	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.241	00:46
Breath Volume		3.046	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.240	00:49
Breath Volume		2.296	00:49
Air Blank		0.000	00:50
Control Test		0.080	00:50
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 07/01/2018 21:34

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEADOWS,SHAWN,M
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MEADOWS,SHAWN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: COSME

DATE/TIME OF TEST: 05/31/2018 01:43:29

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.080	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.150	01:38
Breath Volume		1.546	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.152	01:41
Breath Volume		1.828	01:41
Air Blank		0.000	01:42
Control Test		0.080	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 07/01/2018 21:34

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DONNELY,SHANE
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: MCLELLAN

DATE/TIME OF TEST: 05/02/2018 01:44:53

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:32	
Control Test		0.081	01:32	
Air Blank		0.000	01:33	
Subject Sample	# 1	0.240	01:36	Volume Not Met
Breath Volume		0.113	01:36	
Air Blank		0.000	01:37	
Air Blank		0.000	01:38	
Subject Sample	# 2	0.237	01:40	
Breath Volume		1.343	01:40	
Air Blank		0.000	01:40	
Air Blank		0.000	01:42	
Subject Sample	# 3	0.237	01:43	
Breath Volume		1.644	01:43	
Air Blank		0.000	01:43	
Control Test		0.079	01:44	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:44	

Data Download Date/Time: 05/29/2018 00:59

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLBURN,MEGAN,E
BREATH TEST OPERATOR AGENCY: Casselberry PD
ARRESTING OFFICER: PHILLIPS,RONALD
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/02/2018 07:21:23

Diagnostics Check		OK	07:14
Air Blank		0.000	07:14
Control Test		0.082	07:15
Air Blank		0.000	07:15
Subject Sample	# 1	0.068	07:16
Breath Volume		1.492	07:16
Air Blank		0.000	07:16
Air Blank		0.000	07:18
Subject Sample	# 2	0.065	07:19
Breath Volume		1.230	07:19
Air Blank		0.000	07:20
Control Test		0.082	07:20
Air Blank		0.000	07:21
Diagnostics Check		OK	07:21

Data Download Date/Time: 05/29/2018 00:59

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PHILLIPS,RON
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 05/07/2018 04:44:12

Diagnostics Check		OK	04:37
Air Blank		0.000	04:38
Control Test		0.082	04:38
Air Blank		0.000	04:39
Subject Sample	# 1	0.041	04:39
Breath Volume		2.589	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.040	04:42
Breath Volume		2.437	04:42
Air Blank		0.000	04:43
Control Test		0.081	04:43
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 05/29/2018 00:59

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MOREL,ANTHONY
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: STOCKLEY

DATE/TIME OF TEST: 05/07/2018 20:45:56

Diagnostics Check		OK	20:39
Air Blank		0.000	20:40
Control Test		0.081	20:40
Air Blank		0.000	20:40
Subject Sample	# 1	0.191	20:41
Breath Volume		3.445	20:41
Air Blank		0.000	20:42
Air Blank		0.000	20:43
Subject Sample	# 2	0.201	20:44
Breath Volume		3.425	20:44
Air Blank		0.000	20:45
Control Test		0.082	20:45
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 05/29/2018 00:59

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ANCION,CALIX
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: PENTECOST

DATE/TIME OF TEST: 05/13/2018 01:09:07

Diagnostics Check		OK	01:03
Air Blank		0.000	01:03
Control Test		0.082	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.157	01:04
Breath Volume		1.480	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.160	01:07
Breath Volume		2.582	01:07
Air Blank		0.000	01:08
Control Test		0.083	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 05/29/2018 00:59

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HALL,JONATHAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: COLLIER

DATE/TIME OF TEST: 05/13/2018 05:03:22

Diagnostics Check		OK	04:57
Air Blank		0.000	04:57
Control Test		0.081	04:57
Air Blank		0.000	04:58
Subject Sample	# 1	0.257	04:58
Breath Volume		2.621	04:58
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.259	05:01
Breath Volume		2.609	05:01
Air Blank		0.000	05:02
Control Test		0.081	05:02
Air Blank		0.000	05:03
Diagnostics Check		OK	05:03

Data Download Date/Time: 05/29/2018 00:59

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MOREL,ANTHONY
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BROOKE

DATE/TIME OF TEST: 05/14/2018 00:21:48

Diagnostics Check		OK	00:15
Air Blank		0.000	00:15
Control Test		0.081	00:16
Air Blank		0.000	00:16
Subject Sample	# 1	0.000	00:17
Breath Volume		2.359	00:17
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.000	00:20
Breath Volume		2.156	00:20
Air Blank		0.000	00:20
Control Test		0.081	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 05/29/2018 00:59

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: VENTURA,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: BERRIOS

DATE/TIME OF TEST: 05/15/2018 22:32:31

Diagnostics Check		OK	22:25
Air Blank		0.000	22:26
Control Test		0.081	22:26
Air Blank		0.000	22:27
Subject Sample	# 1	0.139	22:28
Breath Volume		1.343	22:28
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.140	22:31
Breath Volume		1.871	22:31
Air Blank		0.000	22:31
Control Test		0.082	22:31
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 05/29/2018 00:59

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MCDANIEL,JASON
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: INESON

DATE/TIME OF TEST: 05/16/2018 02:25:30

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.081	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.223	02:21
Breath Volume		2.105	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.225	02:23
Breath Volume		2.007	02:23
Air Blank		0.000	02:24
Control Test		0.082	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 05/29/2018 00:59

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MUNN,SCOTT
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SLAUGHTER

DATE/TIME OF TEST: 05/17/2018 00:47:10

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.082	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.205	00:42
Breath Volume		1.769	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.205	00:45
Breath Volume		1.937	00:45
Air Blank		0.000	00:46
Control Test		0.081	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 05/29/2018 00:59

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GRAFFY,THOMAS
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: MARCAN0

DATE/TIME OF TEST: 05/17/2018 04:00:01

Diagnostics Check		OK	03:45	
Air Blank		0.000	03:45	
Control Test		0.081	03:45	
Air Blank		0.000	03:46	
Subject Sample	# 1	0.173	03:49	Volume Not Met
Breath Volume		0.281	03:49	
Air Blank		0.000	03:50	
Air Blank		0.000	03:52	
Subject Sample	# 2	0.170	03:53	
Breath Volume		1.445	03:53	
Air Blank		0.000	03:54	
Air Blank		0.000	03:56	
Subject Sample	# 3	0.172	03:58	
Breath Volume		1.515	03:58	
Air Blank		0.000	03:59	
Control Test		0.080	03:59	
Air Blank		0.000	03:59	
Diagnostics Check		OK	04:00	

Data Download Date/Time: 05/29/2018 00:59

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HILL,SEAN
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/17/2018 23:57:45

Diagnostics Check		OK	23:51
Air Blank		0.000	23:51
Control Test		0.081	23:51
Air Blank		0.000	23:52
Subject Sample	# 1	0.280	23:52
Breath Volume		2.601	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.265	23:56
Breath Volume		2.238	23:56
Air Blank		0.000	23:56
Control Test		0.081	23:57
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 05/29/2018 01:00

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KINKOPF,TIMOTHY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 05/19/2018 01:42:02

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.082	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.171	01:37
Breath Volume		3.148	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.171	01:40
Breath Volume		2.816	01:40
Air Blank		0.000	01:41
Control Test		0.081	01:41
Air Blank		0.000	01:41
Diagnostics Check		OK	01:42

Data Download Date/Time: 05/29/2018 01:00

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: EGER,VINCENT
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/19/2018 05:18:03

Diagnostics Check		OK	05:12
Air Blank		0.000	05:12
Control Test		0.081	05:12
Air Blank		0.000	05:13
Subject Sample	# 1	0.173	05:13
Breath Volume		2.050	05:13
Air Blank		0.000	05:14
Air Blank		0.000	05:16
Subject Sample	# 2	0.174	05:16
Breath Volume		3.421	05:16
Air Blank		0.000	05:17
Control Test		0.081	05:17
Air Blank		0.000	05:17
Diagnostics Check		OK	05:18

Data Download Date/Time: 05/29/2018 01:00

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MCKINNON,FRANCIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: LAZU-FELICIANO

DATE/TIME OF TEST: 05/20/2018 04:34:18

Diagnostics Check		OK	04:28
Air Blank		0.000	04:29
Control Test		0.082	04:29
Air Blank		0.000	04:29
Subject Sample	# 1	0.000	04:30
Breath Volume		2.117	04:30
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.000	04:32
Breath Volume		2.664	04:32
Air Blank		0.000	04:33
Control Test		0.081	04:33
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 05/29/2018 01:00

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TOZZI,SHAUN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: SHOWERS

DATE/TIME OF TEST: 05/21/2018 01:23:52

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.082	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.133	01:19
Breath Volume		1.851	01:19
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.126	01:22
Breath Volume		1.277	01:22
Air Blank		0.000	01:22
Control Test		0.082	01:23
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 05/29/2018 01:00

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HANSARD,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ALVAREZ LEGER

DATE/TIME OF TEST: 05/21/2018 03:47:16

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.081	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.134	03:42
Breath Volume		2.210	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.129	03:45
Breath Volume		2.011	03:45
Air Blank		0.000	03:46
Control Test		0.081	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 05/29/2018 01:00

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RICE,MATHEW
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: PRYER

DATE/TIME OF TEST: 05/22/2018 01:01:43

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.081	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.190	00:57
Breath Volume		1.367	00:57
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.189	01:00
Breath Volume		1.460	01:00
Air Blank		0.000	01:00
Control Test		0.082	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 05/29/2018 01:00

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KRUPA,KYLE
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: FUNKE

DATE/TIME OF TEST: 05/24/2018 23:22:06

Diagnostics Check		OK	23:15
Air Blank		0.000	23:15
Control Test		0.082	23:16
Air Blank		0.000	23:16
Subject Sample	# 1	0.213	23:17
Breath Volume		3.355	23:17
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.221	23:20
Breath Volume		2.046	23:20
Air Blank		0.000	23:21
Control Test		0.083	23:21
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 05/29/2018 01:00

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: WADEMAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ANTOINE

DATE/TIME OF TEST: 05/30/2018 04:26:31

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.081	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.119	04:22
Breath Volume		1.472	04:22
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.121	04:25
Breath Volume		1.179	04:25
Air Blank		0.000	04:25
Control Test		0.080	04:25
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 07/01/2018 21:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KINKOPF,TIM
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: RIAD

DATE/TIME OF TEST: 05/30/2018 05:41:52

Diagnostics Check		OK	05:28	
Air Blank		0.000	05:29	
Control Test		0.081	05:29	
Air Blank		0.000	05:29	
Subject Sample	# 1	0.168	05:31	
Breath Volume		1.183	05:31	
Air Blank		0.000	05:31	
Air Blank		0.000	05:33	
Subject Sample	# 2	0.121	05:36	Volume Not Met
Breath Volume		0.488	05:36	
Air Blank		0.000	05:37	
Air Blank		0.000	05:39	
Subject Sample	# 3	0.161	05:40	
Breath Volume		1.523	05:40	
Air Blank		0.000	05:40	
Control Test		0.079	05:41	
Air Blank		0.000	05:41	
Diagnostics Check		OK	05:41	

Data Download Date/Time: 07/01/2018 21:42

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/03/2018 02:50:15

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.084	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.165	02:45
Breath Volume		1.695	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.158	02:48
Breath Volume		1.398	02:48
Air Blank		0.000	02:49
Control Test		0.082	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 05/03/2018 05:02:04

Diagnostics Check		OK	04:54
Air Blank		0.000	04:54
Control Test		0.083	04:55
Air Blank		0.000	04:55
Subject Sample	# 1	0.197	04:56
Breath Volume		1.515	04:56
Air Blank		0.000	04:57
Air Blank		0.000	04:59
Subject Sample	# 2	0.192	05:00
Breath Volume		1.644	05:00
Air Blank		0.000	05:01
Control Test		0.081	05:01
Air Blank		0.000	05:01
Diagnostics Check		OK	05:02

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: GWINN

DATE/TIME OF TEST: 05/03/2018 23:55:23

Diagnostics Check		OK	23:48
Air Blank		0.000	23:49
Control Test		0.083	23:49
Air Blank		0.000	23:50
Subject Sample	# 1	0.000	23:51
Breath Volume		2.917	23:51
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.000	23:54
Breath Volume		2.371	23:54
Air Blank		0.000	23:54
Control Test		0.082	23:54
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOHR,K,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOHR,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: LEVINSON

DATE/TIME OF TEST: 05/04/2018 01:16:20

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.083	01:10
Air Blank		0.000	01:10
Subject Sample	# 1	0.143	01:11
Breath Volume		2.175	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:13
Subject Sample	# 2	0.136	01:14
Breath Volume		1.750	01:14
Air Blank		0.000	01:15
Control Test		0.082	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 05/05/2018 01:39:55

Diagnostics Check		OK	01:33
Air Blank		0.000	01:34
Control Test		0.083	01:34
Air Blank		0.000	01:35
Subject Sample	# 1	0.000	01:35
Breath Volume		2.054	01:35
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.000	01:38
Breath Volume		2.289	01:38
Air Blank		0.000	01:38
Control Test		0.083	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GRADY

DATE/TIME OF TEST: 05/05/2018 04:15:12

Diagnostics Check		OK	04:12	
Air Blank		0.000	04:12	
Control Test		0.082	04:13	
Air Blank		0.000	04:13	
Subject Sample	# 1		04:13	Subject Test Refused
Breath Volume		0.000	04:13	
Air Blank		0.000	04:14	
Control Test		0.083	04:14	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:15	

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MOSCONE

DATE/TIME OF TEST: 05/05/2018 04:33:43

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.080	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.143	04:28
Breath Volume		1.718	04:28
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.147	04:32
Breath Volume		2.195	04:32
Air Blank		0.000	04:32
Control Test		0.082	04:33
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: VANHOESEN

DATE/TIME OF TEST: 05/06/2018 02:30:56

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.082	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.279	02:25
Breath Volume		1.378	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.281	02:29
Breath Volume		1.449	02:29
Air Blank		0.000	02:29
Control Test		0.081	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: ELLINGTON

DATE/TIME OF TEST: 05/06/2018 04:53:46

Diagnostics Check		OK	04:47
Air Blank		0.000	04:47
Control Test		0.082	04:48
Air Blank		0.000	04:48
Subject Sample	# 1	0.143	04:49
Breath Volume		1.554	04:49
Air Blank		0.000	04:49
Air Blank		0.000	04:51
Subject Sample	# 2	0.142	04:52
Breath Volume		2.406	04:52
Air Blank		0.000	04:52
Control Test		0.081	04:53
Air Blank		0.000	04:53
Diagnostics Check		OK	04:53

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BALDETTI,ANTHONY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 14:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: FEDORCHAK

DATE/TIME OF TEST: 05/06/2018 15:21:48

Diagnostics Check		OK	15:15
Air Blank		0.000	15:16
Control Test		0.082	15:16
Air Blank		0.000	15:17
Subject Sample	# 1	0.000	15:17
Breath Volume		1.988	15:17
Air Blank		0.000	15:17
Air Blank		0.000	15:19
Subject Sample	# 2	0.000	15:20
Breath Volume		2.218	15:20
Air Blank		0.000	15:20
Control Test		0.083	15:21
Air Blank		0.000	15:21
Diagnostics Check		OK	15:21

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:22

SUBJECT'S LAST NAME: SOUTO

DATE/TIME OF TEST: 05/11/2018 12:50:10

Diagnostics Check		OK	12:47	
Air Blank		0.000	12:47	
Control Test		0.083	12:47	
Air Blank		0.000	12:48	
Subject Sample	# 1		12:48	Subject Test Refused
Breath Volume		0.000	12:48	
Air Blank		0.000	12:49	
Control Test		0.082	12:49	
Air Blank		0.000	12:50	
Diagnostics Check		OK	12:50	

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: PEMBERTON

DATE/TIME OF TEST: 05/13/2018 02:10:58

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.082	02:05
Air Blank		0.000	02:05
Subject Sample	# 1	0.130	02:06
Breath Volume		1.816	02:06
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.129	02:09
Breath Volume		1.632	02:09
Air Blank		0.000	02:10
Control Test		0.082	02:10
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEIB,MELANIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: GOSNELL

DATE/TIME OF TEST: 05/13/2018 23:47:51

Diagnostics Check		OK	23:38
Air Blank		0.000	23:39
Control Test		0.083	23:39
Air Blank		0.000	23:39
Subject Sample	# 1	0.233	23:40
Breath Volume		2.410	23:40
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.211	23:43
Breath Volume		1.390	23:43
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 3	0.224	23:46
Breath Volume		2.410	23:46
Air Blank		0.000	23:46
Control Test		0.082	23:47
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: SALGADO

DATE/TIME OF TEST: 05/14/2018 00:28:02

Diagnostics Check		OK	00:21
Air Blank		0.000	00:22
Control Test		0.082	00:22
Air Blank		0.000	00:23
Subject Sample	# 1	0.183	00:23
Breath Volume		1.679	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.170	00:26
Breath Volume		1.253	00:26
Air Blank		0.000	00:27
Control Test		0.082	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:28

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CASTELEIRO,ANTHONY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: MORRILL

DATE/TIME OF TEST: 05/15/2018 20:01:20

Diagnostics Check		OK	19:57	
Air Blank		0.000	19:58	
Control Test		0.082	19:58	
Air Blank		0.000	19:59	
Subject Sample	# 1		19:59	Subject Test Refused
Breath Volume		0.000	19:59	
Air Blank		0.000	20:00	
Control Test		0.082	20:00	
Air Blank		0.000	20:01	
Diagnostics Check		OK	20:01	

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 05/16/2018 18:38:03

Diagnostics Check		OK	18:32
Air Blank		0.000	18:32
Control Test		0.082	18:32
Air Blank		0.000	18:33
Subject Sample	# 1	0.000	18:33
Breath Volume		2.648	18:33
Air Blank		0.000	18:34
Air Blank		0.000	18:36
Subject Sample	# 2	0.000	18:36
Breath Volume		2.492	18:36
Air Blank		0.000	18:37
Control Test		0.082	18:37
Air Blank		0.000	18:37
Diagnostics Check		OK	18:38

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 05/16/2018 22:02:13

Diagnostics Check		OK	21:54
Air Blank		0.000	21:55
Control Test		0.083	21:55
Air Blank		0.000	21:55
Subject Sample	# 1	0.132	21:57
Breath Volume		1.792	21:57
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.133	22:00
Breath Volume		2.007	22:00
Air Blank		0.000	22:01
Control Test		0.081	22:01
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: MCCORMICK

DATE/TIME OF TEST: 05/17/2018 01:46:32

Diagnostics Check		OK	01:43	
Air Blank		0.000	01:44	
Control Test		0.082	01:44	
Air Blank		0.000	01:44	
Subject Sample	# 1		01:45	Subject Test Refused
Breath Volume		0.000	01:45	
Air Blank		0.000	01:45	
Control Test		0.082	01:45	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMAMTRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: JUNTUNEN

DATE/TIME OF TEST: 05/17/2018 05:05:37

Diagnostics Check		OK	04:57	
Air Blank		0.000	04:57	
Control Test		0.082	04:57	
Air Blank		0.000	04:58	
Subject Sample	# 1		05:01	No Sample Provided
Breath Volume		0.000	05:01	
Air Blank		0.000	05:01	
Air Blank		0.000	05:03	
Subject Sample	# 2		05:04	Subject Test Refused
Breath Volume		0.000	05:04	
Air Blank		0.000	05:04	
Control Test		0.082	05:05	
Air Blank		0.000	05:05	
Diagnostics Check		OK	05:05	

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONGMIRE,BENJAMIN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LONGMIRE,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: POSEY

DATE/TIME OF TEST: 05/18/2018 02:15:17

Diagnostics Check		OK	02:10	
Air Blank		0.000	02:11	
Control Test		0.082	02:11	
Air Blank		0.000	02:12	
Subject Sample	# 1		02:13	Subject Test Refused
Breath Volume		0.000	02:13	
Air Blank		0.000	02:14	
Control Test		0.082	02:14	
Air Blank		0.000	02:15	
Diagnostics Check		OK	02:15	

Data Download Date/Time: 05/31/2018 01:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/18/2018 03:16:42

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.081	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.090	03:12
Breath Volume		2.027	03:12
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.093	03:15
Breath Volume		2.335	03:15
Air Blank		0.000	03:15
Control Test		0.082	03:16
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: FISHALOW

DATE/TIME OF TEST: 05/18/2018 04:16:42

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.082	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.102	04:12
Breath Volume		2.476	04:12
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.100	04:15
Breath Volume		2.371	04:15
Air Blank		0.000	04:15
Control Test		0.081	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: JASSO

DATE/TIME OF TEST: 05/19/2018 03:58:19

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.083	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.000	03:53
Breath Volume		2.425	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.000	03:56
Breath Volume		2.109	03:56
Air Blank		0.000	03:57
Control Test		0.082	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FRASER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 05/19/2018 05:18:56

Diagnostics Check		OK	05:12
Air Blank		0.000	05:13
Control Test		0.083	05:13
Air Blank		0.000	05:14
Subject Sample	# 1	0.141	05:14
Breath Volume		4.320	05:14
Air Blank		0.000	05:15
Air Blank		0.000	05:16
Subject Sample	# 2	0.143	05:17
Breath Volume		3.136	05:17
Air Blank		0.000	05:18
Control Test		0.082	05:18
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: GAWRON

DATE/TIME OF TEST: 05/20/2018 00:10:28

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.082	00:05
Air Blank		0.000	00:05
Subject Sample	# 1	0.000	00:06
Breath Volume		3.062	00:06
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.000	00:09
Breath Volume		1.640	00:09
Air Blank		0.000	00:09
Control Test		0.082	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 05/20/2018 00:51:02

Diagnostics Check		OK	00:44
Air Blank		0.000	00:44
Control Test		0.082	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.133	00:46
Breath Volume		1.644	00:46
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.133	00:49
Breath Volume		2.300	00:49
Air Blank		0.000	00:50
Control Test		0.082	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:51

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,COURTNEY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/21/2018 00:44:45

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:42	
Control Test		0.083	00:42	
Air Blank		0.000	00:42	
Subject Sample	# 1		00:43	Subject Test Refused
Breath Volume		0.000	00:43	
Air Blank		0.000	00:43	
Control Test		0.082	00:44	
Air Blank		0.000	00:44	
Diagnostics Check		OK	00:44	

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GEOGHEGAN,LEAH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: DEUPREE

DATE/TIME OF TEST: 05/21/2018 01:43:31

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:32	
Control Test		0.082	01:32	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:36	No Sample Provided
Breath Volume		0.000	01:36	
Air Blank		0.000	01:36	
Air Blank		0.000	01:38	
Subject Sample	# 2		01:42	No Sample Provided
Breath Volume		0.000	01:42	
Air Blank		0.000	01:42	
Control Test		0.082	01:42	
Air Blank		0.000	01:43	
Diagnostics Check		OK	01:43	

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GEOGHEGAN,LEAH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: DEUPREE

DATE/TIME OF TEST: 05/21/2018 01:50:27

Diagnostics Check		OK	01:47	
Air Blank		0.000	01:48	
Control Test		0.083	01:48	
Air Blank		0.000	01:48	
Subject Sample	# 1		01:49	Subject Test Refused
Breath Volume		0.000	01:49	
Air Blank		0.000	01:49	
Control Test		0.082	01:49	
Air Blank		0.000	01:50	
Diagnostics Check		OK	01:50	

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ANTHONY,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 05/23/2018 20:10:21

Diagnostics Check		OK	20:02
Air Blank		0.000	20:02
Control Test		0.082	20:02
Air Blank		0.000	20:03
Subject Sample	# 1	0.210	20:05
Breath Volume		2.765	20:05
Air Blank		0.000	20:06
Air Blank		0.000	20:08
Subject Sample	# 2	0.201	20:08
Breath Volume		1.390	20:08
Air Blank		0.000	20:09
Control Test		0.080	20:09
Air Blank		0.000	20:10
Diagnostics Check		OK	20:10

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: CARRIGAN

DATE/TIME OF TEST: 05/24/2018 01:55:27

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:52	
Control Test		0.083	01:53	
Air Blank		0.000	01:53	
Subject Sample	# 1		01:54	Subject Test Refused
Breath Volume		0.000	01:54	
Air Blank		0.000	01:54	
Control Test		0.083	01:54	
Air Blank		0.000	01:55	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCKENZIE,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 14:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: GORNSON

DATE/TIME OF TEST: 05/24/2018 15:28:40

Diagnostics Check		OK	15:25	
Air Blank		0.000	15:26	
Control Test		0.083	15:26	
Air Blank		0.000	15:27	
Subject Sample	# 1		15:27	Subject Test Refused
Breath Volume		0.000	15:27	
Air Blank		0.000	15:27	
Control Test		0.083	15:28	
Air Blank		0.000	15:28	
Diagnostics Check		OK	15:28	

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: FICOCELLI

DATE/TIME OF TEST: 05/25/2018 02:08:22

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:59	
Control Test		0.082	01:59	
Air Blank		0.000	02:00	
Subject Sample	# 1		02:03	No Sample Provided
Breath Volume		0.000	02:03	
Air Blank		0.000	02:03	
Air Blank		0.000	02:05	
Subject Sample	# 2		02:06	Subject Test Refused
Breath Volume		0.000	02:06	
Air Blank		0.000	02:07	
Control Test		0.082	02:07	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: TORELLI

DATE/TIME OF TEST: 05/25/2018 23:32:02

Diagnostics Check		OK	23:24
Air Blank		0.000	23:25
Control Test		0.082	23:25
Air Blank		0.000	23:25
Subject Sample	# 1	0.197	23:27
Breath Volume		1.781	23:27
Air Blank		0.000	23:28
Air Blank		0.000	23:29
Subject Sample	# 2	0.189	23:30
Breath Volume		1.582	23:30
Air Blank		0.000	23:31
Control Test		0.081	23:31
Air Blank		0.000	23:31
Diagnostics Check		OK	23:32

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: HENNESSY

DATE/TIME OF TEST: 05/26/2018 04:49:21

Diagnostics Check		OK	04:41
Air Blank		0.000	04:42
Control Test		0.082	04:42
Air Blank		0.000	04:43
Subject Sample	# 1	0.219	04:44
Breath Volume		2.382	04:44
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.221	04:47
Breath Volume		2.550	04:47
Air Blank		0.000	04:48
Control Test		0.081	04:48
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 05/31/2018 01:16

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: COSTA

DATE/TIME OF TEST: 05/26/2018 20:31:32

Diagnostics Check		OK	20:28	
Air Blank		0.000	20:29	
Control Test		0.083	20:29	
Air Blank		0.000	20:29	
Subject Sample	# 1		20:30	Subject Test Refused
Breath Volume		0.000	20:30	
Air Blank		0.000	20:30	
Control Test		0.082	20:30	
Air Blank		0.000	20:31	
Diagnostics Check		OK	20:31	

Data Download Date/Time: 05/31/2018 01:17

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: KOGELMANN

DATE/TIME OF TEST: 05/27/2018 21:33:41

Diagnostics Check		OK	21:25	
Air Blank		0.000	21:25	
Control Test		0.083	21:25	
Air Blank		0.000	21:26	
Subject Sample	# 1		21:29	No Sample Provided
Breath Volume		0.000	21:29	
Air Blank		0.000	21:30	
Air Blank		0.000	21:32	
Subject Sample	# 2		21:32	Subject Test Refused
Breath Volume		0.000	21:32	
Air Blank		0.000	21:32	
Control Test		0.082	21:33	
Air Blank		0.000	21:33	
Diagnostics Check		OK	21:33	

Data Download Date/Time: 05/31/2018 01:17

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCKENZIE,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: HUMBARGER

DATE/TIME OF TEST: 05/28/2018 09:52:08

Diagnostics Check		OK	09:46
Air Blank		0.000	09:46
Control Test		0.082	09:46
Air Blank		0.000	09:47
Subject Sample	# 1	0.047	09:47
Breath Volume		2.980	09:47
Air Blank		0.000	09:48
Air Blank		0.000	09:50
Subject Sample	# 2	0.047	09:50
Breath Volume		3.195	09:50
Air Blank		0.000	09:51
Control Test		0.082	09:51
Air Blank		0.000	09:52
Diagnostics Check		OK	09:52

Data Download Date/Time: 05/31/2018 01:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: GEORG

DATE/TIME OF TEST: 05/29/2018 02:34:31

Diagnostics Check		OK	02:29	
Air Blank		0.000	02:29	
Control Test		0.083	02:30	
Air Blank		0.000	02:30	
Subject Sample	# 1		02:33	Subject Test Refused
Breath Volume		0.000	02:33	
Air Blank		0.000	02:33	
Control Test		0.082	02:33	
Air Blank		0.000	02:34	
Diagnostics Check		OK	02:34	

Data Download Date/Time: 05/31/2018 01:17

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,S,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLE,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: DURANT

DATE/TIME OF TEST: 05/01/2018 04:07:56

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:57	
Control Test		0.080	03:57	
Air Blank		0.000	03:57	
Subject Sample	# 1		04:00	No Sample Provided
Breath Volume		0.000	04:00	
Air Blank		0.000	04:01	
Air Blank		0.000	04:03	
Subject Sample	# 2		04:06	No Sample Provided
Breath Volume		0.000	04:06	
Air Blank		0.000	04:07	
Control Test		0.080	04:07	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 05/30/2018 20:21

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,JAMES,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILHELM,JAMES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: FESSLER

DATE/TIME OF TEST: 05/13/2018 05:00:28

Diagnostics Check		OK	04:56	
Air Blank		0.000	04:57	
Control Test		0.081	04:57	
Air Blank		0.000	04:58	
Subject Sample	# 1		04:59	Subject Test Refused
Breath Volume		0.000	04:59	
Air Blank		0.000	04:59	
Control Test		0.081	04:59	
Air Blank			05:00	RFI Detect
Air Blank		0.000	05:00	

Data Download Date/Time: 05/30/2018 20:21

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,JAMES,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILHELM,JAMES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: FESSLER

DATE/TIME OF TEST: 05/13/2018 05:10:55

Diagnostics Check		OK	05:07	
Air Blank		0.000	05:08	
Control Test		0.081	05:08	
Air Blank		0.000	05:09	
Subject Sample	# 1		05:09	Subject Test Refused
Breath Volume		0.000	05:09	
Air Blank		0.000	05:10	
Control Test		0.080	05:10	
Air Blank		0.000	05:10	
Diagnostics Check		OK	05:10	

Data Download Date/Time: 05/30/2018 20:21

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,JAMES,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DAVIS,JAMES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: ATTARDI

DATE/TIME OF TEST: 05/13/2018 19:55:06

Diagnostics Check		OK	19:47
Air Blank		0.000	19:47
Control Test		0.080	19:48
Air Blank		0.000	19:48
Subject Sample	# 1	0.000	19:50
Breath Volume		2.089	19:50
Air Blank		0.000	19:50
Air Blank		0.000	19:52
Subject Sample	# 2	0.000	19:53
Breath Volume		1.789	19:53
Air Blank		0.000	19:54
Control Test		0.081	19:54
Air Blank		0.000	19:55
Diagnostics Check		OK	19:55

Data Download Date/Time: 05/30/2018 20:21

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 05/18/2018 00:12:33

Diagnostics Check		OK	00:05
Air Blank		0.000	00:06
Control Test		0.081	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.156	00:07
Breath Volume		1.226	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.162	00:11
Breath Volume		1.773	00:11
Air Blank		0.000	00:11
Control Test		0.079	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 05/30/2018 20:21

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,S,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLE,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: FETZNER

DATE/TIME OF TEST: 05/25/2018 03:10:28

Diagnostics Check		OK	02:59	
Air Blank		0.000	02:59	
Control Test		0.081	02:59	
Air Blank		0.000	03:00	
Subject Sample	# 1		03:03	No Sample Provided
Breath Volume		0.000	03:03	
Air Blank		0.000	03:03	
Air Blank		0.000	03:05	
Subject Sample	# 2		03:09	No Sample Provided
Breath Volume		0.000	03:09	
Air Blank		0.000	03:09	
Control Test		0.081	03:09	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 05/30/2018 20:21

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SZYMANSKI,JACEK,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMIN,NAZIR
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: VOIGHT

DATE/TIME OF TEST: 05/28/2018 00:53:31

Diagnostics Check		OK	00:50	
Air Blank		0.000	00:51	
Control Test		0.081	00:51	
Air Blank		0.000	00:51	
Subject Sample	# 1		00:52	Subject Test Refused
Breath Volume		0.000	00:52	
Air Blank		0.000	00:52	
Control Test		0.081	00:52	
Air Blank		0.000	00:53	
Diagnostics Check		OK	00:53	

Data Download Date/Time: 05/30/2018 20:21

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,STEPHAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLE,STEPHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 05/29/2018 01:57:01

Diagnostics Check		OK	01:50
Air Blank		0.000	01:51
Control Test		0.079	01:51
Air Blank		0.000	01:52
Subject Sample	# 1	0.190	01:52
Breath Volume		2.316	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:54
Subject Sample	# 2	0.197	01:55
Breath Volume		2.187	01:55
Air Blank		0.000	01:56
Control Test		0.079	01:56
Air Blank		0.000	01:56
Diagnostics Check		OK	01:57

Data Download Date/Time: 05/30/2018 20:21

Printed: 07 August 2018

COBRA ©1997-2014 CMI, Inc.