

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,ANDREW,L
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WEST,JAMES
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: MAIONE

DATE/TIME OF TEST: 06/01/2018 03:28:08

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.187	03:23
Breath Volume	3.500	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.178	03:26
Breath Volume	3.484	03:26
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 06/21/2018 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 06/01/2018 23:02:43

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.128	22:58
Breath Volume	3.277	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.130	23:01
Breath Volume	3.289	23:01
Air Blank	0.000	23:01
Control Test	0.079	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 06/21/2018 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DIXON, WESLEY, B

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: DIXON, WESLEY

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06717080A6

DRY GAS EXPIRY DATE: 05/05/2019

VIOLATION CODE: DUI

ARREST TIME: 07:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 06/02/2018 09:46:33

Diagnostics Check	OK	09:40
Air Blank	0.000	09:40
Control Test	0.080	09:41
Air Blank	0.000	09:41
Subject Sample # 1	0.000	09:42
Breath Volume	1.250	09:42
Air Blank	0.000	09:42
Air Blank	0.000	09:44
Subject Sample # 2	0.000	09:45
Breath Volume	3.582	09:45
Air Blank	0.000	09:45
Control Test	0.079	09:45
Air Blank	0.000	09:46
Diagnostics Check	OK	09:46

Data Download Date/Time: 06/21/2018 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODWARD,KYLER,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WOODWARD,KYLER
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: IRIAS-LOPEZ

DATE/TIME OF TEST: 06/05/2018 00:25:22

Diagnostics Check	OK	00:22	
Air Blank	0.000	00:22	
Control Test	0.080	00:23	
Air Blank	0.000	00:23	
Subject Sample	# 1	00:24	Subject Test Refused
Breath Volume	0.000	00:24	
Air Blank	0.000	00:24	
Control Test	0.080	00:24	
Air Blank	0.000	00:25	
Diagnostics Check	OK	00:25	

Data Download Date/Time: 06/21/2018 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: N425010757160

DATE/TIME OF TEST: 06/08/2018 23:13:51

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.142	23:09
Breath Volume	1.578	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:11
Subject Sample # 2	0.130	23:12
Breath Volume	1.261	23:12
Air Blank	0.000	23:12
Control Test	0.080	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 06/21/2018 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVERETT, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: EVERETT, MATTHEW
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/09/2018 19:41:30

Diagnostics Check	OK	19:35
Air Blank	0.000	19:35
Control Test	0.079	19:36
Air Blank	0.000	19:36
Subject Sample # 1	0.270	19:37
Breath Volume	2.746	19:37
Air Blank	0.000	19:37
Air Blank	0.000	19:39
Subject Sample # 2	0.270	19:40
Breath Volume	2.683	19:40
Air Blank	0.000	19:40
Control Test	0.079	19:40
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 06/21/2018 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 06/10/2018 05:09:14

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.079	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.193	05:04
Breath Volume		1.230	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.188	05:07
Breath Volume		1.382	05:07
Air Blank		0.000	05:08
Control Test		0.078	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 06/21/2018 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LECKINGER, MICHAEL, E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DEAN, DONALD
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 06/10/2018 23:36:00

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.079	23:34	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:35	
Control Test	0.080	23:35	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 06/21/2018 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 06/15/2018 03:45:55

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.090	03:41
Breath Volume	2.828	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.085	03:44
Breath Volume	2.003	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 06/21/2018 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 06/17/2018 02:31:55

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.163	02:27
Breath Volume	1.707	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.159	02:30
Breath Volume	1.820	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 06/21/2018 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERESSE,A,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BRANNON,BILL
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:01

SUBJECT'S LAST NAME: MITCHEM

DATE/TIME OF TEST: 06/18/2018 12:30:21

Diagnostics Check	OK	12:24
Air Blank	0.000	12:24
Control Test	0.079	12:24
Air Blank	0.000	12:25
Subject Sample # 1	0.180	12:25
Breath Volume	4.218	12:25
Air Blank	0.000	12:26
Air Blank	0.000	12:28
Subject Sample # 2	0.171	12:28
Breath Volume	4.421	12:28
Air Blank	0.000	12:29
Control Test	0.079	12:29
Air Blank	0.000	12:30
Diagnostics Check	OK	12:30

Data Download Date/Time: 06/21/2018 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: O'CONNOR,STEPHEN,L

BREATH TEST OPERATOR AGENCY: Leon County SO

ARRESTING OFFICER: O'CONNOR,STEPHEN

ARRESTING OFFICER AGENCY: Leon County SO

LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06717080A6

DRY GAS EXPIRY DATE: 05/05/2019

VIOLATION CODE: DUI

ARREST TIME: 16:22

GENDER: Female

OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 06/21/2018 17:35:09

Diagnostics Check	OK	17:28
Air Blank	0.000	17:29
Control Test	0.079	17:29
Air Blank	0.000	17:30
Subject Sample # 1	0.177	17:30
Breath Volume	1.125	17:30
Air Blank	0.000	17:31
Air Blank	0.000	17:33
Subject Sample # 2	0.180	17:33
Breath Volume	1.351	17:33
Air Blank	0.000	17:34
Control Test	0.079	17:34
Air Blank	0.000	17:35
Diagnostics Check	OK	17:35

Data Download Date/Time: 07/30/2018 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: LALLI,TAREN
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: GIBBONS

DATE/TIME OF TEST: 06/22/2018 02:52:04

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.161	02:47
Breath Volume	2.722	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.161	02:50
Breath Volume	1.542	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:52

Data Download Date/Time: 07/30/2018 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: BOYCE

DATE/TIME OF TEST: 06/23/2018 00:57:59

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.079	00:50
Air Blank		0.000	00:50
Subject Sample	# 1	0.253	00:53
Breath Volume		1.226	00:53
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.263	00:56
Breath Volume		1.417	00:56
Air Blank		0.000	00:57
Control Test		0.078	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 07/30/2018 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: RORY

DATE/TIME OF TEST: 06/23/2018 03:28:07

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.104	03:23
Breath Volume	1.964	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.096	03:26
Breath Volume	1.136	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 07/30/2018 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: TODD,ROBERT
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 06/23/2018 05:10:52

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.080	05:03
Air Blank	0.000	05:04
Subject Sample # 1	0.213	05:05
Breath Volume	1.109	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:07
Subject Sample # 2	0.195	05:09
Breath Volume	1.113	05:09
Air Blank	0.000	05:09
Control Test	0.079	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 07/30/2018 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 06/26/2018 02:05:35

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.000	02:01
Breath Volume	2.386	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.000	02:04
Breath Volume	2.207	02:04
Air Blank	0.000	02:04
Control Test	0.080	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 07/30/2018 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVERETT, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: EVERETT, MATTHEW
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: GIBBS

DATE/TIME OF TEST: 06/27/2018 00:43:50

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.079	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.218	00:39
Breath Volume	2.375	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:41
Subject Sample # 2	0.231	00:42
Breath Volume	1.882	00:42
Air Blank	0.000	00:42
Control Test	0.078	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 07/30/2018 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVERETT, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: EVERETT, MATTHEW
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 06/27/2018 21:20:59

Diagnostics Check	OK	21:14
Air Blank	0.000	21:15
Control Test	0.080	21:15
Air Blank	0.000	21:16
Subject Sample # 1	0.172	21:16
Breath Volume	1.929	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.170	21:19
Breath Volume	2.261	21:19
Air Blank	0.000	21:20
Control Test	0.079	21:20
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 07/30/2018 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K

BREATH TEST OPERATOR AGENCY: Tallahassee PD

ARRESTING OFFICER: NORTHWAY,GEORGIA

ARRESTING OFFICER AGENCY: Tallahassee PD

LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06717080A6

DRY GAS EXPIRY DATE: 05/05/2019

VIOLATION CODE: DUI

ARREST TIME: 04:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:41

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 06/28/2018 06:15:52

Diagnostics Check	OK	06:08
Air Blank	0.000	06:09
Control Test	0.079	06:09
Air Blank	0.000	06:10
Subject Sample # 1	0.181	06:11
Breath Volume	2.757	06:11
Air Blank	0.000	06:11
Air Blank	0.000	06:13
Subject Sample # 2	0.170	06:14
Breath Volume	2.492	06:14
Air Blank	0.000	06:14
Control Test	0.078	06:15
Air Blank	0.000	06:15
Diagnostics Check	OK	06:15

Data Download Date/Time: 07/30/2018 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: AZRAK,PAUL,M

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: AZRAK,PAUL

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06717080A6

DRY GAS EXPIRY DATE: 05/05/2019

VIOLATION CODE: DUI

ARREST TIME: 01:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 06/29/2018 03:00:04

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.213	02:55
Breath Volume	2.335	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.201	02:58
Breath Volume	1.886	02:58
Air Blank	0.000	02:59
Control Test	0.078	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	03:00

Data Download Date/Time: 07/30/2018 11:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSCITTO,C,E
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: DAVIDSON,J
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ANGARITA

DATE/TIME OF TEST: 06/30/2018 03:52:26

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.171	03:45
Breath Volume	1.640	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.148	03:48
Breath Volume	1.578	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 3	0.153	03:50
Breath Volume	1.406	03:50
Air Blank	0.000	03:51
Control Test	0.078	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 07/30/2018 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER, DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER, DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 06/30/2018 04:14:00

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.102	04:09
Breath Volume	2.152	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.097	04:12
Breath Volume	1.632	04:12
Air Blank	0.000	04:13
Control Test	0.078	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:14

Data Download Date/Time: 07/30/2018 11:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 840875

DRY GAS EXPIRY DATE: 09/19/2020

VIOLATION CODE: DUI

ARREST TIME: 23:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: GADSON

DATE/TIME OF TEST: 06/01/2018 00:00:24

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.078	23:55
Air Blank	0.000	23:55
Subject Sample # 1	0.054	23:56
Breath Volume	3.054	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.055	23:58
Breath Volume	1.917	23:58
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 06/07/2018 19:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: DIAZ-LLANOS

DATE/TIME OF TEST: 06/03/2018 02:21:36

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.082	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.241	02:17
Breath Volume		2.300	02:17
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.236	02:20
Breath Volume		1.535	02:20
Air Blank		0.000	02:20
Control Test		0.081	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 06/07/2018 19:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DONAHUE

DATE/TIME OF TEST: 06/22/2018 02:56:27

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.081	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.149	02:52
Breath Volume		1.746	02:52
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.144	02:55
Breath Volume		1.621	02:55
Air Blank		0.000	02:55
Control Test		0.078	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 07/18/2018 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 840875

DRY GAS EXPIRY DATE: 09/19/2020

VIOLATION CODE: DUI

ARREST TIME: 02:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: WHYTE

DATE/TIME OF TEST: 06/23/2018 03:22:50

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.083	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.076	03:18
Breath Volume	2.402	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.077	03:21
Breath Volume	1.980	03:21
Air Blank	0.000	03:21
Control Test	0.081	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 07/18/2018 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIDSON, JEREMIAH, D
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: TAYLOR, RICHARD
ARRESTING OFFICER AGENCY: Valparaiso PD
LAST AGENCY INSPECTION: 05/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: DYLEWSKI

DATE/TIME OF TEST: 06/01/2018 22:31:30

Diagnostics Check	OK	22:28	
Air Blank	0.000	22:29	
Control Test	0.081	22:29	
Air Blank	0.000	22:30	
Subject Sample	# 1	22:31	Sequence Aborted
Breath Volume	0.000	22:31	
Air Blank	0.000	22:31	

Data Download Date/Time: 08/29/2018 17:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIDSON, JEREMIAH, D
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: TAYLOR, RICHARD
ARRESTING OFFICER AGENCY: Valparaiso PD
LAST AGENCY INSPECTION: 05/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: DYLEWSKI

DATE/TIME OF TEST: 06/01/2018 22:36:59

Diagnostics Check	OK	22:33	
Air Blank	0.000	22:34	
Control Test	0.080	22:34	
Air Blank	0.000	22:35	
Subject Sample	# 1	22:35	Subject Test Refused
Breath Volume	0.000	22:35	
Air Blank	0.000	22:36	
Control Test	0.082	22:36	
Air Blank	0.000	22:36	
Diagnostics Check	OK	22:36	

Data Download Date/Time: 08/29/2018 17:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLARD,ELIJAH,D
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: DAVIDSON,JEREMIAH
ARRESTING OFFICER AGENCY: Valparaiso PD
LAST AGENCY INSPECTION: 05/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CARMICAL

DATE/TIME OF TEST: 06/03/2018 01:57:00

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:52	
Control Test	0.081	01:53	
Air Blank	0.000	01:53	
Subject Sample	# 1	01:55	Subject Test Refused
Breath Volume	0.000	01:55	
Air Blank	0.000	01:56	
Control Test	0.081	01:56	
Air Blank	0.000	01:56	
Diagnostics Check	OK	01:57	

Data Download Date/Time: 08/29/2018 17:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIDSON, JEREMIAH, D
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: POLLARD, ELIJAH
ARRESTING OFFICER AGENCY: Valparaiso PD
LAST AGENCY INSPECTION: 05/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: GALLEGOS

DATE/TIME OF TEST: 06/09/2018 05:38:18

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.080	05:33
Air Blank	0.000	05:33
Subject Sample # 1	0.182	05:33
Breath Volume	3.210	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.182	05:36
Breath Volume	3.660	05:36
Air Blank	0.000	05:37
Control Test	0.079	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 08/29/2018 17:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TONEY,CRAIG,P
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: TONEY,CRAIG
ARRESTING OFFICER AGENCY: Valparaiso PD
LAST AGENCY INSPECTION: 05/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 06/12/2018 20:10:26

Diagnostics Check		OK	20:04
Air Blank		0.000	20:04
Control Test		0.080	20:04
Air Blank		0.000	20:05
Subject Sample	# 1	0.230	20:05
Breath Volume		2.710	20:05
Air Blank		0.000	20:06
Air Blank		0.000	20:08
Subject Sample	# 2	0.229	20:08
Breath Volume		2.820	20:08
Air Blank		0.000	20:09
Control Test		0.080	20:09
Air Blank		0.000	20:10
Diagnostics Check		OK	20:10

Data Download Date/Time: 08/29/2018 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: POLLARD,ELIJAH
ARRESTING OFFICER AGENCY: Valparaiso PD
LAST AGENCY INSPECTION: 05/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 06/23/2018 23:33:27

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.218	23:29
Breath Volume	3.593	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.205	23:31
Breath Volume	1.683	23:31
Air Blank	0.000	23:32
Control Test	0.080	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 08/29/2018 17:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 06/03/2018 00:02:35

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.078	23:55
Air Blank	0.000	23:55
Subject Sample # 2	0.089	23:58
Breath Volume	1.363	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.088	00:01
Breath Volume	1.289	00:01
Air Blank	0.000	00:01
Control Test	0.077	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: PACKWOOD

DATE/TIME OF TEST: 06/03/2018 00:55:10

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.078	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.158	00:49
Breath Volume	1.441	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:51
Subject Sample # 2	0.149	00:53
Breath Volume	1.406	00:53
Air Blank	0.000	00:54
Control Test	0.078	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARY,ADAM,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GARY,ADAM
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: WALLER

DATE/TIME OF TEST: 06/03/2018 02:52:01

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.078	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.032	02:47
Breath Volume	3.742	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.033	02:50
Breath Volume	3.671	02:50
Air Blank	0.000	02:51
Control Test	0.078	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:52

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBOWSKI,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GROGAN,STEVEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:07

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 06/03/2018 06:34:26

Diagnostics Check	OK	06:28
Air Blank	0.000	06:28
Control Test	0.079	06:28
Air Blank	0.000	06:29
Subject Sample # 1	0.112	06:29
Breath Volume	1.343	06:29
Air Blank	0.000	06:30
Air Blank	0.000	06:32
Subject Sample # 2	0.109	06:32
Breath Volume	1.992	06:32
Air Blank	0.000	06:33
Control Test	0.079	06:33
Air Blank	0.000	06:34
Diagnostics Check	OK	06:34

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: FETTERHOFF,AUGUSTUS
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: PEAKE

DATE/TIME OF TEST: 06/04/2018 01:31:53

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:19	
Control Test	0.079	01:19	
Air Blank	0.000	01:20	
Subject Sample	# 1 0.213	01:21	
Breath Volume	1.199	01:21	
Air Blank	0.000	01:22	
Air Blank	0.000	01:24	
Subject Sample	# 2 0.159	01:27	Volume Not Met
Breath Volume	1.117	01:27	
Air Blank	0.000	01:28	
Air Blank	0.000	01:29	
Subject Sample	# 3 0.216	01:30	
Breath Volume	1.222	01:30	
Air Blank	0.000	01:30	
Control Test	0.074	01:31	Control Outside Tolerance
Air Blank	0.000	01:31	

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: FETTERHOFF,AUGUSTUS
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: PEAKE

DATE/TIME OF TEST: 06/04/2018 01:42:16

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.077	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.185	01:36
Breath Volume	1.562	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.173	01:40
Breath Volume	1.113	01:40
Air Blank	0.000	01:41
Control Test	0.075	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: JOSEY

DATE/TIME OF TEST: 06/07/2018 03:38:50

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.054	03:34
Breath Volume	1.644	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.056	03:37
Breath Volume	2.359	03:37
Air Blank	0.000	03:37
Control Test	0.079	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARACLE, MATTHEW, D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LINDBLOM, ROBERT
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/09/2018 06:07:27

Diagnostics Check	OK	06:01
Air Blank	0.000	06:01
Control Test	0.079	06:01
Air Blank	0.000	06:02
Subject Sample # 1	0.215	06:02
Breath Volume	3.359	06:02
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 2	0.204	06:05
Breath Volume	2.789	06:05
Air Blank	0.000	06:06
Control Test	0.078	06:06
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLASHAW,HUNTER,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DILLASHAW,HUNTER
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: MCLAIN

DATE/TIME OF TEST: 06/10/2018 01:30:22

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.181	01:25
Breath Volume	1.480	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.186	01:28
Breath Volume	1.453	01:28
Air Blank	0.000	01:29
Control Test	0.077	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: PICKENS

DATE/TIME OF TEST: 06/15/2018 20:58:41

Diagnostics Check	OK	20:51
Air Blank	0.000	20:52
Control Test	0.078	20:52
Air Blank	0.000	20:52
Subject Sample # 1	0.091	20:54
Breath Volume	5.796	20:54
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.093	20:57
Breath Volume	4.414	20:57
Air Blank	0.000	20:57
Control Test	0.079	20:58
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROEBER, ANTHONY, L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROEBER, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 06/16/2018 03:09:12

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:57	
Control Test		0.079	02:57	
Air Blank		0.000	02:58	
Subject Sample	# 1	0.175	03:01	Volume Not Met
Breath Volume		0.445	03:01	
Air Blank		0.000	03:02	
Air Blank		0.000	03:04	
Subject Sample	# 2	0.189	03:07	Volume Not Met
Breath Volume		0.117	03:07	
Air Blank		0.000	03:08	
Control Test		0.076	03:08	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROEBER, ANTHONY, L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROEBER, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 06/16/2018 03:21:30

Diagnostics Check		OK	03:11	
Air Blank		0.000	03:11	
Control Test		0.076	03:12	
Air Blank		0.000	03:12	
Subject Sample	# 1	0.185	03:15	Volume Not Met
Breath Volume		0.164	03:15	
Air Blank		0.000	03:16	
Air Blank		0.000	03:18	
Subject Sample	# 2		03:19	Subject Test Refused
Breath Volume		0.105	03:19	
Air Blank		0.000	03:20	
Control Test		0.074	03:20	Control Outside Tolerance
Air Blank		0.000	03:21	

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: MIZELL

DATE/TIME OF TEST: 06/17/2018 21:09:49

Diagnostics Check	OK	21:03
Air Blank	0.000	21:04
Control Test	0.079	21:04
Air Blank	0.000	21:04
Subject Sample # 1	0.119	21:05
Breath Volume	2.199	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:07
Subject Sample # 2	0.115	21:08
Breath Volume	1.898	21:08
Air Blank	0.000	21:08
Control Test	0.078	21:09
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,SIR,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MORRIS,SIR
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SPARROW

DATE/TIME OF TEST: 06/18/2018 00:13:12

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:10	
Control Test	0.079	00:11	
Air Blank	0.000	00:11	
Subject Sample	# 1	00:11	Subject Test Refused
Breath Volume	0.000	00:11	
Air Blank	0.000	00:12	
Control Test	0.079	00:12	
Air Blank	0.000	00:13	
Diagnostics Check	OK	00:13	

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARY,ADAM,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GARY,ADAM
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/21/2018 21:34:37

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.078	21:29
Air Blank	0.000	21:29
Subject Sample # 1	0.082	21:30
Breath Volume	1.496	21:30
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.075	21:33
Breath Volume	2.964	21:33
Air Blank	0.000	21:33
Control Test	0.078	21:34
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: PEEKS

DATE/TIME OF TEST: 06/23/2018 23:39:37

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.080	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.000	23:34
Breath Volume	1.230	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.000	23:38
Breath Volume	1.902	23:38
Air Blank	0.000	23:38
Control Test	0.080	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,BRANDT,J
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: KESSINGER,BRANDT
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: GARNER

DATE/TIME OF TEST: 06/27/2018 01:29:33

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:28	
Control Test	0.074	01:28	Control Outside Tolerance
Air Blank	0.000	01:29	

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER, BRANDT, J
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: KESSINGER, BRANDT
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: GARNER

DATE/TIME OF TEST: 06/27/2018 01:36:13

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:32	
Control Test	0.077	01:33	
Air Blank	0.000	01:33	
Subject Sample	# 1 0.318	01:34	
Breath Volume	6.285	01:34	
Air Blank		01:34	Improper Sample
Air Blank	0.000	01:35	
Control Test	0.077	01:35	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,BRANDT,J
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: KESSINGER,BRANDT
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: GARNER

DATE/TIME OF TEST: 06/27/2018 01:44:49

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.076	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.322	01:40
Breath Volume	4.832	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:42
Subject Sample # 2	0.309	01:43
Breath Volume	4.804	01:43
Air Blank	0.000	01:43
Control Test	0.075	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLASHAW,HUNTER,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DILLASHAW,HUNTER
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: SCROGGINS

DATE/TIME OF TEST: 06/29/2018 05:21:21

Diagnostics Check	OK	05:15
Air Blank	0.000	05:15
Control Test	0.078	05:16
Air Blank	0.000	05:16
Subject Sample # 1	0.000	05:17
Breath Volume	1.570	05:17
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample # 2	0.000	05:19
Breath Volume	1.667	05:19
Air Blank	0.000	05:20
Control Test	0.078	05:20
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 06/29/2018 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIPTON,SHANE,B
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: SCOTT,TRAVIS
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 06/17/2018 20:16:26

Diagnostics Check	OK	20:10
Air Blank	0.000	20:10
Control Test	0.078	20:11
Air Blank	0.000	20:11
Subject Sample # 1	0.101	20:11
Breath Volume	1.796	20:11
Air Blank	0.000	20:12
Air Blank	0.000	20:14
Subject Sample # 2	0.100	20:14
Breath Volume	1.949	20:14
Air Blank	0.000	20:15
Control Test	0.078	20:15
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

Data Download Date/Time: 07/07/2018 16:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,BRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SABO,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 06/01/2018 02:11:29

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.078	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.155	02:07
Breath Volume	1.183	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.152	02:09
Breath Volume	2.199	02:09
Air Blank	0.000	02:10
Control Test	0.078	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 06/25/2018 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHARDSON, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PINO-DIAZ, VICTOR
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: O'LEARY

DATE/TIME OF TEST: 06/09/2018 23:33:24

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:22	
Control Test	0.079	23:22	
Air Blank	0.000	23:23	
Subject Sample	# 1	23:26	No Sample Provided
Breath Volume	0.000	23:26	
Air Blank	0.000	23:26	
Air Blank	0.000	23:28	
Subject Sample	# 2	23:32	No Sample Provided
Breath Volume	0.000	23:32	
Air Blank	0.000	23:32	
Control Test	0.080	23:32	
Air Blank	0.000	23:33	
Diagnostics Check	OK	23:33	

Data Download Date/Time: 06/25/2018 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRINSKO,THOMAS,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 06/10/2018 01:47:18

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:38	
Control Test	0.078	01:38	
Air Blank	0.000	01:39	
Subject Sample	# 1 0.195	01:42	Volume Not Met
Breath Volume	0.246	01:42	
Air Blank	0.000	01:43	
Air Blank	0.000	01:45	
Subject Sample	# 2	01:45	Subject Test Refused
Breath Volume	0.000	01:45	
Air Blank	0.000	01:46	
Control Test	0.077	01:46	
Air Blank	0.000	01:47	
Diagnostics Check	OK	01:47	

Data Download Date/Time: 06/25/2018 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,ERICA,M
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LINK,ERICA
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/11/2018 21:15:26

Diagnostics Check	OK	21:04	
Air Blank	0.000	21:04	
Control Test	0.078	21:04	
Air Blank	0.000	21:05	
Subject Sample	# 1	21:08	No Sample Provided
Breath Volume	0.000	21:08	
Air Blank	0.000	21:08	
Air Blank	0.000	21:10	
Subject Sample	# 2	21:14	No Sample Provided
Breath Volume	0.000	21:14	
Air Blank	0.000	21:14	
Control Test	0.080	21:14	
Air Blank	0.000	21:15	
Diagnostics Check	OK	21:15	

Data Download Date/Time: 06/25/2018 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARA,ALEXANDER,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: LARA,ALEXANDER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: LENNON LAUX

DATE/TIME OF TEST: 06/14/2018 19:59:01

Diagnostics Check	OK	19:52
Air Blank	0.000	19:53
Control Test	0.078	19:53
Air Blank	0.000	19:53
Subject Sample # 1	0.253	19:54
Breath Volume	2.089	19:54
Air Blank	0.000	19:55
Air Blank	0.000	19:56
Subject Sample # 2	0.242	19:57
Breath Volume	2.375	19:57
Air Blank	0.000	19:58
Control Test	0.079	19:58
Air Blank	0.000	19:58
Diagnostics Check	OK	19:59

Data Download Date/Time: 06/25/2018 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNS, WILLIAM, A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: GUYTON, JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 06/17/2018 21:03:27

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:01	
Control Test	0.079	21:01	
Air Blank	0.000	21:01	
Subject Sample	# 1	21:02	Subject Test Refused
Breath Volume	0.000	21:02	
Air Blank	0.000	21:02	
Control Test	0.079	21:02	
Air Blank	0.000	21:03	
Diagnostics Check	OK	21:03	

Data Download Date/Time: 06/25/2018 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHARDSON,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: ACOSTA,ALEX
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: DAVALOS

DATE/TIME OF TEST: 06/24/2018 20:59:05

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:48	
Control Test	0.078	20:48	
Air Blank	0.000	20:49	
Subject Sample	# 1	20:52	No Sample Provided
Breath Volume	0.000	20:52	
Air Blank	0.000	20:52	
Air Blank	0.000	20:54	
Subject Sample	# 2	20:57	No Sample Provided
Breath Volume	0.000	20:57	
Air Blank	0.000	20:58	
Control Test	0.079	20:58	
Air Blank	0.000	20:59	
Diagnostics Check	OK	20:59	

Data Download Date/Time: 06/25/2018 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ESNEUL,MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: GRINOLDS

DATE/TIME OF TEST: 06/06/2018 19:32:32

Diagnostics Check		OK	19:16	
Air Blank		0.000	19:16	
Control Test		0.080	19:17	
Air Blank		0.000	19:17	
Subject Sample	# 1	0.000	19:20	Volume Not Met
Breath Volume		0.402	19:20	
Air Blank		0.000	19:21	
Air Blank		0.000	19:23	
Subject Sample	# 2	0.000	19:25	
Breath Volume		1.585	19:25	
Air Blank		0.000	19:26	
Air Blank		0.000	19:28	
Subject Sample	# 3	0.000	19:31	Volume Not Met
Breath Volume		0.394	19:31	
Air Blank		0.000	19:31	
Control Test		0.080	19:31	
Air Blank		0.000	19:32	
Diagnostics Check		OK	19:32	

Data Download Date/Time: 06/28/2018 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CARLSON,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 06/15/2018 21:19:45

Diagnostics Check	OK	21:15	
Air Blank	0.000	21:16	
Control Test	0.080	21:16	
Air Blank	0.000	21:17	
Subject Sample	# 1	21:18	Subject Test Refused
Breath Volume	0.000	21:18	
Air Blank	0.000	21:18	
Control Test	0.079	21:19	
Air Blank	0.000	21:19	
Diagnostics Check	OK	21:19	

Data Download Date/Time: 06/28/2018 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ESNEUL,MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: RAHN

DATE/TIME OF TEST: 06/17/2018 05:28:02

Diagnostics Check	OK	05:21
Air Blank	0.000	05:22
Control Test	0.081	05:22
Air Blank	0.000	05:22
Subject Sample # 1	0.217	05:23
Breath Volume	3.062	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:25
Subject Sample # 2	0.224	05:26
Breath Volume	3.363	05:26
Air Blank	0.000	05:27
Control Test	0.080	05:27
Air Blank	0.000	05:27
Diagnostics Check	OK	05:28

Data Download Date/Time: 06/28/2018 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,CLEVELAND,D
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: GANUS,GLEN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 06/09/2018 03:04:16

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.078	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.278	02:59
Breath Volume	3.710	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.278	03:02
Breath Volume	2.789	03:02
Air Blank	0.000	03:03
Control Test	0.075	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 07/25/2018 17:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,CLEVELAND,D
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: HARRIS,EDWARD
ARRESTING OFFICER AGENCY: Lafayette County SO
LAST AGENCY INSPECTION: 05/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: HAMILL

DATE/TIME OF TEST: 06/14/2018 23:04:58

Diagnostics Check	OK	22:58
Air Blank	0.000	22:59
Control Test	0.079	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.277	23:00
Breath Volume	2.332	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:02
Subject Sample # 2	0.264	23:03
Breath Volume	2.539	23:03
Air Blank	0.000	23:04
Control Test	0.077	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 07/25/2018 17:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAS,ALFRED,A
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: DEAS,ALFRED
ARRESTING OFFICER AGENCY: Lafayette County SO
LAST AGENCY INSPECTION: 05/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: PARTEN

DATE/TIME OF TEST: 06/16/2018 05:39:46

Diagnostics Check	OK	05:33
Air Blank	0.000	05:34
Control Test	0.079	05:34
Air Blank	0.000	05:34
Subject Sample # 1	0.225	05:35
Breath Volume	1.664	05:35
Air Blank	0.000	05:36
Air Blank	0.000	05:37
Subject Sample # 2	0.211	05:38
Breath Volume	1.601	05:38
Air Blank	0.000	05:38
Control Test	0.077	05:39
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

Data Download Date/Time: 07/25/2018 17:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAS,ALFRED,A
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: DEAS,ALFRED
ARRESTING OFFICER AGENCY: Lafayette County SO
LAST AGENCY INSPECTION: 05/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/25/2018 22:21:56

Diagnostics Check	OK	22:15
Air Blank	0.000	22:15
Control Test	0.078	22:16
Air Blank	0.000	22:16
Subject Sample # 1	0.247	22:17
Breath Volume	1.953	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:19
Subject Sample # 2	0.246	22:20
Breath Volume	1.441	22:20
Air Blank	0.000	22:21
Control Test	0.076	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 07/25/2018 17:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATES,BRIAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BAURES,BRANDYN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: PARROTT

DATE/TIME OF TEST: 06/03/2018 01:01:48

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.082	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.177	00:57
Breath Volume	2.031	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.182	01:00
Breath Volume	1.945	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 06/29/2018 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: MILLIGAN,CONNIE
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 05/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: DAVIS JR

DATE/TIME OF TEST: 06/06/2018 22:42:25

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.082	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.209	22:37
Breath Volume	3.027	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.199	22:40
Breath Volume	2.199	22:40
Air Blank	0.000	22:41
Control Test	0.080	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 06/29/2018 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: MILLIGAN,CONNIE
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 05/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MILEWSKI

DATE/TIME OF TEST: 06/08/2018 03:57:04

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.081	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.178	03:52
Breath Volume	1.265	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.178	03:55
Breath Volume	2.281	03:55
Air Blank	0.000	03:56
Control Test	0.081	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:57

Data Download Date/Time: 06/29/2018 23:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELAN, MATTHEW, S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HERNANDEZ, ALEXIS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: MCFARLAN

DATE/TIME OF TEST: 06/14/2018 00:00:17

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.082	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.071	23:55
Breath Volume	1.285	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.072	23:58
Breath Volume	1.593	23:58
Air Blank	0.000	23:59
Control Test	0.082	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 06/29/2018 23:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELAN, MATTHEW, S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HERNANDEZ, ALEXIS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: R400542838390

DATE/TIME OF TEST: 06/14/2018 23:21:11

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.082	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.000	23:16
Breath Volume	3.078	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.000	23:19
Breath Volume	3.453	23:19
Air Blank	0.000	23:20
Control Test	0.080	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 06/29/2018 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLIE,JOHN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WOODS,CODY
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 05/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: TREJO RAMIREZ

DATE/TIME OF TEST: 06/16/2018 03:31:51

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.081	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.186	03:27
Breath Volume	5.542	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:29
Subject Sample # 2	0.181	03:30
Breath Volume	6.433	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 06/29/2018 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: MILLIGAN,CONNIE
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 05/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: KENNEY

DATE/TIME OF TEST: 06/23/2018 03:17:05

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.082	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.245	03:12
Breath Volume		2.484	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:14
Subject Sample	# 2	0.231	03:15
Breath Volume		1.960	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:16
Diagnostics Check		OK	03:17

Data Download Date/Time: 06/29/2018 23:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,ALFRED,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BEMBRY,JOSHUA
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: BUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: TAYLOR DICK

DATE/TIME OF TEST: 06/23/2018 23:10:25

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.081	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.046	23:06
Breath Volume	3.476	23:06
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.045	23:08
Breath Volume	3.500	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 06/29/2018 23:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATES,BRIAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BATES,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/28/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 08:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 06/29/2018 10:41:49

Diagnostics Check	OK	10:35
Air Blank	0.000	10:36
Control Test	0.082	10:36
Air Blank	0.000	10:37
Subject Sample # 1	0.000	10:37
Breath Volume	2.449	10:37
Air Blank	0.000	10:38
Air Blank	0.000	10:40
Subject Sample # 2	0.000	10:40
Breath Volume	1.988	10:40
Air Blank	0.000	10:40
Control Test	0.082	10:41
Air Blank	0.000	10:41
Diagnostics Check	OK	10:41

Data Download Date/Time: 06/29/2018 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT, JOSHUA, N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: JOOST, MARK
ARRESTING OFFICER AGENCY: Madison PD
LAST AGENCY INSPECTION: 05/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/02/2018 04:34:49

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:18	
Control Test	0.080	04:19	
Air Blank	0.000	04:19	
Subject Sample # 1	0.220	04:21	
Breath Volume	1.660	04:21	
Air Blank	0.000	04:22	
Air Blank	0.000	04:24	
Subject Sample # 2	0.166	04:27	Volume Not Met
Breath Volume	0.843	04:27	
Air Blank	0.000	04:28	
Air Blank	0.000	04:30	
Subject Sample # 3	0.198	04:33	Volume Not Met
Breath Volume	0.015	04:33	
Air Blank	0.000	04:33	
Control Test	0.074	04:34	Control Outside Tolerance
Air Blank	0.000	04:34	

Data Download Date/Time: 06/11/2018 10:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,DAVID,H
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SANDERS,ROBERT
ARRESTING OFFICER AGENCY: Madison PD
LAST AGENCY INSPECTION: 06/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 06/14/2018 00:36:45

Diagnostics Check	OK	00:22	
Air Blank	0.000	00:22	
Control Test	0.081	00:22	
Air Blank	0.000	00:23	
Subject Sample	# 1 0.154	00:24	
Breath Volume	1.316	00:24	
Air Blank	0.000	00:25	
Air Blank	0.000	00:27	
Subject Sample	# 2 0.125	00:30	Volume Not Met
Breath Volume	0.113	00:30	
Air Blank	0.000	00:31	
Air Blank	0.000	00:33	
Subject Sample	# 3 0.159	00:35	
Breath Volume	1.164	00:35	
Air Blank	0.000	00:35	
Control Test	0.079	00:36	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 07/17/2018 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALGOUT,DAVID,G
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: THAKOR,S
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ARGENTINE

DATE/TIME OF TEST: 06/07/2018 02:49:46

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1 0.137	02:48	Volume Not Met
Breath Volume	1.519	02:48	
Air Blank		02:48	Purge Fail
Air Blank		02:49	Purge Fail

Data Download Date/Time: 07/01/2018 22:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALGOUT, DAVID, G
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: THAKOR, S
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ARGENTINE

DATE/TIME OF TEST: 06/07/2018 02:57:29

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.079	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.131	02:53
Breath Volume	1.808	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.132	02:55
Breath Volume	1.632	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 07/01/2018 22:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:38

SUBJECT'S LAST NAME: GATLIN

DATE/TIME OF TEST: 06/07/2018 14:05:13

Diagnostics Check	OK	13:58
Air Blank	0.000	13:59
Control Test	0.083	13:59
Air Blank	0.000	14:00
Subject Sample # 1	0.000	14:00
Breath Volume	2.296	14:00
Air Blank	0.000	14:01
Air Blank	0.000	14:03
Subject Sample # 2	0.000	14:03
Breath Volume	3.218	14:03
Air Blank	0.000	14:04
Control Test	0.082	14:04
Air Blank	0.000	14:05
Diagnostics Check	OK	14:05

Data Download Date/Time: 07/01/2018 22:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSGROVE,DUSTIN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PARISI,TIMOTHY
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 06/07/2018 17:46:22

Diagnostics Check	OK	17:39
Air Blank	0.000	17:39
Control Test	0.082	17:39
Air Blank	0.000	17:40
Subject Sample # 1	0.259	17:40
Breath Volume	1.632	17:40
Air Blank	0.000	17:41
Air Blank	0.000	17:43
Subject Sample # 2	0.239	17:44
Breath Volume	2.871	17:44
Air Blank	0.000	17:45
Control Test	0.082	17:45
Air Blank	0.000	17:46
Diagnostics Check	OK	17:46

Data Download Date/Time: 07/01/2018 22:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENSON,DAVID,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: MCFADDEN,RYAN
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: SAPP

DATE/TIME OF TEST: 06/09/2018 01:36:34

Diagnostics Check		OK	01:24	
Air Blank		0.000	01:24	
Control Test		0.083	01:25	
Air Blank		0.000	01:25	
Subject Sample	# 1	0.107	01:28	Volume Not Met
Breath Volume		0.949	01:28	
Air Blank		0.000	01:29	
Air Blank		0.000	01:31	
Subject Sample	# 2	0.111	01:31	
Breath Volume		1.222	01:31	
Air Blank		0.000	01:32	
Air Blank		0.000	01:34	
Subject Sample	# 3	0.109	01:34	
Breath Volume		1.785	01:34	
Air Blank		0.000	01:35	
Control Test		0.081	01:35	
Air Blank		0.000	01:36	
Diagnostics Check		OK	01:36	

Data Download Date/Time: 07/01/2018 22:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELAN, MATTHEW, S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: OGELSBY, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: HENDRICKSON

DATE/TIME OF TEST: 06/10/2018 05:42:30

Diagnostics Check	OK	05:35
Air Blank	0.000	05:36
Control Test	0.083	05:36
Air Blank	0.000	05:37
Subject Sample # 1	0.142	05:37
Breath Volume	3.726	05:37
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample # 2	0.136	05:40
Breath Volume	3.523	05:40
Air Blank	0.000	05:41
Control Test	0.082	05:41
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

Data Download Date/Time: 07/01/2018 22:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELAN, MATTHEW, S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HERNANDEZ, ALEXIS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SHEPPARD

DATE/TIME OF TEST: 06/11/2018 00:28:54

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.083	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:27	
Control Test	0.083	00:28	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 07/01/2018 22:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENSON,DAVID,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WATFORD,TROY
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PANOZZO

DATE/TIME OF TEST: 06/24/2018 02:44:07

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.083	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:43	
Control Test	0.082	02:43	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 07/01/2018 22:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,JERRY,D
BREATH TEST OPERATOR AGENCY: Washington County SO
ARRESTING OFFICER: FRENCH,CURTIS
ARRESTING OFFICER AGENCY: Washington County SO
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 08:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: CRIPES

DATE/TIME OF TEST: 06/07/2018 22:24:18

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.081	22:17
Air Blank	0.000	22:17
Subject Sample # 1	0.241	22:18
Breath Volume	2.058	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.222	22:22
Breath Volume	1.839	22:22
Air Blank	0.000	22:23
Control Test	0.078	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 06/14/2018 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 10:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: NORRED
DATE/TIME OF TEST: 06/11/2018 11:47:58

Diagnostics Check	OK	11:40
Air Blank	0.000	11:41
Control Test	0.081	11:41
Air Blank	0.000	11:42
Subject Sample # 1	0.000	11:42
Breath Volume	1.207	11:42
Air Blank	0.000	11:43
Air Blank	0.000	11:45
Subject Sample # 2	0.000	11:46
Breath Volume	1.261	11:46
Air Blank	0.000	11:47
Control Test	0.080	11:47
Air Blank	0.000	11:47
Diagnostics Check	OK	11:47

Data Download Date/Time: 06/14/2018 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH,JASON,K
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FRENCH,JASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 06/30/2018 22:50:22

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.081	22:45
Air Blank	0.000	22:45
Subject Sample # 1	0.098	22:46
Breath Volume	1.796	22:46
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.104	22:48
Breath Volume	1.800	22:48
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 07/07/2018 17:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: STROTHER

DATE/TIME OF TEST: 06/02/2018 19:18:16

Diagnostics Check	OK	19:11
Air Blank	0.000	19:12
Control Test	0.081	19:12
Air Blank	0.000	19:13
Subject Sample # 1	0.123	19:13
Breath Volume	3.933	19:13
Air Blank	0.000	19:14
Air Blank	0.000	19:16
Subject Sample # 2	0.120	19:16
Breath Volume	3.773	19:16
Air Blank	0.000	19:17
Control Test	0.083	19:17
Air Blank	0.000	19:18
Diagnostics Check	OK	19:18

Data Download Date/Time: 06/15/2018 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIDD,JERMEL,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: ADAMS,BRYAN
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: AVILA

DATE/TIME OF TEST: 06/03/2018 16:46:36

Diagnostics Check	OK	16:40
Air Blank	0.000	16:41
Control Test	0.082	16:41
Air Blank	0.000	16:42
Subject Sample # 1	0.000	16:42
Breath Volume	5.113	16:42
Air Blank	0.000	16:42
Air Blank	0.000	16:44
Subject Sample # 2	0.000	16:45
Breath Volume	4.726	16:45
Air Blank	0.000	16:45
Control Test	0.083	16:46
Air Blank	0.000	16:46
Diagnostics Check	OK	16:46

Data Download Date/Time: 06/15/2018 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIERS, BLAINE, T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WEIERS, BLAINE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: KERN

DATE/TIME OF TEST: 06/09/2018 01:06:35

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.082	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.107	01:01
Breath Volume	1.335	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.108	01:05
Breath Volume	1.632	01:05
Air Blank	0.000	01:05
Control Test	0.082	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 06/15/2018 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: ROBINSON,ANDREW
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: KATZ

DATE/TIME OF TEST: 06/09/2018 18:56:36

Diagnostics Check	OK	18:50
Air Blank	0.000	18:50
Control Test	0.081	18:51
Air Blank	0.000	18:51
Subject Sample # 1	0.083	18:52
Breath Volume	2.785	18:52
Air Blank	0.000	18:52
Air Blank	0.000	18:54
Subject Sample # 2	0.083	18:55
Breath Volume	2.250	18:55
Air Blank	0.000	18:55
Control Test	0.082	18:56
Air Blank	0.000	18:56
Diagnostics Check	OK	18:56

Data Download Date/Time: 06/15/2018 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 06/09/2018 20:24:42

Diagnostics Check	OK	20:18
Air Blank	0.000	20:18
Control Test	0.082	20:19
Air Blank	0.000	20:19
Subject Sample # 1	0.129	20:20
Breath Volume	2.539	20:20
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.132	20:23
Breath Volume	1.929	20:23
Air Blank	0.000	20:23
Control Test	0.082	20:24
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

Data Download Date/Time: 06/15/2018 18:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIERS,BLAINE,T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WEIERS,BLAINE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: KEMBOI

DATE/TIME OF TEST: 06/10/2018 19:31:02

Diagnostics Check	OK	19:23
Air Blank	0.000	19:23
Control Test	0.082	19:24
Air Blank	0.000	19:24
Subject Sample # 1	0.271	19:25
Breath Volume	2.125	19:25
Air Blank	0.000	19:26
Air Blank	0.000	19:27
Subject Sample # 2	0.255	19:29
Breath Volume	2.609	19:29
Air Blank	0.000	19:30
Control Test	0.082	19:30
Air Blank	0.000	19:30
Diagnostics Check	OK	19:31

Data Download Date/Time: 06/15/2018 18:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIERS,BLAINE,T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WEIERS,BLAINE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: RAUSCHE

DATE/TIME OF TEST: 06/11/2018 01:17:15

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.081	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.200	01:12
Breath Volume		1.675	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.208	01:15
Breath Volume		1.867	01:15
Air Blank		0.000	01:16
Control Test		0.082	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 06/15/2018 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 06/12/2018 02:33:56

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.082	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.133	02:29
Breath Volume	3.828	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample # 2	0.130	02:32
Breath Volume	4.148	02:32
Air Blank	0.000	02:33
Control Test	0.082	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 06/15/2018 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: JOINER

DATE/TIME OF TEST: 06/13/2018 00:42:33

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.082	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.160	00:38
Breath Volume	2.609	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.158	00:41
Breath Volume	2.941	00:41
Air Blank	0.000	00:41
Control Test	0.082	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 06/15/2018 18:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: KRUEGER

DATE/TIME OF TEST: 06/16/2018 03:07:56

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.082	03:07	
Air Blank		03:07	RFI Detect
Air Blank	0.000	03:07	

Data Download Date/Time: 07/15/2018 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: KRUEGER

DATE/TIME OF TEST: 06/16/2018 03:17:16

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.076	03:12
Breath Volume	2.832	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.074	03:15
Breath Volume	3.277	03:15
Air Blank	0.000	03:16
Control Test	0.082	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 07/15/2018 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIERS, BLAINE, T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WEIERS, BLAINE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 06/20/2018 00:05:25

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.082	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.000	00:00
Breath Volume	3.187	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.000	00:04
Breath Volume	2.421	00:04
Air Blank	0.000	00:04
Control Test	0.083	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 07/15/2018 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: BRAMMER

DATE/TIME OF TEST: 06/20/2018 21:22:36

Diagnostics Check	OK	21:16
Air Blank	0.000	21:16
Control Test	0.083	21:16
Air Blank	0.000	21:17
Subject Sample # 1	0.135	21:18
Breath Volume	3.257	21:18
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.133	21:21
Breath Volume	3.875	21:21
Air Blank	0.000	21:21
Control Test	0.083	21:22
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 07/15/2018 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: HIPOLITO

DATE/TIME OF TEST: 06/21/2018 00:15:27

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.082	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.181	00:10
Breath Volume	3.089	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.181	00:13
Breath Volume	2.269	00:13
Air Blank	0.000	00:14
Control Test	0.082	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 07/15/2018 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTALDO,JOHN,J
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: CASTALDO,JOHN
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 06/23/2018 20:35:40

Diagnostics Check	OK	20:29
Air Blank	0.000	20:29
Control Test	0.082	20:29
Air Blank	0.000	20:30
Subject Sample # 1	0.152	20:30
Breath Volume	3.449	20:30
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.160	20:34
Breath Volume	2.882	20:34
Air Blank	0.000	20:34
Control Test	0.083	20:35
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 07/15/2018 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ATWELL II

DATE/TIME OF TEST: 06/24/2018 03:39:48

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.082	03:37	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:38	Subject Test Refused
Breath Volume	0.000	03:38	
Air Blank	0.000	03:38	
Control Test	0.082	03:39	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 07/15/2018 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIERS, BLAINE, T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WEIERS, BLAINE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: DELAMORA

DATE/TIME OF TEST: 06/24/2018 03:56:05

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.082	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.213	03:51
Breath Volume		2.769	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:53
Subject Sample	# 2	0.218	03:54
Breath Volume		2.804	03:54
Air Blank		0.000	03:55
Control Test		0.083	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 07/15/2018 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: RAWLINGS

DATE/TIME OF TEST: 06/26/2018 04:08:32

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.082	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Control Test	0.082	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 07/15/2018 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: PROVAN

DATE/TIME OF TEST: 06/26/2018 22:50:21

Diagnostics Check	OK	22:35	
Air Blank	0.000	22:35	
Control Test	0.082	22:36	
Air Blank	0.000	22:36	
Subject Sample	# 1 0.176	22:37	
Breath Volume	1.550	22:37	
Air Blank	0.000	22:38	
Air Blank	0.000	22:39	
Subject Sample	# 2 0.151	22:43	Volume Not Met
Breath Volume	1.054	22:43	
Air Blank	0.000	22:43	
Air Blank	0.000	22:45	
Subject Sample	# 3 0.176	22:48	Volume Not Met
Breath Volume	0.554	22:48	
Air Blank	0.000	22:49	
Control Test	0.082	22:49	
Air Blank	0.000	22:50	
Diagnostics Check	OK	22:50	

Data Download Date/Time: 07/15/2018 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMS,KYLE,N
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SIMMS,KYLE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: NORTHRUP

DATE/TIME OF TEST: 06/26/2018 23:21:31

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.081	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.000	23:17
Breath Volume	1.718	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.000	23:20
Breath Volume	2.109	23:20
Air Blank	0.000	23:20
Control Test	0.082	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 07/15/2018 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: GOOCH

DATE/TIME OF TEST: 06/27/2018 01:39:23

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.082	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.124	01:34
Breath Volume	1.421	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.123	01:37
Breath Volume	1.394	01:37
Air Blank	0.000	01:38
Control Test	0.081	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 07/15/2018 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: WALLER

DATE/TIME OF TEST: 06/29/2018 22:23:37

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.082	22:17
Air Blank	0.000	22:17
Subject Sample # 1	0.177	22:18
Breath Volume	1.785	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.170	22:22
Breath Volume	1.898	22:22
Air Blank	0.000	22:22
Control Test	0.083	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 07/15/2018 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEATY,BRYAN,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: NELSON,JEREMIAH
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: DAKIS

DATE/TIME OF TEST: 06/29/2018 20:16:56

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.081	20:11
Air Blank	0.000	20:12
Subject Sample # 1	0.000	20:12
Breath Volume	2.410	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.000	20:15
Breath Volume	2.570	20:15
Air Blank	0.000	20:15
Control Test	0.082	20:16
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

Data Download Date/Time: 07/12/2018 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEATY,BRYAN,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: CUEVAS,BENJAMIN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 06/30/2018 01:31:56

Diagnostics Check	Fail	01:31	Voltage/Current Test
Air Blank	0.000	01:31	

Data Download Date/Time: 07/12/2018 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEATY,BRYAN,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: CUEVAS, JR,BENJAMIN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 06/30/2018 01:33:42

Diagnostics Check	Fail	01:33	Voltage/Current Test
Air Blank	0.000	01:33	

Data Download Date/Time: 07/12/2018 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEATY,BRYAN,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: CUEVAS, JR,BENJAMIN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 06/30/2018 01:36:13

Diagnostics Check	Fail	01:35	Voltage/Current Test
Air Blank	0.000	01:36	

Data Download Date/Time: 07/12/2018 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HUNTER,RANDY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 06/02/2018 00:19:33

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.080	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.081	00:15
Breath Volume	3.246	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.080	00:18
Breath Volume	2.757	00:18
Air Blank	0.000	00:18
Control Test	0.081	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 06/15/2018 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUEDE,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: SCHALK

DATE/TIME OF TEST: 06/03/2018 06:01:47

Diagnostics Check	OK	05:55
Air Blank	0.000	05:56
Control Test	0.079	05:56
Air Blank	0.000	05:56
Subject Sample # 1	0.123	05:57
Breath Volume	1.855	05:57
Air Blank	0.000	05:58
Air Blank	0.000	05:59
Subject Sample # 2	0.121	06:00
Breath Volume	1.253	06:00
Air Blank	0.000	06:00
Control Test	0.078	06:01
Air Blank	0.000	06:01
Diagnostics Check	OK	06:01

Data Download Date/Time: 06/15/2018 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMLIN,BRETT,W
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HAMLIN,BRETT
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/04/2018 09:12:18

Diagnostics Check		OK	09:05
Air Blank		0.000	09:06
Control Test		0.079	09:06
Air Blank		0.000	09:06
Subject Sample	# 1	0.328	09:07
Breath Volume		1.222	09:07
Air Blank		0.000	09:08
Air Blank		0.000	09:10
Subject Sample	# 2	0.339	09:10
Breath Volume		1.375	09:10
Air Blank		0.000	09:11
Control Test		0.078	09:11
Air Blank		0.000	09:12
Diagnostics Check		OK	09:12

Data Download Date/Time: 06/15/2018 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: FORGUES

DATE/TIME OF TEST: 06/04/2018 22:08:06

Diagnostics Check	OK	22:01
Air Blank	0.000	22:02
Control Test	0.079	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.000	22:03
Breath Volume	2.507	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.000	22:06
Breath Volume	2.289	22:06
Air Blank	0.000	22:07
Control Test	0.079	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 06/15/2018 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAGLIONI,DAVID,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: CAGLIONI,DAVID
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: HATCH

DATE/TIME OF TEST: 06/06/2018 08:07:14

Diagnostics Check	OK	08:01
Air Blank	0.000	08:01
Control Test	0.080	08:01
Air Blank	0.000	08:02
Subject Sample # 1	0.000	08:02
Breath Volume	3.871	08:02
Air Blank	0.000	08:03
Air Blank	0.000	08:05
Subject Sample # 2	0.000	08:05
Breath Volume	3.375	08:05
Air Blank	0.000	08:06
Control Test	0.079	08:06
Air Blank	0.000	08:07
Diagnostics Check	OK	08:07

Data Download Date/Time: 06/15/2018 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WALTENBURG,MARCUS
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 06/06/2018 22:47:14

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.080	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.000	22:42
Breath Volume	2.550	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.000	22:45
Breath Volume	2.691	22:45
Air Blank	0.000	22:46
Control Test	0.078	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 06/15/2018 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WALTENBURG,MARCUS
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: VAZQUEZ RAMIREZ

DATE/TIME OF TEST: 06/07/2018 20:55:20

Diagnostics Check	OK	20:49
Air Blank	0.000	20:49
Control Test	0.080	20:49
Air Blank	0.000	20:50
Subject Sample # 1	0.032	20:50
Breath Volume	3.156	20:50
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.028	20:53
Breath Volume	2.515	20:53
Air Blank	0.000	20:54
Control Test	0.078	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 06/15/2018 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 06/07/2018 22:00:16

Diagnostics Check	OK	21:53
Air Blank	0.000	21:54
Control Test	0.080	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.112	21:55
Breath Volume	1.417	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.110	21:58
Breath Volume	1.441	21:58
Air Blank	0.000	21:59
Control Test	0.079	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 06/15/2018 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVIL,LUXERNE,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: VILLANO,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 06/08/2018 16:27:07

Diagnostics Check	OK	16:19
Air Blank	0.000	16:20
Control Test	0.080	16:20
Air Blank	0.000	16:21
Subject Sample # 1	0.070	16:22
Breath Volume	1.453	16:22
Air Blank	0.000	16:23
Air Blank	0.000	16:25
Subject Sample # 2	0.073	16:25
Breath Volume	1.121	16:25
Air Blank	0.000	16:26
Control Test	0.079	16:26
Air Blank	0.000	16:27
Diagnostics Check	OK	16:27

Data Download Date/Time: 06/15/2018 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DEBLIECK,BENJAMIN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 06/08/2018 23:46:45

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.079	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.071	23:42
Breath Volume	1.117	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.067	23:45
Breath Volume	2.011	23:45
Air Blank	0.000	23:45
Control Test	0.079	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 06/15/2018 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WALTENBURG,MARCUS
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: SIRAK

DATE/TIME OF TEST: 06/10/2018 00:55:32

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.146	00:50
Breath Volume	1.589	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.142	00:53
Breath Volume	1.445	00:53
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 06/15/2018 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HALL,RICHARD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 06/10/2018 05:20:29

Diagnostics Check	OK	05:13
Air Blank	0.000	05:14
Control Test	0.080	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.000	05:15
Breath Volume	2.691	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.000	05:19
Breath Volume	2.640	05:19
Air Blank	0.000	05:19
Control Test	0.079	05:19
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 06/15/2018 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYER,SEBONAIRE,
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: MAURANCY,FRANNY
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BOYER

DATE/TIME OF TEST: 06/11/2018 03:01:39

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.150	02:57
Breath Volume	1.738	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.145	03:00
Breath Volume	1.628	03:00
Air Blank	0.000	03:00
Control Test	0.077	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 06/15/2018 08:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRE,GAETGENS,
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: MAURANCY,FRANNY
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BOYER

DATE/TIME OF TEST: 06/11/2018 03:12:34

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.154	03:07
Breath Volume	1.960	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.147	03:10
Breath Volume	1.898	03:10
Air Blank	0.000	03:11
Control Test	0.077	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 06/15/2018 08:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTEBURG,MARCUS,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: ELLISON,MATTHEW
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 06/16/2018 00:07:23

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.080	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.156	00:02
Breath Volume	1.386	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.158	00:05
Breath Volume	2.398	00:05
Air Blank	0.000	00:06
Control Test	0.077	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 07/12/2018 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 06/19/2018 05:20:08

Diagnostics Check	OK	05:13
Air Blank	0.000	05:14
Control Test	0.080	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.000	05:15
Breath Volume	1.460	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.000	05:18
Breath Volume	1.441	05:18
Air Blank	0.000	05:19
Control Test	0.079	05:19
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 07/12/2018 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 06/21/2018 02:56:20

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.162	02:51
Breath Volume	1.562	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.150	02:54
Breath Volume	2.679	02:54
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 07/12/2018 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 06/22/2018 02:12:47

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.073	02:07
Breath Volume	1.687	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.072	02:11
Breath Volume	2.347	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 07/12/2018 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,JAMES,B
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SHERMETARO,MATRIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: BUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: TIDWELL

DATE/TIME OF TEST: 06/23/2018 23:14:34

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:11	
Control Test	0.080	23:11	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:13	Subject Test Refused
Breath Volume	0.000	23:13	
Air Blank	0.000	23:13	
Control Test	0.080	23:13	
Air Blank	0.000	23:14	
Diagnostics Check	OK	23:14	

Data Download Date/Time: 07/12/2018 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: SANCHEZ ARRIBENO

DATE/TIME OF TEST: 06/24/2018 20:54:19

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.080	20:48
Air Blank	0.000	20:49
Subject Sample # 1	0.321	20:49
Breath Volume	1.230	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.311	20:52
Breath Volume	1.648	20:52
Air Blank	0.000	20:53
Control Test	0.079	20:53
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 07/12/2018 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 06/25/2018 22:20:07

Diagnostics Check		OK	22:08	
Air Blank		0.000	22:09	
Control Test		0.080	22:09	
Air Blank		0.000	22:10	
Subject Sample	# 1	0.000	22:13	Volume Not Met
Breath Volume		0.234	22:13	
Air Blank		0.000	22:13	
Air Blank		0.000	22:15	
Subject Sample	# 2		22:18	Subject Test Refused
Breath Volume		0.093	22:18	
Air Blank		0.000	22:19	
Control Test		0.080	22:19	
Air Blank		0.000	22:20	
Diagnostics Check		OK	22:20	

Data Download Date/Time: 07/12/2018 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: DARREL

DATE/TIME OF TEST: 06/26/2018 03:11:27

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.031	03:06
Breath Volume	1.953	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.026	03:09
Breath Volume	1.265	03:09
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 07/12/2018 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODRUFF,MARK,D
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: WOODRUFF,MARK
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: DEBERRY

DATE/TIME OF TEST: 06/26/2018 05:26:43

Diagnostics Check		OK	05:20
Air Blank		0.000	05:20
Control Test		0.080	05:21
Air Blank		0.000	05:21
Subject Sample	# 1	0.174	05:22
Breath Volume		2.246	05:22
Air Blank		0.000	05:22
Air Blank		0.000	05:24
Subject Sample	# 2	0.173	05:25
Breath Volume		1.750	05:25
Air Blank		0.000	05:25
Control Test		0.079	05:26
Air Blank		0.000	05:26
Diagnostics Check		OK	05:26

Data Download Date/Time: 07/12/2018 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELPH,JASON,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: PAQUETTE,JASON
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:06

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/26/2018 09:32:26

Diagnostics Check	OK	09:26
Air Blank	0.000	09:26
Control Test	0.080	09:27
Air Blank	0.000	09:27
Subject Sample # 1	0.000	09:28
Breath Volume	1.472	09:28
Air Blank	0.000	09:28
Air Blank	0.000	09:30
Subject Sample # 2	0.000	09:31
Breath Volume	1.457	09:31
Air Blank	0.000	09:31
Control Test	0.079	09:31
Air Blank	0.000	09:32
Diagnostics Check	OK	09:32

Data Download Date/Time: 07/12/2018 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGEL,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MANGEL,TIMOTHY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: BURROWS

DATE/TIME OF TEST: 06/26/2018 17:30:50

Diagnostics Check	OK	17:24
Air Blank	0.000	17:24
Control Test	0.081	17:25
Air Blank	0.000	17:25
Subject Sample # 1	0.000	17:26
Breath Volume	1.679	17:26
Air Blank	0.000	17:26
Air Blank	0.000	17:28
Subject Sample # 2	0.000	17:29
Breath Volume	1.378	17:29
Air Blank	0.000	17:29
Control Test	0.079	17:30
Air Blank	0.000	17:30
Diagnostics Check	OK	17:30

Data Download Date/Time: 07/12/2018 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: GORDINEER,CHRISTOPHER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: BAYER

DATE/TIME OF TEST: 06/26/2018 18:37:41

Diagnostics Check	OK	18:31
Air Blank	0.000	18:31
Control Test	0.080	18:32
Air Blank	0.000	18:32
Subject Sample # 1	0.000	18:33
Breath Volume	1.699	18:33
Air Blank	0.000	18:33
Air Blank	0.000	18:35
Subject Sample # 2	0.000	18:36
Breath Volume	1.707	18:36
Air Blank	0.000	18:36
Control Test	0.080	18:37
Air Blank	0.000	18:37
Diagnostics Check	OK	18:37

Data Download Date/Time: 07/12/2018 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGEL,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MANGEL,TIMOTHY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: INMON

DATE/TIME OF TEST: 06/26/2018 22:15:20

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.080	22:09
Air Blank	0.000	22:09
Subject Sample # 1	0.099	22:10
Breath Volume	1.152	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:12
Subject Sample # 2	0.101	22:13
Breath Volume	1.605	22:13
Air Blank	0.000	22:14
Control Test	0.079	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 07/12/2018 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMLIN,BRETT,W
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: ALFARO,ARTHUR
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:58

SUBJECT'S LAST NAME: MUDESTO

DATE/TIME OF TEST: 06/27/2018 15:58:34

Diagnostics Check	OK	15:51
Air Blank	0.000	15:52
Control Test	0.079	15:52
Air Blank	0.000	15:52
Subject Sample # 1	0.196	15:53
Breath Volume	1.679	15:53
Air Blank	0.000	15:54
Air Blank	0.000	15:56
Subject Sample # 2	0.191	15:57
Breath Volume	1.628	15:57
Air Blank	0.000	15:57
Control Test	0.079	15:57
Air Blank	0.000	15:58
Diagnostics Check	OK	15:58

Data Download Date/Time: 07/12/2018 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: CUEVAS,BENJAMIN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 06/30/2018 02:22:17

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.120	02:17
Breath Volume	2.332	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.113	02:20
Breath Volume	1.769	02:20
Air Blank	0.000	02:21
Control Test	0.078	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 07/12/2018 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,PAUL,G
BREATH TEST OPERATOR AGENCY: Gulf County SO
ARRESTING OFFICER: YOUNG,PAUL
ARRESTING OFFICER AGENCY: Gulf County SO
LAST AGENCY INSPECTION: 05/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: SHEARER

DATE/TIME OF TEST: 06/05/2018 07:21:41

Diagnostics Check	OK	07:14
Air Blank	0.000	07:15
Control Test	0.077	07:15
Air Blank	0.000	07:16
Subject Sample # 1	0.152	07:16
Breath Volume	3.457	07:16
Air Blank	0.000	07:17
Air Blank	0.000	07:19
Subject Sample # 2	0.137	07:20
Breath Volume	2.339	07:20
Air Blank	0.000	07:20
Control Test	0.077	07:21
Air Blank	0.000	07:21
Diagnostics Check	OK	07:21

Data Download Date/Time: 06/27/2018 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,SAMUEL,W
BREATH TEST OPERATOR AGENCY: Jasper PD
ARRESTING OFFICER: SMITH,LORNE
ARRESTING OFFICER AGENCY: Hamilton County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: AMAYA RODRIGUEZ

DATE/TIME OF TEST: 06/12/2018 03:34:11

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.078	03:26	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:29	No Sample Provided
Breath Volume	0.000	03:29	
Air Blank	0.000	03:30	
Air Blank	0.000	03:32	
Subject Sample	# 2	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:33	
Control Test	0.077	03:33	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 06/27/2018 01:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,SAMUEL,W
BREATH TEST OPERATOR AGENCY: Jasper PD
ARRESTING OFFICER: MCDONALD,SAMUEL
ARRESTING OFFICER AGENCY: Jasper PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:27

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 06/17/2018 07:54:17

Diagnostics Check	OK	07:47
Air Blank	0.000	07:47
Control Test	0.077	07:48
Air Blank	0.000	07:48
Subject Sample # 1	0.065	07:49
Breath Volume	1.578	07:49
Air Blank	0.000	07:49
Air Blank	0.000	07:51
Subject Sample # 2	0.069	07:52
Breath Volume	2.589	07:52
Air Blank	0.000	07:53
Control Test	0.077	07:53
Air Blank	0.000	07:54
Diagnostics Check	OK	07:54

Data Download Date/Time: 06/27/2018 01:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FULTON,DARIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FULTON,DARIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:44

SUBJECT'S LAST NAME: LANGSTON

DATE/TIME OF TEST: 06/25/2018 08:08:23

Diagnostics Check	OK	08:04	
Air Blank	0.000	08:04	
Control Test	0.080	08:05	
Air Blank	0.000	08:05	
Subject Sample	# 1	08:07	Interferent Detect
Breath Volume	3.003	08:07	
Air Blank	0.000	08:07	
Control Test	0.072	08:07	Interferent Detect
Air Blank	0.000	08:08	

Data Download Date/Time: 07/25/2018 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FULTON,DARIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FULTON,DARIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:44

SUBJECT'S LAST NAME: LANGSTON

DATE/TIME OF TEST: 06/25/2018 08:16:54

Diagnostics Check	OK	08:10
Air Blank	0.000	08:10
Control Test	0.080	08:10
Air Blank	0.000	08:11
Subject Sample # 1	0.131	08:12
Breath Volume	2.253	08:12
Air Blank	0.000	08:12
Air Blank	0.000	08:14
Subject Sample # 2	0.124	08:15
Breath Volume	1.480	08:15
Air Blank	0.000	08:16
Control Test	0.081	08:16
Air Blank	0.000	08:16
Diagnostics Check	OK	08:16

Data Download Date/Time: 07/25/2018 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE, CHARLES, M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE, CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: BOYINGTON

DATE/TIME OF TEST: 06/30/2018 17:07:27

Diagnostics Check	OK	17:01
Air Blank	0.000	17:01
Control Test	0.076	17:01
Air Blank	0.000	17:02
Subject Sample # 1	0.000	17:03
Breath Volume	2.125	17:03
Air Blank	0.000	17:03
Air Blank	0.000	17:05
Subject Sample # 2	0.000	17:06
Breath Volume	1.968	17:06
Air Blank	0.000	17:06
Control Test	0.076	17:06
Air Blank	0.000	17:07
Diagnostics Check	OK	17:07

Data Download Date/Time: 07/25/2018 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMONT,GREGORY,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: LAMONT,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 06/01/2018 12:56:51

Diagnostics Check	OK	12:50
Air Blank	0.000	12:50
Control Test	0.079	12:51
Air Blank	0.000	12:51
Subject Sample # 1	0.210	12:52
Breath Volume	2.023	12:52
Air Blank	0.000	12:53
Air Blank	0.000	12:54
Subject Sample # 2	0.208	12:55
Breath Volume	2.136	12:55
Air Blank	0.000	12:55
Control Test	0.080	12:56
Air Blank	0.000	12:56
Diagnostics Check	OK	12:56

Data Download Date/Time: 06/25/2018 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMONT,GREGORY,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DOUGLAS,JASON
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 12:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 06/01/2018 13:17:51

Diagnostics Check	OK	13:14	
Air Blank	0.000	13:14	
Control Test	0.078	13:15	
Air Blank	0.000	13:15	
Subject Sample	# 1 0.184	13:16	
Breath Volume	3.949	13:16	
Air Blank		13:17	Purge Fail
Air Blank		13:17	Purge Fail

Data Download Date/Time: 06/25/2018 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMONT,GREGORY,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DOUGLAS,JASON
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 12:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 06/01/2018 13:26:11

Diagnostics Check	OK	13:19
Air Blank	0.000	13:20
Control Test	0.079	13:20
Air Blank	0.000	13:21
Subject Sample # 1	0.181	13:21
Breath Volume	4.121	13:21
Air Blank	0.000	13:22
Air Blank	0.000	13:24
Subject Sample # 2	0.174	13:24
Breath Volume	4.207	13:24
Air Blank	0.000	13:25
Control Test	0.080	13:25
Air Blank	0.000	13:26
Diagnostics Check	OK	13:26

Data Download Date/Time: 06/25/2018 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLAZZO,C,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SOLLAZZO,C
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DUNNE

DATE/TIME OF TEST: 06/02/2018 02:13:44

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.080	02:11	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:12	
Control Test	0.080	02:13	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 06/25/2018 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GALLER,JESSICA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 06/07/2018 20:45:03

Diagnostics Check	OK	20:38
Air Blank	0.000	20:39
Control Test	0.080	20:39
Air Blank	0.000	20:39
Subject Sample # 1	0.231	20:40
Breath Volume	1.242	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:42
Subject Sample # 2	0.228	20:43
Breath Volume	2.496	20:43
Air Blank	0.000	20:44
Control Test	0.079	20:44
Air Blank	0.000	20:44
Diagnostics Check	OK	20:45

Data Download Date/Time: 06/25/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UMANA,RAUL,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: UMANA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 06/09/2018 01:32:31

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:26	
Control Test	0.080	01:26	
Air Blank	0.000	01:27	
Subject Sample # 1	0.224	01:28	
Breath Volume	2.636	01:28	
Air Blank	0.000	01:28	
Air Blank	0.000	01:30	
Subject Sample # 2	0.211	01:30	
Breath Volume	1.644	01:30	
Air Blank	0.000	01:31	
Control Test	0.071	01:31	Control Outside Tolerance
Air Blank	0.000	01:32	

Data Download Date/Time: 06/25/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: DOMINGUEZ LOPEZ

DATE/TIME OF TEST: 06/10/2018 04:47:42

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.080	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.064	04:43
Breath Volume	2.878	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.063	04:46
Breath Volume	1.910	04:46
Air Blank	0.000	04:46
Control Test	0.081	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 06/25/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: BALLESTEROS GARCIA

DATE/TIME OF TEST: 06/10/2018 04:56:50

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:53	
Control Test	0.080	04:53	
Air Blank	0.000	04:53	
Subject Sample	# 1 0.254	04:55	
Breath Volume	2.605	04:55	
Air Blank		04:55	RFI Detect
Air Blank	0.000	04:55	
Control Test	0.082	04:56	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:56	

Data Download Date/Time: 06/25/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: BALLESTEROS GARCIA

DATE/TIME OF TEST: 06/10/2018 05:05:36

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.080	05:00
Air Blank	0.000	05:00
Subject Sample # 1	0.256	05:01
Breath Volume	1.984	05:01
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.252	05:04
Breath Volume	1.804	05:04
Air Blank	0.000	05:04
Control Test	0.080	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 06/25/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,RANDALL,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: REED,RANDALL
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/10/2018 06:39:36

Diagnostics Check	OK	06:31
Air Blank	0.000	06:31
Control Test	0.080	06:32
Air Blank	0.000	06:32
Subject Sample # 1	0.031	06:33
Breath Volume	2.046	06:33
Air Blank	0.000	06:33
Air Blank	0.000	06:35
Subject Sample # 2	0.026	06:38
Breath Volume	1.457	06:38
Air Blank	0.000	06:38
Control Test	0.080	06:39
Air Blank	0.000	06:39
Diagnostics Check	OK	06:39

Data Download Date/Time: 06/25/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,ALYSSA,D
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: GARCIA,ALYSSA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 13:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:29

SUBJECT'S LAST NAME: HEMM

DATE/TIME OF TEST: 06/15/2018 14:06:14

Diagnostics Check	OK	14:00
Air Blank	0.000	14:00
Control Test	0.080	14:00
Air Blank	0.000	14:01
Subject Sample # 1	0.000	14:01
Breath Volume	2.183	14:01
Air Blank	0.000	14:02
Air Blank	0.000	14:04
Subject Sample # 2	0.000	14:04
Breath Volume	2.484	14:04
Air Blank	0.000	14:05
Control Test	0.081	14:05
Air Blank	0.000	14:06
Diagnostics Check	OK	14:06

Data Download Date/Time: 06/25/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GALLER,JESSICA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/16/2018 03:10:26

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:02	
Control Test	0.080	03:02	
Air Blank	0.000	03:03	
Subject Sample	# 1	03:06	No Sample Provided
Breath Volume	0.000	03:06	
Air Blank	0.000	03:06	
Air Blank	0.000	03:08	
Subject Sample	# 2	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Control Test	0.080	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 06/25/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MARK,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: STINEHOUR,A
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 64476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: MASTASCUSA

DATE/TIME OF TEST: 06/16/2018 17:16:36

Diagnostics Check	OK	17:10
Air Blank	0.000	17:10
Control Test	0.080	17:10
Air Blank	0.000	17:11
Subject Sample # 1	0.085	17:12
Breath Volume	1.859	17:12
Air Blank	0.000	17:12
Air Blank	0.000	17:14
Subject Sample # 2	0.085	17:15
Breath Volume	1.796	17:15
Air Blank	0.000	17:15
Control Test	0.080	17:16
Air Blank	0.000	17:16
Diagnostics Check	OK	17:16

Data Download Date/Time: 06/25/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,JOSEPH,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SILVA,JEAN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: MURPHY-GRIFFAW

DATE/TIME OF TEST: 06/01/2018 00:23:01

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.176	00:18
Breath Volume	2.031	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:20
Subject Sample # 2	0.177	00:21
Breath Volume	1.535	00:21
Air Blank	0.000	00:22
Control Test	0.077	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:23

Data Download Date/Time: 06/28/2018 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,JOSEPH,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: DODGE,ADAM
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 06/09/2018 02:00:53

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.078	01:58	
Air Blank	0.000	01:59	
Subject Sample	# 1	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	01:59	
Control Test	0.079	02:00	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 06/28/2018 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,JOSEPH,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SLAY,ADRIAN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 06/10/2018 02:30:57

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:28	
Control Test	0.079	02:28	
Air Blank	0.000	02:29	
Subject Sample	# 1	02:29	Subject Test Refused
Breath Volume	0.000	02:29	
Air Blank	0.000	02:30	
Control Test	0.079	02:30	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 06/28/2018 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,MARSHA,K
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NAVARRO,B
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: COOGAN

DATE/TIME OF TEST: 06/12/2018 21:05:20

Diagnostics Check	OK	21:02	
Air Blank	0.000	21:02	
Control Test	0.079	21:03	
Air Blank	0.000	21:03	
Subject Sample	# 1	21:03	Subject Test Refused
Breath Volume	0.000	21:03	
Air Blank	0.000	21:04	
Control Test	0.079	21:04	
Air Blank	0.000	21:05	
Diagnostics Check	OK	21:05	

Data Download Date/Time: 06/28/2018 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,MARSHA,K
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NAVARRO,BENJAMIN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: FORBERG

DATE/TIME OF TEST: 06/16/2018 00:13:40

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.079	00:08
Air Blank	0.000	00:08
Subject Sample # 1	0.202	00:09
Breath Volume	1.523	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.201	00:11
Breath Volume	1.406	00:11
Air Blank	0.000	00:12
Control Test	0.077	00:13
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 06/28/2018 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS, MARSHA, K
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: DOLAN, N
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: INTRAKAMHANG

DATE/TIME OF TEST: 06/17/2018 22:37:42

Diagnostics Check	OK	22:31
Air Blank	0.000	22:31
Control Test	0.079	22:32
Air Blank	0.000	22:32
Subject Sample # 1	0.096	22:33
Breath Volume	1.269	22:33
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.096	22:36
Breath Volume	2.316	22:36
Air Blank	0.000	22:36
Control Test	0.078	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 06/28/2018 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,MARSHA,K
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LAVALLE,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: FLACCO

DATE/TIME OF TEST: 06/17/2018 23:11:00

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.079	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.172	23:05
Breath Volume	1.277	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:07
Subject Sample # 2	0.183	23:09
Breath Volume	1.750	23:09
Air Blank	0.000	23:10
Control Test	0.079	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:11

Data Download Date/Time: 06/28/2018 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JARROD,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BARD,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 06/22/2018 03:53:06

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:50	
Control Test	0.080	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 06/28/2018 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BUTLER,ELIZABETH
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:04

SUBJECT'S LAST NAME: HONEYCUTT

DATE/TIME OF TEST: 06/24/2018 06:37:10

Diagnostics Check	OK	06:34	
Air Blank	0.000	06:34	
Control Test	0.079	06:34	
Air Blank	0.000	06:35	
Subject Sample	# 1	06:35	Subject Test Refused
Breath Volume	0.000	06:35	
Air Blank	0.000	06:36	
Control Test	0.078	06:36	
Air Blank	0.000	06:37	
Diagnostics Check	OK	06:37	

Data Download Date/Time: 06/28/2018 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL,JARROD,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NAVARRO,B
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: MCCLAIN

DATE/TIME OF TEST: 06/26/2018 01:18:36

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.197	01:13
Breath Volume	2.406	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.196	01:16
Breath Volume	1.757	01:16
Air Blank	0.000	01:17
Control Test	0.078	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 06/28/2018 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL,JARROD,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: PETTIT,ROBERT
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MAYO

DATE/TIME OF TEST: 06/30/2018 00:42:29

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.076	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.166	00:37
Breath Volume	2.566	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.163	00:40
Breath Volume	2.378	00:40
Air Blank	0.000	00:41
Control Test	0.076	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 07/20/2018 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAN,MICHAEL,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FERNADEZ,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: KRANICK

DATE/TIME OF TEST: 06/02/2018 02:30:07

Diagnostics Check		OK	02:20	
Air Blank		0.000	02:20	
Control Test		0.079	02:21	
Air Blank		0.000	02:21	
Subject Sample	# 1	0.206	02:24	Volume Not Met
Breath Volume		0.117	02:24	
Air Blank		0.000	02:25	
Air Blank		0.000	02:27	
Subject Sample	# 2		02:28	Subject Test Refused
Breath Volume		0.000	02:28	
Air Blank		0.000	02:29	
Control Test		0.078	02:29	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 06/29/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGUE,VINCE,H
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TITUS,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: SCHMEIDER

DATE/TIME OF TEST: 06/03/2018 15:44:26

Diagnostics Check	OK	15:41	
Air Blank	0.000	15:41	
Control Test	0.079	15:42	
Air Blank	0.000	15:42	
Subject Sample	# 1	15:43	Subject Test Refused
Breath Volume	0.000	15:43	
Air Blank	0.000	15:43	
Control Test	0.079	15:43	
Air Blank	0.000	15:44	
Diagnostics Check	OK	15:44	

Data Download Date/Time: 06/29/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGUE,VINCE,H
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: CHESHIRE,N
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: INGOLD

DATE/TIME OF TEST: 06/06/2018 14:46:32

Diagnostics Check	OK	14:40
Air Blank	0.000	14:40
Control Test	0.081	14:40
Air Blank	0.000	14:41
Subject Sample # 1	0.000	14:42
Breath Volume	1.960	14:42
Air Blank	0.000	14:42
Air Blank	0.000	14:44
Subject Sample # 2	0.000	14:45
Breath Volume	2.429	14:45
Air Blank	0.000	14:45
Control Test	0.080	14:45
Air Blank	0.000	14:46
Diagnostics Check	OK	14:46

Data Download Date/Time: 06/29/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CHASE,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: CLOUD,TOD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 06/08/2018 18:46:07

Diagnostics Check	OK	18:39	
Air Blank	0.000	18:39	
Control Test	0.079	18:39	
Air Blank	0.000	18:40	
Subject Sample	# 1 0.265	18:41	
Breath Volume	1.472	18:41	
Air Blank	0.000	18:41	
Air Blank	0.000	18:43	
Subject Sample	# 2	18:44	Subject Test Refused
Breath Volume	0.769	18:44	
Air Blank	0.000	18:45	
Control Test	0.079	18:45	
Air Blank	0.000	18:46	
Diagnostics Check	OK	18:46	

Data Download Date/Time: 06/29/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CHASE,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DAVENPORT,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 06/09/2018 00:11:03

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:08	
Control Test	0.080	00:08	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:09	Subject Test Refused
Breath Volume	0.000	00:09	
Air Blank	0.000	00:10	
Control Test	0.080	00:10	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:11	

Data Download Date/Time: 06/29/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: CRANE,RYAN
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: BUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 06/09/2018 20:43:02

Diagnostics Check	OK	20:35
Air Blank	0.000	20:35
Control Test	0.079	20:36
Air Blank	0.000	20:36
Subject Sample # 1	0.170	20:38
Breath Volume	3.937	20:38
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.175	20:41
Breath Volume	3.023	20:41
Air Blank	0.000	20:42
Control Test	0.079	20:42
Air Blank	0.000	20:42
Diagnostics Check	OK	20:43

Data Download Date/Time: 06/29/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CHASE,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: QUINN,DEP.
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MORTON

DATE/TIME OF TEST: 06/11/2018 04:40:54

Diagnostics Check	OK	04:37	
Air Blank	0.000	04:38	
Control Test	0.079	04:38	
Air Blank	0.000	04:39	
Subject Sample	# 1	04:39	Subject Test Refused
Breath Volume	0.000	04:39	
Air Blank	0.000	04:39	
Control Test	0.079	04:40	
Air Blank	0.000	04:40	
Diagnostics Check	OK	04:40	

Data Download Date/Time: 06/29/2018 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGUE,VINCE,H
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TITUS,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:14

SUBJECT'S LAST NAME: OLEKSIAK

DATE/TIME OF TEST: 06/11/2018 11:41:45

Diagnostics Check		OK	11:35
Air Blank		0.000	11:35
Control Test		0.079	11:35
Air Blank		0.000	11:36
Subject Sample	# 1	0.342	11:37
Breath Volume		1.277	11:37
Air Blank		0.000	11:37
Air Blank		0.000	11:39
Subject Sample	# 2	0.344	11:40
Breath Volume		2.054	11:40
Air Blank		0.000	11:40
Control Test		0.078	11:41
Air Blank		0.000	11:41
Diagnostics Check		OK	11:41

Data Download Date/Time: 06/29/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAN,MICHAEL,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 06/11/2018 20:15:10

Diagnostics Check	OK	20:09	
Air Blank	0.000	20:09	
Control Test	0.079	20:09	
Air Blank	0.000	20:10	
Subject Sample	# 1 0.209	20:13	Volume Not Met
Breath Volume	0.929	20:13	
Air Blank		20:14	Purge Fail
Air Blank		20:15	Purge Fail

Data Download Date/Time: 06/29/2018 09:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAN,MICHAEL,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WARD,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: BUSH

DATE/TIME OF TEST: 06/12/2018 00:08:10

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:04	
Control Test	0.080	00:04	
Air Blank	0.000	00:04	
Subject Sample	# 1	00:06	Subject Test Refused
Breath Volume	0.000	00:06	
Air Blank	0.000	00:07	
Control Test	0.082	00:07	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 06/29/2018 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER, MICHAEL, D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: NA, NA
ARRESTING OFFICER AGENCY: CCA Citrus
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: OTHER
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 06/19/2018 08:01:05

Diagnostics Check	OK	07:54
Air Blank	0.000	07:55
Control Test	0.080	07:55
Air Blank	0.000	07:55
Subject Sample # 1	0.000	07:56
Breath Volume	2.250	07:56
Air Blank	0.000	07:57
Air Blank	0.000	07:59
Subject Sample # 2	0.000	07:59
Breath Volume	2.109	07:59
Air Blank	0.000	08:00
Control Test	0.080	08:00
Air Blank	0.000	08:00
Diagnostics Check	OK	08:01

Data Download Date/Time: 06/29/2018 09:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CHASE,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: AMATO,ANTHONEY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: KINCHEN

DATE/TIME OF TEST: 06/20/2018 04:04:49

Diagnostics Check		OK	03:49	
Air Blank		0.000	03:50	
Control Test		0.079	03:50	
Air Blank		0.000	03:51	
Subject Sample	# 1	0.137	03:54	Volume Not Met
Breath Volume		0.710	03:54	
Air Blank		0.000	03:54	
Air Blank		0.000	03:56	
Subject Sample	# 2	0.162	03:57	
Breath Volume		1.132	03:57	
Air Blank		0.000	03:58	
Air Blank		0.000	04:00	
Subject Sample	# 3	0.102	04:03	Volume Not Met
Breath Volume		0.230	04:03	
Air Blank		0.000	04:03	
Control Test		0.080	04:04	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 06/29/2018 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BOSWELL,A
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: BAGLEY

DATE/TIME OF TEST: 06/24/2018 01:23:30

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.000	01:18
Breath Volume	2.937	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.000	01:22
Breath Volume	3.335	01:22
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 06/29/2018 09:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CHASE,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DAVENPORT,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: STOCKER

DATE/TIME OF TEST: 06/24/2018 21:47:33

Diagnostics Check	OK	21:44	
Air Blank	0.000	21:45	
Control Test	0.080	21:45	
Air Blank	0.000	21:45	
Subject Sample	# 1	21:46	Subject Test Refused
Breath Volume	0.000	21:46	
Air Blank	0.000	21:46	
Control Test	0.079	21:46	
Air Blank	0.000	21:47	
Diagnostics Check	OK	21:47	

Data Download Date/Time: 06/29/2018 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGUE,VINCE,H
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOWARD,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: MOBERLEY

DATE/TIME OF TEST: 06/26/2018 17:39:37

Diagnostics Check		OK	17:31
Air Blank		0.000	17:32
Control Test		0.080	17:32
Air Blank		0.000	17:32
Subject Sample	# 1	0.220	17:34
Breath Volume		1.460	17:34
Air Blank		0.000	17:34
Air Blank		0.000	17:36
Subject Sample	# 2	0.215	17:37
Breath Volume		1.332	17:37
Air Blank		0.000	17:38
Control Test		0.080	17:39
Air Blank		0.000	17:39
Diagnostics Check		OK	17:39

Data Download Date/Time: 06/29/2018 09:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTIGLIONE,SALVATORE,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CASTIGLIONE,SALVATORE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 06/22/2018 00:07:19

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.079	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.172	00:02
Breath Volume	1.687	00:02
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.168	00:05
Breath Volume	1.832	00:05
Air Blank	0.000	00:06
Control Test	0.079	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 06/25/2018 13:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GALLER,JESSICA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: SPARKS

DATE/TIME OF TEST: 06/22/2018 04:43:42

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.079	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.122	04:39
Breath Volume	3.656	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.115	04:42
Breath Volume	2.687	04:42
Air Blank	0.000	04:42
Control Test	0.077	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 06/25/2018 13:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 06/22/2018 19:56:01

Diagnostics Check	OK	19:50
Air Blank	0.000	19:50
Control Test	0.079	19:50
Air Blank	0.000	19:51
Subject Sample # 1	0.000	19:51
Breath Volume	1.734	19:51
Air Blank	0.000	19:52
Air Blank	0.000	19:54
Subject Sample # 2	0.000	19:54
Breath Volume	1.320	19:54
Air Blank	0.000	19:55
Control Test	0.079	19:55
Air Blank	0.000	19:55
Diagnostics Check	OK	19:56

Data Download Date/Time: 06/25/2018 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTHER,EMILY,S
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: GENTHER,EMILY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: KEILING

DATE/TIME OF TEST: 06/25/2018 03:23:09

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.079	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.262	03:18
Breath Volume		4.042	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.269	03:21
Breath Volume		2.714	03:21
Air Blank		0.000	03:22
Control Test		0.079	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 06/25/2018 13:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER, MATTHEW, D
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BAKER, MATTHEW
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: FIGHERA

DATE/TIME OF TEST: 06/25/2018 03:45:37

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.079	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.216	03:41
Breath Volume		1.597	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.214	03:43
Breath Volume		2.800	03:43
Air Blank		0.000	03:44
Control Test		0.078	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 06/25/2018 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARNEST, WILLIAM, M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: EARNEST, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: ATKINS

DATE/TIME OF TEST: 06/25/2018 04:42:46

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.080	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.198	04:38
Breath Volume	1.687	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.186	04:41
Breath Volume	1.433	04:41
Air Blank	0.000	04:41
Control Test	0.079	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 06/25/2018 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTHER,EMILY,S
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: GENTHER,EMILY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 06/27/2018 21:22:57

Diagnostics Check		OK	21:09	
Air Blank		0.000	21:09	
Control Test		0.080	21:09	
Air Blank		0.000	21:10	
Subject Sample	# 1	0.122	21:13	Volume Not Met
Breath Volume		0.402	21:13	
Air Blank		0.000	21:14	
Air Blank		0.000	21:16	
Subject Sample	# 2	0.139	21:17	
Breath Volume		1.515	21:17	
Air Blank		0.000	21:18	
Air Blank		0.000	21:19	
Subject Sample	# 3	0.137	21:21	
Breath Volume		1.664	21:21	
Air Blank		0.000	21:22	
Control Test		0.075	21:22	
Air Blank		0.000	21:22	
Diagnostics Check		OK	21:22	

Data Download Date/Time: 07/21/2018 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELRUSSO,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DELRUSSO,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 06/29/2018 20:20:09

Diagnostics Check	OK	20:13
Air Blank	0.000	20:14
Control Test	0.080	20:14
Air Blank	0.000	20:15
Subject Sample # 1	0.145	20:15
Breath Volume	2.734	20:15
Air Blank	0.000	20:16
Air Blank	0.000	20:18
Subject Sample # 2	0.143	20:18
Breath Volume	2.953	20:18
Air Blank	0.000	20:19
Control Test	0.079	20:19
Air Blank	0.000	20:20
Diagnostics Check	OK	20:20

Data Download Date/Time: 07/21/2018 15:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WILLIAMS,RONNIE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/29/2018 22:46:06

Diagnostics Check		OK	22:33	
Air Blank		0.000	22:34	
Control Test		0.079	22:34	
Air Blank		0.000	22:34	
Subject Sample	# 1	0.274	22:35	No .020 agreement
Breath Volume		1.984	22:35	
Air Blank		0.000	22:36	
Air Blank		0.000	22:38	
Subject Sample	# 2	0.239	22:38	No .020 agreement
Breath Volume		3.398	22:38	
Air Blank		0.000	22:39	
Air Blank		0.000	22:41	
Subject Sample	# 3	0.227	22:44	Volume Not Met
Breath Volume		0.125	22:44	
Air Blank		0.000	22:45	
Control Test		0.079	22:45	
Air Blank		0.000	22:46	
Diagnostics Check		OK	22:46	

Data Download Date/Time: 07/21/2018 15:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WILLIAMS,RONNIE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/29/2018 22:56:10

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.078	22:49
Air Blank	0.000	22:49
Subject Sample # 1	0.269	22:51
Breath Volume	1.144	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.269	22:54
Breath Volume	1.367	22:54
Air Blank	0.000	22:55
Control Test	0.077	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 07/21/2018 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: SHORT

DATE/TIME OF TEST: 06/30/2018 02:34:36

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.079	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.165	02:30
Breath Volume	2.101	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.164	02:33
Breath Volume	3.855	02:33
Air Blank	0.000	02:33
Control Test	0.079	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 07/21/2018 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINES,ELDON,L
BREATH TEST OPERATOR AGENCY: Dunnellon PD
ARRESTING OFFICER: RAINES,ELDON
ARRESTING OFFICER AGENCY: Dunnellon PD
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: SARMENTO

DATE/TIME OF TEST: 06/30/2018 05:22:14

Diagnostics Check	OK	05:15
Air Blank	0.000	05:16
Control Test	0.079	05:16
Air Blank	0.000	05:17
Subject Sample # 1	0.137	05:17
Breath Volume	2.230	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.132	05:20
Breath Volume	2.503	05:20
Air Blank	0.000	05:21
Control Test	0.078	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 07/21/2018 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS, RONALD, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM, R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MELO

DATE/TIME OF TEST: 06/01/2018 04:29:13

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.079	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.111	04:23
Breath Volume	1.312	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.114	04:27
Breath Volume	1.199	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 06/01/2018 21:26:00

Diagnostics Check	OK	21:18
Air Blank	0.000	21:19
Control Test	0.079	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.193	21:20
Breath Volume	1.984	21:20
Air Blank	0.000	21:21
Air Blank	0.000	21:23
Subject Sample # 2	0.197	21:24
Breath Volume	1.710	21:24
Air Blank	0.000	21:25
Control Test	0.079	21:25
Air Blank	0.000	21:25
Diagnostics Check	OK	21:26

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS, RONALD, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 06/02/2018 03:43:05

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.171	03:38
Breath Volume	1.222	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.174	03:41
Breath Volume	1.457	03:41
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/03/2018 01:10:47

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.080	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.150	01:04
Breath Volume		1.441	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.150	01:09
Breath Volume		1.792	01:09
Air Blank		0.000	01:09
Control Test		0.080	01:10
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: FIGUEROA GONZALEZ

DATE/TIME OF TEST: 06/03/2018 04:58:18

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.078	04:52
Air Blank	0.000	04:53
Subject Sample # 1	0.114	04:53
Breath Volume	1.218	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.114	04:56
Breath Volume	1.539	04:56
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHERMAN,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:21

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 06/03/2018 16:48:28

Diagnostics Check	OK	16:41
Air Blank	0.000	16:41
Control Test	0.079	16:41
Air Blank	0.000	16:42
Subject Sample # 1	0.027	16:43
Breath Volume	1.242	16:43
Air Blank	0.000	16:44
Air Blank	0.000	16:46
Subject Sample # 2	0.027	16:47
Breath Volume	1.257	16:47
Air Blank	0.000	16:47
Control Test	0.079	16:47
Air Blank	0.000	16:48
Diagnostics Check	OK	16:48

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRIZENDINE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 06/06/2018 01:42:23

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.065	01:37
Breath Volume	2.535	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.063	01:40
Breath Volume	2.218	01:40
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS, RONALD, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: MAKSYMETZ

DATE/TIME OF TEST: 06/07/2018 04:35:10

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.130	04:30
Breath Volume	4.148	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.132	04:33
Breath Volume	3.652	04:33
Air Blank	0.000	04:34
Control Test	0.079	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/08/2018 00:57:29

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:41	
Control Test		0.079	00:41	
Air Blank		0.000	00:42	
Subject Sample	# 1	0.084	00:45	Volume Not Met
Breath Volume		0.730	00:45	
Air Blank		0.000	00:46	
Air Blank		0.000	00:47	
Subject Sample	# 2	0.096	00:50	
Breath Volume		1.347	00:50	
Air Blank		0.000	00:51	
Air Blank		0.000	00:52	
Subject Sample	# 3	0.077	00:56	Volume Not Met
Breath Volume		0.062	00:56	
Air Blank		0.000	00:56	
Control Test		0.078	00:56	
Air Blank		0.000	00:57	
Diagnostics Check		OK	00:57	

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/08/2018 01:19:52

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.079	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 1 0.078	01:15	Volume Not Met
Breath Volume	1.007	01:15	
Air Blank	0.000	01:16	
Air Blank	0.000	01:17	
Subject Sample	# 2	01:18	Subject Test Refused
Breath Volume	0.000	01:18	
Air Blank	0.000	01:18	
Control Test	0.078	01:19	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: TRACY

DATE/TIME OF TEST: 06/08/2018 02:30:56

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.178	02:26
Breath Volume	1.902	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.179	02:29
Breath Volume	2.058	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAULINO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: MARTINEZ-CUEVAS

DATE/TIME OF TEST: 06/09/2018 00:21:46

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:16
Subject Sample # 1	0.099	00:17
Breath Volume	1.355	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.099	00:20
Breath Volume	1.585	00:20
Air Blank	0.000	00:20
Control Test	0.079	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEINBERG, T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: MARTINEZ-ANTONIO

DATE/TIME OF TEST: 06/10/2018 00:44:07

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.258	00:39
Breath Volume	3.121	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.251	00:42
Breath Volume	2.609	00:42
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS, RONALD, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 06/10/2018 04:37:41

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.080	04:31
Air Blank	0.000	04:31
Subject Sample # 1	0.144	04:32
Breath Volume	2.015	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:34
Subject Sample # 2	0.146	04:36
Breath Volume	1.898	04:36
Air Blank	0.000	04:36
Control Test	0.079	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: HERNANDEZ GOMEZ

DATE/TIME OF TEST: 06/10/2018 21:42:22

Diagnostics Check	OK	21:35
Air Blank	0.000	21:35
Control Test	0.080	21:36
Air Blank	0.000	21:36
Subject Sample # 1	0.220	21:37
Breath Volume	2.503	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.219	21:40
Breath Volume	2.296	21:40
Air Blank	0.000	21:41
Control Test	0.078	21:41
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS, RONALD, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIERRA, P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: GARRISON

DATE/TIME OF TEST: 06/11/2018 04:51:39

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.079	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.186	04:47
Breath Volume	3.101	04:47
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.185	04:50
Breath Volume	3.949	04:50
Air Blank	0.000	04:50
Control Test	0.078	04:51
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS, RONALD, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 06/11/2018 23:55:14

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.080	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.209	23:50
Breath Volume	1.132	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.226	23:53
Breath Volume	2.539	23:53
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KAMIKA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIERRA,P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: VICTORIO

DATE/TIME OF TEST: 06/12/2018 00:49:14

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.079	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.177	00:44
Breath Volume	1.417	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:46
Subject Sample # 2	0.182	00:47
Breath Volume	2.460	00:47
Air Blank	0.000	00:48
Control Test	0.079	00:48
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: NEWBERG

DATE/TIME OF TEST: 06/12/2018 04:55:40

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.080	04:49
Air Blank	0.000	04:49
Subject Sample # 1	0.217	04:51
Breath Volume	2.320	04:51
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.216	04:54
Breath Volume	2.203	04:54
Air Blank	0.000	04:54
Control Test	0.079	04:55
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: SORRELLS

DATE/TIME OF TEST: 06/14/2018 03:10:43

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.248	03:05
Breath Volume		2.285	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.239	03:09
Breath Volume		1.304	03:09
Air Blank		0.000	03:09
Control Test		0.079	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KAMIKA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: PEREZ MENDEZ

DATE/TIME OF TEST: 06/15/2018 01:13:44

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:59	
Control Test	0.079	00:59	
Air Blank	0.000	00:59	
Subject Sample	# 1 0.232	01:02	
Breath Volume	1.175	01:02	
Air Blank	0.000	01:03	
Air Blank	0.000	01:05	
Subject Sample	# 2 0.195	01:08	Volume Not Met
Breath Volume	0.410	01:08	
Air Blank	0.000	01:08	
Air Blank	0.000	01:10	
Subject Sample	# 3 0.222	01:12	
Breath Volume	2.039	01:12	
Air Blank	0.000	01:12	
Control Test	0.078	01:13	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HUMPHREY

DATE/TIME OF TEST: 06/16/2018 02:18:12

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.080	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.209	02:13
Breath Volume	1.589	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.190	02:16
Breath Volume	1.519	02:16
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 06/16/2018 05:36:41

Diagnostics Check	OK	05:30
Air Blank	0.000	05:30
Control Test	0.079	05:31
Air Blank	0.000	05:31
Subject Sample # 1	0.144	05:32
Breath Volume	3.312	05:32
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.144	05:35
Breath Volume	3.292	05:35
Air Blank	0.000	05:35
Control Test	0.079	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HILES,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MACINDOE

DATE/TIME OF TEST: 06/17/2018 03:06:43

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.174	03:02
Breath Volume	1.128	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.172	03:05
Breath Volume	1.347	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: GONZALEZ OXLAJ

DATE/TIME OF TEST: 06/17/2018 03:36:19

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.270	03:30
Breath Volume	1.253	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.262	03:34
Breath Volume	2.046	03:34
Air Blank	0.000	03:35
Control Test	0.077	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: MARTINEZ LOPEZ

DATE/TIME OF TEST: 06/17/2018 19:25:18

Diagnostics Check	OK	19:11	
Air Blank	0.000	19:11	
Control Test	0.080	19:12	
Air Blank	0.000	19:12	
Subject Sample	# 1	19:15	No Sample Provided
Breath Volume	0.000	19:15	
Air Blank	0.000	19:16	
Air Blank	0.000	19:18	
Subject Sample	# 2	0.223	19:20
Breath Volume	2.519	19:20	
Air Blank	0.000	19:21	
Air Blank	0.000	19:23	
Subject Sample	# 3	0.219	19:23
Breath Volume	2.281	19:23	
Air Blank	0.000	19:24	
Control Test	0.079	19:24	
Air Blank	0.000	19:25	
Diagnostics Check	OK	19:25	

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIERRA,P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: SANCHEZ RODRIGUEZ

DATE/TIME OF TEST: 06/18/2018 20:41:48

Diagnostics Check	OK	20:34
Air Blank	0.000	20:34
Control Test	0.081	20:35
Air Blank	0.000	20:35
Subject Sample # 1	0.000	20:37
Breath Volume	1.195	20:37
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.000	20:40
Breath Volume	2.226	20:40
Air Blank	0.000	20:40
Control Test	0.079	20:41
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: LEMUS

DATE/TIME OF TEST: 06/18/2018 22:11:49

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.080	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.144	22:07
Breath Volume	2.695	22:07
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.138	22:10
Breath Volume	2.363	22:10
Air Blank	0.000	22:10
Control Test	0.079	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 06/19/2018 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: GALLAS

DATE/TIME OF TEST: 06/01/2018 01:15:50

Diagnostics Check		OK	01:06
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.156	01:09
Breath Volume		1.843	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.151	01:14
Breath Volume		1.207	01:14
Air Blank		0.000	01:14
Control Test		0.079	01:15
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 06/19/2018 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CHASSANG

DATE/TIME OF TEST: 06/02/2018 02:47:30

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.081	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.100	02:42
Breath Volume		1.343	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.101	02:45
Breath Volume		1.578	02:45
Air Blank		0.000	02:46
Control Test		0.081	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 06/19/2018 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 06/03/2018 04:16:49

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.081	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.140	04:10
Breath Volume	1.800	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.143	04:15
Breath Volume	2.476	04:15
Air Blank	0.000	04:15
Control Test	0.079	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 06/19/2018 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: HUGHES II

DATE/TIME OF TEST: 06/03/2018 05:04:26

Diagnostics Check		OK	04:58
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Subject Sample	# 1	0.134	05:00
Breath Volume		1.285	05:00
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.146	05:02
Breath Volume		1.367	05:02
Air Blank		0.000	05:03
Control Test		0.080	05:03
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 06/19/2018 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS, RONALD, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: LETCHWORTH

DATE/TIME OF TEST: 06/05/2018 21:15:37

Diagnostics Check	OK	21:09
Air Blank	0.000	21:10
Control Test	0.080	21:10
Air Blank	0.000	21:10
Subject Sample # 1	0.000	21:11
Breath Volume	1.457	21:11
Air Blank	0.000	21:11
Air Blank	0.000	21:13
Subject Sample # 2	0.000	21:14
Breath Volume	2.390	21:14
Air Blank	0.000	21:14
Control Test	0.080	21:15
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time: 06/19/2018 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: VAN HOF

DATE/TIME OF TEST: 06/08/2018 04:41:27

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.152	04:36
Breath Volume	1.859	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.143	04:39
Breath Volume	1.265	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 06/19/2018 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: MONTOYA

DATE/TIME OF TEST: 06/09/2018 01:01:27

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.081	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.158	00:56
Breath Volume	1.589	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:58
Subject Sample # 2	0.173	00:59
Breath Volume	1.257	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 06/19/2018 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/09/2018 03:48:35

Diagnostics Check		OK	03:35	
Air Blank		0.000	03:36	
Control Test		0.081	03:36	
Air Blank		0.000	03:36	
Subject Sample	# 1	0.152	03:40	Volume Not Met
Breath Volume		1.011	03:40	
Air Blank		0.000	03:40	
Air Blank		0.000	03:42	
Subject Sample	# 2	0.167	03:43	
Breath Volume		1.207	03:43	
Air Blank		0.000	03:43	
Air Blank		0.000	03:45	
Subject Sample	# 3	0.162	03:46	
Breath Volume		1.292	03:46	
Air Blank		0.000	03:47	
Control Test		0.080	03:47	
Air Blank		0.000	03:48	
Diagnostics Check		OK	03:48	

Data Download Date/Time: 06/19/2018 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: GOSHA DIXON

DATE/TIME OF TEST: 06/13/2018 01:48:11

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.198	01:42
Breath Volume	1.765	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.192	01:46
Breath Volume	1.917	01:46
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 06/19/2018 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS, V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: MALDONADO CRUZ

DATE/TIME OF TEST: 06/15/2018 05:11:08

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.081	05:05
Air Blank	0.000	05:06
Subject Sample # 1	0.185	05:06
Breath Volume	2.183	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.198	05:09
Breath Volume	1.402	05:09
Air Blank	0.000	05:10
Control Test	0.080	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 06/19/2018 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MUSUMECI

DATE/TIME OF TEST: 06/16/2018 03:29:27

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.081	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.000	03:25
Breath Volume	2.328	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.000	03:28
Breath Volume	4.835	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 06/19/2018 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NASSBAUMER,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: GILES

DATE/TIME OF TEST: 06/16/2018 05:49:39

Diagnostics Check	OK	05:38	
Air Blank	0.000	05:38	
Control Test	0.080	05:39	
Air Blank	0.000	05:39	
Subject Sample	# 1	05:42	No Sample Provided
Breath Volume	0.000	05:42	
Air Blank	0.000	05:43	
Air Blank	0.000	05:45	
Subject Sample	# 2	05:48	No Sample Provided
Breath Volume	0.000	05:48	
Air Blank	0.000	05:48	
Control Test	0.080	05:49	
Air Blank	0.000	05:49	
Diagnostics Check	OK	05:49	

Data Download Date/Time: 06/19/2018 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NUSSBAUMER,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: GILES

DATE/TIME OF TEST: 06/16/2018 06:14:16

Diagnostics Check	OK	06:06
Air Blank	0.000	06:07
Control Test	0.081	06:07
Air Blank	0.000	06:08
Subject Sample # 1	0.126	06:09
Breath Volume	1.394	06:09
Air Blank	0.000	06:10
Air Blank	0.000	06:11
Subject Sample # 2	0.130	06:12
Breath Volume	1.507	06:12
Air Blank	0.000	06:13
Control Test	0.080	06:13
Air Blank	0.000	06:14
Diagnostics Check	OK	06:14

Data Download Date/Time: 06/19/2018 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: APOLINAR VILLADA

DATE/TIME OF TEST: 06/17/2018 03:54:38

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.159	03:49
Breath Volume	2.691	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.171	03:53
Breath Volume	2.539	03:53
Air Blank	0.000	03:53
Control Test	0.080	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 06/19/2018 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PALFY,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: HERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 06/17/2018 05:26:55

Diagnostics Check	OK	05:20
Air Blank	0.000	05:20
Control Test	0.080	05:20
Air Blank	0.000	05:21
Subject Sample # 1	0.110	05:22
Breath Volume	1.292	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:24
Subject Sample # 2	0.111	05:25
Breath Volume	1.414	05:25
Air Blank	0.000	05:25
Control Test	0.080	05:26
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 06/19/2018 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAISSE,A
ARRESTING OFFICER AGENCY: Tampa Airport PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: GROOVER

DATE/TIME OF TEST: 06/18/2018 22:51:41

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.081	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.213	22:47
Breath Volume	2.042	22:47
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.226	22:50
Breath Volume	2.113	22:50
Air Blank	0.000	22:50
Control Test	0.080	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 06/19/2018 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PALKO

DATE/TIME OF TEST: 06/01/2018 03:26:52

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.079	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Control Test	0.079	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 06/07/2018 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: AVILA,JEREMIAH
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:17

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 06/02/2018 11:00:56

Diagnostics Check	OK	10:54
Air Blank	0.000	10:55
Control Test	0.079	10:55
Air Blank	0.000	10:55
Subject Sample # 1	0.000	10:56
Breath Volume	2.730	10:56
Air Blank	0.000	10:56
Air Blank	0.000	10:58
Subject Sample # 2	0.000	10:59
Breath Volume	1.578	10:59
Air Blank	0.000	11:00
Control Test	0.079	11:00
Air Blank	0.000	11:00
Diagnostics Check	OK	11:00

Data Download Date/Time: 06/07/2018 22:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLAP,JASON,N
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: MORROW,ART
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: BUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 06/02/2018 20:13:55

Diagnostics Check	OK	20:07
Air Blank	0.000	20:07
Control Test	0.079	20:07
Air Blank	0.000	20:08
Subject Sample # 1	0.214	20:09
Breath Volume	1.375	20:09
Air Blank	0.000	20:10
Air Blank	0.000	20:11
Subject Sample # 2	0.203	20:12
Breath Volume	1.613	20:12
Air Blank	0.000	20:12
Control Test	0.079	20:13
Air Blank	0.000	20:13
Diagnostics Check	OK	20:13

Data Download Date/Time: 06/07/2018 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/07/2018 01:17:03

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.000	01:12
Breath Volume	2.804	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.000	01:15
Breath Volume	3.265	01:15
Air Blank	0.000	01:16
Control Test	0.079	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:17

Data Download Date/Time: 06/07/2018 22:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORGDARD,BRYAN,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: PEREZ,NICHOLAS
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: AVALOS

DATE/TIME OF TEST: 06/12/2018 03:33:51

Diagnostics Check		OK	03:21	
Air Blank		0.000	03:22	
Control Test		0.079	03:22	
Air Blank		0.000	03:23	
Subject Sample	# 1	0.159	03:26	Volume Not Met
Breath Volume		0.597	03:26	
Air Blank		0.000	03:27	
Air Blank		0.000	03:28	
Subject Sample	# 2	0.165	03:31	Volume Not Met
Breath Volume		0.781	03:31	
Air Blank		0.000	03:32	
Control Test		0.080	03:33	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:33	

Data Download Date/Time: 07/06/2018 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWARD,BRYAN,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: PEREZ,NICHOLAS
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: AVALOS

DATE/TIME OF TEST: 06/12/2018 03:40:50

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.078	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:39	Subject Test Refused
Breath Volume	0.363	03:39	
Air Blank	0.000	03:39	
Control Test	0.079	03:40	
Air Blank	0.000	03:40	
Diagnostics Check	OK	03:40	

Data Download Date/Time: 07/06/2018 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORGDARD,BRYAN,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: APONTE,JOSH
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: GOLEY

DATE/TIME OF TEST: 06/13/2018 03:12:12

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:09	
Control Test	0.079	03:09	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:10	Subject Test Refused
Breath Volume	0.000	03:10	
Air Blank	0.000	03:11	
Control Test	0.078	03:11	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 07/06/2018 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BELL,ETHAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: TRANIELLO

DATE/TIME OF TEST: 06/15/2018 23:06:04

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:01	
Control Test	0.079	23:01	
Air Blank	0.000	23:02	
Subject Sample	# 1	23:04	Subject Test Refused
Breath Volume	0.000	23:04	
Air Blank	0.000	23:05	
Control Test	0.079	23:05	
Air Blank	0.000	23:05	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 07/06/2018 23:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REIMUNDY,B,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: APONTE,J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: WEBER

DATE/TIME OF TEST: 06/17/2018 04:28:08

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.079	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.085	04:23
Breath Volume	2.394	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:25
Subject Sample # 2	0.085	04:26
Breath Volume	1.390	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 07/06/2018 23:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: NAPIER

DATE/TIME OF TEST: 06/23/2018 03:20:17

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.167	03:15
Breath Volume	2.199	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.167	03:18
Breath Volume	2.285	03:18
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 07/06/2018 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARPER,JIMMIE,L
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE,KRIS
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: KNUTELSKY

DATE/TIME OF TEST: 06/23/2018 05:31:28

Diagnostics Check	OK	05:24
Air Blank	0.000	05:25
Control Test	0.079	05:25
Air Blank	0.000	05:25
Subject Sample # 1	0.142	05:26
Breath Volume	3.394	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.138	05:29
Breath Volume	1.468	05:29
Air Blank	0.000	05:30
Control Test	0.079	05:30
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 07/06/2018 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARPER,JIMMIE,L
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: DUBI,JOHN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/24/2018 03:23:08

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:20	
Control Test	0.079	03:20	
Air Blank	0.000	03:21	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:22	
Control Test	0.079	03:22	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 07/06/2018 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WURST,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WURST,JOSHUA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: JOYCE

DATE/TIME OF TEST: 06/08/2018 02:26:46

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.177	02:21
Breath Volume	1.437	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.179	02:25
Breath Volume	3.371	02:25
Air Blank	0.000	02:25
Control Test	0.079	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 06/22/2018 18:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,JAMES,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: POKIPALA,JONAH
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HOLNESS

DATE/TIME OF TEST: 06/09/2018 04:18:09

Diagnostics Check	OK	04:16	
Air Blank	0.000	04:17	
Control Test	0.080	04:17	
Air Blank		04:17	RFI Detect
Air Blank	0.000	04:18	

Data Download Date/Time: 06/22/2018 18:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,JAMES,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: POKIPALA,JONAH
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HOLNESS

DATE/TIME OF TEST: 06/09/2018 04:25:54

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.000	04:21
Breath Volume	2.625	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.000	04:24
Breath Volume	3.273	04:24
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 06/22/2018 18:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,JAMES,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: POKIPALA,JONAH
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: HOLNESS

DATE/TIME OF TEST: 06/09/2018 04:34:55

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.000	04:29
Breath Volume	2.699	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.000	04:33
Breath Volume	2.898	04:33
Air Blank	0.000	04:34
Control Test	0.079	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 06/22/2018 18:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,JAMES,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JONES,JAMES
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 06/14/2018 22:44:29

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.080	22:39
Air Blank	0.000	22:39
Subject Sample # 1	0.200	22:40
Breath Volume	2.300	22:40
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.193	22:42
Breath Volume	1.742	22:42
Air Blank	0.000	22:43
Control Test	0.079	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 06/22/2018 18:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILLMAN,JAMES,L
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: POKIPALA,JONAH
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: LESNIAK

DATE/TIME OF TEST: 06/24/2018 18:49:11

Diagnostics Check	OK	18:41
Air Blank	0.000	18:41
Control Test	0.079	18:42
Air Blank	0.000	18:42
Subject Sample # 1	0.126	18:44
Breath Volume	2.613	18:44
Air Blank	0.000	18:45
Air Blank	0.000	18:46
Subject Sample # 2	0.120	18:47
Breath Volume	1.605	18:47
Air Blank	0.000	18:48
Control Test	0.078	18:48
Air Blank	0.000	18:49
Diagnostics Check	OK	18:49

Data Download Date/Time: 07/06/2018 18:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDYETTE,COREY,L
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: BIEBERITZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: GRACE

DATE/TIME OF TEST: 06/02/2018 23:42:29

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:37
Air Blank	0.000	23:37
Subject Sample # 1	0.115	23:38
Breath Volume	1.542	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.113	23:41
Breath Volume	1.519	23:41
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 06/27/2018 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTZONG,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: HAMSTRA,JOSEPH
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: SWEINHART

DATE/TIME OF TEST: 06/13/2018 05:56:25

Diagnostics Check	OK	05:50
Air Blank	0.000	05:50
Control Test	0.079	05:51
Air Blank	0.000	05:51
Subject Sample # 1	0.132	05:52
Breath Volume	2.675	05:52
Air Blank	0.000	05:52
Air Blank	0.000	05:54
Subject Sample # 2	0.131	05:54
Breath Volume	3.289	05:54
Air Blank	0.000	05:55
Control Test	0.079	05:55
Air Blank	0.000	05:56
Diagnostics Check	OK	05:56

Data Download Date/Time: 06/27/2018 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUTTON,KERRY,D
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: FUGATE,KATIE
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 06/16/2018 04:13:58

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:11	
Control Test	0.078	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:12	Subject Test Refused
Breath Volume	0.000	04:12	
Air Blank	0.000	04:13	
Control Test	0.079	04:13	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 06/27/2018 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSBAUM,K,
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: NUSBAUM,K
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:53

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 06/16/2018 13:25:08

Diagnostics Check	OK	13:18
Air Blank	0.000	13:19
Control Test	0.080	13:19
Air Blank	0.000	13:20
Subject Sample # 1	0.054	13:20
Breath Volume	2.031	13:20
Air Blank	0.000	13:21
Air Blank	0.000	13:23
Subject Sample # 2	0.051	13:23
Breath Volume	1.664	13:23
Air Blank	0.000	13:24
Control Test	0.079	13:24
Air Blank	0.000	13:25
Diagnostics Check	OK	13:25

Data Download Date/Time: 06/27/2018 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD, MATTHEW, E
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: MAYFIELD, MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MCCOMBS

DATE/TIME OF TEST: 06/23/2018 04:03:25

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.078	04:01	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:02	Subject Test Refused
Breath Volume	0.000	04:02	
Air Blank	0.000	04:02	
Control Test	0.078	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 06/27/2018 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD, MATTHEW, E
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: MAYFIELD, MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BOZEMAN

DATE/TIME OF TEST: 06/28/2018 03:14:20

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.182	03:09
Breath Volume	4.621	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.177	03:12
Breath Volume	3.671	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 08/09/2018 01:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,PERRY,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CARVALHO,ASHLEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: LORENZI

DATE/TIME OF TEST: 06/02/2018 00:22:08

Diagnostics Check	OK	00:09	
Air Blank	0.000	00:10	
Control Test	0.082	00:10	
Air Blank	0.000	00:10	
Subject Sample # 1	0.133	00:12	
Breath Volume	1.480	00:12	
Air Blank	0.000	00:12	
Air Blank	0.000	00:14	
Subject Sample # 2	0.124	00:17	Volume Not Met
Breath Volume	0.250	00:17	
Air Blank	0.000	00:18	
Air Blank	0.000	00:20	
Subject Sample # 3		00:20	Subject Test Refused
Breath Volume	0.000	00:20	
Air Blank	0.000	00:21	
Control Test	0.079	00:21	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 06/06/2018 10:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FESSEL,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BAEZ,ANDY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:39

SUBJECT'S LAST NAME: MCINTYRE

DATE/TIME OF TEST: 06/04/2018 08:20:19

Diagnostics Check	OK	08:14
Air Blank	0.000	08:14
Control Test	0.081	08:14
Air Blank	0.000	08:15
Subject Sample # 1	0.000	08:16
Breath Volume	4.062	08:16
Air Blank	0.000	08:16
Air Blank	0.000	08:18
Subject Sample # 2	0.000	08:18
Breath Volume	3.578	08:18
Air Blank	0.000	08:19
Control Test	0.081	08:19
Air Blank	0.000	08:20
Diagnostics Check	OK	08:20

Data Download Date/Time: 06/06/2018 10:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARRILLO,N,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,J
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: DEISS

DATE/TIME OF TEST: 06/05/2018 03:00:24

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.196	02:56
Breath Volume	1.476	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.190	02:58
Breath Volume	1.628	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 06/06/2018 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARRILLO,N,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,J
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: HUTCHISON

DATE/TIME OF TEST: 06/08/2018 21:54:02

Diagnostics Check	OK	21:48
Air Blank	0.000	21:48
Control Test	0.080	21:48
Air Blank	0.000	21:49
Subject Sample # 1	0.174	21:49
Breath Volume	2.210	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.171	21:52
Breath Volume	1.773	21:52
Air Blank	0.000	21:53
Control Test	0.080	21:53
Air Blank	0.000	21:53
Diagnostics Check	OK	21:54

Data Download Date/Time: 07/18/2018 09:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOBLER,AMANDA,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: TOBLER,AMANDA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: FABLE

DATE/TIME OF TEST: 06/09/2018 02:51:00

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.081	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.139	02:44
Breath Volume	1.843	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.149	02:49
Breath Volume	1.257	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:51

Data Download Date/Time: 07/18/2018 09:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,SARAH,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,SARAH
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: GRODZINSKI

DATE/TIME OF TEST: 06/13/2018 22:53:26

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.079	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.124	22:49
Breath Volume	1.527	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.128	22:51
Breath Volume	1.433	22:51
Air Blank	0.000	22:52
Control Test	0.079	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 07/18/2018 09:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSH,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: VAZQUEZ,HUMBERTO
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 06/14/2018 03:39:28

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.081	03:26	
Air Blank	0.000	03:26	
Subject Sample	# 1 0.165	03:27	
Breath Volume	1.230	03:27	
Air Blank	0.000	03:27	
Air Blank	0.000	03:29	
Subject Sample	# 2 0.150	03:32	Volume Not Met
Breath Volume	0.253	03:32	
Air Blank	0.000	03:33	
Air Blank	0.000	03:35	
Subject Sample	# 3	03:37	Subject Test Refused
Breath Volume	0.031	03:37	
Air Blank	0.000	03:38	
Control Test	0.078	03:38	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 07/18/2018 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSH,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: VAZQUEZ,HUMBERTO
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 06/14/2018 03:51:10

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:43	
Control Test		0.079	03:43	
Air Blank		0.000	03:44	
Subject Sample	# 1	0.133	03:47	Volume Not Met
Breath Volume		0.132	03:47	
Air Blank		0.000	03:47	
Air Blank		0.000	03:49	
Subject Sample	# 2		03:49	Subject Test Refused
Breath Volume		0.000	03:49	
Air Blank		0.000	03:50	
Control Test		0.078	03:50	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:51	

Data Download Date/Time: 07/18/2018 09:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLAK,ALLAN,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: KOLAK,ALLAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: JARVIS

DATE/TIME OF TEST: 06/22/2018 19:25:57

Diagnostics Check	OK	19:19
Air Blank	0.000	19:20
Control Test	0.078	19:20
Air Blank	0.000	19:21
Subject Sample # 1	0.085	19:21
Breath Volume	1.625	19:21
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.082	19:24
Breath Volume	1.542	19:24
Air Blank	0.000	19:25
Control Test	0.079	19:25
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 07/18/2018 09:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLAK,ALLAN,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: KOLAK,ALLAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 19:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 06/22/2018 19:42:04

Diagnostics Check	OK	19:36	
Air Blank	0.000	19:36	
Control Test	0.079	19:37	
Air Blank	0.000	19:37	
Subject Sample	# 1 0.103	19:38	
Breath Volume	1.261	19:38	
Air Blank	0.000	19:38	
Air Blank	0.000	19:40	
Subject Sample	# 2	19:40	Subject Test Refused
Breath Volume	0.000	19:40	
Air Blank	0.000	19:41	
Control Test	0.079	19:41	
Air Blank	0.000	19:41	
Diagnostics Check	OK	19:42	

Data Download Date/Time: 07/18/2018 09:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLAK,ALLAN,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: KOLAK,ALLAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: PORTAGALLO

DATE/TIME OF TEST: 06/22/2018 19:58:44

Diagnostics Check	OK	19:51	
Air Blank	0.000	19:52	
Control Test	0.080	19:52	
Air Blank	0.000	19:53	
Subject Sample # 1	0.025	19:54	
Breath Volume	1.343	19:54	
Air Blank	0.000	19:54	
Air Blank	0.000	19:56	
Subject Sample # 2		19:57	Subject Test Refused
Breath Volume	0.000	19:57	
Air Blank	0.000	19:57	
Control Test	0.079	19:58	
Air Blank	0.000	19:58	
Diagnostics Check	OK	19:58	

Data Download Date/Time: 07/18/2018 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,SARAH,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,SARAH
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: RALEY

DATE/TIME OF TEST: 06/23/2018 00:40:07

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.080	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.157	00:35
Breath Volume	1.472	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.146	00:38
Breath Volume	1.492	00:38
Air Blank	0.000	00:39
Control Test	0.081	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 07/18/2018 09:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAU,J.,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,J.
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 06/24/2018 02:42:12

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.137	02:37
Breath Volume	1.898	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.138	02:40
Breath Volume	2.292	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 07/18/2018 09:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALFONSO,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RODRIGUEZ,ALFONSO
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 06/28/2018 23:52:18

Diagnostics Check		OK	23:42	
Air Blank		0.000	23:42	
Control Test		0.080	23:42	
Air Blank		0.000	23:43	
Subject Sample	# 1	0.252	23:46	Volume Not Met
Breath Volume		0.218	23:46	
Air Blank		0.000	23:47	
Air Blank		0.000	23:48	
Subject Sample	# 2		23:50	Subject Test Refused
Breath Volume		0.152	23:50	
Air Blank		0.000	23:51	
Control Test		0.079	23:51	
Air Blank		0.000	23:52	
Diagnostics Check		OK	23:52	

Data Download Date/Time: 07/18/2018 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: ZAPPIA

DATE/TIME OF TEST: 06/06/2018 01:44:13

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.078	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.262	01:39
Breath Volume	3.835	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.279	01:42
Breath Volume	3.820	01:42
Air Blank	0.000	01:43
Control Test	0.077	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 06/29/2018 08:04