

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HILL,DANIEL,A  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HRADECKY,CORBIN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 05/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/03/2018 21:26:53

---

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:11	
Control Test	0.080	21:11	
Air Blank	0.000	21:11	
Subject Sample # 1	0.115	21:13	
Breath Volume	1.511	21:13	
Air Blank	0.000	21:14	
Air Blank	0.000	21:16	
Subject Sample # 2	0.099	21:19	Volume Not Met
Breath Volume	1.476	21:19	
Air Blank	0.000	21:20	
Air Blank	0.000	21:22	
Subject Sample # 3	0.098	21:25	Volume Not Met
Breath Volume	0.894	21:25	
Air Blank	0.000	21:25	
Control Test	0.082	21:26	
Air Blank	0.000	21:26	
Diagnostics Check	OK	21:26	

---

Data Download Date/Time: 06/06/2018 14:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWOGGER, EDWARD, A  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: DANIELS, WILLIAM  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: ZAPRALA

DATE/TIME OF TEST: 06/13/2018 22:04:08

---

Diagnostics Check	OK	22:00	
Air Blank	0.000	22:01	
Control Test	0.080	22:01	
Air Blank	0.000	22:01	
Subject Sample	# 1	22:02	Subject Test Refused
Breath Volume	0.000	22:02	
Air Blank	0.000	22:03	
Control Test	0.081	22:03	
Air Blank	0.000	22:04	
Diagnostics Check	OK	22:04	

---

Data Download Date/Time: 07/12/2018 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWOGGER, EDWARD, A  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: SWOGGER, EDWARD  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: MARX

DATE/TIME OF TEST: 06/19/2018 18:44:13

---

Diagnostics Check	OK	18:35
Air Blank	0.000	18:36
Control Test	0.081	18:36
Air Blank	0.000	18:36
Subject Sample # 1	0.145	18:38
Breath Volume	1.164	18:38
Air Blank	0.000	18:39
Air Blank	0.000	18:40
Subject Sample # 2	0.145	18:42
Breath Volume	1.238	18:42
Air Blank	0.000	18:43
Control Test	0.082	18:43
Air Blank	0.000	18:44
Diagnostics Check	OK	18:44

---

Data Download Date/Time: 07/12/2018 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWOGGER, EDWARD, A  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: WALRAVEN, D  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 06/23/2018 02:44:10

---

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.081	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.093	02:39
Breath Volume	1.906	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.088	02:42
Breath Volume	1.910	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

---

Data Download Date/Time: 07/12/2018 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWOGGER, EDWARD, A  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: SWOGGER, EDWARD  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: DUMONT

DATE/TIME OF TEST: 06/24/2018 21:55:27

---

Diagnostics Check	OK	21:52	
Air Blank	0.000	21:53	
Control Test	0.081	21:53	
Air Blank	0.000	21:53	
Subject Sample	# 1	21:54	Subject Test Refused
Breath Volume	0.000	21:54	
Air Blank	0.000	21:54	
Control Test	0.081	21:54	
Air Blank	0.000	21:55	
Diagnostics Check	OK	21:55	

---

Data Download Date/Time: 07/12/2018 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VIERGUTZ,ERIC,W  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: RIVAS,JUSTIN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/06/2018 21:28:53

---

Diagnostics Check	OK	21:22
Air Blank	0.000	21:23
Control Test	0.081	21:23
Air Blank	0.000	21:23
Subject Sample # 1	0.066	21:24
Breath Volume	2.722	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:26
Subject Sample # 2	0.062	21:27
Breath Volume	1.953	21:27
Air Blank	0.000	21:27
Control Test	0.080	21:28
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

---

Data Download Date/Time: 06/08/2018 14:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VIERGUTZ,ERIC,W  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: VIERGUTZ,ERIC  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: CRISLER

DATE/TIME OF TEST: 06/06/2018 21:47:09

---

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:36	
Control Test	0.081	21:36	
Air Blank	0.000	21:37	
Subject Sample	# 1	21:40	No Sample Provided
Breath Volume	0.000	21:40	
Air Blank	0.000	21:40	
Air Blank	0.000	21:42	
Subject Sample	# 2	21:45	No Sample Provided
Breath Volume	0.000	21:45	
Air Blank	0.000	21:46	
Control Test	0.080	21:46	
Air Blank	0.000	21:47	
Diagnostics Check	OK	21:47	

---

Data Download Date/Time: 06/08/2018 14:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,SHAWN,W  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: FAJARDO,MARIO  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 05/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: O DANIEL

DATE/TIME OF TEST: 06/07/2018 19:02:43

---

Diagnostics Check	OK	18:55
Air Blank	0.000	18:55
Control Test	0.081	18:56
Air Blank	0.000	18:56
Subject Sample # 1	0.315	18:57
Breath Volume	1.843	18:57
Air Blank	0.000	18:58
Air Blank	0.000	18:59
Subject Sample # 2	0.296	19:01
Breath Volume	1.644	19:01
Air Blank	0.000	19:01
Control Test	0.080	19:02
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

---

Data Download Date/Time: 06/08/2018 14:00



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEAN,PAUL,D  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: TORRES,JOEL  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: VERA

DATE/TIME OF TEST: 06/09/2018 02:45:21

---

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.119	02:40
Breath Volume	2.109	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.120	02:43
Breath Volume	1.734	02:43
Air Blank	0.000	02:44
Control Test	0.081	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

---

Data Download Date/Time: 07/13/2018 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHROTH,REWA,P  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: PETRU,MIROS  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 18:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 06/10/2018 19:25:49

---

Diagnostics Check	OK	19:15
Air Blank	0.000	19:16
Control Test	0.081	19:16
Air Blank	0.000	19:17
Subject Sample # 1	0.179	19:17
Breath Volume	2.566	19:17
Air Blank	0.000	19:18
Air Blank	0.000	19:19
Subject Sample # 2	0.158	19:21
Breath Volume	2.367	19:21
Air Blank	0.000	19:22
Air Blank	0.000	19:23
Subject Sample # 3	0.157	19:24
Breath Volume	2.949	19:24
Air Blank	0.000	19:24
Control Test	0.077	19:25
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

---

Data Download Date/Time: 07/13/2018 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,SHAWN,W  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: MARTIN,SHAWN  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: ARGUETA

DATE/TIME OF TEST: 06/11/2018 19:59:25

---

Diagnostics Check	OK	19:52
Air Blank	0.000	19:53
Control Test	0.081	19:53
Air Blank	0.000	19:53
Subject Sample # 1	0.149	19:54
Breath Volume	3.886	19:54
Air Blank	0.000	19:55
Air Blank	0.000	19:57
Subject Sample # 2	0.150	19:57
Breath Volume	2.718	19:57
Air Blank	0.000	19:58
Control Test	0.080	19:58
Air Blank	0.000	19:59
Diagnostics Check	OK	19:59

---

Data Download Date/Time: 07/13/2018 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARRILLO,JULIO,C  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: TORRES,JOEL  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BESENFELDER

DATE/TIME OF TEST: 06/13/2018 23:23:01

---

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:20	
Control Test	0.081	23:20	
Air Blank	0.000	23:21	
Subject Sample	# 1	23:21	Subject Test Refused
Breath Volume	0.000	23:21	
Air Blank	0.000	23:22	
Control Test	0.081	23:22	
Air Blank	0.000	23:22	
Diagnostics Check	OK	23:23	

---

Data Download Date/Time: 07/13/2018 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'KEEFE,PATRICIA,M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: TORRES,JOEL  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/14/2018 05:47:20

---

Diagnostics Check	OK	05:41
Air Blank	0.000	05:41
Control Test	0.082	05:41
Air Blank	0.000	05:42
Subject Sample # 1	0.000	05:42
Breath Volume	2.468	05:42
Air Blank	0.000	05:43
Air Blank	0.000	05:45
Subject Sample # 2	0.000	05:45
Breath Volume	2.949	05:45
Air Blank	0.000	05:46
Control Test	0.081	05:46
Air Blank	0.000	05:47
Diagnostics Check	OK	05:47

---

Data Download Date/Time: 07/13/2018 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'KEEFE,PATRICIA,M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: TORRES,JOEL  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: LISMORE

DATE/TIME OF TEST: 06/15/2018 04:28:26

---

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:25	
Control Test	0.081	04:25	
Air Blank	0.000	04:25	
Subject Sample	# 1	04:27	Subject Test Refused
Breath Volume	0.000	04:27	
Air Blank	0.000	04:27	
Control Test	0.081	04:27	
Air Blank	0.000	04:28	
Diagnostics Check	OK	04:28	

---

Data Download Date/Time: 07/13/2018 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,SHAWN,W  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: FAJARDO,MARIO  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: BARTH

DATE/TIME OF TEST: 06/17/2018 20:14:55

---

Diagnostics Check	OK	20:07
Air Blank	0.000	20:07
Control Test	0.081	20:07
Air Blank	0.000	20:08
Subject Sample # 1	0.071	20:09
Breath Volume	2.777	20:09
Air Blank	0.000	20:10
Air Blank	0.000	20:12
Subject Sample # 2	0.062	20:13
Breath Volume	1.691	20:13
Air Blank	0.000	20:14
Control Test	0.081	20:14
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

---

Data Download Date/Time: 07/13/2018 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KILMURRAY,CHRISTOPHER,P  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: KILMURRAY,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: SMART

DATE/TIME OF TEST: 06/17/2018 22:36:14

---

Diagnostics Check	OK	22:33	
Air Blank	0.000	22:33	
Control Test	0.082	22:34	
Air Blank	0.000	22:34	
Subject Sample	# 1	22:34	Subject Test Refused
Breath Volume	0.000	22:34	
Air Blank	0.000	22:35	
Control Test	0.081	22:35	
Air Blank	0.000	22:36	
Diagnostics Check	OK	22:36	

---

Data Download Date/Time: 07/13/2018 13:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KILMURRAY,CHRISTOPHER,P  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: FARR,JASON  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: ORTIZ VAZQUEZ

DATE/TIME OF TEST: 06/22/2018 03:17:54

---

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:14	
Control Test	0.081	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:16	Subject Test Refused
Breath Volume	0.000	03:16	
Air Blank	0.000	03:16	
Control Test	0.081	03:17	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

---

Data Download Date/Time: 07/13/2018 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'KEEFE,PATRICIA,M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: RIOS,JOEL  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: PENNENGA

DATE/TIME OF TEST: 06/22/2018 23:03:05

---

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:59	
Control Test	0.081	22:59	
Air Blank	0.000	23:00	
Subject Sample	# 1	23:01	Subject Test Refused
Breath Volume	0.000	23:01	
Air Blank	0.000	23:02	
Control Test	0.081	23:02	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

---

Data Download Date/Time: 07/13/2018 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'KEEFE,PATRICIA,M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: TORRES,JOEL  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 06/23/2018 04:19:54

---

Diagnostics Check		OK	04:08	
Air Blank		0.000	04:08	
Control Test		0.082	04:09	
Air Blank		0.000	04:09	
Subject Sample	# 1	0.143	04:12	Volume Not Met
Breath Volume		1.175	04:12	
Air Blank		0.000	04:13	
Air Blank		0.000	04:15	
Subject Sample	# 2	0.119	04:18	Volume Not Met
Breath Volume		0.136	04:18	
Air Blank		0.000	04:18	
Control Test		0.080	04:19	
Air Blank		0.000	04:19	
Diagnostics Check		OK	04:19	

---

Data Download Date/Time: 07/13/2018 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'KEEFE,PATRICIA,M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: TORRES,JOEL  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 06/23/2018 04:30:40

---

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.162	04:24
Breath Volume	1.710	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.158	04:29
Breath Volume	2.738	04:29
Air Blank	0.000	04:29
Control Test	0.081	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

---

Data Download Date/Time: 07/13/2018 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: HERNANDEZ,JOSE  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 06/24/2018 03:17:45

---

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.081	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.218	03:12
Breath Volume		2.621	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.212	03:16
Breath Volume		2.171	03:16
Air Blank		0.000	03:16
Control Test		0.081	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

---

Data Download Date/Time: 07/13/2018 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,CLINT,R  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: WILLIAMS,CLINT  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 18:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: UZ

DATE/TIME OF TEST: 06/30/2018 19:39:03

---

Diagnostics Check	OK	19:36	
Air Blank	0.000	19:36	
Control Test	0.081	19:36	
Air Blank	0.000	19:37	
Subject Sample	# 1	19:37	Subject Test Refused
Breath Volume	0.000	19:37	
Air Blank	0.000	19:38	
Control Test	0.081	19:38	
Air Blank	0.000	19:38	
Diagnostics Check	OK	19:39	

---

Data Download Date/Time: 07/13/2018 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOBLEY,BRITTANY,R  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: THOMPSON,WILLIAM  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 06/30/2018 19:49:59

---

Diagnostics Check	OK	19:44	
Air Blank	0.000	19:45	
Control Test	0.080	19:45	
Air Blank	0.000	19:46	
Subject Sample	# 1 0.171	19:47	
Breath Volume	3.671	19:47	
Air Blank	0.000	19:47	
Air Blank		19:49	Ambient Fail
Air Blank	0.000	19:49	

---

Data Download Date/Time: 07/13/2018 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOBLEY,BRITTANY,R  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: THOMPSON,WILLIAM  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 823252  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 06/30/2018 20:01:12

---

Diagnostics Check	OK	19:54
Air Blank	0.000	19:54
Control Test	0.080	19:55
Air Blank	0.000	19:55
Subject Sample # 1	0.150	19:55
Breath Volume	1.730	19:55
Air Blank	0.000	19:56
Air Blank	0.000	19:58
Subject Sample # 2	0.157	19:59
Breath Volume	2.757	19:59
Air Blank	0.000	20:00
Control Test	0.080	20:00
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

---

Data Download Date/Time: 07/13/2018 13:24



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,RAUL,  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: RODRIGUEZ,RAUL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1870563  
DRY GAS EXPIRY DATE: 06/17/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: GALLARDO

DATE/TIME OF TEST: 06/03/2018 02:14:34

---

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:06	
Control Test	0.081	02:06	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:10	No Sample Provided
Breath Volume	0.000	02:10	
Air Blank	0.000	02:11	
Air Blank	0.000	02:13	
Subject Sample	# 2	02:13	Subject Test Refused
Breath Volume	0.000	02:13	
Air Blank	0.000	02:13	
Control Test	0.082	02:13	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:14	

---

Data Download Date/Time: 06/20/2018 10:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAYS,DARRYL,A  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: RODRIGUEZ,RAUL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1870563  
DRY GAS EXPIRY DATE: 06/17/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: ASICONA ASICONA

DATE/TIME OF TEST: 06/16/2018 02:11:20

---

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.083	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.192	02:06
Breath Volume	1.335	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.178	02:09
Breath Volume	1.242	02:09
Air Blank	0.000	02:10
Control Test	0.083	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

---

Data Download Date/Time: 06/20/2018 10:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SINCORE,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: SINCORE,MICHAEL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1870563  
DRY GAS EXPIRY DATE: 04/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 06/17/2018 04:30:07

---

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.082	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.224	04:25
Breath Volume		2.699	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:27
Subject Sample	# 2	0.240	04:28
Breath Volume		2.957	04:28
Air Blank		0.000	04:29
Control Test		0.082	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

---

Data Download Date/Time: 06/20/2018 10:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: DELGADO,MIGUEL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: LATORTUE

DATE/TIME OF TEST: 06/02/2018 03:24:51

---

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:18	
Control Test	0.079	03:18	
Air Blank	0.000	03:19	
Subject Sample # 1	0.115	03:20	
Breath Volume	3.730	03:20	
Air Blank	0.000	03:20	
Air Blank	0.000	03:22	
Subject Sample # 2	0.118	03:23	
Breath Volume	3.699	03:23	
Air Blank	0.000	03:23	
Control Test	0.075	03:24	Interferent Detect
Air Blank	0.000	03:24	

---

Data Download Date/Time: 06/29/2018 11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: DELGADO,MIGUEL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: LATORTUE

DATE/TIME OF TEST: 06/02/2018 03:55:08

---

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:53	
Subject Sample	# 1	03:53	Interferent Detect
Breath Volume	1.789	03:53	
Air Blank	0.000	03:54	
Control Test	0.076	03:54	
Air Blank	0.000	03:55	
Diagnostics Check	OK	03:55	

---

Data Download Date/Time: 06/29/2018 11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: DELGADO,MIGUEL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: LATORTUE

DATE/TIME OF TEST: 06/02/2018 04:16:31

---

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.077	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.108	04:11
Breath Volume	3.132	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.099	04:15
Breath Volume	1.363	04:15
Air Blank	0.000	04:15
Control Test	0.078	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

---

Data Download Date/Time: 06/29/2018 11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CURTIS,WILLIAM,C  
BREATH TEST OPERATOR AGENCY: Aventura PD  
ARRESTING OFFICER: CURTIS,WILLIAM  
ARRESTING OFFICER AGENCY: Aventura PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 07:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:01

SUBJECT'S LAST NAME: BETETA-SILVA

DATE/TIME OF TEST: 06/02/2018 07:33:07

---

Diagnostics Check	OK	07:31	
Air Blank	0.000	07:32	
Control Test	0.072	07:32	Control Outside Tolerance
Air Blank	0.000	07:33	

---

Data Download Date/Time: 06/29/2018 11:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CURTIS,WILLIAM,C  
BREATH TEST OPERATOR AGENCY: Aventura PD  
ARRESTING OFFICER: CURTIS,WILLIAM  
ARRESTING OFFICER AGENCY: Aventura PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 07:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:01

SUBJECT'S LAST NAME: BETETA-SILVA

DATE/TIME OF TEST: 06/02/2018 07:41:50

---

Diagnostics Check	OK	07:34
Air Blank	0.000	07:35
Control Test	0.075	07:35
Air Blank	0.000	07:36
Subject Sample # 1	0.221	07:37
Breath Volume	1.687	07:37
Air Blank	0.000	07:37
Air Blank	0.000	07:39
Subject Sample # 2	0.219	07:40
Breath Volume	1.378	07:40
Air Blank	0.000	07:40
Control Test	0.080	07:41
Air Blank	0.000	07:41
Diagnostics Check	OK	07:41

---

Data Download Date/Time: 06/29/2018 11:11



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AGINS,LYLE,P  
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD  
ARRESTING OFFICER: AGINS,LYLE  
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 07:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:46

SUBJECT'S LAST NAME: GUNTER

DATE/TIME OF TEST: 06/05/2018 08:12:39

---

Diagnostics Check	OK	08:06
Air Blank	0.000	08:07
Control Test	0.081	08:07
Air Blank	0.000	08:08
Subject Sample # 1	0.178	08:08
Breath Volume	1.320	08:08
Air Blank	0.000	08:08
Air Blank	0.000	08:10
Subject Sample # 2	0.176	08:11
Breath Volume	1.132	08:11
Air Blank	0.000	08:11
Control Test	0.078	08:12
Air Blank	0.000	08:12
Diagnostics Check	OK	08:12

---

Data Download Date/Time: 06/29/2018 11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: DELGADO,MIGUEL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 05:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: JACQUES

DATE/TIME OF TEST: 06/09/2018 06:24:45

---

Diagnostics Check	OK	06:21	
Air Blank	0.000	06:22	
Control Test	0.080	06:22	
Air Blank	0.000	06:23	
Subject Sample	# 1	06:23	Subject Test Refused
Breath Volume	0.000	06:23	
Air Blank	0.000	06:23	
Control Test	0.080	06:24	
Air Blank	0.000	06:24	
Diagnostics Check	OK	06:24	

---

Data Download Date/Time: 06/29/2018 11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCWILLIAMS,TERRENCE,R  
BREATH TEST OPERATOR AGENCY: Miami Shores PD  
ARRESTING OFFICER: MCWILLIAMS,TERRENCE  
ARRESTING OFFICER AGENCY: Miami Shores PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 15:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/10/2018 15:50:34

---

Diagnostics Check	OK	15:47	
Air Blank	0.000	15:48	
Control Test	0.077	15:48	
Air Blank	0.000	15:49	
Subject Sample	# 1	15:49	Subject Test Refused
Breath Volume	0.000	15:49	
Air Blank	0.000	15:49	
Control Test	0.083	15:49	
Air Blank	0.000	15:50	
Diagnostics Check	OK	15:50	

---

Data Download Date/Time: 06/29/2018 11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCWILLIAMS,TERRENCE,  
BREATH TEST OPERATOR AGENCY: Miami Shores PD  
ARRESTING OFFICER: MCWILLIAMS,TERRENCE  
ARRESTING OFFICER AGENCY: Miami Shores PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 15:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/10/2018 15:55:04

---

Diagnostics Check	OK	15:52	
Air Blank	0.000	15:52	
Control Test	0.078	15:52	
Air Blank	0.000	15:53	
Subject Sample	# 1	15:53	Subject Test Refused
Breath Volume	0.000	15:53	
Air Blank	0.000	15:54	
Control Test	0.079	15:54	
Air Blank	0.000	15:54	
Diagnostics Check	OK	15:55	

---

Data Download Date/Time: 06/29/2018 11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HELFNER,ISAAC,S  
BREATH TEST OPERATOR AGENCY: Aventura PD  
ARRESTING OFFICER: HELFNER,ISAAC  
ARRESTING OFFICER AGENCY: Aventura PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: DENIS

DATE/TIME OF TEST: 06/17/2018 01:49:49

---

Diagnostics Check		OK	01:38
Air Blank		0.000	01:38
Control Test		0.079	01:39
Air Blank		0.000	01:39
Subject Sample	# 1	0.184	01:40
Breath Volume		1.414	01:40
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.157	01:43
Breath Volume		1.589	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 3	0.183	01:48
Breath Volume		1.835	01:48
Air Blank		0.000	01:48
Control Test		0.076	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

---

Data Download Date/Time: 06/29/2018 11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VARGHESE, JIBY,  
BREATH TEST OPERATOR AGENCY: Aventura PD  
ARRESTING OFFICER: ALFARO, CHRISTOPHER  
ARRESTING OFFICER AGENCY: Aventura PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DEMCHENKO

DATE/TIME OF TEST: 06/22/2018 02:34:47

---

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.074	02:34	Control Outside Tolerance
Air Blank	0.000	02:34	

---

Data Download Date/Time: 06/29/2018 11:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VARGHESE,JIBY,  
BREATH TEST OPERATOR AGENCY: Aventura PD  
ARRESTING OFFICER: ALFARO,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Aventura PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DEMCHENKO

DATE/TIME OF TEST: 06/22/2018 02:45:06

---

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:38	
Control Test	0.077	02:38	
Air Blank	0.000	02:39	
Subject Sample	# 1 0.396	02:40	
Breath Volume	3.582	02:40	
Air Blank	0.000	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 2	02:43	RFI Detect
Breath Volume	0.000	02:43	
Air Blank	0.000	02:44	
Control Test	0.075	02:44	
Air Blank		02:44	RFI Detect
Air Blank		02:45	RFI Detect

---

Data Download Date/Time: 06/29/2018 11:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: LEON,NEKODA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 06/22/2018 05:22:58

---

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.077	05:17
Air Blank	0.000	05:17
Subject Sample # 1	0.142	05:17
Breath Volume	2.355	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.144	05:21
Breath Volume	1.136	05:21
Air Blank	0.000	05:22
Control Test	0.079	05:22
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

---

Data Download Date/Time: 06/29/2018 11:11



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KONETZ,JONATHAN,D  
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD  
ARRESTING OFFICER: KONETZ,JONATHAN  
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: LEVENETS

DATE/TIME OF TEST: 06/26/2018 01:28:45

---

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:21	
Control Test	0.077	01:21	
Air Blank	0.000	01:22	
Subject Sample # 1	0.215	01:23	
Breath Volume	1.378	01:23	
Air Blank	0.000	01:24	
Air Blank	0.000	01:26	
Subject Sample # 2	0.224	01:27	
Breath Volume	1.667	01:27	
Air Blank	0.000	01:27	
Control Test	0.074	01:28	Control Outside Tolerance
Air Blank	0.000	01:28	

---

Data Download Date/Time: 06/29/2018 11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,STEVEN,L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: ADAMS,STEVEN  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MIRABAL-LEON

DATE/TIME OF TEST: 06/04/2018 02:58:17

---

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.194	02:53
Breath Volume	1.156	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.194	02:56
Breath Volume	2.210	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

---

Data Download Date/Time: 06/29/2018 11:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HODGES,TITUS,R  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: GEBBIA,SALVATORE  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: MUNROE

DATE/TIME OF TEST: 06/05/2018 21:18:09

---

Diagnostics Check	OK	21:10
Air Blank	0.000	21:11
Control Test	0.079	21:11
Air Blank	0.000	21:12
Subject Sample # 1	0.242	21:13
Breath Volume	1.386	21:13
Air Blank	0.000	21:14
Air Blank	0.000	21:15
Subject Sample # 2	0.257	21:16
Breath Volume	1.761	21:16
Air Blank	0.000	21:17
Control Test	0.079	21:17
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

---

Data Download Date/Time: 06/29/2018 11:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA,CHRISTOPHER,C  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: RIVERA,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/05/2018 21:58:51

---

Diagnostics Check	OK	21:50
Air Blank	0.000	21:50
Control Test	0.080	21:51
Air Blank	0.000	21:51
Subject Sample # 1	0.230	21:52
Breath Volume	1.144	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.223	21:57
Breath Volume	1.683	21:57
Air Blank	0.000	21:57
Control Test	0.078	21:58
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

---

Data Download Date/Time: 06/29/2018 11:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GATTORNO,ALEXANDER,F  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: GATTORNO,ALEXANDER  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 07:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: AQUINO

DATE/TIME OF TEST: 06/10/2018 09:15:52

---

Diagnostics Check		OK	09:08
Air Blank		0.000	09:09
Control Test		0.080	09:09
Air Blank		0.000	09:09
Subject Sample	# 1	0.246	09:10
Breath Volume		2.656	09:10
Air Blank		0.000	09:11
Air Blank		0.000	09:13
Subject Sample	# 2	0.249	09:14
Breath Volume		1.882	09:14
Air Blank		0.000	09:14
Control Test		0.079	09:15
Air Blank		0.000	09:15
Diagnostics Check		OK	09:15

---

Data Download Date/Time: 06/29/2018 11:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HODGES,TITUS,R  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: HODGES,TITUS  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/24/2018 03:47:22

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.212	03:42
Breath Volume	1.417	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.213	03:45
Breath Volume	1.156	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

---

Data Download Date/Time: 06/29/2018 11:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, MAURICE, M  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: HENRY, MAURICE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: CARDENAS

DATE/TIME OF TEST: 06/01/2018 01:45:40

---

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.081	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.000	01:40
Breath Volume	1.488	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.000	01:44
Breath Volume	1.976	01:44
Air Blank	0.000	01:44
Control Test	0.079	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

---

Data Download Date/Time: 06/07/2018 14:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORENO,FRANCISCO,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: MORENO,FRANCISCO  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 07:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:24

SUBJECT'S LAST NAME: CASTROGUZMAN

DATE/TIME OF TEST: 06/20/2018 07:51:16

---

Diagnostics Check	OK	07:44
Air Blank	0.000	07:44
Control Test	0.080	07:45
Air Blank	0.000	07:45
Subject Sample # 1	0.130	07:46
Breath Volume	2.429	07:46
Air Blank	0.000	07:46
Air Blank	0.000	07:48
Subject Sample # 2	0.126	07:49
Breath Volume	1.957	07:49
Air Blank	0.000	07:50
Control Test	0.079	07:50
Air Blank	0.000	07:51
Diagnostics Check	OK	07:51

---

Data Download Date/Time: 07/10/2018 12:30



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,EDWIN,M  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FERRO,LISA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 09:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: DE JESUS

DATE/TIME OF TEST: 06/20/2018 10:37:55

---

Diagnostics Check	OK	10:31
Air Blank	0.000	10:31
Control Test	0.080	10:32
Air Blank	0.000	10:32
Subject Sample # 1	0.297	10:33
Breath Volume	1.406	10:33
Air Blank	0.000	10:33
Air Blank	0.000	10:35
Subject Sample # 2	0.303	10:36
Breath Volume	2.359	10:36
Air Blank	0.000	10:37
Control Test	0.079	10:37
Air Blank	0.000	10:37
Diagnostics Check	OK	10:37

---

Data Download Date/Time: 07/10/2018 12:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,E,M  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PEREZ,EDWIN  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 08:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:36

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/21/2018 09:03:26

---

Diagnostics Check		OK	08:56
Air Blank		0.000	08:56
Control Test		0.080	08:57
Air Blank		0.000	08:57
Subject Sample	# 1	0.169	08:58
Breath Volume		2.617	08:58
Air Blank		0.000	08:59
Air Blank		0.000	09:01
Subject Sample	# 2	0.165	09:01
Breath Volume		2.296	09:01
Air Blank		0.000	09:02
Control Test		0.079	09:02
Air Blank		0.000	09:03
Diagnostics Check		OK	09:03

---

Data Download Date/Time: 07/10/2018 12:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HODGES,TITUS,R  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: HODGES,TITUS  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 06/28/2018 18:31:33

---

Diagnostics Check	OK	18:23
Air Blank	0.000	18:23
Control Test	0.080	18:24
Air Blank	0.000	18:24
Subject Sample # 1	0.223	18:26
Breath Volume	1.695	18:26
Air Blank	0.000	18:27
Air Blank	0.000	18:29
Subject Sample # 2	0.226	18:30
Breath Volume	2.269	18:30
Air Blank	0.000	18:30
Control Test	0.079	18:30
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

---

Data Download Date/Time: 07/10/2018 12:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,ADRIAN,  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: GOMEZ,RUBEN  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: BUI  
ARREST TIME: 18:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 06/06/2018 19:40:03

---

Diagnostics Check	OK	19:36	
Air Blank	0.000	19:36	
Control Test	0.079	19:37	
Air Blank	0.000	19:37	
Subject Sample	# 1	19:38	Subject Test Refused
Breath Volume	0.000	19:38	
Air Blank	0.000	19:39	
Control Test	0.080	19:39	
Air Blank	0.000	19:39	
Diagnostics Check	OK	19:40	

---

Data Download Date/Time: 06/18/2018 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAPANES,R,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: TAPANES,R  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: RICARDO

DATE/TIME OF TEST: 06/07/2018 02:37:16

---

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.080	02:33	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:36	
Control Test	0.080	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

---

Data Download Date/Time: 06/18/2018 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CASTILLO,CHRISTIAN  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: VEGA DELGADO

DATE/TIME OF TEST: 06/11/2018 22:33:40

---

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.079	22:27
Air Blank	0.000	22:27
Subject Sample # 1	0.273	22:28
Breath Volume	2.738	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.254	22:32
Breath Volume	3.941	22:32
Air Blank	0.000	22:32
Control Test	0.079	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

---

Data Download Date/Time: 06/18/2018 11:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHILPOT,C,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PHILPOT,C  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: BRITZIUS

DATE/TIME OF TEST: 06/14/2018 03:33:54

---

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.168	03:29
Breath Volume	3.101	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:31
Subject Sample # 2	0.163	03:32
Breath Volume	2.531	03:32
Air Blank	0.000	03:32
Control Test	0.078	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

---

Data Download Date/Time: 06/18/2018 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHILPOT,C,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PHILPOT,C  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BELLINGER

DATE/TIME OF TEST: 06/30/2018 03:53:21

---

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.080	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

---

Data Download Date/Time: 07/12/2018 09:46



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NADAL,JOSE,V  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: NADAL,JOSE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: ALFONSO

DATE/TIME OF TEST: 06/02/2018 04:06:45

---

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:03	
Control Test	0.080	04:03	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:05	Subject Test Refused
Breath Volume	0.000	04:05	
Air Blank	0.000	04:05	
Control Test	0.080	04:06	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

---

Data Download Date/Time: 06/18/2018 11:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAPANES,R,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: VINAS,ERNESTO  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/06/2018 01:56:06

---

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.080	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1 0.200	01:54	
Breath Volume	1.585	01:54	
Air Blank		01:55	Purge Fail
Air Blank		01:56	Purge Fail

---

Data Download Date/Time: 06/18/2018 11:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAPANES,R,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: VINAS,E  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/06/2018 02:05:22

---

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.079	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.208	02:00
Breath Volume		1.652	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.199	02:03
Breath Volume		2.429	02:03
Air Blank		0.000	02:04
Control Test		0.079	02:04
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

---

Data Download Date/Time: 06/18/2018 11:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VARGAS,D,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: VARGAS,D  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 06/07/2018 00:33:22

---

Diagnostics Check		OK	00:26
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:28
Subject Sample	# 1	0.218	00:28
Breath Volume		1.191	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.231	00:31
Breath Volume		1.332	00:31
Air Blank		0.000	00:32
Control Test		0.080	00:32
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

---

Data Download Date/Time: 06/18/2018 11:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VARGAS,D,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: VARGAS,D  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: DE JESUS

DATE/TIME OF TEST: 06/10/2018 04:10:04

---

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:07	
Control Test	0.079	04:07	
Air Blank	0.000	04:08	
Subject Sample	# 1	04:08	Subject Test Refused
Breath Volume	0.000	04:08	
Air Blank	0.000	04:09	
Control Test	0.080	04:09	
Air Blank	0.000	04:09	
Diagnostics Check	OK	04:10	

---

Data Download Date/Time: 06/18/2018 11:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAWSON, GLADSTON,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CRUZ, JOEL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 05/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: LABANINO

DATE/TIME OF TEST: 06/10/2018 22:46:03

---

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:37	
Control Test	0.079	22:38	
Air Blank	0.000	22:38	
Subject Sample	# 1	22:41	No Sample Provided
Breath Volume	0.000	22:41	
Air Blank	0.000	22:42	
Air Blank	0.000	22:44	
Subject Sample	# 2	22:44	Subject Test Refused
Breath Volume	0.000	22:44	
Air Blank	0.000	22:45	
Control Test	0.079	22:45	
Air Blank	0.000	22:45	
Diagnostics Check	OK	22:46	

---

Data Download Date/Time: 06/18/2018 11:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NADAL,JOSE,V  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: NADAL,JOSE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CORONADO SURT

DATE/TIME OF TEST: 06/23/2018 04:27:19

---

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.080	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.178	04:21
Breath Volume	2.359	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.176	04:25
Breath Volume	4.703	04:25
Air Blank	0.000	04:26
Control Test	0.082	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

---

Data Download Date/Time: 07/12/2018 11:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOO,B,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: HOO,BRANDON  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 17:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: JOFRE VALENZUELA

DATE/TIME OF TEST: 06/23/2018 18:44:42

---

Diagnostics Check	OK	18:44	
Air Blank		18:44	RFI Detect
Air Blank		18:44	RFI Detect

---

Data Download Date/Time: 07/12/2018 11:32



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOO,BRANDON,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: HOO,BRANDON  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 17:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: JOFRE VALENZUELA

DATE/TIME OF TEST: 06/23/2018 18:48:28

---

Diagnostics Check	OK	18:47	
Air Blank		18:48	RFI Detect
Air Blank		18:48	RFI Detect

---

Data Download Date/Time: 07/12/2018 11:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOO,BRANDON,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: HOO,BRANDON  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 17:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: JOFRE VALENZUELA

DATE/TIME OF TEST: 06/23/2018 18:54:58

---

Diagnostics Check	OK	18:52	
Air Blank	0.000	18:52	
Control Test	0.079	18:52	
Air Blank	0.000	18:53	
Subject Sample	# 1	18:53	Subject Test Refused
Breath Volume	0.000	18:53	
Air Blank	0.000	18:54	
Control Test	0.079	18:54	
Air Blank	0.000	18:54	
Diagnostics Check	OK	18:54	

---

Data Download Date/Time: 07/12/2018 11:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NADAL,JOSE,V  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: MALDONADO,MIGUEL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: BETANCOURT REYES

DATE/TIME OF TEST: 06/25/2018 01:15:10

---

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.226	01:10
Breath Volume	2.558	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:12
Subject Sample # 2	0.222	01:13
Breath Volume	2.902	01:13
Air Blank	0.000	01:14
Control Test	0.080	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

---

Data Download Date/Time: 07/12/2018 11:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NADAL,JOSE,V  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: NADAL,JOSE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A2  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 06:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: SALES

DATE/TIME OF TEST: 06/25/2018 07:09:45

---

Diagnostics Check	OK	07:03
Air Blank	0.000	07:03
Control Test	0.079	07:03
Air Blank	0.000	07:04
Subject Sample # 1	0.247	07:05
Breath Volume	2.015	07:05
Air Blank	0.000	07:05
Air Blank	0.000	07:07
Subject Sample # 2	0.253	07:08
Breath Volume	3.523	07:08
Air Blank	0.000	07:08
Control Test	0.080	07:09
Air Blank	0.000	07:09
Diagnostics Check	OK	07:09

---

Data Download Date/Time: 07/12/2018 11:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAYA,JOSE,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: MAYA,JOSE  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 05/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12317080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: TECUN

DATE/TIME OF TEST: 06/07/2018 03:18:30

---

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.079	03:16	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:17	
Control Test	0.078	03:17	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

---

Data Download Date/Time: 06/26/2018 23:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAYA,JOSE,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: ALADIN,K  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 05/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12317080A3  
DRY GAS EXPIRY DATE: 07/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: OROZCO RUIZ

DATE/TIME OF TEST: 06/08/2018 20:43:41

---

Diagnostics Check	OK	20:34
Air Blank	0.000	20:35
Control Test	0.079	20:35
Air Blank	0.000	20:36
Subject Sample # 1	0.104	20:37
Breath Volume	1.144	20:37
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.106	20:42
Breath Volume	1.550	20:42
Air Blank	0.000	20:42
Control Test	0.076	20:43
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

---

Data Download Date/Time: 06/26/2018 23:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIFER,JUSTIN,S  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: WINTON,ADAM  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 17:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: DALTON  
DATE/TIME OF TEST: 06/01/2018 18:38:07

---

Diagnostics Check	OK	18:31
Air Blank	0.000	18:32
Control Test	0.080	18:32
Air Blank	0.000	18:32
Subject Sample # 1	0.161	18:33
Breath Volume	3.714	18:33
Air Blank	0.000	18:34
Air Blank	0.000	18:35
Subject Sample # 2	0.165	18:36
Breath Volume	4.105	18:36
Air Blank	0.000	18:37
Control Test	0.079	18:37
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

---

Data Download Date/Time: 06/13/2018 18:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO:	FWC	Date or Time Changed:	No
SOFTWARE VERSION:	8100.27	DRY GAS STANDARD LOT #:	507129
BREATH TEST OPERATOR:	WILKINSON,WILLIAM,T	DRY GAS EXPIRY DATE:	08/04/2019
BREATH TEST OPERATOR AGENCY:	FFWCC (Panama City)	VIOLATION CODE:	BUI
ARRESTING OFFICER:	JOHNSTON,JOE	ARREST TIME:	19:20
ARRESTING OFFICER AGENCY:	FFWCC (Panama City)	GENDER:	Male
LAST AGENCY INSPECTION:	05/09/2018	OBSERVATION PERIOD BEGAN:	19:20

SUBJECT'S LAST NAME:	NECAISE	DATE/TIME OF TEST:	06/01/2018	19:59:34
----------------------	---------	--------------------	------------	----------

---

Diagnostics Check	OK	19:56	
Air Blank	0.000	19:57	
Control Test	0.080	19:57	
Air Blank	0.000	19:57	
Subject Sample	# 1	19:58	Subject Test Refused
Breath Volume	0.000	19:58	
Air Blank	0.000	19:58	
Control Test	0.080	19:58	
Air Blank	0.000	19:59	
Diagnostics Check	OK	19:59	

---

Data Download Date/Time: 06/13/2018 18:27



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEIN,PAUL,M  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: HEIN,PAUL  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 20:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: LASATER

DATE/TIME OF TEST: 06/01/2018 21:29:13

---

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:25	
Control Test	0.080	21:26	
Air Blank	0.000	21:26	
Subject Sample	# 1	21:27	Subject Test Refused
Breath Volume	0.000	21:27	
Air Blank	0.000	21:28	
Control Test	0.081	21:28	
Air Blank	0.000	21:29	
Diagnostics Check	OK	21:29	

---

Data Download Date/Time: 06/13/2018 18:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NICHOLS,HEATH,C  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: MOLNAR,JARROD  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 14:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:16

SUBJECT'S LAST NAME: NAVARRE

DATE/TIME OF TEST: 06/02/2018 14:44:07

---

Diagnostics Check	OK	14:40	
Air Blank	0.000	14:41	
Control Test	0.080	14:41	
Air Blank	0.000	14:42	
Subject Sample	# 1	14:42	Subject Test Refused
Breath Volume	0.000	14:42	
Air Blank	0.000	14:43	
Control Test	0.081	14:43	
Air Blank	0.000	14:44	
Diagnostics Check	OK	14:44	

---

Data Download Date/Time: 06/13/2018 18:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBB,MATT,  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: CARTER',J  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 14:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: WESTERFIELD

DATE/TIME OF TEST: 06/02/2018 14:50:23

---

Diagnostics Check	OK	14:47	
Air Blank	0.000	14:47	
Control Test	0.080	14:48	
Air Blank	0.000	14:48	
Subject Sample	# 1	14:49	Subject Test Refused
Breath Volume	0.000	14:49	
Air Blank	0.000	14:49	
Control Test	0.081	14:49	
Air Blank	0.000	14:50	
Diagnostics Check	OK	14:50	

---

Data Download Date/Time: 06/13/2018 18:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NICHOLS,HEATH,C  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: MALTAIS,ANDREW  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 14:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:51

SUBJECT'S LAST NAME: STAFFORD

DATE/TIME OF TEST: 06/02/2018 15:56:17

---

Diagnostics Check	OK	15:48
Air Blank	0.000	15:49
Control Test	0.081	15:49
Air Blank	0.000	15:49
Subject Sample # 1	0.127	15:51
Breath Volume	2.343	15:51
Air Blank	0.000	15:52
Air Blank	0.000	15:54
Subject Sample # 2	0.130	15:54
Breath Volume	2.820	15:54
Air Blank	0.000	15:55
Control Test	0.076	15:55
Air Blank	0.000	15:56
Diagnostics Check	OK	15:56

---

Data Download Date/Time: 06/13/2018 18:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NICHOLS,HEATH,C  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: WHITE,REID  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 06/02/2018 16:46:28

---

Diagnostics Check	OK	16:39
Air Blank	0.000	16:40
Control Test	0.079	16:40
Air Blank	0.000	16:41
Subject Sample # 1	0.124	16:41
Breath Volume	1.644	16:41
Air Blank	0.000	16:42
Air Blank	0.000	16:44
Subject Sample # 2	0.112	16:44
Breath Volume	2.929	16:44
Air Blank	0.000	16:45
Control Test	0.079	16:45
Air Blank	0.000	16:46
Diagnostics Check	OK	16:46

---

Data Download Date/Time: 06/13/2018 18:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBB,MATT,  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: WINTON,A  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 16:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:23

SUBJECT'S LAST NAME: SENS

DATE/TIME OF TEST: 06/02/2018 17:00:14

---

Diagnostics Check	OK	16:57	
Air Blank	0.000	16:57	
Control Test	0.079	16:58	
Air Blank	0.000	16:58	
Subject Sample	# 1	16:58	Subject Test Refused
Breath Volume	0.000	16:58	
Air Blank	0.000	16:59	
Control Test	0.080	16:59	
Air Blank	0.000	17:00	
Diagnostics Check	OK	17:00	

---

Data Download Date/Time: 06/13/2018 18:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KLEIS,KEVIN,W  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: KLEIS,KEVIN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 17:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/02/2018 18:04:35

---

Diagnostics Check	OK	17:49	
Air Blank	0.000	17:50	
Control Test	0.080	17:50	
Air Blank	0.000	17:51	
Subject Sample # 1	0.109	17:52	
Breath Volume	1.207	17:52	
Air Blank	0.000	17:53	
Air Blank	0.000	17:54	
Subject Sample # 2	0.079	17:57	Volume Not Met
Breath Volume	0.722	17:57	
Air Blank	0.000	17:58	
Air Blank	0.000	18:00	
Subject Sample # 3	0.087	18:03	Volume Not Met
Breath Volume	0.316	18:03	
Air Blank		18:04	RFI Detect
Air Blank	0.000	18:04	

---

Data Download Date/Time: 06/13/2018 18:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KLEIS,KEVIN,W  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: KLEIS,KEVIN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 17:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/02/2018 18:23:18

---

Diagnostics Check	OK	18:07	
Air Blank	0.000	18:07	
Control Test	0.078	18:07	
Air Blank	0.000	18:08	
Subject Sample	# 1 0.088	18:10	
Breath Volume	1.113	18:10	
Air Blank	0.000	18:11	
Air Blank	0.000	18:12	
Subject Sample	# 2 0.080	18:16	Volume Not Met
Breath Volume	0.101	18:16	
Air Blank	0.000	18:16	
Air Blank	0.000	18:18	
Subject Sample	# 3 0.084	18:21	Volume Not Met
Breath Volume	0.183	18:21	
Air Blank	0.000	18:22	
Control Test	0.076	18:22	
Air Blank	0.000	18:23	
Diagnostics Check	OK	18:23	

---

Data Download Date/Time: 06/13/2018 18:27





# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIFER,JUSTIN,S  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: PIFER,JUSTIN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 17:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: KATSIAS

DATE/TIME OF TEST: 06/08/2018 18:31:01

---

Diagnostics Check	OK	18:24
Air Blank	0.000	18:24
Control Test	0.080	18:24
Air Blank	0.000	18:25
Subject Sample # 1	0.155	18:26
Breath Volume	3.582	18:26
Air Blank	0.000	18:26
Air Blank	0.000	18:28
Subject Sample # 2	0.161	18:29
Breath Volume	3.851	18:29
Air Blank	0.000	18:30
Control Test	0.078	18:30
Air Blank	0.000	18:30
Diagnostics Check	OK	18:31

---

Data Download Date/Time: 06/13/2018 18:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOLNAR,JARROD,T  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: MOLNAR,JARROD  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 15:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:27

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 06/09/2018 16:18:10

---

Diagnostics Check	OK	16:12
Air Blank	0.000	16:12
Control Test	0.080	16:12
Air Blank	0.000	16:13
Subject Sample # 1	0.205	16:13
Breath Volume	4.890	16:13
Air Blank	0.000	16:14
Air Blank	0.000	16:16
Subject Sample # 2	0.198	16:16
Breath Volume	4.234	16:16
Air Blank	0.000	16:17
Control Test	0.078	16:17
Air Blank	0.000	16:18
Diagnostics Check	OK	16:18

---

Data Download Date/Time: 06/13/2018 18:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIFER,JUSTIN,S  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: HOMAN,MICHAEL  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 17:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:51

SUBJECT'S LAST NAME: THOMPSON      DATE/TIME OF TEST: 06/09/2018      18:40:05

---

Diagnostics Check	OK	18:31
Air Blank	0.000	18:32
Control Test	0.080	18:32
Air Blank	0.000	18:33
Subject Sample # 1	0.080	18:33
Breath Volume	1.570	18:33
Air Blank	0.000	18:34
Air Blank	0.000	18:36
Subject Sample # 2	0.081	18:38
Breath Volume	1.371	18:38
Air Blank	0.000	18:39
Control Test	0.079	18:39
Air Blank	0.000	18:40
Diagnostics Check	OK	18:40

---

Data Download Date/Time: 06/13/2018 18:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOLNAR,JARROD,T  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: CORBIN,BRITTON  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 18:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: BLANKENSHIP  
DATE/TIME OF TEST: 06/09/2018 18:53:31

---

Diagnostics Check	OK	18:45
Air Blank	0.000	18:46
Control Test	0.079	18:46
Air Blank	0.000	18:47
Subject Sample # 1	0.088	18:49
Breath Volume	1.578	18:49
Air Blank	0.000	18:49
Air Blank	0.000	18:51
Subject Sample # 2	0.087	18:51
Breath Volume	3.265	18:51
Air Blank	0.000	18:52
Control Test	0.078	18:52
Air Blank	0.000	18:53
Diagnostics Check	OK	18:53

---

Data Download Date/Time: 06/13/2018 18:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOLNAR,JARROD,T  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: PIFER,JUSTIN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 21:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 06/09/2018 22:08:19

---

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:05	
Control Test	0.080	22:06	
Air Blank	0.000	22:06	
Subject Sample	# 1	22:06	Subject Test Refused
Breath Volume	0.000	22:06	
Air Blank	0.000	22:07	
Control Test	0.082	22:07	
Air Blank	0.000	22:08	
Diagnostics Check	OK	22:08	

---

Data Download Date/Time: 06/13/2018 18:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIFER,JUSTIN,S  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: PIFER,JUSTIN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 21:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 06/09/2018 23:00:34

---

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.081	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.147	22:55
Breath Volume	1.726	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.148	22:58
Breath Volume	1.632	22:58
Air Blank	0.000	22:59
Control Test	0.079	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

---

Data Download Date/Time: 06/13/2018 18:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOLNAR,JARROD,T  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: PIFER,JUSTIN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 16:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:46

SUBJECT'S LAST NAME: FELICIANO GARCIA

DATE/TIME OF TEST: 06/10/2018 17:24:49

---

Diagnostics Check	OK	17:18
Air Blank	0.000	17:19
Control Test	0.081	17:19
Air Blank	0.000	17:20
Subject Sample # 1	0.055	17:20
Breath Volume	1.457	17:20
Air Blank	0.000	17:21
Air Blank	0.000	17:22
Subject Sample # 2	0.055	17:23
Breath Volume	2.441	17:23
Air Blank	0.000	17:23
Control Test	0.079	17:24
Air Blank	0.000	17:24
Diagnostics Check	OK	17:24

---

Data Download Date/Time: 06/13/2018 18:27



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MALTAIS,ANDREW,W  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: CORBIN,BRITTON  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 16:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 06/16/2018 17:17:38

---

Diagnostics Check	OK	17:14	
Air Blank	0.000	17:15	
Control Test	0.079	17:15	
Air Blank	0.000	17:16	
Subject Sample	# 1	17:16	Subject Test Refused
Breath Volume	0.000	17:16	
Air Blank	0.000	17:16	
Control Test	0.079	17:17	
Air Blank	0.000	17:17	
Diagnostics Check	OK	17:17	

---

Data Download Date/Time: 07/23/2018 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOLNAR,JARROD,T  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: WAGNER,ROBERT  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 18:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: VICE

DATE/TIME OF TEST: 06/16/2018 18:58:02

---

Diagnostics Check	OK	18:55	
Air Blank	0.000	18:55	
Control Test	0.081	18:55	
Air Blank	0.000	18:56	
Subject Sample	# 1	18:56	Subject Test Refused
Breath Volume	0.000	18:56	
Air Blank	0.000	18:57	
Control Test	0.081	18:57	
Air Blank	0.000	18:57	
Diagnostics Check	OK	18:58	

---

Data Download Date/Time: 07/23/2018 15:22



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIFER,JUSTIN,S  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: MOLNAR,JARROD  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 16:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 06/17/2018 17:18:09

---

Diagnostics Check	OK	17:09	
Air Blank	0.000	17:09	
Control Test	0.080	17:09	
Air Blank	0.000	17:10	
Subject Sample # 1	0.020	17:13	
Breath Volume	1.976	17:13	
Air Blank	0.000	17:13	
Air Blank	0.000	17:15	
Subject Sample # 2	0.014	17:17	
Breath Volume	1.582	17:17	
Air Blank		17:17	RFI Detect
Air Blank	0.000	17:18	

---

Data Download Date/Time: 07/23/2018 15:22



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORBIN,BRITTON,L  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: CORBITT,KYLE  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 17:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 06/18/2018 18:08:37

---

Diagnostics Check	OK	18:05	
Air Blank	0.000	18:06	
Control Test	0.081	18:06	
Air Blank	0.000	18:06	
Subject Sample	# 1	18:07	Subject Test Refused
Breath Volume	0.000	18:07	
Air Blank	0.000	18:07	
Control Test	0.080	18:08	
Air Blank	0.000	18:08	
Diagnostics Check	OK	18:08	

---

Data Download Date/Time: 07/23/2018 15:22



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORBIN,BRITTON,L  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: CORBIN,BRITTON  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 13:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 06/23/2018 14:25:04

---

Diagnostics Check	OK	14:22	
Air Blank	0.000	14:22	
Control Test	0.080	14:22	
Air Blank	0.000	14:23	
Subject Sample	# 1	14:23	Subject Test Refused
Breath Volume	0.000	14:23	
Air Blank	0.000	14:24	
Control Test	0.080	14:24	
Air Blank	0.000	14:24	
Diagnostics Check	OK	14:25	

---

Data Download Date/Time: 07/23/2018 15:22





**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENDERSON,DARRELL,L  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: CORBIN,BRITTON  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: ZDZIERA

DATE/TIME OF TEST: 06/24/2018 00:16:55

---

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:05	
Control Test	0.080	00:06	
Air Blank	0.000	00:06	
Subject Sample	# 1	00:09	No Sample Provided
Breath Volume	0.000	00:09	
Air Blank	0.000	00:10	
Air Blank	0.000	00:12	
Subject Sample	# 2	00:15	No Sample Provided
Breath Volume	0.000	00:15	
Air Blank	0.000	00:16	
Control Test	0.081	00:16	
Air Blank	0.000	00:16	
Diagnostics Check	OK	00:16	

---

Data Download Date/Time: 07/23/2018 15:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JARVIS,CHARLES,L  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: PIFER,JUSTIN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 17:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 06/24/2018 18:14:44

---

Diagnostics Check	OK	18:08
Air Blank	0.000	18:08
Control Test	0.081	18:09
Air Blank	0.000	18:09
Subject Sample # 1	0.169	18:10
Breath Volume	2.386	18:10
Air Blank	0.000	18:10
Air Blank	0.000	18:12
Subject Sample # 2	0.173	18:13
Breath Volume	2.769	18:13
Air Blank	0.000	18:13
Control Test	0.079	18:14
Air Blank	0.000	18:14
Diagnostics Check	OK	18:14

---

Data Download Date/Time: 07/23/2018 15:22





# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JARVIS,CHARLES,L  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: CORBIN,BRITTON  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 20:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 06/29/2018 21:12:23

---

Diagnostics Check	OK	21:09	
Air Blank	0.000	21:09	
Control Test	0.080	21:10	
Air Blank	0.000	21:10	
Subject Sample	# 1	21:11	Subject Test Refused
Breath Volume	0.000	21:11	
Air Blank	0.000	21:11	
Control Test	0.081	21:11	
Air Blank	0.000	21:12	
Diagnostics Check	OK	21:12	

---

Data Download Date/Time: 07/23/2018 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REAVES,GEORGE,W  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: PIFER,JUSTIN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 16:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: REAVES

DATE/TIME OF TEST: 06/30/2018 17:12:53

---

Diagnostics Check	OK	17:06
Air Blank	0.000	17:07
Control Test	0.081	17:07
Air Blank	0.000	17:07
Subject Sample # 1	0.178	17:08
Breath Volume	2.968	17:08
Air Blank	0.000	17:08
Air Blank	0.000	17:10
Subject Sample # 2	0.182	17:11
Breath Volume	2.503	17:11
Air Blank	0.000	17:11
Control Test	0.080	17:12
Air Blank	0.000	17:12
Diagnostics Check	OK	17:12

---

Data Download Date/Time: 07/23/2018 15:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JERNIGAN,DAVID,L  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: CORBITT,KYLE  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 16:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: AMAR

DATE/TIME OF TEST: 06/30/2018 17:27:26

---

Diagnostics Check	OK	17:24	
Air Blank	0.000	17:25	
Control Test	0.080	17:25	
Air Blank	0.000	17:25	
Subject Sample	# 1	17:26	Subject Test Refused
Breath Volume	0.000	17:26	
Air Blank	0.000	17:26	
Control Test	0.080	17:26	
Air Blank	0.000	17:27	
Diagnostics Check	OK	17:27	

---

Data Download Date/Time: 07/23/2018 15:22







# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JERNIGAN,DAVID,L  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: WAGNER,ANDY  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 19:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: LAUREL

DATE/TIME OF TEST: 06/30/2018 20:31:44

---

Diagnostics Check	OK	20:25
Air Blank	0.000	20:25
Control Test	0.081	20:26
Air Blank	0.000	20:26
Subject Sample # 1	0.074	20:27
Breath Volume	2.101	20:27
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.079	20:30
Breath Volume	3.230	20:30
Air Blank	0.000	20:30
Control Test	0.080	20:31
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

---

Data Download Date/Time: 07/23/2018 15:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JERNIGAN,DAVID,L  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: CULBERSON,DARYL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 507129  
DRY GAS EXPIRY DATE: 08/04/2019  
VIOLATION CODE: BUI  
ARREST TIME: 19:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 06/30/2018 21:02:26

---

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.081	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.093	20:57
Breath Volume	3.046	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.093	21:00
Breath Volume	3.359	21:00
Air Blank	0.000	21:01
Control Test	0.080	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

---

Data Download Date/Time: 07/23/2018 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BASFORD,TRAVIS,L  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: BASFORD,TRAVIS  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 477627  
DRY GAS EXPIRY DATE: 06/29/2019  
VIOLATION CODE: BUI  
ARREST TIME: 19:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 06/17/2018 20:21:52

---

Diagnostics Check	OK	20:09	
Air Blank	0.000	20:10	
Control Test	0.080	20:10	
Air Blank	0.000	20:11	
Subject Sample	# 1	20:14	No Sample Provided
Breath Volume	0.000	20:14	
Air Blank	0.000	20:14	
Air Blank	0.000	20:16	
Subject Sample	# 2	0.235	20:17
Breath Volume	2.300	20:17	
Air Blank	0.000	20:17	
Air Blank	0.000	20:19	
Subject Sample	# 3	0.239	20:20
Breath Volume	3.304	20:20	
Air Blank	0.000	20:20	
Control Test	0.079	20:21	
Air Blank	0.000	20:21	
Diagnostics Check	OK	20:21	

---

Data Download Date/Time: 07/23/2018 15:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAYES,JARED,A  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: GERBER,OLIVER  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 477627  
DRY GAS EXPIRY DATE: 06/29/2019  
VIOLATION CODE: BUI  
ARREST TIME: 19:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: WHITE  
DATE/TIME OF TEST: 06/30/2018 20:03:45

---

Diagnostics Check	OK	20:00	
Air Blank	0.000	20:01	
Control Test	0.077	20:01	
Air Blank	0.000	20:02	
Subject Sample	# 1	20:02	Subject Test Refused
Breath Volume	0.000	20:02	
Air Blank	0.000	20:02	
Control Test	0.080	20:03	
Air Blank	0.000	20:03	
Diagnostics Check	OK	20:03	

---

Data Download Date/Time: 07/23/2018 15:13



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,ROBERT,B  
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)  
ARRESTING OFFICER: FOX,JAMES  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09117080A2  
DRY GAS EXPIRY DATE: 06/05/2019  
VIOLATION CODE: BUI  
ARREST TIME: 16:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:56

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 06/30/2018 18:06:09

---

Diagnostics Check	OK	18:00
Air Blank	0.000	18:00
Control Test	0.080	18:00
Air Blank	0.000	18:01
Subject Sample # 1	0.185	18:01
Breath Volume	1.746	18:01
Air Blank	0.000	18:02
Air Blank	0.000	18:04
Subject Sample # 2	0.190	18:04
Breath Volume	2.667	18:04
Air Blank	0.000	18:05
Control Test	0.079	18:05
Air Blank	0.000	18:06
Diagnostics Check	OK	18:06

---

Data Download Date/Time: 07/19/2018 10:36



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,ALEX,F  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: MARTINEZ,ALEX  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 05/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A1  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: BUI  
ARREST TIME: 20:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: STENBACK

DATE/TIME OF TEST: 06/10/2018 22:02:36

---

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.077	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.093	21:58
Breath Volume	1.332	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.088	22:01
Breath Volume	2.550	22:01
Air Blank	0.000	22:01
Control Test	0.077	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

---

Data Download Date/Time: 06/14/2018 21:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: PULASKI,DAMON  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 06/14/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A1  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: BUI  
ARREST TIME: 22:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: MAEDER

DATE/TIME OF TEST: 06/23/2018 22:43:04

---

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.080	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.156	22:38
Breath Volume	1.695	22:38
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.153	22:41
Breath Volume	2.843	22:41
Air Blank	0.000	22:42
Control Test	0.081	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:43

---

Data Download Date/Time: 07/18/2018 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,ALEX,F  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: MARTINEZ,ALEX  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 05/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A1  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: BUI  
ARREST TIME: 20:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: MORERA FIGUEREDO

DATE/TIME OF TEST: 06/09/2018 21:30:49

---

Diagnostics Check	OK	21:24
Air Blank	0.000	21:25
Control Test	0.079	21:25
Air Blank	0.000	21:25
Subject Sample # 1	0.137	21:26
Breath Volume	2.402	21:26
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.139	21:29
Breath Volume	1.796	21:29
Air Blank	0.000	21:29
Control Test	0.079	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

---

Data Download Date/Time: 06/14/2018 22:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,MALACHI,C  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: VANTREES,THOMAS  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 05/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A1  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: BUI  
ARREST TIME: 20:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: LEKAKIS

DATE/TIME OF TEST: 06/10/2018 21:26:25

---

Diagnostics Check	OK	21:20
Air Blank	0.000	21:20
Control Test	0.080	21:20
Air Blank	0.000	21:21
Subject Sample # 1	0.131	21:21
Breath Volume	1.468	21:21
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.134	21:24
Breath Volume	1.660	21:24
Air Blank	0.000	21:25
Control Test	0.079	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

---

Data Download Date/Time: 06/14/2018 22:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,MALACHI,C  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: PASS,JUSTIN  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/14/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A1  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: BUI  
ARREST TIME: 16:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: DELAPAZ

DATE/TIME OF TEST: 06/30/2018 17:14:26

---

Diagnostics Check	OK	17:08
Air Blank	0.000	17:08
Control Test	0.079	17:08
Air Blank	0.000	17:09
Subject Sample # 1	0.132	17:09
Breath Volume	3.339	17:09
Air Blank	0.000	17:10
Air Blank	0.000	17:12
Subject Sample # 2	0.132	17:12
Breath Volume	2.554	17:12
Air Blank	0.000	17:13
Control Test	0.079	17:13
Air Blank	0.000	17:14
Diagnostics Check	OK	17:14

---

Data Download Date/Time: 07/18/2018 12:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,MALACHI,C  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: WILKINS,MALACHI  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 06/14/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A1  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: BUI  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: MONROE

DATE/TIME OF TEST: 06/30/2018 22:20:00

---

Diagnostics Check	OK	22:13
Air Blank	0.000	22:14
Control Test	0.079	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.037	22:15
Breath Volume	2.320	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:17
Subject Sample # 2	0.037	22:18
Breath Volume	3.582	22:18
Air Blank	0.000	22:19
Control Test	0.079	22:19
Air Blank	0.000	22:19
Diagnostics Check	OK	22:20

---

Data Download Date/Time: 07/18/2018 12:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MATTSON,CHRIS,J  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: MATTSON,CHRIS  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 666491  
DRY GAS EXPIRY DATE: 12/30/2019  
VIOLATION CODE: BUI  
ARREST TIME: 16:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: MCCALLISTER

DATE/TIME OF TEST: 06/23/2018 16:47:23

---

Diagnostics Check	OK	16:40
Air Blank	0.000	16:41
Control Test	0.082	16:41
Air Blank	0.000	16:42
Subject Sample # 1	0.139	16:42
Breath Volume	2.269	16:42
Air Blank	0.000	16:43
Air Blank	0.000	16:45
Subject Sample # 2	0.126	16:45
Breath Volume	1.468	16:45
Air Blank	0.000	16:46
Control Test	0.081	16:46
Air Blank	0.000	16:47
Diagnostics Check	OK	16:47

---

Data Download Date/Time: 06/27/2018 18:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUNCAN,N,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CALHOUN,K  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: RHOADS

DATE/TIME OF TEST: 06/01/2018 22:58:44

---

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.000	22:54
Breath Volume	2.613	22:54
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.000	22:57
Breath Volume	2.535	22:57
Air Blank	0.000	22:57
Control Test	0.079	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

---

Data Download Date/Time: 06/06/2018 09:11



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WITT,KELLY,  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: WITT,KELLY  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: KYLE

DATE/TIME OF TEST: 06/02/2018 02:36:12

---

Diagnostics Check		OK	02:29
Air Blank		0.000	02:30
Control Test		0.079	02:30
Air Blank		0.000	02:31
Subject Sample	# 1	0.206	02:31
Breath Volume		2.437	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.202	02:34
Breath Volume		2.066	02:34
Air Blank		0.000	02:35
Control Test		0.077	02:35
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

---

Data Download Date/Time: 06/06/2018 09:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPUZZI,KRISTIN,G  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: CAPUZZI,K  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 13:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:22

SUBJECT'S LAST NAME: DUNBAR

DATE/TIME OF TEST: 06/02/2018 14:09:06

---

Diagnostics Check	OK	13:57	
Air Blank	0.000	13:57	
Control Test	0.079	13:58	
Air Blank	0.000	13:58	
Subject Sample	# 1	14:01	No Sample Provided
Breath Volume	0.000	14:01	
Air Blank	0.000	14:02	
Air Blank	0.000	14:04	
Subject Sample	# 2	14:04	
Breath Volume	1.652	14:04	
Air Blank	0.000	14:05	
Air Blank	0.000	14:07	
Subject Sample	# 3	14:07	RFI Detect
Breath Volume	0.000	14:07	
Air Blank	0.000	14:08	
Control Test	0.078	14:08	
Air Blank	0.000	14:09	
Diagnostics Check	OK	14:09	

---

Data Download Date/Time: 06/06/2018 09:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, YOEL,  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: GARCIA, YOEL  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 06/02/2018 21:09:47

---

Diagnostics Check	OK	21:03
Air Blank	0.000	21:04
Control Test	0.079	21:04
Air Blank	0.000	21:05
Subject Sample # 1	0.000	21:05
Breath Volume	3.250	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:08
Subject Sample # 2	0.000	21:08
Breath Volume	3.964	21:08
Air Blank	0.000	21:08
Control Test	0.079	21:09
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

---

Data Download Date/Time: 06/06/2018 09:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALL,CHRISTINE,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HALL,CHRISTINE  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: CISNEROS BELMARES

DATE/TIME OF TEST: 06/03/2018 00:38:20

---

Diagnostics Check	OK	00:31
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.227	00:33
Breath Volume	2.445	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.213	00:36
Breath Volume	1.589	00:36
Air Blank	0.000	00:37
Control Test	0.078	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

---

Data Download Date/Time: 06/06/2018 09:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALL,CHRISTINE,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LISIK,SARAH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: GRELA MARTINEZ

DATE/TIME OF TEST: 06/03/2018 02:11:29

---

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.078	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.183	02:07
Breath Volume	1.320	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.192	02:09
Breath Volume	1.382	02:09
Air Blank	0.000	02:10
Control Test	0.078	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

---

Data Download Date/Time: 06/06/2018 09:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHILKE,CARL,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SCHILKE,CARL  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: MARIETTA

DATE/TIME OF TEST: 06/04/2018 18:15:24

---

Diagnostics Check	OK	18:08
Air Blank	0.000	18:09
Control Test	0.079	18:09
Air Blank	0.000	18:10
Subject Sample # 1	0.147	18:10
Breath Volume	3.339	18:10
Air Blank	0.000	18:11
Air Blank	0.000	18:13
Subject Sample # 2	0.146	18:13
Breath Volume	3.660	18:13
Air Blank	0.000	18:14
Control Test	0.079	18:14
Air Blank	0.000	18:15
Diagnostics Check	OK	18:15

---

Data Download Date/Time: 06/06/2018 09:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHILKE,CARL,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SCHILKE,CARL  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: BRISTOL

DATE/TIME OF TEST: 06/05/2018 00:04:50

---

Diagnostics Check	OK	23:57
Air Blank	0.000	23:58
Control Test	0.079	23:58
Air Blank	0.000	23:59
Subject Sample # 2	0.160	23:59
Breath Volume	4.222	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.163	00:03
Breath Volume	4.769	00:03
Air Blank	0.000	00:03
Control Test	0.078	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

---

Data Download Date/Time: 06/06/2018 09:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANPELT,JAMES,J  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LINDHOLM,J  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: FELTON

DATE/TIME OF TEST: 06/06/2018 04:01:15

---

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.132	03:56
Breath Volume	2.320	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:58
Subject Sample # 2	0.120	03:59
Breath Volume	1.519	03:59
Air Blank	0.000	04:00
Control Test	0.078	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

---

Data Download Date/Time: 06/06/2018 09:11



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURLEY,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HARRISON,W  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 14:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:08

SUBJECT'S LAST NAME: HAUSS

DATE/TIME OF TEST: 06/06/2018 16:56:27

---

Diagnostics Check	OK	16:44	
Air Blank	0.000	16:45	
Control Test	0.079	16:45	
Air Blank	0.000	16:45	
Subject Sample	# 1	16:48	No Sample Provided
Breath Volume	0.000	16:48	
Air Blank	0.000	16:49	
Air Blank	0.000	16:51	
Subject Sample	# 2	0.196	16:51
Breath Volume	3.238	16:51	
Air Blank	0.000	16:52	
Air Blank	0.000	16:54	
Subject Sample	# 3	0.197	16:54
Breath Volume	4.226	16:54	
Air Blank	0.000	16:55	
Control Test	0.079	16:55	
Air Blank	0.000	16:56	
Diagnostics Check	OK	16:56	

---

Data Download Date/Time: 07/18/2018 08:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,JOSIAH,R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SMITH,JOSIAH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: NUNEZ CARRASQUILLA

DATE/TIME OF TEST: 06/06/2018 23:47:13

---

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.079	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.206	23:42
Breath Volume	3.308	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:44
Subject Sample # 2	0.199	23:45
Breath Volume	3.667	23:45
Air Blank	0.000	23:46
Control Test	0.078	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

---

Data Download Date/Time: 07/18/2018 08:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STROBEL, RAYMOND, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: STROBEL, RAYMOND  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 09:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: OTTING

DATE/TIME OF TEST: 06/08/2018 10:55:38

---

Diagnostics Check	OK	10:48
Air Blank	0.000	10:49
Control Test	0.079	10:49
Air Blank	0.000	10:50
Subject Sample # 1	0.000	10:50
Breath Volume	1.640	10:50
Air Blank	0.000	10:51
Air Blank	0.000	10:53
Subject Sample # 2	0.000	10:54
Breath Volume	2.000	10:54
Air Blank	0.000	10:54
Control Test	0.079	10:54
Air Blank	0.000	10:55
Diagnostics Check	OK	10:55

---

Data Download Date/Time: 07/18/2018 08:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEYERS,F,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CAIL,M  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 06/09/2018 03:47:39

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.124	03:42
Breath Volume	3.136	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.123	03:46
Breath Volume	3.765	03:46
Air Blank	0.000	03:46
Control Test	0.078	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

---

Data Download Date/Time: 07/18/2018 08:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANPELT,J,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: TORRES,S  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: ROGER

DATE/TIME OF TEST: 06/10/2018 05:27:54

---

Diagnostics Check	OK	05:21
Air Blank	0.000	05:21
Control Test	0.079	05:22
Air Blank	0.000	05:22
Subject Sample # 1	0.041	05:23
Breath Volume	1.742	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:25
Subject Sample # 2	0.044	05:26
Breath Volume	2.730	05:26
Air Blank	0.000	05:26
Control Test	0.079	05:27
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

---

Data Download Date/Time: 07/18/2018 08:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHILKE,CARL,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BRODBECK,JAMES  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: BENAVIDES

DATE/TIME OF TEST: 06/10/2018 20:54:59

---

Diagnostics Check	OK	20:48
Air Blank	0.000	20:49
Control Test	0.079	20:49
Air Blank	0.000	20:50
Subject Sample # 1	0.105	20:50
Breath Volume	5.574	20:50
Air Blank	0.000	20:51
Air Blank	0.000	20:52
Subject Sample # 2	0.113	20:53
Breath Volume	3.476	20:53
Air Blank	0.000	20:54
Control Test	0.078	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

---

Data Download Date/Time: 07/18/2018 08:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSEPH,G  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: RYAN,MATTHEW  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1081080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: KOENIG

DATE/TIME OF TEST: 06/13/2018 02:20:41

---

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.079	02:14
Air Blank		0.000	02:14
Subject Sample	# 1	0.278	02:15
Breath Volume		1.601	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.260	02:19
Breath Volume		1.683	02:19
Air Blank		0.000	02:19
Control Test		0.079	02:20
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

---

Data Download Date/Time: 07/18/2018 08:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,JOSIAH,R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SMITH,JOSIAH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1081080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 06/13/2018 02:29:56

---

Diagnostics Check		OK	02:23
Air Blank		0.000	02:24
Control Test		0.078	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.108	02:25
Breath Volume		1.574	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:27
Subject Sample	# 2	0.106	02:28
Breath Volume		1.199	02:28
Air Blank		0.000	02:28
Control Test		0.079	02:29
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

---

Data Download Date/Time: 07/18/2018 08:54



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEYERS,F,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GREGORY,T  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1081080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: CORDERO DAVILA

DATE/TIME OF TEST: 06/14/2018 05:10:50

---

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.080	05:05
Air Blank	0.000	05:05
Subject Sample # 1	0.126	05:06
Breath Volume	3.480	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:08
Subject Sample # 2	0.125	05:09
Breath Volume	1.785	05:09
Air Blank	0.000	05:09
Control Test	0.078	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

---

Data Download Date/Time: 07/18/2018 08:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BONSALL,MATTHEW,A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BONSALL,MATTHEW  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1081080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 06/16/2018 00:36:21

---

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.080	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.112	00:31
Breath Volume	2.605	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.106	00:34
Breath Volume	2.210	00:34
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

---

Data Download Date/Time: 07/18/2018 08:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSEPH,G  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: KEEN,JULIAN  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1081080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 16:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: STODDARD

DATE/TIME OF TEST: 06/17/2018 20:46:42

---

Diagnostics Check	OK	20:40
Air Blank	0.000	20:40
Control Test	0.080	20:40
Air Blank	0.000	20:41
Subject Sample # 1	0.137	20:42
Breath Volume	2.609	20:42
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample # 2	0.134	20:45
Breath Volume	2.152	20:45
Air Blank	0.000	20:45
Control Test	0.079	20:46
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

---

Data Download Date/Time: 07/18/2018 08:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,JOSIAH,R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SMITH,JOSIAH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1081080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: LANGFORD

DATE/TIME OF TEST: 06/17/2018 21:34:56

---

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.079	21:29
Air Blank	0.000	21:29
Subject Sample # 1	0.000	21:30
Breath Volume	2.457	21:30
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.000	21:33
Breath Volume	2.667	21:33
Air Blank	0.000	21:34
Control Test	0.079	21:34
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

---

Data Download Date/Time: 07/18/2018 08:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSEPH,G  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SELLERS, TIFFANY  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1081080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: GRILL

DATE/TIME OF TEST: 06/21/2018 01:45:49

---

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.127	01:41
Breath Volume	1.937	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.124	01:44
Breath Volume	1.750	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

---

Data Download Date/Time: 07/18/2018 08:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,JOSIAH,R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SMITH,JOSIAH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1081080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 06/21/2018 01:57:25

---

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.130	01:52
Breath Volume	2.976	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.127	01:55
Breath Volume	2.917	01:55
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

---

Data Download Date/Time: 07/18/2018 08:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WITT,KELLY,  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: WITT,KELLY  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1081080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: FAUSTINO

DATE/TIME OF TEST: 06/22/2018 20:19:41

---

Diagnostics Check	OK	20:12
Air Blank	0.000	20:13
Control Test	0.080	20:13
Air Blank	0.000	20:14
Subject Sample # 1	0.202	20:15
Breath Volume	2.132	20:15
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.199	20:18
Breath Volume	1.875	20:18
Air Blank	0.000	20:18
Control Test	0.078	20:19
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

---

Data Download Date/Time: 07/18/2018 08:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLS,CHARLES,S  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: HIDEG,JARED  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1081080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: CANTRELL

DATE/TIME OF TEST: 06/23/2018 21:12:25

---

Diagnostics Check	OK	21:03
Air Blank	0.000	21:03
Control Test	0.079	21:03
Air Blank	0.000	21:04
Subject Sample # 1	0.108	21:05
Breath Volume	1.945	21:05
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.111	21:10
Breath Volume	2.039	21:10
Air Blank	0.000	21:11
Control Test	0.079	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

---

Data Download Date/Time: 07/18/2018 08:54



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WITT,KELLY,  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: WITT,KELLY  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1081080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: SMALL

DATE/TIME OF TEST: 06/23/2018 22:51:23

---

Diagnostics Check	OK	22:44
Air Blank	0.000	22:45
Control Test	0.078	22:45
Air Blank	0.000	22:46
Subject Sample # 1	0.153	22:46
Breath Volume	3.746	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:48
Subject Sample # 2	0.154	22:49
Breath Volume	3.359	22:49
Air Blank	0.000	22:50
Control Test	0.078	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

---

Data Download Date/Time: 07/18/2018 08:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PERRELLA,STEPHEN,L  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: CAPOTE,DUVAN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1081080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 09:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: PERRELLA

DATE/TIME OF TEST: 06/24/2018 02:24:54

---

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.194	02:20
Breath Volume	1.718	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample # 2	0.185	02:23
Breath Volume	2.464	02:23
Air Blank	0.000	02:23
Control Test	0.078	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

---

Data Download Date/Time: 07/18/2018 08:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPOTE,DUVAN,  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: HIDEG,JARED  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1081080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/25/2018 04:04:32

---

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.111	03:59
Breath Volume	1.578	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.109	04:02
Breath Volume	1.382	04:02
Air Blank	0.000	04:03
Control Test	0.077	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

---

Data Download Date/Time: 07/18/2018 08:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPOTE,DUVAN,  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: CAPOTE,DUVAN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1081080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 09:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: HAGAN

DATE/TIME OF TEST: 06/25/2018 10:39:01

---

Diagnostics Check	OK	10:32
Air Blank	0.000	10:33
Control Test	0.079	10:33
Air Blank	0.000	10:34
Subject Sample # 1	0.000	10:34
Breath Volume	2.320	10:34
Air Blank	0.000	10:35
Air Blank	0.000	10:37
Subject Sample # 2	0.000	10:37
Breath Volume	1.984	10:37
Air Blank	0.000	10:38
Control Test	0.079	10:38
Air Blank	0.000	10:38
Diagnostics Check	OK	10:39

---

Data Download Date/Time: 07/18/2018 08:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BONSALL,MATTHEW,A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BONSALL,MATTHEW  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1081080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: DRAVES

DATE/TIME OF TEST: 06/25/2018 18:34:13

---

Diagnostics Check	OK	18:24
Air Blank	0.000	18:24
Control Test	0.079	18:25
Air Blank	0.000	18:25
Subject Sample # 1	0.231	18:26
Breath Volume	3.625	18:26
Air Blank	0.000	18:27
Air Blank	0.000	18:28
Subject Sample # 2	0.204	18:29
Breath Volume	1.847	18:29
Air Blank	0.000	18:30
Air Blank	0.000	18:31
Subject Sample # 3	0.229	18:32
Breath Volume	2.648	18:32
Air Blank	0.000	18:33
Control Test	0.078	18:33
Air Blank	0.000	18:34
Diagnostics Check	OK	18:34

---

Data Download Date/Time: 07/18/2018 08:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,JOSIAH,R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SMITH,JOSIAH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1081080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: SCHUMACHER

DATE/TIME OF TEST: 06/26/2018 20:35:06

---

Diagnostics Check	OK	20:28
Air Blank	0.000	20:28
Control Test	0.079	20:28
Air Blank	0.000	20:29
Subject Sample # 1	0.000	20:30
Breath Volume	2.406	20:30
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample # 2	0.000	20:33
Breath Volume	2.167	20:33
Air Blank	0.000	20:34
Control Test	0.080	20:34
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

---

Data Download Date/Time: 07/18/2018 08:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GLEASON, WILLIAM, L  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: GLEASON, WILLIAM  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 06/15/2018 07:59:04

---

Diagnostics Check	OK	07:52
Air Blank	0.000	07:53
Control Test	0.078	07:53
Air Blank	0.000	07:54
Subject Sample # 1	0.000	07:54
Breath Volume	2.328	07:54
Air Blank	0.000	07:55
Air Blank	0.000	07:57
Subject Sample # 2	0.000	07:57
Breath Volume	2.554	07:57
Air Blank	0.000	07:58
Control Test	0.079	07:58
Air Blank	0.000	07:58
Diagnostics Check	OK	07:59

---

Data Download Date/Time: 07/26/2018 20:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GLEASON, WILLIAM, L  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: HEBB, PAUL  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: HAZELL

DATE/TIME OF TEST: 06/15/2018 13:25:21

---

Diagnostics Check	OK	13:19
Air Blank	0.000	13:19
Control Test	0.079	13:19
Air Blank	0.000	13:20
Subject Sample # 1	0.000	13:20
Breath Volume	1.171	13:20
Air Blank	0.000	13:21
Air Blank	0.000	13:23
Subject Sample # 2	0.000	13:24
Breath Volume	2.312	13:24
Air Blank	0.000	13:24
Control Test	0.079	13:24
Air Blank	0.000	13:25
Diagnostics Check	OK	13:25

---

Data Download Date/Time: 07/26/2018 20:00



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER,M,H  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: BAKER,M  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 11:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:14

SUBJECT'S LAST NAME: HOPPES

DATE/TIME OF TEST: 06/17/2018 12:34:16

---

Diagnostics Check	OK	12:27
Air Blank	0.000	12:28
Control Test	0.080	12:28
Air Blank	0.000	12:29
Subject Sample # 1	0.000	12:29
Breath Volume	3.449	12:29
Air Blank	0.000	12:30
Air Blank	0.000	12:32
Subject Sample # 2	0.000	12:32
Breath Volume	3.253	12:32
Air Blank	0.000	12:33
Control Test	0.079	12:33
Air Blank	0.000	12:34
Diagnostics Check	OK	12:34

---

Data Download Date/Time: 07/26/2018 20:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: WINSTEAD,JUSTIN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/21/2018 23:12:10

---

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.000	23:07
Breath Volume	3.218	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.000	23:10
Breath Volume	2.984	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

---

Data Download Date/Time: 07/26/2018 20:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,NICHOLAS,W  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: MATTHEWS,G  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: SANCHEZ-LARA

DATE/TIME OF TEST: 06/26/2018 04:27:33

---

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:23	
Control Test	0.079	04:23	
Air Blank	0.000	04:23	
Subject Sample	# 1 0.223	04:25	
Breath Volume	1.300	04:25	
Air Blank		04:26	Purge Fail
Air Blank	0.000	04:27	

---

Data Download Date/Time: 07/26/2018 20:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,NICHOLAS,W  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: MATTHEWS,G  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SANCHEZ-LARA

DATE/TIME OF TEST: 06/26/2018 04:35:38

---

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.077	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.212	04:30
Breath Volume	2.164	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.204	04:34
Breath Volume	1.695	04:34
Air Blank	0.000	04:34
Control Test	0.078	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

---

Data Download Date/Time: 07/26/2018 20:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWELL,NATHANAEL,A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: JENKINS,JOSEPH  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: HIATT

DATE/TIME OF TEST: 06/28/2018 03:19:24

---

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.078	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.301	03:14
Breath Volume	3.222	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.300	03:17
Breath Volume	1.761	03:17
Air Blank	0.000	03:18
Control Test	0.077	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

---

Data Download Date/Time: 07/26/2018 20:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRISCOLL,JAMES,P  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: DRISCOLL,JAMES  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: SANTIAGO LOPEZ

DATE/TIME OF TEST: 06/08/2018 01:24:40

---

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.079	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.173	01:20
Breath Volume	1.722	01:20
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.171	01:23
Breath Volume	2.011	01:23
Air Blank	0.000	01:23
Control Test	0.078	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

---

Data Download Date/Time: 07/02/2018 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OLMSTEAD,J,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: DERRIG,J  
ARRESTING OFFICER AGENCY: Marco Island PD  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 06/08/2018 01:36:57

---

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.080	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.159	01:32
Breath Volume	1.917	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:34
Subject Sample # 2	0.161	01:35
Breath Volume	1.179	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

---

Data Download Date/Time: 07/02/2018 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUTIERREZ,I,T  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: RODRIGEZ,A  
ARRESTING OFFICER AGENCY: Naples PD  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: DILICH

DATE/TIME OF TEST: 06/08/2018 04:49:18

---

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.079	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.199	04:44
Breath Volume	1.988	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.193	04:47
Breath Volume	1.312	04:47
Air Blank	0.000	04:48
Control Test	0.079	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

---

Data Download Date/Time: 07/02/2018 15:29



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ASHLINE,C,J  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: DEAR,D  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: JOLLY

DATE/TIME OF TEST: 06/08/2018 23:22:34

---

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.080	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.078	23:18
Breath Volume	1.578	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.076	23:21
Breath Volume	2.117	23:21
Air Blank	0.000	23:21
Control Test	0.079	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

---

Data Download Date/Time: 07/02/2018 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA,A,N  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: DOMINGUEZ,J  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 14:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: TRANA

DATE/TIME OF TEST: 06/09/2018 16:12:09

---

Diagnostics Check	OK	16:05
Air Blank	0.000	16:05
Control Test	0.080	16:05
Air Blank	0.000	16:06
Subject Sample # 1	0.181	16:06
Breath Volume	2.097	16:06
Air Blank	0.000	16:07
Air Blank	0.000	16:09
Subject Sample # 2	0.179	16:10
Breath Volume	1.664	16:10
Air Blank	0.000	16:11
Control Test	0.079	16:11
Air Blank	0.000	16:12
Diagnostics Check	OK	16:12

---

Data Download Date/Time: 07/02/2018 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORR,C,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: MARTIN,J  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 05:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:19

SUBJECT'S LAST NAME: MENDOZA PEREZ

DATE/TIME OF TEST: 06/10/2018 08:08:16

---

Diagnostics Check		OK	07:54	
Air Blank		0.000	07:54	
Control Test		0.080	07:55	
Air Blank		0.000	07:55	
Subject Sample	# 1	0.101	07:58	Volume Not Met
Breath Volume		0.175	07:58	
Air Blank		0.000	07:59	
Air Blank		0.000	08:01	
Subject Sample	# 2	0.097	08:03	
Breath Volume		2.320	08:03	
Air Blank		0.000	08:04	
Air Blank		0.000	08:06	
Subject Sample	# 3	0.107	08:06	
Breath Volume		1.636	08:06	
Air Blank		0.000	08:07	
Control Test		0.077	08:07	
Air Blank		0.000	08:08	
Diagnostics Check		OK	08:08	

---

Data Download Date/Time: 07/02/2018 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALES,Y,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: DOMINGUEZ,J  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 06/10/2018 19:36:44

---

Diagnostics Check	OK	19:29
Air Blank	0.000	19:30
Control Test	0.079	19:30
Air Blank	0.000	19:31
Subject Sample # 1	0.185	19:31
Breath Volume	1.597	19:31
Air Blank	0.000	19:32
Air Blank	0.000	19:33
Subject Sample # 2	0.173	19:35
Breath Volume	1.843	19:35
Air Blank	0.000	19:35
Control Test	0.080	19:36
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

---

Data Download Date/Time: 07/02/2018 15:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLOPS,COLTON,M  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: MCGILVREY,J  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BALLOU

DATE/TIME OF TEST: 06/10/2018 20:09:07

---

Diagnostics Check	OK	20:03
Air Blank	0.000	20:03
Control Test	0.080	20:03
Air Blank	0.000	20:04
Subject Sample # 1	0.106	20:04
Breath Volume	2.589	20:04
Air Blank	0.000	20:05
Air Blank	0.000	20:07
Subject Sample # 2	0.106	20:07
Breath Volume	2.277	20:07
Air Blank	0.000	20:08
Control Test	0.079	20:08
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

---

Data Download Date/Time: 07/02/2018 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUGO,JONATHAN,B  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: STEVENSON,C  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GRIFFITHS

DATE/TIME OF TEST: 06/11/2018 04:18:02

---

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.155	04:11
Breath Volume	1.289	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.160	04:16
Breath Volume	2.089	04:16
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:18

---

Data Download Date/Time: 07/02/2018 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CURTSINGER,C,A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: WILLIAMS,B  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: FRIAS

DATE/TIME OF TEST: 06/11/2018 19:08:37

---

Diagnostics Check		OK	18:56
Air Blank		0.000	18:57
Control Test		0.079	18:57
Air Blank		0.000	18:57
Subject Sample	# 1	0.224	18:59
Breath Volume		1.527	18:59
Air Blank		0.000	19:00
Air Blank		0.000	19:02
Subject Sample	# 2	0.252	19:02
Breath Volume		1.628	19:02
Air Blank		0.000	19:03
Air Blank		0.000	19:05
Subject Sample	# 3	0.244	19:06
Breath Volume		1.304	19:06
Air Blank		0.000	19:07
Control Test		0.078	19:08
Air Blank		0.000	19:08
Diagnostics Check		OK	19:08

---

Data Download Date/Time: 07/02/2018 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTO,SAMUEL,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: DISARRO,B  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 14:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: PAPPAS

DATE/TIME OF TEST: 06/12/2018 16:01:39

---

Diagnostics Check	OK	15:55
Air Blank	0.000	15:55
Control Test	0.079	15:56
Air Blank	0.000	15:56
Subject Sample # 1	0.171	15:57
Breath Volume	1.597	15:57
Air Blank	0.000	15:57
Air Blank	0.000	15:59
Subject Sample # 2	0.166	16:00
Breath Volume	1.652	16:00
Air Blank	0.000	16:00
Control Test	0.080	16:01
Air Blank	0.000	16:01
Diagnostics Check	OK	16:01

---

Data Download Date/Time: 07/02/2018 15:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KENNEDY,DANIEL  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: SAKAMOTO

DATE/TIME OF TEST: 06/13/2018 00:16:45

---

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:05	
Control Test		0.080	00:05	
Air Blank		0.000	00:06	
Subject Sample	# 1	0.110	00:09	Volume Not Met
Breath Volume		0.851	00:09	
Air Blank		0.000	00:10	
Air Blank		0.000	00:12	
Subject Sample	# 2	0.117	00:15	Volume Not Met
Breath Volume		0.195	00:15	
Air Blank		0.000	00:15	
Control Test		0.079	00:16	
Air Blank		0.000	00:16	
Diagnostics Check		OK	00:16	

---

Data Download Date/Time: 07/02/2018 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KENNEDY,DANIEL  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: SAKAMOTO

DATE/TIME OF TEST: 06/13/2018 00:27:30

---

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.079	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.122	00:22
Breath Volume		1.625	00:22
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.120	00:25
Breath Volume		1.203	00:25
Air Blank		0.000	00:26
Control Test		0.080	00:26
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

---

Data Download Date/Time: 07/02/2018 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKS,D,M  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: STUDZINSKI,K  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: LAWRENSON

DATE/TIME OF TEST: 06/13/2018 02:21:21

---

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.243	02:16
Breath Volume	2.394	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.239	02:19
Breath Volume	2.085	02:19
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

---

Data Download Date/Time: 07/02/2018 15:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVAS,W,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: JOHNSON,D  
ARRESTING OFFICER AGENCY: Marco Island PD  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: NICANOR CAAL

DATE/TIME OF TEST: 06/14/2018 20:05:41

---

Diagnostics Check		OK	19:58
Air Blank		0.000	19:59
Control Test		0.080	19:59
Air Blank		0.000	20:00
Subject Sample	# 1	0.250	20:00
Breath Volume		1.464	20:00
Air Blank		0.000	20:01
Air Blank		0.000	20:03
Subject Sample	# 2	0.248	20:04
Breath Volume		1.492	20:04
Air Blank		0.000	20:04
Control Test		0.080	20:05
Air Blank		0.000	20:05
Diagnostics Check		OK	20:05

---

Data Download Date/Time: 07/02/2018 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PLUSA,KYLE,A  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: PLUSA,KYLE  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: BUI  
ARREST TIME: 17:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: CSOGI

DATE/TIME OF TEST: 06/15/2018 18:23:57

---

Diagnostics Check	OK	18:21	
Air Blank	0.000	18:21	
Control Test	0.080	18:21	
Air Blank	0.000	18:22	
Subject Sample	# 1	18:22	Subject Test Refused
Breath Volume	0.000	18:22	
Air Blank	0.000	18:23	
Control Test	0.080	18:23	
Air Blank	0.000	18:23	
Diagnostics Check	OK	18:23	

---

Data Download Date/Time: 07/02/2018 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROGERS, MATTHEW, L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CRUZ, I  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: CORNELIUS

DATE/TIME OF TEST: 06/15/2018 22:11:02

---

Diagnostics Check	OK	22:04
Air Blank	0.000	22:05
Control Test	0.080	22:05
Air Blank	0.000	22:06
Subject Sample # 1	0.175	22:06
Breath Volume	3.308	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:08
Subject Sample # 2	0.182	22:09
Breath Volume	2.226	22:09
Air Blank	0.000	22:10
Control Test	0.079	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:11

---

Data Download Date/Time: 07/02/2018 15:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELIAS,D,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: WOODS,H  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: CAMACHO

DATE/TIME OF TEST: 06/16/2018 01:03:13

---

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.080	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.161	00:58
Breath Volume	1.917	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:00
Subject Sample # 2	0.158	01:01
Breath Volume	1.792	01:01
Air Blank	0.000	01:02
Control Test	0.079	01:02
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

---

Data Download Date/Time: 07/02/2018 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOODMAN,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KITCHENHOFF,M.  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HASS

DATE/TIME OF TEST: 06/16/2018 02:13:34

---

Diagnostics Check		OK	02:02	
Air Blank		0.000	02:02	
Control Test		0.080	02:02	
Air Blank		0.000	02:03	
Subject Sample	# 1	0.116	02:06	Volume Not Met
Breath Volume		0.136	02:06	
Air Blank		0.000	02:07	
Air Blank		0.000	02:08	
Subject Sample	# 2	0.126	02:12	Volume Not Met
Breath Volume		0.105	02:12	
Air Blank		0.000	02:12	
Control Test		0.079	02:12	
Air Blank		0.000	02:13	
Diagnostics Check		OK	02:13	

---

Data Download Date/Time: 07/02/2018 15:29



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOODMAN,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KITCHENHOFF,M.  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HASS

DATE/TIME OF TEST: 06/16/2018 02:28:38

---

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:17	
Control Test		0.079	02:17	
Air Blank		0.000	02:18	
Subject Sample	# 1	0.130	02:21	Volume Not Met
Breath Volume		0.035	02:21	
Air Blank		0.000	02:22	
Air Blank		0.000	02:23	
Subject Sample	# 2	0.136	02:27	Volume Not Met
Breath Volume		0.066	02:27	
Air Blank		0.000	02:27	
Control Test		0.079	02:28	
Air Blank		0.000	02:28	
Diagnostics Check		OK	02:28	

---

Data Download Date/Time: 07/02/2018 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOODMAN,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KITCHENHOFF,M.  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HASS

DATE/TIME OF TEST: 06/16/2018 02:45:40

---

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:30	
Control Test		0.079	02:31	
Air Blank		0.000	02:31	
Subject Sample	# 1	0.127	02:34	Volume Not Met
Breath Volume		0.660	02:34	
Air Blank		0.000	02:35	
Air Blank		0.000	02:37	
Subject Sample	# 2	0.129	02:38	
Breath Volume		1.261	02:38	
Air Blank		0.000	02:39	
Air Blank		0.000	02:40	
Subject Sample	# 3	0.129	02:44	Volume Not Met
Breath Volume		0.730	02:44	
Air Blank		0.000	02:44	
Control Test		0.079	02:45	
Air Blank		0.000	02:45	
Diagnostics Check		OK	02:45	

---

Data Download Date/Time: 07/02/2018 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOODMAN,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KITCHENHOFF,M.  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HASS

DATE/TIME OF TEST: 06/16/2018 03:04:02

---

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:50	
Control Test	0.079	02:50	
Air Blank	0.000	02:51	
Subject Sample	# 1 0.115	02:52	
Breath Volume	1.140	02:52	
Air Blank	0.000	02:53	
Air Blank	0.000	02:55	
Subject Sample	# 2 0.110	02:58	Volume Not Met
Breath Volume	1.082	02:58	
Air Blank	0.000	02:59	
Air Blank	0.000	03:01	
Subject Sample	# 3 0.123	03:02	
Breath Volume	1.191	03:02	
Air Blank	0.000	03:03	
Control Test	0.079	03:03	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:04	

---

Data Download Date/Time: 07/02/2018 15:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANNORAT,ADELY,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: RODRIGUEZ,A  
ARRESTING OFFICER AGENCY: Naples PD  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 06/16/2018 03:35:53

---

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.079	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.099	03:31
Breath Volume	2.269	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:33
Subject Sample # 2	0.086	03:34
Breath Volume	2.820	03:34
Air Blank	0.000	03:34
Control Test	0.079	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

---

Data Download Date/Time: 07/02/2018 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMPOLO,J,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CAMPOLO,J  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: RISSLER

DATE/TIME OF TEST: 06/18/2018 00:57:21

---

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.080	00:52
Air Blank		0.000	00:52
Subject Sample	# 1	0.179	00:52
Breath Volume		1.105	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.181	00:55
Breath Volume		1.722	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

---

Data Download Date/Time: 07/02/2018 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOGAN,T,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: MILLER,K  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: GAVIRIA

DATE/TIME OF TEST: 06/20/2018 23:38:08

---

Diagnostics Check	OK	23:31
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.169	23:33
Breath Volume	3.113	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.170	23:36
Breath Volume	3.882	23:36
Air Blank	0.000	23:37
Control Test	0.080	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

---

Data Download Date/Time: 07/02/2018 15:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KENNEDY,DANIEL  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: SEISS

DATE/TIME OF TEST: 06/21/2018 23:31:28

---

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.210	23:26
Breath Volume	2.542	23:26
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.208	23:29
Breath Volume	2.222	23:29
Air Blank	0.000	23:30
Control Test	0.079	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

---

Data Download Date/Time: 07/02/2018 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,H,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: FANO,D  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: PEABODY

DATE/TIME OF TEST: 06/22/2018 17:25:21

---

Diagnostics Check	OK	17:19
Air Blank	0.000	17:19
Control Test	0.080	17:19
Air Blank	0.000	17:20
Subject Sample # 1	0.257	17:20
Breath Volume	2.039	17:20
Air Blank	0.000	17:21
Air Blank	0.000	17:23
Subject Sample # 2	0.248	17:23
Breath Volume	1.964	17:23
Air Blank	0.000	17:24
Control Test	0.079	17:24
Air Blank	0.000	17:25
Diagnostics Check	OK	17:25

---

Data Download Date/Time: 07/02/2018 15:30



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOEL,YVES,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: SCHLICHTE,J  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: JEFFERIES

DATE/TIME OF TEST: 06/23/2018 02:35:29

---

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.080	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.260	02:31
Breath Volume		1.683	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.269	02:33
Breath Volume		1.613	02:33
Air Blank		0.000	02:34
Control Test		0.079	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

---

Data Download Date/Time: 07/02/2018 15:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BATSON,MITCHELL,J  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: ZERPA,E  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: URBAN

DATE/TIME OF TEST: 06/23/2018 03:56:25

---

Diagnostics Check		OK	03:41	
Air Blank		0.000	03:42	
Control Test		0.081	03:42	
Air Blank		0.000	03:43	
Subject Sample	# 1	0.239	03:46	Volume Not Met
Breath Volume		0.289	03:46	
Air Blank		0.000	03:46	
Air Blank		0.000	03:48	
Subject Sample	# 2	0.238	03:49	
Breath Volume		1.792	03:49	
Air Blank		0.000	03:49	
Air Blank		0.000	03:51	
Subject Sample	# 3	0.215	03:54	Volume Not Met
Breath Volume		0.218	03:54	
Air Blank		0.000	03:55	
Control Test		0.079	03:55	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

---

Data Download Date/Time: 07/02/2018 15:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,ZACHARY,H  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: ZERPA,E  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: LAWLER

DATE/TIME OF TEST: 06/23/2018 19:30:29

---

Diagnostics Check		OK	19:18
Air Blank		0.000	19:18
Control Test		0.080	19:19
Air Blank		0.000	19:19
Subject Sample	# 1	0.288	19:22
Breath Volume		2.175	19:22
Air Blank		0.000	19:22
Air Blank		0.000	19:24
Subject Sample	# 2	0.318	19:24
Breath Volume		1.769	19:24
Air Blank		0.000	19:25
Air Blank		0.000	19:27
Subject Sample	# 3	0.310	19:28
Breath Volume		2.757	19:28
Air Blank		0.000	19:29
Control Test		0.077	19:29
Air Blank		0.000	19:30
Diagnostics Check		OK	19:30

---

Data Download Date/Time: 07/02/2018 15:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALES,Y,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: REED,  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: FOURNIER ORAMA

DATE/TIME OF TEST: 06/23/2018 22:43:15

---

Diagnostics Check	OK	22:36
Air Blank	0.000	22:37
Control Test	0.080	22:37
Air Blank	0.000	22:38
Subject Sample # 1	0.092	22:38
Breath Volume	2.945	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.083	22:41
Breath Volume	2.664	22:41
Air Blank	0.000	22:42
Control Test	0.080	22:42
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

---

Data Download Date/Time: 07/02/2018 15:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUONPANE,VINCENT,D  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: WOODS,H  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: WIND

DATE/TIME OF TEST: 06/25/2018 22:02:08

---

Diagnostics Check	OK	21:53	
Air Blank	0.000	21:54	
Control Test	0.080	21:54	
Air Blank	0.000	21:55	
Subject Sample	# 1 0.069	21:58	Volume Not Met
Breath Volume	0.292	21:58	
Air Blank	0.000	21:58	
Air Blank	0.000	22:00	
Subject Sample	# 2	22:00	Subject Test Refused
Breath Volume	0.000	22:00	
Air Blank	0.000	22:01	
Control Test	0.080	22:01	
Air Blank	0.000	22:02	
Diagnostics Check	OK	22:02	

---

Data Download Date/Time: 07/02/2018 15:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARP,K,M  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: BENOIT,J  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 05:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:17

SUBJECT'S LAST NAME: ESTANQUE

DATE/TIME OF TEST: 06/27/2018 07:50:59

---

Diagnostics Check	OK	07:44
Air Blank	0.000	07:45
Control Test	0.080	07:45
Air Blank	0.000	07:46
Subject Sample # 1	0.000	07:46
Breath Volume	2.105	07:46
Air Blank	0.000	07:47
Air Blank	0.000	07:49
Subject Sample # 2	0.000	07:49
Breath Volume	2.613	07:49
Air Blank	0.000	07:50
Control Test	0.080	07:50
Air Blank	0.000	07:50
Diagnostics Check	OK	07:50

---

Data Download Date/Time: 07/02/2018 15:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KENNEDY,DANIEL  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: TRELLES LEMES

DATE/TIME OF TEST: 06/30/2018 19:51:40

---

Diagnostics Check	OK	19:45
Air Blank	0.000	19:45
Control Test	0.080	19:45
Air Blank	0.000	19:46
Subject Sample # 1	0.251	19:47
Breath Volume	2.871	19:47
Air Blank	0.000	19:47
Air Blank	0.000	19:49
Subject Sample # 2	0.252	19:50
Breath Volume	3.953	19:50
Air Blank	0.000	19:50
Control Test	0.079	19:51
Air Blank	0.000	19:51
Diagnostics Check	OK	19:51

---

Data Download Date/Time: 07/02/2018 15:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OLMSTEAD, JEREMY,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: ARBOGAST, J  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: BUI  
ARREST TIME: 19:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: PRYOR

DATE/TIME OF TEST: 06/30/2018 22:02:17

---

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.080	21:55
Air Blank	0.000	21:56
Subject Sample # 1	0.145	21:56
Breath Volume	1.476	21:56
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.158	22:00
Breath Volume	3.074	22:00
Air Blank	0.000	22:01
Control Test	0.079	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

---

Data Download Date/Time: 07/02/2018 15:30



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KENNEDY,DANIEL  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: BRAND

DATE/TIME OF TEST: 06/01/2018 23:14:58

---

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:03	
Control Test	0.079	23:03	
Air Blank	0.000	23:04	
Subject Sample	# 1	23:07	No Sample Provided
Breath Volume	0.000	23:07	
Air Blank	0.000	23:07	
Air Blank	0.000	23:09	
Subject Sample	# 2	0.239	23:10
Breath Volume	2.382	23:10	
Air Blank	0.000	23:11	
Air Blank	0.000	23:12	
Subject Sample	# 3	0.235	23:13
Breath Volume	2.734	23:13	
Air Blank	0.000	23:14	
Control Test	0.080	23:14	
Air Blank	0.000	23:14	
Diagnostics Check	OK	23:14	

---

Data Download Date/Time: 06/05/2018 17:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUONPANE,V,D  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: PEREZ,M  
ARRESTING OFFICER AGENCY: Naples PD  
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 06/02/2018 00:45:40

---

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.080	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.067	00:41
Breath Volume	3.355	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.060	00:44
Breath Volume	2.375	00:44
Air Blank	0.000	00:44
Control Test	0.078	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

---

Data Download Date/Time: 06/05/2018 17:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIBEY,AMANDA,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: WOODS,H  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: SAMANICH

DATE/TIME OF TEST: 06/02/2018 01:41:29

---

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.079	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.000	01:37
Breath Volume	2.394	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.000	01:40
Breath Volume	1.605	01:40
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

---

Data Download Date/Time: 06/05/2018 17:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SACKMANN,ANDREW,L  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: SACKMANN,ANDREW  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: SCHECHTER

DATE/TIME OF TEST: 06/02/2018 03:24:22

---

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.118	03:19
Breath Volume	3.679	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.114	03:22
Breath Volume	3.003	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

---

Data Download Date/Time: 06/05/2018 17:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARGO,TIMOTHY,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: WARGO,TIMOTHY  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 06/02/2018 05:38:08

---

Diagnostics Check	OK	05:30
Air Blank	0.000	05:31
Control Test	0.080	05:31
Air Blank	0.000	05:32
Subject Sample # 1	0.143	05:33
Breath Volume	2.480	05:33
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.131	05:36
Breath Volume	1.824	05:36
Air Blank	0.000	05:37
Control Test	0.079	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

---

Data Download Date/Time: 06/05/2018 17:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORNELIUS,TIM,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CORNELIUS,TIM  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: MATOS BARRERO

DATE/TIME OF TEST: 06/02/2018 22:56:00

---

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.080	22:49
Air Blank	0.000	22:49
Subject Sample # 1	0.176	22:50
Breath Volume	2.015	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.180	22:54
Breath Volume	1.464	22:54
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:56

---

Data Download Date/Time: 06/05/2018 17:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KENNEDY,DANIEL  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 06/03/2018 01:20:01

---

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.129	01:15
Breath Volume	2.628	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.127	01:18
Breath Volume	2.789	01:18
Air Blank	0.000	01:19
Control Test	0.079	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:20

---

Data Download Date/Time: 06/05/2018 17:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRISCOLL,JAMES,P  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: DRISCOLL,JAMES  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: LENGYEL

DATE/TIME OF TEST: 06/03/2018 01:41:02

---

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:28	
Control Test	0.079	01:28	
Air Blank	0.000	01:29	
Subject Sample	# 1 0.269	01:29	
Breath Volume	3.277	01:29	
Air Blank	0.000	01:30	
Air Blank	0.000	01:32	
Subject Sample	# 2 0.271	01:35	Volume Not Met
Breath Volume	0.000	01:35	
Air Blank	0.000	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 3 0.259	01:39	
Breath Volume	2.429	01:39	
Air Blank	0.000	01:40	
Control Test	0.076	01:40	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:41	

---

Data Download Date/Time: 06/05/2018 17:02



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,ONDREA,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: GROSSKLAS,D  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: MARTINEZ CARO

DATE/TIME OF TEST: 06/03/2018 21:06:59

---

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.079	20:59
Air Blank	0.000	21:00
Subject Sample # 1	0.155	21:02
Breath Volume	1.851	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:04
Subject Sample # 2	0.163	21:05
Breath Volume	1.355	21:05
Air Blank	0.000	21:06
Control Test	0.078	21:06
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

---

Data Download Date/Time: 06/05/2018 17:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,KEVIN,J  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: RUSSEL,D  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 05/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A2  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: VALENCIA

DATE/TIME OF TEST: 06/05/2018 04:08:57

---

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.000	04:04
Breath Volume	2.085	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.000	04:07
Breath Volume	1.773	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

---

Data Download Date/Time: 06/05/2018 17:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEONARD,ALEYA,D  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: KOCHANOWSKI

DATE/TIME OF TEST: 06/02/2018 01:10:29

---

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:05	
Control Test	0.081	01:05	
Air Blank	0.000	01:06	
Subject Sample	# 1	01:09	Subject Test Refused
Breath Volume	0.000	01:09	
Air Blank	0.000	01:09	
Control Test	0.082	01:09	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:10	

---

Data Download Date/Time: 06/29/2018 06:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEONARD,ALEYA,D  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: KOCHANOWSKI

DATE/TIME OF TEST: 06/02/2018 01:37:23

---

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.082	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.118	01:32
Breath Volume	2.578	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:34
Subject Sample # 2	0.119	01:35
Breath Volume	1.792	01:35
Air Blank	0.000	01:36
Control Test	0.082	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

---

Data Download Date/Time: 06/29/2018 06:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEONARD,ALEYA,D  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: BERNARD,NICK  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BUEHLER

DATE/TIME OF TEST: 06/02/2018 03:27:29

---

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.082	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.119	03:22
Breath Volume	3.593	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.122	03:26
Breath Volume	3.246	03:26
Air Blank	0.000	03:26
Control Test	0.082	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

---

Data Download Date/Time: 06/29/2018 06:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FISHER,SCOTT,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: COTE,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BIENKOWSKI

DATE/TIME OF TEST: 06/02/2018 04:15:07

---

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.081	04:02	
Air Blank	0.000	04:02	
Subject Sample # 1	0.205	04:03	
Breath Volume	2.039	04:03	
Air Blank	0.000	04:03	
Air Blank	0.000	04:05	
Subject Sample # 2	0.177	04:08	Volume Not Met
Breath Volume	0.820	04:08	
Air Blank	0.000	04:09	
Air Blank	0.000	04:11	
Subject Sample # 3	0.191	04:14	Volume Not Met
Breath Volume	0.113	04:14	
Air Blank		04:14	Improper Sample
Air Blank	0.000	04:15	

---

Data Download Date/Time: 06/29/2018 06:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FISHER,SCOTT,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: WHITING,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 06/02/2018 23:29:47

---

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.081	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.146	23:25
Breath Volume	2.250	23:25
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.164	23:28
Breath Volume	3.765	23:28
Air Blank	0.000	23:28
Control Test	0.082	23:29
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

---

Data Download Date/Time: 06/29/2018 06:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PROSPER,BENSON,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: RICHARDSON,B  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 06/05/2018 00:57:26

---

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.081	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.000	00:52
Breath Volume	1.648	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.000	00:56
Breath Volume	2.183	00:56
Air Blank	0.000	00:56
Control Test	0.082	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

---

Data Download Date/Time: 06/29/2018 06:56



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,DONNA,M  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: LAZARSKI,KRISTA  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: ALLSHOUSE

DATE/TIME OF TEST: 06/06/2018 19:45:55

---

Diagnostics Check	OK	19:38
Air Blank	0.000	19:38
Control Test	0.082	19:39
Air Blank	0.000	19:39
Subject Sample # 1	0.087	19:41
Breath Volume	1.355	19:41
Air Blank	0.000	19:42
Air Blank	0.000	19:43
Subject Sample # 2	0.095	19:44
Breath Volume	1.277	19:44
Air Blank	0.000	19:45
Control Test	0.081	19:45
Air Blank	0.000	19:45
Diagnostics Check	OK	19:45

---

Data Download Date/Time: 06/29/2018 06:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMERON,CELESTE,M  
BREATH TEST OPERATOR AGENCY: Punta Gorda PD  
ARRESTING OFFICER: BURGA,F  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: HARDIN

DATE/TIME OF TEST: 06/08/2018 06:29:23

---

Diagnostics Check	OK	06:23
Air Blank	0.000	06:23
Control Test	0.082	06:23
Air Blank	0.000	06:24
Subject Sample # 1	0.212	06:24
Breath Volume	1.671	06:24
Air Blank	0.000	06:25
Air Blank	0.000	06:27
Subject Sample # 2	0.209	06:27
Breath Volume	2.320	06:27
Air Blank	0.000	06:28
Control Test	0.081	06:28
Air Blank	0.000	06:29
Diagnostics Check	OK	06:29

---

Data Download Date/Time: 06/29/2018 06:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PROSPER,BENSON,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: RIVERA,J  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: THIGPEN JR

DATE/TIME OF TEST: 06/09/2018 02:59:02

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.083	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.088	02:54
Breath Volume	1.589	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:56
Subject Sample # 2	0.088	02:57
Breath Volume	2.070	02:57
Air Blank	0.000	02:58
Control Test	0.082	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:59

---

Data Download Date/Time: 06/29/2018 06:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMERON,CELESTE,M  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: VERBIEST,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 08:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 06/09/2018 10:29:41

---

Diagnostics Check	OK	10:23
Air Blank	0.000	10:24
Control Test	0.082	10:24
Air Blank	0.000	10:25
Subject Sample # 1	0.000	10:25
Breath Volume	2.789	10:25
Air Blank	0.000	10:26
Air Blank	0.000	10:28
Subject Sample # 2	0.000	10:28
Breath Volume	2.746	10:28
Air Blank	0.000	10:28
Control Test	0.081	10:29
Air Blank	0.000	10:29
Diagnostics Check	OK	10:29

---

Data Download Date/Time: 06/29/2018 06:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMERON,CELESTE,M  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: CIOLINO,L  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 07:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:49

SUBJECT'S LAST NAME: ANACASIS

DATE/TIME OF TEST: 06/10/2018 08:17:32

---

Diagnostics Check	OK	08:10
Air Blank	0.000	08:10
Control Test	0.082	08:10
Air Blank	0.000	08:11
Subject Sample # 1	0.083	08:12
Breath Volume	1.320	08:12
Air Blank	0.000	08:13
Air Blank	0.000	08:14
Subject Sample # 2	0.078	08:15
Breath Volume	1.582	08:15
Air Blank	0.000	08:16
Control Test	0.081	08:16
Air Blank	0.000	08:17
Diagnostics Check	OK	08:17

---

Data Download Date/Time: 06/29/2018 06:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,DONNA,M  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: PARKHILL,GREG  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 12:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 06/13/2018 01:48:30

---

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.082	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.185	01:44
Breath Volume	3.765	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.189	01:47
Breath Volume	3.324	01:47
Air Blank	0.000	01:47
Control Test	0.081	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

---

Data Download Date/Time: 06/29/2018 06:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT, BRADLEY, W  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: MORAGHAN, JOHN  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 10:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:31

SUBJECT'S LAST NAME: AGIS

DATE/TIME OF TEST: 06/13/2018 12:00:51

---

Diagnostics Check	OK	11:54
Air Blank	0.000	11:55
Control Test	0.082	11:55
Air Blank	0.000	11:55
Subject Sample # 1	0.000	11:56
Breath Volume	1.964	11:56
Air Blank	0.000	11:57
Air Blank	0.000	11:59
Subject Sample # 2	0.000	11:59
Breath Volume	1.125	11:59
Air Blank	0.000	11:59
Control Test	0.082	12:00
Air Blank	0.000	12:00
Diagnostics Check	OK	12:00

---

Data Download Date/Time: 06/29/2018 06:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: THABIT,S  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: PEFFLEY

DATE/TIME OF TEST: 06/15/2018 00:42:49

---

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.082	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.192	00:38
Breath Volume	4.078	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.180	00:41
Breath Volume	1.710	00:41
Air Blank	0.000	00:41
Control Test	0.082	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

---

Data Download Date/Time: 06/29/2018 06:56



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEONARD,ALEYA,D  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: COX,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: TAMBURRI

DATE/TIME OF TEST: 06/16/2018 00:17:17

---

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.082	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.107	00:12
Breath Volume	2.003	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:14
Subject Sample # 2	0.107	00:15
Breath Volume	2.804	00:15
Air Blank	0.000	00:16
Control Test	0.082	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

---

Data Download Date/Time: 06/29/2018 06:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FISHER,SCOTT,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: COX,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: RANGER

DATE/TIME OF TEST: 06/16/2018 23:49:39

---

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.082	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.239	23:45
Breath Volume	1.925	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.257	23:48
Breath Volume	1.890	23:48
Air Blank	0.000	23:48
Control Test	0.082	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

---

Data Download Date/Time: 06/29/2018 06:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FISHER,SCOTT,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: ARTHUR,TROOPER  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: MELNICHENKO

DATE/TIME OF TEST: 06/17/2018 18:47:33

---

Diagnostics Check	OK	18:38
Air Blank	0.000	18:39
Control Test	0.081	18:39
Air Blank	0.000	18:39
Subject Sample # 1	0.313	18:40
Breath Volume	3.832	18:40
Air Blank	0.000	18:40
Air Blank	0.000	18:42
Subject Sample # 2	0.258	18:43
Breath Volume	1.253	18:43
Air Blank	0.000	18:43
Air Blank	0.000	18:45
Subject Sample # 3	0.271	18:46
Breath Volume	1.320	18:46
Air Blank	0.000	18:46
Control Test	0.081	18:46
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

---

Data Download Date/Time: 06/29/2018 06:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEONARD,ALEYA,D  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: LAZARSKI,KRISTA  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: ROBLES

DATE/TIME OF TEST: 06/18/2018 03:53:04

---

Diagnostics Check		OK	03:39	
Air Blank		0.000	03:40	
Control Test		0.083	03:40	
Air Blank		0.000	03:40	
Subject Sample	# 1	0.082	03:44	Volume Not Met
Breath Volume		0.722	03:44	
Air Blank		0.000	03:44	
Air Blank		0.000	03:46	
Subject Sample	# 2	0.089	03:47	
Breath Volume		2.011	03:47	
Air Blank		0.000	03:48	
Air Blank		0.000	03:50	
Subject Sample	# 3	0.106	03:51	
Breath Volume		1.160	03:51	
Air Blank		0.000	03:52	
Control Test		0.080	03:52	
Air Blank		0.000	03:52	
Diagnostics Check		OK	03:53	

---

Data Download Date/Time: 06/29/2018 06:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUCKER,DEPUTY,S  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: NELSON,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 08:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: REAGAN

DATE/TIME OF TEST: 06/20/2018 09:55:19

---

Diagnostics Check	OK	09:49
Air Blank	0.000	09:49
Control Test	0.082	09:49
Air Blank	0.000	09:50
Subject Sample # 1	0.155	09:50
Breath Volume	2.988	09:50
Air Blank	0.000	09:51
Air Blank	0.000	09:53
Subject Sample # 2	0.155	09:53
Breath Volume	2.976	09:53
Air Blank	0.000	09:54
Control Test	0.081	09:54
Air Blank	0.000	09:55
Diagnostics Check	OK	09:55

---

Data Download Date/Time: 06/29/2018 06:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FISHER,SCOTT,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: SARVER,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: TEJADA

DATE/TIME OF TEST: 06/21/2018 02:10:15

---

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.081	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.249	02:05
Breath Volume	1.539	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.259	02:08
Breath Volume	1.312	02:08
Air Blank	0.000	02:09
Control Test	0.082	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

---

Data Download Date/Time: 06/29/2018 06:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: WEDGEWOOD,M  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: FOLEY

DATE/TIME OF TEST: 06/23/2018 02:34:38

---

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.082	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.164	02:29
Breath Volume	1.558	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.163	02:33
Breath Volume	1.207	02:33
Air Blank	0.000	02:33
Control Test	0.081	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

---

Data Download Date/Time: 06/29/2018 06:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAFFARAS,CELINA,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: LIBBY,J  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: JEAN-BAPTISTE

DATE/TIME OF TEST: 06/24/2018 05:34:04

---

Diagnostics Check	OK	05:27
Air Blank	0.000	05:27
Control Test	0.081	05:27
Air Blank	0.000	05:28
Subject Sample # 1	0.164	05:29
Breath Volume	2.109	05:29
Air Blank	0.000	05:30
Air Blank	0.000	05:31
Subject Sample # 2	0.157	05:32
Breath Volume	1.953	05:32
Air Blank	0.000	05:33
Control Test	0.082	05:33
Air Blank	0.000	05:33
Diagnostics Check	OK	05:34

---

Data Download Date/Time: 06/29/2018 06:57



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PROSPER,BENSON,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GERHART,P  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: OMAHEN

DATE/TIME OF TEST: 06/24/2018 21:46:00

---

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.082	21:39
Air Blank	0.000	21:40
Subject Sample # 1	0.095	21:41
Breath Volume	2.062	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.092	21:44
Breath Volume	1.621	21:44
Air Blank	0.000	21:45
Control Test	0.081	21:45
Air Blank	0.000	21:45
Diagnostics Check	OK	21:46

---

Data Download Date/Time: 06/29/2018 06:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EGAN,MELSSA',  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: MATT WEDGEWOOD,MATT  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: DELANEY

DATE/TIME OF TEST: 06/25/2018 01:51:47

---

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.082	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.152	01:47
Breath Volume	1.964	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.141	01:50
Breath Volume	1.906	01:50
Air Blank	0.000	01:50
Control Test	0.082	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

---

Data Download Date/Time: 06/29/2018 06:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PLATZ,KEVIN,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: PLATZ,KEVIN  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: ADMIN  
ARREST TIME: 22:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 06/26/2018 01:10:59

---

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.082	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.000	01:06
Breath Volume	3.160	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.000	01:09
Breath Volume	2.601	01:09
Air Blank	0.000	01:10
Control Test	0.082	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

---

Data Download Date/Time: 06/29/2018 06:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURROWS,ALBERT,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: BALTIMORE,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: ADMIN  
ARREST TIME: 22:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: SCHEER

DATE/TIME OF TEST: 06/26/2018 09:41:15

---

Diagnostics Check	OK	09:35
Air Blank	0.000	09:35
Control Test	0.082	09:35
Air Blank	0.000	09:36
Subject Sample # 1	0.000	09:36
Breath Volume	2.515	09:36
Air Blank	0.000	09:37
Air Blank	0.000	09:39
Subject Sample # 2	0.000	09:39
Breath Volume	1.921	09:39
Air Blank	0.000	09:40
Control Test	0.082	09:40
Air Blank	0.000	09:41
Diagnostics Check	OK	09:41

---

Data Download Date/Time: 06/29/2018 06:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PLATZ,KEVIN,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: LAZARSKI,KRISTA  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: WOODALL

DATE/TIME OF TEST: 06/26/2018 19:44:33

---

Diagnostics Check	OK	19:36
Air Blank	0.000	19:36
Control Test	0.082	19:37
Air Blank	0.000	19:37
Subject Sample # 1	0.249	19:39
Breath Volume	2.234	19:39
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.247	19:43
Breath Volume	2.664	19:43
Air Blank	0.000	19:43
Control Test	0.081	19:43
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

---

Data Download Date/Time: 06/29/2018 06:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PROSPER,BENSON,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: FARLEY,J  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: KILLEEN

DATE/TIME OF TEST: 06/28/2018 01:21:52

---

Diagnostics Check	OK	01:14
Air Blank	0.000	01:15
Control Test	0.082	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.225	01:17
Breath Volume	1.210	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:19
Subject Sample # 2	0.242	01:20
Breath Volume	1.566	01:20
Air Blank	0.000	01:20
Control Test	0.082	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

---

Data Download Date/Time: 06/29/2018 06:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAFFARAS,CELINA,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: FARLEY,J  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: CORRELL

DATE/TIME OF TEST: 06/29/2018 03:18:50

---

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.082	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.175	03:14
Breath Volume	2.117	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:16
Subject Sample # 2	0.182	03:17
Breath Volume	1.515	03:17
Air Blank	0.000	03:17
Control Test	0.082	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

---

Data Download Date/Time: 06/29/2018 06:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURROWS,DEPUTY,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: BURROWS,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: ADMIN  
ARREST TIME: 23:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: BLOSSER

DATE/TIME OF TEST: 06/30/2018 10:43:13

---

Diagnostics Check	OK	10:36
Air Blank	0.000	10:37
Control Test	0.082	10:37
Air Blank	0.000	10:37
Subject Sample # 1	0.130	10:38
Breath Volume	1.550	10:38
Air Blank	0.000	10:39
Air Blank	0.000	10:41
Subject Sample # 2	0.139	10:41
Breath Volume	1.757	10:41
Air Blank	0.000	10:42
Control Test	0.080	10:42
Air Blank	0.000	10:43
Diagnostics Check	OK	10:43

---

Data Download Date/Time: 07/22/2018 04:16



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURROWS,DEPUTY,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: BURROWS,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: ADMIN  
ARREST TIME: 23:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: BLOSSER

DATE/TIME OF TEST: 06/30/2018 12:01:29

---

Diagnostics Check	OK	11:54
Air Blank	0.000	11:55
Control Test	0.082	11:55
Air Blank	0.000	11:56
Subject Sample # 1	0.120	11:56
Breath Volume	1.492	11:56
Air Blank	0.000	11:57
Air Blank	0.000	11:59
Subject Sample # 2	0.126	11:59
Breath Volume	1.238	11:59
Air Blank	0.000	12:00
Control Test	0.082	12:00
Air Blank	0.000	12:01
Diagnostics Check	OK	12:01

---

Data Download Date/Time: 07/22/2018 04:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMPBELL,J,H  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: CAMPBELL,J  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 13:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: BLOSSER

DATE/TIME OF TEST: 06/30/2018 13:59:54

---

Diagnostics Check	OK	13:52
Air Blank	0.000	13:53
Control Test	0.082	13:53
Air Blank	0.000	13:54
Subject Sample # 1	0.088	13:55
Breath Volume	1.371	13:55
Air Blank	0.000	13:55
Air Blank	0.000	13:57
Subject Sample # 2	0.097	13:58
Breath Volume	1.292	13:58
Air Blank	0.000	13:59
Control Test	0.082	13:59
Air Blank	0.000	13:59
Diagnostics Check	OK	13:59

---

Data Download Date/Time: 07/22/2018 04:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURROWS,DEPUTY,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: BURROWS,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: ADMIN  
ARREST TIME: 23:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: BLOSSER

DATE/TIME OF TEST: 06/30/2018 16:01:38

---

Diagnostics Check	OK	15:55
Air Blank	0.000	15:55
Control Test	0.082	15:56
Air Blank	0.000	15:56
Subject Sample # 1	0.063	15:57
Breath Volume	1.468	15:57
Air Blank	0.000	15:57
Air Blank	0.000	15:59
Subject Sample # 2	0.063	16:00
Breath Volume	2.207	16:00
Air Blank	0.000	16:00
Control Test	0.081	16:01
Air Blank	0.000	16:01
Diagnostics Check	OK	16:01

---

Data Download Date/Time: 07/22/2018 04:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURROWS,DEPUTY,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: BURROWS,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: ADMIN  
ARREST TIME: 23:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: BLOSSER

DATE/TIME OF TEST: 06/30/2018 17:22:21

---

Diagnostics Check	OK	17:16
Air Blank	0.000	17:16
Control Test	0.082	17:17
Air Blank	0.000	17:17
Subject Sample # 1	0.041	17:18
Breath Volume	1.656	17:18
Air Blank	0.000	17:18
Air Blank	0.000	17:20
Subject Sample # 2	0.040	17:20
Breath Volume	1.589	17:20
Air Blank	0.000	17:21
Control Test	0.082	17:21
Air Blank	0.000	17:22
Diagnostics Check	OK	17:22

---

Data Download Date/Time: 07/22/2018 04:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONDRAGON,EDWIN,F  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: MATHIEU,REGINALD  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 06/11/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A2  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: WATFORD III

DATE/TIME OF TEST: 06/16/2018 19:08:55

---

Diagnostics Check	OK	19:06	
Air Blank	0.000	19:06	
Control Test	0.079	19:06	
Air Blank	0.000	19:07	
Subject Sample	# 1	19:07	Subject Test Refused
Breath Volume	0.000	19:07	
Air Blank	0.000	19:07	
Control Test	0.079	19:08	
Air Blank	0.000	19:08	
Diagnostics Check	OK	19:08	

---

Data Download Date/Time: 07/02/2018 07:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONDRAGON,EDWIN,F  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: RODRIGUEZ,J.A.  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 06/11/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A2  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: MANCIA-PADILLA

DATE/TIME OF TEST: 06/24/2018 05:37:06

---

Diagnostics Check	OK	05:34	
Air Blank	0.000	05:34	
Control Test	0.078	05:35	
Air Blank	0.000	05:35	
Subject Sample	# 1	05:35	Subject Test Refused
Breath Volume	0.000	05:35	
Air Blank	0.000	05:36	
Control Test	0.079	05:36	
Air Blank	0.000	05:37	
Diagnostics Check	OK	05:37	

---

Data Download Date/Time: 07/02/2018 07:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KELLER,DAVID,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: KELLER,DAVID  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 06/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416090A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: BRAVO

DATE/TIME OF TEST: 06/19/2018 02:03:41

---

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.078	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.149	01:59
Breath Volume	2.894	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.148	02:02
Breath Volume	2.832	02:02
Air Blank	0.000	02:02
Control Test	0.079	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

---

Data Download Date/Time: 07/31/2018 13:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KELLER, DAVID,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: KELLER, DAVID  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 06/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416090A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BURNSIDE

DATE/TIME OF TEST: 06/22/2018 03:16:17

---

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.079	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.168	03:12
Breath Volume	1.457	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.171	03:14
Breath Volume	1.449	03:14
Air Blank	0.000	03:15
Control Test	0.081	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

---

Data Download Date/Time: 07/31/2018 13:05



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMM,JAMES,L  
BREATH TEST OPERATOR AGENCY: Franklin County SO  
ARRESTING OFFICER: MILLER,KRISTEN  
ARRESTING OFFICER AGENCY: Franklin County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:47

SUBJECT'S LAST NAME: MARANICH

DATE/TIME OF TEST: 06/19/2018 18:26:13

---

Diagnostics Check	OK	18:19
Air Blank	0.000	18:20
Control Test	0.079	18:20
Air Blank	0.000	18:21
Subject Sample # 1	0.118	18:21
Breath Volume	2.320	18:21
Air Blank	0.000	18:22
Air Blank	0.000	18:24
Subject Sample # 2	0.105	18:24
Breath Volume	2.179	18:24
Air Blank	0.000	18:25
Control Test	0.079	18:25
Air Blank	0.000	18:26
Diagnostics Check	OK	18:26

---

Data Download Date/Time: 07/16/2018 10:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARD,JAMES,H  
BREATH TEST OPERATOR AGENCY: Franklin County SO  
ARRESTING OFFICER: BRYANT,STELLA  
ARRESTING OFFICER AGENCY: Franklin County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: DUI  
ARREST TIME: 11:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 06/24/2018 12:36:52

---

Diagnostics Check	OK	12:30
Air Blank	0.000	12:30
Control Test	0.079	12:31
Air Blank	0.000	12:31
Subject Sample # 1	0.236	12:32
Breath Volume	2.687	12:32
Air Blank	0.000	12:33
Air Blank	0.000	12:34
Subject Sample # 2	0.223	12:35
Breath Volume	3.511	12:35
Air Blank	0.000	12:35
Control Test	0.077	12:36
Air Blank	0.000	12:36
Diagnostics Check	OK	12:36

---

Data Download Date/Time: 07/16/2018 10:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANDERSON,GENE,A  
BREATH TEST OPERATOR AGENCY: Franklin County SO  
ARRESTING OFFICER: OSBURN,TRAVIS  
ARRESTING OFFICER AGENCY: Franklin County SO  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 06/29/2018 20:28:28

---

Diagnostics Check	OK	20:22
Air Blank	0.000	20:22
Control Test	0.077	20:22
Air Blank	0.000	20:23
Subject Sample # 1	0.239	20:23
Breath Volume	3.308	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.246	20:26
Breath Volume	4.027	20:26
Air Blank	0.000	20:27
Control Test	0.077	20:27
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

---

Data Download Date/Time: 07/16/2018 10:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLEIGHER,JOHN,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: WILCOX,STEPHANIE  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14816080A3  
DRY GAS EXPIRY DATE: 07/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: PIPPIN

DATE/TIME OF TEST: 06/23/2018 19:54:55

---

Diagnostics Check	OK	19:42	
Air Blank	0.000	19:43	
Control Test	0.077	19:43	
Air Blank	0.000	19:44	
Subject Sample	# 1	19:47	No Sample Provided
Breath Volume	0.000	19:47	
Air Blank	0.000	19:47	
Air Blank	0.000	19:49	
Subject Sample	# 2	0.081	19:50
Breath Volume	2.738	19:50	
Air Blank	0.000	19:50	
Air Blank	0.000	19:52	
Subject Sample	# 3	0.077	19:53
Breath Volume	3.031	19:53	
Air Blank	0.000	19:54	
Control Test	0.077	19:54	
Air Blank	0.000	19:54	
Diagnostics Check	OK	19:54	

---

Data Download Date/Time: 07/18/2018 10:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RABUN, MATTHEW, K  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: RABUN, MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 05/23/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14816080A3  
DRY GAS EXPIRY DATE: 07/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: PAPPAS

DATE/TIME OF TEST: 06/04/2018 19:41:32

---

Diagnostics Check	OK	19:35
Air Blank	0.000	19:35
Control Test	0.078	19:35
Air Blank	0.000	19:36
Subject Sample # 1	0.164	19:36
Breath Volume	1.761	19:36
Air Blank	0.000	19:37
Air Blank	0.000	19:39
Subject Sample # 2	0.157	19:39
Breath Volume	1.515	19:39
Air Blank	0.000	19:40
Control Test	0.077	19:40
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

---

Data Download Date/Time: 06/19/2018 10:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAKER,RUEL,W  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: SWINDELL,SEAN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 06/14/2018 17:14:26

---

Diagnostics Check	OK	17:13	
Air Blank		17:13	RFI Detect
Air Blank	0.000	17:14	

---

Data Download Date/Time: 07/11/2018 11:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAKER,RUEL,A  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: SWINDELL,SEAN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 06/14/2018 17:23:53

---

Diagnostics Check		OK	17:17
Air Blank		0.000	17:17
Control Test		0.081	17:18
Air Blank		0.000	17:18
Subject Sample	# 1	0.362	17:19
Breath Volume		2.792	17:19
Air Blank		0.000	17:20
Air Blank		0.000	17:21
Subject Sample	# 2	0.370	17:22
Breath Volume		2.824	17:22
Air Blank		0.000	17:22
Control Test		0.081	17:23
Air Blank		0.000	17:23
Diagnostics Check		OK	17:23

---

Data Download Date/Time: 07/11/2018 11:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPANGLER,GARY,A  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SCHILLING,JOHN  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/04/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 06/21/2018 21:30:06

---

Diagnostics Check	OK	21:26	
Air Blank	0.000	21:26	
Control Test	0.082	21:26	
Air Blank	0.000	21:27	
Subject Sample	# 1	21:28	Subject Test Refused
Breath Volume	0.000	21:28	
Air Blank	0.000	21:29	
Control Test	0.082	21:29	
Air Blank	0.000	21:30	
Diagnostics Check	OK	21:30	

---

Data Download Date/Time: 07/11/2018 11:34



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TALLMAN,JOHN,R  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: TALLMAN,JOHN  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: MCKINNEY

DATE/TIME OF TEST: 06/02/2018 05:01:43

---

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.081	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.170	04:57
Breath Volume	1.167	04:57
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.168	05:00
Breath Volume	1.234	05:00
Air Blank	0.000	05:00
Control Test	0.079	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

---

Data Download Date/Time: 06/12/2018 10:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEMPLIN,RICHARD,P  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: TEMPLIN,RICHARD  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: CARLSON

DATE/TIME OF TEST: 06/10/2018 19:17:49

---

Diagnostics Check	OK	19:10
Air Blank	0.000	19:10
Control Test	0.081	19:11
Air Blank	0.000	19:11
Subject Sample # 1	0.142	19:13
Breath Volume	2.371	19:13
Air Blank	0.000	19:13
Air Blank	0.000	19:15
Subject Sample # 2	0.144	19:16
Breath Volume	2.812	19:16
Air Blank	0.000	19:16
Control Test	0.080	19:17
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

---

Data Download Date/Time: 06/12/2018 10:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FINUFF,DANIEL,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: FINUFF,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14816080A3  
DRY GAS EXPIRY DATE: 07/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: MANCIA HERNANDEZ

DATE/TIME OF TEST: 06/02/2018 21:15:52

---

Diagnostics Check	OK	21:09
Air Blank	0.000	21:09
Control Test	0.079	21:10
Air Blank	0.000	21:10
Subject Sample # 1	0.235	21:10
Breath Volume	2.304	21:10
Air Blank	0.000	21:11
Air Blank	0.000	21:13
Subject Sample # 2	0.238	21:14
Breath Volume	2.324	21:14
Air Blank	0.000	21:14
Control Test	0.079	21:15
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

---

Data Download Date/Time: 06/20/2018 10:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FINUFF,DANIEL,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: FINUFF,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 05/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14816080A3  
DRY GAS EXPIRY DATE: 07/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 17:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: KYSER

DATE/TIME OF TEST: 06/17/2018 19:05:05

---

Diagnostics Check	OK	18:59
Air Blank	0.000	18:59
Control Test	0.080	18:59
Air Blank	0.000	19:00
Subject Sample # 1	0.000	19:00
Breath Volume	1.785	19:00
Air Blank	0.000	19:01
Air Blank	0.000	19:03
Subject Sample # 2	0.000	19:03
Breath Volume	1.933	19:03
Air Blank	0.000	19:04
Control Test	0.080	19:04
Air Blank	0.000	19:05
Diagnostics Check	OK	19:05

---

Data Download Date/Time: 06/20/2018 10:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEEKMAN,JACOB,C  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: BOCCIO,GLORIA  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 06/09/2018 04:52:34

---

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.080	04:45
Air Blank	0.000	04:45
Subject Sample # 1	0.132	04:46
Breath Volume	2.699	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.134	04:50
Breath Volume	2.757	04:50
Air Blank	0.000	04:51
Control Test	0.080	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

---

Data Download Date/Time: 06/18/2018 18:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEEKMAN,JACOB,C  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: BEEKMAN,JACOB  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: GARCIAGAONAHERNANDEZ

DATE/TIME OF TEST: 06/10/2018 03:39:53

---

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.081	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.157	03:35
Breath Volume	2.011	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.165	03:38
Breath Volume	2.339	03:38
Air Blank	0.000	03:38
Control Test	0.080	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

---

Data Download Date/Time: 06/18/2018 18:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VELEZ,H,  
BREATH TEST OPERATOR AGENCY: Apopka PD  
ARRESTING OFFICER: JEREZ,IRVIN  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 13:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:13

SUBJECT'S LAST NAME: GURKSNIS

DATE/TIME OF TEST: 06/16/2018 14:42:06

---

Diagnostics Check	OK	14:34
Air Blank	0.000	14:35
Control Test	0.077	14:35
Air Blank	0.000	14:36
Subject Sample # 1	0.253	14:37
Breath Volume	1.750	14:37
Air Blank	0.000	14:38
Air Blank	0.000	14:39
Subject Sample # 2	0.263	14:40
Breath Volume	2.433	14:40
Air Blank	0.000	14:41
Control Test	0.077	14:41
Air Blank	0.000	14:42
Diagnostics Check	OK	14:42

---

Data Download Date/Time: 06/29/2018 11:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARNHORST, EDDIE, J

BREATH TEST OPERATOR AGENCY: Apopka PD

ARRESTING OFFICER: CALAMO, KATIE

ARRESTING OFFICER AGENCY: Apopka PD

LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 17:16

GENDER: Female

OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: BALKCOM

DATE/TIME OF TEST: 06/19/2018 18:07:49

---

Diagnostics Check	OK	17:55
Air Blank	0.000	17:56
Control Test	0.075	17:56
Air Blank	0.000	17:57
Subject Sample # 1	0.279	17:58
Breath Volume	2.242	17:58
Air Blank	0.000	17:59
Air Blank	0.000	18:01
Subject Sample # 2	0.250	18:02
Breath Volume	1.320	18:02
Air Blank	0.000	18:03
Air Blank	0.000	18:05
Subject Sample # 3	0.280	18:06
Breath Volume	1.148	18:06
Air Blank	0.000	18:06
Control Test	0.076	18:07
Air Blank	0.000	18:07
Diagnostics Check	OK	18:07

---

Data Download Date/Time: 06/29/2018 11:46



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VELEZ,H,  
BREATH TEST OPERATOR AGENCY: Apopka PD  
ARRESTING OFFICER: CINTRON,K  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 14:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: DONAHUE

DATE/TIME OF TEST: 06/23/2018 15:35:53

---

Diagnostics Check	OK	15:28
Air Blank	0.000	15:29
Control Test	0.076	15:29
Air Blank	0.000	15:30
Subject Sample # 1	0.188	15:30
Breath Volume	3.027	15:30
Air Blank	0.000	15:31
Air Blank	0.000	15:33
Subject Sample # 2	0.179	15:34
Breath Volume	2.199	15:34
Air Blank	0.000	15:34
Control Test	0.075	15:35
Air Blank	0.000	15:35
Diagnostics Check	OK	15:35

---

Data Download Date/Time: 06/29/2018 11:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: GANDIA-SANCHEZ

DATE/TIME OF TEST: 06/03/2018 23:03:06

---

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:53	
Control Test	0.078	22:53	
Air Blank	0.000	22:53	
Subject Sample	# 1	22:57	No Sample Provided
Breath Volume	0.000	22:57	
Air Blank	0.000	22:57	
Air Blank	0.000	22:59	
Subject Sample	# 2	23:01	Subject Test Refused
Breath Volume	0.000	23:01	
Air Blank	0.000	23:02	
Control Test	0.079	23:02	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

---

Data Download Date/Time: 06/05/2018 21:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SAYERS,MICHELLE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 05/03/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 05:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: GONZALEZ GONZALEZ

DATE/TIME OF TEST: 06/05/2018 05:57:26

---

Diagnostics Check	OK	05:49
Air Blank	0.000	05:50
Control Test	0.078	05:50
Air Blank	0.000	05:51
Subject Sample # 1	0.045	05:52
Breath Volume	2.816	05:52
Air Blank	0.000	05:53
Air Blank	0.000	05:55
Subject Sample # 2	0.044	05:55
Breath Volume	4.468	05:55
Air Blank	0.000	05:56
Control Test	0.079	05:56
Air Blank	0.000	05:57
Diagnostics Check	OK	05:57

---

Data Download Date/Time: 06/05/2018 21:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: CAMPOS,MICHAEL  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/07/2018 04:47:54

---

Diagnostics Check		OK	04:35	
Air Blank		0.000	04:35	
Control Test		0.079	04:35	
Air Blank		0.000	04:36	
Subject Sample	# 1	0.138	04:39	Volume Not Met
Breath Volume		0.710	04:39	
Air Blank		0.000	04:40	
Air Blank		0.000	04:41	
Subject Sample	# 2	0.161	04:42	
Breath Volume		1.636	04:42	
Air Blank		0.000	04:43	
Air Blank		0.000	04:44	
Subject Sample	# 3	0.167	04:46	
Breath Volume		1.167	04:46	
Air Blank		0.000	04:46	
Control Test		0.078	04:47	
Air Blank		0.000	04:47	
Diagnostics Check		OK	04:47	

---

Data Download Date/Time: 06/18/2018 18:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 05:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: VAZQUEZ-GOMEZ

DATE/TIME OF TEST: 06/07/2018 06:04:28

---

Diagnostics Check	OK	05:55
Air Blank	0.000	05:56
Control Test	0.079	05:56
Air Blank	0.000	05:56
Subject Sample # 1	0.000	05:59
Breath Volume	1.164	05:59
Air Blank	0.000	06:00
Air Blank	0.000	06:02
Subject Sample # 2	0.000	06:03
Breath Volume	1.238	06:03
Air Blank	0.000	06:03
Control Test	0.079	06:03
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

---

Data Download Date/Time: 06/18/2018 18:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: AKINS,JUSTIN  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 06/08/2018 05:31:18

---

Diagnostics Check	OK	05:25
Air Blank	0.000	05:25
Control Test	0.079	05:25
Air Blank	0.000	05:26
Subject Sample # 1	0.135	05:26
Breath Volume	3.179	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.136	05:29
Breath Volume	3.074	05:29
Air Blank	0.000	05:30
Control Test	0.078	05:30
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

---

Data Download Date/Time: 06/18/2018 18:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORUJO,PEDRO,L  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: CORUJO,PEDRO  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: DIAZ MELENDEZ

DATE/TIME OF TEST: 06/09/2018 03:19:12

---

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:16	
Control Test	0.078	03:16	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:18	
Control Test	0.078	03:18	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

---

Data Download Date/Time: 06/18/2018 18:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORUJO,PEDRO,L  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: JOSEPH,PASCAL  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: ZAYAS POZO

DATE/TIME OF TEST: 06/09/2018 03:30:21

---

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.078	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.166	03:25
Breath Volume	2.773	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.163	03:28
Breath Volume	3.121	03:28
Air Blank	0.000	03:29
Control Test	0.078	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

---

Data Download Date/Time: 06/18/2018 18:31



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FORD, GLENN, W  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: MINERVINO, KRISTIN  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 07:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:42

SUBJECT'S LAST NAME: RUIZ ROSADO

DATE/TIME OF TEST: 06/10/2018 08:11:10

---

Diagnostics Check	OK	08:03
Air Blank	0.000	08:04
Control Test	0.079	08:04
Air Blank	0.000	08:05
Subject Sample # 1	0.094	08:05
Breath Volume	1.832	08:05
Air Blank	0.000	08:06
Air Blank	0.000	08:08
Subject Sample # 2	0.092	08:09
Breath Volume	1.707	08:09
Air Blank	0.000	08:10
Control Test	0.078	08:10
Air Blank	0.000	08:11
Diagnostics Check	OK	08:11

---

Data Download Date/Time: 06/18/2018 18:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FIGUERAS COLON,LUIS,A  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: ROSCOE,JAMES  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 12:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: FIGUERAS COLON

DATE/TIME OF TEST: 06/11/2018 13:47:42

---

Diagnostics Check	OK	13:41
Air Blank	0.000	13:41
Control Test	0.078	13:42
Air Blank	0.000	13:42
Subject Sample # 1	0.000	13:43
Breath Volume	1.691	13:43
Air Blank	0.000	13:43
Air Blank	0.000	13:45
Subject Sample # 2	0.000	13:46
Breath Volume	2.214	13:46
Air Blank	0.000	13:46
Control Test	0.078	13:47
Air Blank	0.000	13:47
Diagnostics Check	OK	13:47

---

Data Download Date/Time: 06/18/2018 18:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVA,EMILIANO,R  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: SILVA,EMILIANO  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: LEACHMAN

DATE/TIME OF TEST: 06/15/2018 04:39:14

---

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:36	
Control Test	0.079	04:36	
Air Blank	0.000	04:37	
Subject Sample	# 1	04:37	Subject Test Refused
Breath Volume	0.000	04:37	
Air Blank	0.000	04:38	
Control Test	0.078	04:38	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

---

Data Download Date/Time: 06/18/2018 18:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYONS,DANIEL,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: RAYES,RICHARD  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 06/16/2018 04:10:56

---

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.078	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.205	04:06
Breath Volume	2.089	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.187	04:09
Breath Volume	1.382	04:09
Air Blank	0.000	04:10
Control Test	0.077	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

---

Data Download Date/Time: 06/18/2018 18:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVA,EMILIANO,R  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: SILVA,EMILIANO  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/05/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: MCINTOSH

DATE/TIME OF TEST: 06/17/2018 05:19:52

---

Diagnostics Check	OK	05:16	
Air Blank	0.000	05:17	
Control Test	0.079	05:17	
Air Blank	0.000	05:17	
Subject Sample	# 1	05:18	Subject Test Refused
Breath Volume	0.000	05:18	
Air Blank	0.000	05:18	
Control Test	0.078	05:19	
Air Blank	0.000	05:19	
Diagnostics Check	OK	05:19	

---

Data Download Date/Time: 06/18/2018 18:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SAYERS,MICHELLE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: VIRUET

DATE/TIME OF TEST: 06/19/2018 01:57:03

---

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.000	01:52
Breath Volume	2.804	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.000	01:55
Breath Volume	2.230	01:55
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:57

---

Data Download Date/Time: 07/02/2018 21:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLLIER,THOMAS,L  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: MARCELLA,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 08:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:41

SUBJECT'S LAST NAME: GARCIA CARRENO

DATE/TIME OF TEST: 06/20/2018 10:35:19

---

Diagnostics Check	OK	10:29
Air Blank	0.000	10:29
Control Test	0.079	10:29
Air Blank	0.000	10:30
Subject Sample # 1	0.200	10:30
Breath Volume	3.656	10:30
Air Blank	0.000	10:31
Air Blank	0.000	10:33
Subject Sample # 2	0.203	10:33
Breath Volume	3.914	10:33
Air Blank	0.000	10:34
Control Test	0.077	10:34
Air Blank	0.000	10:35
Diagnostics Check	OK	10:35

---

Data Download Date/Time: 07/02/2018 21:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: GARRETT,G  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: VELAZQUEZ-ROSADO

DATE/TIME OF TEST: 06/24/2018 03:08:05

---

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.078	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.105	03:03
Breath Volume	1.734	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.101	03:06
Breath Volume	1.773	03:06
Air Blank	0.000	03:07
Control Test	0.077	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

---

Data Download Date/Time: 07/02/2018 21:51



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: MOHABIR

DATE/TIME OF TEST: 06/25/2018 22:42:30

---

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.078	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.158	22:37
Breath Volume	1.457	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.144	22:40
Breath Volume	1.546	22:40
Air Blank	0.000	22:41
Control Test	0.077	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

---

Data Download Date/Time: 07/02/2018 21:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SAYERS,MICHELLE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: SANTORIELLO

DATE/TIME OF TEST: 06/27/2018 21:15:50

---

Diagnostics Check	OK	21:09
Air Blank	0.000	21:09
Control Test	0.079	21:10
Air Blank	0.000	21:10
Subject Sample # 1	0.123	21:11
Breath Volume	2.042	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:13
Subject Sample # 2	0.126	21:14
Breath Volume	1.855	21:14
Air Blank	0.000	21:14
Control Test	0.078	21:15
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

---

Data Download Date/Time: 07/02/2018 21:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYONS,DANIEL,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: MENDELSON,GARRETT  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 06/30/2018 22:22:24

---

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.079	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.168	22:17
Breath Volume	2.398	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.154	22:20
Breath Volume	1.859	22:20
Air Blank	0.000	22:21
Control Test	0.079	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

---

Data Download Date/Time: 07/02/2018 21:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHEINER,JAMES,R  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: SCHEINER,JAMES  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: HUFFAKER

DATE/TIME OF TEST: 06/17/2018 19:37:01

---

Diagnostics Check	OK	19:30
Air Blank	0.000	19:31
Control Test	0.078	19:31
Air Blank	0.000	19:32
Subject Sample # 1	0.279	19:32
Breath Volume	2.101	19:32
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample # 2	0.289	19:35
Breath Volume	2.691	19:35
Air Blank	0.000	19:36
Control Test	0.078	19:36
Air Blank	0.000	19:36
Diagnostics Check	OK	19:37

---

Data Download Date/Time: 06/26/2018 16:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: THOMPSON,BILLY  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: GARREN

DATE/TIME OF TEST: 06/02/2018 22:20:28

---

Diagnostics Check	OK	22:11
Air Blank	0.000	22:12
Control Test	0.081	22:12
Air Blank	0.000	22:12
Subject Sample # 1	0.229	22:14
Breath Volume	4.640	22:14
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.223	22:18
Breath Volume	5.398	22:18
Air Blank	0.000	22:19
Control Test	0.081	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

---

Data Download Date/Time: 06/29/2018 20:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEDERMAN,CASEY,L  
BREATH TEST OPERATOR AGENCY: Titusville PD  
ARRESTING OFFICER: DEVOSS,EZEKIEL  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: EASLEY

DATE/TIME OF TEST: 06/10/2018 03:14:35

---

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.209	03:09
Breath Volume	1.253	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.207	03:12
Breath Volume	1.609	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

---

Data Download Date/Time: 06/29/2018 20:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J  
BREATH TEST OPERATOR AGENCY: Titusville PD  
ARRESTING OFFICER: WEAVER,JACOB  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: SHEPHEARD

DATE/TIME OF TEST: 06/14/2018 12:54:06

---

Diagnostics Check	OK	12:47
Air Blank	0.000	12:47
Control Test	0.080	12:47
Air Blank	0.000	12:48
Subject Sample # 1	0.000	12:49
Breath Volume	1.250	12:49
Air Blank	0.000	12:49
Air Blank	0.000	12:51
Subject Sample # 2	0.000	12:52
Breath Volume	1.507	12:52
Air Blank	0.000	12:53
Control Test	0.081	12:53
Air Blank	0.000	12:54
Diagnostics Check	OK	12:54

---

Data Download Date/Time: 06/29/2018 20:18