Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TOZZI,SHAUN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 06/30/2018 22:54:40

Diagnostics Check		OK	22:47
Air Blank		0.000	22:48
Control Test		0.078	22:48
Air Blank		0.000	22:48
Subject Sample	# 1	0.215	22:49
Breath Volume		1.304	22:49
Air Blank		0.000	22:50
Air Blank		0.000	22:52
Subject Sample	# 2	0.223	22:53
Breath Volume		1.468	22:53
Air Blank		0.000	22:53
Control Test		0.080	22:54
Air Blank		0.000	22:54
Diagnostics Check		OK	22:54

Data Download Date/Time: 07/01/2018 21:24

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REHDER,PATRICK,W
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: REZNY,LAUREN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 05/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: GASIER

DATE/TIME OF TEST: 06/18/2018 20:20:34

Diagnostics Check		OK	20:14
Air Blank		0.000	20:14
Control Test		0.079	20:15
Air Blank		0.000	20:15
Subject Sample	# 1	0.000	20:16
Breath Volume		3.570	20:16
Air Blank		0.000	20:16
Air Blank		0.000	20:18
Subject Sample	# 2	0.000	20:19
Breath Volume		2.691	20:19
Air Blank		0.000	20:19
Control Test		0.080	20:19
Air Blank		0.000	20:20
Diagnostics Check		OK	20:20

Data Download Date/Time: 07/01/2018 21:48

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEADOWS,SHAWN,M
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MEADOWS,SHAWN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: BORKOWSKI

DATE/TIME OF TEST: 06/26/2018 20:10:48

Diagnostics Check		OK	20:04
Air Blank		0.000	20:04
Control Test		0.078	20:05
Air Blank		0.000	20:05
Subject Sample	# 1	0.206	20:06
Breath Volume		1.355	20:06
Air Blank		0.000	20:06
Air Blank		0.000	20:08
Subject Sample	# 2	0.221	20:09
Breath Volume		2.433	20:09
Air Blank		0.000	20:09
Control Test		0.079	20:10
Air Blank		0.000	20:10
Diagnostics Check		OK	20:10

Data Download Date/Time: 07/01/2018 21:48

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KINKOPF,TIMOTHY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/29/2018 22:58:48

Diagnostics Check	OK	22:57	
Air Blank	0.000	22:57	
Control Test	0.101	22:58	Control Outside Tolerance
Air Blank	0.000	22:58	

Data Download Date/Time: 07/01/2018 21:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KINKOPF,TIMOTHY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/29/2018 23:06:05

Diagnostics Check		OK	23:00
Air Blank		0.000	23:00
Control Test		0.079	23:00
Air Blank		0.000	23:01
Subject Sample	# 1	0.135	23:01
Breath Volume		1.941	23:01
Air Blank		0.000	23:02
Air Blank		0.000	23:04
Subject Sample	# 2	0.142	23:04
Breath Volume		2.214	23:04
Air Blank		0.000	23:05
Control Test		0.079	23:05
Air Blank		0.000	23:06
Diagnostics Check		OK	23:06

Data Download Date/Time: 07/01/2018 21:48

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: HERNANDEZ,L
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 05/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: LAWHORN

DATE/TIME OF TEST: 06/22/2018 22:49:35

Diagnostics Check		OK	22:43
Air Blank		0.000	22:43
Control Test		0.078	22:44
Air Blank		0.000	22:44
Subject Sample	# 1	0.088	22:45
Breath Volume		1.406	22:45
Air Blank		0.000	22:45
Air Blank		0.000	22:47
Subject Sample	# 2	0.080	22:48
Breath Volume		1.523	22:48
Air Blank		0.000	22:48
Control Test		0.077	22:49
Air Blank		0.000	22:49
Diagnostics Check		OK	22:49

Data Download Date/Time: 06/26/2018 21:54

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: GONZALES,ROMAN
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: DAVIES

DATE/TIME OF TEST: 06/22/2018 23:22:06

Diagnostics Check		OK	23:19	
Air Blank		0.000	23:19	
Control Test		0.079	23:20	
Air Blank		0.000	23:20	
Subject Sample	# 1		23:20	Subject Test Refused
Breath Volume		0.000	23:20	
Air Blank		0.000	23:21	
Control Test		0.078	23:21	
Air Blank		0.000	23:22	
Diagnostics Check		OK	23:22	

Data Download Date/Time: 06/26/2018 21:54

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT,MATTHEW,G
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: BENNETT,MATTHEW
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 05/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/03/2018 01:50:43

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.076	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.222	01:45
Breath Volume		1.941	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.217	01:49
Breath Volume		2.449	01:49
Air Blank		0.000	01:49
Control Test		0.075	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 06/29/2018 15:17

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT,MATTHEW,G
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: BENNETT,MATTHEW
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 05/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DEVOLL

DATE/TIME OF TEST: 06/17/2018 02:58:45

Diagnostics Check		OK	02:51
Air Blank		0.000	02:52
Control Test		0.078	02:52
Air Blank		0.000	02:52
Subject Sample	# 1	0.185	02:54
Breath Volume		1.410	02:54
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.187	02:57
Breath Volume		1.828	02:57
Air Blank		0.000	02:57
Control Test		0.076	02:58
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 06/29/2018 15:17

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSMAN,JOHN,D
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: HOUSMAN,JOHN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 05/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/19/2018 00:03:28

Diagnostics Check		OK	23:57
Air Blank		0.000	23:57
Control Test		0.078	23:58
Air Blank		0.000	23:58
Subject Sample	# 2	0.000	23:59
Breath Volume		1.597	23:59
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.000	00:02
Breath Volume		1.628	00:02
Air Blank		0.000	00:02
Control Test		0.077	00:02
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 06/29/2018 15:17

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDA,ROBERT,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: ZUNIGA,MANUEL
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: NEELEY

DATE/TIME OF TEST: 06/09/2018 22:56:19

Diagnostics Check		OK	22:49
Air Blank		0.000	22:50
Control Test		0.081	22:50
Air Blank		0.000	22:50
Subject Sample	# 1	0.000	22:51
Breath Volume		1.117	22:51
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.000	22:54
Breath Volume		1.121	22:54
Air Blank		0.000	22:55
Control Test		0.082	22:55
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 06/26/2018 16:13

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,JOHN,T
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: DERAVIL,JEAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/30/2018 21:53:45

Diagnostics Check		OK	21:50	
Air Blank		0.000	21:51	
Control Test		0.082	21:51	
Air Blank		0.000	21:52	
Subject Sample	# 1		21:52	Subject Test Refused
Breath Volume		0.000	21:52	
Air Blank		0.000	21:52	
Control Test		0.082	21:53	
Air Blank		0.000	21:53	
Diagnostics Check		OK	21:53	

Data Download Date/Time: 07/29/2018 13:49

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SZYMANSKI,JACEK,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: OREILLY,TERRENCE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: FULKMAN

DATE/TIME OF TEST: 06/01/2018 00:07:05

Diagnostics Check		OK	00:04	
Air Blank		0.000	00:04	
Control Test		0.079	00:05	
Air Blank		0.000	00:05	
Subject Sample	# 1		00:05	Subject Test Refused
Breath Volume		0.000	00:05	
Air Blank		0.000	00:06	
Control Test		0.079	00:06	
Air Blank		0.000	00:06	
Diagnostics Check		OK	00:07	

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: TRICOU

DATE/TIME OF TEST: 06/01/2018 03:47:45

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.079	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.201	03:42
Breath Volume		1.394	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.204	03:46
Breath Volume		2.277	03:46
Air Blank		0.000	03:46
Control Test		0.078	03:47
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: YANEZ MARADIAGA

DATE/TIME OF TEST: 06/03/2018 03:12:47

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:10	
Control Test		0.079	03:10	
Air Blank		0.000	03:11	
Subject Sample	# 1		03:11	Subject Test Refused
Breath Volume		0.000	03:11	
Air Blank		0.000	03:11	
Control Test		0.079	03:12	
Air Blank		0.000	03:12	
Diagnostics Check		OK	03:12	

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 06/03/2018 04:07:28

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.080	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.191	04:03
Breath Volume		2.046	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.187	04:05
Breath Volume		1.242	04:05
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HEATON,RICHARD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: SELLAR

DATE/TIME OF TEST: 06/03/2018 23:06:25

Diagnostics Check		OK	23:03	
Air Blank		0.000	23:03	
Control Test		0.079	23:04	
Air Blank		0.000	23:04	
Subject Sample	# 1		23:05	Subject Test Refused
Breath Volume		0.000	23:05	
Air Blank		0.000	23:05	
Control Test		0.079	23:05	
Air Blank		0.000	23:06	
Diagnostics Check		OK	23:06	

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,SCOTT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRILL,SHANNON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 07:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:04

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 06/05/2018 08:35:06

Diagnostics Check		OK	08:28
Air Blank		0.000	08:29
Control Test		0.079	08:29
Air Blank		0.000	08:29
Subject Sample	# 1	0.000	08:30
Breath Volume		1.457	08:30
Air Blank		0.000	08:30
Air Blank		0.000	08:32
Subject Sample	# 2	0.000	08:33
Breath Volume		1.488	08:33
Air Blank		0.000	08:34
Control Test		0.080	08:34
Air Blank		0.000	08:35
Diagnostics Check		OK	08:35

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: SCHWETER

DATE/TIME OF TEST: 06/06/2018 23:57:08

Diagnostics Check		OK	23:54	
Air Blank		0.000	23:54	
Control Test		0.080	23:55	
Air Blank		0.000	23:55	
Subject Sample	# 1		23:55	Subject Test Refused
Breath Volume		0.000	23:55	
Air Blank		0.000	23:56	
Control Test		0.080	23:56	
Air Blank		0.000	23:57	
Diagnostics Check		OK	23:57	

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CAIN

DATE/TIME OF TEST: 06/07/2018 02:18:33

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:15	
Control Test		0.080	02:16	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:17	Subject Test Refused
Breath Volume		0.000	02:17	
Air Blank		0.000	02:17	
Control Test		0.079	02:17	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: GABRICI

DATE/TIME OF TEST: 06/07/2018 03:54:36

Diagnostics Check		OK	03:51	
Air Blank		0.000	03:52	
Control Test		0.079	03:52	
Air Blank		0.000	03:53	
Subject Sample	# 1		03:53	Subject Test Refused
Breath Volume		0.000	03:53	
Air Blank		0.000	03:53	
Control Test		0.080	03:54	
Air Blank		0.000	03:54	
Diagnostics Check		OK	03:54	

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/08/2018 04:51:49

Diagnostics Check		OK	04:41
Air Blank		0.000	04:42
Control Test		0.079	04:42
Air Blank		0.000	04:42
Subject Sample	# 1	0.222	04:43
Breath Volume		2.089	04:43
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.195	04:46
Breath Volume		1.753	04:46
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 3	0.209	04:50
Breath Volume		1.921	04:50
Air Blank		0.000	04:50
Control Test		0.078	04:51
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 06/09/2018 00:18:17

Diagnostics Check		OK	00:12
Air Blank		0.000	00:12
Control Test		0.078	00:13
Air Blank		0.000	00:13
Subject Sample	# 1	0.146	00:13
Breath Volume		2.925	00:13
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.141	00:16
Breath Volume		2.253	00:16
Air Blank		0.000	00:17
Control Test		0.078	00:17
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: SANCHEZ-VELAZQUEZ

DATE/TIME OF TEST: 06/09/2018 22:38:03

Diagnostics Check		OK	22:26	
Air Blank		0.000	22:27	
Control Test		0.079	22:27	
Air Blank		0.000	22:27	
Subject Sample	# 1		22:31	No Sample Provided
Breath Volume		0.000	22:31	
Air Blank		0.000	22:31	
Air Blank		0.000	22:33	
Subject Sample	# 2		22:36	No Sample Provided
Breath Volume		0.000	22:36	
Air Blank		0.000	22:37	
Control Test		0.079	22:37	
Air Blank		0.000	22:37	
Diagnostics Check		OK	22:38	

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: LANCIA

DATE/TIME OF TEST: 06/10/2018 03:31:26

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:18	
Control Test		0.077	03:19	
Air Blank		0.000	03:19	
Subject Sample	# 1		03:22	No Sample Provided
Breath Volume		0.000	03:22	
Air Blank		0.000	03:23	
Air Blank		0.000	03:25	
Subject Sample	# 2	0.214	03:26	
Breath Volume		1.378	03:26	
Air Blank		0.000	03:26	
Air Blank		0.000	03:28	
Subject Sample	# 3	0.215	03:29	
Breath Volume		2.195	03:29	
Air Blank		0.000	03:30	
Control Test		0.078	03:30	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SZIMA

DATE/TIME OF TEST: 06/10/2018 04:05:52

Diagnostics Check		OK	03:51	
Air Blank		0.000	03:52	
Control Test		0.078	03:52	
Air Blank		0.000	03:53	
Subject Sample	# 1		03:56	No Sample Provided
Breath Volume		0.000	03:56	
Air Blank		0.000	03:56	
Air Blank		0.000	03:58	
Subject Sample	# 2	0.214	04:00	
Breath Volume		1.421	04:00	
Air Blank		0.000	04:01	
Air Blank		0.000	04:03	
Subject Sample	# 3	0.219	04:04	
Breath Volume		2.121	04:04	
Air Blank		0.000	04:04	
Control Test		0.077	04:05	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:05	

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: DENAMEN

DATE/TIME OF TEST: 06/12/2018 23:28:40

Diagnostics Check		OK	23:21
Air Blank		0.000	23:22
Control Test		0.079	23:22
Air Blank		0.000	23:22
Subject Sample	# 1	0.132	23:24
Breath Volume		2.058	23:24
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.145	23:27
Breath Volume		3.085	23:27
Air Blank		0.000	23:27
Control Test		0.080	23:28
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,M,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CONNOR,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: DUMOUCHEL

DATE/TIME OF TEST: 06/13/2018 18:05:51

Diagnostics Check		OK	17:59
Air Blank		0.000	17:59
Control Test		0.079	18:00
Air Blank		0.000	18:00
Subject Sample	# 1	0.061	18:01
Breath Volume		1.679	18:01
Air Blank		0.000	18:02
Air Blank		0.000	18:03
Subject Sample	# 2	0.058	18:04
Breath Volume		1.226	18:04
Air Blank		0.000	18:04
Control Test		0.079	18:05
Air Blank		0.000	18:05
Diagnostics Check		OK	18:05

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/13/2018 22:32:22

Diagnostics Check		OK	22:29	
Air Blank		0.000	22:29	
Control Test		0.078	22:30	
Air Blank		0.000	22:30	
Subject Sample	# 1		22:31	Subject Test Refused
Breath Volume		0.000	22:31	
Air Blank		0.000	22:31	
Control Test		0.079	22:31	
Air Blank		0.000	22:32	
Diagnostics Check		OK	22:32	

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GABRILLO

DATE/TIME OF TEST: 06/15/2018 03:32:12

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.079	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.000	03:27
Breath Volume		2.464	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.000	03:30
Breath Volume		1.558	03:30
Air Blank		0.000	03:31
Control Test		0.079	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 06/16/2018 02:16:09

Diagnostics Check		OK	02:09
Air Blank		0.000	02:10
Control Test		0.078	02:10
Air Blank		0.000	02:11
Subject Sample	# 1	0.334	02:11
Breath Volume		1.476	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.352	02:14
Breath Volume		1.500	02:14
Air Blank		0.000	02:15
Control Test		0.079	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: BLEVINS

DATE/TIME OF TEST: 06/16/2018 04:03:54

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.144	03:59
Breath Volume		3.152	03:59
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.136	04:02
Breath Volume		2.402	04:02
Air Blank		0.000	04:02
Control Test		0.079	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: FERBEE

DATE/TIME OF TEST: 06/16/2018 04:34:34

Diagnostics Check		OK	04:28
Air Blank		0.000	04:28
Control Test		0.079	04:29
Air Blank		0.000	04:29
Subject Sample	# 1	0.045	04:30
Breath Volume		2.324	04:30
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.043	04:33
Breath Volume		1.496	04:33
Air Blank		0.000	04:33
Control Test		0.079	04:33
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CONSTANTINE

DATE/TIME OF TEST: 06/17/2018 02:40:18

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.078	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.195	02:35
Breath Volume		1.914	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.182	02:38
Breath Volume		1.433	02:38
Air Blank		0.000	02:39
Control Test		0.078	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 06/17/2018 03:59:20

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:56	
Control Test		0.079	03:57	
Air Blank		0.000	03:57	
Subject Sample	# 1		03:57	Subject Test Refused
Breath Volume		0.000	03:57	
Air Blank		0.000	03:58	
Control Test		0.079	03:58	
Air Blank		0.000	03:59	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SZYMANSKI,JACEK,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DONOHUE,STEVEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: BEN MIM

DATE/TIME OF TEST: 06/19/2018 03:31:22

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:29	
Control Test		0.079	03:29	
Air Blank		0.000	03:29	
Subject Sample	# 1		03:30	Subject Test Refused
Breath Volume		0.000	03:30	
Air Blank		0.000	03:30	
Control Test		0.079	03:30	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE,RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE,RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 06/21/2018 01:29:00

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:16	
Control Test		0.079	01:16	
Air Blank		0.000	01:16	
Subject Sample	# 1	0.099	01:20	Volume Not Met
Breath Volume		0.457	01:20	
Air Blank		0.000	01:20	
Air Blank		0.000	01:22	
Subject Sample	# 2	0.114	01:24	
Breath Volume		1.375	01:24	
Air Blank		0.000	01:25	
Air Blank		0.000	01:26	
Subject Sample	# 3	0.113	01:27	
Breath Volume		1.226	01:27	
Air Blank		0.000	01:28	
Control Test		0.078	01:28	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:29	

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: HERNANDEZ LOPEZ

DATE/TIME OF TEST: 06/21/2018 03:31:03

Diagnostics Check		OK	03:24
Air Blank		0.000	03:24
Control Test		0.080	03:24
Air Blank		0.000	03:25
Subject Sample	# 1	0.188	03:26
Breath Volume		3.480	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.189	03:29
Breath Volume		3.187	03:29
Air Blank		0.000	03:30
Control Test		0.079	03:30
Air Blank		0.000	03:30
Diagnostics Check		OK	03:31

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: DASSATT

DATE/TIME OF TEST: 06/21/2018 04:26:44

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.079	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.170	04:22
Breath Volume		2.210	04:22
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.161	04:25
Breath Volume		2.000	04:25
Air Blank		0.000	04:25
Control Test		0.079	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/21/2018 20:37:14

Diagnostics Check		OK	20:30
Air Blank		0.000	20:31
Control Test		0.080	20:31
Air Blank		0.000	20:32
Subject Sample	# 1	0.112	20:32
Breath Volume		1.410	20:32
Air Blank		0.000	20:33
Air Blank		0.000	20:35
Subject Sample	# 2	0.113	20:35
Breath Volume		1.359	20:35
Air Blank		0.000	20:36
Control Test		0.080	20:36
Air Blank		0.000	20:37
Diagnostics Check		OK	20:37

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: CARABELAS

DATE/TIME OF TEST: 06/22/2018 00:50:35

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.030	00:46
Breath Volume		1.542	00:46
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.030	00:49
Breath Volume		1.714	00:49
Air Blank		0.000	00:49
Control Test		0.078	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: ECKLUND

DATE/TIME OF TEST: 06/22/2018 03:12:33

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:56	
Control Test		0.080	02:57	
Air Blank		0.000	02:57	
Subject Sample	# 1	0.042	02:59	
Breath Volume		1.367	02:59	
Air Blank		0.000	03:00	
Air Blank		0.000	03:02	
Subject Sample	# 2		03:05	No Sample Provided
Breath Volume		0.000	03:05	
Air Blank		0.000	03:06	
Air Blank		0.000	03:08	
Subject Sample	# 3		03:11	No Sample Provided
Breath Volume		0.000	03:11	
Air Blank		0.000	03:11	
Control Test		0.080	03:11	
Air Blank		0.000	03:12	
Diagnostics Check		OK	03:12	

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: CRAWFORD-ROGERS

DATE/TIME OF TEST: 06/22/2018 03:58:44

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.111	03:53
Breath Volume		1.468	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.127	03:57
Breath Volume		2.066	03:57
Air Blank		0.000	03:57
Control Test		0.080	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: GATTO

DATE/TIME OF TEST: 06/22/2018 23:14:45

Diagnostics Check		OK	23:08
Air Blank		0.000	23:08
Control Test		0.080	23:09
Air Blank		0.000	23:09
Subject Sample	# 1	0.137	23:10
Breath Volume		1.738	23:10
Air Blank		0.000	23:10
Air Blank		0.000	23:12
Subject Sample	# 2	0.140	23:13
Breath Volume		1.414	23:13
Air Blank		0.000	23:13
Control Test		0.080	23:14
Air Blank		0.000	23:14
Diagnostics Check		OK	23:14

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GIBSON,DARREN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/25/2018 18:08:58

Diagnostics Check		OK	18:06	
Air Blank		0.000	18:06	
Control Test		0.079	18:06	
Air Blank		0.000	18:07	
Subject Sample	# 1		18:07	Subject Test Refused
Breath Volume		0.000	18:07	
Air Blank		0.000	18:08	
Control Test		0.079	18:08	
Air Blank		0.000	18:08	
Diagnostics Check		OK	18:08	

Data Download Date/Time: 06/28/2018 23:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMITH,JEREMY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 06/26/2018 02:15:55

Diagnostics Check		OK	02:04	
Air Blank		0.000	02:04	
Control Test		0.079	02:05	
Air Blank		0.000	02:05	
Subject Sample	# 1		02:08	No Sample Provided
Breath Volume		0.000	02:08	
Air Blank		0.000	02:09	
Air Blank		0.000	02:11	
Subject Sample	# 2		02:14	No Sample Provided
Breath Volume		0.000	02:14	
Air Blank		0.000	02:15	
Control Test		0.079	02:15	
Air Blank		0.000	02:15	
Diagnostics Check		OK	02:15	

Data Download Date/Time: 06/28/2018 23:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMITH,JEREMY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 06/26/2018 02:30:29

Diagnostics Check		OK	02:27	
Air Blank		0.000	02:27	
Control Test		0.079	02:28	
Air Blank		0.000	02:28	
Subject Sample	# 1		02:29	Subject Test Refused
Breath Volume		0.000	02:29	
Air Blank		0.000	02:29	
Control Test		0.079	02:29	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 06/28/2018 23:10

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 06/27/2018 19:49:37

Diagnostics Check		OK	19:43
Air Blank		0.000	19:44
Control Test		0.080	19:44
Air Blank		0.000	19:44
Subject Sample	# 1	0.000	19:45
Breath Volume		2.917	19:45
Air Blank		0.000	19:45
Air Blank		0.000	19:47
Subject Sample	# 2	0.000	19:48
Breath Volume		2.519	19:48
Air Blank		0.000	19:48
Control Test		0.080	19:49
Air Blank		0.000	19:49
Diagnostics Check		OK	19:49

Data Download Date/Time: 06/28/2018 23:10

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ANZALONE

DATE/TIME OF TEST: 06/28/2018 02:12:16

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.033	02:07
Breath Volume		1.710	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.033	02:10
Breath Volume		1.722	02:10
Air Blank		0.000	02:11
Control Test		0.080	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 06/28/2018 23:10

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: VITALE

DATE/TIME OF TEST: 06/28/2018 03:03:02

Diagnostics Check		OK	03:00	
Air Blank		0.000	03:00	
Control Test		0.079	03:01	
Air Blank		0.000	03:01	
Subject Sample	# 1		03:01	Subject Test Refused
Breath Volume		0.000	03:01	
Air Blank		0.000	03:02	
Control Test		0.079	03:02	
Air Blank		0.000	03:02	
Diagnostics Check		OK	03:03	

Data Download Date/Time: 06/28/2018 23:10

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/29/2018 03:34:10

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:22	
Control Test		0.079	03:23	
Air Blank		0.000	03:23	
Subject Sample	# 1	0.181	03:26	Volume Not Met
Breath Volume		0.105	03:26	
Air Blank		0.000	03:27	
Air Blank		0.000	03:29	
Subject Sample	# 2	0.178	03:32	Volume Not Met
Breath Volume		0.015	03:32	
Air Blank		0.000	03:33	
Control Test		0.080	03:33	
Air Blank		0.000	03:34	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 07/25/2018 23:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/29/2018 03:48:08

Diagnostics Check		OK	03:36	
Air Blank		0.000	03:37	
Control Test		0.078	03:37	
Air Blank		0.000	03:37	
Subject Sample	# 1	0.200	03:41	Volume Not Met
Breath Volume		0.742	03:41	
Air Blank		0.000	03:41	
Air Blank		0.000	03:43	
Subject Sample	# 2		03:46	Subject Test Refused
Breath Volume		0.312	03:46	
Air Blank		0.000	03:47	
Control Test		0.078	03:47	
Air Blank		0.000	03:48	
Diagnostics Check		OK	03:48	

Data Download Date/Time: 07/25/2018 23:21

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: RINGWOOD

DATE/TIME OF TEST: 06/29/2018 04:19:48

Diagnostics Check		OK	04:08	
Air Blank		0.000	04:08	
Control Test		0.078	04:09	
Air Blank		0.000	04:09	
Subject Sample	# 1		04:12	No Sample Provided
Breath Volume		0.000	04:12	
Air Blank		0.000	04:13	
Air Blank		0.000	04:15	
Subject Sample	# 2		04:18	No Sample Provided
Breath Volume		0.000	04:18	
Air Blank		0.000	04:18	
Control Test		0.079	04:19	
Air Blank		0.000	04:19	
Diagnostics Check		OK	04:19	

Data Download Date/Time: 07/25/2018 23:21

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: BROADWELL

DATE/TIME OF TEST: 06/30/2018 05:04:10

Diagnostics Check		OK	04:57
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Subject Sample	# 1	0.000	04:59
Breath Volume		2.218	04:59
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.000	05:02
Breath Volume		2.277	05:02
Air Blank		0.000	05:03
Control Test		0.079	05:03
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 07/25/2018 23:21

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,STEPHAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLE,STEPHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: MOISUC

DATE/TIME OF TEST: 06/01/2018 20:19:29

Diagnostics Check		OK	20:13
Air Blank		0.000	20:13
Control Test		0.077	20:13
Air Blank		0.000	20:14
Subject Sample	# 1	0.197	20:14
Breath Volume		3.898	20:14
Air Blank		0.000	20:15
Air Blank		0.000	20:17
Subject Sample	# 2	0.201	20:17
Breath Volume		2.871	20:17
Air Blank		0.000	20:18
Control Test		0.077	20:18
Air Blank		0.000	20:19
Diagnostics Check		OK	20:19

Data Download Date/Time: 06/28/2018 20:14

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUSCH,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 14:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: LABORDE

DATE/TIME OF TEST: 06/02/2018 16:18:02

Diagnostics Check		OK	16:11
Air Blank		0.000	16:11
Control Test		0.079	16:11
Air Blank		0.000	16:12
Subject Sample	# 1	0.000	16:13
Breath Volume		2.222	16:13
Air Blank		0.000	16:13
Air Blank		0.000	16:15
Subject Sample	# 2	0.000	16:16
Breath Volume		1.574	16:16
Air Blank		0.000	16:17
Control Test		0.078	16:17
Air Blank		0.000	16:17
Diagnostics Check		OK	16:18

Data Download Date/Time: 06/28/2018 20:14

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,STEPHAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLE,STEPHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: BARKLEY

DATE/TIME OF TEST: 06/04/2018 05:07:47

Diagnostics Check		OK	05:02
Air Blank		0.000	05:02
Control Test		0.078	05:02
Air Blank		0.000	05:03
Subject Sample	# 1	0.025	05:03
Breath Volume		1.316	05:03
Air Blank		0.000	05:04
Air Blank		0.000	05:06
Subject Sample	# 2	0.024	05:06
Breath Volume		1.613	05:06
Air Blank		0.000	05:06
Control Test		0.078	05:07
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 06/28/2018 20:14

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,JESSICA,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FRASER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: PESSAS

DATE/TIME OF TEST: 06/13/2018 03:58:39

Diagnostics Check	OK	03:57	
Air Blank		03:58	RFI Detect
Air Blank	0.000	03:58	

Data Download Date/Time: 06/28/2018 20:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,JESSICA,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FRASER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: PESSAS

DATE/TIME OF TEST: 06/13/2018 04:07:29

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.078	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.155	04:02
Breath Volume		1.523	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.167	04:05
Breath Volume		1.238	04:05
Air Blank		0.000	04:06
Control Test		0.076	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 06/28/2018 20:14

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,JESSICA,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SKALKO,RYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 06/17/2018 23:32:14

Diagnostics Check		OK	23:29	
Air Blank		0.000	23:29	
Control Test		0.078	23:29	
Air Blank		0.000	23:30	
Subject Sample	# 1		23:30	Subject Test Refused
Breath Volume		0.000	23:30	
Air Blank		0.000	23:31	
Control Test		0.077	23:31	
Air Blank		0.000	23:32	
Diagnostics Check		OK	23:32	

Data Download Date/Time: 06/28/2018 20:14

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE JESUS,JULIA,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DE JESUS,JULIA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 08:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: KATAMAY

DATE/TIME OF TEST: 06/21/2018 09:38:21

Diagnostics Check		OK	09:35	
Air Blank		0.000	09:35	
Control Test		0.078	09:36	
Air Blank		0.000	09:36	
Subject Sample	# 1		09:36	Subject Test Refused
Breath Volume		0.000	09:36	
Air Blank		0.000	09:37	
Control Test		0.078	09:37	
Air Blank		0.000	09:38	
Diagnostics Check		OK	09:38	

Data Download Date/Time: 06/28/2018 20:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,GABRIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLE,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/27/2018 02:13:52

Diagnostics Check		OK	02:02	
Air Blank		0.000	02:02	
Control Test		0.077	02:03	
Air Blank		0.000	02:03	
Subject Sample	# 1		02:06	No Sample Provided
Breath Volume		0.000	02:06	
Air Blank		0.000	02:07	
Air Blank		0.000	02:09	
Subject Sample	# 2		02:12	No Sample Provided
Breath Volume		0.000	02:12	
Air Blank		0.000	02:12	
Control Test		0.078	02:13	
Air Blank		0.000	02:13	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 06/28/2018 20:14

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: AL-NASSER

DATE/TIME OF TEST: 06/28/2018 22:45:05

Diagnostics Check		OK	22:42	
Air Blank		0.000	22:42	
Control Test		0.078	22:43	
Air Blank		0.000	22:43	
Subject Sample	# 1		22:43	Subject Test Refused
Breath Volume		0.000	22:43	
Air Blank		0.000	22:44	
Control Test		0.078	22:44	
Air Blank		0.000	22:45	
Diagnostics Check		OK	22:45	

Data Download Date/Time: 07/25/2018 22:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 10:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: MCGLOWN

DATE/TIME OF TEST: 06/14/2018 11:29:15

Diagnostics Check		OK	11:23
Air Blank		0.000	11:23
Control Test		0.081	11:24
Air Blank		0.000	11:24
Subject Sample	# 1	0.000	11:25
Breath Volume		1.808	11:25
Air Blank		0.000	11:25
Air Blank		0.000	11:27
Subject Sample	# 2	0.000	11:27
Breath Volume		1.765	11:27
Air Blank		0.000	11:28
Control Test		0.081	11:28
Air Blank		0.000	11:29
Diagnostics Check		OK	11:29

Data Download Date/Time: 07/08/2018 19:15

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,WILLIAM,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TUCKER,WILLIAM
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 06/16/2018 03:59:56

Diagnostics Check		OK	03:53
Air Blank		0.000	03:53
Control Test		0.082	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.061	03:55
Breath Volume		2.906	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:57
Subject Sample	# 2	0.063	03:58
Breath Volume		3.093	03:58
Air Blank		0.000	03:59
Control Test		0.082	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 07/08/2018 19:15

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBLINO,NICK,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ARKUSZESKI,DAVID
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: ZAMOR

DATE/TIME OF TEST: 06/22/2018 00:48:45

Diagnostics Check		OK	00:45	
Air Blank		0.000	00:46	
Control Test		0.081	00:46	
Air Blank		0.000	00:47	
Subject Sample	# 1		00:47	Subject Test Refused
Breath Volume		0.000	00:47	
Air Blank		0.000	00:47	
Control Test		0.083	00:48	
Air Blank		0.000	00:48	
Diagnostics Check		OK	00:48	

Data Download Date/Time: 07/08/2018 19:15

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,WILLIAM,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TUCKER,WILLIAM
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: ACUNA-POPOCA

DATE/TIME OF TEST: 06/30/2018 05:23:47

Diagnostics Check		OK	05:17
Air Blank		0.000	05:17
Control Test		0.081	05:17
Air Blank		0.000	05:18
Subject Sample	# 1	0.175	05:18
Breath Volume		1.421	05:18
Air Blank		0.000	05:19
Air Blank		0.000	05:21
Subject Sample	# 2	0.167	05:22
Breath Volume		1.410	05:22
Air Blank		0.000	05:22
Control Test		0.081	05:23
Air Blank		0.000	05:23
Diagnostics Check		OK	05:23

Data Download Date/Time: 07/08/2018 19:15

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBLINO,NICK,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ARKUSZESKI,DAVID
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: MORA

DATE/TIME OF TEST: 06/30/2018 05:33:08

Diagnostics Check		OK	05:26
Air Blank		0.000	05:27
Control Test		0.081	05:27
Air Blank		0.000	05:28
Subject Sample	# 1	0.137	05:28
Breath Volume		2.539	05:28
Air Blank		0.000	05:29
Air Blank		0.000	05:31
Subject Sample	# 2	0.126	05:31
Breath Volume		3.019	05:31
Air Blank		0.000	05:32
Control Test		0.080	05:32
Air Blank		0.000	05:33
Diagnostics Check		OK	05:33

Data Download Date/Time: 07/08/2018 19:15

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACEY,JOHN,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: LEONARD,BRIAN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: DELORME

DATE/TIME OF TEST: 06/02/2018 04:22:38

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.080	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.188	04:17
Breath Volume		4.398	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.183	04:21
Breath Volume		4.527	04:21
Air Blank		0.000	04:21
Control Test		0.078	04:22
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 07/16/2018 21:49

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,WILLIAM,L
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: HICKS,WILLIAM
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: DESMOND

DATE/TIME OF TEST: 06/05/2018 01:02:44

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.079	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.133	00:57
Breath Volume		1.968	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.152	01:01
Breath Volume		3.605	01:01
Air Blank		0.000	01:01
Control Test		0.079	01:02
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 07/16/2018 21:49

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACEY,JOHN,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: JONES,JOSEPH
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GRASSO

DATE/TIME OF TEST: 06/07/2018 00:56:41

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.000	00:52
Breath Volume		2.921	00:52
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.000	00:55
Breath Volume		2.582	00:55
Air Blank		0.000	00:55
Control Test		0.079	00:56
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 07/16/2018 21:49

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,WILLIAM,L
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: MILIAN,SHAWN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: ALBINI

DATE/TIME OF TEST: 06/09/2018 06:12:45

Diagnostics Check		OK	06:06
Air Blank		0.000	06:07
Control Test		0.080	06:07
Air Blank		0.000	06:07
Subject Sample	# 1	0.000	06:08
Breath Volume		3.925	06:08
Air Blank		0.000	06:08
Air Blank		0.000	06:10
Subject Sample	# 2	0.000	06:11
Breath Volume		4.074	06:11
Air Blank		0.000	06:11
Control Test		0.079	06:12
Air Blank		0.000	06:12
Diagnostics Check		OK	06:12

Data Download Date/Time: 07/16/2018 21:49

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACEY,JOHN,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: LEONARD,BRIAN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: WILLMAN

DATE/TIME OF TEST: 06/11/2018 20:21:03

Diagnostics Check		OK	20:15
Air Blank		0.000	20:15
Control Test		0.079	20:15
Air Blank		0.000	20:16
Subject Sample	# 1	0.104	20:16
Breath Volume		2.289	20:16
Air Blank		0.000	20:17
Air Blank		0.000	20:19
Subject Sample	# 2	0.106	20:19
Breath Volume		2.242	20:19
Air Blank		0.000	20:20
Control Test		0.078	20:20
Air Blank		0.000	20:20
Diagnostics Check		OK	20:21

Data Download Date/Time: 07/16/2018 21:49

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREY,ROBERT,P
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: BUCKLIN,S
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: WITHERS

DATE/TIME OF TEST: 06/02/2018 02:51:17

Diagnostics Check		OK	02:45
Air Blank		0.000	02:45
Control Test		0.080	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.000	02:47
Breath Volume		1.453	02:47
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.000	02:49
Breath Volume		1.453	02:49
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 06/27/2018 18:33

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/16/2018 05:26:12

Diagnostics Check		OK	05:19
Air Blank		0.000	05:20
Control Test		0.081	05:20
Air Blank		0.000	05:21
Subject Sample	# 1	0.225	05:21
Breath Volume		2.410	05:21
Air Blank		0.000	05:22
Air Blank		0.000	05:24
Subject Sample	# 2	0.212	05:24
Breath Volume		1.804	05:24
Air Blank		0.000	05:25
Control Test		0.079	05:25
Air Blank		0.000	05:26
Diagnostics Check		OK	05:26

Data Download Date/Time: 06/27/2018 18:33

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREY,ROBERT,P
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: VURRO,M
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SOYKA

DATE/TIME OF TEST: 06/29/2018 00:11:09

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.079	00:05
Air Blank		0.000	00:05
Subject Sample	# 1	0.000	00:06
Breath Volume		1.773	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.000	00:09
Breath Volume		2.265	00:09
Air Blank		0.000	00:10
Control Test		0.080	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 07/26/2018 19:41

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 06/08/2018 20:58:59

Diagnostics Check		OK	20:56	
Air Blank		0.000	20:56	
Control Test		0.078	20:56	
Air Blank		0.000	20:57	
Subject Sample	# 1		20:57	Subject Test Refused
Breath Volume		0.000	20:57	
Air Blank		0.000	20:58	
Control Test		0.078	20:58	
Air Blank		0.000	20:58	
Diagnostics Check		OK	20:58	

Data Download Date/Time: 06/20/2018 14:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JACOBSEN,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: DONOHUE

DATE/TIME OF TEST: 06/10/2018 06:11:28

Diagnostics Check		OK	06:04
Air Blank		0.000	06:05
Control Test		0.078	06:05
Air Blank		0.000	06:06
Subject Sample	# 1	0.135	06:07
Breath Volume		3.027	06:07
Air Blank		0.000	06:07
Air Blank		0.000	06:09
Subject Sample	# 2	0.138	06:09
Breath Volume		2.246	06:09
Air Blank		0.000	06:10
Control Test		0.082	06:10
Air Blank		0.000	06:11
Diagnostics Check		OK	06:11

Data Download Date/Time: 06/20/2018 14:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 06/12/2018 23:51:03

Diagnostics Check		OK	23:45
Air Blank		0.000	23:45
Control Test		0.079	23:45
Air Blank		0.000	23:46
Subject Sample	# 1	0.110	23:46
Breath Volume		2.328	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.106	23:49
Breath Volume		1.816	23:49
Air Blank		0.000	23:50
Control Test		0.081	23:50
Air Blank		0.000	23:50
Diagnostics Check		OK	23:51

Data Download Date/Time: 06/20/2018 14:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: SISSON

DATE/TIME OF TEST: 06/13/2018 18:27:30

Diagnostics Check		OK	18:15	
Air Blank		0.000	18:16	
Control Test		0.078	18:16	
Air Blank		0.000	18:17	
Subject Sample	# 1	0.000	18:17	
Breath Volume		3.792	18:17	
Air Blank		0.000	18:18	
Air Blank		0.000	18:20	
Subject Sample	# 2		18:23	No Sample Provided
Breath Volume		0.000	18:23	
Air Blank		0.000	18:23	
Air Blank		0.000	18:25	
Subject Sample	# 3	0.000	18:26	
Breath Volume		3.414	18:26	
Air Blank		0.000	18:26	
Control Test		0.082	18:26	Interferent Detect
Air Blank		0.000	18:27	

Data Download Date/Time: 06/20/2018 14:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: MCCABE

DATE/TIME OF TEST: 06/14/2018 16:59:24

Diagnostics Check		OK	16:53
Air Blank		0.000	16:54
Control Test		0.078	16:54
Air Blank		0.000	16:54
Subject Sample	# 1	0.000	16:55
Breath Volume		1.753	16:55
Air Blank		0.000	16:55
Air Blank		0.000	16:57
Subject Sample	# 2	0.000	16:58
Breath Volume		1.191	16:58
Air Blank		0.000	16:58
Control Test		0.078	16:58
Air Blank		0.000	16:59
Diagnostics Check		OK	16:59

Data Download Date/Time: 06/20/2018 14:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 06/15/2018 00:50:20

Diagnostics Check		OK	00:43
Air Blank		0.000	00:44
Control Test		0.078	00:44
Air Blank		0.000	00:45
Subject Sample	# 1	0.137	00:46
Breath Volume		2.980	00:46
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.129	00:48
Breath Volume		1.933	00:48
Air Blank		0.000	00:49
Control Test		0.082	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 06/20/2018 14:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: MOON

DATE/TIME OF TEST: 06/15/2018 05:00:07

Diagnostics Check		OK	04:54
Air Blank		0.000	04:54
Control Test		0.078	04:54
Air Blank		0.000	04:55
Subject Sample	# 1	0.128	04:55
Breath Volume		3.855	04:55
Air Blank		0.000	04:56
Air Blank		0.000	04:58
Subject Sample	# 2	0.124	04:58
Breath Volume		4.078	04:58
Air Blank		0.000	04:59
Control Test		0.083	04:59
Air Blank		0.000	05:00
Diagnostics Check		OK	05:00

Data Download Date/Time: 06/20/2018 14:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,BENJAMIN,F
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: MASTERS-SAMRETH,D
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: BAER

DATE/TIME OF TEST: 06/16/2018 08:16:18

Diagnostics Check		OK	08:10
Air Blank		0.000	08:10
Control Test		0.079	08:11
Air Blank		0.000	08:11
Subject Sample	# 1	0.000	08:11
Breath Volume		2.375	08:11
Air Blank		0.000	08:12
Air Blank		0.000	08:14
Subject Sample	# 2	0.000	08:14
Breath Volume		2.761	08:14
Air Blank		0.000	08:15
Control Test		0.082	08:15
Air Blank		0.000	08:16
Diagnostics Check		OK	08:16

Data Download Date/Time: 06/20/2018 14:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: CROMPTON

DATE/TIME OF TEST: 06/17/2018 04:40:39

Diagnostics Check		OK	04:34
Air Blank		0.000	04:35
Control Test		0.079	04:35
Air Blank		0.000	04:35
Subject Sample	# 1	0.114	04:36
Breath Volume		2.140	04:36
Air Blank		0.000	04:37
Air Blank		0.000	04:38
Subject Sample	# 2	0.116	04:39
Breath Volume		2.304	04:39
Air Blank		0.000	04:39
Control Test		0.081	04:40
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 06/20/2018 14:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COUCH,WILLIAM,K
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: COUCH,WILLIAM
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 06/17/2018 18:48:27

Diagnostics Check		OK	18:37	
Air Blank		0.000	18:37	
Control Test		0.079	18:37	
Air Blank		0.000	18:38	
Subject Sample	# 1		18:41	No Sample Provided
Breath Volume		0.000	18:41	
Air Blank		0.000	18:41	
Air Blank		0.000	18:43	
Subject Sample	# 2		18:47	No Sample Provided
Breath Volume		0.000	18:47	
Air Blank		0.000	18:47	
Control Test		0.079	18:47	
Air Blank		0.000	18:48	
Diagnostics Check		OK	18:48	

Data Download Date/Time: 06/20/2018 14:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JACOBSEN,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: CINTRON

DATE/TIME OF TEST: 06/19/2018 03:40:37

Diagnostics Check	OK	03:39	
Air Blank		03:40	RFI Detect
Air Blank	0.000	03:40	

Data Download Date/Time: 06/20/2018 14:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JACOBSEN,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: CINTRON

DATE/TIME OF TEST: 06/19/2018 03:46:57

Diagnostics Check		OK	03:44	
Air Blank		0.000	03:44	
Control Test		0.079	03:44	
Air Blank		0.000	03:45	
Subject Sample	# 1		03:45	Subject Test Refused
Breath Volume		0.000	03:45	
Air Blank		0.000	03:46	
Control Test		0.079	03:46	
Air Blank		0.000	03:46	
Diagnostics Check		OK	03:46	

Data Download Date/Time: 06/20/2018 14:02

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: SACHS

DATE/TIME OF TEST: 06/20/2018 22:22:20

Diagnostics Check		OK	22:15
Air Blank		0.000	22:15
Control Test		0.079	22:15
Air Blank		0.000	22:16
Subject Sample	# 1	0.064	22:18
Breath Volume		1.414	22:18
Air Blank		0.000	22:18
Air Blank		0.000	22:20
Subject Sample	# 2	0.060	22:20
Breath Volume		1.359	22:20
Air Blank		0.000	22:21
Control Test		0.080	22:21
Air Blank		0.000	22:22
Diagnostics Check		OK	22:22

Data Download Date/Time: 07/13/2018 12:11

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: KARP

DATE/TIME OF TEST: 06/22/2018 16:36:47

Diagnostics Check		OK	16:30
Air Blank		0.000	16:31
Control Test		0.078	16:31
Air Blank		0.000	16:32
Subject Sample	# 1	0.000	16:32
Breath Volume		2.496	16:32
Air Blank		0.000	16:32
Air Blank		0.000	16:35
Subject Sample	# 2	0.000	16:35
Breath Volume		2.535	16:35
Air Blank		0.000	16:35
Control Test		0.080	16:36
Air Blank		0.000	16:36
Diagnostics Check		OK	16:36

Data Download Date/Time: 07/13/2018 12:11

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: RAPER

DATE/TIME OF TEST: 06/22/2018 23:59:21

Diagnostics Check		OK	23:56	
Air Blank		0.000	23:56	
Control Test		0.079	23:57	
Air Blank		0.000	23:57	
Subject Sample	# 1		23:58	Subject Test Refused
Breath Volume		0.000	23:58	
Air Blank		0.000	23:58	
Control Test		0.079	23:58	
Air Blank		0.000	23:59	
Diagnostics Check		OK	23:59	

Data Download Date/Time: 07/13/2018 12:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: BULAMBO

DATE/TIME OF TEST: 06/24/2018 02:59:44

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.078	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.151	02:54
Breath Volume		1.226	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.147	02:58
Breath Volume		1.207	02:58
Air Blank		0.000	02:58
Control Test		0.081	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 07/13/2018 12:11

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: PAULSON

DATE/TIME OF TEST: 06/25/2018 23:00:22

Diagnostics Check		OK	22:51
Air Blank		0.000	22:51
Control Test		0.078	22:52
Air Blank		0.000	22:52
Subject Sample	# 1	0.117	22:55
Breath Volume		1.296	22:55
Air Blank		0.000	22:56
Air Blank		0.000	22:58
Subject Sample	# 2	0.117	22:58
Breath Volume		1.386	22:58
Air Blank		0.000	22:59
Control Test		0.083	22:59
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 07/13/2018 12:11

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAWARD,MORGANA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HOLM,MATTHEW
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: WISEMAN

DATE/TIME OF TEST: 06/27/2018 04:10:01

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:07	
Control Test		0.078	04:07	
Air Blank		0.000	04:08	
Subject Sample	# 1		04:08	Subject Test Refused
Breath Volume		0.000	04:08	
Air Blank		0.000	04:09	
Control Test		0.078	04:09	
Air Blank		0.000	04:09	
Diagnostics Check		OK	04:10	

Data Download Date/Time: 07/13/2018 12:11

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: BORDWELL

DATE/TIME OF TEST: 06/27/2018 23:45:21

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.078	23:39
Air Blank		0.000	23:39
Subject Sample	# 1	0.188	23:40
Breath Volume		2.144	23:40
Air Blank		0.000	23:41
Air Blank		0.000	23:43
Subject Sample	# 2	0.180	23:43
Breath Volume		1.390	23:43
Air Blank		0.000	23:44
Control Test		0.081	23:44
Air Blank		0.000	23:45
Diagnostics Check		OK	23:45

Data Download Date/Time: 07/13/2018 12:11

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:21

SUBJECT'S LAST NAME: SHELLHAMER

DATE/TIME OF TEST: 06/28/2018 18:44:33

Diagnostics Check		OK	18:41	
Air Blank		0.000	18:42	
Control Test		0.078	18:42	
Air Blank		0.000	18:43	
Subject Sample	# 1		18:43	Subject Test Refused
Breath Volume		0.000	18:43	
Air Blank		0.000	18:43	
Control Test		0.078	18:43	
Air Blank		0.000	18:44	
Diagnostics Check		OK	18:44	

Data Download Date/Time: 07/13/2018 12:11

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,BENJAMIN,F
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SIMPKINS,BENJAMIN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:26

SUBJECT'S LAST NAME: PRIDE

DATE/TIME OF TEST: 06/30/2018 06:50:56

Diagnostics Check		OK	06:46	
Air Blank		0.000	06:47	
Control Test		0.078	06:47	
Air Blank		0.000	06:47	
Subject Sample	# 1	0.165	06:49	
Breath Volume		1.671	06:49	
Air Blank			06:50	Purge Fail
Air Blank		0.000	06:50	

Data Download Date/Time: 07/13/2018 12:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,BENJAMIN,F
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SIMPKINS,BENJAMIN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:26

SUBJECT'S LAST NAME: PRIDE

DATE/TIME OF TEST: 06/30/2018 06:58:58

Diagnostics Check		OK	06:52
Air Blank		0.000	06:52
Control Test		0.081	06:53
Air Blank		0.000	06:53
Subject Sample	# 1	0.166	06:54
Breath Volume		2.078	06:54
Air Blank		0.000	06:55
Air Blank		0.000	06:56
Subject Sample	# 2	0.155	06:57
Breath Volume		1.570	06:57
Air Blank		0.000	06:58
Control Test		0.080	06:58
Air Blank		0.000	06:58
Diagnostics Check		OK	06:58

Data Download Date/Time: 07/13/2018 12:11

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGOODIAL,TRISTON,T
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: DISANO

DATE/TIME OF TEST: 06/01/2018 00:33:02

Diagnostics Check		OK	00:30	
Air Blank		0.000	00:30	
Control Test		0.079	00:30	
Air Blank		0.000	00:31	
Subject Sample	# 1		00:31	Subject Test Refused
Breath Volume		0.000	00:31	
Air Blank		0.000	00:32	
Control Test		0.079	00:32	
Air Blank		0.000	00:32	
Diagnostics Check		OK	00:33	

Data Download Date/Time: 06/29/2018 09:20

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: M422421692910

DATE/TIME OF TEST: 06/03/2018 02:27:35

Diagnostics Check		OK	02:24	
Air Blank		0.000	02:25	
Control Test		0.077	02:25	
Air Blank		0.000	02:26	
Subject Sample	# 1		02:26	Subject Test Refused
Breath Volume		0.000	02:26	
Air Blank		0.000	02:26	
Control Test		0.078	02:27	
Air Blank		0.000	02:27	
Diagnostics Check		OK	02:27	

Data Download Date/Time: 06/27/2018 17:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: GERMAIN

DATE/TIME OF TEST: 06/13/2018 01:48:11

Diagnostics Check		OK	01:41
Air Blank		0.000	01:41
Control Test		0.079	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.079	01:43
Breath Volume		2.136	01:43
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.084	01:46
Breath Volume		3.011	01:46
Air Blank		0.000	01:47
Control Test		0.079	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 06/27/2018 17:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/14/2018 01:10:20

Diagnostics Check		OK	01:07	
Air Blank		0.000	01:07	
Control Test		0.079	01:08	
Air Blank		0.000	01:08	
Subject Sample	# 1		01:08	Subject Test Refused
Breath Volume		0.000	01:08	
Air Blank		0.000	01:09	
Control Test		0.078	01:09	
Air Blank		0.000	01:10	
Diagnostics Check		OK	01:10	

Data Download Date/Time: 06/27/2018 17:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,GAVIN,O
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: DARDEN

DATE/TIME OF TEST: 06/21/2018 02:48:57

Diagnostics Check		OK	02:39	
Air Blank		0.000	02:39	
Control Test		0.078	02:40	
Air Blank		0.000	02:40	
Subject Sample	# 1		02:44	No Sample Provided
Breath Volume		0.000	02:44	
Air Blank		0.000	02:44	
Air Blank		0.000	02:46	
Subject Sample	# 2		02:47	Subject Test Refused
Breath Volume		0.000	02:47	
Air Blank		0.000	02:48	
Control Test		0.078	02:48	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:48	

Data Download Date/Time: 06/27/2018 17:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: SOLANO

DATE/TIME OF TEST: 06/22/2018 00:13:32

Diagnostics Check		OK	00:10	
Air Blank		0.000	00:11	
Control Test		0.079	00:11	
Air Blank		0.000	00:11	
Subject Sample	# 1		00:12	Subject Test Refused
Breath Volume		0.000	00:12	
Air Blank		0.000	00:12	
Control Test		0.078	00:12	
Air Blank		0.000	00:13	
Diagnostics Check		OK	00:13	

Data Download Date/Time: 06/27/2018 17:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASSILAKIS,GEORGE,E
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: VASSILAKIS,GEORGE
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BALDWIN

DATE/TIME OF TEST: 06/23/2018 03:32:50

Diagnostics Check		OK	03:26
Air Blank		0.000	03:27
Control Test		0.079	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.076	03:28
Breath Volume		1.179	03:28
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.074	03:31
Breath Volume		1.328	03:31
Air Blank		0.000	03:31
Control Test		0.078	03:32
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 06/27/2018 17:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CROCKER,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: MCGEE 3

DATE/TIME OF TEST: 06/02/2018 02:38:54

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.080	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.118	02:34
Breath Volume		1.523	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.112	02:37
Breath Volume		1.695	02:37
Air Blank		0.000	02:37
Control Test		0.079	02:38
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 06/19/2018 15:03

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSEN,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: VINCE

DATE/TIME OF TEST: 06/04/2018 02:37:05

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.141	02:32
Breath Volume		2.042	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.137	02:35
Breath Volume		1.425	02:35
Air Blank		0.000	02:36
Control Test		0.079	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 06/19/2018 15:03

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 06/08/2018 01:24:06

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.079	01:18
Air Blank		0.000	01:19
Subject Sample	# 1	0.000	01:19
Breath Volume		4.542	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.000	01:22
Breath Volume		4.031	01:22
Air Blank		0.000	01:23
Control Test		0.079	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 06/19/2018 15:04

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: MCCLAIN

DATE/TIME OF TEST: 06/16/2018 18:42:24

Diagnostics Check		OK	18:37	
Air Blank		0.000	18:37	
Control Test		0.079	18:37	
Air Blank		0.000	18:38	
Subject Sample	# 1	0.276	18:38	
Breath Volume		1.988	18:38	
Air Blank		0.000	18:39	
Air Blank			18:41	RFI Detect
Air Blank		0.000	18:41	
Control Test		0.079	18:41	
Air Blank		0.000	18:42	
Diagnostics Check		OK	18:42	

Data Download Date/Time: 06/19/2018 15:04

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: MCCLAIN

DATE/TIME OF TEST: 06/16/2018 18:52:51

Diagnostics Check		OK	18:46
Air Blank		0.000	18:47
Control Test		0.078	18:47
Air Blank		0.000	18:47
Subject Sample	# 1	0.275	18:48
Breath Volume		1.875	18:48
Air Blank		0.000	18:49
Air Blank		0.000	18:50
Subject Sample	# 2	0.279	18:51
Breath Volume		1.460	18:51
Air Blank		0.000	18:51
Control Test		0.078	18:52
Air Blank		0.000	18:52
Diagnostics Check		OK	18:52

Data Download Date/Time: 06/19/2018 15:04

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: OLVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: TOBLER

DATE/TIME OF TEST: 06/16/2018 22:50:47

Diagnostics Check		OK	22:44
Air Blank		0.000	22:45
Control Test		0.079	22:45
Air Blank		0.000	22:45
Subject Sample	# 1	0.190	22:46
Breath Volume		2.050	22:46
Air Blank		0.000	22:47
Air Blank		0.000	22:48
Subject Sample	# 2	0.183	22:49
Breath Volume		1.289	22:49
Air Blank		0.000	22:49
Control Test		0.079	22:50
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 06/19/2018 15:04

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: HINTON

DATE/TIME OF TEST: 06/17/2018 19:14:09

Diagnostics Check		OK	19:11	
Air Blank		0.000	19:11	
Control Test		0.079	19:12	
Air Blank		0.000	19:12	
Subject Sample	# 1		19:12	Subject Test Refused
Breath Volume		0.000	19:12	
Air Blank		0.000	19:13	
Control Test		0.079	19:13	
Air Blank		0.000	19:14	
Diagnostics Check		OK	19:14	

Data Download Date/Time: 06/19/2018 15:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: PEEPLES

DATE/TIME OF TEST: 06/17/2018 20:29:28

Diagnostics Check		OK	20:23
Air Blank		0.000	20:23
Control Test		0.080	20:24
Air Blank		0.000	20:24
Subject Sample	# 1	0.000	20:25
Breath Volume		1.355	20:25
Air Blank		0.000	20:25
Air Blank		0.000	20:27
Subject Sample	# 2	0.000	20:28
Breath Volume		1.273	20:28
Air Blank		0.000	20:28
Control Test		0.079	20:28
Air Blank		0.000	20:29
Diagnostics Check		OK	20:29

Data Download Date/Time: 06/19/2018 15:04

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: CLIFFORD

DATE/TIME OF TEST: 06/17/2018 22:01:17

Diagnostics Check		OK	21:51
Air Blank		0.000	21:52
Control Test		0.079	21:52
Air Blank		0.000	21:53
Subject Sample	# 1	0.072	21:55
Breath Volume		1.718	21:55
Air Blank		0.000	21:56
Air Blank		0.000	21:58
Subject Sample	# 2	0.067	21:59
Breath Volume		1.175	21:59
Air Blank		0.000	22:00
Control Test		0.079	22:00
Air Blank		0.000	22:01
Diagnostics Check		OK	22:01

Data Download Date/Time: 06/19/2018 15:04

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,E,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WOOLDRIDGE,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: MCLEAN

DATE/TIME OF TEST: 06/19/2018 17:38:41

Diagnostics Check		OK	17:32
Air Blank		0.000	17:32
Control Test		0.080	17:32
Air Blank		0.000	17:33
Subject Sample	# 1	0.176	17:34
Breath Volume		2.964	17:34
Air Blank		0.000	17:34
Air Blank		0.000	17:36
Subject Sample	# 2	0.179	17:37
Breath Volume		1.554	17:37
Air Blank		0.000	17:37
Control Test		0.079	17:38
Air Blank		0.000	17:38
Diagnostics Check		OK	17:38

Data Download Date/Time: 07/06/2018 12:24

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROEBRT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: WINDHAM

DATE/TIME OF TEST: 06/24/2018 01:59:43

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:57	
Control Test		0.079	01:57	
Air Blank		0.000	01:58	
Subject Sample	# 1		01:58	Subject Test Refused
Breath Volume		0.000	01:58	
Air Blank		0.000	01:58	
Control Test		0.078	01:59	
Air Blank		0.000	01:59	
Diagnostics Check		OK	01:59	

Data Download Date/Time: 07/06/2018 12:24

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSEN,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 06/29/2018 03:31:26

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:29	
Control Test		0.079	03:29	
Air Blank		0.000	03:29	
Subject Sample	# 1		03:30	Subject Test Refused
Breath Volume		0.000	03:30	
Air Blank		0.000	03:30	
Control Test		0.078	03:30	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 07/06/2018 12:24

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: STIVERS,H
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: INMAN

DATE/TIME OF TEST: 06/02/2018 01:12:55

Diagnostics Check		OK	01:06
Air Blank		0.000	01:06
Control Test		0.081	01:07
Air Blank		0.000	01:07
Subject Sample	# 1	0.163	01:08
Breath Volume		2.078	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:10
Subject Sample	# 2	0.160	01:11
Breath Volume		1.671	01:11
Air Blank		0.000	01:12
Control Test		0.081	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 06/27/2018 19:18

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ALEXANDER,D
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: WILKINS

DATE/TIME OF TEST: 06/03/2018 21:04:34

Diagnostics Check		OK	20:57
Air Blank		0.000	20:57
Control Test		0.081	20:57
Air Blank		0.000	20:58
Subject Sample	# 1	0.161	21:00
Breath Volume		1.433	21:00
Air Blank		0.000	21:00
Air Blank		0.000	21:02
Subject Sample	# 2	0.158	21:03
Breath Volume		1.785	21:03
Air Blank		0.000	21:03
Control Test		0.080	21:04
Air Blank		0.000	21:04
Diagnostics Check		OK	21:04

Data Download Date/Time: 06/27/2018 19:18

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LANIER,H
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 05/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: AKINS

DATE/TIME OF TEST: 06/06/2018 20:51:59

Diagnostics Check		OK	20:44
Air Blank		0.000	20:44
Control Test		0.080	20:44
Air Blank		0.000	20:45
Subject Sample	# 1	0.262	20:47
Breath Volume		1.640	20:47
Air Blank		0.000	20:47
Air Blank		0.000	20:49
Subject Sample	# 2	0.254	20:50
Breath Volume		1.402	20:50
Air Blank		0.000	20:51
Control Test		0.079	20:51
Air Blank		0.000	20:51
Diagnostics Check		OK	20:51

Data Download Date/Time: 06/27/2018 19:18

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BASSLER,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: PHEBUS

DATE/TIME OF TEST: 06/07/2018 03:35:48

Diagnostics Check		OK	03:29
Air Blank		0.000	03:30
Control Test		0.080	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.000	03:31
Breath Volume		1.824	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.000	03:34
Breath Volume		2.042	03:34
Air Blank		0.000	03:34
Control Test		0.081	03:35
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 06/27/2018 19:18

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GRACKI,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: TIGAR

DATE/TIME OF TEST: 06/17/2018 02:42:45

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:31	
Control Test		0.079	02:32	
Air Blank		0.000	02:32	
Subject Sample	# 1		02:35	No Sample Provided
Breath Volume		0.000	02:35	
Air Blank		0.000	02:36	
Air Blank		0.000	02:38	
Subject Sample	# 2		02:41	No Sample Provided
Breath Volume		0.000	02:41	
Air Blank		0.000	02:41	
Control Test		0.081	02:42	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:42	

Data Download Date/Time: 06/27/2018 19:18

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HOULE,A
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 05/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: JARQUIN

DATE/TIME OF TEST: 06/23/2018 04:13:16

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.081	04:05
Air Blank		0.000	04:06
Subject Sample	# 1	0.161	04:07
Breath Volume		1.101	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.164	04:11
Breath Volume		2.191	04:11
Air Blank		0.000	04:12
Control Test		0.080	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 06/27/2018 19:18

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROOKS,E,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: EHRENFELD,G
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: BEEGLE

DATE/TIME OF TEST: 06/30/2018 17:46:27

Diagnostics Check		OK	17:37	
Air Blank		0.000	17:38	
Control Test		0.080	17:38	
Air Blank		0.000	17:38	
Subject Sample	# 1		17:42	No Sample Provided
Breath Volume		0.000	17:42	
Air Blank		0.000	17:42	
Air Blank		0.000	17:44	
Subject Sample	# 2		17:45	Subject Test Refused
Breath Volume		0.000	17:45	
Air Blank		0.000	17:45	
Control Test		0.080	17:45	
Air Blank		0.000	17:46	
Diagnostics Check		OK	17:46	

Data Download Date/Time: 07/25/2018 23:32

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 06/01/2018 05:21:01

Diagnostics Check		OK	05:13
Air Blank		0.000	05:14
Control Test		0.082	05:14
Air Blank		0.000	05:15
Subject Sample	# 1	0.060	05:16
Breath Volume		1.265	05:16
Air Blank		0.000	05:16
Air Blank		0.000	05:18
Subject Sample	# 2	0.062	05:19
Breath Volume		1.652	05:19
Air Blank		0.000	05:20
Control Test		0.082	05:20
Air Blank		0.000	05:20
Diagnostics Check		OK	05:21

Data Download Date/Time: 06/21/2018 11:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: AKERS,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 15:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: SPINELLI

DATE/TIME OF TEST: 06/01/2018 18:03:16

Diagnostics Check		OK	17:56
Air Blank		0.000	17:56
Control Test		0.082	17:57
Air Blank		0.000	17:57
Subject Sample	# 1	0.282	17:58
Breath Volume		1.761	17:58
Air Blank		0.000	17:59
Air Blank		0.000	18:01
Subject Sample	# 2	0.282	18:01
Breath Volume		1.816	18:01
Air Blank		0.000	18:02
Control Test		0.081	18:02
Air Blank		0.000	18:03
Diagnostics Check		OK	18:03

Data Download Date/Time: 06/21/2018 11:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: MUNOZ-CERVETTO

DATE/TIME OF TEST: 06/02/2018 02:28:25

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.081	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.174	02:24
Breath Volume		1.449	02:24
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.187	02:26
Breath Volume		1.234	02:26
Air Blank		0.000	02:27
Control Test		0.081	02:27
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 06/21/2018 11:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: DALY

DATE/TIME OF TEST: 06/07/2018 01:40:25

Diagnostics Check		OK	01:34
Air Blank		0.000	01:34
Control Test		0.081	01:34
Air Blank		0.000	01:35
Subject Sample	# 1	0.000	01:36
Breath Volume		3.683	01:36
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.000	01:39
Breath Volume		1.828	01:39
Air Blank		0.000	01:39
Control Test		0.081	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 06/21/2018 11:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOUGHRAN,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SCHMERSEY

DATE/TIME OF TEST: 06/07/2018 02:05:14

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.081	01:59
Air Blank		0.000	02:00
Subject Sample	# 1	0.112	02:00
Breath Volume		1.414	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.132	02:03
Breath Volume		1.914	02:03
Air Blank		0.000	02:04
Control Test		0.081	02:04
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 06/21/2018 11:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: BEDOYA-FERNANDEZ

DATE/TIME OF TEST: 06/07/2018 21:54:52

Diagnostics Check		OK	21:47
Air Blank		0.000	21:47
Control Test		0.081	21:48
Air Blank		0.000	21:48
Subject Sample	# 1	0.167	21:50
Breath Volume		1.687	21:50
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.159	21:53
Breath Volume		1.925	21:53
Air Blank		0.000	21:53
Control Test		0.081	21:54
Air Blank		0.000	21:54
Diagnostics Check		OK	21:54

Data Download Date/Time: 06/21/2018 11:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BARTLETT,AMY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: VICENTE

DATE/TIME OF TEST: 06/07/2018 23:46:58

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.081	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.256	23:42
Breath Volume		1.953	23:42
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.252	23:45
Breath Volume		3.148	23:45
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 06/21/2018 11:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: KERN

DATE/TIME OF TEST: 06/08/2018 01:38:00

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.081	01:26	
Air Blank		0.000	01:27	
Subject Sample	# 1		01:30	No Sample Provided
Breath Volume		0.000	01:30	
Air Blank		0.000	01:30	
Air Blank		0.000	01:32	
Subject Sample	# 2	0.093	01:33	
Breath Volume		1.832	01:33	
Air Blank		0.000	01:34	
Air Blank		0.000	01:36	
Subject Sample	# 3	0.094	01:36	
Breath Volume		1.636	01:36	
Air Blank		0.000	01:37	
Control Test		0.080	01:37	
Air Blank		0.000	01:37	
Diagnostics Check		OK	01:38	

Data Download Date/Time: 06/21/2018 11:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOUGHRAN,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 06/08/2018 04:09:07

Diagnostics Check		OK	03:57	
Air Blank		0.000	03:58	
Control Test		0.081	03:58	
Air Blank		0.000	03:59	
Subject Sample	# 1		04:02	No Sample Provided
Breath Volume		0.000	04:02	
Air Blank		0.000	04:02	
Air Blank		0.000	04:04	
Subject Sample	# 2		04:07	No Sample Provided
Breath Volume		0.000	04:07	
Air Blank		0.000	04:08	
Control Test		0.081	04:08	
Air Blank		0.000	04:09	
Diagnostics Check		OK	04:09	

Data Download Date/Time: 06/21/2018 11:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 06/09/2018 01:31:02

Diagnostics Check		OK	01:23
Air Blank		0.000	01:24
Control Test		0.079	01:24
Air Blank		0.000	01:25
Subject Sample	# 1	0.154	01:26
Breath Volume		1.976	01:26
Air Blank		0.000	01:26
Air Blank		0.000	01:28
Subject Sample	# 2	0.147	01:29
Breath Volume		2.183	01:29
Air Blank		0.000	01:30
Control Test		0.079	01:30
Air Blank		0.000	01:30
Diagnostics Check		OK	01:31

Data Download Date/Time: 06/21/2018 11:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: REID,FRANCEEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:36

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/09/2018 07:09:54

Diagnostics Check		OK	07:03
Air Blank		0.000	07:03
Control Test		0.081	07:03
Air Blank		0.000	07:04
Subject Sample	# 1	0.140	07:05
Breath Volume		3.566	07:05
Air Blank		0.000	07:05
Air Blank		0.000	07:07
Subject Sample	# 2	0.135	07:08
Breath Volume		3.503	07:08
Air Blank		0.000	07:09
Control Test		0.082	07:09
Air Blank		0.000	07:09
Diagnostics Check		OK	07:09

Data Download Date/Time: 06/21/2018 11:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: RAMIREZ GONZALEZ

DATE/TIME OF TEST: 06/10/2018 03:30:24

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.082	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.206	03:25
Breath Volume		1.500	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.209	03:28
Breath Volume		2.101	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 06/21/2018 11:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: IRELAN,ANDREW
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: VILLATORO

DATE/TIME OF TEST: 06/10/2018 05:15:51

Diagnostics Check		OK	05:09
Air Blank		0.000	05:09
Control Test		0.081	05:10
Air Blank		0.000	05:10
Subject Sample	# 1	0.147	05:11
Breath Volume		2.082	05:11
Air Blank		0.000	05:12
Air Blank		0.000	05:13
Subject Sample	# 2	0.141	05:14
Breath Volume		1.902	05:14
Air Blank		0.000	05:14
Control Test		0.081	05:15
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 06/21/2018 11:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DHAITI,DWAYNE
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: LOISEAU

DATE/TIME OF TEST: 06/10/2018 18:01:00

Diagnostics Check		OK	17:54
Air Blank		0.000	17:54
Control Test		0.082	17:55
Air Blank		0.000	17:55
Subject Sample	# 1	0.196	17:56
Breath Volume		1.988	17:56
Air Blank		0.000	17:57
Air Blank		0.000	17:59
Subject Sample	# 2	0.204	17:59
Breath Volume		2.027	17:59
Air Blank		0.000	18:00
Control Test		0.081	18:00
Air Blank		0.000	18:00
Diagnostics Check		OK	18:01

Data Download Date/Time: 06/21/2018 11:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STRZALOWSKI,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 06/11/2018 02:32:10

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.081	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.150	02:27
Breath Volume		1.242	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.140	02:30
Breath Volume		1.175	02:30
Air Blank		0.000	02:31
Control Test		0.081	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 06/21/2018 11:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: JASMIN

DATE/TIME OF TEST: 06/13/2018 03:47:26

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.081	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.222	03:42
Breath Volume		1.851	03:42
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.217	03:45
Breath Volume		1.617	03:45
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 06/21/2018 11:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUNA,KRISTINA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: MOJARRANGO CORTES

DATE/TIME OF TEST: 06/14/2018 04:03:36

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:50	
Control Test		0.081	03:50	
Air Blank		0.000	03:51	
Subject Sample	# 1	0.147	03:53	
Breath Volume		1.667	03:53	
Air Blank		0.000	03:54	
Air Blank		0.000	03:56	
Subject Sample	# 2		03:59	No Sample Provided
Breath Volume		0.000	03:59	
Air Blank		0.000	03:59	
Air Blank		0.000	04:01	
Subject Sample	# 3	0.150	04:02	
Breath Volume		1.667	04:02	
Air Blank		0.000	04:02	
Control Test		0.081	04:03	
Air Blank		0.000	04:03	
Diagnostics Check		OK	04:03	

Data Download Date/Time: 06/21/2018 11:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FLEURANT,RODSON
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: BARTOLAZZI

DATE/TIME OF TEST: 06/14/2018 20:31:11

Diagnostics Check		OK	20:24
Air Blank		0.000	20:24
Control Test		0.081	20:25
Air Blank		0.000	20:25
Subject Sample	# 1	0.025	20:26
Breath Volume		1.101	20:26
Air Blank		0.000	20:27
Air Blank		0.000	20:29
Subject Sample	# 2	0.020	20:29
Breath Volume		1.960	20:29
Air Blank		0.000	20:30
Control Test		0.081	20:30
Air Blank		0.000	20:31
Diagnostics Check		OK	20:31

Data Download Date/Time: 06/21/2018 11:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MYLOTT

DATE/TIME OF TEST: 06/15/2018 01:46:01

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.081	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.107	01:41
Breath Volume		3.832	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:43
Subject Sample	# 2	0.110	01:44
Breath Volume		3.785	01:44
Air Blank		0.000	01:45
Control Test		0.080	01:45
Air Blank		0.000	01:45
Diagnostics Check		OK	01:46

Data Download Date/Time: 06/21/2018 11:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: MONTIEL

DATE/TIME OF TEST: 06/15/2018 02:58:31

Diagnostics Check		OK	02:52
Air Blank		0.000	02:52
Control Test		0.081	02:52
Air Blank		0.000	02:53
Subject Sample	# 1	0.019	02:54
Breath Volume		2.609	02:54
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.017	02:57
Breath Volume		3.070	02:57
Air Blank		0.000	02:57
Control Test		0.081	02:57
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 06/21/2018 11:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 06/15/2018 05:05:42

Diagnostics Check		OK	04:59
Air Blank		0.000	04:59
Control Test		0.081	04:59
Air Blank		0.000	05:00
Subject Sample	# 1	0.240	05:01
Breath Volume		2.355	05:01
Air Blank		0.000	05:01
Air Blank		0.000	05:03
Subject Sample	# 2	0.239	05:04
Breath Volume		2.121	05:04
Air Blank		0.000	05:04
Control Test		0.082	05:05
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 06/21/2018 11:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: NODARSE

DATE/TIME OF TEST: 06/15/2018 05:14:44

Diagnostics Check		OK	05:08
Air Blank		0.000	05:08
Control Test		0.080	05:09
Air Blank		0.000	05:09
Subject Sample	# 1	0.173	05:10
Breath Volume		1.820	05:10
Air Blank		0.000	05:10
Air Blank		0.000	05:12
Subject Sample	# 2	0.169	05:13
Breath Volume		1.640	05:13
Air Blank		0.000	05:13
Control Test		0.081	05:14
Air Blank		0.000	05:14
Diagnostics Check		OK	05:14

Data Download Date/Time: 06/21/2018 11:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 06/15/2018 21:59:40

Diagnostics Check		OK	21:44	
Air Blank		0.000	21:44	
Control Test		0.081	21:45	
Air Blank		0.000	21:45	
Subject Sample	# 1	0.160	21:48	Volume Not Met
Breath Volume		0.078	21:48	
Air Blank		0.000	21:49	
Air Blank		0.000	21:51	
Subject Sample	# 2	0.160	21:52	
Breath Volume		1.527	21:52	
Air Blank		0.000	21:53	
Air Blank		0.000	21:55	
Subject Sample	# 3	0.124	21:58	Volume Not Met
Breath Volume		0.347	21:58	
Air Blank		0.000	21:58	
Control Test		0.081	21:59	
Air Blank		0.000	21:59	
Diagnostics Check		OK	21:59	

Data Download Date/Time: 06/21/2018 11:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 06/15/2018 22:14:35

Diagnostics Check		OK	22:01	
Air Blank		0.000	22:02	
Control Test		0.080	22:02	
Air Blank		0.000	22:03	
Subject Sample	# 1	0.157	22:04	
Breath Volume		1.570	22:04	
Air Blank		0.000	22:04	
Air Blank		0.000	22:06	
Subject Sample	# 2	0.141	22:10	Volume Not Met
Breath Volume		0.015	22:10	
Air Blank		0.000	22:10	
Air Blank		0.000	22:12	
Subject Sample	# 3	0.166	22:13	
Breath Volume		1.464	22:13	
Air Blank		0.000	22:13	
Control Test		0.080	22:14	
Air Blank		0.000	22:14	
Diagnostics Check		OK	22:14	

Data Download Date/Time: 06/21/2018 11:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 06/16/2018 00:38:00

Diagnostics Check		OK	00:31
Air Blank		0.000	00:32
Control Test		0.081	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.179	00:33
Breath Volume		2.761	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.174	00:36
Breath Volume		1.296	00:36
Air Blank		0.000	00:37
Control Test		0.080	00:37
Air Blank		0.000	00:37
Diagnostics Check		OK	00:38

Data Download Date/Time: 06/21/2018 11:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CHILDRESS,JACOB
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: PENADO

DATE/TIME OF TEST: 06/16/2018 00:58:16

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.080	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.185	00:52
Breath Volume		1.980	00:52
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.174	00:56
Breath Volume		2.324	00:56
Air Blank		0.000	00:57
Control Test		0.082	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 06/21/2018 11:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: JOANIS

DATE/TIME OF TEST: 06/17/2018 00:58:19

Diagnostics Check		OK	00:46	
Air Blank		0.000	00:46	
Control Test		0.081	00:47	
Air Blank		0.000	00:47	
Subject Sample	# 1	0.134	00:50	Volume Not Met
Breath Volume		0.484	00:50	
Air Blank		0.000	00:51	
Air Blank		0.000	00:53	
Subject Sample	# 2	0.155	00:53	
Breath Volume		2.437	00:53	
Air Blank		0.000	00:54	
Air Blank		0.000	00:56	
Subject Sample	# 3	0.155	00:56	
Breath Volume		2.601	00:56	
Air Blank		0.000	00:57	
Control Test		0.081	00:57	
Air Blank		0.000	00:58	
Diagnostics Check		OK	00:58	

Data Download Date/Time: 06/21/2018 11:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SCHOLLER

DATE/TIME OF TEST: 06/17/2018 03:49:45

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.081	03:43
Air Blank		0.000	03:43
Subject Sample	# 1	0.158	03:45
Breath Volume		1.101	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:47
Subject Sample	# 2	0.149	03:48
Breath Volume		1.144	03:48
Air Blank		0.000	03:48
Control Test		0.080	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 06/21/2018 11:47

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOUGHRAN,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: PIERCY

DATE/TIME OF TEST: 06/18/2018 00:00:03

Diagnostics Check		OK	23:53
Air Blank		0.000	23:53
Control Test		0.082	23:54
Air Blank		0.000	23:54
Subject Sample	# 1	0.202	23:55
Breath Volume		1.683	23:55
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.190	23:58
Breath Volume		1.554	23:58
Air Blank		0.000	23:59
Control Test		0.081	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	00:00

Data Download Date/Time: 06/21/2018 11:47

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DEDEJ,MEDJON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 07:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: DESROSIERS

DATE/TIME OF TEST: 06/18/2018 10:00:12

Diagnostics Check		OK	09:50
Air Blank		0.000	09:51
Control Test		0.081	09:51
Air Blank		0.000	09:52
Subject Sample	# 1	0.146	09:55
Breath Volume		1.742	09:55
Air Blank		0.000	09:55
Air Blank		0.000	09:57
Subject Sample	# 2	0.151	09:58
Breath Volume		1.175	09:58
Air Blank		0.000	09:59
Control Test		0.081	09:59
Air Blank		0.000	10:00
Diagnostics Check		OK	10:00

Data Download Date/Time: 06/21/2018 11:47

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ST PREUX,ALEXANDER
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SAHADEO

DATE/TIME OF TEST: 06/19/2018 01:41:50

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.081	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.200	01:37
Breath Volume		1.472	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:39
Subject Sample	# 2	0.199	01:40
Breath Volume		1.921	01:40
Air Blank		0.000	01:40
Control Test		0.081	01:41
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 06/21/2018 11:47

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: DUTCHER

DATE/TIME OF TEST: 06/20/2018 23:08:43

Diagnostics Check		OK	23:02
Air Blank		0.000	23:03
Control Test		0.082	23:03
Air Blank		0.000	23:03
Subject Sample	# 1	0.106	23:04
Breath Volume		3.171	23:04
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.101	23:07
Breath Volume		2.234	23:07
Air Blank		0.000	23:07
Control Test		0.082	23:08
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 06/21/2018 11:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WALKER,ELONZO
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: BLANCHARD

DATE/TIME OF TEST: 06/21/2018 00:25:23

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.081	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.169	00:20
Breath Volume		2.179	00:20
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.175	00:23
Breath Volume		2.953	00:23
Air Blank		0.000	00:24
Control Test		0.081	00:24
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 06/21/2018 11:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JAY
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 06/21/2018 02:08:31

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:56	
Control Test		0.082	01:56	
Air Blank		0.000	01:56	
Subject Sample	# 1	0.206	01:58	No .020 agreement
Breath Volume		1.203	01:58	
Air Blank		0.000	01:58	
Air Blank		0.000	02:00	
Subject Sample	# 2	0.143	02:03	Volume Not Met
Breath Volume		0.445	02:03	
Air Blank		0.000	02:04	
Air Blank		0.000	02:06	
Subject Sample	# 3	0.183	02:07	No .020 agreement
Breath Volume		2.484	02:07	
Air Blank		0.000	02:07	
Control Test		0.080	02:07	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

Data Download Date/Time: 06/21/2018 11:47

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JAY
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 06/21/2018 02:17:35

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.081	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.197	02:13
Breath Volume		2.343	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.204	02:16
Breath Volume		2.472	02:16
Air Blank		0.000	02:16
Control Test		0.081	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 06/21/2018 11:47

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RAZI,UMAIR
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: WAGAMON

DATE/TIME OF TEST: 06/21/2018 04:41:22

Diagnostics Check		OK	04:33
Air Blank		0.000	04:33
Control Test		0.081	04:33
Air Blank		0.000	04:34
Subject Sample	# 1	0.110	04:36
Breath Volume		1.843	04:36
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.113	04:39
Breath Volume		1.824	04:39
Air Blank		0.000	04:40
Control Test		0.081	04:40
Air Blank		0.000	04:41
Diagnostics Check		OK	04:41

Data Download Date/Time: 06/21/2018 11:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: HOLLINGSWORTH

DATE/TIME OF TEST: 06/21/2018 05:38:31

Diagnostics Check		OK	05:30
Air Blank		0.000	05:30
Control Test		0.080	05:30
Air Blank		0.000	05:31
Subject Sample	# 1	0.175	05:31
Breath Volume		1.308	05:31
Air Blank		0.000	05:32
Air Blank		0.000	05:34
Subject Sample	# 2	0.186	05:37
Breath Volume		1.531	05:37
Air Blank		0.000	05:37
Control Test		0.081	05:37
Air Blank		0.000	05:38
Diagnostics Check		OK	05:38

Data Download Date/Time: 06/21/2018 11:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: ABRAHAM

DATE/TIME OF TEST: 06/01/2018 03:21:23

Diagnostics Check		OK	03:14
Air Blank		0.000	03:15
Control Test		0.079	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.000	03:17
Breath Volume		1.183	03:17
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.000	03:20
Breath Volume		1.722	03:20
Air Blank		0.000	03:20
Control Test		0.080	03:20
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 06/21/2018 19:00

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: IGO

DATE/TIME OF TEST: 06/02/2018 21:42:17

Diagnostics Check		OK	21:35
Air Blank		0.000	21:36
Control Test		0.080	21:36
Air Blank		0.000	21:37
Subject Sample	# 1	0.211	21:37
Breath Volume		3.675	21:37
Air Blank		0.000	21:38
Air Blank		0.000	21:40
Subject Sample	# 2	0.208	21:40
Breath Volume		3.507	21:40
Air Blank		0.000	21:41
Control Test		0.078	21:41
Air Blank		0.000	21:42
Diagnostics Check		OK	21:42

Data Download Date/Time: 06/21/2018 19:00

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: REGAN,E
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BRYDSON

DATE/TIME OF TEST: 06/02/2018 22:29:44

Diagnostics Check		OK	22:20	
Air Blank		0.000	22:20	
Control Test		0.079	22:21	
Air Blank		0.000	22:21	
Subject Sample	# 1	0.138	22:24	Volume Not Met
Breath Volume		0.035	22:24	
Air Blank		0.000	22:25	
Air Blank		0.000	22:27	
Subject Sample	# 2		22:28	Subject Test Refused
Breath Volume		0.195	22:28	
Air Blank		0.000	22:28	
Control Test		0.081	22:29	
Air Blank		0.000	22:29	
Diagnostics Check		OK	22:29	

Data Download Date/Time: 06/21/2018 19:00

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 06/08/2018 00:31:57

Diagnostics Check		OK	00:25
Air Blank		0.000	00:26
Control Test		0.079	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.000	00:27
Breath Volume		2.300	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.000	00:30
Breath Volume		1.875	00:30
Air Blank		0.000	00:31
Control Test		0.080	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 06/21/2018 19:00

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,ANGEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALVARADO,ANGEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 06/12/2018 00:36:07

Diagnostics Check		OK	00:29
Air Blank		0.000	00:29
Control Test		0.079	00:30
Air Blank		0.000	00:30
Subject Sample	# 1	0.096	00:31
Breath Volume		3.171	00:31
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.085	00:34
Breath Volume		2.054	00:34
Air Blank		0.000	00:35
Control Test		0.079	00:35
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 06/21/2018 19:00

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MOKHIBER

DATE/TIME OF TEST: 06/13/2018 03:57:41

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.078	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.176	03:53
Breath Volume		3.667	03:53
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.176	03:56
Breath Volume		3.148	03:56
Air Blank		0.000	03:56
Control Test		0.077	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 06/21/2018 19:00

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: MCHUGH

DATE/TIME OF TEST: 06/15/2018 03:09:31

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:58	
Control Test		0.080	02:58	
Air Blank		0.000	02:59	
Subject Sample	# 1	0.133	03:02	Volume Not Met
Breath Volume		0.101	03:02	
Air Blank		0.000	03:02	
Air Blank		0.000	03:04	
Subject Sample	# 2	0.120	03:07	Volume Not Met
Breath Volume		0.878	03:07	
Air Blank		0.000	03:08	
Control Test		0.078	03:08	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 06/21/2018 19:00

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,ANGEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALVARADO,ANGEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: KROMPHOLD

DATE/TIME OF TEST: 06/15/2018 22:04:18

Diagnostics Check		OK	21:57
Air Blank		0.000	21:58
Control Test		0.079	21:58
Air Blank		0.000	21:59
Subject Sample	# 1	0.000	21:59
Breath Volume		2.007	21:59
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.000	22:02
Breath Volume		2.031	22:02
Air Blank		0.000	22:03
Control Test		0.079	22:03
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 06/21/2018 19:00

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: SCHIRATO

DATE/TIME OF TEST: 06/16/2018 00:29:59

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.080	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.135	00:24
Breath Volume		1.496	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:26
Subject Sample	# 2	0.130	00:28
Breath Volume		1.628	00:28
Air Blank		0.000	00:29
Control Test		0.078	00:29
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 06/21/2018 19:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BUEY

DATE/TIME OF TEST: 06/16/2018 22:58:51

Diagnostics Check		OK	22:52
Air Blank		0.000	22:52
Control Test		0.079	22:53
Air Blank		0.000	22:53
Subject Sample	# 1	0.240	22:54
Breath Volume		3.050	22:54
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.226	22:57
Breath Volume		1.480	22:57
Air Blank		0.000	22:57
Control Test		0.080	22:58
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 06/21/2018 19:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: SANFORD

DATE/TIME OF TEST: 06/28/2018 02:06:06

Diagnostics Check		OK	01:59
Air Blank		0.000	02:00
Control Test		0.079	02:00
Air Blank		0.000	02:01
Subject Sample	# 1	0.098	02:01
Breath Volume		1.871	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.107	02:04
Breath Volume		1.453	02:04
Air Blank		0.000	02:05
Control Test		0.079	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 07/11/2018 14:51

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 06/28/2018 02:42:58

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.081	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.173	02:38
Breath Volume		2.160	02:38
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.163	02:41
Breath Volume		1.960	02:41
Air Blank		0.000	02:42
Control Test		0.079	02:42
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 07/11/2018 14:51

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: JOLINE

DATE/TIME OF TEST: 06/30/2018 01:32:36

Diagnostics Check		OK	01:26
Air Blank		0.000	01:27
Control Test		0.079	01:27
Air Blank		0.000	01:27
Subject Sample	# 1	0.000	01:28
Breath Volume		2.675	01:28
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.000	01:31
Breath Volume		2.312	01:31
Air Blank		0.000	01:31
Control Test		0.080	01:32
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 07/11/2018 14:51

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: ANDERSON,TAYLOR
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 05/24/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: MC KEOUGH

DATE/TIME OF TEST: 06/21/2018 19:39:05

Diagnostics Check		OK	19:36	
Air Blank		0.000	19:36	
Control Test		0.078	19:37	
Air Blank		0.000	19:37	
Subject Sample	# 1		19:37	Subject Test Refused
Breath Volume		0.000	19:37	
Air Blank		0.000	19:38	
Control Test		0.078	19:38	
Air Blank		0.000	19:38	
Diagnostics Check		OK	19:39	

Data Download Date/Time: 06/27/2018 16:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURDOCK,TROY,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: MURDOCK,TROY
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 05/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: MINNIS

DATE/TIME OF TEST: 06/24/2018 05:03:30

Diagnostics Check		OK	04:56
Air Blank		0.000	04:57
Control Test		0.079	04:57
Air Blank		0.000	04:58
Subject Sample	# 1	0.176	04:59
Breath Volume		3.210	04:59
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.173	05:01
Breath Volume		2.562	05:01
Air Blank		0.000	05:02
Control Test		0.078	05:02
Air Blank		0.000	05:03
Diagnostics Check		OK	05:03

Data Download Date/Time: 06/27/2018 16:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KABAN,SEAN,W
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: KABAN,SEAN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 05/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: GLANSEN

DATE/TIME OF TEST: 06/25/2018 00:10:42

Diagnostics Check		OK	23:59	
Air Blank		0.000	23:59	
Control Test		0.079	23:59	
Air Blank		0.000	00:00	
Subject Sample	# 1	0.098	00:03	Volume Not Met
Breath Volume		0.398	00:03	
Air Blank		0.000	00:04	
Air Blank		0.000	00:06	
Subject Sample	# 2	0.088	00:09	Volume Not Met
Breath Volume		0.347	00:09	
Air Blank		0.000	00:09	
Control Test		0.079	00:10	
Air Blank		0.000	00:10	
Diagnostics Check		OK	00:10	

Data Download Date/Time: 06/27/2018 16:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KABAN,SEAN,W
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: KABAN,SEAN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 05/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: GLANSEN

DATE/TIME OF TEST: 06/25/2018 00:22:03

Diagnostics Check		OK	00:13	
Air Blank		0.000	00:14	
Control Test		0.078	00:14	
Air Blank		0.000	00:14	
Subject Sample	# 1	0.089	00:18	Volume Not Met
Breath Volume		0.691	00:18	
Air Blank		0.000	00:18	
Air Blank		0.000	00:20	
Subject Sample	# 2		00:20	Subject Test Refused
Breath Volume		0.000	00:20	
Air Blank		0.000	00:21	
Control Test		0.078	00:21	
Air Blank		0.000	00:21	
Diagnostics Check		OK	00:22	

Data Download Date/Time: 06/27/2018 16:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: VIERA

DATE/TIME OF TEST: 06/06/2018 23:20:50

Diagnostics Check		OK	23:06	
Air Blank		0.000	23:07	
Control Test		0.079	23:07	
Air Blank		0.000	23:07	
Subject Sample	# 1		23:11	No Sample Provided
Breath Volume		0.000	23:11	
Air Blank		0.000	23:11	
Air Blank		0.000	23:13	
Subject Sample	# 2	0.195	23:14	
Breath Volume		1.753	23:14	
Air Blank		0.000	23:15	
Air Blank		0.000	23:17	
Subject Sample	# 3	0.187	23:19	
Breath Volume		2.414	23:19	
Air Blank		0.000	23:19	
Control Test		0.078	23:20	
Air Blank		0.000	23:20	
Diagnostics Check		OK	23:20	

Data Download Date/Time: 06/07/2018 09:11

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAIA,MICHAEL,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: ARMAS

DATE/TIME OF TEST: 06/07/2018 03:09:35

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.078	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.196	03:05
Breath Volume		1.386	03:05
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.196	03:08
Breath Volume		1.738	03:08
Air Blank		0.000	03:08
Control Test		0.078	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 06/07/2018 09:11

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAIA,MICHAEL,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: SAMPSON

DATE/TIME OF TEST: 06/07/2018 03:33:59

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:31	
Control Test		0.076	03:31	
Air Blank		0.000	03:32	
Subject Sample	# 1		03:32	Subject Test Refused
Breath Volume		0.000	03:32	
Air Blank		0.000	03:33	
Control Test		0.079	03:33	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:33	

Data Download Date/Time: 06/07/2018 09:11

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEZ,ALEX
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: BUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 06/07/2018 22:02:03

Diagnostics Check		OK	21:55
Air Blank		0.000	21:56
Control Test		0.078	21:56
Air Blank		0.000	21:56
Subject Sample	# 1	0.117	21:57
Breath Volume		4.484	21:57
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.119	22:00
Breath Volume		4.402	22:00
Air Blank		0.000	22:01
Control Test		0.078	22:01
Air Blank		0.000	22:01
Diagnostics Check		OK	22:02

Data Download Date/Time: 07/11/2018 10:10

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/09/2018 03:57:10

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:54	
Control Test		0.079	03:54	
Air Blank		0.000	03:55	
Subject Sample	# 1	0.203	03:55	
Breath Volume		1.488	03:55	
Air Blank			03:55	RFI Detect
Air Blank		0.000	03:56	
Control Test		0.079	03:56	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 07/11/2018 10:10

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/09/2018 04:04:49

Diagnostics Check		OK	03:58
Air Blank		0.000	03:59
Control Test		0.078	03:59
Air Blank		0.000	04:00
Subject Sample	# 1	0.199	04:00
Breath Volume		1.519	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:02
Subject Sample	# 2	0.193	04:03
Breath Volume		1.695	04:03
Air Blank		0.000	04:03
Control Test		0.077	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 07/11/2018 10:10

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONZ,BRIAN,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: BATTAGLIA

DATE/TIME OF TEST: 06/10/2018 18:59:41

Diagnostics Check		OK	18:56	
Air Blank		0.000	18:57	
Control Test		0.078	18:57	
Air Blank		0.000	18:57	
Subject Sample	# 1		18:58	Subject Test Refused
Breath Volume		0.000	18:58	
Air Blank		0.000	18:58	
Control Test		0.078	18:59	
Air Blank		0.000	18:59	
Diagnostics Check		OK	18:59	

Data Download Date/Time: 07/11/2018 10:10

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 06/14/2018 00:23:33

Diagnostics Check		OK	00:17
Air Blank		0.000	00:17
Control Test		0.079	00:18
Air Blank		0.000	00:18
Subject Sample	# 1	0.246	00:19
Breath Volume		2.574	00:19
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.239	00:22
Breath Volume		1.894	00:22
Air Blank		0.000	00:22
Control Test		0.079	00:22
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 07/11/2018 10:10

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: WORKS

DATE/TIME OF TEST: 06/14/2018 00:58:51

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.078	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.216	00:54
Breath Volume		1.421	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.211	00:57
Breath Volume		1.746	00:57
Air Blank		0.000	00:57
Control Test		0.077	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 07/11/2018 10:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: DOYLE

DATE/TIME OF TEST: 06/14/2018 04:16:26

Diagnostics Check		OK	04:13	
Air Blank		0.000	04:13	
Control Test		0.078	04:14	
Air Blank		0.000	04:14	
Subject Sample	# 1		04:15	Subject Test Refused
Breath Volume		0.000	04:15	
Air Blank		0.000	04:15	
Control Test		0.078	04:15	
Air Blank		0.000	04:16	
Diagnostics Check		OK	04:16	

Data Download Date/Time: 07/11/2018 10:10

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/15/2018 03:40:40

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.078	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.169	03:35
Breath Volume		1.343	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.170	03:39
Breath Volume		1.648	03:39
Air Blank		0.000	03:39
Control Test		0.078	03:40
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 07/11/2018 10:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SIRENI,ERIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: PARKS

DATE/TIME OF TEST: 06/15/2018 22:56:28

Diagnostics Check		OK	22:49
Air Blank		0.000	22:49
Control Test		0.078	22:50
Air Blank		0.000	22:50
Subject Sample	# 1	0.041	22:52
Breath Volume		1.957	22:52
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.044	22:55
Breath Volume		2.093	22:55
Air Blank		0.000	22:55
Control Test		0.078	22:55
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 07/11/2018 10:10

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: ROSSO

DATE/TIME OF TEST: 06/16/2018 04:19:46

Diagnostics Check		OK	04:16	
Air Blank		0.000	04:16	
Control Test		0.077	04:16	
Air Blank		0.000	04:17	
Subject Sample	# 1		04:18	Subject Test Refused
Breath Volume		0.000	04:18	
Air Blank		0.000	04:18	
Control Test		0.077	04:19	
Air Blank		0.000	04:19	
Diagnostics Check		OK	04:19	

Data Download Date/Time: 07/11/2018 10:10

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/16/2018 20:19:26

Diagnostics Check		OK	20:11
Air Blank		0.000	20:12
Control Test		0.079	20:12
Air Blank		0.000	20:13
Subject Sample	# 1	0.207	20:14
Breath Volume		1.453	20:14
Air Blank		0.000	20:14
Air Blank		0.000	20:16
Subject Sample	# 2	0.224	20:17
Breath Volume		1.640	20:17
Air Blank		0.000	20:18
Control Test		0.078	20:18
Air Blank		0.000	20:19
Diagnostics Check		OK	20:19

Data Download Date/Time: 07/11/2018 10:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLOTA,V
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: BUCHNER

DATE/TIME OF TEST: 06/24/2018 00:53:13

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.079	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.077	00:47
Breath Volume		1.867	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.073	00:51
Breath Volume		2.906	00:51
Air Blank		0.000	00:52
Control Test		0.078	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 07/11/2018 10:10

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONZ,BRIAN,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KRONZ,BRIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 06/24/2018 01:57:03

Diagnostics Check		OK	01:51
Air Blank		0.000	01:51
Control Test		0.078	01:51
Air Blank		0.000	01:52
Subject Sample	# 1	0.144	01:52
Breath Volume		1.292	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.143	01:55
Breath Volume		1.238	01:55
Air Blank		0.000	01:56
Control Test		0.078	01:56
Air Blank		0.000	01:56
Diagnostics Check		OK	01:57

Data Download Date/Time: 07/11/2018 10:10

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: WEBER

DATE/TIME OF TEST: 06/24/2018 05:10:58

Diagnostics Check		OK	05:04
Air Blank		0.000	05:05
Control Test		0.078	05:05
Air Blank		0.000	05:05
Subject Sample	# 1	0.198	05:06
Breath Volume		2.261	05:06
Air Blank		0.000	05:07
Air Blank		0.000	05:08
Subject Sample	# 2	0.199	05:09
Breath Volume		3.554	05:09
Air Blank		0.000	05:10
Control Test		0.078	05:10
Air Blank		0.000	05:10
Diagnostics Check		OK	05:10

Data Download Date/Time: 07/11/2018 10:10

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EMILY,K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: STONE,JARED
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 06/24/2018 23:38:53

Diagnostics Check		OK	23:36	
Air Blank		0.000	23:36	
Control Test		0.079	23:36	
Air Blank		0.000	23:37	
Subject Sample	# 1		23:37	Subject Test Refused
Breath Volume		0.000	23:37	
Air Blank		0.000	23:37	
Control Test		0.078	23:38	
Air Blank		0.000	23:38	
Diagnostics Check		OK	23:38	

Data Download Date/Time: 07/11/2018 10:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONZ,BRIAN,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SAIA,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 06/25/2018 01:27:06

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.078	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.162	01:22
Breath Volume		1.609	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:24
Subject Sample	# 2	0.159	01:25
Breath Volume		2.070	01:25
Air Blank		0.000	01:26
Control Test		0.078	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 07/11/2018 10:10

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HAUSAM,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 06/27/2018 02:36:08

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.076	02:28
Air Blank		0.000	02:28
Subject Sample	# 1	0.231	02:30
Breath Volume		2.464	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.221	02:34
Breath Volume		1.695	02:34
Air Blank		0.000	02:35
Control Test		0.076	02:35
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 07/11/2018 10:10

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: KARAC

DATE/TIME OF TEST: 06/28/2018 00:18:15

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:15	
Control Test		0.079	00:15	
Air Blank		0.000	00:16	
Subject Sample	# 1		00:16	Subject Test Refused
Breath Volume		0.000	00:16	
Air Blank		0.000	00:17	
Control Test		0.078	00:17	
Air Blank		0.000	00:18	
Diagnostics Check		OK	00:18	

Data Download Date/Time: 07/11/2018 10:10

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: ADAMS-MILLIGAN

DATE/TIME OF TEST: 06/28/2018 02:31:36

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:29	
Control Test		0.078	02:29	
Air Blank		0.000	02:30	
Subject Sample	# 1		02:30	Subject Test Refused
Breath Volume		0.000	02:30	
Air Blank		0.000	02:30	
Control Test		0.078	02:31	
Air Blank		0.000	02:31	
Diagnostics Check		OK	02:31	

Data Download Date/Time: 07/11/2018 10:11

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 06/28/2018 04:36:11

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.078	04:27
Air Blank		0.000	04:27
Subject Sample	# 1	0.194	04:30
Breath Volume		1.757	04:30
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.195	04:34
Breath Volume		1.492	04:34
Air Blank		0.000	04:35
Control Test		0.077	04:35
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 07/11/2018 10:11

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE JR,SCOTT,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIERCE JR,SCOTT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: WOJCIK

DATE/TIME OF TEST: 06/29/2018 10:46:30

Diagnostics Check		OK	10:30	
Air Blank		0.000	10:30	
Control Test		0.079	10:30	
Air Blank		0.000	10:31	
Subject Sample	# 1		10:34	No Sample Provided
Breath Volume		0.000	10:34	
Air Blank		0.000	10:35	
Air Blank		0.000	10:37	
Subject Sample	# 2	0.000	10:40	
Breath Volume		1.671	10:40	
Air Blank		0.000	10:40	
Air Blank		0.000	10:42	
Subject Sample	# 3	0.000	10:45	
Breath Volume		1.585	10:45	
Air Blank		0.000	10:45	
Control Test		0.079	10:45	
Air Blank		0.000	10:46	
Diagnostics Check		OK	10:46	

Data Download Date/Time: 07/11/2018 10:11

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONZ,BRIAN,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: HORGAN

DATE/TIME OF TEST: 06/30/2018 19:53:20

Diagnostics Check		OK	19:47
Air Blank		0.000	19:47
Control Test		0.078	19:48
Air Blank		0.000	19:48
Subject Sample	# 1	0.258	19:48
Breath Volume		2.136	19:48
Air Blank		0.000	19:49
Air Blank		0.000	19:51
Subject Sample	# 2	0.269	19:51
Breath Volume		2.488	19:51
Air Blank		0.000	19:52
Control Test		0.077	19:52
Air Blank		0.000	19:53
Diagnostics Check		OK	19:53

Data Download Date/Time: 07/11/2018 10:11

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDONA,K,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: LAURENDINE

DATE/TIME OF TEST: 06/27/2018 20:29:44

Diagnostics Check		OK	20:22
Air Blank		0.000	20:23
Control Test		0.080	20:23
Air Blank		0.000	20:24
Subject Sample	# 1	0.146	20:24
Breath Volume		2.914	20:24
Air Blank		0.000	20:25
Air Blank		0.000	20:27
Subject Sample	# 2	0.130	20:28
Breath Volume		1.722	20:28
Air Blank		0.000	20:28
Control Test		0.080	20:29
Air Blank		0.000	20:29
Diagnostics Check		OK	20:29

Data Download Date/Time: 08/03/2018 08:13

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,B,K
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/30/2018 00:03:35

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:01	
Control Test		0.080	00:01	
Air Blank		0.000	00:02	
Subject Sample	# 1		00:02	Subject Test Refused
Breath Volume		0.000	00:02	
Air Blank		0.000	00:02	
Control Test		0.080	00:03	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 08/03/2018 08:13

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOLARD,R,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LACKSON,ROBERT
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:14

SUBJECT'S LAST NAME: SWAFFORD

DATE/TIME OF TEST: 06/30/2018 12:54:19

Diagnostics Check		OK	12:47
Air Blank		0.000	12:47
Control Test		0.079	12:47
Air Blank		0.000	12:48
Subject Sample	# 1	0.215	12:48
Breath Volume		2.164	12:48
Air Blank		0.000	12:49
Air Blank		0.000	12:51
Subject Sample	# 2	0.205	12:52
Breath Volume		1.500	12:52
Air Blank		0.000	12:53
Control Test		0.078	12:53
Air Blank		0.000	12:54
Diagnostics Check		OK	12:54

Data Download Date/Time: 08/03/2018 08:13

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SUSAN,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRISON,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: SCHLENKERT

DATE/TIME OF TEST: 06/02/2018 01:29:02

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.079	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.086	01:24
Breath Volume		1.605	01:24
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.087	01:27
Breath Volume		2.300	01:27
Air Blank		0.000	01:28
Control Test		0.079	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:29

Data Download Date/Time: 06/29/2018 08:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/02/2018 03:22:14

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.108	03:17
Breath Volume		1.421	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.108	03:20
Breath Volume		1.781	03:20
Air Blank		0.000	03:21
Control Test		0.077	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 06/29/2018 08:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SUSAN,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRISON,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/02/2018 04:23:26

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.078	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.156	04:18
Breath Volume		1.742	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.148	04:21
Breath Volume		1.242	04:21
Air Blank		0.000	04:22
Control Test		0.078	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 06/29/2018 08:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SUSAN,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,E
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: POUGH

DATE/TIME OF TEST: 06/03/2018 02:39:04

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:36	
Control Test		0.079	02:37	
Air Blank		0.000	02:37	
Subject Sample	# 1		02:37	Subject Test Refused
Breath Volume		0.000	02:37	
Air Blank		0.000	02:38	
Control Test		0.079	02:38	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 06/29/2018 08:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTH,KEVIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: CARRANZA,R
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 06/04/2018 22:30:26

Diagnostics Check		OK	22:23
Air Blank		0.000	22:23
Control Test		0.078	22:24
Air Blank		0.000	22:24
Subject Sample	# 1	0.000	22:25
Breath Volume		2.890	22:25
Air Blank		0.000	22:26
Air Blank		0.000	22:28
Subject Sample	# 2	0.000	22:29
Breath Volume		2.523	22:29
Air Blank		0.000	22:29
Control Test		0.079	22:29
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 06/29/2018 08:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORT,K,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WORTH,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SPICER

DATE/TIME OF TEST: 06/05/2018 22:04:25

Diagnostics Check		OK	21:58
Air Blank		0.000	21:58
Control Test		0.080	21:58
Air Blank		0.000	21:59
Subject Sample	# 1	0.134	21:59
Breath Volume		2.328	21:59
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.135	22:02
Breath Volume		2.304	22:02
Air Blank		0.000	22:03
Control Test		0.079	22:03
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 06/29/2018 08:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTH,K,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: LOW,B
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: RECTOR

DATE/TIME OF TEST: 06/06/2018 01:58:07

Diagnostics Check		OK	01:51
Air Blank		0.000	01:51
Control Test		0.080	01:51
Air Blank		0.000	01:52
Subject Sample	# 1	0.159	01:53
Breath Volume		1.878	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:55
Subject Sample	# 2	0.147	01:56
Breath Volume		1.167	01:56
Air Blank		0.000	01:57
Control Test		0.079	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 06/29/2018 08:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEEFE,JAYME,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BOGUSKY,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 06/07/2018 00:20:09

Diagnostics Check		OK	00:13
Air Blank		0.000	00:13
Control Test		0.079	00:14
Air Blank		0.000	00:14
Subject Sample	# 1	0.062	00:15
Breath Volume		1.382	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.065	00:18
Breath Volume		2.261	00:18
Air Blank		0.000	00:19
Control Test		0.077	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 06/29/2018 08:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,J,V
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARIN,A
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/09/2018 19:29:21

Diagnostics Check		OK	19:22
Air Blank		0.000	19:22
Control Test		0.078	19:22
Air Blank		0.000	19:23
Subject Sample	# 1	0.294	19:24
Breath Volume		1.601	19:24
Air Blank		0.000	19:24
Air Blank		0.000	19:26
Subject Sample	# 2	0.290	19:27
Breath Volume		2.296	19:27
Air Blank		0.000	19:28
Control Test		0.079	19:28
Air Blank		0.000	19:29
Diagnostics Check		OK	19:29

Data Download Date/Time: 06/29/2018 08:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNELL,COREY,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: DOUBERELY,C
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: LABOY DE JONGH

DATE/TIME OF TEST: 06/12/2018 02:38:23

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:35	
Control Test		0.078	02:36	
Air Blank		0.000	02:36	
Subject Sample	# 1		02:37	Subject Test Refused
Breath Volume		0.000	02:37	
Air Blank		0.000	02:37	
Control Test		0.079	02:37	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:38	

Data Download Date/Time: 06/29/2018 08:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: COOPER,GARY
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: GUERRA

DATE/TIME OF TEST: 06/15/2018 23:34:01

Diagnostics Check		OK	23:30	
Air Blank		0.000	23:31	
Control Test		0.079	23:31	
Air Blank		0.000	23:32	
Subject Sample	# 1		23:32	Subject Test Refused
Breath Volume		0.000	23:32	
Air Blank		0.000	23:33	
Control Test		0.078	23:33	
Air Blank		0.000	23:33	
Diagnostics Check		OK	23:34	

Data Download Date/Time: 06/29/2018 08:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BOGUSKY,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: GALVAN

DATE/TIME OF TEST: 06/16/2018 04:34:26

Diagnostics Check		OK	04:31	
Air Blank		0.000	04:31	
Control Test		0.078	04:31	
Air Blank		0.000	04:32	
Subject Sample	# 1		04:33	Subject Test Refused
Breath Volume		0.000	04:33	
Air Blank		0.000	04:33	
Control Test		0.079	04:33	
Air Blank		0.000	04:34	
Diagnostics Check		OK	04:34	

Data Download Date/Time: 06/29/2018 08:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAY,J,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MEE,T
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: S560796620310

DATE/TIME OF TEST: 06/19/2018 00:01:59

Diagnostics Check		OK	23:55
Air Blank		0.000	23:56
Control Test		0.079	23:56
Air Blank		0.000	23:56
Subject Sample	# 2	0.033	23:57
Breath Volume		1.777	23:57
Air Blank		0.000	23:58
Air Blank		0.000	23:59
Subject Sample	# 1	0.033	00:00
Breath Volume		1.828	00:00
Air Blank		0.000	00:01
Control Test		0.079	00:01
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 06/29/2018 08:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROE,ROBERT
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MASSEY

DATE/TIME OF TEST: 06/22/2018 18:55:30

Diagnostics Check		OK	18:49
Air Blank		0.000	18:49
Control Test		0.079	18:50
Air Blank		0.000	18:50
Subject Sample	# 1	0.000	18:51
Breath Volume		2.097	18:51
Air Blank		0.000	18:51
Air Blank		0.000	18:53
Subject Sample	# 2	0.000	18:54
Breath Volume		1.535	18:54
Air Blank		0.000	18:54
Control Test		0.079	18:54
Air Blank		0.000	18:55
Diagnostics Check		OK	18:55

Data Download Date/Time: 06/29/2018 08:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARIN,ADRIAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MASI

DATE/TIME OF TEST: 06/25/2018 02:07:25

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.078	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.027	02:02
Breath Volume		3.136	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.027	02:05
Breath Volume		2.675	02:05
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 06/29/2018 08:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEEFE,JAYME,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BENNETTE

DATE/TIME OF TEST: 06/27/2018 03:16:11

Diagnostics Check		OK	03:09
Air Blank		0.000	03:10
Control Test		0.078	03:10
Air Blank		0.000	03:10
Subject Sample	# 1	0.082	03:11
Breath Volume		2.636	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:13
Subject Sample	# 2	0.081	03:14
Breath Volume		2.867	03:14
Air Blank		0.000	03:15
Control Test		0.079	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 06/29/2018 08:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,J,V
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARIN,A
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/28/2018 04:34:27

Diagnostics Check		OK	04:27
Air Blank		0.000	04:27
Control Test		0.079	04:28
Air Blank		0.000	04:28
Subject Sample	# 1	0.156	04:29
Breath Volume		3.199	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.152	04:32
Breath Volume		1.824	04:32
Air Blank		0.000	04:33
Control Test		0.078	04:33
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 06/29/2018 08:31

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BURRESS,RYAN
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: HAWTHORNE

DATE/TIME OF TEST: 06/30/2018 19:37:13

Diagnostics Check		OK	19:30
Air Blank		0.000	19:31
Control Test		0.079	19:31
Air Blank		0.000	19:31
Subject Sample	# 1	0.234	19:32
Breath Volume		2.003	19:32
Air Blank		0.000	19:33
Air Blank		0.000	19:35
Subject Sample	# 2	0.218	19:35
Breath Volume		1.429	19:35
Air Blank		0.000	19:36
Control Test		0.078	19:36
Air Blank		0.000	19:37
Diagnostics Check		OK	19:37

Data Download Date/Time: 07/28/2018 22:25

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JESSUP,BRIAN,C
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: STRUNK,JOHN
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MISHKIN

DATE/TIME OF TEST: 06/02/2018 01:08:26

Diagnostics Check		OK	00:57	
Air Blank		0.000	00:57	
Control Test		0.079	00:57	
Air Blank		0.000	00:58	
Subject Sample	# 1		01:01	No Sample Provided
Breath Volume		0.000	01:01	
Air Blank		0.000	01:01	
Air Blank		0.000	01:03	
Subject Sample	# 2		01:07	No Sample Provided
Breath Volume		0.000	01:07	
Air Blank		0.000	01:07	
Control Test		0.080	01:07	
Air Blank		0.000	01:08	
Diagnostics Check		OK	01:08	

Data Download Date/Time: 06/29/2018 18:14

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: STUEHRENBERG,ERIK
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: TOPKIN

DATE/TIME OF TEST: 06/14/2018 00:18:05

Diagnostics Check		OK	00:12
Air Blank		0.000	00:12
Control Test		0.079	00:12
Air Blank		0.000	00:13
Subject Sample	# 1	0.000	00:13
Breath Volume		1.898	00:13
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.000	00:16
Breath Volume		4.910	00:16
Air Blank		0.000	00:17
Control Test		0.079	00:17
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 06/29/2018 18:14

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STUEHRENBERG,ERIK,C
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: STUEHRENBERG,ERIK
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: HULTIN

DATE/TIME OF TEST: 06/15/2018 00:35:40

Diagnostics Check		OK	00:29
Air Blank		0.000	00:29
Control Test		0.081	00:30
Air Blank		0.000	00:30
Subject Sample	# 1	0.115	00:31
Breath Volume		1.199	00:31
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.122	00:34
Breath Volume		1.574	00:34
Air Blank		0.000	00:34
Control Test		0.079	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 06/29/2018 18:14

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DI CINTIO,PATRICK,N
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: CORREA,MATTHEW
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 14:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: ADYDAN

DATE/TIME OF TEST: 06/24/2018 17:24:59

Diagnostics Check		OK	17:18
Air Blank		0.000	17:19
Control Test		0.081	17:19
Air Blank		0.000	17:19
Subject Sample	# 1	0.089	17:20
Breath Volume		2.503	17:20
Air Blank		0.000	17:21
Air Blank		0.000	17:22
Subject Sample	# 2	0.090	17:23
Breath Volume		2.304	17:23
Air Blank		0.000	17:24
Control Test		0.080	17:24
Air Blank		0.000	17:24
Diagnostics Check		OK	17:24

Data Download Date/Time: 06/29/2018 18:14

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,SHELIA,R
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: YOUNG,JOSHUA
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: FREDERICK

DATE/TIME OF TEST: 06/27/2018 03:25:06

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.077	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.230	03:20
Breath Volume		1.492	03:20
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.250	03:23
Breath Volume		1.875	03:23
Air Blank		0.000	03:24
Control Test		0.079	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 06/29/2018 18:14

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALMEIDA,CAMILO,J
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: ALVARADO DAMASO

DATE/TIME OF TEST: 06/27/2018 20:37:55

Diagnostics Check		OK	20:30
Air Blank		0.000	20:31
Control Test		0.080	20:31
Air Blank		0.000	20:32
Subject Sample	# 1	0.202	20:33
Breath Volume		1.367	20:33
Air Blank		0.000	20:34
Air Blank		0.000	20:35
Subject Sample	# 2	0.197	20:36
Breath Volume		1.105	20:36
Air Blank		0.000	20:37
Control Test		0.078	20:37
Air Blank		0.000	20:37
Diagnostics Check		OK	20:37

Data Download Date/Time: 07/16/2018 08:27

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MOSS,JUSTIN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/30/2018 03:20:22

Diagnostics Check		OK	03:14
Air Blank		0.000	03:14
Control Test		0.080	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.000	03:16
Breath Volume		2.636	03:16
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.000	03:19
Breath Volume		2.152	03:19
Air Blank		0.000	03:19
Control Test		0.079	03:19
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 07/16/2018 08:27

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JASON,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: MORENO,STEPHEN
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 05/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: CREWS

DATE/TIME OF TEST: 06/09/2018 05:15:16

Diagnostics Check		OK	05:03
Air Blank		0.000	05:04
Control Test		0.078	05:04
Air Blank		0.000	05:04
Subject Sample	# 1	0.178	05:06
Breath Volume		1.453	05:06
Air Blank		0.000	05:07
Air Blank		0.000	05:09
Subject Sample	# 2	0.200	05:10
Breath Volume		1.199	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 3	0.201	05:13
Breath Volume		2.089	05:13
Air Blank		0.000	05:14
Control Test		0.077	05:14
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 06/13/2018 14:41

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JASON,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSHUA
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 05/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: CLAXTON

DATE/TIME OF TEST: 06/11/2018 01:42:02

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.078	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.220	01:37
Breath Volume		1.964	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.203	01:40
Breath Volume		1.390	01:40
Air Blank		0.000	01:41
Control Test		0.078	01:41
Air Blank		0.000	01:41
Diagnostics Check		OK	01:42

Data Download Date/Time: 06/13/2018 14:41

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JASON,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSHUA
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 06/16/2018 02:58:16

Diagnostics Check		OK	02:55	
Air Blank		0.000	02:55	
Control Test		0.079	02:56	
Air Blank		0.000	02:56	
Subject Sample	# 1		02:56	Subject Test Refused
Breath Volume		0.000	02:56	
Air Blank		0.000	02:57	
Control Test		0.079	02:57	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 07/09/2018 07:25

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: SMITH,JASON
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 06/17/2018 22:59:17

Diagnostics Check		OK	22:56	
Air Blank		0.000	22:56	
Control Test		0.078	22:57	
Air Blank		0.000	22:57	
Subject Sample	# 1		22:57	Subject Test Refused
Breath Volume		0.000	22:57	
Air Blank		0.000	22:58	
Control Test		0.079	22:58	
Air Blank		0.000	22:59	
Diagnostics Check		OK	22:59	

Data Download Date/Time: 07/09/2018 07:25

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONINE,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: SUTPHIN,JULIE
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 06/02/2018 23:59:52

Diagnostics Check		OK	23:54	
Air Blank		0.000	23:54	
Control Test		0.079	23:54	
Air Blank		0.000	23:55	
Subject Sample	# 1		23:58	Interferent Detect
Breath Volume		1.417	23:58	
Air Blank			23:59	Purge Fail
Air Blank			23:59	Purge Fail

Data Download Date/Time: 06/27/2018 10:59

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONINE,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: SUTPHIN,JULIE
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 06/03/2018 00:09:55

Diagnostics Check		OK	00:03
Air Blank		0.000	00:03
Control Test		0.078	00:03
Air Blank		0.000	00:04
Subject Sample	# 1	0.210	00:05
Breath Volume		1.480	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:07
Subject Sample	# 2	0.205	00:08
Breath Volume		1.617	00:08
Air Blank		0.000	00:08
Control Test		0.079	00:09
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 06/27/2018 10:59

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORPHEE,MICHAEL,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: ORPHEE,MICHAEL
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 06/08/2018 03:17:09

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.013	03:11
Breath Volume		2.214	03:11
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.015	03:15
Breath Volume		2.566	03:15
Air Blank		0.000	03:16
Control Test		0.081	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 06/27/2018 10:59

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,KIMBERLY,L
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: MCDONALD,KIMBERLY
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 05/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 06/06/2018 20:02:49

Diagnostics Check		OK	19:56
Air Blank		0.000	19:57
Control Test		0.079	19:57
Air Blank		0.000	19:58
Subject Sample	# 1	0.014	19:58
Breath Volume		2.226	19:58
Air Blank		0.000	19:59
Air Blank		0.000	20:01
Subject Sample	# 2	0.014	20:01
Breath Volume		1.972	20:01
Air Blank		0.000	20:01
Control Test		0.079	20:02
Air Blank		0.000	20:02
Diagnostics Check		OK	20:02

Data Download Date/Time: 06/29/2018 09:38

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRAND,E,A
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: MYRAND,E
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 05/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: OLSEN

DATE/TIME OF TEST: 06/12/2018 18:25:43

Diagnostics Check		OK	18:19
Air Blank		0.000	18:19
Control Test		0.079	18:20
Air Blank		0.000	18:20
Subject Sample	# 1	0.088	18:21
Breath Volume		1.640	18:21
Air Blank		0.000	18:21
Air Blank		0.000	18:23
Subject Sample	# 2	0.087	18:24
Breath Volume		1.585	18:24
Air Blank		0.000	18:24
Control Test		0.079	18:25
Air Blank		0.000	18:25
Diagnostics Check		OK	18:25

Data Download Date/Time: 06/29/2018 09:38

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HADDOCK,CLAY,S
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: HADDOCK,CLAY
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 05/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: GINGERICH

DATE/TIME OF TEST: 06/18/2018 23:40:13

Diagnostics Check		OK	23:33
Air Blank		0.000	23:33
Control Test		0.079	23:34
Air Blank		0.000	23:34
Subject Sample	# 1	0.328	23:35
Breath Volume		3.125	23:35
Air Blank		0.000	23:35
Air Blank		0.000	23:37
Subject Sample	# 2	0.320	23:38
Breath Volume		2.777	23:38
Air Blank		0.000	23:39
Control Test		0.079	23:39
Air Blank		0.000	23:40
Diagnostics Check		OK	23:40

Data Download Date/Time: 06/29/2018 09:38

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: CHIRINO,YANDY
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 05/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: RUIZ ALONSO

DATE/TIME OF TEST: 06/24/2018 02:00:17

Diagnostics Check		OK	01:53
Air Blank		0.000	01:54
Control Test		0.075	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.217	01:55
Breath Volume		1.851	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.220	01:58
Breath Volume		1.359	01:58
Air Blank		0.000	01:59
Control Test		0.075	01:59
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 06/26/2018 11:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 05/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: IRIARTE

DATE/TIME OF TEST: 06/25/2018 03:17:49

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:11	
Control Test		0.075	03:11	
Air Blank		0.000	03:12	
Subject Sample	# 1	0.179	03:13	
Breath Volume		1.230	03:13	
Air Blank		0.000	03:13	
Air Blank		0.000	03:15	
Subject Sample	# 2	0.174	03:16	
Breath Volume		1.671	03:16	
Air Blank		0.000	03:16	
Control Test		0.074	03:17	Control Outside Tolerance
Air Blank		0.000	03:17	

Data Download Date/Time: 06/26/2018 11:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 05/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: IRIARTE

DATE/TIME OF TEST: 06/25/2018 03:28:41

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:21	
Control Test		0.075	03:21	
Air Blank		0.000	03:22	
Subject Sample	# 1	0.172	03:24	
Breath Volume		1.437	03:24	
Air Blank		0.000	03:24	
Air Blank		0.000	03:26	
Subject Sample	# 2	0.170	03:27	
Breath Volume		2.003	03:27	
Air Blank		0.000	03:27	
Control Test		0.074	03:28	Control Outside Tolerance
Air Blank		0.000	03:28	

Data Download Date/Time: 06/26/2018 11:09

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGEL,JASON,S
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAZIME,AL-HUSSEIN
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 05/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: AYERS

DATE/TIME OF TEST: 06/03/2018 02:11:04

Diagnostics Check		OK	02:04
Air Blank		0.000	02:05
Control Test		0.078	02:05
Air Blank		0.000	02:05
Subject Sample	# 1	0.164	02:06
Breath Volume		2.261	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.168	02:09
Breath Volume		2.480	02:09
Air Blank		0.000	02:10
Control Test		0.077	02:10
Air Blank		0.000	02:10
Diagnostics Check		OK	02:11

Data Download Date/Time: 06/06/2018 22:45

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HAZIME,AL
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 06/20/2018 23:57:31

Diagnostics Check		OK	23:51
Air Blank		0.000	23:51
Control Test		0.077	23:52
Air Blank		0.000	23:52
Subject Sample	# 1	0.234	23:53
Breath Volume		1.105	23:53
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.243	23:56
Breath Volume		2.304	23:56
Air Blank		0.000	23:56
Control Test		0.077	23:56
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 07/09/2018 22:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHERTY,DANIEL,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HAZIME,A
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: SPARKMAN

DATE/TIME OF TEST: 06/27/2018 03:23:09

Diagnostics Check		OK	03:19	
Air Blank		0.000	03:20	
Control Test		0.077	03:20	
Air Blank		0.000	03:21	
Subject Sample	# 1		03:21	Subject Test Refused
Breath Volume		0.000	03:21	
Air Blank		0.000	03:22	
Control Test		0.077	03:22	
Air Blank		0.000	03:23	
Diagnostics Check		OK	03:23	

Data Download Date/Time: 07/09/2018 22:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDOS,ROBERT,E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: RONDOS,ROBERT
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: ARYAI

DATE/TIME OF TEST: 06/27/2018 21:55:23

Diagnostics Check		OK	21:47
Air Blank		0.000	21:47
Control Test		0.076	21:47
Air Blank		0.000	21:48
Subject Sample	# 1	0.186	21:50
Breath Volume		2.218	21:50
Air Blank		0.000	21:51
Air Blank		0.000	21:53
Subject Sample	# 2	0.205	21:53
Breath Volume		1.535	21:53
Air Blank		0.000	21:54
Control Test		0.075	21:54
Air Blank		0.000	21:55
Diagnostics Check		OK	21:55

Data Download Date/Time: 07/30/2018 12:58

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBSON,JOHN,A
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: MATTHEWS,S
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:26

SUBJECT'S LAST NAME: PALO

DATE/TIME OF TEST: 06/30/2018 06:51:03

Diagnostics Check		OK	06:48	
Air Blank		0.000	06:48	
Control Test		0.078	06:48	
Air Blank		0.000	06:49	
Subject Sample	# 1		06:49	Subject Test Refused
Breath Volume		0.000	06:49	
Air Blank		0.000	06:50	
Control Test		0.078	06:50	
Air Blank		0.000	06:50	
Diagnostics Check		OK	06:51	

Data Download Date/Time: 07/30/2018 12:58

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 11:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:03

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/01/2018 12:38:29

Diagnostics Check		OK	12:32
Air Blank		0.000	12:32
Control Test		0.081	12:33
Air Blank		0.000	12:33
Subject Sample	# 1	0.000	12:34
Breath Volume		1.152	12:34
Air Blank		0.000	12:34
Air Blank		0.000	12:36
Subject Sample	# 2	0.000	12:37
Breath Volume		1.113	12:37
Air Blank		0.000	12:37
Control Test		0.081	12:37
Air Blank		0.000	12:38
Diagnostics Check		OK	12:38

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VICKERS,JAMES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 06/02/2018 00:30:35

Diagnostics Check		OK	00:27	
Air Blank		0.000	00:27	
Control Test		0.082	00:28	
Air Blank		0.000	00:28	
Subject Sample	# 1		00:29	Subject Test Refused
Breath Volume		0.000	00:29	
Air Blank		0.000	00:29	
Control Test		0.081	00:30	
Air Blank		0.000	00:30	
Diagnostics Check		OK	00:30	

Data Download Date/Time: 06/28/2018 22:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: BURNETTE

DATE/TIME OF TEST: 06/02/2018 16:11:44

Diagnostics Check		OK	16:00	
Air Blank		0.000	16:00	
Control Test		0.081	16:01	
Air Blank		0.000	16:01	
Subject Sample	# 1		16:04	No Sample Provided
Breath Volume		0.000	16:04	
Air Blank		0.000	16:05	
Air Blank		0.000	16:07	
Subject Sample	# 2		16:10	No Sample Provided
Breath Volume		0.000	16:10	
Air Blank		0.000	16:10	
Control Test		0.082	16:11	
Air Blank		0.000	16:11	
Diagnostics Check		OK	16:11	

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEARY,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: SPRINGER

DATE/TIME OF TEST: 06/02/2018 17:22:00

Diagnostics Check		OK	17:15
Air Blank		0.000	17:16
Control Test		0.082	17:16
Air Blank		0.000	17:17
Subject Sample	# 1	0.165	17:17
Breath Volume		1.972	17:17
Air Blank		0.000	17:18
Air Blank		0.000	17:19
Subject Sample	# 2	0.162	17:20
Breath Volume		1.554	17:20
Air Blank		0.000	17:21
Control Test		0.081	17:21
Air Blank		0.000	17:21
Diagnostics Check		OK	17:22

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: CONNER

DATE/TIME OF TEST: 06/03/2018 00:41:32

Diagnostics Check		OK	00:38	
Air Blank		0.000	00:38	
Control Test		0.081	00:39	
Air Blank		0.000	00:39	
Subject Sample	# 1		00:40	Subject Test Refused
Breath Volume		0.000	00:40	
Air Blank		0.000	00:40	
Control Test		0.081	00:40	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 06/28/2018 22:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: ISLER

DATE/TIME OF TEST: 06/03/2018 03:29:47

Diagnostics Check		OK	03:26	
Air Blank		0.000	03:27	
Control Test		0.081	03:27	
Air Blank		0.000	03:28	
Subject Sample	# 1		03:28	Subject Test Refused
Breath Volume		0.000	03:28	
Air Blank		0.000	03:28	
Control Test		0.082	03:29	
Air Blank		0.000	03:29	
Diagnostics Check		OK	03:29	

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBY,LARRY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: LEPINE

DATE/TIME OF TEST: 06/03/2018 04:11:31

Diagnostics Check		OK	04:04
Air Blank		0.000	04:05
Control Test		0.081	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.154	04:06
Breath Volume		1.507	04:06
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.155	04:09
Breath Volume		1.976	04:09
Air Blank		0.000	04:10
Control Test		0.081	04:10
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 06/03/2018 04:57:05

Diagnostics Check		OK	04:50
Air Blank		0.000	04:51
Control Test		0.081	04:51
Air Blank		0.000	04:52
Subject Sample	# 1	0.150	04:52
Breath Volume		3.468	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.146	04:55
Breath Volume		1.699	04:55
Air Blank		0.000	04:56
Control Test		0.080	04:56
Air Blank		0.000	04:56
Diagnostics Check		OK	04:57

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 14:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:38

SUBJECT'S LAST NAME: HULL

DATE/TIME OF TEST: 06/03/2018 15:12:44

Diagnostics Check		OK	14:59	
Air Blank		0.000	14:59	
Control Test		0.081	14:59	
Air Blank		0.000	15:00	
Subject Sample	# 1		15:03	No Sample Provided
Breath Volume		0.000	15:03	
Air Blank		0.000	15:04	
Air Blank		0.000	15:06	
Subject Sample	# 2	0.220	15:07	
Breath Volume		1.878	15:07	
Air Blank		0.000	15:08	
Air Blank		0.000	15:10	
Subject Sample	# 3	0.209	15:11	
Breath Volume		1.144	15:11	
Air Blank		0.000	15:11	
Control Test		0.080	15:12	
Air Blank		0.000	15:12	
Diagnostics Check		OK	15:12	

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELSHANS,DEAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DUNNING,RYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: BUI
ARREST TIME: 16:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: LIVESAY

DATE/TIME OF TEST: 06/03/2018 18:27:19

Diagnostics Check		OK	18:20
Air Blank		0.000	18:20
Control Test		0.081	18:20
Air Blank		0.000	18:21
Subject Sample	# 1	0.040	18:22
Breath Volume		1.562	18:22
Air Blank		0.000	18:23
Air Blank		0.000	18:25
Subject Sample	# 2	0.039	18:25
Breath Volume		2.250	18:25
Air Blank		0.000	18:26
Control Test		0.081	18:26
Air Blank		0.000	18:27
Diagnostics Check		OK	18:27

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: LUBE

DATE/TIME OF TEST: 06/03/2018 22:41:46

Diagnostics Check		OK	22:30	
Air Blank		0.000	22:30	
Control Test		0.081	22:31	
Air Blank		0.000	22:31	
Subject Sample	# 1		22:34	No Sample Provided
Breath Volume		0.000	22:34	
Air Blank		0.000	22:35	
Air Blank		0.000	22:37	
Subject Sample	# 2		22:40	No Sample Provided
Breath Volume		0.000	22:40	
Air Blank		0.000	22:40	
Control Test		0.081	22:41	
Air Blank		0.000	22:41	
Diagnostics Check		OK	22:41	

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GODFREY,DEPUTY C.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 06/04/2018 03:57:47

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.081	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.204	03:53
Breath Volume		2.859	03:53
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.220	03:56
Breath Volume		2.523	03:56
Air Blank		0.000	03:56
Control Test		0.080	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 06/05/2018 19:47:35

Diagnostics Check		OK	19:41
Air Blank		0.000	19:41
Control Test		0.082	19:42
Air Blank		0.000	19:42
Subject Sample	# 1	0.204	19:43
Breath Volume		1.589	19:43
Air Blank		0.000	19:43
Air Blank		0.000	19:45
Subject Sample	# 2	0.219	19:46
Breath Volume		1.144	19:46
Air Blank		0.000	19:46
Control Test		0.081	19:46
Air Blank		0.000	19:47
Diagnostics Check		OK	19:47

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: NORDEEN

DATE/TIME OF TEST: 06/06/2018 21:20:32

Diagnostics Check		OK	21:13
Air Blank		0.000	21:14
Control Test		0.082	21:14
Air Blank		0.000	21:14
Subject Sample	# 1	0.271	21:15
Breath Volume		1.773	21:15
Air Blank		0.000	21:16
Air Blank		0.000	21:18
Subject Sample	# 2	0.270	21:19
Breath Volume		2.023	21:19
Air Blank		0.000	21:19
Control Test		0.081	21:19
Air Blank		0.000	21:20
Diagnostics Check		OK	21:20

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: LONES

DATE/TIME OF TEST: 06/07/2018 00:14:50

Diagnostics Check		OK	00:08
Air Blank		0.000	00:09
Control Test		0.080	00:09
Air Blank		0.000	00:09
Subject Sample	# 1	0.071	00:10
Breath Volume		2.710	00:10
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 2	0.071	00:13
Breath Volume		2.695	00:13
Air Blank		0.000	00:13
Control Test		0.080	00:14
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: OSEGUERA

DATE/TIME OF TEST: 06/08/2018 01:02:35

Diagnostics Check		OK	00:59	
Air Blank		0.000	00:59	
Control Test		0.080	00:59	
Air Blank		0.000	01:00	
Subject Sample	# 1		01:01	Subject Test Refused
Breath Volume		0.000	01:01	
Air Blank		0.000	01:01	
Control Test		0.081	01:01	
Air Blank		0.000	01:02	
Diagnostics Check		OK	01:02	

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: MAMUNISKIS

DATE/TIME OF TEST: 06/08/2018 02:52:13

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.081	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.170	02:47
Breath Volume		1.960	02:47
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.154	02:50
Breath Volume		2.015	02:50
Air Blank		0.000	02:51
Control Test		0.080	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COOK,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 06/08/2018 05:00:19

Diagnostics Check		OK	04:50	
Air Blank		0.000	04:50	
Control Test		0.080	04:51	
Air Blank		0.000	04:51	
Subject Sample	# 1		04:54	No Sample Provided
Breath Volume		0.000	04:54	
Air Blank		0.000	04:55	
Air Blank		0.000	04:57	
Subject Sample	# 2		04:58	Subject Test Refused
Breath Volume		0.000	04:58	
Air Blank		0.000	04:59	
Control Test		0.079	04:59	
Air Blank		0.000	05:00	
Diagnostics Check		OK	05:00	

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: TASILLO

DATE/TIME OF TEST: 06/08/2018 23:05:05

Diagnostics Check		OK	22:53	
Air Blank		0.000	22:54	
Control Test		0.081	22:54	
Air Blank		0.000	22:54	
Subject Sample	# 1		22:58	No Sample Provided
Breath Volume		0.000	22:58	
Air Blank		0.000	22:58	
Air Blank		0.000	23:00	
Subject Sample	# 2		23:03	No Sample Provided
Breath Volume		0.000	23:03	
Air Blank		0.000	23:04	
Control Test		0.081	23:04	
Air Blank		0.000	23:04	
Diagnostics Check		OK	23:05	

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: LOMBA

DATE/TIME OF TEST: 06/09/2018 01:19:50

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:17	
Control Test		0.081	01:17	
Air Blank		0.000	01:18	
Subject Sample	# 1		01:18	Subject Test Refused
Breath Volume		0.000	01:18	
Air Blank		0.000	01:18	
Control Test		0.081	01:19	
Air Blank		0.000	01:19	
Diagnostics Check		OK	01:19	

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: DESTEFANO

DATE/TIME OF TEST: 06/09/2018 04:07:13

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:04	
Control Test		0.081	04:04	
Air Blank		0.000	04:05	
Subject Sample	# 1		04:05	Subject Test Refused
Breath Volume		0.000	04:05	
Air Blank		0.000	04:06	
Control Test		0.081	04:06	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: DELEO

DATE/TIME OF TEST: 06/09/2018 04:48:58

Diagnostics Check		OK	04:45	
Air Blank		0.000	04:46	
Control Test		0.080	04:46	
Air Blank		0.000	04:47	
Subject Sample	# 1		04:47	Subject Test Refused
Breath Volume		0.000	04:47	
Air Blank		0.000	04:48	
Control Test		0.081	04:48	
Air Blank		0.000	04:48	
Diagnostics Check		OK	04:48	

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAREY,KEVIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 06/09/2018 08:23:59

Diagnostics Check		OK	08:21	
Air Blank		0.000	08:21	
Control Test		0.081	08:21	
Air Blank		0.000	08:22	
Subject Sample	# 1		08:22	Subject Test Refused
Breath Volume		0.000	08:22	
Air Blank		0.000	08:23	
Control Test		0.081	08:23	
Air Blank		0.000	08:23	
Diagnostics Check		OK	08:23	

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAREY,KEVIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 16:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:36

SUBJECT'S LAST NAME: STEPP

DATE/TIME OF TEST: 06/09/2018 17:05:42

Diagnostics Check		OK	16:59
Air Blank		0.000	16:59
Control Test		0.081	17:00
Air Blank		0.000	17:00
Subject Sample	# 1	0.000	17:01
Breath Volume		1.765	17:01
Air Blank		0.000	17:01
Air Blank		0.000	17:03
Subject Sample	# 2	0.000	17:04
Breath Volume		2.089	17:04
Air Blank		0.000	17:04
Control Test		0.081	17:05
Air Blank		0.000	17:05
Diagnostics Check		OK	17:05

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: THAT

DATE/TIME OF TEST: 06/10/2018 02:47:36

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:36	
Control Test		0.081	02:36	
Air Blank		0.000	02:37	
Subject Sample	# 1		02:40	No Sample Provided
Breath Volume		0.000	02:40	
Air Blank		0.000	02:40	
Air Blank		0.000	02:42	
Subject Sample	# 2	0.183	02:43	
Breath Volume		1.457	02:43	
Air Blank		0.000	02:43	
Air Blank		0.000	02:45	
Subject Sample	# 3	0.176	02:46	
Breath Volume		1.398	02:46	
Air Blank		0.000	02:46	
Control Test		0.079	02:47	
Air Blank		0.000	02:47	
Diagnostics Check		OK	02:47	

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOITTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: KNEPP

DATE/TIME OF TEST: 06/10/2018 04:04:56

Diagnostics Check		OK	03:58
Air Blank		0.000	03:59
Control Test		0.081	03:59
Air Blank		0.000	04:00
Subject Sample	# 1	0.217	04:00
Breath Volume		2.175	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.219	04:03
Breath Volume		2.007	04:03
Air Blank		0.000	04:04
Control Test		0.080	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 06/28/2018 22:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,AARON,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NIXON,AARON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 06/10/2018 20:24:20

Diagnostics Check		OK	20:19	
Air Blank		0.000	20:20	
Control Test		0.080	20:20	
Air Blank		0.000	20:21	
Subject Sample	# 1		20:22	Subject Test Refused
Breath Volume		0.000	20:22	
Air Blank		0.000	20:23	
Control Test		0.080	20:23	
Air Blank		0.000	20:24	
Diagnostics Check		OK	20:24	

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBY,LARRY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: DORANTES HERNANDEZ

DATE/TIME OF TEST: 06/11/2018 19:45:15

Diagnostics Check		OK	19:42	
Air Blank		0.000	19:42	
Control Test		0.081	19:43	
Air Blank		0.000	19:43	
Subject Sample	# 1		19:43	Subject Test Refused
Breath Volume		0.000	19:43	
Air Blank		0.000	19:44	
Control Test		0.081	19:44	
Air Blank		0.000	19:45	
Diagnostics Check		OK	19:45	

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,M,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CONNOR,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: DUMOUCHEL

DATE/TIME OF TEST: 06/13/2018 17:52:03

Diagnostics Check	OK	17:51	
Air Blank		17:51	RFI Detect
Air Blank	0.000	17:52	

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,M,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CONNOR,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: DUMOUCHEL

DATE/TIME OF TEST: 06/13/2018 17:57:00

Diagnostics Check		OK	17:53	
Air Blank		0.000	17:53	
Control Test		0.081	17:54	
Air Blank		0.000	17:54	
Subject Sample	# 1		17:55	RFI Detect
Breath Volume		0.507	17:55	
Air Blank		0.000	17:56	
Control Test		0.081	17:56	
Air Blank		0.000	17:56	
Diagnostics Check		OK	17:57	

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BAUER

DATE/TIME OF TEST: 06/13/2018 22:54:46

Diagnostics Check		OK	22:51	
Air Blank		0.000	22:51	
Control Test		0.081	22:52	
Air Blank		0.000	22:52	
Subject Sample	# 1		22:53	Subject Test Refused
Breath Volume		0.000	22:53	
Air Blank		0.000	22:53	
Control Test		0.081	22:54	
Air Blank		0.000	22:54	
Diagnostics Check		OK	22:54	

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WHIRLE,CHAD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:17

SUBJECT'S LAST NAME: GIBEAULT

DATE/TIME OF TEST: 06/14/2018 11:45:21

Diagnostics Check		OK	11:37
Air Blank		0.000	11:37
Control Test		0.080	11:38
Air Blank		0.000	11:38
Subject Sample	# 1	0.283	11:40
Breath Volume		1.593	11:40
Air Blank		0.000	11:41
Air Blank		0.000	11:43
Subject Sample	# 2	0.297	11:43
Breath Volume		1.988	11:43
Air Blank		0.000	11:44
Control Test		0.079	11:44
Air Blank		0.000	11:45
Diagnostics Check		OK	11:45

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: HAGE

DATE/TIME OF TEST: 06/15/2018 03:48:42

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.082	03:42
Air Blank		0.000	03:43
Subject Sample	# 1	0.147	03:44
Breath Volume		2.761	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.145	03:47
Breath Volume		1.984	03:47
Air Blank		0.000	03:47
Control Test		0.080	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: ZERENDOW

DATE/TIME OF TEST: 06/15/2018 23:55:07

Diagnostics Check		OK	23:48
Air Blank		0.000	23:49
Control Test		0.081	23:49
Air Blank		0.000	23:50
Subject Sample	# 1	0.169	23:50
Breath Volume		2.605	23:50
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.174	23:53
Breath Volume		3.000	23:53
Air Blank		0.000	23:54
Control Test		0.079	23:54
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: JUDD

DATE/TIME OF TEST: 06/16/2018 03:41:10

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:39	
Control Test		0.081	03:39	
Air Blank		0.000	03:40	
Subject Sample	# 1		03:40	RFI Detect
Breath Volume		0.000	03:40	
Air Blank			03:40	RFI Detect
Air Blank		0.000	03:41	

Data Download Date/Time: 06/28/2018 22:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: JUDD

DATE/TIME OF TEST: 06/16/2018 03:45:29

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:43	
Control Test		0.081	03:43	
Air Blank		0.000	03:43	
Subject Sample	# 1		03:44	Subject Test Refused
Breath Volume		0.000	03:44	
Air Blank		0.000	03:44	
Control Test		0.080	03:44	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 06/28/2018 22:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: YARBOROUGH

DATE/TIME OF TEST: 06/16/2018 04:05:39

Diagnostics Check		OK	03:51	
Air Blank		0.000	03:52	
Control Test		0.081	03:52	
Air Blank		0.000	03:52	
Subject Sample	# 1	0.000	03:56	Volume Not Met
Breath Volume		0.242	03:56	
Air Blank		0.000	03:56	
Air Blank		0.000	03:58	
Subject Sample	# 2	0.000	04:00	
Breath Volume		2.453	04:00	
Air Blank		0.000	04:01	
Air Blank		0.000	04:03	
Subject Sample	# 3	0.000	04:04	
Breath Volume		1.253	04:04	
Air Blank		0.000	04:04	
Control Test		0.080	04:05	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:05	

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GOEGHOGAN,L
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: PIPPIN

DATE/TIME OF TEST: 06/16/2018 22:55:47

Diagnostics Check		OK	22:46
Air Blank		0.000	22:47
Control Test		0.081	22:47
Air Blank		0.000	22:48
Subject Sample	# 1	0.206	22:51
Breath Volume		1.656	22:51
Air Blank		0.000	22:52
Air Blank		0.000	22:53
Subject Sample	# 2	0.212	22:54
Breath Volume		1.230	22:54
Air Blank		0.000	22:54
Control Test		0.078	22:55
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 06/28/2018 22:29

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 06/17/2018 02:35:55

Diagnostics Check		OK	02:24	
Air Blank		0.000	02:25	
Control Test		0.081	02:25	
Air Blank		0.000	02:25	
Subject Sample	# 1		02:28	No Sample Provided
Breath Volume		0.000	02:28	
Air Blank		0.000	02:29	
Air Blank		0.000	02:31	
Subject Sample	# 2		02:34	No Sample Provided
Breath Volume		0.000	02:34	
Air Blank		0.000	02:35	
Control Test		0.081	02:35	
Air Blank		0.000	02:35	
Diagnostics Check		OK	02:35	

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COOK,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: AXTELL

DATE/TIME OF TEST: 06/17/2018 03:49:07

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.081	03:43
Air Blank		0.000	03:43
Subject Sample	# 1	0.133	03:44
Breath Volume		1.902	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.145	03:47
Breath Volume		1.835	03:47
Air Blank		0.000	03:48
Control Test		0.080	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: BENDRIS

DATE/TIME OF TEST: 06/17/2018 04:01:41

Diagnostics Check		OK	03:58	
Air Blank		0.000	03:59	
Control Test		0.080	03:59	
Air Blank		0.000	04:00	
Subject Sample	# 1		04:00	Subject Test Refused
Breath Volume		0.000	04:00	
Air Blank		0.000	04:00	
Control Test		0.080	04:01	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOD,COREY,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHULER,JARRET
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: CABRAL-CUPERY

DATE/TIME OF TEST: 06/18/2018 07:27:40

Diagnostics Check		OK	07:21
Air Blank		0.000	07:21
Control Test		0.081	07:21
Air Blank		0.000	07:22
Subject Sample	# 1	0.040	07:23
Breath Volume		1.664	07:23
Air Blank		0.000	07:23
Air Blank		0.000	07:25
Subject Sample	# 2	0.039	07:26
Breath Volume		1.570	07:26
Air Blank		0.000	07:26
Control Test		0.080	07:27
Air Blank		0.000	07:27
Diagnostics Check		OK	07:27

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,AARON,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NIXON,AARON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: FALLON

DATE/TIME OF TEST: 06/19/2018 21:50:51

Diagnostics Check		OK	21:43
Air Blank		0.000	21:43
Control Test		0.081	21:44
Air Blank		0.000	21:44
Subject Sample	# 1	0.152	21:45
Breath Volume		1.433	21:45
Air Blank		0.000	21:46
Air Blank		0.000	21:48
Subject Sample	# 2	0.153	21:49
Breath Volume		1.500	21:49
Air Blank		0.000	21:49
Control Test		0.081	21:50
Air Blank		0.000	21:50
Diagnostics Check		OK	21:50

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: TSE

DATE/TIME OF TEST: 06/21/2018 03:17:28

Diagnostics Check		OK	03:06	
Air Blank		0.000	03:06	
Control Test		0.081	03:06	
Air Blank		0.000	03:07	
Subject Sample	# 1		03:10	No Sample Provided
Breath Volume		0.000	03:10	
Air Blank		0.000	03:10	
Air Blank		0.000	03:12	
Subject Sample	# 2		03:16	No Sample Provided
Breath Volume		0.000	03:16	
Air Blank		0.000	03:16	
Control Test		0.081	03:16	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:17	

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 06/22/2018 00:45:44

Diagnostics Check		OK	00:42	
Air Blank		0.000	00:43	
Control Test		0.081	00:43	
Air Blank		0.000	00:44	
Subject Sample	# 1		00:44	Subject Test Refused
Breath Volume		0.000	00:44	
Air Blank		0.000	00:44	
Control Test		0.081	00:45	
Air Blank		0.000	00:45	
Diagnostics Check		OK	00:45	

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 06/22/2018 01:54:27

Diagnostics Check		OK	01:48
Air Blank		0.000	01:49
Control Test		0.081	01:49
Air Blank		0.000	01:49
Subject Sample	# 1	0.008	01:50
Breath Volume		2.292	01:50
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.008	01:53
Breath Volume		2.148	01:53
Air Blank		0.000	01:53
Control Test		0.082	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: VALENTE

DATE/TIME OF TEST: 06/22/2018 22:55:32

Diagnostics Check		OK	22:52	
Air Blank		0.000	22:53	
Control Test		0.082	22:53	
Air Blank		0.000	22:54	
Subject Sample	# 1		22:54	Subject Test Refused
Breath Volume		0.000	22:54	
Air Blank		0.000	22:54	
Control Test		0.081	22:54	
Air Blank		0.000	22:55	
Diagnostics Check		OK	22:55	

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: KORPELA

DATE/TIME OF TEST: 06/23/2018 01:34:29

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.081	01:28
Air Blank		0.000	01:29
Subject Sample	# 1	0.082	01:30
Breath Volume		1.667	01:30
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.082	01:33
Breath Volume		2.527	01:33
Air Blank		0.000	01:33
Control Test		0.081	01:33
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: CONCEPCION

DATE/TIME OF TEST: 06/23/2018 03:45:04

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.081	03:38
Air Blank		0.000	03:39
Subject Sample	# 1	0.150	03:39
Breath Volume		1.421	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.145	03:43
Breath Volume		1.691	03:43
Air Blank		0.000	03:44
Control Test		0.081	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:45

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: ARNTZEN

DATE/TIME OF TEST: 06/23/2018 05:21:05

Diagnostics Check		OK	05:14
Air Blank		0.000	05:15
Control Test		0.081	05:15
Air Blank		0.000	05:15
Subject Sample	# 1	0.136	05:16
Breath Volume		2.812	05:16
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 2	0.132	05:19
Breath Volume		2.070	05:19
Air Blank		0.000	05:20
Control Test		0.081	05:20
Air Blank		0.000	05:20
Diagnostics Check		OK	05:21

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: GALATUAR

DATE/TIME OF TEST: 06/24/2018 02:35:25

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.081	02:30
Air Blank		0.000	02:30
Subject Sample	# 1	0.129	02:31
Breath Volume		2.898	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.127	02:33
Breath Volume		2.691	02:33
Air Blank		0.000	02:34
Control Test		0.081	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEIB,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MENGLE

DATE/TIME OF TEST: 06/24/2018 03:21:29

Diagnostics Check		OK	03:12	
Air Blank		0.000	03:13	
Control Test		0.081	03:13	
Air Blank		0.000	03:13	
Subject Sample	# 1	0.020	03:16	Volume Not Met
Breath Volume		0.289	03:16	
Air Blank		0.000	03:17	
Air Blank		0.000	03:19	
Subject Sample	# 2		03:20	Subject Test Refused
Breath Volume		0.000	03:20	
Air Blank		0.000	03:20	
Control Test		0.080	03:20	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: CORDERO

DATE/TIME OF TEST: 06/24/2018 04:10:41

Diagnostics Check		OK	04:03
Air Blank		0.000	04:04
Control Test		0.081	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.216	04:05
Breath Volume		1.312	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.227	04:09
Breath Volume		1.652	04:09
Air Blank		0.000	04:09
Control Test		0.081	04:10
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 06/24/2018 05:12:42

Diagnostics Check		OK	05:06
Air Blank		0.000	05:06
Control Test		0.081	05:06
Air Blank		0.000	05:07
Subject Sample	# 1	0.131	05:08
Breath Volume		1.214	05:08
Air Blank		0.000	05:08
Air Blank		0.000	05:10
Subject Sample	# 2	0.133	05:11
Breath Volume		1.285	05:11
Air Blank		0.000	05:11
Control Test		0.081	05:12
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SWINDLE

DATE/TIME OF TEST: 06/25/2018 02:02:47

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.081	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.180	01:58
Breath Volume		3.210	01:58
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.183	02:01
Breath Volume		3.136	02:01
Air Blank		0.000	02:01
Control Test		0.081	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,AARON,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NIXON,AARON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 06:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:32

SUBJECT'S LAST NAME: GRENIER

DATE/TIME OF TEST: 06/26/2018 07:33:42

Diagnostics Check		OK	07:27
Air Blank		0.000	07:27
Control Test		0.082	07:28
Air Blank		0.000	07:28
Subject Sample	# 1	0.168	07:29
Breath Volume		2.738	07:29
Air Blank		0.000	07:29
Air Blank		0.000	07:31
Subject Sample	# 2	0.160	07:32
Breath Volume		2.351	07:32
Air Blank		0.000	07:32
Control Test		0.081	07:33
Air Blank		0.000	07:33
Diagnostics Check		OK	07:33

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBY,LARRY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/26/2018 22:40:31

Diagnostics Check		OK	22:27	
Air Blank		0.000	22:28	
Control Test		0.082	22:28	
Air Blank		0.000	22:29	
Subject Sample	# 1	0.156	22:32	Volume Not Met
Breath Volume		0.222	22:32	
Air Blank		0.000	22:33	
Air Blank		0.000	22:34	
Subject Sample	# 2	0.164	22:35	
Breath Volume		1.632	22:35	
Air Blank		0.000	22:36	
Air Blank		0.000	22:38	
Subject Sample	# 3	0.176	22:38	
Breath Volume		1.132	22:38	
Air Blank		0.000	22:39	
Control Test		0.079	22:39	
Air Blank		0.000	22:40	
Diagnostics Check		OK	22:40	

Data Download Date/Time: 06/28/2018 22:29

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 06/28/2018 01:54:15

Diagnostics Check		OK	01:50	
Air Blank		0.000	01:50	
Control Test		0.082	01:51	
Air Blank		0.000	01:51	
Subject Sample	# 1		01:52	Subject Test Refused
Breath Volume		0.000	01:52	
Air Blank		0.000	01:53	
Control Test		0.082	01:53	
Air Blank		0.000	01:54	
Diagnostics Check		OK	01:54	

Data Download Date/Time: 06/28/2018 22:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: HAMM

DATE/TIME OF TEST: 06/29/2018 03:54:33

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.081	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.152	03:50
Breath Volume		2.234	03:50
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.148	03:53
Breath Volume		1.277	03:53
Air Blank		0.000	03:53
Control Test		0.080	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 07/26/2018 00:37

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANSEN,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: AUVIL

DATE/TIME OF TEST: 06/29/2018 19:53:12

Diagnostics Check		OK	19:45
Air Blank		0.000	19:46
Control Test		0.081	19:46
Air Blank		0.000	19:46
Subject Sample	# 1	0.108	19:47
Breath Volume		1.562	19:47
Air Blank		0.000	19:48
Air Blank		0.000	19:50
Subject Sample	# 2	0.104	19:51
Breath Volume		1.117	19:51
Air Blank		0.000	19:52
Control Test		0.081	19:52
Air Blank		0.000	19:53
Diagnostics Check		OK	19:53

Data Download Date/Time: 07/26/2018 00:37

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: BOHNSACK

DATE/TIME OF TEST: 06/29/2018 22:33:44

Diagnostics Check		OK	22:27
Air Blank		0.000	22:27
Control Test		0.081	22:28
Air Blank		0.000	22:28
Subject Sample	# 1	0.000	22:29
Breath Volume		1.777	22:29
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.000	22:32
Breath Volume		1.929	22:32
Air Blank		0.000	22:32
Control Test		0.081	22:33
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 07/26/2018 00:37

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: PRUEITT

DATE/TIME OF TEST: 06/30/2018 00:14:50

Diagnostics Check		OK	00:08
Air Blank		0.000	00:09
Control Test		0.081	00:09
Air Blank		0.000	00:09
Subject Sample	# 1	0.165	00:10
Breath Volume		2.796	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:12
Subject Sample	# 2	0.175	00:13
Breath Volume		2.675	00:13
Air Blank		0.000	00:13
Control Test		0.080	00:14
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 07/26/2018 00:37

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 06/30/2018 03:56:31

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:53	
Control Test		0.081	03:53	
Air Blank		0.000	03:54	
Subject Sample	# 1		03:55	Subject Test Refused
Breath Volume		0.000	03:55	
Air Blank		0.000	03:55	
Control Test		0.082	03:55	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 07/26/2018 00:37

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: POD

DATE/TIME OF TEST: 06/30/2018 05:20:41

Diagnostics Check		OK	05:12
Air Blank		0.000	05:12
Control Test		0.081	05:13
Air Blank		0.000	05:13
Subject Sample	# 1	0.186	05:16
Breath Volume		1.921	05:16
Air Blank		0.000	05:16
Air Blank		0.000	05:18
Subject Sample	# 2	0.176	05:19
Breath Volume		1.609	05:19
Air Blank		0.000	05:19
Control Test		0.081	05:20
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 07/26/2018 00:37

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,AARON,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NIXON,AARON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG728303
DRY GAS EXPIRY DATE: 10/10/2019
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: MOSS

DATE/TIME OF TEST: 06/30/2018 05:46:57

Diagnostics Check		OK	05:42	
Air Blank		0.000	05:43	
Control Test		0.081	05:43	
Air Blank		0.000	05:44	
Subject Sample	# 1		05:45	Subject Test Refused
Breath Volume		0.000	05:45	
Air Blank		0.000	05:46	
Control Test		0.081	05:46	
Air Blank		0.000	05:46	
Diagnostics Check		OK	05:46	

Data Download Date/Time: 07/26/2018 00:37

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BRETON,JESSIKA
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 06/22/2018 04:29:48

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.079	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.099	04:25
Breath Volume		2.785	04:25
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.100	04:28
Breath Volume		2.421	04:28
Air Blank		0.000	04:28
Control Test		0.079	04:29
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 07/13/2018 11:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: COGBURN

DATE/TIME OF TEST: 06/23/2018 02:43:56

Diagnostics Check		OK	02:37
Air Blank		0.000	02:37
Control Test		0.077	02:38
Air Blank		0.000	02:38
Subject Sample	# 1	0.066	02:39
Breath Volume		2.554	02:39
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.068	02:42
Breath Volume		2.417	02:42
Air Blank		0.000	02:43
Control Test		0.077	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 07/13/2018 11:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ORTIZ,JEFFREY
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 08:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:41

SUBJECT'S LAST NAME: VALENTIN

DATE/TIME OF TEST: 06/23/2018 10:10:23

Diagnostics Check		OK	10:03
Air Blank		0.000	10:04
Control Test		0.078	10:04
Air Blank		0.000	10:05
Subject Sample	# 1	0.160	10:05
Breath Volume		1.167	10:05
Air Blank		0.000	10:06
Air Blank		0.000	10:08
Subject Sample	# 2	0.147	10:08
Breath Volume		1.195	10:08
Air Blank		0.000	10:09
Control Test		0.078	10:09
Air Blank		0.000	10:10
Diagnostics Check		OK	10:10

Data Download Date/Time: 07/13/2018 11:26

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JASON
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: LEWIS JR

DATE/TIME OF TEST: 06/24/2018 00:12:16

Diagnostics Check		OK	00:04
Air Blank		0.000	00:05
Control Test		0.078	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.109	00:07
Breath Volume		1.175	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.100	00:10
Breath Volume		1.210	00:10
Air Blank		0.000	00:11
Control Test		0.077	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 07/13/2018 11:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOUGHRAN,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ZWEIFEL

DATE/TIME OF TEST: 06/24/2018 01:12:29

Diagnostics Check		OK	01:06
Air Blank		0.000	01:06
Control Test		0.079	01:07
Air Blank		0.000	01:07
Subject Sample	# 1	0.000	01:08
Breath Volume		2.371	01:08
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.000	01:11
Breath Volume		1.300	01:11
Air Blank		0.000	01:11
Control Test		0.079	01:11
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 07/13/2018 11:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HENNESSY,JONATHAN
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 06/24/2018 03:54:51

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.078	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.190	03:47
Breath Volume		1.203	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.190	03:53
Breath Volume		1.304	03:53
Air Blank		0.000	03:53
Control Test		0.078	03:54
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 07/13/2018 11:26

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JASON
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: QUIROZ

DATE/TIME OF TEST: 06/24/2018 04:18:49

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.078	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.145	04:11
Breath Volume		1.636	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.142	04:17
Breath Volume		1.308	04:17
Air Blank		0.000	04:17
Control Test		0.078	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 07/13/2018 11:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: GONZALES BENAVIDES

DATE/TIME OF TEST: 06/24/2018 05:54:26

Diagnostics Check		OK	05:47
Air Blank		0.000	05:47
Control Test		0.078	05:48
Air Blank		0.000	05:48
Subject Sample	# 1	0.161	05:49
Breath Volume		2.238	05:49
Air Blank		0.000	05:50
Air Blank		0.000	05:52
Subject Sample	# 2	0.170	05:52
Breath Volume		3.902	05:52
Air Blank		0.000	05:53
Control Test		0.078	05:53
Air Blank		0.000	05:54
Diagnostics Check		OK	05:54

Data Download Date/Time: 07/13/2018 11:26

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:31

SUBJECT'S LAST NAME: MONTES

DATE/TIME OF TEST: 06/24/2018 06:58:12

Diagnostics Check		OK	06:51
Air Blank		0.000	06:52
Control Test		0.078	06:52
Air Blank		0.000	06:53
Subject Sample	# 1	0.120	06:53
Breath Volume		3.203	06:53
Air Blank		0.000	06:54
Air Blank		0.000	06:56
Subject Sample	# 2	0.118	06:56
Breath Volume		2.167	06:56
Air Blank		0.000	06:57
Control Test		0.078	06:57
Air Blank		0.000	06:58
Diagnostics Check		OK	06:58

Data Download Date/Time: 07/13/2018 11:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GOMEZ,OSVALDO
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/25/2018 05:35:25

Diagnostics Check		OK	05:28
Air Blank		0.000	05:28
Control Test		0.078	05:29
Air Blank		0.000	05:29
Subject Sample	# 1	0.138	05:30
Breath Volume		1.707	05:30
Air Blank		0.000	05:31
Air Blank		0.000	05:33
Subject Sample	# 2	0.139	05:33
Breath Volume		2.605	05:33
Air Blank		0.000	05:34
Control Test		0.078	05:34
Air Blank		0.000	05:35
Diagnostics Check		OK	05:35

Data Download Date/Time: 07/13/2018 11:26

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STRZALKOWSKI,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BRASHERS

DATE/TIME OF TEST: 06/27/2018 03:04:03

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.079	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.209	02:59
Breath Volume		1.875	02:59
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.216	03:02
Breath Volume		1.359	03:02
Air Blank		0.000	03:03
Control Test		0.078	03:03
Air Blank		0.000	03:03
Diagnostics Check		OK	03:04

Data Download Date/Time: 07/13/2018 11:26

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOUGHRAN,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: D'ALIA

DATE/TIME OF TEST: 06/27/2018 04:42:54

Diagnostics Check		OK	04:34
Air Blank		0.000	04:34
Control Test		0.079	04:35
Air Blank		0.000	04:35
Subject Sample	# 1	0.000	04:38
Breath Volume		1.187	04:38
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.000	04:41
Breath Volume		1.441	04:41
Air Blank		0.000	04:42
Control Test		0.078	04:42
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 07/13/2018 11:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: MCCLOUD

DATE/TIME OF TEST: 06/29/2018 04:55:43

Diagnostics Check		OK	04:47
Air Blank		0.000	04:47
Control Test		0.079	04:48
Air Blank		0.000	04:48
Subject Sample	# 1	0.169	04:51
Breath Volume		1.550	04:51
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.177	04:54
Breath Volume		1.546	04:54
Air Blank		0.000	04:54
Control Test		0.077	04:55
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 07/13/2018 11:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VASQUEZ,FABIO
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MONZON

DATE/TIME OF TEST: 06/30/2018 00:43:28

Diagnostics Check		OK	00:37
Air Blank		0.000	00:37
Control Test		0.079	00:37
Air Blank		0.000	00:38
Subject Sample	# 1	0.200	00:39
Breath Volume		1.636	00:39
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.190	00:42
Breath Volume		1.226	00:42
Air Blank		0.000	00:42
Control Test		0.078	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 07/13/2018 11:26

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: MELLA

DATE/TIME OF TEST: 06/30/2018 01:23:54

Diagnostics Check		OK	01:17
Air Blank		0.000	01:18
Control Test		0.079	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.094	01:19
Breath Volume		3.972	01:19
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.094	01:22
Breath Volume		4.308	01:22
Air Blank		0.000	01:23
Control Test		0.077	01:23
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 07/13/2018 11:26

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KARNES,BRIAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: RUIZ CARRILLO

DATE/TIME OF TEST: 06/03/2018 07:17:23

Diagnostics Check		OK	07:10
Air Blank		0.000	07:11
Control Test		0.078	07:11
Air Blank		0.000	07:11
Subject Sample	# 1	0.152	07:12
Breath Volume		1.160	07:12
Air Blank		0.000	07:13
Air Blank		0.000	07:15
Subject Sample	# 2	0.154	07:15
Breath Volume		1.250	07:15
Air Blank		0.000	07:16
Control Test		0.079	07:16
Air Blank		0.000	07:17
Diagnostics Check		OK	07:17

Data Download Date/Time: 06/27/2018 21:33

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,RHOBBY,D
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: JENKINS,RHOBBY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: MCCLURG

DATE/TIME OF TEST: 06/03/2018 18:30:21

Diagnostics Check		OK	18:18	
Air Blank		0.000	18:19	
Control Test		0.079	18:19	
Air Blank		0.000	18:20	
Subject Sample	# 1		18:23	No Sample Provided
Breath Volume		0.000	18:23	
Air Blank		0.000	18:23	
Air Blank		0.000	18:25	
Subject Sample	# 2		18:28	No Sample Provided
Breath Volume		0.000	18:28	
Air Blank		0.000	18:29	
Control Test		0.079	18:29	
Air Blank		0.000	18:30	
Diagnostics Check		OK	18:30	

Data Download Date/Time: 06/27/2018 21:33

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: GAMBILL

DATE/TIME OF TEST: 06/09/2018 02:47:08

Diagnostics Check	OK	02:46	
Air Blank		02:46	RFI Detect
Air Blank	0.000	02:47	

Data Download Date/Time: 06/27/2018 21:33

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: GAMBILL

DATE/TIME OF TEST: 06/09/2018 03:00:09

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:49	
Control Test		0.079	02:49	
Air Blank		0.000	02:50	
Subject Sample	# 1		02:53	No Sample Provided
Breath Volume		0.000	02:53	
Air Blank		0.000	02:53	
Air Blank		0.000	02:55	
Subject Sample	# 2		02:58	No Sample Provided
Breath Volume		0.000	02:58	
Air Blank		0.000	02:59	
Control Test		0.079	02:59	
Air Blank		0.000	03:00	
Diagnostics Check		OK	03:00	

Data Download Date/Time: 06/27/2018 21:33

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: GAMBILL

DATE/TIME OF TEST: 06/09/2018 03:09:46

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.079	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.181	03:04
Breath Volume		1.343	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.166	03:08
Breath Volume		2.683	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 06/27/2018 21:33

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: MOZDY

DATE/TIME OF TEST: 06/14/2018 02:32:41

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.153	02:27
Breath Volume		2.066	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.163	02:31
Breath Volume		2.093	02:31
Air Blank		0.000	02:31
Control Test		0.078	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 06/27/2018 21:34

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REED,CHAD
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: SJODIN

DATE/TIME OF TEST: 06/16/2018 19:04:06

Diagnostics Check		OK	18:56
Air Blank		0.000	18:57
Control Test		0.079	18:57
Air Blank		0.000	18:58
Subject Sample	# 1	0.133	18:59
Breath Volume		1.437	18:59
Air Blank		0.000	19:00
Air Blank		0.000	19:01
Subject Sample	# 2	0.128	19:02
Breath Volume		1.968	19:02
Air Blank		0.000	19:03
Control Test		0.079	19:03
Air Blank		0.000	19:04
Diagnostics Check		OK	19:04

Data Download Date/Time: 06/27/2018 21:34

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMONDS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: EDMONDS,RICHARD
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: KOLLER

DATE/TIME OF TEST: 06/23/2018 00:54:14

Diagnostics Check		OK	00:48
Air Blank		0.000	00:48
Control Test		0.079	00:48
Air Blank		0.000	00:49
Subject Sample	# 1	0.191	00:49
Breath Volume		2.636	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.177	00:52
Breath Volume		1.996	00:52
Air Blank		0.000	00:53
Control Test		0.078	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 06/27/2018 21:34

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHMOND,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: DELGADO,ANASTASIA
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: SISCO

DATE/TIME OF TEST: 06/26/2018 10:28:44

Diagnostics Check		OK	10:22
Air Blank		0.000	10:22
Control Test		0.079	10:22
Air Blank		0.000	10:23
Subject Sample	# 1	0.259	10:23
Breath Volume		2.722	10:23
Air Blank		0.000	10:24
Air Blank		0.000	10:26
Subject Sample	# 2	0.258	10:27
Breath Volume		1.589	10:27
Air Blank		0.000	10:27
Control Test		0.078	10:28
Air Blank		0.000	10:28
Diagnostics Check		OK	10:28

Data Download Date/Time: 06/27/2018 21:34

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: OCHOA

DATE/TIME OF TEST: 06/29/2018 23:25:45

Diagnostics Check		OK	23:18
Air Blank		0.000	23:19
Control Test		0.079	23:19
Air Blank		0.000	23:20
Subject Sample	# 1	0.212	23:21
Breath Volume		1.578	23:21
Air Blank		0.000	23:21
Air Blank		0.000	23:23
Subject Sample	# 2	0.199	23:24
Breath Volume		2.023	23:24
Air Blank		0.000	23:24
Control Test		0.079	23:25
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 07/20/2018 20:08

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 06/30/2018 00:28:00

Diagnostics Check		OK	00:25	
Air Blank		0.000	00:25	
Control Test		0.079	00:25	
Air Blank		0.000	00:26	
Subject Sample	# 1		00:26	Subject Test Refused
Breath Volume		0.000	00:26	
Air Blank		0.000	00:27	
Control Test		0.078	00:27	
Air Blank		0.000	00:27	
Diagnostics Check		OK	00:28	

Data Download Date/Time: 07/20/2018 20:08

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHMOND,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ASBURY,CHIRS
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:41

SUBJECT'S LAST NAME: MCGRATH

DATE/TIME OF TEST: 06/30/2018 10:35:06

Diagnostics Check		OK	10:28
Air Blank		0.000	10:29
Control Test		0.079	10:29
Air Blank		0.000	10:29
Subject Sample	# 1	0.325	10:30
Breath Volume		2.976	10:30
Air Blank		0.000	10:31
Air Blank		0.000	10:32
Subject Sample	# 2	0.324	10:33
Breath Volume		2.808	10:33
Air Blank		0.000	10:34
Control Test		0.078	10:34
Air Blank		0.000	10:35
Diagnostics Check		OK	10:35

Data Download Date/Time: 07/20/2018 20:08

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: SINGER

DATE/TIME OF TEST: 06/30/2018 23:23:52

Diagnostics Check		OK	23:17
Air Blank		0.000	23:17
Control Test		0.079	23:18
Air Blank		0.000	23:18
Subject Sample	# 1	0.105	23:19
Breath Volume		2.058	23:19
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.104	23:22
Breath Volume		1.417	23:22
Air Blank		0.000	23:22
Control Test		0.078	23:23
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 07/20/2018 20:08

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,CHRIS,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LAPOINTE,B
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 05/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 06/01/2018 03:53:54

Diagnostics Check		OK	03:37	
Air Blank		0.000	03:37	
Control Test		0.079	03:37	
Air Blank		0.000	03:38	
Subject Sample	# 1	0.209	03:41	
Breath Volume		1.273	03:41	
Air Blank		0.000	03:41	
Air Blank		0.000	03:43	
Subject Sample	# 2	0.187	03:46	Volume Not Met
Breath Volume		0.968	03:46	
Air Blank		0.000	03:47	
Air Blank		0.000	03:49	
Subject Sample	# 3	0.193	03:52	Slope Not Level
Breath Volume		1.156	03:52	
Air Blank		0.000	03:53	
Control Test		0.079	03:53	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 06/27/2018 19:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LAPOINTE,B
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 05/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HUSTED

DATE/TIME OF TEST: 06/10/2018 03:51:14

Diagnostics Check	OK	03:50	
Air Blank		03:50	RFI Detect
Air Blank	0.000	03:51	

Data Download Date/Time: 06/27/2018 19:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LAPOINTE,B
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 05/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HUSTED

DATE/TIME OF TEST: 06/10/2018 04:00:10

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.079	03:55
Breath Volume		3.960	03:55
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.071	03:58
Breath Volume		1.445	03:58
Air Blank		0.000	03:59
Control Test		0.080	03:59
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 06/27/2018 19:07

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,G,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MOSLEY,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: RIVER

DATE/TIME OF TEST: 06/23/2018 10:25:08

Diagnostics Check		OK	10:10	
Air Blank		0.000	10:11	
Control Test		0.079	10:11	
Air Blank		0.000	10:11	
Subject Sample	# 1	0.161	10:14	
Breath Volume		1.273	10:14	
Air Blank		0.000	10:14	
Air Blank		0.000	10:16	
Subject Sample	# 2	0.170	10:19	Volume Not Met
Breath Volume		0.261	10:19	
Air Blank		0.000	10:20	
Air Blank		0.000	10:22	
Subject Sample	# 3	0.161	10:23	
Breath Volume		1.597	10:23	
Air Blank		0.000	10:24	
Control Test		0.078	10:24	
Air Blank		0.000	10:25	
Diagnostics Check		OK	10:25	

Data Download Date/Time: 06/27/2018 19:07

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,CHRIS,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MALLEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: STURCH

DATE/TIME OF TEST: 06/24/2018 02:40:11

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:29	
Control Test		0.079	02:29	
Air Blank		0.000	02:30	
Subject Sample	# 1		02:33	No Sample Provided
Breath Volume		0.000	02:33	
Air Blank		0.000	02:33	
Air Blank		0.000	02:35	
Subject Sample	# 2		02:38	No Sample Provided
Breath Volume		0.000	02:38	
Air Blank		0.000	02:39	
Control Test		0.080	02:39	
Air Blank		0.000	02:40	
Diagnostics Check		OK	02:40	

Data Download Date/Time: 06/27/2018 19:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,G,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BATTON,BATTON
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 10:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: EARLY

DATE/TIME OF TEST: 06/28/2018 11:19:12

Diagnostics Check		OK	11:13
Air Blank		0.000	11:13
Control Test		0.079	11:13
Air Blank		0.000	11:14
Subject Sample	# 1	0.058	11:14
Breath Volume		3.160	11:14
Air Blank		0.000	11:15
Air Blank		0.000	11:17
Subject Sample	# 2	0.056	11:17
Breath Volume		1.964	11:17
Air Blank		0.000	11:18
Control Test		0.079	11:18
Air Blank		0.000	11:19
Diagnostics Check		OK	11:19

Data Download Date/Time: 07/26/2018 05:37

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEITH,HOOLEY,L
BREATH TEST OPERATOR AGENCY: Dade City PD
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 06/19/2018 23:13:39

Diagnostics Check		OK	23:07
Air Blank		0.000	23:07
Control Test		0.078	23:07
Air Blank		0.000	23:08
Subject Sample	# 1	0.203	23:09
Breath Volume		2.105	23:09
Air Blank		0.000	23:09
Air Blank		0.000	23:11
Subject Sample	# 2	0.206	23:12
Breath Volume		2.230	23:12
Air Blank		0.000	23:12
Control Test		0.079	23:13
Air Blank		0.000	23:13
Diagnostics Check		OK	23:13

Data Download Date/Time: 06/20/2018 11:29

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHERTY,DANIEL,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DAUGHERTY,DANIEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: ZARATE RUIZ

DATE/TIME OF TEST: 06/09/2018 02:02:24

Diagnostics Check		OK	01:56
Air Blank		0.000	01:56
Control Test		0.078	01:56
Air Blank		0.000	01:57
Subject Sample	# 1	0.232	01:57
Breath Volume		1.777	01:57
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.228	02:00
Breath Volume		1.976	02:00
Air Blank		0.000	02:01
Control Test		0.076	02:01
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 06/21/2018 19:49

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,JOHN,T
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: VAZQUEZ,EMMANUEL
ARRESTING OFFICER AGENCY: Bowling Green PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 06/17/2018 07:25:35

Diagnostics Check		OK	07:19
Air Blank		0.000	07:19
Control Test		0.076	07:20
Air Blank		0.000	07:20
Subject Sample	# 1	0.148	07:21
Breath Volume		4.437	07:21
Air Blank		0.000	07:21
Air Blank		0.000	07:23
Subject Sample	# 2	0.151	07:24
Breath Volume		3.937	07:24
Air Blank		0.000	07:24
Control Test		0.076	07:24
Air Blank		0.000	07:25
Diagnostics Check		OK	07:25

Data Download Date/Time: 06/26/2018 16:59

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: GAMBLE

DATE/TIME OF TEST: 06/02/2018 22:07:38

Diagnostics Check		OK	22:01
Air Blank		0.000	22:01
Control Test		0.079	22:01
Air Blank		0.000	22:02
Subject Sample	# 1	0.166	22:03
Breath Volume		2.648	22:03
Air Blank		0.000	22:03
Air Blank		0.000	22:05
Subject Sample	# 2	0.154	22:06
Breath Volume		1.925	22:06
Air Blank		0.000	22:06
Control Test		0.078	22:07
Air Blank		0.000	22:07
Diagnostics Check		OK	22:07

Data Download Date/Time: 06/07/2018 09:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: DHEENDSA

DATE/TIME OF TEST: 06/03/2018 02:25:43

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:22	
Control Test		0.078	02:23	
Air Blank		0.000	02:23	
Subject Sample	# 1		02:24	Subject Test Refused
Breath Volume		0.000	02:24	
Air Blank		0.000	02:24	
Control Test		0.079	02:25	
Air Blank		0.000	02:25	
Diagnostics Check		OK	02:25	

Data Download Date/Time: 06/07/2018 09:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: DHEENDSA

DATE/TIME OF TEST: 06/03/2018 02:39:23

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.079	02:30
Air Blank		0.000	02:31
Subject Sample	# 1	0.183	02:31
Breath Volume		1.457	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.205	02:34
Breath Volume		2.621	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 3	0.200	02:37
Breath Volume		2.441	02:37
Air Blank		0.000	02:38
Control Test		0.077	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 06/07/2018 09:01

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: COPAS

DATE/TIME OF TEST: 06/08/2018 03:16:26

Diagnostics Check		OK	03:13	
Air Blank		0.000	03:14	
Control Test		0.079	03:14	
Air Blank		0.000	03:14	
Subject Sample	# 1		03:15	Subject Test Refused
Breath Volume		0.000	03:15	
Air Blank		0.000	03:15	
Control Test		0.078	03:15	
Air Blank		0.000	03:16	
Diagnostics Check		OK	03:16	

Data Download Date/Time: 07/11/2018 10:15

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: SLAWITSCHKA

DATE/TIME OF TEST: 06/10/2018 04:25:37

Diagnostics Check		OK	04:18
Air Blank		0.000	04:18
Control Test		0.079	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.176	04:20
Breath Volume		1.136	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.178	04:24
Breath Volume		1.207	04:24
Air Blank		0.000	04:24
Control Test		0.078	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 07/11/2018 10:15

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: CLENDENING

DATE/TIME OF TEST: 06/16/2018 02:40:45

Diagnostics Check		OK	02:34
Air Blank		0.000	02:34
Control Test		0.078	02:34
Air Blank		0.000	02:35
Subject Sample	# 1	0.219	02:36
Breath Volume		1.984	02:36
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.213	02:39
Breath Volume		2.117	02:39
Air Blank		0.000	02:39
Control Test		0.077	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 07/11/2018 10:15

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 06/16/2018 22:32:58

Diagnostics Check		OK	22:26
Air Blank		0.000	22:27
Control Test		0.078	22:27
Air Blank		0.000	22:28
Subject Sample	# 1	0.243	22:28
Breath Volume		2.035	22:28
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.237	22:31
Breath Volume		2.035	22:31
Air Blank		0.000	22:32
Control Test		0.079	22:32
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 07/11/2018 10:15

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: STOUGH

DATE/TIME OF TEST: 06/23/2018 00:11:08

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.079	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.182	00:06
Breath Volume		2.308	00:06
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.182	00:09
Breath Volume		2.296	00:09
Air Blank		0.000	00:10
Control Test		0.078	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 07/11/2018 10:15

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: CARMONA

DATE/TIME OF TEST: 06/23/2018 02:37:22

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:34	
Control Test		0.078	02:35	
Air Blank		0.000	02:35	
Subject Sample	# 1		02:36	Subject Test Refused
Breath Volume		0.000	02:36	
Air Blank		0.000	02:36	
Control Test		0.078	02:36	
Air Blank		0.000	02:37	
Diagnostics Check		OK	02:37	

Data Download Date/Time: 07/11/2018 10:15

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: MCCLAIN

DATE/TIME OF TEST: 06/28/2018 19:28:51

Diagnostics Check		OK	19:25	
Air Blank		0.000	19:25	
Control Test		0.079	19:26	
Air Blank		0.000	19:26	
Subject Sample	# 1		19:27	Subject Test Refused
Breath Volume		0.000	19:27	
Air Blank		0.000	19:27	
Control Test		0.080	19:28	
Air Blank		0.000	19:28	
Diagnostics Check		OK	19:28	

Data Download Date/Time: 07/11/2018 10:15

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,V
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: STANDLEY

DATE/TIME OF TEST: 06/30/2018 01:16:17

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.079	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.128	01:11
Breath Volume		3.328	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.129	01:14
Breath Volume		2.347	01:14
Air Blank		0.000	01:15
Control Test		0.078	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 07/11/2018 10:15

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/30/2018 23:59:41

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.079	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.224	23:55
Breath Volume		1.390	23:55
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.235	23:58
Breath Volume		1.832	23:58
Air Blank		0.000	23:58
Control Test		0.078	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 07/11/2018 10:15

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: ZEGEER

DATE/TIME OF TEST: 06/01/2018 05:57:39

Diagnostics Check		OK	05:46	
Air Blank		0.000	05:46	
Control Test		0.080	05:47	
Air Blank		0.000	05:47	
Subject Sample	# 1		05:50	No Sample Provided
Breath Volume		0.000	05:50	
Air Blank		0.000	05:51	
Air Blank		0.000	05:53	
Subject Sample	# 2		05:56	No Sample Provided
Breath Volume		0.000	05:56	
Air Blank		0.000	05:56	
Control Test		0.080	05:57	
Air Blank		0.000	05:57	
Diagnostics Check		OK	05:57	

Data Download Date/Time: 06/21/2018 12:51

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/01/2018 23:07:57

Diagnostics Check		OK	23:04	
Air Blank		0.000	23:05	
Control Test		0.080	23:05	
Air Blank		0.000	23:06	
Subject Sample	# 1		23:06	Subject Test Refused
Breath Volume		0.000	23:06	
Air Blank		0.000	23:07	
Control Test		0.080	23:07	
Air Blank		0.000	23:07	
Diagnostics Check		OK	23:07	

Data Download Date/Time: 06/21/2018 12:51

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: FURLONG

DATE/TIME OF TEST: 06/02/2018 20:00:24

Diagnostics Check		OK	19:53
Air Blank		0.000	19:54
Control Test		0.080	19:54
Air Blank		0.000	19:55
Subject Sample	# 1	0.148	19:56
Breath Volume		1.179	19:56
Air Blank		0.000	19:56
Air Blank		0.000	19:58
Subject Sample	# 2	0.150	19:58
Breath Volume		1.214	19:58
Air Blank		0.000	19:59
Control Test		0.080	19:59
Air Blank		0.000	20:00
Diagnostics Check		OK	20:00

Data Download Date/Time: 06/21/2018 12:51

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: STRANO

DATE/TIME OF TEST: 06/03/2018 22:07:18

Diagnostics Check		OK	22:03	
Air Blank		0.000	22:03	
Control Test		0.080	22:04	
Air Blank		0.000	22:04	
Subject Sample	# 1		22:05	Subject Test Refused
Breath Volume		0.000	22:05	
Air Blank		0.000	22:06	
Control Test		0.079	22:06	
Air Blank		0.000	22:07	
Diagnostics Check		OK	22:07	

Data Download Date/Time: 06/21/2018 12:51

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: MILNER

DATE/TIME OF TEST: 06/04/2018 01:15:46

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.080	01:10
Air Blank		0.000	01:10
Subject Sample	# 1	0.142	01:11
Breath Volume		2.097	01:11
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.135	01:14
Breath Volume		1.621	01:14
Air Blank		0.000	01:14
Control Test		0.080	01:15
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 06/21/2018 12:51

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,ROBERT,C
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: COOK,ROBERT
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:52

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 06/06/2018 12:19:45

Diagnostics Check		OK	12:13
Air Blank		0.000	12:13
Control Test		0.081	12:14
Air Blank		0.000	12:14
Subject Sample	# 1	0.000	12:15
Breath Volume		1.457	12:15
Air Blank		0.000	12:15
Air Blank		0.000	12:17
Subject Sample	# 2	0.000	12:18
Breath Volume		1.503	12:18
Air Blank		0.000	12:18
Control Test		0.081	12:19
Air Blank		0.000	12:19
Diagnostics Check		OK	12:19

Data Download Date/Time: 06/21/2018 12:51

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 06/07/2018 00:42:40

Diagnostics Check		OK	00:37	
Air Blank		0.000	00:37	
Control Test		0.081	00:37	
Air Blank		0.000	00:38	
Subject Sample	# 1		00:41	Subject Test Refused
Breath Volume		0.000	00:41	
Air Blank		0.000	00:41	
Control Test		0.080	00:42	
Air Blank		0.000	00:42	
Diagnostics Check		OK	00:42	

Data Download Date/Time: 06/21/2018 12:51

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: MCALEER

DATE/TIME OF TEST: 06/09/2018 02:56:40

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.076	02:52
Breath Volume		2.203	02:52
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.080	02:55
Breath Volume		2.582	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 06/21/2018 12:51

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: RICKER

DATE/TIME OF TEST: 06/09/2018 07:24:26

Diagnostics Check		OK	07:21	
Air Blank		0.000	07:21	
Control Test		0.080	07:22	
Air Blank		0.000	07:22	
Subject Sample	# 1		07:23	Subject Test Refused
Breath Volume		0.000	07:23	
Air Blank		0.000	07:23	
Control Test		0.080	07:23	
Air Blank		0.000	07:24	
Diagnostics Check		OK	07:24	

Data Download Date/Time: 06/21/2018 12:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 06/13/2018 21:31:56

Diagnostics Check		OK	21:25
Air Blank		0.000	21:26
Control Test		0.080	21:26
Air Blank		0.000	21:26
Subject Sample	# 1	0.223	21:27
Breath Volume		2.312	21:27
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.220	21:30
Breath Volume		2.347	21:30
Air Blank		0.000	21:31
Control Test		0.079	21:31
Air Blank		0.000	21:31
Diagnostics Check		OK	21:31

Data Download Date/Time: 06/21/2018 12:51

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: CHEEK

DATE/TIME OF TEST: 06/13/2018 22:45:09

Diagnostics Check		OK	22:42	
Air Blank		0.000	22:42	
Control Test		0.080	22:43	
Air Blank		0.000	22:43	
Subject Sample	# 1		22:43	Subject Test Refused
Breath Volume		0.000	22:43	
Air Blank		0.000	22:44	
Control Test		0.080	22:44	
Air Blank		0.000	22:45	
Diagnostics Check		OK	22:45	

Data Download Date/Time: 06/21/2018 12:51

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: CIEPLY

DATE/TIME OF TEST: 06/14/2018 03:34:03

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:31	
Control Test		0.080	03:31	
Air Blank		0.000	03:32	
Subject Sample	# 1		03:32	Subject Test Refused
Breath Volume		0.000	03:32	
Air Blank		0.000	03:33	
Control Test		0.079	03:33	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 06/21/2018 12:51

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BRADEN

DATE/TIME OF TEST: 06/16/2018 01:39:29

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.080	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.000	01:35
Breath Volume		2.835	01:35
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.000	01:38
Breath Volume		1.421	01:38
Air Blank		0.000	01:38
Control Test		0.080	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 06/21/2018 12:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 06/16/2018 04:35:53

Diagnostics Check		OK	04:28
Air Blank		0.000	04:28
Control Test		0.080	04:29
Air Blank		0.000	04:29
Subject Sample	# 1	0.196	04:30
Breath Volume		1.652	04:30
Air Blank		0.000	04:31
Air Blank		0.000	04:32
Subject Sample	# 2	0.186	04:34
Breath Volume		1.554	04:34
Air Blank		0.000	04:34
Control Test		0.079	04:35
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 06/21/2018 12:51

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: COWAN

DATE/TIME OF TEST: 06/17/2018 01:33:10

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.080	01:25
Air Blank		0.000	01:25
Subject Sample	# 1	0.000	01:28
Breath Volume		2.777	01:28
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.000	01:31
Breath Volume		2.390	01:31
Air Blank		0.000	01:32
Control Test		0.080	01:32
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 06/21/2018 12:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: FEDERICO

DATE/TIME OF TEST: 06/21/2018 03:01:46

Diagnostics Check		OK	02:58	
Air Blank		0.000	02:59	
Control Test		0.080	02:59	
Air Blank		0.000	03:00	
Subject Sample	# 1		03:00	Subject Test Refused
Breath Volume		0.000	03:00	
Air Blank		0.000	03:00	
Control Test		0.081	03:01	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

Data Download Date/Time: 06/21/2018 12:51

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:32

SUBJECT'S LAST NAME: PEREZ-TREJO

DATE/TIME OF TEST: 06/22/2018 07:08:34

Diagnostics Check		OK	07:00
Air Blank		0.000	07:01
Control Test		0.080	07:01
Air Blank		0.000	07:02
Subject Sample	# 1	0.198	07:04
Breath Volume		1.617	07:04
Air Blank		0.000	07:04
Air Blank		0.000	07:06
Subject Sample	# 2	0.187	07:07
Breath Volume		2.031	07:07
Air Blank		0.000	07:07
Control Test		0.079	07:07
Air Blank		0.000	07:08
Diagnostics Check		OK	07:08

Data Download Date/Time: 07/03/2018 13:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: GRONER

DATE/TIME OF TEST: 06/24/2018 01:50:11

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:47	
Control Test		0.080	01:47	
Air Blank		0.000	01:48	
Subject Sample	# 1		01:48	Subject Test Refused
Breath Volume		0.000	01:48	
Air Blank		0.000	01:49	
Control Test		0.079	01:49	
Air Blank		0.000	01:50	
Diagnostics Check		OK	01:50	

Data Download Date/Time: 07/03/2018 13:46

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 06/24/2018 04:41:29

Diagnostics Check		OK	04:28	
Air Blank		0.000	04:28	
Control Test		0.080	04:29	
Air Blank		0.000	04:29	
Subject Sample	# 1	0.184	04:32	Volume Not Met
Breath Volume		0.617	04:32	
Air Blank		0.000	04:33	
Air Blank		0.000	04:35	
Subject Sample	# 2	0.179	04:36	
Breath Volume		1.574	04:36	
Air Blank		0.000	04:36	
Air Blank		0.000	04:38	
Subject Sample	# 3	0.183	04:39	
Breath Volume		1.929	04:39	
Air Blank		0.000	04:40	
Control Test		0.079	04:40	
Air Blank		0.000	04:41	
Diagnostics Check		OK	04:41	

Data Download Date/Time: 07/03/2018 13:47

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,B
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: ASBURY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:31

SUBJECT'S LAST NAME: BOGGS

DATE/TIME OF TEST: 06/24/2018 13:59:02

Diagnostics Check		OK	13:52
Air Blank		0.000	13:52
Control Test		0.079	13:53
Air Blank		0.000	13:53
Subject Sample	# 1	0.212	13:54
Breath Volume		1.656	13:54
Air Blank		0.000	13:54
Air Blank		0.000	13:56
Subject Sample	# 2	0.205	13:57
Breath Volume		1.585	13:57
Air Blank		0.000	13:58
Control Test		0.079	13:58
Air Blank		0.000	13:58
Diagnostics Check		OK	13:59

Data Download Date/Time: 07/03/2018 13:47

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: WESTFALL

DATE/TIME OF TEST: 06/28/2018 23:02:39

Diagnostics Check		OK	22:56
Air Blank		0.000	22:56
Control Test		0.080	22:56
Air Blank		0.000	22:57
Subject Sample	# 1	0.000	22:57
Breath Volume		4.003	22:57
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.000	23:01
Breath Volume		3.992	23:01
Air Blank		0.000	23:01
Control Test		0.080	23:02
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 07/03/2018 13:47

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 06/28/2018 23:49:18

Diagnostics Check		OK	23:41
Air Blank		0.000	23:42
Control Test		0.080	23:42
Air Blank		0.000	23:43
Subject Sample	# 1	0.229	23:44
Breath Volume		1.359	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.236	23:47
Breath Volume		1.199	23:47
Air Blank		0.000	23:48
Control Test		0.079	23:48
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 07/03/2018 13:47

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,ADAM,H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: POTTER,ADAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: SUGGS

DATE/TIME OF TEST: 06/30/2018 23:57:14

Diagnostics Check		OK	23:54	
Air Blank		0.000	23:54	
Control Test		0.077	23:55	
Air Blank		0.000	23:55	
Subject Sample	# 1		23:55	Subject Test Refused
Breath Volume		0.000	23:55	
Air Blank		0.000	23:56	
Control Test		0.078	23:56	
Air Blank		0.000	23:57	
Diagnostics Check		OK	23:57	

Data Download Date/Time: 07/27/2018 10:07

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: BUKATA,MARIE
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 05/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A2
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/02/2018 01:05:21

Diagnostics Check		OK	00:58
Air Blank		0.000	00:59
Control Test		0.076	00:59
Air Blank		0.000	00:59
Subject Sample	# 1	0.000	01:00
Breath Volume		2.558	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.000	01:03
Breath Volume		2.660	01:03
Air Blank		0.000	01:04
Control Test		0.077	01:04
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 06/23/2018 14:44

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 05/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A2
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: VALDEZ

DATE/TIME OF TEST: 06/16/2018 03:17:41

Diagnostics Check		OK	03:06	
Air Blank		0.000	03:06	
Control Test		0.076	03:06	
Air Blank		0.000	03:07	
Subject Sample	# 1	0.233	03:10	Volume Not Met
Breath Volume		0.152	03:10	
Air Blank		0.000	03:11	
Air Blank		0.000	03:12	
Subject Sample	# 2	0.239	03:16	Volume Not Met
Breath Volume		0.699	03:16	
Air Blank		0.000	03:16	
Control Test		0.075	03:17	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:17	

Data Download Date/Time: 06/23/2018 14:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REIDY,CSA,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: WENSYEL,OFC
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 05/26/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14816080A2
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 09:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 06/17/2018 11:42:39

Diagnostics Check		OK	11:30	
Air Blank		0.000	11:31	
Control Test		0.076	11:31	
Air Blank		0.000	11:32	
Subject Sample	# 1		11:35	No Sample Provided
Breath Volume		0.000	11:35	
Air Blank		0.000	11:35	
Air Blank		0.000	11:37	
Subject Sample	# 2	0.000	11:38	
Breath Volume		2.894	11:38	
Air Blank		0.000	11:38	
Air Blank		0.000	11:40	
Subject Sample	# 3	0.000	11:41	
Breath Volume		3.417	11:41	
Air Blank		0.000	11:41	
Control Test		0.076	11:42	
Air Blank		0.000	11:42	
Diagnostics Check		OK	11:42	

Data Download Date/Time: 06/23/2018 14:44

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: STAN,IRIS
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 06/23/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:41

SUBJECT'S LAST NAME: ACCETTURO

DATE/TIME OF TEST: 06/29/2018 11:21:20

Diagnostics Check		OK	11:12
Air Blank		0.000	11:12
Control Test		0.079	11:12
Air Blank		0.000	11:13
Subject Sample	# 1	0.132	11:16
Breath Volume		2.238	11:16
Air Blank		0.000	11:16
Air Blank		0.000	11:18
Subject Sample	# 2	0.124	11:19
Breath Volume		1.601	11:19
Air Blank		0.000	11:20
Control Test		0.079	11:20
Air Blank		0.000	11:21
Diagnostics Check		OK	11:21

Data Download Date/Time: 07/26/2018 15:06

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAXIMILIEN,RONALD,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MAXIMILIEN,RONALD
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: CARNRITE

DATE/TIME OF TEST: 06/02/2018 18:27:31

Diagnostics Check		OK	18:20
Air Blank		0.000	18:20
Control Test		0.081	18:21
Air Blank		0.000	18:21
Subject Sample	# 1	0.248	18:23
Breath Volume		1.761	18:23
Air Blank		0.000	18:23
Air Blank		0.000	18:25
Subject Sample	# 2	0.237	18:25
Breath Volume		1.363	18:25
Air Blank		0.000	18:26
Control Test		0.080	18:26
Air Blank		0.000	18:27
Diagnostics Check		OK	18:27

Data Download Date/Time: 06/29/2018 12:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADCLIFF,MICHAEL,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: AHLBERG

DATE/TIME OF TEST: 06/02/2018 03:32:18

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:29	
Control Test		0.081	03:29	
Air Blank		0.000	03:30	
Subject Sample	# 1		03:30	Subject Test Refused
Breath Volume		0.000	03:30	
Air Blank		0.000	03:31	
Control Test		0.081	03:31	
Air Blank		0.000	03:32	
Diagnostics Check		OK	03:32	

Data Download Date/Time: 06/18/2018 00:30

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: CHANDLER JR

DATE/TIME OF TEST: 06/23/2018 01:14:37

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.081	01:09
Air Blank		0.000	01:09
Subject Sample	# 1	0.154	01:10
Breath Volume		1.695	01:10
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.161	01:13
Breath Volume		1.656	01:13
Air Blank		0.000	01:13
Control Test		0.082	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 07/06/2018 19:04

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GOLDEN,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 06/23/2018 21:29:18

Diagnostics Check		OK	21:24	
Air Blank		0.000	21:24	
Control Test		0.081	21:24	
Air Blank		0.000	21:25	
Subject Sample	# 1		21:27	Subject Test Refused
Breath Volume		0.000	21:27	
Air Blank		0.000	21:28	
Control Test		0.082	21:28	
Air Blank		0.000	21:29	
Diagnostics Check		OK	21:29	

Data Download Date/Time: 07/06/2018 19:04

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOWALSKI,KEVIN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KOWALSKI,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: BRUTON

DATE/TIME OF TEST: 06/27/2018 17:33:35

Diagnostics Check		OK	17:27
Air Blank		0.000	17:27
Control Test		0.081	17:28
Air Blank		0.000	17:28
Subject Sample	# 1	0.269	17:29
Breath Volume		4.207	17:29
Air Blank		0.000	17:29
Air Blank		0.000	17:31
Subject Sample	# 2	0.263	17:32
Breath Volume		2.914	17:32
Air Blank		0.000	17:32
Control Test		0.081	17:33
Air Blank		0.000	17:33
Diagnostics Check		OK	17:33

Data Download Date/Time: 07/06/2018 19:04

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CODY,RICHARD,D
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG616503
DRY GAS EXPIRY DATE: 06/13/2018
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: SCHRACK

DATE/TIME OF TEST: 06/17/2018 04:01:17

Diagnostics Check		OK	03:55
Air Blank		0.000	03:55
Control Test		0.075	03:56
Air Blank		0.000	03:56
Subject Sample	# 1	0.157	03:56
Breath Volume		2.546	03:56
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.156	03:59
Breath Volume		1.542	03:59
Air Blank		0.000	04:00
Control Test		0.075	04:00
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 06/26/2018 13:40

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAGLE,TONY,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: WATTS,SHANNON
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG616503
DRY GAS EXPIRY DATE: 06/13/2018
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 06/23/2018 23:39:21

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:38	
Control Test	0.074	23:38	Control Outside Tolerance
Air Blank	0.000	23:39	

Data Download Date/Time: 06/26/2018 13:40

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAGLE,TONY,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: WATTS,SHANNON
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG616503
DRY GAS EXPIRY DATE: 06/13/2018
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 06/23/2018 23:49:30

Diagnostics Check	OK	23:48	
Air Blank	0.000	23:48	
Control Test	0.074	23:48	Control Outside Tolerance
Air Blank	0.000	23:49	

Data Download Date/Time: 06/26/2018 13:40

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDGECOMB,STEVEN,J
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: COCHRAN,JOHN
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 14:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/25/2018 15:56:15

Diagnostics Check		OK	15:49
Air Blank		0.000	15:50
Control Test		0.077	15:50
Air Blank		0.000	15:51
Subject Sample	# 1	0.236	15:51
Breath Volume		1.898	15:51
Air Blank		0.000	15:52
Air Blank		0.000	15:54
Subject Sample	# 2	0.222	15:54
Breath Volume		2.019	15:54
Air Blank		0.000	15:55
Control Test		0.076	15:55
Air Blank		0.000	15:56
Diagnostics Check		OK	15:56

Data Download Date/Time: 06/26/2018 13:40

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: CALISKAN

DATE/TIME OF TEST: 06/02/2018 01:55:08

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.081	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.107	01:50
Breath Volume		2.496	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.105	01:53
Breath Volume		2.042	01:53
Air Blank		0.000	01:54
Control Test		0.081	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 06/21/2018 18:58

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,RICARDO,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: ORTIZ,RICARDO
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:27

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 06/02/2018 07:57:54

Diagnostics Check		OK	07:51
Air Blank		0.000	07:52
Control Test		0.080	07:52
Air Blank		0.000	07:52
Subject Sample	# 1	0.228	07:53
Breath Volume		1.925	07:53
Air Blank		0.000	07:54
Air Blank		0.000	07:55
Subject Sample	# 2	0.223	07:56
Breath Volume		1.234	07:56
Air Blank		0.000	07:57
Control Test		0.080	07:57
Air Blank		0.000	07:57
Diagnostics Check		OK	07:57

Data Download Date/Time: 06/21/2018 18:58

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEQUEIRA,NATHAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SEQUEIRA,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HARDAKER

DATE/TIME OF TEST: 06/03/2018 00:06:41

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.081	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.000	00:02
Breath Volume		1.480	00:02
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.000	00:05
Breath Volume		1.300	00:05
Air Blank		0.000	00:05
Control Test		0.081	00:06
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 06/21/2018 18:58

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULASKI,DAMON,J
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: BERGWERFF,ROBYN
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: BUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: ALEJANDRO

DATE/TIME OF TEST: 06/03/2018 22:03:30

Diagnostics Check		OK	21:56
Air Blank		0.000	21:57
Control Test		0.080	21:57
Air Blank		0.000	21:58
Subject Sample	# 1	0.161	21:59
Breath Volume		3.359	21:59
Air Blank		0.000	21:59
Air Blank		0.000	22:01
Subject Sample	# 2	0.151	22:02
Breath Volume		3.207	22:02
Air Blank		0.000	22:02
Control Test		0.081	22:02
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 06/21/2018 18:58

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAEZ HERNANDEZ,DENYS,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PAEZ HERNANDEZ,DENYS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 06/06/2018 03:20:11

Diagnostics Check		OK	03:11	
Air Blank		0.000	03:11	
Control Test		0.080	03:12	
Air Blank		0.000	03:12	
Subject Sample	# 1		03:15	No Sample Provided
Breath Volume		0.000	03:15	
Air Blank		0.000	03:16	
Air Blank		0.000	03:18	
Subject Sample	# 2		03:18	Subject Test Refused
Breath Volume		0.000	03:18	
Air Blank		0.000	03:19	
Control Test		0.079	03:19	
Air Blank		0.000	03:20	
Diagnostics Check		OK	03:20	

Data Download Date/Time: 06/21/2018 18:58

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GOODWALD,DARREN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: ALLISTER CASANOVA

DATE/TIME OF TEST: 06/06/2018 05:11:21

Diagnostics Check		OK	05:04
Air Blank		0.000	05:04
Control Test		0.081	05:05
Air Blank		0.000	05:05
Subject Sample	# 1	0.000	05:06
Breath Volume		1.937	05:06
Air Blank		0.000	05:07
Air Blank		0.000	05:09
Subject Sample	# 2	0.000	05:09
Breath Volume		2.199	05:09
Air Blank		0.000	05:10
Control Test		0.082	05:10
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 06/21/2018 18:58

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHANSEN,LLOYD,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SCHULDT,VICTORIA
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: HOLLEY

DATE/TIME OF TEST: 06/08/2018 07:38:47

Diagnostics Check		OK	07:32
Air Blank		0.000	07:32
Control Test		0.081	07:33
Air Blank		0.000	07:33
Subject Sample	# 1	0.090	07:34
Breath Volume		1.562	07:34
Air Blank		0.000	07:34
Air Blank		0.000	07:36
Subject Sample	# 2	0.099	07:37
Breath Volume		1.800	07:37
Air Blank		0.000	07:37
Control Test		0.079	07:38
Air Blank		0.000	07:38
Diagnostics Check		OK	07:38

Data Download Date/Time: 06/21/2018 18:58

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINELLI,ADAM,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CAMPANA,NICHOLAS
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: O KELLY

DATE/TIME OF TEST: 06/10/2018 04:14:30

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.081	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.252	04:10
Breath Volume		2.246	04:10
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.242	04:12
Breath Volume		2.191	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 06/21/2018 18:58

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHANSEN,LLOYD,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LAROCHE,MICHAEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: TOLLIS

DATE/TIME OF TEST: 06/10/2018 07:51:50

Diagnostics Check		OK	07:45
Air Blank		0.000	07:46
Control Test		0.081	07:46
Air Blank		0.000	07:46
Subject Sample	# 1	0.000	07:47
Breath Volume		2.820	07:47
Air Blank		0.000	07:47
Air Blank		0.000	07:49
Subject Sample	# 2	0.000	07:50
Breath Volume		1.687	07:50
Air Blank		0.000	07:50
Control Test		0.079	07:51
Air Blank		0.000	07:51
Diagnostics Check		OK	07:51

Data Download Date/Time: 06/21/2018 18:58

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCE,SHNAYDINE,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SPENCE,SHNAYDINE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 06/11/2018 00:54:17

Diagnostics Check		OK	00:48
Air Blank		0.000	00:48
Control Test		0.081	00:48
Air Blank		0.000	00:49
Subject Sample	# 1	0.175	00:49
Breath Volume		1.445	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.180	00:52
Breath Volume		2.382	00:52
Air Blank		0.000	00:53
Control Test		0.081	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 06/21/2018 18:58

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDDLETON,DANIEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MIDDLETON,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: VELCHOFF

DATE/TIME OF TEST: 06/16/2018 02:14:10

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.082	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.144	02:09
Breath Volume		1.316	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.149	02:12
Breath Volume		1.949	02:12
Air Blank		0.000	02:13
Control Test		0.081	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 06/21/2018 18:58

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEQUEIRA,NATHAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SEQUEIRA,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: IRVIN

DATE/TIME OF TEST: 06/17/2018 22:12:59

Diagnostics Check		OK	22:08	
Air Blank		0.000	22:09	
Control Test		0.080	22:09	
Air Blank		0.000	22:09	
Subject Sample	# 1		22:11	Subject Test Refused
Breath Volume		0.000	22:11	
Air Blank		0.000	22:12	
Control Test		0.080	22:12	
Air Blank		0.000	22:12	
Diagnostics Check		OK	22:12	

Data Download Date/Time: 06/21/2018 18:58

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCE,SHNAYDINE,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SPENCE,SHNAYDINE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 06/20/2018 00:16:46

Diagnostics Check		OK	00:13	
Air Blank		0.000	00:14	
Control Test		0.082	00:14	
Air Blank		0.000	00:15	
Subject Sample	# 1		00:15	Subject Test Refused
Breath Volume		0.000	00:15	
Air Blank		0.000	00:15	
Control Test		0.081	00:16	
Air Blank		0.000	00:16	
Diagnostics Check		OK	00:16	

Data Download Date/Time: 06/21/2018 18:58

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEQUEIRA,NATHAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SEQUEIRA,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:43

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 06/21/2018 01:47:26

Diagnostics Check		OK	01:44	
Air Blank		0.000	01:45	
Control Test		0.082	01:45	
Air Blank		0.000	01:46	
Subject Sample	# 1		01:46	Sequence Aborted
Breath Volume		0.000	01:46	
Air Blank		0.000	01:47	

Data Download Date/Time: 06/21/2018 18:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEQUEIRA,NATHAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SEQUEIRA,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 06/21/2018 02:07:38

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.081	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.175	02:03
Breath Volume		2.199	02:03
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.183	02:06
Breath Volume		1.812	02:06
Air Blank		0.000	02:06
Control Test		0.080	02:07
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 06/21/2018 18:58

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,TYLER,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIRK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MULLENS

DATE/TIME OF TEST: 06/22/2018 00:17:37

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.081	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.165	00:13
Breath Volume		2.687	00:13
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.175	00:16
Breath Volume		1.687	00:16
Air Blank		0.000	00:16
Control Test		0.081	00:17
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 07/11/2018 14:48

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/23/2018 06:00:32

Diagnostics Check		OK	05:52
Air Blank		0.000	05:52
Control Test		0.082	05:53
Air Blank		0.000	05:53
Subject Sample	# 1	0.108	05:55
Breath Volume		1.562	05:55
Air Blank		0.000	05:56
Air Blank		0.000	05:58
Subject Sample	# 2	0.103	05:59
Breath Volume		1.480	05:59
Air Blank		0.000	05:59
Control Test		0.082	05:59
Air Blank		0.000	06:00
Diagnostics Check		OK	06:00

Data Download Date/Time: 07/11/2018 14:48

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: RAMOS BOULOGNE

DATE/TIME OF TEST: 06/23/2018 22:14:56

Diagnostics Check		OK	22:08
Air Blank		0.000	22:08
Control Test		0.081	22:09
Air Blank		0.000	22:09
Subject Sample	# 1	0.132	22:10
Breath Volume		1.425	22:10
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.130	22:13
Breath Volume		1.285	22:13
Air Blank		0.000	22:14
Control Test		0.081	22:14
Air Blank		0.000	22:14
Diagnostics Check		OK	22:14

Data Download Date/Time: 07/11/2018 14:48

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BLANCHARD

DATE/TIME OF TEST: 06/26/2018 23:13:01

Diagnostics Check		OK	23:01	
Air Blank		0.000	23:02	
Control Test		0.081	23:02	
Air Blank		0.000	23:02	
Subject Sample	# 1		23:06	No Sample Provided
Breath Volume		0.000	23:06	
Air Blank		0.000	23:06	
Air Blank		0.000	23:08	
Subject Sample	# 2		23:11	No Sample Provided
Breath Volume		0.000	23:11	
Air Blank		0.000	23:12	
Control Test		0.081	23:12	
Air Blank		0.000	23:12	
Diagnostics Check		OK	23:13	

Data Download Date/Time: 07/11/2018 14:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,TYLER,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIRK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 06/26/2018 23:51:53

Diagnostics Check		OK	23:45
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:46
Subject Sample	# 1	0.166	23:47
Breath Volume		2.023	23:47
Air Blank		0.000	23:48
Air Blank		0.000	23:49
Subject Sample	# 2	0.156	23:50
Breath Volume		1.363	23:50
Air Blank		0.000	23:50
Control Test		0.081	23:51
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 07/11/2018 14:48

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: KAPPLER

DATE/TIME OF TEST: 06/28/2018 23:21:42

Diagnostics Check		OK	23:10	
Air Blank		0.000	23:10	
Control Test		0.081	23:11	
Air Blank		0.000	23:11	
Subject Sample	# 1		23:14	No Sample Provided
Breath Volume		0.000	23:14	
Air Blank		0.000	23:15	
Air Blank		0.000	23:17	
Subject Sample	# 2		23:20	No Sample Provided
Breath Volume		0.000	23:20	
Air Blank		0.000	23:20	
Control Test		0.081	23:21	
Air Blank		0.000	23:21	
Diagnostics Check		OK	23:21	

Data Download Date/Time: 07/11/2018 14:48

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: EDGE

DATE/TIME OF TEST: 06/24/2018 22:08:53

Diagnostics Check		OK	22:06	
Air Blank		0.000	22:06	
Control Test		0.075	22:07	
Air Blank		0.000	22:07	
Subject Sample	# 1		22:08	RFI Detect
Breath Volume		0.328	22:08	
Air Blank			22:08	RFI Detect
Air Blank		0.000	22:08	

Data Download Date/Time: 07/06/2018 18:44

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: EDGE

DATE/TIME OF TEST: 06/24/2018 22:16:43

Diagnostics Check		OK	22:10
Air Blank		0.000	22:10
Control Test		0.081	22:11
Air Blank		0.000	22:11
Subject Sample	# 1	0.250	22:12
Breath Volume		2.320	22:12
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.256	22:15
Breath Volume		3.046	22:15
Air Blank		0.000	22:15
Control Test		0.078	22:16
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 07/06/2018 18:44

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001120

INSTRUMENT REGISTERED TO: FHP TROOP K
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEVEZ,BRYAN,
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ESTEVEZ,BRYAN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 06/01/2018 04:07:38

Diagnostics Check		OK	04:04	
Air Blank		0.000	04:05	
Control Test		0.079	04:05	
Air Blank		0.000	04:06	
Subject Sample	# 1		04:06	Subject Test Refused
Breath Volume		0.000	04:06	
Air Blank		0.000	04:06	
Control Test		0.079	04:07	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 06/29/2018 12:05

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001120

INSTRUMENT REGISTERED TO: FHP TROOP K
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEVEZ,BRIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ESTEVEZ,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 06/03/2018 04:32:26

Diagnostics Check		OK	04:28	
Air Blank		0.000	04:29	
Control Test		0.080	04:29	
Air Blank		0.000	04:30	
Subject Sample	# 1		04:31	Subject Test Refused
Breath Volume		0.000	04:31	
Air Blank		0.000	04:31	
Control Test		0.079	04:31	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:32	

Data Download Date/Time: 06/29/2018 12:05

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001120

INSTRUMENT REGISTERED TO: FHP TROOP K
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUGHIE,DARIN,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: HAUGHIE,DARIN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 06:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: MORALES HERNANDEZ

DATE/TIME OF TEST: 06/03/2018 07:23:10

Diagnostics Check		OK	07:19	
Air Blank		0.000	07:20	
Control Test		0.080	07:20	
Air Blank		0.000	07:21	
Subject Sample	# 1		07:21	Subject Test Refused
Breath Volume		0.000	07:21	
Air Blank		0.000	07:22	
Control Test		0.079	07:22	
Air Blank		0.000	07:23	
Diagnostics Check		OK	07:23	

Data Download Date/Time: 06/29/2018 12:05

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001120

INSTRUMENT REGISTERED TO: FHP TROOP K
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUGHIE,DARIN,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: VARON,MARCOS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 06/15/2018 23:34:54

Diagnostics Check		OK	23:28
Air Blank		0.000	23:28
Control Test		0.079	23:29
Air Blank		0.000	23:29
Subject Sample	# 1	0.111	23:30
Breath Volume		2.371	23:30
Air Blank		0.000	23:31
Air Blank		0.000	23:32
Subject Sample	# 2	0.110	23:33
Breath Volume		1.996	23:33
Air Blank		0.000	23:33
Control Test		0.078	23:34
Air Blank		0.000	23:34
Diagnostics Check		OK	23:34

Data Download Date/Time: 06/29/2018 12:05

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAXIMILIEN,RONALD,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MAXIMILIEN,RONALD
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: CARNRITE

DATE/TIME OF TEST: 06/02/2018 22:15:29

Diagnostics Check		OK	22:03	
Air Blank		0.000	22:04	
Control Test		0.079	22:04	
Air Blank		0.000	22:04	
Subject Sample	# 1	0.220	22:08	Volume Not Met
Breath Volume		0.863	22:08	
Air Blank		0.000	22:08	
Air Blank		0.000	22:10	
Subject Sample	# 2	0.254	22:11	
Breath Volume		1.792	22:11	
Air Blank		0.000	22:12	
Air Blank		0.000	22:13	
Subject Sample	# 3	0.248	22:14	
Breath Volume		1.214	22:14	
Air Blank			22:15	Purge Fail
Air Blank		0.000	22:15	

Data Download Date/Time: 06/29/2018 12:07

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASSO SILVA,NICOLAS,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: LASSO SILVA,NICOLAS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:42

SUBJECT'S LAST NAME: MAYCOCK

DATE/TIME OF TEST: 06/10/2018 16:43:06

Diagnostics Check		OK	16:40	
Air Blank		0.000	16:40	
Control Test		0.076	16:41	
Air Blank		0.000	16:41	
Subject Sample	# 1		16:41	Subject Test Refused
Breath Volume		0.000	16:41	
Air Blank		0.000	16:42	
Control Test		0.079	16:42	
Air Blank		0.000	16:42	
Diagnostics Check		OK	16:43	

Data Download Date/Time: 06/29/2018 12:07

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: FABELO BELLO

DATE/TIME OF TEST: 06/10/2018 19:52:34

Diagnostics Check		OK	19:49	
Air Blank		0.000	19:50	
Control Test		0.079	19:50	
Air Blank		0.000	19:51	
Subject Sample	# 1		19:51	Subject Test Refused
Breath Volume		0.000	19:51	
Air Blank		0.000	19:51	
Control Test		0.079	19:51	
Air Blank		0.000	19:52	
Diagnostics Check		OK	19:52	

Data Download Date/Time: 06/29/2018 12:07

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,JOSE,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: DIAZ,MARCOS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: GOROSITO

DATE/TIME OF TEST: 06/17/2018 06:35:59

Diagnostics Check		OK	06:29
Air Blank		0.000	06:30
Control Test		0.080	06:30
Air Blank		0.000	06:31
Subject Sample	# 1	0.199	06:31
Breath Volume		1.933	06:31
Air Blank		0.000	06:32
Air Blank		0.000	06:34
Subject Sample	# 2	0.199	06:34
Breath Volume		2.210	06:34
Air Blank		0.000	06:35
Control Test		0.079	06:35
Air Blank		0.000	06:35
Diagnostics Check		OK	06:35

Data Download Date/Time: 06/29/2018 12:07

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINTILIEN,BOB,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SAINTILIEN,BOB
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 06/24/2018 22:30:31

Diagnostics Check		OK	22:21
Air Blank		0.000	22:21
Control Test		0.080	22:21
Air Blank		0.000	22:22
Subject Sample	# 1	0.242	22:24
Breath Volume		1.480	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.243	22:28
Breath Volume		2.003	22:28
Air Blank		0.000	22:29
Control Test		0.078	22:29
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 06/29/2018 12:07

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARO,EMILIO,S
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: HARO,EMILIO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: ROMERO ACOSTA

DATE/TIME OF TEST: 06/27/2018 05:23:15

Diagnostics Check		OK	05:18	
Air Blank		0.000	05:19	
Control Test		0.080	05:19	
Air Blank		0.000	05:19	
Subject Sample	# 1		05:21	Subject Test Refused
Breath Volume		0.207	05:21	
Air Blank		0.000	05:22	
Control Test		0.080	05:22	
Air Blank		0.000	05:23	
Diagnostics Check		OK	05:23	

Data Download Date/Time: 06/29/2018 12:07

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TIFFANY,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WILLIAMS,TIFFANY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/01/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:27

SUBJECT'S LAST NAME: HORNE

DATE/TIME OF TEST: 06/11/2018 13:13:17

Diagnostics Check		OK	13:06
Air Blank		0.000	13:07
Control Test		0.079	13:07
Air Blank		0.000	13:08
Subject Sample	# 1	0.318	13:08
Breath Volume		1.460	13:08
Air Blank		0.000	13:09
Air Blank		0.000	13:11
Subject Sample	# 2	0.320	13:11
Breath Volume		1.710	13:11
Air Blank		0.000	13:12
Control Test		0.079	13:12
Air Blank		0.000	13:13
Diagnostics Check		OK	13:13

Data Download Date/Time: 07/02/2018 11:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,ZACHARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: STEPHENS,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/01/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 06/16/2018 18:27:01

Diagnostics Check		OK	18:20
Air Blank		0.000	18:21
Control Test		0.076	18:21
Air Blank		0.000	18:22
Subject Sample	# 1	0.137	18:22
Breath Volume		1.773	18:22
Air Blank		0.000	18:23
Air Blank		0.000	18:25
Subject Sample	# 2	0.136	18:25
Breath Volume		2.007	18:25
Air Blank		0.000	18:26
Control Test		0.075	18:26
Air Blank		0.000	18:26
Diagnostics Check		OK	18:27

Data Download Date/Time: 07/02/2018 11:28

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: KING,BRANDON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 06/15/2018 02:41:34

Diagnostics Check		OK	02:34
Air Blank		0.000	02:35
Control Test		0.077	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.134	02:36
Breath Volume		1.167	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.141	02:40
Breath Volume		1.449	02:40
Air Blank		0.000	02:40
Control Test		0.077	02:40
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 06/18/2018 11:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: SANTANA,JOHN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 06/23/2018 05:39:04

Diagnostics Check		OK	05:32
Air Blank		0.000	05:33
Control Test		0.076	05:33
Air Blank		0.000	05:34
Subject Sample	# 1	0.137	05:34
Breath Volume		2.875	05:34
Air Blank		0.000	05:35
Air Blank		0.000	05:37
Subject Sample	# 2	0.132	05:37
Breath Volume		2.261	05:37
Air Blank		0.000	05:38
Control Test		0.076	05:38
Air Blank		0.000	05:38
Diagnostics Check		OK	05:39

Data Download Date/Time: 07/30/2018 12:39

Printed: 12 September 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: VANZO,MICHAEL
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 06/28/2018 22:08:09

Diagnostics Check		OK	22:02
Air Blank		0.000	22:02
Control Test		0.077	22:02
Air Blank		0.000	22:03
Subject Sample	# 1	0.000	22:03
Breath Volume		1.167	22:03
Air Blank		0.000	22:04
Air Blank		0.000	22:06
Subject Sample	# 2	0.000	22:06
Breath Volume		1.492	22:06
Air Blank		0.000	22:07
Control Test		0.077	22:07
Air Blank		0.000	22:08
Diagnostics Check		OK	22:08

Data Download Date/Time: 07/30/2018 12:39

COBRA ©1997-2014 CMI, Inc.