

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROY,G,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: DEAS,B  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: PERSONENTT

DATE/TIME OF TEST: 07/03/2018 19:36:06

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 19:25 |
| Air Blank         |     | 0.000 | 19:26 |
| Control Test      |     | 0.078 | 19:26 |
| Air Blank         |     | 0.000 | 19:27 |
| Subject Sample    | # 1 | 0.268 | 19:27 |
| Breath Volume     |     | 1.949 | 19:27 |
| Air Blank         |     | 0.000 | 19:28 |
| Air Blank         |     | 0.000 | 19:30 |
| Subject Sample    | # 2 | 0.242 | 19:30 |
| Breath Volume     |     | 1.496 | 19:30 |
| Air Blank         |     | 0.000 | 19:31 |
| Air Blank         |     | 0.000 | 19:33 |
| Subject Sample    | # 3 | 0.245 | 19:34 |
| Breath Volume     |     | 1.324 | 19:34 |
| Air Blank         |     | 0.000 | 19:35 |
| Control Test      |     | 0.078 | 19:35 |
| Air Blank         |     | 0.000 | 19:36 |
| Diagnostics Check |     | OK    | 19:36 |

---

Data Download Date/Time: 07/30/2018 11:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVERETT, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: EVERETT, MATTHEW  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: CARLBERG

DATE/TIME OF TEST: 07/04/2018 22:08:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:01 |
| Air Blank          | 0.000 | 22:02 |
| Control Test       | 0.079 | 22:02 |
| Air Blank          | 0.000 | 22:02 |
| Subject Sample # 1 | 0.218 | 22:04 |
| Breath Volume      | 1.843 | 22:04 |
| Air Blank          | 0.000 | 22:04 |
| Air Blank          | 0.000 | 22:06 |
| Subject Sample # 2 | 0.236 | 22:07 |
| Breath Volume      | 1.632 | 22:07 |
| Air Blank          | 0.000 | 22:07 |
| Control Test       | 0.079 | 22:07 |
| Air Blank          | 0.000 | 22:08 |
| Diagnostics Check  | OK    | 22:08 |

---

Data Download Date/Time: 07/30/2018 11:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: NORTHWAY,GEORGIA  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:11

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 07/05/2018 07:40:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:33 |
| Air Blank          | 0.000 | 07:33 |
| Control Test       | 0.080 | 07:33 |
| Air Blank          | 0.000 | 07:34 |
| Subject Sample # 1 | 0.160 | 07:35 |
| Breath Volume      | 1.132 | 07:35 |
| Air Blank          | 0.000 | 07:36 |
| Air Blank          | 0.000 | 07:38 |
| Subject Sample # 2 | 0.160 | 07:38 |
| Breath Volume      | 1.480 | 07:38 |
| Air Blank          | 0.000 | 07:39 |
| Control Test       | 0.079 | 07:39 |
| Air Blank          | 0.000 | 07:40 |
| Diagnostics Check  | OK    | 07:40 |

---

Data Download Date/Time: 07/30/2018 11:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BULLOCK,JOSHUA,R  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: BLEVINS,AARON  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 07/11/2018 04:25:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:18 |
| Air Blank          | 0.000 | 04:19 |
| Control Test       | 0.079 | 04:19 |
| Air Blank          | 0.000 | 04:20 |
| Subject Sample # 1 | 0.154 | 04:21 |
| Breath Volume      | 1.757 | 04:21 |
| Air Blank          | 0.000 | 04:21 |
| Air Blank          | 0.000 | 04:23 |
| Subject Sample # 2 | 0.144 | 04:24 |
| Breath Volume      | 1.210 | 04:24 |
| Air Blank          | 0.000 | 04:24 |
| Control Test       | 0.079 | 04:24 |
| Air Blank          | 0.000 | 04:25 |
| Diagnostics Check  | OK    | 04:25 |

---

Data Download Date/Time: 07/30/2018 11:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KELLER, DAVID,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: DEAS, BRANDYN  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 07/13/2018 02:52:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:47 |
| Air Blank          | 0.000 | 02:47 |
| Control Test       | 0.080 | 02:47 |
| Air Blank          | 0.000 | 02:48 |
| Subject Sample # 1 | 0.000 | 02:48 |
| Breath Volume      | 1.769 | 02:48 |
| Air Blank          | 0.000 | 02:49 |
| Air Blank          | 0.000 | 02:51 |
| Subject Sample # 2 | 0.000 | 02:51 |
| Breath Volume      | 1.218 | 02:51 |
| Air Blank          | 0.000 | 02:52 |
| Control Test       | 0.080 | 02:52 |
| Air Blank          | 0.000 | 02:52 |
| Diagnostics Check  | OK    | 02:52 |

---

Data Download Date/Time: 07/30/2018 11:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PILKANIS,MARY,C  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: LAFFITTE,TRAVIS  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: POOLE

DATE/TIME OF TEST: 07/20/2018 00:18:40

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 00:11 |
| Air Blank         |     | 0.000 | 00:11 |
| Control Test      |     | 0.080 | 00:11 |
| Air Blank         |     | 0.000 | 00:12 |
| Subject Sample    | # 1 | 0.166 | 00:13 |
| Breath Volume     |     | 1.761 | 00:13 |
| Air Blank         |     | 0.000 | 00:14 |
| Air Blank         |     | 0.000 | 00:16 |
| Subject Sample    | # 2 | 0.160 | 00:17 |
| Breath Volume     |     | 1.332 | 00:17 |
| Air Blank         |     | 0.000 | 00:17 |
| Control Test      |     | 0.080 | 00:18 |
| Air Blank         |     | 0.000 | 00:18 |
| Diagnostics Check |     | OK    | 00:18 |

---

Data Download Date/Time: 07/30/2018 11:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: NORTHWAY,GEORGIA  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: CRIM

DATE/TIME OF TEST: 07/22/2018 07:09:31

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 07:03 |
| Air Blank         |     | 0.000 | 07:03 |
| Control Test      |     | 0.079 | 07:03 |
| Air Blank         |     | 0.000 | 07:04 |
| Subject Sample    | # 1 | 0.148 | 07:05 |
| Breath Volume     |     | 2.269 | 07:05 |
| Air Blank         |     | 0.000 | 07:05 |
| Air Blank         |     | 0.000 | 07:07 |
| Subject Sample    | # 2 | 0.141 | 07:08 |
| Breath Volume     |     | 1.769 | 07:08 |
| Air Blank         |     | 0.000 | 07:08 |
| Control Test      |     | 0.079 | 07:08 |
| Air Blank         |     | 0.000 | 07:09 |
| Diagnostics Check |     | OK    | 07:09 |

---

Data Download Date/Time: 07/30/2018 11:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: CROUCH,CLIFF  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:37

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/22/2018 07:23:25

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 07:16 |
| Air Blank         |     | 0.000 | 07:17 |
| Control Test      |     | 0.079 | 07:17 |
| Air Blank         |     | 0.000 | 07:18 |
| Subject Sample    | # 1 | 0.151 | 07:18 |
| Breath Volume     |     | 2.230 | 07:18 |
| Air Blank         |     | 0.000 | 07:19 |
| Air Blank         |     | 0.000 | 07:21 |
| Subject Sample    | # 2 | 0.152 | 07:21 |
| Breath Volume     |     | 2.121 | 07:21 |
| Air Blank         |     | 0.000 | 07:22 |
| Control Test      |     | 0.078 | 07:22 |
| Air Blank         |     | 0.000 | 07:23 |
| Diagnostics Check |     | OK    | 07:23 |

---

Data Download Date/Time: 07/30/2018 11:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVERETT,M,A  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: EVERETT,MATTHEW  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: PORTILLO RENTERIA

DATE/TIME OF TEST: 07/22/2018 19:12:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:06 |
| Air Blank          | 0.000 | 19:06 |
| Control Test       | 0.079 | 19:06 |
| Air Blank          | 0.000 | 19:07 |
| Subject Sample # 1 | 0.106 | 19:07 |
| Breath Volume      | 1.191 | 19:07 |
| Air Blank          | 0.000 | 19:08 |
| Air Blank          | 0.000 | 19:10 |
| Subject Sample # 2 | 0.111 | 19:11 |
| Breath Volume      | 1.882 | 19:11 |
| Air Blank          | 0.000 | 19:11 |
| Control Test       | 0.079 | 19:11 |
| Air Blank          | 0.000 | 19:12 |
| Diagnostics Check  | OK    | 19:12 |

---

Data Download Date/Time: 07/30/2018 11:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROY,G,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: ROY,G  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: DELLARIA

DATE/TIME OF TEST: 07/27/2018 02:36:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:30 |
| Air Blank          | 0.000 | 02:31 |
| Control Test       | 0.080 | 02:31 |
| Air Blank          | 0.000 | 02:32 |
| Subject Sample # 1 | 0.179 | 02:32 |
| Breath Volume      | 2.496 | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Subject Sample # 2 | 0.176 | 02:35 |
| Breath Volume      | 2.148 | 02:35 |
| Air Blank          | 0.000 | 02:35 |
| Control Test       | 0.079 | 02:36 |
| Air Blank          | 0.000 | 02:36 |
| Diagnostics Check  | OK    | 02:36 |

---

Data Download Date/Time: 07/30/2018 11:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROY,G,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: ROY,G  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: EASOM

DATE/TIME OF TEST: 07/27/2018 22:45:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:39 |
| Air Blank          | 0.000 | 22:39 |
| Control Test       | 0.079 | 22:39 |
| Air Blank          | 0.000 | 22:40 |
| Subject Sample # 1 | 0.170 | 22:40 |
| Breath Volume      | 2.019 | 22:40 |
| Air Blank          | 0.000 | 22:41 |
| Air Blank          | 0.000 | 22:43 |
| Subject Sample # 2 | 0.166 | 22:43 |
| Breath Volume      | 1.890 | 22:43 |
| Air Blank          | 0.000 | 22:44 |
| Control Test       | 0.079 | 22:44 |
| Air Blank          | 0.000 | 22:45 |
| Diagnostics Check  | OK    | 22:45 |

---

Data Download Date/Time: 07/30/2018 11:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROY,G,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: ROY,G  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: AMOROSI

DATE/TIME OF TEST: 07/28/2018 22:05:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:59 |
| Air Blank          | 0.000 | 21:59 |
| Control Test       | 0.079 | 22:00 |
| Air Blank          | 0.000 | 22:00 |
| Subject Sample # 1 | 0.170 | 22:01 |
| Breath Volume      | 2.695 | 22:01 |
| Air Blank          | 0.000 | 22:01 |
| Air Blank          | 0.000 | 22:03 |
| Subject Sample # 2 | 0.172 | 22:04 |
| Breath Volume      | 2.949 | 22:04 |
| Air Blank          | 0.000 | 22:04 |
| Control Test       | 0.079 | 22:04 |
| Air Blank          | 0.000 | 22:05 |
| Diagnostics Check  | OK    | 22:05 |

---

Data Download Date/Time: 07/30/2018 11:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LATHAN,MICHAEL,S

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: REID,T

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 18:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: NEWSOME

DATE/TIME OF TEST: 07/01/2018 19:45:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:37 |
| Air Blank          | 0.000 | 19:38 |
| Control Test       | 0.076 | 19:38 |
| Air Blank          | 0.000 | 19:38 |
| Subject Sample # 1 | 0.082 | 19:40 |
| Breath Volume      | 1.343 | 19:40 |
| Air Blank          | 0.000 | 19:40 |
| Air Blank          | 0.000 | 19:42 |
| Subject Sample # 2 | 0.087 | 19:43 |
| Breath Volume      | 2.078 | 19:43 |
| Air Blank          | 0.000 | 19:44 |
| Control Test       | 0.077 | 19:44 |
| Air Blank          | 0.000 | 19:45 |
| Diagnostics Check  | OK    | 19:45 |

---

Data Download Date/Time: 07/31/2018 11:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LATHAN,MICHAEL,S  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: DAVIDSON,T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 07/02/2018 03:30:50

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:27 |                      |
| Air Blank         | 0.000 | 03:28 |                      |
| Control Test      | 0.076 | 03:28 |                      |
| Air Blank         | 0.000 | 03:29 |                      |
| Subject Sample    | # 1   | 03:29 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:29 |                      |
| Air Blank         | 0.000 | 03:29 |                      |
| Control Test      | 0.076 | 03:30 |                      |
| Air Blank         | 0.000 | 03:30 |                      |
| Diagnostics Check | OK    | 03:30 |                      |

---

Data Download Date/Time: 07/31/2018 11:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRETT,CHASE,A  
BREATH TEST OPERATOR AGENCY: Citrus County SO  
ARRESTING OFFICER: THOMASON,D  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: BUI  
ARREST TIME: 22:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: DENTICI

DATE/TIME OF TEST: 07/03/2018 23:51:19

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 23:44 |
| Air Blank         |     | 0.000 | 23:44 |
| Control Test      |     | 0.077 | 23:45 |
| Air Blank         |     | 0.000 | 23:45 |
| Subject Sample    | # 1 | 0.118 | 23:46 |
| Breath Volume     |     | 2.429 | 23:46 |
| Air Blank         |     | 0.000 | 23:46 |
| Air Blank         |     | 0.000 | 23:48 |
| Subject Sample    | # 2 | 0.115 | 23:49 |
| Breath Volume     |     | 2.734 | 23:49 |
| Air Blank         |     | 0.000 | 23:50 |
| Control Test      |     | 0.076 | 23:50 |
| Air Blank         |     | 0.000 | 23:51 |
| Diagnostics Check |     | OK    | 23:51 |

---

Data Download Date/Time: 07/31/2018 11:56



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRETT,CHASE,A  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: DOBOSIEWICZ,A  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 07/04/2018 00:46:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:40 |
| Air Blank          | 0.000 | 00:40 |
| Control Test       | 0.077 | 00:41 |
| Air Blank          | 0.000 | 00:41 |
| Subject Sample # 1 | 0.085 | 00:42 |
| Breath Volume      | 3.472 | 00:42 |
| Air Blank          | 0.000 | 00:43 |
| Air Blank          | 0.000 | 00:44 |
| Subject Sample # 2 | 0.082 | 00:45 |
| Breath Volume      | 3.554 | 00:45 |
| Air Blank          | 0.000 | 00:46 |
| Control Test       | 0.077 | 00:46 |
| Air Blank          | 0.000 | 00:46 |
| Diagnostics Check  | OK    | 00:46 |

---

Data Download Date/Time: 07/31/2018 11:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARRETT,CHASE,A

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: KNIGHT,C

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 01:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 07/04/2018 02:43:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:36 |
| Air Blank          | 0.000 | 02:37 |
| Control Test       | 0.077 | 02:37 |
| Air Blank          | 0.000 | 02:37 |
| Subject Sample # 1 | 0.220 | 02:38 |
| Breath Volume      | 2.746 | 02:38 |
| Air Blank          | 0.000 | 02:39 |
| Air Blank          | 0.000 | 02:41 |
| Subject Sample # 2 | 0.212 | 02:41 |
| Breath Volume      | 3.476 | 02:41 |
| Air Blank          | 0.000 | 02:42 |
| Control Test       | 0.077 | 02:42 |
| Air Blank          | 0.000 | 02:43 |
| Diagnostics Check  | OK    | 02:43 |

---

Data Download Date/Time: 07/31/2018 11:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FLOYD,DANIEL,Z

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: TAYLOR,DEP

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 17:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: CRAFTON

DATE/TIME OF TEST: 07/04/2018 18:40:37

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 18:33 |
| Air Blank         |     | 0.000 | 18:34 |
| Control Test      |     | 0.076 | 18:34 |
| Air Blank         |     | 0.000 | 18:35 |
| Subject Sample    | # 1 | 0.314 | 18:35 |
| Breath Volume     |     | 2.433 | 18:35 |
| Air Blank         |     | 0.000 | 18:36 |
| Air Blank         |     | 0.000 | 18:38 |
| Subject Sample    | # 2 | 0.294 | 18:39 |
| Breath Volume     |     | 3.753 | 18:39 |
| Air Blank         |     | 0.000 | 18:39 |
| Control Test      |     | 0.076 | 18:40 |
| Air Blank         |     | 0.000 | 18:40 |
| Diagnostics Check |     | OK    | 18:40 |

---

Data Download Date/Time: 07/31/2018 11:56



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARRETT,CHASE,A

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: LAKATIS,K

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 18:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: BERTINE

DATE/TIME OF TEST: 07/08/2018 19:27:49

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:24 |                      |
| Air Blank         | 0.000 | 19:25 |                      |
| Control Test      | 0.077 | 19:25 |                      |
| Air Blank         | 0.000 | 19:26 |                      |
| Subject Sample    | # 1   | 19:26 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:26 |                      |
| Air Blank         | 0.000 | 19:26 |                      |
| Control Test      | 0.077 | 19:27 |                      |
| Air Blank         | 0.000 | 19:27 |                      |
| Diagnostics Check | OK    | 19:27 |                      |

---

Data Download Date/Time: 07/31/2018 11:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: BLOTZ,G

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 16:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/12/2018 18:23:56

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 18:21 |                      |
| Air Blank         | 0.000 | 18:21 |                      |
| Control Test      | 0.077 | 18:21 |                      |
| Air Blank         | 0.000 | 18:22 |                      |
| Subject Sample    | # 1   | 18:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 18:22 |                      |
| Air Blank         | 0.000 | 18:23 |                      |
| Control Test      | 0.077 | 18:23 |                      |
| Air Blank         | 0.000 | 18:23 |                      |
| Diagnostics Check | OK    | 18:23 |                      |

---

Data Download Date/Time: 07/31/2018 11:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FLOYD,DANIEL,Z

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: GERENA,DEP

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 23:23

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: SENECHAL

DATE/TIME OF TEST: 07/14/2018 00:25:35

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:22 |                      |
| Air Blank         | 0.000 | 00:23 |                      |
| Control Test      | 0.077 | 00:23 |                      |
| Air Blank         | 0.000 | 00:24 |                      |
| Subject Sample    | # 1   | 00:24 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:24 |                      |
| Air Blank         | 0.000 | 00:24 |                      |
| Control Test      | 0.078 | 00:24 |                      |
| Air Blank         | 0.000 | 00:25 |                      |
| Diagnostics Check | OK    | 00:25 |                      |

---

Data Download Date/Time: 07/31/2018 11:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRETT,CHASE,A  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: CLOUD,T  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: ERSKINE

DATE/TIME OF TEST: 07/17/2018 18:26:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:19 |
| Air Blank          | 0.000 | 18:19 |
| Control Test       | 0.077 | 18:19 |
| Air Blank          | 0.000 | 18:20 |
| Subject Sample # 1 | 0.000 | 18:21 |
| Breath Volume      | 4.410 | 18:21 |
| Air Blank          | 0.000 | 18:21 |
| Air Blank          | 0.000 | 18:23 |
| Subject Sample # 2 | 0.000 | 18:24 |
| Breath Volume      | 4.832 | 18:24 |
| Air Blank          | 0.000 | 18:25 |
| Control Test       | 0.077 | 18:25 |
| Air Blank          | 0.000 | 18:26 |
| Diagnostics Check  | OK    | 18:26 |

---

Data Download Date/Time: 07/31/2018 11:56



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FOSTER, MICHAEL, D

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: LEJARZAR, JOHN

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: BUI

ARREST TIME: 10:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: THORTON

DATE/TIME OF TEST: 07/21/2018 13:30:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:24 |
| Air Blank          | 0.000 | 13:24 |
| Control Test       | 0.077 | 13:24 |
| Air Blank          | 0.000 | 13:25 |
| Subject Sample # 1 | 0.062 | 13:25 |
| Breath Volume      | 2.417 | 13:25 |
| Air Blank          | 0.000 | 13:26 |
| Air Blank          | 0.000 | 13:28 |
| Subject Sample # 2 | 0.064 | 13:29 |
| Breath Volume      | 3.738 | 13:29 |
| Air Blank          | 0.000 | 13:29 |
| Control Test       | 0.077 | 13:29 |
| Air Blank          | 0.000 | 13:30 |
| Diagnostics Check  | OK    | 13:30 |

---

Data Download Date/Time: 07/31/2018 11:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLOYD,DANIEL,Z  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: WARD,TROOPER  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: LIND

DATE/TIME OF TEST: 07/27/2018 22:49:45

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 22:38 |                |
| Air Blank         |     | 0.000 | 22:38 |                |
| Control Test      |     | 0.077 | 22:38 |                |
| Air Blank         |     | 0.000 | 22:39 |                |
| Subject Sample    | # 1 | 0.099 | 22:42 | Volume Not Met |
| Breath Volume     |     | 0.035 | 22:42 |                |
| Air Blank         |     | 0.000 | 22:43 |                |
| Air Blank         |     | 0.000 | 22:45 |                |
| Subject Sample    | # 2 | 0.133 | 22:48 | Volume Not Met |
| Breath Volume     |     | 0.316 | 22:48 |                |
| Air Blank         |     | 0.000 | 22:48 |                |
| Control Test      |     | 0.077 | 22:49 |                |
| Air Blank         |     | 0.000 | 22:49 |                |
| Diagnostics Check |     | OK    | 22:49 |                |

---

Data Download Date/Time: 07/31/2018 11:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLOYD,DANIEL,Z  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: WARD,TROOPER  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: LIND

DATE/TIME OF TEST: 07/27/2018 23:00:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:52 |
| Air Blank          | 0.000 | 22:53 |
| Control Test       | 0.076 | 22:53 |
| Air Blank          | 0.000 | 22:54 |
| Subject Sample # 1 | 0.170 | 22:55 |
| Breath Volume      | 1.707 | 22:55 |
| Air Blank          | 0.000 | 22:55 |
| Air Blank          | 0.000 | 22:57 |
| Subject Sample # 2 | 0.160 | 22:59 |
| Breath Volume      | 1.437 | 22:59 |
| Air Blank          | 0.000 | 22:59 |
| Control Test       | 0.076 | 23:00 |
| Air Blank          | 0.000 | 23:00 |
| Diagnostics Check  | OK    | 23:00 |

---

Data Download Date/Time: 07/31/2018 11:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FOSTER, MICHAEL, D

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: NA, NA

ARRESTING OFFICER AGENCY: FDOT MCC

LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: COMM VEHIC

ARREST TIME: 00:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: ZWINGE

DATE/TIME OF TEST: 07/30/2018 09:06:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:00 |
| Air Blank          | 0.000 | 09:00 |
| Control Test       | 0.077 | 09:01 |
| Air Blank          | 0.000 | 09:01 |
| Subject Sample # 1 | 0.000 | 09:02 |
| Breath Volume      | 2.671 | 09:02 |
| Air Blank          | 0.000 | 09:02 |
| Air Blank          | 0.000 | 09:04 |
| Subject Sample # 2 | 0.000 | 09:05 |
| Breath Volume      | 2.929 | 09:05 |
| Air Blank          | 0.000 | 09:05 |
| Control Test       | 0.077 | 09:06 |
| Air Blank          | 0.000 | 09:06 |
| Diagnostics Check  | OK    | 09:06 |

---

Data Download Date/Time: 07/31/2018 11:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVAREZ,ROBERTO,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: ALVAREZ,ROBRTO  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: ARCAYA ROMULO

DATE/TIME OF TEST: 07/05/2018 22:37:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:31 |
| Air Blank          | 0.000 | 22:31 |
| Control Test       | 0.079 | 22:31 |
| Air Blank          | 0.000 | 22:32 |
| Subject Sample # 1 | 0.214 | 22:32 |
| Breath Volume      | 2.113 | 22:32 |
| Air Blank          | 0.000 | 22:33 |
| Air Blank          | 0.000 | 22:35 |
| Subject Sample # 2 | 0.223 | 22:35 |
| Breath Volume      | 1.574 | 22:35 |
| Air Blank          | 0.000 | 22:36 |
| Control Test       | 0.079 | 22:36 |
| Air Blank          | 0.000 | 22:37 |
| Diagnostics Check  | OK    | 22:37 |

---

Data Download Date/Time: 07/30/2018 10:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUNT,CHARLES,R  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: HUNT,CHARLES  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: TETI

DATE/TIME OF TEST: 07/25/2018 04:22:32

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 04:16 |
| Air Blank         |     | 0.000 | 04:16 |
| Control Test      |     | 0.079 | 04:16 |
| Air Blank         |     | 0.000 | 04:17 |
| Subject Sample    | # 1 | 0.042 | 04:17 |
| Breath Volume     |     | 1.824 | 04:17 |
| Air Blank         |     | 0.000 | 04:18 |
| Air Blank         |     | 0.000 | 04:20 |
| Subject Sample    | # 2 | 0.040 | 04:21 |
| Breath Volume     |     | 2.191 | 04:21 |
| Air Blank         |     | 0.000 | 04:21 |
| Control Test      |     | 0.078 | 04:21 |
| Air Blank         |     | 0.000 | 04:22 |
| Diagnostics Check |     | OK    | 04:22 |

---

Data Download Date/Time: 07/30/2018 10:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURGESS,MASON,W  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: BURGESS,MASON  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 666491  
DRY GAS EXPIRY DATE: 12/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 07/01/2018 05:05:38

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 04:51 |                    |
| Air Blank         | 0.000 | 04:52 |                    |
| Control Test      | 0.081 | 04:52 |                    |
| Air Blank         | 0.000 | 04:52 |                    |
| Subject Sample    | # 1   | 04:56 | No Sample Provided |
| Breath Volume     | 0.000 | 04:56 |                    |
| Air Blank         | 0.000 | 04:56 |                    |
| Air Blank         | 0.000 | 04:58 |                    |
| Subject Sample    | # 2   | 0.111 | 05:00              |
| Breath Volume     | 1.500 | 05:00 |                    |
| Air Blank         | 0.000 | 05:01 |                    |
| Air Blank         | 0.000 | 05:02 |                    |
| Subject Sample    | # 3   | 0.113 | 05:04              |
| Breath Volume     | 2.484 | 05:04 |                    |
| Air Blank         | 0.000 | 05:04 |                    |
| Control Test      | 0.079 | 05:05 |                    |
| Air Blank         | 0.000 | 05:05 |                    |
| Diagnostics Check | OK    | 05:05 |                    |

---

Data Download Date/Time: 07/31/2018 14:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: ESNEUL,MICHAEL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 666491  
DRY GAS EXPIRY DATE: 12/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/16/2018 05:33:08

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:30 |                      |
| Air Blank         | 0.000 | 05:30 |                      |
| Control Test      | 0.080 | 05:30 |                      |
| Air Blank         | 0.000 | 05:31 |                      |
| Subject Sample    | # 1   | 05:31 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:31 |                      |
| Air Blank         | 0.000 | 05:32 |                      |
| Control Test      | 0.080 | 05:32 |                      |
| Air Blank         | 0.000 | 05:33 |                      |
| Diagnostics Check | OK    | 05:33 |                      |

---

Data Download Date/Time: 07/31/2018 14:47



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,LANCE,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WILLIAMS,LANCE  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 666491  
DRY GAS EXPIRY DATE: 12/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 05:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: NAYLOR

DATE/TIME OF TEST: 07/25/2018 05:57:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:51 |
| Air Blank          | 0.000 | 05:51 |
| Control Test       | 0.080 | 05:51 |
| Air Blank          | 0.000 | 05:52 |
| Subject Sample # 1 | 0.000 | 05:53 |
| Breath Volume      | 1.628 | 05:53 |
| Air Blank          | 0.000 | 05:53 |
| Air Blank          | 0.000 | 05:55 |
| Subject Sample # 2 | 0.000 | 05:56 |
| Breath Volume      | 2.226 | 05:56 |
| Air Blank          | 0.000 | 05:56 |
| Control Test       | 0.079 | 05:56 |
| Air Blank          | 0.000 | 05:57 |
| Diagnostics Check  | OK    | 05:57 |

---

Data Download Date/Time: 07/31/2018 14:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANDREW, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: GORDON, JON  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 666491  
DRY GAS EXPIRY DATE: 12/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 10:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: RECINOS

DATE/TIME OF TEST: 07/29/2018 11:06:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 11:00 |
| Air Blank          | 0.000 | 11:00 |
| Control Test       | 0.080 | 11:01 |
| Air Blank          | 0.000 | 11:01 |
| Subject Sample # 1 | 0.101 | 11:02 |
| Breath Volume      | 3.296 | 11:02 |
| Air Blank          | 0.000 | 11:02 |
| Air Blank          | 0.000 | 11:04 |
| Subject Sample # 2 | 0.102 | 11:05 |
| Breath Volume      | 3.097 | 11:05 |
| Air Blank          | 0.000 | 11:05 |
| Control Test       | 0.079 | 11:06 |
| Air Blank          | 0.000 | 11:06 |
| Diagnostics Check  | OK    | 11:06 |

---

Data Download Date/Time: 07/31/2018 14:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BYRD,BRANDON,C  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: BYRD,BRANDON  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A3  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: WORLEY

DATE/TIME OF TEST: 07/28/2018 01:51:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:44 |
| Air Blank          | 0.000 | 01:45 |
| Control Test       | 0.077 | 01:45 |
| Air Blank          | 0.000 | 01:46 |
| Subject Sample # 1 | 0.144 | 01:46 |
| Breath Volume      | 2.664 | 01:46 |
| Air Blank          | 0.000 | 01:47 |
| Air Blank          | 0.000 | 01:49 |
| Subject Sample # 2 | 0.126 | 01:49 |
| Breath Volume      | 2.429 | 01:49 |
| Air Blank          | 0.000 | 01:50 |
| Control Test       | 0.077 | 01:50 |
| Air Blank          | 0.000 | 01:51 |
| Diagnostics Check  | OK    | 01:51 |

---

Data Download Date/Time: 08/21/2018 11:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALMATEER,TRACY,D  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: PALMATEER,TRACY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A3  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 08:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:07

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/28/2018 09:34:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:27 |
| Air Blank          | 0.000 | 09:27 |
| Control Test       | 0.079 | 09:27 |
| Air Blank          | 0.000 | 09:28 |
| Subject Sample # 1 | 0.147 | 09:29 |
| Breath Volume      | 2.113 | 09:29 |
| Air Blank          | 0.000 | 09:30 |
| Air Blank          | 0.000 | 09:32 |
| Subject Sample # 2 | 0.147 | 09:33 |
| Breath Volume      | 1.984 | 09:33 |
| Air Blank          | 0.000 | 09:33 |
| Control Test       | 0.076 | 09:34 |
| Air Blank          | 0.000 | 09:34 |
| Diagnostics Check  | OK    | 09:34 |

---

Data Download Date/Time: 08/21/2018 11:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDY,JOE,E  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: HARDY,JOE  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 07/07/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/08/2018 20:04:58

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:56 |                      |
| Air Blank         | 0.000 | 19:56 |                      |
| Control Test      | 0.079 | 19:57 |                      |
| Air Blank         | 0.000 | 19:57 |                      |
| Subject Sample    | # 1   | 20:00 | No Sample Provided   |
| Breath Volume     | 0.000 | 20:00 |                      |
| Air Blank         | 0.000 | 20:01 |                      |
| Air Blank         | 0.000 | 20:03 |                      |
| Subject Sample    | # 2   | 20:03 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:03 |                      |
| Air Blank         | 0.000 | 20:04 |                      |
| Control Test      | 0.079 | 20:04 |                      |
| Air Blank         | 0.000 | 20:04 |                      |
| Diagnostics Check | OK    | 20:04 |                      |

---

Data Download Date/Time: 09/17/2018 11:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYLE JR,HARRY,R  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: PYLE JR,HARRY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 07/07/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: HAGER

DATE/TIME OF TEST: 07/19/2018 20:13:54

---

|                    |       |       |                           |
|--------------------|-------|-------|---------------------------|
| Diagnostics Check  | OK    | 20:06 |                           |
| Air Blank          | 0.000 | 20:07 |                           |
| Control Test       | 0.079 | 20:07 |                           |
| Air Blank          | 0.000 | 20:07 |                           |
| Subject Sample # 1 | 0.165 | 20:08 |                           |
| Breath Volume      | 1.882 | 20:08 |                           |
| Air Blank          | 0.000 | 20:09 |                           |
| Air Blank          | 0.000 | 20:11 |                           |
| Subject Sample # 2 | 0.161 | 20:12 |                           |
| Breath Volume      | 1.265 | 20:12 |                           |
| Air Blank          | 0.000 | 20:13 |                           |
| Control Test       | 0.074 | 20:13 | Control Outside Tolerance |
| Air Blank          | 0.000 | 20:13 |                           |

---

Data Download Date/Time: 09/17/2018 11:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KELLER,JERRIT,W  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: SABO,JESSICA  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: BARBOZA

DATE/TIME OF TEST: 07/04/2018 03:47:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:40 |
| Air Blank          | 0.000 | 03:41 |
| Control Test       | 0.079 | 03:41 |
| Air Blank          | 0.000 | 03:41 |
| Subject Sample # 1 | 0.151 | 03:42 |
| Breath Volume      | 1.722 | 03:42 |
| Air Blank          | 0.000 | 03:43 |
| Air Blank          | 0.000 | 03:45 |
| Subject Sample # 2 | 0.156 | 03:45 |
| Breath Volume      | 1.386 | 03:45 |
| Air Blank          | 0.000 | 03:46 |
| Control Test       | 0.078 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Diagnostics Check  | OK    | 03:47 |

---

Data Download Date/Time: 07/30/2018 10:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SABO,JESSICA,L  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: SABO,JESSICA  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 17:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 07/04/2018 19:25:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:17 |
| Air Blank          | 0.000 | 19:18 |
| Control Test       | 0.078 | 19:18 |
| Air Blank          | 0.000 | 19:19 |
| Subject Sample # 1 | 0.167 | 19:19 |
| Breath Volume      | 1.750 | 19:19 |
| Air Blank          | 0.000 | 19:20 |
| Air Blank          | 0.000 | 19:22 |
| Subject Sample # 2 | 0.163 | 19:23 |
| Breath Volume      | 1.585 | 19:23 |
| Air Blank          | 0.000 | 19:24 |
| Control Test       | 0.078 | 19:24 |
| Air Blank          | 0.000 | 19:25 |
| Diagnostics Check  | OK    | 19:25 |

---

Data Download Date/Time: 07/30/2018 10:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVAREZ,ROBERTO,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: ALVAREZ,ROBERTO  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: ZABEL

DATE/TIME OF TEST: 07/05/2018 00:57:41

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:54 |                      |
| Air Blank         | 0.000 | 00:55 |                      |
| Control Test      | 0.078 | 00:55 |                      |
| Air Blank         | 0.000 | 00:56 |                      |
| Subject Sample    | # 1   | 00:56 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:56 |                      |
| Air Blank         | 0.000 | 00:56 |                      |
| Control Test      | 0.078 | 00:57 |                      |
| Air Blank         | 0.000 | 00:57 |                      |
| Diagnostics Check | OK    | 00:57 |                      |

---

Data Download Date/Time: 07/30/2018 10:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KANKAM,DENZEL,K  
BREATH TEST OPERATOR AGENCY: Gainesville PD  
ARRESTING OFFICER: KANKAM,DENZEL  
ARRESTING OFFICER AGENCY: Gainesville PD  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: LOZADA

DATE/TIME OF TEST: 07/07/2018 04:10:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:03 |
| Air Blank          | 0.000 | 04:03 |
| Control Test       | 0.079 | 04:03 |
| Air Blank          | 0.000 | 04:04 |
| Subject Sample # 1 | 0.105 | 04:06 |
| Breath Volume      | 1.218 | 04:06 |
| Air Blank          | 0.000 | 04:06 |
| Air Blank          | 0.000 | 04:08 |
| Subject Sample # 2 | 0.099 | 04:09 |
| Breath Volume      | 1.378 | 04:09 |
| Air Blank          | 0.000 | 04:09 |
| Control Test       | 0.077 | 04:10 |
| Air Blank          | 0.000 | 04:10 |
| Diagnostics Check  | OK    | 04:10 |

---

Data Download Date/Time: 07/30/2018 10:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CABRAL,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: CABRAL,JONATHAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 05:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/12/2018 07:07:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:00 |
| Air Blank          | 0.000 | 07:01 |
| Control Test       | 0.078 | 07:01 |
| Air Blank          | 0.000 | 07:01 |
| Subject Sample # 1 | 0.097 | 07:02 |
| Breath Volume      | 3.933 | 07:02 |
| Air Blank          | 0.000 | 07:03 |
| Air Blank          | 0.000 | 07:04 |
| Subject Sample # 2 | 0.091 | 07:05 |
| Breath Volume      | 3.667 | 07:05 |
| Air Blank          | 0.000 | 07:06 |
| Control Test       | 0.079 | 07:06 |
| Air Blank          | 0.000 | 07:06 |
| Diagnostics Check  | OK    | 07:07 |

---

Data Download Date/Time: 07/30/2018 10:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICHARDSON,MICHAEL,R  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: PARRISH,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: CINTRON

DATE/TIME OF TEST: 07/13/2018 05:08:29

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 05:02 |
| Air Blank         |     | 0.000 | 05:02 |
| Control Test      |     | 0.078 | 05:03 |
| Air Blank         |     | 0.000 | 05:03 |
| Subject Sample    | # 1 | 0.033 | 05:04 |
| Breath Volume     |     | 2.246 | 05:04 |
| Air Blank         |     | 0.000 | 05:04 |
| Air Blank         |     | 0.000 | 05:06 |
| Subject Sample    | # 2 | 0.031 | 05:06 |
| Breath Volume     |     | 2.242 | 05:06 |
| Air Blank         |     | 0.000 | 05:07 |
| Control Test      |     | 0.079 | 05:07 |
| Air Blank         |     | 0.000 | 05:08 |
| Diagnostics Check |     | OK    | 05:08 |

---

Data Download Date/Time: 07/30/2018 10:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: HOWARD,ANTHONY  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: WASYLOW

DATE/TIME OF TEST: 07/15/2018 05:36:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:29 |
| Air Blank          | 0.000 | 05:29 |
| Control Test       | 0.079 | 05:30 |
| Air Blank          | 0.000 | 05:30 |
| Subject Sample # 1 | 0.159 | 05:31 |
| Breath Volume      | 1.898 | 05:31 |
| Air Blank          | 0.000 | 05:32 |
| Air Blank          | 0.000 | 05:34 |
| Subject Sample # 2 | 0.159 | 05:34 |
| Breath Volume      | 2.441 | 05:34 |
| Air Blank          | 0.000 | 05:35 |
| Control Test       | 0.079 | 05:35 |
| Air Blank          | 0.000 | 05:36 |
| Diagnostics Check  | OK    | 05:36 |

---

Data Download Date/Time: 07/30/2018 10:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAIN II,KENNETH,A  
BREATH TEST OPERATOR AGENCY: U of FI PD  
ARRESTING OFFICER: SHUTTERLY,BROOKE  
ARRESTING OFFICER AGENCY: Gainesville PD  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/15/2018 05:56:32

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:53 |                      |
| Air Blank         | 0.000 | 05:54 |                      |
| Control Test      | 0.078 | 05:54 |                      |
| Air Blank         | 0.000 | 05:54 |                      |
| Subject Sample    | # 1   | 05:55 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:55 |                      |
| Air Blank         | 0.000 | 05:55 |                      |
| Control Test      | 0.079 | 05:55 |                      |
| Air Blank         | 0.000 | 05:56 |                      |
| Diagnostics Check | OK    | 05:56 |                      |

---

Data Download Date/Time: 07/30/2018 10:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: HOWARD,ANTHONY  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: BAYLESS

DATE/TIME OF TEST: 07/28/2018 03:18:23

---

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 03:12 |            |
| Air Blank          | 0.000 | 03:12 |            |
| Control Test       | 0.079 | 03:12 |            |
| Air Blank          | 0.000 | 03:13 |            |
| Subject Sample # 1 | 0.115 | 03:14 |            |
| Breath Volume      | 1.781 | 03:14 |            |
| Air Blank          | 0.000 | 03:15 |            |
| Air Blank          |       | 03:17 | RFI Detect |
| Air Blank          | 0.000 | 03:17 |            |
| Control Test       | 0.081 | 03:17 |            |
| Air Blank          | 0.000 | 03:18 |            |
| Diagnostics Check  | OK    | 03:18 |            |

---

Data Download Date/Time: 07/30/2018 10:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: HOWARD,ANTHONY  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: BAYLESS

DATE/TIME OF TEST: 07/28/2018 03:28:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:21 |
| Air Blank          | 0.000 | 03:21 |
| Control Test       | 0.078 | 03:22 |
| Air Blank          | 0.000 | 03:22 |
| Subject Sample # 1 | 0.121 | 03:23 |
| Breath Volume      | 2.394 | 03:23 |
| Air Blank          | 0.000 | 03:24 |
| Air Blank          | 0.000 | 03:25 |
| Subject Sample # 2 | 0.121 | 03:26 |
| Breath Volume      | 2.785 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Control Test       | 0.077 | 03:27 |
| Air Blank          | 0.000 | 03:28 |
| Diagnostics Check  | OK    | 03:28 |

---

Data Download Date/Time: 07/30/2018 10:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: GUYTON,JOHN  
ARRESTING OFFICER AGENCY: Gainesville PD  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/28/2018 22:48:28

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:45 |                      |
| Air Blank         | 0.000 | 22:45 |                      |
| Control Test      | 0.078 | 22:46 |                      |
| Air Blank         | 0.000 | 22:46 |                      |
| Subject Sample    | # 1   | 22:47 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:47 |                      |
| Air Blank         | 0.000 | 22:47 |                      |
| Control Test      | 0.079 | 22:47 |                      |
| Air Blank         | 0.000 | 22:48 |                      |
| Diagnostics Check | OK    | 22:48 |                      |

---

Data Download Date/Time: 07/30/2018 10:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CABRAL,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: CABRAL,JONATHAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 07/31/2018 05:42:14

---

|                   |     |       |       |                   |
|-------------------|-----|-------|-------|-------------------|
| Diagnostics Check |     | OK    | 05:27 |                   |
| Air Blank         |     | 0.000 | 05:28 |                   |
| Control Test      |     | 0.079 | 05:28 |                   |
| Air Blank         |     | 0.000 | 05:29 |                   |
| Subject Sample    | # 1 | 0.157 | 05:30 | No .020 agreement |
| Breath Volume     |     | 1.105 | 05:30 |                   |
| Air Blank         |     | 0.000 | 05:30 |                   |
| Air Blank         |     | 0.000 | 05:32 |                   |
| Subject Sample    | # 2 | 0.138 | 05:35 | Volume Not Met    |
| Breath Volume     |     | 0.070 | 05:35 |                   |
| Air Blank         |     | 0.000 | 05:36 |                   |
| Air Blank         |     | 0.000 | 05:38 |                   |
| Subject Sample    | # 3 | 0.134 | 05:40 | No .020 agreement |
| Breath Volume     |     | 1.542 | 05:40 |                   |
| Air Blank         |     | 0.000 | 05:41 |                   |
| Control Test      |     | 0.078 | 05:41 |                   |
| Air Blank         |     | 0.000 | 05:42 |                   |
| Diagnostics Check |     | OK    | 05:42 |                   |

---

Data Download Date/Time: 08/27/2018 10:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLEMENTS,WILLIAM,J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CLEMENTS,WILIAM  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 07:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: MCAROY

DATE/TIME OF TEST: 07/01/2018 09:12:43

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 09:00 |                |
| Air Blank         |     | 0.000 | 09:01 |                |
| Control Test      |     | 0.080 | 09:01 |                |
| Air Blank         |     | 0.000 | 09:02 |                |
| Subject Sample    | # 1 | 0.000 | 09:05 | Volume Not Met |
| Breath Volume     |     | 0.406 | 09:05 |                |
| Air Blank         |     | 0.000 | 09:05 |                |
| Air Blank         |     | 0.000 | 09:07 |                |
| Subject Sample    | # 2 | 0.000 | 09:08 |                |
| Breath Volume     |     | 1.457 | 09:08 |                |
| Air Blank         |     | 0.000 | 09:08 |                |
| Air Blank         |     | 0.000 | 09:10 |                |
| Subject Sample    | # 3 | 0.000 | 09:11 |                |
| Breath Volume     |     | 1.195 | 09:11 |                |
| Air Blank         |     | 0.000 | 09:11 |                |
| Control Test      |     | 0.080 | 09:12 |                |
| Air Blank         |     | 0.000 | 09:12 |                |
| Diagnostics Check |     | OK    | 09:12 |                |

---

Data Download Date/Time: 07/30/2018 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARD,BENJAMIN,D  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: WARD,BENJAMIN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: VANBERKUM

DATE/TIME OF TEST: 07/02/2018 00:52:57

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:50 |                      |
| Air Blank         | 0.000 | 00:50 |                      |
| Control Test      | 0.081 | 00:50 |                      |
| Air Blank         | 0.000 | 00:51 |                      |
| Subject Sample    | # 1   | 00:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:51 |                      |
| Air Blank         | 0.000 | 00:52 |                      |
| Control Test      | 0.081 | 00:52 |                      |
| Air Blank         | 0.000 | 00:52 |                      |
| Diagnostics Check | OK    | 00:52 |                      |

---

Data Download Date/Time: 07/30/2018 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARVOTH,DANIEL,P  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: HARVOTH,DANIEL  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: CROZIER

DATE/TIME OF TEST: 07/03/2018 22:54:32

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:45 |                      |
| Air Blank         | 0.000 | 22:46 |                      |
| Control Test      | 0.081 | 22:46 |                      |
| Air Blank         | 0.000 | 22:47 |                      |
| Subject Sample    | # 1   | 22:50 | No Sample Provided   |
| Breath Volume     | 0.000 | 22:50 |                      |
| Air Blank         | 0.000 | 22:50 |                      |
| Air Blank         | 0.000 | 22:52 |                      |
| Subject Sample    | # 2   | 22:53 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:53 |                      |
| Air Blank         | 0.000 | 22:53 |                      |
| Control Test      | 0.081 | 22:53 |                      |
| Air Blank         | 0.000 | 22:54 |                      |
| Diagnostics Check | OK    | 22:54 |                      |

---

Data Download Date/Time: 07/30/2018 16:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KAHALLEY,JAMEY,J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: KAHALLEY,JAMEY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: KLOUS

DATE/TIME OF TEST: 07/04/2018 03:48:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:41 |
| Air Blank          | 0.000 | 03:41 |
| Control Test       | 0.080 | 03:42 |
| Air Blank          | 0.000 | 03:42 |
| Subject Sample # 1 | 0.124 | 03:44 |
| Breath Volume      | 2.539 | 03:44 |
| Air Blank          | 0.000 | 03:44 |
| Air Blank          | 0.000 | 03:46 |
| Subject Sample # 2 | 0.123 | 03:46 |
| Breath Volume      | 1.414 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Control Test       | 0.079 | 03:47 |
| Air Blank          | 0.000 | 03:48 |
| Diagnostics Check  | OK    | 03:48 |

---

Data Download Date/Time: 07/30/2018 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAHR,DANIEL,E  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: HAHR,DANIEL  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: BUI  
ARREST TIME: 20:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: STANLEY

DATE/TIME OF TEST: 07/04/2018 21:00:31

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:57 |                      |
| Air Blank         | 0.000 | 20:57 |                      |
| Control Test      | 0.080 | 20:58 |                      |
| Air Blank         | 0.000 | 20:58 |                      |
| Subject Sample    | # 1   | 20:59 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:59 |                      |
| Air Blank         | 0.000 | 20:59 |                      |
| Control Test      | 0.081 | 20:59 |                      |
| Air Blank         | 0.000 | 21:00 |                      |
| Diagnostics Check | OK    | 21:00 |                      |

---

Data Download Date/Time: 07/30/2018 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARD,BENJAMIN,D  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: WARD,BENJAMIN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 07/04/2018 22:49:03

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:46 |                      |
| Air Blank         | 0.000 | 22:46 |                      |
| Control Test      | 0.081 | 22:47 |                      |
| Air Blank         | 0.000 | 22:47 |                      |
| Subject Sample    | # 1   | 22:47 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:47 |                      |
| Air Blank         | 0.000 | 22:48 |                      |
| Control Test      | 0.081 | 22:48 |                      |
| Air Blank         | 0.000 | 22:48 |                      |
| Diagnostics Check | OK    | 22:49 |                      |

---

Data Download Date/Time: 07/30/2018 16:05



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOYE,JAMES,M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MOYE,JAMES  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PREVOST

DATE/TIME OF TEST: 07/05/2018 02:43:16

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:40 |                      |
| Air Blank         | 0.000 | 02:40 |                      |
| Control Test      | 0.080 | 02:41 |                      |
| Air Blank         | 0.000 | 02:41 |                      |
| Subject Sample    | # 1   | 02:41 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:41 |                      |
| Air Blank         | 0.000 | 02:42 |                      |
| Control Test      | 0.081 | 02:42 |                      |
| Air Blank         | 0.000 | 02:43 |                      |
| Diagnostics Check | OK    | 02:43 |                      |

---

Data Download Date/Time: 07/30/2018 16:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KAHALLEY,JAMEY,J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: KAHALLEY,JAMEY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 07/05/2018 02:59:32

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:50 |                      |
| Air Blank         | 0.000 | 02:51 |                      |
| Control Test      | 0.081 | 02:51 |                      |
| Air Blank         | 0.000 | 02:52 |                      |
| Subject Sample    | # 1   | 02:55 | No Sample Provided   |
| Breath Volume     | 0.000 | 02:55 |                      |
| Air Blank         | 0.000 | 02:55 |                      |
| Air Blank         | 0.000 | 02:57 |                      |
| Subject Sample    | # 2   | 02:58 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:58 |                      |
| Air Blank         | 0.000 | 02:58 |                      |
| Control Test      | 0.080 | 02:58 |                      |
| Air Blank         | 0.000 | 02:59 |                      |
| Diagnostics Check | OK    | 02:59 |                      |

---

Data Download Date/Time: 07/30/2018 16:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DANIELS,DAVID,C  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: DANIELS,DAVID  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 07/06/2018 00:18:33

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 00:12 |
| Air Blank         |     | 0.000 | 00:13 |
| Control Test      |     | 0.081 | 00:13 |
| Air Blank         |     | 0.000 | 00:14 |
| Subject Sample    | # 1 | 0.000 | 00:14 |
| Breath Volume     |     | 2.062 | 00:14 |
| Air Blank         |     | 0.000 | 00:14 |
| Air Blank         |     | 0.000 | 00:16 |
| Subject Sample    | # 2 | 0.000 | 00:17 |
| Breath Volume     |     | 2.339 | 00:17 |
| Air Blank         |     | 0.000 | 00:17 |
| Control Test      |     | 0.081 | 00:17 |
| Air Blank         |     | 0.000 | 00:18 |
| Diagnostics Check |     | OK    | 00:18 |

---

Data Download Date/Time: 07/30/2018 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUICK,STEPHEN,W  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: QUICK,STEPHEN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: HORTON

DATE/TIME OF TEST: 07/06/2018 02:34:29

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:25 |                      |
| Air Blank         | 0.000 | 02:26 |                      |
| Control Test      | 0.081 | 02:26 |                      |
| Air Blank         | 0.000 | 02:27 |                      |
| Subject Sample    | # 1   | 02:30 | No Sample Provided   |
| Breath Volume     | 0.000 | 02:30 |                      |
| Air Blank         | 0.000 | 02:30 |                      |
| Air Blank         | 0.000 | 02:32 |                      |
| Subject Sample    | # 2   | 02:33 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:33 |                      |
| Air Blank         | 0.000 | 02:33 |                      |
| Control Test      | 0.081 | 02:33 |                      |
| Air Blank         | 0.000 | 02:34 |                      |
| Diagnostics Check | OK    | 02:34 |                      |

---

Data Download Date/Time: 07/30/2018 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,STEPHEN,F  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CLARK,STEPHEN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 10:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:05

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/07/2018 11:48:18

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 11:44 |                      |
| Air Blank         | 0.000 | 11:45 |                      |
| Control Test      | 0.081 | 11:45 |                      |
| Air Blank         | 0.000 | 11:45 |                      |
| Subject Sample    | # 1   | 11:46 | Subject Test Refused |
| Breath Volume     | 0.000 | 11:46 |                      |
| Air Blank         | 0.000 | 11:47 |                      |
| Control Test      | 0.081 | 11:47 |                      |
| Air Blank         | 0.000 | 11:48 |                      |
| Diagnostics Check | OK    | 11:48 |                      |

---

Data Download Date/Time: 07/30/2018 16:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASCHAL,THOMAS,C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: PASCHAL,THOMAS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/07/2018 20:40:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:33 |
| Air Blank          | 0.000 | 20:34 |
| Control Test       | 0.082 | 20:34 |
| Air Blank          | 0.000 | 20:35 |
| Subject Sample # 1 | 0.214 | 20:36 |
| Breath Volume      | 2.687 | 20:36 |
| Air Blank          | 0.000 | 20:36 |
| Air Blank          | 0.000 | 20:38 |
| Subject Sample # 2 | 0.219 | 20:39 |
| Breath Volume      | 2.378 | 20:39 |
| Air Blank          | 0.000 | 20:39 |
| Control Test       | 0.081 | 20:40 |
| Air Blank          | 0.000 | 20:40 |
| Diagnostics Check  | OK    | 20:40 |

---

Data Download Date/Time: 07/30/2018 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARLSON,JOSHUA,L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CARLSON,JOSHUA  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: COFFMAN

DATE/TIME OF TEST: 07/07/2018 20:57:47

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:53 |                      |
| Air Blank         | 0.000 | 20:54 |                      |
| Control Test      | 0.081 | 20:54 |                      |
| Air Blank         | 0.000 | 20:55 |                      |
| Subject Sample    | # 1   | 20:56 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:56 |                      |
| Air Blank         | 0.000 | 20:56 |                      |
| Control Test      | 0.081 | 20:57 |                      |
| Air Blank         | 0.000 | 20:57 |                      |
| Diagnostics Check | OK    | 20:57 |                      |

---

Data Download Date/Time: 07/30/2018 16:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARTZOG,ROBERT,B  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: HARTZOG,ROBERT  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: LARSSON

DATE/TIME OF TEST: 07/08/2018 17:44:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:38 |
| Air Blank          | 0.000 | 17:39 |
| Control Test       | 0.081 | 17:39 |
| Air Blank          | 0.000 | 17:40 |
| Subject Sample # 1 | 0.000 | 17:40 |
| Breath Volume      | 2.382 | 17:40 |
| Air Blank          | 0.000 | 17:41 |
| Air Blank          | 0.000 | 17:43 |
| Subject Sample # 2 | 0.000 | 17:43 |
| Breath Volume      | 2.515 | 17:43 |
| Air Blank          | 0.000 | 17:44 |
| Control Test       | 0.081 | 17:44 |
| Air Blank          | 0.000 | 17:44 |
| Diagnostics Check  | OK    | 17:44 |

---

Data Download Date/Time: 07/30/2018 16:05



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASCHAL,THOMAS,C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: PASCHAL,THOMAS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: TRACY

DATE/TIME OF TEST: 07/08/2018 19:29:56

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:26 |                      |
| Air Blank         | 0.000 | 19:26 |                      |
| Control Test      | 0.081 | 19:26 |                      |
| Air Blank         | 0.000 | 19:27 |                      |
| Subject Sample    | # 1   | 19:28 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:28 |                      |
| Air Blank         | 0.000 | 19:29 |                      |
| Control Test      | 0.081 | 19:29 |                      |
| Air Blank         | 0.000 | 19:29 |                      |
| Diagnostics Check | OK    | 19:29 |                      |

---

Data Download Date/Time: 07/30/2018 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KAHALLEY,JAMEY,J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: KAHALLEY,JAMEY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 07/08/2018 19:43:15

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:37 |                      |
| Air Blank         | 0.000 | 19:38 |                      |
| Control Test      | 0.082 | 19:38 |                      |
| Air Blank         | 0.000 | 19:38 |                      |
| Subject Sample    | # 1   | 19:41 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:41 |                      |
| Air Blank         | 0.000 | 19:42 |                      |
| Control Test      | 0.081 | 19:42 |                      |
| Air Blank         |       | 19:42 | RFI Detect           |
| Air Blank         | 0.000 | 19:43 |                      |

---

Data Download Date/Time: 07/30/2018 16:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAMIREZ, MICHAEL, E  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: RAMIREZ, MICHAEL  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 07/08/2018 22:36:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:28 |
| Air Blank          | 0.000 | 22:29 |
| Control Test       | 0.081 | 22:29 |
| Air Blank          | 0.000 | 22:29 |
| Subject Sample # 1 | 0.191 | 22:31 |
| Breath Volume      | 1.710 | 22:31 |
| Air Blank          | 0.000 | 22:32 |
| Air Blank          | 0.000 | 22:33 |
| Subject Sample # 2 | 0.201 | 22:34 |
| Breath Volume      | 2.972 | 22:34 |
| Air Blank          | 0.000 | 22:35 |
| Control Test       | 0.078 | 22:35 |
| Air Blank          | 0.000 | 22:36 |
| Diagnostics Check  | OK    | 22:36 |

---

Data Download Date/Time: 07/30/2018 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASCHAL,THOMAS,C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: PASCHAL,PASCHAL  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: RAMSAY

DATE/TIME OF TEST: 07/08/2018 23:13:54

---

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 23:10 |            |
| Air Blank          | 0.000 | 23:11 |            |
| Control Test       | 0.081 | 23:11 |            |
| Air Blank          | 0.000 | 23:11 |            |
| Subject Sample # 1 | 0.144 | 23:12 |            |
| Breath Volume      | 2.058 | 23:12 |            |
| Air Blank          |       | 23:12 | RFI Detect |
| Air Blank          | 0.000 | 23:12 |            |
| Control Test       | 0.081 | 23:13 |            |
| Air Blank          | 0.000 | 23:13 |            |
| Diagnostics Check  | OK    | 23:13 |            |

---

Data Download Date/Time: 07/30/2018 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASCHAL,THOMAS,C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: PASCHAL,THOMAS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: RAMSAY

DATE/TIME OF TEST: 07/08/2018 23:22:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:15 |
| Air Blank          | 0.000 | 23:16 |
| Control Test       | 0.081 | 23:16 |
| Air Blank          | 0.000 | 23:17 |
| Subject Sample # 1 | 0.142 | 23:17 |
| Breath Volume      | 2.152 | 23:17 |
| Air Blank          | 0.000 | 23:18 |
| Air Blank          | 0.000 | 23:20 |
| Subject Sample # 2 | 0.138 | 23:20 |
| Breath Volume      | 2.007 | 23:20 |
| Air Blank          | 0.000 | 23:21 |
| Control Test       | 0.081 | 23:21 |
| Air Blank          | 0.000 | 23:21 |
| Diagnostics Check  | OK    | 23:22 |

---

Data Download Date/Time: 07/30/2018 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CALABRO,ALFRED,J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: YODER,JASON  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/09/2018 20:19:31

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:15 |                      |
| Air Blank         | 0.000 | 20:15 |                      |
| Control Test      | 0.081 | 20:15 |                      |
| Air Blank         | 0.000 | 20:16 |                      |
| Subject Sample    | # 1   | 20:18 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:18 |                      |
| Air Blank         | 0.000 | 20:18 |                      |
| Control Test      | 0.080 | 20:18 |                      |
| Air Blank         | 0.000 | 20:19 |                      |
| Diagnostics Check | OK    | 20:19 |                      |

---

Data Download Date/Time: 07/30/2018 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWELL,DAVID,M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: BERRYMAN,DOUGLAS  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 19:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: SHOVLIN

DATE/TIME OF TEST: 07/10/2018 21:17:46

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 21:04 |                    |
| Air Blank         | 0.000 | 21:04 |                    |
| Control Test      | 0.082 | 21:05 |                    |
| Air Blank         | 0.000 | 21:05 |                    |
| Subject Sample    | # 1   | 21:08 | No Sample Provided |
| Breath Volume     | 0.000 | 21:08 |                    |
| Air Blank         | 0.000 | 21:09 |                    |
| Air Blank         | 0.000 | 21:11 |                    |
| Subject Sample    | # 2   | 0.256 | 21:12              |
| Breath Volume     | 1.562 | 21:12 |                    |
| Air Blank         | 0.000 | 21:13 |                    |
| Air Blank         | 0.000 | 21:15 |                    |
| Subject Sample    | # 3   | 0.252 | 21:16              |
| Breath Volume     | 1.136 | 21:16 |                    |
| Air Blank         | 0.000 | 21:16 |                    |
| Control Test      | 0.081 | 21:17 |                    |
| Air Blank         | 0.000 | 21:17 |                    |
| Diagnostics Check | OK    | 21:17 |                    |

---

Data Download Date/Time: 07/30/2018 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,WILLIAM,E  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: OBANNON,ANTHONY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: DEMARCUS

DATE/TIME OF TEST: 07/13/2018 05:07:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:01 |
| Air Blank          | 0.000 | 05:01 |
| Control Test       | 0.081 | 05:01 |
| Air Blank          | 0.000 | 05:02 |
| Subject Sample # 1 | 0.065 | 05:02 |
| Breath Volume      | 3.066 | 05:02 |
| Air Blank          | 0.000 | 05:03 |
| Air Blank          | 0.000 | 05:05 |
| Subject Sample # 2 | 0.061 | 05:05 |
| Breath Volume      | 4.234 | 05:05 |
| Air Blank          | 0.000 | 05:06 |
| Control Test       | 0.079 | 05:06 |
| Air Blank          | 0.000 | 05:07 |
| Diagnostics Check  | OK    | 05:07 |

---

Data Download Date/Time: 07/30/2018 16:05



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALEXANDER,LLOYD,A  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: ALEXANDER,LLOYD  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 15:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:09

SUBJECT'S LAST NAME: HINOTE

DATE/TIME OF TEST: 07/14/2018 16:36:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:29 |
| Air Blank          | 0.000 | 16:30 |
| Control Test       | 0.082 | 16:30 |
| Air Blank          | 0.000 | 16:31 |
| Subject Sample # 1 | 0.000 | 16:32 |
| Breath Volume      | 2.085 | 16:32 |
| Air Blank          | 0.000 | 16:33 |
| Air Blank          | 0.000 | 16:35 |
| Subject Sample # 2 | 0.000 | 16:35 |
| Breath Volume      | 2.824 | 16:35 |
| Air Blank          | 0.000 | 16:36 |
| Control Test       | 0.082 | 16:36 |
| Air Blank          | 0.000 | 16:36 |
| Diagnostics Check  | OK    | 16:36 |

---

Data Download Date/Time: 07/30/2018 16:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPENCER, RAYMOND, L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: SPENCER, RAYMOND  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: CASILLAS

DATE/TIME OF TEST: 07/15/2018 04:16:00

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:11 |                      |
| Air Blank         | 0.000 | 04:11 |                      |
| Control Test      | 0.081 | 04:12 |                      |
| Air Blank         | 0.000 | 04:12 |                      |
| Subject Sample    | # 1   | 04:14 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:14 |                      |
| Air Blank         | 0.000 | 04:15 |                      |
| Control Test      | 0.082 | 04:15 |                      |
| Air Blank         | 0.000 | 04:15 |                      |
| Diagnostics Check | OK    | 04:16 |                      |

---

Data Download Date/Time: 07/30/2018 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALEXANDER,LLOYD,A  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: SPENCER,RAYMOND  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/17/2018 00:36:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:27 |
| Air Blank          | 0.000 | 00:27 |
| Control Test       | 0.081 | 00:27 |
| Air Blank          | 0.000 | 00:28 |
| Subject Sample # 1 | 0.021 | 00:31 |
| Breath Volume      | 1.621 | 00:31 |
| Air Blank          | 0.000 | 00:31 |
| Air Blank          | 0.000 | 00:33 |
| Subject Sample # 2 | 0.021 | 00:34 |
| Breath Volume      | 2.800 | 00:34 |
| Air Blank          | 0.000 | 00:35 |
| Control Test       | 0.082 | 00:35 |
| Air Blank          | 0.000 | 00:36 |
| Diagnostics Check  | OK    | 00:36 |

---

Data Download Date/Time: 07/30/2018 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOGGETTE,MASON,J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: DOGGETTE,MASON  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 19:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: RIVARD

DATE/TIME OF TEST: 07/18/2018 20:41:25

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 20:29 |                    |
| Air Blank         | 0.000 | 20:30 |                    |
| Control Test      | 0.081 | 20:30 |                    |
| Air Blank         | 0.000 | 20:30 |                    |
| Subject Sample    | # 1   | 20:34 | No Sample Provided |
| Breath Volume     | 0.000 | 20:34 |                    |
| Air Blank         | 0.000 | 20:34 |                    |
| Air Blank         | 0.000 | 20:36 |                    |
| Subject Sample    | # 2   | 20:36 | 0.255              |
| Breath Volume     | 2.769 | 20:36 |                    |
| Air Blank         | 0.000 | 20:37 |                    |
| Air Blank         | 0.000 | 20:39 |                    |
| Subject Sample    | # 3   | 20:39 | 0.256              |
| Breath Volume     | 3.550 | 20:39 |                    |
| Air Blank         | 0.000 | 20:40 |                    |
| Control Test      | 0.081 | 20:40 |                    |
| Air Blank         | 0.000 | 20:41 |                    |
| Diagnostics Check | OK    | 20:41 |                    |

---

Data Download Date/Time: 07/30/2018 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUSLER,DANNY,K  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: TUSLER,DANNY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: BARRY

DATE/TIME OF TEST: 07/19/2018 01:49:40

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:46 |                      |
| Air Blank         | 0.000 | 01:47 |                      |
| Control Test      | 0.081 | 01:47 |                      |
| Air Blank         | 0.000 | 01:48 |                      |
| Subject Sample    | # 1   | 01:48 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:48 |                      |
| Air Blank         | 0.000 | 01:48 |                      |
| Control Test      | 0.081 | 01:49 |                      |
| Air Blank         | 0.000 | 01:49 |                      |
| Diagnostics Check | OK    | 01:49 |                      |

---

Data Download Date/Time: 07/30/2018 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOZARD,CHRYSYAL,D  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: BOZARD,CHRYSYAL  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 12:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:22

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 07/19/2018 13:46:46

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 13:42 |                    |
| Air Blank         | 0.000 | 13:42 |                    |
| Control Test      | 0.081 | 13:43 |                    |
| Air Blank         | 0.000 | 13:43 |                    |
| Subject Sample    | # 1   | 13:45 | Interferent Detect |
| Breath Volume     | 1.964 | 13:45 |                    |
| Air Blank         | 0.000 | 13:45 |                    |
| Control Test      | 0.082 | 13:46 |                    |
| Air Blank         | 0.000 | 13:46 |                    |
| Diagnostics Check | OK    | 13:46 |                    |

---

Data Download Date/Time: 07/30/2018 16:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOZARD,CHRISTAL,D  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: BOZARD,CHRISTAL  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 12:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:22

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 07/19/2018 13:52:18

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 13:48 |                    |
| Air Blank         | 0.000 | 13:49 |                    |
| Control Test      | 0.081 | 13:49 |                    |
| Air Blank         | 0.000 | 13:50 |                    |
| Subject Sample    | # 1   | 13:50 | Interferent Detect |
| Breath Volume     | 1.546 | 13:50 |                    |
| Air Blank         | 0.000 | 13:51 |                    |
| Control Test      | 0.082 | 13:51 |                    |
| Air Blank         | 0.000 | 13:52 |                    |
| Diagnostics Check | OK    | 13:52 |                    |

---

Data Download Date/Time: 07/30/2018 16:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOZARD,CHRYSYAL,D  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: BOZARD,CHRYSYAL  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 12:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:17

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 07/19/2018 14:41:29

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 14:38 |                    |
| Air Blank         | 0.000 | 14:38 |                    |
| Control Test      | 0.081 | 14:38 |                    |
| Air Blank         | 0.000 | 14:39 |                    |
| Subject Sample    | # 1   | 14:39 | Interferent Detect |
| Breath Volume     | 2.304 | 14:39 |                    |
| Air Blank         | 0.000 | 14:40 |                    |
| Control Test      | 0.081 | 14:40 |                    |
| Air Blank         | 0.000 | 14:41 |                    |
| Diagnostics Check | OK    | 14:41 |                    |

---

Data Download Date/Time: 07/30/2018 16:06



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARMON, MICHAEL, A  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: HARMON, MICHAEL  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: GROOMS

DATE/TIME OF TEST: 07/20/2018 23:44:14

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 23:35 |                      |
| Air Blank         | 0.000     | 23:36 |                      |
| Control Test      | 0.081     | 23:36 |                      |
| Air Blank         | 0.000     | 23:37 |                      |
| Subject Sample    | # 1 0.156 | 23:40 | Volume Not Met       |
| Breath Volume     | 0.367     | 23:40 |                      |
| Air Blank         | 0.000     | 23:40 |                      |
| Air Blank         | 0.000     | 23:42 |                      |
| Subject Sample    | # 2       | 23:42 | Subject Test Refused |
| Breath Volume     | 0.000     | 23:42 |                      |
| Air Blank         | 0.000     | 23:43 |                      |
| Control Test      | 0.080     | 23:43 |                      |
| Air Blank         | 0.000     | 23:44 |                      |
| Diagnostics Check | OK        | 23:44 |                      |

---

Data Download Date/Time: 07/30/2018 16:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARMON,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: HARMON,MICHAEL  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: GROOMS

DATE/TIME OF TEST: 07/21/2018 00:16:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:09 |
| Air Blank          | 0.000 | 00:09 |
| Control Test       | 0.080 | 00:10 |
| Air Blank          | 0.000 | 00:10 |
| Subject Sample # 1 | 0.187 | 00:11 |
| Breath Volume      | 1.242 | 00:11 |
| Air Blank          | 0.000 | 00:11 |
| Air Blank          | 0.000 | 00:13 |
| Subject Sample # 2 | 0.180 | 00:14 |
| Breath Volume      | 1.425 | 00:14 |
| Air Blank          | 0.000 | 00:15 |
| Control Test       | 0.081 | 00:15 |
| Air Blank          | 0.000 | 00:16 |
| Diagnostics Check  | OK    | 00:16 |

---

Data Download Date/Time: 07/30/2018 16:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,COREY,M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: JONES,COREY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: THARP

DATE/TIME OF TEST: 07/21/2018 19:11:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:05 |
| Air Blank          | 0.000 | 19:05 |
| Control Test       | 0.081 | 19:06 |
| Air Blank          | 0.000 | 19:06 |
| Subject Sample # 1 | 0.000 | 19:07 |
| Breath Volume      | 1.546 | 19:07 |
| Air Blank          | 0.000 | 19:07 |
| Air Blank          | 0.000 | 19:09 |
| Subject Sample # 2 | 0.000 | 19:10 |
| Breath Volume      | 1.500 | 19:10 |
| Air Blank          | 0.000 | 19:10 |
| Control Test       | 0.080 | 19:11 |
| Air Blank          | 0.000 | 19:11 |
| Diagnostics Check  | OK    | 19:11 |

---

Data Download Date/Time: 07/30/2018 16:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALEXANDER,LLOYD,A  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: ALEXANDER,LLOYD  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/21/2018 22:17:01

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 22:05 |                    |
| Air Blank         | 0.000 | 22:06 |                    |
| Control Test      | 0.081 | 22:06 |                    |
| Air Blank         | 0.000 | 22:06 |                    |
| Subject Sample    | # 1   | 22:10 | No Sample Provided |
| Breath Volume     | 0.000 | 22:10 |                    |
| Air Blank         | 0.000 | 22:10 |                    |
| Air Blank         | 0.000 | 22:12 |                    |
| Subject Sample    | # 2   | 22:15 | No Sample Provided |
| Breath Volume     | 0.000 | 22:15 |                    |
| Air Blank         | 0.000 | 22:16 |                    |
| Control Test      | 0.082 | 22:16 |                    |
| Air Blank         | 0.000 | 22:16 |                    |
| Diagnostics Check | OK    | 22:17 |                    |

---

Data Download Date/Time: 07/30/2018 16:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASCHAL,THOMAS,C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: PASCHAL,THOMAS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 08:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:54

SUBJECT'S LAST NAME: RIJOS RIVERA

DATE/TIME OF TEST: 07/22/2018 10:16:22

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 10:15 |            |
| Air Blank         | 0.000 | 10:15 |            |
| Control Test      | 0.080 | 10:15 | RFI Detect |
| Air Blank         | 0.000 | 10:16 |            |

---

Data Download Date/Time: 07/30/2018 16:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASCHAL,THOMAS,C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: PASCHAL,THOMAS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 08:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:54

SUBJECT'S LAST NAME: RIJOS RIVERA

DATE/TIME OF TEST: 07/22/2018 10:25:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:18 |
| Air Blank          | 0.000 | 10:19 |
| Control Test       | 0.081 | 10:19 |
| Air Blank          | 0.000 | 10:20 |
| Subject Sample # 1 | 0.110 | 10:20 |
| Breath Volume      | 3.500 | 10:20 |
| Air Blank          | 0.000 | 10:21 |
| Air Blank          | 0.000 | 10:23 |
| Subject Sample # 2 | 0.114 | 10:24 |
| Breath Volume      | 4.484 | 10:24 |
| Air Blank          | 0.000 | 10:24 |
| Control Test       | 0.082 | 10:25 |
| Air Blank          | 0.000 | 10:25 |
| Diagnostics Check  | OK    | 10:25 |

---

Data Download Date/Time: 07/30/2018 16:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DESOUSA,CARLOS,J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: DESOUSA,CARLOS  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: DREYER

DATE/TIME OF TEST: 07/24/2018 04:45:58

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:40 |                      |
| Air Blank         | 0.000 | 04:41 |                      |
| Control Test      | 0.081 | 04:41 |                      |
| Air Blank         | 0.000 | 04:41 |                      |
| Subject Sample    | # 1   | 04:44 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:44 |                      |
| Air Blank         | 0.000 | 04:45 |                      |
| Control Test      | 0.082 | 04:45 |                      |
| Air Blank         | 0.000 | 04:45 |                      |
| Diagnostics Check | OK    | 04:45 |                      |

---

Data Download Date/Time: 07/30/2018 16:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COFFELT,DYLAN,H  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: COFFELT,DYLAN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 07/26/2018 01:57:53

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:55 |                      |
| Air Blank         | 0.000 | 01:55 |                      |
| Control Test      | 0.081 | 01:55 |                      |
| Air Blank         | 0.000 | 01:56 |                      |
| Subject Sample    | # 1   | 01:56 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:56 |                      |
| Air Blank         | 0.000 | 01:56 |                      |
| Control Test      | 0.081 | 01:57 |                      |
| Air Blank         | 0.000 | 01:57 |                      |
| Diagnostics Check | OK    | 01:57 |                      |

---

Data Download Date/Time: 07/30/2018 16:06



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSNAK, BRANDEN, M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: RUSNAK, BRANDEN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: BURLEIGH

DATE/TIME OF TEST: 07/26/2018 03:12:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:04 |
| Air Blank          | 0.000 | 03:05 |
| Control Test       | 0.082 | 03:05 |
| Air Blank          | 0.000 | 03:06 |
| Subject Sample # 1 | 0.132 | 03:07 |
| Breath Volume      | 1.171 | 03:07 |
| Air Blank          | 0.000 | 03:07 |
| Air Blank          | 0.000 | 03:09 |
| Subject Sample # 2 | 0.137 | 03:11 |
| Breath Volume      | 2.210 | 03:11 |
| Air Blank          | 0.000 | 03:11 |
| Control Test       | 0.081 | 03:12 |
| Air Blank          | 0.000 | 03:12 |
| Diagnostics Check  | OK    | 03:12 |

---

Data Download Date/Time: 07/30/2018 16:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMPSON,ANDREW,J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: SIMPSON,ANDREW  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: FLOYD

DATE/TIME OF TEST: 07/26/2018 03:59:31

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:55 |                      |
| Air Blank         | 0.000 | 03:56 |                      |
| Control Test      | 0.081 | 03:56 |                      |
| Air Blank         | 0.000 | 03:57 |                      |
| Subject Sample    | # 1   | 03:58 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:58 |                      |
| Air Blank         | 0.000 | 03:58 |                      |
| Control Test      | 0.082 | 03:58 |                      |
| Air Blank         | 0.000 | 03:59 |                      |
| Diagnostics Check | OK    | 03:59 |                      |

---

Data Download Date/Time: 07/30/2018 16:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPENCER, RAYMOND, L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: SPENCER, RAYMOND  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 07/29/2018 04:29:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:22 |
| Air Blank          | 0.000 | 04:22 |
| Control Test       | 0.081 | 04:23 |
| Air Blank          | 0.000 | 04:23 |
| Subject Sample # 1 | 0.142 | 04:24 |
| Breath Volume      | 2.082 | 04:24 |
| Air Blank          | 0.000 | 04:25 |
| Air Blank          | 0.000 | 04:27 |
| Subject Sample # 2 | 0.135 | 04:28 |
| Breath Volume      | 1.492 | 04:28 |
| Air Blank          | 0.000 | 04:28 |
| Control Test       | 0.081 | 04:28 |
| Air Blank          | 0.000 | 04:29 |
| Diagnostics Check  | OK    | 04:29 |

---

Data Download Date/Time: 07/30/2018 16:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARLSON,JOSHUA,L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CARLSON,JOSHUA  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629801  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: IDEIS

DATE/TIME OF TEST: 07/30/2018 00:44:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:36 |
| Air Blank          | 0.000 | 00:36 |
| Control Test       | 0.081 | 00:36 |
| Air Blank          | 0.000 | 00:37 |
| Subject Sample # 1 | 0.152 | 00:39 |
| Breath Volume      | 3.453 | 00:39 |
| Air Blank          | 0.000 | 00:40 |
| Air Blank          | 0.000 | 00:41 |
| Subject Sample # 2 | 0.144 | 00:42 |
| Breath Volume      | 2.667 | 00:42 |
| Air Blank          | 0.000 | 00:43 |
| Control Test       | 0.081 | 00:43 |
| Air Blank          | 0.000 | 00:43 |
| Diagnostics Check  | OK    | 00:44 |

---

Data Download Date/Time: 07/30/2018 16:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A  
BREATH TEST OPERATOR AGENCY: Lake City PD  
ARRESTING OFFICER: MILLIGAN,CONNIE  
ARRESTING OFFICER AGENCY: Lake City PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: JERNIGAN III

DATE/TIME OF TEST: 07/01/2018 17:39:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:32 |
| Air Blank          | 0.000 | 17:32 |
| Control Test       | 0.082 | 17:33 |
| Air Blank          | 0.000 | 17:33 |
| Subject Sample # 1 | 0.192 | 17:34 |
| Breath Volume      | 1.597 | 17:34 |
| Air Blank          | 0.000 | 17:35 |
| Air Blank          | 0.000 | 17:36 |
| Subject Sample # 2 | 0.176 | 17:37 |
| Breath Volume      | 1.449 | 17:37 |
| Air Blank          | 0.000 | 17:38 |
| Control Test       | 0.079 | 17:38 |
| Air Blank          | 0.000 | 17:39 |
| Diagnostics Check  | OK    | 17:39 |

---

Data Download Date/Time: 08/01/2018 02:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAURES, BRANDYN, K  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BAURES, BRANDYN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: LEWIN

DATE/TIME OF TEST: 07/04/2018 15:39:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:33 |
| Air Blank          | 0.000 | 15:33 |
| Control Test       | 0.081 | 15:33 |
| Air Blank          | 0.000 | 15:34 |
| Subject Sample # 1 | 0.236 | 15:35 |
| Breath Volume      | 1.351 | 15:35 |
| Air Blank          | 0.000 | 15:36 |
| Air Blank          | 0.000 | 15:37 |
| Subject Sample # 2 | 0.246 | 15:38 |
| Breath Volume      | 2.089 | 15:38 |
| Air Blank          | 0.000 | 15:38 |
| Control Test       | 0.082 | 15:39 |
| Air Blank          | 0.000 | 15:39 |
| Diagnostics Check  | OK    | 15:39 |

---

Data Download Date/Time: 08/01/2018 02:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A  
BREATH TEST OPERATOR AGENCY: Lake City PD  
ARRESTING OFFICER: MILLIGAN,CONNIE  
ARRESTING OFFICER AGENCY: Lake City PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 19:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: SPRADLEY

DATE/TIME OF TEST: 07/05/2018 20:42:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:35 |
| Air Blank          | 0.000 | 20:36 |
| Control Test       | 0.082 | 20:36 |
| Air Blank          | 0.000 | 20:37 |
| Subject Sample # 1 | 0.225 | 20:37 |
| Breath Volume      | 3.972 | 20:37 |
| Air Blank          | 0.000 | 20:38 |
| Air Blank          | 0.000 | 20:40 |
| Subject Sample # 2 | 0.224 | 20:40 |
| Breath Volume      | 3.011 | 20:40 |
| Air Blank          | 0.000 | 20:41 |
| Control Test       | 0.080 | 20:41 |
| Air Blank          | 0.000 | 20:42 |
| Diagnostics Check  | OK    | 20:42 |

---

Data Download Date/Time: 08/01/2018 02:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHELAN, MATTHEW, S  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: HUGHES, ALFRED  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MARSEE

DATE/TIME OF TEST: 07/06/2018 22:26:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:20 |
| Air Blank          | 0.000 | 22:20 |
| Control Test       | 0.081 | 22:21 |
| Air Blank          | 0.000 | 22:21 |
| Subject Sample # 1 | 0.242 | 22:22 |
| Breath Volume      | 1.828 | 22:22 |
| Air Blank          | 0.000 | 22:22 |
| Air Blank          | 0.000 | 22:24 |
| Subject Sample # 2 | 0.227 | 22:24 |
| Breath Volume      | 1.632 | 22:24 |
| Air Blank          | 0.000 | 22:25 |
| Control Test       | 0.080 | 22:25 |
| Air Blank          | 0.000 | 22:26 |
| Diagnostics Check  | OK    | 22:26 |

---

Data Download Date/Time: 08/01/2018 02:07



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLIE,JOHN,R  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: TILLIE,JOHN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 07:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/15/2018 08:37:33

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 08:31 |
| Air Blank         |     | 0.000 | 08:31 |
| Control Test      |     | 0.081 | 08:31 |
| Air Blank         |     | 0.000 | 08:32 |
| Subject Sample    | # 1 | 0.108 | 08:33 |
| Breath Volume     |     | 4.773 | 08:33 |
| Air Blank         |     | 0.000 | 08:33 |
| Air Blank         |     | 0.000 | 08:35 |
| Subject Sample    | # 2 | 0.089 | 08:35 |
| Breath Volume     |     | 1.789 | 08:35 |
| Air Blank         |     | 0.000 | 08:36 |
| Control Test      |     | 0.081 | 08:36 |
| Air Blank         |     | 0.000 | 08:37 |
| Diagnostics Check |     | OK    | 08:37 |

---

Data Download Date/Time: 08/01/2018 02:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLAY,DARREL,E  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: CLAY,DARREL  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: GEIGER

DATE/TIME OF TEST: 07/17/2018 02:42:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:36 |
| Air Blank          | 0.000 | 02:36 |
| Control Test       | 0.080 | 02:36 |
| Air Blank          | 0.000 | 02:37 |
| Subject Sample # 1 | 0.124 | 02:37 |
| Breath Volume      | 3.035 | 02:37 |
| Air Blank          | 0.000 | 02:38 |
| Air Blank          | 0.000 | 02:40 |
| Subject Sample # 2 | 0.129 | 02:40 |
| Breath Volume      | 2.832 | 02:40 |
| Air Blank          | 0.000 | 02:41 |
| Control Test       | 0.078 | 02:41 |
| Air Blank          | 0.000 | 02:41 |
| Diagnostics Check  | OK    | 02:42 |

---

Data Download Date/Time: 08/01/2018 02:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHELAN, MATTHEW, S  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: PHELAN, MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: STOCKWELL

DATE/TIME OF TEST: 07/24/2018 03:41:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:35 |
| Air Blank          | 0.000 | 03:35 |
| Control Test       | 0.082 | 03:36 |
| Air Blank          | 0.000 | 03:36 |
| Subject Sample # 1 | 0.125 | 03:37 |
| Breath Volume      | 1.882 | 03:37 |
| Air Blank          | 0.000 | 03:38 |
| Air Blank          | 0.000 | 03:39 |
| Subject Sample # 2 | 0.130 | 03:40 |
| Breath Volume      | 1.960 | 03:40 |
| Air Blank          | 0.000 | 03:41 |
| Control Test       | 0.080 | 03:41 |
| Air Blank          | 0.000 | 03:41 |
| Diagnostics Check  | OK    | 03:41 |

---

Data Download Date/Time: 08/01/2018 02:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBINSON,STUART,P  
BREATH TEST OPERATOR AGENCY: Lake City PD  
ARRESTING OFFICER: ROBINSON,STUART  
ARRESTING OFFICER AGENCY: Lake City PD  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 07/26/2018 21:31:24

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:28 |                      |
| Air Blank         | 0.000 | 21:28 |                      |
| Control Test      | 0.082 | 21:29 |                      |
| Air Blank         | 0.000 | 21:29 |                      |
| Subject Sample    | # 1   | 21:30 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:30 |                      |
| Air Blank         | 0.000 | 21:30 |                      |
| Control Test      | 0.081 | 21:30 |                      |
| Air Blank         | 0.000 | 21:31 |                      |
| Diagnostics Check | OK    | 21:31 |                      |

---

Data Download Date/Time: 08/01/2018 02:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: VOLLERTSEN,CRAIG  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 07/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SALINAS

DATE/TIME OF TEST: 07/28/2018 03:41:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:35 |
| Air Blank          | 0.000 | 03:36 |
| Control Test       | 0.079 | 03:36 |
| Air Blank          | 0.000 | 03:36 |
| Subject Sample # 1 | 0.216 | 03:37 |
| Breath Volume      | 3.347 | 03:37 |
| Air Blank          | 0.000 | 03:38 |
| Air Blank          | 0.000 | 03:39 |
| Subject Sample # 2 | 0.210 | 03:40 |
| Breath Volume      | 2.585 | 03:40 |
| Air Blank          | 0.000 | 03:40 |
| Control Test       | 0.078 | 03:41 |
| Air Blank          | 0.000 | 03:41 |
| Diagnostics Check  | OK    | 03:41 |

---

Data Download Date/Time: 08/22/2018 10:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMSON,JACOB,L  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: NICHOLSON,MICHAEL  
ARRESTING OFFICER AGENCY: Live Oak PD  
LAST AGENCY INSPECTION: 07/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: CHERRY

DATE/TIME OF TEST: 07/09/2018 00:55:40

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 00:42 |                |
| Air Blank         |     | 0.000 | 00:42 |                |
| Control Test      |     | 0.082 | 00:42 |                |
| Air Blank         |     | 0.000 | 00:43 |                |
| Subject Sample    | # 1 | 0.115 | 00:45 |                |
| Breath Volume     |     | 2.347 | 00:45 |                |
| Air Blank         |     | 0.000 | 00:46 |                |
| Air Blank         |     | 0.000 | 00:47 |                |
| Subject Sample    | # 2 | 0.110 | 00:50 | Volume Not Met |
| Breath Volume     |     | 0.019 | 00:50 |                |
| Air Blank         |     | 0.000 | 00:51 |                |
| Air Blank         |     | 0.000 | 00:53 |                |
| Subject Sample    | # 3 | 0.110 | 00:53 |                |
| Breath Volume     |     | 1.546 | 00:53 |                |
| Air Blank         |     | 0.000 | 00:54 |                |
| Control Test      |     | 0.080 | 00:55 |                |
| Air Blank         |     | 0.000 | 00:55 |                |
| Diagnostics Check |     | OK    | 00:55 |                |

---

Data Download Date/Time: 08/31/2018 05:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,ALEXIS,  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: HERNANDEZ,ALEXIS  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 07/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 07/23/2018 00:52:18

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 00:45 |
| Air Blank         |     | 0.000 | 00:46 |
| Control Test      |     | 0.083 | 00:46 |
| Air Blank         |     | 0.000 | 00:47 |
| Subject Sample    | # 1 | 0.261 | 00:47 |
| Breath Volume     |     | 1.476 | 00:47 |
| Air Blank         |     | 0.000 | 00:48 |
| Air Blank         |     | 0.000 | 00:50 |
| Subject Sample    | # 2 | 0.245 | 00:50 |
| Breath Volume     |     | 1.835 | 00:50 |
| Air Blank         |     | 0.000 | 00:51 |
| Control Test      |     | 0.081 | 00:51 |
| Air Blank         |     | 0.000 | 00:52 |
| Diagnostics Check |     | OK    | 00:52 |

---

Data Download Date/Time: 08/31/2018 05:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WADFORD,CHARLES,T  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: WADFORD,CHARLES  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 07/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: LINDSEY

DATE/TIME OF TEST: 07/24/2018 08:04:23

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 08:01 |                      |
| Air Blank         | 0.000 | 08:01 |                      |
| Control Test      | 0.082 | 08:02 |                      |
| Air Blank         | 0.000 | 08:02 |                      |
| Subject Sample    | # 1   | 08:03 | Subject Test Refused |
| Breath Volume     | 0.000 | 08:03 |                      |
| Air Blank         | 0.000 | 08:03 |                      |
| Control Test      | 0.082 | 08:03 |                      |
| Air Blank         | 0.000 | 08:04 |                      |
| Diagnostics Check | OK    | 08:04 |                      |

---

Data Download Date/Time: 08/31/2018 05:03



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SINGLETARY, MATHEW, C  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: SINGLETARY, MATHEW  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 07/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: DRAPER

DATE/TIME OF TEST: 07/26/2018 00:48:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:42 |
| Air Blank          | 0.000 | 00:42 |
| Control Test       | 0.081 | 00:42 |
| Air Blank          | 0.000 | 00:43 |
| Subject Sample # 1 | 0.121 | 00:43 |
| Breath Volume      | 5.250 | 00:43 |
| Air Blank          | 0.000 | 00:44 |
| Air Blank          | 0.000 | 00:46 |
| Subject Sample # 2 | 0.122 | 00:46 |
| Breath Volume      | 5.375 | 00:46 |
| Air Blank          | 0.000 | 00:47 |
| Control Test       | 0.082 | 00:47 |
| Air Blank          | 0.000 | 00:48 |
| Diagnostics Check  | OK    | 00:48 |

---

Data Download Date/Time: 08/31/2018 05:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WADFORD,CHALRES,T  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: DAVIS,KEITH  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 07/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A2  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/29/2018 07:29:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:23 |
| Air Blank          | 0.000 | 07:23 |
| Control Test       | 0.082 | 07:23 |
| Air Blank          | 0.000 | 07:24 |
| Subject Sample # 1 | 0.129 | 07:24 |
| Breath Volume      | 3.281 | 07:24 |
| Air Blank          | 0.000 | 07:25 |
| Air Blank          | 0.000 | 07:27 |
| Subject Sample # 2 | 0.126 | 07:27 |
| Breath Volume      | 2.187 | 07:27 |
| Air Blank          | 0.000 | 07:28 |
| Control Test       | 0.082 | 07:28 |
| Air Blank          | 0.000 | 07:29 |
| Diagnostics Check  | OK    | 07:29 |

---

Data Download Date/Time: 08/31/2018 05:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SISK,CARLOS,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SISK,CARLOS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 07/07/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/11/2018 19:59:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:50 |
| Air Blank          | 0.000 | 19:51 |
| Control Test       | 0.080 | 19:51 |
| Air Blank          | 0.000 | 19:52 |
| Subject Sample # 1 | 0.189 | 19:53 |
| Breath Volume      | 3.277 | 19:53 |
| Air Blank          | 0.000 | 19:54 |
| Air Blank          | 0.000 | 19:56 |
| Subject Sample # 2 | 0.189 | 19:57 |
| Breath Volume      | 2.136 | 19:57 |
| Air Blank          | 0.000 | 19:58 |
| Control Test       | 0.080 | 19:58 |
| Air Blank          | 0.000 | 19:59 |
| Diagnostics Check  | OK    | 19:59 |

---

Data Download Date/Time: 08/22/2018 14:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: YOUNGREN,DENNIS  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 684131  
DRY GAS EXPIRY DATE: 01/23/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 07/01/2018 21:29:00

---

|                   |     |       |       |                      |
|-------------------|-----|-------|-------|----------------------|
| Diagnostics Check |     | OK    | 21:18 |                      |
| Air Blank         |     | 0.000 | 21:19 |                      |
| Control Test      |     | 0.081 | 21:19 |                      |
| Air Blank         |     | 0.000 | 21:19 |                      |
| Subject Sample    | # 1 | 0.268 | 21:22 | Volume Not Met       |
| Breath Volume     |     | 0.757 | 21:22 |                      |
| Air Blank         |     | 0.000 | 21:23 |                      |
| Air Blank         |     | 0.000 | 21:25 |                      |
| Subject Sample    | # 2 |       | 21:27 | Subject Test Refused |
| Breath Volume     |     | 0.000 | 21:27 |                      |
| Air Blank         |     | 0.000 | 21:28 |                      |
| Control Test      |     | 0.078 | 21:28 |                      |
| Air Blank         |     | 0.000 | 21:28 |                      |
| Diagnostics Check |     | OK    | 21:29 |                      |

---

Data Download Date/Time: 07/15/2018 10:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEIERS, BLAINE, T  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: WHITE, TRAVIS  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 684131  
DRY GAS EXPIRY DATE: 01/23/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: BILDNER

DATE/TIME OF TEST: 07/11/2018 22:00:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:48 |
| Air Blank          | 0.000 | 21:48 |
| Control Test       | 0.082 | 21:48 |
| Air Blank          | 0.000 | 21:49 |
| Subject Sample # 1 | 0.273 | 21:50 |
| Breath Volume      | 1.144 | 21:50 |
| Air Blank          | 0.000 | 21:50 |
| Air Blank          | 0.000 | 21:52 |
| Subject Sample # 2 | 0.296 | 21:54 |
| Breath Volume      | 1.425 | 21:54 |
| Air Blank          | 0.000 | 21:54 |
| Air Blank          | 0.000 | 21:56 |
| Subject Sample # 3 | 0.281 | 21:58 |
| Breath Volume      | 1.937 | 21:58 |
| Air Blank          | 0.000 | 21:59 |
| Control Test       | 0.084 | 21:59 |
| Air Blank          | 0.000 | 22:00 |
| Diagnostics Check  | OK    | 22:00 |

---

Data Download Date/Time: 07/15/2018 10:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUNTER,JOHN,H  
BREATH TEST OPERATOR AGENCY: OTHER Fed  
ARRESTING OFFICER: HUNTER,JOHN  
ARRESTING OFFICER AGENCY: OTHER Fed  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 684131  
DRY GAS EXPIRY DATE: 01/23/2020  
VIOLATION CODE: OTHER  
ARREST TIME: 20:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: BACULI

DATE/TIME OF TEST: 07/14/2018 22:41:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:35 |
| Air Blank          | 0.000 | 22:35 |
| Control Test       | 0.082 | 22:35 |
| Air Blank          | 0.000 | 22:36 |
| Subject Sample # 1 | 0.000 | 22:37 |
| Breath Volume      | 3.542 | 22:37 |
| Air Blank          | 0.000 | 22:37 |
| Air Blank          | 0.000 | 22:39 |
| Subject Sample # 2 | 0.000 | 22:39 |
| Breath Volume      | 4.210 | 22:39 |
| Air Blank          | 0.000 | 22:40 |
| Control Test       | 0.083 | 22:40 |
| Air Blank          | 0.000 | 22:41 |
| Diagnostics Check  | OK    | 22:41 |

---

Data Download Date/Time: 07/15/2018 10:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: YOUNGREN,DENNIS  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 684131  
DRY GAS EXPIRY DATE: 01/23/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: UNTERREINER

DATE/TIME OF TEST: 07/14/2018 23:25:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:18 |
| Air Blank          | 0.000 | 23:19 |
| Control Test       | 0.081 | 23:19 |
| Air Blank          | 0.000 | 23:20 |
| Subject Sample # 1 | 0.177 | 23:20 |
| Breath Volume      | 2.859 | 23:20 |
| Air Blank          | 0.000 | 23:21 |
| Air Blank          | 0.000 | 23:23 |
| Subject Sample # 2 | 0.180 | 23:23 |
| Breath Volume      | 2.824 | 23:23 |
| Air Blank          | 0.000 | 23:24 |
| Control Test       | 0.081 | 23:24 |
| Air Blank          | 0.000 | 23:25 |
| Diagnostics Check  | OK    | 23:25 |

---

Data Download Date/Time: 07/15/2018 10:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BANKS,SEAN,C  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: BANKS,SEAN  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 684131  
DRY GAS EXPIRY DATE: 01/23/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: DUFFY

DATE/TIME OF TEST: 07/14/2018 23:43:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:34 |
| Air Blank          | 0.000 | 23:35 |
| Control Test       | 0.081 | 23:35 |
| Air Blank          | 0.000 | 23:36 |
| Subject Sample # 1 | 0.157 | 23:38 |
| Breath Volume      | 1.347 | 23:38 |
| Air Blank          | 0.000 | 23:38 |
| Air Blank          | 0.000 | 23:40 |
| Subject Sample # 2 | 0.146 | 23:42 |
| Breath Volume      | 1.917 | 23:42 |
| Air Blank          | 0.000 | 23:42 |
| Control Test       | 0.082 | 23:43 |
| Air Blank          | 0.000 | 23:43 |
| Diagnostics Check  | OK    | 23:43 |

---

Data Download Date/Time: 07/15/2018 10:18



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTALDO,JOHN,J  
BREATH TEST OPERATOR AGENCY: OTHER Fed  
ARRESTING OFFICER: CASTALDO,JOHN  
ARRESTING OFFICER AGENCY: OTHER Fed  
LAST AGENCY INSPECTION: 07/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 684131  
DRY GAS EXPIRY DATE: 01/23/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: BRINKLEY

DATE/TIME OF TEST: 07/15/2018 21:16:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:09 |
| Air Blank          | 0.000 | 21:10 |
| Control Test       | 0.081 | 21:10 |
| Air Blank          | 0.000 | 21:11 |
| Subject Sample # 1 | 0.114 | 21:11 |
| Breath Volume      | 1.671 | 21:11 |
| Air Blank          | 0.000 | 21:12 |
| Air Blank          | 0.000 | 21:14 |
| Subject Sample # 2 | 0.114 | 21:14 |
| Breath Volume      | 1.460 | 21:14 |
| Air Blank          | 0.000 | 21:15 |
| Control Test       | 0.082 | 21:15 |
| Air Blank          | 0.000 | 21:16 |
| Diagnostics Check  | OK    | 21:16 |

---

Data Download Date/Time: 08/15/2018 11:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEIERS,BLAINE,T  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: WEIERS,BLAINE  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 07/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 684131  
DRY GAS EXPIRY DATE: 01/23/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: POST

DATE/TIME OF TEST: 07/21/2018 00:27:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:20 |
| Air Blank          | 0.000 | 00:21 |
| Control Test       | 0.083 | 00:21 |
| Air Blank          | 0.000 | 00:21 |
| Subject Sample # 1 | 0.198 | 00:22 |
| Breath Volume      | 2.132 | 00:22 |
| Air Blank          | 0.000 | 00:23 |
| Air Blank          | 0.000 | 00:24 |
| Subject Sample # 2 | 0.203 | 00:25 |
| Breath Volume      | 2.226 | 00:25 |
| Air Blank          | 0.000 | 00:26 |
| Control Test       | 0.082 | 00:26 |
| Air Blank          | 0.000 | 00:27 |
| Diagnostics Check  | OK    | 00:27 |

---

Data Download Date/Time: 08/15/2018 11:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCAPIN,BRANDON,L  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: SCAPIN,BRANDON  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 07/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 684131  
DRY GAS EXPIRY DATE: 01/23/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: GODFREY

DATE/TIME OF TEST: 07/21/2018 02:24:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:18 |
| Air Blank          | 0.000 | 02:18 |
| Control Test       | 0.082 | 02:19 |
| Air Blank          | 0.000 | 02:19 |
| Subject Sample # 1 | 0.185 | 02:20 |
| Breath Volume      | 1.363 | 02:20 |
| Air Blank          | 0.000 | 02:20 |
| Air Blank          | 0.000 | 02:22 |
| Subject Sample # 2 | 0.188 | 02:23 |
| Breath Volume      | 2.078 | 02:23 |
| Air Blank          | 0.000 | 02:23 |
| Control Test       | 0.081 | 02:24 |
| Air Blank          | 0.000 | 02:24 |
| Diagnostics Check  | OK    | 02:24 |

---

Data Download Date/Time: 08/15/2018 11:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: YOUNGREN,DENNIS  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 07/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 684131  
DRY GAS EXPIRY DATE: 01/23/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: CHESTERMAN

DATE/TIME OF TEST: 07/24/2018 05:06:43

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 05:00 |
| Air Blank         |     | 0.000 | 05:00 |
| Control Test      |     | 0.082 | 05:01 |
| Air Blank         |     | 0.000 | 05:01 |
| Subject Sample    | # 1 | 0.111 | 05:02 |
| Breath Volume     |     | 4.007 | 05:02 |
| Air Blank         |     | 0.000 | 05:02 |
| Air Blank         |     | 0.000 | 05:04 |
| Subject Sample    | # 2 | 0.094 | 05:05 |
| Breath Volume     |     | 2.093 | 05:05 |
| Air Blank         |     | 0.000 | 05:05 |
| Control Test      |     | 0.082 | 05:06 |
| Air Blank         |     | 0.000 | 05:06 |
| Diagnostics Check |     | OK    | 05:06 |

---

Data Download Date/Time: 08/15/2018 11:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOFF,SARA,B  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: WHITE,TRAVIS  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 07/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 684131  
DRY GAS EXPIRY DATE: 01/23/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: LAW

DATE/TIME OF TEST: 07/25/2018 15:17:10

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 15:05 |                |
| Air Blank         | 0.000     | 15:06 |                |
| Control Test      | 0.081     | 15:06 |                |
| Air Blank         | 0.000     | 15:07 |                |
| Subject Sample    | # 1 0.000 | 15:10 | Volume Not Met |
| Breath Volume     | 0.519     | 15:10 |                |
| Air Blank         | 0.000     | 15:10 |                |
| Air Blank         | 0.000     | 15:12 |                |
| Subject Sample    | # 2 0.000 | 15:15 | Volume Not Met |
| Breath Volume     | 0.855     | 15:15 |                |
| Air Blank         | 0.000     | 15:16 |                |
| Control Test      | 0.082     | 15:16 |                |
| Air Blank         | 0.000     | 15:17 |                |
| Diagnostics Check | OK        | 15:17 |                |

---

Data Download Date/Time: 08/15/2018 11:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOFF,SARA,B  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: WHITE,TRAVIS  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 07/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 684131  
DRY GAS EXPIRY DATE: 01/23/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: LAW

DATE/TIME OF TEST: 07/25/2018 15:29:50

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 15:21 |                      |
| Air Blank         | 0.000     | 15:21 |                      |
| Control Test      | 0.082     | 15:22 |                      |
| Air Blank         | 0.000     | 15:22 |                      |
| Subject Sample    | # 1 0.000 | 15:25 | Volume Not Met       |
| Breath Volume     | 0.300     | 15:25 |                      |
| Air Blank         | 0.000     | 15:26 |                      |
| Air Blank         | 0.000     | 15:28 |                      |
| Subject Sample    | # 2       | 15:28 | Subject Test Refused |
| Breath Volume     | 0.000     | 15:28 |                      |
| Air Blank         | 0.000     | 15:28 |                      |
| Control Test      | 0.081     | 15:29 |                      |
| Air Blank         | 0.000     | 15:29 |                      |
| Diagnostics Check | OK        | 15:29 |                      |

---

Data Download Date/Time: 08/15/2018 11:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEIERS,BLAINE,T  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: WEIERS,BLAINE  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 07/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 684131  
DRY GAS EXPIRY DATE: 01/23/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: FENDLASON

DATE/TIME OF TEST: 07/26/2018 00:28:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:22 |
| Air Blank          | 0.000 | 00:22 |
| Control Test       | 0.082 | 00:22 |
| Air Blank          | 0.000 | 00:23 |
| Subject Sample # 1 | 0.075 | 00:24 |
| Breath Volume      | 2.480 | 00:24 |
| Air Blank          | 0.000 | 00:24 |
| Air Blank          | 0.000 | 00:26 |
| Subject Sample # 2 | 0.074 | 00:27 |
| Breath Volume      | 2.472 | 00:27 |
| Air Blank          | 0.000 | 00:27 |
| Control Test       | 0.085 | 00:28 |
| Air Blank          | 0.000 | 00:28 |
| Diagnostics Check  | OK    | 00:28 |

---

Data Download Date/Time: 08/15/2018 11:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: YOUNGREN,DENNIS  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 07/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 684131  
DRY GAS EXPIRY DATE: 01/23/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: STANTON

DATE/TIME OF TEST: 07/29/2018 03:04:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:56 |
| Air Blank          | 0.000 | 02:57 |
| Control Test       | 0.081 | 02:57 |
| Air Blank          | 0.000 | 02:57 |
| Subject Sample # 1 | 0.156 | 03:00 |
| Breath Volume      | 1.519 | 03:00 |
| Air Blank          | 0.000 | 03:00 |
| Air Blank          | 0.000 | 03:02 |
| Subject Sample # 2 | 0.158 | 03:03 |
| Breath Volume      | 2.441 | 03:03 |
| Air Blank          | 0.000 | 03:03 |
| Control Test       | 0.081 | 03:04 |
| Air Blank          | 0.000 | 03:04 |
| Diagnostics Check  | OK    | 03:04 |

---

Data Download Date/Time: 08/15/2018 11:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BANKS,SEAN,C  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: BANKS,SEAN  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 07/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 684131  
DRY GAS EXPIRY DATE: 01/23/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BARR

DATE/TIME OF TEST: 07/30/2018 01:46:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:40 |
| Air Blank          | 0.000 | 01:41 |
| Control Test       | 0.082 | 01:41 |
| Air Blank          | 0.000 | 01:42 |
| Subject Sample # 1 | 0.035 | 01:42 |
| Breath Volume      | 2.089 | 01:42 |
| Air Blank          | 0.000 | 01:43 |
| Air Blank          | 0.000 | 01:44 |
| Subject Sample # 2 | 0.034 | 01:45 |
| Breath Volume      | 2.031 | 01:45 |
| Air Blank          | 0.000 | 01:46 |
| Control Test       | 0.082 | 01:46 |
| Air Blank          | 0.000 | 01:46 |
| Diagnostics Check  | OK    | 01:46 |

---

Data Download Date/Time: 08/15/2018 11:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: WALTENBURG,MARCUS  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A2  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: WALDRON

DATE/TIME OF TEST: 07/13/2018 00:59:01

---

|                   |    |       |            |
|-------------------|----|-------|------------|
| Diagnostics Check | OK | 00:58 |            |
| Air Blank         |    | 00:58 | RFI Detect |
| Air Blank         |    | 00:59 | RFI Detect |

---

Data Download Date/Time: 08/22/2018 08:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: WALTENBURG,MARCUS  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A2  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: WALDRON

DATE/TIME OF TEST: 07/13/2018 01:00:54

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 01:00 |            |
| Air Blank         |       | 01:00 | RFI Detect |
| Air Blank         | 0.000 | 01:00 |            |

---

Data Download Date/Time: 08/22/2018 08:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: WALTENBURG,MARCUS  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A2  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: WALDRON

DATE/TIME OF TEST: 07/13/2018 01:06:07

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 01:02 |            |
| Air Blank         | 0.000 | 01:02 |            |
| Control Test      | 0.081 | 01:02 |            |
| Air Blank         | 0.000 | 01:03 |            |
| Subject Sample    | # 1   | 01:04 | RFI Detect |
| Breath Volume     | 0.726 | 01:04 |            |
| Air Blank         | 0.000 | 01:05 |            |
| Control Test      | 0.081 | 01:05 |            |
| Air Blank         | 0.000 | 01:06 |            |
| Diagnostics Check | OK    | 01:06 |            |

---

Data Download Date/Time: 08/22/2018 08:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: WALTENBURG,MARCUS  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A2  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: WALDRON

DATE/TIME OF TEST: 07/13/2018 01:11:33

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 01:07 |            |
| Air Blank         | 0.000 | 01:07 |            |
| Control Test      | 0.079 | 01:08 |            |
| Air Blank         | 0.000 | 01:08 |            |
| Subject Sample    | # 1   | 01:09 | RFI Detect |
| Breath Volume     | 0.191 | 01:09 |            |
| Air Blank         | 0.000 | 01:10 |            |
| Control Test      | 0.080 | 01:10 |            |
| Air Blank         | 0.000 | 01:11 |            |
| Diagnostics Check | OK    | 01:11 |            |

---

Data Download Date/Time: 08/22/2018 08:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: WALTENBURG,MARCUS  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A2  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: WALDRON

DATE/TIME OF TEST: 07/13/2018 01:27:30

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:23 |                      |
| Air Blank         | 0.000 | 01:24 |                      |
| Control Test      | 0.080 | 01:24 |                      |
| Air Blank         | 0.000 | 01:24 |                      |
| Subject Sample    | # 1   | 01:25 | Subject Test Refused |
| Breath Volume     | 0.015 | 01:25 |                      |
| Air Blank         | 0.000 | 01:26 |                      |
| Control Test      | 0.080 | 01:26 |                      |
| Air Blank         |       | 01:27 | RFI Detect           |
| Air Blank         | 0.000 | 01:27 |                      |

---

Data Download Date/Time: 08/22/2018 08:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLBERT,ERIC,R  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: HUNTER,RANDY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A2  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: LOCKWOOD

DATE/TIME OF TEST: 07/13/2018 22:30:30

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 22:28 |            |
| Air Blank         | 0.000 | 22:29 |            |
| Control Test      | 0.080 | 22:29 |            |
| Air Blank         |       | 22:30 | RFI Detect |
| Air Blank         | 0.000 | 22:30 |            |

---

Data Download Date/Time: 08/22/2018 08:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLBERT,ERIC,R  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: HUNTER,RANDY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A2  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: LOCKWOOD

DATE/TIME OF TEST: 07/13/2018 22:38:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:32 |
| Air Blank          | 0.000 | 22:32 |
| Control Test       | 0.080 | 22:33 |
| Air Blank          | 0.000 | 22:33 |
| Subject Sample # 1 | 0.161 | 22:34 |
| Breath Volume      | 2.164 | 22:34 |
| Air Blank          | 0.000 | 22:34 |
| Air Blank          | 0.000 | 22:36 |
| Subject Sample # 2 | 0.164 | 22:37 |
| Breath Volume      | 4.140 | 22:37 |
| Air Blank          | 0.000 | 22:38 |
| Control Test       | 0.080 | 22:38 |
| Air Blank          | 0.000 | 22:38 |
| Diagnostics Check  | OK    | 22:38 |

---

Data Download Date/Time: 08/22/2018 08:34



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANTON,NICHOLAS,A  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: ANTON,NICHOLAS  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A2  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: OTHER  
ARREST TIME: 01:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HELMAN

DATE/TIME OF TEST: 07/15/2018 02:20:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:10 |
| Air Blank          | 0.000 | 02:11 |
| Control Test       | 0.080 | 02:11 |
| Air Blank          | 0.000 | 02:11 |
| Subject Sample # 1 | 0.176 | 02:13 |
| Breath Volume      | 1.187 | 02:13 |
| Air Blank          | 0.000 | 02:14 |
| Air Blank          | 0.000 | 02:16 |
| Subject Sample # 2 | 0.175 | 02:18 |
| Breath Volume      | 1.140 | 02:18 |
| Air Blank          | 0.000 | 02:19 |
| Control Test       | 0.080 | 02:19 |
| Air Blank          | 0.000 | 02:20 |
| Diagnostics Check  | OK    | 02:20 |

---

Data Download Date/Time: 08/22/2018 08:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEER,BRADLEY,J  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: MEER,BRADLEY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A2  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:31

SUBJECT'S LAST NAME: CHANCEY

DATE/TIME OF TEST: 07/19/2018 16:13:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:06 |
| Air Blank          | 0.000 | 16:06 |
| Control Test       | 0.081 | 16:07 |
| Air Blank          | 0.000 | 16:07 |
| Subject Sample # 1 | 0.139 | 16:08 |
| Breath Volume      | 1.535 | 16:08 |
| Air Blank          | 0.000 | 16:09 |
| Air Blank          | 0.000 | 16:11 |
| Subject Sample # 2 | 0.141 | 16:11 |
| Breath Volume      | 1.445 | 16:11 |
| Air Blank          | 0.000 | 16:12 |
| Control Test       | 0.081 | 16:12 |
| Air Blank          | 0.000 | 16:13 |
| Diagnostics Check  | OK    | 16:13 |

---

Data Download Date/Time: 08/22/2018 08:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: DRAWDY,LAINE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A2  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: DELORME

DATE/TIME OF TEST: 07/22/2018 21:30:45

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 21:25 |            |
| Air Blank         | 0.000     | 21:25 |            |
| Control Test      | 0.081     | 21:25 |            |
| Air Blank         | 0.000     | 21:26 |            |
| Subject Sample    | # 1 0.201 | 21:26 |            |
| Breath Volume     | 1.992     | 21:26 |            |
| Air Blank         | 0.000     | 21:27 |            |
| Air Blank         | 0.000     | 21:29 |            |
| Subject Sample    | # 2 0.204 | 21:29 |            |
| Breath Volume     | 1.878     | 21:29 |            |
| Air Blank         |           | 21:30 | RFI Detect |
| Air Blank         |           | 21:30 | RFI Detect |

---

Data Download Date/Time: 08/22/2018 08:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: WALTENBURG,MARCUS  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A2  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: AMATO

DATE/TIME OF TEST: 07/27/2018 03:47:01

---

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 03:40 |            |
| Air Blank          | 0.000 | 03:41 |            |
| Control Test       | 0.082 | 03:41 |            |
| Air Blank          | 0.000 | 03:41 |            |
| Subject Sample # 1 | 0.000 | 03:42 |            |
| Breath Volume      | 3.851 | 03:42 |            |
| Air Blank          | 0.000 | 03:42 |            |
| Air Blank          | 0.000 | 03:44 |            |
| Subject Sample # 2 | 0.000 | 03:45 |            |
| Breath Volume      | 3.925 | 03:45 |            |
| Air Blank          | 0.000 | 03:45 |            |
| Control Test       | 0.080 | 03:46 |            |
| Air Blank          |       | 03:46 | RFI Detect |
| Air Blank          | 0.000 | 03:47 |            |

---

Data Download Date/Time: 08/22/2018 08:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: WALTENBURG,MARCUS  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A2  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MICHEL

DATE/TIME OF TEST: 07/28/2018 00:50:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:44 |
| Air Blank          | 0.000 | 00:45 |
| Control Test       | 0.082 | 00:45 |
| Air Blank          | 0.000 | 00:45 |
| Subject Sample # 1 | 0.148 | 00:46 |
| Breath Volume      | 3.625 | 00:46 |
| Air Blank          | 0.000 | 00:47 |
| Air Blank          | 0.000 | 00:48 |
| Subject Sample # 2 | 0.144 | 00:49 |
| Breath Volume      | 3.679 | 00:49 |
| Air Blank          | 0.000 | 00:50 |
| Control Test       | 0.078 | 00:50 |
| Air Blank          | 0.000 | 00:50 |
| Diagnostics Check  | OK    | 00:50 |

---

Data Download Date/Time: 08/22/2018 08:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAYFORD,GREGG,R  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: STUBLEY,NATE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A2  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/31/2018 16:48:03

---

|                   |    |       |            |
|-------------------|----|-------|------------|
| Diagnostics Check | OK | 16:47 |            |
| Air Blank         |    | 16:47 | RFI Detect |
| Air Blank         |    | 16:48 | RFI Detect |

---

Data Download Date/Time: 08/22/2018 08:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAYFORD,GREGG,R  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: STUBLEY,NATE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A2  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/31/2018 17:00:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:50 |
| Air Blank          | 0.000 | 16:50 |
| Control Test       | 0.081 | 16:50 |
| Air Blank          | 0.000 | 16:51 |
| Subject Sample # 1 | 0.153 | 16:54 |
| Breath Volume      | 1.609 | 16:54 |
| Air Blank          | 0.000 | 16:55 |
| Air Blank          | 0.000 | 16:56 |
| Subject Sample # 2 | 0.145 | 16:58 |
| Breath Volume      | 1.281 | 16:58 |
| Air Blank          | 0.000 | 16:59 |
| Control Test       | 0.079 | 16:59 |
| Air Blank          | 0.000 | 16:59 |
| Diagnostics Check  | OK    | 17:00 |

---

Data Download Date/Time: 08/22/2018 08:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: CUEVAS,BENJAMIN  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: HABEL

DATE/TIME OF TEST: 07/01/2018 00:01:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:54 |
| Air Blank          | 0.000 | 23:55 |
| Control Test       | 0.079 | 23:55 |
| Air Blank          | 0.000 | 23:56 |
| Subject Sample # 1 | 0.152 | 23:56 |
| Breath Volume      | 2.902 | 23:56 |
| Air Blank          | 0.000 | 23:57 |
| Air Blank          | 0.000 | 23:59 |
| Subject Sample # 2 | 0.143 | 23:59 |
| Breath Volume      | 1.375 | 23:59 |
| Air Blank          | 0.000 | 00:00 |
| Control Test       | 0.078 | 00:00 |
| Air Blank          | 0.000 | 00:01 |
| Diagnostics Check  | OK    | 00:01 |

---

Data Download Date/Time: 07/12/2018 09:03



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMLIN,BRETT,W  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: HAMLIN,BRETT  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BENTO

DATE/TIME OF TEST: 07/01/2018 02:47:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:40 |
| Air Blank          | 0.000 | 02:40 |
| Control Test       | 0.080 | 02:41 |
| Air Blank          | 0.000 | 02:41 |
| Subject Sample # 1 | 0.119 | 02:42 |
| Breath Volume      | 1.582 | 02:42 |
| Air Blank          | 0.000 | 02:43 |
| Air Blank          | 0.000 | 02:45 |
| Subject Sample # 2 | 0.113 | 02:45 |
| Breath Volume      | 1.375 | 02:45 |
| Air Blank          | 0.000 | 02:46 |
| Control Test       | 0.079 | 02:46 |
| Air Blank          | 0.000 | 02:47 |
| Diagnostics Check  | OK    | 02:47 |

---

Data Download Date/Time: 07/12/2018 09:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PATEL,MITHIL,V  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: PATEL,MITHIL  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 11:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: STACK

DATE/TIME OF TEST: 07/02/2018 00:44:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:38 |
| Air Blank          | 0.000 | 00:38 |
| Control Test       | 0.079 | 00:38 |
| Air Blank          | 0.000 | 00:39 |
| Subject Sample # 1 | 0.216 | 00:40 |
| Breath Volume      | 2.390 | 00:40 |
| Air Blank          | 0.000 | 00:40 |
| Air Blank          | 0.000 | 00:42 |
| Subject Sample # 2 | 0.226 | 00:43 |
| Breath Volume      | 1.851 | 00:43 |
| Air Blank          | 0.000 | 00:43 |
| Control Test       | 0.079 | 00:44 |
| Air Blank          | 0.000 | 00:44 |
| Diagnostics Check  | OK    | 00:44 |

---

Data Download Date/Time: 07/12/2018 09:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: RIDLE,EVAN  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: MEJIA RAMOS

DATE/TIME OF TEST: 07/02/2018 04:12:20

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 04:05 |
| Air Blank         |     | 0.000 | 04:05 |
| Control Test      |     | 0.079 | 04:06 |
| Air Blank         |     | 0.000 | 04:06 |
| Subject Sample    | # 1 | 0.148 | 04:07 |
| Breath Volume     |     | 1.324 | 04:07 |
| Air Blank         |     | 0.000 | 04:08 |
| Air Blank         |     | 0.000 | 04:09 |
| Subject Sample    | # 2 | 0.146 | 04:10 |
| Breath Volume     |     | 1.859 | 04:10 |
| Air Blank         |     | 0.000 | 04:11 |
| Control Test      |     | 0.079 | 04:11 |
| Air Blank         |     | 0.000 | 04:12 |
| Diagnostics Check |     | OK    | 04:12 |

---

Data Download Date/Time: 07/12/2018 09:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FENTON, WILLIAM, B  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: FENTON, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: SCHOENWALDER

DATE/TIME OF TEST: 07/04/2018 20:01:30

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 20:00 |            |
| Air Blank         |       | 20:01 | RFI Detect |
| Air Blank         | 0.000 | 20:01 |            |

---

Data Download Date/Time: 07/12/2018 09:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FENTON, WILLIAM, B  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: FENTON, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: SCHOENWALDER

DATE/TIME OF TEST: 07/04/2018 20:04:12

---

|                   |    |       |            |
|-------------------|----|-------|------------|
| Diagnostics Check | OK | 20:03 |            |
| Air Blank         |    | 20:03 | RFI Detect |
| Air Blank         |    | 20:04 | RFI Detect |

---

Data Download Date/Time: 07/12/2018 09:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FENTON, WILLIAM, B  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: FENTON, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: SCHOENWALDER

DATE/TIME OF TEST: 07/04/2018 20:12:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:06 |
| Air Blank          | 0.000 | 20:06 |
| Control Test       | 0.080 | 20:06 |
| Air Blank          | 0.000 | 20:07 |
| Subject Sample # 1 | 0.140 | 20:07 |
| Breath Volume      | 1.113 | 20:07 |
| Air Blank          | 0.000 | 20:08 |
| Air Blank          | 0.000 | 20:10 |
| Subject Sample # 2 | 0.145 | 20:11 |
| Breath Volume      | 2.250 | 20:11 |
| Air Blank          | 0.000 | 20:11 |
| Control Test       | 0.079 | 20:12 |
| Air Blank          | 0.000 | 20:12 |
| Diagnostics Check  | OK    | 20:12 |

---

Data Download Date/Time: 07/12/2018 09:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LIMB,HUNTER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: RONE

DATE/TIME OF TEST: 07/04/2018 21:52:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:46 |
| Air Blank          | 0.000 | 21:46 |
| Control Test       | 0.080 | 21:46 |
| Air Blank          | 0.000 | 21:47 |
| Subject Sample # 1 | 0.172 | 21:47 |
| Breath Volume      | 2.207 | 21:47 |
| Air Blank          | 0.000 | 21:48 |
| Air Blank          | 0.000 | 21:50 |
| Subject Sample # 2 | 0.168 | 21:50 |
| Breath Volume      | 1.683 | 21:50 |
| Air Blank          | 0.000 | 21:51 |
| Control Test       | 0.079 | 21:51 |
| Air Blank          | 0.000 | 21:52 |
| Diagnostics Check  | OK    | 21:52 |

---

Data Download Date/Time: 07/12/2018 09:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUTLER,ALONZO,  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: MEER,BRADLEY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 07/04/2018 23:21:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:15 |
| Air Blank          | 0.000 | 23:15 |
| Control Test       | 0.079 | 23:15 |
| Air Blank          | 0.000 | 23:16 |
| Subject Sample # 1 | 0.231 | 23:17 |
| Breath Volume      | 1.945 | 23:17 |
| Air Blank          | 0.000 | 23:17 |
| Air Blank          | 0.000 | 23:19 |
| Subject Sample # 2 | 0.230 | 23:19 |
| Breath Volume      | 2.156 | 23:19 |
| Air Blank          | 0.000 | 23:20 |
| Control Test       | 0.079 | 23:20 |
| Air Blank          | 0.000 | 23:21 |
| Diagnostics Check  | OK    | 23:21 |

---

Data Download Date/Time: 07/12/2018 09:03



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: WALTENBURG,MARCUS  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: ALBRIGHT

DATE/TIME OF TEST: 07/05/2018 01:53:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:44 |
| Air Blank          | 0.000 | 01:45 |
| Control Test       | 0.080 | 01:45 |
| Air Blank          | 0.000 | 01:46 |
| Subject Sample # 1 | 0.153 | 01:46 |
| Breath Volume      | 2.597 | 01:46 |
| Air Blank          | 0.000 | 01:47 |
| Air Blank          | 0.000 | 01:48 |
| Subject Sample # 2 | 0.126 | 01:49 |
| Breath Volume      | 1.917 | 01:49 |
| Air Blank          | 0.000 | 01:50 |
| Air Blank          | 0.000 | 01:51 |
| Subject Sample # 3 | 0.145 | 01:52 |
| Breath Volume      | 2.640 | 01:52 |
| Air Blank          | 0.000 | 01:52 |
| Control Test       | 0.079 | 01:53 |
| Air Blank          | 0.000 | 01:53 |
| Diagnostics Check  | OK    | 01:53 |

---

Data Download Date/Time: 07/12/2018 09:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LIMB,HUNTER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: BEAULIEU

DATE/TIME OF TEST: 07/05/2018 02:59:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:53 |
| Air Blank          | 0.000 | 02:53 |
| Control Test       | 0.080 | 02:53 |
| Air Blank          | 0.000 | 02:54 |
| Subject Sample # 1 | 0.220 | 02:54 |
| Breath Volume      | 1.972 | 02:54 |
| Air Blank          | 0.000 | 02:55 |
| Air Blank          | 0.000 | 02:57 |
| Subject Sample # 2 | 0.220 | 02:57 |
| Breath Volume      | 1.988 | 02:57 |
| Air Blank          | 0.000 | 02:58 |
| Control Test       | 0.080 | 02:58 |
| Air Blank          | 0.000 | 02:59 |
| Diagnostics Check  | OK    | 02:59 |

---

Data Download Date/Time: 07/12/2018 09:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: WALTENBURG,MARCUS  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 07/05/2018 23:22:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:15 |
| Air Blank          | 0.000 | 23:16 |
| Control Test       | 0.080 | 23:16 |
| Air Blank          | 0.000 | 23:17 |
| Subject Sample # 1 | 0.000 | 23:17 |
| Breath Volume      | 2.191 | 23:17 |
| Air Blank          | 0.000 | 23:18 |
| Air Blank          | 0.000 | 23:20 |
| Subject Sample # 2 | 0.000 | 23:20 |
| Breath Volume      | 2.402 | 23:20 |
| Air Blank          | 0.000 | 23:21 |
| Control Test       | 0.079 | 23:21 |
| Air Blank          | 0.000 | 23:21 |
| Diagnostics Check  | OK    | 23:22 |

---

Data Download Date/Time: 07/12/2018 09:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: NELSON,JEREMIAH  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 07/06/2018 23:41:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:33 |
| Air Blank          | 0.000 | 23:34 |
| Control Test       | 0.081 | 23:34 |
| Air Blank          | 0.000 | 23:34 |
| Subject Sample # 1 | 0.275 | 23:36 |
| Breath Volume      | 2.644 | 23:36 |
| Air Blank          | 0.000 | 23:37 |
| Air Blank          | 0.000 | 23:38 |
| Subject Sample # 2 | 0.269 | 23:39 |
| Breath Volume      | 2.253 | 23:39 |
| Air Blank          | 0.000 | 23:40 |
| Control Test       | 0.078 | 23:40 |
| Air Blank          | 0.000 | 23:41 |
| Diagnostics Check  | OK    | 23:41 |

---

Data Download Date/Time: 07/12/2018 09:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: NELSON,JEREMIAH  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: REID BRATHWAITE

DATE/TIME OF TEST: 07/07/2018 03:29:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:23 |
| Air Blank          | 0.000 | 03:23 |
| Control Test       | 0.080 | 03:24 |
| Air Blank          | 0.000 | 03:24 |
| Subject Sample # 1 | 0.137 | 03:25 |
| Breath Volume      | 1.207 | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Subject Sample # 2 | 0.143 | 03:28 |
| Breath Volume      | 2.050 | 03:28 |
| Air Blank          | 0.000 | 03:28 |
| Control Test       | 0.079 | 03:29 |
| Air Blank          | 0.000 | 03:29 |
| Diagnostics Check  | OK    | 03:29 |

---

Data Download Date/Time: 07/12/2018 09:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALTENBURG,MARCUS,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SHIREMAN,MARK  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BUSH

DATE/TIME OF TEST: 07/08/2018 04:03:53

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 03:57 |
| Air Blank         |     | 0.000 | 03:58 |
| Control Test      |     | 0.080 | 03:58 |
| Air Blank         |     | 0.000 | 03:58 |
| Subject Sample    | # 1 | 0.175 | 03:59 |
| Breath Volume     |     | 2.566 | 03:59 |
| Air Blank         |     | 0.000 | 04:00 |
| Air Blank         |     | 0.000 | 04:01 |
| Subject Sample    | # 2 | 0.169 | 04:02 |
| Breath Volume     |     | 2.613 | 04:02 |
| Air Blank         |     | 0.000 | 04:02 |
| Control Test      |     | 0.079 | 04:03 |
| Air Blank         |     | 0.000 | 04:03 |
| Diagnostics Check |     | OK    | 04:03 |

---

Data Download Date/Time: 07/12/2018 09:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBB,MATT,  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: WEBB,HAL  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A2  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: BUI  
ARREST TIME: 12:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:31

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 07/04/2018 14:51:42

---

|                    |       |       |                           |
|--------------------|-------|-------|---------------------------|
| Diagnostics Check  | OK    | 14:45 |                           |
| Air Blank          | 0.000 | 14:45 |                           |
| Control Test       | 0.076 | 14:45 |                           |
| Air Blank          | 0.000 | 14:46 |                           |
| Subject Sample # 1 | 0.094 | 14:47 |                           |
| Breath Volume      | 4.597 | 14:47 |                           |
| Air Blank          | 0.000 | 14:47 |                           |
| Air Blank          | 0.000 | 14:49 |                           |
| Subject Sample # 2 | 0.094 | 14:50 |                           |
| Breath Volume      | 5.414 | 14:50 |                           |
| Air Blank          | 0.000 | 14:50 |                           |
| Control Test       | 0.074 | 14:51 | Control Outside Tolerance |
| Air Blank          | 0.000 | 14:51 |                           |

---

Data Download Date/Time: 07/25/2018 17:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CURRY,ALISA,D  
BREATH TEST OPERATOR AGENCY: Hamilton County SO  
ARRESTING OFFICER: CZABAN,NICHOLAS  
ARRESTING OFFICER AGENCY: FL Dept of AG  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A1  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GRAVES

DATE/TIME OF TEST: 07/12/2018 04:42:15

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 04:29 |                |
| Air Blank         |     | 0.000 | 04:30 |                |
| Control Test      |     | 0.082 | 04:30 |                |
| Air Blank         |     | 0.000 | 04:31 |                |
| Subject Sample    | # 1 | 0.096 | 04:34 | Volume Not Met |
| Breath Volume     |     | 0.308 | 04:34 |                |
| Air Blank         |     | 0.000 | 04:35 |                |
| Air Blank         |     | 0.000 | 04:36 |                |
| Subject Sample    | # 2 | 0.127 | 04:37 |                |
| Breath Volume     |     | 1.546 | 04:37 |                |
| Air Blank         |     | 0.000 | 04:38 |                |
| Air Blank         |     | 0.000 | 04:39 |                |
| Subject Sample    | # 3 | 0.130 | 04:40 |                |
| Breath Volume     |     | 1.675 | 04:40 |                |
| Air Blank         |     | 0.000 | 04:41 |                |
| Control Test      |     | 0.079 | 04:41 |                |
| Air Blank         |     | 0.000 | 04:42 |                |
| Diagnostics Check |     | OK    | 04:42 |                |

---

Data Download Date/Time: 07/30/2018 01:23



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,DANIEL,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MARTIN,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A1  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 07/14/2018 23:09:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:59 |
| Air Blank          | 0.000 | 22:59 |
| Control Test       | 0.081 | 23:00 |
| Air Blank          | 0.000 | 23:00 |
| Subject Sample # 1 | 0.159 | 23:02 |
| Breath Volume      | 1.609 | 23:02 |
| Air Blank          | 0.000 | 23:03 |
| Air Blank          | 0.000 | 23:05 |
| Subject Sample # 2 | 0.168 | 23:08 |
| Breath Volume      | 1.941 | 23:08 |
| Air Blank          | 0.000 | 23:08 |
| Control Test       | 0.079 | 23:09 |
| Air Blank          | 0.000 | 23:09 |
| Diagnostics Check  | OK    | 23:09 |

---

Data Download Date/Time: 07/30/2018 01:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CURRY,ALISA,D  
BREATH TEST OPERATOR AGENCY: Hamilton County SO  
ARRESTING OFFICER: ANDERSON,JOSHUA  
ARRESTING OFFICER AGENCY: Hamilton County SO  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A1  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: ROHMAN

DATE/TIME OF TEST: 07/17/2018 22:40:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:31 |
| Air Blank          | 0.000 | 22:31 |
| Control Test       | 0.081 | 22:32 |
| Air Blank          | 0.000 | 22:32 |
| Subject Sample # 1 | 0.056 | 22:34 |
| Breath Volume      | 1.113 | 22:34 |
| Air Blank          | 0.000 | 22:34 |
| Air Blank          | 0.000 | 22:36 |
| Subject Sample # 2 | 0.051 | 22:38 |
| Breath Volume      | 1.320 | 22:38 |
| Air Blank          | 0.000 | 22:39 |
| Control Test       | 0.080 | 22:39 |
| Air Blank          | 0.000 | 22:40 |
| Diagnostics Check  | OK    | 22:40 |

---

Data Download Date/Time: 07/30/2018 01:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUSGROVE,DUSTIN,K  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MUSGROVE,DUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A1  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: THORPE

DATE/TIME OF TEST: 07/23/2018 22:01:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:55 |
| Air Blank          | 0.000 | 21:56 |
| Control Test       | 0.081 | 21:56 |
| Air Blank          | 0.000 | 21:57 |
| Subject Sample # 1 | 0.000 | 21:57 |
| Breath Volume      | 2.171 | 21:57 |
| Air Blank          | 0.000 | 21:58 |
| Air Blank          | 0.000 | 22:00 |
| Subject Sample # 2 | 0.000 | 22:00 |
| Breath Volume      | 2.164 | 22:00 |
| Air Blank          | 0.000 | 22:01 |
| Control Test       | 0.081 | 22:01 |
| Air Blank          | 0.000 | 22:01 |
| Diagnostics Check  | OK    | 22:01 |

---

Data Download Date/Time: 07/30/2018 01:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWINDLE, CHARLES, M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SWINDLE, CHARLES  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 390314  
DRY GAS EXPIRY DATE: 04/19/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LACHUT

DATE/TIME OF TEST: 07/05/2018 00:07:32

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 00:06 |            |
| Air Blank         |       | 00:07 | RFI Detect |
| Air Blank         | 0.000 | 00:07 |            |

---

Data Download Date/Time: 07/25/2018 11:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWINDLE, CHARLES, M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SWINDLE, CHARLES  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 390314  
DRY GAS EXPIRY DATE: 04/19/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LACHUT

DATE/TIME OF TEST: 07/05/2018 00:15:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:09 |
| Air Blank          | 0.000 | 00:09 |
| Control Test       | 0.076 | 00:09 |
| Air Blank          | 0.000 | 00:10 |
| Subject Sample # 1 | 0.099 | 00:10 |
| Breath Volume      | 2.156 | 00:10 |
| Air Blank          | 0.000 | 00:11 |
| Air Blank          | 0.000 | 00:13 |
| Subject Sample # 2 | 0.098 | 00:13 |
| Breath Volume      | 2.062 | 00:13 |
| Air Blank          | 0.000 | 00:14 |
| Control Test       | 0.076 | 00:14 |
| Air Blank          | 0.000 | 00:15 |
| Diagnostics Check  | OK    | 00:15 |

---

Data Download Date/Time: 07/25/2018 11:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: VOLLERTSEN,CRAIG  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 390314  
DRY GAS EXPIRY DATE: 04/19/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: KLING

DATE/TIME OF TEST: 07/05/2018 02:32:42

---

|                    |       |       |                 |
|--------------------|-------|-------|-----------------|
| Diagnostics Check  | OK    | 02:27 |                 |
| Air Blank          | 0.000 | 02:27 |                 |
| Control Test       | 0.076 | 02:27 |                 |
| Air Blank          | 0.000 | 02:28 |                 |
| Subject Sample # 1 | 0.223 | 02:29 |                 |
| Breath Volume      | 2.894 | 02:29 |                 |
| Air Blank          | 0.000 | 02:29 |                 |
| Air Blank          | 0.000 | 02:31 |                 |
| Subject Sample # 2 | 0.217 | 02:32 |                 |
| Breath Volume      | 3.109 | 02:32 |                 |
| Air Blank          |       | 02:32 | Improper Sample |
| Air Blank          | 0.000 | 02:32 |                 |

---

Data Download Date/Time: 07/25/2018 11:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: VOLLERTSEN,CRAIG  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 390314  
DRY GAS EXPIRY DATE: 04/19/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: KLING

DATE/TIME OF TEST: 07/05/2018 02:40:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:34 |
| Air Blank          | 0.000 | 02:34 |
| Control Test       | 0.076 | 02:35 |
| Air Blank          | 0.000 | 02:35 |
| Subject Sample # 1 | 0.221 | 02:36 |
| Breath Volume      | 3.187 | 02:36 |
| Air Blank          | 0.000 | 02:36 |
| Air Blank          | 0.000 | 02:38 |
| Subject Sample # 2 | 0.216 | 02:39 |
| Breath Volume      | 1.703 | 02:39 |
| Air Blank          | 0.000 | 02:39 |
| Control Test       | 0.076 | 02:40 |
| Air Blank          | 0.000 | 02:40 |
| Diagnostics Check  | OK    | 02:40 |

---

Data Download Date/Time: 07/25/2018 11:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWINDLE, CHARLES, M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SWINDLE, CHARLES  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 390314  
DRY GAS EXPIRY DATE: 04/19/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: DODGE 2

DATE/TIME OF TEST: 07/13/2018 23:44:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:38 |
| Air Blank          | 0.000 | 23:39 |
| Control Test       | 0.076 | 23:39 |
| Air Blank          | 0.000 | 23:40 |
| Subject Sample # 1 | 0.090 | 23:40 |
| Breath Volume      | 2.703 | 23:40 |
| Air Blank          | 0.000 | 23:41 |
| Air Blank          | 0.000 | 23:43 |
| Subject Sample # 2 | 0.088 | 23:43 |
| Breath Volume      | 3.621 | 23:43 |
| Air Blank          | 0.000 | 23:43 |
| Control Test       | 0.075 | 23:44 |
| Air Blank          | 0.000 | 23:44 |
| Diagnostics Check  | OK    | 23:44 |

---

Data Download Date/Time: 07/25/2018 11:32



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMYRNIOS,GUS,T  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SMYRNIOS,GUS  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 390314  
DRY GAS EXPIRY DATE: 04/19/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: S415424001490

DATE/TIME OF TEST: 07/22/2018 05:37:22

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 05:31 |
| Air Blank         |     | 0.000 | 05:31 |
| Control Test      |     | 0.077 | 05:31 |
| Air Blank         |     | 0.000 | 05:32 |
| Subject Sample    | # 1 | 0.156 | 05:32 |
| Breath Volume     |     | 2.437 | 05:32 |
| Air Blank         |     | 0.000 | 05:33 |
| Air Blank         |     | 0.000 | 05:35 |
| Subject Sample    | # 2 | 0.165 | 05:35 |
| Breath Volume     |     | 2.980 | 05:35 |
| Air Blank         |     | 0.000 | 05:36 |
| Control Test      |     | 0.077 | 05:36 |
| Air Blank         |     | 0.000 | 05:37 |
| Diagnostics Check |     | OK    | 05:37 |

---

Data Download Date/Time: 07/25/2018 11:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEWIS,JARROD,W  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: MCCAIG,AMBER  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A3  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: NIESSEN

DATE/TIME OF TEST: 07/01/2018 22:19:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:13 |
| Air Blank          | 0.000 | 22:13 |
| Control Test       | 0.079 | 22:14 |
| Air Blank          | 0.000 | 22:14 |
| Subject Sample # 1 | 0.186 | 22:15 |
| Breath Volume      | 1.277 | 22:15 |
| Air Blank          | 0.000 | 22:15 |
| Air Blank          | 0.000 | 22:17 |
| Subject Sample # 2 | 0.187 | 22:18 |
| Breath Volume      | 1.398 | 22:18 |
| Air Blank          | 0.000 | 22:18 |
| Control Test       | 0.078 | 22:19 |
| Air Blank          | 0.000 | 22:19 |
| Diagnostics Check  | OK    | 22:19 |

---

Data Download Date/Time: 07/20/2018 14:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLIS,TERENCE,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: BOATROS,P  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 07/20/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 07/24/2018 23:33:23

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:30 |                      |
| Air Blank         | 0.000 | 23:30 |                      |
| Control Test      | 0.079 | 23:31 |                      |
| Air Blank         | 0.000 | 23:31 |                      |
| Subject Sample    | # 1   | 23:31 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:31 |                      |
| Air Blank         | 0.000 | 23:32 |                      |
| Control Test      | 0.078 | 23:32 |                      |
| Air Blank         | 0.000 | 23:33 |                      |
| Diagnostics Check | OK    | 23:33 |                      |

---

Data Download Date/Time: 08/28/2018 21:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONG,ERIC,W  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: WILSON,JUSTIN  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 07/20/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 07/26/2018 23:02:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:54 |
| Air Blank          | 0.000 | 22:55 |
| Control Test       | 0.079 | 22:55 |
| Air Blank          | 0.000 | 22:55 |
| Subject Sample # 1 | 0.183 | 22:56 |
| Breath Volume      | 1.574 | 22:56 |
| Air Blank          | 0.000 | 22:57 |
| Air Blank          | 0.000 | 22:59 |
| Subject Sample # 2 | 0.185 | 23:00 |
| Breath Volume      | 1.808 | 23:00 |
| Air Blank          | 0.000 | 23:01 |
| Control Test       | 0.077 | 23:01 |
| Air Blank          | 0.000 | 23:02 |
| Diagnostics Check  | OK    | 23:02 |

---

Data Download Date/Time: 08/28/2018 21:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLIS,TERENCE,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: BRITT,R  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 07/20/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: DELCY

DATE/TIME OF TEST: 07/29/2018 22:52:37

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 22:40 |                |
| Air Blank         |     | 0.000 | 22:41 |                |
| Control Test      |     | 0.079 | 22:41 |                |
| Air Blank         |     | 0.000 | 22:42 |                |
| Subject Sample    | # 1 | 0.238 | 22:45 | Volume Not Met |
| Breath Volume     |     | 0.367 | 22:45 |                |
| Air Blank         |     | 0.000 | 22:46 |                |
| Air Blank         |     | 0.000 | 22:47 |                |
| Subject Sample    | # 2 | 0.258 | 22:50 | Volume Not Met |
| Breath Volume     |     | 0.300 | 22:50 |                |
| Air Blank         |     | 0.000 | 22:51 |                |
| Control Test      |     | 0.077 | 22:51 |                |
| Air Blank         |     | 0.000 | 22:52 |                |
| Diagnostics Check |     | OK    | 22:52 |                |

---

Data Download Date/Time: 08/28/2018 21:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONG,ERIC,W  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: MANDERS,MARC  
ARRESTING OFFICER AGENCY: Wildwood PD  
LAST AGENCY INSPECTION: 07/20/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: GIVENS

DATE/TIME OF TEST: 07/30/2018 19:44:24

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:39 |                      |
| Air Blank         | 0.000 | 19:40 |                      |
| Control Test      | 0.079 | 19:40 |                      |
| Air Blank         | 0.000 | 19:41 |                      |
| Subject Sample    | # 1   | 19:43 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:43 |                      |
| Air Blank         | 0.000 | 19:43 |                      |
| Control Test      | 0.076 | 19:43 |                      |
| Air Blank         | 0.000 | 19:44 |                      |
| Diagnostics Check | OK    | 19:44 |                      |

---

Data Download Date/Time: 08/28/2018 21:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIVERS,DAWN,T  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: RAFFERTY,CHRIS  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 07/20/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: ZELENT

DATE/TIME OF TEST: 07/31/2018 22:44:38

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:41 |                      |
| Air Blank         | 0.000 | 22:42 |                      |
| Control Test      | 0.079 | 22:42 |                      |
| Air Blank         | 0.000 | 22:43 |                      |
| Subject Sample    | # 1   | 22:43 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:43 |                      |
| Air Blank         | 0.000 | 22:43 |                      |
| Control Test      | 0.079 | 22:44 |                      |
| Air Blank         | 0.000 | 22:44 |                      |
| Diagnostics Check | OK    | 22:44 |                      |

---

Data Download Date/Time: 08/28/2018 21:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLOW,VINESSA,  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: CIULLA,J  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A3  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: DOGGETT

DATE/TIME OF TEST: 07/02/2018 17:10:59

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 17:05 |            |
| Air Blank         | 0.000     | 17:05 |            |
| Control Test      | 0.077     | 17:06 |            |
| Air Blank         | 0.000     | 17:06 |            |
| Subject Sample    | # 1 0.360 | 17:09 |            |
| Breath Volume     | 1.335     | 17:09 |            |
| Air Blank         |           | 17:10 | Purge Fail |
| Air Blank         | 0.000     | 17:10 |            |

---

Data Download Date/Time: 07/20/2018 14:01



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLOW,VINESSA,  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: CIULLA,J  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A3  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: DOGGETT

DATE/TIME OF TEST: 07/02/2018 17:14:11

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 17:12 |                           |
| Air Blank         | 0.000 | 17:13 |                           |
| Control Test      | 0.062 | 17:13 | Control Outside Tolerance |
| Air Blank         | 0.000 | 17:14 |                           |

---

Data Download Date/Time: 07/20/2018 14:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STEPHENS,JOSEPH,C  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: WILSON,JUSTIN  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A3  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 07/07/2018 22:55:07

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:52 |                      |
| Air Blank         | 0.000 | 22:52 |                      |
| Control Test      | 0.076 | 22:52 |                      |
| Air Blank         | 0.000 | 22:53 |                      |
| Subject Sample    | # 1   | 22:53 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:53 |                      |
| Air Blank         | 0.000 | 22:54 |                      |
| Control Test      | 0.076 | 22:54 |                      |
| Air Blank         | 0.000 | 22:55 |                      |
| Diagnostics Check | OK    | 22:55 |                      |

---

Data Download Date/Time: 07/20/2018 14:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: BOUTROS,PHIL  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A3  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 17:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: MCPARLIN

DATE/TIME OF TEST: 07/13/2018 19:10:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:04 |
| Air Blank          | 0.000 | 19:04 |
| Control Test       | 0.077 | 19:05 |
| Air Blank          | 0.000 | 19:05 |
| Subject Sample # 1 | 0.000 | 19:06 |
| Breath Volume      | 2.292 | 19:06 |
| Air Blank          | 0.000 | 19:06 |
| Air Blank          | 0.000 | 19:08 |
| Subject Sample # 2 | 0.000 | 19:09 |
| Breath Volume      | 3.238 | 19:09 |
| Air Blank          | 0.000 | 19:09 |
| Control Test       | 0.077 | 19:10 |
| Air Blank          | 0.000 | 19:10 |
| Diagnostics Check  | OK    | 19:10 |

---

Data Download Date/Time: 07/20/2018 14:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLIS,TERENCE,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: WESCH,C  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A3  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: CALLAHAN JR

DATE/TIME OF TEST: 07/14/2018 03:23:03

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:20 |                      |
| Air Blank         | 0.000 | 03:20 |                      |
| Control Test      | 0.076 | 03:20 |                      |
| Air Blank         | 0.000 | 03:21 |                      |
| Subject Sample    | # 1   | 03:21 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:21 |                      |
| Air Blank         | 0.000 | 03:22 |                      |
| Control Test      | 0.076 | 03:22 |                      |
| Air Blank         | 0.000 | 03:22 |                      |
| Diagnostics Check | OK    | 03:23 |                      |

---

Data Download Date/Time: 07/20/2018 14:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLIS,TERENCE,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: RAMOS,JOSEPH  
ARRESTING OFFICER AGENCY: Wildwood PD  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A3  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: GUY

DATE/TIME OF TEST: 07/15/2018 00:36:07

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 00:26 |
| Air Blank         |     | 0.000 | 00:26 |
| Control Test      |     | 0.077 | 00:26 |
| Air Blank         |     | 0.000 | 00:27 |
| Subject Sample    | # 1 | 0.181 | 00:30 |
| Breath Volume     |     | 1.558 | 00:30 |
| Air Blank         |     | 0.000 | 00:31 |
| Air Blank         |     | 0.000 | 00:33 |
| Subject Sample    | # 2 | 0.176 | 00:34 |
| Breath Volume     |     | 1.640 | 00:34 |
| Air Blank         |     | 0.000 | 00:35 |
| Control Test      |     | 0.075 | 00:35 |
| Air Blank         |     | 0.000 | 00:36 |
| Diagnostics Check |     | OK    | 00:36 |

---

Data Download Date/Time: 07/20/2018 14:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FORINASH,GARY,B  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: RAFFERTY,CHRIS  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A3  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: FORREST

DATE/TIME OF TEST: 07/16/2018 19:12:01

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:09 |                      |
| Air Blank         | 0.000 | 19:09 |                      |
| Control Test      | 0.077 | 19:09 |                      |
| Air Blank         | 0.000 | 19:10 |                      |
| Subject Sample    | # 1   | 19:10 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:10 |                      |
| Air Blank         | 0.000 | 19:11 |                      |
| Control Test      | 0.076 | 19:11 |                      |
| Air Blank         | 0.000 | 19:11 |                      |
| Diagnostics Check | OK    | 19:12 |                      |

---

Data Download Date/Time: 07/20/2018 14:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FORINASH,GARY,B  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: WILSON,JUSTIN  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A3  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: MUNOZPRECIADO

DATE/TIME OF TEST: 07/16/2018 19:47:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:36 |
| Air Blank          | 0.000 | 19:37 |
| Control Test       | 0.076 | 19:37 |
| Air Blank          | 0.000 | 19:38 |
| Subject Sample # 1 | 0.179 | 19:39 |
| Breath Volume      | 2.808 | 19:39 |
| Air Blank          | 0.000 | 19:40 |
| Air Blank          | 0.000 | 19:41 |
| Subject Sample # 2 | 0.158 | 19:42 |
| Breath Volume      | 3.066 | 19:42 |
| Air Blank          | 0.000 | 19:43 |
| Air Blank          | 0.000 | 19:44 |
| Subject Sample # 3 | 0.161 | 19:45 |
| Breath Volume      | 3.089 | 19:45 |
| Air Blank          | 0.000 | 19:46 |
| Control Test       | 0.075 | 19:46 |
| Air Blank          | 0.000 | 19:47 |
| Diagnostics Check  | OK    | 19:47 |

---

Data Download Date/Time: 07/20/2018 14:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STEPHENS,JOSEPH,C  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: RAFFERTY,CHRIS  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A3  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: LACY

DATE/TIME OF TEST: 07/17/2018 01:31:47

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:28 |                      |
| Air Blank         | 0.000 | 01:29 |                      |
| Control Test      | 0.076 | 01:29 |                      |
| Air Blank         | 0.000 | 01:30 |                      |
| Subject Sample    | # 1   | 01:30 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:30 |                      |
| Air Blank         | 0.000 | 01:30 |                      |
| Control Test      | 0.077 | 01:31 |                      |
| Air Blank         | 0.000 | 01:31 |                      |
| Diagnostics Check | OK    | 01:31 |                      |

---

Data Download Date/Time: 07/20/2018 14:01



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORRIS,MARSHA,K  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: BOUTROS,P  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20016080A3  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 12:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:42

SUBJECT'S LAST NAME: KELLY-LAWSON

DATE/TIME OF TEST: 07/19/2018 13:17:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:10 |
| Air Blank          | 0.000 | 13:11 |
| Control Test       | 0.077 | 13:11 |
| Air Blank          | 0.000 | 13:11 |
| Subject Sample # 1 | 0.137 | 13:13 |
| Breath Volume      | 1.433 | 13:13 |
| Air Blank          | 0.000 | 13:14 |
| Air Blank          | 0.000 | 13:15 |
| Subject Sample # 2 | 0.138 | 13:16 |
| Breath Volume      | 1.492 | 13:16 |
| Air Blank          | 0.000 | 13:16 |
| Control Test       | 0.076 | 13:17 |
| Air Blank          | 0.000 | 13:17 |
| Diagnostics Check  | OK    | 13:17 |

---

Data Download Date/Time: 07/20/2018 14:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,BRONSON  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BRITTON

DATE/TIME OF TEST: 07/07/2018 03:28:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:21 |
| Air Blank          | 0.000 | 03:22 |
| Control Test       | 0.076 | 03:22 |
| Air Blank          | 0.000 | 03:23 |
| Subject Sample # 1 | 0.153 | 03:23 |
| Breath Volume      | 2.214 | 03:23 |
| Air Blank          | 0.000 | 03:24 |
| Air Blank          | 0.000 | 03:26 |
| Subject Sample # 2 | 0.154 | 03:26 |
| Breath Volume      | 1.847 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Control Test       | 0.077 | 03:27 |
| Air Blank          | 0.000 | 03:28 |
| Diagnostics Check  | OK    | 03:28 |

---

Data Download Date/Time: 08/15/2018 23:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,P  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,BRONSON  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 07/08/2018 00:01:40

---

|                   |           |       |                           |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK        | 23:55 |                           |
| Air Blank         | 0.000     | 23:55 |                           |
| Control Test      | 0.076     | 23:56 |                           |
| Air Blank         | 0.000     | 23:56 |                           |
| Subject Sample    | # 2 0.279 | 23:57 |                           |
| Breath Volume     | 1.644     | 23:57 |                           |
| Air Blank         | 0.000     | 23:57 |                           |
| Air Blank         | 0.000     | 23:59 |                           |
| Subject Sample    | # 1 0.278 | 00:00 |                           |
| Breath Volume     | 1.539     | 00:00 |                           |
| Air Blank         | 0.000     | 00:00 |                           |
| Control Test      | 0.074     | 00:01 | Control Outside Tolerance |
| Air Blank         | 0.000     | 00:01 |                           |

---

Data Download Date/Time: 08/15/2018 23:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,P  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,BRONSON  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 07/08/2018 00:11:34

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 00:05 |
| Air Blank         |     | 0.000 | 00:05 |
| Control Test      |     | 0.076 | 00:05 |
| Air Blank         |     | 0.000 | 00:06 |
| Subject Sample    | # 1 | 0.284 | 00:06 |
| Breath Volume     |     | 1.613 | 00:06 |
| Air Blank         |     | 0.000 | 00:07 |
| Air Blank         |     | 0.000 | 00:09 |
| Subject Sample    | # 2 | 0.277 | 00:09 |
| Breath Volume     |     | 2.046 | 00:09 |
| Air Blank         |     | 0.000 | 00:10 |
| Control Test      |     | 0.075 | 00:10 |
| Air Blank         |     | 0.000 | 00:11 |
| Diagnostics Check |     | OK    | 00:11 |

---

Data Download Date/Time: 08/15/2018 23:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,BRONSON  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: CORTIJO BURGOS

DATE/TIME OF TEST: 07/08/2018 02:56:40

---

|                    |       |       |                           |
|--------------------|-------|-------|---------------------------|
| Diagnostics Check  | OK    | 02:50 |                           |
| Air Blank          | 0.000 | 02:50 |                           |
| Control Test       | 0.076 | 02:51 |                           |
| Air Blank          | 0.000 | 02:51 |                           |
| Subject Sample # 1 | 0.164 | 02:52 |                           |
| Breath Volume      | 2.128 | 02:52 |                           |
| Air Blank          | 0.000 | 02:52 |                           |
| Air Blank          | 0.000 | 02:54 |                           |
| Subject Sample # 2 | 0.149 | 02:55 |                           |
| Breath Volume      | 1.121 | 02:55 |                           |
| Air Blank          | 0.000 | 02:55 |                           |
| Control Test       | 0.073 | 02:56 | Control Outside Tolerance |
| Air Blank          | 0.000 | 02:56 |                           |

---

Data Download Date/Time: 08/15/2018 23:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,BRONSON  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: CORTIJO BURGOS

DATE/TIME OF TEST: 07/08/2018 03:06:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:59 |
| Air Blank          | 0.000 | 03:00 |
| Control Test       | 0.076 | 03:00 |
| Air Blank          | 0.000 | 03:00 |
| Subject Sample # 1 | 0.160 | 03:01 |
| Breath Volume      | 2.003 | 03:01 |
| Air Blank          | 0.000 | 03:02 |
| Air Blank          | 0.000 | 03:03 |
| Subject Sample # 2 | 0.148 | 03:04 |
| Breath Volume      | 2.015 | 03:04 |
| Air Blank          | 0.000 | 03:05 |
| Control Test       | 0.075 | 03:05 |
| Air Blank          | 0.000 | 03:05 |
| Diagnostics Check  | OK    | 03:06 |

---

Data Download Date/Time: 08/15/2018 23:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANDI  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 07/11/2018 23:26:01

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:23 |                      |
| Air Blank         | 0.000 | 23:23 |                      |
| Control Test      | 0.075 | 23:23 |                      |
| Air Blank         | 0.000 | 23:24 |                      |
| Subject Sample    | # 1   | 23:24 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:24 |                      |
| Air Blank         | 0.000 | 23:25 |                      |
| Control Test      | 0.075 | 23:25 |                      |
| Air Blank         | 0.000 | 23:25 |                      |
| Diagnostics Check | OK    | 23:26 |                      |

---

Data Download Date/Time: 08/15/2018 23:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KETCHUM,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: KINNEY,C  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A2  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: TODORAN

DATE/TIME OF TEST: 07/01/2018 05:41:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:34 |
| Air Blank          | 0.000 | 05:35 |
| Control Test       | 0.080 | 05:35 |
| Air Blank          | 0.000 | 05:35 |
| Subject Sample # 1 | 0.166 | 05:36 |
| Breath Volume      | 3.863 | 05:36 |
| Air Blank          | 0.000 | 05:37 |
| Air Blank          | 0.000 | 05:39 |
| Subject Sample # 2 | 0.156 | 05:40 |
| Breath Volume      | 2.816 | 05:40 |
| Air Blank          | 0.000 | 05:40 |
| Control Test       | 0.079 | 05:40 |
| Air Blank          | 0.000 | 05:41 |
| Diagnostics Check  | OK    | 05:41 |

---

Data Download Date/Time: 07/13/2018 03:59



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FOSTER,ANDREW,D  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: JOHNSON,STEVEN  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A2  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 07/02/2018 23:23:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:17 |
| Air Blank          | 0.000 | 23:18 |
| Control Test       | 0.080 | 23:18 |
| Air Blank          | 0.000 | 23:19 |
| Subject Sample # 1 | 0.225 | 23:19 |
| Breath Volume      | 2.277 | 23:19 |
| Air Blank          | 0.000 | 23:20 |
| Air Blank          | 0.000 | 23:22 |
| Subject Sample # 2 | 0.226 | 23:22 |
| Breath Volume      | 1.601 | 23:22 |
| Air Blank          | 0.000 | 23:23 |
| Control Test       | 0.080 | 23:23 |
| Air Blank          | 0.000 | 23:23 |
| Diagnostics Check  | OK    | 23:23 |

---

Data Download Date/Time: 07/13/2018 03:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KETCHUM,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: SCHEER,JOHN  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A2  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: ALVA FIGUEROA

DATE/TIME OF TEST: 07/14/2018 23:59:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:51 |
| Air Blank          | 0.000 | 23:51 |
| Control Test       | 0.080 | 23:52 |
| Air Blank          | 0.000 | 23:52 |
| Subject Sample # 1 | 0.122 | 23:54 |
| Breath Volume      | 1.710 | 23:54 |
| Air Blank          | 0.000 | 23:55 |
| Air Blank          | 0.000 | 23:57 |
| Subject Sample # 2 | 0.132 | 23:58 |
| Breath Volume      | 1.183 | 23:58 |
| Air Blank          | 0.000 | 23:58 |
| Control Test       | 0.080 | 23:59 |
| Air Blank          | 0.000 | 23:59 |
| Diagnostics Check  | OK    | 23:59 |

---

Data Download Date/Time: 08/15/2018 03:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOUROUMOUSIS,GEORGE,T  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: COOK,JOSH  
ARRESTING OFFICER AGENCY: Mount Dora PD  
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 583457  
DRY GAS EXPIRY DATE: 10/20/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: NEWBERRY

DATE/TIME OF TEST: 07/03/2018 19:32:17

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:29 |                      |
| Air Blank         | 0.000 | 19:29 |                      |
| Control Test      | 0.077 | 19:30 |                      |
| Air Blank         | 0.000 | 19:30 |                      |
| Subject Sample    | # 1   | 19:30 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:30 |                      |
| Air Blank         | 0.000 | 19:31 |                      |
| Control Test      | 0.078 | 19:31 |                      |
| Air Blank         | 0.000 | 19:32 |                      |
| Diagnostics Check | OK    | 19:32 |                      |

---

Data Download Date/Time: 07/31/2018 21:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: MORALES,PABLO  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 07/04/2018 22:21:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:15 |
| Air Blank          | 0.000 | 22:15 |
| Control Test       | 0.079 | 22:15 |
| Air Blank          | 0.000 | 22:16 |
| Subject Sample # 1 | 0.199 | 22:17 |
| Breath Volume      | 2.773 | 22:17 |
| Air Blank          | 0.000 | 22:17 |
| Air Blank          | 0.000 | 22:19 |
| Subject Sample # 2 | 0.203 | 22:20 |
| Breath Volume      | 2.976 | 22:20 |
| Air Blank          | 0.000 | 22:20 |
| Control Test       | 0.078 | 22:21 |
| Air Blank          | 0.000 | 22:21 |
| Diagnostics Check  | OK    | 22:21 |

---

Data Download Date/Time: 07/21/2018 15:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,GARY,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MILLER,GARY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 07/05/2018 00:17:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:09 |
| Air Blank          | 0.000 | 00:09 |
| Control Test       | 0.079 | 00:09 |
| Air Blank          | 0.000 | 00:10 |
| Subject Sample # 1 | 0.000 | 00:10 |
| Breath Volume      | 1.167 | 00:10 |
| Air Blank          | 0.000 | 00:11 |
| Air Blank          | 0.000 | 00:13 |
| Subject Sample # 2 | 0.000 | 00:15 |
| Breath Volume      | 1.441 | 00:15 |
| Air Blank          | 0.000 | 00:16 |
| Control Test       | 0.079 | 00:16 |
| Air Blank          | 0.000 | 00:17 |
| Diagnostics Check  | OK    | 00:17 |

---

Data Download Date/Time: 07/21/2018 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EARNEST, WILLIAM, M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BALL, CODY  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 07/05/2018 02:03:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:56 |
| Air Blank          | 0.000 | 01:56 |
| Control Test       | 0.078 | 01:57 |
| Air Blank          | 0.000 | 01:57 |
| Subject Sample # 1 | 0.123 | 01:58 |
| Breath Volume      | 1.761 | 01:58 |
| Air Blank          | 0.000 | 01:58 |
| Air Blank          | 0.000 | 02:00 |
| Subject Sample # 2 | 0.123 | 02:01 |
| Breath Volume      | 1.542 | 02:01 |
| Air Blank          | 0.000 | 02:02 |
| Control Test       | 0.079 | 02:02 |
| Air Blank          | 0.000 | 02:03 |
| Diagnostics Check  | OK    | 02:03 |

---

Data Download Date/Time: 07/21/2018 15:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOLLAZZO,C,  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: SOLLAZZO,C  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 07/05/2018 22:23:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:13 |
| Air Blank          | 0.000 | 22:13 |
| Control Test       | 0.079 | 22:13 |
| Air Blank          | 0.000 | 22:14 |
| Subject Sample # 1 | 0.194 | 22:17 |
| Breath Volume      | 1.304 | 22:17 |
| Air Blank          | 0.000 | 22:18 |
| Air Blank          | 0.000 | 22:19 |
| Subject Sample # 2 | 0.199 | 22:21 |
| Breath Volume      | 1.500 | 22:21 |
| Air Blank          | 0.000 | 22:22 |
| Control Test       | 0.077 | 22:22 |
| Air Blank          | 0.000 | 22:23 |
| Diagnostics Check  | OK    | 22:23 |

---

Data Download Date/Time: 07/21/2018 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALESE, ANTHONY, W  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: PALESE, ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 07/06/2018 02:45:52

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:42 |                      |
| Air Blank         | 0.000 | 02:42 |                      |
| Control Test      | 0.079 | 02:43 |                      |
| Air Blank         | 0.000 | 02:43 |                      |
| Subject Sample    | # 1   | 02:44 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:44 |                      |
| Air Blank         | 0.000 | 02:44 |                      |
| Control Test      | 0.079 | 02:45 |                      |
| Air Blank         | 0.000 | 02:45 |                      |
| Diagnostics Check | OK    | 02:45 |                      |

---

Data Download Date/Time: 07/21/2018 15:44



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTIGLIONE,S,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: CASTIGLIONE,S  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 07/06/2018 07:47:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:41 |
| Air Blank          | 0.000 | 07:41 |
| Control Test       | 0.079 | 07:42 |
| Air Blank          | 0.000 | 07:42 |
| Subject Sample # 1 | 0.171 | 07:43 |
| Breath Volume      | 1.585 | 07:43 |
| Air Blank          | 0.000 | 07:43 |
| Air Blank          | 0.000 | 07:45 |
| Subject Sample # 2 | 0.160 | 07:45 |
| Breath Volume      | 1.570 | 07:45 |
| Air Blank          | 0.000 | 07:46 |
| Control Test       | 0.078 | 07:46 |
| Air Blank          | 0.000 | 07:47 |
| Diagnostics Check  | OK    | 07:47 |

---

Data Download Date/Time: 07/21/2018 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAHN,P,W  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: HAHN,P  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 07/06/2018 23:24:41

---

|                   |     |       |       |                      |
|-------------------|-----|-------|-------|----------------------|
| Diagnostics Check |     | OK    | 23:14 |                      |
| Air Blank         |     | 0.000 | 23:14 |                      |
| Control Test      |     | 0.080 | 23:14 |                      |
| Air Blank         |     | 0.000 | 23:15 |                      |
| Subject Sample    | # 1 | 0.030 | 23:18 | Volume Not Met       |
| Breath Volume     |     | 0.070 | 23:18 |                      |
| Air Blank         |     | 0.000 | 23:19 |                      |
| Air Blank         |     | 0.000 | 23:20 |                      |
| Subject Sample    | # 2 |       | 23:23 | Subject Test Refused |
| Breath Volume     |     | 0.054 | 23:23 |                      |
| Air Blank         |     | 0.000 | 23:23 |                      |
| Control Test      |     | 0.078 | 23:24 |                      |
| Air Blank         |     | 0.000 | 23:24 |                      |
| Diagnostics Check |     | OK    | 23:24 |                      |

---

Data Download Date/Time: 07/21/2018 15:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERNANDEZ,ALAIN,  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: FERNANDEZ,ALAIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: ADKINS

DATE/TIME OF TEST: 07/07/2018 04:56:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:50 |
| Air Blank          | 0.000 | 04:50 |
| Control Test       | 0.079 | 04:51 |
| Air Blank          | 0.000 | 04:51 |
| Subject Sample # 1 | 0.085 | 04:52 |
| Breath Volume      | 3.554 | 04:52 |
| Air Blank          | 0.000 | 04:52 |
| Air Blank          | 0.000 | 04:54 |
| Subject Sample # 2 | 0.085 | 04:55 |
| Breath Volume      | 2.929 | 04:55 |
| Air Blank          | 0.000 | 04:55 |
| Control Test       | 0.080 | 04:56 |
| Air Blank          | 0.000 | 04:56 |
| Diagnostics Check  | OK    | 04:56 |

---

Data Download Date/Time: 07/21/2018 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EARNEST, WILLIAM, M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: EARNEST, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 05:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 07/07/2018 06:09:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:02 |
| Air Blank          | 0.000 | 06:03 |
| Control Test       | 0.080 | 06:03 |
| Air Blank          | 0.000 | 06:04 |
| Subject Sample # 1 | 0.114 | 06:04 |
| Breath Volume      | 3.476 | 06:04 |
| Air Blank          | 0.000 | 06:05 |
| Air Blank          | 0.000 | 06:07 |
| Subject Sample # 2 | 0.108 | 06:07 |
| Breath Volume      | 3.503 | 06:07 |
| Air Blank          | 0.000 | 06:08 |
| Control Test       | 0.079 | 06:08 |
| Air Blank          | 0.000 | 06:09 |
| Diagnostics Check  | OK    | 06:09 |

---

Data Download Date/Time: 07/21/2018 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,GARY,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MILLER,GARY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: BATALLAS MIELES

DATE/TIME OF TEST: 07/08/2018 00:34:14

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 00:21 |
| Air Blank         |     | 0.000 | 00:22 |
| Control Test      |     | 0.079 | 00:22 |
| Air Blank         |     | 0.000 | 00:22 |
| Subject Sample    | # 1 | 0.294 | 00:23 |
| Breath Volume     |     | 1.218 | 00:23 |
| Air Blank         |     | 0.000 | 00:23 |
| Air Blank         |     | 0.000 | 00:25 |
| Subject Sample    | # 2 | 0.240 | 00:28 |
| Breath Volume     |     | 2.519 | 00:28 |
| Air Blank         |     | 0.000 | 00:29 |
| Air Blank         |     | 0.000 | 00:31 |
| Subject Sample    | # 3 | 0.249 | 00:32 |
| Breath Volume     |     | 1.300 | 00:32 |
| Air Blank         |     | 0.000 | 00:33 |
| Control Test      |     | 0.077 | 00:33 |
| Air Blank         |     | 0.000 | 00:34 |
| Diagnostics Check |     | OK    | 00:34 |

---

Data Download Date/Time: 07/21/2018 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EARNEST, WILLIAM, M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: EARNEST, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: NOTLEY

DATE/TIME OF TEST: 07/09/2018 02:51:32

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:48 |                      |
| Air Blank         | 0.000 | 02:48 |                      |
| Control Test      | 0.079 | 02:48 |                      |
| Air Blank         | 0.000 | 02:49 |                      |
| Subject Sample    | # 1   | 02:50 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:50 |                      |
| Air Blank         | 0.000 | 02:50 |                      |
| Control Test      | 0.079 | 02:50 |                      |
| Air Blank         | 0.000 | 02:51 |                      |
| Diagnostics Check | OK    | 02:51 |                      |

---

Data Download Date/Time: 07/21/2018 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,MICHAEL,T  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: THOMAS,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 07/11/2018 19:29:07

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:26 |                      |
| Air Blank         | 0.000 | 19:26 |                      |
| Control Test      | 0.081 | 19:27 |                      |
| Air Blank         | 0.000 | 19:27 |                      |
| Subject Sample    | # 1   | 19:27 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:27 |                      |
| Air Blank         | 0.000 | 19:28 |                      |
| Control Test      | 0.080 | 19:28 |                      |
| Air Blank         | 0.000 | 19:29 |                      |
| Diagnostics Check | OK    | 19:29 |                      |

---

Data Download Date/Time: 07/21/2018 15:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERNANDEZ,ALAIN,  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: FERNANDEZ,ALAIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 07/11/2018 20:13:43

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:05 |                      |
| Air Blank         | 0.000 | 20:05 |                      |
| Control Test      | 0.080 | 20:06 |                      |
| Air Blank         | 0.000 | 20:06 |                      |
| Subject Sample    | # 1   | 20:09 | No Sample Provided   |
| Breath Volume     | 0.000 | 20:09 |                      |
| Air Blank         | 0.000 | 20:10 |                      |
| Air Blank         | 0.000 | 20:12 |                      |
| Subject Sample    | # 2   | 20:12 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:12 |                      |
| Air Blank         | 0.000 | 20:12 |                      |
| Control Test      | 0.079 | 20:13 |                      |
| Air Blank         | 0.000 | 20:13 |                      |
| Diagnostics Check | OK    | 20:13 |                      |

---

Data Download Date/Time: 07/21/2018 15:44



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELRUSSO,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DELRUSSO,CHRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/14/2018 05:28:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:21 |
| Air Blank          | 0.000 | 05:21 |
| Control Test       | 0.081 | 05:22 |
| Air Blank          | 0.000 | 05:22 |
| Subject Sample # 1 | 0.152 | 05:23 |
| Breath Volume      | 1.230 | 05:23 |
| Air Blank          | 0.000 | 05:24 |
| Air Blank          | 0.000 | 05:26 |
| Subject Sample # 2 | 0.152 | 05:26 |
| Breath Volume      | 1.355 | 05:26 |
| Air Blank          | 0.000 | 05:27 |
| Control Test       | 0.079 | 05:27 |
| Air Blank          | 0.000 | 05:28 |
| Diagnostics Check  | OK    | 05:28 |

---

Data Download Date/Time: 07/21/2018 15:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTIGLIONE,S,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DELRUSSO,C  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: MONDRAGON LARA

DATE/TIME OF TEST: 07/14/2018 05:54:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:44 |
| Air Blank          | 0.000 | 05:45 |
| Control Test       | 0.079 | 05:45 |
| Air Blank          | 0.000 | 05:46 |
| Subject Sample # 1 | 0.152 | 05:47 |
| Breath Volume      | 1.488 | 05:47 |
| Air Blank          | 0.000 | 05:47 |
| Air Blank          | 0.000 | 05:49 |
| Subject Sample # 2 | 0.152 | 05:52 |
| Breath Volume      | 1.152 | 05:52 |
| Air Blank          | 0.000 | 05:53 |
| Control Test       | 0.079 | 05:53 |
| Air Blank          | 0.000 | 05:53 |
| Diagnostics Check  | OK    | 05:54 |

---

Data Download Date/Time: 07/21/2018 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OCASIO,FANNIE,E  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: OCASIO,F  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: OTHER  
ARREST TIME: 11:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 07/14/2018 12:17:14

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 12:16 |                           |
| Air Blank         | 0.000 | 12:16 |                           |
| Control Test      | 0.000 | 12:16 | Control Outside Tolerance |
| Air Blank         |       | 12:17 | Improper Sample           |

---

Data Download Date/Time: 07/21/2018 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OCASIO,F,E  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: OCASIO,F  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: OTHER  
ARREST TIME: 11:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 07/14/2018 12:20:48

---

|                   |    |       |                 |
|-------------------|----|-------|-----------------|
| Diagnostics Check | OK | 12:20 |                 |
| Air Blank         |    | 12:20 | Ambient Fail    |
| Air Blank         |    | 12:20 | Improper Sample |

---

Data Download Date/Time: 07/21/2018 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELRUSSO,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DELRUSSO,CHRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 07/15/2018 00:24:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:18 |
| Air Blank          | 0.000 | 00:18 |
| Control Test       | 0.080 | 00:19 |
| Air Blank          | 0.000 | 00:19 |
| Subject Sample # 1 | 0.033 | 00:20 |
| Breath Volume      | 2.507 | 00:20 |
| Air Blank          | 0.000 | 00:20 |
| Air Blank          | 0.000 | 00:22 |
| Subject Sample # 2 | 0.034 | 00:22 |
| Breath Volume      | 2.332 | 00:22 |
| Air Blank          | 0.000 | 00:23 |
| Control Test       | 0.079 | 00:23 |
| Air Blank          | 0.000 | 00:24 |
| Diagnostics Check  | OK    | 00:24 |

---

Data Download Date/Time: 07/21/2018 15:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAS,ASA,E  
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)  
ARRESTING OFFICER: DIAS,ASA  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: BUI  
ARREST TIME: 19:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: MORRA

DATE/TIME OF TEST: 07/15/2018 21:04:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:57 |
| Air Blank          | 0.000 | 20:58 |
| Control Test       | 0.080 | 20:58 |
| Air Blank          | 0.000 | 20:58 |
| Subject Sample # 1 | 0.180 | 20:59 |
| Breath Volume      | 2.164 | 20:59 |
| Air Blank          | 0.000 | 21:00 |
| Air Blank          | 0.000 | 21:01 |
| Subject Sample # 2 | 0.161 | 21:02 |
| Breath Volume      | 1.402 | 21:02 |
| Air Blank          | 0.000 | 21:03 |
| Control Test       | 0.078 | 21:03 |
| Air Blank          | 0.000 | 21:04 |
| Diagnostics Check  | OK    | 21:04 |

---

Data Download Date/Time: 07/21/2018 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLER,JESSICA,G  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: GALLER,JESSICA  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: ALVAREZ DE JONES

DATE/TIME OF TEST: 07/16/2018 00:05:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:58 |
| Air Blank          | 0.000 | 23:58 |
| Control Test       | 0.080 | 23:59 |
| Air Blank          | 0.000 | 23:59 |
| Subject Sample # 1 | 0.000 | 00:00 |
| Breath Volume      | 1.363 | 00:00 |
| Air Blank          | 0.000 | 00:01 |
| Air Blank          | 0.000 | 00:03 |
| Subject Sample # 2 | 0.000 | 00:03 |
| Breath Volume      | 1.402 | 00:03 |
| Air Blank          | 0.000 | 00:04 |
| Control Test       | 0.079 | 00:04 |
| Air Blank          | 0.000 | 00:05 |
| Diagnostics Check  | OK    | 00:05 |

---

Data Download Date/Time: 07/21/2018 15:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRINGTON,ALEX,M  
BREATH TEST OPERATOR AGENCY: Belleview PD  
ARRESTING OFFICER: HARRINGTON,ALEX  
ARRESTING OFFICER AGENCY: Belleview PD  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HOUDER

DATE/TIME OF TEST: 07/16/2018 03:37:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Control Test       | 0.080 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Subject Sample # 1 | 0.182 | 03:32 |
| Breath Volume      | 1.222 | 03:32 |
| Air Blank          | 0.000 | 03:33 |
| Air Blank          | 0.000 | 03:35 |
| Subject Sample # 2 | 0.185 | 03:35 |
| Breath Volume      | 1.464 | 03:35 |
| Air Blank          | 0.000 | 03:36 |
| Control Test       | 0.079 | 03:36 |
| Air Blank          | 0.000 | 03:37 |
| Diagnostics Check  | OK    | 03:37 |

---

Data Download Date/Time: 07/21/2018 15:44



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNDORFF,TERRY,D  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MUNDORFF,TERRY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 09:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 07/16/2018 10:37:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:31 |
| Air Blank          | 0.000 | 10:31 |
| Control Test       | 0.079 | 10:32 |
| Air Blank          | 0.000 | 10:32 |
| Subject Sample # 1 | 0.000 | 10:33 |
| Breath Volume      | 2.308 | 10:33 |
| Air Blank          | 0.000 | 10:33 |
| Air Blank          | 0.000 | 10:35 |
| Subject Sample # 2 | 0.000 | 10:36 |
| Breath Volume      | 1.960 | 10:36 |
| Air Blank          | 0.000 | 10:36 |
| Control Test       | 0.078 | 10:36 |
| Air Blank          | 0.000 | 10:37 |
| Diagnostics Check  | OK    | 10:37 |

---

Data Download Date/Time: 07/21/2018 15:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: BARTH,TRVOR  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:29

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 07/18/2018 17:21:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:14 |
| Air Blank          | 0.000 | 17:14 |
| Control Test       | 0.079 | 17:14 |
| Air Blank          | 0.000 | 17:15 |
| Subject Sample # 1 | 0.218 | 17:15 |
| Breath Volume      | 1.906 | 17:15 |
| Air Blank          | 0.000 | 17:16 |
| Air Blank          | 0.000 | 17:18 |
| Subject Sample # 2 | 0.203 | 17:19 |
| Breath Volume      | 1.914 | 17:19 |
| Air Blank          | 0.000 | 17:20 |
| Control Test       | 0.079 | 17:20 |
| Air Blank          | 0.000 | 17:21 |
| Diagnostics Check  | OK    | 17:21 |

---

Data Download Date/Time: 07/21/2018 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRYNER,DAVID,B  
BREATH TEST OPERATOR AGENCY: Belleview PD  
ARRESTING OFFICER: BRYNER,DAVID  
ARRESTING OFFICER AGENCY: Belleview PD  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 10:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:04

SUBJECT'S LAST NAME: NEARY

DATE/TIME OF TEST: 07/20/2018 11:34:23

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 11:25 |
| Air Blank         |     | 0.000 | 11:25 |
| Control Test      |     | 0.080 | 11:25 |
| Air Blank         |     | 0.000 | 11:26 |
| Subject Sample    | # 1 | 0.234 | 11:26 |
| Breath Volume     |     | 1.960 | 11:26 |
| Air Blank         |     | 0.000 | 11:27 |
| Air Blank         |     | 0.000 | 11:29 |
| Subject Sample    | # 2 | 0.208 | 11:29 |
| Breath Volume     |     | 1.941 | 11:29 |
| Air Blank         |     | 0.000 | 11:30 |
| Air Blank         |     | 0.000 | 11:32 |
| Subject Sample    | # 3 | 0.218 | 11:32 |
| Breath Volume     |     | 1.507 | 11:32 |
| Air Blank         |     | 0.000 | 11:33 |
| Control Test      |     | 0.079 | 11:33 |
| Air Blank         |     | 0.000 | 11:34 |
| Diagnostics Check |     | OK    | 11:34 |

---

Data Download Date/Time: 07/21/2018 15:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: UMANA,RAUL,J  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: UMANA,RAUL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 07/21/2018 01:42:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:36 |
| Air Blank          | 0.000 | 01:36 |
| Control Test       | 0.080 | 01:37 |
| Air Blank          | 0.000 | 01:37 |
| Subject Sample # 1 | 0.000 | 01:38 |
| Breath Volume      | 2.890 | 01:38 |
| Air Blank          | 0.000 | 01:39 |
| Air Blank          | 0.000 | 01:41 |
| Subject Sample # 2 | 0.000 | 01:41 |
| Breath Volume      | 1.792 | 01:41 |
| Air Blank          | 0.000 | 01:41 |
| Control Test       | 0.079 | 01:42 |
| Air Blank          | 0.000 | 01:42 |
| Diagnostics Check  | OK    | 01:42 |

---

Data Download Date/Time: 07/21/2018 15:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GENTHER,EMILY,S  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: MACDONALD,KAITLYN  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 07/21/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: GURCHARAN

DATE/TIME OF TEST: 07/21/2018 23:44:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:38 |
| Air Blank          | 0.000 | 23:38 |
| Control Test       | 0.080 | 23:39 |
| Air Blank          | 0.000 | 23:39 |
| Subject Sample # 1 | 0.201 | 23:40 |
| Breath Volume      | 3.441 | 23:40 |
| Air Blank          | 0.000 | 23:40 |
| Air Blank          | 0.000 | 23:42 |
| Subject Sample # 2 | 0.209 | 23:43 |
| Breath Volume      | 3.339 | 23:43 |
| Air Blank          | 0.000 | 23:43 |
| Control Test       | 0.080 | 23:44 |
| Air Blank          | 0.000 | 23:44 |
| Diagnostics Check  | OK    | 23:44 |

---

Data Download Date/Time: 08/27/2018 13:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WORFEL,KYLE,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: WORFEL,KYLE  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 07/21/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: FARMER SR

DATE/TIME OF TEST: 07/22/2018 01:07:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:00 |
| Air Blank          | 0.000 | 01:01 |
| Control Test       | 0.080 | 01:01 |
| Air Blank          | 0.000 | 01:01 |
| Subject Sample # 1 | 0.183 | 01:03 |
| Breath Volume      | 1.785 | 01:03 |
| Air Blank          | 0.000 | 01:03 |
| Air Blank          | 0.000 | 01:05 |
| Subject Sample # 2 | 0.179 | 01:06 |
| Breath Volume      | 1.367 | 01:06 |
| Air Blank          | 0.000 | 01:06 |
| Control Test       | 0.080 | 01:07 |
| Air Blank          | 0.000 | 01:07 |
| Diagnostics Check  | OK    | 01:07 |

---

Data Download Date/Time: 08/27/2018 13:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELRUSSO,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DELRUSSO,CHRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 07/21/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: SOSSA

DATE/TIME OF TEST: 07/24/2018 22:55:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:48 |
| Air Blank          | 0.000 | 22:48 |
| Control Test       | 0.080 | 22:48 |
| Air Blank          | 0.000 | 22:49 |
| Subject Sample # 1 | 0.176 | 22:50 |
| Breath Volume      | 1.457 | 22:50 |
| Air Blank          | 0.000 | 22:51 |
| Air Blank          | 0.000 | 22:52 |
| Subject Sample # 2 | 0.164 | 22:54 |
| Breath Volume      | 1.441 | 22:54 |
| Air Blank          | 0.000 | 22:55 |
| Control Test       | 0.079 | 22:55 |
| Air Blank          | 0.000 | 22:55 |
| Diagnostics Check  | OK    | 22:55 |

---

Data Download Date/Time: 08/27/2018 13:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WORFEL,KYLE,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: WORFEL,KYLE  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 07/21/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 07/26/2018 02:56:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:45 |
| Air Blank          | 0.000 | 02:46 |
| Control Test       | 0.080 | 02:46 |
| Air Blank          | 0.000 | 02:46 |
| Subject Sample # 1 | 0.252 | 02:49 |
| Breath Volume      | 1.703 | 02:49 |
| Air Blank          | 0.000 | 02:49 |
| Air Blank          | 0.000 | 02:51 |
| Subject Sample # 2 | 0.231 | 02:52 |
| Breath Volume      | 1.269 | 02:52 |
| Air Blank          | 0.000 | 02:52 |
| Air Blank          | 0.000 | 02:54 |
| Subject Sample # 3 | 0.250 | 02:55 |
| Breath Volume      | 1.468 | 02:55 |
| Air Blank          | 0.000 | 02:55 |
| Control Test       | 0.079 | 02:56 |
| Air Blank          | 0.000 | 02:56 |
| Diagnostics Check  | OK    | 02:56 |

---

Data Download Date/Time: 08/27/2018 13:20



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELRUSSO,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DELRUSSO,CHRISOPHER  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 07/21/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: KNUTSON

DATE/TIME OF TEST: 07/28/2018 19:59:54

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:57 |                      |
| Air Blank         | 0.000 | 19:57 |                      |
| Control Test      | 0.080 | 19:57 |                      |
| Air Blank         | 0.000 | 19:58 |                      |
| Subject Sample    | # 1   | 19:58 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:58 |                      |
| Air Blank         | 0.000 | 19:59 |                      |
| Control Test      | 0.080 | 19:59 |                      |
| Air Blank         | 0.000 | 19:59 |                      |
| Diagnostics Check | OK    | 19:59 |                      |

---

Data Download Date/Time: 08/27/2018 13:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,KAMIKA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: GAFFORD

DATE/TIME OF TEST: 07/12/2018 20:29:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:23 |
| Air Blank          | 0.000 | 20:23 |
| Control Test       | 0.080 | 20:24 |
| Air Blank          | 0.000 | 20:24 |
| Subject Sample # 1 | 0.000 | 20:25 |
| Breath Volume      | 2.464 | 20:25 |
| Air Blank          | 0.000 | 20:25 |
| Air Blank          | 0.000 | 20:27 |
| Subject Sample # 2 | 0.000 | 20:28 |
| Breath Volume      | 2.609 | 20:28 |
| Air Blank          | 0.000 | 20:28 |
| Control Test       | 0.078 | 20:29 |
| Air Blank          | 0.000 | 20:29 |
| Diagnostics Check  | OK    | 20:29 |

---

Data Download Date/Time: 08/07/2018 15:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 07/13/2018 01:50:01

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 01:42 |
| Air Blank         |     | 0.000 | 01:43 |
| Control Test      |     | 0.079 | 01:43 |
| Air Blank         |     | 0.000 | 01:43 |
| Subject Sample    | # 1 | 0.161 | 01:45 |
| Breath Volume     |     | 2.070 | 01:45 |
| Air Blank         |     | 0.000 | 01:45 |
| Air Blank         |     | 0.000 | 01:47 |
| Subject Sample    | # 2 | 0.166 | 01:48 |
| Breath Volume     |     | 2.378 | 01:48 |
| Air Blank         |     | 0.000 | 01:49 |
| Control Test      |     | 0.079 | 01:49 |
| Air Blank         |     | 0.000 | 01:49 |
| Diagnostics Check |     | OK    | 01:50 |

---

Data Download Date/Time: 08/07/2018 15:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FRANKLIN, R  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 07/13/2018 05:36:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:30 |
| Air Blank          | 0.000 | 05:30 |
| Control Test       | 0.079 | 05:31 |
| Air Blank          | 0.000 | 05:31 |
| Subject Sample # 1 | 0.166 | 05:32 |
| Breath Volume      | 2.039 | 05:32 |
| Air Blank          | 0.000 | 05:33 |
| Air Blank          | 0.000 | 05:35 |
| Subject Sample # 2 | 0.177 | 05:35 |
| Breath Volume      | 1.585 | 05:35 |
| Air Blank          | 0.000 | 05:36 |
| Control Test       | 0.080 | 05:36 |
| Air Blank          | 0.000 | 05:36 |
| Diagnostics Check  | OK    | 05:36 |

---

Data Download Date/Time: 08/07/2018 15:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HILES, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: ESSER

DATE/TIME OF TEST: 07/13/2018 20:28:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:22 |
| Air Blank          | 0.000 | 20:22 |
| Control Test       | 0.080 | 20:22 |
| Air Blank          | 0.000 | 20:23 |
| Subject Sample # 1 | 0.000 | 20:23 |
| Breath Volume      | 1.910 | 20:23 |
| Air Blank          | 0.000 | 20:24 |
| Air Blank          | 0.000 | 20:26 |
| Subject Sample # 2 | 0.000 | 20:26 |
| Breath Volume      | 2.507 | 20:26 |
| Air Blank          | 0.000 | 20:27 |
| Control Test       | 0.080 | 20:27 |
| Air Blank          | 0.000 | 20:27 |
| Diagnostics Check  | OK    | 20:28 |

---

Data Download Date/Time: 08/07/2018 15:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HILES,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: ALBALADEJO

DATE/TIME OF TEST: 07/14/2018 01:24:32

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 01:15 |
| Air Blank         |     | 0.000 | 01:15 |
| Control Test      |     | 0.079 | 01:16 |
| Air Blank         |     | 0.000 | 01:16 |
| Subject Sample    | # 1 | 0.275 | 01:19 |
| Breath Volume     |     | 1.656 | 01:19 |
| Air Blank         |     | 0.000 | 01:20 |
| Air Blank         |     | 0.000 | 01:22 |
| Subject Sample    | # 2 | 0.256 | 01:23 |
| Breath Volume     |     | 2.265 | 01:23 |
| Air Blank         |     | 0.000 | 01:23 |
| Control Test      |     | 0.077 | 01:23 |
| Air Blank         |     | 0.000 | 01:24 |
| Diagnostics Check |     | OK    | 01:24 |

---

Data Download Date/Time: 08/07/2018 15:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRANNIS, RONALD, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SIERRA, P  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: FAULKNER JR.

DATE/TIME OF TEST: 07/14/2018 03:21:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:15 |
| Air Blank          | 0.000 | 03:15 |
| Control Test       | 0.079 | 03:16 |
| Air Blank          | 0.000 | 03:16 |
| Subject Sample # 1 | 0.159 | 03:17 |
| Breath Volume      | 1.910 | 03:17 |
| Air Blank          | 0.000 | 03:18 |
| Air Blank          | 0.000 | 03:19 |
| Subject Sample # 2 | 0.169 | 03:20 |
| Breath Volume      | 3.531 | 03:20 |
| Air Blank          | 0.000 | 03:20 |
| Control Test       | 0.079 | 03:21 |
| Air Blank          | 0.000 | 03:21 |
| Diagnostics Check  | OK    | 03:21 |

---

Data Download Date/Time: 08/23/2018 12:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,KAMIKA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: DOMINGUEZ RAMOS

DATE/TIME OF TEST: 07/14/2018 04:00:37

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 03:54 |
| Air Blank         |     | 0.000 | 03:54 |
| Control Test      |     | 0.079 | 03:54 |
| Air Blank         |     | 0.000 | 03:55 |
| Subject Sample    | # 1 | 0.254 | 03:56 |
| Breath Volume     |     | 2.363 | 03:56 |
| Air Blank         |     | 0.000 | 03:56 |
| Air Blank         |     | 0.000 | 03:58 |
| Subject Sample    | # 2 | 0.258 | 03:59 |
| Breath Volume     |     | 2.097 | 03:59 |
| Air Blank         |     | 0.000 | 03:59 |
| Control Test      |     | 0.079 | 04:00 |
| Air Blank         |     | 0.000 | 04:00 |
| Diagnostics Check |     | OK    | 04:00 |

---

Data Download Date/Time: 08/23/2018 12:27



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HILES,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: LAUDER

DATE/TIME OF TEST: 07/15/2018 03:06:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:59 |
| Air Blank          | 0.000 | 02:59 |
| Control Test       | 0.080 | 02:59 |
| Air Blank          | 0.000 | 03:00 |
| Subject Sample # 1 | 0.121 | 03:02 |
| Breath Volume      | 1.457 | 03:02 |
| Air Blank          | 0.000 | 03:02 |
| Air Blank          | 0.000 | 03:04 |
| Subject Sample # 2 | 0.111 | 03:05 |
| Breath Volume      | 1.894 | 03:05 |
| Air Blank          | 0.000 | 03:05 |
| Control Test       | 0.078 | 03:05 |
| Air Blank          | 0.000 | 03:06 |
| Diagnostics Check  | OK    | 03:06 |

---

Data Download Date/Time: 08/23/2018 12:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,REBECCA,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LESHICK,E  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 13:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:27

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 07/15/2018 15:02:07

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 14:48 |                |
| Air Blank         |     | 0.000 | 14:48 |                |
| Control Test      |     | 0.079 | 14:49 |                |
| Air Blank         |     | 0.000 | 14:49 |                |
| Subject Sample    | # 1 | 0.194 | 14:52 | Volume Not Met |
| Breath Volume     |     | 0.949 | 14:52 |                |
| Air Blank         |     | 0.000 | 14:53 |                |
| Air Blank         |     | 0.000 | 14:55 |                |
| Subject Sample    | # 2 | 0.249 | 14:57 |                |
| Breath Volume     |     | 2.371 | 14:57 |                |
| Air Blank         |     | 0.000 | 14:58 |                |
| Air Blank         |     | 0.000 | 15:00 |                |
| Subject Sample    | # 3 | 0.259 | 15:00 |                |
| Breath Volume     |     | 2.113 | 15:00 |                |
| Air Blank         |     | 0.000 | 15:01 |                |
| Control Test      |     | 0.078 | 15:01 |                |
| Air Blank         |     | 0.000 | 15:02 |                |
| Diagnostics Check |     | OK    | 15:02 |                |

---

Data Download Date/Time: 08/23/2018 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,REBECCA,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ANDERSON,J  
ARRESTING OFFICER AGENCY: Temple Terrace PD  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:04

SUBJECT'S LAST NAME: KONDRAD

DATE/TIME OF TEST: 07/15/2018 16:31:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:24 |
| Air Blank          | 0.000 | 16:25 |
| Control Test       | 0.079 | 16:25 |
| Air Blank          | 0.000 | 16:26 |
| Subject Sample # 1 | 0.000 | 16:26 |
| Breath Volume      | 1.371 | 16:26 |
| Air Blank          | 0.000 | 16:27 |
| Air Blank          | 0.000 | 16:29 |
| Subject Sample # 2 | 0.000 | 16:29 |
| Breath Volume      | 1.148 | 16:29 |
| Air Blank          | 0.000 | 16:30 |
| Control Test       | 0.079 | 16:30 |
| Air Blank          | 0.000 | 16:31 |
| Diagnostics Check  | OK    | 16:31 |

---

Data Download Date/Time: 08/23/2018 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: NIEUWENDAM,R  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: ENGESETH

DATE/TIME OF TEST: 07/15/2018 22:10:14

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 21:59 |
| Air Blank         |     | 0.000 | 22:00 |
| Control Test      |     | 0.080 | 22:00 |
| Air Blank         |     | 0.000 | 22:01 |
| Subject Sample    | # 1 | 0.326 | 22:02 |
| Breath Volume     |     | 3.417 | 22:02 |
| Air Blank         |     | 0.000 | 22:03 |
| Air Blank         |     | 0.000 | 22:04 |
| Subject Sample    | # 2 | 0.295 | 22:05 |
| Breath Volume     |     | 2.203 | 22:05 |
| Air Blank         |     | 0.000 | 22:06 |
| Air Blank         |     | 0.000 | 22:08 |
| Subject Sample    | # 3 | 0.328 | 22:08 |
| Breath Volume     |     | 3.300 | 22:08 |
| Air Blank         |     | 0.000 | 22:09 |
| Control Test      |     | 0.078 | 22:09 |
| Air Blank         |     | 0.000 | 22:10 |
| Diagnostics Check |     | OK    | 22:10 |

---

Data Download Date/Time: 08/23/2018 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: NIEUWENDAM,R  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: CARBO

DATE/TIME OF TEST: 07/16/2018 04:36:42

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:28 |                      |
| Air Blank         | 0.000 | 04:28 |                      |
| Control Test      | 0.080 | 04:29 |                      |
| Air Blank         | 0.000 | 04:29 |                      |
| Subject Sample    | # 1   | 04:32 | No Sample Provided   |
| Breath Volume     | 0.000 | 04:32 |                      |
| Air Blank         | 0.000 | 04:33 |                      |
| Air Blank         | 0.000 | 04:35 |                      |
| Subject Sample    | # 2   | 04:35 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:35 |                      |
| Air Blank         | 0.000 | 04:35 |                      |
| Control Test      | 0.079 | 04:36 |                      |
| Air Blank         | 0.000 | 04:36 |                      |
| Diagnostics Check | OK    | 04:36 |                      |

---

Data Download Date/Time: 08/23/2018 12:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN, T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: BURNHAM

DATE/TIME OF TEST: 07/17/2018 18:45:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:39 |
| Air Blank          | 0.000 | 18:40 |
| Control Test       | 0.080 | 18:40 |
| Air Blank          | 0.000 | 18:40 |
| Subject Sample # 1 | 0.031 | 18:41 |
| Breath Volume      | 3.203 | 18:41 |
| Air Blank          | 0.000 | 18:41 |
| Air Blank          | 0.000 | 18:43 |
| Subject Sample # 2 | 0.033 | 18:44 |
| Breath Volume      | 2.492 | 18:44 |
| Air Blank          | 0.000 | 18:44 |
| Control Test       | 0.080 | 18:44 |
| Air Blank          | 0.000 | 18:45 |
| Diagnostics Check  | OK    | 18:45 |

---

Data Download Date/Time: 08/23/2018 12:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,KAMIKA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WOLSTENHOLM,A  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: STOVER

DATE/TIME OF TEST: 07/17/2018 21:29:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:23 |
| Air Blank          | 0.000 | 21:23 |
| Control Test       | 0.080 | 21:24 |
| Air Blank          | 0.000 | 21:24 |
| Subject Sample # 1 | 0.000 | 21:25 |
| Breath Volume      | 2.277 | 21:25 |
| Air Blank          | 0.000 | 21:25 |
| Air Blank          | 0.000 | 21:27 |
| Subject Sample # 2 | 0.000 | 21:28 |
| Breath Volume      | 2.078 | 21:28 |
| Air Blank          | 0.000 | 21:28 |
| Control Test       | 0.079 | 21:28 |
| Air Blank          | 0.000 | 21:29 |
| Diagnostics Check  | OK    | 21:29 |

---

Data Download Date/Time: 08/23/2018 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BELEN,C  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/26/2018 01:27:24

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 01:18 |
| Air Blank         |     | 0.000 | 01:19 |
| Control Test      |     | 0.079 | 01:19 |
| Air Blank         |     | 0.000 | 01:20 |
| Subject Sample    | # 1 | 0.108 | 01:21 |
| Breath Volume     |     | 2.632 | 01:21 |
| Air Blank         |     | 0.000 | 01:22 |
| Air Blank         |     | 0.000 | 01:23 |
| Subject Sample    | # 2 | 0.116 | 01:25 |
| Breath Volume     |     | 1.945 | 01:25 |
| Air Blank         |     | 0.000 | 01:26 |
| Control Test      |     | 0.080 | 01:26 |
| Air Blank         |     | 0.000 | 01:27 |
| Diagnostics Check |     | OK    | 01:27 |

---

Data Download Date/Time: 08/23/2018 12:27



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS,ANGELA,V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SMITH,NONE  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: ADMIN  
ARREST TIME: 16:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:23

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 07/26/2018 16:48:59

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 16:43 |                      |
| Air Blank         | 0.000     | 16:43 |                      |
| Control Test      | 0.079     | 16:44 |                      |
| Air Blank         | 0.000     | 16:44 |                      |
| Subject Sample    | # 1 0.199 | 16:44 |                      |
| Breath Volume     | 2.027     | 16:44 |                      |
| Air Blank         | 0.000     | 16:45 |                      |
| Air Blank         | 0.000     | 16:47 |                      |
| Subject Sample    | # 2       | 16:47 | Subject Test Refused |
| Breath Volume     | 0.000     | 16:47 |                      |
| Air Blank         | 0.000     | 16:48 |                      |
| Control Test      | 0.078     | 16:48 |                      |
| Air Blank         | 0.000     | 16:48 |                      |
| Diagnostics Check | OK        | 16:48 |                      |

---

Data Download Date/Time: 08/23/2018 12:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 07/27/2018 00:14:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:05 |
| Air Blank          | 0.000 | 00:06 |
| Control Test       | 0.080 | 00:06 |
| Air Blank          | 0.000 | 00:06 |
| Subject Sample # 1 | 0.127 | 00:08 |
| Breath Volume      | 2.179 | 00:08 |
| Air Blank          | 0.000 | 00:09 |
| Air Blank          | 0.000 | 00:10 |
| Subject Sample # 2 | 0.130 | 00:12 |
| Breath Volume      | 2.730 | 00:12 |
| Air Blank          | 0.000 | 00:13 |
| Control Test       | 0.080 | 00:13 |
| Air Blank          | 0.000 | 00:14 |
| Diagnostics Check  | OK    | 00:14 |

---

Data Download Date/Time: 08/23/2018 12:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRANNIS, RONALD, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: DONALDSON

DATE/TIME OF TEST: 07/27/2018 04:18:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:12 |
| Air Blank          | 0.000 | 04:12 |
| Control Test       | 0.080 | 04:13 |
| Air Blank          | 0.000 | 04:13 |
| Subject Sample # 1 | 0.086 | 04:14 |
| Breath Volume      | 2.972 | 04:14 |
| Air Blank          | 0.000 | 04:14 |
| Air Blank          | 0.000 | 04:16 |
| Subject Sample # 2 | 0.083 | 04:17 |
| Breath Volume      | 3.195 | 04:17 |
| Air Blank          | 0.000 | 04:18 |
| Control Test       | 0.079 | 04:18 |
| Air Blank          | 0.000 | 04:18 |
| Diagnostics Check  | OK    | 04:18 |

---

Data Download Date/Time: 08/23/2018 12:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BRIZENDINE, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/27/2018 05:22:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:16 |
| Air Blank          | 0.000 | 05:16 |
| Control Test       | 0.079 | 05:16 |
| Air Blank          | 0.000 | 05:17 |
| Subject Sample # 1 | 0.220 | 05:17 |
| Breath Volume      | 2.011 | 05:17 |
| Air Blank          | 0.000 | 05:18 |
| Air Blank          | 0.000 | 05:20 |
| Subject Sample # 2 | 0.220 | 05:20 |
| Breath Volume      | 2.304 | 05:20 |
| Air Blank          | 0.000 | 05:21 |
| Control Test       | 0.078 | 05:21 |
| Air Blank          | 0.000 | 05:22 |
| Diagnostics Check  | OK    | 05:22 |

---

Data Download Date/Time: 08/23/2018 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRANNIS, RONALD, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SIERRA, P  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: ROLOSON

DATE/TIME OF TEST: 07/28/2018 01:51:04

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 01:44 |
| Air Blank         |     | 0.000 | 01:45 |
| Control Test      |     | 0.079 | 01:45 |
| Air Blank         |     | 0.000 | 01:46 |
| Subject Sample    | # 1 | 0.174 | 01:46 |
| Breath Volume     |     | 2.613 | 01:46 |
| Air Blank         |     | 0.000 | 01:47 |
| Air Blank         |     | 0.000 | 01:49 |
| Subject Sample    | # 2 | 0.176 | 01:49 |
| Breath Volume     |     | 2.664 | 01:49 |
| Air Blank         |     | 0.000 | 01:50 |
| Control Test      |     | 0.078 | 01:50 |
| Air Blank         |     | 0.000 | 01:50 |
| Diagnostics Check |     | OK    | 01:51 |

---

Data Download Date/Time: 08/23/2018 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY, D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: BARRETT

DATE/TIME OF TEST: 07/28/2018 04:58:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:51 |
| Air Blank          | 0.000 | 04:52 |
| Control Test       | 0.080 | 04:52 |
| Air Blank          | 0.000 | 04:53 |
| Subject Sample # 1 | 0.172 | 04:53 |
| Breath Volume      | 1.535 | 04:53 |
| Air Blank          | 0.000 | 04:54 |
| Air Blank          | 0.000 | 04:56 |
| Subject Sample # 2 | 0.183 | 04:56 |
| Breath Volume      | 1.542 | 04:56 |
| Air Blank          | 0.000 | 04:57 |
| Control Test       | 0.079 | 04:57 |
| Air Blank          | 0.000 | 04:58 |
| Diagnostics Check  | OK    | 04:58 |

---

Data Download Date/Time: 08/23/2018 12:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HALL, L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 07/28/2018 05:57:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:51 |
| Air Blank          | 0.000 | 05:52 |
| Control Test       | 0.078 | 05:52 |
| Air Blank          | 0.000 | 05:52 |
| Subject Sample # 1 | 0.229 | 05:53 |
| Breath Volume      | 2.070 | 05:53 |
| Air Blank          | 0.000 | 05:53 |
| Air Blank          | 0.000 | 05:55 |
| Subject Sample # 2 | 0.227 | 05:56 |
| Breath Volume      | 1.343 | 05:56 |
| Air Blank          | 0.000 | 05:56 |
| Control Test       | 0.079 | 05:57 |
| Air Blank          | 0.000 | 05:57 |
| Diagnostics Check  | OK    | 05:57 |

---

Data Download Date/Time: 08/23/2018 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRANNIS, RONALD, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SIERRA, P  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: CRUMPJR.

DATE/TIME OF TEST: 07/28/2018 06:36:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:30 |
| Air Blank          | 0.000 | 06:31 |
| Control Test       | 0.079 | 06:31 |
| Air Blank          | 0.000 | 06:32 |
| Subject Sample # 1 | 0.152 | 06:32 |
| Breath Volume      | 3.875 | 06:32 |
| Air Blank          | 0.000 | 06:33 |
| Air Blank          | 0.000 | 06:34 |
| Subject Sample # 2 | 0.151 | 06:35 |
| Breath Volume      | 3.531 | 06:35 |
| Air Blank          | 0.000 | 06:35 |
| Control Test       | 0.078 | 06:36 |
| Air Blank          | 0.000 | 06:36 |
| Diagnostics Check  | OK    | 06:36 |

---

Data Download Date/Time: 08/23/2018 12:28



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DRISCOLL,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: HERNANDEZ DIAZ

DATE/TIME OF TEST: 07/28/2018 19:01:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:55 |
| Air Blank          | 0.000 | 18:56 |
| Control Test       | 0.080 | 18:56 |
| Air Blank          | 0.000 | 18:56 |
| Subject Sample # 1 | 0.250 | 18:57 |
| Breath Volume      | 3.222 | 18:57 |
| Air Blank          | 0.000 | 18:57 |
| Air Blank          | 0.000 | 18:59 |
| Subject Sample # 2 | 0.246 | 19:00 |
| Breath Volume      | 2.398 | 19:00 |
| Air Blank          | 0.000 | 19:00 |
| Control Test       | 0.077 | 19:01 |
| Air Blank          | 0.000 | 19:01 |
| Diagnostics Check  | OK    | 19:01 |

---

Data Download Date/Time: 08/23/2018 12:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HALL,L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: TONEY

DATE/TIME OF TEST: 07/29/2018 04:59:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:52 |
| Air Blank          | 0.000 | 04:53 |
| Control Test       | 0.078 | 04:53 |
| Air Blank          | 0.000 | 04:53 |
| Subject Sample # 1 | 0.129 | 04:54 |
| Breath Volume      | 1.898 | 04:54 |
| Air Blank          | 0.000 | 04:55 |
| Air Blank          | 0.000 | 04:57 |
| Subject Sample # 2 | 0.142 | 04:58 |
| Breath Volume      | 1.710 | 04:58 |
| Air Blank          | 0.000 | 04:58 |
| Control Test       | 0.077 | 04:59 |
| Air Blank          | 0.000 | 04:59 |
| Diagnostics Check  | OK    | 04:59 |

---

Data Download Date/Time: 08/23/2018 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HONEYGHON,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: JUAREZ MENDOZA

DATE/TIME OF TEST: 07/29/2018 23:49:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:43 |
| Air Blank          | 0.000 | 23:43 |
| Control Test       | 0.080 | 23:43 |
| Air Blank          | 0.000 | 23:44 |
| Subject Sample # 1 | 0.155 | 23:44 |
| Breath Volume      | 4.433 | 23:44 |
| Air Blank          | 0.000 | 23:45 |
| Air Blank          | 0.000 | 23:47 |
| Subject Sample # 2 | 0.168 | 23:47 |
| Breath Volume      | 4.386 | 23:47 |
| Air Blank          | 0.000 | 23:48 |
| Control Test       | 0.078 | 23:48 |
| Air Blank          | 0.000 | 23:49 |
| Diagnostics Check  | OK    | 23:49 |

---

Data Download Date/Time: 08/23/2018 12:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GONZALEZ,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: CARABALLO MELENDEZ

DATE/TIME OF TEST: 07/30/2018 19:54:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:46 |
| Air Blank          | 0.000 | 19:46 |
| Control Test       | 0.080 | 19:46 |
| Air Blank          | 0.000 | 19:47 |
| Subject Sample # 1 | 0.244 | 19:49 |
| Breath Volume      | 1.343 | 19:49 |
| Air Blank          | 0.000 | 19:49 |
| Air Blank          | 0.000 | 19:51 |
| Subject Sample # 2 | 0.241 | 19:53 |
| Breath Volume      | 1.535 | 19:53 |
| Air Blank          | 0.000 | 19:53 |
| Control Test       | 0.079 | 19:53 |
| Air Blank          | 0.000 | 19:54 |
| Diagnostics Check  | OK    | 19:54 |

---

Data Download Date/Time: 08/23/2018 12:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:46

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 07/31/2018 07:14:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:07 |
| Air Blank          | 0.000 | 07:08 |
| Control Test       | 0.079 | 07:08 |
| Air Blank          | 0.000 | 07:09 |
| Subject Sample # 1 | 0.048 | 07:09 |
| Breath Volume      | 2.339 | 07:09 |
| Air Blank          | 0.000 | 07:10 |
| Air Blank          | 0.000 | 07:12 |
| Subject Sample # 2 | 0.045 | 07:12 |
| Breath Volume      | 2.144 | 07:12 |
| Air Blank          | 0.000 | 07:13 |
| Control Test       | 0.080 | 07:13 |
| Air Blank          | 0.000 | 07:14 |
| Diagnostics Check  | OK    | 07:14 |

---

Data Download Date/Time: 08/23/2018 12:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS,ANGELA,V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUSWELL,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: ALBRITTON

DATE/TIME OF TEST: 07/31/2018 16:55:52

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 16:48 |
| Air Blank         |     | 0.000 | 16:49 |
| Control Test      |     | 0.080 | 16:49 |
| Air Blank         |     | 0.000 | 16:49 |
| Subject Sample    | # 1 | 0.278 | 16:51 |
| Breath Volume     |     | 1.687 | 16:51 |
| Air Blank         |     | 0.000 | 16:52 |
| Air Blank         |     | 0.000 | 16:53 |
| Subject Sample    | # 2 | 0.281 | 16:54 |
| Breath Volume     |     | 1.414 | 16:54 |
| Air Blank         |     | 0.000 | 16:54 |
| Control Test      |     | 0.078 | 16:55 |
| Air Blank         |     | 0.000 | 16:55 |
| Diagnostics Check |     | OK    | 16:55 |

---

Data Download Date/Time: 08/23/2018 12:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FRANKLIN, R  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A1  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 07/13/2018 01:21:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:15 |
| Air Blank          | 0.000 | 01:16 |
| Control Test       | 0.080 | 01:16 |
| Air Blank          | 0.000 | 01:16 |
| Subject Sample # 1 | 0.038 | 01:17 |
| Breath Volume      | 1.476 | 01:17 |
| Air Blank          | 0.000 | 01:17 |
| Air Blank          | 0.000 | 01:19 |
| Subject Sample # 2 | 0.040 | 01:20 |
| Breath Volume      | 1.339 | 01:20 |
| Air Blank          | 0.000 | 01:21 |
| Control Test       | 0.080 | 01:21 |
| Air Blank          | 0.000 | 01:21 |
| Diagnostics Check  | OK    | 01:21 |

---

Data Download Date/Time: 07/16/2018 09:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: COSTA,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A1  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 05:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:14

SUBJECT'S LAST NAME: WALTON

DATE/TIME OF TEST: 07/14/2018 06:45:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:35 |
| Air Blank          | 0.000 | 06:35 |
| Control Test       | 0.080 | 06:36 |
| Air Blank          | 0.000 | 06:36 |
| Subject Sample # 1 | 0.133 | 06:39 |
| Breath Volume      | 1.347 | 06:39 |
| Air Blank          | 0.000 | 06:40 |
| Air Blank          | 0.000 | 06:41 |
| Subject Sample # 2 | 0.130 | 06:43 |
| Breath Volume      | 1.406 | 06:43 |
| Air Blank          | 0.000 | 06:44 |
| Control Test       | 0.080 | 06:44 |
| Air Blank          | 0.000 | 06:45 |
| Diagnostics Check  | OK    | 06:45 |

---

Data Download Date/Time: 07/16/2018 09:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: COSTA,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A1  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: KOPFLER

DATE/TIME OF TEST: 07/15/2018 00:49:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:42 |
| Air Blank          | 0.000 | 00:43 |
| Control Test       | 0.080 | 00:43 |
| Air Blank          | 0.000 | 00:43 |
| Subject Sample # 1 | 0.212 | 00:44 |
| Breath Volume      | 3.652 | 00:44 |
| Air Blank          | 0.000 | 00:45 |
| Air Blank          | 0.000 | 00:47 |
| Subject Sample # 2 | 0.199 | 00:47 |
| Breath Volume      | 1.621 | 00:47 |
| Air Blank          | 0.000 | 00:48 |
| Control Test       | 0.080 | 00:48 |
| Air Blank          | 0.000 | 00:49 |
| Diagnostics Check  | OK    | 00:49 |

---

Data Download Date/Time: 07/16/2018 09:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REIMUNDY,B,  
BREATH TEST OPERATOR AGENCY: Clermont PD  
ARRESTING OFFICER: FAUST,J  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 07/01/2018 02:17:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:10 |
| Air Blank          | 0.000 | 02:11 |
| Control Test       | 0.079 | 02:11 |
| Air Blank          | 0.000 | 02:12 |
| Subject Sample # 1 | 0.175 | 02:12 |
| Breath Volume      | 1.878 | 02:12 |
| Air Blank          | 0.000 | 02:13 |
| Air Blank          | 0.000 | 02:15 |
| Subject Sample # 2 | 0.172 | 02:15 |
| Breath Volume      | 1.574 | 02:15 |
| Air Blank          | 0.000 | 02:16 |
| Control Test       | 0.080 | 02:16 |
| Air Blank          | 0.000 | 02:17 |
| Diagnostics Check  | OK    | 02:17 |

---

Data Download Date/Time: 07/06/2018 23:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARPER,JIMMIE,L  
BREATH TEST OPERATOR AGENCY: Clermont PD  
ARRESTING OFFICER: KRUSE,KRIS  
ARRESTING OFFICER AGENCY: Clermont PD  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: POLINO

DATE/TIME OF TEST: 07/07/2018 03:00:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:52 |
| Air Blank          | 0.000 | 02:52 |
| Control Test       | 0.079 | 02:53 |
| Air Blank          | 0.000 | 02:53 |
| Subject Sample # 1 | 0.169 | 02:54 |
| Breath Volume      | 2.621 | 02:54 |
| Air Blank          | 0.000 | 02:55 |
| Air Blank          | 0.000 | 02:57 |
| Subject Sample # 2 | 0.162 | 02:58 |
| Breath Volume      | 2.410 | 02:58 |
| Air Blank          | 0.000 | 02:59 |
| Control Test       | 0.080 | 02:59 |
| Air Blank          | 0.000 | 03:00 |
| Diagnostics Check  | OK    | 03:00 |

---

Data Download Date/Time: 08/15/2018 23:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,P  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANDRA  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: NODD

DATE/TIME OF TEST: 07/08/2018 03:11:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:04 |
| Air Blank          | 0.000 | 03:05 |
| Control Test       | 0.080 | 03:05 |
| Air Blank          | 0.000 | 03:06 |
| Subject Sample # 1 | 0.048 | 03:06 |
| Breath Volume      | 1.925 | 03:06 |
| Air Blank          | 0.000 | 03:07 |
| Air Blank          | 0.000 | 03:09 |
| Subject Sample # 2 | 0.048 | 03:09 |
| Breath Volume      | 1.414 | 03:09 |
| Air Blank          | 0.000 | 03:10 |
| Control Test       | 0.080 | 03:10 |
| Air Blank          | 0.000 | 03:11 |
| Diagnostics Check  | OK    | 03:11 |

---

Data Download Date/Time: 08/15/2018 23:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCNEIL,CALVIN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: ALIJAGIC,SAMIR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: MCNEIL

DATE/TIME OF TEST: 07/08/2018 19:44:20

---

|                   |       |       |                  |
|-------------------|-------|-------|------------------|
| Diagnostics Check | OK    | 19:41 |                  |
| Air Blank         | 0.000 | 19:42 |                  |
| Control Test      | 0.079 | 19:42 |                  |
| Air Blank         | 0.000 | 19:42 |                  |
| Subject Sample    | # 1   | 19:43 | Sequence Aborted |
| Breath Volume     | 0.078 | 19:43 |                  |
| Air Blank         | 0.000 | 19:44 |                  |

---

Data Download Date/Time: 08/15/2018 23:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCNEIL,CALVIN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: ALIJAGIC,SAMIR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: MCNEIL

DATE/TIME OF TEST: 07/08/2018 19:52:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:46 |
| Air Blank          | 0.000 | 19:46 |
| Control Test       | 0.078 | 19:46 |
| Air Blank          | 0.000 | 19:47 |
| Subject Sample # 1 | 0.000 | 19:48 |
| Breath Volume      | 1.878 | 19:48 |
| Air Blank          | 0.000 | 19:48 |
| Air Blank          | 0.000 | 19:50 |
| Subject Sample # 2 | 0.000 | 19:51 |
| Breath Volume      | 2.019 | 19:51 |
| Air Blank          | 0.000 | 19:51 |
| Control Test       | 0.079 | 19:51 |
| Air Blank          | 0.000 | 19:52 |
| Diagnostics Check  | OK    | 19:52 |

---

Data Download Date/Time: 08/15/2018 23:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOLLERY,MICHAEL,D  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: DOLLERY,MICHAEL  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: DEVORA

DATE/TIME OF TEST: 07/11/2018 22:15:35

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:12 |                      |
| Air Blank         | 0.000 | 22:12 |                      |
| Control Test      | 0.079 | 22:13 |                      |
| Air Blank         | 0.000 | 22:13 |                      |
| Subject Sample    | # 1   | 22:14 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:14 |                      |
| Air Blank         | 0.000 | 22:14 |                      |
| Control Test      | 0.079 | 22:14 |                      |
| Air Blank         | 0.000 | 22:15 |                      |
| Diagnostics Check | OK    | 22:15 |                      |

---

Data Download Date/Time: 08/15/2018 23:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEAL, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)  
ARRESTING OFFICER: AUBER5, JOHN  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: BUI  
ARREST TIME: 16:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: VANMETER

DATE/TIME OF TEST: 07/15/2018 18:54:13

---

|                   |       |       |                 |
|-------------------|-------|-------|-----------------|
| Diagnostics Check | OK    | 18:52 |                 |
| Air Blank         | 0.000 | 18:53 |                 |
| Control Test      | 0.079 | 18:53 |                 |
| Air Blank         |       | 18:53 | Improper Sample |
| Air Blank         | 0.000 | 18:54 |                 |

---

Data Download Date/Time: 08/15/2018 23:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEAL, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)  
ARRESTING OFFICER: AUBER, JOHN  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: BUI  
ARREST TIME: 16:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: ANMETER

DATE/TIME OF TEST: 07/15/2018 19:02:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:56 |
| Air Blank          | 0.000 | 18:56 |
| Control Test       | 0.075 | 18:57 |
| Air Blank          | 0.000 | 18:57 |
| Subject Sample # 1 | 0.206 | 18:57 |
| Breath Volume      | 2.824 | 18:57 |
| Air Blank          | 0.000 | 18:58 |
| Air Blank          | 0.000 | 19:00 |
| Subject Sample # 2 | 0.203 | 19:00 |
| Breath Volume      | 3.226 | 19:00 |
| Air Blank          | 0.000 | 19:01 |
| Control Test       | 0.078 | 19:01 |
| Air Blank          | 0.000 | 19:02 |
| Diagnostics Check  | OK    | 19:02 |

---

Data Download Date/Time: 08/15/2018 23:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,P  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: PITCHER,DARRYL  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 14:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: IGNOT

DATE/TIME OF TEST: 07/25/2018 16:23:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:16 |
| Air Blank          | 0.000 | 16:16 |
| Control Test       | 0.079 | 16:17 |
| Air Blank          | 0.000 | 16:17 |
| Subject Sample # 1 | 0.196 | 16:18 |
| Breath Volume      | 1.109 | 16:18 |
| Air Blank          | 0.000 | 16:19 |
| Air Blank          | 0.000 | 16:21 |
| Subject Sample # 2 | 0.186 | 16:21 |
| Breath Volume      | 3.082 | 16:21 |
| Air Blank          | 0.000 | 16:22 |
| Control Test       | 0.078 | 16:22 |
| Air Blank          | 0.000 | 16:23 |
| Diagnostics Check  | OK    | 16:23 |

---

Data Download Date/Time: 08/15/2018 23:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WURST,JOSHUA,D  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: POPPELREITER,DERICK  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: LOUDERMILK

DATE/TIME OF TEST: 07/01/2018 00:38:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:32 |
| Air Blank          | 0.000 | 00:33 |
| Control Test       | 0.079 | 00:33 |
| Air Blank          | 0.000 | 00:33 |
| Subject Sample # 1 | 0.267 | 00:34 |
| Breath Volume      | 3.351 | 00:34 |
| Air Blank          | 0.000 | 00:35 |
| Air Blank          | 0.000 | 00:37 |
| Subject Sample # 2 | 0.249 | 00:37 |
| Breath Volume      | 2.675 | 00:37 |
| Air Blank          | 0.000 | 00:38 |
| Control Test       | 0.079 | 00:38 |
| Air Blank          | 0.000 | 00:38 |
| Diagnostics Check  | OK    | 00:38 |

---

Data Download Date/Time: 07/06/2018 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUTTON,KERRY,D  
BREATH TEST OPERATOR AGENCY: Lynn Haven PD  
ARRESTING OFFICER: GROBB,RANDY  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: GIRAITIS

DATE/TIME OF TEST: 07/06/2018 03:20:11

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:17 |                      |
| Air Blank         | 0.000 | 03:17 |                      |
| Control Test      | 0.080 | 03:18 |                      |
| Air Blank         | 0.000 | 03:18 |                      |
| Subject Sample    | # 1   | 03:18 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:18 |                      |
| Air Blank         | 0.000 | 03:19 |                      |
| Control Test      | 0.079 | 03:19 |                      |
| Air Blank         | 0.000 | 03:20 |                      |
| Diagnostics Check | OK    | 03:20 |                      |

---

Data Download Date/Time: 08/09/2018 01:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAYFIELD, MATTHEW, E  
BREATH TEST OPERATOR AGENCY: Lynn Haven PD  
ARRESTING OFFICER: MAYFIELD, MATTHEW  
ARRESTING OFFICER AGENCY: Lynn Haven PD  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: CAMERON

DATE/TIME OF TEST: 07/06/2018 17:29:17

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 17:21 |
| Air Blank         |     | 0.000 | 17:21 |
| Control Test      |     | 0.079 | 17:21 |
| Air Blank         |     | 0.000 | 17:22 |
| Subject Sample    | # 1 | 0.231 | 17:24 |
| Breath Volume     |     | 1.375 | 17:24 |
| Air Blank         |     | 0.000 | 17:25 |
| Air Blank         |     | 0.000 | 17:27 |
| Subject Sample    | # 2 | 0.236 | 17:27 |
| Breath Volume     |     | 1.500 | 17:27 |
| Air Blank         |     | 0.000 | 17:28 |
| Control Test      |     | 0.077 | 17:28 |
| Air Blank         |     | 0.000 | 17:29 |
| Diagnostics Check |     | OK    | 17:29 |

---

Data Download Date/Time: 08/09/2018 01:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAYFIELD, MATTHEW, E  
BREATH TEST OPERATOR AGENCY: Lynn Haven PD  
ARRESTING OFFICER: MAYFIELD, MATTHEW  
ARRESTING OFFICER AGENCY: Lynn Haven PD  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: WILKINSON

DATE/TIME OF TEST: 07/07/2018 20:24:30

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:21 |                      |
| Air Blank         | 0.000 | 20:22 |                      |
| Control Test      | 0.079 | 20:22 |                      |
| Air Blank         | 0.000 | 20:22 |                      |
| Subject Sample    | # 1   | 20:23 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:23 |                      |
| Air Blank         | 0.000 | 20:23 |                      |
| Control Test      | 0.079 | 20:23 |                      |
| Air Blank         | 0.000 | 20:24 |                      |
| Diagnostics Check | OK    | 20:24 |                      |

---

Data Download Date/Time: 08/09/2018 01:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAYFIELD, MATTHEW, E  
BREATH TEST OPERATOR AGENCY: Lynn Haven PD  
ARRESTING OFFICER: MAYFIELD, MATTHEW  
ARRESTING OFFICER AGENCY: Lynn Haven PD  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: SHERRILL

DATE/TIME OF TEST: 07/08/2018 16:13:49

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 16:11 |                      |
| Air Blank         | 0.000 | 16:11 |                      |
| Control Test      | 0.079 | 16:11 |                      |
| Air Blank         | 0.000 | 16:12 |                      |
| Subject Sample    | # 1   | 16:12 | Subject Test Refused |
| Breath Volume     | 0.000 | 16:12 |                      |
| Air Blank         | 0.000 | 16:12 |                      |
| Control Test      | 0.080 | 16:13 |                      |
| Air Blank         | 0.000 | 16:13 |                      |
| Diagnostics Check | OK    | 16:13 |                      |

---

Data Download Date/Time: 08/09/2018 01:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUTTON,KERRY,D  
BREATH TEST OPERATOR AGENCY: Lynn Haven PD  
ARRESTING OFFICER: DUTTON,KERRY  
ARRESTING OFFICER AGENCY: Lynn Haven PD  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 13:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: NALL

DATE/TIME OF TEST: 07/18/2018 14:04:11

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 14:00 |                    |
| Air Blank         | 0.000 | 14:01 |                    |
| Control Test      | 0.079 | 14:01 |                    |
| Air Blank         | 0.000 | 14:02 |                    |
| Subject Sample    | # 1   | 14:02 | Interferent Detect |
| Breath Volume     | 1.656 | 14:02 |                    |
| Air Blank         | 0.000 | 14:03 |                    |
| Control Test      | 0.079 | 14:03 |                    |
| Air Blank         | 0.000 | 14:04 |                    |
| Diagnostics Check | OK    | 14:04 |                    |

---

Data Download Date/Time: 08/09/2018 01:49



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUTTON,KERRY,D  
BREATH TEST OPERATOR AGENCY: Lynn Haven PD  
ARRESTING OFFICER: DUTTON,KERRY  
ARRESTING OFFICER AGENCY: Lynn Haven PD  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 13:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: NALL

DATE/TIME OF TEST: 07/18/2018 14:09:16

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 14:06 |                    |
| Air Blank         | 0.000 | 14:06 |                    |
| Control Test      | 0.078 | 14:06 |                    |
| Air Blank         | 0.000 | 14:07 |                    |
| Subject Sample    | # 1   | 14:07 | Interferent Detect |
| Breath Volume     | 3.394 | 14:07 |                    |
| Air Blank         | 0.000 | 14:08 |                    |
| Control Test      | 0.079 | 14:08 |                    |
| Air Blank         | 0.000 | 14:09 |                    |
| Diagnostics Check | OK    | 14:09 |                    |

---

Data Download Date/Time: 08/09/2018 01:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAYFIELD, MATTHEW, E  
BREATH TEST OPERATOR AGENCY: Lynn Haven PD  
ARRESTING OFFICER: MAYFIELD, MATTHEW  
ARRESTING OFFICER AGENCY: Lynn Haven PD  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: DANNER

DATE/TIME OF TEST: 07/21/2018 02:31:56

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 02:25 |
| Air Blank         |     | 0.000 | 02:26 |
| Control Test      |     | 0.079 | 02:26 |
| Air Blank         |     | 0.000 | 02:26 |
| Subject Sample    | # 1 | 0.118 | 02:27 |
| Breath Volume     |     | 1.226 | 02:27 |
| Air Blank         |     | 0.000 | 02:28 |
| Air Blank         |     | 0.000 | 02:29 |
| Subject Sample    | # 2 | 0.122 | 02:30 |
| Breath Volume     |     | 1.203 | 02:30 |
| Air Blank         |     | 0.000 | 02:31 |
| Control Test      |     | 0.079 | 02:31 |
| Air Blank         |     | 0.000 | 02:31 |
| Diagnostics Check |     | OK    | 02:31 |

---

Data Download Date/Time: 08/09/2018 01:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAYFIELD, MATTHEW, E  
BREATH TEST OPERATOR AGENCY: Lynn Haven PD  
ARRESTING OFFICER: MAYFIELD, MATTHEW  
ARRESTING OFFICER AGENCY: Lynn Haven PD  
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: WHICHARD

DATE/TIME OF TEST: 07/21/2018 23:04:27

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:01 |                      |
| Air Blank         | 0.000 | 23:02 |                      |
| Control Test      | 0.078 | 23:02 |                      |
| Air Blank         | 0.000 | 23:02 |                      |
| Subject Sample    | # 1   | 23:03 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:03 |                      |
| Air Blank         | 0.000 | 23:03 |                      |
| Control Test      | 0.078 | 23:03 |                      |
| Air Blank         | 0.000 | 23:04 |                      |
| Diagnostics Check | OK    | 23:04 |                      |

---

Data Download Date/Time: 08/09/2018 01:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,DANNY,W  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: GRAY,DANNY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: KALUGDAN

DATE/TIME OF TEST: 07/01/2018 02:55:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:48 |
| Air Blank          | 0.000 | 02:49 |
| Control Test       | 0.079 | 02:49 |
| Air Blank          | 0.000 | 02:49 |
| Subject Sample # 1 | 0.042 | 02:50 |
| Breath Volume      | 3.410 | 02:50 |
| Air Blank          | 0.000 | 02:51 |
| Air Blank          | 0.000 | 02:53 |
| Subject Sample # 2 | 0.037 | 02:53 |
| Breath Volume      | 2.160 | 02:53 |
| Air Blank          | 0.000 | 02:54 |
| Control Test       | 0.080 | 02:54 |
| Air Blank          | 0.000 | 02:55 |
| Diagnostics Check  | OK    | 02:55 |

---

Data Download Date/Time: 07/18/2018 09:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINI,RYAN,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: LUNA,CHRISTIAN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ROME

DATE/TIME OF TEST: 07/04/2018 22:05:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:58 |
| Air Blank          | 0.000 | 21:59 |
| Control Test       | 0.080 | 21:59 |
| Air Blank          | 0.000 | 22:00 |
| Subject Sample # 1 | 0.172 | 22:00 |
| Breath Volume      | 1.664 | 22:00 |
| Air Blank          | 0.000 | 22:01 |
| Air Blank          | 0.000 | 22:03 |
| Subject Sample # 2 | 0.171 | 22:03 |
| Breath Volume      | 1.976 | 22:03 |
| Air Blank          | 0.000 | 22:04 |
| Control Test       | 0.080 | 22:04 |
| Air Blank          | 0.000 | 22:05 |
| Diagnostics Check  | OK    | 22:05 |

---

Data Download Date/Time: 07/18/2018 09:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATO,MARK,P  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 13:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:12

SUBJECT'S LAST NAME: LUNDSTROM

DATE/TIME OF TEST: 07/11/2018 14:59:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:53 |
| Air Blank          | 0.000 | 14:53 |
| Control Test       | 0.079 | 14:54 |
| Air Blank          | 0.000 | 14:54 |
| Subject Sample # 1 | 0.313 | 14:55 |
| Breath Volume      | 3.464 | 14:55 |
| Air Blank          | 0.000 | 14:55 |
| Air Blank          | 0.000 | 14:57 |
| Subject Sample # 2 | 0.312 | 14:58 |
| Breath Volume      | 3.558 | 14:58 |
| Air Blank          | 0.000 | 14:58 |
| Control Test       | 0.080 | 14:58 |
| Air Blank          | 0.000 | 14:59 |
| Diagnostics Check  | OK    | 14:59 |

---

Data Download Date/Time: 07/18/2018 09:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,ALFONSO,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: THOMAS,DANIEL  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: HEISS

DATE/TIME OF TEST: 07/13/2018 02:49:20

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:46 |                      |
| Air Blank         | 0.000 | 02:46 |                      |
| Control Test      | 0.080 | 02:47 |                      |
| Air Blank         | 0.000 | 02:47 |                      |
| Subject Sample    | # 1   | 02:47 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:47 |                      |
| Air Blank         | 0.000 | 02:48 |                      |
| Control Test      | 0.081 | 02:48 |                      |
| Air Blank         | 0.000 | 02:49 |                      |
| Diagnostics Check | OK    | 02:49 |                      |

---

Data Download Date/Time: 07/18/2018 09:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZARRILLO,N,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: LOOS,S  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: ROHDE

DATE/TIME OF TEST: 07/17/2018 23:48:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:41 |
| Air Blank          | 0.000 | 23:41 |
| Control Test       | 0.081 | 23:42 |
| Air Blank          | 0.000 | 23:42 |
| Subject Sample # 1 | 0.191 | 23:43 |
| Breath Volume      | 1.277 | 23:43 |
| Air Blank          | 0.000 | 23:44 |
| Air Blank          | 0.000 | 23:46 |
| Subject Sample # 2 | 0.194 | 23:46 |
| Breath Volume      | 1.273 | 23:46 |
| Air Blank          | 0.000 | 23:47 |
| Control Test       | 0.079 | 23:47 |
| Air Blank          | 0.000 | 23:48 |
| Diagnostics Check  | OK    | 23:48 |

---

Data Download Date/Time: 07/18/2018 09:02



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TOBLER,AMANDA,R  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: TOBLER,AMANDA  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/21/2018 04:04:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:57 |
| Air Blank          | 0.000 | 03:57 |
| Control Test       | 0.081 | 03:57 |
| Air Blank          | 0.000 | 03:58 |
| Subject Sample # 1 | 0.093 | 03:59 |
| Breath Volume      | 1.253 | 03:59 |
| Air Blank          | 0.000 | 04:00 |
| Air Blank          | 0.000 | 04:02 |
| Subject Sample # 2 | 0.091 | 04:02 |
| Breath Volume      | 1.835 | 04:02 |
| Air Blank          | 0.000 | 04:03 |
| Control Test       | 0.081 | 04:03 |
| Air Blank          | 0.000 | 04:04 |
| Diagnostics Check  | OK    | 04:04 |

---

Data Download Date/Time: 08/01/2018 09:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATO,MARK,P  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: SUPER,MICHAEL  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 11:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:14

SUBJECT'S LAST NAME: GREENE,II

DATE/TIME OF TEST: 07/22/2018 12:59:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 12:53 |
| Air Blank          | 0.000 | 12:53 |
| Control Test       | 0.081 | 12:53 |
| Air Blank          | 0.000 | 12:54 |
| Subject Sample # 1 | 0.000 | 12:55 |
| Breath Volume      | 4.347 | 12:55 |
| Air Blank          | 0.000 | 12:55 |
| Air Blank          | 0.000 | 12:57 |
| Subject Sample # 2 | 0.000 | 12:58 |
| Breath Volume      | 4.015 | 12:58 |
| Air Blank          | 0.000 | 12:58 |
| Control Test       | 0.081 | 12:58 |
| Air Blank          | 0.000 | 12:59 |
| Diagnostics Check  | OK    | 12:59 |

---

Data Download Date/Time: 08/01/2018 09:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON,SARAH,M  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: JOHNSON,SARAH  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: MISTRETТА

DATE/TIME OF TEST: 07/23/2018 02:03:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Control Test       | 0.081 | 01:58 |
| Air Blank          | 0.000 | 01:58 |
| Subject Sample # 1 | 0.033 | 01:59 |
| Breath Volume      | 3.371 | 01:59 |
| Air Blank          | 0.000 | 01:59 |
| Air Blank          | 0.000 | 02:01 |
| Subject Sample # 2 | 0.031 | 02:02 |
| Breath Volume      | 3.417 | 02:02 |
| Air Blank          | 0.000 | 02:02 |
| Control Test       | 0.080 | 02:03 |
| Air Blank          | 0.000 | 02:03 |
| Diagnostics Check  | OK    | 02:03 |

---

Data Download Date/Time: 08/01/2018 09:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLS,CHARLES,S  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MILLS,CHARLES  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 07/23/2018 17:30:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:24 |
| Air Blank          | 0.000 | 17:24 |
| Control Test       | 0.080 | 17:24 |
| Air Blank          | 0.000 | 17:25 |
| Subject Sample # 1 | 0.115 | 17:25 |
| Breath Volume      | 1.636 | 17:25 |
| Air Blank          | 0.000 | 17:26 |
| Air Blank          | 0.000 | 17:28 |
| Subject Sample # 2 | 0.113 | 17:28 |
| Breath Volume      | 1.726 | 17:28 |
| Air Blank          | 0.000 | 17:29 |
| Control Test       | 0.081 | 17:29 |
| Air Blank          | 0.000 | 17:30 |
| Diagnostics Check  | OK    | 17:30 |

---

Data Download Date/Time: 08/01/2018 09:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINI,RYAN,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MARTINI,RYAN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: OHIME

DATE/TIME OF TEST: 07/23/2018 22:22:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:16 |
| Air Blank          | 0.000 | 22:17 |
| Control Test       | 0.081 | 22:17 |
| Air Blank          | 0.000 | 22:17 |
| Subject Sample # 1 | 0.205 | 22:18 |
| Breath Volume      | 2.839 | 22:18 |
| Air Blank          | 0.000 | 22:19 |
| Air Blank          | 0.000 | 22:20 |
| Subject Sample # 2 | 0.200 | 22:21 |
| Breath Volume      | 3.390 | 22:21 |
| Air Blank          | 0.000 | 22:21 |
| Control Test       | 0.080 | 22:22 |
| Air Blank          | 0.000 | 22:22 |
| Diagnostics Check  | OK    | 22:22 |

---

Data Download Date/Time: 08/01/2018 09:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: CICCHINI,PAUL  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/24/2018 02:26:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:20 |
| Air Blank          | 0.000 | 02:20 |
| Control Test       | 0.081 | 02:21 |
| Air Blank          | 0.000 | 02:21 |
| Subject Sample # 1 | 0.136 | 02:22 |
| Breath Volume      | 3.320 | 02:22 |
| Air Blank          | 0.000 | 02:22 |
| Air Blank          | 0.000 | 02:24 |
| Subject Sample # 2 | 0.135 | 02:25 |
| Breath Volume      | 3.203 | 02:25 |
| Air Blank          | 0.000 | 02:25 |
| Control Test       | 0.080 | 02:26 |
| Air Blank          | 0.000 | 02:26 |
| Diagnostics Check  | OK    | 02:26 |

---

Data Download Date/Time: 08/01/2018 09:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINI,RYAN,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: CHESNUT,JONATHAN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: BEAVER

DATE/TIME OF TEST: 07/25/2018 01:13:01

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 01:06 |
| Air Blank         |     | 0.000 | 01:06 |
| Control Test      |     | 0.081 | 01:07 |
| Air Blank         |     | 0.000 | 01:07 |
| Subject Sample    | # 1 | 0.240 | 01:08 |
| Breath Volume     |     | 1.488 | 01:08 |
| Air Blank         |     | 0.000 | 01:08 |
| Air Blank         |     | 0.000 | 01:10 |
| Subject Sample    | # 2 | 0.222 | 01:11 |
| Breath Volume     |     | 1.179 | 01:11 |
| Air Blank         |     | 0.000 | 01:12 |
| Control Test      |     | 0.080 | 01:12 |
| Air Blank         |     | 0.000 | 01:12 |
| Diagnostics Check |     | OK    | 01:13 |

---

Data Download Date/Time: 08/01/2018 09:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUSKEVICH,CURT,E  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: SCARLATO,M  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 07/26/2018 00:32:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:25 |
| Air Blank          | 0.000 | 00:26 |
| Control Test       | 0.081 | 00:26 |
| Air Blank          | 0.000 | 00:26 |
| Subject Sample # 1 | 0.147 | 00:27 |
| Breath Volume      | 2.617 | 00:27 |
| Air Blank          | 0.000 | 00:28 |
| Air Blank          | 0.000 | 00:30 |
| Subject Sample # 2 | 0.144 | 00:30 |
| Breath Volume      | 3.195 | 00:30 |
| Air Blank          | 0.000 | 00:31 |
| Control Test       | 0.081 | 00:31 |
| Air Blank          | 0.000 | 00:31 |
| Diagnostics Check  | OK    | 00:32 |

---

Data Download Date/Time: 08/01/2018 09:11



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,DANNY,W  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: GRAY,DANNY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: AURA

DATE/TIME OF TEST: 07/28/2018 00:14:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:08 |
| Air Blank          | 0.000 | 00:09 |
| Control Test       | 0.079 | 00:09 |
| Air Blank          | 0.000 | 00:09 |
| Subject Sample # 1 | 0.189 | 00:10 |
| Breath Volume      | 2.394 | 00:10 |
| Air Blank          | 0.000 | 00:10 |
| Air Blank          | 0.000 | 00:12 |
| Subject Sample # 2 | 0.183 | 00:13 |
| Breath Volume      | 1.542 | 00:13 |
| Air Blank          | 0.000 | 00:14 |
| Control Test       | 0.081 | 00:14 |
| Air Blank          | 0.000 | 00:14 |
| Diagnostics Check  | OK    | 00:14 |

---

Data Download Date/Time: 08/01/2018 09:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAUSER,SGT,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: BIANCO,A  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: ORTA SUAREZ

DATE/TIME OF TEST: 07/28/2018 22:22:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:16 |
| Air Blank          | 0.000 | 22:16 |
| Control Test       | 0.080 | 22:16 |
| Air Blank          | 0.000 | 22:17 |
| Subject Sample # 1 | 0.243 | 22:17 |
| Breath Volume      | 2.277 | 22:17 |
| Air Blank          | 0.000 | 22:18 |
| Air Blank          | 0.000 | 22:20 |
| Subject Sample # 2 | 0.241 | 22:21 |
| Breath Volume      | 2.812 | 22:21 |
| Air Blank          | 0.000 | 22:21 |
| Control Test       | 0.080 | 22:22 |
| Air Blank          | 0.000 | 22:22 |
| Diagnostics Check  | OK    | 22:22 |

---

Data Download Date/Time: 08/01/2018 09:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: BARKLEY,JEFFREY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: KINNETT

DATE/TIME OF TEST: 07/29/2018 02:10:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:03 |
| Air Blank          | 0.000 | 02:04 |
| Control Test       | 0.081 | 02:04 |
| Air Blank          | 0.000 | 02:05 |
| Subject Sample # 1 | 0.159 | 02:05 |
| Breath Volume      | 2.742 | 02:05 |
| Air Blank          | 0.000 | 02:06 |
| Air Blank          | 0.000 | 02:08 |
| Subject Sample # 2 | 0.151 | 02:08 |
| Breath Volume      | 2.281 | 02:08 |
| Air Blank          | 0.000 | 02:09 |
| Control Test       | 0.080 | 02:09 |
| Air Blank          | 0.000 | 02:10 |
| Diagnostics Check  | OK    | 02:10 |

---

Data Download Date/Time: 08/01/2018 09:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MEADOWS,KELSEY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CHAVEZ MARRERO

DATE/TIME OF TEST: 07/29/2018 22:49:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:42 |
| Air Blank          | 0.000 | 22:42 |
| Control Test       | 0.080 | 22:42 |
| Air Blank          | 0.000 | 22:43 |
| Subject Sample # 1 | 0.183 | 22:44 |
| Breath Volume      | 1.839 | 22:44 |
| Air Blank          | 0.000 | 22:45 |
| Air Blank          | 0.000 | 22:46 |
| Subject Sample # 2 | 0.172 | 22:47 |
| Breath Volume      | 1.375 | 22:47 |
| Air Blank          | 0.000 | 22:48 |
| Control Test       | 0.080 | 22:48 |
| Air Blank          | 0.000 | 22:49 |
| Diagnostics Check  | OK    | 22:49 |

---

Data Download Date/Time: 08/01/2018 09:12