

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTAMIRANDA,ANTONIO,G
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: VILLANO,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 07/17/2018 05:30:44

Diagnostics Check	OK	05:24
Air Blank	0.000	05:24
Control Test	0.079	05:25
Air Blank	0.000	05:25
Subject Sample # 1	0.171	05:26
Breath Volume	1.757	05:26
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.171	05:29
Breath Volume	2.355	05:29
Air Blank	0.000	05:29
Control Test	0.078	05:30
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 07/21/2018 01:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRA,CORP,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SERRA,CORP
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 07/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: BEALL

DATE/TIME OF TEST: 07/21/2018 02:27:49

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.134	02:23
Breath Volume	2.847	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.133	02:26
Breath Volume	3.230	02:26
Air Blank	0.000	02:26
Control Test	0.078	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 08/13/2018 00:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: RODRIGUEZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BALTAZAR NICOLAS

DATE/TIME OF TEST: 07/01/2018 03:52:40

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.202	03:47
Breath Volume	1.843	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.209	03:51
Breath Volume	1.839	03:51
Air Blank	0.000	03:51
Control Test	0.080	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 07/10/2018 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MURGUIDO,ALEJANDRO
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: LOREDO

DATE/TIME OF TEST: 07/06/2018 23:35:24

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.132	23:30
Breath Volume	2.683	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.135	23:33
Breath Volume	2.390	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 07/10/2018 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: PEREZ GOMEZ

DATE/TIME OF TEST: 07/07/2018 02:35:27

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
Control Test	0.080	02:24	
Air Blank	0.000	02:24	
Subject Sample	# 1 0.105	02:27	Volume Not Met
Breath Volume	1.035	02:27	
Air Blank	0.000	02:28	
Air Blank	0.000	02:30	
Subject Sample	# 2 0.094	02:30	
Breath Volume	1.332	02:30	
Air Blank	0.000	02:31	
Air Blank	0.000	02:33	
Subject Sample	# 3	02:33	RFI Detect
Breath Volume	1.035	02:33	
Air Blank	0.000	02:34	
Control Test	0.078	02:34	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 07/10/2018 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: PEREZ GOMEZ

DATE/TIME OF TEST: 07/07/2018 02:44:18

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.104	02:39
Breath Volume	1.933	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.100	02:42
Breath Volume	1.500	02:42
Air Blank	0.000	02:43
Control Test	0.078	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 07/10/2018 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILPIN, SCHELISA,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ, RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: TUM

DATE/TIME OF TEST: 07/08/2018 00:01:24

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:51	
Control Test	0.080	23:52	
Air Blank	0.000	23:52	
Subject Sample	# 2 0.217	23:55	Volume Not Met
Breath Volume	0.000	23:55	
Air Blank	0.000	23:56	
Air Blank	0.000	23:58	
Subject Sample	# 1	00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:00	
Control Test	0.079	00:00	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 07/10/2018 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SARAVIA BANEGAS

DATE/TIME OF TEST: 07/08/2018 04:30:24

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:22	
Control Test	0.082	04:22	
Air Blank	0.000	04:23	
Subject Sample	# 1	04:26	No Sample Provided
Breath Volume	0.000	04:26	
Air Blank	0.000	04:26	
Air Blank	0.000	04:28	
Subject Sample	# 2	04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:29	
Control Test	0.080	04:29	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 07/10/2018 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: VILLARREAL,JOSE
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: MAYEN CONTRERAS

DATE/TIME OF TEST: 07/08/2018 06:06:17

Diagnostics Check	OK	06:01	
Air Blank	0.000	06:01	
Control Test	0.081	06:02	
Air Blank	0.000	06:02	
Subject Sample # 1	0.239	06:04	
Breath Volume	1.476	06:04	
Air Blank		06:04	RFI Detect
Air Blank	0.000	06:05	
Control Test	0.081	06:05	
Air Blank	0.000	06:06	
Diagnostics Check	OK	06:06	

Data Download Date/Time: 07/10/2018 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: VILLARREAL,JOSE
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: MAYEN CONTRERAS

DATE/TIME OF TEST: 07/08/2018 06:16:46

Diagnostics Check	OK	06:08
Air Blank	0.000	06:08
Control Test	0.079	06:09
Air Blank	0.000	06:09
Subject Sample # 1	0.224	06:10
Breath Volume	1.503	06:10
Air Blank	0.000	06:11
Air Blank	0.000	06:13
Subject Sample # 2	0.225	06:14
Breath Volume	2.980	06:14
Air Blank	0.000	06:15
Control Test	0.080	06:16
Air Blank	0.000	06:16
Diagnostics Check	OK	06:16

Data Download Date/Time: 07/10/2018 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 07/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: BEATTIE

DATE/TIME OF TEST: 07/14/2018 00:22:52

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.081	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.128	00:18
Breath Volume	1.292	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.130	00:21
Breath Volume	1.609	00:21
Air Blank	0.000	00:21
Control Test	0.080	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 08/25/2018 22:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: RODRIGUEZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 07/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: COBO CEDILLO

DATE/TIME OF TEST: 07/15/2018 04:51:08

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.080	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.196	04:45
Breath Volume	1.632	04:45
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.190	04:49
Breath Volume	1.273	04:49
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 08/25/2018 22:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILPIN, SCHELISA,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GILPIN, SCHELISA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 07/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: VASQUEZ RAMIREZ

DATE/TIME OF TEST: 07/15/2018 23:02:38

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.080	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.326	22:58
Breath Volume	1.386	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.314	23:00
Breath Volume	1.222	23:00
Air Blank	0.000	23:01
Control Test	0.079	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 08/25/2018 22:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: DELACRUZ,RICK
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 07/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 07/21/2018 03:39:59

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.224	03:35
Breath Volume	1.230	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.209	03:38
Breath Volume	1.335	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 08/25/2018 22:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILPIN, SCHELISA,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GILPIN, SCHELISA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 07/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: LAYNEZ PACHECO

DATE/TIME OF TEST: 07/24/2018 21:42:49

Diagnostics Check	OK	21:34
Air Blank	0.000	21:34
Control Test	0.080	21:35
Air Blank	0.000	21:35
Subject Sample # 1	0.220	21:36
Breath Volume	2.011	21:36
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.212	21:41
Breath Volume	1.937	21:41
Air Blank	0.000	21:41
Control Test	0.078	21:42
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 08/25/2018 22:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURGUIDO,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ROGERS,WAYNE
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 07/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: VENEGAS

DATE/TIME OF TEST: 07/28/2018 20:08:25

Diagnostics Check	OK	20:00
Air Blank	0.000	20:01
Control Test	0.081	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.190	20:03
Breath Volume	1.828	20:03
Air Blank	0.000	20:04
Air Blank	0.000	20:06
Subject Sample # 2	0.182	20:06
Breath Volume	1.871	20:06
Air Blank	0.000	20:07
Control Test	0.077	20:07
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 08/25/2018 22:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE, WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: COY

DATE/TIME OF TEST: 07/05/2018 00:17:03

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.077	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.116	00:12
Breath Volume	2.757	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.119	00:15
Breath Volume	2.144	00:15
Air Blank	0.000	00:16
Control Test	0.076	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:17

Data Download Date/Time: 07/25/2018 23:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE, WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: HEADLEY

DATE/TIME OF TEST: 07/07/2018 01:38:27

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.077	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.235	01:34
Breath Volume	1.757	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.240	01:36
Breath Volume	1.539	01:36
Air Blank	0.000	01:37
Control Test	0.077	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 07/25/2018 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: BARTH,T
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 07/18/2018 16:11:00

Diagnostics Check	OK	16:10	
Air Blank		16:10	RFI Detect
Air Blank		16:11	RFI Detect

Data Download Date/Time: 07/25/2018 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: BARTH,T
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 07/18/2018 16:13:36

Diagnostics Check	OK	16:12	
Air Blank	0.000	16:12	
Control Test	0.074	16:13	Control Outside Tolerance
Air Blank	0.000	16:13	

Data Download Date/Time: 07/25/2018 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: BARTH, T
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 07/18/2018 16:16:58

Diagnostics Check	OK	16:15	
Air Blank	0.000	16:16	
Control Test	0.074	16:16	Control Outside Tolerance
Air Blank	0.000	16:16	

Data Download Date/Time: 07/25/2018 23:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MASSEO, D
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/20/2018 19:29:27

Diagnostics Check	OK	19:22
Air Blank	0.000	19:22
Control Test	0.077	19:23
Air Blank	0.000	19:23
Subject Sample # 1	0.195	19:24
Breath Volume	3.902	19:24
Air Blank	0.000	19:25
Air Blank	0.000	19:27
Subject Sample # 2	0.197	19:27
Breath Volume	3.882	19:27
Air Blank	0.000	19:28
Control Test	0.075	19:28
Air Blank	0.000	19:29
Diagnostics Check	OK	19:29

Data Download Date/Time: 07/25/2018 23:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE, WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 07/20/2018 23:20:22

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.075	23:14
Air Blank	0.000	23:15
Subject Sample # 1	0.049	23:15
Breath Volume	1.285	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.056	23:18
Breath Volume	3.851	23:18
Air Blank	0.000	23:19
Control Test	0.076	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 07/25/2018 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN, GERALD, E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: DURYE, M
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BRADDOCK

DATE/TIME OF TEST: 07/05/2018 01:38:48

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:36	
Control Test	0.082	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 1	01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:37	
Control Test	0.081	01:38	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 07/06/2018 23:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 07/05/2018 02:18:35

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.081	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.258	02:14
Breath Volume	2.699	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.259	02:16
Breath Volume	2.511	02:16
Air Blank	0.000	02:17
Control Test	0.081	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 07/06/2018 23:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 07/14/2018 04:45:19

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:41	
Control Test	0.082	04:42	
Air Blank	0.000	04:42	
Subject Sample	# 1	04:43	Subject Test Refused
Breath Volume	0.000	04:43	
Air Blank	0.000	04:44	
Control Test	0.081	04:44	
Air Blank	0.000	04:45	
Diagnostics Check	OK	04:45	

Data Download Date/Time: 08/15/2018 23:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,JOHN,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: ORTIZ-GONZALEZ,JOSE
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: DAMIANI

DATE/TIME OF TEST: 07/15/2018 20:31:31

Diagnostics Check	OK	20:24
Air Blank	0.000	20:25
Control Test	0.081	20:25
Air Blank	0.000	20:25
Subject Sample # 1	0.053	20:27
Breath Volume	1.488	20:27
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.048	20:29
Breath Volume	1.140	20:29
Air Blank	0.000	20:30
Control Test	0.078	20:30
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

Data Download Date/Time: 08/15/2018 23:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,ALEC,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: ESTILL

DATE/TIME OF TEST: 07/16/2018 00:29:03

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.081	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:28	
Control Test	0.079	00:28	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:29	

Data Download Date/Time: 08/15/2018 23:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,JOHN,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: FLAMARY,KEITH
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 07/16/2018 01:16:35

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.082	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.145	01:12
Breath Volume	2.117	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.134	01:14
Breath Volume	1.468	01:14
Air Blank	0.000	01:15
Control Test	0.079	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 08/15/2018 23:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, WILLIAM, S
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MCKENZIE, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 07/17/2018 22:39:22

Diagnostics Check	OK	22:32
Air Blank	0.000	22:33
Control Test	0.082	22:33
Air Blank	0.000	22:34
Subject Sample # 1	0.116	22:34
Breath Volume	2.781	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.116	22:37
Breath Volume	2.792	22:37
Air Blank	0.000	22:38
Control Test	0.081	22:38
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 08/15/2018 23:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,JOHN,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: FLAMARY,KEITH
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 07/19/2018 01:47:32

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.082	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.122	01:42
Breath Volume	1.984	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.119	01:45
Breath Volume	1.464	01:45
Air Blank	0.000	01:46
Control Test	0.078	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 08/15/2018 23:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,ALEC,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: LEBO

DATE/TIME OF TEST: 07/19/2018 23:11:52

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.081	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.000	23:07
Breath Volume	2.527	23:07
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.000	23:10
Breath Volume	1.433	23:10
Air Blank	0.000	23:10
Control Test	0.081	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 08/15/2018 23:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, EDWIN,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RODRIGUEZ, EDWIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: FINDLEY

DATE/TIME OF TEST: 07/20/2018 22:47:50

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:33	
Control Test	0.081	22:33	
Air Blank	0.000	22:34	
Subject Sample	# 1 0.311	22:36	
Breath Volume	1.871	22:36	
Air Blank	0.000	22:37	
Air Blank	0.000	22:38	
Subject Sample	# 2 0.251	22:42	Volume Not Met
Breath Volume	0.046	22:42	
Air Blank	0.000	22:43	
Air Blank	0.000	22:44	
Subject Sample	# 3	22:46	Subject Test Refused
Breath Volume	0.914	22:46	
Air Blank	0.000	22:46	
Control Test	0.076	22:47	
Air Blank	0.000	22:47	
Diagnostics Check	OK	22:47	

Data Download Date/Time: 08/15/2018 23:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETTHAND,BREUNA,C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: LETTHAND,BREUNA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 07/22/2018 01:15:51

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.082	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.119	01:10
Breath Volume	2.007	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:12
Subject Sample # 2	0.128	01:14
Breath Volume	2.597	01:14
Air Blank	0.000	01:14
Control Test	0.080	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 08/15/2018 23:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN, GERALD, E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CREECH, J
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 07/24/2018 22:18:31

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:15	
Control Test	0.082	22:15	
Air Blank	0.000	22:16	
Subject Sample	# 1	22:17	Subject Test Refused
Breath Volume	0.000	22:17	
Air Blank	0.000	22:17	
Control Test	0.081	22:17	
Air Blank	0.000	22:18	
Diagnostics Check	OK	22:18	

Data Download Date/Time: 08/15/2018 23:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN, GERALD, E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PRINGLE, J SCOTT
ARRESTING OFFICER AGENCY: Umatilla PD
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/28/2018 05:36:01

Diagnostics Check	OK	05:29
Air Blank	0.000	05:30
Control Test	0.082	05:30
Air Blank	0.000	05:30
Subject Sample # 1	0.000	05:31
Breath Volume	2.019	05:31
Air Blank	0.000	05:31
Air Blank	0.000	05:33
Subject Sample # 2	0.000	05:34
Breath Volume	1.964	05:34
Air Blank	0.000	05:35
Control Test	0.082	05:35
Air Blank	0.000	05:35
Diagnostics Check	OK	05:36

Data Download Date/Time: 08/15/2018 23:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN, GERALD, E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN, GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 07/28/2018 05:49:38

Diagnostics Check	OK	05:43
Air Blank	0.000	05:43
Control Test	0.081	05:44
Air Blank	0.000	05:44
Subject Sample # 1	0.130	05:45
Breath Volume	2.562	05:45
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.122	05:48
Breath Volume	2.820	05:48
Air Blank	0.000	05:48
Control Test	0.080	05:49
Air Blank	0.000	05:49
Diagnostics Check	OK	05:49

Data Download Date/Time: 08/15/2018 23:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KETCHUM, CHARLES, A
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: APONTE, J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: CABALLERO GARCIA

DATE/TIME OF TEST: 07/28/2018 22:56:54

Diagnostics Check	OK	22:50	
Air Blank	0.000	22:51	
Control Test	0.082	22:51	
Air Blank	0.000	22:52	
Subject Sample # 1	0.115	22:52	
Breath Volume	2.582	22:52	
Air Blank	0.000	22:53	
Air Blank	0.000	22:55	
Subject Sample # 2	0.113	22:55	
Breath Volume	2.628	22:55	
Air Blank		22:56	RFI Detect
Air Blank	0.000	22:56	

Data Download Date/Time: 08/15/2018 23:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: LUU

DATE/TIME OF TEST: 07/01/2018 01:28:22

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.077	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.121	01:23
Breath Volume	1.937	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.118	01:26
Breath Volume	1.648	01:26
Air Blank	0.000	01:27
Control Test	0.077	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 07/26/2018 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: MECCA

DATE/TIME OF TEST: 07/10/2018 20:19:53

Diagnostics Check	OK	20:13
Air Blank	0.000	20:13
Control Test	0.078	20:14
Air Blank	0.000	20:14
Subject Sample # 1	0.244	20:14
Breath Volume	1.574	20:14
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.264	20:18
Breath Volume	1.570	20:18
Air Blank	0.000	20:18
Control Test	0.077	20:19
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 07/26/2018 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKIN,SEAN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WILKIN,SEAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: TIDD

DATE/TIME OF TEST: 07/12/2018 03:06:02

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.077	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.246	03:01
Breath Volume	4.707	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.245	03:04
Breath Volume	4.289	03:04
Air Blank	0.000	03:05
Control Test	0.077	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:06

Data Download Date/Time: 07/26/2018 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,DAVID,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: BAYLY,BLAKE
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SMELCER

DATE/TIME OF TEST: 07/03/2018 23:15:28

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.081	23:09
Air Blank	0.000	23:10
Subject Sample # 1	0.050	23:10
Breath Volume	2.257	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.052	23:13
Breath Volume	2.726	23:13
Air Blank	0.000	23:14
Control Test	0.081	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 07/27/2018 21:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/22/2018 23:08:13

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.077	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.162	23:03
Breath Volume	1.726	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.158	23:06
Breath Volume	1.792	23:06
Air Blank	0.000	23:07
Control Test	0.076	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 08/21/2018 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: VAR

DATE/TIME OF TEST: 07/01/2018 00:32:06

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.225	00:25
Breath Volume	2.218	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.220	00:30
Breath Volume	1.843	00:30
Air Blank	0.000	00:31
Control Test	0.077	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WITT,TYLER
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: CESTARO

DATE/TIME OF TEST: 07/01/2018 03:05:33

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.159	02:58
Breath Volume	1.582	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.156	03:03
Breath Volume	1.582	03:03
Air Blank	0.000	03:04
Control Test	0.077	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: CATALAN CORTES

DATE/TIME OF TEST: 07/01/2018 05:27:03

Diagnostics Check	OK	05:20
Air Blank	0.000	05:20
Control Test	0.079	05:21
Air Blank	0.000	05:21
Subject Sample # 1	0.231	05:22
Breath Volume	1.964	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:24
Subject Sample # 2	0.225	05:25
Breath Volume	2.593	05:25
Air Blank	0.000	05:26
Control Test	0.078	05:26
Air Blank	0.000	05:26
Diagnostics Check	OK	05:27

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 07/02/2018 01:43:06

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.079	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.172	01:38
Breath Volume	2.355	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.166	01:41
Breath Volume	1.367	01:41
Air Blank	0.000	01:42
Control Test	0.079	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 07/05/2018 00:30:07

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.121	00:25
Breath Volume	1.925	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.130	00:28
Breath Volume	1.824	00:28
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: WESTBROOK JR

DATE/TIME OF TEST: 07/05/2018 03:52:37

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.106	03:46
Breath Volume	1.574	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.094	03:51
Breath Volume	1.289	03:51
Air Blank	0.000	03:51
Control Test	0.080	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-
RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male

OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 07/05/2018 05:27:16

Diagnostics Check	OK	05:20
Air Blank	0.000	05:20
Control Test	0.080	05:21
Air Blank	0.000	05:21
Subject Sample # 1	0.189	05:22
Breath Volume	2.320	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:24
Subject Sample # 2	0.171	05:25
Breath Volume	1.121	05:25
Air Blank	0.000	05:26
Control Test	0.079	05:26
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-
RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male

OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: ADAMKO

DATE/TIME OF TEST: 07/06/2018 02:23:15

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.081	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.194	02:18
Breath Volume	1.710	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.184	02:21
Breath Volume	1.960	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCMAHON,G
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HANCHARD

DATE/TIME OF TEST: 07/06/2018 03:25:56

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.117	03:21
Breath Volume	1.863	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:23
Subject Sample # 2	0.118	03:24
Breath Volume	2.871	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCCLUSKEY,RYAN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 07/06/2018 03:47:23

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.179	03:42
Breath Volume	3.089	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.172	03:45
Breath Volume	3.699	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: GALEANO

DATE/TIME OF TEST: 07/07/2018 03:53:20

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.201	03:48
Breath Volume	2.144	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.201	03:51
Breath Volume	3.449	03:51
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: LEAHY

DATE/TIME OF TEST: 07/08/2018 00:42:01

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.116	00:36
Breath Volume	1.769	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.125	00:40
Breath Volume	2.750	00:40
Air Blank	0.000	00:41
Control Test	0.080	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:42

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DONDE,JAY
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 07/08/2018 01:36:13

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.184	01:31
Breath Volume	1.160	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.184	01:34
Breath Volume	1.488	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: EUGENE

DATE/TIME OF TEST: 07/08/2018 02:46:57

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.110	02:42
Breath Volume	1.695	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.111	02:45
Breath Volume	3.074	02:45
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WITT,TYLER
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: BONAVIDA JR

DATE/TIME OF TEST: 07/08/2018 03:20:44

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.053	03:15
Breath Volume	1.156	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.053	03:19
Breath Volume	1.511	03:19
Air Blank	0.000	03:19
Control Test	0.079	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AIME,JEAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: CORONA PAREDES

DATE/TIME OF TEST: 07/08/2018 10:47:33

Diagnostics Check	OK	10:40
Air Blank	0.000	10:41
Control Test	0.080	10:41
Air Blank	0.000	10:42
Subject Sample # 1	0.140	10:42
Breath Volume	2.253	10:42
Air Blank	0.000	10:43
Air Blank	0.000	10:45
Subject Sample # 2	0.151	10:45
Breath Volume	3.375	10:45
Air Blank	0.000	10:46
Control Test	0.080	10:46
Air Blank	0.000	10:47
Diagnostics Check	OK	10:47

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: VALDERRAMA

DATE/TIME OF TEST: 07/09/2018 01:35:14

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.080	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.245	01:30
Breath Volume	1.832	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.244	01:33
Breath Volume	1.761	01:33
Air Blank	0.000	01:34
Control Test	0.079	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:18

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/09/2018 06:50:04

Diagnostics Check	OK	06:47	
Air Blank	0.000	06:47	
Control Test	0.080	06:48	
Air Blank	0.000	06:48	
Subject Sample	# 1	06:48	Subject Test Refused
Breath Volume	0.000	06:48	
Air Blank	0.000	06:49	
Control Test	0.079	06:49	
Air Blank	0.000	06:49	
Diagnostics Check	OK	06:50	

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: WALDRON

DATE/TIME OF TEST: 07/09/2018 23:43:53

Diagnostics Check	OK	23:37
Air Blank	0.000	23:38
Control Test	0.080	23:38
Air Blank	0.000	23:39
Subject Sample # 1	0.000	23:39
Breath Volume	1.996	23:39
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.000	23:42
Breath Volume	2.882	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WAGNER,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 07/10/2018 03:45:26

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.113	03:40
Breath Volume	1.832	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.108	03:43
Breath Volume	1.933	03:43
Air Blank	0.000	03:44
Control Test	0.079	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-
RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male

OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 07/12/2018 21:15:03

Diagnostics Check	OK	21:07
Air Blank	0.000	21:08
Control Test	0.080	21:08
Air Blank	0.000	21:09
Subject Sample # 1	0.000	21:10
Breath Volume	1.738	21:10
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample # 2	0.000	21:13
Breath Volume	1.339	21:13
Air Blank	0.000	21:14
Control Test	0.081	21:14
Air Blank	0.000	21:14
Diagnostics Check	OK	21:15

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BOOGREN

DATE/TIME OF TEST: 07/13/2018 02:30:43

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.080	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.106	02:26
Breath Volume	3.089	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.111	02:29
Breath Volume	3.324	02:29
Air Blank	0.000	02:29
Control Test	0.080	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: JONES III

DATE/TIME OF TEST: 07/13/2018 05:05:04

Diagnostics Check	OK	04:58
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.191	05:00
Breath Volume	1.363	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:02
Subject Sample # 2	0.183	05:03
Breath Volume	1.437	05:03
Air Blank	0.000	05:04
Control Test	0.080	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:05

Data Download Date/Time: 07/13/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: QUAILE

DATE/TIME OF TEST: 07/01/2018 01:45:59

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.079	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.254	01:41
Breath Volume	1.292	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:43
Subject Sample # 2	0.251	01:44
Breath Volume	1.652	01:44
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 07/13/2018 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: FURROUGH

DATE/TIME OF TEST: 07/01/2018 03:55:48

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.139	03:51
Breath Volume	2.062	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.136	03:54
Breath Volume	1.804	03:54
Air Blank	0.000	03:54
Control Test	0.080	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 07/13/2018 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: CHASE

DATE/TIME OF TEST: 07/01/2018 05:03:31

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.080	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.130	04:58
Breath Volume	1.691	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:00
Subject Sample # 2	0.132	05:01
Breath Volume	2.656	05:01
Air Blank	0.000	05:02
Control Test	0.079	05:02
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 07/13/2018 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BRODBECK,JAMES
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: NIELSEN

DATE/TIME OF TEST: 07/01/2018 22:05:58

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.080	22:00
Air Blank	0.000	22:00
Subject Sample # 1	0.188	22:01
Breath Volume	4.457	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:03
Subject Sample # 2	0.189	22:04
Breath Volume	4.171	22:04
Air Blank	0.000	22:05
Control Test	0.079	22:05
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 07/13/2018 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DOMINGUEZ,JAYMES
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/02/2018 01:05:55

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.080	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.218	01:01
Breath Volume	1.460	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.230	01:04
Breath Volume	1.851	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:17

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 07/04/2018 15:54:39

Diagnostics Check	OK	15:44
Air Blank	0.000	15:44
Control Test	0.079	15:45
Air Blank	0.000	15:45
Subject Sample # 1	0.227	15:47
Breath Volume	1.285	15:47
Air Blank	0.000	15:48
Air Blank	0.000	15:50
Subject Sample # 2	0.245	15:52
Breath Volume	1.781	15:52
Air Blank	0.000	15:53
Control Test	0.079	15:53
Air Blank	0.000	15:54
Diagnostics Check	OK	15:54

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: RANDALL

DATE/TIME OF TEST: 07/04/2018 23:15:46

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.179	23:10
Breath Volume	1.355	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.182	23:14
Breath Volume	1.683	23:14
Air Blank	0.000	23:14
Control Test	0.080	23:15
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VALERIO,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 07/05/2018 00:56:20

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.079	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.276	00:51
Breath Volume	2.613	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.286	00:54
Breath Volume	2.906	00:54
Air Blank	0.000	00:55
Control Test	0.078	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: BOWMAN

DATE/TIME OF TEST: 07/05/2018 02:43:47

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.079	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.158	02:39
Breath Volume	1.941	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.157	02:42
Breath Volume	2.441	02:42
Air Blank	0.000	02:42
Control Test	0.079	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BEROSIK

DATE/TIME OF TEST: 07/06/2018 03:10:24

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.217	03:05
Breath Volume	1.187	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.205	03:08
Breath Volume	1.617	03:08
Air Blank	0.000	03:09
Control Test	0.079	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: RUSSO

DATE/TIME OF TEST: 07/06/2018 04:02:24

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.063	03:58
Breath Volume	1.742	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.062	04:00
Breath Volume	1.578	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HOBBY,MATTHEW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: ANTONUCCI-CENTI

DATE/TIME OF TEST: 07/06/2018 05:10:34

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.080	05:04
Air Blank	0.000	05:04
Subject Sample # 1	0.000	05:05
Breath Volume	2.261	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.000	05:09
Breath Volume	1.429	05:09
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MITCHELL,CEDRIC
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: GAZDICK JR

DATE/TIME OF TEST: 07/07/2018 01:45:45

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.226	01:41
Breath Volume	3.964	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.229	01:44
Breath Volume	4.007	01:44
Air Blank	0.000	01:44
Control Test	0.079	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: CAYO

DATE/TIME OF TEST: 07/07/2018 23:45:30

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.208	23:36
Breath Volume	1.996	23:36
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.185	23:40
Breath Volume	1.527	23:40
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 3	0.176	23:43
Breath Volume	2.250	23:43
Air Blank	0.000	23:44
Control Test	0.079	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BRODBECK,JAMES
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: METROVICH

DATE/TIME OF TEST: 07/08/2018 21:51:53

Diagnostics Check	OK	21:41
Air Blank	0.000	21:42
Control Test	0.080	21:42
Air Blank	0.000	21:43
Subject Sample # 1	0.166	21:44
Breath Volume	1.230	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.164	21:50
Breath Volume	1.167	21:50
Air Blank	0.000	21:50
Control Test	0.079	21:51
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLOWERS,GARY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BURTON

DATE/TIME OF TEST: 07/09/2018 01:13:42

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:10	
Control Test	0.079	01:11	
Air Blank	0.000	01:11	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:12	
Control Test	0.080	01:13	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: MCCONNELL

DATE/TIME OF TEST: 07/09/2018 01:33:56

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.080	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.180	01:29
Breath Volume	1.222	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:31
Subject Sample # 2	0.186	01:32
Breath Volume	1.234	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: YAGUAL MORAN

DATE/TIME OF TEST: 07/09/2018 02:19:22

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:06	
Control Test	0.080	02:07	
Air Blank	0.000	02:07	
Subject Sample	# 1 0.164	02:10	Volume Not Met
Breath Volume	0.941	02:10	
Air Blank	0.000	02:11	
Air Blank	0.000	02:13	
Subject Sample	# 2 0.178	02:14	
Breath Volume	1.765	02:14	
Air Blank	0.000	02:14	
Air Blank	0.000	02:16	
Subject Sample	# 3 0.177	02:17	
Breath Volume	1.167	02:17	
Air Blank	0.000	02:18	
Control Test	0.079	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ALDOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MENDEZ GALINDO

DATE/TIME OF TEST: 07/09/2018 04:40:45

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.079	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.103	04:36
Breath Volume	1.738	04:36
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.098	04:39
Breath Volume	2.269	04:39
Air Blank	0.000	04:39
Control Test	0.080	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,S,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,DAYRON
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/09/2018 21:20:07

Diagnostics Check	OK	21:13
Air Blank	0.000	21:13
Control Test	0.080	21:14
Air Blank	0.000	21:14
Subject Sample # 1	0.141	21:15
Breath Volume	2.300	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.139	21:18
Breath Volume	2.062	21:18
Air Blank	0.000	21:19
Control Test	0.079	21:19
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 07/10/2018 05:06:22

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.080	05:00
Air Blank	0.000	05:01
Subject Sample # 1	0.105	05:01
Breath Volume	1.273	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.103	05:04
Breath Volume	2.632	05:04
Air Blank	0.000	05:05
Control Test	0.080	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGUERO,CARLOS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: CRAIG

DATE/TIME OF TEST: 07/10/2018 23:21:01

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.080	23:14
Air Blank	0.000	23:15
Subject Sample # 1	0.147	23:15
Breath Volume	2.023	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.158	23:19
Breath Volume	3.460	23:19
Air Blank	0.000	23:20
Control Test	0.079	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:21

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: PIMENTEL

DATE/TIME OF TEST: 07/11/2018 03:40:33

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.161	03:35
Breath Volume	1.328	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.159	03:38
Breath Volume	1.820	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JARRELL,THOMAS
ARRESTING OFFICER AGENCY: Tequesta PD
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 07/11/2018 19:52:23

Diagnostics Check	OK	19:45
Air Blank	0.000	19:45
Control Test	0.079	19:45
Air Blank	0.000	19:46
Subject Sample # 1	0.305	19:47
Breath Volume	1.539	19:47
Air Blank	0.000	19:48
Air Blank	0.000	19:50
Subject Sample # 2	0.313	19:50
Breath Volume	1.875	19:50
Air Blank	0.000	19:51
Control Test	0.078	19:51
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 07/13/2018 03:45:58

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.079	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.196	03:40
Breath Volume	1.785	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.192	03:44
Breath Volume	1.421	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 07/13/2018 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DOMINGUEZ,JAYMES
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: COLAIACOVO

DATE/TIME OF TEST: 07/13/2018 20:07:29

Diagnostics Check	OK	20:00
Air Blank	0.000	20:00
Control Test	0.080	20:00
Air Blank	0.000	20:01
Subject Sample # 1	0.266	20:02
Breath Volume	2.519	20:02
Air Blank	0.000	20:02
Air Blank	0.000	20:04
Subject Sample # 2	0.275	20:05
Breath Volume	2.652	20:05
Air Blank	0.000	20:06
Control Test	0.079	20:06
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 08/10/2018 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 07/13/2018 20:38:23

Diagnostics Check	OK	20:31
Air Blank	0.000	20:32
Control Test	0.079	20:32
Air Blank	0.000	20:33
Subject Sample # 1	0.000	20:33
Breath Volume	3.117	20:33
Air Blank	0.000	20:34
Air Blank	0.000	20:36
Subject Sample # 2	0.000	20:37
Breath Volume	3.363	20:37
Air Blank	0.000	20:37
Control Test	0.079	20:37
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 08/10/2018 16:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: ESKRIDGE

DATE/TIME OF TEST: 07/14/2018 00:31:51

Diagnostics Check	OK	00:24
Air Blank	0.000	00:25
Control Test	0.080	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.142	00:26
Breath Volume	2.968	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.136	00:30
Breath Volume	2.835	00:30
Air Blank	0.000	00:30
Control Test	0.079	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 08/10/2018 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROCHA,LUIS
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: MANEAU

DATE/TIME OF TEST: 07/14/2018 02:35:48

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.134	02:29
Breath Volume	1.187	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.124	02:34
Breath Volume	1.250	02:34
Air Blank	0.000	02:34
Control Test	0.079	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 08/10/2018 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALLEN,NATHAN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 07/14/2018 08:29:53

Diagnostics Check	OK	08:22
Air Blank	0.000	08:22
Control Test	0.080	08:23
Air Blank	0.000	08:23
Subject Sample # 1	0.120	08:24
Breath Volume	1.105	08:24
Air Blank	0.000	08:25
Air Blank	0.000	08:27
Subject Sample # 2	0.122	08:28
Breath Volume	1.484	08:28
Air Blank	0.000	08:28
Control Test	0.078	08:29
Air Blank	0.000	08:29
Diagnostics Check	OK	08:29

Data Download Date/Time: 08/10/2018 16:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HUDSON,DILLAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: BUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: BUENO

DATE/TIME OF TEST: 07/14/2018 22:52:21

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.079	22:40	
Air Blank	0.000	22:40	
Subject Sample	# 1 0.101	22:43	Volume Not Met
Breath Volume	0.015	22:43	
Air Blank	0.000	22:44	
Air Blank	0.000	22:46	
Subject Sample	# 2 0.124	22:47	
Breath Volume	2.554	22:47	
Air Blank	0.000	22:48	
Air Blank	0.000	22:49	
Subject Sample	# 3 0.126	22:50	
Breath Volume	2.945	22:50	
Air Blank	0.000	22:51	
Control Test	0.078	22:51	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 08/10/2018 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/17/2018 00:51:20

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.079	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.167	00:46
Breath Volume	1.515	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:48
Subject Sample # 2	0.165	00:49
Breath Volume	1.148	00:49
Air Blank	0.000	00:50
Control Test	0.078	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 08/10/2018 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GERMAIN,F
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MARCOS

DATE/TIME OF TEST: 07/17/2018 03:41:52

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.079	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.200	03:37
Breath Volume	1.546	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:39
Subject Sample # 2	0.200	03:40
Breath Volume	2.238	03:40
Air Blank	0.000	03:40
Control Test	0.079	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 08/10/2018 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LUCIANO,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 07/18/2018 01:57:23

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.079	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.165	01:53
Breath Volume	2.308	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.173	01:55
Breath Volume	1.710	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 08/10/2018 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-
RODRIGUEZ,LENNDEY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male

OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 07/18/2018 03:53:19

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.200	03:48
Breath Volume	2.917	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.208	03:51
Breath Volume	3.921	03:51
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 08/10/2018 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 07/20/2018 01:03:30

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.209	00:58
Breath Volume	1.667	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.204	01:01
Breath Volume	1.757	01:01
Air Blank	0.000	01:02
Control Test	0.079	01:02
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 08/10/2018 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MACCAROE,DAVID
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: FERRIER

DATE/TIME OF TEST: 07/20/2018 18:57:04

Diagnostics Check	OK	18:50
Air Blank	0.000	18:50
Control Test	0.079	18:51
Air Blank	0.000	18:51
Subject Sample # 1	0.079	18:52
Breath Volume	1.191	18:52
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.081	18:55
Breath Volume	1.468	18:55
Air Blank	0.000	18:56
Control Test	0.079	18:56
Air Blank	0.000	18:56
Diagnostics Check	OK	18:57

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: SCHERER

DATE/TIME OF TEST: 07/21/2018 03:09:04

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.154	03:04
Breath Volume	2.191	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:06
Subject Sample # 2	0.150	03:07
Breath Volume	1.343	03:07
Air Blank	0.000	03:08
Control Test	0.077	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:09

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: BOWMAN

DATE/TIME OF TEST: 07/21/2018 04:42:14

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.079	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.217	04:37
Breath Volume	1.289	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.207	04:40
Breath Volume	1.359	04:40
Air Blank	0.000	04:41
Control Test	0.079	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MACKEY,JAMES
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: MIGUEL PASCUAL

DATE/TIME OF TEST: 07/21/2018 19:14:30

Diagnostics Check	OK	19:06
Air Blank	0.000	19:06
Control Test	0.079	19:07
Air Blank	0.000	19:07
Subject Sample # 1	0.175	19:09
Breath Volume	2.417	19:09
Air Blank	0.000	19:10
Air Blank	0.000	19:12
Subject Sample # 2	0.165	19:12
Breath Volume	1.753	19:12
Air Blank	0.000	19:13
Control Test	0.078	19:13
Air Blank	0.000	19:14
Diagnostics Check	OK	19:14

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LIMING,TIFANI
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SANTOS ANDRES

DATE/TIME OF TEST: 07/22/2018 02:49:57

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.217	02:44
Breath Volume	1.660	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.233	02:48
Breath Volume	2.746	02:48
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DILLARD,DANIEL
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: PALACIOS

DATE/TIME OF TEST: 07/22/2018 06:25:41

Diagnostics Check	OK	06:18
Air Blank	0.000	06:19
Control Test	0.079	06:19
Air Blank	0.000	06:20
Subject Sample # 1	0.191	06:20
Breath Volume	1.585	06:20
Air Blank	0.000	06:21
Air Blank	0.000	06:23
Subject Sample # 2	0.193	06:24
Breath Volume	2.230	06:24
Air Blank	0.000	06:24
Control Test	0.078	06:25
Air Blank	0.000	06:25
Diagnostics Check	OK	06:25

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ,ROBERTO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: NEWBOLD

DATE/TIME OF TEST: 07/23/2018 01:19:38

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.200	01:14
Breath Volume	1.582	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.201	01:17
Breath Volume	1.167	01:17
Air Blank	0.000	01:18
Control Test	0.078	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 07/25/2018 02:41:44

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.080	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.218	02:36
Breath Volume	1.746	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:38
Subject Sample # 2	0.211	02:40
Breath Volume	1.578	02:40
Air Blank	0.000	02:40
Control Test	0.079	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GILLINGS,DAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/25/2018 08:45:51

Diagnostics Check	OK	08:38
Air Blank	0.000	08:38
Control Test	0.079	08:39
Air Blank	0.000	08:39
Subject Sample # 1	0.043	08:40
Breath Volume	1.199	08:40
Air Blank	0.000	08:41
Air Blank	0.000	08:43
Subject Sample # 2	0.043	08:44
Breath Volume	1.253	08:44
Air Blank	0.000	08:44
Control Test	0.079	08:45
Air Blank	0.000	08:45
Diagnostics Check	OK	08:45

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: WEINSTEI

DATE/TIME OF TEST: 07/26/2018 20:12:17

Diagnostics Check	OK	20:04
Air Blank	0.000	20:04
Control Test	0.080	20:04
Air Blank	0.000	20:05
Subject Sample # 1	0.204	20:06
Breath Volume	1.230	20:06
Air Blank	0.000	20:07
Air Blank	0.000	20:09
Subject Sample # 2	0.198	20:10
Breath Volume	1.945	20:10
Air Blank	0.000	20:11
Control Test	0.079	20:11
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HUDSON,DILLAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 07/27/2018 00:54:35

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.080	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.049	00:50
Breath Volume	2.796	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.047	00:53
Breath Volume	2.886	00:53
Air Blank	0.000	00:53
Control Test	0.079	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 07/27/2018 02:38:42

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.080	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.137	02:33
Breath Volume	1.183	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.147	02:37
Breath Volume	1.253	02:37
Air Blank	0.000	02:37
Control Test	0.079	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ACOSTA REYES

DATE/TIME OF TEST: 07/27/2018 03:14:25

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.078	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.152	03:09
Breath Volume	2.019	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.149	03:12
Breath Volume	2.910	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: BONILLA ESPINO

DATE/TIME OF TEST: 07/27/2018 05:22:22

Diagnostics Check	OK	05:15
Air Blank	0.000	05:15
Control Test	0.079	05:16
Air Blank	0.000	05:16
Subject Sample # 1	0.178	05:17
Breath Volume	2.058	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.187	05:20
Breath Volume	2.285	05:20
Air Blank	0.000	05:21
Control Test	0.078	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-
RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female

OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: VAIL LOPEZ

DATE/TIME OF TEST: 07/28/2018 01:30:05

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.226	01:25
Breath Volume	2.183	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.238	01:28
Breath Volume	1.679	01:28
Air Blank	0.000	01:29
Control Test	0.077	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:30

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,DAYRON
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: JEAN BAPTISTE

DATE/TIME OF TEST: 07/28/2018 02:56:37

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.078	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.171	02:51
Breath Volume	2.062	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.163	02:54
Breath Volume	1.566	02:54
Air Blank	0.000	02:55
Control Test	0.078	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: ARIAS GARCIA

DATE/TIME OF TEST: 07/28/2018 03:42:58

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.222	03:38
Breath Volume	2.480	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.225	03:41
Breath Volume	2.628	03:41
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/28/2018 22:29:59

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.079	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.126	22:25
Breath Volume	1.503	22:25
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.124	22:28
Breath Volume	1.234	22:28
Air Blank	0.000	22:29
Control Test	0.079	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: REMMERDEN

DATE/TIME OF TEST: 07/29/2018 04:13:50

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.165	04:09
Breath Volume	2.601	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.165	04:12
Breath Volume	2.582	04:12
Air Blank	0.000	04:12
Control Test	0.079	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: D'ANDREA,ADAM
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:13

SUBJECT'S LAST NAME: TOMAS RAMIREZ

DATE/TIME OF TEST: 07/29/2018 08:07:17

Diagnostics Check	OK	07:58
Air Blank	0.000	07:59
Control Test	0.079	07:59
Air Blank	0.000	08:00
Subject Sample # 1	0.217	08:01
Breath Volume	2.109	08:01
Air Blank	0.000	08:02
Air Blank	0.000	08:03
Subject Sample # 2	0.199	08:05
Breath Volume	2.042	08:05
Air Blank	0.000	08:06
Control Test	0.079	08:06
Air Blank	0.000	08:07
Diagnostics Check	OK	08:07

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: CRUZ ROMERO

DATE/TIME OF TEST: 07/30/2018 01:08:13

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.079	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.122	01:03
Breath Volume	1.289	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.127	01:06
Breath Volume	1.945	01:06
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MAIA,GUILHERME
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:23

SUBJECT'S LAST NAME: STROUSE

DATE/TIME OF TEST: 07/30/2018 16:59:20

Diagnostics Check	OK	16:52
Air Blank	0.000	16:52
Control Test	0.080	16:53
Air Blank	0.000	16:53
Subject Sample # 1	0.000	16:54
Breath Volume	1.542	16:54
Air Blank	0.000	16:54
Air Blank	0.000	16:56
Subject Sample # 2	0.000	16:57
Breath Volume	1.734	16:57
Air Blank	0.000	16:58
Control Test	0.079	16:58
Air Blank	0.000	16:59
Diagnostics Check	OK	16:59

Data Download Date/Time: 08/10/2018 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: EINFALT

DATE/TIME OF TEST: 07/13/2018 01:40:52

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.078	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.189	01:36
Breath Volume	3.339	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:38
Subject Sample # 2	0.176	01:39
Breath Volume	2.457	01:39
Air Blank	0.000	01:39
Control Test	0.082	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 07/18/2018 19:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/14/2018 02:18:53

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.077	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.188	02:14
Breath Volume	3.042	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.185	02:17
Breath Volume	2.996	02:17
Air Blank	0.000	02:17
Control Test	0.081	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 07/18/2018 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: BRISCOE

DATE/TIME OF TEST: 07/15/2018 02:38:55

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.078	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.109	02:34
Breath Volume	1.421	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.110	02:37
Breath Volume	1.308	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 07/18/2018 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: FOURAKER

DATE/TIME OF TEST: 07/19/2018 01:41:58

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.079	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.054	01:37
Breath Volume	2.003	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.051	01:40
Breath Volume	1.613	01:40
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 07/31/2018 21:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BELLEHUMEUR

DATE/TIME OF TEST: 07/20/2018 00:50:32

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:43	
Control Test	0.077	00:44	
Air Blank	0.000	00:44	
Subject Sample	# 1 0.256	00:45	
Breath Volume	2.656	00:45	
Air Blank	0.000	00:45	
Air Blank	0.000	00:47	
Subject Sample	# 2 0.263	00:48	
Breath Volume	2.261	00:48	
Air Blank	0.000	00:49	
Control Test	0.080	00:49	
Air Blank	0.000	00:49	
Diagnostics Check	Fail	00:50	Analytical Stability Test
Air Blank	0.000	00:50	

Data Download Date/Time: 07/31/2018 21:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BELLEHUMEUR

DATE/TIME OF TEST: 07/20/2018 00:53:15

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:52	
Control Test	0.081	00:52	Interferent Detect
Air Blank	0.000	00:53	

Data Download Date/Time: 07/31/2018 21:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 07/21/2018 01:15:06

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.076	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.127	01:10
Breath Volume	4.542	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:12
Subject Sample # 2	0.125	01:13
Breath Volume	4.570	01:13
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 07/31/2018 21:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHILLER,E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: AOLE AOLE COPP

DATE/TIME OF TEST: 07/13/2018 04:15:41

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.081	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.000	04:11
Breath Volume	1.566	04:11
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.000	04:14
Breath Volume	2.042	04:14
Air Blank	0.000	04:14
Control Test	0.081	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: GONCHAROVA

DATE/TIME OF TEST: 07/15/2018 01:43:32

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.081	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.228	01:38
Breath Volume	1.375	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.221	01:41
Breath Volume	1.203	01:41
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: LOPEZ LLANOS

DATE/TIME OF TEST: 07/15/2018 04:44:18

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.081	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.087	04:39
Breath Volume	1.664	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.088	04:42
Breath Volume	3.101	04:42
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: HICIANO CRUZ

DATE/TIME OF TEST: 07/15/2018 05:29:25

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.078	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.214	05:24
Breath Volume	2.144	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:26
Subject Sample # 2	0.215	05:27
Breath Volume	3.675	05:27
Air Blank	0.000	05:28
Control Test	0.078	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: WOLFE

DATE/TIME OF TEST: 07/15/2018 22:34:50

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.080	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.154	22:30
Breath Volume	2.269	22:30
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.158	22:33
Breath Volume	2.226	22:33
Air Blank	0.000	22:33
Control Test	0.079	22:34
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: GATTO JR

DATE/TIME OF TEST: 07/16/2018 05:05:36

Diagnostics Check	OK	04:58
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.129	05:00
Breath Volume	1.730	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.129	05:03
Breath Volume	1.406	05:03
Air Blank	0.000	05:04
Control Test	0.079	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: BUENTIEMPO-WADE

DATE/TIME OF TEST: 07/16/2018 21:09:35

Diagnostics Check	OK	20:58
Air Blank	0.000	20:58
Control Test	0.081	20:58
Air Blank	0.000	20:59
Subject Sample # 1	0.237	21:00
Breath Volume	1.273	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:02
Subject Sample # 2	0.214	21:04
Breath Volume	1.917	21:04
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 3	0.222	21:07
Breath Volume	1.886	21:07
Air Blank	0.000	21:08
Control Test	0.078	21:08
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: JERONIMO

DATE/TIME OF TEST: 07/19/2018 02:18:21

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.081	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.303	02:12
Breath Volume	2.824	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.290	02:16
Breath Volume	2.156	02:16
Air Blank	0.000	02:17
Control Test	0.078	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: PAVILUS

DATE/TIME OF TEST: 07/20/2018 01:02:59

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.080	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.144	00:58
Breath Volume	1.371	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.145	01:01
Breath Volume	1.140	01:01
Air Blank	0.000	01:02
Control Test	0.079	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: CORDERO ALVAREZ

DATE/TIME OF TEST: 07/20/2018 03:05:48

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:52	
Control Test	0.080	02:52	
Air Blank	0.000	02:53	
Subject Sample	# 1 0.025	02:55	
Breath Volume	1.605	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:57	
Subject Sample	# 2 0.020	03:01	Volume Not Met
Breath Volume	0.781	03:01	
Air Blank	0.000	03:01	
Air Blank	0.000	03:03	
Subject Sample	# 3 0.021	03:04	
Breath Volume	1.695	03:04	
Air Blank	0.000	03:04	
Control Test	0.078	03:05	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRIZENDINE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 07/21/2018 01:00:56

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Subject Sample # 1	0.234	00:56
Breath Volume	2.625	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:58
Subject Sample # 2	0.244	00:59
Breath Volume	2.871	00:59
Air Blank	0.000	01:00
Control Test	0.079	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 07/21/2018 02:01:48

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.080	01:52
Air Blank	0.000	01:53
Subject Sample # 1	0.118	01:55
Breath Volume	1.121	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.118	02:00
Breath Volume	1.445	02:00
Air Blank	0.000	02:00
Control Test	0.079	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAULINO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: CASTILLO LOPEZ

DATE/TIME OF TEST: 07/21/2018 23:50:55

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.081	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.179	23:46
Breath Volume	2.386	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.186	23:49
Breath Volume	2.492	23:49
Air Blank	0.000	23:49
Control Test	0.080	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEINBERG, T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: TORRES MUNOZ

DATE/TIME OF TEST: 07/22/2018 00:20:20

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:12	
Control Test	0.081	00:12	
Air Blank	0.000	00:13	
Subject Sample	# 1 0.177	00:16	Volume Not Met
Breath Volume	0.058	00:16	
Air Blank	0.000	00:17	
Air Blank	0.000	00:18	
Subject Sample	# 2	00:18	Subject Test Refused
Breath Volume	0.000	00:18	
Air Blank	0.000	00:19	
Control Test	0.079	00:19	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEINBERG,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: FARRIS

DATE/TIME OF TEST: 07/22/2018 03:48:03

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.216	03:43
Breath Volume	2.285	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:45
Subject Sample # 2	0.211	03:46
Breath Volume	2.707	03:46
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:48

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: LOCHETTO

DATE/TIME OF TEST: 07/22/2018 04:47:13

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.129	04:42
Breath Volume	1.550	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:44
Subject Sample # 2	0.130	04:45
Breath Volume	2.093	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 07/23/2018 04:19:28

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.132	04:14
Breath Volume	1.351	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.129	04:17
Breath Volume	1.648	04:17
Air Blank	0.000	04:18
Control Test	0.080	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/24/2018 04:48:17

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.081	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.232	04:42
Breath Volume	1.113	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.229	04:46
Breath Volume	1.394	04:46
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: SAIZ-LEMIEUX

DATE/TIME OF TEST: 07/25/2018 00:57:47

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.195	00:53
Breath Volume	1.898	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.184	00:56
Breath Volume	1.582	00:56
Air Blank	0.000	00:56
Control Test	0.079	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: HARRINGTON

DATE/TIME OF TEST: 07/25/2018 23:41:28

Diagnostics Check	OK	23:34
Air Blank	0.000	23:35
Control Test	0.080	23:35
Air Blank	0.000	23:36
Subject Sample # 1	0.112	23:36
Breath Volume	2.046	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.109	23:39
Breath Volume	1.601	23:39
Air Blank	0.000	23:40
Control Test	0.080	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: CREEL

DATE/TIME OF TEST: 07/27/2018 02:27:17

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.080	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.000	02:22
Breath Volume	5.195	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.000	02:25
Breath Volume	4.199	02:25
Air Blank	0.000	02:26
Control Test	0.081	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA, A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: DIAZ VICENTE

DATE/TIME OF TEST: 07/28/2018 01:18:49

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.080	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.249	01:13
Breath Volume	1.335	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.240	01:17
Breath Volume	1.371	01:17
Air Blank	0.000	01:17
Control Test	0.080	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS, RONALD, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EDGE, S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: CLEMMER

DATE/TIME OF TEST: 07/28/2018 04:17:10

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.200	04:12
Breath Volume	2.960	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.200	04:15
Breath Volume	3.023	04:15
Air Blank	0.000	04:16
Control Test	0.078	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: LOHAUS JR

DATE/TIME OF TEST: 07/28/2018 20:37:10

Diagnostics Check	OK	20:30
Air Blank	0.000	20:31
Control Test	0.080	20:31
Air Blank	0.000	20:32
Subject Sample # 1	0.171	20:32
Breath Volume	3.464	20:32
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.180	20:35
Breath Volume	3.031	20:35
Air Blank	0.000	20:36
Control Test	0.079	20:36
Air Blank	0.000	20:37
Diagnostics Check	OK	20:37

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: CARROW

DATE/TIME OF TEST: 07/29/2018 04:33:23

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.120	04:28
Breath Volume	2.640	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.121	04:31
Breath Volume	3.644	04:31
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BUZA

DATE/TIME OF TEST: 07/30/2018 00:11:43

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.000	00:07
Breath Volume	2.156	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.000	00:10
Breath Volume	2.640	00:10
Air Blank	0.000	00:10
Control Test	0.081	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: ARGESON

DATE/TIME OF TEST: 07/30/2018 02:46:45

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.269	02:41
Breath Volume	1.535	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.275	02:45
Breath Volume	1.519	02:45
Air Blank	0.000	02:45
Control Test	0.080	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEINBERG,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: BRATCHER

DATE/TIME OF TEST: 07/31/2018 02:59:32

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.081	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.190	02:54
Breath Volume	2.406	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.190	02:57
Breath Volume	2.972	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: LOGUE

DATE/TIME OF TEST: 07/31/2018 04:13:09

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.115	04:08
Breath Volume	1.925	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.115	04:11
Breath Volume	1.863	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 08/07/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAFKJEN,THOMAS,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GAFKJEN,THOMAS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: PERCELL

DATE/TIME OF TEST: 07/02/2018 12:32:01

Diagnostics Check	OK	12:25
Air Blank	0.000	12:26
Control Test	0.081	12:26
Air Blank	0.000	12:27
Subject Sample # 1	0.000	12:27
Breath Volume	4.300	12:27
Air Blank	0.000	12:28
Air Blank	0.000	12:30
Subject Sample # 2	0.000	12:30
Breath Volume	3.976	12:30
Air Blank	0.000	12:31
Control Test	0.079	12:31
Air Blank	0.000	12:31
Diagnostics Check	OK	12:32

Data Download Date/Time: 07/27/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS,JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: BENJAMIN

DATE/TIME OF TEST: 07/05/2018 02:41:02

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.081	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.121	02:36
Breath Volume	1.527	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.126	02:39
Breath Volume	2.660	02:39
Air Blank	0.000	02:40
Control Test	0.078	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:41

Data Download Date/Time: 07/27/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLASHAW,HUNTER,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: TEETERS,OSBORNE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:33

SUBJECT'S LAST NAME: SJUVE-WILLIAMS

DATE/TIME OF TEST: 07/06/2018 18:04:12

Diagnostics Check	OK	17:58
Air Blank	0.000	17:58
Control Test	0.080	17:59
Air Blank	0.000	17:59
Subject Sample # 1	0.000	17:59
Breath Volume	1.902	17:59
Air Blank	0.000	18:00
Air Blank	0.000	18:02
Subject Sample # 2	0.000	18:02
Breath Volume	2.253	18:02
Air Blank	0.000	18:03
Control Test	0.079	18:03
Air Blank	0.000	18:04
Diagnostics Check	OK	18:04

Data Download Date/Time: 07/27/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNCH,CHAD,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LYNCH,CHAD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/08/2018 06:52:04

Diagnostics Check	OK	06:45
Air Blank	0.000	06:46
Control Test	0.081	06:46
Air Blank	0.000	06:46
Subject Sample # 1	0.151	06:47
Breath Volume	2.722	06:47
Air Blank	0.000	06:48
Air Blank	0.000	06:50
Subject Sample # 2	0.150	06:50
Breath Volume	3.156	06:50
Air Blank	0.000	06:51
Control Test	0.080	06:51
Air Blank	0.000	06:51
Diagnostics Check	OK	06:52

Data Download Date/Time: 07/27/2018 09:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENNELL,JOHN,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VENNELL,JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: TORRES RAMIREZ

DATE/TIME OF TEST: 07/08/2018 20:12:44

Diagnostics Check	OK	20:06
Air Blank	0.000	20:06
Control Test	0.080	20:06
Air Blank	0.000	20:07
Subject Sample # 1	0.226	20:08
Breath Volume	3.070	20:08
Air Blank	0.000	20:08
Air Blank	0.000	20:10
Subject Sample # 2	0.232	20:11
Breath Volume	3.359	20:11
Air Blank	0.000	20:11
Control Test	0.079	20:12
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 07/27/2018 09:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TASCHETTA,JOSEPH,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: ISENBERG,KESLEY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: ELKINS

DATE/TIME OF TEST: 07/09/2018 18:04:38

Diagnostics Check	OK	18:01	
Air Blank	0.000	18:02	
Control Test	0.080	18:02	
Air Blank	0.000	18:03	
Subject Sample	# 1	18:03	Subject Test Refused
Breath Volume	0.000	18:03	
Air Blank	0.000	18:03	
Control Test	0.080	18:04	
Air Blank	0.000	18:04	
Diagnostics Check	OK	18:04	

Data Download Date/Time: 07/27/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TASCHETTA,JOSEPH,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: OSBORNE,J
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 07/11/2018 05:10:51

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.081	05:05
Air Blank	0.000	05:05
Subject Sample # 1	0.093	05:06
Breath Volume	1.968	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:08
Subject Sample # 2	0.091	05:09
Breath Volume	1.167	05:09
Air Blank	0.000	05:09
Control Test	0.080	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 07/27/2018 09:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS,JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 07/13/2018 21:28:38

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:25	
Control Test	0.080	21:25	
Air Blank	0.000	21:26	
Subject Sample	# 1	21:27	Subject Test Refused
Breath Volume	0.000	21:27	
Air Blank	0.000	21:27	
Control Test	0.081	21:28	
Air Blank	0.000	21:28	
Diagnostics Check	OK	21:28	

Data Download Date/Time: 07/27/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KAYLAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 07/14/2018 03:03:13

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.081	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.044	02:58
Breath Volume	1.527	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.042	03:01
Breath Volume	1.140	03:01
Air Blank	0.000	03:02
Control Test	0.081	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 07/27/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINH,TAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DINH,TAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: HESTER

DATE/TIME OF TEST: 07/14/2018 19:00:23

Diagnostics Check	OK	18:53
Air Blank	0.000	18:54
Control Test	0.081	18:54
Air Blank	0.000	18:55
Subject Sample # 1	0.112	18:55
Breath Volume	1.335	18:55
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample # 2	0.110	18:58
Breath Volume	1.562	18:58
Air Blank	0.000	18:59
Control Test	0.081	18:59
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

Data Download Date/Time: 07/27/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KNOWLES,CONSTANCE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/14/2018 21:33:48

Diagnostics Check	OK	21:27
Air Blank	0.000	21:27
Control Test	0.081	21:28
Air Blank	0.000	21:28
Subject Sample # 1	0.000	21:29
Breath Volume	4.773	21:29
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.000	21:32
Breath Volume	5.019	21:32
Air Blank	0.000	21:32
Control Test	0.081	21:33
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 07/27/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS,JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 07/15/2018 01:15:09

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.080	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.154	01:10
Breath Volume	2.125	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:12
Subject Sample # 2	0.158	01:13
Breath Volume	1.398	01:13
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 07/27/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAVES,TIMOTHY,F
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: GRAVES,TIMOTHY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: CROMWELL

DATE/TIME OF TEST: 07/15/2018 17:49:26

Diagnostics Check	OK	17:42
Air Blank	0.000	17:43
Control Test	0.081	17:43
Air Blank	0.000	17:43
Subject Sample # 1	0.000	17:44
Breath Volume	1.152	17:44
Air Blank	0.000	17:45
Air Blank	0.000	17:47
Subject Sample # 2	0.000	17:48
Breath Volume	2.765	17:48
Air Blank	0.000	17:48
Control Test	0.081	17:48
Air Blank	0.000	17:49
Diagnostics Check	OK	17:49

Data Download Date/Time: 07/27/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS, JEREMIAH, S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS, JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: BOLLING

DATE/TIME OF TEST: 07/21/2018 04:15:33

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.081	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.095	04:11
Breath Volume	1.539	04:11
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.086	04:13
Breath Volume	1.187	04:13
Air Blank	0.000	04:14
Control Test	0.079	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 07/27/2018 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS, JEREMIAH, S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS, JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: PERCELL

DATE/TIME OF TEST: 07/21/2018 23:48:59

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.000	23:44
Breath Volume	1.687	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.000	23:47
Breath Volume	1.632	23:47
Air Blank	0.000	23:48
Control Test	0.079	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 07/27/2018 09:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SECOR,BROC,A
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SECOR,BROC
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: FRANZ

DATE/TIME OF TEST: 07/28/2018 03:59:39

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.082	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.138	03:53
Breath Volume	1.207	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:55
Subject Sample # 2	0.145	03:57
Breath Volume	2.144	03:57
Air Blank	0.000	03:58
Control Test	0.079	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 08/21/2018 11:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS, JEREMIAH, S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS, JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: SARAVIA MAYA

DATE/TIME OF TEST: 07/29/2018 02:33:34

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.258	02:29
Breath Volume	2.546	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.250	02:31
Breath Volume	2.640	02:31
Air Blank	0.000	02:32
Control Test	0.078	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 08/21/2018 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORENE,AIDAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NORENE,AIDAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: BOWSER

DATE/TIME OF TEST: 07/29/2018 21:29:50

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.081	21:24
Air Blank	0.000	21:25
Subject Sample # 1	0.000	21:25
Breath Volume	2.792	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.000	21:28
Breath Volume	2.757	21:28
Air Blank	0.000	21:28
Control Test	0.081	21:29
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 08/21/2018 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS, JEREMIAH, S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS, JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: VATNE

DATE/TIME OF TEST: 07/30/2018 00:51:33

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.081	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.302	00:46
Breath Volume	4.035	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.303	00:49
Breath Volume	3.648	00:49
Air Blank	0.000	00:50
Control Test	0.078	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 08/21/2018 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: SCHLOSSER

DATE/TIME OF TEST: 07/01/2018 01:07:58

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.078	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.147	01:03
Breath Volume	4.230	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:05
Subject Sample # 2	0.145	01:06
Breath Volume	4.507	01:06
Air Blank	0.000	01:07
Control Test	0.076	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 07/27/2018 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAMUEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: STALLINGS GODWIN

DATE/TIME OF TEST: 07/01/2018 02:56:40

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.078	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.116	02:51
Breath Volume	2.199	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.120	02:55
Breath Volume	1.816	02:55
Air Blank	0.000	02:55
Control Test	0.077	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 07/27/2018 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/01/2018 22:02:41

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.078	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.140	21:58
Breath Volume	1.800	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.149	22:01
Breath Volume	2.507	22:01
Air Blank	0.000	22:01
Control Test	0.077	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: FARRIS

DATE/TIME OF TEST: 07/02/2018 02:53:28

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.077	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.238	02:48
Breath Volume	2.597	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.238	02:51
Breath Volume	2.488	02:51
Air Blank	0.000	02:52
Control Test	0.077	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAMUEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 07/04/2018 23:21:00

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.078	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.160	23:15
Breath Volume	2.925	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.161	23:19
Breath Volume	3.386	23:19
Air Blank	0.000	23:20
Control Test	0.078	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:21

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/05/2018 04:06:32

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.209	04:01
Breath Volume	1.457	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.204	04:04
Breath Volume	1.105	04:04
Air Blank	0.000	04:05
Control Test	0.077	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/05/2018 20:55:08

Diagnostics Check	OK	20:48
Air Blank	0.000	20:49
Control Test	0.079	20:49
Air Blank	0.000	20:50
Subject Sample # 1	0.082	20:50
Breath Volume	2.542	20:50
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.076	20:53
Breath Volume	1.718	20:53
Air Blank	0.000	20:54
Control Test	0.077	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALON,DENNIS,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LACKEY,NICHOLAS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:52

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 07/06/2018 15:24:59

Diagnostics Check	OK	15:13	
Air Blank	0.000	15:13	
Control Test	0.077	15:13	
Air Blank	0.000	15:14	
Subject Sample	# 1 0.123	15:17	Volume Not Met
Breath Volume	0.808	15:17	
Air Blank	0.000	15:18	
Air Blank	0.000	15:20	
Subject Sample	# 2 0.112	15:23	Volume Not Met
Breath Volume	0.902	15:23	
Air Blank	0.000	15:24	
Control Test	0.075	15:24	
Air Blank	0.000	15:24	
Diagnostics Check	OK	15:24	

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALON,DENNIS,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LACKEY,NICHOLAS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:52

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 07/06/2018 15:40:19

Diagnostics Check	OK	15:28	
Air Blank	0.000	15:29	
Control Test	0.075	15:29	
Air Blank	0.000	15:30	
Subject Sample	# 1 0.109	15:33	Volume Not Met
Breath Volume	0.828	15:33	
Air Blank	0.000	15:33	
Air Blank	0.000	15:35	
Subject Sample	# 2 0.109	15:38	Volume Not Met
Breath Volume	0.757	15:38	
Air Blank	0.000	15:39	
Control Test	0.072	15:39	Control Outside Tolerance
Air Blank	0.000	15:40	

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESTIFO, CHARLES, T
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MAYO, TREVOR
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: WARHOLIC

DATE/TIME OF TEST: 07/10/2018 17:40:08

Diagnostics Check	OK	17:37	
Air Blank	0.000	17:37	
Control Test	0.076	17:38	
Air Blank	0.000	17:38	
Subject Sample	# 1	17:39	Sequence Aborted
Breath Volume	0.000	17:39	
Air Blank	0.000	17:40	

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESTIFO,CHARLES,T
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MAYO,TREVOR
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: WARHOLIC

DATE/TIME OF TEST: 07/10/2018 17:44:46

Diagnostics Check	OK	17:43	
Air Blank	0.000	17:43	
Control Test	0.074	17:44	Control Outside Tolerance
Air Blank	0.000	17:44	

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESTIFO, CHARLES, T
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MAYO, TREVOR
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: WARHOLIC

DATE/TIME OF TEST: 07/10/2018 17:54:27

Diagnostics Check	OK	17:47
Air Blank	0.000	17:48
Control Test	0.076	17:48
Air Blank	0.000	17:49
Subject Sample # 1	0.297	17:49
Breath Volume	1.906	17:49
Air Blank	0.000	17:50
Air Blank	0.000	17:52
Subject Sample # 2	0.305	17:52
Breath Volume	3.023	17:52
Air Blank	0.000	17:53
Control Test	0.076	17:53
Air Blank	0.000	17:54
Diagnostics Check	OK	17:54

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGALADO,JORGE,I
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: REGALADO,JORGE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: EUBANKS

DATE/TIME OF TEST: 07/10/2018 19:02:31

Diagnostics Check	OK	18:56	
Air Blank	0.000	18:56	
Control Test	0.077	18:57	
Air Blank	0.000	18:57	
Subject Sample # 1	0.000	18:58	
Breath Volume	4.175	18:58	
Air Blank	0.000	18:58	
Air Blank	0.000	19:00	
Subject Sample # 2	0.000	19:01	
Breath Volume	3.722	19:01	
Air Blank	0.000	19:01	
Control Test	0.074	19:01	Control Outside Tolerance
Air Blank	0.000	19:02	

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: MCKINNEY

DATE/TIME OF TEST: 07/11/2018 02:05:28

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.077	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.181	02:00
Breath Volume	1.507	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.181	02:03
Breath Volume	1.468	02:03
Air Blank	0.000	02:04
Control Test	0.077	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 07/12/2018 01:10:13

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.078	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.110	01:05
Breath Volume	1.796	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.104	01:08
Breath Volume	2.546	01:08
Air Blank	0.000	01:09
Control Test	0.078	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAMUEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 07/12/2018 20:58:26

Diagnostics Check	OK	20:51
Air Blank	0.000	20:52
Control Test	0.076	20:52
Air Blank	0.000	20:52
Subject Sample # 1	0.248	20:53
Breath Volume	1.316	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:55
Subject Sample # 2	0.232	20:56
Breath Volume	1.765	20:56
Air Blank	0.000	20:57
Control Test	0.076	20:57
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACOUR,CANDACE,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY,WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: LACOUR

DATE/TIME OF TEST: 07/12/2018 22:36:16

Diagnostics Check	OK	22:29
Air Blank	0.000	22:30
Control Test	0.078	22:30
Air Blank	0.000	22:31
Subject Sample # 1	0.162	22:31
Breath Volume	3.375	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.166	22:34
Breath Volume	3.828	22:34
Air Blank	0.000	22:35
Control Test	0.077	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAMUEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 07/14/2018 02:47:26

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.078	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.157	02:42
Breath Volume	2.664	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.161	02:45
Breath Volume	2.921	02:45
Air Blank	0.000	02:46
Control Test	0.077	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTE,JACOB,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: COTE,JACOB
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 07/15/2018 01:44:29

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.078	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.181	01:39
Breath Volume	4.933	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.166	01:42
Breath Volume	3.175	01:42
Air Blank	0.000	01:43
Control Test	0.075	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEITE,STEVEN,G
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEITE,STEVEN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 07/16/2018 23:48:34

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:46	
Control Test	0.078	23:46	
Air Blank	0.000	23:47	
Subject Sample	# 1	23:47	Subject Test Refused
Breath Volume	0.000	23:47	
Air Blank	0.000	23:47	
Control Test	0.079	23:47	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/17/2018 22:12:52

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.078	22:07
Air Blank	0.000	22:07
Subject Sample # 1	0.153	22:08
Breath Volume	3.195	22:08
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.154	22:11
Breath Volume	4.292	22:11
Air Blank	0.000	22:11
Control Test	0.076	22:12
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MALINOWSKI

DATE/TIME OF TEST: 07/18/2018 00:57:23

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.078	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.258	00:52
Breath Volume	1.582	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.252	00:55
Breath Volume	2.160	00:55
Air Blank	0.000	00:56
Control Test	0.077	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAMUEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: CARAWAY

DATE/TIME OF TEST: 07/19/2018 23:23:34

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.078	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.128	23:18
Breath Volume	2.023	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.131	23:21
Breath Volume	2.921	23:21
Air Blank	0.000	23:22
Control Test	0.078	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAMUEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 07/20/2018 23:38:45

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.077	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.104	23:34
Breath Volume	3.210	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.103	23:37
Breath Volume	3.410	23:37
Air Blank	0.000	23:37
Control Test	0.078	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAMUEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: CRAFT

DATE/TIME OF TEST: 07/21/2018 04:18:23

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.078	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.000	04:13
Breath Volume	3.906	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.000	04:17
Breath Volume	2.589	04:17
Air Blank	0.000	04:17
Control Test	0.078	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENE,ROBERT,R
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEE,J
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:47

SUBJECT'S LAST NAME: BARTUSH

DATE/TIME OF TEST: 07/22/2018 18:14:11

Diagnostics Check	OK	18:08
Air Blank	0.000	18:08
Control Test	0.077	18:08
Air Blank	0.000	18:09
Subject Sample # 1	0.209	18:09
Breath Volume	3.875	18:09
Air Blank	0.000	18:10
Air Blank	0.000	18:12
Subject Sample # 2	0.192	18:12
Breath Volume	4.464	18:12
Air Blank	0.000	18:13
Control Test	0.076	18:13
Air Blank	0.000	18:14
Diagnostics Check	OK	18:14

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLASHAW,HUNTER,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DILLASHAW,HUNTER
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: BINDER

DATE/TIME OF TEST: 07/23/2018 02:54:56

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.077	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.134	02:50
Breath Volume	2.046	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:52
Subject Sample # 2	0.140	02:53
Breath Volume	2.207	02:53
Air Blank	0.000	02:54
Control Test	0.076	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BYRD,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/25/2018 01:59:49

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.078	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.177	01:55
Breath Volume	1.949	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.189	01:58
Breath Volume	3.898	01:58
Air Blank	0.000	01:58
Control Test	0.076	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 07/27/2018 03:13:47

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.079	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.000	03:09
Breath Volume	4.238	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.000	03:12
Breath Volume	4.085	03:12
Air Blank	0.000	03:12
Control Test	0.079	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 07/27/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERNANDEZ,DIYANGO,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: FERNANDEZ,STEPHANIE

ARRESTING OFFICER AGENCY: South Miami PD

LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 640976

DRY GAS EXPIRY DATE: 11/21/2019

VIOLATION CODE: DUI

ARREST TIME: 01:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 07/01/2018 02:29:40

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.220	02:25
Breath Volume	2.355	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.222	02:27
Breath Volume	1.109	02:27
Air Blank	0.000	02:28
Control Test	0.078	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 07/30/2018 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GUERRERO,OFC O,L

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: GUERRERO,OFC O.

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 640976

DRY GAS EXPIRY DATE: 11/21/2019

VIOLATION CODE: DUI

ARREST TIME: 10:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 11:13

SUBJECT'S LAST NAME: LARGAESPADA

DATE/TIME OF TEST: 07/01/2018 11:46:27

Diagnostics Check	OK	11:39
Air Blank	0.000	11:39
Control Test	0.080	11:39
Air Blank	0.000	11:40
Subject Sample # 1	0.173	11:41
Breath Volume	3.406	11:41
Air Blank	0.000	11:42
Air Blank	0.000	11:44
Subject Sample # 2	0.169	11:44
Breath Volume	3.628	11:44
Air Blank	0.000	11:45
Control Test	0.079	11:45
Air Blank	0.000	11:46
Diagnostics Check	OK	11:46

Data Download Date/Time: 07/30/2018 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERNANDEZ,DIYANGO,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: MENENDEZ,S

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 640976

DRY GAS EXPIRY DATE: 11/21/2019

VIOLATION CODE: DUI

ARREST TIME: 01:50

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: JUAREZ

DATE/TIME OF TEST: 07/02/2018 02:35:11

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.079	02:33	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:33	Subject Test Refused
Breath Volume	0.000	02:33	
Air Blank	0.000	02:34	
Control Test	0.080	02:34	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 07/30/2018 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RAMIREZ,CRISTIAN,A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: TARGIA,FRANK

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 640976

DRY GAS EXPIRY DATE: 11/21/2019

VIOLATION CODE: DUI

ARREST TIME: 04:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: CONTERAS CABRERA

DATE/TIME OF TEST: 07/07/2018 05:09:08

Diagnostics Check	OK	05:08	
Air Blank		05:08	RFI Detect
Air Blank	0.000	05:09	

Data Download Date/Time: 07/30/2018 18:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RAMIREZ,CRISTIAN,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: TARGIA,FRANK

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 640976

DRY GAS EXPIRY DATE: 11/21/2019

VIOLATION CODE: DUI

ARREST TIME: 04:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: CONTERAS CABRERA

DATE/TIME OF TEST: 07/07/2018 05:18:44

Diagnostics Check	OK	05:12
Air Blank	0.000	05:12
Control Test	0.080	05:12
Air Blank	0.000	05:13
Subject Sample # 1	0.067	05:14
Breath Volume	2.378	05:14
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.066	05:17
Breath Volume	2.410	05:17
Air Blank	0.000	05:17
Control Test	0.079	05:18
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 07/30/2018 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROIG,DAWN,M

BREATH TEST OPERATOR AGENCY: Miami Shores PD

ARRESTING OFFICER: ROIG,DAWN

ARRESTING OFFICER AGENCY: Miami Shores PD

LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 640976

DRY GAS EXPIRY DATE: 11/21/2019

VIOLATION CODE: DUI

ARREST TIME: 16:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: ALLENDES

DATE/TIME OF TEST: 07/14/2018 17:52:19

Diagnostics Check	OK	17:40	
Air Blank	0.000	17:41	
Control Test	0.080	17:41	
Air Blank	0.000	17:41	
Subject Sample	# 1	17:44	No Sample Provided
Breath Volume	0.000	17:44	
Air Blank	0.000	17:45	
Air Blank	0.000	17:47	
Subject Sample	# 2	17:50	Volume Not Met
Breath Volume	0.531	17:50	
Air Blank	0.000	17:51	
Control Test	0.081	17:51	
Air Blank	0.000	17:52	
Diagnostics Check	OK	17:52	

Data Download Date/Time: 07/30/2018 18:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROIG,DAWN,M
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: ROIG,DAWN
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: ALLENDES

DATE/TIME OF TEST: 07/14/2018 18:11:09

Diagnostics Check	OK	17:56	
Air Blank	0.000	17:57	
Control Test	0.079	17:57	
Air Blank	0.000	17:57	
Subject Sample	# 1 0.121	18:00	Volume Not Met
Breath Volume	0.589	18:00	
Air Blank	0.000	18:01	
Air Blank	0.000	18:03	
Subject Sample	# 2 0.147	18:05	
Breath Volume	1.281	18:05	
Air Blank	0.000	18:06	
Air Blank	0.000	18:07	
Subject Sample	# 3 0.142	18:09	
Breath Volume	1.375	18:09	
Air Blank	0.000	18:10	
Control Test	0.078	18:10	
Air Blank	0.000	18:11	
Diagnostics Check	OK	18:11	

Data Download Date/Time: 07/30/2018 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OTANO,D,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: OTANO,D
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 08:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:22

SUBJECT'S LAST NAME: ROQUE MALDONADO

DATE/TIME OF TEST: 07/21/2018 09:11:13

Diagnostics Check	OK	09:03
Air Blank	0.000	09:03
Control Test	0.079	09:04
Air Blank	0.000	09:04
Subject Sample # 1	0.174	09:06
Breath Volume	1.976	09:06
Air Blank	0.000	09:07
Air Blank	0.000	09:08
Subject Sample # 2	0.169	09:09
Breath Volume	2.000	09:09
Air Blank	0.000	09:10
Control Test	0.079	09:10
Air Blank	0.000	09:11
Diagnostics Check	OK	09:11

Data Download Date/Time: 07/30/2018 18:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,DAMIAN,N
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CASTILLO,ISMAELIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: LIEBES

DATE/TIME OF TEST: 07/27/2018 21:28:26

Diagnostics Check	OK	21:21
Air Blank	0.000	21:22
Control Test	0.080	21:22
Air Blank	0.000	21:23
Subject Sample # 1	0.018	21:23
Breath Volume	3.285	21:23
Air Blank	0.000	21:24
Air Blank	0.000	21:26
Subject Sample # 2	0.016	21:26
Breath Volume	2.871	21:26
Air Blank	0.000	21:27
Control Test	0.078	21:27
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

Data Download Date/Time: 07/30/2018 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LANFEAR,JAMES,A

BREATH TEST OPERATOR AGENCY: Coral Gables PD

ARRESTING OFFICER: RAMBERT,MICHAEL

ARRESTING OFFICER AGENCY: Coral Gables PD

LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 00:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: TOMLINSON

DATE/TIME OF TEST: 07/01/2018 01:11:47

Diagnostics Check	OK	01:08	
Air Blank	0.000	01:09	
Control Test	0.080	01:09	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:10	Subject Test Refused
Breath Volume	0.000	01:10	
Air Blank	0.000	01:10	
Control Test	0.080	01:11	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 07/30/2018 18:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OTANO,D,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: OTANO,D
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 10:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:43

SUBJECT'S LAST NAME: PONCE AGUILERA

DATE/TIME OF TEST: 07/01/2018 11:23:39

Diagnostics Check	OK	11:20	
Air Blank	0.000	11:20	
Control Test	0.080	11:21	
Air Blank	0.000	11:21	
Subject Sample	# 1	11:22	Subject Test Refused
Breath Volume	0.000	11:22	
Air Blank	0.000	11:22	
Control Test	0.081	11:23	
Air Blank	0.000	11:23	
Diagnostics Check	OK	11:23	

Data Download Date/Time: 07/30/2018 18:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,ANDRE,D
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: JOHNSON,K
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ENRIQUEZ

DATE/TIME OF TEST: 07/02/2018 02:22:25

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:20	
Control Test	0.080	02:20	
Air Blank	0.000	02:20	
Subject Sample	# 1	02:21	Subject Test Refused
Breath Volume	0.000	02:21	
Air Blank	0.000	02:21	
Control Test	0.080	02:21	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 07/30/2018 18:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMINGTON,EARL,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SIMINGTON,EARL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 14:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: DE CAMARGO

DATE/TIME OF TEST: 07/02/2018 18:02:58

Diagnostics Check	OK	17:52	
Air Blank	0.000	17:52	
Control Test	0.080	17:52	
Air Blank	0.000	17:53	
Subject Sample	# 1	17:56	No Sample Provided
Breath Volume	0.000	17:56	
Air Blank	0.000	17:56	
Air Blank	0.000	17:59	
Subject Sample	# 2	18:01	Subject Test Refused
Breath Volume	0.000	18:01	
Air Blank	0.000	18:02	
Control Test	0.081	18:02	
Air Blank	0.000	18:02	
Diagnostics Check	OK	18:02	

Data Download Date/Time: 07/30/2018 18:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BETANCOURT, YANIER,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: ACOSTA, CAROLINA

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 20:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/02/2018 21:03:16

Diagnostics Check	OK	20:54	
Air Blank	0.000	20:55	
Control Test	0.081	20:55	
Air Blank	0.000	20:56	
Subject Sample	# 1	20:59	No Sample Provided
Breath Volume	0.000	20:59	
Air Blank	0.000	20:59	
Air Blank	0.000	21:01	
Subject Sample	# 2	21:01	Subject Test Refused
Breath Volume	0.000	21:01	
Air Blank	0.000	21:02	
Control Test	0.080	21:02	
Air Blank	0.000	21:03	
Diagnostics Check	OK	21:03	

Data Download Date/Time: 07/30/2018 18:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PALOMINO,JULIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/03/2018 03:05:22

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.181	02:59
Breath Volume	1.488	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.174	03:03
Breath Volume	1.972	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 07/30/2018 18:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PALOMINO,JULIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/04/2018 02:34:32

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:29	
Control Test	0.079	02:29	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:33	No Sample Provided
Breath Volume	0.000	02:33	
Air Blank		02:34	Ambient Fail
Air Blank	0.000	02:34	

Data Download Date/Time: 07/30/2018 18:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PALOMINO,JULIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/04/2018 02:42:09

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.080	02:39	
Air Blank	0.000	02:39	
Subject Sample	# 1	02:40	Subject Test Refused
Breath Volume	0.000	02:40	
Air Blank	0.000	02:41	
Control Test	0.080	02:41	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 07/30/2018 18:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,JASIEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 07/07/2018 04:21:16

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:12	
Control Test	0.079	04:12	
Air Blank	0.000	04:13	
Subject Sample	# 1	04:16	No Sample Provided
Breath Volume	0.000	04:16	
Air Blank	0.000	04:16	
Air Blank	0.000	04:18	
Subject Sample	# 2	04:19	Subject Test Refused
Breath Volume	0.000	04:19	
Air Blank	0.000	04:20	
Control Test	0.085	04:20	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 07/30/2018 18:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ACOSTA,CAROLINA,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: ACOSTA,CAROLINA

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 18:32

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 07/07/2018 19:18:53

Diagnostics Check	OK	19:11
Air Blank	0.000	19:12
Control Test	0.080	19:12
Air Blank	0.000	19:13
Subject Sample # 1	0.169	19:13
Breath Volume	1.433	19:13
Air Blank	0.000	19:14
Air Blank	0.000	19:16
Subject Sample # 2	0.167	19:17
Breath Volume	1.304	19:17
Air Blank	0.000	19:17
Control Test	0.081	19:18
Air Blank	0.000	19:18
Diagnostics Check	OK	19:18

Data Download Date/Time: 07/30/2018 18:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,JOSEPH,M
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GOMEZ,JOSEPH
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 07/08/2018 06:50:32

Diagnostics Check	OK	06:47	
Air Blank	0.000	06:47	
Control Test	0.079	06:48	
Air Blank	0.000	06:48	
Subject Sample	# 1	06:49	Subject Test Refused
Breath Volume	0.000	06:49	
Air Blank	0.000	06:49	
Control Test	0.081	06:49	
Air Blank	0.000	06:50	
Diagnostics Check	OK	06:50	

Data Download Date/Time: 07/30/2018 18:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,IVAN,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GARCIA,IVAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: BUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: VERDIAL

DATE/TIME OF TEST: 07/08/2018 20:35:13

Diagnostics Check	OK	20:31	
Air Blank	0.000	20:32	
Control Test	0.081	20:32	
Air Blank	0.000	20:33	
Subject Sample	# 1	20:33	Subject Test Refused
Breath Volume	0.000	20:33	
Air Blank	0.000	20:34	
Control Test	0.081	20:34	
Air Blank	0.000	20:35	
Diagnostics Check	OK	20:35	

Data Download Date/Time: 07/30/2018 18:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BETANCOURT,YANIER,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: ACOSTA,CAROLINA

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 20:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: COMMANDER

DATE/TIME OF TEST: 07/08/2018 21:17:26

Diagnostics Check	OK	21:04	
Air Blank	0.000	21:05	
Control Test	0.079	21:05	
Air Blank	0.000	21:06	
Subject Sample	# 1 0.224	21:09	Volume Not Met
Breath Volume	0.324	21:09	
Air Blank	0.000	21:10	
Air Blank	0.000	21:11	
Subject Sample	# 2 0.302	21:12	
Breath Volume	1.820	21:12	
Air Blank	0.000	21:13	
Air Blank	0.000	21:15	
Subject Sample	# 3	21:16	Subject Test Refused
Breath Volume	0.000	21:16	
Air Blank	0.000	21:16	
Control Test	0.080	21:16	
Air Blank	0.000	21:17	
Diagnostics Check	OK	21:17	

Data Download Date/Time: 07/30/2018 18:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,JASIEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: LANEUX

DATE/TIME OF TEST: 07/10/2018 05:44:14

Diagnostics Check	OK	05:41	
Air Blank	0.000	05:41	
Control Test	0.081	05:42	
Air Blank	0.000	05:42	
Subject Sample	# 1	05:42	Subject Test Refused
Breath Volume	0.000	05:42	
Air Blank	0.000	05:43	
Control Test	0.082	05:43	
Air Blank	0.000	05:44	
Diagnostics Check	OK	05:44	

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARGIA,FRANKA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TARGIA,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: BUSTOS

DATE/TIME OF TEST: 07/11/2018 19:11:47

Diagnostics Check	OK	19:05
Air Blank	0.000	19:05
Control Test	0.082	19:06
Air Blank	0.000	19:06
Subject Sample # 1	0.000	19:07
Breath Volume	1.484	19:07
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample # 2	0.000	19:10
Breath Volume	1.554	19:10
Air Blank	0.000	19:10
Control Test	0.082	19:11
Air Blank	0.000	19:11
Diagnostics Check	OK	19:11

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUKSA,ANDREW,

BREATH TEST OPERATOR AGENCY: Coral Gables PD

ARRESTING OFFICER: SEEGERSEARS,KRISTEN

ARRESTING OFFICER AGENCY: Coral Gables PD

LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 01:40

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: ABREU

DATE/TIME OF TEST: 07/12/2018 03:26:56

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.082	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.173	03:22
Breath Volume	2.152	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample # 2	0.176	03:25
Breath Volume	1.542	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RAMIREZ,CRISTIAN,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: TARGIA,F

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 02:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: TOVAR

DATE/TIME OF TEST: 07/13/2018 03:12:09

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:01	
Control Test	0.083	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:05	No Sample Provided
Breath Volume	0.000	03:05	
Air Blank	0.000	03:05	
Air Blank	0.000	03:07	
Subject Sample	# 2	03:10	No Sample Provided
Breath Volume	0.000	03:10	
Air Blank	0.000	03:11	
Control Test	0.083	03:11	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RAMIREZ,CRISTIAN,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: TARGIA,FRANK

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 02:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: TOVAR

DATE/TIME OF TEST: 07/13/2018 03:20:19

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:17	
Control Test	0.080	03:18	
Air Blank	0.000	03:18	
Subject Sample	# 1	03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:19	
Control Test	0.081	03:19	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OTANO,D,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: OTANO,D
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 13:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:26

SUBJECT'S LAST NAME: CHAVARRIA CASTILLO

DATE/TIME OF TEST: 07/13/2018 14:10:03

Diagnostics Check	OK	14:07	
Air Blank	0.000	14:07	
Control Test	0.081	14:08	
Air Blank	0.000	14:08	
Subject Sample	# 1	14:08	Subject Test Refused
Breath Volume	0.000	14:08	
Air Blank	0.000	14:09	
Control Test	0.082	14:09	
Air Blank	0.000	14:09	
Diagnostics Check	OK	14:10	

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYER,THOMAS,S
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUEZ,ARNIEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: CANDELARIO

DATE/TIME OF TEST: 07/15/2018 19:24:50

Diagnostics Check	OK	19:13	
Air Blank	0.000	19:13	
Control Test	0.082	19:13	
Air Blank	0.000	19:14	
Subject Sample	# 1 0.219	19:17	Volume Not Met
Breath Volume	0.968	19:17	
Air Blank	0.000	19:18	
Air Blank	0.000	19:19	
Subject Sample	# 2 0.220	19:23	Volume Not Met
Breath Volume	0.164	19:23	
Air Blank	0.000	19:23	
Control Test	0.081	19:24	
Air Blank	0.000	19:24	
Diagnostics Check	OK	19:24	

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYER,THOMAS,S
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUEZ,ARNIEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: CANDELARIO

DATE/TIME OF TEST: 07/15/2018 19:33:03

Diagnostics Check	OK	19:27	
Air Blank	0.000	19:28	
Control Test	0.076	19:28	
Air Blank	0.000	19:29	
Subject Sample	# 1	19:31	Subject Test Refused
Breath Volume	0.203	19:31	
Air Blank	0.000	19:32	
Control Test	0.082	19:32	
Air Blank	0.000	19:32	
Diagnostics Check	OK	19:33	

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NAYA,CHRISTOPHER,M

BREATH TEST OPERATOR AGENCY: Coral Gables PD

ARRESTING OFFICER: FLEURIMONT,JEFF

ARRESTING OFFICER AGENCY: Coral Gables PD

LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 23:16

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MASSA

DATE/TIME OF TEST: 07/17/2018 00:27:01

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:21	
Control Test	0.082	00:21	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:25	Slope Not Met
Breath Volume	1.558	00:25	
Air Blank	0.000	00:26	
Control Test	0.078	00:26	
Air Blank	0.000	00:26	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALVEZ,JOSE,R
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,JASIEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: BAZAN

DATE/TIME OF TEST: 07/20/2018 05:54:15

Diagnostics Check	OK	05:38	
Air Blank	0.000	05:39	
Control Test	0.077	05:39	
Air Blank	0.000	05:39	
Subject Sample	# 1 0.259	05:41	
Breath Volume	1.386	05:41	
Air Blank	0.000	05:42	
Air Blank	0.000	05:43	
Subject Sample	# 2 0.227	05:46	Volume Not Met
Breath Volume	0.242	05:46	
Air Blank	0.000	05:47	
Air Blank	0.000	05:49	
Subject Sample	# 3 0.222	05:52	Volume Not Met
Breath Volume	1.089	05:52	
Air Blank	0.000	05:53	
Control Test	0.077	05:53	
Air Blank	0.000	05:54	
Diagnostics Check	OK	05:54	

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PEREZ,CRISTINO,

BREATH TEST OPERATOR AGENCY: Coral Gables PD

ARRESTING OFFICER: GIBSON,GERONIMO

ARRESTING OFFICER AGENCY: Coral Gables PD

LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 00:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: PUJOL

DATE/TIME OF TEST: 07/21/2018 01:40:47

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:38	
Control Test	0.080	01:38	
Air Blank	0.000	01:38	
Subject Sample	# 1	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:39	
Control Test	0.081	01:40	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OTANO,D,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: OTANO,D
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 08:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:22

SUBJECT'S LAST NAME: ROQUE MALDONADO

DATE/TIME OF TEST: 07/21/2018 08:58:11

Diagnostics Check	OK	08:53	
Air Blank	0.000	08:54	
Control Test	0.079	08:54	
Air Blank	0.000	08:54	
Subject Sample	# 1 0.175	08:56	
Breath Volume	1.789	08:56	
Air Blank		08:57	Ambient Fail
Air Blank	0.000	08:58	

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OTANO,D,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: OTANO,D
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 07:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: HOOKER PINEDA

DATE/TIME OF TEST: 07/22/2018 08:35:05

Diagnostics Check	OK	08:24	
Air Blank	0.000	08:24	
Control Test	0.083	08:24	
Air Blank	0.000	08:25	
Subject Sample	# 1 0.263	08:28	Volume Not Met
Breath Volume	0.484	08:28	
Air Blank	0.000	08:29	
Air Blank	0.000	08:31	
Subject Sample	# 2	08:33	Subject Test Refused
Breath Volume	0.101	08:33	
Air Blank	0.000	08:34	
Control Test	0.073	08:34	Control Outside Tolerance
Air Blank	0.000	08:35	

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARGIA,FRANK,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TARGIA,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: FERRO SANCHEZ

DATE/TIME OF TEST: 07/25/2018 20:13:58

Diagnostics Check	OK	20:10	
Air Blank	0.000	20:10	
Control Test	0.079	20:10	
Air Blank	0.000	20:11	
Subject Sample	# 1	20:12	Subject Test Refused
Breath Volume	0.000	20:12	
Air Blank	0.000	20:13	
Control Test	0.063	20:13	Interferent Detect
Air Blank		20:13	Ambient Fail

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,DAMIAN,N
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CASTILLO,ISMAELIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: LIEBES

DATE/TIME OF TEST: 07/27/2018 20:24:34

Diagnostics Check	OK	20:20	
Air Blank	0.000	20:20	
Control Test	0.085	20:20	
Air Blank	0.000	20:21	
Subject Sample	# 1	20:24	Sequence Aborted
Breath Volume	0.000	20:24	
Air Blank	0.000	20:24	

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,DAMIAN,N
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CASTILLO,ISMAELIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: LIEBES

DATE/TIME OF TEST: 07/27/2018 20:57:27

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:46	
Control Test	0.080	20:46	
Air Blank	0.000	20:47	
Subject Sample	# 1	20:50	No Sample Provided
Breath Volume	0.000	20:50	
Air Blank	0.000	20:50	
Air Blank	0.000	20:52	
Subject Sample	# 2	20:56	No Sample Provided
Breath Volume	0.000	20:56	
Air Blank	0.000	20:56	
Control Test	0.080	20:56	
Air Blank	0.000	20:57	
Diagnostics Check	OK	20:57	

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,LEONARDO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DIAZ,LEONARDO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 07:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: BEDOYA

DATE/TIME OF TEST: 07/29/2018 08:48:55

Diagnostics Check	OK	08:45	
Air Blank	0.000	08:45	
Control Test	0.080	08:46	
Air Blank	0.000	08:46	
Subject Sample	# 1	08:47	Subject Test Refused
Breath Volume	0.000	08:47	
Air Blank	0.000	08:48	
Control Test	0.080	08:48	
Air Blank	0.000	08:48	
Diagnostics Check	OK	08:48	

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRERO,O,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MCLAUGHLIN,P
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 12:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:47

SUBJECT'S LAST NAME: SAAVEDRA

DATE/TIME OF TEST: 07/29/2018 13:18:04

Diagnostics Check	OK	13:13	
Air Blank	0.000	13:14	
Control Test	0.080	13:14	
Air Blank	0.000	13:15	
Subject Sample	# 1	13:16	Subject Test Refused
Breath Volume	0.000	13:16	
Air Blank	0.000	13:17	
Control Test	0.080	13:17	
Air Blank	0.000	13:17	
Diagnostics Check	OK	13:18	

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WELLESLEY SHELDON,BODIE,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: O'NEAL,D

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 15:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: WELLESLEY SHELDON

DATE/TIME OF TEST: 07/29/2018 16:24:20

Diagnostics Check	OK	16:13	
Air Blank	0.000	16:14	
Control Test	0.081	16:14	
Air Blank	0.000	16:14	
Subject Sample	# 1	16:18	No Sample Provided
Breath Volume	0.000	16:18	
Air Blank	0.000	16:18	
Air Blank	0.000	16:20	
Subject Sample	# 2	16:22	RFI Detect
Breath Volume	0.000	16:22	
Air Blank	0.000	16:23	
Control Test	0.080	16:23	
Air Blank	0.000	16:24	
Diagnostics Check	OK	16:24	

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ONEAL,D,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ONEAL,DERRICK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: WELLESLEY SHELDON

DATE/TIME OF TEST: 07/29/2018 16:16:15

Diagnostics Check	OK	16:13	
Air Blank	0.000	16:13	
Control Test	0.077	16:14	
Air Blank	0.000	16:14	
Subject Sample	# 1	16:14	Subject Test Refused
Breath Volume	0.000	16:14	
Air Blank	0.000	16:15	
Control Test	0.080	16:15	
Air Blank	0.000	16:16	
Diagnostics Check	OK	16:16	

Data Download Date/Time: 07/30/2018 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: ALZAMORA

DATE/TIME OF TEST: 07/01/2018 05:49:12

Diagnostics Check	OK	05:42
Air Blank	0.000	05:42
Control Test	0.079	05:43
Air Blank	0.000	05:43
Subject Sample # 1	0.155	05:44
Breath Volume	2.199	05:44
Air Blank	0.000	05:45
Air Blank	0.000	05:46
Subject Sample # 2	0.146	05:47
Breath Volume	1.621	05:47
Air Blank	0.000	05:48
Control Test	0.078	05:48
Air Blank	0.000	05:49
Diagnostics Check	OK	05:49

Data Download Date/Time: 07/11/2018 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEOHANE,KUNIKO,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: KEOHANE,KUNIKO
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:31

SUBJECT'S LAST NAME: CASKEY

DATE/TIME OF TEST: 07/02/2018 12:58:33

Diagnostics Check	OK	12:55	
Air Blank	0.000	12:56	
Control Test	0.079	12:56	
Air Blank	0.000	12:57	
Subject Sample	# 1	12:57	Subject Test Refused
Breath Volume	0.000	12:57	
Air Blank	0.000	12:57	
Control Test	0.079	12:57	
Air Blank	0.000	12:58	
Diagnostics Check	OK	12:58	

Data Download Date/Time: 07/11/2018 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONATY,JAY,T
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: CONATY,JAY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: HONEYCUTT

DATE/TIME OF TEST: 07/04/2018 00:52:59

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.079	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 1 0.124	00:49	
Breath Volume	1.582	00:49	
Air Blank	0.000	00:50	
Air Blank	0.000	00:51	
Subject Sample	# 2	00:52	RFI Detect
Breath Volume	0.000	00:52	
Air Blank		00:52	RFI Detect
Air Blank		00:52	RFI Detect

Data Download Date/Time: 07/11/2018 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONATY,JAY,T
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: CONATY,JAY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: HONEYCUTT

DATE/TIME OF TEST: 07/04/2018 01:02:40

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.078	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.122	00:57
Breath Volume	1.437	00:57
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.117	01:00
Breath Volume	1.519	01:00
Air Blank	0.000	01:01
Control Test	0.076	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 07/11/2018 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEOHANE,KUNIKO,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LANE,JOSEPH
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: BLANCO

DATE/TIME OF TEST: 07/07/2018 06:03:46

Diagnostics Check	OK	06:00	
Air Blank	0.000	06:01	
Control Test	0.079	06:01	
Air Blank	0.000	06:02	
Subject Sample	# 1	06:02	Subject Test Refused
Breath Volume	0.000	06:02	
Air Blank	0.000	06:02	
Control Test	0.079	06:03	
Air Blank	0.000	06:03	
Diagnostics Check	OK	06:03	

Data Download Date/Time: 07/11/2018 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 07/12/2018 01:11:07

Diagnostics Check	OK	01:08	
Air Blank	0.000	01:08	
Control Test	0.078	01:09	
Air Blank	0.000	01:09	
Subject Sample	# 1	01:09	Subject Test Refused
Breath Volume	0.000	01:09	
Air Blank	0.000	01:10	
Control Test	0.078	01:10	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 07/15/2018 14:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI, RYAN, P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LANE, JONATHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: BARTON

DATE/TIME OF TEST: 07/12/2018 02:06:26

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:03	
Control Test	0.079	02:04	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:05	Subject Test Refused
Breath Volume	0.000	02:05	
Air Blank	0.000	02:05	
Control Test	0.079	02:05	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

Data Download Date/Time: 07/15/2018 14:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRICKE,KENNETH,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FRICKE,KENNETH
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 07/12/2018 09:55:02

Diagnostics Check	OK	09:52	
Air Blank	0.000	09:52	
Control Test	0.079	09:52	
Air Blank	0.000	09:53	
Subject Sample	# 1	09:53	Subject Test Refused
Breath Volume	0.000	09:53	
Air Blank	0.000	09:54	
Control Test	0.079	09:54	
Air Blank	0.000	09:54	
Diagnostics Check	OK	09:55	

Data Download Date/Time: 07/15/2018 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONATY,JAY,T
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: CONATY,JAY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: NASH

DATE/TIME OF TEST: 07/13/2018 01:15:36

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.079	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.076	01:11
Breath Volume	2.714	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.074	01:13
Breath Volume	2.660	01:13
Air Blank	0.000	01:14
Control Test	0.077	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 07/15/2018 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: LECOUMPTÉ

DATE/TIME OF TEST: 07/13/2018 20:28:22

Diagnostics Check	OK	20:22
Air Blank	0.000	20:22
Control Test	0.079	20:22
Air Blank	0.000	20:23
Subject Sample # 1	0.239	20:23
Breath Volume	2.957	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.223	20:26
Breath Volume	2.328	20:26
Air Blank	0.000	20:27
Control Test	0.078	20:27
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

Data Download Date/Time: 07/15/2018 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,BRIAN,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SAPP,BRIAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: BUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: HAHN

DATE/TIME OF TEST: 07/15/2018 21:42:27

Diagnostics Check	OK	21:31	
Air Blank	0.000	21:31	
Control Test	0.078	21:31	
Air Blank	0.000	21:32	
Subject Sample	# 1	21:35	No Sample Provided
Breath Volume	0.000	21:35	
Air Blank	0.000	21:35	
Air Blank	0.000	21:37	
Subject Sample	# 2	21:41	No Sample Provided
Breath Volume	0.000	21:41	
Air Blank	0.000	21:41	
Control Test	0.078	21:41	
Air Blank	0.000	21:42	
Diagnostics Check	OK	21:42	

Data Download Date/Time: 08/02/2018 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMSON, BRANDEE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: EASON, ZAC
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 07/15/2018 21:57:39

Diagnostics Check	OK	21:49
Air Blank	0.000	21:49
Control Test	0.079	21:49
Air Blank	0.000	21:50
Subject Sample # 1	0.139	21:52
Breath Volume	1.617	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.138	21:55
Breath Volume	1.507	21:55
Air Blank	0.000	21:56
Control Test	0.076	21:57
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 08/02/2018 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,SHAWN,S
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: STEELE,SHAWN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: POFF JR

DATE/TIME OF TEST: 07/16/2018 20:42:39

Diagnostics Check	OK	20:31	
Air Blank	0.000	20:31	
Control Test	0.078	20:31	
Air Blank	0.000	20:32	
Subject Sample	# 1 0.228	20:32	
Breath Volume	1.996	20:32	
Air Blank	0.000	20:33	
Air Blank	0.000	20:35	
Subject Sample	# 2 0.267	20:35	
Breath Volume	1.566	20:35	
Air Blank	0.000	20:36	
Air Blank	0.000	20:38	
Subject Sample	# 3 0.246	20:40	
Breath Volume	3.863	20:40	
Air Blank	0.000	20:41	
Control Test	0.076	20:41	
Air Blank		20:42	RFI Detect
Air Blank	0.000	20:42	

Data Download Date/Time: 08/02/2018 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,SHAWN,S
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: STEELE,SHAWN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: POFF JR

DATE/TIME OF TEST: 07/16/2018 20:49:11

Diagnostics Check	OK	20:44	
Air Blank	0.000	20:44	
Control Test	0.075	20:45	
Air Blank	0.000	20:45	
Subject Sample	# 1	20:47	Subject Test Refused
Breath Volume	0.000	20:47	
Air Blank	0.000	20:48	
Control Test	0.076	20:48	
Air Blank	0.000	20:49	
Diagnostics Check	OK	20:49	

Data Download Date/Time: 08/02/2018 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ALEXANDRA,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DAVIS,ALEXANDRA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: LEIGH

DATE/TIME OF TEST: 07/19/2018 18:29:08

Diagnostics Check	OK	18:26	
Air Blank	0.000	18:26	
Control Test	0.079	18:26	
Air Blank	0.000	18:27	
Subject Sample	# 1	18:27	Subject Test Refused
Breath Volume	0.000	18:27	
Air Blank	0.000	18:28	
Control Test	0.078	18:28	
Air Blank	0.000	18:29	
Diagnostics Check	OK	18:29	

Data Download Date/Time: 08/02/2018 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMSON, BRANDEE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: WILLIAMSON, BRANDEE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: TAMA

DATE/TIME OF TEST: 07/21/2018 20:48:04

Diagnostics Check	OK	20:40
Air Blank	0.000	20:40
Control Test	0.078	20:41
Air Blank	0.000	20:41
Subject Sample # 1	0.163	20:42
Breath Volume	1.308	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:44
Subject Sample # 2	0.182	20:46
Breath Volume	3.015	20:46
Air Blank	0.000	20:47
Control Test	0.078	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:48

Data Download Date/Time: 08/02/2018 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ALEXANDRA,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DAVIS,ALEXANDRA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:46

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/22/2018 16:22:57

Diagnostics Check	OK	16:07	
Air Blank	0.000	16:08	
Control Test	0.079	16:08	
Air Blank	0.000	16:08	
Subject Sample	# 1 0.157	16:11	
Breath Volume	1.824	16:11	
Air Blank	0.000	16:12	
Air Blank	0.000	16:14	
Subject Sample	# 2 0.143	16:17	Volume Not Met
Breath Volume	0.753	16:17	
Air Blank	0.000	16:18	
Air Blank	0.000	16:19	
Subject Sample	# 3 0.155	16:21	
Breath Volume	1.636	16:21	
Air Blank	0.000	16:22	
Control Test	0.076	16:22	
Air Blank	0.000	16:22	
Diagnostics Check	OK	16:22	

Data Download Date/Time: 08/02/2018 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARTIN-REYES,JUAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: EASTER

DATE/TIME OF TEST: 07/25/2018 17:12:42

Diagnostics Check	OK	17:05
Air Blank	0.000	17:06
Control Test	0.079	17:06
Air Blank	0.000	17:06
Subject Sample # 1	0.000	17:07
Breath Volume	1.933	17:07
Air Blank	0.000	17:07
Air Blank	0.000	17:09
Subject Sample # 2	0.000	17:11
Breath Volume	2.636	17:11
Air Blank	0.000	17:11
Control Test	0.079	17:12
Air Blank	0.000	17:12
Diagnostics Check	OK	17:12

Data Download Date/Time: 08/02/2018 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,SHAWN,S
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ARAUJO,CHRISTOPHER
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: BUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: PFEIFFER

DATE/TIME OF TEST: 07/26/2018 02:20:54

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.107	02:16
Breath Volume	2.976	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:18
Subject Sample # 2	0.089	02:19
Breath Volume	2.503	02:19
Air Blank	0.000	02:19
Control Test	0.076	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 08/02/2018 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERDE,SHIRLEY,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: ABRAMSON,SHELBY
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: RIVERO RODRIGUEZ

DATE/TIME OF TEST: 07/05/2018 02:25:51

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:23	
Control Test	0.078	02:23	
Air Blank	0.000	02:24	
Subject Sample	# 1	02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:24	
Control Test	0.078	02:25	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

Data Download Date/Time: 07/12/2018 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRIGA,JUAN,C
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: ABRAMSON,SHELBY
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: LECHADO BUITRAGO

DATE/TIME OF TEST: 07/29/2018 04:02:59

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:51	
Control Test	0.078	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:55	No Sample Provided
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Air Blank	0.000	03:57	
Subject Sample	# 2	0.190	03:58
Breath Volume	1.960	03:58	
Air Blank	0.000	03:59	
Air Blank	0.000	04:00	
Subject Sample	# 3	0.190	04:01
Breath Volume	1.371	04:01	
Air Blank	0.000	04:02	
Control Test	0.076	04:02	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 08/31/2018 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 07/03/2018 23:09:46

Diagnostics Check	OK	23:02
Air Blank	0.000	23:03
Control Test	0.076	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.212	23:05
Breath Volume	2.289	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.204	23:08
Breath Volume	2.082	23:08
Air Blank	0.000	23:08
Control Test	0.076	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 07/13/2018 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: GIGLIOTTI KUNKEN

DATE/TIME OF TEST: 07/04/2018 02:20:55

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.076	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.185	02:16
Breath Volume	2.277	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.192	02:19
Breath Volume	2.277	02:19
Air Blank	0.000	02:19
Control Test	0.076	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 07/13/2018 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: CLAUDIO

DATE/TIME OF TEST: 07/09/2018 01:35:32

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.077	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.164	01:30
Breath Volume	1.460	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.160	01:33
Breath Volume	1.437	01:33
Air Blank	0.000	01:34
Control Test	0.077	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 07/13/2018 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: MC BRIDE

DATE/TIME OF TEST: 07/10/2018 00:13:48

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.076	00:08
Air Blank	0.000	00:08
Subject Sample # 1	0.078	00:09
Breath Volume	2.382	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.077	00:12
Breath Volume	2.933	00:12
Air Blank	0.000	00:12
Control Test	0.076	00:13
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 07/13/2018 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: AKERS,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:02

SUBJECT'S LAST NAME: TEJEIRA

DATE/TIME OF TEST: 07/13/2018 13:33:08

Diagnostics Check	OK	13:26
Air Blank	0.000	13:27
Control Test	0.077	13:27
Air Blank	0.000	13:28
Subject Sample # 1	0.346	13:28
Breath Volume	1.671	13:28
Air Blank	0.000	13:29
Air Blank	0.000	13:31
Subject Sample # 2	0.331	13:31
Breath Volume	1.753	13:31
Air Blank	0.000	13:32
Control Test	0.077	13:32
Air Blank	0.000	13:33
Diagnostics Check	OK	13:33

Data Download Date/Time: 08/29/2018 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: MINAYA

DATE/TIME OF TEST: 07/15/2018 21:00:10

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.076	20:54
Air Blank	0.000	20:54
Subject Sample # 1	0.187	20:55
Breath Volume	2.687	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.179	20:58
Breath Volume	2.468	20:58
Air Blank	0.000	20:59
Control Test	0.076	20:59
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 08/29/2018 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: VOELKER

DATE/TIME OF TEST: 07/15/2018 22:41:51

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:33	
Control Test	0.077	22:33	
Air Blank	0.000	22:34	
Subject Sample	# 1 0.161	22:36	
Breath Volume	1.187	22:36	
Air Blank	0.000	22:37	
Air Blank	0.000	22:39	
Subject Sample	# 2	22:40	RFI Detect
Breath Volume	0.500	22:40	
Air Blank	0.000	22:40	
Control Test	0.075	22:41	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:41	

Data Download Date/Time: 08/29/2018 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: VOELKER

DATE/TIME OF TEST: 07/15/2018 22:51:45

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:49	
Control Test	0.076	22:49	
Air Blank	0.000	22:50	
Subject Sample	# 1	22:50	RFI Detect
Breath Volume	0.000	22:50	
Air Blank	0.000	22:50	
Control Test	0.075	22:51	
Air Blank	0.000	22:51	
Diagnostics Check	OK	22:51	

Data Download Date/Time: 08/29/2018 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: VALDEZ HIRIART

DATE/TIME OF TEST: 07/16/2018 03:15:06

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.076	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.322	03:09
Breath Volume	1.542	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:11
Subject Sample # 2	0.326	03:13
Breath Volume	1.605	03:13
Air Blank	0.000	03:14
Control Test	0.075	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 08/29/2018 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: RODRIGUEZ VELASQUEZ

DATE/TIME OF TEST: 07/22/2018 21:10:31

Diagnostics Check	OK	21:02
Air Blank	0.000	21:02
Control Test	0.077	21:03
Air Blank	0.000	21:03
Subject Sample # 1	0.127	21:05
Breath Volume	1.628	21:05
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.125	21:08
Breath Volume	1.472	21:08
Air Blank	0.000	21:09
Control Test	0.075	21:09
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 08/29/2018 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: MORIGENE

DATE/TIME OF TEST: 07/22/2018 23:44:48

Diagnostics Check	OK	23:37
Air Blank	0.000	23:38
Control Test	0.076	23:38
Air Blank	0.000	23:38
Subject Sample # 1	0.171	23:39
Breath Volume	1.656	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.174	23:43
Breath Volume	1.460	23:43
Air Blank	0.000	23:43
Control Test	0.076	23:44
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 08/29/2018 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:22

SUBJECT'S LAST NAME: MUIRHEAD

DATE/TIME OF TEST: 07/23/2018 06:57:45

Diagnostics Check	OK	06:48
Air Blank	0.000	06:48
Control Test	0.077	06:49
Air Blank	0.000	06:49
Subject Sample # 1	0.185	06:52
Breath Volume	2.394	06:52
Air Blank	0.000	06:53
Air Blank	0.000	06:55
Subject Sample # 2	0.183	06:56
Breath Volume	1.292	06:56
Air Blank	0.000	06:56
Control Test	0.075	06:57
Air Blank	0.000	06:57
Diagnostics Check	OK	06:57

Data Download Date/Time: 08/29/2018 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VALERIO,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: SCHARF

DATE/TIME OF TEST: 07/14/2018 00:42:49

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.277	00:38
Breath Volume	1.320	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.261	00:41
Breath Volume	1.242	00:41
Air Blank	0.000	00:41
Control Test	0.079	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 08/10/2018 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LOMBARDO

DATE/TIME OF TEST: 07/14/2018 02:31:38

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.118	02:26
Breath Volume	3.015	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.115	02:29
Breath Volume	2.187	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 08/10/2018 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CAVE

DATE/TIME OF TEST: 07/14/2018 03:47:07

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.196	03:42
Breath Volume	2.863	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:44
Subject Sample # 2	0.201	03:45
Breath Volume	2.824	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 08/10/2018 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: NAVA

DATE/TIME OF TEST: 07/14/2018 04:12:30

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.166	04:07
Breath Volume	3.472	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.159	04:10
Breath Volume	3.472	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 08/10/2018 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 07/15/2018 01:47:00

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.080	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.000	01:42
Breath Volume	1.210	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.000	01:45
Breath Volume	1.226	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:47

Data Download Date/Time: 08/10/2018 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 07/15/2018 03:15:02

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.080	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.081	03:10
Breath Volume	4.906	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.083	03:13
Breath Volume	4.535	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:15

Data Download Date/Time: 08/10/2018 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: ZORRILLA RODRIGUEZ

DATE/TIME OF TEST: 07/15/2018 06:00:00

Diagnostics Check	OK	05:52
Air Blank	0.000	05:52
Control Test	0.079	05:52
Air Blank	0.000	05:53
Subject Sample # 1	0.137	05:55
Breath Volume	1.226	05:55
Air Blank	0.000	05:55
Air Blank	0.000	05:57
Subject Sample # 2	0.129	05:58
Breath Volume	1.593	05:58
Air Blank	0.000	05:59
Control Test	0.078	05:59
Air Blank	0.000	05:59
Diagnostics Check	OK	06:00

Data Download Date/Time: 08/10/2018 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JOY,DAVID
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 07/15/2018 15:47:45

Diagnostics Check	OK	15:39
Air Blank	0.000	15:40
Control Test	0.081	15:40
Air Blank	0.000	15:41
Subject Sample # 1	0.000	15:41
Breath Volume	2.683	15:41
Air Blank	0.000	15:42
Air Blank	0.000	15:44
Subject Sample # 2	0.000	15:46
Breath Volume	1.898	15:46
Air Blank	0.000	15:46
Control Test	0.081	15:47
Air Blank	0.000	15:47
Diagnostics Check	OK	15:47

Data Download Date/Time: 08/10/2018 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,KEITH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: LAFOND

DATE/TIME OF TEST: 07/15/2018 19:33:06

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.080	19:26
Air Blank	0.000	19:26
Subject Sample # 1	0.172	19:28
Breath Volume	1.160	19:28
Air Blank	0.000	19:29
Air Blank	0.000	19:30
Subject Sample # 2	0.185	19:31
Breath Volume	1.152	19:31
Air Blank	0.000	19:32
Control Test	0.078	19:32
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 08/10/2018 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: LOPEZ PEREZ

DATE/TIME OF TEST: 07/15/2018 21:11:17

Diagnostics Check	OK	21:04
Air Blank	0.000	21:05
Control Test	0.080	21:05
Air Blank	0.000	21:06
Subject Sample # 1	0.275	21:06
Breath Volume	1.636	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.275	21:09
Breath Volume	1.273	21:09
Air Blank	0.000	21:10
Control Test	0.079	21:10
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCCRAY II,MYERS
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: HUTCHESON

DATE/TIME OF TEST: 07/18/2018 01:48:40

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.205	01:43
Breath Volume	2.101	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.207	01:47
Breath Volume	2.984	01:47
Air Blank	0.000	01:47
Control Test	0.079	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DOWLING,DAVID
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: MELLEN JR

DATE/TIME OF TEST: 07/18/2018 21:33:17

Diagnostics Check	OK	21:31	
Air Blank	0.000	21:31	
Control Test	0.080	21:31	
Air Blank	0.000	21:32	
Subject Sample	# 1	21:32	RFI Detect
Breath Volume	0.000	21:32	
Air Blank		21:32	RFI Detect
Air Blank	0.000	21:33	

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DOWLING,DAVID
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: MELLEN JR

DATE/TIME OF TEST: 07/18/2018 21:42:45

Diagnostics Check	OK	21:35
Air Blank	0.000	21:35
Control Test	0.080	21:35
Air Blank	0.000	21:36
Subject Sample # 1	0.217	21:37
Breath Volume	2.593	21:37
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.219	21:41
Breath Volume	2.839	21:41
Air Blank	0.000	21:41
Control Test	0.079	21:42
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHQUITIO
RODRIGUEZ,LENNDEY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male

OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: QUAGLIA

DATE/TIME OF TEST: 07/20/2018 04:55:31

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.081	04:49
Air Blank	0.000	04:49
Subject Sample # 1	0.206	04:50
Breath Volume	1.203	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.213	04:53
Breath Volume	1.960	04:53
Air Blank	0.000	04:54
Control Test	0.078	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: LUNEKE

DATE/TIME OF TEST: 07/21/2018 02:58:27

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.165	02:53
Breath Volume	2.593	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.169	02:56
Breath Volume	2.382	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DENIS,ISMARY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: DELAURA

DATE/TIME OF TEST: 07/21/2018 03:30:34

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.039	03:25
Breath Volume	3.968	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.036	03:28
Breath Volume	2.820	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TRUDEAU,RYAN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MARTIN JIMENEZ

DATE/TIME OF TEST: 07/21/2018 03:50:21

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.155	03:45
Breath Volume	1.425	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.153	03:48
Breath Volume	1.425	03:48
Air Blank	0.000	03:49
Control Test	0.079	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VALAREZO,MEGAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SEARS

DATE/TIME OF TEST: 07/21/2018 23:27:37

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.081	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.115	23:22
Breath Volume	2.386	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:24
Subject Sample # 2	0.116	23:25
Breath Volume	2.343	23:25
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: QUESADA,YOSVANI
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: PONCHOT

DATE/TIME OF TEST: 07/22/2018 01:32:32

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.157	01:27
Breath Volume	2.660	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.147	01:30
Breath Volume	1.250	01:30
Air Blank	0.000	01:31
Control Test	0.079	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GARCIA,KIMBERLY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: JASTRJEMBSKAIA

DATE/TIME OF TEST: 07/22/2018 20:52:07

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:47	
Control Test	0.080	20:47	
Air Blank	0.000	20:48	
Subject Sample	# 1	20:50	RFI Detect
Breath Volume	0.121	20:50	
Air Blank	0.000	20:51	
Control Test	0.081	20:51	
Air Blank	0.000	20:52	
Diagnostics Check	OK	20:52	

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GARCIA,KIMBERLY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: JASTRJEMBSKAIA

DATE/TIME OF TEST: 07/22/2018 21:06:31

Diagnostics Check	OK	20:53
Air Blank	0.000	20:54
Control Test	0.080	20:54
Air Blank	0.000	20:54
Subject Sample # 1	0.142	20:57
Breath Volume	1.449	20:57
Air Blank	0.000	20:58
Air Blank	0.000	20:59
Subject Sample # 2	0.171	21:01
Breath Volume	1.992	21:01
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 3	0.171	21:04
Breath Volume	2.046	21:04
Air Blank	0.000	21:05
Control Test	0.079	21:05
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LOPEZ,BRANDON
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: BARRESI

DATE/TIME OF TEST: 07/23/2018 02:03:46

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.080	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.074	01:59
Breath Volume	2.546	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.073	02:02
Breath Volume	2.832	02:02
Air Blank	0.000	02:02
Control Test	0.081	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CAMACHO

DATE/TIME OF TEST: 07/23/2018 03:52:46

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.222	03:48
Breath Volume	1.988	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.225	03:51
Breath Volume	1.351	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ,ROBERTO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: ZEPETA

DATE/TIME OF TEST: 07/23/2018 21:51:38

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:46	
Control Test	0.081	21:46	
Air Blank	0.000	21:47	
Subject Sample	# 1	21:49	Subject Test Refused
Breath Volume	1.019	21:49	
Air Blank	0.000	21:50	
Control Test	0.081	21:50	
Air Blank	0.000	21:51	
Diagnostics Check	OK	21:51	

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SANDERS II

DATE/TIME OF TEST: 07/25/2018 23:49:32

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.182	23:44
Breath Volume	2.343	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.195	23:47
Breath Volume	3.789	23:47
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO
RODRIGUEZ,LENNDEY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male

OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: SOTELO GONZALEZ

DATE/TIME OF TEST: 07/27/2018 01:12:09

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.277	01:07
Breath Volume	1.152	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.286	01:10
Breath Volume	1.722	01:10
Air Blank	0.000	01:11
Control Test	0.079	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VALERIO,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: YEARTY

DATE/TIME OF TEST: 07/28/2018 02:38:53

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.167	02:34
Breath Volume	2.726	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.163	02:37
Breath Volume	2.574	02:37
Air Blank	0.000	02:37
Control Test	0.079	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: RIVERON GARCIA

DATE/TIME OF TEST: 07/28/2018 03:30:17

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.079	03:16	
Air Blank	0.000	03:16	
Subject Sample	# 1 0.162	03:17	
Breath Volume	1.214	03:17	
Air Blank	0.000	03:18	
Air Blank	0.000	03:19	
Subject Sample	# 2 0.116	03:22	Volume Not Met
Breath Volume	0.562	03:22	
Air Blank	0.000	03:23	
Air Blank	0.000	03:25	
Subject Sample	# 3 0.117	03:28	Volume Not Met
Breath Volume	0.480	03:28	
Air Blank	0.000	03:29	
Control Test	0.077	03:29	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STREIGOLD,DAVID
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MINOR

DATE/TIME OF TEST: 07/29/2018 01:47:14

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.172	01:42
Breath Volume	1.398	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.170	01:45
Breath Volume	2.582	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: HERRERA CUTZ

DATE/TIME OF TEST: 07/29/2018 03:51:14

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.226	03:45
Breath Volume	1.351	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.214	03:49
Breath Volume	1.910	03:49
Air Blank	0.000	03:50
Control Test	0.076	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MIRANDA,EDWIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: JEAN PAUL

DATE/TIME OF TEST: 07/29/2018 06:32:30

Diagnostics Check	OK	06:20	
Air Blank	0.000	06:21	
Control Test	0.080	06:21	
Air Blank	0.000	06:21	
Subject Sample	# 1 0.197	06:25	Volume Not Met
Breath Volume	0.136	06:25	
Air Blank	0.000	06:25	
Air Blank	0.000	06:27	
Subject Sample	# 2 0.180	06:30	Volume Not Met
Breath Volume	0.894	06:30	
Air Blank	0.000	06:31	
Control Test	0.078	06:31	
Air Blank	0.000	06:32	
Diagnostics Check	OK	06:32	

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,KEITH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 07/29/2018 19:25:56

Diagnostics Check	OK	19:19
Air Blank	0.000	19:19
Control Test	0.079	19:19
Air Blank	0.000	19:20
Subject Sample # 1	0.232	19:21
Breath Volume	2.421	19:21
Air Blank	0.000	19:22
Air Blank	0.000	19:23
Subject Sample # 2	0.230	19:24
Breath Volume	2.410	19:24
Air Blank	0.000	19:24
Control Test	0.079	19:25
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FERGUSON,RYAN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: ZIMMER JR

DATE/TIME OF TEST: 07/30/2018 22:13:15

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.080	22:07
Air Blank	0.000	22:07
Subject Sample # 1	0.127	22:08
Breath Volume	3.292	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.116	22:11
Breath Volume	2.832	22:11
Air Blank	0.000	22:12
Control Test	0.080	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: BERNARD

DATE/TIME OF TEST: 07/31/2018 22:48:39

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.080	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.086	22:43
Breath Volume	1.226	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.088	22:46
Breath Volume	1.281	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 08/10/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPURLOCK,ROBERT,V
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPURLOCK,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 07/05/2018 18:19:00

Diagnostics Check	OK	18:13	
Air Blank	0.000	18:13	
Control Test	0.077	18:13	
Air Blank	0.000	18:14	
Subject Sample # 1	0.000	18:14	
Breath Volume	1.597	18:14	
Air Blank	0.000	18:15	
Air Blank	0.000	18:17	
Subject Sample # 2	0.000	18:17	
Breath Volume	1.855	18:17	
Air Blank	0.000	18:18	
Control Test	0.080	18:18	Interferent Detect
Air Blank	0.000	18:19	

Data Download Date/Time: 07/24/2018 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPURLOCK,ROBERT,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPURLOCK,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 07/05/2018 18:21:51

Diagnostics Check	OK	18:20	
Air Blank	0.000	18:20	
Control Test	0.079	18:21	Interferent Detect
Air Blank	0.000	18:21	

Data Download Date/Time: 07/24/2018 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPURLOCK,ROBERT,V
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPURLOCK,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 07/05/2018 18:24:56

Diagnostics Check	OK	18:23	
Air Blank	0.000	18:24	
Control Test	0.080	18:24	Interferent Detect
Air Blank	0.000	18:24	

Data Download Date/Time: 07/24/2018 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPURLOCK,ROBERT,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPURLOCK,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 07/05/2018 19:04:46

Diagnostics Check	OK	18:59	
Air Blank	0.000	18:59	
Control Test	0.077	18:59	
Air Blank	0.000	19:00	
Subject Sample # 1	0.000	19:00	
Breath Volume	1.777	19:00	
Air Blank	0.000	19:01	
Air Blank	0.000	19:03	
Subject Sample # 2	0.000	19:03	
Breath Volume	1.886	19:03	
Air Blank	0.000	19:03	
Control Test	0.081	19:04	Interferent Detect
Air Blank	0.000	19:04	

Data Download Date/Time: 07/24/2018 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 07/05/2018 21:06:24

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:00	
Control Test	0.078	21:00	
Air Blank	0.000	21:01	
Subject Sample # 1	0.000	21:02	
Breath Volume	3.402	21:02	
Air Blank	0.000	21:02	
Air Blank	0.000	21:04	
Subject Sample # 2	0.000	21:04	
Breath Volume	3.000	21:04	
Air Blank	0.000	21:05	
Control Test	0.082	21:05	Interferent Detect
Air Blank	0.000	21:06	

Data Download Date/Time: 07/24/2018 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 07/05/2018 21:26:34

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:21	
Control Test	0.077	21:21	
Air Blank	0.000	21:21	
Subject Sample # 1	0.000	21:22	
Breath Volume	2.179	21:22	
Air Blank	0.000	21:22	
Air Blank	0.000	21:24	
Subject Sample # 2		21:25	Interferent Detect
Breath Volume	2.453	21:25	
Air Blank	0.000	21:25	
Control Test	0.079	21:25	Interferent Detect
Air Blank	0.000	21:26	

Data Download Date/Time: 07/24/2018 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 07/09/2018 16:16:34

Diagnostics Check	OK	16:10
Air Blank	0.000	16:11
Control Test	0.077	16:11
Air Blank	0.000	16:12
Subject Sample # 1	0.000	16:12
Breath Volume	2.207	16:12
Air Blank	0.000	16:12
Air Blank	0.000	16:14
Subject Sample # 2	0.000	16:15
Breath Volume	2.144	16:15
Air Blank	0.000	16:15
Control Test	0.076	16:15
Air Blank	0.000	16:16
Diagnostics Check	OK	16:16

Data Download Date/Time: 07/24/2018 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEVEZ,JOSE,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ESTEVEZ,JOSE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: HYMAN

DATE/TIME OF TEST: 07/21/2018 01:07:26

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.081	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.175	01:02
Breath Volume	3.218	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.160	01:05
Breath Volume	2.574	01:05
Air Blank	0.000	01:06
Control Test	0.076	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 08/21/2018 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE, RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: REFUSE, RONY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 07/21/2018 02:02:03

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.123	01:57
Breath Volume	2.078	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.124	02:00
Breath Volume	1.707	02:00
Air Blank	0.000	02:01
Control Test	0.080	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:02

Data Download Date/Time: 08/21/2018 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIZOT,ALEXANDRA,P
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BIZOT,ALEXANDRA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/21/2018 07:25:44

Diagnostics Check	OK	07:19
Air Blank	0.000	07:19
Control Test	0.080	07:20
Air Blank	0.000	07:20
Subject Sample # 1	0.137	07:21
Breath Volume	2.410	07:21
Air Blank	0.000	07:22
Air Blank	0.000	07:23
Subject Sample # 2	0.148	07:24
Breath Volume	1.117	07:24
Air Blank	0.000	07:24
Control Test	0.078	07:25
Air Blank	0.000	07:25
Diagnostics Check	OK	07:25

Data Download Date/Time: 08/21/2018 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIZOT,ALEXANDRA,P
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BIZOT,ALEXANDRA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/26/2018 02:20:46

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.000	02:16
Breath Volume	1.660	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.000	02:19
Breath Volume	2.246	02:19
Air Blank	0.000	02:19
Control Test	0.078	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 08/21/2018 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIDSON,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: DAVIDSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: BUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:38

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 07/28/2018 18:08:43

Diagnostics Check	OK	18:02
Air Blank	0.000	18:03
Control Test	0.080	18:03
Air Blank	0.000	18:03
Subject Sample # 1	0.000	18:04
Breath Volume	2.882	18:04
Air Blank	0.000	18:04
Air Blank	0.000	18:06
Subject Sample # 2	0.000	18:07
Breath Volume	3.789	18:07
Air Blank	0.000	18:07
Control Test	0.080	18:08
Air Blank	0.000	18:08
Diagnostics Check	OK	18:08

Data Download Date/Time: 08/21/2018 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCLJOSE,F
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TUCCLJOSE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: GIL

DATE/TIME OF TEST: 07/30/2018 00:48:08

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:41	
Control Test	0.081	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1 0.165	00:43	
Breath Volume	3.226	00:43	
Air Blank	0.000	00:44	
Air Blank	0.000	00:46	
Subject Sample	# 2 0.173	00:46	
Breath Volume	3.574	00:46	
Air Blank	0.000	00:47	
Control Test	0.080	00:47	
Air Blank		00:47	RFI Detect
Air Blank	0.000	00:48	

Data Download Date/Time: 08/21/2018 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GWODZ,OFC
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: KASMER

DATE/TIME OF TEST: 07/31/2018 21:39:01

Diagnostics Check	OK	21:36	
Air Blank	0.000	21:36	
Control Test	0.081	21:36	
Air Blank	0.000	21:37	
Subject Sample	# 1	21:37	Subject Test Refused
Breath Volume	0.000	21:37	
Air Blank	0.000	21:38	
Control Test	0.081	21:38	
Air Blank	0.000	21:38	
Diagnostics Check	OK	21:39	

Data Download Date/Time: 08/21/2018 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: LAZARO ROMERO

DATE/TIME OF TEST: 07/01/2018 01:33:47

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.081	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.184	01:28
Breath Volume	1.207	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.169	01:32
Breath Volume	1.152	01:32
Air Blank	0.000	01:32
Control Test	0.080	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 07/12/2018 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: GRIZZARD

DATE/TIME OF TEST: 07/01/2018 02:21:51

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.080	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.236	02:17
Breath Volume	2.382	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.228	02:20
Breath Volume	1.808	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 07/12/2018 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PALFY,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: HOLLINS

DATE/TIME OF TEST: 07/01/2018 03:49:37

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:34	
Control Test	0.080	03:35	
Air Blank	0.000	03:35	
Subject Sample	# 1 0.126	03:38	Volume Not Met
Breath Volume	0.238	03:38	
Air Blank	0.000	03:39	
Air Blank	0.000	03:41	
Subject Sample	# 2 0.125	03:42	
Breath Volume	1.101	03:42	
Air Blank	0.000	03:43	
Air Blank	0.000	03:44	
Subject Sample	# 3 0.120	03:47	
Breath Volume	1.351	03:47	
Air Blank	0.000	03:48	
Control Test	0.079	03:48	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 07/12/2018 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: HURTADO TREJO

DATE/TIME OF TEST: 07/01/2018 21:29:36

Diagnostics Check	OK	21:26	
Air Blank	0.000	21:26	
Control Test	0.080	21:27	
Air Blank	0.000	21:27	
Subject Sample	# 1	21:28	Subject Test Refused
Breath Volume	0.000	21:28	
Air Blank	0.000	21:28	
Control Test	0.080	21:28	
Air Blank	0.000	21:29	
Diagnostics Check	OK	21:29	

Data Download Date/Time: 07/12/2018 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: MORALES-HERNANDEZ

DATE/TIME OF TEST: 07/01/2018 23:34:57

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:22	
Control Test	0.080	23:22	
Air Blank	0.000	23:23	
Subject Sample # 1	0.181	23:24	
Breath Volume	1.945	23:24	
Air Blank	0.000	23:25	
Air Blank	0.000	23:27	
Subject Sample # 2	0.150	23:30	Volume Not Met
Breath Volume	0.539	23:30	
Air Blank	0.000	23:31	
Air Blank	0.000	23:32	
Subject Sample # 3	0.162	23:33	
Breath Volume	1.105	23:33	
Air Blank	0.000	23:34	
Control Test	0.078	23:34	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 07/12/2018 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: SANTIAGO HERNANDEZ

DATE/TIME OF TEST: 07/02/2018 22:24:26

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.202	22:18
Breath Volume	2.019	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:20
Subject Sample # 2	0.197	22:22
Breath Volume	1.308	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 07/12/2018 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRIGHTWELL, R
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SOUZA MATTOS

DATE/TIME OF TEST: 07/04/2018 03:22:56

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.080	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.142	03:18
Breath Volume	1.906	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:20
Subject Sample # 2	0.161	03:21
Breath Volume	1.128	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 07/12/2018 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS, RONALD, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: PORTAL SANCHEZ

DATE/TIME OF TEST: 07/04/2018 04:39:33

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.081	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.171	04:34
Breath Volume	1.968	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.168	04:37
Breath Volume	3.253	04:37
Air Blank	0.000	04:38
Control Test	0.080	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 07/12/2018 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS, RONALD, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 07/05/2018 03:35:02

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.171	03:30
Breath Volume	1.332	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:32
Subject Sample # 2	0.173	03:33
Breath Volume	1.808	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:35

Data Download Date/Time: 07/12/2018 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VELAZQUEZ,B
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: WERMUTH

DATE/TIME OF TEST: 07/06/2018 21:36:25

Diagnostics Check	OK	21:30
Air Blank	0.000	21:30
Control Test	0.081	21:30
Air Blank	0.000	21:31
Subject Sample # 1	0.069	21:31
Breath Volume	1.171	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.067	21:34
Breath Volume	1.128	21:34
Air Blank	0.000	21:35
Control Test	0.081	21:35
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 07/12/2018 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ROLMAN

DATE/TIME OF TEST: 07/07/2018 00:42:09

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.111	00:37
Breath Volume	3.167	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.115	00:40
Breath Volume	3.332	00:40
Air Blank	0.000	00:41
Control Test	0.080	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 07/12/2018 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 07/08/2018 00:39:44

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.081	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.095	00:34
Breath Volume	1.183	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:36
Subject Sample # 2	0.094	00:38
Breath Volume	1.359	00:38
Air Blank	0.000	00:38
Control Test	0.079	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 07/12/2018 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS, RONALD, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI, J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 07/08/2018 04:13:39

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:02	
Control Test	0.081	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 1 0.184	04:06	Volume Not Met
Breath Volume	0.039	04:06	
Air Blank	0.000	04:06	
Air Blank	0.000	04:08	
Subject Sample	# 2 0.182	04:11	Volume Not Met
Breath Volume	0.207	04:11	
Air Blank	0.000	04:12	
Control Test	0.078	04:12	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 07/12/2018 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: FARRIS

DATE/TIME OF TEST: 07/09/2018 02:14:10

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.081	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.141	02:09
Breath Volume	3.566	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:11
Subject Sample # 2	0.139	02:12
Breath Volume	4.058	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 07/12/2018 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ESCOBAR,G
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 07/19/2018 02:28:27

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.189	02:23
Breath Volume	2.058	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.205	02:26
Breath Volume	2.324	02:26
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 08/07/2018 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: WISON

DATE/TIME OF TEST: 07/20/2018 01:38:28

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.145	01:33
Breath Volume	1.601	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.134	01:36
Breath Volume	1.113	01:36
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 08/07/2018 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: JAUREGUI

DATE/TIME OF TEST: 07/20/2018 03:51:11

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.142	03:46
Breath Volume	1.332	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.147	03:49
Breath Volume	1.570	03:49
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 08/07/2018 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIERRA,P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: DIAZ GONZALEZ

DATE/TIME OF TEST: 07/20/2018 04:52:20

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.158	04:47
Breath Volume	1.429	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.160	04:50
Breath Volume	2.207	04:50
Air Blank	0.000	04:51
Control Test	0.079	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 08/07/2018 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 07/21/2018 02:20:31

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.168	02:16
Breath Volume	3.343	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.164	02:18
Breath Volume	3.328	02:18
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 08/07/2018 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: GARCIA VANBUREN

DATE/TIME OF TEST: 07/21/2018 03:46:40

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.094	03:42
Breath Volume	1.156	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.087	03:45
Breath Volume	1.296	03:45
Air Blank	0.000	03:45
Control Test	0.079	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 08/07/2018 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: DE LOS SANTOS

DATE/TIME OF TEST: 07/21/2018 05:25:29

Diagnostics Check	OK	05:18
Air Blank	0.000	05:19
Control Test	0.080	05:19
Air Blank	0.000	05:20
Subject Sample # 1	0.162	05:20
Breath Volume	3.828	05:20
Air Blank	0.000	05:21
Air Blank	0.000	05:23
Subject Sample # 2	0.165	05:23
Breath Volume	3.843	05:23
Air Blank	0.000	05:24
Control Test	0.080	05:24
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

Data Download Date/Time: 08/07/2018 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: LOREDO-HERNANDEZ

DATE/TIME OF TEST: 07/22/2018 02:34:12

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.176	02:29
Breath Volume	2.140	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.163	02:32
Breath Volume	1.386	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 08/07/2018 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: PALM

DATE/TIME OF TEST: 07/25/2018 03:18:52

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.149	03:14
Breath Volume	2.359	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:16
Subject Sample # 2	0.150	03:17
Breath Volume	2.781	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 08/07/2018 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 07/26/2018 04:31:05

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.173	04:26
Breath Volume	2.082	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:28
Subject Sample # 2	0.166	04:29
Breath Volume	1.906	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 08/07/2018 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHILLER, E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: RAMCHARAN

DATE/TIME OF TEST: 07/26/2018 21:04:05

Diagnostics Check	OK	20:57
Air Blank	0.000	20:58
Control Test	0.080	20:58
Air Blank	0.000	20:59
Subject Sample # 1	0.000	20:59
Breath Volume	2.207	20:59
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.000	21:02
Breath Volume	1.492	21:02
Air Blank	0.000	21:03
Control Test	0.080	21:03
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

Data Download Date/Time: 08/07/2018 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KAMIKA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAMPBELL,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: BABY

DATE/TIME OF TEST: 07/28/2018 01:59:25

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.080	01:52
Air Blank	0.000	01:53
Subject Sample # 1	0.100	01:54
Breath Volume	1.734	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.101	01:57
Breath Volume	2.195	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 08/07/2018 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KLOCKSIEBEN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: LEE I

DATE/TIME OF TEST: 07/28/2018 04:47:17

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Subject Sample # 1	0.007	04:42
Breath Volume	2.351	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.000	04:45
Breath Volume	2.339	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 08/07/2018 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KAMIKA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHANNON,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: ZAYAS

DATE/TIME OF TEST: 07/28/2018 05:02:41

Diagnostics Check	OK	04:59	
Air Blank	0.000	04:59	
Control Test	0.080	04:59	
Air Blank	0.000	05:00	
Subject Sample	# 1	05:01	Subject Test Refused
Breath Volume	0.000	05:01	
Air Blank	0.000	05:01	
Control Test	0.080	05:02	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

Data Download Date/Time: 08/07/2018 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KAMIKA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHANNON,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: ZAYAS

DATE/TIME OF TEST: 07/28/2018 05:22:02

Diagnostics Check	OK	05:08	
Air Blank	0.000	05:08	
Control Test	0.080	05:08	
Air Blank	0.000	05:09	
Subject Sample	# 1 0.183	05:10	No .020 agreement
Breath Volume	2.058	05:10	
Air Blank	0.000	05:10	
Air Blank	0.000	05:12	
Subject Sample	# 2 0.161	05:14	No .020 agreement
Breath Volume	1.375	05:14	
Air Blank	0.000	05:15	
Air Blank	0.000	05:17	
Subject Sample	# 3 0.167	05:20	Volume Not Met
Breath Volume	1.148	05:20	
Air Blank	0.000	05:21	
Control Test	0.079	05:21	
Air Blank	0.000	05:21	
Diagnostics Check	OK	05:22	

Data Download Date/Time: 08/07/2018 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KAMIKA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHANNON,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: ZAYAS

DATE/TIME OF TEST: 07/28/2018 05:32:50

Diagnostics Check	OK	05:24
Air Blank	0.000	05:24
Control Test	0.078	05:25
Air Blank	0.000	05:25
Subject Sample # 1	0.170	05:28
Breath Volume	1.468	05:28
Air Blank	0.000	05:29
Air Blank	0.000	05:30
Subject Sample # 2	0.170	05:31
Breath Volume	1.976	05:31
Air Blank	0.000	05:31
Control Test	0.078	05:32
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 08/07/2018 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: MASSON

DATE/TIME OF TEST: 07/28/2018 23:42:39

Diagnostics Check	OK	23:34
Air Blank	0.000	23:35
Control Test	0.080	23:35
Air Blank	0.000	23:36
Subject Sample # 1	0.108	23:38
Breath Volume	1.515	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.099	23:41
Breath Volume	1.222	23:41
Air Blank	0.000	23:41
Control Test	0.079	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 08/07/2018 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 07/29/2018 03:16:29

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.079	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.116	03:11
Breath Volume	1.144	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.122	03:14
Breath Volume	1.449	03:14
Air Blank	0.000	03:15
Control Test	0.078	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 08/07/2018 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: ARLINGTON

DATE/TIME OF TEST: 07/30/2018 21:23:19

Diagnostics Check	OK	21:16
Air Blank	0.000	21:16
Control Test	0.080	21:17
Air Blank	0.000	21:17
Subject Sample # 1	0.087	21:18
Breath Volume	2.191	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:20
Subject Sample # 2	0.088	21:21
Breath Volume	1.488	21:21
Air Blank	0.000	21:22
Control Test	0.080	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

Data Download Date/Time: 08/07/2018 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOGDON,BRAD,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MALHEIROS,NANCY
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 07/04/2018 16:45:18

Diagnostics Check	OK	16:38
Air Blank	0.000	16:39
Control Test	0.080	16:39
Air Blank	0.000	16:40
Subject Sample # 1	0.230	16:40
Breath Volume	1.953	16:40
Air Blank	0.000	16:41
Air Blank	0.000	16:43
Subject Sample # 2	0.236	16:43
Breath Volume	2.027	16:43
Air Blank	0.000	16:44
Control Test	0.079	16:44
Air Blank	0.000	16:45
Diagnostics Check	OK	16:45

Data Download Date/Time: 07/30/2018 19:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESCARTES,PETER,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CAPELA,JAYED
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: DEY

DATE/TIME OF TEST: 07/04/2018 21:43:06

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:40	
Control Test	0.080	21:41	
Air Blank	0.000	21:41	
Subject Sample	# 1	21:41	Subject Test Refused
Breath Volume	0.000	21:41	
Air Blank	0.000	21:42	
Control Test	0.079	21:42	
Air Blank	0.000	21:43	
Diagnostics Check	OK	21:43	

Data Download Date/Time: 07/30/2018 19:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESCARTES,PETER,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DAWSON,MARCUS
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: WEATHERFORD

DATE/TIME OF TEST: 07/05/2018 02:30:02

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.116	02:23
Breath Volume	2.542	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.114	02:28
Breath Volume	1.460	02:28
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:30

Data Download Date/Time: 07/30/2018 19:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESCARTES,PETER,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GADDIE,KYLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: MENSCHIKOV

DATE/TIME OF TEST: 07/05/2018 23:52:24

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:38	
Control Test	0.080	23:38	
Air Blank	0.000	23:38	
Subject Sample	# 1 0.171	23:42	Volume Not Met
Breath Volume	0.093	23:42	
Air Blank	0.000	23:42	
Air Blank	0.000	23:44	
Subject Sample	# 2 0.139	23:47	
Breath Volume	1.359	23:47	
Air Blank	0.000	23:48	
Air Blank	0.000	23:50	
Subject Sample	# 3 0.157	23:50	
Breath Volume	1.570	23:50	
Air Blank	0.000	23:51	
Control Test	0.074	23:51	Control Outside Tolerance
Air Blank	0.000	23:52	

Data Download Date/Time: 07/30/2018 19:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESCARTES,PETER,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: ST JOHN,JORDAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: APSEY

DATE/TIME OF TEST: 07/06/2018 03:22:38

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.080	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.000	03:18
Breath Volume	3.917	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.000	03:21
Breath Volume	3.003	03:21
Air Blank	0.000	03:21
Control Test	0.078	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 07/30/2018 19:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRAY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WRAY,CHISTOPHER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: GREENHALGH

DATE/TIME OF TEST: 07/07/2018 00:20:43

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.080	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.169	00:15
Breath Volume	1.859	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.168	00:19
Breath Volume	1.851	00:19
Air Blank	0.000	00:19
Control Test	0.080	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 07/30/2018 19:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEMP,F,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GADDIE,K
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: LEONETTI

DATE/TIME OF TEST: 07/11/2018 01:50:05

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:47	
Control Test	0.079	01:47	
Air Blank	0.000	01:48	
Subject Sample	# 1	01:48	Subject Test Refused
Breath Volume	0.000	01:48	
Air Blank	0.000	01:49	
Control Test	0.080	01:49	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 07/30/2018 19:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS, RONALD, D
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LEWIS, RONALD
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: DEMEDEIROS

DATE/TIME OF TEST: 07/11/2018 23:21:05

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.080	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.190	23:16
Breath Volume	2.285	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:18
Subject Sample # 2	0.194	23:19
Breath Volume	2.871	23:19
Air Blank	0.000	23:20
Control Test	0.078	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 07/30/2018 19:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRAY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: JACKSON,BRYAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:03

SUBJECT'S LAST NAME: SERRA

DATE/TIME OF TEST: 07/15/2018 14:33:51

Diagnostics Check	OK	14:24
Air Blank	0.000	14:24
Control Test	0.080	14:24
Air Blank	0.000	14:25
Subject Sample # 1	0.197	14:26
Breath Volume	1.609	14:26
Air Blank	0.000	14:26
Air Blank	0.000	14:28
Subject Sample # 2	0.174	14:29
Breath Volume	1.238	14:29
Air Blank	0.000	14:29
Air Blank	0.000	14:31
Subject Sample # 3	0.197	14:32
Breath Volume	2.113	14:32
Air Blank	0.000	14:32
Control Test	0.078	14:33
Air Blank	0.000	14:33
Diagnostics Check	OK	14:33

Data Download Date/Time: 07/30/2018 19:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESCARTES,PETER,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GADDIE,KYLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: GORT

DATE/TIME OF TEST: 07/15/2018 22:03:45

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.079	21:58
Air Blank	0.000	21:58
Subject Sample # 1	0.197	21:59
Breath Volume	2.050	21:59
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.197	22:02
Breath Volume	2.136	22:02
Air Blank	0.000	22:02
Control Test	0.078	22:03
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 07/30/2018 19:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: MILLARD

DATE/TIME OF TEST: 07/15/2018 22:52:21

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.117	22:47
Breath Volume	2.437	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.116	22:50
Breath Volume	2.343	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 07/30/2018 19:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRAY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WRAY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 07/20/2018 21:54:07

Diagnostics Check	OK	21:42	
Air Blank	0.000	21:43	
Control Test	0.079	21:43	
Air Blank	0.000	21:44	
Subject Sample	# 1	21:47	No Sample Provided
Breath Volume	0.000	21:47	
Air Blank	0.000	21:47	
Air Blank	0.000	21:49	
Subject Sample	# 2	21:52	No Sample Provided
Breath Volume	0.000	21:52	
Air Blank	0.000	21:53	
Control Test	0.080	21:53	
Air Blank	0.000	21:54	
Diagnostics Check	OK	21:54	

Data Download Date/Time: 07/30/2018 19:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEMP,F,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: EMRICH,I
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 07/28/2018 00:22:05

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.081	00:16
Air Blank	0.000	00:16
Subject Sample # 1	0.148	00:17
Breath Volume	2.765	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.143	00:20
Breath Volume	1.796	00:20
Air Blank	0.000	00:21
Control Test	0.078	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 07/30/2018 19:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEMP,F,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: FORD,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 07/28/2018 01:05:41

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.080	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.074	01:01
Breath Volume	1.320	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.068	01:04
Breath Volume	1.156	01:04
Air Blank	0.000	01:04
Control Test	0.078	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 07/30/2018 19:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHEWNING,AUSTIN,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CHEWNING,AUSTIN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: MCFADDIN

DATE/TIME OF TEST: 07/28/2018 21:06:41

Diagnostics Check	OK	21:03	
Air Blank	0.000	21:04	
Control Test	0.080	21:04	
Air Blank	0.000	21:05	
Subject Sample	# 1	21:05	Subject Test Refused
Breath Volume	0.000	21:05	
Air Blank	0.000	21:05	
Control Test	0.080	21:06	
Air Blank	0.000	21:06	
Diagnostics Check	OK	21:06	

Data Download Date/Time: 07/30/2018 19:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISSI,DAN,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HORNE,A
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 07/02/2018 00:24:11

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.079	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.036	00:19
Breath Volume	2.906	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.034	00:22
Breath Volume	3.000	00:22
Air Blank	0.000	00:23
Control Test	0.078	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENTHAL,DAVID,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/02/2018 12:59:42

Diagnostics Check	OK	12:55	
Air Blank	0.000	12:55	
Control Test	0.079	12:56	
Air Blank	0.000	12:56	
Subject Sample	# 1 0.296	12:58	
Breath Volume	1.574	12:58	
Air Blank		12:59	Purge Fail
Air Blank	0.000	12:59	

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENTHAL,DAVID,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/02/2018 13:03:21

Diagnostics Check	OK	13:01	
Air Blank	0.000	13:02	
Control Test	0.074	13:02	Control Outside Tolerance
Air Blank	0.000	13:03	

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENTHAL,DAVID,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/02/2018 13:18:33

Diagnostics Check	OK	13:05	
Air Blank	0.000	13:06	
Control Test	0.077	13:06	
Air Blank	0.000	13:07	
Subject Sample	# 1 0.279	13:08	
Breath Volume	1.882	13:08	
Air Blank	0.000	13:09	
Air Blank	0.000	13:10	
Subject Sample	# 2 0.265	13:13	Volume Not Met
Breath Volume	0.085	13:13	
Air Blank	0.000	13:14	
Air Blank	0.000	13:16	
Subject Sample	# 3	13:17	Subject Test Refused
Breath Volume	0.000	13:17	
Air Blank	0.000	13:17	
Control Test	0.074	13:17	Control Outside Tolerance
Air Blank	0.000	13:18	

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARE,BRIAN,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: PARE,BRIAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 08:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: BRAY

DATE/TIME OF TEST: 07/03/2018 08:38:50

Diagnostics Check	OK	08:32	
Air Blank	0.000	08:33	
Control Test	0.079	08:33	
Air Blank	0.000	08:34	
Subject Sample # 1	0.000	08:34	
Breath Volume	1.257	08:34	
Air Blank	0.000	08:35	
Air Blank	0.000	08:37	
Subject Sample # 2		08:37	Subject Test Refused
Breath Volume	0.000	08:37	
Air Blank	0.000	08:37	
Control Test	0.079	08:38	
Air Blank	0.000	08:38	
Diagnostics Check	OK	08:38	

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISSI,DANIEL,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HORNE,A
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: ALMOGUER

DATE/TIME OF TEST: 07/04/2018 04:50:02

Diagnostics Check	OK	04:47	
Air Blank	0.000	04:47	
Control Test	0.078	04:47	
Air Blank	0.000	04:48	
Subject Sample	# 1	04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:49	
Control Test	0.078	04:49	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:50	

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,BRITTANIE,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SAAVEDRA,ALONSO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: MOLER

DATE/TIME OF TEST: 07/05/2018 02:53:09

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.079	02:51	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:52	
Control Test	0.079	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN CAMP,JEFFERY,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: PAYNE,JONATHAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: RUBNER

DATE/TIME OF TEST: 07/05/2018 21:21:44

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:19	
Control Test	0.079	21:19	
Air Blank	0.000	21:20	
Subject Sample	# 1	21:20	Subject Test Refused
Breath Volume	0.000	21:20	
Air Blank	0.000	21:20	
Control Test	0.079	21:21	
Air Blank	0.000	21:21	
Diagnostics Check	OK	21:21	

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEEN,ALYIA,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,STEPHEN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: DAVID

DATE/TIME OF TEST: 07/06/2018 02:34:59

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.079	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.069	02:30
Breath Volume	2.773	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:32
Subject Sample # 2	0.062	02:33
Breath Volume	2.308	02:33
Air Blank	0.000	02:34
Control Test	0.075	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN CAMP,JEFFERY,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: PAYNE,JONATHAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: WILSEY

DATE/TIME OF TEST: 07/06/2018 19:38:21

Diagnostics Check	OK	19:35	
Air Blank	0.000	19:35	
Control Test	0.079	19:36	
Air Blank	0.000	19:36	
Subject Sample	# 1	19:36	Subject Test Refused
Breath Volume	0.000	19:36	
Air Blank	0.000	19:37	
Control Test	0.079	19:37	
Air Blank	0.000	19:38	
Diagnostics Check	OK	19:38	

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFALKO,TRAVIS,K
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SAVADERA,ALONSO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BLANCHARD

DATE/TIME OF TEST: 07/06/2018 23:09:47

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.079	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.000	23:05
Breath Volume	3.253	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.000	23:08
Breath Volume	2.683	23:08
Air Blank	0.000	23:08
Control Test	0.080	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFALKO,TRAVIS,K
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,STEPHEN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 07/13/2018 02:14:12

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:10	
Control Test	0.079	02:10	
Air Blank	0.000	02:11	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:13	
Control Test	0.080	02:13	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:14	

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFALK, TRAVIS, K
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON, STEPHEN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: MOISE

DATE/TIME OF TEST: 07/13/2018 20:57:03

Diagnostics Check	OK	20:49
Air Blank	0.000	20:50
Control Test	0.079	20:50
Air Blank	0.000	20:51
Subject Sample # 1	0.236	20:51
Breath Volume	2.097	20:51
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.233	20:55
Breath Volume	1.988	20:55
Air Blank	0.000	20:56
Control Test	0.078	20:56
Air Blank	0.000	20:56
Diagnostics Check	OK	20:57

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFALKO,TRAVIS,K
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,STEPHEN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: PERRI

DATE/TIME OF TEST: 07/14/2018 02:29:07

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:25	
Control Test	0.078	02:25	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:28	
Control Test	0.079	02:28	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CALHOUN,KEVIN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:18

SUBJECT'S LAST NAME: CODDINGTON

DATE/TIME OF TEST: 07/14/2018 07:05:23

Diagnostics Check	OK	06:59
Air Blank	0.000	06:59
Control Test	0.080	06:59
Air Blank	0.000	07:00
Subject Sample # 1	0.000	07:00
Breath Volume	5.441	07:00
Air Blank	0.000	07:01
Air Blank	0.000	07:03
Subject Sample # 2	0.000	07:04
Breath Volume	5.781	07:04
Air Blank	0.000	07:04
Control Test	0.078	07:04
Air Blank	0.000	07:05
Diagnostics Check	OK	07:05

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISSI,DAN,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,S
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: SIZEMORE

DATE/TIME OF TEST: 07/15/2018 20:16:36

Diagnostics Check	OK	20:09
Air Blank	0.000	20:10
Control Test	0.079	20:10
Air Blank	0.000	20:11
Subject Sample # 1	0.218	20:11
Breath Volume	2.593	20:11
Air Blank	0.000	20:12
Air Blank	0.000	20:14
Subject Sample # 2	0.224	20:14
Breath Volume	2.953	20:14
Air Blank	0.000	20:15
Control Test	0.078	20:15
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONG,KENNETH,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SAAVEDRA,ALONSO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: FILORAMO

DATE/TIME OF TEST: 07/21/2018 00:08:44

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.079	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.000	00:04
Breath Volume	2.265	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.000	00:07
Breath Volume	2.933	00:07
Air Blank	0.000	00:07
Control Test	0.079	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONG,KENNETH,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VAN CAMP,JEFFERY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 07/21/2018 22:46:49

Diagnostics Check	OK	22:43	
Air Blank	0.000	22:44	
Control Test	0.079	22:44	
Air Blank	0.000	22:45	
Subject Sample	# 1	22:45	Subject Test Refused
Breath Volume	0.000	22:45	
Air Blank	0.000	22:45	
Control Test	0.080	22:46	
Air Blank	0.000	22:46	
Diagnostics Check	OK	22:46	

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISSI,DAN,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VANCAMP,J
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MATSON

DATE/TIME OF TEST: 07/23/2018 01:47:18

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:44	
Control Test	0.079	01:45	
Air Blank	0.000	01:45	
Subject Sample	# 1	01:45	Subject Test Refused
Breath Volume	0.000	01:45	
Air Blank	0.000	01:46	
Control Test	0.079	01:46	
Air Blank	0.000	01:47	
Diagnostics Check	OK	01:47	

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISSI,D,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VANCAMP,J
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: GROBLER

DATE/TIME OF TEST: 07/27/2018 01:00:09

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:57	
Control Test	0.078	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1	00:58	Subject Test Refused
Breath Volume	0.000	00:58	
Air Blank	0.000	00:59	
Control Test	0.079	00:59	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISSI,DAN,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,S
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: OKUN

DATE/TIME OF TEST: 07/27/2018 02:05:28

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.073	02:00
Breath Volume	1.812	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.073	02:03
Breath Volume	1.597	02:03
Air Blank	0.000	02:04
Control Test	0.079	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 07/27/2018 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISSOON,STEPHEN,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VANCAMP,JEFFERY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ACEVEDO GUILLEN

DATE/TIME OF TEST: 07/28/2018 00:25:10

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.079	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.212	00:16
Breath Volume	1.925	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:18
Subject Sample # 2	0.190	00:19
Breath Volume	1.375	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 3	0.200	00:23
Breath Volume	1.792	00:23
Air Blank	0.000	00:24
Control Test	0.078	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 08/17/2018 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN CAMP,JEFFERY,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: REISSI,DANIEL
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 07/28/2018 01:33:19

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.078	01:31	
Air Blank	0.000	01:31	
Subject Sample	# 1	01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:32	
Control Test	0.079	01:32	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 08/17/2018 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISSI,DAN,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,S
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: DIBARTOLOMEO

DATE/TIME OF TEST: 07/28/2018 03:04:30

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.144	02:59
Breath Volume	2.640	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.137	03:02
Breath Volume	2.058	03:02
Air Blank	0.000	03:03
Control Test	0.077	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 08/17/2018 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISSI,DAN,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,S
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: TURNER DRINKWATER

DATE/TIME OF TEST: 07/29/2018 00:58:49

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:47	
Control Test	0.079	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:51	No Sample Provided
Breath Volume	0.000	00:51	
Air Blank	0.000	00:51	
Air Blank	0.000	00:53	
Subject Sample	# 2	0.155	00:54
Breath Volume	1.808	00:54	
Air Blank	0.000	00:55	
Air Blank	0.000	00:56	
Subject Sample	# 3	0.163	00:57
Breath Volume	1.472	00:57	
Air Blank	0.000	00:57	
Control Test	0.078	00:58	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

Data Download Date/Time: 08/17/2018 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: FLORES BUENO

DATE/TIME OF TEST: 07/22/2018 03:56:12

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:50	
Control Test	0.077	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1 0.179	03:51	
Breath Volume	3.671	03:51	
Air Blank	0.000	03:52	
Air Blank	0.000	03:54	
Subject Sample	# 2 0.190	03:54	
Breath Volume	4.257	03:54	
Air Blank	0.000	03:55	
Control Test	0.077	03:55	
Air Blank		03:55	RFI Detect
Air Blank	0.000	03:56	

Data Download Date/Time: 07/31/2018 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: FLORES BUENO

DATE/TIME OF TEST: 07/22/2018 04:02:37

Diagnostics Check	OK	03:59	
Air Blank	0.000	03:59	
Control Test	0.077	04:00	
Air Blank	0.000	04:00	
Subject Sample	# 1 0.182	04:00	
Breath Volume	4.160	04:00	
Air Blank		04:01	RFI Detect
Air Blank	0.000	04:01	
Control Test	0.076	04:02	
Air Blank		04:02	RFI Detect
Air Blank	0.000	04:02	

Data Download Date/Time: 07/31/2018 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: GALVES

DATE/TIME OF TEST: 07/03/2018 00:13:49

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.080	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.149	00:09
Breath Volume	1.101	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.152	00:12
Breath Volume	1.519	00:12
Air Blank	0.000	00:12
Control Test	0.079	00:13
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 08/03/2018 06:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: AYUB

DATE/TIME OF TEST: 07/03/2018 02:46:38

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:45	
Control Test	0.080	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 08/03/2018 06:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MENENDEZ

DATE/TIME OF TEST: 07/04/2018 23:17:12

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.081	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.118	23:12
Breath Volume	1.871	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.126	23:15
Breath Volume	2.343	23:15
Air Blank	0.000	23:16
Control Test	0.081	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 08/03/2018 06:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: LASSETER

DATE/TIME OF TEST: 07/07/2018 20:28:30

Diagnostics Check	OK	20:22
Air Blank	0.000	20:22
Control Test	0.077	20:22
Air Blank	0.000	20:23
Subject Sample # 1	0.192	20:23
Breath Volume	1.960	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.202	20:26
Breath Volume	1.480	20:26
Air Blank	0.000	20:27
Control Test	0.078	20:27
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

Data Download Date/Time: 08/01/2018 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PEREZ BARTOLON

DATE/TIME OF TEST: 07/08/2018 03:14:51

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.078	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.207	03:09
Breath Volume	1.457	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:11
Subject Sample # 2	0.195	03:13
Breath Volume	1.132	03:13
Air Blank	0.000	03:13
Control Test	0.077	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 08/01/2018 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: KEPNER

DATE/TIME OF TEST: 07/08/2018 21:33:55

Diagnostics Check	OK	21:28	
Air Blank	0.000	21:29	
Control Test	0.079	21:29	
Air Blank	0.000	21:29	
Subject Sample	# 1 0.166	21:30	
Breath Volume	3.675	21:30	
Air Blank	0.000	21:31	
Air Blank	0.000	21:32	
Subject Sample	# 2 0.170	21:33	
Breath Volume	3.406	21:33	
Air Blank		21:33	RFI Detect
Air Blank		21:33	RFI Detect

Data Download Date/Time: 08/01/2018 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: KEPNER

DATE/TIME OF TEST: 07/08/2018 21:44:02

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.078	21:38
Air Blank	0.000	21:38
Subject Sample # 1	0.166	21:39
Breath Volume	3.175	21:39
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.172	21:42
Breath Volume	3.328	21:42
Air Blank	0.000	21:43
Control Test	0.078	21:43
Air Blank	0.000	21:43
Diagnostics Check	OK	21:44

Data Download Date/Time: 08/01/2018 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: OSSMAN

DATE/TIME OF TEST: 07/04/2018 18:53:51

Diagnostics Check	OK	18:40	
Air Blank	0.000	18:41	
Control Test	0.082	18:41	
Air Blank	0.000	18:41	
Subject Sample	# 1 0.177	18:45	Volume Not Met
Breath Volume	0.000	18:45	
Air Blank	0.000	18:45	
Air Blank	0.000	18:47	
Subject Sample	# 2 0.196	18:49	
Breath Volume	1.894	18:49	
Air Blank	0.000	18:49	
Air Blank	0.000	18:51	
Subject Sample	# 3 0.185	18:52	
Breath Volume	1.480	18:52	
Air Blank	0.000	18:52	
Control Test	0.078	18:53	
Air Blank	0.000	18:53	
Diagnostics Check	OK	18:53	

Data Download Date/Time: 07/23/2018 16:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: DZIB MOO

DATE/TIME OF TEST: 07/14/2018 01:43:33

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.142	01:38
Breath Volume	1.726	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.149	01:41
Breath Volume	2.421	01:41
Air Blank	0.000	01:42
Control Test	0.079	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 07/23/2018 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 07/14/2018 02:59:47

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.080	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.100	02:55
Breath Volume	2.062	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.092	02:58
Breath Volume	1.878	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 07/23/2018 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: RAMIREZ CASTRO

DATE/TIME OF TEST: 07/15/2018 03:36:23

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:28	
Control Test	0.081	03:28	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:32	No Sample Provided
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Air Blank	0.000	03:34	
Subject Sample	# 2	03:34	Subject Test Refused
Breath Volume	0.000	03:34	
Air Blank	0.000	03:35	
Control Test	0.081	03:35	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:36	

Data Download Date/Time: 07/23/2018 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: FARRELL

DATE/TIME OF TEST: 07/16/2018 02:07:50

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.127	02:03
Breath Volume	2.464	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.123	02:06
Breath Volume	1.628	02:06
Air Blank	0.000	02:06
Control Test	0.080	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 07/23/2018 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: FLANDERS

DATE/TIME OF TEST: 07/24/2018 01:40:37

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.102	01:36
Breath Volume	2.304	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.096	01:38
Breath Volume	1.460	01:38
Air Blank	0.000	01:39
Control Test	0.081	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 09/04/2018 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,L,R
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: DIAZ,L
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: CONTRERAS-CEJA

DATE/TIME OF TEST: 07/01/2018 05:05:38

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.080	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.120	05:00
Breath Volume	1.847	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.123	05:03
Breath Volume	2.601	05:03
Air Blank	0.000	05:04
Control Test	0.078	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 07/01/2018 05:14:36

Diagnostics Check	OK	05:11	
Air Blank	0.000	05:12	
Control Test	0.079	05:12	
Air Blank	0.000	05:13	
Subject Sample	# 1	05:13	Subject Test Refused
Breath Volume	0.000	05:13	
Air Blank	0.000	05:13	
Control Test	0.080	05:14	
Air Blank	0.000	05:14	
Diagnostics Check	OK	05:14	

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,AARON,K
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: GONZALES,AARON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: BUI
ARREST TIME: 16:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 07/01/2018 17:57:28

Diagnostics Check	OK	17:50
Air Blank	0.000	17:51
Control Test	0.079	17:51
Air Blank	0.000	17:52
Subject Sample # 1	0.152	17:52
Breath Volume	3.261	17:52
Air Blank	0.000	17:53
Air Blank	0.000	17:55
Subject Sample # 2	0.156	17:55
Breath Volume	3.214	17:55
Air Blank	0.000	17:56
Control Test	0.077	17:56
Air Blank	0.000	17:57
Diagnostics Check	OK	17:57

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON, RONALD, E
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: PETERSON, RONALD
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: MCCARTNEY

DATE/TIME OF TEST: 07/01/2018 19:37:07

Diagnostics Check	OK	19:30
Air Blank	0.000	19:30
Control Test	0.079	19:31
Air Blank	0.000	19:31
Subject Sample # 1	0.256	19:32
Breath Volume	2.726	19:32
Air Blank	0.000	19:33
Air Blank	0.000	19:34
Subject Sample # 2	0.255	19:35
Breath Volume	2.554	19:35
Air Blank	0.000	19:36
Control Test	0.078	19:36
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: FIGUERO BARAHONA

DATE/TIME OF TEST: 07/02/2018 04:37:16

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.079	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.218	04:32
Breath Volume	1.113	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:34
Subject Sample # 2	0.233	04:35
Breath Volume	1.539	04:35
Air Blank	0.000	04:36
Control Test	0.079	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,AARON,K
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: DALTON,JASON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: BUI
ARREST TIME: 16:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/04/2018 18:36:54

Diagnostics Check	OK	18:24
Air Blank	0.000	18:24
Control Test	0.079	18:25
Air Blank	0.000	18:25
Subject Sample # 1	0.178	18:28
Breath Volume	1.695	18:28
Air Blank	0.000	18:29
Air Blank	0.000	18:30
Subject Sample # 2	0.210	18:31
Breath Volume	1.117	18:31
Air Blank	0.000	18:32
Air Blank	0.000	18:34
Subject Sample # 3	0.207	18:35
Breath Volume	1.296	18:35
Air Blank	0.000	18:35
Control Test	0.077	18:36
Air Blank	0.000	18:36
Diagnostics Check	OK	18:36

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 12:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CHIMAL

DATE/TIME OF TEST: 07/05/2018 01:50:26

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:48	
Control Test	0.079	01:48	
Air Blank	0.000	01:48	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:49	
Control Test	0.080	01:49	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: PALUCKI

DATE/TIME OF TEST: 07/05/2018 02:19:11

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:16	
Control Test	0.079	02:17	
Air Blank	0.000	02:17	
Subject Sample	# 1	02:17	Subject Test Refused
Breath Volume	0.000	02:17	
Air Blank	0.000	02:18	
Control Test	0.079	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: SANCHEZ GOMEZ

DATE/TIME OF TEST: 07/06/2018 04:39:36

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.098	04:35
Breath Volume	3.105	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.093	04:37
Breath Volume	2.839	04:37
Air Blank	0.000	04:38
Control Test	0.080	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TAYLOR,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/06/2018 22:12:00

Diagnostics Check	OK	22:00	
Air Blank	0.000	22:00	
Control Test	0.079	22:01	
Air Blank	0.000	22:01	
Subject Sample	# 1 0.081	22:04	Volume Not Met
Breath Volume	0.113	22:04	
Air Blank	0.000	22:05	
Air Blank	0.000	22:07	
Subject Sample	# 2 0.071	22:10	Volume Not Met
Breath Volume	0.054	22:10	
Air Blank	0.000	22:11	
Control Test	0.077	22:11	
Air Blank	0.000	22:11	
Diagnostics Check	OK	22:12	

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,AARON,K
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: GONZALES,AARON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: BUI
ARREST TIME: 17:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: DIETZ

DATE/TIME OF TEST: 07/07/2018 19:15:55

Diagnostics Check	OK	19:08
Air Blank	0.000	19:09
Control Test	0.079	19:09
Air Blank	0.000	19:10
Subject Sample # 1	0.143	19:11
Breath Volume	3.875	19:11
Air Blank	0.000	19:11
Air Blank	0.000	19:13
Subject Sample # 2	0.143	19:14
Breath Volume	3.406	19:14
Air Blank	0.000	19:14
Control Test	0.079	19:15
Air Blank	0.000	19:15
Diagnostics Check	OK	19:15

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIZOT,ALEXANDRA,P
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BIZOT,ALEXANDRA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: MOSHKOLA

DATE/TIME OF TEST: 07/08/2018 06:28:23

Diagnostics Check	OK	06:25	
Air Blank	0.000	06:25	
Control Test	0.080	06:26	
Air Blank	0.000	06:26	
Subject Sample	# 1	06:27	Subject Test Refused
Breath Volume	0.000	06:27	
Air Blank	0.000	06:27	
Control Test	0.080	06:27	
Air Blank	0.000	06:28	
Diagnostics Check	OK	06:28	

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: GOMEZ LOPEZ

DATE/TIME OF TEST: 07/14/2018 00:08:13

Diagnostics Check	OK	23:59	
Air Blank	0.000	23:59	
Control Test	0.080	00:00	
Air Blank	0.000	00:00	
Subject Sample	# 1 0.114	00:03	Volume Not Met
Breath Volume	0.156	00:03	
Air Blank	0.000	00:04	
Air Blank	0.000	00:06	
Subject Sample	# 2	00:06	Subject Test Refused
Breath Volume	0.171	00:06	
Air Blank	0.000	00:07	
Control Test	0.076	00:07	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DAVIDSON,CHRIS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: DINSDALE

DATE/TIME OF TEST: 07/14/2018 00:25:52

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.079	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.091	00:21
Breath Volume	3.156	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:23
Subject Sample # 2	0.090	00:24
Breath Volume	4.320	00:24
Air Blank	0.000	00:24
Control Test	0.078	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: URUCHIMA,ANGEL,R
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: URUCHIMA,ANGEL
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: STACEY

DATE/TIME OF TEST: 07/15/2018 00:26:57

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.080	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.137	00:22
Breath Volume	3.070	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:24
Subject Sample # 2	0.120	00:25
Breath Volume	1.273	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CASAY ALVAREZ

DATE/TIME OF TEST: 07/15/2018 01:48:33

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.165	01:43
Breath Volume	2.585	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.158	01:46
Breath Volume	2.585	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: URUCHIMA,ANGEL,R
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: WILSON,MATTHEW
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: GONZALEZ-SALAZAR

DATE/TIME OF TEST: 07/15/2018 04:56:00

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.080	04:49
Air Blank	0.000	04:49
Subject Sample # 1	0.195	04:50
Breath Volume	1.968	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.187	04:54
Breath Volume	2.082	04:54
Air Blank	0.000	04:55
Control Test	0.078	04:55
Air Blank	0.000	04:55
Diagnostics Check	OK	04:56

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCLJOSE,F
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TUCCLJOSE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: CUXEVA

DATE/TIME OF TEST: 07/15/2018 05:22:13

Diagnostics Check	OK	05:15
Air Blank	0.000	05:16
Control Test	0.078	05:16
Air Blank	0.000	05:17
Subject Sample # 1	0.139	05:17
Breath Volume	2.718	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.141	05:20
Breath Volume	2.675	05:20
Air Blank	0.000	05:21
Control Test	0.079	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 07/16/2018 01:51:37

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:46	
Control Test	0.080	01:47	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.207	01:49	
Air Blank	0.000	01:50	
Control Test	0.081	01:50	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 07/19/2018 00:09:09

Diagnostics Check	OK	00:02
Air Blank	0.000	00:03
Control Test	0.079	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.147	00:04
Breath Volume	1.492	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:06
Subject Sample # 2	0.141	00:07
Breath Volume	1.109	00:07
Air Blank	0.000	00:08
Control Test	0.077	00:08
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 07/24/2018 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,FABIO,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: PISCITELLI

DATE/TIME OF TEST: 07/01/2018 05:07:48

Diagnostics Check	OK	04:56	
Air Blank	0.000	04:56	
Control Test	0.078	04:57	
Air Blank	0.000	04:57	
Subject Sample	# 1	05:00	No Sample Provided
Breath Volume	0.000	05:00	
Air Blank	0.000	05:01	
Air Blank	0.000	05:03	
Subject Sample	# 2	05:06	No Sample Provided
Breath Volume	0.000	05:06	
Air Blank	0.000	05:06	
Control Test	0.078	05:07	
Air Blank	0.000	05:07	
Diagnostics Check	OK	05:07	

Data Download Date/Time: 07/31/2018 17:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,FABIO,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MORNAN

DATE/TIME OF TEST: 07/05/2018 03:37:21

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.078	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.106	03:31
Breath Volume	1.492	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:33
Subject Sample # 2	0.106	03:35
Breath Volume	1.871	03:35
Air Blank	0.000	03:36
Control Test	0.075	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 07/31/2018 17:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,FABIO,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: BRITO DE ALMEIDA

DATE/TIME OF TEST: 07/06/2018 04:19:32

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.225	04:14
Breath Volume	3.566	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.215	04:17
Breath Volume	1.101	04:17
Air Blank	0.000	04:18
Control Test	0.077	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 07/31/2018 17:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,FABIO,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: GONZALEZ ROJAS

DATE/TIME OF TEST: 07/07/2018 02:45:26

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.078	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.076	02:40
Breath Volume	1.863	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.069	02:43
Breath Volume	1.535	02:43
Air Blank	0.000	02:44
Control Test	0.075	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 07/31/2018 17:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,FABIO,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: COLLET

DATE/TIME OF TEST: 07/08/2018 00:38:54

Diagnostics Check	OK	00:31
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Subject Sample # 1	0.116	00:33
Breath Volume	1.500	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.120	00:37
Breath Volume	1.140	00:37
Air Blank	0.000	00:37
Control Test	0.077	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 07/31/2018 17:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,FABIO,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: GRAJALES

DATE/TIME OF TEST: 07/08/2018 04:10:19

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.078	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.148	04:05
Breath Volume	3.539	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.134	04:08
Breath Volume	2.000	04:08
Air Blank	0.000	04:09
Control Test	0.075	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 07/31/2018 17:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,FABIO,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SMITHIE

DATE/TIME OF TEST: 07/13/2018 00:22:50

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.078	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.271	00:18
Breath Volume	2.781	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:20
Subject Sample # 2	0.268	00:21
Breath Volume	2.949	00:21
Air Blank	0.000	00:21
Control Test	0.076	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 07/31/2018 17:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,FABIO,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: TORRESFONSECA

DATE/TIME OF TEST: 07/15/2018 04:03:02

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.077	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.209	03:57
Breath Volume	3.152	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.192	04:01
Breath Volume	2.718	04:01
Air Blank	0.000	04:02
Control Test	0.077	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:03

Data Download Date/Time: 07/31/2018 17:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: POOLE

DATE/TIME OF TEST: 07/08/2018 02:09:45

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.079	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.042	02:05
Breath Volume	1.933	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:07
Subject Sample # 2	0.042	02:08
Breath Volume	2.304	02:08
Air Blank	0.000	02:08
Control Test	0.079	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 07/31/2018 02:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 07/25/2018 03:02:51

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:55	
Control Test	0.082	02:55	
Air Blank	0.000	02:56	
Subject Sample	# 1 0.253	02:57	
Breath Volume	1.199	02:57	
Air Blank	0.000	02:58	
Air Blank	0.000	03:00	
Subject Sample	# 2 0.243	03:01	
Breath Volume	1.480	03:01	
Air Blank		03:02	Ambient Fail
Air Blank		03:02	Purge Fail

Data Download Date/Time: 08/29/2018 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: TORRES-CRUZ

DATE/TIME OF TEST: 07/01/2018 01:25:49

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.079	01:19
Air Blank	0.000	01:20
Subject Sample # 1	0.165	01:21
Breath Volume	2.703	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:23
Subject Sample # 2	0.159	01:24
Breath Volume	1.968	01:24
Air Blank	0.000	01:24
Control Test	0.077	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 07/24/2018 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 07/01/2018 04:27:25

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.079	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.000	04:23
Breath Volume	2.906	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.000	04:26
Breath Volume	3.171	04:26
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 07/24/2018 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 07/01/2018 16:20:48

Diagnostics Check	OK	16:17	
Air Blank	0.000	16:18	
Control Test	0.080	16:18	
Air Blank	0.000	16:19	
Subject Sample	# 1	16:19	Subject Test Refused
Breath Volume	0.000	16:19	
Air Blank	0.000	16:19	
Control Test	0.079	16:20	
Air Blank	0.000	16:20	
Diagnostics Check	OK	16:20	

Data Download Date/Time: 07/24/2018 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 07/02/2018 00:32:30

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:30	
Control Test	0.080	00:30	
Air Blank	0.000	00:30	
Subject Sample	# 1	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:31	
Control Test	0.080	00:31	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 07/24/2018 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEGGETT,JARED,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: LEGGETT,JARED
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: DUBON HERNANDEZ

DATE/TIME OF TEST: 07/02/2018 23:25:23

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.079	23:19
Air Blank	0.000	23:19
Subject Sample # 1	0.164	23:20
Breath Volume	2.816	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.163	23:23
Breath Volume	2.804	23:23
Air Blank	0.000	23:24
Control Test	0.079	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 07/24/2018 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: MUGAVERO

DATE/TIME OF TEST: 07/04/2018 22:39:56

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:37	
Control Test	0.079	22:37	
Air Blank	0.000	22:38	
Subject Sample	# 1	22:38	Subject Test Refused
Breath Volume	0.000	22:38	
Air Blank	0.000	22:39	
Control Test	0.079	22:39	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 07/24/2018 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEGGETT,JARED,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: LEGGETT,JARED
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MATZIR MORO

DATE/TIME OF TEST: 07/07/2018 02:53:27

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.080	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.203	02:49
Breath Volume	3.835	02:49
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.201	02:51
Breath Volume	3.144	02:51
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 07/24/2018 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: WILKINS

DATE/TIME OF TEST: 07/16/2018 02:21:43

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.061	02:17
Breath Volume	3.804	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:19
Subject Sample # 2	0.058	02:20
Breath Volume	3.492	02:20
Air Blank	0.000	02:20
Control Test	0.077	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 07/24/2018 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEGGETT,JARED,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: LEGGETT,JARED
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: AVILA-LOPEZ

DATE/TIME OF TEST: 07/17/2018 23:58:08

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.081	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.066	23:53
Breath Volume	2.867	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.066	23:56
Breath Volume	2.949	23:56
Air Blank	0.000	23:57
Control Test	0.079	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 07/24/2018 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,DANIEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WATSON,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 07/19/2018 05:05:25

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.080	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.123	04:59
Breath Volume	1.613	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.128	05:03
Breath Volume	1.746	05:03
Air Blank	0.000	05:04
Control Test	0.079	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 07/24/2018 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 840875
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R	DRY GAS EXPIRY DATE: 09/19/2020
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: MONTGOMERY,KENNETH	ARREST TIME: 03:13
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Female
LAST AGENCY INSPECTION: 06/07/2018	OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BLACKFORD DATE/TIME OF TEST: 07/05/2018 03:40:23

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.123	03:35
Breath Volume	2.335	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.124	03:38
Breath Volume	2.695	03:38
Air Blank	0.000	03:39
Control Test	0.078	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 07/18/2018 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY, KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GOFF

DATE/TIME OF TEST: 07/06/2018 00:48:22

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:43	
Control Test	0.078	00:43	
Air Blank	0.000	00:44	
Subject Sample # 1	0.239	00:44	
Breath Volume	1.660	00:44	
Air Blank	0.000	00:45	
Air Blank	0.000	00:46	
Subject Sample # 2	0.224	00:47	
Breath Volume	1.968	00:47	
Air Blank		00:47	RFI Detect
Air Blank	0.000	00:48	

Data Download Date/Time: 07/18/2018 18:50

Subject Test Electronic Data

DATE/TIME OF TEST: 07/06/2018 00:58:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: ADGER

DATE/TIME OF TEST: 07/18/2018 21:02:04

Diagnostics Check	OK	20:55
Air Blank	0.000	20:56
Control Test	0.080	20:56
Air Blank	0.000	20:56
Subject Sample # 1	0.154	20:57
Breath Volume	1.347	20:57
Air Blank	0.000	20:58
Air Blank	0.000	20:59
Subject Sample # 2	0.151	21:00
Breath Volume	1.292	21:00
Air Blank	0.000	21:01
Control Test	0.077	21:01
Air Blank	0.000	21:01
Diagnostics Check	OK	21:02

Data Download Date/Time: 07/31/2018 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 840875
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R	DRY GAS EXPIRY DATE: 09/19/2020
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: MONTGOMERY,KENNETH	ARREST TIME: 00:44
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Male
LAST AGENCY INSPECTION: 07/18/2018	OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MELAKIAN DATE/TIME OF TEST: 07/20/2018 01:02:08

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.077	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.143	00:57
Breath Volume	2.269	00:57
Air Blank	0.000	00:58
Air Blank	0.000	00:59
Subject Sample # 2	0.157	01:00
Breath Volume	2.796	01:00
Air Blank	0.000	01:01
Control Test	0.078	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 07/31/2018 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 840875
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R	DRY GAS EXPIRY DATE: 09/19/2020
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: MONTGOMERY,KENNETH	ARREST TIME: 04:06
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Female
LAST AGENCY INSPECTION: 07/18/2018	OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: SOLOMON DATE/TIME OF TEST: 07/20/2018 04:21:41

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.078	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.111	04:17
Breath Volume	2.417	04:17
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.118	04:20
Breath Volume	2.160	04:20
Air Blank	0.000	04:20
Control Test	0.078	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 07/31/2018 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: RUANE

DATE/TIME OF TEST: 07/23/2018 23:35:07

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.078	23:34	
Air Blank		23:34	RFI Detect
Air Blank	0.000	23:35	

Data Download Date/Time: 07/31/2018 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: RUANE

DATE/TIME OF TEST: 07/23/2018 23:38:41

Diagnostics Check	OK	23:38	
Air Blank		23:38	RFI Detect
Air Blank	0.000	23:38	

Data Download Date/Time: 07/31/2018 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY, KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: RUANE

DATE/TIME OF TEST: 07/23/2018 23:41:01

Diagnostics Check	OK	23:40	
Air Blank		23:40	RFI Detect
Air Blank	0.000	23:41	

Data Download Date/Time: 07/31/2018 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 840875
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R	DRY GAS EXPIRY DATE: 09/19/2020
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: MONTGOMERY,KENNETH	ARREST TIME: 23:27
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Male
LAST AGENCY INSPECTION: 07/18/2018	OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: RUANE	DATE/TIME OF TEST: 07/23/2018	23:49:45
----------------------------	-------------------------------	----------

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.215	23:45
Breath Volume	1.464	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.221	23:48
Breath Volume	1.945	23:48
Air Blank	0.000	23:48
Control Test	0.077	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 07/31/2018 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 840875
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R	DRY GAS EXPIRY DATE: 09/19/2020
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: MONTGOMERY,KENNETH	ARREST TIME: 03:09
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Female
LAST AGENCY INSPECTION: 07/18/2018	OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MCGREGOR DATE/TIME OF TEST: 07/25/2018 03:26:59

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.077	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.089	03:22
Breath Volume	1.191	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.094	03:25
Breath Volume	2.839	03:25
Air Blank	0.000	03:26
Control Test	0.077	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 07/31/2018 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HEALY, GREGORY, B

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: HEALY, GREGORY

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 840875

DRY GAS EXPIRY DATE: 09/19/2020

VIOLATION CODE: DUI

ARREST TIME: 21:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: ENGLAND

DATE/TIME OF TEST: 07/20/2018 21:42:31

Diagnostics Check	OK	21:36
Air Blank	0.000	21:36
Control Test	0.081	21:37
Air Blank	0.000	21:37
Subject Sample # 1	0.039	21:37
Breath Volume	3.136	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.039	21:41
Breath Volume	3.085	21:41
Air Blank	0.000	21:41
Control Test	0.079	21:41
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 08/01/2018 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: Yes
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 34416080A2
BREATH TEST OPERATOR: DAVIS,KAYLAN,A	DRY GAS EXPIRY DATE: 02/05/2019
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: DURHAM,JAMES	ARREST TIME: 20:38
ARRESTING OFFICER AGENCY: FHP Troop A	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2018	OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: JOHNSON DATE/TIME OF TEST: 07/04/2018 21:48:55

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:47	
Control Test	0.079	21:47	
Air Blank	0.000	21:48	
Subject Sample	# 1	21:48	RFI Detect
Breath Volume	0.000	21:48	
Air Blank		21:48	RFI Detect
Air Blank	0.000	21:48	

Data Download Date/Time: 07/29/2018 18:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 34416080A2
BREATH TEST OPERATOR: DAVIS,KAYLAN,A	DRY GAS EXPIRY DATE: 02/05/2019
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: DURHAM,JAMES	ARREST TIME: 20:38
ARRESTING OFFICER AGENCY: FHP Troop A	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2018	OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: JOHNSON	DATE/TIME OF TEST: 07/04/2018	21:58:09
------------------------------	-------------------------------	----------

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.079	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.000	21:53
Breath Volume	4.195	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.000	21:56
Breath Volume	3.937	21:56
Air Blank	0.000	21:57
Control Test	0.079	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 07/29/2018 18:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HENNESSY,JONATHAN
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 07/01/2018 04:24:01

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.080	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.196	04:19
Breath Volume	2.031	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.191	04:22
Breath Volume	1.750	04:22
Air Blank	0.000	04:23
Control Test	0.078	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:24

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BUDISH,HARRISON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 06:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:36

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 07/03/2018 07:04:08

Diagnostics Check	OK	06:57
Air Blank	0.000	06:58
Control Test	0.080	06:58
Air Blank	0.000	06:59
Subject Sample # 1	0.063	06:59
Breath Volume	1.500	06:59
Air Blank	0.000	07:00
Air Blank	0.000	07:02
Subject Sample # 2	0.060	07:02
Breath Volume	1.117	07:02
Air Blank	0.000	07:03
Control Test	0.079	07:03
Air Blank	0.000	07:04
Diagnostics Check	OK	07:04

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MCGILLICUTTY,STEVEN
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: FAGAN

DATE/TIME OF TEST: 07/04/2018 01:30:12

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.080	01:19	
Air Blank	0.000	01:19	
Subject Sample	# 1 0.228	01:22	Volume Not Met
Breath Volume	0.117	01:22	
Air Blank	0.000	01:23	
Air Blank	0.000	01:25	
Subject Sample	# 2 0.223	01:28	Volume Not Met
Breath Volume	0.054	01:28	
Air Blank	0.000	01:29	
Control Test	0.079	01:29	
Air Blank	0.000	01:30	
Diagnostics Check	OK	01:30	

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MCGILLICUTTY,STEVEN
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: FAGAN

DATE/TIME OF TEST: 07/04/2018 01:49:31

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:35	
Control Test	0.077	01:35	
Air Blank	0.000	01:36	
Subject Sample	# 1 0.218	01:39	Volume Not Met
Breath Volume	0.117	01:39	
Air Blank	0.000	01:39	
Air Blank	0.000	01:41	
Subject Sample	# 2 0.219	01:43	
Breath Volume	1.273	01:43	
Air Blank	0.000	01:44	
Air Blank	0.000	01:45	
Subject Sample	# 3 0.221	01:47	
Breath Volume	1.414	01:47	
Air Blank	0.000	01:48	
Control Test	0.076	01:48	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STRZALKOWSKI,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: FULTZ

DATE/TIME OF TEST: 07/04/2018 03:50:00

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.078	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.217	03:45
Breath Volume	3.089	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.211	03:48
Breath Volume	1.832	03:48
Air Blank	0.000	03:49
Control Test	0.077	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:50

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LEATHERS,GERALD
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: ARTEAGA

DATE/TIME OF TEST: 07/04/2018 22:46:19

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.079	22:37
Air Blank	0.000	22:37
Subject Sample # 1	0.175	22:38
Breath Volume	2.644	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.153	22:41
Breath Volume	1.417	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 3	0.171	22:44
Breath Volume	2.832	22:44
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: LANDERS

DATE/TIME OF TEST: 07/05/2018 00:58:43

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.078	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.171	00:54
Breath Volume	1.507	00:54
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.167	00:57
Breath Volume	1.375	00:57
Air Blank	0.000	00:57
Control Test	0.078	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,YVETTE,D
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: BEAUBRUN

DATE/TIME OF TEST: 07/05/2018 18:57:16

Diagnostics Check	OK	18:56	
Air Blank	0.000	18:56	
		18:56	RFI Detect
Air Blank	0.000	18:57	

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, YVETTE, D
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP, JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: BEAUBRUN

DATE/TIME OF TEST: 07/05/2018 19:15:07

Diagnostics Check	OK	19:03	
Air Blank	0.000	19:03	
Control Test	0.078	19:04	
Air Blank	0.000	19:04	
Subject Sample	# 1 0.142	19:07	Volume Not Met
Breath Volume	0.082	19:07	
Air Blank	0.000	19:08	
Air Blank	0.000	19:10	
Subject Sample	# 2 0.166	19:13	Volume Not Met
Breath Volume	0.105	19:13	
Air Blank	0.000	19:14	
Control Test	0.077	19:14	
Air Blank	0.000	19:15	
Diagnostics Check	OK	19:15	

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: WERNER

DATE/TIME OF TEST: 07/05/2018 22:09:12

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.080	22:03
Air Blank	0.000	22:03
Subject Sample # 1	0.165	22:04
Breath Volume	1.878	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:06
Subject Sample # 2	0.167	22:07
Breath Volume	3.042	22:07
Air Blank	0.000	22:08
Control Test	0.078	22:08
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: COLUMNA HERNANDEZ

DATE/TIME OF TEST: 07/06/2018 03:16:51

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.080	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.044	03:12
Breath Volume	2.039	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.043	03:15
Breath Volume	2.328	03:15
Air Blank	0.000	03:15
Control Test	0.080	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: SITNICK

DATE/TIME OF TEST: 07/07/2018 00:12:39

Diagnostics Check	OK	23:59	
Air Blank	0.000	00:00	
Control Test	0.081	00:00	
Air Blank	0.000	00:01	
Subject Sample	# 1 0.165	00:04	Volume Not Met
Breath Volume	0.355	00:04	
Air Blank	0.000	00:05	
Air Blank	0.000	00:06	
Subject Sample	# 2 0.155	00:07	
Breath Volume	1.792	00:07	
Air Blank	0.000	00:08	
Air Blank	0.000	00:10	
Subject Sample	# 3 0.166	00:11	
Breath Volume	2.164	00:11	
Air Blank	0.000	00:11	
Control Test	0.079	00:12	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: DEVINE

DATE/TIME OF TEST: 07/07/2018 01:01:31

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.203	00:56
Breath Volume	1.890	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.213	00:59
Breath Volume	2.136	00:59
Air Blank	0.000	01:00
Control Test	0.079	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: O'NEILL

DATE/TIME OF TEST: 07/07/2018 02:25:39

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.080	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.231	02:21
Breath Volume	1.507	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.241	02:24
Breath Volume	2.613	02:24
Air Blank	0.000	02:24
Control Test	0.077	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TRABUE,THOMAS
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 07/07/2018 04:16:10

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.134	04:11
Breath Volume	2.601	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample # 2	0.129	04:14
Breath Volume	1.886	04:14
Air Blank	0.000	04:15
Control Test	0.076	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 07/07/2018 06:13:56

Diagnostics Check	OK	06:07
Air Blank	0.000	06:07
Control Test	0.079	06:08
Air Blank	0.000	06:08
Subject Sample # 1	0.122	06:09
Breath Volume	3.042	06:09
Air Blank	0.000	06:10
Air Blank	0.000	06:11
Subject Sample # 2	0.122	06:12
Breath Volume	3.250	06:12
Air Blank	0.000	06:13
Control Test	0.078	06:13
Air Blank	0.000	06:13
Diagnostics Check	OK	06:13

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 07/08/2018 05:11:45

Diagnostics Check	OK	05:05
Air Blank	0.000	05:06
Control Test	0.080	05:06
Air Blank	0.000	05:06
Subject Sample # 1	0.146	05:07
Breath Volume	2.007	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.145	05:10
Breath Volume	2.101	05:10
Air Blank	0.000	05:10
Control Test	0.079	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 06:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: GALLANT

DATE/TIME OF TEST: 07/08/2018 07:24:52

Diagnostics Check	OK	07:18
Air Blank	0.000	07:19
Control Test	0.080	07:19
Air Blank	0.000	07:19
Subject Sample # 1	0.040	07:20
Breath Volume	1.257	07:20
Air Blank	0.000	07:20
Air Blank	0.000	07:22
Subject Sample # 2	0.043	07:23
Breath Volume	1.695	07:23
Air Blank	0.000	07:23
Control Test	0.079	07:24
Air Blank	0.000	07:24
Diagnostics Check	OK	07:24

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CHILDRESS,JACOB
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 07/10/2018 22:20:08

Diagnostics Check	OK	22:13
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.113	22:15
Breath Volume	1.851	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.114	22:18
Breath Volume	1.500	22:18
Air Blank	0.000	22:19
Control Test	0.078	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: ZELEDON

DATE/TIME OF TEST: 07/12/2018 23:30:16

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.080	23:24
Air Blank	0.000	23:25
Subject Sample # 1	0.151	23:25
Breath Volume	3.125	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.148	23:28
Breath Volume	3.593	23:28
Air Blank	0.000	23:29
Control Test	0.078	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GIL,HANNIE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GIBBONS

DATE/TIME OF TEST: 07/13/2018 02:57:22

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.248	02:52
Breath Volume	1.367	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.238	02:55
Breath Volume	1.738	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 07/13/2018 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BRETON,JESSIKA
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/15/2018 07:16:26

Diagnostics Check	OK	07:09
Air Blank	0.000	07:10
Control Test	0.079	07:10
Air Blank	0.000	07:11
Subject Sample # 1	0.120	07:11
Breath Volume	2.015	07:11
Air Blank	0.000	07:12
Air Blank	0.000	07:14
Subject Sample # 2	0.112	07:14
Breath Volume	1.523	07:14
Air Blank	0.000	07:15
Control Test	0.079	07:15
Air Blank	0.000	07:16
Diagnostics Check	OK	07:16

Data Download Date/Time: 08/29/2018 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOUGHRAN,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: BEEMER

DATE/TIME OF TEST: 07/16/2018 03:28:14

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.128	03:23
Breath Volume	2.480	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:25
Subject Sample # 2	0.122	03:26
Breath Volume	1.535	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 08/29/2018 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STRZALKOWSKI,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 07/17/2018 01:29:02

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.081	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.152	01:24
Breath Volume	2.246	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.142	01:27
Breath Volume	1.378	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:29

Data Download Date/Time: 08/29/2018 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PLANTE,MILTON
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: STRIDIRON JR

DATE/TIME OF TEST: 07/19/2018 04:35:26

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.080	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.186	04:30
Breath Volume	1.289	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.184	04:33
Breath Volume	1.242	04:33
Air Blank	0.000	04:34
Control Test	0.077	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 08/29/2018 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 07/20/2018 02:12:52

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.079	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.198	02:06
Breath Volume	1.585	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.198	02:11
Breath Volume	1.421	02:11
Air Blank	0.000	02:11
Control Test	0.079	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 08/29/2018 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DEQUARTO,NICHOLAS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MENDEZ DIAZ

DATE/TIME OF TEST: 07/21/2018 01:46:00

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.202	01:40
Breath Volume	2.351	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.191	01:44
Breath Volume	1.304	01:44
Air Blank	0.000	01:45
Control Test	0.077	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 08/29/2018 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: COLETTA

DATE/TIME OF TEST: 07/21/2018 06:08:18

Diagnostics Check	OK	05:56	
Air Blank	0.000	05:57	
Control Test	0.079	05:57	
Air Blank	0.000	05:57	
Subject Sample	# 1 0.123	06:01	Volume Not Met
Breath Volume	0.039	06:01	
Air Blank	0.000	06:01	
Air Blank	0.000	06:03	
Subject Sample	# 2 0.116	06:06	Volume Not Met
Breath Volume	0.417	06:06	
Air Blank	0.000	06:07	
Control Test	0.078	06:07	
Air Blank	0.000	06:08	
Diagnostics Check	OK	06:08	

Data Download Date/Time: 08/29/2018 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CLAIRE,CASEY
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 06:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:02

SUBJECT'S LAST NAME: TRUJILLO DABOIN

DATE/TIME OF TEST: 07/22/2018 07:49:08

Diagnostics Check	OK	07:41
Air Blank	0.000	07:42
Control Test	0.080	07:42
Air Blank	0.000	07:43
Subject Sample # 1	0.192	07:44
Breath Volume	1.597	07:44
Air Blank	0.000	07:45
Air Blank	0.000	07:46
Subject Sample # 2	0.191	07:47
Breath Volume	2.253	07:47
Air Blank	0.000	07:48
Control Test	0.080	07:48
Air Blank	0.000	07:49
Diagnostics Check	OK	07:49

Data Download Date/Time: 08/29/2018 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HENNESSY,JONATHAN
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: OLIVO BURGOS

DATE/TIME OF TEST: 07/23/2018 00:27:47

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.078	00:22
Air Blank	0.000	00:22
Subject Sample # 1	0.164	00:23
Breath Volume	3.542	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.168	00:26
Breath Volume	3.992	00:26
Air Blank	0.000	00:26
Control Test	0.078	00:27
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 08/29/2018 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 07/23/2018 02:46:07

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.000	02:41
Breath Volume	1.460	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.000	02:44
Breath Volume	1.335	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 08/29/2018 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: THOMAZ,ANTONIO
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: JEANLOUIS

DATE/TIME OF TEST: 07/23/2018 20:41:32

Diagnostics Check	OK	20:34
Air Blank	0.000	20:35
Control Test	0.081	20:35
Air Blank	0.000	20:35
Subject Sample # 1	0.188	20:37
Breath Volume	1.183	20:37
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.183	20:39
Breath Volume	1.625	20:39
Air Blank	0.000	20:40
Control Test	0.079	20:40
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 08/29/2018 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SPEZIALE,ADAM
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: ARCE ZAPATA

DATE/TIME OF TEST: 07/24/2018 21:56:17

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.079	21:49
Air Blank	0.000	21:50
Subject Sample # 1	0.208	21:51
Breath Volume	1.605	21:51
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.204	21:54
Breath Volume	2.269	21:54
Air Blank	0.000	21:55
Control Test	0.079	21:55
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 08/29/2018 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: COUTAIN

DATE/TIME OF TEST: 07/27/2018 04:30:51

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.164	04:25
Breath Volume	2.218	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.160	04:29
Breath Volume	1.800	04:29
Air Blank	0.000	04:29
Control Test	0.079	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 08/29/2018 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: RAMIREZ-PORTILLO

DATE/TIME OF TEST: 07/27/2018 05:00:27

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:47	
Control Test	0.078	04:47	
Air Blank	0.000	04:48	
Subject Sample	# 1 0.180	04:51	Volume Not Met
Breath Volume	0.742	04:51	
Air Blank	0.000	04:52	
Air Blank	0.000	04:53	
Subject Sample	# 2 0.194	04:54	
Breath Volume	1.125	04:54	
Air Blank	0.000	04:54	
Air Blank	0.000	04:56	
Subject Sample	# 3 0.207	04:58	
Breath Volume	1.179	04:58	
Air Blank	0.000	04:59	
Control Test	0.079	04:59	
Air Blank	0.000	05:00	
Diagnostics Check	OK	05:00	

Data Download Date/Time: 08/29/2018 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: LAURENT

DATE/TIME OF TEST: 07/29/2018 01:16:46

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.216	01:12
Breath Volume	1.269	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.220	01:15
Breath Volume	1.363	01:15
Air Blank	0.000	01:15
Control Test	0.080	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 08/29/2018 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MCNAMARA

DATE/TIME OF TEST: 07/31/2018 01:50:28

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.225	01:44
Breath Volume	2.656	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.224	01:48
Breath Volume	3.156	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 08/29/2018 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 07/31/2018 03:54:47

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:41	
Control Test	0.080	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1 0.146	03:44	
Breath Volume	1.488	03:44	
Air Blank	0.000	03:44	
Air Blank	0.000	03:46	
Subject Sample	# 2 0.145	03:49	Volume Not Met
Breath Volume	0.906	03:49	
Air Blank	0.000	03:50	
Air Blank	0.000	03:52	
Subject Sample	# 3 0.159	03:53	
Breath Volume	2.792	03:53	
Air Blank	0.000	03:53	
Control Test	0.076	03:54	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 08/29/2018 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BERMAN,MARC
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: JOHANSSON

DATE/TIME OF TEST: 07/31/2018 21:31:00

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.080	21:22
Air Blank	0.000	21:22
Subject Sample # 1	0.162	21:23
Breath Volume	1.207	21:23
Air Blank	0.000	21:24
Air Blank	0.000	21:25
Subject Sample # 2	0.183	21:26
Breath Volume	3.191	21:26
Air Blank	0.000	21:27
Air Blank	0.000	21:28
Subject Sample # 3	0.181	21:29
Breath Volume	2.718	21:29
Air Blank	0.000	21:30
Control Test	0.079	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:31

Data Download Date/Time: 08/29/2018 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPOTE,DUVAN,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CAPOTE,DUVAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: FELIPE

DATE/TIME OF TEST: 07/08/2018 22:06:26

Diagnostics Check	OK	22:00
Air Blank	0.000	22:00
Control Test	0.079	22:00
Air Blank	0.000	22:01
Subject Sample # 1	0.210	22:01
Breath Volume	2.515	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.192	22:04
Breath Volume	1.832	22:04
Air Blank	0.000	22:05
Control Test	0.076	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 08/01/2018 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JOSE,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LOPEZ,JOSE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 07/15/2018 03:55:47

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.079	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.100	03:51
Breath Volume	1.417	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.100	03:54
Breath Volume	2.210	03:54
Air Blank	0.000	03:54
Control Test	0.079	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 08/01/2018 15:17