Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,YOSNIEL,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: LORENZO,YOSNIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: BILLIE

DATE/TIME OF TEST: 08/02/2018 22:32:56

Diagnostics Check		OK	22:26
Air Blank		0.000	22:26
Control Test		0.080	22:27
Air Blank		0.000	22:27
Subject Sample	# 1	0.204	22:28
Breath Volume		2.968	22:28
Air Blank		0.000	22:29
Air Blank		0.000	22:30
Subject Sample	# 2	0.212	22:31
Breath Volume		3.937	22:31
Air Blank		0.000	22:31
Control Test		0.079	22:32
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 08/14/2018 10:09

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: ST THOMAS,JACK
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 08/04/2018 03:52:11

Diagnostics Check		OK	03:45
Air Blank		0.000	03:45
Control Test		0.080	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.189	03:47
Breath Volume		1.937	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:49
Subject Sample	# 2	0.184	03:50
Breath Volume		2.054	03:50
Air Blank		0.000	03:51
Control Test		0.077	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 08/14/2018 10:09

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,YOSNIEL,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: LORENZO,YOSNIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: TINOCO

DATE/TIME OF TEST: 08/07/2018 19:25:03

Diagnostics Check		OK	19:16
Air Blank		0.000	19:17
Control Test		0.079	19:17
Air Blank		0.000	19:18
Subject Sample	# 1	0.202	19:20
Breath Volume		2.214	19:20
Air Blank		0.000	19:21
Air Blank		0.000	19:22
Subject Sample	# 2	0.185	19:23
Breath Volume		2.015	19:23
Air Blank		0.000	19:24
Control Test		0.076	19:24
Air Blank		0.000	19:24
Diagnostics Check		OK	19:25

Data Download Date/Time: 08/14/2018 10:09

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MARTINEZ CASANUEVA

DATE/TIME OF TEST: 08/08/2018 04:38:49

Diagnostics Check		OK	04:35	
Air Blank		0.000	04:36	
Control Test		0.080	04:36	
Air Blank		0.000	04:37	
Subject Sample	# 1		04:37	Subject Test Refused
Breath Volume		0.000	04:37	
Air Blank		0.000	04:37	
Control Test		0.080	04:38	
Air Blank		0.000	04:38	
Diagnostics Check		OK	04:38	

Data Download Date/Time: 08/14/2018 10:09

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 08/17/2018 19:57:19

Diagnostics Check		OK	19:54	
Air Blank		0.000	19:54	
Control Test		0.079	19:55	
Air Blank		0.000	19:55	
Subject Sample	# 1		19:55	Subject Test Refused
Breath Volume		0.000	19:55	
Air Blank		0.000	19:56	
Control Test		0.080	19:56	
Air Blank		0.000	19:57	
Diagnostics Check		OK	19:57	

Data Download Date/Time: 09/07/2018 10:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ELDRIDGE 3

DATE/TIME OF TEST: 08/17/2018 22:01:21

Diagnostics Check		OK	21:58	
Air Blank		0.000	21:58	
Control Test		0.081	21:59	
Air Blank		0.000	21:59	
Subject Sample	# 1		21:59	Subject Test Refused
Breath Volume		0.000	21:59	
Air Blank		0.000	22:00	
Control Test		0.081	22:00	
Air Blank		0.000	22:01	
Diagnostics Check		OK	22:01	

Data Download Date/Time: 09/07/2018 10:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:18

SUBJECT'S LAST NAME: BRETTHAUER

DATE/TIME OF TEST: 08/19/2018 07:43:53

Diagnostics Check		OK	07:40	
Air Blank		0.000	07:41	
Control Test		0.080	07:41	
Air Blank		0.000	07:42	
Subject Sample	# 1		07:42	Subject Test Refused
Breath Volume		0.000	07:42	
Air Blank		0.000	07:42	
Control Test		0.080	07:43	
Air Blank		0.000	07:43	
Diagnostics Check		OK	07:43	

Data Download Date/Time: 09/07/2018 10:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORGES,SEAN,P
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BORGES,SEAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 09:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: CUARE GIL

DATE/TIME OF TEST: 08/19/2018 10:33:43

Diagnostics Check		OK	10:30	
Air Blank		0.000	10:31	
Control Test		0.079	10:31	
Air Blank		0.000	10:31	
Subject Sample	# 1		10:32	Subject Test Refused
Breath Volume		0.000	10:32	
Air Blank		0.000	10:32	
Control Test		0.081	10:33	
Air Blank		0.000	10:33	
Diagnostics Check		OK	10:33	

Data Download Date/Time: 09/07/2018 10:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: PRICE,RITA
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BRICENO-URDANETA

DATE/TIME OF TEST: 08/20/2018 03:08:22

Diagnostics Check		OK	03:01
Air Blank		0.000	03:01
Control Test		0.080	03:02
Air Blank		0.000	03:02
Subject Sample	# 1	0.185	03:03
Breath Volume		1.632	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:05
Subject Sample	# 2	0.176	03:06
Breath Volume		1.906	03:06
Air Blank		0.000	03:07
Control Test		0.078	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 09/07/2018 10:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: MORON

DATE/TIME OF TEST: 08/23/2018 00:23:01

Diagnostics Check		OK	00:15
Air Blank		0.000	00:15
Control Test		0.079	00:15
Air Blank		0.000	00:16
Subject Sample	# 1	0.252	00:17
Breath Volume		1.570	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:19
Subject Sample	# 2	0.263	00:21
Breath Volume		1.679	00:21
Air Blank		0.000	00:22
Control Test		0.076	00:22
Air Blank		0.000	00:22
Diagnostics Check		OK	00:23

Data Download Date/Time: 09/07/2018 10:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 08/23/2018 22:41:00

Diagnostics Check		OK	22:34
Air Blank		0.000	22:34
Control Test		0.080	22:34
Air Blank		0.000	22:35
Subject Sample	# 1	0.000	22:36
Breath Volume		1.742	22:36
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.000	22:39
Breath Volume		1.324	22:39
Air Blank		0.000	22:40
Control Test		0.079	22:40
Air Blank		0.000	22:40
Diagnostics Check		OK	22:41

Data Download Date/Time: 09/07/2018 10:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: YAGUARAMAY-AGUIAR

DATE/TIME OF TEST: 08/24/2018 02:24:25

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.185	02:19
Breath Volume		1.406	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.197	02:22
Breath Volume		1.167	02:22
Air Blank		0.000	02:23
Control Test		0.079	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 09/07/2018 10:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENCARNACION,H,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: ENCARNACION,H
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CHICA

DATE/TIME OF TEST: 08/29/2018 02:17:20

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:05	
Control Test		0.080	02:06	
Air Blank		0.000	02:06	
Subject Sample	# 1	0.102	02:09	Volume Not Met
Breath Volume		0.164	02:09	
Air Blank		0.000	02:10	
Air Blank		0.000	02:12	
Subject Sample	# 2	0.099	02:15	Volume Not Met
Breath Volume		0.894	02:15	
Air Blank		0.000	02:16	
Control Test		0.077	02:16	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 09/07/2018 10:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENCARNACION,H,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: ENCARNACION,H
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CHICA

DATE/TIME OF TEST: 08/29/2018 02:34:51

Diagnostics Check		OK	02:19	
Air Blank		0.000	02:19	
Control Test		0.077	02:20	
Air Blank		0.000	02:20	
Subject Sample	# 1	0.097	02:21	
Breath Volume		1.863	02:21	
Air Blank		0.000	02:22	
Air Blank		0.000	02:24	
Subject Sample	# 2	0.095	02:27	Volume Not Met
Breath Volume		0.312	02:27	
Air Blank		0.000	02:28	
Air Blank		0.000	02:29	
Subject Sample	# 3	0.080	02:32	Volume Not Met
Breath Volume		0.480	02:32	
Air Blank		0.000	02:33	
Control Test		0.077	02:34	
Air Blank		0.000	02:34	
Diagnostics Check		OK	02:34	

Data Download Date/Time: 09/07/2018 10:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENCARNACION,H,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: ENCARNACION,H
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CHICA

DATE/TIME OF TEST: 08/29/2018 02:46:06

Diagnostics Check		OK	02:38
Air Blank		0.000	02:39
Control Test		0.076	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.095	02:41
Breath Volume		2.296	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:43
Subject Sample	# 2	0.092	02:44
Breath Volume		3.605	02:44
Air Blank		0.000	02:45
Control Test		0.077	02:45
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 09/07/2018 10:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENCARNACION,H,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: ENCARNACION,H
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: STARRETT

DATE/TIME OF TEST: 08/31/2018 00:21:15

Diagnostics Check		OK	00:14
Air Blank		0.000	00:15
Control Test		0.080	00:15
Air Blank		0.000	00:16
Subject Sample	# 1	0.000	00:16
Breath Volume		1.207	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.000	00:19
Breath Volume		2.027	00:19
Air Blank		0.000	00:20
Control Test		0.079	00:20
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 09/07/2018 10:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 08/01/2018 05:12:33

Diagnostics Check		OK	05:06
Air Blank		0.000	05:06
Control Test		0.080	05:07
Air Blank		0.000	05:07
Subject Sample	# 1	0.055	05:08
Breath Volume		3.699	05:08
Air Blank		0.000	05:08
Air Blank		0.000	05:10
Subject Sample	# 2	0.057	05:11
Breath Volume		3.988	05:11
Air Blank		0.000	05:11
Control Test		0.080	05:11
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: POUTRE

DATE/TIME OF TEST: 08/02/2018 02:37:51

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.081	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.010	02:33
Breath Volume		2.214	02:33
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.010	02:36
Breath Volume		1.730	02:36
Air Blank		0.000	02:36
Control Test		0.080	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 08/04/2018 03:54:44

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.081	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.089	03:50
Breath Volume		1.187	03:50
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.088	03:53
Breath Volume		1.136	03:53
Air Blank		0.000	03:53
Control Test		0.080	03:54
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: ROMEU

DATE/TIME OF TEST: 08/05/2018 01:48:38

Diagnostics Check		OK	01:42
Air Blank		0.000	01:43
Control Test		0.080	01:43
Air Blank		0.000	01:43
Subject Sample	# 1	0.036	01:44
Breath Volume		4.113	01:44
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.036	01:47
Breath Volume		4.601	01:47
Air Blank		0.000	01:47
Control Test		0.081	01:48
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 08/05/2018 02:56:30

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.081	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.164	02:52
Breath Volume		2.000	02:52
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.159	02:55
Breath Volume		1.554	02:55
Air Blank		0.000	02:55
Control Test		0.081	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSEN,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: TERKOWSKI

DATE/TIME OF TEST: 08/06/2018 00:46:18

Diagnostics Check		OK	00:43	
Air Blank		0.000	00:43	
Control Test		0.080	00:44	
Air Blank		0.000	00:44	
Subject Sample	# 1		00:44	Subject Test Refused
Breath Volume		0.000	00:44	
Air Blank		0.000	00:45	
Control Test		0.080	00:45	
Air Blank		0.000	00:46	
Diagnostics Check		OK	00:46	

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 08/06/2018 03:16:49

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:14	
Control Test		0.080	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:15	Subject Test Refused
Breath Volume		0.000	03:15	
Air Blank		0.000	03:15	
Control Test		0.080	03:16	
Air Blank		0.000	03:16	
Diagnostics Check		OK	03:16	

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: FANKHAUSER

DATE/TIME OF TEST: 08/06/2018 21:19:24

Diagnostics Check		OK	21:13
Air Blank		0.000	21:13
Control Test		0.080	21:13
Air Blank		0.000	21:14
Subject Sample	# 1	0.219	21:15
Breath Volume		2.406	21:15
Air Blank		0.000	21:15
Air Blank		0.000	21:17
Subject Sample	# 2	0.227	21:17
Breath Volume		2.175	21:17
Air Blank		0.000	21:18
Control Test		0.080	21:18
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: HEARD

DATE/TIME OF TEST: 08/07/2018 01:55:57

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:53	
Control Test		0.081	01:53	
Air Blank		0.000	01:54	
Subject Sample	# 1		01:54	Subject Test Refused
Breath Volume		0.000	01:54	
Air Blank		0.000	01:55	
Control Test		0.080	01:55	
Air Blank		0.000	01:55	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: GALLIGHER

DATE/TIME OF TEST: 08/09/2018 01:41:31

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.081	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.045	01:36
Breath Volume		2.820	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.042	01:40
Breath Volume		2.363	01:40
Air Blank		0.000	01:40
Control Test		0.080	01:40
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 08/10/2018 02:00:24

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.080	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.104	01:55
Breath Volume		1.554	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.104	01:58
Breath Volume		2.695	01:58
Air Blank		0.000	01:59
Control Test		0.081	01:59
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: OQUINN

DATE/TIME OF TEST: 08/11/2018 01:22:33

Diagnostics Check		OK	01:19	
Air Blank		0.000	01:20	
Control Test		0.080	01:20	
Air Blank		0.000	01:20	
Subject Sample	# 1		01:21	Subject Test Refused
Breath Volume		0.000	01:21	
Air Blank		0.000	01:21	
Control Test		0.079	01:21	
Air Blank		0.000	01:22	
Diagnostics Check		OK	01:22	

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 08/11/2018 03:17:20

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.080	03:12
Air Blank		0.000	03:12
Subject Sample	# 1	0.207	03:13
Breath Volume		2.269	03:13
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.197	03:15
Breath Volume		1.261	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 08/11/2018 22:20:25

Diagnostics Check		OK	22:17	
Air Blank		0.000	22:17	
Control Test		0.080	22:18	
Air Blank		0.000	22:18	
Subject Sample	# 1		22:19	Subject Test Refused
Breath Volume		0.000	22:19	
Air Blank		0.000	22:19	
Control Test		0.079	22:19	
Air Blank		0.000	22:20	
Diagnostics Check		OK	22:20	

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: KARWUR

DATE/TIME OF TEST: 08/12/2018 02:08:02

Diagnostics Check		OK	02:01
Air Blank		0.000	02:02
Control Test		0.080	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.125	02:03
Breath Volume		3.105	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:05
Subject Sample	# 2	0.127	02:06
Breath Volume		1.593	02:06
Air Blank		0.000	02:07
Control Test		0.080	02:07
Air Blank		0.000	02:07
Diagnostics Check		OK	02:08

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 08/12/2018 02:39:03

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:35	
Control Test		0.080	02:36	
Air Blank		0.000	02:36	
Subject Sample	# 1	0.232	02:37	
Breath Volume		1.824	02:37	
Air Blank			02:37	RFI Detect
Air Blank		0.000	02:38	
Control Test		0.080	02:38	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 08/12/2018 02:45:55

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.080	02:40
Air Blank		0.000	02:41
Subject Sample	# 1	0.230	02:41
Breath Volume		1.109	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.232	02:44
Breath Volume		1.175	02:44
Air Blank		0.000	02:45
Control Test		0.080	02:45
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: RUBIANO DE MAZZINI

DATE/TIME OF TEST: 08/12/2018 04:46:56

Diagnostics Check		OK	04:39
Air Blank		0.000	04:39
Control Test		0.080	04:40
Air Blank		0.000	04:40
Subject Sample	# 1	0.181	04:41
Breath Volume		1.652	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.180	04:45
Breath Volume		1.687	04:45
Air Blank		0.000	04:46
Control Test		0.080	04:46
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PATTERSON,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: EVERETT

DATE/TIME OF TEST: 08/12/2018 06:12:20

Diagnostics Check		OK	06:06
Air Blank		0.000	06:06
Control Test		0.080	06:06
Air Blank		0.000	06:07
Subject Sample	# 1	0.156	06:07
Breath Volume		1.375	06:07
Air Blank		0.000	06:08
Air Blank		0.000	06:10
Subject Sample	# 2	0.146	06:10
Breath Volume		1.226	06:10
Air Blank		0.000	06:11
Control Test		0.081	06:11
Air Blank		0.000	06:12
Diagnostics Check		OK	06:12

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: AGUILAR CARRERA

DATE/TIME OF TEST: 08/13/2018 00:51:49

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.080	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.135	00:47
Breath Volume		1.222	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:49
Subject Sample	# 2	0.123	00:50
Breath Volume		1.136	00:50
Air Blank		0.000	00:50
Control Test		0.080	00:51
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: GALLEON

DATE/TIME OF TEST: 08/13/2018 03:51:21

Diagnostics Check		OK	03:45
Air Blank		0.000	03:45
Control Test		0.080	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.216	03:46
Breath Volume		2.347	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.218	03:49
Breath Volume		2.601	03:49
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CRESPO RUIZ

DATE/TIME OF TEST: 08/16/2018 22:40:11

Diagnostics Check		OK	22:31	
Air Blank		0.000	22:32	
Control Test		0.081	22:32	
Air Blank		0.000	22:32	
Subject Sample	# 1		22:36	No Sample Provided
Breath Volume		0.000	22:36	
Air Blank		0.000	22:36	
Air Blank		0.000	22:38	
Subject Sample	# 2		22:38	Subject Test Refused
Breath Volume		0.000	22:38	
Air Blank		0.000	22:39	
Control Test		0.081	22:39	
Air Blank		0.000	22:40	
Diagnostics Check		OK	22:40	

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 08/17/2018 00:22:16

Diagnostics Check		OK	00:12
Air Blank		0.000	00:13
Control Test		0.081	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.263	00:14
Breath Volume		1.800	00:14
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.238	00:17
Breath Volume		1.183	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 3	0.248	00:20
Breath Volume		2.378	00:20
Air Blank		0.000	00:21
Control Test		0.080	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 08/17/2018 11:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: TOUPIN

DATE/TIME OF TEST: 08/17/2018 03:44:27

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:39
Subject Sample	# 1	0.148	03:40
Breath Volume		3.007	03:40
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.155	03:42
Breath Volume		2.964	03:42
Air Blank		0.000	03:43
Control Test		0.080	03:43
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: DILORETO

DATE/TIME OF TEST: 08/17/2018 04:30:24

Diagnostics Check		OK	04:27	
Air Blank		0.000	04:28	
Control Test		0.081	04:28	
Air Blank		0.000	04:28	
Subject Sample	# 1		04:29	Subject Test Refused
Breath Volume		0.000	04:29	
Air Blank		0.000	04:29	
Control Test		0.081	04:29	
Air Blank		0.000	04:30	
Diagnostics Check		OK	04:30	

Data Download Date/Time: 08/17/2018 11:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNER,CHARLES,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOREL-SEPULVEDA,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: TREJO

DATE/TIME OF TEST: 08/17/2018 21:40:51

Diagnostics Check		OK	21:38	
Air Blank		0.000	21:38	
Control Test		0.081	21:38	
Air Blank		0.000	21:39	
Subject Sample	# 1		21:39	Subject Test Refused
Breath Volume		0.000	21:39	
Air Blank		0.000	21:39	
Control Test		0.080	21:40	
Air Blank		0.000	21:40	
Diagnostics Check		OK	21:40	

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: GONZOL

DATE/TIME OF TEST: 08/18/2018 02:31:03

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.081	02:24
Air Blank		0.000	02:25
Subject Sample	# 1	0.115	02:26
Breath Volume		1.101	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.122	02:29
Breath Volume		1.597	02:29
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:31

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARDWELL,M
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: PITTS

DATE/TIME OF TEST: 08/18/2018 03:27:38

Diagnostics Check		OK	03:24	
Air Blank		0.000	03:25	
Control Test		0.080	03:25	
Air Blank		0.000	03:26	
Subject Sample	# 1		03:26	Subject Test Refused
Breath Volume		0.000	03:26	
Air Blank		0.000	03:26	
Control Test		0.080	03:27	
Air Blank		0.000	03:27	
Diagnostics Check		OK	03:27	

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SMITH,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 08/18/2018 04:32:26

Diagnostics Check		OK	04:29	
Air Blank		0.000	04:30	
Control Test		0.081	04:30	
Air Blank		0.000	04:30	
Subject Sample	# 1		04:31	Subject Test Refused
Breath Volume		0.000	04:31	
Air Blank		0.000	04:31	
Control Test		0.081	04:31	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:32	

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 08/19/2018 00:24:24

Diagnostics Check		OK	00:18
Air Blank		0.000	00:18
Control Test		0.080	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.076	00:20
Breath Volume		1.250	00:20
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.071	00:22
Breath Volume		1.265	00:22
Air Blank		0.000	00:23
Control Test		0.080	00:23
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: UY

DATE/TIME OF TEST: 08/19/2018 03:52:11

Diagnostics Check		OK	03:49	
Air Blank		0.000	03:49	
Control Test		0.080	03:50	
Air Blank		0.000	03:50	
Subject Sample	# 1		03:50	Subject Test Refused
Breath Volume		0.000	03:50	
Air Blank		0.000	03:51	
Control Test		0.080	03:51	
Air Blank		0.000	03:52	
Diagnostics Check		OK	03:52	

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: LUNDY

DATE/TIME OF TEST: 08/19/2018 04:37:57

Diagnostics Check		OK	04:31
Air Blank		0.000	04:31
Control Test		0.080	04:32
Air Blank		0.000	04:32
Subject Sample	# 1	0.164	04:33
Breath Volume		3.101	04:33
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.165	04:36
Breath Volume		2.847	04:36
Air Blank		0.000	04:37
Control Test		0.080	04:37
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: GANTLING

DATE/TIME OF TEST: 08/20/2018 05:53:50

Diagnostics Check		OK	05:50	
Air Blank		0.000	05:51	
Control Test		0.080	05:51	
Air Blank		0.000	05:52	
Subject Sample	# 1		05:52	Subject Test Refused
Breath Volume		0.000	05:52	
Air Blank		0.000	05:52	
Control Test		0.080	05:53	
Air Blank		0.000	05:53	
Diagnostics Check		OK	05:53	

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 08/23/2018 02:36:47

Diagnostics Check		OK	02:33	
Air Blank		0.000	02:34	
Control Test		0.081	02:34	
Air Blank		0.000	02:35	
Subject Sample	# 1		02:35	Subject Test Refused
Breath Volume		0.000	02:35	
Air Blank		0.000	02:35	
Control Test		0.081	02:36	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:36	

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 08/25/2018 02:51:31

Diagnostics Check		OK	02:45
Air Blank		0.000	02:45
Control Test		0.080	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.162	02:47
Breath Volume		2.386	02:47
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.161	02:50
Breath Volume		1.710	02:50
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSEN,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: WALLER

DATE/TIME OF TEST: 08/25/2018 04:51:53

Diagnostics Check		OK	04:45
Air Blank		0.000	04:46
Control Test		0.080	04:46
Air Blank		0.000	04:46
Subject Sample	# 1	0.155	04:47
Breath Volume		1.347	04:47
Air Blank		0.000	04:48
Air Blank		0.000	04:49
Subject Sample	# 2	0.149	04:50
Breath Volume		1.921	04:50
Air Blank		0.000	04:50
Control Test		0.080	04:51
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,F,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARDWELL,M
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:57

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 08/25/2018 07:29:25

Diagnostics Check		OK	07:19
Air Blank		0.000	07:19
Control Test		0.081	07:20
Air Blank		0.000	07:20
Subject Sample	# 1	0.243	07:21
Breath Volume		2.308	07:21
Air Blank		0.000	07:22
Air Blank		0.000	07:24
Subject Sample	# 2	0.270	07:24
Breath Volume		2.097	07:24
Air Blank		0.000	07:25
Air Blank		0.000	07:27
Subject Sample	# 3	0.241	07:27
Breath Volume		1.273	07:27
Air Blank		0.000	07:28
Control Test		0.078	07:28
Air Blank		0.000	07:29
Diagnostics Check		OK	07:29

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,R,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: SUASTEGUI

DATE/TIME OF TEST: 08/25/2018 20:32:47

Diagnostics Check		OK	20:29	
Air Blank		0.000	20:30	
Control Test		0.080	20:30	
Air Blank		0.000	20:31	
Subject Sample	# 1		20:31	Subject Test Refused
Breath Volume		0.000	20:31	
Air Blank		0.000	20:31	
Control Test		0.080	20:32	
Air Blank		0.000	20:32	
Diagnostics Check		OK	20:32	

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 08/26/2018 02:25:52

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:23	
Control Test		0.081	02:23	
Air Blank		0.000	02:24	
Subject Sample	# 1		02:24	Subject Test Refused
Breath Volume		0.000	02:24	
Air Blank		0.000	02:24	
Control Test		0.081	02:25	
Air Blank		0.000	02:25	
Diagnostics Check		OK	02:25	

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 08/26/2018 04:37:23

Diagnostics Check		OK	04:31
Air Blank		0.000	04:31
Control Test		0.080	04:31
Air Blank		0.000	04:32
Subject Sample	# 1	0.131	04:32
Breath Volume		2.644	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.141	04:35
Breath Volume		3.699	04:35
Air Blank		0.000	04:36
Control Test		0.080	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,R,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: COLLIER

DATE/TIME OF TEST: 08/26/2018 21:38:13

Diagnostics Check		OK	21:35	
Air Blank		0.000	21:35	
Control Test		0.080	21:36	
Air Blank		0.000	21:36	
Subject Sample	# 1		21:36	Subject Test Refused
Breath Volume		0.000	21:36	
Air Blank		0.000	21:37	
Control Test		0.080	21:37	
Air Blank		0.000	21:38	
Diagnostics Check		OK	21:38	

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: NIGATU

DATE/TIME OF TEST: 08/26/2018 23:56:36

Diagnostics Check		OK	23:47	
Air Blank		0.000	23:47	
Control Test		0.080	23:48	
Air Blank		0.000	23:48	
Subject Sample	# 1	0.151	23:51	Volume Not Met
Breath Volume		0.808	23:51	
Air Blank		0.000	23:52	
Air Blank		0.000	23:54	
Subject Sample	# 2		23:55	Subject Test Refused
Breath Volume		1.285	23:55	
Air Blank		0.000	23:55	
Control Test		0.080	23:56	
Air Blank		0.000	23:56	
Diagnostics Check		OK	23:56	

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ADLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 08/27/2018 01:27:55

Diagnostics Check		OK	01:21
Air Blank		0.000	01:22
Control Test		0.080	01:22
Air Blank		0.000	01:23
Subject Sample	# 1	0.125	01:23
Breath Volume		1.523	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:25
Subject Sample	# 2	0.120	01:26
Breath Volume		2.300	01:26
Air Blank		0.000	01:27
Control Test		0.080	01:27
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 09/06/2018 14:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: SWORD

DATE/TIME OF TEST: 08/27/2018 02:40:06

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:34
Subject Sample	# 1	0.113	02:34
Breath Volume		1.492	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.109	02:38
Breath Volume		1.339	02:38
Air Blank		0.000	02:39
Control Test		0.081	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: SOUTH

DATE/TIME OF TEST: 08/27/2018 03:42:34

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.092	03:37
Breath Volume		1.632	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.089	03:41
Breath Volume		1.132	03:41
Air Blank		0.000	03:41
Control Test		0.080	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: PHAM

DATE/TIME OF TEST: 08/28/2018 01:33:17

Diagnostics Check		OK	01:26
Air Blank		0.000	01:27
Control Test		0.081	01:27
Air Blank		0.000	01:28
Subject Sample	# 1	0.163	01:28
Breath Volume		1.117	01:28
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.155	01:31
Breath Volume		1.605	01:31
Air Blank		0.000	01:32
Control Test		0.080	01:32
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSEN,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 08/28/2018 02:37:11

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.102	02:32
Breath Volume		2.582	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.096	02:35
Breath Volume		1.820	02:35
Air Blank		0.000	02:36
Control Test		0.080	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CAMP

DATE/TIME OF TEST: 08/30/2018 01:48:39

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.081	01:43
Air Blank		0.000	01:43
Subject Sample	# 1	0.087	01:44
Breath Volume		2.460	01:44
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.090	01:47
Breath Volume		2.613	01:47
Air Blank		0.000	01:47
Control Test		0.081	01:48
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PAUL,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: HELLER

DATE/TIME OF TEST: 08/30/2018 05:44:18

Diagnostics Check		OK	05:37
Air Blank		0.000	05:37
Control Test		0.080	05:38
Air Blank		0.000	05:38
Subject Sample	# 1	0.154	05:39
Breath Volume		1.906	05:39
Air Blank		0.000	05:40
Air Blank		0.000	05:41
Subject Sample	# 2	0.145	05:42
Breath Volume		1.792	05:42
Air Blank		0.000	05:43
Control Test		0.081	05:43
Air Blank		0.000	05:44
Diagnostics Check		OK	05:44

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 08/31/2018 00:01:20

Diagnostics Check		OK	23:54
Air Blank		0.000	23:55
Control Test		0.080	23:55
Air Blank		0.000	23:56
Subject Sample	# 1	0.239	23:56
Breath Volume		1.441	23:56
Air Blank		0.000	23:57
Air Blank		0.000	23:58
Subject Sample	# 2	0.228	23:59
Breath Volume		1.476	23:59
Air Blank		0.000	00:00
Control Test		0.079	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 09/06/2018 14:20

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDONA,K,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 08/27/2018 20:17:12

Diagnostics Check		OK	20:14	
Air Blank		0.000	20:14	
Control Test		0.081	20:15	
Air Blank		0.000	20:15	
Subject Sample	# 1		20:15	Subject Test Refused
Breath Volume		0.000	20:15	
Air Blank		0.000	20:16	
Control Test		0.081	20:16	
Air Blank		0.000	20:17	
Diagnostics Check		OK	20:17	

Data Download Date/Time: 09/24/2018 12:10

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: ONEAL

DATE/TIME OF TEST: 08/29/2018 00:16:07

Diagnostics Check		OK	00:09
Air Blank		0.000	00:09
Control Test		0.082	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.074	00:11
Breath Volume		2.390	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.075	00:14
Breath Volume		2.585	00:14
Air Blank		0.000	00:15
Control Test		0.080	00:15
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 09/24/2018 12:10

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDONA,K,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MALBONE

DATE/TIME OF TEST: 08/29/2018 02:18:33

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:15	
Control Test		0.080	02:16	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:17	Subject Test Refused
Breath Volume		0.000	02:17	
Air Blank		0.000	02:17	
Control Test		0.079	02:17	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 09/24/2018 12:10

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: PITTS

DATE/TIME OF TEST: 08/30/2018 20:23:57

Diagnostics Check		OK	20:16	
Air Blank		0.000	20:16	
Control Test		0.081	20:17	
Air Blank		0.000	20:17	
Subject Sample	# 1	0.070	20:19	
Breath Volume		1.441	20:19	
Air Blank		0.000	20:19	
Air Blank		0.000	20:21	
Subject Sample	# 2		20:22	Subject Test Refused
Breath Volume		0.000	20:22	
Air Blank		0.000	20:23	
Control Test		0.080	20:23	
Air Blank		0.000	20:23	
Diagnostics Check		OK	20:23	

Data Download Date/Time: 09/24/2018 12:10

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MOYERS,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: KEEN

DATE/TIME OF TEST: 08/31/2018 03:10:46

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:08	
Control Test		0.081	03:08	
Air Blank		0.000	03:09	
Subject Sample	# 1		03:09	Subject Test Refused
Breath Volume		0.000	03:09	
Air Blank		0.000	03:09	
Control Test		0.081	03:10	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 09/24/2018 12:10

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDONA,K,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: THOMASON,P
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: WINTERS

DATE/TIME OF TEST: 08/31/2018 22:47:18

Diagnostics Check		OK	22:44	
Air Blank		0.000	22:44	
Control Test		0.081	22:45	
Air Blank		0.000	22:45	
Subject Sample	# 1		22:45	Subject Test Refused
Breath Volume		0.000	22:45	
Air Blank		0.000	22:46	
Control Test		0.081	22:46	
Air Blank		0.000	22:47	
Diagnostics Check		OK	22:47	

Data Download Date/Time: 09/24/2018 12:10

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAINEY,JIMMIE,D
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: HEATH,MARC
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 08/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: LILLY

DATE/TIME OF TEST: 08/25/2018 23:32:40

Diagnostics Check		OK	23:25
Air Blank		0.000	23:25
Control Test		0.082	23:26
Air Blank		0.000	23:26
Subject Sample	# 1	0.254	23:27
Breath Volume		3.527	23:27
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.250	23:31
Breath Volume		3.195	23:31
Air Blank		0.000	23:31
Control Test		0.081	23:32
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 09/25/2018 20:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: RAULERSON

DATE/TIME OF TEST: 08/26/2018 00:45:10

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.082	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.000	00:40
Breath Volume		2.386	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.000	00:43
Breath Volume		1.210	00:43
Air Blank		0.000	00:44
Control Test		0.081	00:44
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 09/25/2018 20:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 08/07/2020
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 08/11/2018 22:05:55

Diagnostics Check	OK	22:04	
Air Blank	0.000	22:05	
Control Test	0.069	22:05	Control Outside Tolerance
Air Blank	0.000	22:05	

Data Download Date/Time: 08/31/2018 15:15

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 08/07/2020
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 08/11/2018 22:09:29

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:08	
Control Test	0.070	22:08	Control Outside Tolerance
Air Blank	0.000	22:09	

Data Download Date/Time: 08/31/2018 15:15

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 08/07/2020
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: COLMAN

DATE/TIME OF TEST: 08/11/2018 22:14:34

Diagnostics Check	OK	22:13	
Air Blank	0.000	22:13	
Control Test	0.072	22:13	Control Outside Tolerance
Air Blank	0.000	22:14	

Data Download Date/Time: 08/31/2018 15:15

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,MARCUS,G
BREATH TEST OPERATOR AGENCY: Union County SO
ARRESTING OFFICER: LEE,WILLIE
ARRESTING OFFICER AGENCY: Union County SO
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 08/07/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 14:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 08/14/2018 14:49:38

Diagnostics Check		OK	14:46	
Air Blank		0.000	14:46	
Control Test		0.080	14:47	
Air Blank		0.000	14:47	
Subject Sample	# 1	0.063	14:48	
Breath Volume		2.351	14:48	
Air Blank			14:48	RFI Detect
Air Blank		0.000	14:48	
Control Test		0.080	14:49	
Air Blank		0.000	14:49	
Diagnostics Check		OK	14:49	

Data Download Date/Time: 08/31/2018 15:16

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TERRELL,G
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: HICKMAN,SEAN
ARRESTING OFFICER AGENCY: Lawtey PD
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: BANZIN

DATE/TIME OF TEST: 08/01/2018 00:10:37

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.079	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.140	00:06
Breath Volume		1.375	00:06
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.135	00:09
Breath Volume		1.289	00:09
Air Blank		0.000	00:09
Control Test		0.078	00:10
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 08/29/2018 16:26

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 08/11/2018 23:15:18

Diagnostics Check		OK	23:11	
Air Blank		0.000	23:11	
Control Test		0.079	23:11	
Air Blank		0.000	23:12	
Subject Sample	# 1		23:13	Subject Test Refused
Breath Volume		0.000	23:13	
Air Blank		0.000	23:14	
Control Test		0.074	23:14	Control Outside Tolerance
Air Blank		0.000	23:15	

Data Download Date/Time: 08/29/2018 16:26

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRELL,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: KING,PAUL
ARRESTING OFFICER AGENCY: Starke PD
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GILLIAM

DATE/TIME OF TEST: 08/26/2018 04:01:36

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:56	
Control Test		0.077	03:57	
Air Blank		0.000	03:57	
Subject Sample	# 1		04:00	Subject Test Refused
Breath Volume		0.000	04:00	
Air Blank		0.000	04:00	
Control Test		0.078	04:01	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 08/29/2018 16:26

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAEYAERT,JEFFREY,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: NAEYAERT,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 14:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: CRIBBS

DATE/TIME OF TEST: 08/26/2018 15:03:25

Diagnostics Check		OK	15:00	
Air Blank		0.000	15:01	
Control Test		0.079	15:01	
Air Blank		0.000	15:01	
Subject Sample	# 1		15:02	Sequence Aborted
Breath Volume		0.000	15:02	
Air Blank		0.000	15:03	

Data Download Date/Time: 08/29/2018 16:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAEYAERT,JEFF,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: NAEYAERT,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 14:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: CRIBBS

DATE/TIME OF TEST: 08/26/2018 15:18:43

Diagnostics Check		OK	15:04	
Air Blank		0.000	15:05	
Control Test		0.078	15:05	
Air Blank		0.000	15:06	
Subject Sample	# 1	0.168	15:09	Volume Not Met
Breath Volume		0.398	15:09	
Air Blank		0.000	15:09	
Air Blank		0.000	15:11	
Subject Sample	# 2	0.164	15:14	
Breath Volume		1.335	15:14	
Air Blank		0.000	15:14	
Air Blank		0.000	15:16	
Subject Sample	# 3	0.175	15:17	
Breath Volume		1.824	15:17	
Air Blank		0.000	15:17	
Control Test		0.077	15:18	
Air Blank		0.000	15:18	
Diagnostics Check		OK	15:18	

Data Download Date/Time: 08/29/2018 16:26

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,CHARLES,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: PIERCE,CHARLES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 08/31/2018 20:27:35

Diagnostics Check		OK	20:16	
Air Blank		0.000	20:16	
Control Test		0.077	20:16	
Air Blank		0.000	20:17	
Subject Sample	# 1		20:20	No Sample Provided
Breath Volume		0.000	20:20	
Air Blank		0.000	20:21	
Air Blank		0.000	20:23	
Subject Sample	# 2		20:26	No Sample Provided
Breath Volume		0.000	20:26	
Air Blank		0.000	20:26	
Control Test		0.077	20:27	
Air Blank		0.000	20:27	
Diagnostics Check		OK	20:27	

Data Download Date/Time: 09/21/2018 21:41

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFELT,DYLAN,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: COFFELT,DYLAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: SALZANO

DATE/TIME OF TEST: 08/01/2018 01:34:52

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:32	
Control Test		0.079	01:32	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:33	Subject Test Refused
Breath Volume		0.000	01:33	
Air Blank		0.000	01:33	
Control Test		0.079	01:34	
Air Blank		0.000	01:34	
Diagnostics Check		OK	01:34	

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRIBB,BRIAN,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: NAGY,COLE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: ATCHISON

DATE/TIME OF TEST: 08/03/2018 19:06:44

Diagnostics Check		OK	19:00
Air Blank		0.000	19:01
Control Test		0.078	19:01
Air Blank		0.000	19:01
Subject Sample	# 1	0.225	19:02
Breath Volume		1.847	19:02
Air Blank		0.000	19:03
Air Blank		0.000	19:04
Subject Sample	# 2	0.212	19:05
Breath Volume		1.750	19:05
Air Blank		0.000	19:05
Control Test		0.077	19:06
Air Blank		0.000	19:06
Diagnostics Check		OK	19:06

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMON,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARMON,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: BRANTLEY

DATE/TIME OF TEST: 08/04/2018 20:03:58

Diagnostics Check		OK	20:00	
Air Blank		0.000	20:00	
Control Test		0.079	20:01	
Air Blank		0.000	20:01	
Subject Sample	# 1		20:02	Subject Test Refused
Breath Volume		0.000	20:02	
Air Blank		0.000	20:03	
Control Test		0.080	20:03	
Air Blank		0.000	20:03	
Diagnostics Check		OK	20:03	

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVOTH,DANIEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARVOTH,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: BLOIS

DATE/TIME OF TEST: 08/05/2018 01:58:13

Diagnostics Check		OK	01:51
Air Blank		0.000	01:52
Control Test		0.080	01:52
Air Blank		0.000	01:52
Subject Sample	# 1	0.000	01:53
Breath Volume		1.800	01:53
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.000	01:56
Breath Volume		1.335	01:56
Air Blank		0.000	01:57
Control Test		0.079	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,WILLIAM,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ADAMS,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: LEBLANC

DATE/TIME OF TEST: 08/05/2018 06:37:40

Diagnostics Check		OK	06:33	
Air Blank		0.000	06:34	
Control Test		0.079	06:34	
Air Blank		0.000	06:35	
Subject Sample	# 1		06:36	Subject Test Refused
Breath Volume		0.000	06:36	
Air Blank		0.000	06:36	
Control Test		0.079	06:37	
Air Blank		0.000	06:37	
Diagnostics Check		OK	06:37	

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEINER,ZACHARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: STEINER,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 08/08/2018 19:32:26

Diagnostics Check		OK	19:26
Air Blank		0.000	19:26
Control Test		0.080	19:26
Air Blank		0.000	19:27
Subject Sample	# 1	0.150	19:28
Breath Volume		3.304	19:28
Air Blank		0.000	19:28
Air Blank		0.000	19:30
Subject Sample	# 2	0.141	19:30
Breath Volume		2.265	19:30
Air Blank		0.000	19:31
Control Test		0.079	19:31
Air Blank		0.000	19:32
Diagnostics Check		OK	19:32

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRIBB,BRIAN,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CRIBB,BRIAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 08/09/2018 01:08:28

Diagnostics Check		OK	01:05	
Air Blank		0.000	01:06	
Control Test		0.080	01:06	
Air Blank		0.000	01:06	
Subject Sample	# 1		01:07	Subject Test Refused
Breath Volume		0.000	01:07	
Air Blank		0.000	01:07	
Control Test		0.080	01:07	
Air Blank		0.000	01:08	
Diagnostics Check		OK	01:08	

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,ANNA,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: JONES,ANNA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: CRAINE

DATE/TIME OF TEST: 08/09/2018 22:43:42

Diagnostics Check		OK	22:37
Air Blank		0.000	22:37
Control Test		0.080	22:37
Air Blank		0.000	22:38
Subject Sample	# 1	0.235	22:39
Breath Volume		2.445	22:39
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.237	22:42
Breath Volume		2.742	22:42
Air Blank		0.000	22:42
Control Test		0.079	22:43
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTZOG,ROBERT,
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARTZOG,ROBERT
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: VENERY

DATE/TIME OF TEST: 08/11/2018 01:04:23

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:01	
Control Test		0.079	01:01	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:03	Subject Test Refused
Breath Volume		0.000	01:03	
Air Blank		0.000	01:03	
Control Test		0.079	01:03	
Air Blank		0.000	01:04	
Diagnostics Check		OK	01:04	

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,CHAD,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: FLOYD,CHAD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: SEEVERS

DATE/TIME OF TEST: 08/12/2018 03:50:21

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:46	
Control Test		0.079	03:47	
Air Blank		0.000	03:47	
Subject Sample	# 1		03:48	Subject Test Refused
Breath Volume		0.000	03:48	
Air Blank		0.000	03:49	
Control Test		0.080	03:49	
Air Blank		0.000	03:50	
Diagnostics Check		OK	03:50	

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/12/2018 05:52:38

Diagnostics Check		OK	05:45
Air Blank		0.000	05:46
Control Test		0.080	05:46
Air Blank		0.000	05:46
Subject Sample	# 1	0.075	05:47
Breath Volume		2.195	05:47
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.079	05:51
Breath Volume		2.910	05:51
Air Blank		0.000	05:51
Control Test		0.079	05:52
Air Blank		0.000	05:52
Diagnostics Check		OK	05:52

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVOTH,DANIEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARVOTH,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: MUNGLE

DATE/TIME OF TEST: 08/12/2018 18:11:12

Diagnostics Check		OK	18:04
Air Blank		0.000	18:04
Control Test		0.079	18:05
Air Blank		0.000	18:05
Subject Sample	# 1	0.173	18:06
Breath Volume		1.957	18:06
Air Blank		0.000	18:07
Air Blank		0.000	18:09
Subject Sample	# 2	0.174	18:09
Breath Volume		1.312	18:09
Air Blank		0.000	18:10
Control Test		0.078	18:10
Air Blank		0.000	18:11
Diagnostics Check		OK	18:11

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOZARD,CHRYSTAL,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOZARD,CHRYSTAL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 08/13/2018 04:04:37

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:53	
Control Test		0.079	03:54	
Air Blank		0.000	03:54	
Subject Sample	# 1		03:57	No Sample Provided
Breath Volume		0.000	03:57	
Air Blank		0.000	03:58	
Air Blank		0.000	04:00	
Subject Sample	# 2		04:03	No Sample Provided
Breath Volume		0.000	04:03	
Air Blank		0.000	04:03	
Control Test		0.079	04:04	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPLING,JEFF,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: APPLING,JEFF
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 11:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:28

SUBJECT'S LAST NAME: BARLOW

DATE/TIME OF TEST: 08/13/2018 12:54:58

Diagnostics Check		OK	12:48
Air Blank		0.000	12:49
Control Test		0.079	12:49
Air Blank		0.000	12:50
Subject Sample	# 1	0.285	12:50
Breath Volume		2.582	12:50
Air Blank		0.000	12:51
Air Blank		0.000	12:52
Subject Sample	# 2	0.288	12:53
Breath Volume		2.847	12:53
Air Blank		0.000	12:54
Control Test		0.078	12:54
Air Blank		0.000	12:54
Diagnostics Check		OK	12:54

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEINER,ZACHARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: STEINER,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:32

SUBJECT'S LAST NAME: KIMMONS

DATE/TIME OF TEST: 08/13/2018 16:58:51

Diagnostics Check		OK	16:52
Air Blank		0.000	16:53
Control Test		0.080	16:53
Air Blank		0.000	16:53
Subject Sample	# 1	0.000	16:54
Breath Volume		4.042	16:54
Air Blank		0.000	16:54
Air Blank		0.000	16:56
Subject Sample	# 2	0.000	16:57
Breath Volume		3.992	16:57
Air Blank		0.000	16:57
Control Test		0.080	16:58
Air Blank		0.000	16:58
Diagnostics Check		OK	16:58

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAHALLEY,JAMEY,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: KAHALLEY,JAMEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 08/18/2018 02:26:18

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:22	
Control Test		0.079	02:23	
Air Blank		0.000	02:23	
Subject Sample	# 1		02:24	Subject Test Refused
Breath Volume		0.000	02:24	
Air Blank		0.000	02:25	
Control Test		0.080	02:25	
Air Blank		0.000	02:26	
Diagnostics Check		OK	02:26	

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CLEMENTS,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: DYER

DATE/TIME OF TEST: 08/18/2018 20:54:21

Diagnostics Check		OK	20:47
Air Blank		0.000	20:47
Control Test		0.080	20:47
Air Blank		0.000	20:48
Subject Sample	# 1	0.155	20:49
Breath Volume		1.691	20:49
Air Blank		0.000	20:50
Air Blank		0.000	20:51
Subject Sample	# 2	0.157	20:52
Breath Volume		2.000	20:52
Air Blank		0.000	20:53
Control Test		0.078	20:53
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 08/19/2018 02:28:02

Diagnostics Check		OK	02:24	
Air Blank		0.000	02:25	
Control Test		0.079	02:25	
Air Blank		0.000	02:26	
Subject Sample	# 1		02:26	Subject Test Refused
Breath Volume		0.000	02:26	
Air Blank		0.000	02:27	
Control Test		0.080	02:27	
Air Blank		0.000	02:27	
Diagnostics Check		OK	02:28	

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASCHAL,THOMAS,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PASCHAL,THOMAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 06:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/19/2018 07:46:48

Diagnostics Check		OK	07:40
Air Blank		0.000	07:40
Control Test		0.080	07:41
Air Blank		0.000	07:41
Subject Sample	# 1	0.132	07:42
Breath Volume		3.070	07:42
Air Blank		0.000	07:42
Air Blank		0.000	07:44
Subject Sample	# 2	0.135	07:45
Breath Volume		1.921	07:45
Air Blank		0.000	07:45
Control Test		0.079	07:46
Air Blank		0.000	07:46
Diagnostics Check		OK	07:46

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERHUNE,JEREMY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TERHUNE,JEREMY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: STINSON

DATE/TIME OF TEST: 08/19/2018 22:51:51

Diagnostics Check		OK	22:45
Air Blank		0.000	22:46
Control Test		0.080	22:46
Air Blank		0.000	22:46
Subject Sample	# 1	0.000	22:47
Breath Volume		4.195	22:47
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.000	22:50
Breath Volume		2.628	22:50
Air Blank		0.000	22:50
Control Test		0.078	22:51
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOZARD,CHRYSTAL,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOZARD,CHRYSTAL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: BABB

DATE/TIME OF TEST: 08/20/2018 04:47:13

Diagnostics Check		OK	04:35	
Air Blank		0.000	04:36	
Control Test		0.080	04:36	
Air Blank		0.000	04:37	
Subject Sample	# 1		04:40	No Sample Provided
Breath Volume		0.000	04:40	
Air Blank		0.000	04:40	
Air Blank		0.000	04:42	
Subject Sample	# 2		04:45	No Sample Provided
Breath Volume		0.000	04:45	
Air Blank		0.000	04:46	
Control Test		0.080	04:46	
Air Blank		0.000	04:47	
Diagnostics Check		OK	04:47	

Data Download Date/Time: 08/30/2018 18:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWELL,DAVID,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROWELL,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: PATE

DATE/TIME OF TEST: 08/21/2018 06:20:56

Diagnostics Check		OK	06:15	
Air Blank		0.000	06:15	
Control Test		0.080	06:16	
Air Blank		0.000	06:16	
Subject Sample	# 1		06:19	RFI Detect
Breath Volume		0.000	06:19	
Air Blank		0.000	06:20	
Control Test		0.080	06:20	
Air Blank		0.000	06:20	
Diagnostics Check		OK	06:20	

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWELL,DAVID,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROWELL,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: PATE

DATE/TIME OF TEST: 08/21/2018 06:28:16

Diagnostics Check		OK	06:22
Air Blank		0.000	06:22
Control Test		0.080	06:23
Air Blank		0.000	06:23
Subject Sample	# 1	0.000	06:24
Breath Volume		4.113	06:24
Air Blank		0.000	06:24
Air Blank		0.000	06:26
Subject Sample	# 2	0.000	06:26
Breath Volume		2.976	06:26
Air Blank		0.000	06:27
Control Test		0.080	06:27
Air Blank		0.000	06:28
Diagnostics Check		OK	06:28

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASCHAL,THOMAS,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PASCHAL,THOMAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 06:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:02

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 08/22/2018 07:23:31

Diagnostics Check	OK	07:22	
Air Blank	0.000	07:22	
Control Test	0.000	07:22	RFI Detect
Air Blank	0.000	07:23	

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASCHAL,THOMAS,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PASCHAL,THOMAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 06:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:02

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 08/22/2018 07:31:05

Diagnostics Check		OK	07:25
Air Blank		0.000	07:25
Control Test		0.080	07:25
Air Blank		0.000	07:26
Subject Sample	# 1	0.000	07:26
Breath Volume		2.359	07:26
Air Blank		0.000	07:27
Air Blank		0.000	07:29
Subject Sample	# 2	0.000	07:29
Breath Volume		2.449	07:29
Air Blank		0.000	07:30
Control Test		0.080	07:30
Air Blank		0.000	07:31
Diagnostics Check		OK	07:31

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORENE,AIDAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NORENE,AIDAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: USATEGUI

DATE/TIME OF TEST: 08/23/2018 20:10:30

Diagnostics Check		OK	20:04
Air Blank		0.000	20:04
Control Test		0.080	20:05
Air Blank		0.000	20:05
Subject Sample	# 1	0.165	20:06
Breath Volume		2.320	20:06
Air Blank		0.000	20:06
Air Blank		0.000	20:08
Subject Sample	# 2	0.157	20:09
Breath Volume		1.480	20:09
Air Blank		0.000	20:09
Control Test		0.079	20:09
Air Blank		0.000	20:10
Diagnostics Check		OK	20:10

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,LLOYD,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ALEXANDER,LLOYD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: KLINGER

DATE/TIME OF TEST: 08/27/2018 18:02:21

Diagnostics Check		OK	17:55
Air Blank		0.000	17:56
Control Test		0.080	17:56
Air Blank		0.000	17:57
Subject Sample	# 1	0.000	17:57
Breath Volume		2.488	17:57
Air Blank		0.000	17:58
Air Blank		0.000	18:00
Subject Sample	# 2	0.000	18:00
Breath Volume		2.972	18:00
Air Blank		0.000	18:01
Control Test		0.079	18:01
Air Blank		0.000	18:02
Diagnostics Check		OK	18:02

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DALGLEISH,KEVIN,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: PACE,KAITLYN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 08/28/2018 05:36:16

Diagnostics Check		OK	05:29
Air Blank		0.000	05:29
Control Test		0.079	05:30
Air Blank		0.000	05:30
Subject Sample	# 1	0.124	05:31
Breath Volume		3.226	05:31
Air Blank		0.000	05:32
Air Blank		0.000	05:34
Subject Sample	# 2	0.118	05:34
Breath Volume		2.167	05:34
Air Blank		0.000	05:35
Control Test		0.079	05:35
Air Blank		0.000	05:36
Diagnostics Check		OK	05:36

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIRCH,JESSIE,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 08/29/2018 01:09:10

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:04	
Control Test		0.081	01:05	
Air Blank		0.000	01:05	
Subject Sample	# 1		01:07	Subject Test Refused
Breath Volume		0.000	01:07	
Air Blank		0.000	01:08	
Control Test		0.079	01:08	
Air Blank		0.000	01:09	
Diagnostics Check		OK	01:09	

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOHN,F
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: YOUNG,JOHN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG715202
DRY GAS EXPIRY DATE: 06/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: RANDOLPH

DATE/TIME OF TEST: 08/29/2018 02:42:22

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.078	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.105	02:37
Breath Volume		2.859	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.102	02:40
Breath Volume		2.593	02:40
Air Blank		0.000	02:41
Control Test		0.078	02:41
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 08/30/2018 18:45

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DALLAS,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JONES,DALLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:11

SUBJECT'S LAST NAME: PETRONACI

DATE/TIME OF TEST: 08/06/2018 17:59:41

Diagnostics Check		OK	17:53
Air Blank		0.000	17:54
Control Test		0.082	17:54
Air Blank		0.000	17:54
Subject Sample	# 1	0.130	17:55
Breath Volume		2.539	17:55
Air Blank		0.000	17:55
Air Blank		0.000	17:57
Subject Sample	# 2	0.122	17:58
Breath Volume		2.292	17:58
Air Blank		0.000	17:58
Control Test		0.081	17:59
Air Blank		0.000	17:59
Diagnostics Check		OK	17:59

Data Download Date/Time: 09/18/2018 12:28

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,WILFREDO,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ARROYO,WILFREDO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: CHAC

DATE/TIME OF TEST: 08/11/2018 06:24:47

Diagnostics Check		OK	06:21	
Air Blank		0.000	06:22	
Control Test		0.083	06:22	
Air Blank		0.000	06:23	
Subject Sample	# 1		06:23	RFI Detect
Breath Volume		0.000	06:23	
Air Blank		0.000	06:23	
Control Test		0.082	06:24	
Air Blank		0.000	06:24	
Diagnostics Check		OK	06:24	

Data Download Date/Time: 09/18/2018 12:28

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,WILFREDO,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ARROYO,WILFREDO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: CHAC

DATE/TIME OF TEST: 08/11/2018 06:28:58

Diagnostics Check	OK	06:27	
Air Blank	0.000	06:27	
Control Test	0.082	06:28	
Air Blank		06:28	RFI Detect
Air Blank	0.000	06:28	

Data Download Date/Time: 09/18/2018 12:28

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,WILFREDO,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ARROYO,WILFREDO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: CHAC

DATE/TIME OF TEST: 08/11/2018 06:40:03

Diagnostics Check		OK	06:33
Air Blank		0.000	06:34
Control Test		0.083	06:34
Air Blank		0.000	06:35
Subject Sample	# 1	0.163	06:35
Breath Volume		3.902	06:35
Air Blank		0.000	06:36
Air Blank		0.000	06:38
Subject Sample	# 2	0.166	06:38
Breath Volume		4.148	06:38
Air Blank		0.000	06:39
Control Test		0.081	06:39
Air Blank		0.000	06:39
Diagnostics Check		OK	06:40

Data Download Date/Time: 09/18/2018 12:28

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CHRISTY,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: TODD,CALEB
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 08/29/2018 21:05:50

Diagnostics Check		OK	20:56
Air Blank		0.000	20:57
Control Test		0.080	20:57
Air Blank		0.000	20:58
Subject Sample	# 1	0.207	21:00
Breath Volume		3.230	21:00
Air Blank		0.000	21:01
Air Blank		0.000	21:02
Subject Sample	# 2	0.189	21:04
Breath Volume		2.949	21:04
Air Blank		0.000	21:04
Control Test		0.078	21:05
Air Blank		0.000	21:05
Diagnostics Check		OK	21:05

Data Download Date/Time: 09/28/2018 11:25

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,BRAD,
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PIERCE,BRAD
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: TRAVASOS

DATE/TIME OF TEST: 08/30/2018 23:28:15

Diagnostics Check		OK	23:25	
Air Blank		0.000	23:25	
Control Test		0.081	23:26	
Air Blank		0.000	23:26	
Subject Sample	# 1		23:27	Sequence Aborted
Breath Volume		0.000	23:27	
Air Blank		0.000	23:28	

Data Download Date/Time: 09/28/2018 11:26

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,BRAD,
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PIERCE,BRAD
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 08/31/2018 23:31:39

Diagnostics Check		OK	23:25
Air Blank		0.000	23:25
Control Test		0.080	23:26
Air Blank		0.000	23:26
Subject Sample	# 1	0.246	23:27
Breath Volume		3.089	23:27
Air Blank		0.000	23:27
Air Blank		0.000	23:29
Subject Sample	# 2	0.241	23:30
Breath Volume		2.167	23:30
Air Blank		0.000	23:30
Control Test		0.080	23:31
Air Blank		0.000	23:31
Diagnostics Check		OK	23:31

Data Download Date/Time: 09/28/2018 11:26

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MCGOWIN,JAKE
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 08/02/2018 00:27:08

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:24	
Control Test		0.079	00:25	
Air Blank		0.000	00:25	
Subject Sample	# 1		00:25	Subject Test Refused
Breath Volume		0.000	00:25	
Air Blank		0.000	00:26	
Control Test		0.079	00:26	
Air Blank		0.000	00:27	
Diagnostics Check		OK	00:27	

Data Download Date/Time: 08/29/2018 13:05

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD,MATTHEW,E
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: MAYFIELD,MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: MESSICK

DATE/TIME OF TEST: 08/04/2018 21:44:58

Diagnostics Check		OK	21:38
Air Blank		0.000	21:39
Control Test		0.080	21:39
Air Blank		0.000	21:39
Subject Sample	# 1	0.028	21:40
Breath Volume		3.562	21:40
Air Blank		0.000	21:41
Air Blank		0.000	21:42
Subject Sample	# 2	0.028	21:43
Breath Volume		3.292	21:43
Air Blank		0.000	21:44
Control Test		0.081	21:44
Air Blank		0.000	21:44
Diagnostics Check		OK	21:44

Data Download Date/Time: 08/29/2018 13:05

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CHRISTY,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BOAN,CPL
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: BLAKEY

DATE/TIME OF TEST: 08/07/2018 02:24:54

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.079	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.218	02:20
Breath Volume		2.750	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:22
Subject Sample	# 2	0.207	02:23
Breath Volume		2.625	02:23
Air Blank		0.000	02:24
Control Test		0.078	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 08/29/2018 13:05

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: YOUNG,THOMAS
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: STIEBER

DATE/TIME OF TEST: 08/07/2018 05:08:06

Diagnostics Check		OK	05:05	
Air Blank		0.000	05:05	
Control Test		0.080	05:06	
Air Blank		0.000	05:06	
Subject Sample	# 1		05:06	Subject Test Refused
Breath Volume		0.000	05:06	
Air Blank		0.000	05:07	
Control Test		0.080	05:07	
Air Blank		0.000	05:08	
Diagnostics Check		OK	05:08	

Data Download Date/Time: 08/29/2018 13:05

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: CLACK,JASMIN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 07:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: DILLON

DATE/TIME OF TEST: 08/11/2018 08:36:44

Diagnostics Check		OK	08:30
Air Blank		0.000	08:30
Control Test		0.079	08:31
Air Blank		0.000	08:31
Subject Sample	# 1	0.103	08:32
Breath Volume		3.125	08:32
Air Blank		0.000	08:32
Air Blank		0.000	08:34
Subject Sample	# 2	0.107	08:35
Breath Volume		3.250	08:35
Air Blank		0.000	08:35
Control Test		0.079	08:36
Air Blank		0.000	08:36
Diagnostics Check		OK	08:36

Data Download Date/Time: 08/29/2018 13:05

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JESSE,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PEAVY,DAVID
ARRESTING OFFICER AGENCY: Parker PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: TARBOX

DATE/TIME OF TEST: 08/11/2018 11:24:06

Diagnostics Check		OK	11:21	
Air Blank		0.000	11:21	
Control Test		0.079	11:21	
Air Blank		0.000	11:22	
Subject Sample	# 1		11:22	Subject Test Refused
Breath Volume		0.000	11:22	
Air Blank		0.000	11:23	
Control Test		0.079	11:23	
Air Blank		0.000	11:24	
Diagnostics Check		OK	11:24	

Data Download Date/Time: 08/29/2018 13:05

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: J525175922460

DATE/TIME OF TEST: 08/18/2018 03:02:12

Diagnostics Check		OK	02:54
Air Blank		0.000	02:55
Control Test		0.080	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.197	02:57
Breath Volume		2.300	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.199	03:00
Breath Volume		2.218	03:00
Air Blank		0.000	03:01
Control Test		0.080	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 08/29/2018 13:05

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL,CHAD,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HARRELL,CHAD
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:21

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/18/2018 17:47:40

Diagnostics Check		OK	17:44	
Air Blank		0.000	17:45	
Control Test		0.079	17:45	
Air Blank		0.000	17:46	
Subject Sample	# 1		17:46	Subject Test Refused
Breath Volume		0.000	17:46	
Air Blank		0.000	17:46	
Control Test		0.080	17:47	
Air Blank		0.000	17:47	
Diagnostics Check		OK	17:47	

Data Download Date/Time: 08/29/2018 13:05

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CHRISTY,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: CIESLIK,JESSE
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 08/22/2018 02:29:29

Diagnostics Check		OK	02:26	
Air Blank		0.000	02:27	
Control Test		0.080	02:27	
Air Blank		0.000	02:27	
Subject Sample	# 1		02:28	Subject Test Refused
Breath Volume		0.000	02:28	
Air Blank		0.000	02:28	
Control Test		0.079	02:28	
Air Blank		0.000	02:29	
Diagnostics Check		OK	02:29	

Data Download Date/Time: 08/29/2018 13:05

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CHRISTY,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BOAN,W
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 08/25/2018 18:05:47

Diagnostics Check		OK	18:02	
Air Blank		0.000	18:03	
Control Test		0.080	18:03	
Air Blank		0.000	18:03	
Subject Sample	# 1		18:04	Subject Test Refused
Breath Volume		0.000	18:04	
Air Blank		0.000	18:04	
Control Test		0.080	18:05	
Air Blank		0.000	18:05	
Diagnostics Check		OK	18:05	

Data Download Date/Time: 08/29/2018 13:05

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CHRISTY,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: YOUNG,THOMAS
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 08/26/2018 03:49:08

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.080	03:41	
Air Blank		0.000	03:41	
Subject Sample	# 1		03:45	No Sample Provided
Breath Volume		0.000	03:45	
Air Blank		0.000	03:45	
Air Blank		0.000	03:47	
Subject Sample	# 2		03:47	Subject Test Refused
Breath Volume		0.000	03:47	
Air Blank		0.000	03:48	
Control Test		0.079	03:48	
Air Blank		0.000	03:49	
Diagnostics Check		OK	03:49	

Data Download Date/Time: 08/29/2018 13:05

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CHRISTY,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PARKER,CHRISTY
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: HALE

DATE/TIME OF TEST: 08/26/2018 22:00:38

Diagnostics Check		OK	21:46	
Air Blank		0.000	21:46	
Control Test		0.080	21:46	
Air Blank		0.000	21:47	
Subject Sample	# 1	0.187	21:50	Volume Not Met
Breath Volume		0.187	21:50	
Air Blank		0.000	21:51	
Air Blank		0.000	21:52	
Subject Sample	# 2	0.186	21:55	
Breath Volume		2.214	21:55	
Air Blank		0.000	21:56	
Air Blank		0.000	21:58	
Subject Sample	# 3	0.222	21:58	
Breath Volume		1.195	21:58	
Air Blank		0.000	21:59	
Control Test		0.072	22:00	Control Outside Tolerance
Air Blank		0.000	22:00	

Data Download Date/Time: 08/29/2018 13:05

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DAVID,O
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JONES,CHAISE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: JEROME

DATE/TIME OF TEST: 08/01/2018 22:35:17

Diagnostics Check		OK	22:28
Air Blank		0.000	22:29
Control Test		0.080	22:29
Air Blank		0.000	22:30
Subject Sample	# 1	0.149	22:30
Breath Volume		2.210	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.140	22:33
Breath Volume		2.242	22:33
Air Blank		0.000	22:34
Control Test		0.080	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 08/10/2018 16:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEHEAD,JEFFREY,A
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: KAISER,ANTHONY
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 08/08/2018 22:01:13

Diagnostics Check		OK	21:54
Air Blank		0.000	21:54
Control Test		0.079	21:55
Air Blank		0.000	21:55
Subject Sample	# 1	0.300	21:56
Breath Volume		2.375	21:56
Air Blank		0.000	21:57
Air Blank		0.000	21:59
Subject Sample	# 2	0.287	21:59
Breath Volume		2.371	21:59
Air Blank		0.000	22:00
Control Test		0.078	22:00
Air Blank		0.000	22:01
Diagnostics Check		OK	22:01

Data Download Date/Time: 08/10/2018 16:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGHANSE,KYLE,O
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: YOUNGHANSE,KYLE
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 08/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: RECENDIZ

DATE/TIME OF TEST: 08/12/2018 04:21:50

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.079	04:16
Air Blank		0.000	04:16
Subject Sample	# 1	0.177	04:17
Breath Volume		1.894	04:17
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.176	04:20
Breath Volume		1.488	04:20
Air Blank		0.000	04:20
Control Test		0.081	04:21
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 09/22/2018 23:13

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYDOS,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCCORMICK,BLAKE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 15:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: RENTON

DATE/TIME OF TEST: 08/21/2018 16:45:45

Diagnostics Check		OK	16:37
Air Blank		0.000	16:37
Control Test		0.079	16:38
Air Blank		0.000	16:38
Subject Sample	# 1	0.132	16:40
Breath Volume		1.828	16:40
Air Blank		0.000	16:41
Air Blank		0.000	16:42
Subject Sample	# 2	0.134	16:44
Breath Volume		1.960	16:44
Air Blank		0.000	16:44
Control Test		0.080	16:45
Air Blank		0.000	16:45
Diagnostics Check		OK	16:45

Data Download Date/Time: 09/22/2018 23:13

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DAVID,O
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SANDERS,DAVID
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: BOGDAHN

DATE/TIME OF TEST: 08/23/2018 23:22:26

Diagnostics Check		OK	23:15
Air Blank		0.000	23:16
Control Test		0.079	23:16
Air Blank		0.000	23:17
Subject Sample	# 1	0.183	23:17
Breath Volume		3.066	23:17
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.180	23:20
Breath Volume		2.984	23:20
Air Blank		0.000	23:21
Control Test		0.079	23:21
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 09/22/2018 23:13

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,CONNOR,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GUNN,CONNOR
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: ATKINS

DATE/TIME OF TEST: 08/05/2018 05:41:03

Diagnostics Check		OK	05:34
Air Blank		0.000	05:35
Control Test		0.080	05:35
Air Blank		0.000	05:35
Subject Sample	# 1	0.175	05:36
Breath Volume		2.488	05:36
Air Blank		0.000	05:37
Air Blank		0.000	05:38
Subject Sample	# 2	0.170	05:39
Breath Volume		1.671	05:39
Air Blank		0.000	05:40
Control Test		0.080	05:40
Air Blank		0.000	05:40
Diagnostics Check		OK	05:41

Data Download Date/Time: 08/10/2018 16:30

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCHULTZ,JEREMY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: HOLLON

DATE/TIME OF TEST: 08/10/2018 21:45:00

Diagnostics Check		OK	21:38
Air Blank		0.000	21:38
Control Test		0.079	21:39
Air Blank		0.000	21:39
Subject Sample	# 1	0.192	21:40
Breath Volume		2.964	21:40
Air Blank		0.000	21:41
Air Blank		0.000	21:42
Subject Sample	# 2	0.188	21:43
Breath Volume		2.050	21:43
Air Blank		0.000	21:44
Control Test		0.078	21:44
Air Blank		0.000	21:44
Diagnostics Check		OK	21:45

Data Download Date/Time: 09/22/2018 23:03

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,CONNOR,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PULLINGS,ALLEN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BREAUX

DATE/TIME OF TEST: 08/23/2018 00:53:39

Diagnostics Check		OK	00:50	
Air Blank		0.000	00:51	
Control Test		0.079	00:51	
Air Blank		0.000	00:51	
Subject Sample	# 1		00:52	Subject Test Refused
Breath Volume		0.000	00:52	
Air Blank		0.000	00:52	
Control Test		0.079	00:53	
Air Blank		0.000	00:53	
Diagnostics Check		OK	00:53	

Data Download Date/Time: 09/22/2018 23:03

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 08/25/2018 20:06:27

Diagnostics Check		OK	20:00
Air Blank		0.000	20:00
Control Test		0.080	20:01
Air Blank		0.000	20:01
Subject Sample	# 1	0.202	20:01
Breath Volume		2.894	20:01
Air Blank		0.000	20:02
Air Blank		0.000	20:04
Subject Sample	# 2	0.206	20:04
Breath Volume		2.507	20:04
Air Blank		0.000	20:05
Control Test		0.081	20:05
Air Blank		0.000	20:06
Diagnostics Check		OK	20:06

Data Download Date/Time: 09/22/2018 23:03

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DAVID,O
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SANDERS,DAVID
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: EMERY

DATE/TIME OF TEST: 08/25/2018 21:39:16

Diagnostics Check		OK	21:29
Air Blank		0.000	21:30
Control Test		0.080	21:30
Air Blank		0.000	21:31
Subject Sample	# 1	0.114	21:33
Breath Volume		1.519	21:33
Air Blank		0.000	21:34
Air Blank		0.000	21:36
Subject Sample	# 2	0.115	21:37
Breath Volume		2.558	21:37
Air Blank		0.000	21:38
Control Test		0.080	21:38
Air Blank		0.000	21:39
Diagnostics Check		OK	21:39

Data Download Date/Time: 09/22/2018 23:04

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: GOMEZ GOMEZ

DATE/TIME OF TEST: 08/25/2018 22:10:39

Diagnostics Check		OK	22:00	
Air Blank		0.000	22:01	
Control Test		0.079	22:01	
Air Blank		0.000	22:01	
Subject Sample	# 1		22:05	No Sample Provided
Breath Volume		0.000	22:05	
Air Blank		0.000	22:05	
Air Blank		0.000	22:07	
Subject Sample	# 2		22:09	Subject Test Refused
Breath Volume		0.000	22:09	
Air Blank		0.000	22:09	
Control Test		0.082	22:10	
Air Blank		0.000	22:10	
Diagnostics Check		OK	22:10	

Data Download Date/Time: 09/22/2018 23:04

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: ELIAS MORENO

DATE/TIME OF TEST: 08/04/2018 02:51:26

Diagnostics Check		OK	02:44
Air Blank		0.000	02:45
Control Test		0.079	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.220	02:47
Breath Volume		3.000	02:47
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.214	02:49
Breath Volume		2.531	02:49
Air Blank		0.000	02:50
Control Test		0.079	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 08/07/2018 11:26

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: SANTILLAN

DATE/TIME OF TEST: 08/05/2018 03:39:34

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.080	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.151	03:35
Breath Volume		4.066	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.151	03:38
Breath Volume		4.031	03:38
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 08/07/2018 11:26

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: HUDDLESTON

DATE/TIME OF TEST: 08/24/2018 21:18:14

Diagnostics Check	OK	21:17	
Air Blank		21:18	RFI Detect
Air Blank		21:18	RFI Detect

Data Download Date/Time: 09/12/2018 10:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: HUDDLESTON

DATE/TIME OF TEST: 08/24/2018 21:26:09

Diagnostics Check		OK	21:20
Air Blank		0.000	21:20
Control Test		0.079	21:21
Air Blank		0.000	21:21
Subject Sample	# 1	0.000	21:21
Breath Volume		3.050	21:21
Air Blank		0.000	21:22
Air Blank		0.000	21:24
Subject Sample	# 2	0.000	21:24
Breath Volume		1.402	21:24
Air Blank		0.000	21:25
Control Test		0.080	21:25
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 09/12/2018 10:48

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLAND,MARK,W
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 15:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: OLVERA-SANCHEZ

DATE/TIME OF TEST: 08/29/2018 16:13:38

Diagnostics Check		OK	16:00	
Air Blank		0.000	16:01	
Control Test		0.079	16:01	
Air Blank		0.000	16:02	
Subject Sample	# 1	0.386	16:05	Volume Not Met
Breath Volume		0.015	16:05	
Air Blank		0.000	16:05	
Air Blank		0.000	16:07	
Subject Sample	# 2	0.417	16:08	No .020 agreement
Breath Volume		1.328	16:08	
Air Blank		0.000	16:09	
Air Blank		0.000	16:11	
Subject Sample	# 3	0.443	16:12	No .020 agreement
Breath Volume		1.765	16:12	
Air Blank		0.000	16:12	
Control Test		0.077	16:13	
Air Blank		0.000	16:13	
Diagnostics Check		OK	16:13	

Data Download Date/Time: 09/12/2018 10:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,SALVADOR,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: GARCIA,SALVADOR
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: CORTEZ

DATE/TIME OF TEST: 08/03/2018 19:23:41

Diagnostics Check		OK	19:17
Air Blank		0.000	19:17
Control Test		0.079	19:18
Air Blank		0.000	19:18
Subject Sample	# 1	0.174	19:19
Breath Volume		2.425	19:19
Air Blank		0.000	19:19
Air Blank		0.000	19:21
Subject Sample	# 2	0.165	19:22
Breath Volume		1.996	19:22
Air Blank		0.000	19:22
Control Test		0.078	19:23
Air Blank		0.000	19:23
Diagnostics Check		OK	19:23

Data Download Date/Time: 08/27/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUEDE,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: LEEFLANG,DANIEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: KRIMSLEY

DATE/TIME OF TEST: 08/05/2018 04:38:12

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.080	04:33
Air Blank		0.000	04:33
Subject Sample	# 1	0.000	04:34
Breath Volume		1.687	04:34
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.000	04:36
Breath Volume		1.359	04:36
Air Blank		0.000	04:37
Control Test		0.079	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 08/27/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INSUASTY,SANDRA,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: GASPAR,BENJAMIN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 08/10/2018 04:47:08

Diagnostics Check		OK	04:40
Air Blank		0.000	04:40
Control Test		0.078	04:41
Air Blank		0.000	04:41
Subject Sample	# 1	0.181	04:42
Breath Volume		1.429	04:42
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.184	04:45
Breath Volume		2.093	04:45
Air Blank		0.000	04:46
Control Test		0.077	04:46
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 08/27/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OJEDA,NELSON,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: WILLIAMS,MICHAEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MASUCCI

DATE/TIME OF TEST: 08/21/2018 00:48:56

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.079	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.191	00:44
Breath Volume		1.644	00:44
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.192	00:47
Breath Volume		2.324	00:47
Air Blank		0.000	00:48
Control Test		0.078	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 08/27/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIERNAN,STEPHANYE,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: REYNOLDS,MATHEW
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: KIERNAN

DATE/TIME OF TEST: 08/22/2018 04:43:39

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.079	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.000	04:39
Breath Volume		1.761	04:39
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.000	04:42
Breath Volume		1.414	04:42
Air Blank		0.000	04:42
Control Test		0.078	04:43
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 08/27/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUEDE,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MCINERNEY,JESSE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: CAULEY

DATE/TIME OF TEST: 08/26/2018 04:04:46

Diagnostics Check		OK	03:49	
Air Blank		0.000	03:49	
Control Test		0.079	03:49	
Air Blank		0.000	03:50	
Subject Sample	# 1	0.194	03:53	Volume Not Met
Breath Volume		0.074	03:53	
Air Blank		0.000	03:54	
Air Blank		0.000	03:56	
Subject Sample	# 2	0.186	03:58	
Breath Volume		1.519	03:58	
Air Blank		0.000	03:59	
Air Blank		0.000	04:01	
Subject Sample	# 3	0.199	04:03	
Breath Volume		1.199	04:03	
Air Blank		0.000	04:03	
Control Test		0.076	04:04	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 08/27/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRING,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HERRING,TIMOTHY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MULROONEY

DATE/TIME OF TEST: 08/29/2018 05:34:20

Diagnostics Check		OK	05:27
Air Blank		0.000	05:27
Control Test		0.079	05:27
Air Blank		0.000	05:28
Subject Sample	# 1	0.104	05:29
Breath Volume		2.000	05:29
Air Blank		0.000	05:30
Air Blank		0.000	05:32
Subject Sample	# 2	0.106	05:32
Breath Volume		2.621	05:32
Air Blank		0.000	05:33
Control Test		0.078	05:33
Air Blank		0.000	05:34
Diagnostics Check		OK	05:34

Data Download Date/Time: 09/24/2018 11:49

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,SALVADOR,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: GARCIA,SALVADOR
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: HUFF

DATE/TIME OF TEST: 08/31/2018 22:59:57

Diagnostics Check		OK	22:56	
Air Blank		0.000	22:56	
Control Test		0.079	22:57	
Air Blank		0.000	22:57	
Subject Sample	# 1		22:58	Subject Test Refused
Breath Volume		0.210	22:58	
Air Blank		0.000	22:59	
Control Test		0.078	22:59	
Air Blank		0.000	22:59	
Diagnostics Check		OK	22:59	

Data Download Date/Time: 09/24/2018 11:49

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANTORSKI,MARK,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEREZ,CRISTAL
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: BRANDARIS

DATE/TIME OF TEST: 08/01/2018 01:03:09

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:00	
Control Test		0.079	01:00	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:01	Subject Test Refused
Breath Volume		0.000	01:01	
Air Blank		0.000	01:02	
Control Test		0.080	01:02	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 08/17/2018 05:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,LATONYA,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: WARD,ANDREW
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 08/03/2018 00:38:52

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.079	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.244	00:34
Breath Volume		2.058	00:34
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.246	00:37
Breath Volume		1.652	00:37
Air Blank		0.000	00:37
Control Test		0.078	00:38
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 08/17/2018 05:30

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GIBSON,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 13:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:35

SUBJECT'S LAST NAME: PERONARD

DATE/TIME OF TEST: 08/04/2018 14:09:50

Diagnostics Check		OK	14:03
Air Blank		0.000	14:03
Control Test		0.079	14:03
Air Blank		0.000	14:04
Subject Sample	# 1	0.264	14:05
Breath Volume		2.230	14:05
Air Blank		0.000	14:05
Air Blank		0.000	14:07
Subject Sample	# 2	0.261	14:08
Breath Volume		3.054	14:08
Air Blank		0.000	14:08
Control Test		0.078	14:09
Air Blank		0.000	14:09
Diagnostics Check		OK	14:09

Data Download Date/Time: 08/17/2018 05:30

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 15:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: MANIERRE

DATE/TIME OF TEST: 08/06/2018 18:22:29

Diagnostics Check		OK	18:19	
Air Blank		0.000	18:20	
Control Test		0.080	18:20	
Air Blank		0.000	18:20	
Subject Sample	# 1		18:21	Subject Test Refused
Breath Volume		0.000	18:21	
Air Blank		0.000	18:21	
Control Test		0.079	18:21	
Air Blank		0.000	18:22	
Diagnostics Check		OK	18:22	

Data Download Date/Time: 08/17/2018 05:30

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LIGHTSEY,D
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: KOUYOUMJIAN

DATE/TIME OF TEST: 08/08/2018 00:24:49

Diagnostics Check		OK	00:15
Air Blank		0.000	00:16
Control Test		0.079	00:16
Air Blank		0.000	00:16
Subject Sample	# 1	0.190	00:18
Breath Volume		1.679	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.192	00:23
Breath Volume		1.234	00:23
Air Blank		0.000	00:23
Control Test		0.078	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 08/17/2018 05:30

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GALLAGHER,M
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: GALBRAITH

DATE/TIME OF TEST: 08/08/2018 23:04:56

Diagnostics Check		OK	22:58
Air Blank		0.000	22:59
Control Test		0.079	22:59
Air Blank		0.000	23:00
Subject Sample	# 1	0.147	23:00
Breath Volume		1.488	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:02
Subject Sample	# 2	0.154	23:03
Breath Volume		1.671	23:03
Air Blank		0.000	23:04
Control Test		0.079	23:04
Air Blank		0.000	23:04
Diagnostics Check		OK	23:04

Data Download Date/Time: 08/17/2018 05:30

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JURSIK,M
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: SHEPHERD

DATE/TIME OF TEST: 08/10/2018 21:46:25

Diagnostics Check		OK	21:38
Air Blank		0.000	21:39
Control Test		0.079	21:39
Air Blank		0.000	21:39
Subject Sample	# 1	0.244	21:41
Breath Volume		1.546	21:41
Air Blank		0.000	21:41
Air Blank		0.000	21:43
Subject Sample	# 2	0.243	21:44
Breath Volume		1.464	21:44
Air Blank		0.000	21:45
Control Test		0.078	21:45
Air Blank		0.000	21:46
Diagnostics Check		OK	21:46

Data Download Date/Time: 08/17/2018 05:30

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LIGHTSEY,D
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 08/11/2018 20:38:39

Diagnostics Check		OK	20:32
Air Blank		0.000	20:33
Control Test		0.079	20:33
Air Blank		0.000	20:34
Subject Sample	# 1	0.000	20:34
Breath Volume		3.152	20:34
Air Blank		0.000	20:34
Air Blank		0.000	20:36
Subject Sample	# 2	0.000	20:37
Breath Volume		3.585	20:37
Air Blank		0.000	20:37
Control Test		0.079	20:38
Air Blank		0.000	20:38
Diagnostics Check		OK	20:38

Data Download Date/Time: 08/17/2018 05:30

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,BRIAN,P
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: GREEN,BRIAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: DICKSON

DATE/TIME OF TEST: 08/11/2018 23:02:36

Diagnostics Check		OK	22:59	
Air Blank		0.000	23:00	
Control Test		0.079	23:00	
Air Blank		0.000	23:01	
Subject Sample	# 1		23:01	Subject Test Refused
Breath Volume		0.000	23:01	
Air Blank		0.000	23:01	
Control Test		0.080	23:02	
Air Blank		0.000	23:02	
Diagnostics Check		OK	23:02	

Data Download Date/Time: 08/17/2018 05:30

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GETCHELL,ERIC
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: GRANHOLM

DATE/TIME OF TEST: 08/12/2018 04:29:20

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.079	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.180	04:24
Breath Volume		2.511	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.173	04:27
Breath Volume		1.687	04:27
Air Blank		0.000	04:28
Control Test		0.079	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 08/17/2018 05:30

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GETCHELL,ERIC
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: DAWKINS

DATE/TIME OF TEST: 08/18/2018 02:42:21

Diagnostics Check		OK	02:39	
Air Blank		0.000	02:39	
Control Test		0.079	02:40	
Air Blank		0.000	02:40	
Subject Sample	# 1		02:40	Subject Test Refused
Breath Volume		0.000	02:40	
Air Blank		0.000	02:41	
Control Test		0.079	02:41	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:42	

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,LATONYA,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: NEGRON,SHAWN
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 11:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DE COSA JR

DATE/TIME OF TEST: 08/18/2018 03:29:44

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.079	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.000	03:25
Breath Volume		1.664	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.000	03:28
Breath Volume		1.671	03:28
Air Blank		0.000	03:28
Control Test		0.079	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HERNANDEZ,DANIEL
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BAJANA

DATE/TIME OF TEST: 08/18/2018 03:55:01

Diagnostics Check		OK	03:47
Air Blank		0.000	03:48
Control Test		0.080	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.170	03:49
Breath Volume		1.656	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.158	03:53
Breath Volume		1.289	03:53
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:54
Diagnostics Check		OK	03:55

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,LATONYA,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DONSIFE,MATHEW
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: RORICK

DATE/TIME OF TEST: 08/18/2018 04:26:16

Diagnostics Check		OK	04:17
Air Blank		0.000	04:18
Control Test		0.079	04:18
Air Blank		0.000	04:19
Subject Sample	# 1	0.121	04:20
Breath Volume		2.457	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:22
Subject Sample	# 2	0.114	04:24
Breath Volume		1.660	04:24
Air Blank		0.000	04:25
Control Test		0.079	04:25
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:07

SUBJECT'S LAST NAME: GATZKE

DATE/TIME OF TEST: 08/18/2018 18:39:46

Diagnostics Check		OK	18:32
Air Blank		0.000	18:32
Control Test		0.078	18:32
Air Blank		0.000	18:33
Subject Sample	# 1	0.257	18:35
Breath Volume		1.125	18:35
Air Blank		0.000	18:36
Air Blank		0.000	18:37
Subject Sample	# 2	0.269	18:38
Breath Volume		1.195	18:38
Air Blank		0.000	18:38
Control Test		0.077	18:39
Air Blank		0.000	18:39
Diagnostics Check		OK	18:39

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,BRIAN,P
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: GREEN,BRIAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: LAVIGNE

DATE/TIME OF TEST: 08/21/2018 21:26:40

Diagnostics Check		OK	21:15	
Air Blank		0.000	21:15	
Control Test		0.079	21:15	
Air Blank		0.000	21:16	
Subject Sample	# 1		21:19	No Sample Provided
Breath Volume		0.000	21:19	
Air Blank		0.000	21:20	
Air Blank		0.000	21:22	
Subject Sample	# 2	0.064	21:25	Volume Not Met
Breath Volume		0.968	21:25	
Air Blank		0.000	21:25	
Control Test		0.079	21:26	
Air Blank		0.000	21:26	
Diagnostics Check		OK	21:26	

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,BRIAN,P
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: GREEN,BRIAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: LAVIGNE

DATE/TIME OF TEST: 08/21/2018 21:32:26

Diagnostics Check		OK	21:28	
Air Blank		0.000	21:29	
Control Test		0.079	21:29	
Air Blank		0.000	21:29	
Subject Sample	# 1		21:30	Subject Test Refused
Breath Volume		0.000	21:30	
Air Blank		0.000	21:31	
Control Test		0.079	21:31	
Air Blank		0.000	21:32	
Diagnostics Check		OK	21:32	

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: NIEVES,L
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: KENDLE

DATE/TIME OF TEST: 08/23/2018 17:21:49

Diagnostics Check		OK	17:13
Air Blank		0.000	17:13
Control Test		0.079	17:14
Air Blank		0.000	17:14
Subject Sample	# 1	0.235	17:15
Breath Volume		1.468	17:15
Air Blank		0.000	17:16
Air Blank		0.000	17:18
Subject Sample	# 2	0.224	17:20
Breath Volume		1.828	17:20
Air Blank		0.000	17:20
Control Test		0.079	17:21
Air Blank		0.000	17:21
Diagnostics Check		OK	17:21

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: WHITE,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: BASART

DATE/TIME OF TEST: 08/24/2018 23:01:58

Diagnostics Check		OK	22:51
Air Blank		0.000	22:51
Control Test		0.079	22:51
Air Blank		0.000	22:52
Subject Sample	# 1	0.223	22:53
Breath Volume		1.906	22:53
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.200	22:56
Breath Volume		1.398	22:56
Air Blank		0.000	22:56
Air Blank		0.000	22:58
Subject Sample	# 3	0.222	23:00
Breath Volume		1.441	23:00
Air Blank		0.000	23:01
Control Test		0.078	23:01
Air Blank		0.000	23:01
Diagnostics Check		OK	23:01

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: WARD,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: BRAZINA

DATE/TIME OF TEST: 08/24/2018 23:49:07

Diagnostics Check		OK	23:43
Air Blank		0.000	23:43
Control Test		0.079	23:43
Air Blank		0.000	23:44
Subject Sample	# 1	0.186	23:44
Breath Volume		2.628	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.182	23:47
Breath Volume		2.632	23:47
Air Blank		0.000	23:48
Control Test		0.079	23:48
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JOERGER,S
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: BYNUM

DATE/TIME OF TEST: 08/25/2018 05:27:19

Diagnostics Check	OK	05:26	
Air Blank		05:27	RFI Detect
Air Blank		05:27	RFI Detect

Data Download Date/Time: 09/14/2018 03:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JOERGER,S
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: BYNUM

DATE/TIME OF TEST: 08/25/2018 05:31:26

Diagnostics Check		OK	05:29	
Air Blank		0.000	05:30	
Control Test		0.079	05:30	
Air Blank		0.000	05:31	
Subject Sample	# 1		05:31	RFI Detect
Breath Volume		0.000	05:31	
Air Blank			05:31	RFI Detect
Air Blank			05:31	RFI Detect

Data Download Date/Time: 09/14/2018 03:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JOERGER,S
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: BYNUM

DATE/TIME OF TEST: 08/25/2018 05:39:55

Diagnostics Check		OK	05:33	
Air Blank		0.000	05:34	
Control Test		0.079	05:34	
Air Blank		0.000	05:35	
Subject Sample	# 1	0.163	05:35	
Breath Volume		1.632	05:35	
Air Blank		0.000	05:36	
Air Blank		0.000	05:38	
Subject Sample	# 2		05:38	RFI Detect
Breath Volume		0.000	05:38	
Air Blank		0.000	05:39	
Control Test		0.078	05:39	
Air Blank		0.000	05:39	
Diagnostics Check		OK	05:39	

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JOERGER,S
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: BYNUM

DATE/TIME OF TEST: 08/25/2018 05:44:42

Diagnostics Check		OK	05:42	
Air Blank		0.000	05:42	
Control Test		0.079	05:43	
Air Blank		0.000	05:43	
Subject Sample	# 1		05:44	RFI Detect
Breath Volume		0.000	05:44	
Air Blank			05:44	RFI Detect
Air Blank		0.000	05:44	

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JOERGER,S
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: BYNUM

DATE/TIME OF TEST: 08/25/2018 06:29:43

Diagnostics Check		OK	06:23
Air Blank		0.000	06:23
Control Test		0.079	06:24
Air Blank		0.000	06:24
Subject Sample	# 1	0.153	06:25
Breath Volume		1.980	06:25
Air Blank		0.000	06:25
Air Blank		0.000	06:27
Subject Sample	# 2	0.158	06:28
Breath Volume		1.828	06:28
Air Blank		0.000	06:28
Control Test		0.079	06:29
Air Blank		0.000	06:29
Diagnostics Check		OK	06:29

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SINGH,T
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HOOD

DATE/TIME OF TEST: 08/25/2018 06:37:29

Diagnostics Check		OK	06:34	
Air Blank		0.000	06:35	
Control Test		0.079	06:35	
Air Blank		0.000	06:35	
Subject Sample	# 1		06:36	Subject Test Refused
Breath Volume		0.000	06:36	
Air Blank		0.000	06:36	
Control Test		0.078	06:36	
Air Blank		0.000	06:37	
Diagnostics Check		OK	06:37	

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YORK,JIMMY,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GETCHELL,ERIC
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: DOBSON

DATE/TIME OF TEST: 08/25/2018 23:23:27

Diagnostics Check		OK	23:16
Air Blank		0.000	23:17
Control Test		0.079	23:17
Air Blank		0.000	23:17
Subject Sample	# 1	0.174	23:18
Breath Volume		1.667	23:18
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.180	23:21
Breath Volume		1.101	23:21
Air Blank		0.000	23:22
Control Test		0.079	23:22
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SINGH,TEGPREET
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: SALVINO

DATE/TIME OF TEST: 08/26/2018 01:30:20

Diagnostics Check		OK	01:21
Air Blank		0.000	01:21
Control Test		0.079	01:22
Air Blank		0.000	01:22
Subject Sample	# 1	0.185	01:25
Breath Volume		1.277	01:25
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.173	01:28
Breath Volume		1.253	01:28
Air Blank		0.000	01:29
Control Test		0.079	01:29
Air Blank		0.000	01:30
Diagnostics Check		OK	01:30

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JOERGER,S
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 08/26/2018 04:19:40

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:07	
Control Test		0.079	04:07	
Air Blank		0.000	04:08	
Subject Sample	# 1	0.159	04:09	
Breath Volume		1.507	04:09	
Air Blank		0.000	04:10	
Air Blank		0.000	04:12	
Subject Sample	# 2	0.149	04:15	Volume Not Met
Breath Volume		1.078	04:15	
Air Blank		0.000	04:16	
Air Blank		0.000	04:18	
Subject Sample	# 3		04:18	Subject Test Refused
Breath Volume		0.000	04:18	
Air Blank		0.000	04:18	
Control Test		0.078	04:19	
Air Blank		0.000	04:19	
Diagnostics Check		OK	04:19	

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TANGEL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DELEON,TEA
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: MCCOOL

DATE/TIME OF TEST: 08/27/2018 15:05:49

Diagnostics Check		OK	15:03	
Air Blank		0.000	15:03	
Control Test		0.079	15:03	
Air Blank		0.000	15:04	
Subject Sample	# 1		15:04	Subject Test Refused
Breath Volume		0.000	15:04	
Air Blank		0.000	15:04	
Control Test		0.079	15:05	
Air Blank		0.000	15:05	
Diagnostics Check		OK	15:05	

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,QUINTON,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: 1723,M
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 17:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 08/29/2018 18:18:11

Diagnostics Check		OK	18:10
Air Blank		0.000	18:10
Control Test		0.079	18:11
Air Blank		0.000	18:11
Subject Sample	# 1	0.121	18:13
Breath Volume		3.570	18:13
Air Blank		0.000	18:14
Air Blank		0.000	18:16
Subject Sample	# 2	0.119	18:16
Breath Volume		3.734	18:16
Air Blank		0.000	18:17
Control Test		0.078	18:17
Air Blank		0.000	18:18
Diagnostics Check		OK	18:18

Data Download Date/Time: 09/14/2018 03:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDILL,CHERI,L
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: COX,RILEY
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 07/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: SUTPHIN

DATE/TIME OF TEST: 08/16/2018 20:09:15

Diagnostics Check		OK	20:02
Air Blank		0.000	20:03
Control Test		0.081	20:03
Air Blank		0.000	20:04
Subject Sample	# 1	0.337	20:04
Breath Volume		1.898	20:04
Air Blank		0.000	20:05
Air Blank		0.000	20:07
Subject Sample	# 2	0.329	20:07
Breath Volume		1.121	20:07
Air Blank		0.000	20:08
Control Test		0.080	20:08
Air Blank		0.000	20:09
Diagnostics Check		OK	20:09

Data Download Date/Time: 08/21/2018 21:13

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JOHN,P
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BAKER,DUSTIN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 07/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 08/19/2018 20:40:29

Diagnostics Check		OK	20:34
Air Blank		0.000	20:34
Control Test		0.082	20:34
Air Blank		0.000	20:35
Subject Sample	# 1	0.071	20:36
Breath Volume		4.152	20:36
Air Blank		0.000	20:36
Air Blank		0.000	20:38
Subject Sample	# 2	0.071	20:39
Breath Volume		4.355	20:39
Air Blank		0.000	20:39
Control Test		0.081	20:39
Air Blank		0.000	20:40
Diagnostics Check		OK	20:40

Data Download Date/Time: 08/21/2018 21:13

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMAYA,PETER,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: AMAYA,PETER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/23/2018 03:58:42

Diagnostics Check		OK	03:55	
Air Blank		0.000	03:56	
Control Test		0.081	03:56	
Air Blank		0.000	03:56	
Subject Sample	# 1		03:57	Subject Test Refused
Breath Volume		0.000	03:57	
Air Blank		0.000	03:57	
Control Test		0.081	03:58	
Air Blank		0.000	03:58	
Diagnostics Check		OK	03:58	

Data Download Date/Time: 09/24/2018 19:32

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PERRY,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SQUIRES

DATE/TIME OF TEST: 08/05/2018 02:41:24

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:34	
Control Test		0.079	02:35	
Air Blank		0.000	02:35	
Subject Sample	# 1	0.080	02:36	
Breath Volume		2.734	02:36	
Air Blank		0.000	02:36	
Air Blank		0.000	02:38	
Subject Sample	# 2	0.074	02:39	
Breath Volume		2.011	02:39	
Air Blank		0.000	02:40	
Control Test		0.078	02:40	
Air Blank			02:40	RFI Detect
Air Blank		0.000	02:41	

Data Download Date/Time: 08/21/2018 07:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PERRY,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SQUIRES

DATE/TIME OF TEST: 08/05/2018 02:52:02

Diagnostics Check		OK	02:45
Air Blank		0.000	02:46
Control Test		0.078	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.071	02:47
Breath Volume		2.125	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:49
Subject Sample	# 2	0.068	02:50
Breath Volume		1.679	02:50
Air Blank		0.000	02:51
Control Test		0.078	02:51
Air Blank		0.000	02:51
Diagnostics Check		OK	02:52

Data Download Date/Time: 08/21/2018 07:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BAUTISTA CALDERON

DATE/TIME OF TEST: 08/05/2018 03:04:20

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.078	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.182	02:59
Breath Volume		1.347	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:01
Subject Sample	# 2	0.180	03:02
Breath Volume		1.574	03:02
Air Blank		0.000	03:03
Control Test		0.077	03:03
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 08/21/2018 07:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 08/07/2018 03:53:09

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:50	
Control Test		0.079	03:51	
Air Blank		0.000	03:51	
Subject Sample	# 1		03:51	Subject Test Refused
Breath Volume		0.000	03:51	
Air Blank		0.000	03:52	
Control Test		0.079	03:52	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 08/21/2018 07:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: STOLL,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 08/07/2018 23:52:32

Diagnostics Check		OK	23:42	
Air Blank		0.000	23:42	
Control Test		0.078	23:43	
Air Blank		0.000	23:43	
Subject Sample	# 1		23:46	No Sample Provided
Breath Volume		0.000	23:46	
Air Blank		0.000	23:47	
Air Blank		0.000	23:49	
Subject Sample	# 2		23:51	Subject Test Refused
Breath Volume		0.000	23:51	
Air Blank		0.000	23:51	
Control Test		0.079	23:51	
Air Blank		0.000	23:52	
Diagnostics Check		OK	23:52	

Data Download Date/Time: 08/21/2018 07:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: STOLL,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 08/08/2018 00:01:20

Diagnostics Check		OK	23:55
Air Blank		0.000	23:55
Control Test		0.079	23:55
Air Blank		0.000	23:56
Subject Sample	# 1	0.118	23:56
Breath Volume		3.984	23:56
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 2	0.115	23:59
Breath Volume		4.402	23:59
Air Blank		0.000	00:00
Control Test		0.077	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 08/21/2018 07:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SUNDERLAND,OFFICER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: GAROFALO

DATE/TIME OF TEST: 08/08/2018 04:40:15

Diagnostics Check		OK	04:29
Air Blank		0.000	04:30
Control Test		0.078	04:30
Air Blank		0.000	04:31
Subject Sample	# 1	0.093	04:33
Breath Volume		1.128	04:33
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.095	04:38
Breath Volume		2.171	04:38
Air Blank		0.000	04:39
Control Test		0.079	04:39
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 08/21/2018 07:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: OSTROSKI,NICOLE
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: WARMAN

DATE/TIME OF TEST: 08/09/2018 06:29:36

Diagnostics Check		OK	06:22
Air Blank		0.000	06:23
Control Test		0.078	06:23
Air Blank		0.000	06:24
Subject Sample	# 1	0.160	06:24
Breath Volume		2.753	06:24
Air Blank		0.000	06:25
Air Blank		0.000	06:27
Subject Sample	# 2	0.164	06:27
Breath Volume		2.417	06:27
Air Blank		0.000	06:28
Control Test		0.077	06:28
Air Blank		0.000	06:29
Diagnostics Check		OK	06:29

Data Download Date/Time: 08/21/2018 07:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUBRANO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: VERMILLION,B
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BARTRAM

DATE/TIME OF TEST: 08/10/2018 00:59:49

Diagnostics Check		OK	00:52
Air Blank		0.000	00:53
Control Test		0.079	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.000	00:55
Breath Volume		1.468	00:55
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.000	00:58
Breath Volume		1.742	00:58
Air Blank		0.000	00:58
Control Test		0.079	00:59
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 08/21/2018 07:36

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: OSTROSKI,OFFICER
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: FRANNCISCO

DATE/TIME OF TEST: 08/10/2018 03:18:52

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.078	03:13
Air Blank		0.000	03:13
Subject Sample	# 1	0.136	03:14
Breath Volume		1.687	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:16
Subject Sample	# 2	0.139	03:17
Breath Volume		2.523	03:17
Air Blank		0.000	03:17
Control Test		0.079	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 08/21/2018 07:36

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLYNN,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CHRISTIE,J
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: RYAN-OVESEN

DATE/TIME OF TEST: 08/11/2018 00:35:18

Diagnostics Check		OK	00:22	
Air Blank		0.000	00:23	
Control Test		0.079	00:23	
Air Blank		0.000	00:23	
Subject Sample	# 1	0.115	00:27	Volume Not Met
Breath Volume		0.191	00:27	
Air Blank		0.000	00:27	
Air Blank		0.000	00:29	
Subject Sample	# 2	0.149	00:30	
Breath Volume		2.410	00:30	
Air Blank		0.000	00:31	
Air Blank		0.000	00:32	
Subject Sample	# 3	0.146	00:33	
Breath Volume		1.910	00:33	
Air Blank		0.000	00:34	
Control Test		0.074	00:34	Control Outside Tolerance
Air Blank		0.000	00:35	

Data Download Date/Time: 08/21/2018 07:36

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLYNN,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CHRISTIE,J
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: RYAN-OVESEN

DATE/TIME OF TEST: 08/11/2018 00:47:02

Diagnostics Check		OK	00:38
Air Blank		0.000	00:38
Control Test		0.075	00:39
Air Blank		0.000	00:39
Subject Sample	# 1	0.162	00:42
Breath Volume		1.367	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:44
Subject Sample	# 2	0.151	00:45
Breath Volume		1.527	00:45
Air Blank		0.000	00:46
Control Test		0.076	00:46
Air Blank		0.000	00:46
Diagnostics Check		OK	00:47

Data Download Date/Time: 08/21/2018 07:36

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,STEFANIE,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: HUDGINS,G
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: UNGER

DATE/TIME OF TEST: 08/11/2018 04:00:00

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:57	
Control Test		0.078	03:57	
Air Blank		0.000	03:57	
Subject Sample	# 1		03:58	Subject Test Refused
Breath Volume		0.000	03:58	
Air Blank		0.000	03:59	
Control Test		0.077	03:59	
Air Blank		0.000	03:59	
Diagnostics Check		OK	04:00	

Data Download Date/Time: 08/21/2018 07:36

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,HALSTON,L
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: LIMA,VANIR
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 08/11/2018 20:43:27

Diagnostics Check		OK	20:36
Air Blank		0.000	20:37
Control Test		0.079	20:37
Air Blank		0.000	20:38
Subject Sample	# 1	0.173	20:38
Breath Volume		4.328	20:38
Air Blank		0.000	20:39
Air Blank		0.000	20:41
Subject Sample	# 2	0.161	20:41
Breath Volume		3.238	20:41
Air Blank		0.000	20:42
Control Test		0.078	20:42
Air Blank		0.000	20:43
Diagnostics Check		OK	20:43

Data Download Date/Time: 08/21/2018 07:36

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,HALSTON,L
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: COX,DAVID
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: AGOVIDA

DATE/TIME OF TEST: 08/12/2018 00:32:24

Diagnostics Check		OK	00:24
Air Blank		0.000	00:25
Control Test		0.078	00:25
Air Blank		0.000	00:25
Subject Sample	# 1	0.107	00:27
Breath Volume		1.835	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.106	00:30
Breath Volume		2.339	00:30
Air Blank		0.000	00:31
Control Test		0.079	00:31
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 08/21/2018 07:36

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: GRIMES

DATE/TIME OF TEST: 08/13/2018 19:58:09

Diagnostics Check		OK	19:51
Air Blank		0.000	19:52
Control Test		0.078	19:52
Air Blank		0.000	19:53
Subject Sample	# 1	0.198	19:53
Breath Volume		3.765	19:53
Air Blank		0.000	19:54
Air Blank		0.000	19:56
Subject Sample	# 2	0.206	19:56
Breath Volume		3.738	19:56
Air Blank		0.000	19:57
Control Test		0.077	19:57
Air Blank		0.000	19:58
Diagnostics Check		OK	19:58

Data Download Date/Time: 08/21/2018 07:36

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: CAMBRIA,STEVEN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MILNER

DATE/TIME OF TEST: 08/13/2018 21:48:50

Diagnostics Check		OK	21:42
Air Blank		0.000	21:42
Control Test		0.078	21:43
Air Blank		0.000	21:43
Subject Sample	# 1	0.247	21:44
Breath Volume		1.816	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:46
Subject Sample	# 2	0.230	21:47
Breath Volume		2.105	21:47
Air Blank		0.000	21:47
Control Test		0.077	21:48
Air Blank		0.000	21:48
Diagnostics Check		OK	21:48

Data Download Date/Time: 08/21/2018 07:36

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: VERMILLION,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/14/2018 21:47:46

Diagnostics Check		OK	21:41
Air Blank		0.000	21:42
Control Test		0.079	21:42
Air Blank		0.000	21:42
Subject Sample	# 1	0.098	21:43
Breath Volume		3.253	21:43
Air Blank		0.000	21:44
Air Blank		0.000	21:45
Subject Sample	# 2	0.098	21:46
Breath Volume		2.625	21:46
Air Blank		0.000	21:46
Control Test		0.078	21:47
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 08/21/2018 07:36

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PERKINS,W
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 08/18/2018 01:21:00

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.079	01:14
Air Blank		0.000	01:15
Subject Sample	# 1	0.172	01:16
Breath Volume		1.644	01:16
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.165	01:19
Breath Volume		1.460	01:19
Air Blank		0.000	01:20
Control Test		0.079	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:21

Data Download Date/Time: 08/21/2018 07:36

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MIKLUSCAK,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CARMONAESPARZA

DATE/TIME OF TEST: 08/18/2018 02:05:38

Diagnostics Check		OK	01:59
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:00
Subject Sample	# 1	0.156	02:01
Breath Volume		3.640	02:01
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.157	02:03
Breath Volume		2.769	02:03
Air Blank		0.000	02:04
Control Test		0.078	02:05
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 08/21/2018 07:36

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CARTER,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: STUCKMEYER

DATE/TIME OF TEST: 08/18/2018 20:27:46

Diagnostics Check		OK	20:14	
Air Blank		0.000	20:15	
Control Test		0.080	20:15	
Air Blank		0.000	20:16	
Subject Sample	# 1	0.000	20:16	
Breath Volume		1.265	20:16	
Air Blank		0.000	20:17	
Air Blank		0.000	20:19	
Subject Sample	# 2	0.000	20:22	Volume Not Met
Breath Volume		0.035	20:22	
Air Blank		0.000	20:23	
Air Blank		0.000	20:25	
Subject Sample	# 3		20:26	Subject Test Refused
Breath Volume		1.085	20:26	
Air Blank		0.000	20:26	
Control Test		0.078	20:27	
Air Blank		0.000	20:27	
Diagnostics Check		OK	20:27	

Data Download Date/Time: 08/21/2018 07:36

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BALES,TIM
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 08/18/2018 23:52:54

Diagnostics Check		OK	23:46
Air Blank		0.000	23:46
Control Test		0.079	23:47
Air Blank		0.000	23:47
Subject Sample	# 1	0.030	23:48
Breath Volume		1.894	23:48
Air Blank		0.000	23:48
Air Blank		0.000	23:50
Subject Sample	# 2	0.029	23:51
Breath Volume		2.378	23:51
Air Blank		0.000	23:51
Control Test		0.079	23:52
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 08/21/2018 07:36

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: STOLL,A
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 08/19/2018 00:10:46

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.077	00:02
Air Blank		0.000	00:02
Subject Sample	# 1	0.235	00:03
Breath Volume		2.246	00:03
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.200	00:06
Breath Volume		1.550	00:06
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 3	0.232	00:09
Breath Volume		1.925	00:09
Air Blank		0.000	00:09
Control Test		0.077	00:10
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 08/21/2018 07:36

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,D,A
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: ELMORE,D
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: SLOAN

DATE/TIME OF TEST: 08/19/2018 03:52:23

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:47
Subject Sample	# 1	0.256	03:47
Breath Volume		2.960	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.244	03:50
Breath Volume		2.710	03:50
Air Blank		0.000	03:51
Control Test		0.077	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 08/21/2018 07:36

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLYNN,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BLOMGREN,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MASKEWIT

DATE/TIME OF TEST: 08/21/2018 00:33:46

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.079	00:28
Air Blank		0.000	00:28
Subject Sample	# 1	0.097	00:29
Breath Volume		3.269	00:29
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.098	00:32
Breath Volume		3.085	00:32
Air Blank		0.000	00:32
Control Test		0.078	00:33
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 08/21/2018 07:36

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHROTH,B
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: OCONNOR

DATE/TIME OF TEST: 08/21/2018 05:41:42

Diagnostics Check		OK	05:35
Air Blank		0.000	05:35
Control Test		0.079	05:36
Air Blank		0.000	05:36
Subject Sample	# 1	0.218	05:37
Breath Volume		3.542	05:37
Air Blank		0.000	05:37
Air Blank		0.000	05:39
Subject Sample	# 2	0.210	05:40
Breath Volume		3.601	05:40
Air Blank		0.000	05:40
Control Test		0.077	05:41
Air Blank		0.000	05:41
Diagnostics Check		OK	05:41

Data Download Date/Time: 08/21/2018 07:36

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHROTH,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SAPORITO

DATE/TIME OF TEST: 08/23/2018 01:52:32

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.079	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.162	01:47
Breath Volume		2.394	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.152	01:50
Breath Volume		1.632	01:50
Air Blank		0.000	01:51
Control Test		0.077	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 09/30/2018 07:47

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLYNN,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DREZNIN,M
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: RISPOLI

DATE/TIME OF TEST: 08/25/2018 00:34:18

Diagnostics Check		OK	00:28
Air Blank		0.000	00:28
Control Test		0.080	00:28
Air Blank		0.000	00:29
Subject Sample	# 1	0.089	00:29
Breath Volume		3.464	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.085	00:32
Breath Volume		1.800	00:32
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

Data Download Date/Time: 09/30/2018 07:47

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATTHEW
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: PALOMAA

DATE/TIME OF TEST: 08/25/2018 01:48:16

Diagnostics Check		OK	01:41
Air Blank		0.000	01:42
Control Test		0.079	01:42
Air Blank		0.000	01:43
Subject Sample	# 1	0.247	01:43
Breath Volume		4.253	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.248	01:46
Breath Volume		3.699	01:46
Air Blank		0.000	01:47
Control Test		0.078	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 09/30/2018 07:47

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: VARGAS CONTRERAS

DATE/TIME OF TEST: 08/25/2018 01:59:05

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.078	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.215	01:53
Breath Volume		1.882	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.208	01:57
Breath Volume		2.027	01:57
Air Blank		0.000	01:58
Control Test		0.078	01:58
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 09/30/2018 07:48

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCI,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TUCCI,JOSE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: ZAPATA CASAS

DATE/TIME OF TEST: 08/25/2018 04:32:16

Diagnostics Check		OK	04:20	
Air Blank		0.000	04:21	
Control Test		0.079	04:21	
Air Blank		0.000	04:21	
Subject Sample	# 1	0.149	04:25	Volume Not Met
Breath Volume		0.101	04:25	
Air Blank		0.000	04:25	
Air Blank		0.000	04:27	
Subject Sample	# 2	0.145	04:30	Volume Not Met
Breath Volume		0.089	04:30	
Air Blank		0.000	04:31	
Control Test		0.076	04:31	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:32	

Data Download Date/Time: 09/30/2018 07:48

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCI,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TUCCI,JOSE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: KNICELEY

DATE/TIME OF TEST: 08/26/2018 01:32:03

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.080	01:25
Air Blank		0.000	01:25
Subject Sample	# 1	0.239	01:26
Breath Volume		5.101	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.257	01:30
Breath Volume		4.710	01:30
Air Blank		0.000	01:31
Control Test		0.078	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:32

Data Download Date/Time: 09/30/2018 07:48

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MOFFAT,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 08/26/2018 23:12:32

Diagnostics Check		OK	23:05
Air Blank		0.000	23:06
Control Test		0.081	23:06
Air Blank		0.000	23:07
Subject Sample	# 1	0.382	23:07
Breath Volume		3.042	23:07
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.378	23:10
Breath Volume		1.433	23:10
Air Blank		0.000	23:11
Control Test		0.078	23:11
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 09/30/2018 07:48

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUBRANO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CARDWELL,C
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: DIMASCIO

DATE/TIME OF TEST: 08/28/2018 00:26:06

Diagnostics Check		OK	00:19
Air Blank		0.000	00:19
Control Test		0.080	00:20
Air Blank		0.000	00:20
Subject Sample	# 1	0.207	00:21
Breath Volume		3.691	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:23
Subject Sample	# 2	0.205	00:24
Breath Volume		3.644	00:24
Air Blank		0.000	00:25
Control Test		0.078	00:25
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 09/30/2018 07:48

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOLADVANDI,M
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: FYLONENKO

DATE/TIME OF TEST: 08/28/2018 03:26:59

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.080	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.140	03:22
Breath Volume		5.996	03:22
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.142	03:25
Breath Volume		5.058	03:25
Air Blank		0.000	03:26
Control Test		0.079	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 09/30/2018 07:48

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,SEAN,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ANDREANO,JACK
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 08/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 08/05/2018 00:40:19

Diagnostics Check		OK	00:31
Air Blank		0.000	00:32
Control Test		0.081	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.164	00:34
Breath Volume		2.265	00:34
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.165	00:38
Breath Volume		1.312	00:38
Air Blank		0.000	00:39
Control Test		0.079	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 09/25/2018 08:21

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,RONNIE,
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ABDELMESSIH,ADAM
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 08/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 13:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:42

SUBJECT'S LAST NAME: MCKEOWN

DATE/TIME OF TEST: 08/08/2018 14:29:09

Diagnostics Check		OK	14:22
Air Blank		0.000	14:23
Control Test		0.080	14:23
Air Blank		0.000	14:24
Subject Sample	# 1	0.129	14:24
Breath Volume		2.160	14:24
Air Blank		0.000	14:25
Air Blank		0.000	14:27
Subject Sample	# 2	0.139	14:27
Breath Volume		2.566	14:27
Air Blank		0.000	14:28
Control Test		0.080	14:28
Air Blank		0.000	14:29
Diagnostics Check		OK	14:29

Data Download Date/Time: 09/25/2018 08:21

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,RONNIE,
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: CORTES,VICTOR
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 08/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 10:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:28

SUBJECT'S LAST NAME: HUBER

DATE/TIME OF TEST: 08/09/2018 11:10:55

Diagnostics Check		OK	11:05
Air Blank		0.000	11:05
Control Test		0.081	11:05
Air Blank		0.000	11:06
Subject Sample	# 1	0.013	11:06
Breath Volume		2.011	11:06
Air Blank		0.000	11:07
Air Blank		0.000	11:09
Subject Sample	# 2	0.015	11:09
Breath Volume		2.343	11:09
Air Blank		0.000	11:10
Control Test		0.081	11:10
Air Blank		0.000	11:10
Diagnostics Check		OK	11:10

Data Download Date/Time: 09/25/2018 08:21

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POST,DAN,H
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: CORTES,VICTOR
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 08/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 08:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:27

SUBJECT'S LAST NAME: GUILLEN-JIMENEZ

DATE/TIME OF TEST: 08/10/2018 09:04:33

Diagnostics Check		OK	08:58
Air Blank		0.000	08:58
Control Test		0.080	08:58
Air Blank		0.000	08:59
Subject Sample	# 1	0.148	08:59
Breath Volume		1.796	08:59
Air Blank		0.000	09:00
Air Blank		0.000	09:02
Subject Sample	# 2	0.164	09:02
Breath Volume		2.859	09:02
Air Blank		0.000	09:03
Control Test		0.080	09:03
Air Blank		0.000	09:04
Diagnostics Check		OK	09:04

Data Download Date/Time: 09/25/2018 08:21

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIEVES,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: GRIEVES,MATTHEW
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 08/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: PARR

DATE/TIME OF TEST: 08/12/2018 03:09:50

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:06	
Control Test		0.079	03:06	
Air Blank		0.000	03:07	
Subject Sample	# 1		03:08	Subject Test Refused
Breath Volume		0.000	03:08	
Air Blank		0.000	03:08	
Control Test		0.081	03:09	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 09/25/2018 08:21

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ALLRED,JAMES
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 08/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 08/16/2018 23:24:25

Diagnostics Check		OK	23:18	
Air Blank		0.000	23:19	
Control Test		0.081	23:19	
Air Blank		0.000	23:20	
Subject Sample	# 1		23:22	Subject Test Refused
Breath Volume		0.000	23:22	
Air Blank		0.000	23:23	
Control Test		0.081	23:23	
Air Blank		0.000	23:24	
Diagnostics Check		OK	23:24	

Data Download Date/Time: 09/25/2018 08:21

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISKELL,ADAM,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: WILLIAMS,AUSTIN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 08/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: CZESNAKOWICZ

DATE/TIME OF TEST: 08/19/2018 00:36:55

Diagnostics Check		OK	00:33	
Air Blank		0.000	00:34	
Control Test		0.080	00:34	
Air Blank		0.000	00:35	
Subject Sample	# 1		00:35	Slope Not Met
Breath Volume		0.929	00:35	
Air Blank		0.000	00:36	
Control Test		0.081	00:36	
Air Blank		0.000	00:36	
Diagnostics Check		OK	00:36	

Data Download Date/Time: 09/25/2018 08:21

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISKELL,ADAM,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: WILLIAMS,AUSTIN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 08/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: CZESNAKOWICZ

DATE/TIME OF TEST: 08/19/2018 00:47:37

Diagnostics Check		OK	00:41
Air Blank		0.000	00:42
Control Test		0.080	00:42
Air Blank		0.000	00:42
Subject Sample	# 1	0.326	00:43
Breath Volume		2.085	00:43
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.321	00:46
Breath Volume		2.285	00:46
Air Blank		0.000	00:46
Control Test		0.079	00:47
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 09/25/2018 08:21

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: WADE,CODY
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 08/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: GARCIA DIAZ

DATE/TIME OF TEST: 08/22/2018 01:57:39

Diagnostics Check		OK	01:50
Air Blank		0.000	01:51
Control Test		0.080	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.116	01:53
Breath Volume		2.167	01:53
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.105	01:56
Breath Volume		2.546	01:56
Air Blank		0.000	01:56
Control Test		0.078	01:57
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 09/25/2018 08:21

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP],STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ALLRED,JAMES
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 08/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: GOOD

DATE/TIME OF TEST: 08/22/2018 05:49:36

Diagnostics Check		OK	05:46	
Air Blank		0.000	05:47	
Control Test		0.081	05:47	
Air Blank		0.000	05:47	
Subject Sample	# 1		05:48	Subject Test Refused
Breath Volume		0.000	05:48	
Air Blank		0.000	05:48	
Control Test		0.080	05:49	
Air Blank		0.000	05:49	
Diagnostics Check		OK	05:49	

Data Download Date/Time: 09/25/2018 08:21

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: STROUP,STEVE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 08/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: WINSHIP

DATE/TIME OF TEST: 08/26/2018 03:09:42

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.080	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.194	03:05
Breath Volume		5.417	03:05
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.192	03:08
Breath Volume		5.316	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 09/25/2018 08:21

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,SEAN,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ALLRED,JAMES
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 08/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: VILLALON

DATE/TIME OF TEST: 08/31/2018 23:26:05

Diagnostics Check		OK	23:19
Air Blank		0.000	23:19
Control Test		0.081	23:19
Air Blank		0.000	23:20
Subject Sample	# 1	0.230	23:21
Breath Volume		2.406	23:21
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.238	23:24
Breath Volume		2.441	23:24
Air Blank		0.000	23:25
Control Test		0.078	23:25
Air Blank		0.000	23:25
Diagnostics Check		OK	23:26

Data Download Date/Time: 09/25/2018 08:21

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: VAZQUEZ JR

DATE/TIME OF TEST: 08/07/2018 23:46:15

Diagnostics Check		OK	23:39
Air Blank		0.000	23:40
Control Test		0.076	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.145	23:41
Breath Volume		2.789	23:41
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.150	23:44
Breath Volume		3.871	23:44
Air Blank		0.000	23:45
Control Test		0.077	23:45
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 08/21/2018 08:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TUCCI,JOSE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: CARRERA

DATE/TIME OF TEST: 08/21/2018 23:52:58

Diagnostics Check		OK	23:46
Air Blank		0.000	23:47
Control Test		0.078	23:47
Air Blank		0.000	23:48
Subject Sample	# 1	0.216	23:48
Breath Volume		3.566	23:48
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.214	23:51
Breath Volume		2.675	23:51
Air Blank		0.000	23:52
Control Test		0.076	23:52
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 09/30/2018 07:49

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SGT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: ROLON

DATE/TIME OF TEST: 08/24/2018 00:53:36

Diagnostics Check		OK	00:47
Air Blank		0.000	00:47
Control Test		0.078	00:48
Air Blank		0.000	00:48
Subject Sample	# 1	0.000	00:49
Breath Volume		1.304	00:49
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.000	00:52
Breath Volume		1.390	00:52
Air Blank		0.000	00:52
Control Test		0.078	00:53
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 09/30/2018 07:49

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 08/17/2018 02:03:16

Diagnostics Check		OK	01:56
Air Blank		0.000	01:57
Control Test		0.078	01:57
Air Blank		0.000	01:58
Subject Sample	# 1	0.189	01:58
Breath Volume		2.632	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.178	02:01
Breath Volume		1.378	02:01
Air Blank		0.000	02:02
Control Test		0.079	02:02
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 08/21/2018 09:28

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: SPURGEON

DATE/TIME OF TEST: 08/17/2018 19:52:06

Diagnostics Check		OK	19:45
Air Blank		0.000	19:45
Control Test		0.078	19:46
Air Blank		0.000	19:46
Subject Sample	# 1	0.088	19:47
Breath Volume		3.414	19:47
Air Blank		0.000	19:48
Air Blank		0.000	19:50
Subject Sample	# 2	0.082	19:50
Breath Volume		2.222	19:50
Air Blank		0.000	19:51
Control Test		0.078	19:51
Air Blank		0.000	19:52
Diagnostics Check		OK	19:52

Data Download Date/Time: 08/21/2018 09:28

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: MOULTON

DATE/TIME OF TEST: 08/26/2018 02:09:19

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.190	02:04
Breath Volume		4.046	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.183	02:07
Breath Volume		4.214	02:07
Air Blank		0.000	02:08
Control Test		0.078	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 09/30/2018 08:42

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MARC,E
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SMITH,MARC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: DINICOLA

DATE/TIME OF TEST: 08/02/2018 00:07:57

Diagnostics Check		OK	00:01
Air Blank		0.000	00:02
Control Test		0.081	00:02
Air Blank		0.000	00:02
Subject Sample	# 1	0.253	00:03
Breath Volume		1.425	00:03
Air Blank		0.000	00:04
Air Blank		0.000	00:05
Subject Sample	# 2	0.242	00:06
Breath Volume		1.238	00:06
Air Blank		0.000	00:07
Control Test		0.080	00:07
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 08/28/2018 01:46

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MARC,E
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SOUSA,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 08/02/2018 03:53:44

Diagnostics Check		OK	03:46
Air Blank		0.000	03:47
Control Test		0.080	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.205	03:49
Breath Volume		1.933	03:49
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.208	03:52
Breath Volume		2.183	03:52
Air Blank		0.000	03:52
Control Test		0.081	03:53
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 08/28/2018 01:46

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: DONAHUE

DATE/TIME OF TEST: 08/05/2018 02:22:03

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.081	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.115	02:17
Breath Volume		2.402	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.115	02:20
Breath Volume		2.429	02:20
Air Blank		0.000	02:21
Control Test		0.082	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:22

Data Download Date/Time: 08/28/2018 01:46

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JATEFF,TIFFANY,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: JATEFF,TIFFANY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: TODD

DATE/TIME OF TEST: 08/08/2018 20:08:22

Diagnostics Check		OK	20:02
Air Blank		0.000	20:02
Control Test		0.081	20:02
Air Blank		0.000	20:03
Subject Sample	# 1	0.286	20:03
Breath Volume		3.964	20:03
Air Blank		0.000	20:04
Air Blank		0.000	20:06
Subject Sample	# 2	0.269	20:06
Breath Volume		1.808	20:06
Air Blank		0.000	20:07
Control Test		0.081	20:07
Air Blank		0.000	20:08
Diagnostics Check		OK	20:08

Data Download Date/Time: 08/28/2018 01:46

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MARC,E
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SMITH,MARC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: DE BRITO

DATE/TIME OF TEST: 08/27/2018 03:16:36

Diagnostics Check		OK	03:03	
Air Blank		0.000	03:03	
Control Test		0.081	03:03	
Air Blank		0.000	03:04	
Subject Sample	# 1	0.088	03:07	Volume Not Met
Breath Volume		0.484	03:07	
Air Blank		0.000	03:08	
Air Blank		0.000	03:09	
Subject Sample	# 2	0.096	03:11	
Breath Volume		1.132	03:11	
Air Blank		0.000	03:11	
Air Blank		0.000	03:13	
Subject Sample	# 3	0.092	03:15	
Breath Volume		1.226	03:15	
Air Blank		0.000	03:15	
Control Test		0.080	03:16	
Air Blank		0.000	03:16	
Diagnostics Check		OK	03:16	

Data Download Date/Time: 08/28/2018 01:46

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,CHANNING,L
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: LEATHEROW,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 10:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:05

SUBJECT'S LAST NAME: SMITHLING

DATE/TIME OF TEST: 08/28/2018 11:52:40

Diagnostics Check		OK	11:46
Air Blank		0.000	11:47
Control Test		0.081	11:47
Air Blank		0.000	11:48
Subject Sample	# 1	0.000	11:48
Breath Volume		1.859	11:48
Air Blank		0.000	11:48
Air Blank		0.000	11:50
Subject Sample	# 2	0.000	11:51
Breath Volume		2.246	11:51
Air Blank		0.000	11:51
Control Test		0.081	11:52
Air Blank		0.000	11:52
Diagnostics Check		OK	11:52

Data Download Date/Time: 09/06/2018 04:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MARC,E
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SMITH,MARC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: MCGURRIN

DATE/TIME OF TEST: 08/29/2018 01:20:12

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.081	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.239	01:14
Breath Volume		1.265	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.235	01:18
Breath Volume		1.136	01:18
Air Blank		0.000	01:19
Control Test		0.080	01:19
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 09/06/2018 04:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TOMLINSON,LUCAS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 08/01/2018 23:41:19

Diagnostics Check		OK	23:34
Air Blank		0.000	23:34
Control Test		0.079	23:35
Air Blank		0.000	23:35
Subject Sample	# 1	0.153	23:36
Breath Volume		3.183	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.156	23:39
Breath Volume		3.363	23:39
Air Blank		0.000	23:40
Control Test		0.080	23:40
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 08/31/2018 07:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,HENRY,E
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORAGHAN,JOHN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: MCCULLY

DATE/TIME OF TEST: 08/05/2018 15:23:31

Diagnostics Check		OK	15:17
Air Blank		0.000	15:17
Control Test		0.079	15:18
Air Blank		0.000	15:18
Subject Sample	# 1	0.173	15:19
Breath Volume		3.066	15:19
Air Blank		0.000	15:19
Air Blank		0.000	15:21
Subject Sample	# 2	0.177	15:22
Breath Volume		3.246	15:22
Air Blank		0.000	15:22
Control Test		0.078	15:22
Air Blank		0.000	15:23
Diagnostics Check		OK	15:23

Data Download Date/Time: 08/31/2018 07:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRONE-CABRERA,ANTHONY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FARLEY,JOSPEH
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: LYNGZEIDETSON

DATE/TIME OF TEST: 08/06/2018 15:18:31

Diagnostics Check		OK	15:12
Air Blank		0.000	15:13
Control Test		0.080	15:13
Air Blank		0.000	15:13
Subject Sample	# 1	0.000	15:14
Breath Volume		2.261	15:14
Air Blank		0.000	15:14
Air Blank		0.000	15:16
Subject Sample	# 2	0.000	15:17
Breath Volume		2.523	15:17
Air Blank		0.000	15:17
Control Test		0.080	15:17
Air Blank		0.000	15:18
Diagnostics Check		OK	15:18

Data Download Date/Time: 08/31/2018 07:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TOMLINSON,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MORTON

DATE/TIME OF TEST: 08/07/2018 22:43:17

Diagnostics Check		OK	22:36
Air Blank		0.000	22:36
Control Test		0.080	22:36
Air Blank		0.000	22:37
Subject Sample	# 1	0.193	22:38
Breath Volume		1.710	22:38
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.193	22:41
Breath Volume		2.132	22:41
Air Blank		0.000	22:42
Control Test		0.080	22:42
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 08/31/2018 07:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAFFARS,CELINA,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SWARTZENTRUBER,B
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 08/09/2018 03:21:27

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.080	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.141	03:17
Breath Volume		2.906	03:17
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.144	03:19
Breath Volume		2.867	03:19
Air Blank		0.000	03:20
Control Test		0.080	03:20
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 08/31/2018 07:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MUSGROVE,DAVID
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: MCGOWAN

DATE/TIME OF TEST: 08/10/2018 21:38:23

Diagnostics Check		OK	21:23	
Air Blank		0.000	21:24	
Control Test		0.080	21:24	
Air Blank		0.000	21:25	
Subject Sample	# 1	0.154	21:28	Volume Not Met
Breath Volume		0.113	21:28	
Air Blank		0.000	21:28	
Air Blank		0.000	21:30	
Subject Sample	# 2	0.160	21:32	
Breath Volume		1.128	21:32	
Air Blank		0.000	21:32	
Air Blank		0.000	21:34	
Subject Sample	# 3	0.156	21:36	
Breath Volume		1.113	21:36	
Air Blank		0.000	21:37	
Control Test		0.079	21:37	
Air Blank		0.000	21:38	
Diagnostics Check		OK	21:38	

Data Download Date/Time: 08/31/2018 07:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JONES,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 08/11/2018 01:20:41

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.144	01:14
Breath Volume		1.140	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.143	01:19
Breath Volume		1.367	01:19
Air Blank		0.000	01:19
Control Test		0.080	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 08/31/2018 07:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,DONNA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SARVER,S
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: REISS

DATE/TIME OF TEST: 08/11/2018 02:58:53

Diagnostics Check		OK	02:51
Air Blank		0.000	02:52
Control Test		0.079	02:52
Air Blank		0.000	02:53
Subject Sample	# 1	0.107	02:54
Breath Volume		2.636	02:54
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.104	02:57
Breath Volume		2.406	02:57
Air Blank		0.000	02:57
Control Test		0.079	02:58
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 08/31/2018 07:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ROSE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 08/11/2018 03:18:48

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.079	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.189	03:14
Breath Volume		4.863	03:14
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.199	03:17
Breath Volume		4.699	03:17
Air Blank		0.000	03:17
Control Test		0.079	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 08/31/2018 07:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: COX,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 08/12/2018 00:02:15

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.080	23:53
Air Blank		0.000	23:54
Subject Sample	# 2	0.159	23:56
Breath Volume		1.902	23:56
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 1	0.147	00:00
Breath Volume		1.347	00:00
Air Blank		0.000	00:01
Control Test		0.079	00:01
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 08/31/2018 07:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FREITAS,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 08/12/2018 01:24:11

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.080	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.128	01:19
Breath Volume		2.511	01:19
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.126	01:22
Breath Volume		2.750	01:22
Air Blank		0.000	01:23
Control Test		0.080	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 08/31/2018 07:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JONES,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 08/12/2018 03:46:11

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.080	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.120	03:41
Breath Volume		1.277	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.124	03:44
Breath Volume		1.410	03:44
Air Blank		0.000	03:45
Control Test		0.079	03:45
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 08/31/2018 07:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLATZ,KEVIN,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: COX,DAVID
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: PLATA AYALA

DATE/TIME OF TEST: 08/17/2018 01:06:51

Diagnostics Check		OK	00:58
Air Blank		0.000	00:59
Control Test		0.080	00:59
Air Blank		0.000	00:59
Subject Sample	# 1	0.208	01:01
Breath Volume		1.636	01:01
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.209	01:05
Breath Volume		2.019	01:05
Air Blank		0.000	01:05
Control Test		0.079	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 08/31/2018 07:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LAZARSKI,KRISTA
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: SAINT CYR

DATE/TIME OF TEST: 08/17/2018 01:30:46

Diagnostics Check		OK	01:21
Air Blank		0.000	01:22
Control Test		0.079	01:22
Air Blank		0.000	01:23
Subject Sample	# 1	0.133	01:24
Breath Volume		1.261	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:26
Subject Sample	# 2	0.135	01:29
Breath Volume		1.843	01:29
Air Blank		0.000	01:29
Control Test		0.080	01:30
Air Blank		0.000	01:30
Diagnostics Check		OK	01:30

Data Download Date/Time: 08/31/2018 07:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEPARD,C,A
BREATH TEST OPERATOR AGENCY: Punta Gorda PD
ARRESTING OFFICER: TRUFAN,J
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:23

SUBJECT'S LAST NAME: EMMERLING

DATE/TIME OF TEST: 08/18/2018 11:57:46

Diagnostics Check		OK	11:50
Air Blank		0.000	11:50
Control Test		0.080	11:50
Air Blank		0.000	11:51
Subject Sample	# 1	0.128	11:52
Breath Volume		2.113	11:52
Air Blank		0.000	11:53
Air Blank		0.000	11:55
Subject Sample	# 2	0.115	11:56
Breath Volume		1.789	11:56
Air Blank		0.000	11:56
Control Test		0.077	11:57
Air Blank		0.000	11:57
Diagnostics Check		OK	11:57

Data Download Date/Time: 08/31/2018 07:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JONES,JASON
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 08/27/2018 00:50:40

Diagnostics Check		OK	00:44
Air Blank		0.000	00:44
Control Test		0.080	00:44
Air Blank		0.000	00:45
Subject Sample	# 1	0.000	00:46
Breath Volume		1.667	00:46
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.000	00:49
Breath Volume		1.429	00:49
Air Blank		0.000	00:49
Control Test		0.080	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 08/31/2018 07:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 08/04/2018 03:56:14

Diagnostics Check		OK	03:48
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:49
Subject Sample	# 1	0.154	03:51
Breath Volume		1.292	03:51
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.154	03:54
Breath Volume		1.312	03:54
Air Blank		0.000	03:55
Control Test		0.078	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 08/16/2018 22:00

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: GLUECK

DATE/TIME OF TEST: 08/05/2018 02:30:52

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.100	02:26
Breath Volume		3.652	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:28
Subject Sample	# 2	0.099	02:29
Breath Volume		3.789	02:29
Air Blank		0.000	02:29
Control Test		0.079	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 08/16/2018 22:00

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 08/10/2018 00:55:41

Diagnostics Check		OK	00:48
Air Blank		0.000	00:48
Control Test		0.079	00:49
Air Blank		0.000	00:49
Subject Sample	# 1	0.152	00:51
Breath Volume		2.804	00:51
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.150	00:54
Breath Volume		2.691	00:54
Air Blank		0.000	00:54
Control Test		0.078	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 08/16/2018 22:00

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANKAM,DENZEL,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KANKAM,DENZEL
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: STONEROCK

DATE/TIME OF TEST: 08/11/2018 03:54:26

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.079	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.205	03:49
Breath Volume		1.765	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.209	03:52
Breath Volume		1.683	03:52
Air Blank		0.000	03:53
Control Test		0.079	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 08/16/2018 22:00

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEIMIG,EMORIE,I
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KEIMIG,EMORIE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 08/16/2018 22:34:08

Diagnostics Check		OK	22:24	
Air Blank		0.000	22:24	
Control Test		0.079	22:25	
Air Blank		0.000	22:25	
Subject Sample	# 1	0.242	22:26	
Breath Volume		1.148	22:26	
Air Blank		0.000	22:27	
Air Blank		0.000	22:29	
Subject Sample	# 2	0.240	22:32	
Breath Volume		1.875	22:32	
Air Blank		0.000	22:33	
Control Test		0.079	22:33	
Air Blank			22:33	RFI Detect
Air Blank		0.000	22:34	

Data Download Date/Time: 09/17/2018 23:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEIMIG,EMORIE,I
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KEIMIG,EMORIE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 08/16/2018 22:51:02

Diagnostics Check		OK	22:38	
Air Blank		0.000	22:38	
Control Test		0.077	22:38	
Air Blank		0.000	22:39	
Subject Sample	# 1	0.247	22:42	Volume Not Met
Breath Volume		0.902	22:42	
Air Blank		0.000	22:43	
Air Blank		0.000	22:44	
Subject Sample	# 2	0.228	22:45	
Breath Volume		1.953	22:45	
Air Blank		0.000	22:46	
Air Blank		0.000	22:48	
Subject Sample	# 3	0.245	22:49	
Breath Volume		1.316	22:49	
Air Blank		0.000	22:50	
Control Test		0.077	22:50	
Air Blank		0.000	22:50	
Diagnostics Check		OK	22:51	

Data Download Date/Time: 09/17/2018 23:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: DILLION,JUSTIN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: TYREE

DATE/TIME OF TEST: 08/17/2018 23:48:04

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.079	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.177	23:42
Breath Volume		2.058	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.173	23:46
Breath Volume		2.054	23:46
Air Blank		0.000	23:47
Control Test		0.078	23:47
Air Blank		0.000	23:47
Diagnostics Check		OK	23:48

Data Download Date/Time: 09/17/2018 23:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: WELDY

DATE/TIME OF TEST: 08/18/2018 02:41:38

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.079	02:35
Air Blank		0.000	02:36
Subject Sample	# 1	0.128	02:37
Breath Volume		3.996	02:37
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.131	02:40
Breath Volume		3.949	02:40
Air Blank		0.000	02:40
Control Test		0.079	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 09/17/2018 23:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: PURINGTON,PATRICK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 08/18/2018 06:19:28

Diagnostics Check		OK	06:12
Air Blank		0.000	06:13
Control Test		0.079	06:13
Air Blank		0.000	06:14
Subject Sample	# 1	0.211	06:14
Breath Volume		1.648	06:14
Air Blank		0.000	06:15
Air Blank		0.000	06:17
Subject Sample	# 2	0.204	06:17
Breath Volume		1.824	06:17
Air Blank		0.000	06:18
Control Test		0.078	06:18
Air Blank		0.000	06:19
Diagnostics Check		OK	06:19

Data Download Date/Time: 09/17/2018 23:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: LASSITER

DATE/TIME OF TEST: 08/19/2018 03:51:38

Diagnostics Check		OK	03:45
Air Blank		0.000	03:45
Control Test		0.079	03:45
Air Blank		0.000	03:46
Subject Sample	# 1	0.215	03:46
Breath Volume		2.464	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.208	03:50
Breath Volume		3.277	03:50
Air Blank		0.000	03:50
Control Test		0.078	03:51
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 09/17/2018 23:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: PERDUE

DATE/TIME OF TEST: 08/25/2018 04:21:21

Diagnostics Check		OK	04:05	
Air Blank		0.000	04:06	
Control Test		0.079	04:06	
Air Blank		0.000	04:07	
Subject Sample	# 1	0.000	04:10	Volume Not Met
Breath Volume		0.167	04:10	
Air Blank		0.000	04:11	
Air Blank		0.000	04:12	
Subject Sample	# 2	0.224	04:14	
Breath Volume		1.269	04:14	
Air Blank		0.000	04:15	
Air Blank		0.000	04:17	
Subject Sample	# 3	0.208	04:19	
Breath Volume		1.750	04:19	
Air Blank		0.000	04:20	
Control Test		0.078	04:20	
Air Blank		0.000	04:21	
Diagnostics Check		OK	04:21	

Data Download Date/Time: 09/17/2018 23:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANKAM,DENZEL,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KANKAM,DENZEL
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: ADAIR

DATE/TIME OF TEST: 08/25/2018 18:59:05

Diagnostics Check		OK	18:47	
Air Blank		0.000	18:48	
Control Test		0.079	18:48	
Air Blank		0.000	18:48	
Subject Sample	# 1		18:52	No Sample Provided
Breath Volume		0.000	18:52	
Air Blank		0.000	18:52	
Air Blank		0.000	18:54	
Subject Sample	# 2		18:57	No Sample Provided
Breath Volume		0.000	18:57	
Air Blank		0.000	18:58	
Control Test		0.079	18:58	
Air Blank		0.000	18:59	
Diagnostics Check		OK	18:59	

Data Download Date/Time: 09/17/2018 23:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,TRAVIS,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TURNER,KENDRICK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: POLLOCK

DATE/TIME OF TEST: 08/26/2018 00:10:30

Diagnostics Check		OK	00:01	
Air Blank		0.000	00:01	
Control Test		0.079	00:01	
Air Blank		0.000	00:02	
Subject Sample	# 1	0.259	00:03	No .020 agreement
Breath Volume		2.734	00:03	
Air Blank		0.000	00:03	
Air Blank		0.000	00:05	
Subject Sample	# 2	0.280	00:05	No .020 agreement
Breath Volume		2.703	00:05	
Air Blank		0.000	00:06	
Air Blank		0.000	00:08	
Subject Sample	# 3	0.235	00:08	No .020 agreement
Breath Volume		1.707	00:08	
Air Blank		0.000	00:09	
Control Test		0.078	00:09	
Air Blank		0.000	00:10	
Diagnostics Check		OK	00:10	

Data Download Date/Time: 09/17/2018 23:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,KENDRICK,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TURNER,KENDRICK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: POLLOCK

DATE/TIME OF TEST: 08/26/2018 00:20:13

Diagnostics Check		OK	00:16	
Air Blank		0.000	00:16	
Control Test		0.078	00:17	
Air Blank		0.000	00:17	
Subject Sample	# 1	0.267	00:18	
Breath Volume		2.039	00:18	
Air Blank			00:18	RFI Detect
Air Blank		0.000	00:19	
Control Test		0.079	00:19	
Air Blank		0.000	00:20	
Diagnostics Check		OK	00:20	

Data Download Date/Time: 09/17/2018 23:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,KENDRICK,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TURNER,KENDRICK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: POLLOCK

DATE/TIME OF TEST: 08/26/2018 00:32:43

Diagnostics Check		OK	00:26
Air Blank		0.000	00:27
Control Test		0.079	00:27
Air Blank		0.000	00:27
Subject Sample	# 1	0.263	00:28
Breath Volume		2.468	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:30
Subject Sample	# 2	0.249	00:31
Breath Volume		1.304	00:31
Air Blank		0.000	00:31
Control Test		0.078	00:32
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 09/17/2018 23:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,KENDRICK,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TURNER,KENDRICK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: FUENMAYOR PICO

DATE/TIME OF TEST: 08/26/2018 22:46:38

Diagnostics Check		OK	22:39
Air Blank		0.000	22:40
Control Test		0.079	22:40
Air Blank		0.000	22:41
Subject Sample	# 1	0.214	22:42
Breath Volume		1.214	22:42
Air Blank		0.000	22:42
Air Blank		0.000	22:44
Subject Sample	# 2	0.219	22:45
Breath Volume		1.207	22:45
Air Blank		0.000	22:45
Control Test		0.078	22:46
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 09/17/2018 23:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUNN,MARVIN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: THERIOT

DATE/TIME OF TEST: 08/28/2018 02:39:18

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:36	
Control Test		0.079	02:36	
Air Blank		0.000	02:37	
Subject Sample	# 1		02:38	RFI Detect
Breath Volume		0.000	02:38	
Air Blank			02:38	RFI Detect
Air Blank		0.000	02:39	

Data Download Date/Time: 09/17/2018 23:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUNN,MARVIN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: THERIOT

DATE/TIME OF TEST: 08/28/2018 02:52:40

Diagnostics Check		OK	02:41	
Air Blank		0.000	02:41	
Control Test		0.079	02:42	
Air Blank		0.000	02:42	
Subject Sample	# 1		02:45	No Sample Provided
Breath Volume		0.000	02:45	
Air Blank		0.000	02:46	
Air Blank		0.000	02:48	
Subject Sample	# 2		02:51	No Sample Provided
Breath Volume		0.000	02:51	
Air Blank		0.000	02:51	
Control Test		0.079	02:52	
Air Blank		0.000	02:52	
Diagnostics Check		OK	02:52	

Data Download Date/Time: 09/17/2018 23:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: NEEL

DATE/TIME OF TEST: 08/31/2018 03:39:47

Diagnostics Check		OK	03:33
Air Blank		0.000	03:34
Control Test		0.079	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.151	03:35
Breath Volume		2.222	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:37
Subject Sample	# 2	0.151	03:38
Breath Volume		2.820	03:38
Air Blank		0.000	03:38
Control Test		0.079	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 09/17/2018 23:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,SCOTT,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 08/11/2018 23:34:21

Diagnostics Check		OK	23:31	
Air Blank		0.000	23:31	
Control Test		0.078	23:32	
Air Blank		0.000	23:32	
Subject Sample	# 1		23:32	Subject Test Refused
Breath Volume		0.000	23:32	
Air Blank		0.000	23:33	
Control Test		0.079	23:33	
Air Blank		0.000	23:34	
Diagnostics Check		OK	23:34	

Data Download Date/Time: 08/29/2018 19:49

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINKOFER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: WILSON,B
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: RYCKEWAERT

DATE/TIME OF TEST: 08/14/2018 00:07:20

Diagnostics Check		OK	00:00
Air Blank		0.000	00:01
Control Test		0.079	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.190	00:02
Breath Volume		1.757	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.175	00:05
Breath Volume		1.566	00:05
Air Blank		0.000	00:06
Control Test		0.079	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 08/29/2018 19:49

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINKOFER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: LONG,K
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: BURTON

DATE/TIME OF TEST: 08/18/2018 03:11:19

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.080	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.103	03:06
Breath Volume		2.125	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:08
Subject Sample	# 2	0.099	03:09
Breath Volume		1.617	03:09
Air Blank		0.000	03:10
Control Test		0.079	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 08/29/2018 19:49

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESHER,TIFFANY,A
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: ADAMS,MATT
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 14:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/21/2018 15:51:44

Diagnostics Check		OK	15:45
Air Blank		0.000	15:45
Control Test		0.080	15:46
Air Blank		0.000	15:46
Subject Sample	# 1	0.219	15:47
Breath Volume		1.777	15:47
Air Blank		0.000	15:47
Air Blank		0.000	15:49
Subject Sample	# 2	0.220	15:50
Breath Volume		1.703	15:50
Air Blank		0.000	15:50
Control Test		0.078	15:51
Air Blank		0.000	15:51
Diagnostics Check		OK	15:51

Data Download Date/Time: 08/29/2018 19:49

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,SCOTT,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: STUBBS

DATE/TIME OF TEST: 08/26/2018 04:05:34

Diagnostics Check		OK	03:58
Air Blank		0.000	03:59
Control Test		0.079	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.238	04:00
Breath Volume		3.140	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.240	04:03
Breath Volume		4.171	04:03
Air Blank		0.000	04:04
Control Test		0.078	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 08/29/2018 19:49

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOOS,JAMES,P
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: TUER,DAVID
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: COULIBALY

DATE/TIME OF TEST: 08/10/2018 22:06:13

Diagnostics Check		OK	21:59
Air Blank		0.000	22:00
Control Test		0.078	22:00
Air Blank		0.000	22:01
Subject Sample	# 1	0.169	22:01
Breath Volume		1.453	22:01
Air Blank		0.000	22:02
Air Blank		0.000	22:04
Subject Sample	# 2	0.175	22:04
Breath Volume		2.976	22:04
Air Blank		0.000	22:05
Control Test		0.079	22:05
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 08/30/2018 11:58

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 08/01/2018 04:47:16

Diagnostics Check		OK	04:40
Air Blank		0.000	04:41
Control Test		0.081	04:41
Air Blank		0.000	04:42
Subject Sample	# 1	0.181	04:42
Breath Volume		2.808	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.178	04:45
Breath Volume		4.289	04:45
Air Blank		0.000	04:46
Control Test		0.081	04:46
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 08/14/2018 07:13

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: BARLEY

DATE/TIME OF TEST: 08/08/2018 01:56:00

Diagnostics Check		OK	01:48
Air Blank		0.000	01:49
Control Test		0.081	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.304	01:50
Breath Volume		2.093	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:52
Subject Sample	# 2	0.306	01:54
Breath Volume		1.445	01:54
Air Blank		0.000	01:55
Control Test		0.081	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:56

Data Download Date/Time: 08/14/2018 07:13

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WARNER,MICHAEL
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: ALVAREZ ROSALES

DATE/TIME OF TEST: 08/10/2018 04:32:36

Diagnostics Check		OK	04:26
Air Blank		0.000	04:27
Control Test		0.081	04:27
Air Blank		0.000	04:27
Subject Sample	# 1	0.148	04:28
Breath Volume		2.363	04:28
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.148	04:31
Breath Volume		1.445	04:31
Air Blank		0.000	04:31
Control Test		0.081	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 08/14/2018 07:13

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOONDA,SHUAIB
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: HURST

DATE/TIME OF TEST: 08/13/2018 03:37:20

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.081	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.196	03:32
Breath Volume		1.234	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.196	03:35
Breath Volume		1.324	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 08/14/2018 07:13

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 08/17/2018 04:58:27

Diagnostics Check		OK	04:52
Air Blank		0.000	04:52
Control Test		0.081	04:53
Air Blank		0.000	04:53
Subject Sample	# 1	0.201	04:54
Breath Volume		1.652	04:54
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.210	04:56
Breath Volume		1.468	04:56
Air Blank		0.000	04:57
Control Test		0.080	04:57
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 08/28/2018 12:17

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MENDOZA,ANGEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: GAY

DATE/TIME OF TEST: 08/17/2018 06:16:51

Diagnostics Check		OK	06:08
Air Blank		0.000	06:09
Control Test		0.081	06:09
Air Blank		0.000	06:09
Subject Sample	# 1	0.174	06:12
Breath Volume		1.378	06:12
Air Blank		0.000	06:12
Air Blank		0.000	06:14
Subject Sample	# 2	0.184	06:15
Breath Volume		1.152	06:15
Air Blank		0.000	06:15
Control Test		0.079	06:16
Air Blank		0.000	06:16
Diagnostics Check		OK	06:16

Data Download Date/Time: 08/28/2018 12:17

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KENNEDY,BRIAN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/23/2018 03:18:57

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.082	03:13
Air Blank		0.000	03:13
Subject Sample	# 1	0.029	03:14
Breath Volume		1.468	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:16
Subject Sample	# 2	0.026	03:17
Breath Volume		1.289	03:17
Air Blank		0.000	03:18
Control Test		0.081	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 08/28/2018 12:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FERRIS,JEFFREY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: BAPTISTE

DATE/TIME OF TEST: 08/25/2018 03:33:39

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.081	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.348	03:29
Breath Volume		1.121	03:29
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.343	03:32
Breath Volume		1.652	03:32
Air Blank		0.000	03:32
Control Test		0.080	03:33
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 08/28/2018 12:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMEY,MICHAEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: LOPEZ QUINONES

DATE/TIME OF TEST: 08/28/2018 05:31:14

Diagnostics Check		OK	05:25
Air Blank		0.000	05:25
Control Test		0.081	05:25
Air Blank		0.000	05:26
Subject Sample	# 1	0.146	05:26
Breath Volume		2.273	05:26
Air Blank		0.000	05:27
Air Blank		0.000	05:29
Subject Sample	# 2	0.131	05:29
Breath Volume		1.441	05:29
Air Blank		0.000	05:30
Control Test		0.081	05:30
Air Blank		0.000	05:31
Diagnostics Check		OK	05:31

Data Download Date/Time: 08/28/2018 12:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KREKEL,ALEC
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: LEIGH JR

DATE/TIME OF TEST: 08/30/2018 02:32:55

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.081	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.177	02:28
Breath Volume		1.554	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.182	02:31
Breath Volume		2.519	02:31
Air Blank		0.000	02:32
Control Test		0.081	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 09/06/2018 13:13

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBBIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: FORTSON

DATE/TIME OF TEST: 08/30/2018 04:14:53

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.080	04:05
Air Blank		0.000	04:06
Subject Sample	# 1	0.161	04:07
Breath Volume		1.621	04:07
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.182	04:10
Breath Volume		2.636	04:10
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 3	0.188	04:13
Breath Volume		1.890	04:13
Air Blank		0.000	04:13
Control Test		0.080	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 09/06/2018 13:13

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: LABOY-CABAN

DATE/TIME OF TEST: 08/31/2018 03:34:35

Diagnostics Check		OK	03:27
Air Blank		0.000	03:28
Control Test		0.081	03:28
Air Blank		0.000	03:29
Subject Sample	# 1	0.110	03:30
Breath Volume		2.464	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.109	03:33
Breath Volume		2.820	03:33
Air Blank		0.000	03:33
Control Test		0.080	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 09/06/2018 13:13

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ROE,B
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 08/11/2018 16:39:47

Diagnostics Check		OK	16:33
Air Blank		0.000	16:34
Control Test		0.081	16:34
Air Blank		0.000	16:34
Subject Sample	# 1	0.267	16:35
Breath Volume		1.316	16:35
Air Blank		0.000	16:36
Air Blank		0.000	16:37
Subject Sample	# 2	0.263	16:38
Breath Volume		1.242	16:38
Air Blank		0.000	16:38
Control Test		0.080	16:39
Air Blank		0.000	16:39
Diagnostics Check		OK	16:39

Data Download Date/Time: 08/25/2018 15:29

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 08/15/2018 22:28:58

Diagnostics Check		OK	22:20
Air Blank		0.000	22:20
Control Test		0.080	22:20
Air Blank		0.000	22:21
Subject Sample	# 1	0.272	22:23
Breath Volume		2.324	22:23
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.269	22:27
Breath Volume		3.875	22:27
Air Blank		0.000	22:28
Control Test		0.080	22:28
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 08/25/2018 15:29

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CRUZ,C
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 07/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 08/17/2018 06:07:32

Diagnostics Check		OK	06:01
Air Blank		0.000	06:01
Control Test		0.080	06:01
Air Blank		0.000	06:02
Subject Sample	# 1	0.124	06:03
Breath Volume		2.429	06:03
Air Blank		0.000	06:03
Air Blank		0.000	06:05
Subject Sample	# 2	0.119	06:06
Breath Volume		1.585	06:06
Air Blank		0.000	06:06
Control Test		0.079	06:06
Air Blank		0.000	06:07
Diagnostics Check		OK	06:07

Data Download Date/Time: 08/25/2018 15:29

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAHMSEN,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/25/2018 22:59:28

Diagnostics Check		OK	22:53
Air Blank		0.000	22:53
Control Test		0.081	22:54
Air Blank		0.000	22:54
Subject Sample	# 1	0.228	22:55
Breath Volume		1.238	22:55
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.230	22:58
Breath Volume		2.621	22:58
Air Blank		0.000	22:58
Control Test		0.079	22:58
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 09/26/2018 03:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CHECO,E
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 08/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: SCHUMACHER

DATE/TIME OF TEST: 08/26/2018 14:41:15

Diagnostics Check		OK	14:35	
Air Blank		0.000	14:36	
Control Test		0.080	14:36	
Air Blank		0.000	14:36	
Subject Sample	# 1	0.326	14:37	
Breath Volume		2.050	14:37	
Air Blank		0.000	14:37	
Air Blank		0.000	14:39	
Subject Sample	# 2	0.326	14:40	
Breath Volume		1.187	14:40	
Air Blank		0.000	14:40	
Control Test		0.080	14:40	
Air Blank			14:41	RFI Detect
Air Blank			14:41	RFI Detect

Data Download Date/Time: 09/26/2018 03:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CHECO,E
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 08/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: SCHUMACHER

DATE/TIME OF TEST: 08/26/2018 14:49:23

Diagnostics Check		OK	14:43
Air Blank		0.000	14:43
Control Test		0.079	14:44
Air Blank		0.000	14:44
Subject Sample	# 1	0.325	14:45
Breath Volume		1.601	14:45
Air Blank		0.000	14:45
Air Blank		0.000	14:47
Subject Sample	# 2	0.315	14:47
Breath Volume		1.308	14:47
Air Blank		0.000	14:48
Control Test		0.079	14:48
Air Blank		0.000	14:49
Diagnostics Check		OK	14:49

Data Download Date/Time: 09/26/2018 03:32

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIESLINSKI,JOSEPH,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SHIPPY,DYLAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: MESSER

DATE/TIME OF TEST: 08/24/2018 01:03:11

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.078	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.183	00:55
Breath Volume		3.152	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.161	00:58
Breath Volume		1.687	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 3	0.181	01:01
Breath Volume		4.062	01:01
Air Blank		0.000	01:02
Control Test		0.078	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 09/05/2018 09:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ROSS,CHEYANNE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: WILDS

DATE/TIME OF TEST: 08/25/2018 01:21:58

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.078	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.187	01:17
Breath Volume		1.679	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.180	01:20
Breath Volume		1.761	01:20
Air Blank		0.000	01:21
Control Test		0.077	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 09/05/2018 09:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILLAK,MICHAEL,O
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FILLAK,MICHAEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/26/2018 00:24:32

Diagnostics Check		OK	00:15
Air Blank		0.000	00:16
Control Test		0.078	00:16
Air Blank		0.000	00:17
Subject Sample	# 1	0.208	00:18
Breath Volume		3.125	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.201	00:23
Breath Volume		1.582	00:23
Air Blank		0.000	00:23
Control Test		0.078	00:23
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 09/05/2018 09:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GRAY,DANNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: GECIONIS

DATE/TIME OF TEST: 08/26/2018 04:32:25

Diagnostics Check		OK	04:29	
Air Blank		0.000	04:29	
Control Test		0.077	04:30	
Air Blank		0.000	04:30	
Subject Sample	# 1		04:31	Subject Test Refused
Breath Volume		0.000	04:31	
Air Blank		0.000	04:31	
Control Test		0.078	04:31	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:32	

Data Download Date/Time: 09/05/2018 09:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIESLINSKI,JOSEPH,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SHIPPY,DYLAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 08/28/2018 01:31:28

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.078	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.000	01:27
Breath Volume		2.773	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.000	01:30
Breath Volume		2.964	01:30
Air Blank		0.000	01:30
Control Test		0.078	01:30
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 09/05/2018 09:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: CASSIDY,ASHTON
ARRESTING OFFICER AGENCY: Palatka PD
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: \LESTER

DATE/TIME OF TEST: 08/16/2018 05:44:59

Diagnostics Check		OK	05:33	
Air Blank		0.000	05:34	
Control Test		0.078	05:34	
Air Blank		0.000	05:34	
Subject Sample	# 1		05:38	No Sample Provided
Breath Volume		0.000	05:38	
Air Blank		0.000	05:38	
Air Blank		0.000	05:40	
Subject Sample	# 2		05:43	No Sample Provided
Breath Volume		0.000	05:43	
Air Blank		0.000	05:44	
Control Test		0.080	05:44	
Air Blank		0.000	05:44	
Diagnostics Check		OK	05:44	

Data Download Date/Time: 09/04/2018 08:58

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRING,JAMES,E
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SENA,JOSEPH
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A2
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 08/11/2018 00:19:04

Diagnostics Check		OK	00:12
Air Blank		0.000	00:13
Control Test		0.079	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.178	00:14
Breath Volume		1.195	00:14
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.181	00:17
Breath Volume		3.136	00:17
Air Blank		0.000	00:18
Control Test		0.078	00:18
Air Blank		0.000	00:18
Diagnostics Check		OK	00:19

Data Download Date/Time: 09/03/2018 15:48

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JOSE,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LOPEZ,JOSE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A2
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: MCCARTNEY

DATE/TIME OF TEST: 08/24/2018 20:27:12

Diagnostics Check		OK	20:20
Air Blank		0.000	20:21
Control Test		0.078	20:21
Air Blank		0.000	20:22
Subject Sample	# 1	0.089	20:22
Breath Volume		2.261	20:22
Air Blank		0.000	20:23
Air Blank		0.000	20:25
Subject Sample	# 2	0.091	20:25
Breath Volume		3.539	20:25
Air Blank		0.000	20:26
Control Test		0.078	20:26
Air Blank		0.000	20:27
Diagnostics Check		OK	20:27

Data Download Date/Time: 09/03/2018 15:48

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JOSE,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LOPEZ,JOSE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A2
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: PEREZ VASQUEZ

DATE/TIME OF TEST: 08/25/2018 22:05:54

Diagnostics Check		OK	21:59
Air Blank		0.000	22:00
Control Test		0.079	22:00
Air Blank		0.000	22:01
Subject Sample	# 1	0.146	22:01
Breath Volume		1.937	22:01
Air Blank		0.000	22:02
Air Blank		0.000	22:04
Subject Sample	# 2	0.145	22:04
Breath Volume		1.914	22:04
Air Blank		0.000	22:04
Control Test		0.078	22:05
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 09/03/2018 15:48

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JOSE,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LOPEZ,JOSE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A2
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: CARMONA

DATE/TIME OF TEST: 08/26/2018 21:58:26

Diagnostics Check		OK	21:55	
Air Blank		0.000	21:56	
Control Test		0.078	21:56	
Air Blank		0.000	21:56	
Subject Sample	# 1		21:57	Subject Test Refused
Breath Volume		0.000	21:57	
Air Blank		0.000	21:57	
Control Test		0.079	21:57	
Air Blank		0.000	21:58	
Diagnostics Check		OK	21:58	

Data Download Date/Time: 09/03/2018 15:48

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: FRES,LUIS
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:36

SUBJECT'S LAST NAME: BULLINGER

DATE/TIME OF TEST: 08/14/2018 20:08:41

Diagnostics Check		OK	20:02
Air Blank		0.000	20:03
Control Test		0.078	20:03
Air Blank		0.000	20:03
Subject Sample	# 1	0.174	20:04
Breath Volume		2.019	20:04
Air Blank		0.000	20:04
Air Blank		0.000	20:06
Subject Sample	# 2	0.169	20:07
Breath Volume		1.375	20:07
Air Blank		0.000	20:07
Control Test		0.078	20:08
Air Blank		0.000	20:08
Diagnostics Check		OK	20:08

Data Download Date/Time: 08/21/2018 09:15

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: GRENON

DATE/TIME OF TEST: 08/25/2018 03:51:02

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.079	03:41	
Air Blank		0.000	03:41	
Subject Sample	# 1		03:45	No Sample Provided
Breath Volume		0.000	03:45	
Air Blank		0.000	03:45	
Air Blank		0.000	03:47	
Subject Sample	# 2		03:49	Subject Test Refused
Breath Volume		0.000	03:49	
Air Blank		0.000	03:50	
Control Test		0.079	03:50	
Air Blank		0.000	03:50	
Diagnostics Check		OK	03:51	

Data Download Date/Time: 09/10/2018 10:02

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PALOMINO,JULIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 08/11/2018 05:47:33

Diagnostics Check		OK	05:38
Air Blank		0.000	05:38
Control Test		0.081	05:39
Air Blank		0.000	05:39
Subject Sample	# 1	0.197	05:42
Breath Volume		2.027	05:42
Air Blank		0.000	05:43
Air Blank		0.000	05:45
Subject Sample	# 2	0.202	05:45
Breath Volume		2.503	05:45
Air Blank		0.000	05:46
Control Test		0.078	05:46
Air Blank		0.000	05:47
Diagnostics Check		OK	05:47

Data Download Date/Time: 08/29/2018 20:11

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: BORRERO

DATE/TIME OF TEST: 08/19/2018 05:36:18

Diagnostics Check		OK	05:29
Air Blank		0.000	05:30
Control Test		0.082	05:30
Air Blank		0.000	05:30
Subject Sample	# 1	0.118	05:31
Breath Volume		2.312	05:31
Air Blank		0.000	05:32
Air Blank		0.000	05:34
Subject Sample	# 2	0.115	05:34
Breath Volume		2.457	05:34
Air Blank		0.000	05:35
Control Test		0.080	05:35
Air Blank		0.000	05:36
Diagnostics Check		OK	05:36

Data Download Date/Time: 08/29/2018 20:11

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/20/2018 04:03:05

Diagnostics Check		OK	03:51	
Air Blank		0.000	03:51	
Control Test		0.079	03:52	
Air Blank		0.000	03:52	
Subject Sample	# 1	0.135	03:55	Volume Not Met
Breath Volume		0.332	03:55	
Air Blank		0.000	03:56	
Air Blank		0.000	03:58	
Subject Sample	# 2	0.140	04:01	Volume Not Met
Breath Volume		0.132	04:01	
Air Blank		0.000	04:02	
Control Test		0.078	04:02	
Air Blank		0.000	04:03	
Diagnostics Check		OK	04:03	

Data Download Date/Time: 08/29/2018 20:11

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/20/2018 04:14:07

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.078	04:05
Air Blank		0.000	04:06
Subject Sample	# 1	0.155	04:08
Breath Volume		1.265	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:10
Subject Sample	# 2	0.135	04:12
Breath Volume		1.261	04:12
Air Blank		0.000	04:13
Control Test		0.078	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 08/29/2018 20:11

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,ANDRE,D
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MATHIS,ANDRE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 08/29/2018 00:17:22

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.082	00:12
Air Blank		0.000	00:12
Subject Sample	# 1	0.171	00:12
Breath Volume		1.953	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.164	00:15
Breath Volume		2.070	00:15
Air Blank		0.000	00:16
Control Test		0.080	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 08/29/2018 20:11

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TIMOTHY,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WHITE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: ORBE

DATE/TIME OF TEST: 08/12/2018 20:13:45

Diagnostics Check		OK	20:10	
Air Blank		0.000	20:11	
Control Test		0.081	20:11	
Air Blank		0.000	20:12	
Subject Sample	# 1		20:12	RFI Detect
Breath Volume		0.000	20:12	
Air Blank		0.000	20:12	
Control Test		0.081	20:13	
Air Blank		0.000	20:13	
Diagnostics Check		OK	20:13	

Data Download Date/Time: 08/14/2018 14:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TIMOTHY,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WHITE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: ORBE

DATE/TIME OF TEST: 08/12/2018 20:23:12

Diagnostics Check		OK	20:15
Air Blank		0.000	20:15
Control Test		0.081	20:16
Air Blank		0.000	20:16
Subject Sample	# 1	0.243	20:18
Breath Volume		2.953	20:18
Air Blank		0.000	20:19
Air Blank		0.000	20:21
Subject Sample	# 2	0.241	20:21
Breath Volume		2.828	20:21
Air Blank		0.000	20:22
Control Test		0.079	20:22
Air Blank		0.000	20:23
Diagnostics Check		OK	20:23

Data Download Date/Time: 08/14/2018 14:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ALAMEDA

DATE/TIME OF TEST: 08/16/2018 03:52:22

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.081	03:46
Air Blank		0.000	03:47
Subject Sample	# 1	0.162	03:47
Breath Volume		2.914	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.164	03:50
Breath Volume		2.804	03:50
Air Blank		0.000	03:51
Control Test		0.082	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 09/24/2018 14:06

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPROUSE,JOHN,S
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SPROUSE,JOHN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MCCLURE

DATE/TIME OF TEST: 08/22/2018 03:02:08

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.081	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.205	02:57
Breath Volume		4.273	02:57
Air Blank		0.000	02:58
Air Blank		0.000	02:59
Subject Sample	# 2	0.204	03:00
Breath Volume		4.375	03:00
Air Blank		0.000	03:01
Control Test		0.077	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 09/24/2018 14:06

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPROUSE,JOHN,S
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SPROUSE,JOHN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ZHAO

DATE/TIME OF TEST: 08/31/2018 03:17:21

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.081	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.134	03:12
Breath Volume		2.058	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.133	03:15
Breath Volume		1.765	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 09/24/2018 14:06

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: PUPO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 08/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ARROLIGA

DATE/TIME OF TEST: 08/16/2018 02:13:52

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.081	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.211	02:09
Breath Volume		1.277	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.215	02:12
Breath Volume		1.929	02:12
Air Blank		0.000	02:12
Control Test		0.082	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 09/20/2018 11:07

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

INSTRUMENT REGISTERED TO: MICCOSUKEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MISTY,L
BREATH TEST OPERATOR AGENCY: Miccosukee PD
ARRESTING OFFICER: GRECO,STEVEN
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 08/09/2018 03:08:11

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:05	
Control Test		0.077	03:06	
Air Blank		0.000	03:06	
Subject Sample	# 1		03:06	Subject Test Refused
Breath Volume		0.000	03:06	
Air Blank		0.000	03:07	
Control Test		0.077	03:07	
Air Blank		0.000	03:08	
Diagnostics Check		OK	03:08	

Data Download Date/Time: 08/31/2018 12:40

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

INSTRUMENT REGISTERED TO: MICCOSUKEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBO,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Miccosukee PD
ARRESTING OFFICER: ALBO,JOSHUA
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 07/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: FAJARDO

DATE/TIME OF TEST: 08/11/2018 01:35:44

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:20	
Control Test		0.077	01:21	
Air Blank		0.000	01:21	
Subject Sample	# 1	0.147	01:22	
Breath Volume		2.152	01:22	
Air Blank		0.000	01:23	
Air Blank		0.000	01:25	
Subject Sample	# 2	0.109	01:28	Volume Not Met
Breath Volume		0.683	01:28	
Air Blank		0.000	01:29	
Air Blank		0.000	01:30	
Subject Sample	# 3	0.103	01:34	Volume Not Met
Breath Volume		0.562	01:34	
Air Blank		0.000	01:34	
Control Test		0.075	01:35	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 08/31/2018 12:40

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

INSTRUMENT REGISTERED TO: MICCOSUKEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MISTY,L
BREATH TEST OPERATOR AGENCY: Miccosukee PD
ARRESTING OFFICER: DOMINGUEZ,HORACIO
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: ANAYA SANCHEZ

DATE/TIME OF TEST: 08/31/2018 19:50:58

Diagnostics Check		OK	19:44
Air Blank		0.000	19:44
Control Test		0.078	19:44
Air Blank		0.000	19:45
Subject Sample	# 1	0.123	19:45
Breath Volume		2.167	19:45
Air Blank		0.000	19:46
Air Blank		0.000	19:48
Subject Sample	# 2	0.116	19:49
Breath Volume		1.585	19:49
Air Blank		0.000	19:50
Control Test		0.077	19:50
Air Blank		0.000	19:50
Diagnostics Check		OK	19:50

Data Download Date/Time: 09/13/2018 08:15

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001714

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEWART,DILLON,J
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: STEWART,DILLON
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: NEVILLE

DATE/TIME OF TEST: 08/13/2018 01:54:47

Diagnostics Check		OK	01:47
Air Blank		0.000	01:48
Control Test		0.076	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.218	01:50
Breath Volume		2.824	01:50
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.215	01:53
Breath Volume		3.105	01:53
Air Blank		0.000	01:53
Control Test		0.075	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 08/31/2018 15:03

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001714

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEFFIELD,KHRIS,D
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: ZARATE,JESSICA
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ANDREASEN

DATE/TIME OF TEST: 08/13/2018 02:41:09

Diagnostics Check		OK	02:34
Air Blank		0.000	02:35
Control Test		0.078	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.021	02:36
Breath Volume		3.468	02:36
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.019	02:39
Breath Volume		3.273	02:39
Air Blank		0.000	02:40
Control Test		0.078	02:40
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 08/31/2018 15:03

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001714

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKHAM,MARCUS,W
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: BAIN,KENNETH
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: SEARS

DATE/TIME OF TEST: 08/16/2018 03:03:51

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.077	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.166	02:59
Breath Volume		2.906	02:59
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.167	03:02
Breath Volume		2.507	03:02
Air Blank		0.000	03:02
Control Test		0.077	03:03
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 08/31/2018 15:03

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: ABSHIRE

DATE/TIME OF TEST: 08/02/2018 22:04:44

Diagnostics Check	Fail	22:04	Temperature Regulation Test
Air Blank	0.000	22:04	

Data Download Date/Time: 08/08/2018 19:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: ABSHIRE

DATE/TIME OF TEST: 08/02/2018 22:07:17

Diagnostics Check	Fail	22:06	Temperature Regulation Test
Air Blank	0.000	22:07	

Data Download Date/Time: 08/08/2018 19:12

COBRA ©1997-2014 CMI, Inc.