Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BARDY

DATE/TIME OF TEST: 08/11/2018 00:06:05

Diagnostics Check		OK	23:51	
Air Blank		0.000	23:51	
Control Test		0.078	23:52	
Air Blank		0.000	23:52	
Subject Sample	# 3	0.219	23:55	
Breath Volume		1.804	23:55	
Air Blank		0.000	23:56	
Air Blank		0.000	23:57	
Subject Sample	# 1		00:01	No Sample Provided
Breath Volume		0.000	00:01	
Air Blank		0.000	00:01	
Air Blank		0.000	00:03	
Subject Sample	# 2	0.218	00:04	
Breath Volume		1.335	00:04	
Air Blank		0.000	00:05	
Control Test		0.079	00:05	
Air Blank		0.000	00:05	
Diagnostics Check		OK	00:06	

Data Download Date/Time: 08/28/2018 13:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 08/11/2018 03:26:58

Diagnostics Check		OK	03:19	
Air Blank		0.000	03:19	
Control Test		0.079	03:19	
Air Blank		0.000	03:20	
Subject Sample	# 1	0.106	03:21	
Breath Volume		1.582	03:21	
Air Blank		0.000	03:22	
Air Blank		0.000	03:24	
Subject Sample	# 2	0.101	03:25	
Breath Volume		1.398	03:25	
Air Blank		0.000	03:26	
Control Test		0.078	03:26	RFI Detect
Air Blank		0.000	03:26	

Data Download Date/Time: 08/28/2018 13:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 08/11/2018 20:13:11

Diagnostics Check		OK	20:07
Air Blank		0.000	20:07
Control Test		0.078	20:07
Air Blank		0.000	20:08
Subject Sample	# 1	0.000	20:08
Breath Volume		2.355	20:08
Air Blank		0.000	20:09
Air Blank		0.000	20:11
Subject Sample	# 2	0.000	20:11
Breath Volume		1.984	20:11
Air Blank		0.000	20:12
Control Test		0.079	20:12
Air Blank		0.000	20:13
Diagnostics Check		OK	20:13

Data Download Date/Time: 08/28/2018 13:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: CRAWLEY

DATE/TIME OF TEST: 08/16/2018 20:43:35

Diagnostics Check		OK	20:36
Air Blank		0.000	20:37
Control Test		0.078	20:37
Air Blank		0.000	20:38
Subject Sample	# 1	0.276	20:38
Breath Volume		2.003	20:38
Air Blank		0.000	20:39
Air Blank		0.000	20:41
Subject Sample	# 2	0.291	20:41
Breath Volume		1.648	20:41
Air Blank		0.000	20:42
Control Test		0.078	20:42
Air Blank		0.000	20:43
Diagnostics Check		OK	20:43

Data Download Date/Time: 08/28/2018 13:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: FARR

DATE/TIME OF TEST: 08/19/2018 02:17:47

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.082	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.030	02:13
Breath Volume		1.644	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.031	02:16
Breath Volume		1.703	02:16
Air Blank		0.000	02:16
Control Test		0.079	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 08/28/2018 13:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 07/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GRULLON

DATE/TIME OF TEST: 08/25/2018 03:21:45

Diagnostics Check		OK	03:15	
Air Blank		0.000	03:15	
Control Test		0.078	03:16	
Air Blank		0.000	03:16	
Subject Sample	# 1	0.109	03:17	
Breath Volume		2.464	03:17	
Air Blank		0.000	03:17	
Air Blank		0.000	03:19	
Subject Sample	# 2	0.106	03:20	
Breath Volume		1.363	03:20	
Air Blank			03:21	RFI Detect
Air Blank		0.000	03:21	

Data Download Date/Time: 08/28/2018 13:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: DEWILD

DATE/TIME OF TEST: 08/06/2018 01:35:05

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.081	01:28
Air Blank		0.000	01:29
Subject Sample	# 1	0.206	01:30
Breath Volume		3.156	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:32
Subject Sample	# 2	0.198	01:33
Breath Volume		1.855	01:33
Air Blank		0.000	01:34
Control Test		0.080	01:34
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 08/21/2018 11:53

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MARCANTONIO

DATE/TIME OF TEST: 08/08/2018 01:23:36

Diagnostics Check		OK	01:16
Air Blank		0.000	01:16
Control Test		0.081	01:17
Air Blank		0.000	01:17
Subject Sample	# 1	0.030	01:18
Breath Volume		3.449	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.029	01:22
Breath Volume		2.605	01:22
Air Blank		0.000	01:22
Control Test		0.080	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 08/21/2018 11:53

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFC,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BIANCA,OFC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: BEDNARSKI

DATE/TIME OF TEST: 08/09/2018 22:27:41

Diagnostics Check		OK	22:21
Air Blank		0.000	22:21
Control Test		0.081	22:21
Air Blank		0.000	22:22
Subject Sample	# 1	0.388	22:22
Breath Volume		2.699	22:22
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.378	22:25
Breath Volume		2.320	22:25
Air Blank		0.000	22:26
Control Test		0.079	22:27
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 08/21/2018 11:53

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: SUGGS

DATE/TIME OF TEST: 08/12/2018 06:43:02

Diagnostics Check		OK	06:35
Air Blank		0.000	06:35
Control Test		0.080	06:36
Air Blank		0.000	06:36
Subject Sample	# 1	0.080	06:38
Breath Volume		2.265	06:38
Air Blank		0.000	06:38
Air Blank		0.000	06:40
Subject Sample	# 2	0.079	06:41
Breath Volume		1.613	06:41
Air Blank		0.000	06:42
Control Test		0.080	06:42
Air Blank		0.000	06:42
Diagnostics Check		OK	06:43

Data Download Date/Time: 08/21/2018 11:53

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/20/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: KENDLE

DATE/TIME OF TEST: 08/13/2018 22:03:14

Diagnostics Check		OK	21:55
Air Blank		0.000	21:56
Control Test		0.080	21:56
Air Blank		0.000	21:57
Subject Sample	# 1	0.000	21:58
Breath Volume		1.648	21:58
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.000	22:01
Breath Volume		1.824	22:01
Air Blank		0.000	22:02
Control Test		0.081	22:02
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 08/21/2018 11:53

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OBRIEN,DANIEL,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: SUTTON,MARY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 08/16/2018 23:02:22

Diagnostics Check		OK	22:54
Air Blank		0.000	22:54
Control Test		0.080	22:54
Air Blank		0.000	22:55
Subject Sample	# 1	0.197	22:57
Breath Volume		2.382	22:57
Air Blank		0.000	22:58
Air Blank		0.000	22:59
Subject Sample	# 2	0.205	23:00
Breath Volume		2.042	23:00
Air Blank		0.000	23:01
Control Test		0.079	23:01
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 08/21/2018 11:53

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BENWAY

DATE/TIME OF TEST: 08/22/2018 00:27:51

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.079	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.142	00:22
Breath Volume		2.562	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.125	00:26
Breath Volume		1.558	00:26
Air Blank		0.000	00:26
Control Test		0.079	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 09/27/2018 20:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: KLOTZ

DATE/TIME OF TEST: 08/23/2018 03:05:38

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:48	
Control Test		0.078	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 1	0.062	02:52	Volume Not Met
Breath Volume		0.574	02:52	
Air Blank		0.000	02:53	
Air Blank		0.000	02:55	
Subject Sample	# 2	0.072	02:58	
Breath Volume		2.156	02:58	
Air Blank		0.000	02:59	
Air Blank		0.000	03:00	
Subject Sample	# 3	0.038	03:04	Volume Not Met
Breath Volume		0.601	03:04	
Air Blank		0.000	03:04	
Control Test		0.077	03:05	
Air Blank		0.000	03:05	
Diagnostics Check		OK	03:05	

Data Download Date/Time: 09/27/2018 20:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 08/23/2018 03:31:28

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.079	03:24
Air Blank		0.000	03:24
Subject Sample	# 1	0.148	03:26
Breath Volume		2.953	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.159	03:29
Breath Volume		2.351	03:29
Air Blank		0.000	03:30
Control Test		0.078	03:30
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 09/27/2018 20:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 08/23/2018 22:56:16

Diagnostics Check		OK	22:46	
Air Blank		0.000	22:47	
Control Test		0.079	22:47	
Air Blank		0.000	22:48	
Subject Sample	# 1	0.195	22:50	
Breath Volume		1.316	22:50	
Air Blank		0.000	22:50	
Air Blank		0.000	22:52	
Subject Sample	# 2	0.203	22:55	
Breath Volume		1.351	22:55	
Air Blank			22:55	RFI Detect
Air Blank		0.000	22:56	

Data Download Date/Time: 09/27/2018 20:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 08/23/2018 23:09:05

Diagnostics Check		OK	23:00
Air Blank		0.000	23:00
Control Test		0.077	23:00
Air Blank		0.000	23:01
Subject Sample	# 1	0.199	23:03
Breath Volume		1.277	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:05
Subject Sample	# 2	0.196	23:07
Breath Volume		1.515	23:07
Air Blank		0.000	23:08
Control Test		0.078	23:08
Air Blank		0.000	23:08
Diagnostics Check		OK	23:09

Data Download Date/Time: 09/27/2018 20:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: ASHLEY

DATE/TIME OF TEST: 08/26/2018 02:51:07

Diagnostics Check		OK	02:43
Air Blank		0.000	02:44
Control Test		0.079	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.198	02:46
Breath Volume		2.511	02:46
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.195	02:49
Breath Volume		2.144	02:49
Air Blank		0.000	02:50
Control Test		0.078	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 09/27/2018 20:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: CHRISTENSEN

DATE/TIME OF TEST: 08/07/2018 02:46:17

Diagnostics Check		OK	02:43	
Air Blank		0.000	02:43	
Control Test		0.079	02:43	
Air Blank		0.000	02:44	
Subject Sample	# 1		02:44	RFI Detect
Breath Volume		0.000	02:44	
Air Blank		0.000	02:45	
Control Test		0.079	02:45	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:46	

Data Download Date/Time: 08/21/2018 10:02

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: CHRISTENSEN

DATE/TIME OF TEST: 08/07/2018 02:50:54

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:48	
Control Test		0.079	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 1		02:49	Subject Test Refused
Breath Volume		0.000	02:49	
Air Blank		0.000	02:49	
Control Test		0.079	02:50	
Air Blank		0.000	02:50	
Diagnostics Check		OK	02:50	

Data Download Date/Time: 08/21/2018 10:02

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 08/07/2018 20:59:39

Diagnostics Check		OK	20:56	
Air Blank		0.000	20:57	
Control Test		0.078	20:57	
Air Blank		0.000	20:58	
Subject Sample	# 1		20:58	Subject Test Refused
Breath Volume		0.000	20:58	
Air Blank		0.000	20:58	
Control Test		0.078	20:59	
Air Blank		0.000	20:59	
Diagnostics Check		OK	20:59	

Data Download Date/Time: 08/21/2018 10:02

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: TERRELL

DATE/TIME OF TEST: 08/08/2018 19:25:23

Diagnostics Check		OK	19:21	
Air Blank		0.000	19:21	
Control Test		0.079	19:22	
Air Blank		0.000	19:22	
Subject Sample	# 1		19:24	Subject Test Refused
Breath Volume		0.000	19:24	
Air Blank		0.000	19:24	
Control Test		0.079	19:24	
Air Blank		0.000	19:25	
Diagnostics Check		OK	19:25	

Data Download Date/Time: 08/21/2018 10:02

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 08/10/2018 17:16:34

Diagnostics Check	OK	17:16	
Air Blank		17:16	RFI Detect
Air Blank		17:16	RFI Detect

Data Download Date/Time: 08/21/2018 10:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 08/10/2018 17:22:24

Diagnostics Check		OK	17:19	
Air Blank		0.000	17:20	
Control Test		0.079	17:20	
Air Blank		0.000	17:20	
Subject Sample	# 1		17:21	Subject Test Refused
Breath Volume		0.000	17:21	
Air Blank		0.000	17:21	
Control Test		0.079	17:21	
Air Blank		0.000	17:22	
Diagnostics Check		OK	17:22	

Data Download Date/Time: 08/21/2018 10:02

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: RHODES

DATE/TIME OF TEST: 08/11/2018 21:45:46

Diagnostics Check		OK	21:39
Air Blank		0.000	21:39
Control Test		0.079	21:40
Air Blank		0.000	21:40
Subject Sample	# 1	0.021	21:41
Breath Volume		1.976	21:41
Air Blank		0.000	21:41
Air Blank		0.000	21:43
Subject Sample	# 2	0.019	21:44
Breath Volume		3.343	21:44
Air Blank		0.000	21:44
Control Test		0.077	21:45
Air Blank		0.000	21:45
Diagnostics Check		OK	21:45

Data Download Date/Time: 08/21/2018 10:02

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 15:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 08/14/2018 16:39:12

Diagnostics Check		OK	16:32
Air Blank		0.000	16:33
Control Test		0.079	16:33
Air Blank		0.000	16:34
Subject Sample	# 1	0.000	16:34
Breath Volume		3.109	16:34
Air Blank		0.000	16:35
Air Blank		0.000	16:37
Subject Sample	# 2	0.000	16:37
Breath Volume		3.121	16:37
Air Blank		0.000	16:38
Control Test		0.077	16:38
Air Blank		0.000	16:39
Diagnostics Check		OK	16:39

Data Download Date/Time: 08/21/2018 10:02

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 08/20/2018 22:58:05

Diagnostics Check		OK	22:52
Air Blank		0.000	22:52
Control Test		0.079	22:52
Air Blank		0.000	22:53
Subject Sample	# 1	0.000	22:53
Breath Volume		2.445	22:53
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.000	22:56
Breath Volume		2.515	22:56
Air Blank		0.000	22:57
Control Test		0.077	22:57
Air Blank		0.000	22:57
Diagnostics Check		OK	22:58

Data Download Date/Time: 08/21/2018 10:02

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: BLANTON

DATE/TIME OF TEST: 08/21/2018 02:33:11

Diagnostics Check	OK	02:32	
Air Blank		02:33	RFI Detect
Air Blank		02:33	RFI Detect

Data Download Date/Time: 08/21/2018 10:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: PUICHI

DATE/TIME OF TEST: 08/25/2018 03:02:15

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:50	
Control Test		0.079	02:51	
Air Blank		0.000	02:51	
Subject Sample	# 1	0.243	02:54	Volume Not Met
Breath Volume		0.410	02:54	
Air Blank		0.000	02:55	
Air Blank		0.000	02:57	
Subject Sample	# 2	0.206	03:00	Volume Not Met
Breath Volume		0.085	03:00	
Air Blank		0.000	03:01	
Control Test		0.077	03:01	
Air Blank		0.000	03:02	
Diagnostics Check		OK	03:02	

Data Download Date/Time: 09/25/2018 09:21

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 16:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: MERRICK

DATE/TIME OF TEST: 08/26/2018 17:12:12

Diagnostics Check		OK	17:09	
Air Blank		0.000	17:09	
Control Test		0.079	17:10	
Air Blank		0.000	17:10	
Subject Sample	# 1		17:10	Subject Test Refused
Breath Volume		0.000	17:10	
Air Blank		0.000	17:11	
Control Test		0.080	17:11	
Air Blank		0.000	17:12	
Diagnostics Check		OK	17:12	

Data Download Date/Time: 09/25/2018 09:21

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: KOVALSKI

DATE/TIME OF TEST: 08/27/2018 16:43:27

Diagnostics Check	OK	16:43	
Air Blank		16:43	RFI Detect
Air Blank		16:43	RFI Detect

Data Download Date/Time: 09/25/2018 09:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: KOVALSKI

DATE/TIME OF TEST: 08/27/2018 16:48:39

Diagnostics Check		OK	16:45	
Air Blank		0.000	16:46	
Control Test		0.079	16:46	
Air Blank		0.000	16:46	
Subject Sample	# 1		16:47	Subject Test Refused
Breath Volume		0.000	16:47	
Air Blank		0.000	16:47	
Control Test		0.079	16:48	
Air Blank		0.000	16:48	
Diagnostics Check		OK	16:48	

Data Download Date/Time: 09/25/2018 09:21

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: SEGOBIANO-HERNANDEZ

DATE/TIME OF TEST: 08/31/2018 23:19:47

Diagnostics Check		OK	23:13
Air Blank		0.000	23:13
Control Test		0.079	23:13
Air Blank		0.000	23:14
Subject Sample	# 1	0.155	23:15
Breath Volume		1.882	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:17
Subject Sample	# 2	0.153	23:18
Breath Volume		2.296	23:18
Air Blank		0.000	23:18
Control Test		0.077	23:19
Air Blank		0.000	23:19
Diagnostics Check		OK	23:19

Data Download Date/Time: 09/25/2018 09:21

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LANE,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: SUKOVICH

DATE/TIME OF TEST: 08/01/2018 02:14:15

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.080	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.000	02:10
Breath Volume		2.781	02:10
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.000	02:12
Breath Volume		2.875	02:12
Air Blank		0.000	02:13
Control Test		0.081	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 08/17/2018 11:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,M
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 08/03/2018 02:28:59

Diagnostics Check		OK	02:24	
Air Blank		0.000	02:25	
Control Test		0.080	02:25	
Air Blank		0.000	02:26	
Subject Sample	# 1		02:27	Subject Test Refused
Breath Volume		0.000	02:27	
Air Blank		0.000	02:28	
Control Test		0.081	02:28	
Air Blank		0.000	02:28	
Diagnostics Check		OK	02:28	

Data Download Date/Time: 08/17/2018 11:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILBERGER,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WEBER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 08/03/2018 23:00:43

Diagnostics Check		OK	22:54
Air Blank		0.000	22:54
Control Test		0.081	22:54
Air Blank		0.000	22:55
Subject Sample	# 1	0.000	22:56
Breath Volume		3.304	22:56
Air Blank		0.000	22:56
Air Blank		0.000	22:58
Subject Sample	# 2	0.000	22:59
Breath Volume		2.984	22:59
Air Blank		0.000	22:59
Control Test		0.082	23:00
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 08/17/2018 11:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: STIBER

DATE/TIME OF TEST: 08/08/2018 04:35:37

Diagnostics Check		OK	04:31	
Air Blank		0.000	04:31	
Control Test		0.081	04:32	
Air Blank		0.000	04:32	
Subject Sample	# 1		04:34	Subject Test Refused
Breath Volume		0.000	04:34	
Air Blank		0.000	04:34	
Control Test		0.082	04:35	
Air Blank		0.000	04:35	
Diagnostics Check		OK	04:35	

Data Download Date/Time: 08/17/2018 11:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILBERGER,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: DORMADY

DATE/TIME OF TEST: 08/09/2018 21:09:46

Diagnostics Check		OK	21:03
Air Blank		0.000	21:03
Control Test		0.081	21:03
Air Blank		0.000	21:04
Subject Sample	# 1	0.133	21:05
Breath Volume		2.082	21:05
Air Blank		0.000	21:05
Air Blank		0.000	21:07
Subject Sample	# 2	0.125	21:08
Breath Volume		1.679	21:08
Air Blank		0.000	21:08
Control Test		0.080	21:09
Air Blank		0.000	21:09
Diagnostics Check		OK	21:09

Data Download Date/Time: 08/17/2018 11:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILBERGER,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TIRRI,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:23

SUBJECT'S LAST NAME: POLIZZI

DATE/TIME OF TEST: 08/10/2018 15:50:53

Diagnostics Check		OK	15:43
Air Blank		0.000	15:43
Control Test		0.081	15:44
Air Blank		0.000	15:44
Subject Sample	# 1	0.180	15:46
Breath Volume		2.011	15:46
Air Blank		0.000	15:46
Air Blank		0.000	15:48
Subject Sample	# 2	0.183	15:49
Breath Volume		1.207	15:49
Air Blank		0.000	15:49
Control Test		0.080	15:50
Air Blank		0.000	15:50
Diagnostics Check		OK	15:50

Data Download Date/Time: 08/17/2018 11:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RUSSELL,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: LIGHTFOOT

DATE/TIME OF TEST: 08/14/2018 03:14:54

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.080	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.127	03:10
Breath Volume		2.707	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.121	03:13
Breath Volume		2.917	03:13
Air Blank		0.000	03:13
Control Test		0.080	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 08/17/2018 11:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PLATT,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: PAWELA

DATE/TIME OF TEST: 08/14/2018 22:58:23

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:55	
Control Test		0.081	22:56	
Air Blank		0.000	22:56	
Subject Sample	# 1		22:57	Subject Test Refused
Breath Volume		0.000	22:57	
Air Blank		0.000	22:57	
Control Test		0.080	22:57	
Air Blank		0.000	22:58	
Diagnostics Check		OK	22:58	

Data Download Date/Time: 08/17/2018 11:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: EIMICKE

DATE/TIME OF TEST: 08/15/2018 01:47:54

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.080	01:42
Air Blank		0.000	01:43
Subject Sample	# 1	0.159	01:43
Breath Volume		1.593	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.158	01:46
Breath Volume		1.339	01:46
Air Blank		0.000	01:46
Control Test		0.081	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 08/17/2018 11:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: ERRICO

DATE/TIME OF TEST: 08/15/2018 02:29:27

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.081	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.000	02:25
Breath Volume		4.156	02:25
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.000	02:28
Breath Volume		3.371	02:28
Air Blank		0.000	02:28
Control Test		0.081	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 08/17/2018 11:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARPER,T
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: RAVELO

DATE/TIME OF TEST: 08/15/2018 05:21:08

Diagnostics Check		OK	05:15
Air Blank		0.000	05:15
Control Test		0.081	05:15
Air Blank		0.000	05:16
Subject Sample	# 1	0.213	05:16
Breath Volume		1.250	05:16
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 2	0.208	05:19
Breath Volume		1.617	05:19
Air Blank		0.000	05:20
Control Test		0.080	05:20
Air Blank		0.000	05:21
Diagnostics Check		OK	05:21

Data Download Date/Time: 08/17/2018 11:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,E,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HANSON,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: DICKSON

DATE/TIME OF TEST: 08/17/2018 19:52:30

Diagnostics Check		OK	19:46
Air Blank		0.000	19:46
Control Test		0.081	19:46
Air Blank		0.000	19:47
Subject Sample	# 1	0.261	19:47
Breath Volume		3.316	19:47
Air Blank		0.000	19:48
Air Blank		0.000	19:50
Subject Sample	# 2	0.263	19:50
Breath Volume		3.738	19:50
Air Blank		0.000	19:51
Control Test		0.081	19:51
Air Blank		0.000	19:52
Diagnostics Check		OK	19:52

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,E,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CROCKER,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: VOLL

DATE/TIME OF TEST: 08/17/2018 20:59:08

Diagnostics Check		OK	20:52
Air Blank		0.000	20:52
Control Test		0.081	20:52
Air Blank		0.000	20:53
Subject Sample	# 1	0.025	20:54
Breath Volume		1.328	20:54
Air Blank		0.000	20:54
Air Blank		0.000	20:56
Subject Sample	# 2	0.023	20:57
Breath Volume		1.468	20:57
Air Blank		0.000	20:58
Control Test		0.081	20:58
Air Blank		0.000	20:59
Diagnostics Check		OK	20:59

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,E,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: MALINAG

DATE/TIME OF TEST: 08/17/2018 23:01:16

Diagnostics Check		OK	22:58	
Air Blank		0.000	22:58	
Control Test		0.081	22:59	
Air Blank		0.000	22:59	
Subject Sample	# 1		22:59	Subject Test Refused
Breath Volume		0.000	22:59	
Air Blank		0.000	23:00	
Control Test		0.081	23:00	
Air Blank		0.000	23:01	
Diagnostics Check		OK	23:01	

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KELLY,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 08/18/2018 10:01:59

Diagnostics Check		OK	09:50	
Air Blank		0.000	09:50	
Control Test		0.081	09:51	
Air Blank		0.000	09:51	
Subject Sample	# 1	0.117	09:54	Volume Not Met
Breath Volume		0.253	09:54	
Air Blank		0.000	09:55	
Air Blank		0.000	09:57	
Subject Sample	# 2	0.138	10:00	Volume Not Met
Breath Volume		0.015	10:00	
Air Blank		0.000	10:01	
Control Test		0.080	10:01	
Air Blank		0.000	10:01	
Diagnostics Check		OK	10:01	

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILBERGER,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WOLFORD,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 08/18/2018 16:30:32

Diagnostics Check		OK	16:27	
Air Blank		0.000	16:28	
Control Test		0.080	16:28	
Air Blank		0.000	16:28	
Subject Sample	# 1		16:29	Subject Test Refused
Breath Volume		0.000	16:29	
Air Blank		0.000	16:29	
Control Test		0.081	16:29	
Air Blank		0.000	16:30	
Diagnostics Check		OK	16:30	

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILBERGER,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ALBERT,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: GRIFFITHS

DATE/TIME OF TEST: 08/18/2018 19:20:01

Diagnostics Check		OK	19:13
Air Blank		0.000	19:13
Control Test		0.080	19:14
Air Blank		0.000	19:14
Subject Sample	# 1	0.348	19:15
Breath Volume		4.097	19:15
Air Blank		0.000	19:15
Air Blank		0.000	19:17
Subject Sample	# 2	0.334	19:18
Breath Volume		3.671	19:18
Air Blank		0.000	19:19
Control Test		0.080	19:19
Air Blank		0.000	19:19
Diagnostics Check		OK	19:20

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,I
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 08/19/2018 16:24:16

Diagnostics Check		OK	16:17
Air Blank		0.000	16:17
Control Test		0.080	16:18
Air Blank		0.000	16:18
Subject Sample	# 1	0.231	16:19
Breath Volume		2.117	16:19
Air Blank		0.000	16:20
Air Blank		0.000	16:22
Subject Sample	# 2	0.227	16:22
Breath Volume		2.148	16:22
Air Blank		0.000	16:23
Control Test		0.080	16:23
Air Blank		0.000	16:24
Diagnostics Check		OK	16:24

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALDREP,B
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: JOLLY

DATE/TIME OF TEST: 08/19/2018 20:14:23

Diagnostics Check		OK	20:07
Air Blank		0.000	20:08
Control Test		0.080	20:08
Air Blank		0.000	20:08
Subject Sample	# 1	0.249	20:09
Breath Volume		1.378	20:09
Air Blank		0.000	20:10
Air Blank		0.000	20:12
Subject Sample	# 2	0.250	20:12
Breath Volume		1.468	20:12
Air Blank		0.000	20:13
Control Test		0.080	20:13
Air Blank		0.000	20:14
Diagnostics Check		OK	20:14

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,E,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 08/19/2018 20:54:07

Diagnostics Check		OK	20:51	
Air Blank		0.000	20:51	
Control Test		0.080	20:51	
Air Blank		0.000	20:52	
Subject Sample	# 1		20:52	Subject Test Refused
Breath Volume		0.000	20:52	
Air Blank		0.000	20:53	
Control Test		0.080	20:53	
Air Blank		0.000	20:54	
Diagnostics Check		OK	20:54	

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILBERGER,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MORIARTY

DATE/TIME OF TEST: 08/20/2018 22:27:40

Diagnostics Check		OK	22:20
Air Blank		0.000	22:21
Control Test		0.081	22:21
Air Blank		0.000	22:22
Subject Sample	# 1	0.000	22:23
Breath Volume		2.585	22:23
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.000	22:26
Breath Volume		2.816	22:26
Air Blank		0.000	22:26
Control Test		0.081	22:27
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILBERGER,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: ELHASSAN

DATE/TIME OF TEST: 08/21/2018 21:54:53

Diagnostics Check		OK	21:48
Air Blank		0.000	21:48
Control Test		0.081	21:49
Air Blank		0.000	21:49
Subject Sample	# 1	0.000	21:50
Breath Volume		4.558	21:50
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.000	21:53
Breath Volume		4.246	21:53
Air Blank		0.000	21:53
Control Test		0.080	21:54
Air Blank		0.000	21:54
Diagnostics Check		OK	21:54

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WEBER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/22/2018 01:55:30

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.081	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.051	01:51
Breath Volume		3.652	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.051	01:54
Breath Volume		3.886	01:54
Air Blank		0.000	01:54
Control Test		0.081	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: LEHRIAN

DATE/TIME OF TEST: 08/23/2018 01:13:33

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.081	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.187	01:08
Breath Volume		3.097	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.184	01:12
Breath Volume		2.839	01:12
Air Blank		0.000	01:12
Control Test		0.081	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 08/23/2018 04:13:23

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.081	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.144	04:07
Breath Volume		1.843	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.150	04:11
Breath Volume		1.250	04:11
Air Blank		0.000	04:12
Control Test		0.080	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DRABEK,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HAYS

DATE/TIME OF TEST: 08/24/2018 02:01:35

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.080	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.000	01:57
Breath Volume		2.160	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.000	02:00
Breath Volume		1.609	02:00
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: LIRANZO

DATE/TIME OF TEST: 08/24/2018 02:46:36

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.081	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.140	02:42
Breath Volume		2.351	02:42
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.140	02:45
Breath Volume		2.589	02:45
Air Blank		0.000	02:45
Control Test		0.082	02:45
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARDWELL,M
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: GENTRY

DATE/TIME OF TEST: 08/24/2018 05:23:20

Diagnostics Check		OK	05:20	
Air Blank		0.000	05:20	
Control Test		0.081	05:21	
Air Blank		0.000	05:21	
Subject Sample	# 1		05:21	Subject Test Refused
Breath Volume		0.000	05:21	
Air Blank		0.000	05:22	
Control Test		0.081	05:22	
Air Blank		0.000	05:23	
Diagnostics Check		OK	05:23	

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: MAYEN

DATE/TIME OF TEST: 08/25/2018 02:39:04

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.081	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.206	02:34
Breath Volume		1.894	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.211	02:37
Breath Volume		1.339	02:37
Air Blank		0.000	02:38
Control Test		0.080	02:38
Air Blank		0.000	02:38
Diagnostics Check		OK	02:39

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARMENTER,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MCCARY

DATE/TIME OF TEST: 08/25/2018 23:45:04

Diagnostics Check		OK	23:42	
Air Blank		0.000	23:42	
Control Test		0.080	23:42	
Air Blank		0.000	23:43	
Subject Sample	# 1		23:43	Subject Test Refused
Breath Volume		0.000	23:43	
Air Blank		0.000	23:44	
Control Test		0.081	23:44	
Air Blank		0.000	23:44	
Diagnostics Check		OK	23:45	

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBBINS,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: MATEO MORLAES

DATE/TIME OF TEST: 08/26/2018 04:34:16

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.081	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.160	04:29
Breath Volume		2.074	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.159	04:32
Breath Volume		1.621	04:32
Air Blank		0.000	04:33
Control Test		0.080	04:33
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SEMBACH,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:42

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 08/26/2018 06:05:54

Diagnostics Check		OK	06:03	
Air Blank		0.000	06:03	
Control Test		0.081	06:03	
Air Blank		0.000	06:04	
Subject Sample	# 1		06:04	Subject Test Refused
Breath Volume		0.000	06:04	
Air Blank		0.000	06:05	
Control Test		0.080	06:05	
Air Blank		0.000	06:05	
Diagnostics Check		OK	06:05	

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILBERGER,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: STANCIL

DATE/TIME OF TEST: 08/29/2018 23:12:13

Diagnostics Check		OK	23:09	
Air Blank		0.000	23:09	
Control Test		0.081	23:09	
Air Blank		0.000	23:10	
Subject Sample	# 1		23:10	Subject Test Refused
Breath Volume		0.000	23:10	
Air Blank		0.000	23:11	
Control Test		0.081	23:11	
Air Blank		0.000	23:12	
Diagnostics Check		OK	23:12	

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILBERGER,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: POULOS

DATE/TIME OF TEST: 08/30/2018 20:19:32

Diagnostics Check		OK	20:13
Air Blank		0.000	20:13
Control Test		0.082	20:13
Air Blank		0.000	20:14
Subject Sample	# 1	0.000	20:14
Breath Volume		2.410	20:14
Air Blank		0.000	20:15
Air Blank		0.000	20:17
Subject Sample	# 2	0.000	20:18
Breath Volume		2.421	20:18
Air Blank		0.000	20:18
Control Test		0.081	20:18
Air Blank		0.000	20:19
Diagnostics Check		OK	20:19

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LANE,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: CASON

DATE/TIME OF TEST: 08/31/2018 00:18:40

Diagnostics Check		OK	00:12
Air Blank		0.000	00:13
Control Test		0.081	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.169	00:14
Breath Volume		2.011	00:14
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.166	00:17
Breath Volume		2.707	00:17
Air Blank		0.000	00:17
Control Test		0.080	00:18
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/31/2018 02:58:06

Diagnostics Check		OK	02:55	
Air Blank		0.000	02:55	
Control Test		0.082	02:55	
Air Blank		0.000	02:56	
Subject Sample	# 1		02:56	Subject Test Refused
Breath Volume		0.000	02:56	
Air Blank		0.000	02:57	
Control Test		0.080	02:57	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILBERGER,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 08/31/2018 19:23:30

Diagnostics Check		OK	19:17
Air Blank		0.000	19:17
Control Test		0.081	19:17
Air Blank		0.000	19:18
Subject Sample	# 1	0.201	19:18
Breath Volume		2.277	19:18
Air Blank		0.000	19:19
Air Blank		0.000	19:21
Subject Sample	# 2	0.210	19:21
Breath Volume		2.410	19:21
Air Blank		0.000	19:22
Control Test		0.081	19:22
Air Blank		0.000	19:23
Diagnostics Check		OK	19:23

Data Download Date/Time: 09/06/2018 14:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

INSTRUMENT REGISTERED TO: NAS JACKSONVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUSINO,LUIS,H
BREATH TEST OPERATOR AGENCY: NAS Jacksonville PD
ARRESTING OFFICER: JUSINO,LUIS
ARRESTING OFFICER AGENCY: NAS Jacksonville PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:55

SUBJECT'S LAST NAME: MONTES

DATE/TIME OF TEST: 08/19/2018 09:29:56

Diagnostics Check		OK	09:22	
Air Blank		0.000	09:22	
Control Test		0.077	09:23	
Air Blank		0.000	09:23	
Subject Sample	# 1	0.225	09:25	
Breath Volume		2.863	09:25	
Air Blank		0.000	09:26	
Air Blank		0.000	09:27	
Subject Sample	# 2	0.227	09:28	
Breath Volume		1.785	09:28	
Air Blank		0.000	09:29	
Control Test		0.071	09:29	Control Outside Tolerance
Air Blank		0.000	09:29	

Data Download Date/Time: 08/19/2018 17:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

INSTRUMENT REGISTERED TO: NAS JACKSONVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUSINO,LUIS,H
BREATH TEST OPERATOR AGENCY: NAS Jacksonville PD
ARRESTING OFFICER: JUSINO,LUIS
ARRESTING OFFICER AGENCY: NAS Jacksonville PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: MONTES

DATE/TIME OF TEST: 08/19/2018 09:49:49

Diagnostics Check	OK	09:48	
Air Blank		09:49	Ambient Fail
Air Blank	0.000	09:49	

Data Download Date/Time: 08/19/2018 17:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

INSTRUMENT REGISTERED TO: NAS JACKSONVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUSINO,LUIS,H
BREATH TEST OPERATOR AGENCY: NAS Jacksonville PD
ARRESTING OFFICER: JUSINO,LUIS
ARRESTING OFFICER AGENCY: NAS Jacksonville PD
LAST AGENCY INSPECTION: 07/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: MONTES

DATE/TIME OF TEST: 08/19/2018 09:58:14

Diagnostics Check		OK	09:51
Air Blank		0.000	09:51
Control Test		0.077	09:51
Air Blank		0.000	09:52
Subject Sample	# 1	0.211	09:53
Breath Volume		2.019	09:53
Air Blank		0.000	09:54
Air Blank		0.000	09:56
Subject Sample	# 2	0.214	09:56
Breath Volume		1.347	09:56
Air Blank		0.000	09:57
Control Test		0.076	09:57
Air Blank		0.000	09:58
Diagnostics Check		OK	09:58

Data Download Date/Time: 08/19/2018 17:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGODA,JOSHUA,S
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: VEGODA,JOSHUA
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: ECKERT

DATE/TIME OF TEST: 08/04/2018 18:35:10

Diagnostics Check		OK	18:28
Air Blank		0.000	18:28
Control Test		0.080	18:28
Air Blank		0.000	18:29
Subject Sample	# 1	0.210	18:30
Breath Volume		2.523	18:30
Air Blank		0.000	18:30
Air Blank		0.000	18:32
Subject Sample	# 2	0.203	18:33
Breath Volume		2.871	18:33
Air Blank		0.000	18:34
Control Test		0.080	18:34
Air Blank		0.000	18:35
Diagnostics Check		OK	18:35

Data Download Date/Time: 08/22/2018 10:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: WHITTAKER

DATE/TIME OF TEST: 08/05/2018 02:04:23

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.081	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.140	02:00
Breath Volume		1.316	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.149	02:02
Breath Volume		1.453	02:02
Air Blank		0.000	02:03
Control Test		0.080	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 08/22/2018 10:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 06:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:23

SUBJECT'S LAST NAME: CHAVIANO

DATE/TIME OF TEST: 08/05/2018 06:46:18

Diagnostics Check		OK	06:43	
Air Blank		0.000	06:43	
Control Test		0.079	06:44	
Air Blank		0.000	06:44	
Subject Sample	# 1		06:44	Subject Test Refused
Breath Volume		0.000	06:44	
Air Blank		0.000	06:45	
Control Test		0.080	06:45	
Air Blank		0.000	06:46	
Diagnostics Check		OK	06:46	

Data Download Date/Time: 08/22/2018 10:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 06:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:23

SUBJECT'S LAST NAME: CHAVIANO

DATE/TIME OF TEST: 08/05/2018 07:00:50

Diagnostics Check		OK	06:53
Air Blank		0.000	06:53
Control Test		0.080	06:54
Air Blank		0.000	06:54
Subject Sample	# 1	0.157	06:56
Breath Volume		2.199	06:56
Air Blank		0.000	06:56
Air Blank		0.000	06:58
Subject Sample	# 2	0.150	06:59
Breath Volume		1.511	06:59
Air Blank		0.000	06:59
Control Test		0.080	07:00
Air Blank		0.000	07:00
Diagnostics Check		OK	07:00

Data Download Date/Time: 08/22/2018 10:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONGE,GRETTEL,M
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: RODRIGUEZ,MONICA
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 08/08/2018 23:09:40

Diagnostics Check		OK	22:57	
Air Blank		0.000	22:58	
Control Test		0.080	22:58	
Air Blank		0.000	22:58	
Subject Sample	# 1		23:02	No Sample Provided
Breath Volume		0.000	23:02	
Air Blank		0.000	23:02	
Air Blank		0.000	23:04	
Subject Sample	# 2	0.200	23:05	
Breath Volume		2.355	23:05	
Air Blank		0.000	23:05	
Air Blank		0.000	23:07	
Subject Sample	# 3	0.195	23:08	
Breath Volume		3.910	23:08	
Air Blank		0.000	23:08	
Control Test		0.079	23:09	
Air Blank		0.000	23:09	
Diagnostics Check		OK	23:09	

Data Download Date/Time: 08/22/2018 10:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUMPHRY,KERRY,R
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: VASILIEVICH,M
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: BONATTI

DATE/TIME OF TEST: 08/10/2018 05:25:11

Diagnostics Check		OK	05:22	
Air Blank		0.000	05:22	
Control Test		0.079	05:23	
Air Blank		0.000	05:23	
Subject Sample	# 1		05:23	Subject Test Refused
Breath Volume		0.000	05:23	
Air Blank		0.000	05:24	
Control Test		0.080	05:24	
Air Blank		0.000	05:25	
Diagnostics Check		OK	05:25	

Data Download Date/Time: 08/22/2018 10:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: BUENO

DATE/TIME OF TEST: 08/11/2018 04:51:08

Diagnostics Check		OK	04:48	
Air Blank		0.000	04:48	
Control Test		0.080	04:49	
Air Blank		0.000	04:49	
Subject Sample	# 1		04:49	Subject Test Refused
Breath Volume		0.000	04:49	
Air Blank		0.000	04:50	
Control Test		0.080	04:50	
Air Blank		0.000	04:51	
Diagnostics Check		OK	04:51	

Data Download Date/Time: 08/22/2018 10:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SOSA,JONATHAN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 08/13/2018 00:35:57

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.080	00:27
Air Blank		0.000	00:27
Subject Sample	# 1	0.156	00:30
Breath Volume		1.128	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.139	00:34
Breath Volume		1.492	00:34
Air Blank		0.000	00:35
Control Test		0.078	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 08/22/2018 10:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: BEY

DATE/TIME OF TEST: 08/18/2018 04:12:52

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:10	
Control Test		0.080	04:10	
Air Blank		0.000	04:11	
Subject Sample	# 1		04:11	Subject Test Refused
Breath Volume		0.000	04:11	
Air Blank		0.000	04:11	
Control Test		0.080	04:12	
Air Blank		0.000	04:12	
Diagnostics Check		OK	04:12	

Data Download Date/Time: 08/22/2018 10:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: KOUME

DATE/TIME OF TEST: 08/18/2018 06:26:45

Diagnostics Check		OK	06:23	
Air Blank		0.000	06:24	
Control Test		0.080	06:24	
Air Blank		0.000	06:25	
Subject Sample	# 1		06:25	Subject Test Refused
Breath Volume		0.000	06:25	
Air Blank		0.000	06:25	
Control Test		0.079	06:26	
Air Blank		0.000	06:26	
Diagnostics Check		OK	06:26	

Data Download Date/Time: 08/22/2018 10:52

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAZONOV,VLADISLAV,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: KLEMENTS,ALEXANDRA
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 09:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: ALBRAK

DATE/TIME OF TEST: 08/18/2018 10:53:47

Diagnostics Check		OK	10:40	
Air Blank		0.000	10:41	
Control Test		0.080	10:41	
Air Blank		0.000	10:42	
Subject Sample	# 1	0.086	10:45	Volume Not Met
Breath Volume		0.558	10:45	
Air Blank		0.000	10:45	
Air Blank		0.000	10:47	
Subject Sample	# 2	0.081	10:49	
Breath Volume		1.875	10:49	
Air Blank		0.000	10:49	
Air Blank		0.000	10:51	
Subject Sample	# 3	0.080	10:52	
Breath Volume		1.222	10:52	
Air Blank		0.000	10:52	
Control Test		0.079	10:53	
Air Blank		0.000	10:53	
Diagnostics Check		OK	10:53	

Data Download Date/Time: 08/22/2018 10:52

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SOSA,JONATHAN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: FRIAS II

DATE/TIME OF TEST: 08/20/2018 03:02:02

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.079	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.153	02:57
Breath Volume		1.921	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.150	03:00
Breath Volume		2.152	03:00
Air Blank		0.000	03:01
Control Test		0.080	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:02

Data Download Date/Time: 08/22/2018 10:52

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,RICARDO,A
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CASTILLO,RICARDO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: MONROY

DATE/TIME OF TEST: 08/21/2018 05:42:22

Diagnostics Check		OK	05:35
Air Blank		0.000	05:36
Control Test		0.080	05:36
Air Blank		0.000	05:37
Subject Sample	# 1	0.124	05:37
Breath Volume		3.808	05:37
Air Blank		0.000	05:38
Air Blank		0.000	05:40
Subject Sample	# 2	0.126	05:40
Breath Volume		2.011	05:40
Air Blank		0.000	05:41
Control Test		0.080	05:41
Air Blank		0.000	05:42
Diagnostics Check		OK	05:42

Data Download Date/Time: 08/22/2018 10:52

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CALLON,TONY
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PERALTA

DATE/TIME OF TEST: 08/25/2018 01:54:44

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:46	
Control Test		0.080	01:46	
Air Blank		0.000	01:47	
Subject Sample	# 1		01:50	No Sample Provided
Breath Volume		0.000	01:50	
Air Blank		0.000	01:50	
Air Blank		0.000	01:52	
Subject Sample	# 2		01:53	Subject Test Refused
Breath Volume		0.000	01:53	
Air Blank		0.000	01:53	
Control Test		0.079	01:54	
Air Blank		0.000	01:54	
Diagnostics Check		OK	01:54	

Data Download Date/Time: 09/25/2018 10:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SOSA,JONATHAN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 08/25/2018 03:37:04

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:33	
Control Test		0.080	03:33	
Air Blank		0.000	03:34	
Subject Sample	# 1	0.154	03:35	
Breath Volume		2.121	03:35	
Air Blank			03:35	Improper Sample
Air Blank		0.000	03:36	
Control Test		0.080	03:36	
Air Blank		0.000	03:36	
Diagnostics Check		OK	03:37	

Data Download Date/Time: 09/25/2018 10:07

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SOSA,JONATHAN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 08/25/2018 03:46:43

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.080	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.150	03:42
Breath Volume		1.445	03:42
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.150	03:45
Breath Volume		1.144	03:45
Air Blank		0.000	03:45
Control Test		0.081	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 09/25/2018 10:07

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MELENDEZ,DANIEL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: JIMENEZ GARCIA

DATE/TIME OF TEST: 08/25/2018 05:16:30

Diagnostics Check		OK	05:10
Air Blank		0.000	05:11
Control Test		0.078	05:11
Air Blank		0.000	05:11
Subject Sample	# 1	0.235	05:12
Breath Volume		1.859	05:12
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.230	05:15
Breath Volume		1.585	05:15
Air Blank		0.000	05:15
Control Test		0.080	05:15
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 09/25/2018 10:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGODA,JOSHUA,S
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: VIDAL,ELIZABETH
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: SCHAEFER

DATE/TIME OF TEST: 08/25/2018 19:58:27

Diagnostics Check		OK	19:55	
Air Blank		0.000	19:55	
Control Test		0.080	19:55	
Air Blank		0.000	19:56	
Subject Sample	# 1		19:57	Subject Test Refused
Breath Volume		0.000	19:57	
Air Blank		0.000	19:57	
Control Test		0.080	19:57	
Air Blank		0.000	19:58	
Diagnostics Check		OK	19:58	

Data Download Date/Time: 09/25/2018 10:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,PEDRO,A
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: VIDAL,ELIZABETH
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 11:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:07

SUBJECT'S LAST NAME: CUBILLOS-RESTREPO

DATE/TIME OF TEST: 08/26/2018 12:36:14

Diagnostics Check		OK	12:29
Air Blank		0.000	12:29
Control Test		0.080	12:30
Air Blank		0.000	12:30
Subject Sample	# 1	0.220	12:31
Breath Volume		2.273	12:31
Air Blank		0.000	12:32
Air Blank		0.000	12:34
Subject Sample	# 2	0.212	12:34
Breath Volume		2.179	12:34
Air Blank		0.000	12:35
Control Test		0.078	12:35
Air Blank		0.000	12:36
Diagnostics Check		OK	12:36

Data Download Date/Time: 09/25/2018 10:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/26/2018 18:17:44

Diagnostics Check		OK	18:10
Air Blank		0.000	18:10
Control Test		0.080	18:11
Air Blank		0.000	18:11
Subject Sample	# 1	0.092	18:12
Breath Volume		1.140	18:12
Air Blank		0.000	18:13
Air Blank		0.000	18:15
Subject Sample	# 2	0.082	18:16
Breath Volume		1.382	18:16
Air Blank		0.000	18:16
Control Test		0.080	18:17
Air Blank		0.000	18:17
Diagnostics Check		OK	18:17

Data Download Date/Time: 09/25/2018 10:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONFESSORE,STEVE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: ARDILA,LIZBETH
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ANTANES MILIAN

DATE/TIME OF TEST: 08/27/2018 23:06:05

Diagnostics Check		OK	22:59
Air Blank		0.000	22:59
Control Test		0.079	22:59
Air Blank		0.000	23:00
Subject Sample	# 1	0.209	23:00
Breath Volume		3.207	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.206	23:04
Breath Volume		3.839	23:04
Air Blank		0.000	23:05
Control Test		0.079	23:05
Air Blank		0.000	23:05
Diagnostics Check		OK	23:06

Data Download Date/Time: 09/25/2018 10:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESTER,REGINALD,D
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PALACIOS,J
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: MARANUK

DATE/TIME OF TEST: 08/30/2018 17:27:39

Diagnostics Check		OK	17:20
Air Blank		0.000	17:21
Control Test		0.081	17:21
Air Blank		0.000	17:22
Subject Sample	# 1	0.156	17:22
Breath Volume		1.941	17:22
Air Blank		0.000	17:23
Air Blank		0.000	17:25
Subject Sample	# 2	0.145	17:26
Breath Volume		1.656	17:26
Air Blank		0.000	17:26
Control Test		0.080	17:27
Air Blank		0.000	17:27
Diagnostics Check		OK	17:27

Data Download Date/Time: 09/25/2018 10:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONGE,GRETTEL,M
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SOSA,JONATHAN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: CAVE

DATE/TIME OF TEST: 08/31/2018 00:58:45

Diagnostics Check		OK	00:51
Air Blank		0.000	00:52
Control Test		0.080	00:52
Air Blank		0.000	00:53
Subject Sample	# 1	0.181	00:54
Breath Volume		4.089	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.172	00:57
Breath Volume		3.968	00:57
Air Blank		0.000	00:57
Control Test		0.078	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 09/25/2018 10:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,DANIELLA,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: COLLADO

DATE/TIME OF TEST: 08/31/2018 21:02:24

Diagnostics Check		OK	20:56
Air Blank		0.000	20:56
Control Test		0.080	20:57
Air Blank		0.000	20:57
Subject Sample	# 1	0.000	20:58
Breath Volume		1.839	20:58
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.000	21:01
Breath Volume		1.765	21:01
Air Blank		0.000	21:01
Control Test		0.080	21:01
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 09/25/2018 10:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,AILEEN,B
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: HERNANDEZ,OSDANY
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 08/19/2018 05:11:38

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.078	05:04
Air Blank		0.000	05:04
Subject Sample	# 1	0.214	05:05
Breath Volume		1.191	05:05
Air Blank		0.000	05:06
Air Blank		0.000	05:07
Subject Sample	# 2	0.213	05:09
Breath Volume		1.734	05:09
Air Blank		0.000	05:10
Control Test		0.077	05:11
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 09/19/2018 10:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROJAS,GABRIEL,L
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: GARZON,HAROLD
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: COLLANTE

DATE/TIME OF TEST: 08/31/2018 02:10:58

Diagnostics Check	OK	02:10	
Air Blank		02:10	RFI Detect
Air Blank	0.000	02:10	

Data Download Date/Time: 09/19/2018 10:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROJAS,GABRIEL,L
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: GARZON,HAROLD
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 08/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: COLLANTE

DATE/TIME OF TEST: 08/31/2018 02:18:10

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:15	
Control Test		0.078	02:16	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:16	Subject Test Refused
Breath Volume		0.000	02:16	
Air Blank		0.000	02:17	
Control Test		0.078	02:17	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 09/19/2018 10:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHERTY,DANIEL,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: DOCTOR

DATE/TIME OF TEST: 08/08/2018 02:16:13

Diagnostics Check		OK	02:09
Air Blank		0.000	02:09
Control Test		0.080	02:10
Air Blank		0.000	02:10
Subject Sample	# 1	0.132	02:11
Breath Volume		1.691	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.130	02:14
Breath Volume		2.074	02:14
Air Blank		0.000	02:15
Control Test		0.078	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 09/26/2018 21:58

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NARCISSE,ANORCE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: NARCISSE,ANORCE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 11:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: PINON

DATE/TIME OF TEST: 08/03/2018 01:45:40

Diagnostics Check	OK	01:45	
Air Blank		01:45	RFI Detect
Air Blank	0.000	01:45	

Data Download Date/Time: 08/03/2018 06:17

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: DUNBAR

DATE/TIME OF TEST: 08/07/2018 21:24:20

Diagnostics Check		OK	21:17
Air Blank		0.000	21:17
Control Test		0.077	21:17
Air Blank		0.000	21:18
Subject Sample	# 1	0.198	21:19
Breath Volume		1.449	21:19
Air Blank		0.000	21:20
Air Blank		0.000	21:22
Subject Sample	# 2	0.200	21:22
Breath Volume		2.187	21:22
Air Blank		0.000	21:23
Control Test		0.077	21:23
Air Blank		0.000	21:24
Diagnostics Check		OK	21:24

Data Download Date/Time: 09/06/2018 14:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 08/08/2018 03:57:34

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.078	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.089	03:52
Breath Volume		1.246	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:54
Subject Sample	# 2	0.098	03:55
Breath Volume		1.992	03:55
Air Blank		0.000	03:56
Control Test		0.078	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 09/06/2018 14:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHILLER,E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 08/08/2018 22:52:37

Diagnostics Check		OK	22:49	
Air Blank		0.000	22:49	
Control Test		0.078	22:50	
Air Blank		0.000	22:50	
Subject Sample	# 1		22:51	Slope Not Met
Breath Volume		0.949	22:51	
Air Blank		0.000	22:51	
Control Test		0.078	22:52	
Air Blank		0.000	22:52	
Diagnostics Check		OK	22:52	

Data Download Date/Time: 09/06/2018 14:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHILLER,E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 08/08/2018 23:20:34

Diagnostics Check		OK	23:14
Air Blank		0.000	23:14
Control Test		0.078	23:15
Air Blank		0.000	23:15
Subject Sample	# 1	0.297	23:16
Breath Volume		2.085	23:16
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.311	23:18
Breath Volume		2.054	23:18
Air Blank		0.000	23:19
Control Test		0.078	23:19
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 09/06/2018 14:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS-COLON,S
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: PABLO-REYES

DATE/TIME OF TEST: 08/10/2018 01:53:21

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.078	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.171	01:48
Breath Volume		1.402	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.183	01:51
Breath Volume		2.144	01:51
Air Blank		0.000	01:52
Control Test		0.077	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 09/06/2018 14:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/11/2018 01:54:58

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.079	01:49
Air Blank		0.000	01:49
Subject Sample	# 1	0.095	01:50
Breath Volume		4.031	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:52
Subject Sample	# 2	0.099	01:53
Breath Volume		4.234	01:53
Air Blank		0.000	01:54
Control Test		0.078	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 09/06/2018 14:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 08/11/2018 03:49:28

Diagnostics Check		OK	03:42
Air Blank		0.000	03:43
Control Test		0.079	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.095	03:44
Breath Volume		3.421	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.100	03:47
Breath Volume		3.949	03:47
Air Blank		0.000	03:48
Control Test		0.078	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 09/06/2018 14:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: CANION III

DATE/TIME OF TEST: 08/11/2018 04:43:32

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.077	04:37
Air Blank		0.000	04:38
Subject Sample	# 1	0.090	04:39
Breath Volume		1.757	04:39
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.095	04:41
Breath Volume		2.332	04:41
Air Blank		0.000	04:42
Control Test		0.077	04:42
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 09/06/2018 14:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BODAMER,E
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/12/2018 00:17:36

Diagnostics Check		OK	00:13	
Air Blank		0.000	00:14	
Control Test		0.078	00:14	
Air Blank		0.000	00:15	
Subject Sample	# 1		00:16	Subject Test Refused
Breath Volume		0.000	00:16	
Air Blank		0.000	00:16	
Control Test		0.078	00:17	
Air Blank		0.000	00:17	
Diagnostics Check		OK	00:17	

Data Download Date/Time: 09/06/2018 14:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: CLINTON-HARRIS

DATE/TIME OF TEST: 08/12/2018 04:18:33

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.079	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.133	04:13
Breath Volume		1.437	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.139	04:16
Breath Volume		2.695	04:16
Air Blank		0.000	04:17
Control Test		0.078	04:17
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 09/06/2018 14:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: MINAHAN

DATE/TIME OF TEST: 08/12/2018 19:31:08

Diagnostics Check		OK	19:24
Air Blank		0.000	19:24
Control Test		0.078	19:25
Air Blank		0.000	19:25
Subject Sample	# 1	0.236	19:26
Breath Volume		1.699	19:26
Air Blank		0.000	19:26
Air Blank		0.000	19:28
Subject Sample	# 2	0.251	19:29
Breath Volume		1.867	19:29
Air Blank		0.000	19:30
Control Test		0.078	19:30
Air Blank		0.000	19:31
Diagnostics Check		OK	19:31

Data Download Date/Time: 09/06/2018 14:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 08/15/2018 02:50:06

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.078	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.007	02:45
Breath Volume		1.554	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.000	02:48
Breath Volume		1.644	02:48
Air Blank		0.000	02:49
Control Test		0.077	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 09/06/2018 14:33

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 08/15/2018 04:13:20

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:02	
Control Test		0.078	04:02	
Air Blank		0.000	04:03	
Subject Sample	# 1	0.088	04:06	Volume Not Met
Breath Volume		0.402	04:06	
Air Blank		0.000	04:06	
Air Blank		0.000	04:08	
Subject Sample	# 2		04:11	No Sample Provided
Breath Volume		0.000	04:11	
Air Blank		0.000	04:12	
Control Test		0.078	04:12	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: TRUSCHEL

DATE/TIME OF TEST: 08/15/2018 20:42:25

Diagnostics Check		OK	20:35
Air Blank		0.000	20:36
Control Test		0.078	20:36
Air Blank		0.000	20:37
Subject Sample	# 1	0.133	20:37
Breath Volume		2.785	20:37
Air Blank		0.000	20:38
Air Blank		0.000	20:40
Subject Sample	# 2	0.122	20:40
Breath Volume		1.492	20:40
Air Blank		0.000	20:41
Control Test		0.078	20:41
Air Blank		0.000	20:42
Diagnostics Check		OK	20:42

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: AERUMALIKARA KURIAN

DATE/TIME OF TEST: 08/16/2018 04:08:24

Diagnostics Check		OK	03:55	
Air Blank		0.000	03:55	
Control Test		0.078	03:56	
Air Blank		0.000	03:56	
Subject Sample	# 1	0.164	03:59	Volume Not Met
Breath Volume		0.273	03:59	
Air Blank		0.000	04:00	
Air Blank		0.000	04:02	
Subject Sample	# 2	0.167	04:03	
Breath Volume		1.511	04:03	
Air Blank		0.000	04:04	
Air Blank		0.000	04:05	
Subject Sample	# 3	0.159	04:06	
Breath Volume		2.199	04:06	
Air Blank		0.000	04:07	
Control Test		0.077	04:07	
Air Blank		0.000	04:08	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BALOGH

DATE/TIME OF TEST: 08/17/2018 00:27:55

Diagnostics Check		OK	00:20
Air Blank		0.000	00:21
Control Test		0.078	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.148	00:22
Breath Volume		1.554	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.135	00:26
Breath Volume		1.386	00:26
Air Blank		0.000	00:27
Control Test		0.078	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 08/17/2018 05:10:55

Diagnostics Check		OK	05:04
Air Blank		0.000	05:04
Control Test		0.079	05:05
Air Blank		0.000	05:05
Subject Sample	# 1	0.149	05:06
Breath Volume		1.683	05:06
Air Blank		0.000	05:07
Air Blank		0.000	05:08
Subject Sample	# 2	0.158	05:09
Breath Volume		2.191	05:09
Air Blank		0.000	05:10
Control Test		0.078	05:10
Air Blank		0.000	05:10
Diagnostics Check		OK	05:10

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/18/2018 00:12:23

Diagnostics Check		OK	00:05
Air Blank		0.000	00:06
Control Test		0.079	00:06
Air Blank		0.000	00:07
Subject Sample	# 1	0.164	00:07
Breath Volume		2.382	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:09
Subject Sample	# 2	0.161	00:10
Breath Volume		2.812	00:10
Air Blank		0.000	00:11
Control Test		0.077	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: FILEP

DATE/TIME OF TEST: 08/18/2018 04:44:39

Diagnostics Check		OK	04:35
Air Blank		0.000	04:35
Control Test		0.078	04:35
Air Blank		0.000	04:36
Subject Sample	# 1	0.207	04:37
Breath Volume		2.960	04:37
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.181	04:40
Breath Volume		1.648	04:40
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 3	0.182	04:43
Breath Volume		1.175	04:43
Air Blank		0.000	04:43
Control Test		0.077	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NEIPERT,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: VEJAR-BARRERA

DATE/TIME OF TEST: 08/19/2018 01:31:57

Diagnostics Check		OK	01:24
Air Blank		0.000	01:24
Control Test		0.077	01:25
Air Blank		0.000	01:25
Subject Sample	# 1	0.149	01:27
Breath Volume		2.382	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.150	01:30
Breath Volume		2.187	01:30
Air Blank		0.000	01:31
Control Test		0.078	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 08/19/2018 04:14:41

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:03	
Control Test		0.078	04:03	
Air Blank		0.000	04:04	
Subject Sample	# 1		04:07	No Sample Provided
Breath Volume		0.000	04:07	
Air Blank		0.000	04:07	
Air Blank		0.000	04:09	
Subject Sample	# 2	0.226	04:12	Volume Not Met
Breath Volume		0.015	04:12	
Air Blank		0.000	04:13	
Control Test		0.078	04:14	
Air Blank		0.000	04:14	
Diagnostics Check		OK	04:14	

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 08/19/2018 04:27:34

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.077	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.224	04:20
Breath Volume		1.617	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.235	04:25
Breath Volume		2.496	04:25
Air Blank		0.000	04:26
Control Test		0.076	04:26
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: RENZA

DATE/TIME OF TEST: 08/20/2018 00:19:48

Diagnostics Check		OK	00:13
Air Blank		0.000	00:14
Control Test		0.078	00:14
Air Blank		0.000	00:14
Subject Sample	# 1	0.168	00:15
Breath Volume		1.562	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:17
Subject Sample	# 2	0.171	00:18
Breath Volume		1.324	00:18
Air Blank		0.000	00:18
Control Test		0.078	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 08/20/2018 01:12:46

Diagnostics Check		OK	01:04
Air Blank		0.000	01:05
Control Test		0.078	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.230	01:07
Breath Volume		1.843	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.243	01:11
Breath Volume		1.714	01:11
Air Blank		0.000	01:11
Control Test		0.077	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAULINO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/20/2018 05:06:50

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.077	05:00
Air Blank		0.000	05:01
Subject Sample	# 1	0.168	05:02
Breath Volume		2.500	05:02
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.167	05:05
Breath Volume		2.164	05:05
Air Blank		0.000	05:05
Control Test		0.077	05:06
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 08/21/2018 00:52:52

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.078	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.188	00:48
Breath Volume		2.480	00:48
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.187	00:51
Breath Volume		2.632	00:51
Air Blank		0.000	00:51
Control Test		0.077	00:52
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: HARDWICK

DATE/TIME OF TEST: 08/21/2018 19:59:49

Diagnostics Check		OK	19:53
Air Blank		0.000	19:53
Control Test		0.078	19:54
Air Blank		0.000	19:54
Subject Sample	# 1	0.181	19:55
Breath Volume		2.101	19:55
Air Blank		0.000	19:55
Air Blank		0.000	19:57
Subject Sample	# 2	0.172	19:58
Breath Volume		1.976	19:58
Air Blank		0.000	19:58
Control Test		0.078	19:59
Air Blank		0.000	19:59
Diagnostics Check		OK	19:59

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRIZENDINE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MILANES MORALES

DATE/TIME OF TEST: 08/23/2018 01:06:59

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.079	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.027	01:02
Breath Volume		1.636	01:02
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.027	01:05
Breath Volume		2.128	01:05
Air Blank		0.000	01:06
Control Test		0.078	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 08/23/2018 03:59:03

Diagnostics Check		OK	03:52
Air Blank		0.000	03:52
Control Test		0.078	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.190	03:54
Breath Volume		1.648	03:54
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.191	03:57
Breath Volume		3.062	03:57
Air Blank		0.000	03:58
Control Test		0.078	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:59

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PYLES

DATE/TIME OF TEST: 08/24/2018 01:53:28

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.078	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.172	01:48
Breath Volume		1.812	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:50
Subject Sample	# 2	0.181	01:51
Breath Volume		2.859	01:51
Air Blank		0.000	01:52
Control Test		0.078	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: CUTINI

DATE/TIME OF TEST: 08/25/2018 00:50:52

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.078	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.210	00:46
Breath Volume		1.847	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:48
Subject Sample	# 2	0.195	00:49
Breath Volume		1.687	00:49
Air Blank		0.000	00:49
Control Test		0.078	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: HUFFMAN

DATE/TIME OF TEST: 08/25/2018 02:15:19

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:01	
Control Test		0.078	02:02	
Air Blank		0.000	02:02	
Subject Sample	# 1	0.149	02:05	Volume Not Met
Breath Volume		0.542	02:05	
Air Blank		0.000	02:06	
Air Blank		0.000	02:08	
Subject Sample	# 2	0.156	02:10	
Breath Volume		1.597	02:10	
Air Blank		0.000	02:10	
Air Blank		0.000	02:12	
Subject Sample	# 3	0.167	02:13	
Breath Volume		2.070	02:13	
Air Blank		0.000	02:14	
Control Test		0.077	02:14	
Air Blank		0.000	02:15	
Diagnostics Check		OK	02:15	

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: HOLDSWORTH

DATE/TIME OF TEST: 08/25/2018 02:41:48

Diagnostics Check		OK	02:34
Air Blank		0.000	02:35
Control Test		0.078	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.116	02:36
Breath Volume		1.601	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.119	02:40
Breath Volume		1.406	02:40
Air Blank		0.000	02:40
Control Test		0.077	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HANVILLE

DATE/TIME OF TEST: 08/25/2018 04:04:07

Diagnostics Check		OK	03:51	
Air Blank		0.000	03:51	
Control Test		0.078	03:51	
Air Blank		0.000	03:52	
Subject Sample	# 1		03:55	No Sample Provided
Breath Volume		0.000	03:55	
Air Blank		0.000	03:55	
Air Blank		0.000	03:57	
Subject Sample	# 2	0.150	03:59	
Breath Volume		2.277	03:59	
Air Blank		0.000	04:00	
Air Blank		0.000	04:01	
Subject Sample	# 3	0.157	04:02	
Breath Volume		2.675	04:02	
Air Blank		0.000	04:03	
Control Test		0.077	04:03	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 08/25/2018 05:52:50

Diagnostics Check		OK	05:45
Air Blank		0.000	05:46
Control Test		0.078	05:46
Air Blank		0.000	05:47
Subject Sample	# 1	0.045	05:48
Breath Volume		1.527	05:48
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.035	05:51
Breath Volume		1.140	05:51
Air Blank		0.000	05:51
Control Test		0.078	05:52
Air Blank		0.000	05:52
Diagnostics Check		OK	05:52

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HILES,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: SILCOX

DATE/TIME OF TEST: 08/25/2018 23:20:45

Diagnostics Check		OK	23:14
Air Blank		0.000	23:14
Control Test		0.078	23:15
Air Blank		0.000	23:15
Subject Sample	# 1	0.007	23:16
Breath Volume		1.285	23:16
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.007	23:19
Breath Volume		1.589	23:19
Air Blank		0.000	23:19
Control Test		0.078	23:20
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 08/26/2018 02:20:31

Diagnostics Check		OK	02:12
Air Blank		0.000	02:12
Control Test		0.078	02:13
Air Blank		0.000	02:13
Subject Sample	# 1	0.210	02:15
Breath Volume		1.148	02:15
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.219	02:18
Breath Volume		1.562	02:18
Air Blank		0.000	02:19
Control Test		0.076	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: EGGERS

DATE/TIME OF TEST: 08/26/2018 03:34:58

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.077	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.158	03:30
Breath Volume		1.593	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:32
Subject Sample	# 2	0.154	03:33
Breath Volume		1.375	03:33
Air Blank		0.000	03:34
Control Test		0.077	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: LAGUNA IV

DATE/TIME OF TEST: 08/26/2018 06:12:05

Diagnostics Check		OK	06:04
Air Blank		0.000	06:05
Control Test		0.078	06:05
Air Blank		0.000	06:06
Subject Sample	# 1	0.134	06:06
Breath Volume		1.343	06:06
Air Blank		0.000	06:07
Air Blank		0.000	06:09
Subject Sample	# 2	0.139	06:10
Breath Volume		2.566	06:10
Air Blank		0.000	06:11
Control Test		0.078	06:11
Air Blank		0.000	06:12
Diagnostics Check		OK	06:12

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 08/27/2018 00:23:14

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.078	00:17
Air Blank		0.000	00:17
Subject Sample	# 1	0.123	00:18
Breath Volume		2.535	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.115	00:21
Breath Volume		1.285	00:21
Air Blank		0.000	00:22
Control Test		0.078	00:22
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: EWING

DATE/TIME OF TEST: 08/27/2018 04:11:20

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.078	04:05
Air Blank		0.000	04:06
Subject Sample	# 1	0.065	04:06
Breath Volume		1.828	04:06
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.066	04:09
Breath Volume		2.062	04:09
Air Blank		0.000	04:10
Control Test		0.078	04:10
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 08/28/2018 04:35:02

Diagnostics Check		OK	04:28
Air Blank		0.000	04:28
Control Test		0.078	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.192	04:30
Breath Volume		1.222	04:30
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.201	04:33
Breath Volume		2.144	04:33
Air Blank		0.000	04:34
Control Test		0.077	04:34
Air Blank		0.000	04:34
Diagnostics Check		OK	04:35

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: FIKE

DATE/TIME OF TEST: 08/29/2018 23:38:36

Diagnostics Check		OK	23:23	
Air Blank		0.000	23:24	
Control Test		0.079	23:24	
Air Blank		0.000	23:25	
Subject Sample	# 1	0.198	23:28	Volume Not Met
Breath Volume		0.722	23:28	
Air Blank		0.000	23:28	
Air Blank		0.000	23:30	
Subject Sample	# 2	0.208	23:33	
Breath Volume		1.144	23:33	
Air Blank		0.000	23:33	
Air Blank		0.000	23:35	
Subject Sample	# 3	0.204	23:36	
Breath Volume		1.140	23:36	
Air Blank		0.000	23:37	
Control Test		0.077	23:37	
Air Blank		0.000	23:38	
Diagnostics Check		OK	23:38	

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: TOOTS

DATE/TIME OF TEST: 08/30/2018 22:20:40

Diagnostics Check		OK	22:13
Air Blank		0.000	22:14
Control Test		0.078	22:14
Air Blank		0.000	22:15
Subject Sample	# 1	0.235	22:16
Breath Volume		3.652	22:16
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.217	22:19
Breath Volume		1.699	22:19
Air Blank		0.000	22:19
Control Test		0.077	22:20
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: ROMANELLO

DATE/TIME OF TEST: 08/30/2018 23:23:08

Diagnostics Check		OK	23:14
Air Blank		0.000	23:14
Control Test		0.077	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.000	23:18
Breath Volume		1.187	23:18
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.000	23:21
Breath Volume		1.210	23:21
Air Blank		0.000	23:22
Control Test		0.078	23:22
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 09/06/2018 14:34

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 08/02/2018 22:46:16

Diagnostics Check		OK	22:39
Air Blank		0.000	22:40
Control Test		0.081	22:40
Air Blank		0.000	22:40
Subject Sample	# 1	0.116	22:41
Breath Volume		3.152	22:41
Air Blank		0.000	22:42
Air Blank		0.000	22:44
Subject Sample	# 2	0.111	22:44
Breath Volume		2.738	22:44
Air Blank		0.000	22:45
Control Test		0.080	22:45
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 08/07/2018 13:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: VEERAMACHANENI

DATE/TIME OF TEST: 08/04/2018 03:55:42

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.080	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.170	03:48
Breath Volume		1.164	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.156	03:54
Breath Volume		1.125	03:54
Air Blank		0.000	03:54
Control Test		0.077	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 08/07/2018 13:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULLIVAN,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: PENIX

DATE/TIME OF TEST: 08/05/2018 01:31:17

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.081	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.173	01:24
Breath Volume		1.386	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.165	01:29
Breath Volume		1.453	01:29
Air Blank		0.000	01:30
Control Test		0.080	01:30
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 08/07/2018 13:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: HAHN

DATE/TIME OF TEST: 08/05/2018 04:19:29

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.080	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.188	04:14
Breath Volume		2.042	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.188	04:17
Breath Volume		2.652	04:17
Air Blank		0.000	04:18
Control Test		0.080	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 08/07/2018 13:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: JEAN SIMON

DATE/TIME OF TEST: 08/06/2018 00:14:57

Diagnostics Check		OK	00:08
Air Blank		0.000	00:08
Control Test		0.081	00:09
Air Blank		0.000	00:09
Subject Sample	# 1	0.162	00:10
Breath Volume		1.546	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:12
Subject Sample	# 2	0.169	00:13
Breath Volume		2.929	00:13
Air Blank		0.000	00:14
Control Test		0.081	00:14
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 08/07/2018 13:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: DIAZ-BLANCO

DATE/TIME OF TEST: 08/06/2018 04:32:45

Diagnostics Check		OK	04:23
Air Blank		0.000	04:24
Control Test		0.080	04:24
Air Blank		0.000	04:25
Subject Sample	# 1	0.141	04:27
Breath Volume		2.843	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.128	04:31
Breath Volume		1.406	04:31
Air Blank		0.000	04:31
Control Test		0.079	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 08/07/2018 13:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JOSEPH,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: SCARPINATO

DATE/TIME OF TEST: 08/07/2018 02:47:54

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.080	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.200	02:43
Breath Volume		2.738	02:43
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.206	02:46
Breath Volume		2.402	02:46
Air Blank		0.000	02:46
Control Test		0.079	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 08/07/2018 13:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: CASTRO JR.

DATE/TIME OF TEST: 08/08/2018 00:07:11

Diagnostics Check		OK	23:54	
Air Blank		0.000	23:55	
Control Test		0.080	23:55	
Air Blank		0.000	23:55	
Subject Sample	# 3	0.076	23:59	Volume Not Met
Breath Volume		0.125	23:59	
Air Blank		0.000	23:59	
Air Blank		0.000	00:01	
Subject Sample	# 1	0.080	00:02	
Breath Volume		1.417	00:02	
Air Blank		0.000	00:02	
Air Blank		0.000	00:04	
Subject Sample	# 2	0.080	00:05	
Breath Volume		1.605	00:05	
Air Blank		0.000	00:06	
Control Test		0.080	00:06	
Air Blank		0.000	00:07	
Diagnostics Check		OK	00:07	

Data Download Date/Time: 09/06/2018 14:37

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 08/08/2018 03:53:29

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:50	
Control Test		0.080	03:50	
Air Blank		0.000	03:51	
Subject Sample	# 1		03:52	Subject Test Refused
Breath Volume		0.000	03:52	
Air Blank		0.000	03:52	
Control Test		0.079	03:52	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 09/06/2018 14:37

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: ELBERT

DATE/TIME OF TEST: 08/08/2018 22:45:59

Diagnostics Check		OK	22:39
Air Blank		0.000	22:39
Control Test		0.079	22:40
Air Blank		0.000	22:40
Subject Sample	# 1	0.236	22:41
Breath Volume		1.382	22:41
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.226	22:44
Breath Volume		1.453	22:44
Air Blank		0.000	22:45
Control Test		0.080	22:45
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 09/06/2018 14:37

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHILLER,E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: BARRERA

DATE/TIME OF TEST: 08/09/2018 03:38:49

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.080	03:37	
Air Blank		03:38	RFI Detect
Air Blank	0.000	03:38	

Data Download Date/Time: 09/06/2018 14:37

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHILLER,E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: BARRERA

DATE/TIME OF TEST: 08/09/2018 03:47:24

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.080	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.118	03:43
Breath Volume		1.203	03:43
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.117	03:45
Breath Volume		1.144	03:45
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 09/06/2018 14:37

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 08/09/2018 21:30:59

Diagnostics Check		OK	21:23
Air Blank		0.000	21:23
Control Test		0.081	21:24
Air Blank		0.000	21:24
Subject Sample	# 1	0.176	21:25
Breath Volume		1.742	21:25
Air Blank		0.000	21:26
Air Blank		0.000	21:27
Subject Sample	# 2	0.178	21:29
Breath Volume		2.230	21:29
Air Blank		0.000	21:30
Control Test		0.080	21:30
Air Blank		0.000	21:30
Diagnostics Check		OK	21:30

Data Download Date/Time: 09/06/2018 14:37

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: WHERRY

DATE/TIME OF TEST: 08/10/2018 01:43:13

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.080	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.080	01:38
Breath Volume		1.937	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.078	01:41
Breath Volume		1.453	01:41
Air Blank		0.000	01:42
Control Test		0.080	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 09/06/2018 14:37

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SKIPPER,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: ADMIN

DATE/TIME OF TEST: 08/10/2018 10:55:12

Diagnostics Check		OK	10:52	
Air Blank		0.000	10:52	
Control Test		0.080	10:52	
Air Blank		0.000	10:53	
Subject Sample	# 1		10:53	Slope Not Met
Breath Volume		0.675	10:53	
Air Blank		0.000	10:54	
Control Test		0.079	10:54	
Air Blank		0.000	10:55	
Diagnostics Check		OK	10:55	

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SKIPPER,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: ADMIN

DATE/TIME OF TEST: 08/10/2018 11:03:01

Diagnostics Check		OK	10:56	
Air Blank		0.000	10:56	
Control Test		0.080	10:57	
Air Blank		0.000	10:57	
Subject Sample	# 1	0.330	10:59	
Breath Volume		1.132	10:59	
Air Blank		0.000	10:59	
Air Blank		0.000	11:01	
Subject Sample	# 2		11:01	Subject Test Refused
Breath Volume		0.000	11:01	
Air Blank		0.000	11:02	
Control Test		0.077	11:02	
Air Blank		0.000	11:02	
Diagnostics Check		OK	11:03	

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARRELL,Q
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:39

SUBJECT'S LAST NAME: MARTINEZ MARTINEZ

DATE/TIME OF TEST: 08/10/2018 17:06:52

Diagnostics Check		OK	17:00
Air Blank		0.000	17:00
Control Test		0.080	17:01
Air Blank		0.000	17:01
Subject Sample	# 1	0.310	17:02
Breath Volume		1.542	17:02
Air Blank		0.000	17:03
Air Blank		0.000	17:04
Subject Sample	# 2	0.321	17:05
Breath Volume		1.304	17:05
Air Blank		0.000	17:05
Control Test		0.079	17:06
Air Blank		0.000	17:06
Diagnostics Check		OK	17:06

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 08/11/2018 01:46:35

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.082	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.183	01:41
Breath Volume		1.792	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.197	01:44
Breath Volume		2.476	01:44
Air Blank		0.000	01:45
Control Test		0.079	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: DUHART

DATE/TIME OF TEST: 08/11/2018 02:50:51

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.080	02:43
Air Blank		0.000	02:43
Subject Sample	# 1	0.135	02:45
Breath Volume		2.167	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:47
Subject Sample	# 2	0.136	02:49
Breath Volume		2.785	02:49
Air Blank		0.000	02:49
Control Test		0.080	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHOLL,V
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: JESSUP

DATE/TIME OF TEST: 08/11/2018 04:36:18

Diagnostics Check		OK	04:27
Air Blank		0.000	04:27
Control Test		0.080	04:27
Air Blank		0.000	04:28
Subject Sample	# 1	0.092	04:30
Breath Volume		2.035	04:30
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.095	04:34
Breath Volume		2.285	04:34
Air Blank		0.000	04:35
Control Test		0.080	04:35
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: KITCHENER

DATE/TIME OF TEST: 08/11/2018 04:56:26

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.081	04:51
Air Blank		0.000	04:51
Subject Sample	# 1	0.000	04:52
Breath Volume		2.562	04:52
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.000	04:55
Breath Volume		2.601	04:55
Air Blank		0.000	04:55
Control Test		0.080	04:55
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: FLACH

DATE/TIME OF TEST: 08/11/2018 22:17:06

Diagnostics Check		OK	22:03	
Air Blank		0.000	22:04	
Control Test		0.080	22:04	
Air Blank		0.000	22:04	
Subject Sample	# 1	0.078	22:08	Volume Not Met
Breath Volume		0.234	22:08	
Air Blank		0.000	22:08	
Air Blank		0.000	22:10	
Subject Sample	# 2	0.105	22:11	
Breath Volume		1.132	22:11	
Air Blank		0.000	22:12	
Air Blank		0.000	22:14	
Subject Sample	# 3	0.105	22:15	
Breath Volume		1.113	22:15	
Air Blank		0.000	22:16	
Control Test		0.078	22:16	
Air Blank		0.000	22:17	
Diagnostics Check		OK	22:17	

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: ALMANZAR

DATE/TIME OF TEST: 08/12/2018 01:41:01

Diagnostics Check		OK	01:29	
Air Blank		0.000	01:29	
Control Test		0.080	01:30	
Air Blank		0.000	01:30	
Subject Sample	# 1	0.109	01:33	Volume Not Met
Breath Volume		0.625	01:33	
Air Blank		0.000	01:34	
Air Blank		0.000	01:36	
Subject Sample	# 2	0.126	01:39	Volume Not Met
Breath Volume		0.289	01:39	
Air Blank		0.000	01:40	
Control Test		0.079	01:40	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:41	

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: ALMANZAR

DATE/TIME OF TEST: 08/12/2018 01:53:20

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.120	01:46
Breath Volume		1.136	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:48
Subject Sample	# 2	0.127	01:51
Breath Volume		1.808	01:51
Air Blank		0.000	01:52
Control Test		0.078	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: STAGGS JR

DATE/TIME OF TEST: 08/12/2018 04:23:30

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:15	
Control Test		0.079	04:15	
Air Blank		0.000	04:16	
Subject Sample	# 1	0.119	04:16	
Breath Volume		1.484	04:16	
Air Blank		0.000	04:17	
Air Blank		0.000	04:19	
Subject Sample	# 2		04:22	Subject Test Refused
Breath Volume		0.000	04:22	
Air Blank		0.000	04:22	
Control Test		0.079	04:22	
Air Blank		0.000	04:23	
Diagnostics Check		OK	04:23	

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: DIBENARDO

DATE/TIME OF TEST: 08/13/2018 01:35:43

Diagnostics Check		OK	01:29
Air Blank		0.000	01:29
Control Test		0.080	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.078	01:31
Breath Volume		1.152	01:31
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.084	01:34
Breath Volume		3.136	01:34
Air Blank		0.000	01:34
Control Test		0.080	01:35
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 08/14/2018 00:25:41

Diagnostics Check		OK	00:17	
Air Blank		0.000	00:17	
Control Test		0.080	00:18	
Air Blank		0.000	00:18	
Subject Sample	# 1		00:21	No Sample Provided
Breath Volume		0.000	00:21	
Air Blank		0.000	00:22	
Air Blank		0.000	00:24	
Subject Sample	# 2		00:24	Subject Test Refused
Breath Volume		0.000	00:24	
Air Blank		0.000	00:24	
Control Test		0.079	00:25	
Air Blank		0.000	00:25	
Diagnostics Check		OK	00:25	

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCRUGGS,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: BAIZER

DATE/TIME OF TEST: 08/14/2018 21:40:26

Diagnostics Check		OK	21:33
Air Blank		0.000	21:34
Control Test		0.080	21:34
Air Blank		0.000	21:34
Subject Sample	# 1	0.077	21:35
Breath Volume		2.058	21:35
Air Blank		0.000	21:36
Air Blank		0.000	21:38
Subject Sample	# 2	0.073	21:38
Breath Volume		1.195	21:38
Air Blank		0.000	21:39
Control Test		0.080	21:39
Air Blank		0.000	21:40
Diagnostics Check		OK	21:40

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: MARTINEZ PINEDA

DATE/TIME OF TEST: 08/15/2018 23:15:49

Diagnostics Check		OK	23:09
Air Blank		0.000	23:09
Control Test		0.080	23:10
Air Blank		0.000	23:10
Subject Sample	# 1	0.166	23:11
Breath Volume		4.488	23:11
Air Blank		0.000	23:11
Air Blank		0.000	23:13
Subject Sample	# 2	0.171	23:14
Breath Volume		4.375	23:14
Air Blank		0.000	23:14
Control Test		0.080	23:15
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 08/16/2018 02:38:56

Diagnostics Check		OK	02:31
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.144	02:34
Breath Volume		2.210	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.134	02:37
Breath Volume		1.613	02:37
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: CIROLI

DATE/TIME OF TEST: 08/17/2018 01:49:51

Diagnostics Check		OK	01:41
Air Blank		0.000	01:42
Control Test		0.081	01:42
Air Blank		0.000	01:43
Subject Sample	# 1	0.000	01:45
Breath Volume		1.898	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.000	01:48
Breath Volume		1.433	01:48
Air Blank		0.000	01:48
Control Test		0.080	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: HAMILTON JR

DATE/TIME OF TEST: 08/17/2018 04:48:01

Diagnostics Check		OK	04:41
Air Blank		0.000	04:42
Control Test		0.081	04:42
Air Blank		0.000	04:42
Subject Sample	# 1	0.119	04:43
Breath Volume		5.402	04:43
Air Blank		0.000	04:44
Air Blank		0.000	04:45
Subject Sample	# 2	0.118	04:46
Breath Volume		4.785	04:46
Air Blank		0.000	04:47
Control Test		0.079	04:47
Air Blank		0.000	04:47
Diagnostics Check		OK	04:48

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: SALIVA

DATE/TIME OF TEST: 08/17/2018 20:14:06

Diagnostics Check		OK	20:01	
Air Blank		0.000	20:02	
Control Test		0.080	20:02	
Air Blank		0.000	20:03	
Subject Sample	# 1	0.112	20:06	Volume Not Met
Breath Volume		0.554	20:06	
Air Blank		0.000	20:06	
Air Blank		0.000	20:08	
Subject Sample	# 2	0.120	20:09	
Breath Volume		1.863	20:09	
Air Blank		0.000	20:10	
Air Blank		0.000	20:12	
Subject Sample	# 3	0.117	20:12	
Breath Volume		1.132	20:12	
Air Blank		0.000	20:13	
Control Test		0.079	20:13	
Air Blank		0.000	20:14	
Diagnostics Check		OK	20:14	

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUSTEK,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: BROUGHTON JR

DATE/TIME OF TEST: 08/18/2018 04:32:32

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.080	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.140	04:27
Breath Volume		1.316	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.137	04:30
Breath Volume		1.351	04:30
Air Blank		0.000	04:31
Control Test		0.079	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: ESTRADA ESCAMILLA

DATE/TIME OF TEST: 08/18/2018 05:48:18

Diagnostics Check		OK	05:41
Air Blank		0.000	05:41
Control Test		0.080	05:42
Air Blank		0.000	05:42
Subject Sample	# 1	0.082	05:43
Breath Volume		1.488	05:43
Air Blank		0.000	05:44
Air Blank		0.000	05:46
Subject Sample	# 2	0.086	05:46
Breath Volume		2.566	05:46
Air Blank		0.000	05:47
Control Test		0.081	05:47
Air Blank		0.000	05:48
Diagnostics Check		OK	05:48

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: CALIHUA-TEHUINTLE

DATE/TIME OF TEST: 08/18/2018 23:00:04

Diagnostics Check		OK	22:51
Air Blank		0.000	22:51
Control Test		0.081	22:52
Air Blank		0.000	22:52
Subject Sample	# 1	0.183	22:54
Breath Volume		1.425	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.186	22:58
Breath Volume		2.156	22:58
Air Blank		0.000	22:59
Control Test		0.079	22:59
Air Blank		0.000	22:59
Diagnostics Check		OK	23:00

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: BACA-ROSAS

DATE/TIME OF TEST: 08/19/2018 01:21:17

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.080	01:15
Air Blank		0.000	01:16
Subject Sample	# 1	0.168	01:16
Breath Volume		3.003	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.168	01:19
Breath Volume		3.000	01:19
Air Blank		0.000	01:20
Control Test		0.080	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 08/19/2018 03:28:30

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.080	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.206	03:23
Breath Volume		1.183	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.214	03:26
Breath Volume		2.105	03:26
Air Blank		0.000	03:27
Control Test		0.079	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: KOMARA

DATE/TIME OF TEST: 08/19/2018 04:11:51

Diagnostics Check		OK	04:01
Air Blank		0.000	04:02
Control Test		0.080	04:02
Air Blank		0.000	04:03
Subject Sample	# 1	0.158	04:05
Breath Volume		1.210	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:07
Subject Sample	# 2	0.150	04:10
Breath Volume		1.367	04:10
Air Blank		0.000	04:10
Control Test		0.078	04:11
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: WAGGONER III

DATE/TIME OF TEST: 08/19/2018 21:57:50

Diagnostics Check		OK	21:50
Air Blank		0.000	21:51
Control Test		0.080	21:51
Air Blank		0.000	21:51
Subject Sample	# 1	0.229	21:53
Breath Volume		1.519	21:53
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.215	21:56
Breath Volume		2.363	21:56
Air Blank		0.000	21:56
Control Test		0.079	21:57
Air Blank		0.000	21:57
Diagnostics Check		OK	21:57

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIERRA,P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: GONZALEZ LA PARRA

DATE/TIME OF TEST: 08/20/2018 04:15:40

Diagnostics Check		OK	04:08
Air Blank		0.000	04:09
Control Test		0.080	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.195	04:10
Breath Volume		1.550	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.183	04:14
Breath Volume		2.511	04:14
Air Blank		0.000	04:14
Control Test		0.079	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: NATZEL

DATE/TIME OF TEST: 08/20/2018 23:16:03

Diagnostics Check		OK	23:08
Air Blank		0.000	23:09
Control Test		0.081	23:09
Air Blank		0.000	23:10
Subject Sample	# 1	0.130	23:11
Breath Volume		3.636	23:11
Air Blank		0.000	23:12
Air Blank		0.000	23:13
Subject Sample	# 2	0.119	23:14
Breath Volume		1.617	23:14
Air Blank		0.000	23:15
Control Test		0.080	23:15
Air Blank		0.000	23:15
Diagnostics Check		OK	23:16

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: DIAZ RODRIGUEZ

DATE/TIME OF TEST: 08/20/2018 23:37:03

Diagnostics Check		OK	23:30
Air Blank		0.000	23:31
Control Test		0.080	23:31
Air Blank		0.000	23:32
Subject Sample	# 1	0.137	23:32
Breath Volume		2.109	23:32
Air Blank		0.000	23:33
Air Blank		0.000	23:34
Subject Sample	# 2	0.137	23:35
Breath Volume		2.781	23:35
Air Blank		0.000	23:36
Control Test		0.079	23:36
Air Blank		0.000	23:36
Diagnostics Check		OK	23:37

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: DAUGHTRY

DATE/TIME OF TEST: 08/22/2018 02:16:50

Diagnostics Check		OK	02:09
Air Blank		0.000	02:10
Control Test		0.080	02:10
Air Blank		0.000	02:10
Subject Sample	# 1	0.177	02:11
Breath Volume		1.363	02:11
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.173	02:15
Breath Volume		1.917	02:15
Air Blank		0.000	02:15
Control Test		0.080	02:16
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: KILPPEL

DATE/TIME OF TEST: 08/22/2018 03:24:37

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.276	03:19
Breath Volume		1.839	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.281	03:22
Breath Volume		1.550	03:22
Air Blank		0.000	03:23
Control Test		0.077	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: PEPPER- MENDOSA

DATE/TIME OF TEST: 08/22/2018 05:12:50

Diagnostics Check		OK	04:57	
Air Blank		0.000	04:58	
Control Test		0.078	04:58	
Air Blank		0.000	04:59	
Subject Sample	# 1	0.076	05:02	Volume Not Met
Breath Volume		0.367	05:02	
Air Blank		0.000	05:02	
Air Blank		0.000	05:04	
Subject Sample	# 2	0.111	05:07	
Breath Volume		2.480	05:07	
Air Blank		0.000	05:08	
Air Blank		0.000	05:09	
Subject Sample	# 3	0.115	05:11	
Breath Volume		2.605	05:11	
Air Blank		0.000	05:11	
Control Test		0.078	05:12	
Air Blank		0.000	05:12	
Diagnostics Check		OK	05:12	

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: BOWERMAN

DATE/TIME OF TEST: 08/22/2018 23:42:44

Diagnostics Check		OK	23:35
Air Blank		0.000	23:36
Control Test		0.080	23:36
Air Blank		0.000	23:37
Subject Sample	# 1	0.104	23:37
Breath Volume		1.933	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.103	23:41
Breath Volume		1.496	23:41
Air Blank		0.000	23:41
Control Test		0.079	23:42
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: TREVINO

DATE/TIME OF TEST: 08/24/2018 02:05:46

Diagnostics Check		OK	01:57
Air Blank		0.000	01:57
Control Test		0.081	01:58
Air Blank		0.000	01:58
Subject Sample	# 1	0.260	01:59
Breath Volume		2.109	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.246	02:04
Breath Volume		1.835	02:04
Air Blank		0.000	02:04
Control Test		0.080	02:05
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: VAIL

DATE/TIME OF TEST: 08/24/2018 03:33:05

Diagnostics Check		OK	03:26
Air Blank		0.000	03:27
Control Test		0.080	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.268	03:28
Breath Volume		2.527	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:30
Subject Sample	# 2	0.277	03:31
Breath Volume		2.175	03:31
Air Blank		0.000	03:32
Control Test		0.079	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHILLER,E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 08/24/2018 06:03:00

Diagnostics Check		OK	05:57
Air Blank		0.000	05:57
Control Test		0.080	05:57
Air Blank		0.000	05:58
Subject Sample	# 1	0.000	05:58
Breath Volume		1.644	05:58
Air Blank		0.000	05:59
Air Blank		0.000	06:01
Subject Sample	# 2	0.000	06:01
Breath Volume		2.042	06:01
Air Blank		0.000	06:02
Control Test		0.080	06:02
Air Blank		0.000	06:02
Diagnostics Check		OK	06:03

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: CASTILLO OCHOA

DATE/TIME OF TEST: 08/24/2018 23:36:13

Diagnostics Check		OK	23:29
Air Blank		0.000	23:29
Control Test		0.080	23:30
Air Blank		0.000	23:30
Subject Sample	# 1	0.149	23:31
Breath Volume		2.414	23:31
Air Blank		0.000	23:32
Air Blank		0.000	23:34
Subject Sample	# 2	0.147	23:34
Breath Volume		2.929	23:34
Air Blank		0.000	23:35
Control Test		0.080	23:35
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: COHEN II

DATE/TIME OF TEST: 08/25/2018 03:51:38

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.080	03:45
Air Blank		0.000	03:46
Subject Sample	# 1	0.179	03:46
Breath Volume		4.968	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.186	03:50
Breath Volume		4.691	03:50
Air Blank		0.000	03:50
Control Test		0.079	03:51
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: ROSARIO-SANCHEZ

DATE/TIME OF TEST: 08/25/2018 04:51:51

Diagnostics Check		OK	04:44
Air Blank		0.000	04:44
Control Test		0.079	04:45
Air Blank		0.000	04:45
Subject Sample	# 1	0.140	04:47
Breath Volume		1.187	04:47
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.138	04:50
Breath Volume		2.203	04:50
Air Blank		0.000	04:50
Control Test		0.079	04:51
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CANETE-VILLAFRANCA,I
ARRESTING OFFICER AGENCY: Tampa Airport PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:47

SUBJECT'S LAST NAME: MACKENZIE

DATE/TIME OF TEST: 08/25/2018 14:16:12

Diagnostics Check		OK	14:09
Air Blank		0.000	14:09
Control Test		0.081	14:10
Air Blank		0.000	14:10
Subject Sample	# 1	0.237	14:11
Breath Volume		2.941	14:11
Air Blank		0.000	14:12
Air Blank		0.000	14:13
Subject Sample	# 2	0.232	14:14
Breath Volume		2.605	14:14
Air Blank		0.000	14:15
Control Test		0.079	14:15
Air Blank		0.000	14:16
Diagnostics Check		OK	14:16

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: KIDDER

DATE/TIME OF TEST: 08/26/2018 03:05:29

Diagnostics Check		OK	02:57
Air Blank		0.000	02:58
Control Test		0.081	02:58
Air Blank		0.000	02:59
Subject Sample	# 1	0.210	03:00
Breath Volume		1.527	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.202	03:03
Breath Volume		1.238	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: ALHARBI

DATE/TIME OF TEST: 08/26/2018 04:12:19

Diagnostics Check		OK	04:09	
Air Blank		0.000	04:09	
Control Test		0.079	04:10	
Air Blank		0.000	04:10	
Subject Sample	# 1		04:10	Subject Test Refused
Breath Volume		0.000	04:10	
Air Blank		0.000	04:11	
Control Test		0.079	04:11	
Air Blank		0.000	04:12	
Diagnostics Check		OK	04:12	

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: JIMINEZ-GOMEZ

DATE/TIME OF TEST: 08/26/2018 04:46:11

Diagnostics Check		OK	04:34	
Air Blank		0.000	04:34	
Control Test		0.080	04:35	
Air Blank		0.000	04:35	
Subject Sample	# 1	0.136	04:38	Volume Not Met
Breath Volume		0.238	04:38	
Air Blank		0.000	04:39	
Air Blank		0.000	04:41	
Subject Sample	# 2	0.124	04:44	Volume Not Met
Breath Volume		0.335	04:44	
Air Blank		0.000	04:45	
Control Test		0.079	04:45	
Air Blank		0.000	04:46	
Diagnostics Check		OK	04:46	

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: JIMENEZ-GOMEZ

DATE/TIME OF TEST: 08/26/2018 05:00:41

Diagnostics Check		OK	04:53
Air Blank		0.000	04:54
Control Test		0.078	04:54
Air Blank		0.000	04:55
Subject Sample	# 1	0.155	04:56
Breath Volume		1.554	04:56
Air Blank		0.000	04:56
Air Blank		0.000	04:58
Subject Sample	# 2	0.143	04:59
Breath Volume		1.160	04:59
Air Blank		0.000	04:59
Control Test		0.078	05:00
Air Blank		0.000	05:00
Diagnostics Check		OK	05:00

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:34

SUBJECT'S LAST NAME: MARIN

DATE/TIME OF TEST: 08/26/2018 07:02:14

Diagnostics Check		OK	06:55
Air Blank		0.000	06:56
Control Test		0.079	06:56
Air Blank		0.000	06:57
Subject Sample	# 1	0.126	06:57
Breath Volume		2.886	06:57
Air Blank		0.000	06:58
Air Blank		0.000	07:00
Subject Sample	# 2	0.129	07:00
Breath Volume		2.750	07:00
Air Blank		0.000	07:01
Control Test		0.079	07:01
Air Blank		0.000	07:02
Diagnostics Check		OK	07:02

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: DALLAS

DATE/TIME OF TEST: 08/27/2018 01:10:02

Diagnostics Check		OK	01:03
Air Blank		0.000	01:03
Control Test		0.080	01:04
Air Blank		0.000	01:04
Subject Sample	# 1	0.038	01:05
Breath Volume		3.789	01:05
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.039	01:08
Breath Volume		3.769	01:08
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:10

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: NENTON

DATE/TIME OF TEST: 08/27/2018 20:06:52

Diagnostics Check		OK	20:00
Air Blank		0.000	20:01
Control Test		0.080	20:01
Air Blank		0.000	20:01
Subject Sample	# 1	0.195	20:02
Breath Volume		2.679	20:02
Air Blank		0.000	20:03
Air Blank		0.000	20:04
Subject Sample	# 2	0.193	20:05
Breath Volume		2.058	20:05
Air Blank		0.000	20:05
Control Test		0.079	20:06
Air Blank		0.000	20:06
Diagnostics Check		OK	20:06

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 08/28/2018 00:59:04

Diagnostics Check		OK	00:53
Air Blank		0.000	00:53
Control Test		0.080	00:53
Air Blank		0.000	00:54
Subject Sample	# 1	0.245	00:54
Breath Volume		1.457	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.244	00:57
Breath Volume		1.757	00:57
Air Blank		0.000	00:58
Control Test		0.080	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:59

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/28/2018 04:10:50

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:02	
Control Test		0.080	04:03	
Air Blank		0.000	04:03	
Subject Sample	# 1	0.164	04:06	Volume Not Met
Breath Volume		0.703	04:06	
Air Blank		0.000	04:07	
Air Blank		0.000	04:09	
Subject Sample	# 2		04:09	Subject Test Refused
Breath Volume		0.000	04:09	
Air Blank		0.000	04:09	
Control Test		0.078	04:10	
Air Blank		0.000	04:10	
Diagnostics Check		OK	04:10	

Data Download Date/Time: 09/06/2018 14:38

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: MULDREW

DATE/TIME OF TEST: 08/29/2018 01:26:30

Diagnostics Check		OK	01:19
Air Blank		0.000	01:20
Control Test		0.081	01:20
Air Blank		0.000	01:21
Subject Sample	# 1	0.000	01:22
Breath Volume		2.636	01:22
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.000	01:25
Breath Volume		1.550	01:25
Air Blank		0.000	01:25
Control Test		0.080	01:25
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 09/06/2018 14:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: LUGO-FELIX

DATE/TIME OF TEST: 08/29/2018 04:05:52

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:54	
Control Test		0.080	03:54	
Air Blank		0.000	03:55	
Subject Sample	# 1	0.149	03:58	Volume Not Met
Breath Volume		0.000	03:58	
Air Blank		0.000	03:58	
Air Blank		0.000	04:00	
Subject Sample	# 2	0.144	04:01	
Breath Volume		1.984	04:01	
Air Blank		0.000	04:01	
Air Blank		0.000	04:03	
Subject Sample	# 3	0.139	04:04	
Breath Volume		1.527	04:04	
Air Blank		0.000	04:04	
Control Test		0.080	04:05	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:05	

Data Download Date/Time: 09/06/2018 14:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: WAGGONER III

DATE/TIME OF TEST: 08/29/2018 20:40:35

Diagnostics Check		OK	20:34
Air Blank		0.000	20:34
Control Test		0.080	20:35
Air Blank		0.000	20:35
Subject Sample	# 1	0.192	20:36
Breath Volume		2.425	20:36
Air Blank		0.000	20:36
Air Blank		0.000	20:38
Subject Sample	# 2	0.191	20:38
Breath Volume		1.808	20:38
Air Blank		0.000	20:39
Control Test		0.080	20:39
Air Blank		0.000	20:40
Diagnostics Check		OK	20:40

Data Download Date/Time: 09/06/2018 14:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: BROWN JR

DATE/TIME OF TEST: 08/30/2018 00:41:56

Diagnostics Check		OK	00:34
Air Blank		0.000	00:34
Control Test		0.081	00:35
Air Blank		0.000	00:35
Subject Sample	# 1	0.139	00:37
Breath Volume		3.316	00:37
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.142	00:40
Breath Volume		3.507	00:40
Air Blank		0.000	00:41
Control Test		0.079	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 09/06/2018 14:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: TAPIA-CRUZ

DATE/TIME OF TEST: 08/30/2018 20:40:26

Diagnostics Check		OK	20:33
Air Blank		0.000	20:34
Control Test		0.079	20:34
Air Blank		0.000	20:35
Subject Sample	# 1	0.153	20:35
Breath Volume		1.484	20:35
Air Blank		0.000	20:36
Air Blank		0.000	20:38
Subject Sample	# 2	0.158	20:38
Breath Volume		1.273	20:38
Air Blank		0.000	20:39
Control Test		0.080	20:39
Air Blank		0.000	20:40
Diagnostics Check		OK	20:40

Data Download Date/Time: 09/06/2018 14:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 08/31/2018 02:04:19

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.080	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.182	01:59
Breath Volume		2.578	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.180	02:02
Breath Volume		2.656	02:02
Air Blank		0.000	02:03
Control Test		0.080	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 09/06/2018 14:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRAL,JONATHAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SABO,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 08/05/2018 02:45:01

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
		02:44	RFI Detect
Air Blank		02:45	RFI Detect

Data Download Date/Time: 08/17/2018 17:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRAL,JONATHAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SABO,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 08/05/2018 02:48:53

Diagnostics Check	OK	02:48	
Air Blank		02:48	RFI Detect
Air Blank		02:48	RFI Detect

Data Download Date/Time: 08/17/2018 17:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRAL,JONATHAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SABO,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 08/05/2018 02:52:05

Diagnostics Check	OK	02:51	
Air Blank		02:51	RFI Detect
Air Blank	0.000	02:52	

Data Download Date/Time: 08/17/2018 17:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRAL,JONATHAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SABO,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 08/05/2018 03:03:54

Diagnostics Check	OK	03:03	
Air Blank		03:03	RFI Detect
Air Blank		03:03	RFI Detect

Data Download Date/Time: 08/17/2018 17:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRAL,JONATHAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MURPHY,JAVON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: WHITEHEAD

DATE/TIME OF TEST: 08/20/2018 04:16:50

Diagnostics Check		OK	04:13	
Air Blank		0.000	04:14	
Control Test		0.078	04:14	
Air Blank		0.000	04:15	
Subject Sample	# 1		04:15	Subject Test Refused
Breath Volume		0.000	04:15	
Air Blank		0.000	04:15	
Control Test		0.078	04:16	
Air Blank		0.000	04:16	
Diagnostics Check		OK	04:16	

Data Download Date/Time: 09/25/2018 15:10

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTAN,FRANK,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: QUINTANA,FRANK
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: PASCUAL

DATE/TIME OF TEST: 08/11/2018 23:07:18

Diagnostics Check		OK	23:00
Air Blank		0.000	23:01
Control Test		0.081	23:01
Air Blank		0.000	23:02
Subject Sample	# 1	0.216	23:02
Breath Volume		2.410	23:02
Air Blank		0.000	23:03
Air Blank		0.000	23:04
Subject Sample	# 2	0.214	23:05
Breath Volume		1.828	23:05
Air Blank		0.000	23:06
Control Test		0.080	23:06
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 08/30/2018 10:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/18/2018 01:10:42

Diagnostics Check		OK	01:04
Air Blank		0.000	01:05
Control Test		0.079	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.142	01:06
Breath Volume		1.789	01:06
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.137	01:09
Breath Volume		1.355	01:09
Air Blank		0.000	01:09
Control Test		0.078	01:10
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 08/29/2018 15:13

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: BARRICK

DATE/TIME OF TEST: 08/19/2018 17:02:14

Diagnostics Check		OK	16:56
Air Blank		0.000	16:56
Control Test		0.079	16:56
Air Blank		0.000	16:57
Subject Sample	# 1	0.219	16:57
Breath Volume		1.769	16:57
Air Blank		0.000	16:58
Air Blank		0.000	17:00
Subject Sample	# 2	0.222	17:00
Breath Volume		2.011	17:00
Air Blank		0.000	17:01
Control Test		0.080	17:01
Air Blank		0.000	17:02
Diagnostics Check		OK	17:02

Data Download Date/Time: 08/29/2018 15:13

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: STEER

DATE/TIME OF TEST: 08/19/2018 21:24:45

Diagnostics Check		OK	21:18
Air Blank		0.000	21:18
Control Test		0.078	21:19
Air Blank		0.000	21:19
Subject Sample	# 1	0.198	21:20
Breath Volume		2.460	21:20
Air Blank		0.000	21:20
Air Blank		0.000	21:22
Subject Sample	# 2	0.196	21:23
Breath Volume		2.714	21:23
Air Blank		0.000	21:23
Control Test		0.080	21:24
Air Blank		0.000	21:24
Diagnostics Check		OK	21:24

Data Download Date/Time: 08/29/2018 15:13

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/16/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 08/03/2018 22:32:47

Diagnostics Check		OK	22:29	
Air Blank		0.000	22:30	
Control Test		0.080	22:30	
Air Blank		0.000	22:31	
Subject Sample	# 1		22:31	Subject Test Refused
Breath Volume		0.000	22:31	
Air Blank		0.000	22:31	
Control Test		0.080	22:32	
Air Blank		0.000	22:32	
Diagnostics Check		OK	22:32	

Data Download Date/Time: 08/20/2018 10:58

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 08/03/2018 22:58:05

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:55	
Control Test		0.079	22:56	
Air Blank		0.000	22:56	
Subject Sample	# 1		22:56	Subject Test Refused
Breath Volume		0.000	22:56	
Air Blank		0.000	22:57	
Control Test		0.080	22:57	
Air Blank		0.000	22:57	
Diagnostics Check		OK	22:58	

Data Download Date/Time: 08/20/2018 10:58

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ANDUJAR

DATE/TIME OF TEST: 08/03/2018 23:07:32

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.081	23:02
Air Blank		0.000	23:02
Subject Sample	# 1	0.130	23:03
Breath Volume		2.480	23:03
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.134	23:05
Breath Volume		2.542	23:05
Air Blank		0.000	23:06
Control Test		0.080	23:06
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 08/20/2018 10:58

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 08/04/2018 22:10:45

Diagnostics Check		OK	22:07	
Air Blank		0.000	22:08	
Control Test		0.080	22:08	
Air Blank		0.000	22:09	
Subject Sample	# 1		22:09	Subject Test Refused
Breath Volume		0.000	22:09	
Air Blank		0.000	22:09	
Control Test		0.079	22:10	
Air Blank		0.000	22:10	
Diagnostics Check		OK	22:10	

Data Download Date/Time: 08/20/2018 10:58

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: DYE

DATE/TIME OF TEST: 08/08/2018 16:40:43

Diagnostics Check		OK	16:34
Air Blank		0.000	16:35
Control Test		0.079	16:35
Air Blank		0.000	16:36
Subject Sample	# 1	0.000	16:36
Breath Volume		2.125	16:36
Air Blank		0.000	16:36
Air Blank		0.000	16:38
Subject Sample	# 2	0.000	16:39
Breath Volume		1.960	16:39
Air Blank		0.000	16:39
Control Test		0.079	16:40
Air Blank		0.000	16:40
Diagnostics Check		OK	16:40

Data Download Date/Time: 08/20/2018 10:58

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MOSS,JUSTIN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: CASAREZ

DATE/TIME OF TEST: 08/12/2018 20:40:32

Diagnostics Check		OK	20:28	
Air Blank		0.000	20:29	
Control Test		0.079	20:29	
Air Blank		0.000	20:29	
Subject Sample	# 1	0.263	20:30	
Breath Volume		3.890	20:30	
Air Blank		0.000	20:30	
Air Blank		0.000	20:32	
Subject Sample	# 2		20:35	No Sample Provided
Breath Volume		0.000	20:35	
Air Blank		0.000	20:36	
Air Blank		0.000	20:38	
Subject Sample	# 3	0.246	20:38	
Breath Volume		3.246	20:38	
Air Blank		0.000	20:39	
Control Test		0.080	20:39	
Air Blank		0.000	20:40	
Diagnostics Check		OK	20:40	

Data Download Date/Time: 08/20/2018 10:58

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 08/18/2018 02:19:37

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.080	02:14
Air Blank		0.000	02:14
Subject Sample	# 1	0.136	02:15
Breath Volume		2.894	02:15
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.133	02:18
Breath Volume		3.097	02:18
Air Blank		0.000	02:18
Control Test		0.079	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 08/20/2018 10:58

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: STROUP

DATE/TIME OF TEST: 08/19/2018 02:47:23

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.079	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.111	02:42
Breath Volume		2.128	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.116	02:45
Breath Volume		1.851	02:45
Air Blank		0.000	02:46
Control Test		0.080	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 08/20/2018 10:58

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: DENTON

DATE/TIME OF TEST: 08/19/2018 03:50:22

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.079	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.091	03:46
Breath Volume		2.894	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.092	03:48
Breath Volume		2.003	03:48
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 08/20/2018 10:58

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 08/19/2018 04:27:54

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.080	04:20
Air Blank		0.000	04:21
Subject Sample	# 1	0.139	04:22
Breath Volume		2.035	04:22
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.141	04:26
Breath Volume		1.925	04:26
Air Blank		0.000	04:26
Control Test		0.079	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 08/20/2018 10:58

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: BLACKBURN,MAX
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: LIEBISCH

DATE/TIME OF TEST: 08/26/2018 04:08:35

Diagnostics Check		OK	03:57	
Air Blank		0.000	03:57	
Control Test		0.079	03:57	
Air Blank		0.000	03:58	
Subject Sample	# 1		04:01	No Sample Provided
Breath Volume		0.000	04:01	
Air Blank		0.000	04:02	
Air Blank		0.000	04:04	
Subject Sample	# 2		04:07	No Sample Provided
Breath Volume		0.000	04:07	
Air Blank		0.000	04:07	
Control Test		0.080	04:07	
Air Blank		0.000	04:08	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 09/24/2018 11:52

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 08/15/2018 00:45:33

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.079	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.191	00:41
Breath Volume		1.972	00:41
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.193	00:44
Breath Volume		1.613	00:44
Air Blank		0.000	00:44
Control Test		0.077	00:44
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 09/06/2018 23:26

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BEYDA

DATE/TIME OF TEST: 08/18/2018 22:03:35

Diagnostics Check		OK	21:56
Air Blank		0.000	21:56
Control Test		0.078	21:56
Air Blank		0.000	21:57
Subject Sample	# 1	0.169	21:58
Breath Volume		1.703	21:58
Air Blank		0.000	21:59
Air Blank		0.000	22:01
Subject Sample	# 2	0.151	22:02
Breath Volume		1.722	22:02
Air Blank		0.000	22:02
Control Test		0.077	22:02
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 09/06/2018 23:26

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: REAL-ALOMAR

DATE/TIME OF TEST: 08/19/2018 03:21:51

Diagnostics Check		OK	03:14
Air Blank		0.000	03:15
Control Test		0.078	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.158	03:17
Breath Volume		2.371	03:17
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.150	03:20
Breath Volume		1.875	03:20
Air Blank		0.000	03:20
Control Test		0.075	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 09/06/2018 23:26

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: RAMOS-MARTINEZ

DATE/TIME OF TEST: 08/31/2018 22:17:42

Diagnostics Check		OK	22:09
Air Blank		0.000	22:10
Control Test		0.077	22:10
Air Blank		0.000	22:11
Subject Sample	# 1	0.180	22:11
Breath Volume		1.355	22:11
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.178	22:16
Breath Volume		1.285	22:16
Air Blank		0.000	22:16
Control Test		0.076	22:17
Air Blank		0.000	22:17
Diagnostics Check		OK	22:17

Data Download Date/Time: 09/06/2018 23:26

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUELLO,JAVIER,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: PUELLO,JAVIER
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: STROTHER

DATE/TIME OF TEST: 08/01/2018 02:53:23

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:40	
Control Test		0.080	02:40	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:44	No Sample Provided
Breath Volume		0.000	02:44	
Air Blank		0.000	02:45	
Air Blank		0.000	02:47	
Subject Sample	# 2	0.196	02:48	
Breath Volume		2.425	02:48	
Air Blank		0.000	02:49	
Air Blank		0.000	02:51	
Subject Sample	# 3	0.181	02:51	
Breath Volume		1.898	02:51	
Air Blank		0.000	02:52	
Control Test		0.074	02:52	Control Outside Tolerance
Air Blank		0.000	02:53	

Data Download Date/Time: 08/02/2018 09:00

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DAVID,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CAIN,DAVID
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: RENNA

DATE/TIME OF TEST: 08/01/2018 19:58:36

Diagnostics Check		OK	19:50
Air Blank		0.000	19:51
Control Test		0.080	19:51
Air Blank		0.000	19:52
Subject Sample	# 1	0.000	19:54
Breath Volume		2.246	19:54
Air Blank		0.000	19:54
Air Blank		0.000	19:56
Subject Sample	# 2	0.000	19:57
Breath Volume		2.695	19:57
Air Blank		0.000	19:57
Control Test		0.081	19:58
Air Blank		0.000	19:58
Diagnostics Check		OK	19:58

Data Download Date/Time: 08/02/2018 09:00

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNHAM,J,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURMHAM,J
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: ENGLISH

DATE/TIME OF TEST: 08/01/2018 20:55:36

Diagnostics Check		OK	20:48
Air Blank		0.000	20:49
Control Test		0.081	20:49
Air Blank		0.000	20:49
Subject Sample	# 1	0.170	20:51
Breath Volume		4.746	20:51
Air Blank		0.000	20:51
Air Blank		0.000	20:53
Subject Sample	# 2	0.163	20:54
Breath Volume		3.523	20:54
Air Blank		0.000	20:54
Control Test		0.080	20:55
Air Blank		0.000	20:55
Diagnostics Check		OK	20:55

Data Download Date/Time: 08/02/2018 09:00

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,ROBERT,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NAYLOR,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 08/02/2018 01:55:20

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:52	
Control Test		0.080	01:52	
Air Blank		0.000	01:53	
Subject Sample	# 1		01:53	Subject Test Refused
Breath Volume		0.000	01:53	
Air Blank		0.000	01:54	
Control Test		0.080	01:54	
Air Blank		0.000	01:55	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 08/02/2018 09:00

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIFER,JUSTIN,S
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: PIFER,JUSTIN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: BUI
ARREST TIME: 16:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:53

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 08/02/2018 17:20:46

Diagnostics Check		OK	17:14
Air Blank		0.000	17:14
Control Test		0.080	17:15
Air Blank		0.000	17:15
Subject Sample	# 1	0.143	17:16
Breath Volume		4.523	17:16
Air Blank		0.000	17:16
Air Blank		0.000	17:18
Subject Sample	# 2	0.140	17:19
Breath Volume		4.359	17:19
Air Blank		0.000	17:19
Control Test		0.079	17:20
Air Blank		0.000	17:20
Diagnostics Check		OK	17:20

Data Download Date/Time: 09/27/2018 09:07

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUELLO,JAVIER,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: PUELLO,JAVIER
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: PATRICK

DATE/TIME OF TEST: 08/02/2018 22:43:55

Diagnostics Check		OK	22:41	
Air Blank		0.000	22:41	
Control Test		0.081	22:41	
Air Blank		0.000	22:42	
Subject Sample	# 1		22:42	Subject Test Refused
Breath Volume		0.000	22:42	
Air Blank		0.000	22:43	
Control Test		0.080	22:43	
Air Blank		0.000	22:43	
Diagnostics Check		OK	22:43	

Data Download Date/Time: 09/27/2018 09:07

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 08/03/2018 22:14:52

Diagnostics Check		OK	22:08
Air Blank		0.000	22:08
Control Test		0.080	22:09
Air Blank		0.000	22:09
Subject Sample	# 1	0.124	22:10
Breath Volume		2.589	22:10
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.126	22:13
Breath Volume		2.023	22:13
Air Blank		0.000	22:13
Control Test		0.080	22:14
Air Blank		0.000	22:14
Diagnostics Check		OK	22:14

Data Download Date/Time: 09/27/2018 09:07

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/04/2018 04:07:31

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:02	
Control Test		0.080	04:02	
Air Blank		0.000	04:03	
Subject Sample	# 1		04:05	Subject Test Refused
Breath Volume		0.273	04:05	
Air Blank		0.000	04:06	
Control Test		0.081	04:06	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 09/27/2018 09:07

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,ROYCE,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BROOKS,ROYCE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/04/2018 04:46:16

Diagnostics Check		OK	04:37	
Air Blank		0.000	04:38	
Control Test		0.079	04:38	
Air Blank		0.000	04:39	
Subject Sample	# 1		04:42	No Sample Provided
Breath Volume		0.000	04:42	
Air Blank		0.000	04:42	
Air Blank		0.000	04:44	
Subject Sample	# 2		04:44	Subject Test Refused
Breath Volume		0.000	04:44	
Air Blank		0.000	04:45	
Control Test		0.080	04:45	
Air Blank		0.000	04:46	
Diagnostics Check		OK	04:46	

Data Download Date/Time: 09/27/2018 09:07

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILIAMS,LANCE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,LANCE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: DORE

DATE/TIME OF TEST: 08/05/2018 06:02:49

Diagnostics Check		OK	05:51	
Air Blank		0.000	05:51	
Control Test		0.080	05:51	
Air Blank		0.000	05:52	
Subject Sample	# 1		05:55	No Sample Provided
Breath Volume		0.000	05:55	
Air Blank		0.000	05:56	
Air Blank		0.000	05:58	
Subject Sample	# 2	0.140	06:01	Volume Not Met
Breath Volume		0.152	06:01	
Air Blank		0.000	06:01	
Control Test		0.078	06:02	
Air Blank		0.000	06:02	
Diagnostics Check		OK	06:02	

Data Download Date/Time: 09/27/2018 09:07

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,LANCE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMSL,LANCE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: D600060685900

DATE/TIME OF TEST: 08/05/2018 06:40:20

Diagnostics Check		OK	06:32	
Air Blank		0.000	06:33	
Control Test		0.079	06:33	
Air Blank		0.000	06:33	
Subject Sample	# 1	0.125	06:37	Volume Not Met
Breath Volume		0.015	06:37	
Air Blank		0.000	06:37	
Air Blank			06:39	Ambient Fail
Air Blank		0.000	06:40	

Data Download Date/Time: 09/27/2018 09:07

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,RUSSELL,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COOK,RUSSELL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: GIBBS

DATE/TIME OF TEST: 08/07/2018 21:11:39

Diagnostics Check		OK	20:56	
Air Blank		0.000	20:57	
Control Test		0.080	20:57	
Air Blank		0.000	20:58	
Subject Sample	# 1	0.161	20:58	
Breath Volume		1.230	20:58	
Air Blank		0.000	20:59	
Air Blank		0.000	21:01	
Subject Sample	# 2	0.132	21:04	Volume Not Met
Breath Volume		0.285	21:04	
Air Blank		0.000	21:05	
Air Blank		0.000	21:06	
Subject Sample	# 3	0.130	21:09	Volume Not Met
Breath Volume		1.027	21:09	
Air Blank		0.000	21:10	
Control Test		0.078	21:11	
Air Blank		0.000	21:11	
Diagnostics Check		OK	21:11	

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LECOCQ,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: BIRCHFIELD

DATE/TIME OF TEST: 08/10/2018 03:21:48

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:19	
Control Test		0.080	03:19	
Air Blank		0.000	03:19	
Subject Sample	# 1		03:20	Subject Test Refused
Breath Volume		0.000	03:20	
Air Blank		0.000	03:20	
Control Test		0.081	03:21	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORTON,JAMES,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MONARCH,BRIAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: GUPTON

DATE/TIME OF TEST: 08/11/2018 02:57:08

Diagnostics Check		OK	02:45	
Air Blank		0.000	02:46	
Control Test		0.081	02:46	
Air Blank		0.000	02:47	
Subject Sample	# 1		02:50	No Sample Provided
Breath Volume		0.000	02:50	
Air Blank		0.000	02:50	
Air Blank		0.000	02:52	
Subject Sample	# 2		02:55	No Sample Provided
Breath Volume		0.000	02:55	
Air Blank		0.000	02:56	
Control Test		0.080	02:56	
Air Blank		0.000	02:57	
Diagnostics Check		OK	02:57	

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORTON,JAMES,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAMILTON,RAY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: WELLBORN

DATE/TIME OF TEST: 08/12/2018 00:49:14

Diagnostics Check		OK	00:37	
Air Blank		0.000	00:38	
Control Test		0.080	00:38	
Air Blank		0.000	00:39	
Subject Sample	# 1		00:42	No Sample Provided
Breath Volume		0.000	00:42	
Air Blank		0.000	00:42	
Air Blank		0.000	00:44	
Subject Sample	# 2		00:47	No Sample Provided
Breath Volume		0.000	00:47	
Air Blank		0.000	00:48	
Control Test		0.080	00:48	
Air Blank		0.000	00:49	
Diagnostics Check		OK	00:49	

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNHAM,J,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURNHAM,J
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: LETONA

DATE/TIME OF TEST: 08/12/2018 19:22:45

Diagnostics Check		OK	19:14	
Air Blank		0.000	19:14	
Control Test		0.081	19:14	
Air Blank		0.000	19:15	
Subject Sample	# 1		19:18	No Sample Provided
Breath Volume		0.000	19:18	
Air Blank		0.000	19:19	
Air Blank		0.000	19:21	
Subject Sample	# 2		19:21	Subject Test Refused
Breath Volume		0.000	19:21	
Air Blank		0.000	19:21	
Control Test		0.080	19:22	
Air Blank		0.000	19:22	
Diagnostics Check		OK	19:22	

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILDE,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: RUNYEN

DATE/TIME OF TEST: 08/14/2018 02:55:15

Diagnostics Check		OK	02:46
Air Blank		0.000	02:47
Control Test		0.080	02:47
Air Blank		0.000	02:48
Subject Sample	# 1	0.176	02:50
Breath Volume		2.097	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.172	02:53
Breath Volume		1.324	02:53
Air Blank		0.000	02:54
Control Test		0.079	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: POGOZELSKI,CHRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: NANTASUAWAN

DATE/TIME OF TEST: 08/14/2018 23:24:45

Diagnostics Check		OK	23:21	
Air Blank		0.000	23:22	
Control Test		0.080	23:22	
Air Blank		0.000	23:23	
Subject Sample	# 1		23:23	Subject Test Refused
Breath Volume		0.000	23:23	
Air Blank		0.000	23:23	
Control Test		0.080	23:24	
Air Blank		0.000	23:24	
Diagnostics Check		OK	23:24	

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,JAMES,E
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MARTIN,JAMES
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: TABOR

DATE/TIME OF TEST: 08/15/2018 03:46:01

Diagnostics Check		OK	03:41	
Air Blank		0.000	03:41	
Control Test		0.081	03:42	
Air Blank		0.000	03:42	
Subject Sample	# 1		03:44	Subject Test Refused
Breath Volume		0.000	03:44	
Air Blank		0.000	03:45	
Control Test		0.081	03:45	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:46	

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,LANCE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,LANCE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: DUBOIS

DATE/TIME OF TEST: 08/15/2018 23:30:38

Diagnostics Check		OK	23:24
Air Blank		0.000	23:25
Control Test		0.081	23:25
Air Blank		0.000	23:25
Subject Sample	# 1	0.000	23:26
Breath Volume		3.136	23:26
Air Blank		0.000	23:26
Air Blank		0.000	23:28
Subject Sample	# 2	0.000	23:29
Breath Volume		2.714	23:29
Air Blank		0.000	23:29
Control Test		0.080	23:30
Air Blank		0.000	23:30
Diagnostics Check		OK	23:30

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,ROBERT,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NAYLOR,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: B632175942860

DATE/TIME OF TEST: 08/16/2018 01:59:20

Diagnostics Check		OK	01:53
Air Blank		0.000	01:53
Control Test		0.081	01:54
Air Blank		0.000	01:54
Subject Sample	# 1	0.179	01:55
Breath Volume		2.425	01:55
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.176	01:57
Breath Volume		1.636	01:57
Air Blank		0.000	01:58
Control Test		0.079	01:58
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,ROBERT,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MACBETH,MATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: BREEN

DATE/TIME OF TEST: 08/16/2018 02:24:33

Diagnostics Check		OK	02:17
Air Blank		0.000	02:18
Control Test		0.081	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.215	02:19
Breath Volume		2.191	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.211	02:23
Breath Volume		1.218	02:23
Air Blank		0.000	02:23
Control Test		0.080	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILHELM,TONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: QUEEN

DATE/TIME OF TEST: 08/17/2018 23:10:42

Diagnostics Check		OK	23:04
Air Blank		0.000	23:05
Control Test		0.080	23:05
Air Blank		0.000	23:05
Subject Sample	# 1	0.160	23:06
Breath Volume		1.222	23:06
Air Blank		0.000	23:07
Air Blank		0.000	23:08
Subject Sample	# 2	0.157	23:09
Breath Volume		1.492	23:09
Air Blank		0.000	23:09
Control Test		0.078	23:10
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: MINIACI

DATE/TIME OF TEST: 08/18/2018 01:31:06

Diagnostics Check		OK	01:24
Air Blank		0.000	01:24
Control Test		0.081	01:25
Air Blank		0.000	01:25
Subject Sample	# 1	0.183	01:26
Breath Volume		1.894	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:28
Subject Sample	# 2	0.180	01:29
Breath Volume		1.371	01:29
Air Blank		0.000	01:30
Control Test		0.078	01:30
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEENEY,ERIC,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FEENEY,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 08/18/2018 02:33:36

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:31	
Control Test		0.081	02:31	
Air Blank		0.000	02:31	
Subject Sample	# 1		02:32	Subject Test Refused
Breath Volume		0.000	02:32	
Air Blank		0.000	02:32	
Control Test		0.080	02:33	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,JAMES,E
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MARTIN,JAMES
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 08/18/2018 22:25:19

Diagnostics Check		OK	22:20	
Air Blank		0.000	22:21	
Control Test		0.080	22:21	
Air Blank		0.000	22:21	
Subject Sample	# 1		22:23	Subject Test Refused
Breath Volume		0.000	22:23	
Air Blank		0.000	22:24	
Control Test		0.080	22:24	
Air Blank		0.000	22:25	
Diagnostics Check		OK	22:25	

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: MIXON

DATE/TIME OF TEST: 08/18/2018 23:04:18

Diagnostics Check		OK	22:58
Air Blank		0.000	22:58
Control Test		0.080	22:58
Air Blank		0.000	22:59
Subject Sample	# 1	0.193	22:59
Breath Volume		1.765	22:59
Air Blank		0.000	23:00
Air Blank		0.000	23:02
Subject Sample	# 2	0.177	23:02
Breath Volume		1.683	23:02
Air Blank		0.000	23:03
Control Test		0.080	23:03
Air Blank		0.000	23:04
Diagnostics Check		OK	23:04

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNHAM,J,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURNHAM,J
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CARTWRIGHT

DATE/TIME OF TEST: 08/20/2018 22:47:49

Diagnostics Check		OK	22:41
Air Blank		0.000	22:42
Control Test		0.081	22:42
Air Blank		0.000	22:43
Subject Sample	# 1	0.000	22:43
Breath Volume		1.660	22:43
Air Blank		0.000	22:44
Air Blank		0.000	22:46
Subject Sample	# 2	0.000	22:46
Breath Volume		1.957	22:46
Air Blank		0.000	22:46
Control Test		0.080	22:47
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNHAM,J,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURNHAM,J
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: BARTLETT

DATE/TIME OF TEST: 08/22/2018 05:01:16

Diagnostics Check		OK	04:54
Air Blank		0.000	04:54
Control Test		0.080	04:54
Air Blank		0.000	04:55
Subject Sample	# 1	0.178	04:56
Breath Volume		1.429	04:56
Air Blank		0.000	04:57
Air Blank		0.000	04:58
Subject Sample	# 2	0.177	04:59
Breath Volume		1.820	04:59
Air Blank		0.000	05:00
Control Test		0.080	05:00
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: RIPLEY

DATE/TIME OF TEST: 08/23/2018 00:56:22

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.081	00:50
Air Blank		0.000	00:50
Subject Sample	# 1	0.129	00:51
Breath Volume		2.644	00:51
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.124	00:54
Breath Volume		2.167	00:54
Air Blank		0.000	00:55
Control Test		0.079	00:55
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,ROYCE,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BROOKS,ROYCE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: VANCE

DATE/TIME OF TEST: 08/23/2018 23:48:58

Diagnostics Check		OK	23:42
Air Blank		0.000	23:43
Control Test		0.081	23:43
Air Blank		0.000	23:44
Subject Sample	# 1	0.000	23:44
Breath Volume		3.644	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.000	23:47
Breath Volume		4.433	23:47
Air Blank		0.000	23:48
Control Test		0.079	23:48
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DON,
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BAILEY,RONALD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:47

SUBJECT'S LAST NAME: STOTTS

DATE/TIME OF TEST: 08/24/2018 15:51:20

Diagnostics Check		OK	15:45
Air Blank		0.000	15:45
Control Test		0.080	15:45
Air Blank		0.000	15:46
Subject Sample	# 1	0.236	15:46
Breath Volume		1.984	15:46
Air Blank		0.000	15:47
Air Blank		0.000	15:49
Subject Sample	# 2	0.231	15:49
Breath Volume		2.273	15:49
Air Blank		0.000	15:50
Control Test		0.079	15:50
Air Blank		0.000	15:51
Diagnostics Check		OK	15:51

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNHAM,J,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURNHAM,J
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: MICHAELS

DATE/TIME OF TEST: 08/25/2018 06:28:50

Diagnostics Check		OK	06:21
Air Blank		0.000	06:22
Control Test		0.081	06:22
Air Blank		0.000	06:23
Subject Sample	# 1	0.197	06:24
Breath Volume		2.773	06:24
Air Blank		0.000	06:24
Air Blank		0.000	06:26
Subject Sample	# 2	0.201	06:27
Breath Volume		1.597	06:27
Air Blank		0.000	06:27
Control Test		0.081	06:28
Air Blank		0.000	06:28
Diagnostics Check		OK	06:28

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONEY,PATRICK,S
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MOONEY,PATRICK
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: AKINS

DATE/TIME OF TEST: 08/27/2018 23:46:54

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.081	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.238	23:42
Breath Volume		2.210	23:42
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.223	23:45
Breath Volume		1.121	23:45
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,DUSTAN,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RICE,DUSTAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:46

SUBJECT'S LAST NAME: HANNIGAN CANGIANO

DATE/TIME OF TEST: 08/28/2018 16:20:27

Diagnostics Check		OK	16:09	
Air Blank		0.000	16:10	
Control Test		0.080	16:10	
Air Blank		0.000	16:10	
Subject Sample	# 1	0.142	16:13	Volume Not Met
Breath Volume		0.492	16:13	
Air Blank		0.000	16:14	
Air Blank		0.000	16:16	
Subject Sample	# 2		16:19	Subject Test Refused
Breath Volume		0.000	16:19	
Air Blank		0.000	16:19	
Control Test		0.080	16:19	
Air Blank		0.000	16:20	
Diagnostics Check		OK	16:20	

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NABULSI,MOOTAZ,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NABULSI,MOOTAZ
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: SHEESLEY

DATE/TIME OF TEST: 08/29/2018 03:09:53

Diagnostics Check		OK	03:01	
Air Blank		0.000	03:01	
Control Test		0.080	03:02	
Air Blank		0.000	03:02	
Subject Sample	# 1		03:05	No Sample Provided
Breath Volume		0.000	03:05	
Air Blank		0.000	03:06	
Air Blank		0.000	03:08	
Subject Sample	# 2		03:08	Subject Test Refused
Breath Volume		0.000	03:08	
Air Blank		0.000	03:08	
Control Test		0.081	03:09	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,ROBERT,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PRUSICK,ZACHERY
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 08/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 08/31/2018 05:48:43

Diagnostics Check		OK	05:42
Air Blank		0.000	05:42
Control Test		0.080	05:43
Air Blank		0.000	05:43
Subject Sample	# 1	0.212	05:44
Breath Volume		3.535	05:44
Air Blank		0.000	05:44
Air Blank		0.000	05:46
Subject Sample	# 2	0.212	05:47
Breath Volume		3.234	05:47
Air Blank		0.000	05:47
Control Test		0.081	05:48
Air Blank		0.000	05:48
Diagnostics Check		OK	05:48

Data Download Date/Time: 09/27/2018 09:08

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEWCOMB,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: NEWCOMB,WILLIAM
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 08/12/2018 17:29:07

Diagnostics Check		OK	17:25	
Air Blank		0.000	17:26	
Control Test		0.079	17:26	
Air Blank		0.000	17:27	
Subject Sample	# 1		17:27	Subject Test Refused
Breath Volume		0.000	17:27	
Air Blank		0.000	17:28	
Control Test		0.079	17:28	
Air Blank		0.000	17:29	
Diagnostics Check		OK	17:29	

Data Download Date/Time: 09/10/2018 09:29

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 08/11/2018 00:02:17

Diagnostics Check		OK	23:56
Air Blank		0.000	23:56
Control Test		0.076	23:56
Air Blank		0.000	23:57
Subject Sample	# 2	0.192	23:57
Breath Volume		1.710	23:57
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.185	00:00
Breath Volume		1.921	00:00
Air Blank		0.000	00:01
Control Test		0.075	00:01
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 08/21/2018 09:35

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PEDERSON,S
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ZUBIAGA

DATE/TIME OF TEST: 08/11/2018 03:31:51

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.077	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.201	03:27
Breath Volume		3.734	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:29
Subject Sample	# 2	0.188	03:30
Breath Volume		3.984	03:30
Air Blank		0.000	03:30
Control Test		0.076	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 08/21/2018 09:35

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: RAMIREZ-MAXIMO

DATE/TIME OF TEST: 08/31/2018 01:47:56

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.190	01:43
Breath Volume		1.339	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:45
Subject Sample	# 2	0.188	01:46
Breath Volume		1.195	01:46
Air Blank		0.000	01:47
Control Test		0.075	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 09/30/2018 09:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: PINEDA,MIGUEL
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 08/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: GALLO

DATE/TIME OF TEST: 08/22/2018 02:01:13

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:58	
Control Test		0.079	01:59	
Air Blank		0.000	01:59	
Subject Sample	# 1		01:59	Subject Test Refused
Breath Volume		0.000	01:59	
Air Blank		0.000	02:00	
Control Test		0.078	02:00	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 09/20/2018 11:39

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: RIVERA BATZ

DATE/TIME OF TEST: 08/09/2018 00:00:00

Diagnostics Check		OK	23:50
Air Blank		0.000	23:51
Control Test		0.082	23:51
Air Blank		0.000	23:51
Subject Sample	# 1	0.125	23:52
Breath Volume		1.867	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.111	23:58
Breath Volume		1.367	23:58
Air Blank		0.000	23:59
Control Test		0.083	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	00:00

Data Download Date/Time: 09/06/2018 15:00

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: SALAZAR-PEREZ

DATE/TIME OF TEST: 08/09/2018 01:52:14

Diagnostics Check		OK	01:45
Air Blank		0.000	01:45
Control Test		0.081	01:46
Air Blank		0.000	01:46
Subject Sample	# 1	0.106	01:47
Breath Volume		2.070	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.110	01:50
Breath Volume		2.144	01:50
Air Blank		0.000	01:51
Control Test		0.081	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 09/06/2018 15:00

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HILES,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: FOLSOM

DATE/TIME OF TEST: 08/10/2018 03:39:11

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.081	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.195	03:34
Breath Volume		2.472	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.185	03:37
Breath Volume		1.703	03:37
Air Blank		0.000	03:38
Control Test		0.081	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: DELAYE

DATE/TIME OF TEST: 08/10/2018 04:18:09

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.081	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.227	04:13
Breath Volume		2.500	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.213	04:16
Breath Volume		1.167	04:16
Air Blank		0.000	04:17
Control Test		0.081	04:17
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: CATES

DATE/TIME OF TEST: 08/11/2018 02:35:52

Diagnostics Check		OK	02:29
Air Blank		0.000	02:30
Control Test		0.082	02:30
Air Blank		0.000	02:30
Subject Sample	# 1	0.199	02:31
Breath Volume		3.429	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.201	02:34
Breath Volume		3.136	02:34
Air Blank		0.000	02:34
Control Test		0.082	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: FELIPE RAMIREZ

DATE/TIME OF TEST: 08/11/2018 05:28:06

Diagnostics Check		OK	05:19
Air Blank		0.000	05:20
Control Test		0.082	05:20
Air Blank		0.000	05:21
Subject Sample	# 1	0.176	05:22
Breath Volume		1.281	05:22
Air Blank		0.000	05:22
Air Blank		0.000	05:24
Subject Sample	# 2	0.167	05:26
Breath Volume		1.441	05:26
Air Blank		0.000	05:27
Control Test		0.083	05:27
Air Blank		0.000	05:28
Diagnostics Check		OK	05:28

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ALFORD

DATE/TIME OF TEST: 08/12/2018 00:43:10

Diagnostics Check		OK	00:35
Air Blank		0.000	00:36
Control Test		0.081	00:36
Air Blank		0.000	00:36
Subject Sample	# 1	0.152	00:38
Breath Volume		3.093	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.155	00:41
Breath Volume		2.957	00:41
Air Blank		0.000	00:42
Control Test		0.081	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MALEC

DATE/TIME OF TEST: 08/12/2018 02:22:14

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.082	02:15
Air Blank		0.000	02:16
Subject Sample	# 1	0.162	02:17
Breath Volume		1.726	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:19
Subject Sample	# 2	0.158	02:20
Breath Volume		1.726	02:20
Air Blank		0.000	02:21
Control Test		0.082	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 08/12/2018 04:10:16

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.081	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.028	04:05
Breath Volume		1.156	04:05
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.026	04:08
Breath Volume		1.187	04:08
Air Blank		0.000	04:09
Control Test		0.082	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PAYES JR

DATE/TIME OF TEST: 08/13/2018 02:37:03

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.081	02:31
Air Blank		0.000	02:31
Subject Sample	# 1	0.205	02:32
Breath Volume		1.343	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:34
Subject Sample	# 2	0.203	02:35
Breath Volume		1.390	02:35
Air Blank		0.000	02:36
Control Test		0.081	02:36
Air Blank		0.000	02:36
Diagnostics Check		OK	02:37

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCH,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 08/14/2018 23:48:21

Diagnostics Check		OK	23:41
Air Blank		0.000	23:41
Control Test		0.081	23:42
Air Blank		0.000	23:42
Subject Sample	# 1	0.130	23:43
Breath Volume		3.402	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.128	23:46
Breath Volume		4.101	23:46
Air Blank		0.000	23:47
Control Test		0.081	23:47
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: AVIRETT

DATE/TIME OF TEST: 08/17/2018 02:24:32

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.081	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.101	02:20
Breath Volume		2.207	02:20
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.098	02:22
Breath Volume		2.382	02:22
Air Blank		0.000	02:23
Control Test		0.081	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHILLER,E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 08/18/2018 00:35:06

Diagnostics Check		OK	00:28
Air Blank		0.000	00:28
Control Test		0.082	00:28
Air Blank		0.000	00:29
Subject Sample	# 1	0.088	00:30
Breath Volume		1.406	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:32
Subject Sample	# 2	0.088	00:33
Breath Volume		1.566	00:33
Air Blank		0.000	00:34
Control Test		0.081	00:34
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: MATOS-VAZQUEZ

DATE/TIME OF TEST: 08/19/2018 02:02:28

Diagnostics Check		OK	01:56
Air Blank		0.000	01:56
Control Test		0.082	01:56
Air Blank		0.000	01:57
Subject Sample	# 1	0.128	01:57
Breath Volume		2.875	01:57
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.127	02:00
Breath Volume		2.992	02:00
Air Blank		0.000	02:01
Control Test		0.081	02:01
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: RAPOSO

DATE/TIME OF TEST: 08/19/2018 02:48:30

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:37	
Control Test		0.082	02:37	
Air Blank		0.000	02:38	
Subject Sample	# 1		02:41	No Sample Provided
Breath Volume		0.000	02:41	
Air Blank		0.000	02:41	
Air Blank		0.000	02:44	
Subject Sample	# 2		02:47	No Sample Provided
Breath Volume		0.000	02:47	
Air Blank		0.000	02:47	
Control Test		0.082	02:47	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:48	

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: DOMINGO-DE LA CRUZ

DATE/TIME OF TEST: 08/19/2018 03:14:11

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.081	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.218	03:09
Breath Volume		1.789	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.227	03:12
Breath Volume		2.613	03:12
Air Blank		0.000	03:13
Control Test		0.082	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: FORSYTHE

DATE/TIME OF TEST: 08/19/2018 04:42:52

Diagnostics Check		OK	04:35
Air Blank		0.000	04:35
Control Test		0.081	04:35
Air Blank		0.000	04:36
Subject Sample	# 1	0.135	04:38
Breath Volume		2.171	04:38
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.135	04:41
Breath Volume		2.296	04:41
Air Blank		0.000	04:41
Control Test		0.081	04:42
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: REID,W
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: CALZADA

DATE/TIME OF TEST: 08/20/2018 02:23:44

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.081	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.107	02:18
Breath Volume		1.812	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.118	02:22
Breath Volume		2.328	02:22
Air Blank		0.000	02:22
Control Test		0.081	02:23
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: MCELVEEN

DATE/TIME OF TEST: 08/20/2018 03:34:33

Diagnostics Check		OK	03:27
Air Blank		0.000	03:28
Control Test		0.081	03:28
Air Blank		0.000	03:29
Subject Sample	# 1	0.231	03:29
Breath Volume		3.390	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.227	03:32
Breath Volume		3.023	03:32
Air Blank		0.000	03:33
Control Test		0.081	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ZELAYA MENDOZA

DATE/TIME OF TEST: 08/22/2018 01:07:00

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.082	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.180	01:02
Breath Volume		1.722	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:04
Subject Sample	# 2	0.174	01:05
Breath Volume		1.304	01:05
Air Blank		0.000	01:06
Control Test		0.081	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:07

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WHITCOMB,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: BENTLEY

DATE/TIME OF TEST: 08/22/2018 15:33:23

Diagnostics Check		OK	15:27
Air Blank		0.000	15:27
Control Test		0.081	15:27
Air Blank		0.000	15:28
Subject Sample	# 1	0.000	15:29
Breath Volume		4.539	15:29
Air Blank		0.000	15:29
Air Blank		0.000	15:31
Subject Sample	# 2	0.000	15:32
Breath Volume		3.875	15:32
Air Blank		0.000	15:32
Control Test		0.083	15:32
Air Blank		0.000	15:33
Diagnostics Check		OK	15:33

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: BLEVINS

DATE/TIME OF TEST: 08/22/2018 21:53:09

Diagnostics Check		OK	21:44
Air Blank		0.000	21:44
Control Test		0.082	21:44
Air Blank		0.000	21:45
Subject Sample	# 1	0.182	21:45
Breath Volume		1.871	21:45
Air Blank		0.000	21:46
Air Blank		0.000	21:48
Subject Sample	# 2	0.182	21:51
Breath Volume		2.859	21:51
Air Blank		0.000	21:52
Control Test		0.081	21:52
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MARTINS

DATE/TIME OF TEST: 08/24/2018 03:13:18

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:10	
Control Test		0.082	03:11	
Air Blank		0.000	03:11	
Subject Sample	# 1		03:12	Sequence Aborted
Breath Volume		0.000	03:12	
Air Blank		0.000	03:13	

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MARTINS

DATE/TIME OF TEST: 08/24/2018 03:27:29

Diagnostics Check		OK	03:19
Air Blank		0.000	03:20
Control Test		0.082	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.080	03:22
Breath Volume		1.203	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.081	03:25
Breath Volume		1.246	03:25
Air Blank		0.000	03:26
Control Test		0.081	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: BEAHM

DATE/TIME OF TEST: 08/25/2018 02:14:05

Diagnostics Check		OK	02:07
Air Blank		0.000	02:08
Control Test		0.081	02:08
Air Blank		0.000	02:09
Subject Sample	# 1	0.239	02:09
Breath Volume		2.113	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:11
Subject Sample	# 2	0.254	02:12
Breath Volume		2.125	02:12
Air Blank		0.000	02:13
Control Test		0.082	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: CLEMENTS

DATE/TIME OF TEST: 08/25/2018 03:06:14

Diagnostics Check		OK	03:01	
Air Blank		0.000	03:02	
Control Test		0.082	03:02	
Air Blank		0.000	03:02	
Subject Sample	# 1		03:04	Subject Test Refused
Breath Volume		0.000	03:04	
Air Blank		0.000	03:05	
Control Test		0.082	03:05	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:06	

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: HEBDEN

DATE/TIME OF TEST: 08/26/2018 00:40:57

Diagnostics Check		OK	00:38	
Air Blank		0.000	00:38	
Control Test		0.082	00:38	
Air Blank		0.000	00:39	
Subject Sample	# 1		00:39	Subject Test Refused
Breath Volume		0.000	00:39	
Air Blank		0.000	00:40	
Control Test		0.081	00:40	
Air Blank		0.000	00:40	
Diagnostics Check		OK	00:40	

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: HEBDEN

DATE/TIME OF TEST: 08/26/2018 00:58:12

Diagnostics Check		OK	00:48
Air Blank		0.000	00:48
Control Test		0.081	00:48
Air Blank		0.000	00:49
Subject Sample	# 1	0.157	00:52
Breath Volume		1.441	00:52
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.148	00:56
Breath Volume		2.015	00:56
Air Blank		0.000	00:57
Control Test		0.081	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 08/26/2018 02:04:00

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.080	01:58
Air Blank		0.000	01:58
Subject Sample	# 1	0.138	01:59
Breath Volume		2.976	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:01
Subject Sample	# 2	0.136	02:02
Breath Volume		3.062	02:02
Air Blank		0.000	02:03
Control Test		0.081	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:04

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: CRUZ MOLINERO

DATE/TIME OF TEST: 08/26/2018 04:19:28

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.081	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.177	04:14
Breath Volume		2.027	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.178	04:17
Breath Volume		2.039	04:17
Air Blank		0.000	04:18
Control Test		0.081	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PALFY,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: EPRIGHT

DATE/TIME OF TEST: 08/27/2018 02:13:00

Diagnostics Check		OK	02:06
Air Blank		0.000	02:07
Control Test		0.081	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.000	02:08
Breath Volume		1.441	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.000	02:11
Breath Volume		2.214	02:11
Air Blank		0.000	02:12
Control Test		0.081	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:13

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: DEERING II

DATE/TIME OF TEST: 08/27/2018 04:05:58

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.081	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.000	04:01
Breath Volume		2.910	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.000	04:04
Breath Volume		2.839	04:04
Air Blank		0.000	04:05
Control Test		0.081	04:05
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: DORSEY

DATE/TIME OF TEST: 08/28/2018 00:31:16

Diagnostics Check		OK	00:19	
Air Blank		0.000	00:19	
Control Test		0.081	00:20	
Air Blank		0.000	00:20	
Subject Sample	# 1	0.198	00:23	Slope Not Level
Breath Volume		1.167	00:23	
Air Blank		0.000	00:24	
Air Blank		0.000	00:26	
Subject Sample	# 2	0.268	00:26	
Breath Volume		1.347	00:26	
Air Blank		0.000	00:27	
Air Blank		0.000	00:29	
Subject Sample	# 3	0.277	00:29	
Breath Volume		1.285	00:29	
Air Blank		0.000	00:30	
Control Test		0.080	00:30	
Air Blank		0.000	00:31	
Diagnostics Check		OK	00:31	

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAULINO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: SAFFRI

DATE/TIME OF TEST: 08/29/2018 04:14:36

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.081	04:08
Air Blank		0.000	04:09
Subject Sample	# 1	0.179	04:10
Breath Volume		2.117	04:10
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.185	04:13
Breath Volume		1.992	04:13
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: PERTEE

DATE/TIME OF TEST: 08/30/2018 05:12:26

Diagnostics Check		OK	05:04
Air Blank		0.000	05:05
Control Test		0.081	05:05
Air Blank		0.000	05:06
Subject Sample	# 1	0.094	05:07
Breath Volume		2.449	05:07
Air Blank		0.000	05:08
Air Blank		0.000	05:10
Subject Sample	# 2	0.092	05:10
Breath Volume		2.226	05:10
Air Blank		0.000	05:11
Control Test		0.081	05:11
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HOWARD,ROBERT
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:47

SUBJECT'S LAST NAME: DEPAUL

DATE/TIME OF TEST: 08/30/2018 15:17:16

Diagnostics Check		OK	15:09
Air Blank		0.000	15:10
Control Test		0.081	15:10
Air Blank		0.000	15:10
Subject Sample	# 1	0.135	15:12
Breath Volume		2.949	15:12
Air Blank		0.000	15:13
Air Blank		0.000	15:14
Subject Sample	# 2	0.120	15:15
Breath Volume		1.328	15:15
Air Blank		0.000	15:16
Control Test		0.081	15:16
Air Blank		0.000	15:17
Diagnostics Check		OK	15:17

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: PENLEY

DATE/TIME OF TEST: 08/31/2018 02:16:24

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.082	02:10
Air Blank		0.000	02:11
Subject Sample	# 1	0.185	02:11
Breath Volume		3.261	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.184	02:14
Breath Volume		2.980	02:14
Air Blank		0.000	02:15
Control Test		0.081	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 09/06/2018 15:01

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADILLA,ANTHONY,D
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CARTAYA,GUILLERMO
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: BUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 08/05/2018 20:22:19

Diagnostics Check		OK	20:18	
Air Blank		0.000	20:19	
Control Test		0.082	20:19	
Air Blank		0.000	20:19	
Subject Sample	# 1		20:20	Subject Test Refused
Breath Volume		0.000	20:20	
Air Blank		0.000	20:21	
Control Test		0.082	20:21	
Air Blank		0.000	20:22	
Diagnostics Check		OK	20:22	

Data Download Date/Time: 08/20/2018 22:57

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ADRIAN,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: GOMEZ,RUBEN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: BUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 08/12/2018 21:22:54

Diagnostics Check		OK	21:16
Air Blank		0.000	21:17
Control Test		0.082	21:17
Air Blank		0.000	21:17
Subject Sample	# 1	0.082	21:18
Breath Volume		3.785	21:18
Air Blank		0.000	21:19
Air Blank		0.000	21:20
Subject Sample	# 2	0.069	21:21
Breath Volume		1.117	21:21
Air Blank		0.000	21:21
Control Test		0.082	21:22
Air Blank		0.000	21:22
Diagnostics Check		OK	21:22

Data Download Date/Time: 08/20/2018 22:57

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,GABRIEL,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BURNS,JEFF
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 07/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: BUI
ARREST TIME: 14:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: KRAMER

DATE/TIME OF TEST: 08/18/2018 15:17:14

Diagnostics Check		OK	15:14	
Air Blank		0.000	15:14	
Control Test		0.081	15:15	
Air Blank		0.000	15:15	
Subject Sample	# 1		15:15	Subject Test Refused
Breath Volume		0.000	15:15	
Air Blank		0.000	15:16	
Control Test		0.081	15:16	
Air Blank		0.000	15:17	
Diagnostics Check		OK	15:17	

Data Download Date/Time: 08/20/2018 22:57

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSCULLUELA,BRYAN,M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CARTAYA,GUILLERMO
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: BUI
ARREST TIME: 14:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:26

SUBJECT'S LAST NAME: TOLSTOV

DATE/TIME OF TEST: 08/25/2018 15:23:20

Diagnostics Check		OK	15:20	
Air Blank		0.000	15:20	
Control Test		0.082	15:21	
Air Blank		0.000	15:21	
Subject Sample	# 1		15:21	Subject Test Refused
Breath Volume		0.000	15:21	
Air Blank		0.000	15:22	
Control Test		0.082	15:22	
Air Blank		0.000	15:23	
Diagnostics Check		OK	15:23	

Data Download Date/Time: 09/26/2018 20:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEIRCE,HAROLD,F
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KELLY,WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: PRIETO

DATE/TIME OF TEST: 08/01/2018 06:36:48

Diagnostics Check		OK	06:30
Air Blank		0.000	06:30
Control Test		0.080	06:31
Air Blank		0.000	06:31
Subject Sample	# 1	0.122	06:32
Breath Volume		1.640	06:32
Air Blank		0.000	06:32
Air Blank		0.000	06:34
Subject Sample	# 2	0.121	06:35
Breath Volume		1.871	06:35
Air Blank		0.000	06:35
Control Test		0.080	06:36
Air Blank		0.000	06:36
Diagnostics Check		OK	06:36

Data Download Date/Time: 08/19/2018 15:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLITO,BRAD,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: HOAGLAN,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: WHITSON

DATE/TIME OF TEST: 08/03/2018 15:29:32

Diagnostics Check		OK	15:23
Air Blank		0.000	15:23
Control Test		0.080	15:23
Air Blank		0.000	15:24
Subject Sample	# 1	0.000	15:25
Breath Volume		1.562	15:25
Air Blank		0.000	15:25
Air Blank		0.000	15:27
Subject Sample	# 2	0.000	15:28
Breath Volume		1.167	15:28
Air Blank		0.000	15:28
Control Test		0.079	15:28
Air Blank		0.000	15:29
Diagnostics Check		OK	15:29

Data Download Date/Time: 08/19/2018 15:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTA,ANGEL,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MATHIS,JOHN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: PIERSON

DATE/TIME OF TEST: 08/03/2018 20:51:38

Diagnostics Check		OK	20:44
Air Blank		0.000	20:44
Control Test		0.079	20:45
Air Blank		0.000	20:45
Subject Sample	# 1	0.166	20:47
Breath Volume		2.535	20:47
Air Blank		0.000	20:47
Air Blank		0.000	20:49
Subject Sample	# 2	0.157	20:50
Breath Volume		1.300	20:50
Air Blank		0.000	20:50
Control Test		0.078	20:51
Air Blank		0.000	20:51
Diagnostics Check		OK	20:51

Data Download Date/Time: 08/19/2018 15:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTA,ANGEL,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WHITLOCK,DANIEL
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 08/03/2018 22:48:10

Diagnostics Check		OK	22:41
Air Blank		0.000	22:42
Control Test		0.079	22:42
Air Blank		0.000	22:42
Subject Sample	# 1	0.277	22:43
Breath Volume		1.136	22:43
Air Blank		0.000	22:44
Air Blank		0.000	22:46
Subject Sample	# 2	0.271	22:46
Breath Volume		1.167	22:46
Air Blank		0.000	22:47
Control Test		0.079	22:47
Air Blank		0.000	22:48
Diagnostics Check		OK	22:48

Data Download Date/Time: 08/19/2018 15:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,DANIEL,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RICHIE,F
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: LOWERS

DATE/TIME OF TEST: 08/04/2018 01:49:26

Diagnostics Check		OK	01:42
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.148	01:44
Breath Volume		1.503	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.146	01:47
Breath Volume		1.171	01:47
Air Blank		0.000	01:48
Control Test		0.078	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 08/19/2018 15:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,CHRISTINA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MATHIS,JOHN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: RHODES

DATE/TIME OF TEST: 08/05/2018 19:36:36

Diagnostics Check		OK	19:30
Air Blank		0.000	19:30
Control Test		0.079	19:31
Air Blank		0.000	19:31
Subject Sample	# 1	0.175	19:32
Breath Volume		2.085	19:32
Air Blank		0.000	19:32
Air Blank		0.000	19:34
Subject Sample	# 2	0.172	19:35
Breath Volume		2.007	19:35
Air Blank		0.000	19:35
Control Test		0.079	19:36
Air Blank		0.000	19:36
Diagnostics Check		OK	19:36

Data Download Date/Time: 08/19/2018 15:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,DANIEL,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SIMMONS,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: TASSONE

DATE/TIME OF TEST: 08/06/2018 04:57:15

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.080	04:51
Air Blank		0.000	04:51
Subject Sample	# 1	0.169	04:52
Breath Volume		1.351	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.167	04:55
Breath Volume		1.175	04:55
Air Blank		0.000	04:56
Control Test		0.080	04:56
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 08/19/2018 15:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: LATTY,VICTOR
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/06/2018 22:37:03

Diagnostics Check		OK	22:33	
Air Blank		0.000	22:34	
Control Test		0.080	22:34	
Air Blank		0.000	22:35	
Subject Sample	# 1		22:35	Subject Test Refused
Breath Volume		0.000	22:35	
Air Blank		0.000	22:36	
Control Test		0.080	22:36	
Air Blank		0.000	22:36	
Diagnostics Check		OK	22:37	

Data Download Date/Time: 08/19/2018 15:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIMPSON,RACHEL,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KENT,NATHAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 08/12/2018 16:53:07

Diagnostics Check		OK	16:50	
Air Blank		0.000	16:50	
Control Test		0.081	16:51	
Air Blank		0.000	16:51	
Subject Sample	# 1		16:51	Subject Test Refused
Breath Volume		0.000	16:51	
Air Blank		0.000	16:52	
Control Test		0.080	16:52	
Air Blank		0.000	16:53	
Diagnostics Check		OK	16:53	

Data Download Date/Time: 08/19/2018 15:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTA,ANGEL,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: GREIS,DAVID
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 08/14/2018 21:43:17

Diagnostics Check		OK	21:40	
Air Blank		0.000	21:40	
Control Test		0.080	21:41	
Air Blank		0.000	21:41	
Subject Sample	# 1		21:41	Subject Test Refused
Breath Volume		0.000	21:41	
Air Blank		0.000	21:42	
Control Test		0.079	21:42	
Air Blank		0.000	21:43	
Diagnostics Check		OK	21:43	

Data Download Date/Time: 08/19/2018 15:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MATHIS,JON
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: NEWBERRY

DATE/TIME OF TEST: 08/16/2018 01:21:53

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:19	
Control Test		0.080	01:19	
Air Blank		0.000	01:20	
Subject Sample	# 1		01:20	Subject Test Refused
Breath Volume		0.000	01:20	
Air Blank		0.000	01:20	
Control Test		0.080	01:21	
Air Blank		0.000	01:21	
Diagnostics Check		OK	01:21	

Data Download Date/Time: 08/19/2018 15:31

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIMPSON,RACHEL,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SMITH,BRIAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 08/20/2018 16:37:40

Diagnostics Check		OK	16:30
Air Blank		0.000	16:31
Control Test		0.080	16:31
Air Blank		0.000	16:31
Subject Sample	# 1	0.093	16:32
Breath Volume		1.140	16:32
Air Blank		0.000	16:33
Air Blank		0.000	16:35
Subject Sample	# 2	0.090	16:36
Breath Volume		1.140	16:36
Air Blank		0.000	16:36
Control Test		0.079	16:37
Air Blank		0.000	16:37
Diagnostics Check		OK	16:37

Data Download Date/Time: 09/15/2018 11:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTA,ANGEL,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 08/27/2018 23:44:44

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.081	23:39
Air Blank		0.000	23:39
Subject Sample	# 1	0.000	23:40
Breath Volume		1.289	23:40
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.000	23:43
Breath Volume		1.410	23:43
Air Blank		0.000	23:43
Control Test		0.080	23:44
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 09/15/2018 11:16

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTA,ANGEL,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A3
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: PULEO

DATE/TIME OF TEST: 08/31/2018 23:28:40

Diagnostics Check		OK	23:25	
Air Blank		0.000	23:26	
Control Test		0.080	23:26	
Air Blank		0.000	23:27	
Subject Sample	# 1		23:27	Subject Test Refused
Breath Volume		0.000	23:27	
Air Blank		0.000	23:27	
Control Test		0.079	23:28	
Air Blank		0.000	23:28	
Diagnostics Check		OK	23:28	

Data Download Date/Time: 09/15/2018 11:16

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSIAH,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH,JOSIAH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: MUNIZ

DATE/TIME OF TEST: 08/08/2018 01:36:16

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.083	01:28
Air Blank		0.000	01:29
Subject Sample	# 1	0.070	01:30
Breath Volume		1.316	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.075	01:34
Breath Volume		2.347	01:34
Air Blank		0.000	01:35
Control Test		0.082	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 09/05/2018 09:54

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: TRASK,J
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 08/19/2018 02:06:51

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:01	
Control Test		0.082	02:02	
Air Blank		0.000	02:02	
Subject Sample	# 1	0.231	02:05	
Breath Volume		1.812	02:05	
Air Blank			02:06	Purge Fail
Air Blank		0.000	02:06	

Data Download Date/Time: 09/05/2018 09:54

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: TRASK,J
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 08/19/2018 02:16:41

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.077	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.221	02:10
Breath Volume		1.308	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.228	02:14
Breath Volume		1.613	02:14
Air Blank		0.000	02:15
Control Test		0.077	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 09/05/2018 09:54

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: APARICIO

DATE/TIME OF TEST: 08/19/2018 21:12:26

Diagnostics Check		OK	21:05
Air Blank		0.000	21:06
Control Test		0.082	21:06
Air Blank		0.000	21:07
Subject Sample	# 1	0.222	21:07
Breath Volume		2.406	21:07
Air Blank		0.000	21:08
Air Blank		0.000	21:10
Subject Sample	# 2	0.225	21:10
Breath Volume		2.445	21:10
Air Blank		0.000	21:11
Control Test		0.081	21:11
Air Blank		0.000	21:12
Diagnostics Check		OK	21:12

Data Download Date/Time: 09/05/2018 09:54

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEINER,RYAN,T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: CARTER,C
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 08/21/2018 20:34:36

Diagnostics Check		OK	20:28
Air Blank		0.000	20:28
Control Test		0.082	20:29
Air Blank		0.000	20:29
Subject Sample	# 1	0.000	20:30
Breath Volume		3.152	20:30
Air Blank		0.000	20:30
Air Blank		0.000	20:32
Subject Sample	# 2	0.000	20:33
Breath Volume		2.914	20:33
Air Blank		0.000	20:33
Control Test		0.083	20:33
Air Blank		0.000	20:34
Diagnostics Check		OK	20:34

Data Download Date/Time: 09/05/2018 09:54

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDIN,JEREMY,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: RALL,THOMAS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:42

SUBJECT'S LAST NAME: BECKER

DATE/TIME OF TEST: 08/22/2018 12:54:50

Diagnostics Check		OK	12:48
Air Blank		0.000	12:48
Control Test		0.082	12:48
Air Blank		0.000	12:49
Subject Sample	# 1	0.000	12:49
Breath Volume		1.496	12:49
Air Blank		0.000	12:50
Air Blank		0.000	12:52
Subject Sample	# 2	0.000	12:53
Breath Volume		1.414	12:53
Air Blank		0.000	12:53
Control Test		0.080	12:54
Air Blank		0.000	12:54
Diagnostics Check		OK	12:54

Data Download Date/Time: 09/05/2018 09:54

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAEFER,KENDALL,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SCHAEFER,KENDALL
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: ITZEP BATEN

DATE/TIME OF TEST: 08/27/2018 14:30:43

Diagnostics Check		OK	14:23
Air Blank		0.000	14:24
Control Test		0.081	14:24
Air Blank		0.000	14:24
Subject Sample	# 1	0.317	14:25
Breath Volume		2.339	14:25
Air Blank		0.000	14:26
Air Blank		0.000	14:28
Subject Sample	# 2	0.315	14:28
Breath Volume		1.625	14:28
Air Blank		0.000	14:29
Control Test		0.080	14:30
Air Blank		0.000	14:30
Diagnostics Check		OK	14:30

Data Download Date/Time: 09/05/2018 09:54

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEINER,RYAN,T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: HERNANDEZ,Y
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/31/2018 05:26:42

Diagnostics Check		OK	05:22	
Air Blank		0.000	05:23	
Control Test		0.080	05:23	
Air Blank		0.000	05:24	
Subject Sample	# 1		05:25	Subject Test Refused
Breath Volume		0.000	05:25	
Air Blank		0.000	05:25	
Control Test		0.081	05:26	
Air Blank		0.000	05:26	
Diagnostics Check		OK	05:26	

Data Download Date/Time: 09/05/2018 09:54

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,LANDON,L
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: EDWARDS,LANDON
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: DEGROFF

DATE/TIME OF TEST: 08/02/2018 02:05:47

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.081	01:59
Air Blank		0.000	02:00
Subject Sample	# 1	0.000	02:01
Breath Volume		5.550	02:01
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.000	02:04
Breath Volume		4.839	02:04
Air Blank		0.000	02:04
Control Test		0.079	02:05
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 08/22/2018 21:15

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WARD,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: LAFOY

DATE/TIME OF TEST: 08/03/2018 22:18:25

Diagnostics Check		OK	22:11
Air Blank		0.000	22:11
Control Test		0.081	22:12
Air Blank		0.000	22:12
Subject Sample	# 1	0.196	22:13
Breath Volume		2.535	22:13
Air Blank		0.000	22:14
Air Blank		0.000	22:15
Subject Sample	# 2	0.189	22:16
Breath Volume		1.917	22:16
Air Blank		0.000	22:17
Control Test		0.080	22:17
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 08/22/2018 21:15

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAWCETT,PETER,T
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: GASPARD,FERTZ
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: TABACO

DATE/TIME OF TEST: 08/05/2018 23:04:59

Diagnostics Check		OK	22:57
Air Blank		0.000	22:58
Control Test		0.080	22:58
Air Blank		0.000	22:59
Subject Sample	# 1	0.076	23:00
Breath Volume		1.328	23:00
Air Blank		0.000	23:00
Air Blank		0.000	23:02
Subject Sample	# 2	0.082	23:03
Breath Volume		1.437	23:03
Air Blank		0.000	23:04
Control Test		0.080	23:04
Air Blank		0.000	23:04
Diagnostics Check		OK	23:04

Data Download Date/Time: 08/22/2018 21:15

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MCGRORY,ZACH
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: FARELLA

DATE/TIME OF TEST: 08/10/2018 02:46:17

Diagnostics Check		OK	02:43	
Air Blank		0.000	02:43	
Control Test		0.079	02:44	
Air Blank		0.000	02:44	
Subject Sample	# 1		02:44	Subject Test Refused
Breath Volume		0.000	02:44	
Air Blank		0.000	02:45	
Control Test		0.080	02:45	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:46	

Data Download Date/Time: 08/22/2018 21:15

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMPE,CORTLAND,B
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: LAMPE,CORTLAND
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 08/17/2018 17:33:48

Diagnostics Check		OK	17:26
Air Blank		0.000	17:27
Control Test		0.081	17:27
Air Blank		0.000	17:27
Subject Sample	# 1	0.083	17:29
Breath Volume		4.414	17:29
Air Blank		0.000	17:29
Air Blank		0.000	17:31
Subject Sample	# 2	0.086	17:32
Breath Volume		4.300	17:32
Air Blank		0.000	17:32
Control Test		0.080	17:33
Air Blank		0.000	17:33
Diagnostics Check		OK	17:33

Data Download Date/Time: 08/22/2018 21:15

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: CARMAN,TAYLOR
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: YOCKEY

DATE/TIME OF TEST: 08/19/2018 02:21:27

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.080	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.198	02:15
Breath Volume		1.183	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:17
Subject Sample	# 2	0.200	02:19
Breath Volume		2.300	02:19
Air Blank		0.000	02:20
Control Test		0.079	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 08/22/2018 21:15

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,LANDON,L
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: PEREZ,JAMIE
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 07/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: EGAN

DATE/TIME OF TEST: 08/22/2018 00:06:00

Diagnostics Check		OK	23:52	
Air Blank		0.000	23:53	
Control Test		0.080	23:53	
Air Blank		0.000	23:54	
Subject Sample	# 3	0.173	23:55	
Breath Volume		1.757	23:55	
Air Blank		0.000	23:55	
Air Blank		0.000	23:57	
Subject Sample	# 1	0.144	00:00	Volume Not Met
Breath Volume		1.019	00:00	
Air Blank		0.000	00:01	
Air Blank		0.000	00:03	
Subject Sample	# 2		00:04	Subject Test Refused
Breath Volume		0.144	00:04	
Air Blank		0.000	00:05	
Control Test		0.076	00:05	
Air Blank		0.000	00:05	
Diagnostics Check		OK	00:06	

Data Download Date/Time: 08/22/2018 21:15

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JASON,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: HUNTER,MARK
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: LOVELADY JR

DATE/TIME OF TEST: 08/01/2018 03:34:36

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.080	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.118	03:29
Breath Volume		2.347	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.107	03:32
Breath Volume		1.546	03:32
Air Blank		0.000	03:33
Control Test		0.079	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 08/09/2018 09:57

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JASON,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: MITCHELL,JARELL
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: BARNARD

DATE/TIME OF TEST: 08/04/2018 22:37:34

Diagnostics Check		OK	22:34	
Air Blank		0.000	22:35	
Control Test		0.080	22:35	
Air Blank		0.000	22:35	
Subject Sample	# 1		22:36	Subject Test Refused
Breath Volume		0.000	22:36	
Air Blank		0.000	22:36	
Control Test		0.080	22:36	
Air Blank		0.000	22:37	
Diagnostics Check		OK	22:37	

Data Download Date/Time: 08/09/2018 09:57

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JASON,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSHUA
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 08/05/2018 01:31:04

Diagnostics Check		OK	01:28	
Air Blank		0.000	01:28	
Control Test		0.081	01:28	
Air Blank		0.000	01:29	
Subject Sample	# 1		01:29	Subject Test Refused
Breath Volume		0.000	01:29	
Air Blank		0.000	01:30	
Control Test		0.080	01:30	
Air Blank		0.000	01:30	
Diagnostics Check		OK	01:31	

Data Download Date/Time: 08/09/2018 09:57

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAZURYK,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD,RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: BOHL

DATE/TIME OF TEST: 08/06/2018 20:12:18

Diagnostics Check		OK	20:05
Air Blank		0.000	20:05
Control Test		0.080	20:05
Air Blank		0.000	20:06
Subject Sample	# 1	0.279	20:06
Breath Volume		1.894	20:06
Air Blank		0.000	20:07
Air Blank		0.000	20:09
Subject Sample	# 2	0.260	20:10
Breath Volume		1.273	20:10
Air Blank		0.000	20:11
Control Test		0.078	20:11
Air Blank		0.000	20:12
Diagnostics Check		OK	20:12

Data Download Date/Time: 08/09/2018 09:57

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAZURYK,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD,RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: RICHTER

DATE/TIME OF TEST: 08/07/2018 22:41:15

Diagnostics Check		OK	22:33	
Air Blank		0.000	22:33	
Control Test		0.079	22:34	
Air Blank		0.000	22:34	
Subject Sample	# 1	0.205	22:35	
Breath Volume		2.468	22:35	
Air Blank		0.000	22:37	
Air Blank		0.000	22:38	
Subject Sample	# 2	0.194	22:39	
Breath Volume		3.218	22:39	
Air Blank		0.000	22:40	
Control Test		0.071	22:40	Control Outside Tolerance
Air Blank		0.000	22:41	

Data Download Date/Time: 08/09/2018 09:57

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAZURYK,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD,RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: RICHTER

DATE/TIME OF TEST: 08/07/2018 22:58:03

Diagnostics Check		OK	22:54	
Air Blank		0.000	22:55	
Control Test		0.075	22:55	
Air Blank		0.000	22:56	
Subject Sample	# 1	0.196	22:56	
Breath Volume		2.617	22:56	
Air Blank			22:57	Purge Fail
Air Blank		0.000	22:58	

Data Download Date/Time: 08/09/2018 09:57

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAZURYK,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD,RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: RICHTER

DATE/TIME OF TEST: 08/07/2018 23:25:27

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:24	
Control Test	0.070	23:24	Control Outside Tolerance
Air Blank	0.000	23:25	

Data Download Date/Time: 08/09/2018 09:57

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JASON,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSHUA
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 08/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: HOOPER

DATE/TIME OF TEST: 08/15/2018 00:49:50

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.080	00:44
Air Blank		0.000	00:44
Subject Sample	# 1	0.190	00:45
Breath Volume		1.949	00:45
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.193	00:48
Breath Volume		2.062	00:48
Air Blank		0.000	00:48
Control Test		0.079	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 09/25/2018 15:42

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: LIKOURENTZOS

DATE/TIME OF TEST: 08/01/2018 05:42:51

Diagnostics Check		OK	05:39	
Air Blank		0.000	05:40	
Control Test		0.079	05:40	
Air Blank		0.000	05:41	
Subject Sample	# 1		05:41	Subject Test Refused
Breath Volume		0.000	05:41	
Air Blank		0.000	05:41	
Control Test		0.079	05:42	
Air Blank		0.000	05:42	
Diagnostics Check		OK	05:42	

Data Download Date/Time: 08/14/2018 09:57

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MENDEZ-RUIZ

DATE/TIME OF TEST: 08/02/2018 03:05:36

Diagnostics Check		OK	02:58
Air Blank		0.000	02:59
Control Test		0.079	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.071	03:01
Breath Volume		1.125	03:01
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.073	03:04
Breath Volume		1.992	03:04
Air Blank		0.000	03:04
Control Test		0.080	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 08/14/2018 09:57

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: RODRIGUEZ REBOYRAS

DATE/TIME OF TEST: 08/03/2018 02:22:13

Diagnostics Check		OK	02:13
Air Blank		0.000	02:14
Control Test		0.080	02:14
Air Blank		0.000	02:14
Subject Sample	# 1	0.076	02:17
Breath Volume		1.441	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.084	02:20
Breath Volume		1.628	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 08/14/2018 09:57

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ABRAMSON,SHELBY
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 07/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: LIZARDE

DATE/TIME OF TEST: 08/03/2018 05:25:41

Diagnostics Check		OK	05:19
Air Blank		0.000	05:20
Control Test		0.079	05:20
Air Blank		0.000	05:20
Subject Sample	# 1	0.191	05:21
Breath Volume		3.445	05:21
Air Blank		0.000	05:21
Air Blank		0.000	05:23
Subject Sample	# 2	0.198	05:24
Breath Volume		2.828	05:24
Air Blank		0.000	05:24
Control Test		0.079	05:25
Air Blank		0.000	05:25
Diagnostics Check		OK	05:25

Data Download Date/Time: 08/14/2018 09:57

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: REYNOSO

DATE/TIME OF TEST: 08/04/2018 03:11:38

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.142	03:06
Breath Volume		1.871	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.140	03:10
Breath Volume		1.480	03:10
Air Blank		0.000	03:10
Control Test		0.078	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 08/14/2018 09:57

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ARJONA-ARGOTE

DATE/TIME OF TEST: 08/05/2018 01:56:05

Diagnostics Check		OK	01:49
Air Blank		0.000	01:50
Control Test		0.080	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.183	01:51
Breath Volume		2.187	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.193	01:54
Breath Volume		2.351	01:54
Air Blank		0.000	01:55
Control Test		0.079	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:56

Data Download Date/Time: 08/14/2018 09:57

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/15/2018 01:30:44

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.078	01:23
Air Blank		0.000	01:24
Subject Sample	# 1	0.188	01:25
Breath Volume		1.453	01:25
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.193	01:29
Breath Volume		1.203	01:29
Air Blank		0.000	01:29
Control Test		0.078	01:30
Air Blank		0.000	01:30
Diagnostics Check		OK	01:30

Data Download Date/Time: 09/06/2018 13:17

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 08/16/2018 03:46:28

Diagnostics Check		OK	03:43	
Air Blank		0.000	03:44	
Control Test		0.079	03:44	
Air Blank		0.000	03:44	
Subject Sample	# 1		03:45	Subject Test Refused
Breath Volume		0.000	03:45	
Air Blank		0.000	03:45	
Control Test		0.078	03:45	
Air Blank		0.000	03:46	
Diagnostics Check		OK	03:46	

Data Download Date/Time: 09/06/2018 13:17

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: SOCORRO

DATE/TIME OF TEST: 08/18/2018 02:58:42

Diagnostics Check		OK	02:55	
Air Blank		0.000	02:56	
Control Test		0.077	02:56	
Air Blank		0.000	02:57	
Subject Sample	# 1		02:57	Subject Test Refused
Breath Volume		0.000	02:57	
Air Blank		0.000	02:57	
Control Test		0.077	02:58	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 09/06/2018 13:17

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: APOSTOL

DATE/TIME OF TEST: 08/18/2018 23:02:32

Diagnostics Check		OK	22:59	
Air Blank		0.000	23:00	
Control Test		0.078	23:00	
Air Blank		0.000	23:00	
Subject Sample	# 1		23:01	Subject Test Refused
Breath Volume		0.000	23:01	
Air Blank		0.000	23:01	
Control Test		0.078	23:01	
Air Blank		0.000	23:02	
Diagnostics Check		OK	23:02	

Data Download Date/Time: 09/06/2018 13:17

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 08/19/2018 02:47:35

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.078	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.189	02:43
Breath Volume		2.062	02:43
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.197	02:46
Breath Volume		1.546	02:46
Air Blank		0.000	02:46
Control Test		0.077	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 09/06/2018 13:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ROQUE-GONZALEZ

DATE/TIME OF TEST: 08/25/2018 02:24:39

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.077	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.225	02:20
Breath Volume		2.867	02:20
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.228	02:23
Breath Volume		2.988	02:23
Air Blank		0.000	02:23
Control Test		0.076	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 09/06/2018 13:17

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 08/25/2018 05:39:13

Diagnostics Check		OK	05:36	
Air Blank		0.000	05:36	
Control Test		0.078	05:37	
Air Blank		0.000	05:37	
Subject Sample	# 1		05:37	Subject Test Refused
Breath Volume		0.000	05:37	
Air Blank		0.000	05:38	
Control Test		0.078	05:38	
Air Blank		0.000	05:39	
Diagnostics Check		OK	05:39	

Data Download Date/Time: 09/06/2018 13:17

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: CHEVERE

DATE/TIME OF TEST: 08/30/2018 03:32:55

Diagnostics Check		OK	03:26
Air Blank		0.000	03:27
Control Test		0.078	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.172	03:28
Breath Volume		2.035	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:30
Subject Sample	# 2	0.172	03:31
Breath Volume		2.046	03:31
Air Blank		0.000	03:31
Control Test		0.079	03:32
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 09/06/2018 13:17

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: WEDDERBURN

DATE/TIME OF TEST: 08/31/2018 03:34:26

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:32	
Control Test		0.078	03:32	
Air Blank		0.000	03:32	
Subject Sample	# 1		03:33	Subject Test Refused
Breath Volume		0.000	03:33	
Air Blank		0.000	03:33	
Control Test		0.078	03:33	
Air Blank		0.000	03:34	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 09/06/2018 13:17

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: MARTINEZ-SOTO

DATE/TIME OF TEST: 08/31/2018 23:35:59

Diagnostics Check		OK	23:30
Air Blank		0.000	23:30
Control Test		0.078	23:30
Air Blank		0.000	23:31
Subject Sample	# 1	0.061	23:31
Breath Volume		2.335	23:31
Air Blank		0.000	23:32
Air Blank		0.000	23:34
Subject Sample	# 2	0.059	23:34
Breath Volume		2.125	23:34
Air Blank		0.000	23:35
Control Test		0.077	23:35
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 09/06/2018 13:17

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAPANES,R,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: ARBELAEZ SALAZAR

DATE/TIME OF TEST: 08/16/2018 04:24:09

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.079	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.147	04:19
Breath Volume		2.250	04:19
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.139	04:22
Breath Volume		2.746	04:22
Air Blank		0.000	04:23
Control Test		0.079	04:23
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 09/06/2018 13:11

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAPANES,R,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PHILPOT,C
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: FINEGAN

DATE/TIME OF TEST: 08/19/2018 02:57:42

Diagnostics Check		OK	02:51
Air Blank		0.000	02:51
Control Test		0.079	02:51
Air Blank		0.000	02:52
Subject Sample	# 1	0.176	02:53
Breath Volume		3.585	02:53
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.165	02:56
Breath Volume		3.683	02:56
Air Blank		0.000	02:56
Control Test		0.079	02:57
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 09/06/2018 13:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILPOT,C,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PHILPOT,C
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/23/2018 02:52:34

Diagnostics Check		OK	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.141	02:47
Breath Volume		1.289	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.144	02:50
Breath Volume		1.414	02:50
Air Blank		0.000	02:51
Control Test		0.079	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 09/06/2018 13:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILPOT,C,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PHILPOT,C
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CARDONA

DATE/TIME OF TEST: 08/25/2018 02:19:14

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.079	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 1		02:17	Subject Test Refused
Breath Volume		0.000	02:17	
Air Blank		0.000	02:18	
Control Test		0.079	02:18	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 09/06/2018 13:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILPOT,C,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PHILPOT,C
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 08/26/2018 02:30:28

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.079	02:23
Air Blank		0.000	02:23
Subject Sample	# 1	0.114	02:25
Breath Volume		1.414	02:25
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.112	02:28
Breath Volume		1.359	02:28
Air Blank		0.000	02:29
Control Test		0.079	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 09/06/2018 13:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: CHRISTIAN

DATE/TIME OF TEST: 08/27/2018 02:02:55

Diagnostics Check		OK	01:54
Air Blank		0.000	01:55
Control Test		0.078	01:55
Air Blank		0.000	01:55
Subject Sample	# 1	0.262	01:57
Breath Volume		2.066	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.263	02:01
Breath Volume		2.171	02:01
Air Blank		0.000	02:01
Control Test		0.077	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 09/06/2018 13:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: KYELI

DATE/TIME OF TEST: 08/02/2018 03:54:00

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.078	03:42
Air Blank		0.000	03:42
Subject Sample	# 1	0.147	03:45
Breath Volume		1.195	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.178	03:48
Breath Volume		2.441	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 3	0.168	03:52
Breath Volume		2.386	03:52
Air Blank		0.000	03:53
Control Test		0.077	03:53
Air Blank		0.000	03:53
Diagnostics Check		OK	03:54

Data Download Date/Time: 08/14/2018 09:55

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 08/03/2018 01:32:17

Diagnostics Check		OK	01:26
Air Blank		0.000	01:26
Control Test		0.078	01:26
Air Blank		0.000	01:27
Subject Sample	# 1	0.153	01:27
Breath Volume		2.277	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.152	01:30
Breath Volume		1.906	01:30
Air Blank		0.000	01:31
Control Test		0.078	01:31
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 08/14/2018 09:55

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 08/11/2018 08:17:00

Diagnostics Check		OK	08:13	
Air Blank		0.000	08:13	
Control Test		0.078	08:13	
Air Blank		0.000	08:14	
Subject Sample	# 1		08:15	Subject Test Refused
Breath Volume		0.000	08:15	
Air Blank		0.000	08:16	
Control Test		0.078	08:16	
Air Blank		0.000	08:16	
Diagnostics Check		OK	08:17	

Data Download Date/Time: 08/14/2018 09:55

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: APONTE JR

DATE/TIME OF TEST: 08/16/2018 00:52:21

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.078	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.172	00:47
Breath Volume		3.730	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.167	00:50
Breath Volume		3.589	00:50
Air Blank		0.000	00:51
Control Test		0.079	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 09/06/2018 13:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: ABREU

DATE/TIME OF TEST: 08/25/2018 03:31:56

Diagnostics Check		OK	03:24
Air Blank		0.000	03:25
Control Test		0.078	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.124	03:27
Breath Volume		1.726	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:29
Subject Sample	# 2	0.127	03:30
Breath Volume		1.285	03:30
Air Blank		0.000	03:30
Control Test		0.077	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 09/06/2018 13:18

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MOSS,JUSTIN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: POTTS

DATE/TIME OF TEST: 08/21/2018 02:25:15

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.078	02:19
Air Blank		0.000	02:20
Subject Sample	# 1	0.209	02:20
Breath Volume		2.289	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.209	02:23
Breath Volume		2.929	02:23
Air Blank		0.000	02:24
Control Test		0.076	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 09/24/2018 10:29

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: PAULI

DATE/TIME OF TEST: 08/23/2018 04:04:11

Diagnostics Check		OK	03:57
Air Blank		0.000	03:58
Control Test		0.077	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.121	03:59
Breath Volume		1.515	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.124	04:02
Breath Volume		1.101	04:02
Air Blank		0.000	04:03
Control Test		0.077	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 09/24/2018 10:29

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MOSS,JUSTIN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: HILLIMAN

DATE/TIME OF TEST: 08/26/2018 03:55:13

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:54	
Control Test	0.079	03:54	
Air Blank		03:54	RFI Detect
Air Blank	0.000	03:55	

Data Download Date/Time: 09/24/2018 10:29

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MOSS,JUSTIN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: HILLIMAN

DATE/TIME OF TEST: 08/26/2018 04:08:58

Diagnostics Check		OK	03:57	
Air Blank		0.000	03:57	
Control Test		0.078	03:58	
Air Blank		0.000	03:58	
Subject Sample	# 1		04:01	No Sample Provided
Breath Volume		0.000	04:01	
Air Blank		0.000	04:02	
Air Blank		0.000	04:04	
Subject Sample	# 2		04:07	No Sample Provided
Breath Volume		0.000	04:07	
Air Blank		0.000	04:08	
Control Test		0.079	04:08	
Air Blank		0.000	04:08	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 09/24/2018 10:29

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,DAVID,L
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: GUPTILL,DAVID
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: OTHER
ARREST TIME: 13:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:36

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 08/31/2018 14:11:22

Diagnostics Check		OK	14:05
Air Blank		0.000	14:05
Control Test		0.078	14:06
Air Blank		0.000	14:06
Subject Sample	# 1	0.000	14:07
Breath Volume		2.835	14:07
Air Blank		0.000	14:07
Air Blank		0.000	14:09
Subject Sample	# 2	0.000	14:09
Breath Volume		2.992	14:09
Air Blank		0.000	14:10
Control Test		0.077	14:10
Air Blank		0.000	14:11
Diagnostics Check		OK	14:11

Data Download Date/Time: 09/24/2018 10:29

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MASSI

DATE/TIME OF TEST: 08/03/2018 22:23:13

Diagnostics Check		OK	22:11	
Air Blank		0.000	22:12	
Control Test		0.079	22:12	
Air Blank		0.000	22:12	
Subject Sample	# 1	0.132	22:16	Volume Not Met
Breath Volume		0.867	22:16	
Air Blank		0.000	22:16	
Air Blank		0.000	22:18	
Subject Sample	# 2	0.137	22:21	Volume Not Met
Breath Volume		0.324	22:21	
Air Blank		0.000	22:22	
Control Test		0.079	22:22	
Air Blank		0.000	22:23	
Diagnostics Check		OK	22:23	

Data Download Date/Time: 08/08/2018 21:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 08/04/2018 19:41:43

Diagnostics Check		OK	19:34
Air Blank		0.000	19:34
Control Test		0.078	19:35
Air Blank		0.000	19:35
Subject Sample	# 1	0.222	19:37
Breath Volume		1.500	19:37
Air Blank		0.000	19:37
Air Blank		0.000	19:39
Subject Sample	# 2	0.226	19:40
Breath Volume		1.421	19:40
Air Blank		0.000	19:40
Control Test		0.077	19:41
Air Blank		0.000	19:41
Diagnostics Check		OK	19:41

Data Download Date/Time: 08/08/2018 21:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: RAMOS-MEJIAS

DATE/TIME OF TEST: 08/05/2018 02:12:54

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.087	02:08
Breath Volume		2.078	02:08
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.087	02:11
Breath Volume		2.613	02:11
Air Blank		0.000	02:12
Control Test		0.080	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 08/08/2018 21:23

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: RAMOS JR

DATE/TIME OF TEST: 08/10/2018 01:58:35

Diagnostics Check		OK	01:51
Air Blank		0.000	01:51
Control Test		0.080	01:52
Air Blank		0.000	01:52
Subject Sample	# 1	0.152	01:53
Breath Volume		2.460	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.144	01:57
Breath Volume		2.089	01:57
Air Blank		0.000	01:57
Control Test		0.078	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 09/06/2018 23:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: BOSTON

DATE/TIME OF TEST: 08/13/2018 20:34:22

Diagnostics Check		OK	20:25
Air Blank		0.000	20:26
Control Test		0.079	20:26
Air Blank		0.000	20:26
Subject Sample	# 1	0.214	20:29
Breath Volume		1.148	20:29
Air Blank		0.000	20:30
Air Blank		0.000	20:31
Subject Sample	# 2	0.207	20:32
Breath Volume		1.117	20:32
Air Blank		0.000	20:33
Control Test		0.078	20:33
Air Blank		0.000	20:34
Diagnostics Check		OK	20:34

Data Download Date/Time: 09/06/2018 23:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: STELZER,NICHOLAS
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 08/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CARVAJAL-CAYCEDO

DATE/TIME OF TEST: 08/15/2018 00:33:44

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.079	00:27
Air Blank		0.000	00:28
Subject Sample	# 1	0.066	00:29
Breath Volume		3.632	00:29
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.062	00:32
Breath Volume		2.050	00:32
Air Blank		0.000	00:32
Control Test		0.078	00:33
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 09/06/2018 23:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: MCBRIDE-THOMPSON

DATE/TIME OF TEST: 08/23/2018 21:59:20

Diagnostics Check		OK	21:47
Air Blank		0.000	21:47
Control Test		0.080	21:47
Air Blank		0.000	21:48
Subject Sample	# 1	0.327	21:49
Breath Volume		2.078	21:49
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.305	21:54
Breath Volume		2.042	21:54
Air Blank		0.000	21:55
Air Blank		0.000	21:56
Subject Sample	# 3	0.324	21:57
Breath Volume		2.832	21:57
Air Blank		0.000	21:58
Control Test		0.076	21:58
Air Blank		0.000	21:59
Diagnostics Check		OK	21:59

Data Download Date/Time: 09/06/2018 23:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/08/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ENOEX

DATE/TIME OF TEST: 08/29/2018 03:52:53

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.079	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.125	03:48
Breath Volume		1.671	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:50
Subject Sample	# 2	0.106	03:51
Breath Volume		1.726	03:51
Air Blank		0.000	03:51
Control Test		0.078	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 09/06/2018 23:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SANCHEZ SANTOS

DATE/TIME OF TEST: 08/21/2018 21:57:12

Diagnostics Check		OK	21:51
Air Blank		0.000	21:51
Control Test		0.078	21:51
Air Blank		0.000	21:52
Subject Sample	# 1	0.176	21:52
Breath Volume		2.742	21:52
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.168	21:55
Breath Volume		2.628	21:55
Air Blank		0.000	21:56
Control Test		0.078	21:56
Air Blank		0.000	21:57
Diagnostics Check		OK	21:57

Data Download Date/Time: 09/30/2018 08:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: HOCKENSMITH

DATE/TIME OF TEST: 08/23/2018 23:13:12

Diagnostics Check		OK	23:04	
Air Blank		0.000	23:05	
Control Test		0.078	23:05	
Air Blank		0.000	23:06	
Subject Sample	# 1		23:09	No Sample Provided
Breath Volume		0.000	23:09	
Air Blank		0.000	23:09	
Air Blank		0.000	23:11	
Subject Sample	# 2		23:11	Subject Test Refused
Breath Volume		0.000	23:11	
Air Blank		0.000	23:12	
Control Test		0.079	23:12	
Air Blank		0.000	23:13	
Diagnostics Check		OK	23:13	

Data Download Date/Time: 09/30/2018 08:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 08/30/2018 00:51:41

Diagnostics Check		OK	00:42	
Air Blank		0.000	00:43	
Control Test		0.080	00:43	
Air Blank		0.000	00:44	
Subject Sample	# 1		00:47	No Sample Provided
Breath Volume		0.000	00:47	
Air Blank		0.000	00:48	
Air Blank		0.000	00:49	
Subject Sample	# 2		00:50	Subject Test Refused
Breath Volume		0.000	00:50	
Air Blank		0.000	00:50	
Control Test		0.077	00:51	
Air Blank		0.000	00:51	
Diagnostics Check		OK	00:51	

Data Download Date/Time: 09/30/2018 08:51

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,TABARIE,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,TABARIE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: PENIX

DATE/TIME OF TEST: 08/05/2018 00:23:11

Diagnostics Check		OK	00:11	
Air Blank		0.000	00:11	
Control Test		0.078	00:12	
Air Blank		0.000	00:12	
Subject Sample	# 1	0.168	00:15	Volume Not Met
Breath Volume		0.230	00:15	
Air Blank		0.000	00:16	
Air Blank		0.000	00:18	
Subject Sample	# 2	0.179	00:21	Volume Not Met
Breath Volume		0.078	00:21	
Air Blank		0.000	00:21	
Control Test		0.077	00:22	
Air Blank			00:22	RFI Detect
Air Blank		0.000	00:23	

Data Download Date/Time: 08/29/2018 15:00

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,TABARIE,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,TABARIE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BATTAGLIA

DATE/TIME OF TEST: 08/25/2018 03:02:34

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.077	02:57
Air Blank		0.000	02:57
Subject Sample	# 1	0.132	02:58
Breath Volume		2.191	02:58
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.131	03:01
Breath Volume		1.367	03:01
Air Blank		0.000	03:01
Control Test		0.078	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 08/29/2018 15:00

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,TABARIE,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,TABARIE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: PEREZ ORTIZ

DATE/TIME OF TEST: 08/25/2018 05:15:23

Diagnostics Check		OK	05:08
Air Blank		0.000	05:08
Control Test		0.077	05:09
Air Blank		0.000	05:09
Subject Sample	# 1	0.198	05:10
Breath Volume		1.523	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:12
Subject Sample	# 2	0.192	05:13
Breath Volume		1.375	05:13
Air Blank		0.000	05:14
Control Test		0.077	05:14
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 08/29/2018 15:00

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANCO,JUAN,S
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: GOMEZ,RUBEN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: BUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: HERNANDEZ-PEREZ

DATE/TIME OF TEST: 08/18/2018 21:51:50

Diagnostics Check		OK	21:47	
Air Blank		0.000	21:48	
Control Test		0.080	21:48	
Air Blank		0.000	21:49	
Subject Sample	# 1		21:50	Subject Test Refused
Breath Volume		0.000	21:50	
Air Blank		0.000	21:50	
Control Test		0.081	21:51	
Air Blank		0.000	21:51	
Diagnostics Check		OK	21:51	

Data Download Date/Time: 08/22/2018 12:28

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,GABRIEL,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: FERNANDEZ,GABRIEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2016
VIOLATION CODE: BUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: PITTER

DATE/TIME OF TEST: 08/19/2018 21:38:44

Diagnostics Check		OK	21:35	
Air Blank		0.000	21:36	
Control Test		0.080	21:36	
Air Blank		0.000	21:37	
Subject Sample	# 1		21:37	Subject Test Refused
Breath Volume		0.000	21:37	
Air Blank		0.000	21:37	
Control Test		0.081	21:38	
Air Blank		0.000	21:38	
Diagnostics Check		OK	21:38	

Data Download Date/Time: 08/22/2018 12:28

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNAUD,GEORGE,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: GOMEZ,RUBEN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2016
VIOLATION CODE: BUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: POMARICO

DATE/TIME OF TEST: 08/25/2018 20:03:53

Diagnostics Check		OK	19:54	
Air Blank		0.000	19:54	
Control Test		0.080	19:54	
Air Blank		0.000	19:55	
Subject Sample	# 1		19:58	No Sample Provided
Breath Volume		0.000	19:58	
Air Blank		0.000	19:59	
Air Blank		0.000	20:01	
Subject Sample	# 2		20:02	Subject Test Refused
Breath Volume		0.000	20:02	
Air Blank		0.000	20:02	
Control Test		0.081	20:03	
Air Blank		0.000	20:03	
Diagnostics Check		OK	20:03	

Data Download Date/Time: 09/25/2018 10:50

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ADRIAN,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PEREZ,ADRIAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 08/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2016
VIOLATION CODE: BUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: RODRIGUEZ HERNANDEZ

DATE/TIME OF TEST: 08/26/2018 20:43:14

Diagnostics Check		OK	20:36
Air Blank		0.000	20:37
Control Test		0.081	20:37
Air Blank		0.000	20:38
Subject Sample	# 1	0.117	20:38
Breath Volume		1.363	20:38
Air Blank		0.000	20:39
Air Blank		0.000	20:41
Subject Sample	# 2	0.114	20:41
Breath Volume		1.269	20:41
Air Blank		0.000	20:42
Control Test		0.080	20:42
Air Blank		0.000	20:43
Diagnostics Check		OK	20:43

Data Download Date/Time: 09/25/2018 10:50

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: WHEALTON

DATE/TIME OF TEST: 08/15/2018 00:56:34

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.079	00:50
Air Blank		0.000	00:51
Subject Sample	# 1	0.193	00:51
Breath Volume		2.722	00:51
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.190	00:54
Breath Volume		2.371	00:54
Air Blank		0.000	00:55
Control Test		0.078	00:55
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 09/17/2018 15:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: STOCKWELL

DATE/TIME OF TEST: 08/15/2018 23:29:11

Diagnostics Check		OK	23:21
Air Blank		0.000	23:21
Control Test		0.080	23:21
Air Blank		0.000	23:22
Subject Sample	# 1	0.191	23:23
Breath Volume		1.683	23:23
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.198	23:27
Breath Volume		1.812	23:27
Air Blank		0.000	23:28
Control Test		0.078	23:28
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 09/17/2018 15:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADE,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Sewall's Point PD
ARRESTING OFFICER: WADE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sewall's Point PD
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 08/17/2018 17:11:21

Diagnostics Check		OK	17:03	
Air Blank		0.000	17:03	
Control Test		0.079	17:04	
Air Blank		0.000	17:04	
Subject Sample	# 1	0.187	17:05	
Breath Volume		1.500	17:05	
Air Blank		0.000	17:06	
Air Blank		0.000	17:08	
Subject Sample	# 2		17:09	Subject Test Refused
Breath Volume		0.414	17:09	
Air Blank		0.000	17:10	
Control Test		0.079	17:10	
Air Blank		0.000	17:11	
Diagnostics Check		OK	17:11	

Data Download Date/Time: 09/17/2018 15:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISHAM,MATTHEW,H
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ISHAM,MATTHEW
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 08/18/2018 04:26:15

Diagnostics Check		OK	04:16	
Air Blank		0.000	04:17	
Control Test		0.077	04:17	
Air Blank		0.000	04:18	
Subject Sample	# 1		04:21	No Sample Provided
Breath Volume		0.000	04:21	
Air Blank		0.000	04:21	
Air Blank		0.000	04:23	
Subject Sample	# 2		04:24	Subject Test Refused
Breath Volume		0.000	04:24	
Air Blank		0.000	04:25	
Control Test		0.081	04:25	
Air Blank		0.000	04:26	
Diagnostics Check		OK	04:26	

Data Download Date/Time: 09/17/2018 15:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENOVIL,GEDEON,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRENOVIL,GEDEON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: BILLDTRO

DATE/TIME OF TEST: 08/19/2018 01:13:25

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:12	
Control Test	0.000	01:12	RFI Detect
Air Blank	0.000	01:13	

Data Download Date/Time: 09/17/2018 15:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENOVIL,GEDEON,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRENOVIL,GEDEON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: BILLDTRO

DATE/TIME OF TEST: 08/19/2018 01:25:59

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.079	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.288	01:17
Breath Volume		2.273	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.255	01:21
Breath Volume		1.218	01:21
Air Blank		0.000	01:22
Air Blank		0.000	01:23
Subject Sample	# 3	0.289	01:24
Breath Volume		3.148	01:24
Air Blank		0.000	01:25
Control Test		0.077	01:25
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 09/17/2018 15:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/19/2018 22:20:16

Diagnostics Check		OK	22:13
Air Blank		0.000	22:13
Control Test		0.079	22:14
Air Blank		0.000	22:14
Subject Sample	# 1	0.127	22:15
Breath Volume		3.554	22:15
Air Blank		0.000	22:15
Air Blank		0.000	22:17
Subject Sample	# 2	0.122	22:18
Breath Volume		2.773	22:18
Air Blank		0.000	22:19
Control Test		0.079	22:19
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 09/17/2018 15:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FENTON,WILLIAM,B
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FENTON,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: POST

DATE/TIME OF TEST: 08/21/2018 19:34:28

Diagnostics Check	OK	19:32	
Air Blank	0.000	19:33	
Control Test	0.084	19:33	
Air Blank		19:33	Ambient Fail
Air Blank		19:34	Ambient Fail

Data Download Date/Time: 09/17/2018 15:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FENTON,WILLIAM,B
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FENTON,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: POST

DATE/TIME OF TEST: 08/21/2018 19:41:45

Diagnostics Check		OK	19:38	
Air Blank		0.000	19:39	
Control Test		0.080	19:39	
Air Blank		0.000	19:40	
Subject Sample	# 1		19:40	Subject Test Refused
Breath Volume		0.000	19:40	
Air Blank		0.000	19:40	
Control Test		0.080	19:41	
Air Blank		0.000	19:41	
Diagnostics Check		OK	19:41	

Data Download Date/Time: 09/17/2018 15:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MASSE

DATE/TIME OF TEST: 08/22/2018 03:25:18

Diagnostics Check		OK	03:13	
Air Blank		0.000	03:14	
Control Test		0.079	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:18	No Sample Provided
Breath Volume		0.000	03:18	
Air Blank		0.000	03:18	
Air Blank		0.000	03:20	
Subject Sample	# 2		03:23	No Sample Provided
Breath Volume		0.000	03:23	
Air Blank		0.000	03:24	
Control Test		0.080	03:24	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 09/17/2018 15:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MASSE

DATE/TIME OF TEST: 08/22/2018 03:49:42

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:38	
Control Test		0.080	03:39	
Air Blank		0.000	03:39	
Subject Sample	# 1		03:42	No Sample Provided
Breath Volume		0.000	03:42	
Air Blank		0.000	03:43	
Air Blank		0.000	03:45	
Subject Sample	# 2		03:48	No Sample Provided
Breath Volume		0.000	03:48	
Air Blank		0.000	03:48	
Control Test		0.079	03:49	
Air Blank		0.000	03:49	
Diagnostics Check		OK	03:49	

Data Download Date/Time: 09/17/2018 15:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MASSE

DATE/TIME OF TEST: 08/22/2018 03:57:55

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:52	
Control Test		0.077	03:53	
Air Blank		0.000	03:53	
Subject Sample	# 1	0.224	03:56	
Breath Volume		1.320	03:56	
Air Blank			03:57	Purge Fail
Air Blank		0.000	03:57	

Data Download Date/Time: 09/17/2018 15:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MASSE

DATE/TIME OF TEST: 08/22/2018 04:09:49

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:00	
Control Test		0.075	04:00	
Air Blank		0.000	04:01	
Subject Sample	# 1	0.216	04:04	
Breath Volume		1.367	04:04	
Air Blank		0.000	04:05	
Air Blank		0.000	04:06	
Subject Sample	# 2	0.197	04:07	
Breath Volume		1.898	04:07	
Air Blank		0.000	04:08	
Control Test		0.072	04:09	Control Outside Tolerance
Air Blank		0.000	04:09	

Data Download Date/Time: 09/17/2018 15:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENOVIL,GEDEON,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRENOVIL,GEDEON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:13

SUBJECT'S LAST NAME: POINTER

DATE/TIME OF TEST: 08/25/2018 15:46:14

Diagnostics Check		OK	15:33	
Air Blank		0.000	15:33	
Control Test		0.079	15:34	
Air Blank		0.000	15:34	
Subject Sample	# 1		15:37	No Sample Provided
Breath Volume		0.000	15:37	
Air Blank		0.000	15:38	
Air Blank		0.000	15:40	
Subject Sample	# 2	0.278	15:41	
Breath Volume		3.699	15:41	
Air Blank		0.000	15:42	
Air Blank		0.000	15:44	
Subject Sample	# 3	0.262	15:44	
Breath Volume		2.125	15:44	
Air Blank		0.000	15:45	
Control Test		0.075	15:45	
Air Blank		0.000	15:46	
Diagnostics Check		OK	15:46	

Data Download Date/Time: 09/17/2018 15:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: MENDINHALL

DATE/TIME OF TEST: 08/25/2018 19:46:01

Diagnostics Check		OK	19:39
Air Blank		0.000	19:39
Control Test		0.079	19:39
Air Blank		0.000	19:40
Subject Sample	# 1	0.203	19:41
Breath Volume		1.718	19:41
Air Blank		0.000	19:41
Air Blank		0.000	19:43
Subject Sample	# 2	0.195	19:44
Breath Volume		1.800	19:44
Air Blank		0.000	19:45
Control Test		0.079	19:45
Air Blank		0.000	19:45
Diagnostics Check		OK	19:46

Data Download Date/Time: 09/17/2018 15:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: MOSGROVE

DATE/TIME OF TEST: 08/26/2018 00:24:36

Diagnostics Check		OK	00:11	
Air Blank		0.000	00:12	
Control Test		0.079	00:12	
Air Blank		0.000	00:13	
Subject Sample	# 1	0.205	00:16	Volume Not Met
Breath Volume		0.957	00:16	
Air Blank		0.000	00:16	
Air Blank		0.000	00:18	
Subject Sample	# 2	0.233	00:19	No .020 agreement
Breath Volume		2.300	00:19	
Air Blank		0.000	00:19	
Air Blank		0.000	00:21	
Subject Sample	# 3	0.202	00:22	No .020 agreement
Breath Volume		1.617	00:22	
Air Blank		0.000	00:23	
Control Test		0.078	00:23	
Air Blank		0.000	00:24	
Diagnostics Check		OK	00:24	

Data Download Date/Time: 09/17/2018 15:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: LORENZO

DATE/TIME OF TEST: 08/26/2018 23:58:55

Diagnostics Check		OK	23:50
Air Blank		0.000	23:51
Control Test		0.080	23:51
Air Blank		0.000	23:52
Subject Sample	# 1	0.171	23:53
Breath Volume		2.054	23:53
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.170	23:57
Breath Volume		2.125	23:57
Air Blank		0.000	23:57
Control Test		0.078	23:58
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 09/17/2018 15:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,DARIUS,J
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: HOLMES,DARIUS
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: LYTTLE

DATE/TIME OF TEST: 08/29/2018 04:12:15

Diagnostics Check		OK	04:04
Air Blank		0.000	04:05
Control Test		0.079	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.129	04:07
Breath Volume		1.410	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:09
Subject Sample	# 2	0.131	04:10
Breath Volume		1.539	04:10
Air Blank		0.000	04:11
Control Test		0.078	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 09/17/2018 15:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTY,J,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CUTY,J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: RAMIREZ MAS

DATE/TIME OF TEST: 08/31/2018 19:35:44

Diagnostics Check		OK	19:31	
Air Blank		0.000	19:31	
Control Test		0.080	19:32	
Air Blank		0.000	19:32	
Subject Sample	# 1	0.299	19:34	
Breath Volume		1.167	19:34	
Air Blank			19:35	Purge Fail
Air Blank		0.000	19:35	

Data Download Date/Time: 09/17/2018 15:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTY,J,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CUTY,J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: RAMIREZ MAS

DATE/TIME OF TEST: 08/31/2018 19:45:05

Diagnostics Check		OK	19:37
Air Blank		0.000	19:37
Control Test		0.077	19:38
Air Blank		0.000	19:38
Subject Sample	# 1	0.295	19:40
Breath Volume		1.203	19:40
Air Blank		0.000	19:41
Air Blank		0.000	19:42
Subject Sample	# 2	0.292	19:43
Breath Volume		1.285	19:43
Air Blank		0.000	19:44
Control Test		0.077	19:44
Air Blank		0.000	19:45
Diagnostics Check		OK	19:45

Data Download Date/Time: 09/17/2018 15:12

Printed: 16 October 2018

COBRA ©1997-2014 CMI, Inc.