

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEGGETT,JARED,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CARR,JAMES
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 14:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: BOOKER

DATE/TIME OF TEST: 09/24/2018 15:54:58

Diagnostics Check	OK	15:47
Air Blank	0.000	15:48
Control Test	0.080	15:48
Air Blank	0.000	15:48
Subject Sample # 1	0.264	15:50
Breath Volume	2.289	15:50
Air Blank	0.000	15:51
Air Blank	0.000	15:52
Subject Sample # 2	0.253	15:53
Breath Volume	1.839	15:53
Air Blank	0.000	15:54
Control Test	0.078	15:54
Air Blank	0.000	15:54
Diagnostics Check	OK	15:54

Data Download Date/Time: 10/23/2018 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEGGETT,JARED,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: LEGGETT,JARED
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: ESAYAN

DATE/TIME OF TEST: 09/24/2018 23:32:22

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.080	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.116	23:27
Breath Volume	1.394	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.117	23:30
Breath Volume	3.625	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 10/23/2018 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FESTER,NINA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/24/2018 23:42:31

Diagnostics Check	OK	23:34
Air Blank	0.000	23:35
Control Test	0.080	23:35
Air Blank	0.000	23:36
Subject Sample # 1	0.165	23:37
Breath Volume	1.312	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:39
Subject Sample # 2	0.171	23:40
Breath Volume	1.441	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 10/23/2018 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 09/27/2018 04:15:27

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:12	
Control Test	0.080	04:13	
Air Blank	0.000	04:13	
Subject Sample	# 1	04:14	Subject Test Refused
Breath Volume	0.000	04:14	
Air Blank	0.000	04:14	
Control Test	0.080	04:14	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 10/23/2018 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VONADOR,JERARD,E
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: VONADOR,JERARD
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: SCHURMAN

DATE/TIME OF TEST: 09/29/2018 07:28:03

Diagnostics Check	OK	07:21
Air Blank	0.000	07:21
Control Test	0.080	07:21
Air Blank	0.000	07:22
Subject Sample # 1	0.227	07:23
Breath Volume	3.101	07:23
Air Blank	0.000	07:23
Air Blank	0.000	07:25
Subject Sample # 2	0.210	07:26
Breath Volume	2.503	07:26
Air Blank	0.000	07:27
Control Test	0.079	07:27
Air Blank	0.000	07:27
Diagnostics Check	OK	07:28

Data Download Date/Time: 10/23/2018 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEGGETT,JARED,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: LEGGETT,JARED
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: SCOLARO

DATE/TIME OF TEST: 09/30/2018 23:59:53

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.000	23:55
Breath Volume	2.183	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.000	23:58
Breath Volume	2.082	23:58
Air Blank	0.000	23:58
Control Test	0.080	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 10/23/2018 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: CANTHA PALMA

DATE/TIME OF TEST: 09/01/2018 02:30:23

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.099	02:25
Breath Volume	1.699	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.100	02:28
Breath Volume	1.777	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 09/06/2018 15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: BARRETT

DATE/TIME OF TEST: 09/01/2018 05:22:27

Diagnostics Check	OK	05:15
Air Blank	0.000	05:15
Control Test	0.080	05:16
Air Blank	0.000	05:16
Subject Sample # 1	0.156	05:17
Breath Volume	1.171	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.162	05:20
Breath Volume	1.992	05:20
Air Blank	0.000	05:21
Control Test	0.080	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 09/06/2018 15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 09/02/2018 04:11:23

Diagnostics Check		OK	04:04
Air Blank		0.000	04:05
Control Test		0.080	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.223	04:06
Breath Volume		1.296	04:06
Air Blank		0.000	04:07
Air Blank		0.000	04:08
Subject Sample	# 2	0.228	04:09
Breath Volume		1.738	04:09
Air Blank		0.000	04:10
Control Test		0.080	04:10
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 09/06/2018 15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN, R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: SAINT LOUIS

DATE/TIME OF TEST: 09/02/2018 04:47:28

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.080	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.195	04:42
Breath Volume	2.664	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.180	04:45
Breath Volume	1.222	04:45
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 09/06/2018 15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: TRINCHET

DATE/TIME OF TEST: 09/02/2018 05:02:53

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.080	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.152	04:58
Breath Volume	3.136	04:58
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.154	05:01
Breath Volume	3.253	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 09/06/2018 15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: TORPEY-BISHOP

DATE/TIME OF TEST: 09/03/2018 03:24:46

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.100	03:19
Breath Volume	1.187	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.109	03:23
Breath Volume	1.230	03:23
Air Blank	0.000	03:23
Control Test	0.080	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 09/06/2018 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: STAFFORD

DATE/TIME OF TEST: 09/05/2018 04:13:14

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.080	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.144	04:08
Breath Volume	1.445	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.130	04:11
Breath Volume	1.531	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 09/06/2018 15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHILLER,E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: ROEMELING

DATE/TIME OF TEST: 09/06/2018 05:16:20

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.000	05:11
Breath Volume	3.566	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.000	05:14
Breath Volume	2.929	05:14
Air Blank	0.000	05:15
Control Test	0.080	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 09/06/2018 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: BOGHICH

DATE/TIME OF TEST: 09/07/2018 06:41:03

Diagnostics Check	OK	06:33
Air Blank	0.000	06:34
Control Test	0.080	06:34
Air Blank	0.000	06:35
Subject Sample # 1	0.217	06:36
Breath Volume	2.781	06:36
Air Blank	0.000	06:37
Air Blank	0.000	06:38
Subject Sample # 2	0.213	06:39
Breath Volume	2.246	06:39
Air Blank	0.000	06:40
Control Test	0.080	06:40
Air Blank	0.000	06:40
Diagnostics Check	OK	06:41

Data Download Date/Time: 10/04/2018 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: STAVISH

DATE/TIME OF TEST: 09/08/2018 06:08:28

Diagnostics Check	OK	05:59
Air Blank	0.000	06:00
Control Test	0.081	06:00
Air Blank	0.000	06:00
Subject Sample # 1	0.195	06:03
Breath Volume	1.335	06:03
Air Blank	0.000	06:04
Air Blank	0.000	06:06
Subject Sample # 2	0.190	06:06
Breath Volume	1.207	06:06
Air Blank	0.000	06:07
Control Test	0.076	06:07
Air Blank	0.000	06:08
Diagnostics Check	OK	06:08

Data Download Date/Time: 10/04/2018 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: SANTOS GALEAS

DATE/TIME OF TEST: 09/09/2018 01:13:10

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.079	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.156	01:07
Breath Volume	1.949	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.154	01:11
Breath Volume	2.664	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 10/04/2018 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEDINA,P
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: SEPULVEDA III

DATE/TIME OF TEST: 09/09/2018 04:11:24

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.080	04:05
Air Blank		0.000	04:06
Subject Sample	# 1	0.228	04:06
Breath Volume		2.250	04:06
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.227	04:09
Breath Volume		1.644	04:09
Air Blank		0.000	04:10
Control Test		0.080	04:10
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: STRANGE

DATE/TIME OF TEST: 09/09/2018 22:45:47

Diagnostics Check	OK	22:34	
Air Blank	0.000	22:34	
Control Test	0.080	22:35	
Air Blank	0.000	22:35	
Subject Sample	# 1	22:38	No Sample Provided
Breath Volume	0.000	22:38	
Air Blank	0.000	22:39	
Air Blank	0.000	22:41	
Subject Sample	# 2	22:44	No Sample Provided
Breath Volume	0.000	22:44	
Air Blank	0.000	22:44	
Control Test	0.080	22:45	
Air Blank	0.000	22:45	
Diagnostics Check	OK	22:45	

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: STRANGE

DATE/TIME OF TEST: 09/09/2018 22:50:59

Diagnostics Check	OK	22:47	
Air Blank	0.000	22:47	
Control Test	0.080	22:48	
Air Blank	0.000	22:48	
Subject Sample	# 1	22:49	Subject Test Refused
Breath Volume	0.000	22:49	
Air Blank	0.000	22:50	
Control Test	0.080	22:50	
Air Blank	0.000	22:50	
Diagnostics Check	OK	22:50	

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 09/11/2018 05:41:29

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.080	05:32
Air Blank	0.000	05:33
Subject Sample # 1	0.219	05:33
Breath Volume	2.308	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.185	05:36
Breath Volume	1.324	05:36
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 3	0.215	05:39
Breath Volume	3.992	05:39
Air Blank	0.000	05:40
Control Test	0.080	05:40
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: DEABREU

DATE/TIME OF TEST: 09/12/2018 03:29:49

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.080	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.180	03:25
Breath Volume	1.804	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.173	03:28
Breath Volume	1.230	03:28
Air Blank	0.000	03:28
Control Test	0.079	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: MATTESON

DATE/TIME OF TEST: 09/14/2018 03:56:52

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.081	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.000	03:52
Breath Volume	3.136	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.000	03:55
Breath Volume	3.113	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/15/2018 02:15:04

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.081	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.115	02:10
Breath Volume	1.410	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:12
Subject Sample # 2	0.120	02:13
Breath Volume	2.316	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:15

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: MCCALL

DATE/TIME OF TEST: 09/15/2018 03:15:56

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.081	03:10
Air Blank	0.000	03:10
Subject Sample # 1	0.139	03:11
Breath Volume	1.117	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.155	03:14
Breath Volume	2.683	03:14
Air Blank	0.000	03:15
Control Test	0.081	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLORES,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: POLSTER

DATE/TIME OF TEST: 09/15/2018 05:37:28

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.081	05:31
Air Blank	0.000	05:32
Subject Sample # 1	0.129	05:32
Breath Volume	3.593	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.130	05:35
Breath Volume	3.843	05:35
Air Blank	0.000	05:36
Control Test	0.080	05:36
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROSALES,F
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: ISHIE

DATE/TIME OF TEST: 09/15/2018 22:48:23

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:37	
Control Test	0.081	22:37	
Air Blank	0.000	22:37	
Subject Sample	# 1 0.141	22:41	Volume Not Met
Breath Volume	0.015	22:41	
Air Blank	0.000	22:41	
Air Blank	0.000	22:43	
Subject Sample	# 2 0.177	22:46	Volume Not Met
Breath Volume	0.117	22:46	
Air Blank	0.000	22:47	
Control Test	0.075	22:47	
Air Blank	0.000	22:48	
Diagnostics Check	OK	22:48	

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: DE INGENIIS

DATE/TIME OF TEST: 09/16/2018 02:09:13

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:55	
Control Test		0.080	01:55	
Air Blank		0.000	01:56	
Subject Sample	# 1	0.254	01:57	No .020 agreement
Breath Volume		1.128	01:57	
Air Blank		0.000	01:58	
Air Blank		0.000	02:00	
Subject Sample	# 2	0.232	02:03	Volume Not Met
Breath Volume		0.597	02:03	
Air Blank		0.000	02:04	
Air Blank		0.000	02:05	
Subject Sample	# 3	0.275	02:07	No .020 agreement
Breath Volume		1.250	02:07	
Air Blank		0.000	02:08	
Control Test		0.077	02:08	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: ROYSTON

DATE/TIME OF TEST: 09/16/2018 02:49:07

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.104	02:44
Breath Volume	3.585	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:46
Subject Sample # 2	0.104	02:47
Breath Volume	3.687	02:47
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: PUGA

DATE/TIME OF TEST: 09/16/2018 04:44:17

Diagnostics Check	OK	04:38
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.077	04:39
Breath Volume	3.273	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.072	04:42
Breath Volume	2.750	04:42
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHILLER, E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: JONES II

DATE/TIME OF TEST: 09/16/2018 23:03:05

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.080	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.119	22:58
Breath Volume	1.507	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:00
Subject Sample # 2	0.115	23:01
Breath Volume	1.128	23:01
Air Blank	0.000	23:02
Control Test	0.080	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:03

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: OVANDO ROBLES

DATE/TIME OF TEST: 09/19/2018 23:13:41

Diagnostics Check	OK	23:06
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.125	23:09
Breath Volume	1.695	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.119	23:12
Breath Volume	1.230	23:12
Air Blank	0.000	23:12
Control Test	0.079	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: DURANT

DATE/TIME OF TEST: 09/20/2018 04:08:19

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.125	04:02
Breath Volume	1.972	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:04
Subject Sample # 2	0.118	04:06
Breath Volume	1.542	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI, J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: CECIL

DATE/TIME OF TEST: 09/20/2018 22:28:13

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.080	22:22
Air Blank	0.000	22:23
Subject Sample # 1	0.138	22:23
Breath Volume	3.597	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.141	22:26
Breath Volume	3.656	22:26
Air Blank	0.000	22:27
Control Test	0.079	22:27
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: BARRIENTOS

DATE/TIME OF TEST: 09/23/2018 01:29:17

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.080	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.127	01:24
Breath Volume	2.152	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:26
Subject Sample # 2	0.130	01:27
Breath Volume	2.289	01:27
Air Blank	0.000	01:28
Control Test	0.079	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 09/23/2018 02:19:00

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.127	02:14
Breath Volume	3.339	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.133	02:17
Breath Volume	3.687	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:19

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: MAYORGA

DATE/TIME OF TEST: 09/23/2018 03:38:35

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.183	03:33
Breath Volume	3.371	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.179	03:36
Breath Volume	3.054	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: CHAGALA-GUEVARA

DATE/TIME OF TEST: 09/23/2018 19:51:11

Diagnostics Check	OK	19:44
Air Blank	0.000	19:45
Control Test	0.080	19:45
Air Blank	0.000	19:46
Subject Sample # 1	0.188	19:46
Breath Volume	2.007	19:46
Air Blank	0.000	19:47
Air Blank	0.000	19:49
Subject Sample # 2	0.186	19:49
Breath Volume	1.238	19:49
Air Blank	0.000	19:50
Control Test	0.079	19:50
Air Blank	0.000	19:51
Diagnostics Check	OK	19:51

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: ENDERS

DATE/TIME OF TEST: 09/23/2018 23:37:01

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.226	23:32
Breath Volume	3.371	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:34
Subject Sample # 2	0.215	23:35
Breath Volume	1.296	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:37

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: ATEN

DATE/TIME OF TEST: 09/25/2018 04:34:15

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.079	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.154	04:29
Breath Volume		2.062	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:31
Subject Sample	# 2	0.158	04:32
Breath Volume		3.003	04:32
Air Blank		0.000	04:33
Control Test		0.080	04:33
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 09/25/2018 22:47:13

Diagnostics Check	OK	22:35	
Air Blank	0.000	22:36	
Control Test	0.080	22:36	
Air Blank	0.000	22:37	
Subject Sample	# 1	22:40	No Sample Provided
Breath Volume	0.000	22:40	
Air Blank	0.000	22:40	
Air Blank	0.000	22:42	
Subject Sample	# 2	22:45	No Sample Provided
Breath Volume	0.000	22:45	
Air Blank	0.000	22:46	
Control Test	0.080	22:46	
Air Blank	0.000	22:47	
Diagnostics Check	OK	22:47	

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 09/25/2018 23:04:34

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:52	
Control Test	0.080	22:53	
Air Blank	0.000	22:53	
Subject Sample	# 1	22:56	No Sample Provided
Breath Volume	0.000	22:56	
Air Blank	0.000	22:57	
Air Blank	0.000	22:59	
Subject Sample	# 2	0.126	22:59
Breath Volume	1.109	22:59	
Air Blank	0.000	23:00	
Air Blank	0.000	23:02	
Subject Sample	# 3	0.124	23:02
Breath Volume	2.667	23:02	
Air Blank	0.000	23:03	
Control Test	0.080	23:03	
Air Blank	0.000	23:04	
Diagnostics Check	OK	23:04	

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: PINZON

DATE/TIME OF TEST: 09/28/2018 05:42:44

Diagnostics Check	OK	05:32
Air Blank	0.000	05:33
Control Test	0.080	05:33
Air Blank	0.000	05:33
Subject Sample # 1	0.093	05:36
Breath Volume	1.410	05:36
Air Blank	0.000	05:37
Air Blank	0.000	05:38
Subject Sample # 2	0.091	05:41
Breath Volume	1.449	05:41
Air Blank	0.000	05:41
Control Test	0.080	05:42
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: PURVIS

DATE/TIME OF TEST: 09/28/2018 20:13:57

Diagnostics Check	OK	20:07
Air Blank	0.000	20:07
Control Test	0.080	20:08
Air Blank	0.000	20:08
Subject Sample # 1	0.000	20:09
Breath Volume	1.996	20:09
Air Blank	0.000	20:09
Air Blank	0.000	20:11
Subject Sample # 2	0.000	20:12
Breath Volume	1.906	20:12
Air Blank	0.000	20:13
Control Test	0.080	20:13
Air Blank	0.000	20:13
Diagnostics Check	OK	20:13

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 09/29/2018 04:00:55

Diagnostics Check	OK	03:54
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.102	03:56
Breath Volume	3.808	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:58
Subject Sample # 2	0.091	03:59
Breath Volume	3.910	03:59
Air Blank	0.000	03:59
Control Test	0.080	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: RAJENDRAN

DATE/TIME OF TEST: 09/30/2018 01:04:39

Diagnostics Check		OK	00:57
Air Blank		0.000	00:58
Control Test		0.081	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.136	00:59
Breath Volume		1.191	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.130	01:03
Breath Volume		1.171	01:03
Air Blank		0.000	01:03
Control Test		0.080	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: RIVERA GUTIERREZ

DATE/TIME OF TEST: 09/30/2018 04:25:07

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.080	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.186	04:20
Breath Volume	2.042	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.184	04:23
Breath Volume	2.210	04:23
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLORES,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: GRULLON MONTAN

DATE/TIME OF TEST: 09/30/2018 05:20:32

Diagnostics Check	OK	05:13
Air Blank	0.000	05:13
Control Test	0.081	05:14
Air Blank	0.000	05:14
Subject Sample # 1	0.149	05:16
Breath Volume	1.425	05:16
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.143	05:18
Breath Volume	2.191	05:18
Air Blank	0.000	05:19
Control Test	0.080	05:19
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MONTESINO

DATE/TIME OF TEST: 09/30/2018 23:12:05

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.081	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.080	23:07
Breath Volume	2.265	23:07
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.082	23:10
Breath Volume	2.070	23:10
Air Blank	0.000	23:11
Control Test	0.080	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 10/04/2018 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,D,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CONWAY,P
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 09/01/2018 03:06:56

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.080	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.109	03:02
Breath Volume	1.898	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.118	03:05
Breath Volume	1.722	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 09/18/2018 22:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,MITA,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GREEN,SETH
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: BRERETON

DATE/TIME OF TEST: 09/01/2018 09:06:38

Diagnostics Check		OK	09:00
Air Blank		0.000	09:00
Control Test		0.079	09:01
Air Blank		0.000	09:01
Subject Sample	# 1	0.126	09:02
Breath Volume		2.273	09:02
Air Blank		0.000	09:02
Air Blank		0.000	09:04
Subject Sample	# 2	0.130	09:05
Breath Volume		1.828	09:05
Air Blank		0.000	09:05
Control Test		0.079	09:06
Air Blank		0.000	09:06
Diagnostics Check		OK	09:06

Data Download Date/Time: 09/18/2018 22:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CONWAY,PHILIP
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: OHLRICH

DATE/TIME OF TEST: 09/04/2018 03:20:27

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.154	03:15
Breath Volume	3.800	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.148	03:18
Breath Volume	3.617	03:18
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 09/18/2018 22:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESCARTES,PETER,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MEEHAN,SHANE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: VILHARD

DATE/TIME OF TEST: 09/09/2018 17:51:39

Diagnostics Check	OK	17:44
Air Blank	0.000	17:44
Control Test	0.080	17:44
Air Blank	0.000	17:45
Subject Sample # 1	0.124	17:46
Breath Volume	2.070	17:46
Air Blank	0.000	17:47
Air Blank	0.000	17:48
Subject Sample # 2	0.113	17:49
Breath Volume	2.183	17:49
Air Blank	0.000	17:50
Control Test	0.076	17:51
Air Blank	0.000	17:51
Diagnostics Check	OK	17:51

Data Download Date/Time: 09/18/2018 22:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINEY,CRISTA,N
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PRATHER,JASON
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: CAMP

DATE/TIME OF TEST: 09/10/2018 22:16:05

Diagnostics Check	OK	22:07
Air Blank	0.000	22:07
Control Test	0.080	22:08
Air Blank	0.000	22:08
Subject Sample # 1	0.000	22:09
Breath Volume	1.117	22:09
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample # 2	0.000	22:14
Breath Volume	1.320	22:14
Air Blank	0.000	22:15
Control Test	0.077	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:16

Data Download Date/Time: 09/18/2018 22:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINEY,CRISTA,N
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: OLDHAM,TAYLOR
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: EUBANKS JR.

DATE/TIME OF TEST: 09/11/2018 06:18:29

Diagnostics Check	OK	06:12
Air Blank	0.000	06:12
Control Test	0.080	06:13
Air Blank	0.000	06:13
Subject Sample # 1	0.000	06:14
Breath Volume	5.488	06:14
Air Blank	0.000	06:14
Air Blank	0.000	06:16
Subject Sample # 2	0.000	06:17
Breath Volume	5.386	06:17
Air Blank	0.000	06:17
Control Test	0.077	06:17
Air Blank	0.000	06:18
Diagnostics Check	OK	06:18

Data Download Date/Time: 09/18/2018 22:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH,E,W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: O'BARR,J
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: DEANE

DATE/TIME OF TEST: 09/13/2018 21:17:49

Diagnostics Check	OK	21:11
Air Blank	0.000	21:11
Control Test	0.080	21:12
Air Blank	0.000	21:12
Subject Sample # 1	0.164	21:13
Breath Volume	2.105	21:13
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.172	21:16
Breath Volume	2.457	21:16
Air Blank	0.000	21:16
Control Test	0.079	21:17
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 09/18/2018 22:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,MITA,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MCCARTHY,EVAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:19

SUBJECT'S LAST NAME: VANLEER

DATE/TIME OF TEST: 09/14/2018 14:45:14

Diagnostics Check	OK	14:42	
Air Blank	0.000	14:42	
Control Test	0.079	14:43	
Air Blank	0.000	14:43	
Subject Sample	# 1	14:43	Subject Test Refused
Breath Volume	0.000	14:43	
Air Blank	0.000	14:44	
Control Test	0.079	14:44	
Air Blank	0.000	14:45	
Diagnostics Check	OK	14:45	

Data Download Date/Time: 09/18/2018 22:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VINCI-BUSTOS,ROSANNA,L
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: JACKSON,BRYAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MAINES

DATE/TIME OF TEST: 09/18/2018 01:07:16

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.159	01:02
Breath Volume	3.636	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.150	01:05
Breath Volume	1.789	01:05
Air Blank	0.000	01:06
Control Test	0.075	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 09/18/2018 22:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GADDIE,KYLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: HYMAN

DATE/TIME OF TEST: 09/18/2018 19:16:15

Diagnostics Check	OK	19:08
Air Blank	0.000	19:09
Control Test	0.080	19:09
Air Blank	0.000	19:10
Subject Sample # 1	0.000	19:11
Breath Volume	1.777	19:11
Air Blank	0.000	19:12
Air Blank	0.000	19:14
Subject Sample # 2	0.000	19:14
Breath Volume	3.085	19:14
Air Blank	0.000	19:15
Control Test	0.080	19:15
Air Blank	0.000	19:16
Diagnostics Check	OK	19:16

Data Download Date/Time: 09/18/2018 22:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINEY,CRISTA,N
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: OLDHAM,TAYLOR
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 09/26/2018 03:10:28

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.080	03:08	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Control Test	0.080	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 10/29/2018 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WALLIS,RICHARD
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: KESSINGER

DATE/TIME OF TEST: 09/29/2018 02:31:52

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:29	
Control Test	0.080	02:29	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:30	Subject Test Refused
Breath Volume	0.000	02:30	
Air Blank	0.000	02:30	
Control Test	0.080	02:31	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

Data Download Date/Time: 10/29/2018 17:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CARTER,B
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 09/29/2018 23:50:58

Diagnostics Check	OK	23:44
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.207	23:46
Breath Volume	2.242	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:48
Subject Sample # 2	0.205	23:49
Breath Volume	1.757	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 10/29/2018 17:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA, MICHAEL, A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LIMA, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ARITA RODENZO

DATE/TIME OF TEST: 09/30/2018 22:08:49

Diagnostics Check		OK	21:55	
Air Blank		0.000	21:55	
Control Test		0.080	21:56	
Air Blank		0.000	21:56	
Subject Sample	# 1	0.156	21:59	Slope Not Level
Breath Volume		1.312	21:59	
Air Blank		0.000	22:00	
Air Blank		0.000	22:02	
Subject Sample	# 2	0.171	22:04	
Breath Volume		2.183	22:04	
Air Blank		0.000	22:04	
Air Blank		0.000	22:06	
Subject Sample	# 3	0.176	22:07	
Breath Volume		2.574	22:07	
Air Blank		0.000	22:07	
Control Test		0.078	22:08	
Air Blank		0.000	22:08	
Diagnostics Check		OK	22:08	

Data Download Date/Time: 10/29/2018 17:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHMID,ERNEST,J
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: LOPEZ,TIMOTHY
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 08/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: HAMELERS-IRONS

DATE/TIME OF TEST: 09/02/2018 11:40:11

Diagnostics Check	OK	11:33
Air Blank	0.000	11:33
Control Test	0.075	11:34
Air Blank	0.000	11:34
Subject Sample # 1	0.160	11:35
Breath Volume	5.523	11:35
Air Blank	0.000	11:36
Air Blank	0.000	11:37
Subject Sample # 2	0.152	11:38
Breath Volume	4.246	11:38
Air Blank	0.000	11:39
Control Test	0.076	11:39
Air Blank	0.000	11:40
Diagnostics Check	OK	11:40

Data Download Date/Time: 09/09/2018 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROONQUIST,DONALD,R
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: LOPEZ,TIM
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 09/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: MUNIZ TORRES

DATE/TIME OF TEST: 09/27/2018 21:18:09

Diagnostics Check	OK	21:16	
Air Blank	0.000	21:17	
Control Test	0.068	21:17	Control Outside Tolerance
Air Blank	0.000	21:18	

Data Download Date/Time: 10/14/2018 15:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROONQUIST,DON,R
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: LOPEZ,TIM
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 09/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: MUNIZ TORRES

DATE/TIME OF TEST: 09/27/2018 21:30:48

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.075	21:24
Air Blank	0.000	21:24
Subject Sample # 1	0.159	21:25
Breath Volume	2.707	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.145	21:28
Breath Volume	1.804	21:28
Air Blank	0.000	21:29
Control Test	0.075	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 10/14/2018 15:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROONQUIST,DON,R
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: LOPEZ,TIMOTHY
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 09/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 09/27/2018 21:42:08

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:41	
Control Test	0.074	21:41	Control Outside Tolerance
Air Blank	0.000	21:42	

Data Download Date/Time: 10/14/2018 15:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROONQUIST,DON,R
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: LOPEZ,TIMMOTHY
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 09/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 09/27/2018 21:53:08

Diagnostics Check	OK	21:45	
Air Blank	0.000	21:46	
Control Test	0.077	21:46	
Air Blank	0.000	21:47	
Subject Sample # 1	0.154	21:48	
Breath Volume	2.523	21:48	
Air Blank	0.000	21:49	
Air Blank	0.000	21:50	
Subject Sample # 2	0.161	21:51	
Breath Volume	2.542	21:51	
Air Blank	0.000	21:52	
Control Test	0.072	21:52	Control Outside Tolerance
Air Blank	0.000	21:53	

Data Download Date/Time: 10/14/2018 15:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEEN,ALYIA,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: REILLY,GEORGE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: STUART

DATE/TIME OF TEST: 09/01/2018 04:52:50

Diagnostics Check		OK	04:41	
Air Blank		0.000	04:41	
Control Test		0.079	04:41	
Air Blank		0.000	04:42	
Subject Sample	# 1	0.101	04:45	Volume Not Met
Breath Volume		0.117	04:45	
Air Blank		0.000	04:46	
Air Blank		0.000	04:47	
Subject Sample	# 2	0.142	04:51	Volume Not Met
Breath Volume		0.023	04:51	
Air Blank		0.000	04:51	
Control Test		0.075	04:52	
Air Blank		0.000	04:52	
Diagnostics Check		OK	04:52	

Data Download Date/Time: 09/25/2018 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEEN,ALYIA,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: REILLY,GEORGE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: STUART

DATE/TIME OF TEST: 09/01/2018 05:13:39

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:59	
Control Test	0.076	04:59	
Air Blank	0.000	04:59	
Subject Sample	# 1 0.157	05:01	
Breath Volume	1.777	05:01	
Air Blank	0.000	05:02	
Air Blank	0.000	05:03	
Subject Sample	# 2 0.128	05:07	Volume Not Met
Breath Volume	0.613	05:07	
Air Blank	0.000	05:07	
Air Blank	0.000	05:09	
Subject Sample	# 3 0.151	05:11	
Breath Volume	1.765	05:11	
Air Blank	0.000	05:12	
Control Test	0.075	05:13	
Air Blank	0.000	05:13	
Diagnostics Check	OK	05:13	

Data Download Date/Time: 09/25/2018 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEEN,ALYIA,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: RENTERIA,BARTOLO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: PACE

DATE/TIME OF TEST: 09/06/2018 03:35:24

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.148	03:30
Breath Volume	2.472	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:32
Subject Sample # 2	0.145	03:33
Breath Volume	1.730	03:33
Air Blank	0.000	03:34
Control Test	0.076	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 09/25/2018 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINO,ROBERT,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BACHAR,B
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: LINO

DATE/TIME OF TEST: 09/11/2018 22:45:16

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.078	22:43	
Air Blank	0.000	22:43	
Subject Sample	# 1	22:43	Subject Test Refused
Breath Volume	0.000	22:43	
Air Blank	0.000	22:44	
Control Test	0.078	22:44	
Air Blank	0.000	22:45	
Diagnostics Check	OK	22:45	

Data Download Date/Time: 09/25/2018 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISSI,DAN,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HORNE,A
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: FLORES-HERNANDEZ

DATE/TIME OF TEST: 09/12/2018 04:22:42

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:20	
Control Test	0.078	04:20	
Air Blank	0.000	04:21	
Subject Sample	# 1	04:21	Subject Test Refused
Breath Volume	0.000	04:21	
Air Blank	0.000	04:21	
Control Test	0.079	04:22	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:22	

Data Download Date/Time: 09/25/2018 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEEN,ALYIA,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CASAS,STEPHANIE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/12/2018 23:07:44

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.079	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.048	23:02
Breath Volume	1.535	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:04
Subject Sample # 2	0.047	23:06
Breath Volume	1.617	23:06
Air Blank	0.000	23:06
Control Test	0.080	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 09/25/2018 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONG,KENNETH,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CASAS,JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SEEPERSAUD

DATE/TIME OF TEST: 09/13/2018 01:52:13

Diagnostics Check		OK	01:45
Air Blank		0.000	01:45
Control Test		0.080	01:46
Air Blank		0.000	01:46
Subject Sample	# 1	0.243	01:47
Breath Volume		2.460	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.236	01:50
Breath Volume		1.492	01:50
Air Blank		0.000	01:51
Control Test		0.079	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 09/25/2018 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,BRITTANIE,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ANJOS,MICHAEL
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: GONZALEZ PEREZ

DATE/TIME OF TEST: 09/13/2018 19:16:45

Diagnostics Check	OK	19:13	
Air Blank	0.000	19:14	
Control Test	0.079	19:14	
Air Blank	0.000	19:15	
Subject Sample	# 1	19:15	Subject Test Refused
Breath Volume	0.000	19:15	
Air Blank	0.000	19:15	
Control Test	0.079	19:16	
Air Blank	0.000	19:16	
Diagnostics Check	OK	19:16	

Data Download Date/Time: 09/25/2018 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,BRITTANIE,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: RENTERIA,BARTOLO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: STOREY

DATE/TIME OF TEST: 09/14/2018 22:37:20

Diagnostics Check	OK	22:34	
Air Blank	0.000	22:34	
Control Test	0.079	22:35	
Air Blank	0.000	22:35	
Subject Sample	# 1	22:35	Subject Test Refused
Breath Volume	0.000	22:35	
Air Blank	0.000	22:36	
Control Test	0.079	22:36	
Air Blank	0.000	22:37	
Diagnostics Check	OK	22:37	

Data Download Date/Time: 09/25/2018 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONG,KENNETH,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: RAFALKO,TRAVIS
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BRODERICK

DATE/TIME OF TEST: 09/21/2018 03:07:49

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.079	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.198	03:02
Breath Volume		2.695	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.188	03:06
Breath Volume		2.210	03:06
Air Blank		0.000	03:06
Control Test		0.079	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 09/25/2018 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,GARY,M
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: KRIGGER,ROHALIO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: CARRIER

DATE/TIME OF TEST: 09/21/2018 07:11:19

Diagnostics Check	OK	07:08	
Air Blank	0.000	07:08	
Control Test	0.079	07:09	
Air Blank	0.000	07:09	
Subject Sample	# 1	07:09	Subject Test Refused
Breath Volume	0.000	07:09	
Air Blank	0.000	07:10	
Control Test	0.079	07:10	
Air Blank	0.000	07:11	
Diagnostics Check	OK	07:11	

Data Download Date/Time: 09/25/2018 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFALKO,TRAVIS,K
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ALVAREZ,AMANDA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: TORRENTE

DATE/TIME OF TEST: 09/22/2018 00:20:14

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:17	
Control Test	0.079	00:17	
Air Blank	0.000	00:18	
Subject Sample	# 1	00:18	Subject Test Refused
Breath Volume	0.000	00:18	
Air Blank	0.000	00:19	
Control Test	0.080	00:19	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

Data Download Date/Time: 09/25/2018 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFALKO,TRAVIS,K
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BACHAR,BRAIN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: WINGEIR

DATE/TIME OF TEST: 09/23/2018 04:08:40

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.080	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.168	04:04
Breath Volume		2.093	04:04
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.165	04:06
Breath Volume		2.132	04:06
Air Blank		0.000	04:07
Control Test		0.078	04:08
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 09/25/2018 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONG,KENNETH,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: RICCIARDI,AMANDA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: LAZENBY

DATE/TIME OF TEST: 09/29/2018 17:54:20

Diagnostics Check	OK	17:51	
Air Blank	0.000	17:51	
Control Test	0.078	17:52	
Air Blank	0.000	17:52	
Subject Sample	# 1	17:52	Subject Test Refused
Breath Volume	0.000	17:52	
Air Blank	0.000	17:53	
Control Test	0.078	17:53	
Air Blank	0.000	17:54	
Diagnostics Check	OK	17:54	

Data Download Date/Time: 11/14/2018 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN CAMP,JEFFERY,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VAN CAMP,JEFFERY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: LONANO

DATE/TIME OF TEST: 09/30/2018 22:31:52

Diagnostics Check	OK	22:21
Air Blank	0.000	22:22
Control Test	0.077	22:22
Air Blank	0.000	22:23
Subject Sample # 1	0.081	22:25
Breath Volume	1.589	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:27
Subject Sample # 2	0.074	22:30
Breath Volume	2.101	22:30
Air Blank	0.000	22:30
Control Test	0.075	22:31
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 11/14/2018 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/30/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: JANER

DATE/TIME OF TEST: 09/01/2018 02:15:03

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:12	
Control Test	0.079	02:12	
Air Blank	0.000	02:13	
Subject Sample	# 1	02:13	Subject Test Refused
Breath Volume	0.000	02:13	
Air Blank	0.000	02:14	
Control Test	0.080	02:14	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:15	

Data Download Date/Time: 09/30/2018 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: VICIOSO

DATE/TIME OF TEST: 09/09/2018 03:58:15

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.080	03:56	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:57	
Control Test	0.079	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 09/30/2018 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/12/2018 00:42:42

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.079	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.178	00:37
Breath Volume	1.542	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.189	00:41
Breath Volume	1.304	00:41
Air Blank	0.000	00:41
Control Test	0.080	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 09/30/2018 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: BONET TORRES

DATE/TIME OF TEST: 09/13/2018 02:48:02

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.080	02:45	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:47	
Control Test	0.080	02:47	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:48	

Data Download Date/Time: 09/30/2018 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 09/13/2018 23:44:01

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:41	
Control Test	0.080	23:41	
Air Blank	0.000	23:42	
Subject Sample	# 1	23:42	Subject Test Refused
Breath Volume	0.000	23:42	
Air Blank	0.000	23:43	
Control Test	0.080	23:43	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:44	

Data Download Date/Time: 09/30/2018 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: VARGAS,DALE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: ZAPATA CORREA

DATE/TIME OF TEST: 09/22/2018 04:08:46

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.080	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.224	04:04
Breath Volume		1.367	04:04
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.224	04:07
Breath Volume		1.996	04:07
Air Blank		0.000	04:07
Control Test		0.079	04:08
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 09/30/2018 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMARAL,WILKER,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: AMARAL,WILKER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 09/01/2018 04:43:26

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:41	
Control Test	0.078	04:41	
Air Blank	0.000	04:41	
Subject Sample	# 1	04:42	Subject Test Refused
Breath Volume	0.000	04:42	
Air Blank	0.000	04:42	
Control Test	0.078	04:42	
Air Blank	0.000	04:43	
Diagnostics Check	OK	04:43	

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRILL,JACOB,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MERRILL,JACOB
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 09/03/2018 05:16:44

Diagnostics Check	OK	05:05	
Air Blank	0.000	05:05	
Control Test	0.079	05:06	
Air Blank	0.000	05:06	
Subject Sample	# 1	05:09	No Sample Provided
Breath Volume	0.000	05:09	
Air Blank	0.000	05:10	
Air Blank	0.000	05:12	
Subject Sample	# 2	05:15	No Sample Provided
Breath Volume	0.000	05:15	
Air Blank	0.000	05:15	
Control Test	0.079	05:16	
Air Blank	0.000	05:16	
Diagnostics Check	OK	05:16	

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADOVA,BRIAN,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FESTER,NINA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MENDEZ SANTOS

DATE/TIME OF TEST: 09/06/2018 03:23:03

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:20	
Control Test	0.079	03:20	
Air Blank	0.000	03:21	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:22	
Control Test	0.078	03:22	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRILL,JACOB,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: ADAME

DATE/TIME OF TEST: 09/07/2018 03:39:23

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:37	
Control Test	0.079	03:37	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:38	Subject Test Refused
Breath Volume	0.000	03:38	
Air Blank	0.000	03:38	
Control Test	0.079	03:38	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: DAHLQUIST

DATE/TIME OF TEST: 09/07/2018 23:46:43

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.079	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.160	23:42
Breath Volume	4.128	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.161	23:45
Breath Volume	4.265	23:45
Air Blank	0.000	23:45
Control Test	0.076	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: SANCHEZ-RAMIREZ

DATE/TIME OF TEST: 09/08/2018 02:11:55

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.078	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.160	02:07
Breath Volume	2.781	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:09
Subject Sample # 2	0.156	02:10
Breath Volume	2.480	02:10
Air Blank	0.000	02:10
Control Test	0.078	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CONDE ALTAMIRANO

DATE/TIME OF TEST: 09/08/2018 02:19:22

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:16	
Control Test	0.078	02:17	
Air Blank	0.000	02:17	
Subject Sample	# 1	02:17	Subject Test Refused
Breath Volume	0.000	02:17	
Air Blank	0.000	02:18	
Control Test	0.078	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VONADOR, JERARD, E
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: DESANTIS, ADAM
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: GUERRA

DATE/TIME OF TEST: 09/08/2018 05:54:09

Diagnostics Check	OK	05:42		
Air Blank	0.000	05:42		
Control Test	0.079	05:43		
Air Blank	0.000	05:43		
Subject Sample	# 1	05:46	No Sample Provided	
Breath Volume	0.000	05:46		
Air Blank	0.000	05:47		
Air Blank	0.000	05:49		
Subject Sample	# 2	0.113	05:52	Volume Not Met
Breath Volume	1.093	05:52		
Air Blank	0.000	05:53		
Control Test	0.079	05:53		
Air Blank	0.000	05:54		
Diagnostics Check	OK	05:54		

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VONADOR, JERARD, E
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: DESANTIS, ADAM
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: GUERRA

DATE/TIME OF TEST: 09/08/2018 06:04:43

Diagnostics Check	OK	05:56
Air Blank	0.000	05:56
Control Test	0.077	05:57
Air Blank	0.000	05:57
Subject Sample # 1	0.120	05:58
Breath Volume	1.218	05:58
Air Blank	0.000	05:59
Air Blank	0.000	06:00
Subject Sample # 2	0.111	06:03
Breath Volume	1.757	06:03
Air Blank	0.000	06:03
Control Test	0.077	06:04
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: GARCIA-RAMIREZ

DATE/TIME OF TEST: 09/08/2018 06:12:03

Diagnostics Check	OK	06:08	
Air Blank	0.000	06:09	
Control Test	0.078	06:09	
Air Blank	0.000	06:10	
Subject Sample	# 1	06:10	Subject Test Refused
Breath Volume	0.000	06:10	
Air Blank	0.000	06:11	
Control Test	0.078	06:11	
Air Blank	0.000	06:11	
Diagnostics Check	OK	06:12	

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: GOMEZ VALLEJO

DATE/TIME OF TEST: 09/09/2018 04:04:31

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.190	03:57
Breath Volume	1.257	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.200	04:02
Breath Volume	1.386	04:02
Air Blank	0.000	04:03
Control Test	0.078	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 09/09/2018 19:16:27

Diagnostics Check	OK	19:13	
Air Blank	0.000	19:14	
Control Test	0.079	19:14	
Air Blank	0.000	19:14	
Subject Sample	# 1	19:15	Subject Test Refused
Breath Volume	0.000	19:15	
Air Blank	0.000	19:15	
Control Test	0.079	19:15	
Air Blank	0.000	19:16	
Diagnostics Check	OK	19:16	

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 09/10/2018 00:08:32

Diagnostics Check		OK	00:02
Air Blank		0.000	00:02
Control Test		0.078	00:02
Air Blank		0.000	00:03
Subject Sample	# 1	0.155	00:03
Breath Volume		2.003	00:03
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.150	00:06
Breath Volume		2.078	00:06
Air Blank		0.000	00:07
Control Test		0.078	00:07
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,DANIEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WATSON,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: TORRE

DATE/TIME OF TEST: 09/10/2018 00:49:44

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.079	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.052	00:45
Breath Volume	2.832	00:45
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.049	00:48
Breath Volume	2.199	00:48
Air Blank	0.000	00:48
Control Test	0.077	00:49
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 09/10/2018 01:27:15

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.078	01:19
Air Blank	0.000	01:20
Subject Sample # 1	0.062	01:21
Breath Volume	1.835	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.059	01:25
Breath Volume	1.234	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEGGETT,JARED,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: LEGGETT,JARED
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 09/11/2018 23:32:12

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.078	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.182	23:27
Breath Volume	3.140	23:27
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.190	23:30
Breath Volume	2.437	23:30
Air Blank	0.000	23:31
Control Test	0.078	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA-MORALES, ENRIQUE,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY, CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: BARRERA-MORALES

DATE/TIME OF TEST: 09/14/2018 04:42:24

Diagnostics Check	OK	04:33	
Air Blank	0.000	04:34	
Control Test	0.079	04:34	
Air Blank	0.000	04:35	
Subject Sample	# 1	04:38	No Sample Provided
Breath Volume	0.000	04:38	
Air Blank	0.000	04:38	
Air Blank	0.000	04:40	
Subject Sample	# 2	04:41	Subject Test Refused
Breath Volume	0.000	04:41	
Air Blank	0.000	04:41	
Control Test	0.079	04:41	
Air Blank	0.000	04:42	
Diagnostics Check	OK	04:42	

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEGGETT,JARED,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: LEGGETT,JARED
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: ONEAL

DATE/TIME OF TEST: 09/14/2018 17:39:30

Diagnostics Check	OK	17:33
Air Blank	0.000	17:33
Control Test	0.078	17:33
Air Blank	0.000	17:34
Subject Sample # 1	0.000	17:35
Breath Volume	2.886	17:35
Air Blank	0.000	17:35
Air Blank	0.000	17:37
Subject Sample # 2	0.000	17:38
Breath Volume	2.246	17:38
Air Blank	0.000	17:38
Control Test	0.079	17:38
Air Blank	0.000	17:39
Diagnostics Check	OK	17:39

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEGGETT,JARED,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GONZALEZ,KEVIN
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: JIMENEZ-PEREZ

DATE/TIME OF TEST: 09/15/2018 00:28:07

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.078	00:22
Air Blank	0.000	00:22
Subject Sample # 1	0.140	00:23
Breath Volume	1.953	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.138	00:26
Breath Volume	1.781	00:26
Air Blank	0.000	00:27
Control Test	0.077	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEGGETT,JARED,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: LEGGETT,JARED
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: PLOTKIN

DATE/TIME OF TEST: 09/16/2018 00:22:56

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.123	00:18
Breath Volume	3.617	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.124	00:21
Breath Volume	3.312	00:21
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: NICE

DATE/TIME OF TEST: 09/18/2018 03:51:31

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.079	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.142	03:46
Breath Volume	3.527	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.125	03:49
Breath Volume	1.207	03:49
Air Blank	0.000	03:50
Control Test	0.076	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: MENA LUNA

DATE/TIME OF TEST: 09/18/2018 04:14:28

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:12	
Control Test	0.078	04:12	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:13	Subject Test Refused
Breath Volume	0.000	04:13	
Air Blank	0.000	04:13	
Control Test	0.078	04:13	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: DURHAM

DATE/TIME OF TEST: 09/19/2018 00:51:33

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.078	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.184	00:47
Breath Volume	1.976	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.188	00:49
Breath Volume	2.019	00:49
Air Blank	0.000	00:50
Control Test	0.079	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,GERRY,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMITH,GERRY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 16:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: VICKERS

DATE/TIME OF TEST: 09/21/2018 19:33:39

Diagnostics Check	OK	19:30	
Air Blank	0.000	19:31	
Control Test	0.079	19:31	
Air Blank	0.000	19:32	
Subject Sample	# 1	19:32	Subject Test Refused
Breath Volume	0.000	19:32	
Air Blank	0.000	19:32	
Control Test	0.079	19:33	
Air Blank	0.000	19:33	
Diagnostics Check	OK	19:33	

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CASTILLO DUARTE

DATE/TIME OF TEST: 09/22/2018 02:34:42

Diagnostics Check		OK	02:23
Air Blank		0.000	02:24
Control Test		0.079	02:24
Air Blank		0.000	02:25
Subject Sample	# 1	0.164	02:27
Breath Volume		2.312	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:29
Subject Sample	# 2	0.160	02:33
Breath Volume		2.132	02:33
Air Blank		0.000	02:33
Control Test		0.078	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: PORTILLO COLINDRES

DATE/TIME OF TEST: 09/22/2018 03:08:34

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.078	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.097	03:03
Breath Volume	1.304	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.101	03:06
Breath Volume	2.699	03:06
Air Blank	0.000	03:07
Control Test	0.078	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GRAFMAN

DATE/TIME OF TEST: 09/22/2018 03:23:31

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.078	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.000	03:19
Breath Volume	2.070	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.000	03:22
Breath Volume	1.511	03:22
Air Blank	0.000	03:22
Control Test	0.076	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: QUIROZ AMAYA

DATE/TIME OF TEST: 09/22/2018 05:36:15

Diagnostics Check	OK	05:33	
Air Blank	0.000	05:33	
Control Test	0.077	05:33	
Air Blank	0.000	05:34	
Subject Sample	# 1	05:34	Subject Test Refused
Breath Volume	0.000	05:34	
Air Blank	0.000	05:35	
Control Test	0.078	05:35	
Air Blank	0.000	05:36	
Diagnostics Check	OK	05:36	

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SOSNA

DATE/TIME OF TEST: 09/23/2018 00:17:04

Diagnostics Check	OK	00:10
Air Blank	0.000	00:11
Control Test	0.078	00:11
Air Blank	0.000	00:12
Subject Sample # 1	0.198	00:12
Breath Volume	2.953	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:14
Subject Sample # 2	0.197	00:15
Breath Volume	3.296	00:15
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:17

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: CONTRERAS MARISCAL

DATE/TIME OF TEST: 09/23/2018 04:03:41

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.161	03:59
Breath Volume	1.613	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.167	04:02
Breath Volume	1.234	04:02
Air Blank	0.000	04:02
Control Test	0.079	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: STEELE

DATE/TIME OF TEST: 09/23/2018 22:08:39

Diagnostics Check	OK	22:03	
Air Blank	0.000	22:03	
Control Test	0.079	22:04	
Air Blank	0.000	22:04	
Subject Sample	# 1	22:07	Subject Test Refused
Breath Volume	0.000	22:07	
Air Blank	0.000	22:07	
Control Test	0.079	22:08	
Air Blank	0.000	22:08	
Diagnostics Check	OK	22:08	

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: STEELE

DATE/TIME OF TEST: 09/24/2018 00:51:25

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.078	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.225	00:46
Breath Volume	2.546	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:48
Subject Sample # 2	0.214	00:49
Breath Volume	2.410	00:49
Air Blank	0.000	00:50
Control Test	0.078	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 09/25/2018 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: WEGHORN

DATE/TIME OF TEST: 09/27/2018 04:19:54

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.078	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.154	04:14
Breath Volume	3.718	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.151	04:18
Breath Volume	3.722	04:18
Air Blank	0.000	04:18
Control Test	0.078	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 10/23/2018 09:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: BOZARTH

DATE/TIME OF TEST: 09/28/2018 00:27:45

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:21	
Control Test	0.079	00:21	
Air Blank	0.000	00:21	
Subject Sample # 1	0.194	00:22	
Breath Volume	3.433	00:22	
Air Blank	0.000	00:23	
Air Blank	0.000	00:25	
Subject Sample # 2	0.201	00:26	
Breath Volume	3.144	00:26	
Air Blank	0.000	00:26	
Control Test	0.074	00:27	Control Outside Tolerance
Air Blank	0.000	00:27	

Data Download Date/Time: 10/23/2018 09:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTWELL,J,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CANTWELL,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 12:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: HULL

DATE/TIME OF TEST: 09/28/2018 02:00:45

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.078	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.000	01:56
Breath Volume	2.898	01:56
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.000	01:59
Breath Volume	2.128	01:59
Air Blank	0.000	01:59
Control Test	0.077	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 10/23/2018 09:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIZOT,ALEXANDRA,P
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BIZOT,ALEXANDRA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MARTINEZ VARGAS

DATE/TIME OF TEST: 09/29/2018 01:07:49

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:05	
Control Test	0.077	01:05	
Air Blank	0.000	01:06	
Subject Sample	# 1	01:06	Subject Test Refused
Breath Volume	0.000	01:06	
Air Blank	0.000	01:06	
Control Test	0.077	01:07	
Air Blank	0.000	01:07	
Diagnostics Check	OK	01:07	

Data Download Date/Time: 10/23/2018 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GWODZ,OFC
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: LAY-RODRIGUEZ

DATE/TIME OF TEST: 09/29/2018 05:16:38

Diagnostics Check		OK	05:09
Air Blank		0.000	05:10
Control Test		0.077	05:10
Air Blank		0.000	05:11
Subject Sample	# 1	0.138	05:11
Breath Volume		2.652	05:11
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.127	05:14
Breath Volume		1.710	05:14
Air Blank		0.000	05:15
Control Test		0.077	05:16
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 10/23/2018 09:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SMITH, MARC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 09/05/2018 03:06:06

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.078	03:03	
Air Blank	0.000	03:03	
Subject Sample	# 1	03:04	Interferent Detect
Breath Volume	2.046	03:04	
Air Blank	0.000	03:05	
Control Test	0.078	03:05	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:06	

Data Download Date/Time: 09/06/2018 04:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SMITH,MARC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 09/05/2018 03:10:53

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.078	03:08	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:09	Interferent Detect
Breath Volume	2.007	03:09	
Air Blank	0.000	03:10	
Control Test	0.078	03:10	
Air Blank		03:10	RFI Detect
Air Blank		03:10	RFI Detect

Data Download Date/Time: 09/06/2018 04:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WARD,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: WINSTON

DATE/TIME OF TEST: 09/13/2018 01:46:38

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:31	
Control Test	0.077	01:31	
Air Blank	0.000	01:32	
Subject Sample	# 1 0.150	01:35	Volume Not Met
Breath Volume	0.261	01:35	
Air Blank	0.000	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 2 0.178	01:39	
Breath Volume	1.628	01:39	
Air Blank	0.000	01:40	
Air Blank	0.000	01:41	
Subject Sample	# 3 0.158	01:44	Volume Not Met
Breath Volume	0.207	01:44	
Air Blank	0.000	01:45	
Control Test	0.074	01:45	Control Outside Tolerance
Air Blank	0.000	01:46	

Data Download Date/Time: 09/27/2018 20:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 09/26/2018 01:03:47

Diagnostics Check	OK	00:56
Air Blank	0.000	00:57
Control Test	0.076	00:57
Air Blank	0.000	00:57
Subject Sample # 1	0.213	00:58
Breath Volume	1.378	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.211	01:01
Breath Volume	1.601	01:01
Air Blank	0.000	01:02
Control Test	0.078	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 09/27/2018 20:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: HURT,ALEXANDER
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 10/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 09/30/2018 03:47:30

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:44	
Control Test	0.079	03:45	
Air Blank	0.000	03:45	
Subject Sample	# 1	03:46	Subject Test Refused
Breath Volume	0.000	03:46	
Air Blank	0.000	03:46	
Control Test	0.079	03:46	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 10/22/2018 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,AARON,K
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: GONZALES,AARON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: BUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 09/02/2018 22:54:08

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.080	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.148	22:49
Breath Volume	3.621	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.142	22:52
Breath Volume	2.078	22:52
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 09/25/2018 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON, RONALD, E
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: BURCHETT, JASON
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 09/13/2018 21:21:38

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:18	
Control Test	0.080	21:19	
Air Blank	0.000	21:19	
Subject Sample	# 1	21:20	Subject Test Refused
Breath Volume	0.000	21:20	
Air Blank	0.000	21:20	
Control Test	0.080	21:21	
Air Blank	0.000	21:21	
Diagnostics Check	OK	21:21	

Data Download Date/Time: 09/25/2018 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTWELL,J,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CANTWELL,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: TORRES GONZALEZ

DATE/TIME OF TEST: 09/18/2018 20:26:27

Diagnostics Check	OK	20:19
Air Blank	0.000	20:20
Control Test	0.079	20:20
Air Blank	0.000	20:21
Subject Sample # 1	0.283	20:21
Breath Volume	3.167	20:21
Air Blank	0.000	20:22
Air Blank	0.000	20:24
Subject Sample # 2	0.275	20:24
Breath Volume	1.699	20:24
Air Blank	0.000	20:25
Control Test	0.076	20:25
Air Blank	0.000	20:26
Diagnostics Check	OK	20:26

Data Download Date/Time: 09/25/2018 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: CALVO-SANCHEZ

DATE/TIME OF TEST: 09/23/2018 01:44:16

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.079	01:39	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:42	Subject Test Refused
Breath Volume	0.042	01:42	
Air Blank	0.000	01:43	
Control Test	0.080	01:43	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 09/25/2018 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: FRANCISCO MARCOS

DATE/TIME OF TEST: 09/23/2018 02:25:02

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.212	02:19
Breath Volume	2.515	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.221	02:23
Breath Volume	1.628	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:25

Data Download Date/Time: 09/25/2018 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON, RONALD, E
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: PETERSON, RONALD
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: MCFADDEN

DATE/TIME OF TEST: 09/23/2018 03:12:51

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.143	03:08
Breath Volume	4.453	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.134	03:11
Breath Volume	3.667	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 09/25/2018 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/01/2018 00:41:11

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:35	
Control Test	0.080	00:35	
Air Blank	0.000	00:36	
Subject Sample # 1	0.090	00:36	
Breath Volume	2.406	00:36	
Air Blank	0.000	00:37	
Air Blank	0.000	00:39	
Subject Sample # 2	0.093	00:39	
Breath Volume	2.324	00:39	
Air Blank	0.000	00:40	
Control Test	0.082	00:40	Interferent Detect
Air Blank	0.000	00:41	

Data Download Date/Time: 09/24/2018 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: MARCOS

DATE/TIME OF TEST: 09/08/2018 19:40:09

Diagnostics Check	OK	19:33
Air Blank	0.000	19:34
Control Test	0.082	19:34
Air Blank	0.000	19:35
Subject Sample # 1	0.190	19:35
Breath Volume	2.039	19:35
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample # 2	0.191	19:38
Breath Volume	1.523	19:38
Air Blank	0.000	19:39
Control Test	0.081	19:39
Air Blank	0.000	19:40
Diagnostics Check	OK	19:40

Data Download Date/Time: 09/24/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: CORREA BELTRAN

DATE/TIME OF TEST: 09/09/2018 03:14:11

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:11	
Control Test	0.083	03:12	
Air Blank	0.000	03:12	
Subject Sample	# 1	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:13	
Control Test	0.083	03:13	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 09/24/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: DAIGLE

DATE/TIME OF TEST: 09/10/2018 03:08:09

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.116	03:03
Breath Volume	4.011	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.116	03:06
Breath Volume	2.808	03:06
Air Blank	0.000	03:07
Control Test	0.082	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 09/24/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: TRANQUILINO

DATE/TIME OF TEST: 09/22/2018 00:46:20

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:40	
Control Test	0.082	00:41	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:44	Subject Test Refused
Breath Volume	0.179	00:44	
Air Blank	0.000	00:45	
Control Test	0.080	00:45	
Air Blank	0.000	00:46	
Diagnostics Check	OK	00:46	

Data Download Date/Time: 09/24/2018 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: HUERTA-GALINDO

DATE/TIME OF TEST: 09/24/2018 02:43:11

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.083	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.224	02:38
Breath Volume	2.453	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.225	02:41
Breath Volume	1.578	02:41
Air Blank	0.000	02:42
Control Test	0.082	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 09/24/2018 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: TATE

DATE/TIME OF TEST: 09/16/2018 06:07:02

Diagnostics Check	OK	06:05	
Air Blank	0.000	06:06	
Control Test	0.082	06:06	Interferent Detect
Air Blank	0.000	06:07	

Data Download Date/Time: 10/08/2018 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: TATE

DATE/TIME OF TEST: 09/16/2018 06:15:18

Diagnostics Check	OK	06:08
Air Blank	0.000	06:09
Control Test	0.080	06:09
Air Blank	0.000	06:10
Subject Sample # 1	0.146	06:10
Breath Volume	3.546	06:10
Air Blank	0.000	06:11
Air Blank	0.000	06:13
Subject Sample # 2	0.147	06:13
Breath Volume	2.539	06:13
Air Blank	0.000	06:14
Control Test	0.082	06:14
Air Blank	0.000	06:15
Diagnostics Check	OK	06:15

Data Download Date/Time: 10/08/2018 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 09/03/2018 21:17:07

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.078	21:11
Air Blank	0.000	21:11
Subject Sample # 1	0.175	21:12
Breath Volume	1.449	21:12
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.179	21:15
Breath Volume	1.375	21:15
Air Blank	0.000	21:16
Control Test	0.077	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 09/13/2018 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: STALLINGS

DATE/TIME OF TEST: 09/04/2018 23:33:23

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.079	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.000	23:29
Breath Volume	3.511	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.000	23:32
Breath Volume	2.652	23:32
Air Blank	0.000	23:32
Control Test	0.078	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 09/13/2018 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: TENORIO

DATE/TIME OF TEST: 09/07/2018 23:48:20

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.078	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.205	23:43
Breath Volume	1.828	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.200	23:46
Breath Volume	1.984	23:46
Air Blank	0.000	23:47
Control Test	0.076	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 09/13/2018 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: SOSA TORRES

DATE/TIME OF TEST: 09/01/2018 02:16:42

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.083	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.000	02:12
Breath Volume	2.515	02:12
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.000	02:15
Breath Volume	3.140	02:15
Air Blank	0.000	02:15
Control Test	0.083	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 09/10/2018 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VALDEZ,MICHAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: BERNAL-FLEITES

DATE/TIME OF TEST: 09/01/2018 03:22:46

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:20	
Control Test	0.083	03:20	
Air Blank	0.000	03:21	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:21	
Control Test	0.082	03:22	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 09/10/2018 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NADAL,JOSE,V
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LOUBRIEL,YARIAXYLYNE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/03/2018 01:39:23

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.084	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.121	01:34
Breath Volume	1.246	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.126	01:37
Breath Volume	2.117	01:37
Air Blank	0.000	01:38
Control Test	0.083	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 09/10/2018 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNIZ,ROBERTO,C
BREATH TEST OPERATOR AGENCY: Miccosukee PD
ARRESTING OFFICER: GRECO,STEVEN
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 09/20/2018 23:39:57

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.082	23:34
Air Blank	0.000	23:34
Subject Sample # 1	0.000	23:35
Breath Volume	3.824	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.000	23:38
Breath Volume	3.625	23:38
Air Blank	0.000	23:39
Control Test	0.081	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 10/18/2018 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MISTY,L
BREATH TEST OPERATOR AGENCY: Miccosukee PD
ARRESTING OFFICER: GRECO,STEVEN
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: BENEDIT

DATE/TIME OF TEST: 09/24/2018 22:03:34

Diagnostics Check	OK	22:00	
Air Blank	0.000	22:00	
Control Test	0.082	22:01	
Air Blank	0.000	22:01	
Subject Sample	# 1	22:02	Subject Test Refused
Breath Volume	0.000	22:02	
Air Blank	0.000	22:02	
Control Test	0.082	22:02	
Air Blank	0.000	22:03	
Diagnostics Check	OK	22:03	

Data Download Date/Time: 10/18/2018 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HENNESSY,JONATHAN
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:22

SUBJECT'S LAST NAME: GABRIEL

DATE/TIME OF TEST: 09/01/2018 06:58:57

Diagnostics Check		OK	06:46	
Air Blank		0.000	06:46	
Control Test		0.079	06:47	
Air Blank		0.000	06:47	
Subject Sample	# 1	0.080	06:50	Volume Not Met
Breath Volume		0.039	06:50	
Air Blank		0.000	06:51	
Air Blank		0.000	06:53	
Subject Sample	# 2	0.087	06:53	
Breath Volume		1.582	06:53	
Air Blank		0.000	06:54	
Air Blank		0.000	06:56	
Subject Sample	# 3	0.095	06:57	
Breath Volume		1.425	06:57	
Air Blank		0.000	06:58	
Control Test		0.078	06:58	
Air Blank		0.000	06:58	
Diagnostics Check		OK	06:58	

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RODRIGUEZ,J
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: YOUSUF

DATE/TIME OF TEST: 09/02/2018 01:27:41

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.000	01:23
Breath Volume	1.496	01:23
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.000	01:26
Breath Volume	2.398	01:26
Air Blank	0.000	01:26
Control Test	0.080	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOUGHRAN,KEVIN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: DUFFIN

DATE/TIME OF TEST: 09/02/2018 05:04:58

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.080	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.219	04:59
Breath Volume	1.425	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.215	05:03
Breath Volume	1.246	05:03
Air Blank	0.000	05:04
Control Test	0.078	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 09/02/2018 19:34:00

Diagnostics Check	OK	19:27
Air Blank	0.000	19:27
Control Test	0.080	19:28
Air Blank	0.000	19:28
Subject Sample # 1	0.227	19:29
Breath Volume	2.082	19:29
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.215	19:32
Breath Volume	1.476	19:32
Air Blank	0.000	19:33
Control Test	0.080	19:33
Air Blank	0.000	19:33
Diagnostics Check	OK	19:34

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CARO

DATE/TIME OF TEST: 09/03/2018 03:19:11

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:13
Subject Sample	# 1	0.143	03:14
Breath Volume		1.757	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:16
Subject Sample	# 2	0.146	03:17
Breath Volume		2.445	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOONE,MICHAEL
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: SCHAFFER

DATE/TIME OF TEST: 09/03/2018 04:43:07

Diagnostics Check		OK	04:35
Air Blank		0.000	04:35
Control Test		0.080	04:36
Air Blank		0.000	04:36
Subject Sample	# 1	0.105	04:38
Breath Volume		1.460	04:38
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.102	04:41
Breath Volume		2.351	04:41
Air Blank		0.000	04:42
Control Test		0.078	04:42
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, YVETTE, D
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FINDLAN, WILLIAM
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: BELLONE

DATE/TIME OF TEST: 09/03/2018 18:25:14

Diagnostics Check	OK	18:17
Air Blank	0.000	18:18
Control Test	0.079	18:18
Air Blank	0.000	18:19
Subject Sample # 1	0.000	18:20
Breath Volume	2.300	18:20
Air Blank	0.000	18:20
Air Blank	0.000	18:22
Subject Sample # 2	0.000	18:23
Breath Volume	1.976	18:23
Air Blank	0.000	18:24
Control Test	0.080	18:24
Air Blank	0.000	18:25
Diagnostics Check	OK	18:25

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, YVETTE, D
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MEJIA, ILEANA
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 09/03/2018 18:57:01

Diagnostics Check		OK	18:44	
Air Blank		0.000	18:45	
Control Test		0.079	18:45	
Air Blank		0.000	18:46	
Subject Sample	# 1	0.244	18:47	No .020 agreement
Breath Volume		3.132	18:47	
Air Blank		0.000	18:48	
Air Blank		0.000	18:49	
Subject Sample	# 2	0.212	18:52	No .020 agreement
Breath Volume		2.332	18:52	
Air Blank		0.000	18:53	
Air Blank		0.000	18:54	
Subject Sample	# 3	0.186	18:55	No .020 agreement
Breath Volume		1.230	18:55	
Air Blank		0.000	18:56	
Control Test		0.076	18:56	
Air Blank		0.000	18:56	
Diagnostics Check		OK	18:57	

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, YVETTE, D
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MEJIA, ILEANA
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 09/03/2018 19:09:45

Diagnostics Check	OK	19:01
Air Blank	0.000	19:01
Control Test	0.078	19:02
Air Blank	0.000	19:02
Subject Sample # 1	0.220	19:04
Breath Volume	2.605	19:04
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample # 2	0.222	19:08
Breath Volume	2.488	19:08
Air Blank	0.000	19:08
Control Test	0.078	19:09
Air Blank	0.000	19:09
Diagnostics Check	OK	19:09

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MCGUIRE,THOMAS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: GROSSMAN

DATE/TIME OF TEST: 09/04/2018 20:54:55

Diagnostics Check	OK	20:47
Air Blank	0.000	20:47
Control Test	0.079	20:48
Air Blank	0.000	20:48
Subject Sample # 1	0.079	20:50
Breath Volume	1.292	20:50
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.070	20:53
Breath Volume	1.167	20:53
Air Blank	0.000	20:54
Control Test	0.078	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, YVETTE, D
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BERGER, MICHAEL
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 10:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 09/05/2018 12:00:52

Diagnostics Check		OK	11:54
Air Blank		0.000	11:54
Control Test		0.080	11:54
Air Blank		0.000	11:55
Subject Sample	# 1	0.108	11:56
Breath Volume		2.230	11:56
Air Blank		0.000	11:56
Air Blank		0.000	11:58
Subject Sample	# 2	0.104	11:59
Breath Volume		2.128	11:59
Air Blank		0.000	11:59
Control Test		0.080	12:00
Air Blank		0.000	12:00
Diagnostics Check		OK	12:00

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KOTT,DOUGLAS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: MURCIA

DATE/TIME OF TEST: 09/06/2018 01:44:47

Diagnostics Check		OK	01:38
Air Blank		0.000	01:38
Control Test		0.079	01:39
Air Blank		0.000	01:39
Subject Sample	# 1	0.135	01:40
Breath Volume		4.992	01:40
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.133	01:43
Breath Volume		4.093	01:43
Air Blank		0.000	01:43
Control Test		0.078	01:44
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: SPENCE

DATE/TIME OF TEST: 09/07/2018 01:16:30

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.177	01:11
Breath Volume	1.738	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:13
Subject Sample # 2	0.168	01:14
Breath Volume	1.484	01:14
Air Blank	0.000	01:15
Control Test	0.078	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: HURTADO MOLINA

DATE/TIME OF TEST: 09/08/2018 03:09:13

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.170	03:04
Breath Volume	2.113	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.164	03:07
Breath Volume	1.398	03:07
Air Blank	0.000	03:08
Control Test	0.077	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DRAYTON

DATE/TIME OF TEST: 09/09/2018 02:07:44

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.000	02:03
Breath Volume	1.140	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.000	02:06
Breath Volume	1.214	02:06
Air Blank	0.000	02:06
Control Test	0.079	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SGT BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: ROCCO

DATE/TIME OF TEST: 09/09/2018 02:56:03

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.105	02:50
Breath Volume	1.156	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.105	02:54
Breath Volume	1.457	02:54
Air Blank	0.000	02:55
Control Test	0.078	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:56

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FERNANDEZ,JAMES
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: PUCCIO

DATE/TIME OF TEST: 09/09/2018 19:07:07

Diagnostics Check		OK	18:55	
Air Blank		0.000	18:55	
Control Test		0.080	18:56	
Air Blank		0.000	18:56	
Subject Sample	# 1	0.150	18:59	Volume Not Met
Breath Volume		0.082	18:59	
Air Blank		0.000	19:00	
Air Blank		0.000	19:02	
Subject Sample	# 2	0.234	19:05	Volume Not Met
Breath Volume		0.050	19:05	
Air Blank		0.000	19:06	
Control Test		0.077	19:06	
Air Blank		0.000	19:07	
Diagnostics Check		OK	19:07	

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GILL,HANNIE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 09/10/2018 03:42:39

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:30	
Control Test	0.080	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:34	No Sample Provided
Breath Volume	0.000	03:34	
Air Blank	0.000	03:34	
Air Blank	0.000	03:36	
Subject Sample	# 2	0.168	03:37
Breath Volume	2.574	03:37	
Air Blank	0.000	03:38	
Air Blank	0.000	03:40	
Subject Sample	# 3	0.163	03:41
Breath Volume	1.425	03:41	
Air Blank	0.000	03:41	
Control Test	0.078	03:42	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BERMAN,MARC
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MERRILL

DATE/TIME OF TEST: 09/12/2018 00:04:53

Diagnostics Check	OK	23:57
Air Blank	0.000	23:58
Control Test	0.078	23:58
Air Blank	0.000	23:59
Subject Sample # 1	0.153	00:00
Breath Volume	2.109	00:00
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 2	0.148	00:03
Breath Volume	2.906	00:03
Air Blank	0.000	00:03
Control Test	0.079	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 09/19/2018 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STRZALKOWSKI,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: ENGLEHARDT

DATE/TIME OF TEST: 09/13/2018 00:16:01

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.079	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.000	00:11
Breath Volume	2.050	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.000	00:14
Breath Volume	1.230	00:14
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:16

Data Download Date/Time: 09/19/2018 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DEJESUS,CARLOS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: VASSELL

DATE/TIME OF TEST: 09/13/2018 00:54:05

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.079	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.175	00:49
Breath Volume	3.109	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:51
Subject Sample # 2	0.174	00:52
Breath Volume	3.312	00:52
Air Blank	0.000	00:53
Control Test	0.079	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 09/19/2018 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,DANNY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: TEVALAN GONZALEZ

DATE/TIME OF TEST: 09/14/2018 00:25:26

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:16
Subject Sample # 1	0.127	00:20
Breath Volume	1.253	00:20
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.130	00:23
Breath Volume	1.328	00:23
Air Blank	0.000	00:24
Control Test	0.076	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 09/19/2018 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JASON
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: MAGNUS

DATE/TIME OF TEST: 09/14/2018 02:14:45

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.079	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.198	02:10
Breath Volume	2.125	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.195	02:13
Breath Volume	1.593	02:13
Air Blank	0.000	02:13
Control Test	0.079	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 09/19/2018 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: CHAMBERLAIN

DATE/TIME OF TEST: 09/14/2018 03:48:04

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.079	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.214	03:43
Breath Volume	3.253	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:45
Subject Sample # 2	0.206	03:46
Breath Volume	3.152	03:46
Air Blank	0.000	03:47
Control Test	0.078	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:48

Data Download Date/Time: 09/19/2018 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TRABUE,THOMAS
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: KHUAL

DATE/TIME OF TEST: 09/15/2018 01:13:10

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.136	01:08
Breath Volume	2.558	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:10
Subject Sample # 2	0.142	01:11
Breath Volume	2.125	01:11
Air Blank	0.000	01:12
Control Test	0.078	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 09/19/2018 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JASON
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 09/15/2018 02:23:28

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.159	02:18
Breath Volume	1.292	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.163	02:21
Breath Volume	1.726	02:21
Air Blank	0.000	02:22
Control Test	0.077	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 09/19/2018 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/15/2018 03:38:05

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.078	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.196	03:33
Breath Volume	1.507	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.189	03:36
Breath Volume	1.601	03:36
Air Blank	0.000	03:37
Control Test	0.077	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 09/19/2018 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 09/15/2018 06:22:25

Diagnostics Check	OK	06:16
Air Blank	0.000	06:16
Control Test	0.081	06:17
Air Blank	0.000	06:17
Subject Sample # 1	0.121	06:18
Breath Volume	2.535	06:18
Air Blank	0.000	06:18
Air Blank	0.000	06:20
Subject Sample # 2	0.115	06:20
Breath Volume	1.878	06:20
Air Blank	0.000	06:21
Control Test	0.079	06:21
Air Blank	0.000	06:22
Diagnostics Check	OK	06:22

Data Download Date/Time: 09/19/2018 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FLEURANT,RODSON
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 11:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:31

SUBJECT'S LAST NAME: PIERRE CHARLES

DATE/TIME OF TEST: 09/15/2018 13:03:01

Diagnostics Check	OK	12:56
Air Blank	0.000	12:57
Control Test	0.081	12:57
Air Blank	0.000	12:57
Subject Sample # 1	0.107	12:58
Breath Volume	1.750	12:58
Air Blank	0.000	12:59
Air Blank	0.000	13:00
Subject Sample # 2	0.104	13:01
Breath Volume	2.199	13:01
Air Blank	0.000	13:02
Control Test	0.080	13:02
Air Blank	0.000	13:02
Diagnostics Check	OK	13:03

Data Download Date/Time: 09/19/2018 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SGT. BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: POBIAK

DATE/TIME OF TEST: 09/16/2018 01:21:46

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:10	
Control Test		0.079	01:10	
Air Blank		0.000	01:11	
Subject Sample	# 1	0.033	01:14	Volume Not Met
Breath Volume		0.039	01:14	
Air Blank		0.000	01:15	
Air Blank		0.000	01:17	
Subject Sample	# 2	0.033	01:20	Volume Not Met
Breath Volume		0.773	01:20	
Air Blank		0.000	01:20	
Control Test		0.080	01:21	
Air Blank		0.000	01:21	
Diagnostics Check		OK	01:21	

Data Download Date/Time: 09/19/2018 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SGT. BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: POBIAK

DATE/TIME OF TEST: 09/16/2018 01:32:10

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.041	01:26
Breath Volume	1.554	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.039	01:30
Breath Volume	1.152	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 09/19/2018 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CRUZ,DANIEL
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 08:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 09/16/2018 11:29:43

Diagnostics Check	OK	11:22
Air Blank	0.000	11:23
Control Test	0.079	11:23
Air Blank	0.000	11:24
Subject Sample # 1	0.200	11:25
Breath Volume	1.167	11:25
Air Blank	0.000	11:25
Air Blank	0.000	11:27
Subject Sample # 2	0.189	11:28
Breath Volume	1.382	11:28
Air Blank	0.000	11:28
Control Test	0.078	11:29
Air Blank	0.000	11:29
Diagnostics Check	OK	11:29

Data Download Date/Time: 09/19/2018 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, YVETTE, D
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WASHINGTON, YVETTE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 12:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:24

SUBJECT'S LAST NAME: ENGLE

DATE/TIME OF TEST: 09/17/2018 14:02:42

Diagnostics Check	OK	13:56
Air Blank	0.000	13:56
Control Test	0.079	13:56
Air Blank	0.000	13:57
Subject Sample # 1	0.000	13:58
Breath Volume	2.406	13:58
Air Blank	0.000	13:58
Air Blank	0.000	14:00
Subject Sample # 2	0.000	14:01
Breath Volume	2.839	14:01
Air Blank	0.000	14:01
Control Test	0.080	14:02
Air Blank	0.000	14:02
Diagnostics Check	OK	14:02

Data Download Date/Time: 09/19/2018 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WALBERT,A
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: JOHN

DATE/TIME OF TEST: 09/01/2018 02:30:30

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.130	02:25
Breath Volume	1.613	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.132	02:28
Breath Volume	2.144	02:28
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 09/03/2018 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AXELSON,B,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCCOY,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: KHAN

DATE/TIME OF TEST: 09/01/2018 03:31:30

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.172	03:26
Breath Volume	2.808	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.173	03:29
Breath Volume	1.789	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 09/03/2018 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARBOGAST, JOSHUA, S
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: ARBOGAST, JOSHUA
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: CANTOR

DATE/TIME OF TEST: 09/02/2018 01:00:41

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.081	00:54
Air Blank	0.000	00:55
Subject Sample # 1	0.155	00:55
Breath Volume	1.292	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.157	00:59
Breath Volume	2.679	00:59
Air Blank	0.000	00:59
Control Test	0.080	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 09/03/2018 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ZACHARY,H
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEAR,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 09/02/2018 03:35:01

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.081	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.202	03:29
Breath Volume	2.148	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.197	03:33
Breath Volume	1.796	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:35

Data Download Date/Time: 09/03/2018 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELANEY,PATRICK,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: FANO,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 09/02/2018 14:55:41

Diagnostics Check	OK	14:48
Air Blank	0.000	14:49
Control Test	0.080	14:49
Air Blank	0.000	14:50
Subject Sample # 1	0.235	14:51
Breath Volume	2.093	14:51
Air Blank	0.000	14:51
Air Blank	0.000	14:53
Subject Sample # 2	0.232	14:54
Breath Volume	2.207	14:54
Air Blank	0.000	14:54
Control Test	0.079	14:55
Air Blank	0.000	14:55
Diagnostics Check	OK	14:55

Data Download Date/Time: 09/03/2018 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGIS,D,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WALBERT,A
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 08/01/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: NUGENT

DATE/TIME OF TEST: 09/03/2018 02:32:02

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:18	
Control Test	0.082	02:18	
Air Blank	0.000	02:19	
Subject Sample	# 1 0.164	02:20	
Breath Volume	1.492	02:20	
Air Blank	0.000	02:21	
Air Blank	0.000	02:22	
Subject Sample	# 2	02:25	No Sample Provided
Breath Volume	0.000	02:25	
Air Blank	0.000	02:26	
Air Blank	0.000	02:28	
Subject Sample	# 3	02:31	No Sample Provided
Breath Volume	0.000	02:31	
Air Blank		02:31	Sequence Aborted
Air Blank	0.000	02:32	

Data Download Date/Time: 09/03/2018 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUONPANE,V,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STUDZINSKI,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/03/2018 21:53:22

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.081	21:47
Air Blank	0.000	21:47
Subject Sample # 1	0.224	21:48
Breath Volume	1.714	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.212	21:51
Breath Volume	1.316	21:51
Air Blank	0.000	21:52
Control Test	0.080	21:52
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,JAMES,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DRISCOLL,JAMES
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: LECLERC

DATE/TIME OF TEST: 09/04/2018 22:22:00

Diagnostics Check	OK	22:15
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:16
Subject Sample # 1	0.000	22:17
Breath Volume	2.199	22:17
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.000	22:20
Breath Volume	2.175	22:20
Air Blank	0.000	22:21
Control Test	0.080	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:22

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOEL,YVES,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROWLEY,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: GABRIEL

DATE/TIME OF TEST: 09/06/2018 05:18:04

Diagnostics Check		OK	05:10
Air Blank		0.000	05:10
Control Test		0.081	05:11
Air Blank		0.000	05:11
Subject Sample	# 1	0.188	05:12
Breath Volume		1.441	05:12
Air Blank		0.000	05:13
Air Blank		0.000	05:15
Subject Sample	# 2	0.191	05:16
Breath Volume		1.632	05:16
Air Blank		0.000	05:17
Control Test		0.078	05:17
Air Blank		0.000	05:17
Diagnostics Check		OK	05:18

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLMSTEAD,J,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PETERKIN,T
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: HARKER

DATE/TIME OF TEST: 09/08/2018 04:36:58

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.082	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.173	04:28
Breath Volume	1.714	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.194	04:32
Breath Volume	1.355	04:32
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 3	0.194	04:35
Breath Volume	1.132	04:35
Air Blank	0.000	04:36
Control Test	0.081	04:36
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,I,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STAFFORD,J
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: TROLLOPE

DATE/TIME OF TEST: 09/09/2018 00:46:25

Diagnostics Check		OK	00:39
Air Blank		0.000	00:40
Control Test		0.082	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.285	00:41
Breath Volume		2.667	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.281	00:44
Breath Volume		2.246	00:44
Air Blank		0.000	00:45
Control Test		0.075	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HUDSON, C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: CULMER

DATE/TIME OF TEST: 09/09/2018 03:17:07

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.082	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.198	03:12
Breath Volume	1.519	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:14
Subject Sample # 2	0.216	03:15
Breath Volume	2.214	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELIAS,D,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PEREZ,M
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: WHITELEATHER

DATE/TIME OF TEST: 09/09/2018 05:17:27

Diagnostics Check	OK	05:03	
Air Blank	0.000	05:03	
Control Test	0.081	05:04	
Air Blank	0.000	05:04	
Subject Sample	# 1 0.147	05:07	Volume Not Met
Breath Volume	0.898	05:07	
Air Blank	0.000	05:08	
Air Blank	0.000	05:10	
Subject Sample	# 2 0.152	05:11	
Breath Volume	1.589	05:11	
Air Blank	0.000	05:12	
Air Blank	0.000	05:13	
Subject Sample	# 3	05:16	Sequence Aborted
Breath Volume	0.796	05:16	
Air Blank	0.000	05:17	

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNORAT,ADELY,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: FORD,L
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: SCHLLACI

DATE/TIME OF TEST: 09/10/2018 06:16:59

Diagnostics Check	OK	06:09
Air Blank	0.000	06:09
Control Test	0.081	06:09
Air Blank	0.000	06:10
Subject Sample # 1	0.252	06:10
Breath Volume	1.437	06:10
Air Blank	0.000	06:11
Air Blank	0.000	06:13
Subject Sample # 2	0.259	06:15
Breath Volume	1.593	06:15
Air Blank	0.000	06:15
Control Test	0.080	06:16
Air Blank	0.000	06:16
Diagnostics Check	OK	06:16

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DESIMONE,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 09/10/2018 21:20:25

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.080	21:14
Air Blank	0.000	21:14
Subject Sample # 1	0.290	21:15
Breath Volume	1.699	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.290	21:18
Breath Volume	2.183	21:18
Air Blank	0.000	21:19
Control Test	0.079	21:19
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVAS,W,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GREISHEIMER,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BENOIT

DATE/TIME OF TEST: 09/11/2018 00:49:46

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.080	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.103	00:44
Breath Volume		1.578	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.103	00:48
Breath Volume		1.554	00:48
Air Blank		0.000	00:48
Control Test		0.080	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNARD,NICK,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BERNARD,NICK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: DE JESUS TORRES

DATE/TIME OF TEST: 09/11/2018 20:50:49

Diagnostics Check	OK	20:43
Air Blank	0.000	20:43
Control Test	0.080	20:44
Air Blank	0.000	20:44
Subject Sample # 1	0.127	20:45
Breath Volume	1.539	20:45
Air Blank	0.000	20:46
Air Blank	0.000	20:47
Subject Sample # 2	0.139	20:49
Breath Volume	1.144	20:49
Air Blank	0.000	20:49
Control Test	0.080	20:50
Air Blank	0.000	20:50
Diagnostics Check	OK	20:50

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOODMAN,ROBERT,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: JACKSON,D.
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: BERGEN

DATE/TIME OF TEST: 09/13/2018 02:22:27

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.081	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.212	02:17
Breath Volume		1.500	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:19
Subject Sample	# 2	0.205	02:20
Breath Volume		1.761	02:20
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONGORA, ANTONIO, A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CRISP, DAVID
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: GARCIA SANTOYO

DATE/TIME OF TEST: 09/14/2018 21:36:38

Diagnostics Check	OK	21:29
Air Blank	0.000	21:30
Control Test	0.082	21:30
Air Blank	0.000	21:31
Subject Sample # 1	0.210	21:31
Breath Volume	2.011	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.202	21:34
Breath Volume	1.675	21:34
Air Blank	0.000	21:35
Control Test	0.080	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DESIMONE,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: POULSEN

DATE/TIME OF TEST: 09/15/2018 02:39:35

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.082	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.133	02:35
Breath Volume	2.300	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.131	02:37
Breath Volume	1.593	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON',N
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: ROQUE

DATE/TIME OF TEST: 09/15/2018 04:01:32

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:50	
Control Test	0.081	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1 0.175	03:54	Volume Not Met
Breath Volume	0.617	03:54	
Air Blank	0.000	03:55	
Air Blank	0.000	03:56	
Subject Sample	# 2 0.170	03:59	Volume Not Met
Breath Volume	0.824	03:59	
Air Blank	0.000	04:00	
Control Test	0.079	04:00	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONGORA, ANTONIO, A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CARPENTER, C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: LOPEZ HERNANDEZ

DATE/TIME OF TEST: 09/15/2018 22:55:50

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.082	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.190	22:49
Breath Volume	1.738	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:51
Subject Sample # 2	0.197	22:53
Breath Volume	1.996	22:53
Air Blank	0.000	22:54
Control Test	0.077	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVAS,W,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CASTILLA,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/17/2018 01:44:44

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.081	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.141	01:40
Breath Volume	1.105	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.137	01:43
Breath Volume	1.308	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,D,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PULIZZOTTO,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 09/18/2018 00:05:13

Diagnostics Check		OK	23:58
Air Blank		0.000	23:59
Control Test		0.081	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.190	00:00
Breath Volume		2.906	00:00
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.196	00:03
Breath Volume		2.992	00:03
Air Blank		0.000	00:04
Control Test		0.080	00:04
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIBEY,AMANDA,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DICARLO,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: VALVERDE

DATE/TIME OF TEST: 09/19/2018 03:44:40

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.082	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.127	03:37
Breath Volume	1.941	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.125	03:42
Breath Volume	1.359	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVAS,W,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CRISP,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 09/20/2018 03:43:23

Diagnostics Check		OK	03:35
Air Blank		0.000	03:35
Control Test		0.081	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.164	03:38
Breath Volume		1.609	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:40
Subject Sample	# 2	0.159	03:41
Breath Volume		1.980	03:41
Air Blank		0.000	03:42
Control Test		0.079	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLINGS,EMANUEL,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HOWELL,HUNTER
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: MOORS

DATE/TIME OF TEST: 09/20/2018 20:44:31

Diagnostics Check		OK	20:28	
Air Blank		0.000	20:29	
Control Test		0.081	20:29	
Air Blank		0.000	20:30	
Subject Sample	# 1	0.103	20:33	Volume Not Met
Breath Volume		1.093	20:33	
Air Blank		0.000	20:34	
Air Blank		0.000	20:35	
Subject Sample	# 2	0.102	20:38	
Breath Volume		1.367	20:38	
Air Blank		0.000	20:39	
Air Blank		0.000	20:41	
Subject Sample	# 3	0.099	20:42	
Breath Volume		1.722	20:42	
Air Blank		0.000	20:43	
Control Test		0.079	20:43	
Air Blank		0.000	20:44	
Diagnostics Check		OK	20:44	

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN,PIERRE,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BENOIT,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 09/22/2018 16:24:06

Diagnostics Check		OK	16:17
Air Blank		0.000	16:18
Control Test		0.081	16:18
Air Blank		0.000	16:18
Subject Sample	# 1	0.140	16:19
Breath Volume		2.382	16:19
Air Blank		0.000	16:20
Air Blank		0.000	16:21
Subject Sample	# 2	0.136	16:22
Breath Volume		2.273	16:22
Air Blank		0.000	16:23
Control Test		0.080	16:23
Air Blank		0.000	16:24
Diagnostics Check		OK	16:24

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERMAN,BRENT,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: FERRO,L
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: MORALESORTIZ

DATE/TIME OF TEST: 09/22/2018 22:02:24

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.081	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.146	21:57
Breath Volume	2.117	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.165	22:00
Breath Volume	2.320	22:00
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,JAMES,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DRISCOLL,JAMES
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: BOLIVAR

DATE/TIME OF TEST: 09/23/2018 05:08:35

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.081	05:03
Air Blank	0.000	05:03
Subject Sample # 1	0.184	05:04
Breath Volume	3.683	05:04
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.186	05:06
Breath Volume	3.707	05:06
Air Blank	0.000	05:07
Control Test	0.081	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: BESING

DATE/TIME OF TEST: 09/24/2018 23:51:16

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.082	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.179	23:46
Breath Volume	1.117	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:48
Subject Sample # 2	0.162	23:49
Breath Volume	1.464	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONGORA,ANTONIO,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: FORD,LEAH
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: PEREZ GARCIA

DATE/TIME OF TEST: 09/25/2018 23:06:08

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.082	22:58
Air Blank	0.000	22:59
Subject Sample # 1	0.213	23:00
Breath Volume	1.117	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.231	23:04
Breath Volume	1.769	23:04
Air Blank	0.000	23:05
Control Test	0.079	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: ARTINO

DATE/TIME OF TEST: 09/26/2018 01:57:03

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.081	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.187	01:52
Breath Volume	2.050	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.177	01:55
Breath Volume	1.804	01:55
Air Blank	0.000	01:56
Control Test	0.081	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:57

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERTILUS,FREDELYNE,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VON SOOSTEN,N
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 09/26/2018 21:29:59

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.081	21:24
Air Blank	0.000	21:24
Subject Sample # 1	0.080	21:25
Breath Volume	3.488	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:27
Subject Sample # 2	0.075	21:28
Breath Volume	1.722	21:28
Air Blank	0.000	21:29
Control Test	0.081	21:29
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOEL,YVES,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DESIMONE,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: MCNELIS

DATE/TIME OF TEST: 09/29/2018 00:04:28

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:02	
Control Test	0.081	00:02	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:03	Subject Test Refused
Breath Volume	0.000	00:03	
Air Blank	0.000	00:03	
Control Test	0.081	00:03	
Air Blank	0.000	00:04	
Diagnostics Check	OK	00:04	

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEAR,DALE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 09/29/2018 02:13:10

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.082	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.149	02:08
Breath Volume	1.464	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:10
Subject Sample # 2	0.142	02:11
Breath Volume	1.640	02:11
Air Blank	0.000	02:12
Control Test	0.081	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,Y,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DISARRO,.
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: MENDOZACRUZ

DATE/TIME OF TEST: 09/29/2018 04:07:50

Diagnostics Check		OK	03:51	
Air Blank		0.000	03:52	
Control Test		0.081	03:52	
Air Blank		0.000	03:53	
Subject Sample	# 1	0.174	03:56	Volume Not Met
Breath Volume		0.074	03:56	
Air Blank		0.000	03:57	
Air Blank		0.000	03:58	
Subject Sample	# 2	0.208	04:01	
Breath Volume		1.289	04:01	
Air Blank		0.000	04:02	
Air Blank		0.000	04:03	
Subject Sample	# 3	0.199	04:05	
Breath Volume		1.628	04:05	
Air Blank		0.000	04:06	
Control Test		0.077	04:07	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRUBIN,L,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: GAWLINSKI

DATE/TIME OF TEST: 09/29/2018 18:11:47

Diagnostics Check	OK	18:04
Air Blank	0.000	18:05
Control Test	0.081	18:05
Air Blank	0.000	18:05
Subject Sample # 1	0.180	18:06
Breath Volume	1.554	18:06
Air Blank	0.000	18:07
Air Blank	0.000	18:09
Subject Sample # 2	0.176	18:10
Breath Volume	1.531	18:10
Air Blank	0.000	18:10
Control Test	0.079	18:11
Air Blank	0.000	18:11
Diagnostics Check	OK	18:11

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 09/29/2018 22:14:09

Diagnostics Check	OK	22:07
Air Blank	0.000	22:07
Control Test	0.081	22:08
Air Blank	0.000	22:08
Subject Sample # 1	0.000	22:09
Breath Volume	2.167	22:09
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.000	22:12
Breath Volume	1.214	22:12
Air Blank	0.000	22:13
Control Test	0.081	22:13
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

Data Download Date/Time: 10/02/2018 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SALAZAR MENDIZABAL

DATE/TIME OF TEST: 09/30/2018 03:26:05

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.081	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.197	03:21
Breath Volume	1.203	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:23
Subject Sample # 2	0.207	03:24
Breath Volume	1.605	03:24
Air Blank	0.000	03:25
Control Test	0.078	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 10/02/2018 14:11