

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: IRIZARRY, CHRISTOPHER  
ARRESTING OFFICER AGENCY: Sanford PD  
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1789170  
DRY GAS EXPIRY DATE: 01/14/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 10/26/2018 22:36:14

---

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.078	22:30
Air Blank	0.000	22:31
Subject Sample # 1	0.220	22:31
Breath Volume	2.285	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.208	22:34
Breath Volume	1.656	22:34
Air Blank	0.000	22:35
Control Test	0.077	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

---

Data Download Date/Time: 10/27/2018 04:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: DELCASTILLO, BRET  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1789170  
DRY GAS EXPIRY DATE: 01/14/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 10/27/2018 01:30:25

---

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:18	
Control Test	0.077	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1 0.263	01:19	
Breath Volume	1.128	01:19	
Air Blank	0.000	01:20	
Air Blank	0.000	01:22	
Subject Sample	# 2 0.245	01:25	Slope Not Level
Breath Volume	1.562	01:25	
Air Blank	0.000	01:26	
Air Blank	0.000	01:27	
Subject Sample	# 3 0.251	01:28	
Breath Volume	1.203	01:28	
Air Blank	0.000	01:29	
Control Test	0.078	01:29	
Air Blank	0.000	01:30	
Diagnostics Check	OK	01:30	

---

Data Download Date/Time: 10/27/2018 04:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: DELCASTILLO, BRET  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: STREETER

DATE/TIME OF TEST: 10/27/2018 21:03:49

---

Diagnostics Check	OK	20:54
Air Blank	0.000	20:54
Control Test	0.079	20:55
Air Blank	0.000	20:55
Subject Sample # 1	0.282	20:56
Breath Volume	1.140	20:56
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample # 2	0.267	21:02
Breath Volume	1.765	21:02
Air Blank	0.000	21:02
Control Test	0.079	21:03
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

---

Data Download Date/Time: 12/02/2018 21:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: BURNS, CHRIS  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 10/28/2018 00:56:56

---

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.196	00:52
Breath Volume	1.601	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:54
Subject Sample # 2	0.191	00:55
Breath Volume	2.804	00:55
Air Blank	0.000	00:56
Control Test	0.080	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

---

Data Download Date/Time: 12/02/2018 21:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MOREL, ANTHONY  
ARRESTING OFFICER AGENCY: Winter Springs PD  
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: SANCHEZ MALDONADO

DATE/TIME OF TEST: 10/29/2018 02:21:24

---

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.206	02:16
Breath Volume	2.613	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.204	02:19
Breath Volume	1.882	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

---

Data Download Date/Time: 12/02/2018 21:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,JAMES,P  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: HERNANDEZ,LUIS  
ARRESTING OFFICER AGENCY: Cocoa PD  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: GOMEZ RAMIREZ

DATE/TIME OF TEST: 10/13/2018 22:32:59

---

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.080	22:27
Air Blank	0.000	22:27
Subject Sample # 1	0.126	22:28
Breath Volume	2.285	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:30
Subject Sample # 2	0.123	22:31
Breath Volume	1.703	22:31
Air Blank	0.000	22:32
Control Test	0.079	22:32
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

---

Data Download Date/Time: 10/25/2018 21:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HOLBROOK,NICHOLAS,B

BREATH TEST OPERATOR AGENCY: Rockledge PD

ARRESTING OFFICER: TORRES,H

ARRESTING OFFICER AGENCY: Rockledge PD

LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02618080A3

DRY GAS EXPIRY DATE: 03/05/2020

VIOLATION CODE: DUI

ARREST TIME: 16:20

GENDER: Female

OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: KEMP

DATE/TIME OF TEST: 10/20/2018 17:53:36

---

Diagnostics Check	OK	17:50	
Air Blank	0.000	17:51	
Control Test	0.078	17:51	
Air Blank	0.000	17:52	
Subject Sample	# 1	17:52	Subject Test Refused
Breath Volume	0.000	17:52	
Air Blank	0.000	17:52	
Control Test	0.079	17:53	
Air Blank	0.000	17:53	
Diagnostics Check	OK	17:53	

---

Data Download Date/Time: 10/25/2018 21:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRIEBELL,DON,R  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: TRIEBELL,DON  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 09/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: PAULINO

DATE/TIME OF TEST: 10/05/2018 04:14:37

---

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.076	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.187	04:10
Breath Volume	1.972	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.188	04:13
Breath Volume	1.558	04:13
Air Blank	0.000	04:13
Control Test	0.075	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

---

Data Download Date/Time: 10/28/2018 12:20



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANDERSON,ERIC,G  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: COCCHIOLA,AUSTIN  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 09/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: WERLEMAN

DATE/TIME OF TEST: 10/08/2018 03:47:43

---

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.076	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.109	03:43
Breath Volume	2.199	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.108	03:46
Breath Volume	3.558	03:46
Air Blank	0.000	03:46
Control Test	0.075	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

---

Data Download Date/Time: 10/28/2018 12:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLAY,AMBERLY,L  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: MARTINEZ,PHILIP  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 09/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 10/10/2018 23:42:54

---

Diagnostics Check		OK	23:36
Air Blank		0.000	23:37
Control Test		0.076	23:37
Air Blank		0.000	23:37
Subject Sample	# 1	0.164	23:38
Breath Volume		2.160	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:40
Subject Sample	# 2	0.157	23:41
Breath Volume		1.398	23:41
Air Blank		0.000	23:42
Control Test		0.075	23:42
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

---

Data Download Date/Time: 10/28/2018 12:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRIEBELL,DON,R  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: TRIEBELL,DON  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 09/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: ROOKS

DATE/TIME OF TEST: 10/12/2018 21:07:10

---

Diagnostics Check	OK	21:00
Air Blank	0.000	21:01
Control Test	0.077	21:01
Air Blank	0.000	21:02
Subject Sample # 1	0.000	21:02
Breath Volume	1.972	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.000	21:05
Breath Volume	1.691	21:05
Air Blank	0.000	21:06
Control Test	0.076	21:06
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

---

Data Download Date/Time: 10/28/2018 12:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRIEBELL,DON,R  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: COCCHIOLA,AUSTIN  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 09/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GRISBAUM

DATE/TIME OF TEST: 10/14/2018 01:54:39

---

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.076	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.197	01:50
Breath Volume	2.679	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.195	01:53
Breath Volume	3.140	01:53
Air Blank	0.000	01:53
Control Test	0.076	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

---

Data Download Date/Time: 10/28/2018 12:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSMAN,JOHN,D  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: HOUSMAN,JOHN  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 09/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: ZAHNER

DATE/TIME OF TEST: 10/14/2018 19:09:35

---

Diagnostics Check	OK	19:04	
Air Blank	0.000	19:04	
Control Test	0.076	19:05	
Air Blank	0.000	19:05	
Subject Sample	# 1	19:08	Subject Test Refused
Breath Volume	0.000	19:08	
Air Blank	0.000	19:08	
Control Test	0.077	19:09	
Air Blank	0.000	19:09	
Diagnostics Check	OK	19:09	

---

Data Download Date/Time: 10/28/2018 12:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRIEBELL,DON,R  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: COCCHIOLA,AUSTIN  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 09/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: FINEMAN

DATE/TIME OF TEST: 10/15/2018 02:39:38

---

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.076	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.187	02:34
Breath Volume		1.703	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.193	02:38
Breath Volume		2.402	02:38
Air Blank		0.000	02:38
Control Test		0.075	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

---

Data Download Date/Time: 10/28/2018 12:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANDERSON,ERIC,G  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: COCCHIOLA,AUSTIN  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 10/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/31/2018 01:27:10

---

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:21	
Control Test	0.077	01:22	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:25	Subject Test Refused
Breath Volume	0.015	01:25	
Air Blank	0.000	01:26	
Control Test	0.077	01:26	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:27	

---

Data Download Date/Time: 11/29/2018 06:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASSICK,PATRICIA,A  
BREATH TEST OPERATOR AGENCY: Hardee County SO  
ARRESTING OFFICER: EVERSOLE,DONNIE  
ARRESTING OFFICER AGENCY: Hardee County SO  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GARDUNO

DATE/TIME OF TEST: 10/06/2018 00:38:09

---

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.082	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.061	00:33
Breath Volume	1.945	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.061	00:36
Breath Volume	1.906	00:36
Air Blank	0.000	00:37
Control Test	0.082	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

---

Data Download Date/Time: 10/31/2018 00:04



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTWRIGHT JR,GARY,W  
BREATH TEST OPERATOR AGENCY: Hardee County SO  
ARRESTING OFFICER: HIPPOLYTE, VENEL  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: FLEURIMOND

DATE/TIME OF TEST: 10/26/2018 22:44:07

---

Diagnostics Check	OK	22:37
Air Blank	0.000	22:37
Control Test	0.081	22:37
Air Blank	0.000	22:38
Subject Sample # 1	0.171	22:39
Breath Volume	1.953	22:39
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.187	22:42
Breath Volume	1.488	22:42
Air Blank	0.000	22:43
Control Test	0.081	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

---

Data Download Date/Time: 10/31/2018 00:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRANDA,ROBERT,W  
BREATH TEST OPERATOR AGENCY: Hardee County SO  
ARRESTING OFFICER: HIPPOLYTE, VENEL  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: SALINAS-GUTIERREZ

DATE/TIME OF TEST: 10/28/2018 07:25:05

---

Diagnostics Check		OK	07:17
Air Blank		0.000	07:17
Control Test		0.080	07:17
Air Blank		0.000	07:18
Subject Sample	# 1	0.105	07:19
Breath Volume		1.265	07:19
Air Blank		0.000	07:20
Air Blank		0.000	07:22
Subject Sample	# 2	0.097	07:23
Breath Volume		1.386	07:23
Air Blank		0.000	07:24
Control Test		0.080	07:24
Air Blank		0.000	07:24
Diagnostics Check		OK	07:25

---

Data Download Date/Time: 10/31/2018 00:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: ANGELICCHI,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: VILLALOBOS

DATE/TIME OF TEST: 10/01/2018 18:35:35

---

Diagnostics Check	OK	18:27
Air Blank	0.000	18:28
Control Test	0.079	18:28
Air Blank	0.000	18:28
Subject Sample # 1	0.381	18:30
Breath Volume	2.578	18:30
Air Blank	0.000	18:31
Air Blank	0.000	18:33
Subject Sample # 2	0.379	18:34
Breath Volume	2.414	18:34
Air Blank	0.000	18:34
Control Test	0.080	18:34
Air Blank	0.000	18:35
Diagnostics Check	OK	18:35

---

Data Download Date/Time: 10/24/2018 22:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANTHONY, MICHAEL,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CAMPION, KRISTIE  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: GIBBONS

DATE/TIME OF TEST: 10/06/2018 02:31:12

---

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:22	
Control Test		0.079	02:22	
Air Blank		0.000	02:23	
Subject Sample	# 1	0.130	02:26	Volume Not Met
Breath Volume		0.000	02:26	
Air Blank		0.000	02:27	
Air Blank		0.000	02:29	
Subject Sample	# 2		02:29	Subject Test Refused
Breath Volume		0.000	02:29	
Air Blank		0.000	02:30	
Control Test		0.079	02:30	
Air Blank		0.000	02:31	
Diagnostics Check		OK	02:31	

---

Data Download Date/Time: 10/24/2018 22:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, WILLIAM, T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SMITH, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 10/11/2018 04:39:21

---

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:28	
Control Test	0.080	04:28	
Air Blank	0.000	04:29	
Subject Sample	# 1	04:32	No Sample Provided
Breath Volume	0.000	04:32	
Air Blank	0.000	04:32	
Air Blank	0.000	04:34	
Subject Sample	# 2	04:37	No Sample Provided
Breath Volume	0.000	04:37	
Air Blank	0.000	04:38	
Control Test	0.080	04:38	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

---

Data Download Date/Time: 10/24/2018 22:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KASELAK,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: RICOTTILLI,JOSHUA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: MALGADEY

DATE/TIME OF TEST: 10/12/2018 01:48:48

---

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.096	01:44
Breath Volume	2.312	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.093	01:47
Breath Volume	2.363	01:47
Air Blank	0.000	01:47
Control Test	0.080	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

---

Data Download Date/Time: 10/24/2018 22:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KASELAK,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MULLINS,JOHN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 10/12/2018 02:58:01

---

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:55	
Control Test	0.080	02:55	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:57	
Control Test	0.081	02:57	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:58	

---

Data Download Date/Time: 10/24/2018 22:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: AMATRUDA,CHRIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: FRIAS

DATE/TIME OF TEST: 10/12/2018 03:27:12

---

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.120	03:22
Breath Volume	1.425	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.136	03:25
Breath Volume	1.347	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

---

Data Download Date/Time: 10/24/2018 22:39



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, WILLIAM, T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SMITH, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: PURTILL

DATE/TIME OF TEST: 10/12/2018 04:34:42

---

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:31	
Control Test	0.081	04:31	
Air Blank	0.000	04:32	
Subject Sample	# 1	04:33	Subject Test Refused
Breath Volume	0.000	04:33	
Air Blank	0.000	04:33	
Control Test	0.080	04:34	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:34	

---

Data Download Date/Time: 10/24/2018 22:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONGMIRE,BENJAMIN,E  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: LONGMIRE,BENJAMIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: LOVERA

DATE/TIME OF TEST: 10/15/2018 03:13:39

---

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.079	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.117	03:09
Breath Volume	3.648	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.117	03:12
Breath Volume	2.394	03:12
Air Blank	0.000	03:12
Control Test	0.080	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

---

Data Download Date/Time: 10/24/2018 22:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: PUPKE,DEPUTY L.  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: MCMAHON

DATE/TIME OF TEST: 10/16/2018 00:03:50

---

Diagnostics Check	OK	23:59	
Air Blank	0.000	00:00	
Control Test	0.080	00:00	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:02	Subject Test Refused
Breath Volume	0.000	00:02	
Air Blank	0.000	00:02	
Control Test	0.080	00:03	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

---

Data Download Date/Time: 10/24/2018 22:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/20/2018 23:24:27

---

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.079	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.138	23:19
Breath Volume	1.207	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.149	23:22
Breath Volume	1.300	23:22
Air Blank	0.000	23:23
Control Test	0.079	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

---

Data Download Date/Time: 10/24/2018 22:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: AMATRUDA,CHRIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: HEIGHTON

DATE/TIME OF TEST: 10/21/2018 02:34:28

---

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.118	02:30
Breath Volume	2.035	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.119	02:32
Breath Volume	1.902	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

---

Data Download Date/Time: 10/24/2018 22:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORGAN,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: BEHERNHOUER,IAN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 10/21/2018 02:52:19

---

Diagnostics Check	OK	02:46	
Air Blank	0.000	02:47	
Control Test	0.079	02:47	
Air Blank	0.000	02:48	
Subject Sample	# 1 0.209	02:50	
Breath Volume	2.582	02:50	
Air Blank		02:51	Purge Fail
Air Blank		02:52	Purge Fail

---

Data Download Date/Time: 10/24/2018 22:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORGAN,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: BEHERENHOUSER,IAN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 10/21/2018 02:55:33

---

Diagnostics Check	OK	02:54	
Air Blank		02:55	Ambient Fail
Air Blank	0.000	02:55	

---

Data Download Date/Time: 10/24/2018 22:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORGAN,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: BEHERENHOUSER,IAN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 10/21/2018 02:58:43

---

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.026	02:58	Control Outside Tolerance
Air Blank	0.000	02:58	

---

Data Download Date/Time: 10/24/2018 22:39



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORGAN,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: BEHERENOUSER,IAN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 10/21/2018 03:05:18

---

Diagnostics Check	OK	03:03	
Air Blank	0.000	03:04	
Control Test	0.037	03:04	Control Outside Tolerance
Air Blank	0.000	03:05	

---

Data Download Date/Time: 10/24/2018 22:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORGAN,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: BEHERNHouser,IAN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 10/21/2018 03:08:50

---

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:07	
Control Test	0.049	03:08	Control Outside Tolerance
Air Blank	0.000	03:08	

---

Data Download Date/Time: 10/24/2018 22:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EDGE,JESSICA,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: HANSEN,DUSTIN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: CHASE

DATE/TIME OF TEST: 10/05/2018 19:20:42

---

Diagnostics Check	OK	19:17	
Air Blank	0.000	19:18	
Control Test	0.079	19:18	
Air Blank	0.000	19:19	
Subject Sample	# 1	19:19	Subject Test Refused
Breath Volume	0.000	19:19	
Air Blank	0.000	19:19	
Control Test	0.080	19:20	
Air Blank	0.000	19:20	
Diagnostics Check	OK	19:20	

---

Data Download Date/Time: 10/24/2018 20:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLE,STEPHAN,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: HOLE,STEPHAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 10/07/2018 02:56:33

---

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.078	02:45	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:49	No Sample Provided
Breath Volume	0.000	02:49	
Air Blank	0.000	02:50	
Air Blank	0.000	02:52	
Subject Sample	# 2	02:55	No Sample Provided
Breath Volume	0.000	02:55	
Air Blank	0.000	02:55	
Control Test	0.080	02:55	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

---

Data Download Date/Time: 10/24/2018 20:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLE,STEPHAN,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: HOLE,STEPHAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: NORTON

DATE/TIME OF TEST: 10/07/2018 23:54:34

---

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.080	23:43	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:47	No Sample Provided
Breath Volume	0.000	23:47	
Air Blank	0.000	23:48	
Air Blank	0.000	23:50	
Subject Sample	# 2	23:53	No Sample Provided
Breath Volume	0.000	23:53	
Air Blank	0.000	23:53	
Control Test	0.080	23:53	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

---

Data Download Date/Time: 10/24/2018 20:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLE,STEPHAN,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: HOLE,STEPHAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: NORTON

DATE/TIME OF TEST: 10/08/2018 00:07:59

---

Diagnostics Check		OK	00:01
Air Blank		0.000	00:02
Control Test		0.079	00:02
Air Blank		0.000	00:03
Subject Sample	# 1	0.106	00:03
Breath Volume		2.910	00:03
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.104	00:06
Breath Volume		3.105	00:06
Air Blank		0.000	00:07
Control Test		0.079	00:07
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

---

Data Download Date/Time: 10/24/2018 20:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLE,STEPHAN,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: VAN WORP,J  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: HOPE

DATE/TIME OF TEST: 10/08/2018 01:02:59

---

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.079	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.190	00:58
Breath Volume	2.050	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.174	01:01
Breath Volume	1.472	01:01
Air Blank	0.000	01:02
Control Test	0.079	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

---

Data Download Date/Time: 10/24/2018 20:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EDGE,JESSICA,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: ROGERS,RICKEY  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: MCLIN

DATE/TIME OF TEST: 10/12/2018 05:29:11

---

Diagnostics Check	OK	05:26	
Air Blank	0.000	05:26	
Control Test	0.080	05:26	
Air Blank	0.000	05:27	
Subject Sample	# 1	05:27	Subject Test Refused
Breath Volume	0.000	05:27	
Air Blank	0.000	05:28	
Control Test	0.080	05:28	
Air Blank	0.000	05:29	
Diagnostics Check	OK	05:29	

---

Data Download Date/Time: 10/24/2018 20:49



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REEVES,MARK,O  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: REEVES,MARK  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 10/20/2018 16:00:10

---

Diagnostics Check	OK	15:57	
Air Blank	0.000	15:57	
Control Test	0.081	15:58	
Air Blank	0.000	15:58	
Subject Sample	# 1	15:58	Subject Test Refused
Breath Volume	0.000	15:58	
Air Blank	0.000	15:59	
Control Test	0.080	15:59	
Air Blank	0.000	16:00	
Diagnostics Check	OK	16:00	

---

Data Download Date/Time: 10/24/2018 20:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORGAN,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: BEHERNHOUER,IAN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 10/21/2018 03:21:23

---

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.079	03:16	
Air Blank	0.000	03:16	
Subject Sample	# 1 0.174	03:19	Volume Not Met
Breath Volume	0.410	03:19	
Air Blank		03:20	Purge Fail
Air Blank		03:21	Purge Fail

---

Data Download Date/Time: 10/24/2018 20:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORGAN,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: BEHRENSHOUSER,RICHARD  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 10/21/2018 03:26:55

---

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.042	03:26	Control Outside Tolerance
Air Blank	0.000	03:26	

---

Data Download Date/Time: 10/24/2018 20:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MULLINS,J  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: SIPKO

DATE/TIME OF TEST: 10/21/2018 04:41:45

---

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:40	
Control Test	0.049	04:41	Control Outside Tolerance
Air Blank	0.000	04:41	

---

Data Download Date/Time: 10/24/2018 20:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEAL,ADAM,M  
BREATH TEST OPERATOR AGENCY: Indialantic PD  
ARRESTING OFFICER: BEAL,ADAM  
ARRESTING OFFICER AGENCY: Indialantic PD  
LAST AGENCY INSPECTION: 09/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/14/2018 23:54:17

---

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.080	23:48
Air Blank	0.000	23:48
Subject Sample # 1	0.178	23:49
Breath Volume	2.515	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:51
Subject Sample # 2	0.166	23:52
Breath Volume	3.714	23:52
Air Blank	0.000	23:53
Control Test	0.079	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

---

Data Download Date/Time: 10/25/2018 13:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FOURNIER,TIMOTHY,R  
BREATH TEST OPERATOR AGENCY: Indialantic PD  
ARRESTING OFFICER: FOURNIER,TIMOTHY  
ARRESTING OFFICER AGENCY: Indialantic PD  
LAST AGENCY INSPECTION: 09/29/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: REDD

DATE/TIME OF TEST: 10/21/2018 19:20:51

---

Diagnostics Check		OK	19:13
Air Blank		0.000	19:13
Control Test		0.078	19:14
Air Blank		0.000	19:14
Subject Sample	# 1	0.223	19:15
Breath Volume		2.621	19:15
Air Blank		0.000	19:16
Air Blank		0.000	19:18
Subject Sample	# 2	0.221	19:19
Breath Volume		1.886	19:19
Air Blank		0.000	19:19
Control Test		0.079	19:20
Air Blank		0.000	19:20
Diagnostics Check		OK	19:20

---

Data Download Date/Time: 10/25/2018 13:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAGER,KEVIN,M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: ARMBRUSTER,H  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 09/11/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 684131  
DRY GAS EXPIRY DATE: 01/23/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: SHARRETT

DATE/TIME OF TEST: 10/01/2018 15:04:04

---

Diagnostics Check	OK	14:58	
Air Blank	0.000	14:59	
Control Test	0.083	14:59	
Air Blank	0.000	15:00	
Subject Sample	# 1	15:02	Subject Test Refused
Breath Volume	0.332	15:02	
Air Blank	0.000	15:03	
Control Test	0.080	15:03	
Air Blank	0.000	15:03	
Diagnostics Check	OK	15:04	

---

Data Download Date/Time: 10/30/2018 08:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WILSON,KALEB  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 09/11/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 684131  
DRY GAS EXPIRY DATE: 01/23/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: CORNEJO-JIMENEZ

DATE/TIME OF TEST: 10/28/2018 03:07:08

---

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.082	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.199	03:02
Breath Volume	3.269	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.202	03:05
Breath Volume	3.703	03:05
Air Blank	0.000	03:06
Control Test	0.083	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

---

Data Download Date/Time: 10/30/2018 08:54



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUTLER,ADAM,S  
BREATH TEST OPERATOR AGENCY: West Melbourne PD  
ARRESTING OFFICER: VURRO,MARC  
ARRESTING OFFICER AGENCY: West Melbourne PD  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: TYNAN

DATE/TIME OF TEST: 10/05/2018 02:51:27

---

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.183	02:46
Breath Volume	5.582	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.186	02:49
Breath Volume	4.707	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

---

Data Download Date/Time: 10/31/2018 08:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RYBA,LAURIE,A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: RUTHERFORD,CHRIS  
ARRESTING OFFICER AGENCY: West Melbourne PD  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 07:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:54

SUBJECT'S LAST NAME: TINELL

DATE/TIME OF TEST: 10/06/2018 08:24:29

---

Diagnostics Check	OK	08:18
Air Blank	0.000	08:18
Control Test	0.080	08:19
Air Blank	0.000	08:19
Subject Sample # 1	0.116	08:20
Breath Volume	3.320	08:20
Air Blank	0.000	08:20
Air Blank	0.000	08:22
Subject Sample # 2	0.118	08:22
Breath Volume	2.835	08:22
Air Blank	0.000	08:23
Control Test	0.079	08:23
Air Blank	0.000	08:24
Diagnostics Check	OK	08:24

---

Data Download Date/Time: 10/31/2018 08:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OBYRNE,JAMES,W  
BREATH TEST OPERATOR AGENCY: West Melbourne PD  
ARRESTING OFFICER: RUTHERFORD,CHRISTOPHER  
ARRESTING OFFICER AGENCY: West Melbourne PD  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: PROCTOR

DATE/TIME OF TEST: 10/16/2018 16:39:22

---

Diagnostics Check	OK	16:33
Air Blank	0.000	16:33
Control Test	0.079	16:34
Air Blank	0.000	16:34
Subject Sample # 1	0.179	16:34
Breath Volume	2.589	16:34
Air Blank	0.000	16:35
Air Blank	0.000	16:37
Subject Sample # 2	0.181	16:37
Breath Volume	2.515	16:37
Air Blank	0.000	16:38
Control Test	0.080	16:38
Air Blank	0.000	16:39
Diagnostics Check	OK	16:39

---

Data Download Date/Time: 10/31/2018 08:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HICKS,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: West Melbourne PD  
ARRESTING OFFICER: VAUGHN,JOHN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: JAMIESON

DATE/TIME OF TEST: 10/19/2018 00:11:28

---

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.081	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.229	00:07
Breath Volume	1.886	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.234	00:09
Breath Volume	2.480	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

---

Data Download Date/Time: 10/31/2018 08:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HICKS,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: West Melbourne PD  
ARRESTING OFFICER: DUMMER,LUKE  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/27/2018 02:39:52

---

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.147	02:35
Breath Volume	1.699	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:37
Subject Sample # 2	0.146	02:38
Breath Volume	2.000	02:38
Air Blank	0.000	02:38
Control Test	0.079	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

---

Data Download Date/Time: 10/31/2018 08:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROLPH,AMANDA,L

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: ROLPH,AMANDA

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 20:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 10/04/2018 21:44:08

---

Diagnostics Check	OK	21:36
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.024	21:39
Breath Volume	1.398	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.023	21:42
Breath Volume	1.304	21:42
Air Blank	0.000	21:43
Control Test	0.080	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

---

Data Download Date/Time: 10/24/2018 13:11



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMILTON,J,W  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: HAMILTON,J  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LOFFLER

DATE/TIME OF TEST: 10/08/2018 04:27:06

---

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.081	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.173	04:22
Breath Volume	2.035	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.168	04:25
Breath Volume	2.230	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

---

Data Download Date/Time: 10/24/2018 13:11



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAMILTON,J,W

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HAMILTON,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 00:40

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: MOYENO SOTO

DATE/TIME OF TEST: 10/14/2018 01:35:05

---

Diagnostics Check		OK	01:27
Air Blank		0.000	01:28
Control Test		0.081	01:28
Air Blank		0.000	01:29
Subject Sample	# 1	0.163	01:29
Breath Volume		1.707	01:29
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.179	01:33
Breath Volume		1.750	01:33
Air Blank		0.000	01:34
Control Test		0.080	01:34
Air Blank		0.000	01:34
Diagnostics Check		OK	01:35

---

Data Download Date/Time: 10/24/2018 13:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPAULDING,CODY,R  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: SPAULDING,CODY  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: ELLINGSGARD

DATE/TIME OF TEST: 10/17/2018 17:50:07

---

Diagnostics Check	OK	17:43
Air Blank	0.000	17:44
Control Test	0.080	17:44
Air Blank	0.000	17:45
Subject Sample # 1	0.295	17:45
Breath Volume	2.777	17:45
Air Blank	0.000	17:46
Air Blank	0.000	17:48
Subject Sample # 2	0.286	17:48
Breath Volume	3.082	17:48
Air Blank	0.000	17:49
Control Test	0.078	17:49
Air Blank	0.000	17:50
Diagnostics Check	OK	17:50

---

Data Download Date/Time: 10/24/2018 13:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROLPH,AMANDA,L  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: ROLPH,AMANDA  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BYARS

DATE/TIME OF TEST: 10/17/2018 23:57:53

---

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.080	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.129	23:53
Breath Volume	2.500	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.132	23:56
Breath Volume	1.671	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

---

Data Download Date/Time: 10/24/2018 13:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FREY,ROBERT,P  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: WILLIAMS,B  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CAPION

DATE/TIME OF TEST: 10/19/2018 03:36:46

---

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.232	03:32
Breath Volume	1.839	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.248	03:35
Breath Volume	1.593	03:35
Air Blank	0.000	03:35
Control Test	0.079	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

---

Data Download Date/Time: 10/24/2018 13:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROLPH,AMANDA,L  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: ROLPH,AMANDA  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 10/24/2018 04:41:18

---

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.095	04:36
Breath Volume	1.632	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.099	04:39
Breath Volume	1.324	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

---

Data Download Date/Time: 10/24/2018 13:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAND,JARRETT,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: EANES,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: PADILLA MONTOYA

DATE/TIME OF TEST: 10/01/2018 03:01:38

---

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.081	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.150	02:56
Breath Volume	3.140	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.149	02:59
Breath Volume	2.082	02:59
Air Blank	0.000	03:00
Control Test	0.081	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

---

Data Download Date/Time: 10/17/2018 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTTON, BRADEN, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BROWN, T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 10/02/2018 03:15:36

---

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.238	03:10
Breath Volume		3.675	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.240	03:13
Breath Volume		3.535	03:13
Air Blank		0.000	03:14
Control Test		0.081	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

---

Data Download Date/Time: 10/17/2018 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAND,JARRETT,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: TORRES,W  
ARRESTING OFFICER AGENCY: Neptune Beach PD  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: BOUGIS

DATE/TIME OF TEST: 10/02/2018 04:41:38

---

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:39	
Control Test	0.081	04:39	
Air Blank	0.000	04:40	
Subject Sample	# 1	04:40	Subject Test Refused
Breath Volume	0.000	04:40	
Air Blank	0.000	04:40	
Control Test	0.081	04:41	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

---

Data Download Date/Time: 10/17/2018 15:39



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BENNETT,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: BALTIMORE

DATE/TIME OF TEST: 10/03/2018 21:36:38

---

Diagnostics Check	OK	21:29
Air Blank	0.000	21:30
Control Test	0.082	21:30
Air Blank	0.000	21:31
Subject Sample # 1	0.241	21:31
Breath Volume	2.492	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.242	21:34
Breath Volume	2.687	21:34
Air Blank	0.000	21:35
Control Test	0.079	21:35
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

---

Data Download Date/Time: 10/17/2018 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: COOK,L  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 12:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:08

SUBJECT'S LAST NAME: DUFFY

DATE/TIME OF TEST: 10/05/2018 15:36:01

---

Diagnostics Check	OK	15:29
Air Blank	0.000	15:30
Control Test	0.081	15:30
Air Blank	0.000	15:31
Subject Sample # 1	0.000	15:31
Breath Volume	2.609	15:31
Air Blank	0.000	15:31
Air Blank	0.000	15:33
Subject Sample # 2	0.000	15:34
Breath Volume	1.917	15:34
Air Blank	0.000	15:35
Control Test	0.081	15:35
Air Blank	0.000	15:35
Diagnostics Check	OK	15:36

---

Data Download Date/Time: 10/17/2018 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BAKER,D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:04

SUBJECT'S LAST NAME: UDEKWU

DATE/TIME OF TEST: 10/05/2018 16:32:41

---

Diagnostics Check	OK	16:25
Air Blank	0.000	16:26
Control Test	0.081	16:26
Air Blank	0.000	16:27
Subject Sample # 1	0.267	16:27
Breath Volume	1.859	16:27
Air Blank	0.000	16:28
Air Blank	0.000	16:30
Subject Sample # 2	0.259	16:30
Breath Volume	2.324	16:30
Air Blank	0.000	16:31
Control Test	0.081	16:32
Air Blank	0.000	16:32
Diagnostics Check	OK	16:32

---

Data Download Date/Time: 10/17/2018 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: STANLEY,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BASSETT

DATE/TIME OF TEST: 10/06/2018 00:45:24

---

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.081	00:42	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:44	
Control Test	0.081	00:44	
Air Blank	0.000	00:45	
Diagnostics Check	OK	00:45	

---

Data Download Date/Time: 10/17/2018 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DEW,P  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: LOVATO

DATE/TIME OF TEST: 10/06/2018 03:42:25

---

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.081	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.123	03:37
Breath Volume	1.148	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.122	03:40
Breath Volume	1.316	03:40
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

---

Data Download Date/Time: 10/17/2018 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERTT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MOELLER,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 10/07/2018 03:52:24

---

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.170	03:47
Breath Volume	3.636	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.175	03:50
Breath Volume	3.550	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

---

Data Download Date/Time: 10/17/2018 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SORRELLS,M  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: PLATZ

DATE/TIME OF TEST: 10/07/2018 06:28:05

---

Diagnostics Check	OK	06:21
Air Blank	0.000	06:21
Control Test	0.082	06:22
Air Blank	0.000	06:22
Subject Sample # 1	0.129	06:23
Breath Volume	2.070	06:23
Air Blank	0.000	06:24
Air Blank	0.000	06:25
Subject Sample # 2	0.136	06:26
Breath Volume	1.285	06:26
Air Blank	0.000	06:27
Control Test	0.081	06:27
Air Blank	0.000	06:27
Diagnostics Check	OK	06:28

---

Data Download Date/Time: 10/17/2018 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FISETTE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 10/07/2018 21:15:37

---

Diagnostics Check	OK	21:09
Air Blank	0.000	21:09
Control Test	0.082	21:09
Air Blank	0.000	21:10
Subject Sample # 1	0.088	21:10
Breath Volume	2.589	21:10
Air Blank	0.000	21:11
Air Blank	0.000	21:13
Subject Sample # 2	0.091	21:13
Breath Volume	2.734	21:13
Air Blank	0.000	21:14
Control Test	0.081	21:14
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

---

Data Download Date/Time: 10/17/2018 15:39



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MOORE,A  
ARRESTING OFFICER AGENCY: Atlantic Beach PD  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: DE JESUS

DATE/TIME OF TEST: 10/08/2018 02:09:43

---

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.090	02:05
Breath Volume	1.960	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.085	02:08
Breath Volume	1.746	02:08
Air Blank	0.000	02:08
Control Test	0.081	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

---

Data Download Date/Time: 10/17/2018 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DIAZ,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: VLIEK

DATE/TIME OF TEST: 10/08/2018 02:51:32

---

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.081	02:46
Breath Volume	1.445	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.081	02:49
Breath Volume	1.726	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

---

Data Download Date/Time: 10/17/2018 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIQUENZ,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 10/08/2018 05:38:18

---

Diagnostics Check	OK	05:35	
Air Blank	0.000	05:35	
Control Test	0.081	05:35	
Air Blank	0.000	05:36	
Subject Sample	# 1	05:36	Subject Test Refused
Breath Volume	0.000	05:36	
Air Blank	0.000	05:37	
Control Test	0.081	05:37	
Air Blank	0.000	05:38	
Diagnostics Check	OK	05:38	

---

Data Download Date/Time: 10/17/2018 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,J,T  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: GARCIA,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: PELLETIER

DATE/TIME OF TEST: 10/08/2018 18:04:47

---

Diagnostics Check	OK	18:01	
Air Blank	0.000	18:02	
Control Test	0.081	18:02	
Air Blank	0.000	18:03	
Subject Sample	# 1	18:03	Subject Test Refused
Breath Volume	0.000	18:03	
Air Blank	0.000	18:03	
Control Test	0.081	18:04	
Air Blank	0.000	18:04	
Diagnostics Check	OK	18:04	

---

Data Download Date/Time: 10/17/2018 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MARRERO,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: CHRISMAN

DATE/TIME OF TEST: 10/09/2018 20:40:27

---

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:37	
Control Test	0.082	20:38	
Air Blank	0.000	20:38	
Subject Sample	# 1	20:39	Subject Test Refused
Breath Volume	0.000	20:39	
Air Blank	0.000	20:39	
Control Test	0.081	20:39	
Air Blank	0.000	20:40	
Diagnostics Check	OK	20:40	

---

Data Download Date/Time: 10/17/2018 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BASS,H,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BROWN,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: ODOL

DATE/TIME OF TEST: 10/12/2018 01:06:37

---

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:04	
Control Test	0.082	01:04	
Air Blank	0.000	01:04	
Subject Sample	# 1	01:05	Subject Test Refused
Breath Volume	0.000	01:05	
Air Blank	0.000	01:05	
Control Test	0.081	01:05	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

---

Data Download Date/Time: 10/17/2018 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BASS,H,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: BEACH-JOHNSON

DATE/TIME OF TEST: 10/12/2018 04:49:43

---

Diagnostics Check	OK	04:42
Air Blank	0.000	04:43
Control Test	0.081	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.144	04:44
Breath Volume	3.632	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.135	04:48
Breath Volume	1.609	04:48
Air Blank	0.000	04:48
Control Test	0.080	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

---

Data Download Date/Time: 10/17/2018 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CRIBBINS,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: ADMIN  
ARREST TIME: 20:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 10/12/2018 21:10:04

---

Diagnostics Check	OK	21:00
Air Blank	0.000	21:01
Control Test	0.081	21:01
Air Blank	0.000	21:02
Subject Sample # 1	0.175	21:05
Breath Volume	2.175	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:07
Subject Sample # 2	0.165	21:08
Breath Volume	1.136	21:08
Air Blank	0.000	21:09
Control Test	0.077	21:09
Air Blank	0.000	21:09
Diagnostics Check	OK	21:10

---

Data Download Date/Time: 10/17/2018 15:40



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: COEN,S  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:59

SUBJECT'S LAST NAME: FORBESS

DATE/TIME OF TEST: 10/13/2018 16:22:39

---

Diagnostics Check	OK	16:19	
Air Blank	0.000	16:20	
Control Test	0.081	16:20	
Air Blank	0.000	16:21	
Subject Sample	# 1	16:21	Subject Test Refused
Breath Volume	0.000	16:21	
Air Blank	0.000	16:21	
Control Test	0.082	16:22	
Air Blank	0.000	16:22	
Diagnostics Check	OK	16:22	

---

Data Download Date/Time: 10/17/2018 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FISETTE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: TAMAYO-CITALAN

DATE/TIME OF TEST: 10/13/2018 21:53:57

---

Diagnostics Check	OK	21:47
Air Blank	0.000	21:47
Control Test	0.081	21:47
Air Blank	0.000	21:48
Subject Sample # 1	0.139	21:49
Breath Volume	1.589	21:49
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.141	21:52
Breath Volume	1.703	21:52
Air Blank	0.000	21:53
Control Test	0.081	21:53
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

---

Data Download Date/Time: 10/17/2018 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/14/2018 20:48:12

---

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.081	20:42
Air Blank	0.000	20:42
Subject Sample # 1	0.203	20:43
Breath Volume	2.390	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:45
Subject Sample # 2	0.213	20:46
Breath Volume	2.230	20:46
Air Blank	0.000	20:47
Control Test	0.080	20:47
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

---

Data Download Date/Time: 10/17/2018 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MOELLER,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 10/14/2018 21:27:27

---

Diagnostics Check	OK	21:20
Air Blank	0.000	21:21
Control Test	0.081	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.034	21:22
Breath Volume	2.226	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.033	21:25
Breath Volume	1.835	21:25
Air Blank	0.000	21:26
Control Test	0.082	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

---

Data Download Date/Time: 10/17/2018 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURCELL,M,N  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 10/14/2018 22:15:34

---

Diagnostics Check	OK	22:12	
Air Blank	0.000	22:13	
Control Test	0.082	22:13	
Air Blank	0.000	22:13	
Subject Sample	# 1	22:14	Subject Test Refused
Breath Volume	0.000	22:14	
Air Blank	0.000	22:14	
Control Test	0.081	22:14	
Air Blank	0.000	22:15	
Diagnostics Check	OK	22:15	

---

Data Download Date/Time: 10/17/2018 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MOELLER,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: AL-ALI RAYMAT

DATE/TIME OF TEST: 10/15/2018 00:36:50

---

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.081	00:30
Air Blank		0.000	00:31
Subject Sample	# 1	0.151	00:31
Breath Volume		2.617	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.151	00:35
Breath Volume		2.589	00:35
Air Blank		0.000	00:35
Control Test		0.081	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

---

Data Download Date/Time: 10/17/2018 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WALL,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 10/15/2018 17:31:58

---

Diagnostics Check	OK	17:24
Air Blank	0.000	17:25
Control Test	0.081	17:25
Air Blank	0.000	17:26
Subject Sample # 1	0.283	17:27
Breath Volume	1.753	17:27
Air Blank	0.000	17:27
Air Blank	0.000	17:29
Subject Sample # 2	0.279	17:30
Breath Volume	1.820	17:30
Air Blank	0.000	17:30
Control Test	0.079	17:31
Air Blank	0.000	17:31
Diagnostics Check	OK	17:31

---

Data Download Date/Time: 10/17/2018 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CARDWELL,M  
ARRESTING OFFICER AGENCY: Atlantic Beach PD  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: GAYLE

DATE/TIME OF TEST: 10/15/2018 19:13:36

---

Diagnostics Check	OK	19:07
Air Blank	0.000	19:07
Control Test	0.080	19:07
Air Blank	0.000	19:08
Subject Sample # 1	0.233	19:08
Breath Volume	1.726	19:08
Air Blank	0.000	19:09
Air Blank	0.000	19:11
Subject Sample # 2	0.215	19:11
Breath Volume	1.410	19:11
Air Blank	0.000	19:12
Control Test	0.081	19:12
Air Blank	0.000	19:13
Diagnostics Check	OK	19:13

---

Data Download Date/Time: 10/17/2018 15:40



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SORRELLS,M  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MEEK

DATE/TIME OF TEST: 10/17/2018 03:30:11

---

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:27	
Control Test	0.081	03:27	
Air Blank	0.000	03:28	
Subject Sample	# 1	03:28	Subject Test Refused
Breath Volume	0.000	03:28	
Air Blank	0.000	03:29	
Control Test	0.081	03:29	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

---

Data Download Date/Time: 10/17/2018 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: TORRES, W  
ARRESTING OFFICER AGENCY: Neptune Beach PD  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: CHALKER

DATE/TIME OF TEST: 10/18/2018 04:35:26

---

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.081	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.143	04:30
Breath Volume	1.171	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.123	04:33
Breath Volume	1.519	04:33
Air Blank	0.000	04:34
Control Test	0.079	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

---

Data Download Date/Time: 11/06/2018 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOSTON,D,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SORRELLS,M  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: KOURIA

DATE/TIME OF TEST: 10/19/2018 04:19:37

---

Diagnostics Check	OK	04:16	
Air Blank	0.000	04:17	
Control Test	0.082	04:17	
Air Blank	0.000	04:18	
Subject Sample	# 1	04:18	Subject Test Refused
Breath Volume	0.000	04:18	
Air Blank	0.000	04:18	
Control Test	0.081	04:18	
Air Blank	0.000	04:19	
Diagnostics Check	OK	04:19	

---

Data Download Date/Time: 11/06/2018 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAND,JARRETT,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BROWN,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: AMROL

DATE/TIME OF TEST: 10/20/2018 00:35:05

---

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:32	
Control Test	0.081	00:32	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:34	
Control Test	0.081	00:34	
Air Blank	0.000	00:35	
Diagnostics Check	OK	00:35	

---

Data Download Date/Time: 11/06/2018 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAND,JARRETT,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: EANES,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: SOLIS

DATE/TIME OF TEST: 10/20/2018 03:25:46

---

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.081	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.155	03:21
Breath Volume	3.039	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.152	03:24
Breath Volume	1.738	03:24
Air Blank	0.000	03:24
Control Test	0.081	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

---

Data Download Date/Time: 11/06/2018 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAND,JARRETT,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,H  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 10/20/2018 05:27:56

---

Diagnostics Check	OK	05:25	
Air Blank	0.000	05:25	
Control Test	0.081	05:25	
Air Blank	0.000	05:26	
Subject Sample	# 1	05:26	Subject Test Refused
Breath Volume	0.000	05:26	
Air Blank	0.000	05:26	
Control Test	0.081	05:27	
Air Blank	0.000	05:27	
Diagnostics Check	OK	05:27	

---

Data Download Date/Time: 11/06/2018 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAND,JARRETT,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: EANES,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: CASTRO RAMOS

DATE/TIME OF TEST: 10/20/2018 23:34:46

---

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.160	23:30
Breath Volume	3.132	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.168	23:33
Breath Volume	3.015	23:33
Air Blank	0.000	23:33
Control Test	0.080	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

---

Data Download Date/Time: 11/06/2018 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAND,JARRETT,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MARRERO,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 10/21/2018 03:05:16

---

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:02	
Control Test	0.081	03:03	
Air Blank	0.000	03:03	
Subject Sample	# 1	03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	
Control Test	0.081	03:04	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

---

Data Download Date/Time: 11/06/2018 12:18



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAND,JARRETT,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BROWN,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 10/23/2018 00:49:48

---

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.082	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.000	00:45
Breath Volume	1.566	00:45
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.000	00:48
Breath Volume	2.640	00:48
Air Blank	0.000	00:48
Control Test	0.081	00:49
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

---

Data Download Date/Time: 11/06/2018 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLS,R,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MOELLER,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: GUYER

DATE/TIME OF TEST: 10/23/2018 19:28:51

---

Diagnostics Check	OK	19:21
Air Blank	0.000	19:21
Control Test	0.081	19:22
Air Blank	0.000	19:22
Subject Sample # 1	0.313	19:23
Breath Volume	2.304	19:23
Air Blank	0.000	19:24
Air Blank	0.000	19:25
Subject Sample # 2	0.300	19:26
Breath Volume	1.820	19:26
Air Blank	0.000	19:27
Control Test	0.078	19:28
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

---

Data Download Date/Time: 11/06/2018 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLS,R,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: STANLEY,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/23/2018 22:55:53

---

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.081	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.235	22:50
Breath Volume	2.554	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.234	22:54
Breath Volume	2.792	22:54
Air Blank	0.000	22:54
Control Test	0.079	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

---

Data Download Date/Time: 11/06/2018 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: EANES,J,L  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: MATTEO

DATE/TIME OF TEST: 10/27/2018 19:36:32

---

Diagnostics Check	OK	19:28
Air Blank	0.000	19:28
Control Test	0.080	19:28
Air Blank	0.000	19:29
Subject Sample # 1	0.188	19:31
Breath Volume	1.996	19:31
Air Blank	0.000	19:31
Air Blank	0.000	19:33
Subject Sample # 2	0.190	19:34
Breath Volume	1.636	19:34
Air Blank	0.000	19:35
Control Test	0.079	19:35
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

---

Data Download Date/Time: 11/06/2018 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FUTCH,G.B  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: RADFORD

DATE/TIME OF TEST: 10/27/2018 21:07:57

---

Diagnostics Check		OK	21:00
Air Blank		0.000	21:01
Control Test		0.080	21:01
Air Blank		0.000	21:01
Subject Sample	# 1	0.204	21:02
Breath Volume		1.726	21:02
Air Blank		0.000	21:03
Air Blank		0.000	21:05
Subject Sample	# 2	0.191	21:06
Breath Volume		1.960	21:06
Air Blank		0.000	21:06
Control Test		0.080	21:07
Air Blank		0.000	21:07
Diagnostics Check		OK	21:07

---

Data Download Date/Time: 11/06/2018 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DYER,W  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 10/28/2018 22:24:47

---

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.081	22:19
Air Blank	0.000	22:19
Subject Sample # 1	0.154	22:20
Breath Volume	3.289	22:20
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.154	22:23
Breath Volume	3.121	22:23
Air Blank	0.000	22:23
Control Test	0.081	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

---

Data Download Date/Time: 11/06/2018 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BROWN,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 10/29/2018 21:39:14

---

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.082	21:32
Air Blank	0.000	21:33
Subject Sample # 1	0.087	21:34
Breath Volume	1.750	21:34
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.086	21:37
Breath Volume	1.914	21:37
Air Blank	0.000	21:38
Control Test	0.081	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

---

Data Download Date/Time: 11/06/2018 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEVERETT,I,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: O'CALLAGHAN,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: MAS

DATE/TIME OF TEST: 10/30/2018 06:05:14

---

Diagnostics Check	OK	05:57
Air Blank	0.000	05:58
Control Test	0.082	05:58
Air Blank	0.000	05:59
Subject Sample # 1	0.000	06:00
Breath Volume	2.410	06:00
Air Blank	0.000	06:00
Air Blank	0.000	06:02
Subject Sample # 2	0.000	06:03
Breath Volume	1.410	06:03
Air Blank	0.000	06:04
Control Test	0.082	06:04
Air Blank	0.000	06:05
Diagnostics Check	OK	06:05

---

Data Download Date/Time: 11/06/2018 12:18



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,TAYLOR,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: EANES,J,L  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: BESHESSE

DATE/TIME OF TEST: 10/30/2018 21:13:05

---

Diagnostics Check	OK	21:06
Air Blank	0.000	21:07
Control Test	0.082	21:07
Air Blank	0.000	21:07
Subject Sample # 1	0.107	21:08
Breath Volume	2.617	21:08
Air Blank	0.000	21:09
Air Blank	0.000	21:10
Subject Sample # 2	0.102	21:11
Breath Volume	2.742	21:11
Air Blank	0.000	21:12
Control Test	0.082	21:12
Air Blank	0.000	21:12
Diagnostics Check	OK	21:13

---

Data Download Date/Time: 11/06/2018 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROLLESTON,JACOB,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ROLLESTON,JACOB  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A2  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: ROZEMA

DATE/TIME OF TEST: 10/02/2018 21:39:26

---

Diagnostics Check	OK	21:36	
Air Blank	0.000	21:36	
Control Test	0.078	21:37	
Air Blank	0.000	21:37	
Subject Sample	# 1	21:38	Subject Test Refused
Breath Volume	0.000	21:38	
Air Blank	0.000	21:38	
Control Test	0.078	21:38	
Air Blank	0.000	21:39	
Diagnostics Check	OK	21:39	

---

Data Download Date/Time: 10/09/2018 08:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROLLESTON,JACOB,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ROLLESTON,JACOB  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A2  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: PITZER

DATE/TIME OF TEST: 10/04/2018 20:02:12

---

Diagnostics Check	OK	19:56
Air Blank	0.000	19:56
Control Test	0.078	19:57
Air Blank	0.000	19:57
Subject Sample # 1	0.064	19:57
Breath Volume	2.640	19:57
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample # 2	0.061	20:00
Breath Volume	2.558	20:00
Air Blank	0.000	20:01
Control Test	0.080	20:01
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

---

Data Download Date/Time: 10/09/2018 08:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEAWARD,MORGANA,A  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: SEAWARD,MORGANA  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A2  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: MUNIZ SUAREZ

DATE/TIME OF TEST: 10/07/2018 04:24:07

---

Diagnostics Check	OK	04:14	
Air Blank	0.000	04:14	
Control Test	0.079	04:15	
Air Blank	0.000	04:15	
Subject Sample	# 1 0.108	04:16	
Breath Volume	1.406	04:16	
Air Blank	0.000	04:17	
Air Blank	0.000	04:19	
Subject Sample	# 2 0.119	04:22	Volume Not Met
Breath Volume	0.246	04:22	
Air Blank		04:23	Purge Fail
Air Blank		04:24	Purge Fail

---

Data Download Date/Time: 10/09/2018 08:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLEMAN,JOHN,H  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: COLEMAN,JOHN  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A2  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:19

SUBJECT'S LAST NAME: HAMBLIN

DATE/TIME OF TEST: 10/08/2018 17:40:38

---

Diagnostics Check	OK	17:39	
Air Blank	0.000	17:39	
Control Test	0.074	17:40	Control Outside Tolerance
Air Blank	0.000	17:40	

---

Data Download Date/Time: 10/09/2018 08:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLEMAN,JOHN,H  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: COLEMAN,JOHN  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A2  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: HAMBLIN

DATE/TIME OF TEST: 10/08/2018 17:46:55

---

Diagnostics Check	OK	17:44	
Air Blank	0.000	17:44	
Control Test	0.075	17:44	
Air Blank	0.000	17:45	
Subject Sample	# 1	17:45	Subject Test Refused
Breath Volume	0.000	17:45	
Air Blank	0.000	17:46	
Control Test	0.075	17:46	
Air Blank	0.000	17:46	
Diagnostics Check	OK	17:46	

---

Data Download Date/Time: 10/09/2018 08:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEAWARD,MORGANA,A  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: PRINGLE,JOSHUA  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 10/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A2  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: MACINTYRE

DATE/TIME OF TEST: 10/11/2018 23:27:02

---

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:22	
Control Test	0.078	23:22	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:25	Interferent Detect
Breath Volume	1.175	23:25	
Air Blank		23:26	Purge Fail
Air Blank		23:27	Purge Fail

---

Data Download Date/Time: 11/02/2018 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEAWARD,MORGANA,A  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: PRINGLE,JOSHUA  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 10/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A2  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MACINTYRE

DATE/TIME OF TEST: 10/11/2018 23:58:35

---

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.000	23:53
Breath Volume	2.011	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.000	23:57
Breath Volume	1.128	23:57
Air Blank	0.000	23:57
Control Test	0.081	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

---

Data Download Date/Time: 11/02/2018 12:18



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COUCH,WILLIAM,K  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: PETROVICH,DALLAS  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 10/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A2  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: HO

DATE/TIME OF TEST: 10/21/2018 22:53:43

---

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.079	22:43	
Air Blank	0.000	22:43	
Subject Sample	# 1	22:46	No Sample Provided
Breath Volume	0.000	22:46	
Air Blank	0.000	22:47	
Air Blank	0.000	22:49	
Subject Sample	# 2	22:52	No Sample Provided
Breath Volume	0.000	22:52	
Air Blank	0.000	22:52	
Control Test	0.079	22:53	
Air Blank	0.000	22:53	
Diagnostics Check	OK	22:53	

---

Data Download Date/Time: 11/02/2018 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIBLEY,JAMES,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: SKINNER,SEAN  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 10/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A2  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: SIBLEY

DATE/TIME OF TEST: 10/28/2018 05:26:36

---

Diagnostics Check	OK	05:22	
Air Blank	0.000	05:22	
Control Test	0.077	05:23	
Air Blank	0.000	05:23	
Subject Sample # 1	0.197	05:24	
Breath Volume	1.164	05:24	
Air Blank		05:25	Purge Fail
Air Blank	0.000	05:26	

---

Data Download Date/Time: 11/02/2018 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUCHENEY,WESLEY,A  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: SKINNER,SEAN  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 10/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A2  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: SIBLEY

DATE/TIME OF TEST: 10/28/2018 05:44:20

---

Diagnostics Check		OK	05:30	
Air Blank		0.000	05:31	
Control Test		0.077	05:31	
Air Blank		0.000	05:32	
Subject Sample	# 1	0.148	05:35	Slope Not Level
Breath Volume		1.246	05:35	
Air Blank		0.000	05:36	
Air Blank		0.000	05:37	
Subject Sample	# 2	0.180	05:38	
Breath Volume		1.316	05:38	
Air Blank		0.000	05:39	
Air Blank		0.000	05:41	
Subject Sample	# 3		05:42	Subject Test Refused
Breath Volume		0.000	05:42	
Air Blank		0.000	05:43	
Control Test		0.078	05:43	
Air Blank		0.000	05:44	
Diagnostics Check		OK	05:44	

---

Data Download Date/Time: 11/02/2018 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROLLESTON,JACOB,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ROLLESTON,JACOB  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 10/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A2  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: COTE

DATE/TIME OF TEST: 10/29/2018 23:03:58

---

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:52	
Control Test	0.078	22:53	
Air Blank	0.000	22:53	
Subject Sample	# 1 0.180	22:53	
Breath Volume	1.425	22:53	
Air Blank	0.000	22:54	
Air Blank	0.000	22:56	
Subject Sample	# 2 0.155	22:59	Volume Not Met
Breath Volume	1.023	22:59	
Air Blank	0.000	23:00	
Air Blank	0.000	23:02	
Subject Sample	# 3 0.170	23:02	
Breath Volume	1.492	23:02	
Air Blank	0.000	23:03	
Control Test	0.080	23:03	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

---

Data Download Date/Time: 11/02/2018 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEAWARD,MORGANA,A  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: SCORDINO,DOUGLAS  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 10/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A2  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 10/30/2018 01:48:44

---

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:46	
Control Test	0.079	01:46	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:47	Subject Test Refused
Breath Volume	0.000	01:47	
Air Blank	0.000	01:47	
Control Test	0.079	01:48	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

---

Data Download Date/Time: 11/02/2018 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOLBICKA,LAWRENCE,R  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: KOLBICKA,LAWRENCE  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 10/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A2  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 07:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: NEWBERRY

DATE/TIME OF TEST: 10/30/2018 08:17:01

---

Diagnostics Check	OK	08:10
Air Blank	0.000	08:11
Control Test	0.079	08:11
Air Blank	0.000	08:12
Subject Sample # 1	0.186	08:12
Breath Volume	1.683	08:12
Air Blank	0.000	08:13
Air Blank	0.000	08:15
Subject Sample # 2	0.189	08:15
Breath Volume	2.437	08:15
Air Blank	0.000	08:16
Control Test	0.081	08:16
Air Blank	0.000	08:16
Diagnostics Check	OK	08:17

---

Data Download Date/Time: 11/02/2018 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 10/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A2  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: PENK

DATE/TIME OF TEST: 10/31/2018 13:17:21

---

Diagnostics Check	OK	13:11	
Air Blank	0.000	13:11	
Control Test	0.079	13:12	
Air Blank	0.000	13:12	
Subject Sample # 1	0.000	13:13	
Breath Volume	2.390	13:13	
Air Blank	0.000	13:13	
Air Blank	0.000	13:15	
Subject Sample # 2	0.000	13:16	
Breath Volume	1.859	13:16	
Air Blank	0.000	13:16	
Control Test	0.082	13:16	Interferent Detect
Air Blank	0.000	13:17	

---

Data Download Date/Time: 11/02/2018 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 10/09/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32316080A2  
DRY GAS EXPIRY DATE: 01/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: PENK

DATE/TIME OF TEST: 10/31/2018 13:24:36

---

Diagnostics Check	OK	13:18
Air Blank	0.000	13:19
Control Test	0.079	13:19
Air Blank	0.000	13:19
Subject Sample # 1	0.000	13:20
Breath Volume	1.511	13:20
Air Blank	0.000	13:20
Air Blank	0.000	13:22
Subject Sample # 2	0.000	13:23
Breath Volume	1.398	13:23
Air Blank	0.000	13:23
Control Test	0.081	13:24
Air Blank	0.000	13:24
Diagnostics Check	OK	13:24

---

Data Download Date/Time: 11/02/2018 12:19



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: ESCOBAR,ALEJANDRO  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 09/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: ANTOINE

DATE/TIME OF TEST: 10/07/2018 00:25:22

---

Diagnostics Check		OK	00:17
Air Blank		0.000	00:17
Control Test		0.079	00:18
Air Blank		0.000	00:18
Subject Sample	# 1	0.240	00:20
Breath Volume		1.281	00:20
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.251	00:23
Breath Volume		1.320	00:23
Air Blank		0.000	00:24
Control Test		0.078	00:24
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

---

Data Download Date/Time: 10/17/2018 14:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: ESCOBAR,ALEJANDRO  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 09/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 10/07/2018 03:33:37

---

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.080	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.079	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

---

Data Download Date/Time: 10/17/2018 14:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLMAN,JOSEPH,D  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: GALLMAN,JOSEPH  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: HENRIGUEZCHACON

DATE/TIME OF TEST: 10/22/2018 20:37:08

---

Diagnostics Check	OK	20:33	
Air Blank	0.000	20:33	
Control Test	0.080	20:34	
Air Blank	0.000	20:34	
Subject Sample	# 1	20:35	Subject Test Refused
Breath Volume	0.000	20:35	
Air Blank	0.000	20:36	
Control Test	0.080	20:36	
Air Blank	0.000	20:37	
Diagnostics Check	OK	20:37	

---

Data Download Date/Time: 11/21/2018 15:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORALES,DAVID,J  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: ABRAM,JUSTIN  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: GARCIA LOPEZ

DATE/TIME OF TEST: 10/28/2018 01:29:15

---

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:26	
Control Test	0.079	01:26	
Air Blank	0.000	01:27	
Subject Sample	# 1	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:28	
Control Test	0.080	01:28	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

---

Data Download Date/Time: 11/21/2018 15:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ, M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 10/01/2018 02:41:04

---

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.080	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.162	02:35
Breath Volume	1.941	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.164	02:39
Breath Volume	1.414	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:41

---

Data Download Date/Time: 10/17/2018 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BYRNE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 10/03/2018 04:08:30

---

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.080	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Control Test	0.079	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

---

Data Download Date/Time: 10/17/2018 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: IRRA,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: LAZO

DATE/TIME OF TEST: 10/04/2018 01:01:09

---

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:58	
Control Test	0.080	00:59	
Air Blank	0.000	00:59	
Subject Sample	# 1	00:59	Subject Test Refused
Breath Volume	0.000	00:59	
Air Blank	0.000	01:00	
Control Test	0.079	01:00	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:01	

---

Data Download Date/Time: 10/17/2018 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SPLATT,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: AMAYA PORTILLO

DATE/TIME OF TEST: 10/04/2018 05:20:32

---

Diagnostics Check	OK	05:17	
Air Blank	0.000	05:18	
Control Test	0.079	05:18	
Air Blank	0.000	05:18	
Subject Sample	# 1	05:19	Subject Test Refused
Breath Volume	0.000	05:19	
Air Blank	0.000	05:19	
Control Test	0.079	05:19	
Air Blank	0.000	05:20	
Diagnostics Check	OK	05:20	

---

Data Download Date/Time: 10/17/2018 15:21



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DEVERGER

DATE/TIME OF TEST: 10/06/2018 02:16:55

---

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:14	
Control Test	0.080	02:14	
Air Blank	0.000	02:15	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:16	
Control Test	0.079	02:16	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

---

Data Download Date/Time: 10/17/2018 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HILLIER,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: MOLLE

DATE/TIME OF TEST: 10/06/2018 04:18:03

---

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.148	04:13
Breath Volume	2.304	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:15
Subject Sample # 2	0.154	04:16
Breath Volume	2.308	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:18

---

Data Download Date/Time: 10/17/2018 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POWELL, EDWARD, A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DOUGHFMAN, D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 10/07/2018 03:47:06

---

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.079	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.144	03:42
Breath Volume		1.335	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.141	03:45
Breath Volume		2.089	03:45
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

---

Data Download Date/Time: 10/17/2018 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POWELL, EDWARD, A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KASSIS, F  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: SWINTON

DATE/TIME OF TEST: 10/07/2018 04:43:05

---

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.048	04:36
Breath Volume	1.226	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.049	04:41
Breath Volume	1.394	04:41
Air Blank	0.000	04:42
Control Test	0.079	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

---

Data Download Date/Time: 10/17/2018 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/08/2018 00:15:28

---

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:13	
Control Test	0.080	00:13	
Air Blank	0.000	00:13	
Subject Sample	# 1	00:14	Subject Test Refused
Breath Volume	0.000	00:14	
Air Blank	0.000	00:14	
Control Test	0.080	00:14	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

---

Data Download Date/Time: 10/17/2018 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DOUGHFMAN,D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 10/08/2018 02:50:45

---

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:48	
Control Test	0.080	02:48	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:49	Subject Test Refused
Breath Volume	0.000	02:49	
Air Blank	0.000	02:49	
Control Test	0.079	02:50	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

---

Data Download Date/Time: 10/17/2018 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MOELLER,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: HINES

DATE/TIME OF TEST: 10/08/2018 03:42:10

---

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.089	03:37
Breath Volume	1.289	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.090	03:40
Breath Volume	1.253	03:40
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

---

Data Download Date/Time: 10/17/2018 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WEBER,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: JAY

DATE/TIME OF TEST: 10/09/2018 05:02:27

---

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.080	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.000	04:58
Breath Volume	1.824	04:58
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.000	05:01
Breath Volume	2.046	05:01
Air Blank	0.000	05:01
Control Test	0.079	05:01
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

---

Data Download Date/Time: 10/17/2018 15:21



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MARRERO, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 10/10/2018 02:09:59

---

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:07	
Control Test	0.080	02:07	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:08	Slope Not Met
Breath Volume	2.382	02:08	
Air Blank	0.000	02:09	
Control Test	0.080	02:09	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

---

Data Download Date/Time: 10/17/2018 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MARRERO, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 10/10/2018 02:18:57

---

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.224	02:14
Breath Volume	2.871	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.221	02:17
Breath Volume	2.066	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

---

Data Download Date/Time: 10/17/2018 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 10/13/2018 01:35:22

---

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.080	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.076	01:31
Breath Volume	1.765	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.076	01:33
Breath Volume	1.812	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

---

Data Download Date/Time: 10/17/2018 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FISETTE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: HAYNES

DATE/TIME OF TEST: 10/14/2018 02:08:34

---

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.079	02:02
Air Blank	0.000	02:03
Subject Sample # 1	0.238	02:04
Breath Volume	3.113	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.238	02:06
Breath Volume	2.910	02:06
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

---

Data Download Date/Time: 10/17/2018 15:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DOUGHMAN,D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: WATFORD

DATE/TIME OF TEST: 10/14/2018 04:58:54

---

Diagnostics Check	OK	04:52
Air Blank	0.000	04:53
Control Test	0.079	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.122	04:54
Breath Volume	2.406	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:56
Subject Sample # 2	0.124	04:57
Breath Volume	2.105	04:57
Air Blank	0.000	04:57
Control Test	0.079	04:58
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

---

Data Download Date/Time: 10/17/2018 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SWEAT,G  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:36

SUBJECT'S LAST NAME: HORNE

DATE/TIME OF TEST: 10/14/2018 08:06:51

---

Diagnostics Check	OK	08:04	
Air Blank	0.000	08:04	
Control Test	0.080	08:04	
Air Blank	0.000	08:05	
Subject Sample	# 1	08:05	Subject Test Refused
Breath Volume	0.000	08:05	
Air Blank	0.000	08:05	
Control Test	0.079	08:06	
Air Blank	0.000	08:06	
Diagnostics Check	OK	08:06	

---

Data Download Date/Time: 10/17/2018 15:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BRIDWELL,C  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CORBETT

DATE/TIME OF TEST: 10/14/2018 23:56:43

---

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.000	23:52
Breath Volume	1.769	23:52
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.000	23:55
Breath Volume	1.492	23:55
Air Blank	0.000	23:55
Control Test	0.080	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

---

Data Download Date/Time: 10/17/2018 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: HAKES

DATE/TIME OF TEST: 10/16/2018 00:00:17

---

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.080	23:58	
Air Blank	0.000	23:58	
Subject Sample	# 1	23:58	Subject Test Refused
Breath Volume	0.000	23:58	
Air Blank	0.000	23:59	
Control Test	0.079	23:59	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

---

Data Download Date/Time: 10/17/2018 15:22



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: PRITCHARD,T  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 10/04/2018 04:00:59

---

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.206	03:55
Breath Volume	1.515	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.202	03:59
Breath Volume	2.269	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

---

Data Download Date/Time: 10/26/2018 17:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNOW,C,D  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: PRITCHARD,T  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 10/09/2018 00:03:32

---

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.079	23:57	
Air Blank	0.000	23:58	
Subject Sample	# 2 0.220	23:59	
Breath Volume	1.640	23:59	
Air Blank	0.000	23:59	
Air Blank	0.000	00:01	
Subject Sample	# 1	00:02	Slope Not Met
Breath Volume	1.988	00:02	
Air Blank	0.000	00:02	
Control Test	0.079	00:02	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

---

Data Download Date/Time: 10/26/2018 17:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNOW,C,D  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: PRITCHARD,T  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 10/09/2018 00:12:43

---

Diagnostics Check		OK	00:06
Air Blank		0.000	00:06
Control Test		0.078	00:07
Air Blank		0.000	00:07
Subject Sample	# 1	0.244	00:08
Breath Volume		1.796	00:08
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.235	00:11
Breath Volume		1.851	00:11
Air Blank		0.000	00:11
Control Test		0.079	00:12
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

---

Data Download Date/Time: 10/26/2018 17:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNOW,C,D  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: THEUS,J  
ARRESTING OFFICER AGENCY: Orange Park PD  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: SCHULTZ

DATE/TIME OF TEST: 10/12/2018 23:56:23

---

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.080	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.210	23:51
Breath Volume	2.285	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.197	23:54
Breath Volume	1.371	23:54
Air Blank	0.000	23:55
Control Test	0.080	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

---

Data Download Date/Time: 10/26/2018 17:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVARADO,ANGEL,L  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HAMMELMAN,TYLER  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/02/2018 04:23:20

---

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.061	04:18
Breath Volume	2.132	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.063	04:21
Breath Volume	2.308	04:21
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

---

Data Download Date/Time: 10/10/2018 18:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNON, MATTHEW, F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HANNON, MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: DEPERNO

DATE/TIME OF TEST: 10/08/2018 21:38:12

---

Diagnostics Check	OK	21:31
Air Blank	0.000	21:32
Control Test	0.080	21:32
Air Blank	0.000	21:33
Subject Sample # 1	0.165	21:33
Breath Volume	1.691	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.167	21:36
Breath Volume	2.707	21:36
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

---

Data Download Date/Time: 10/10/2018 18:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNON, MATTHEW, F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HANNON, MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: DONAHOE

DATE/TIME OF TEST: 10/15/2018 06:36:25

---

Diagnostics Check	OK	06:32	
Air Blank	0.000	06:33	
Control Test	0.080	06:33	
Air Blank	0.000	06:34	
Subject Sample	# 1	06:35	Subject Test Refused
Breath Volume	0.000	06:35	
Air Blank	0.000	06:35	
Control Test	0.081	06:35	
Air Blank	0.000	06:36	
Diagnostics Check	OK	06:36	

---

Data Download Date/Time: 11/26/2018 19:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNON, MATTHEW, F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HANNON, MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 10/19/2018 02:45:38

---

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.190	02:40
Breath Volume	2.796	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.194	02:44
Breath Volume	1.917	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

---

Data Download Date/Time: 11/26/2018 19:09



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINTERS,LANE,A  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: HAMMELMAN,TYLER  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: KRAUSS

DATE/TIME OF TEST: 10/26/2018 04:49:37

---

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.080	04:44
Air Blank	0.000	04:44
Subject Sample # 1	0.064	04:45
Breath Volume	3.628	04:45
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.061	04:48
Breath Volume	3.843	04:48
Air Blank	0.000	04:48
Control Test	0.080	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

---

Data Download Date/Time: 11/26/2018 19:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NIXON,BARRY,  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: NIXON,BARRY  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: CISEK

DATE/TIME OF TEST: 10/28/2018 03:02:48

---

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.068	02:58
Breath Volume	2.355	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.066	03:01
Breath Volume	1.218	03:01
Air Blank	0.000	03:01
Control Test	0.081	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

---

Data Download Date/Time: 11/26/2018 19:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CONNELL,JACOB,B  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: CONNELL,JACOB  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: DALESSANDRO

DATE/TIME OF TEST: 10/31/2018 22:52:43

---

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.081	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.000	22:48
Breath Volume	1.867	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.000	22:51
Breath Volume	1.621	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

---

Data Download Date/Time: 11/26/2018 19:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAYE, MAKINA, A  
BREATH TEST OPERATOR AGENCY: Coral Springs PD  
ARRESTING OFFICER: JOHNSON, JAMES  
ARRESTING OFFICER AGENCY: Coral Springs PD  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A1  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: VERTUS

DATE/TIME OF TEST: 10/01/2018 23:19:11

---

Diagnostics Check	OK	23:14	
Air Blank	0.000	23:14	
Control Test	0.078	23:15	
Air Blank	0.000	23:15	
Subject Sample	# 1	23:17	RFI Detect
Breath Volume	0.000	23:17	
Air Blank	0.000	23:18	
Control Test	0.077	23:18	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:19	

---

Data Download Date/Time: 10/29/2018 15:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MURDOCK,TROY,J  
BREATH TEST OPERATOR AGENCY: Coral Springs PD  
ARRESTING OFFICER: MURDOCK,TROY  
ARRESTING OFFICER AGENCY: Coral Springs PD  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A1  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:48

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 10/05/2018 07:35:20

---

Diagnostics Check		OK	07:27
Air Blank		0.000	07:28
Control Test		0.079	07:28
Air Blank		0.000	07:29
Subject Sample	# 1	0.194	07:30
Breath Volume		2.906	07:30
Air Blank		0.000	07:30
Air Blank		0.000	07:32
Subject Sample	# 2	0.176	07:33
Breath Volume		2.480	07:33
Air Blank		0.000	07:34
Control Test		0.079	07:34
Air Blank		0.000	07:35
Diagnostics Check		OK	07:35

---

Data Download Date/Time: 10/29/2018 15:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAYE, MAKINA, A  
BREATH TEST OPERATOR AGENCY: Coral Springs PD  
ARRESTING OFFICER: CAJUSTE, ALISHIA  
ARRESTING OFFICER AGENCY: Coral Springs PD  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A1  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: QUINTERO

DATE/TIME OF TEST: 10/07/2018 03:01:14

---

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.078	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.214	02:56
Breath Volume	2.554	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.216	02:59
Breath Volume	1.312	02:59
Air Blank	0.000	03:00
Control Test	0.078	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

---

Data Download Date/Time: 10/29/2018 15:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KRONZ,BRIAN,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: LAMAR,ZACHARY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 10/03/2018 22:41:50

---

Diagnostics Check	OK	22:35
Air Blank	0.000	22:36
Control Test	0.079	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.136	22:37
Breath Volume	1.152	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.138	22:40
Breath Volume	1.300	22:40
Air Blank	0.000	22:40
Control Test	0.078	22:41
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

---

Data Download Date/Time: 10/15/2018 08:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDO,AARON,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ZOLLOTA,VLADIMIR  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: GILCREAST

DATE/TIME OF TEST: 10/04/2018 22:53:39

---

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.079	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.193	22:47
Breath Volume	1.734	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.177	22:52
Breath Volume	1.460	22:52
Air Blank	0.000	22:52
Control Test	0.078	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

---

Data Download Date/Time: 10/15/2018 08:44



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: WALDO,AARON  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: ROBITZSCH

DATE/TIME OF TEST: 10/05/2018 02:10:18

---

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.078	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.212	02:05
Breath Volume	1.484	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.220	02:08
Breath Volume	1.253	02:08
Air Blank	0.000	02:09
Control Test	0.075	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

---

Data Download Date/Time: 10/15/2018 08:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,BRENT,D  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: CLARK,BRENT  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: RAHDERT

DATE/TIME OF TEST: 10/05/2018 18:16:14

---

Diagnostics Check	OK	18:12	
Air Blank	0.000	18:13	
Control Test	0.078	18:13	
Air Blank	0.000	18:13	
Subject Sample # 1	0.364	18:14	
Breath Volume	3.464	18:14	
Air Blank		18:14	RFI Detect
Air Blank	0.000	18:15	
Control Test	0.079	18:15	
Air Blank	0.000	18:16	
Diagnostics Check	OK	18:16	

---

Data Download Date/Time: 10/15/2018 08:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,BRENT,D  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: CLARK,BRENT  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: RAHDERT

DATE/TIME OF TEST: 10/05/2018 18:24:42

---

Diagnostics Check	OK	18:18
Air Blank	0.000	18:18
Control Test	0.078	18:19
Air Blank	0.000	18:19
Subject Sample # 1	0.353	18:19
Breath Volume	2.042	18:19
Air Blank	0.000	18:20
Air Blank	0.000	18:22
Subject Sample # 2	0.352	18:23
Breath Volume	3.332	18:23
Air Blank	0.000	18:23
Control Test	0.078	18:24
Air Blank	0.000	18:24
Diagnostics Check	OK	18:24

---

Data Download Date/Time: 10/15/2018 08:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDO,AARON,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ZOLLOTA,VLADIMIR  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: FOUNTAIN

DATE/TIME OF TEST: 10/06/2018 01:31:08

---

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.077	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.193	01:25
Breath Volume	2.277	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.209	01:29
Breath Volume	2.199	01:29
Air Blank	0.000	01:30
Control Test	0.078	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

---

Data Download Date/Time: 10/15/2018 08:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAIA,MICHAEL,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: SAIA,MICHAEL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 05:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: TIGER

DATE/TIME OF TEST: 10/07/2018 05:25:39

---

Diagnostics Check	OK	05:21	
Air Blank	0.000	05:22	
Control Test	0.077	05:22	
Air Blank	0.000	05:22	
Subject Sample	# 1	05:24	Subject Test Refused
Breath Volume	0.000	05:24	
Air Blank	0.000	05:24	
Control Test	0.080	05:25	
Air Blank	0.000	05:25	
Diagnostics Check	OK	05:25	

---

Data Download Date/Time: 10/15/2018 08:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: CAPRA,JOHN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:38

SUBJECT'S LAST NAME: STATON

DATE/TIME OF TEST: 10/07/2018 18:04:45

---

Diagnostics Check	OK	17:58
Air Blank	0.000	17:59
Control Test	0.079	17:59
Air Blank	0.000	17:59
Subject Sample # 1	0.177	18:00
Breath Volume	2.636	18:00
Air Blank	0.000	18:00
Air Blank	0.000	18:02
Subject Sample # 2	0.191	18:03
Breath Volume	3.367	18:03
Air Blank	0.000	18:03
Control Test	0.078	18:04
Air Blank	0.000	18:04
Diagnostics Check	OK	18:04

---

Data Download Date/Time: 10/15/2018 08:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: MARTINEZ,LUIS  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: HESKETT

DATE/TIME OF TEST: 10/08/2018 00:37:25

---

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.079	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.236	00:33
Breath Volume	3.035	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.233	00:35
Breath Volume	2.281	00:35
Air Blank	0.000	00:36
Control Test	0.078	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

---

Data Download Date/Time: 10/15/2018 08:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDO,AARON,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ZOLLOTA,VLADIMIR  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: UNGER

DATE/TIME OF TEST: 10/10/2018 23:52:34

---

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:49	
Control Test	0.079	23:50	
Air Blank	0.000	23:50	
Subject Sample	# 1	23:51	Subject Test Refused
Breath Volume	0.000	23:51	
Air Blank	0.000	23:51	
Control Test	0.079	23:51	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

---

Data Download Date/Time: 10/15/2018 08:44



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: WALDO,AARON  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ELMASSARI

DATE/TIME OF TEST: 10/11/2018 03:44:17

---

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.078	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:40	No Sample Provided
Breath Volume	0.000	03:40	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample	# 2	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:43	
Control Test	0.079	03:43	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

---

Data Download Date/Time: 10/15/2018 08:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUM,ZACKARY,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ZOLLOTA,V  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: ZEITLER

DATE/TIME OF TEST: 10/12/2018 00:00:17

---

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.078	23:54
Air Blank	0.000	23:54
Subject Sample # 1	0.162	23:55
Breath Volume	3.082	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.160	23:58
Breath Volume	3.113	23:58
Air Blank	0.000	23:59
Control Test	0.079	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

---

Data Download Date/Time: 10/15/2018 08:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDO,AARON,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: KARAYIANES,MICHAEL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: MORALES PEREZ

DATE/TIME OF TEST: 10/12/2018 05:25:27

---

Diagnostics Check	OK	05:15	
Air Blank	0.000	05:16	
Control Test	0.080	05:16	
Air Blank	0.000	05:17	
Subject Sample	# 1	05:20	No Sample Provided
Breath Volume	0.000	05:20	
Air Blank	0.000	05:20	
Air Blank	0.000	05:22	
Subject Sample	# 2	05:24	Subject Test Refused
Breath Volume	0.000	05:24	
Air Blank	0.000	05:24	
Control Test	0.079	05:24	
Air Blank	0.000	05:25	
Diagnostics Check	OK	05:25	

---

Data Download Date/Time: 10/15/2018 08:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KRONZ,BRIAN,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: HINES,PAUL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: KRYSTOPA

DATE/TIME OF TEST: 10/12/2018 18:31:02

---

Diagnostics Check	OK	18:23	
Air Blank	0.000	18:24	
Control Test	0.079	18:24	
Air Blank	0.000	18:24	
Subject Sample # 1	0.265	18:25	
Breath Volume	1.398	18:25	
Air Blank	0.000	18:26	
Air Blank	0.000	18:28	
Subject Sample # 2		18:29	Subject Test Refused
Breath Volume	0.000	18:29	
Air Blank	0.000	18:30	
Control Test	0.076	18:30	
Air Blank	0.000	18:30	
Diagnostics Check	OK	18:31	

---

Data Download Date/Time: 10/15/2018 08:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: CRUM,ZACKARY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: BOOKER

DATE/TIME OF TEST: 10/12/2018 21:32:58

---

Diagnostics Check	OK	21:26
Air Blank	0.000	21:27
Control Test	0.078	21:27
Air Blank	0.000	21:27
Subject Sample # 1	0.233	21:28
Breath Volume	1.898	21:28
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.243	21:31
Breath Volume	2.050	21:31
Air Blank	0.000	21:32
Control Test	0.077	21:32
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

---

Data Download Date/Time: 10/15/2018 08:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUM,ZACKARY,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: WALDO,A  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: FUSCO

DATE/TIME OF TEST: 10/12/2018 23:03:41

---

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.080	22:58
Air Blank	0.000	22:58
Subject Sample # 1	0.000	22:59
Breath Volume	3.531	22:59
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.000	23:02
Breath Volume	2.367	23:02
Air Blank	0.000	23:02
Control Test	0.080	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

---

Data Download Date/Time: 10/15/2018 08:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ZOLLOTA,VLADIMIR  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 10/13/2018 00:17:17

---

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:14	
Control Test	0.079	00:14	
Air Blank	0.000	00:15	
Subject Sample	# 1	00:15	Subject Test Refused
Breath Volume	0.000	00:15	
Air Blank	0.000	00:16	
Control Test	0.079	00:16	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

---

Data Download Date/Time: 10/15/2018 08:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUM,ZACKARY,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: KARAYIANES,M  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 10/13/2018 01:12:07

---

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.077	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.256	01:07
Breath Volume		3.855	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.268	01:10
Breath Volume		4.406	01:10
Air Blank		0.000	01:11
Control Test		0.077	01:11
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

---

Data Download Date/Time: 10/15/2018 08:44



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: WALDO,AARON  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 10/13/2018 03:24:49

---

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.078	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.179	03:20
Breath Volume	2.074	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.188	03:23
Breath Volume	1.777	03:23
Air Blank	0.000	03:23
Control Test	0.078	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

---

Data Download Date/Time: 10/15/2018 08:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,BRENT,D  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: HANES,PAUL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: BIGGS

DATE/TIME OF TEST: 10/13/2018 21:39:13

---

Diagnostics Check	OK	21:32
Air Blank	0.000	21:33
Control Test	0.079	21:33
Air Blank	0.000	21:33
Subject Sample # 1	0.000	21:34
Breath Volume	2.937	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.000	21:37
Breath Volume	2.664	21:37
Air Blank	0.000	21:38
Control Test	0.080	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

---

Data Download Date/Time: 10/15/2018 08:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEISKOPF, MICHAEL, D  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: MCKENZIE, RONALD  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:54

SUBJECT'S LAST NAME: CRAGO JR

DATE/TIME OF TEST: 10/16/2018 13:28:01

---

Diagnostics Check	OK	13:24	
Air Blank	0.000	13:25	
Control Test	0.077	13:25	
Air Blank	0.000	13:26	
Subject Sample	# 1	13:26	Subject Test Refused
Breath Volume	0.000	13:26	
Air Blank	0.000	13:27	
Control Test	0.078	13:27	
Air Blank	0.000	13:27	
Diagnostics Check	OK	13:28	

---

Data Download Date/Time: 11/20/2018 10:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUM,ZACKARY,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: KARAYIANES,M  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: KVITA

DATE/TIME OF TEST: 10/17/2018 18:38:07

---

Diagnostics Check	OK	18:31
Air Blank	0.000	18:31
Control Test	0.078	18:32
Air Blank	0.000	18:32
Subject Sample # 1	0.227	18:33
Breath Volume	3.914	18:33
Air Blank	0.000	18:33
Air Blank	0.000	18:35
Subject Sample # 2	0.243	18:36
Breath Volume	3.058	18:36
Air Blank	0.000	18:37
Control Test	0.078	18:37
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

---

Data Download Date/Time: 11/20/2018 10:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDO,AARON,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: KARAYIANES,MICHAEL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 10/18/2018 02:39:21

---

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.119	02:34
Breath Volume	1.558	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample # 2	0.125	02:37
Breath Volume	2.191	02:37
Air Blank	0.000	02:38
Control Test	0.077	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

---

Data Download Date/Time: 11/20/2018 10:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: CRUM,ZACKARY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: KACZMARCZYK

DATE/TIME OF TEST: 10/19/2018 01:21:28

---

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.078	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.204	01:13
Breath Volume	4.058	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:15
Subject Sample # 2	0.180	01:16
Breath Volume	2.187	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 3	0.185	01:19
Breath Volume	2.140	01:19
Air Blank	0.000	01:20
Control Test	0.078	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

---

Data Download Date/Time: 11/20/2018 10:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUM,ZACKARY,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ZOLLOTA,V  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: DOEPEL

DATE/TIME OF TEST: 10/19/2018 19:07:06

---

Diagnostics Check	OK	19:03	
Air Blank	0.000	19:04	
Control Test	0.079	19:04	
Air Blank	0.000	19:05	
Subject Sample	# 1	19:05	RFI Detect
Breath Volume	0.000	19:05	
Air Blank	0.000	19:06	
Control Test	0.079	19:06	
Air Blank	0.000	19:07	
Diagnostics Check	OK	19:07	

---

Data Download Date/Time: 11/20/2018 10:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUM,ZACKARY,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ZOLLOTA,V  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: DOEPEL

DATE/TIME OF TEST: 10/19/2018 19:16:40

---

Diagnostics Check	OK	19:10
Air Blank	0.000	19:10
Control Test	0.079	19:10
Air Blank	0.000	19:11
Subject Sample # 1	0.000	19:12
Breath Volume	1.953	19:12
Air Blank	0.000	19:12
Air Blank	0.000	19:14
Subject Sample # 2	0.000	19:15
Breath Volume	2.000	19:15
Air Blank	0.000	19:15
Control Test	0.079	19:16
Air Blank	0.000	19:16
Diagnostics Check	OK	19:16

---

Data Download Date/Time: 11/20/2018 10:12



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: KARAYIANES,MICHAEL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 10/20/2018 00:24:34

---

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:21	
Control Test	0.079	00:22	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:23	
Control Test	0.079	00:23	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

---

Data Download Date/Time: 11/20/2018 10:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUM,ZACKARY,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ZOLLOTA,V  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SUPPER

DATE/TIME OF TEST: 10/20/2018 03:32:25

---

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.078	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.192	03:27
Breath Volume	3.066	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.186	03:30
Breath Volume	2.703	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

---

Data Download Date/Time: 11/20/2018 10:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,BRENT,D  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: GODSALL,DANIEL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 10/20/2018 22:37:12

---

Diagnostics Check	OK	22:30
Air Blank	0.000	22:31
Control Test	0.078	22:31
Air Blank	0.000	22:32
Subject Sample # 1	0.150	22:32
Breath Volume	3.441	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.149	22:35
Breath Volume	3.093	22:35
Air Blank	0.000	22:36
Control Test	0.078	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

---

Data Download Date/Time: 11/20/2018 10:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MIRANDA,RICHARD,  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: SNOWDEN,HENRY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MANATT

DATE/TIME OF TEST: 10/21/2018 01:47:19

---

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.160	01:42
Breath Volume	3.347	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.152	01:45
Breath Volume	1.875	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

---

Data Download Date/Time: 11/20/2018 10:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCKENZIE, RONALD, W  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: THOMAS, KRISTEN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: EISENBIEL

DATE/TIME OF TEST: 10/24/2018 16:37:02

---

Diagnostics Check	OK	16:30
Air Blank	0.000	16:31
Control Test	0.079	16:31
Air Blank	0.000	16:32
Subject Sample # 1	0.128	16:32
Breath Volume	2.824	16:32
Air Blank	0.000	16:33
Air Blank	0.000	16:35
Subject Sample # 2	0.128	16:35
Breath Volume	3.039	16:35
Air Blank	0.000	16:36
Control Test	0.079	16:36
Air Blank	0.000	16:36
Diagnostics Check	OK	16:37

---

Data Download Date/Time: 11/20/2018 10:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUM,ZACKARY,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ZOLLOTA,V  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HUFFMAN

DATE/TIME OF TEST: 10/29/2018 02:15:08

---

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:11	
Control Test	0.079	02:12	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:13	Subject Test Refused
Breath Volume	0.000	02:13	
Air Blank	0.000	02:14	
Control Test	0.079	02:14	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:15	

---

Data Download Date/Time: 11/20/2018 10:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: CRUM,Z  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: FIEDLER

DATE/TIME OF TEST: 10/29/2018 03:56:57

---

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:48	
Control Test	0.079	03:49	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:52	No Sample Provided
Breath Volume	0.000	03:52	
Air Blank	0.000	03:53	
Air Blank	0.000	03:55	
Subject Sample	# 2	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:56	
Control Test	0.079	03:56	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

---

Data Download Date/Time: 11/20/2018 10:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIGGINBOTHAM,B,K  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: BREWER,B  
ARRESTING OFFICER AGENCY: FL Dept of AG  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: RHUE

DATE/TIME OF TEST: 10/06/2018 04:28:47

---

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:26	
Control Test	0.081	04:26	
Air Blank	0.000	04:26	
Subject Sample	# 1	04:27	Subject Test Refused
Breath Volume	0.000	04:27	
Air Blank	0.000	04:27	
Control Test	0.080	04:28	
Air Blank	0.000	04:28	
Diagnostics Check	OK	04:28	

---

Data Download Date/Time: 10/28/2018 10:03



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMPBELL,D,R  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: GORDON,G  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: BORRAS

DATE/TIME OF TEST: 10/07/2018 02:51:23

---

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.154	02:46
Breath Volume	1.484	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:48
Subject Sample # 2	0.157	02:49
Breath Volume	1.828	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

---

Data Download Date/Time: 10/28/2018 10:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MATHIS,A,J  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: REYNOLDS,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 10/08/2018 03:21:54

---

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:21	
Control Test	0.081	03:21	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

---

Data Download Date/Time: 10/28/2018 10:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCRACKEN,J,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: FERRY,T  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: MOISAN

DATE/TIME OF TEST: 10/09/2018 18:54:49

---

Diagnostics Check	OK	18:48
Air Blank	0.000	18:48
Control Test	0.080	18:48
Air Blank	0.000	18:49
Subject Sample # 1	0.135	18:49
Breath Volume	1.316	18:49
Air Blank	0.000	18:50
Air Blank	0.000	18:52
Subject Sample # 2	0.140	18:53
Breath Volume	1.640	18:53
Air Blank	0.000	18:53
Control Test	0.081	18:54
Air Blank	0.000	18:54
Diagnostics Check	OK	18:54

---

Data Download Date/Time: 10/28/2018 10:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCracken,J,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: WALTON,C  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: OSTEN

DATE/TIME OF TEST: 10/10/2018 00:26:58

---

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:24	
Control Test	0.081	00:24	
Air Blank	0.000	00:25	
Subject Sample	# 1	00:25	Subject Test Refused
Breath Volume	0.000	00:25	
Air Blank	0.000	00:26	
Control Test	0.081	00:26	
Air Blank	0.000	00:26	
Diagnostics Check	OK	00:26	

---

Data Download Date/Time: 10/28/2018 10:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOOLARD,RAYMOND,C  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: WOOLARD,C  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 09:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:39

SUBJECT'S LAST NAME: MAGYARI

DATE/TIME OF TEST: 10/10/2018 10:14:40

---

Diagnostics Check	OK	09:59	
Air Blank	0.000	09:59	
Control Test	0.079	09:59	
Air Blank	0.000	10:00	
Subject Sample # 1	0.151	10:02	
Breath Volume	2.589	10:02	
Air Blank	0.000	10:02	
Air Blank	0.000	10:04	
Subject Sample # 2	0.121	10:07	Volume Not Met
Breath Volume	0.179	10:07	
Air Blank	0.000	10:08	
Air Blank	0.000	10:10	
Subject Sample # 3		10:13	Subject Test Refused
Breath Volume	0.078	10:13	
Air Blank	0.000	10:13	
Control Test	0.081	10:14	
Air Blank	0.000	10:14	
Diagnostics Check	OK	10:14	

---

Data Download Date/Time: 10/28/2018 10:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCRACKEN,J,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUNTER,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 10/13/2018 01:09:02

---

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.081	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.121	01:04
Breath Volume	1.257	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.117	01:07
Breath Volume	1.304	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:09

---

Data Download Date/Time: 10/28/2018 10:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRES,J,L  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUNTER,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 10/13/2018 03:21:56

---

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.081	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.068	03:17
Breath Volume	1.941	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.070	03:20
Breath Volume	1.582	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

---

Data Download Date/Time: 10/28/2018 10:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRES,J,L  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: KNIGHT,B  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: HAILEY

DATE/TIME OF TEST: 10/13/2018 05:20:44

---

Diagnostics Check	OK	05:14
Air Blank	0.000	05:15
Control Test	0.080	05:15
Air Blank	0.000	05:15
Subject Sample # 1	0.016	05:16
Breath Volume	1.183	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:18
Subject Sample # 2	0.015	05:19
Breath Volume	1.796	05:19
Air Blank	0.000	05:19
Control Test	0.080	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

---

Data Download Date/Time: 10/28/2018 10:04



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRES,J,L  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUNTER,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: MULDOON

DATE/TIME OF TEST: 10/13/2018 20:50:25

---

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:48	
Control Test	0.080	20:48	
Air Blank	0.000	20:48	
Subject Sample	# 1	20:49	Subject Test Refused
Breath Volume	0.000	20:49	
Air Blank	0.000	20:49	
Control Test	0.080	20:49	
Air Blank	0.000	20:50	
Diagnostics Check	OK	20:50	

---

Data Download Date/Time: 10/28/2018 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRES,J,L  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: THOMASON,P  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: STELMACH

DATE/TIME OF TEST: 10/13/2018 21:22:52

---

Diagnostics Check	OK	21:16
Air Blank	0.000	21:17
Control Test	0.079	21:17
Air Blank	0.000	21:17
Subject Sample # 1	0.278	21:18
Breath Volume	1.546	21:18
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.266	21:21
Breath Volume	1.492	21:21
Air Blank	0.000	21:21
Control Test	0.079	21:22
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

---

Data Download Date/Time: 10/28/2018 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCracken,J,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: 2C16,HUNTER  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 10/14/2018 01:02:55

---

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:00	
Control Test	0.080	01:00	
Air Blank	0.000	01:01	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:02	
Control Test	0.080	01:02	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

---

Data Download Date/Time: 10/28/2018 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCRACKEN,J,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUNTER,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DANA

DATE/TIME OF TEST: 10/15/2018 02:07:15

---

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.080	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.256	02:02
Breath Volume		2.601	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.246	02:05
Breath Volume		2.894	02:05
Air Blank		0.000	02:06
Control Test		0.078	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

---

Data Download Date/Time: 10/28/2018 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEALY,GREGORY,B  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: HEALY,GREGORY  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: BLALOCK

DATE/TIME OF TEST: 10/16/2018 01:55:50

---

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.080	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.276	01:51
Breath Volume		1.730	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.284	01:54
Breath Volume		1.906	01:54
Air Blank		0.000	01:54
Control Test		0.079	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

---

Data Download Date/Time: 10/28/2018 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCRACKEN,J,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUNTER,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: COURCHENE

DATE/TIME OF TEST: 10/17/2018 23:56:33

---

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.080	23:51
Air Blank		0.000	23:51
Subject Sample	# 1	0.000	23:52
Breath Volume		1.906	23:52
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.000	23:55
Breath Volume		1.984	23:55
Air Blank		0.000	23:55
Control Test		0.080	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

---

Data Download Date/Time: 10/28/2018 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRES,J,L  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: TERRY,A  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: MORRISSEY

DATE/TIME OF TEST: 10/18/2018 00:44:20

---

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:42	
Control Test	0.080	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:42	Subject Test Refused
Breath Volume	0.000	00:42	
Air Blank	0.000	00:43	
Control Test	0.081	00:43	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

---

Data Download Date/Time: 10/28/2018 10:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCRACKEN,J,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUNTER,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: DIAMAURO

DATE/TIME OF TEST: 10/18/2018 19:21:45

---

Diagnostics Check	OK	19:16	
Air Blank	0.000	19:17	
Control Test	0.078	19:17	
Air Blank	0.000	19:18	
Subject Sample	# 1	19:20	Subject Test Refused
Breath Volume	0.109	19:20	
Air Blank	0.000	19:20	
Control Test	0.079	19:21	
Air Blank	0.000	19:21	
Diagnostics Check	OK	19:21	

---

Data Download Date/Time: 10/28/2018 10:04



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRES,J,L  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: JACKSON,R  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 10/19/2018 00:51:23

---

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.000	00:47
Breath Volume	1.425	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.000	00:50
Breath Volume	1.839	00:50
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

---

Data Download Date/Time: 10/28/2018 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCRACKEN,J,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUNTER,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: CURLEY

DATE/TIME OF TEST: 10/19/2018 03:49:32

---

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.081	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.000	03:45
Breath Volume		1.765	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.000	03:48
Breath Volume		1.910	03:48
Air Blank		0.000	03:48
Control Test		0.080	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

---

Data Download Date/Time: 10/28/2018 10:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOOLARD,COLE,C  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: YEATMAN,B  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 07:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 10/20/2018 08:15:18

---

Diagnostics Check		OK	08:01	
Air Blank		0.000	08:02	
Control Test		0.079	08:02	
Air Blank		0.000	08:02	
Subject Sample	# 1	0.215	08:06	Volume Not Met
Breath Volume		0.917	08:06	
Air Blank		0.000	08:06	
Air Blank		0.000	08:08	
Subject Sample	# 2	0.227	08:10	
Breath Volume		1.339	08:10	
Air Blank		0.000	08:11	
Air Blank		0.000	08:13	
Subject Sample	# 3	0.217	08:13	
Breath Volume		1.304	08:13	
Air Blank		0.000	08:14	
Control Test		0.077	08:14	
Air Blank		0.000	08:15	
Diagnostics Check		OK	08:15	

---

Data Download Date/Time: 10/28/2018 10:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRES,J,L  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUNTER,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: CLIFTON

DATE/TIME OF TEST: 10/26/2018 23:47:22

---

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:46	
Control Test	0.080	23:46	RFI Detect
Air Blank	0.000	23:47	

---

Data Download Date/Time: 10/28/2018 10:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRES,J,L  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUNTER,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: CLIFTON

DATE/TIME OF TEST: 10/26/2018 23:53:54

---

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:51	
Control Test	0.080	23:51	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:53	
Control Test	0.081	23:53	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

---

Data Download Date/Time: 10/28/2018 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRES,J,L  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUNTER,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: DAVID

DATE/TIME OF TEST: 10/27/2018 02:44:24

---

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.081	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.000	02:40
Breath Volume	1.617	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.000	02:43
Breath Volume	2.191	02:43
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

---

Data Download Date/Time: 10/28/2018 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBB,CATHERINE,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: KNIGHT,G  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 10/27/2018 22:52:22

---

Diagnostics Check	OK	22:41	
Air Blank	0.000	22:41	
Control Test	0.081	22:41	
Air Blank	0.000	22:42	
Subject Sample	# 1	22:45	No Sample Provided
Breath Volume	0.000	22:45	
Air Blank	0.000	22:45	
Air Blank	0.000	22:47	
Subject Sample	# 2	22:51	No Sample Provided
Breath Volume	0.000	22:51	
Air Blank	0.000	22:51	
Control Test	0.080	22:51	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

---

Data Download Date/Time: 10/28/2018 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEALY,GREGORY,B  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: HEALY,GREGORY  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 10/28/2018 00:36:38

---

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.081	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.148	00:32
Breath Volume	3.207	00:32
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.151	00:35
Breath Volume	3.695	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

---

Data Download Date/Time: 10/28/2018 10:04



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBB,CATHERINE,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: COPHER,C  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 11:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: HANSON

DATE/TIME OF TEST: 10/28/2018 01:30:32

---

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.080	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.192	01:26
Breath Volume	1.414	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.197	01:29
Breath Volume	2.472	01:29
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

---

Data Download Date/Time: 10/28/2018 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBB,CATHERINE,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUNTER,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/24/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HANNANS

DATE/TIME OF TEST: 10/28/2018 02:48:41

---

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.081	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.061	02:44
Breath Volume	2.617	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.064	02:47
Breath Volume	3.148	02:47
Air Blank	0.000	02:47
Control Test	0.080	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

---

Data Download Date/Time: 10/28/2018 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PECK,JOHN,R  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: TROUT,STUART  
ARRESTING OFFICER AGENCY: Lake Placid PD  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: CHIU

DATE/TIME OF TEST: 10/08/2018 16:31:59

---

Diagnostics Check	OK	16:23
Air Blank	0.000	16:23
Control Test	0.079	16:23
Air Blank	0.000	16:24
Subject Sample # 1	0.275	16:27
Breath Volume	1.871	16:27
Air Blank	0.000	16:27
Air Blank	0.000	16:29
Subject Sample # 2	0.277	16:30
Breath Volume	1.675	16:30
Air Blank	0.000	16:31
Control Test	0.077	16:31
Air Blank	0.000	16:31
Diagnostics Check	OK	16:31

---

Data Download Date/Time: 10/20/2018 12:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MURPHY,JAMES,  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: GONZALEZ,SANDRA  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: COURT ORD  
ARREST TIME: 11:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:39

SUBJECT'S LAST NAME: FILEHNE

DATE/TIME OF TEST: 10/09/2018 12:09:39

---

Diagnostics Check	OK	12:03
Air Blank	0.000	12:04
Control Test	0.079	12:04
Air Blank	0.000	12:04
Subject Sample # 1	0.000	12:05
Breath Volume	1.894	12:05
Air Blank	0.000	12:05
Air Blank	0.000	12:07
Subject Sample # 2	0.000	12:08
Breath Volume	1.871	12:08
Air Blank	0.000	12:08
Control Test	0.079	12:09
Air Blank	0.000	12:09
Diagnostics Check	OK	12:09

---

Data Download Date/Time: 10/20/2018 12:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KEEFE,JAYME,W  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: RITCHIE,SHELBY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: HENSON

DATE/TIME OF TEST: 10/12/2018 02:01:46

---

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:59	
Control Test	0.079	01:59	
Air Blank	0.000	01:59	
Subject Sample	# 1	02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:00	
Control Test	0.079	02:01	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

---

Data Download Date/Time: 10/20/2018 12:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROE,ROBERT,D  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: MARIN,ADRIAN  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GEORGES

DATE/TIME OF TEST: 10/14/2018 00:57:07

---

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.000	00:52
Breath Volume	3.550	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.000	00:55
Breath Volume	3.292	00:55
Air Blank	0.000	00:56
Control Test	0.079	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

---

Data Download Date/Time: 10/20/2018 12:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARLOW,ROYCE,B  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: MARLEY,TREY  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: AARON

DATE/TIME OF TEST: 10/15/2018 20:07:59

---

Diagnostics Check	OK	20:00
Air Blank	0.000	20:01
Control Test	0.079	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.261	20:03
Breath Volume	1.199	20:03
Air Blank	0.000	20:04
Air Blank	0.000	20:05
Subject Sample # 2	0.258	20:06
Breath Volume	1.363	20:06
Air Blank	0.000	20:07
Control Test	0.079	20:07
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

---

Data Download Date/Time: 10/20/2018 12:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGOODIAL,TRISTON,T  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: RAGOODIAL,TRISTON  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: COURT ORD  
ARREST TIME: 10:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/17/2018 10:49:03

---

Diagnostics Check	OK	10:42
Air Blank	0.000	10:43
Control Test	0.080	10:43
Air Blank	0.000	10:44
Subject Sample # 1	0.000	10:44
Breath Volume	1.960	10:44
Air Blank	0.000	10:45
Air Blank	0.000	10:47
Subject Sample # 2	0.000	10:47
Breath Volume	1.519	10:47
Air Blank	0.000	10:48
Control Test	0.080	10:48
Air Blank	0.000	10:48
Diagnostics Check	OK	10:49

---

Data Download Date/Time: 10/20/2018 12:42



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROE,ROBERT,D  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: BECERRA,JOSE  
ARRESTING OFFICER AGENCY: Lake Placid PD  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:59

SUBJECT'S LAST NAME: TUCCIO

DATE/TIME OF TEST: 10/17/2018 18:26:00

---

Diagnostics Check	OK	18:19
Air Blank	0.000	18:20
Control Test	0.078	18:20
Air Blank	0.000	18:20
Subject Sample # 1	0.065	18:21
Breath Volume	1.167	18:21
Air Blank	0.000	18:21
Air Blank	0.000	18:23
Subject Sample # 2	0.062	18:24
Breath Volume	1.222	18:24
Air Blank	0.000	18:25
Control Test	0.080	18:25
Air Blank	0.000	18:25
Diagnostics Check	OK	18:26

---

Data Download Date/Time: 10/20/2018 12:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KEEFE,JAYME,W  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: ORTEGA,YULIER  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 10/20/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: DIAZ PEREZ

DATE/TIME OF TEST: 10/20/2018 21:09:12

---

Diagnostics Check		OK	21:02
Air Blank		0.000	21:02
Control Test		0.080	21:02
Air Blank		0.000	21:03
Subject Sample	# 1	0.239	21:03
Breath Volume		2.312	21:03
Air Blank		0.000	21:04
Air Blank		0.000	21:06
Subject Sample	# 2	0.246	21:07
Breath Volume		2.460	21:07
Air Blank		0.000	21:08
Control Test		0.079	21:08
Air Blank		0.000	21:09
Diagnostics Check		OK	21:09

---

Data Download Date/Time: 11/27/2018 07:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRISON,DUSTIN,J  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: HARRISON,DUSTIN  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 10/20/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: COURT ORD  
ARREST TIME: 09:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: TUCCIO

DATE/TIME OF TEST: 10/22/2018 09:49:58

---

Diagnostics Check	OK	09:43
Air Blank	0.000	09:44
Control Test	0.079	09:44
Air Blank	0.000	09:44
Subject Sample # 1	0.000	09:45
Breath Volume	1.394	09:45
Air Blank	0.000	09:45
Air Blank	0.000	09:47
Subject Sample # 2	0.000	09:48
Breath Volume	1.613	09:48
Air Blank	0.000	09:49
Control Test	0.079	09:49
Air Blank	0.000	09:49
Diagnostics Check	OK	09:49

---

Data Download Date/Time: 11/27/2018 07:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBLES,J,V  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: MARIN,A  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 10/20/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: PALMORE III

DATE/TIME OF TEST: 10/24/2018 03:27:50

---

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.079	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.068	03:23
Breath Volume	1.750	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.065	03:26
Breath Volume	2.476	03:26
Air Blank	0.000	03:26
Control Test	0.079	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

---

Data Download Date/Time: 11/27/2018 07:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGOODIAL,TRISTON,T  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: MARIN,ADRIAN  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 10/20/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: JAIMES

DATE/TIME OF TEST: 10/26/2018 20:55:32

---

Diagnostics Check	OK	20:49
Air Blank	0.000	20:49
Control Test	0.078	20:49
Air Blank	0.000	20:50
Subject Sample # 1	0.000	20:51
Breath Volume	3.453	20:51
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.000	20:54
Breath Volume	2.941	20:54
Air Blank	0.000	20:54
Control Test	0.079	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

---

Data Download Date/Time: 11/27/2018 07:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGOODIAL,TRISTON,T  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: MARIN,ADRIAN  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 10/20/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 10/27/2018 03:12:05

---

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.235	03:07
Breath Volume	3.007	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.238	03:10
Breath Volume	2.878	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

---

Data Download Date/Time: 11/27/2018 07:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGOODIAL,TRISTON,T  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: MARIN,ADRIAN  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 10/20/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: WUDTKE

DATE/TIME OF TEST: 10/27/2018 23:14:44

---

Diagnostics Check	OK	23:07
Air Blank	0.000	23:08
Control Test	0.079	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.221	23:10
Breath Volume	2.203	23:10
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.207	23:13
Breath Volume	1.601	23:13
Air Blank	0.000	23:13
Control Test	0.079	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

---

Data Download Date/Time: 11/27/2018 07:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAINES,DAVID,K  
BREATH TEST OPERATOR AGENCY: Davie PD  
ARRESTING OFFICER: JESSUP,BRIAN  
ARRESTING OFFICER AGENCY: Davie PD  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG728303  
DRY GAS EXPIRY DATE: 10/10/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 10/12/2018 22:46:20

---

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.081	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.169	22:41
Breath Volume	3.007	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.162	22:44
Breath Volume	2.531	22:44
Air Blank	0.000	22:45
Control Test	0.081	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

---

Data Download Date/Time: 10/16/2018 17:13



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,JASON,E  
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD  
ARRESTING OFFICER: MORENO,STEPHEN  
ARRESTING OFFICER AGENCY: Fernandina Beach PD  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A7  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: BERNARDING

DATE/TIME OF TEST: 10/07/2018 01:20:44

---

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:18	
Control Test	0.078	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Control Test	0.078	01:20	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

---

Data Download Date/Time: 10/23/2018 12:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORPHEE, MICHAEL,  
BREATH TEST OPERATOR AGENCY: Winter Haven PD  
ARRESTING OFFICER: AVERY, JONATHAN  
ARRESTING OFFICER AGENCY: Winter Haven PD  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 10/05/2018 18:58:21

---

Diagnostics Check	OK	18:53	
Air Blank	0.000	18:53	
Control Test	0.079	18:53	
Air Blank	0.000	18:54	
Subject Sample	# 1	18:57	Interferent Detect
Breath Volume	1.566	18:57	
Air Blank		18:57	Purge Fail
Air Blank	0.000	18:58	

---

Data Download Date/Time: 10/29/2018 15:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORPHEE, MICHAEL,  
BREATH TEST OPERATOR AGENCY: Winter Haven PD  
ARRESTING OFFICER: AVERY, JONATHAN  
ARRESTING OFFICER AGENCY: Winter Haven PD  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 10/05/2018 19:09:58

---

Diagnostics Check	OK	19:01
Air Blank	0.000	19:01
Control Test	0.080	19:02
Air Blank	0.000	19:02
Subject Sample # 1	0.069	19:03
Breath Volume	1.441	19:03
Air Blank	0.000	19:04
Air Blank	0.000	19:05
Subject Sample # 2	0.071	19:08
Breath Volume	1.210	19:08
Air Blank	0.000	19:08
Control Test	0.083	19:09
Air Blank	0.000	19:09
Diagnostics Check	OK	19:09

---

Data Download Date/Time: 10/29/2018 15:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEBSON,BRET,E  
BREATH TEST OPERATOR AGENCY: Winter Haven PD  
ARRESTING OFFICER: HEBSON,BRET  
ARRESTING OFFICER AGENCY: Winter Haven PD  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: COURT ORD  
ARREST TIME: 16:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: HITLEY

DATE/TIME OF TEST: 10/15/2018 16:45:55

---

Diagnostics Check	OK	16:39
Air Blank	0.000	16:39
Control Test	0.079	16:40
Air Blank	0.000	16:40
Subject Sample # 1	0.000	16:41
Breath Volume	2.199	16:41
Air Blank	0.000	16:41
Air Blank	0.000	16:43
Subject Sample # 2	0.000	16:44
Breath Volume	2.226	16:44
Air Blank	0.000	16:44
Control Test	0.080	16:45
Air Blank	0.000	16:45
Diagnostics Check	OK	16:45

---

Data Download Date/Time: 10/29/2018 15:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CALLAWAY, BRADLEY, L  
BREATH TEST OPERATOR AGENCY: Winter Haven PD  
ARRESTING OFFICER: CALLAWAY, BRADLEY  
ARRESTING OFFICER AGENCY: Winter Haven PD  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 09:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:26

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 10/16/2018 09:57:03

---

Diagnostics Check	OK	09:50
Air Blank	0.000	09:51
Control Test	0.078	09:51
Air Blank	0.000	09:52
Subject Sample # 1	0.000	09:52
Breath Volume	1.640	09:52
Air Blank	0.000	09:53
Air Blank	0.000	09:55
Subject Sample # 2	0.000	09:55
Breath Volume	1.824	09:55
Air Blank	0.000	09:56
Control Test	0.079	09:56
Air Blank	0.000	09:56
Diagnostics Check	OK	09:57

---

Data Download Date/Time: 10/29/2018 15:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORPHEE,MICHAEL,  
BREATH TEST OPERATOR AGENCY: Winter Haven PD  
ARRESTING OFFICER: ORPHEE,MICHAEL  
ARRESTING OFFICER AGENCY: Winter Haven PD  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 10/21/2018 02:22:39

---

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:19	
Control Test	0.079	02:19	
Air Blank	0.000	02:20	
Subject Sample	# 1	02:21	Subject Test Refused
Breath Volume	0.000	02:21	
Air Blank	0.000	02:21	
Control Test	0.078	02:21	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

---

Data Download Date/Time: 10/29/2018 15:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARTER,TIMOTHY,R

BREATH TEST OPERATOR AGENCY: Gulfport PD

ARRESTING OFFICER: CARTER,TIMOTHY

ARRESTING OFFICER AGENCY: Gulfport PD

LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A2

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 23:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: BOLAN

DATE/TIME OF TEST: 10/04/2018 00:24:26

---

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:21	
Control Test	0.081	00:21	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:23	
Control Test	0.079	00:23	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

---

Data Download Date/Time: 10/31/2018 03:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VANDENBERG,MICHAEL,J

BREATH TEST OPERATOR AGENCY: Gulfport PD

ARRESTING OFFICER: CLAGUE,ERIC

ARRESTING OFFICER AGENCY: Gulfport PD

LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A2

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 04:58

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: QUARTERMAN

DATE/TIME OF TEST: 10/29/2018 05:37:25

---

Diagnostics Check	OK	05:34	
Air Blank	0.000	05:34	
Control Test	0.081	05:35	
Air Blank	0.000	05:35	
Subject Sample	# 1	05:36	Subject Test Refused
Breath Volume	0.000	05:36	
Air Blank	0.000	05:36	
Control Test	0.080	05:36	
Air Blank	0.000	05:37	
Diagnostics Check	OK	05:37	

---

Data Download Date/Time: 10/31/2018 03:11



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCONNELL,J,W  
BREATH TEST OPERATOR AGENCY: New Port Richey PD  
ARRESTING OFFICER: CROWELL,D  
ARRESTING OFFICER AGENCY: New Port Richey PD  
LAST AGENCY INSPECTION: 10/21/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A6  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: OTHER  
ARREST TIME: 19:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: COUTRE-WOODS

DATE/TIME OF TEST: 10/25/2018 21:22:11

---

Diagnostics Check	OK	21:15
Air Blank	0.000	21:15
Control Test	0.079	21:16
Air Blank	0.000	21:16
Subject Sample # 1	0.216	21:17
Breath Volume	3.160	21:17
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.205	21:20
Breath Volume	2.070	21:20
Air Blank	0.000	21:21
Control Test	0.079	21:21
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

---

Data Download Date/Time: 11/18/2018 23:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHELL,MATTHEW,T  
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD  
ARRESTING OFFICER: SCHELL,MATTHEW  
ARRESTING OFFICER AGENCY: Tarpon Springs PD  
LAST AGENCY INSPECTION: 09/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A3  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: CHEKRYGIN

DATE/TIME OF TEST: 10/02/2018 20:06:06

---

Diagnostics Check	OK	19:57
Air Blank	0.000	19:58
Control Test	0.078	19:58
Air Blank	0.000	19:59
Subject Sample # 1	0.000	20:00
Breath Volume	1.312	20:00
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample # 2	0.000	20:04
Breath Volume	1.546	20:04
Air Blank	0.000	20:05
Control Test	0.078	20:05
Air Blank	0.000	20:06
Diagnostics Check	OK	20:06

---

Data Download Date/Time: 10/29/2018 14:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RONDOS,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD  
ARRESTING OFFICER: RONDOS,ROBERT  
ARRESTING OFFICER AGENCY: Tarpon Springs PD  
LAST AGENCY INSPECTION: 09/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A3  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: BARBERI

DATE/TIME OF TEST: 10/09/2018 19:16:37

---

Diagnostics Check	OK	19:10
Air Blank	0.000	19:10
Control Test	0.078	19:11
Air Blank	0.000	19:11
Subject Sample # 1	0.411	19:12
Breath Volume	1.171	19:12
Air Blank	0.000	19:12
Air Blank	0.000	19:14
Subject Sample # 2	0.415	19:15
Breath Volume	1.425	19:15
Air Blank	0.000	19:15
Control Test	0.076	19:16
Air Blank	0.000	19:16
Diagnostics Check	OK	19:16

---

Data Download Date/Time: 10/29/2018 14:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHELL, MATTHEW, T  
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD  
ARRESTING OFFICER: GIBSON, S  
ARRESTING OFFICER AGENCY: Tarpon Springs PD  
LAST AGENCY INSPECTION: 09/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A3  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: NEELY

DATE/TIME OF TEST: 10/20/2018 00:45:08

---

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:32	
Control Test	0.077	00:32	
Air Blank	0.000	00:33	
Subject Sample	# 1 0.091	00:34	
Breath Volume	1.375	00:34	
Air Blank	0.000	00:35	
Air Blank	0.000	00:37	
Subject Sample	# 2 0.075	00:40	Volume Not Met
Breath Volume	0.621	00:40	
Air Blank	0.000	00:41	
Air Blank	0.000	00:43	
Subject Sample	# 3 0.087	00:43	
Breath Volume	1.304	00:43	
Air Blank	0.000	00:44	
Control Test	0.077	00:44	
Air Blank	0.000	00:45	
Diagnostics Check	OK	00:45	

---

Data Download Date/Time: 10/29/2018 14:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RONDOS,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD  
ARRESTING OFFICER: RONDOS,ROBERT  
ARRESTING OFFICER AGENCY: Tarpon Springs PD  
LAST AGENCY INSPECTION: 09/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A3  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: GANIER

DATE/TIME OF TEST: 10/26/2018 19:24:48

---

Diagnostics Check	OK	19:17
Air Blank	0.000	19:17
Control Test	0.077	19:18
Air Blank	0.000	19:18
Subject Sample # 1	0.224	19:19
Breath Volume	1.488	19:19
Air Blank	0.000	19:20
Air Blank	0.000	19:22
Subject Sample # 2	0.231	19:23
Breath Volume	2.042	19:23
Air Blank	0.000	19:23
Control Test	0.076	19:24
Air Blank	0.000	19:24
Diagnostics Check	OK	19:24

---

Data Download Date/Time: 10/29/2018 14:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANEY,DAMON,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LANEY,DAMON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: WHALEN

DATE/TIME OF TEST: 10/01/2018 03:39:18

---

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.081	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.123	03:34
Breath Volume	1.832	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.125	03:37
Breath Volume	1.972	03:37
Air Blank	0.000	03:38
Control Test	0.081	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

---

Data Download Date/Time: 10/24/2018 21:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBY,LARRY,K  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: JACOBY,LARRY  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: KURTZ

DATE/TIME OF TEST: 10/03/2018 05:13:44

---

Diagnostics Check	OK	05:07
Air Blank	0.000	05:07
Control Test	0.082	05:08
Air Blank	0.000	05:08
Subject Sample # 1	0.109	05:09
Breath Volume	1.246	05:09
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.103	05:12
Breath Volume	2.734	05:12
Air Blank	0.000	05:12
Control Test	0.081	05:13
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

---

Data Download Date/Time: 10/24/2018 21:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANTHONY, MICHAEL,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: RICOTTILLI, JOSHUA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: SLATER

DATE/TIME OF TEST: 10/04/2018 04:38:29

---

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.082	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.103	04:33
Breath Volume	1.554	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.099	04:36
Breath Volume	2.476	04:36
Air Blank	0.000	04:37
Control Test	0.081	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

---

Data Download Date/Time: 10/24/2018 21:51



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANTHONY, MICHAEL,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: RICOTTILLI, JOSHUA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 10/04/2018 23:16:19

---

Diagnostics Check	OK	23:09
Air Blank	0.000	23:10
Control Test	0.081	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.148	23:11
Breath Volume	2.937	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.149	23:14
Breath Volume	2.578	23:14
Air Blank	0.000	23:15
Control Test	0.081	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

---

Data Download Date/Time: 10/24/2018 21:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANTHONY, MICHAEL,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WEDE, BRENNEN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: NOLTON

DATE/TIME OF TEST: 10/05/2018 01:10:48

---

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:08	
Control Test	0.082	01:08	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:09	Subject Test Refused
Breath Volume	0.000	01:09	
Air Blank	0.000	01:09	
Control Test	0.082	01:10	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:10	

---

Data Download Date/Time: 10/24/2018 21:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANTHONY, MICHAEL,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: RICOTTILLI, JOSHUA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: DONALDSON

DATE/TIME OF TEST: 10/05/2018 03:20:55

---

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:09	
Control Test		0.082	03:10	
Air Blank		0.000	03:10	
Subject Sample	# 1	0.097	03:13	Volume Not Met
Breath Volume		0.773	03:13	
Air Blank		0.000	03:14	
Air Blank		0.000	03:16	
Subject Sample	# 2	0.088	03:19	Volume Not Met
Breath Volume		0.804	03:19	
Air Blank		0.000	03:20	
Control Test		0.081	03:20	
Air Blank		0.000	03:20	
Diagnostics Check		OK	03:20	

---

Data Download Date/Time: 10/24/2018 21:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANTHONY, MICHAEL,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MASON, GERALD  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 10/05/2018 04:45:02

---

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.082	04:37
Air Blank		0.000	04:38
Subject Sample	# 1	0.225	04:39
Breath Volume		1.402	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.225	04:43
Breath Volume		1.199	04:43
Air Blank		0.000	04:44
Control Test		0.081	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:45

---

Data Download Date/Time: 10/24/2018 21:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANTHONY, MICHAEL,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WEDE, BRENNEN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: MACKIN

DATE/TIME OF TEST: 10/06/2018 02:23:45

---

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.082	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.031	02:19
Breath Volume	2.500	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.030	02:22
Breath Volume	2.753	02:22
Air Blank	0.000	02:22
Control Test	0.082	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

---

Data Download Date/Time: 10/24/2018 21:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANTHONY, MICHAEL,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WEDE, BRENNEN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: SNELL

DATE/TIME OF TEST: 10/06/2018 22:24:32

---

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.081	22:18
Air Blank	0.000	22:18
Subject Sample # 1	0.162	22:19
Breath Volume	1.656	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.164	22:23
Breath Volume	1.703	22:23
Air Blank	0.000	22:23
Control Test	0.081	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

---

Data Download Date/Time: 10/24/2018 21:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANTHONY, MICHAEL,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: RICOTTILLI, JOSHUA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: SPALLA

DATE/TIME OF TEST: 10/07/2018 00:22:39

---

Diagnostics Check		OK	00:11	
Air Blank		0.000	00:11	
Control Test		0.082	00:11	
Air Blank		0.000	00:12	
Subject Sample	# 1	0.196	00:15	Volume Not Met
Breath Volume		0.382	00:15	
Air Blank		0.000	00:16	
Air Blank		0.000	00:17	
Subject Sample	# 2	0.216	00:21	Volume Not Met
Breath Volume		0.390	00:21	
Air Blank		0.000	00:21	
Control Test		0.080	00:22	
Air Blank		0.000	00:22	
Diagnostics Check		OK	00:22	

---

Data Download Date/Time: 10/24/2018 21:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANTHONY, MICHAEL,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WEDE, BRENNEN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: FUMEA

DATE/TIME OF TEST: 10/07/2018 01:52:06

---

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.081	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.000	01:47
Breath Volume	2.839	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.000	01:50
Breath Volume	2.457	01:50
Air Blank	0.000	01:51
Control Test	0.082	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

---

Data Download Date/Time: 10/24/2018 21:51



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBY,LARRY,K  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: JACOBY,LARRY  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HIMES

DATE/TIME OF TEST: 10/07/2018 03:07:15

---

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.081	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.023	03:02
Breath Volume		1.316	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.023	03:05
Breath Volume		1.523	03:05
Air Blank		0.000	03:06
Control Test		0.081	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

---

Data Download Date/Time: 10/24/2018 21:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PERSAUD,COMPTON,D  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CAMPION,K  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: GUNNING

DATE/TIME OF TEST: 10/08/2018 03:00:53

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.083	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.108	02:55
Breath Volume	1.753	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.104	02:59
Breath Volume	1.589	02:59
Air Blank	0.000	02:59
Control Test	0.082	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

---

Data Download Date/Time: 10/24/2018 21:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANEY,D,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LESANE,A  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: BOGGS JR

DATE/TIME OF TEST: 10/09/2018 04:51:17

---

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.082	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.214	04:46
Breath Volume	1.488	04:46
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.223	04:49
Breath Volume	2.218	04:49
Air Blank	0.000	04:50
Control Test	0.081	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

---

Data Download Date/Time: 10/24/2018 21:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANTHONY, MICHAEL,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: COOK, CHRISTOPHER  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: STEIN

DATE/TIME OF TEST: 10/10/2018 23:57:59

---

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.082	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.189	23:52
Breath Volume	1.960	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.189	23:56
Breath Volume	1.890	23:56
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

---

Data Download Date/Time: 10/24/2018 21:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,WILLIAM,T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SMITH,WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 10/12/2018 22:09:13

---

Diagnostics Check	OK	22:02
Air Blank	0.000	22:03
Control Test	0.083	22:03
Air Blank	0.000	22:04
Subject Sample # 1	0.374	22:04
Breath Volume	2.839	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.365	22:07
Breath Volume	3.093	22:07
Air Blank	0.000	22:08
Control Test	0.081	22:08
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

---

Data Download Date/Time: 10/24/2018 21:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIFFIN,KYLE,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: GRILL,SHANNON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: LOPEMAN

DATE/TIME OF TEST: 10/13/2018 04:50:21

---

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:47	
Control Test	0.083	04:47	
Air Blank	0.000	04:47	
Subject Sample	# 1	04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:49	
Control Test	0.083	04:49	
Air Blank	0.000	04:50	
Diagnostics Check	OK	04:50	

---

Data Download Date/Time: 10/24/2018 21:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUZ,KIOMY,L  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: CRUZ,KIOMY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 08:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:53

SUBJECT'S LAST NAME: DZIEDZIC

DATE/TIME OF TEST: 10/13/2018 10:23:53

---

Diagnostics Check	OK	10:21	
Air Blank	0.000	10:21	
Control Test	0.083	10:21	
Air Blank	0.000	10:22	
Subject Sample	# 1	10:22	Subject Test Refused
Breath Volume	0.000	10:22	
Air Blank	0.000	10:22	
Control Test	0.083	10:23	
Air Blank	0.000	10:23	
Diagnostics Check	OK	10:23	

---

Data Download Date/Time: 10/24/2018 21:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KASELAK,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: RICHAU,AUBREY  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 11:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:37

SUBJECT'S LAST NAME: HOLLANDSWORTH

DATE/TIME OF TEST: 10/13/2018 12:04:42

---

Diagnostics Check	OK	11:58
Air Blank	0.000	11:58
Control Test	0.083	11:58
Air Blank	0.000	11:59
Subject Sample # 1	0.197	11:59
Breath Volume	1.675	11:59
Air Blank	0.000	12:00
Air Blank	0.000	12:02
Subject Sample # 2	0.188	12:03
Breath Volume	1.210	12:03
Air Blank	0.000	12:03
Control Test	0.082	12:04
Air Blank	0.000	12:04
Diagnostics Check	OK	12:04

---

Data Download Date/Time: 10/24/2018 21:51



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,WILLIAM,T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SMITH,WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: KATZ

DATE/TIME OF TEST: 10/14/2018 02:30:10

---

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.083	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.000	02:25
Breath Volume	2.511	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.000	02:28
Breath Volume	2.777	02:28
Air Blank	0.000	02:29
Control Test	0.083	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

---

Data Download Date/Time: 10/24/2018 21:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANEY,DAMON,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LANEY,DAMON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: OSBOURN

DATE/TIME OF TEST: 10/14/2018 04:53:03

---

Diagnostics Check	OK	04:47
Air Blank	0.000	04:47
Control Test	0.082	04:47
Air Blank	0.000	04:48
Subject Sample # 1	0.169	04:48
Breath Volume	2.656	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.166	04:51
Breath Volume	1.445	04:51
Air Blank	0.000	04:52
Control Test	0.082	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:53

---

Data Download Date/Time: 10/24/2018 21:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: DONOHUE, EDWARD  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 10/14/2018 20:20:20

---

Diagnostics Check	OK	20:14
Air Blank	0.000	20:14
Control Test	0.082	20:15
Air Blank	0.000	20:15
Subject Sample # 1	0.181	20:16
Breath Volume	1.726	20:16
Air Blank	0.000	20:16
Air Blank	0.000	20:18
Subject Sample # 2	0.180	20:18
Breath Volume	1.292	20:18
Air Blank	0.000	20:19
Control Test	0.082	20:19
Air Blank	0.000	20:20
Diagnostics Check	OK	20:20

---

Data Download Date/Time: 10/24/2018 21:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,WILLIAM,T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SMITH,WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 10/15/2018 03:24:02

---

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.083	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.140	03:19
Breath Volume	2.402	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.145	03:22
Breath Volume	2.468	03:22
Air Blank	0.000	03:23
Control Test	0.083	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:24

---

Data Download Date/Time: 10/24/2018 21:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADA, RAYMOND, O  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: ADA, RAYMOND  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 13:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: WHITEHEAD

DATE/TIME OF TEST: 10/16/2018 15:23:41

---

Diagnostics Check	OK	15:12	
Air Blank	0.000	15:12	
Control Test	0.083	15:13	
Air Blank	0.000	15:13	
Subject Sample	# 1	15:16	No Sample Provided
Breath Volume	0.000	15:16	
Air Blank	0.000	15:17	
Air Blank	0.000	15:19	
Subject Sample	# 2	15:22	No Sample Provided
Breath Volume	0.000	15:22	
Air Blank	0.000	15:22	
Control Test	0.083	15:23	
Air Blank	0.000	15:23	
Diagnostics Check	OK	15:23	

---

Data Download Date/Time: 10/24/2018 21:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: AMATRUDA,CHRIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: BUBAR

DATE/TIME OF TEST: 10/17/2018 22:09:47

---

Diagnostics Check	OK	21:58	
Air Blank	0.000	21:58	
Control Test	0.084	21:59	
Air Blank	0.000	21:59	
Subject Sample	# 1	22:02	No Sample Provided
Breath Volume	0.000	22:02	
Air Blank	0.000	22:03	
Air Blank	0.000	22:05	
Subject Sample	# 2	22:08	No Sample Provided
Breath Volume	0.000	22:08	
Air Blank	0.000	22:08	
Control Test	0.083	22:09	
Air Blank	0.000	22:09	
Diagnostics Check	OK	22:09	

---

Data Download Date/Time: 10/24/2018 21:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: RICOTTILLI,J  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MITCHELL JR

DATE/TIME OF TEST: 10/18/2018 02:27:40

---

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.084	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.028	02:23
Breath Volume	1.730	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.029	02:26
Breath Volume	2.941	02:26
Air Blank	0.000	02:26
Control Test	0.084	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

---

Data Download Date/Time: 10/24/2018 21:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: AMATRUDA,CHRIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: ZACHMAN

DATE/TIME OF TEST: 10/18/2018 03:27:43

---

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:16	
Control Test	0.082	03:17	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:20	No Sample Provided
Breath Volume	0.000	03:20	
Air Blank	0.000	03:21	
Air Blank	0.000	03:23	
Subject Sample	# 2	03:26	No Sample Provided
Breath Volume	0.000	03:26	
Air Blank	0.000	03:26	
Control Test	0.082	03:27	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

---

Data Download Date/Time: 10/24/2018 21:52



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANGANARO,JESSICA,H  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: MANGANARO,JESSICA  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: GOETZ

DATE/TIME OF TEST: 10/18/2018 20:10:15

---

Diagnostics Check	OK	20:07	
Air Blank	0.000	20:07	
Control Test	0.082	20:08	
Air Blank	0.000	20:08	
Subject Sample	# 1	20:08	Subject Test Refused
Breath Volume	0.000	20:08	
Air Blank	0.000	20:09	
Control Test	0.082	20:09	
Air Blank	0.000	20:10	
Diagnostics Check	OK	20:10	

---

Data Download Date/Time: 10/24/2018 21:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, WILLIAM, T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SMITH, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: SLOBODA

DATE/TIME OF TEST: 10/19/2018 00:17:08

---

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:06	
Control Test	0.082	00:06	
Air Blank	0.000	00:07	
Subject Sample	# 1	00:10	No Sample Provided
Breath Volume	0.000	00:10	
Air Blank	0.000	00:10	
Air Blank	0.000	00:12	
Subject Sample	# 2	00:15	No Sample Provided
Breath Volume	0.000	00:15	
Air Blank	0.000	00:16	
Control Test	0.082	00:16	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

---

Data Download Date/Time: 10/24/2018 21:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: RICOTTILLI,J  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: KOLB

DATE/TIME OF TEST: 10/19/2018 01:01:33

---

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:58	
Control Test	0.082	00:59	
Air Blank	0.000	00:59	
Subject Sample	# 1	01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:00	
Control Test	0.082	01:00	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:01	

---

Data Download Date/Time: 10/24/2018 21:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WEDE,B  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: WHITNEY

DATE/TIME OF TEST: 10/19/2018 02:58:27

---

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.082	02:56	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:57	
Control Test	0.082	02:57	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

---

Data Download Date/Time: 10/24/2018 21:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOWATT,NATHAN,E  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MOWATT,NATHAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 10/19/2018 03:36:41

---

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:34	
Control Test	0.082	03:34	
Air Blank	0.000	03:35	
Subject Sample	# 1	03:35	Subject Test Refused
Breath Volume	0.000	03:35	
Air Blank	0.000	03:35	
Control Test	0.082	03:36	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:36	

---

Data Download Date/Time: 10/24/2018 21:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: FIELDEN,E  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: COUTRE

DATE/TIME OF TEST: 10/20/2018 02:33:59

---

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.082	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.134	02:29
Breath Volume	2.015	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.133	02:32
Breath Volume	1.597	02:32
Air Blank	0.000	02:33
Control Test	0.081	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

---

Data Download Date/Time: 10/24/2018 21:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOWATT,NATHAN,E  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MOWATT,NATHAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: HYYPIO

DATE/TIME OF TEST: 10/20/2018 03:22:47

---

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:20	
Control Test	0.082	03:20	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:21	
Control Test	0.081	03:22	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

---

Data Download Date/Time: 10/24/2018 21:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: AMATRUDA,CHRIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 10/20/2018 04:45:49

---

Diagnostics Check	OK	04:34	
Air Blank	0.000	04:34	
Control Test	0.082	04:35	
Air Blank	0.000	04:35	
Subject Sample	# 1	04:38	No Sample Provided
Breath Volume	0.000	04:38	
Air Blank	0.000	04:39	
Air Blank	0.000	04:41	
Subject Sample	# 2	04:44	No Sample Provided
Breath Volume	0.000	04:44	
Air Blank	0.000	04:44	
Control Test	0.082	04:45	
Air Blank	0.000	04:45	
Diagnostics Check	OK	04:45	

---

Data Download Date/Time: 10/24/2018 21:52



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANTHONY, MICHAEL,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WEDE, BRENNEN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 10/20/2018 05:27:39

---

Diagnostics Check	OK	05:24	
Air Blank	0.000	05:25	
Control Test	0.082	05:25	
Air Blank	0.000	05:26	
Subject Sample	# 1	05:26	Subject Test Refused
Breath Volume	0.000	05:26	
Air Blank	0.000	05:26	
Control Test	0.081	05:27	
Air Blank	0.000	05:27	
Diagnostics Check	OK	05:27	

---

Data Download Date/Time: 10/24/2018 21:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KASELAK,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: KASELAK,STEPHEN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: BRITTON

DATE/TIME OF TEST: 10/20/2018 23:21:15

---

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.081	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.212	23:16
Breath Volume	2.757	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.205	23:19
Breath Volume	2.863	23:19
Air Blank	0.000	23:20
Control Test	0.082	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

---

Data Download Date/Time: 10/24/2018 21:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORGAN,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: BEHERENHOUSER,IAN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 10/21/2018 02:41:05

---

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.081	02:27	
Air Blank	0.000	02:27	
Subject Sample	# 1 0.191	02:30	Volume Not Met
Breath Volume	0.363	02:30	
Air Blank	0.000	02:31	
Air Blank	0.000	02:33	
Subject Sample	# 2 0.210	02:33	
Breath Volume	1.687	02:33	
Air Blank	0.000	02:34	
Air Blank	0.000	02:36	
Subject Sample	# 3 0.183	02:39	Volume Not Met
Breath Volume	0.144	02:39	
Air Blank		02:40	Purge Fail
Air Blank		02:41	Purge Fail

---

Data Download Date/Time: 10/24/2018 21:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WEDE,B  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: ARANY

DATE/TIME OF TEST: 10/21/2018 03:35:02

---

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:34	
Control Test	0.055	03:34	Control Outside Tolerance
Air Blank	0.000	03:35	

---

Data Download Date/Time: 10/24/2018 21:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WEDE,B  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/12/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: ARANY

DATE/TIME OF TEST: 10/21/2018 03:38:42

---

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.060	03:38	Control Outside Tolerance
Air Blank	0.000	03:38	

---

Data Download Date/Time: 10/24/2018 21:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HERNANDEZ,EDDY  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 09/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MCLYMONT

DATE/TIME OF TEST: 10/02/2018 03:05:26

---

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.080	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.148	03:01
Breath Volume		2.218	03:01
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.145	03:03
Breath Volume		1.816	03:03
Air Blank		0.000	03:04
Control Test		0.080	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

---

Data Download Date/Time: 10/17/2018 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: OSORIO,MIGUEL  
ARRESTING OFFICER AGENCY: Pembroke Pines PD  
LAST AGENCY INSPECTION: 09/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 07:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:58

SUBJECT'S LAST NAME: PIEDRA

DATE/TIME OF TEST: 10/04/2018 09:28:38

---

Diagnostics Check	OK	09:22
Air Blank	0.000	09:23
Control Test	0.080	09:23
Air Blank	0.000	09:23
Subject Sample # 1	0.000	09:24
Breath Volume	4.015	09:24
Air Blank	0.000	09:24
Air Blank	0.000	09:26
Subject Sample # 2	0.000	09:27
Breath Volume	3.898	09:27
Air Blank	0.000	09:27
Control Test	0.079	09:28
Air Blank	0.000	09:28
Diagnostics Check	OK	09:28

---

Data Download Date/Time: 10/17/2018 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: STRZALKOWSKI,SEAN  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 09/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: GUTHRIE

DATE/TIME OF TEST: 10/05/2018 01:50:24

---

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.169	01:45
Breath Volume	2.566	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.151	01:48
Breath Volume	1.769	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

---

Data Download Date/Time: 10/17/2018 15:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ORTUNO,CARLOS  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 09/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/06/2018 05:24:38

---

Diagnostics Check	OK	05:18
Air Blank	0.000	05:19
Control Test	0.080	05:19
Air Blank	0.000	05:19
Subject Sample # 1	0.188	05:20
Breath Volume	2.328	05:20
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.184	05:23
Breath Volume	1.804	05:23
Air Blank	0.000	05:23
Control Test	0.079	05:24
Air Blank	0.000	05:24
Diagnostics Check	OK	05:24

---

Data Download Date/Time: 10/17/2018 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: KEATING,KAYLA  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 09/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 10/07/2018 04:19:43

---

Diagnostics Check		OK	04:13
Air Blank		0.000	04:14
Control Test		0.079	04:14
Air Blank		0.000	04:14
Subject Sample	# 1	0.205	04:15
Breath Volume		1.359	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.211	04:18
Breath Volume		1.468	04:18
Air Blank		0.000	04:18
Control Test		0.079	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

---

Data Download Date/Time: 10/17/2018 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: KELLIHER,MICHAEL  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 09/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 10/09/2018 00:31:26

---

Diagnostics Check		OK	00:17	
Air Blank		0.000	00:17	
Control Test		0.079	00:17	
Air Blank		0.000	00:18	
Subject Sample	# 1	0.086	00:21	Volume Not Met
Breath Volume		0.371	00:21	
Air Blank		0.000	00:22	
Air Blank		0.000	00:24	
Subject Sample	# 2	0.100	00:25	
Breath Volume		1.152	00:25	
Air Blank		0.000	00:26	
Air Blank		0.000	00:28	
Subject Sample	# 3	0.097	00:29	
Breath Volume		1.128	00:29	
Air Blank		0.000	00:30	
Control Test		0.078	00:30	
Air Blank		0.000	00:31	
Diagnostics Check		OK	00:31	

---

Data Download Date/Time: 10/17/2018 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: FITHIAN,JASON  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 09/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 10/11/2018 00:56:50

---

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.000	00:52
Breath Volume	2.734	00:52
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.000	00:55
Breath Volume	2.488	00:55
Air Blank	0.000	00:55
Control Test	0.080	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

---

Data Download Date/Time: 10/17/2018 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: FITHIAN,JASON  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 09/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: SIMMS

DATE/TIME OF TEST: 10/13/2018 00:05:31

---

Diagnostics Check		OK	23:58
Air Blank		0.000	23:59
Control Test		0.080	23:59
Air Blank		0.000	23:59
Subject Sample	# 1	0.204	00:00
Breath Volume		1.265	00:00
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.191	00:04
Breath Volume		1.773	00:04
Air Blank		0.000	00:04
Control Test		0.079	00:04
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

---

Data Download Date/Time: 10/17/2018 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: KELLIHER,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 09/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 10/13/2018 04:56:00

---

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:49	
Control Test	0.080	04:50	
Air Blank	0.000	04:50	
Subject Sample	# 1 0.199	04:51	
Breath Volume	2.039	04:51	
Air Blank	0.000	04:52	
Air Blank	0.000	04:54	
Subject Sample	# 2	04:54	Slope Not Met
Breath Volume	1.382	04:54	
Air Blank	0.000	04:55	
Control Test	0.079	04:55	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:56	

---

Data Download Date/Time: 10/17/2018 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: KERLLIHER,ROBERT  
ARRESTING OFFICER AGENCY: FL Gulf Coast UPD  
LAST AGENCY INSPECTION: 09/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 10/13/2018 05:24:58

---

Diagnostics Check		OK	05:18
Air Blank		0.000	05:19
Control Test		0.080	05:19
Air Blank		0.000	05:20
Subject Sample	# 1	0.192	05:20
Breath Volume		1.484	05:20
Air Blank		0.000	05:21
Air Blank		0.000	05:23
Subject Sample	# 2	0.194	05:23
Breath Volume		1.773	05:23
Air Blank		0.000	05:24
Control Test		0.079	05:24
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

---

Data Download Date/Time: 10/17/2018 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: SAPP,JOSHUA  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 09/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: HANLEY

DATE/TIME OF TEST: 10/13/2018 06:40:00

---

Diagnostics Check	OK	06:33
Air Blank	0.000	06:34
Control Test	0.080	06:34
Air Blank	0.000	06:35
Subject Sample # 1	0.168	06:35
Breath Volume	3.515	06:35
Air Blank	0.000	06:36
Air Blank	0.000	06:38
Subject Sample # 2	0.168	06:38
Breath Volume	3.378	06:38
Air Blank	0.000	06:39
Control Test	0.080	06:39
Air Blank	0.000	06:39
Diagnostics Check	OK	06:40

---

Data Download Date/Time: 10/17/2018 15:48



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LOUGHRAN,KEVIN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 09/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: LOZA AILLAUD

DATE/TIME OF TEST: 10/14/2018 00:30:33

---

Diagnostics Check		OK	00:23
Air Blank		0.000	00:23
Control Test		0.079	00:24
Air Blank		0.000	00:24
Subject Sample	# 1	0.241	00:25
Breath Volume		1.535	00:25
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.238	00:29
Breath Volume		1.558	00:29
Air Blank		0.000	00:29
Control Test		0.080	00:29
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

---

Data Download Date/Time: 10/17/2018 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LOUGHRAN,KEVIN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 09/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: POLK

DATE/TIME OF TEST: 10/14/2018 03:54:14

---

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.079	03:45
Air Blank		0.000	03:46
Subject Sample	# 1	0.246	03:46
Breath Volume		2.007	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.214	03:49
Breath Volume		1.542	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 3	0.231	03:52
Breath Volume		2.066	03:52
Air Blank		0.000	03:53
Control Test		0.079	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

---

Data Download Date/Time: 10/17/2018 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: KELLIHER,MICHAEL  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 09/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: FRANCIQUE

DATE/TIME OF TEST: 10/14/2018 04:30:01

---

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.079	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.219	04:25
Breath Volume		1.203	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:27
Subject Sample	# 2	0.210	04:28
Breath Volume		1.242	04:28
Air Blank		0.000	04:29
Control Test		0.078	04:29
Air Blank		0.000	04:29
Diagnostics Check		OK	04:30

---

Data Download Date/Time: 10/17/2018 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: SAPP,JOSHUA  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 09/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: ADONIS

DATE/TIME OF TEST: 10/14/2018 05:08:28

---

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.079	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.157	05:01
Breath Volume	1.476	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.153	05:06
Breath Volume	1.160	05:06
Air Blank	0.000	05:07
Control Test	0.079	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

---

Data Download Date/Time: 10/17/2018 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: CARNEIRO,JUSTIN  
ARRESTING OFFICER AGENCY: Davie PD  
LAST AGENCY INSPECTION: 09/19/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/16/2018 19:42:50

---

Diagnostics Check	OK	19:36
Air Blank	0.000	19:37
Control Test	0.080	19:37
Air Blank	0.000	19:37
Subject Sample # 1	0.171	19:38
Breath Volume	2.179	19:38
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.171	19:41
Breath Volume	2.343	19:41
Air Blank	0.000	19:41
Control Test	0.080	19:42
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

---

Data Download Date/Time: 10/17/2018 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ROJAS,RANDY  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: NAPPI

DATE/TIME OF TEST: 10/19/2018 01:00:36

---

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.186	00:55
Breath Volume	1.261	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.197	00:59
Breath Volume	1.699	00:59
Air Blank	0.000	00:59
Control Test	0.080	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

---

Data Download Date/Time: 11/14/2018 14:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: STRZALKOWSKI,SEAN  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: HURTADO

DATE/TIME OF TEST: 10/19/2018 03:54:22

---

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:53	
Control Test	0.080	03:53	
Air Blank		03:53	RFI Detect
Air Blank	0.000	03:54	

---

Data Download Date/Time: 11/14/2018 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LOUGHRAN,KEVIN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 10/19/2018 04:39:41

---

Diagnostics Check		OK	04:28	
Air Blank		0.000	04:28	
Control Test		0.080	04:28	
Air Blank		0.000	04:29	
Subject Sample	# 1	0.108	04:32	Volume Not Met
Breath Volume		0.074	04:32	
Air Blank		0.000	04:33	
Air Blank		0.000	04:35	
Subject Sample	# 2	0.107	04:38	Volume Not Met
Breath Volume		0.109	04:38	
Air Blank		0.000	04:38	
Control Test		0.079	04:39	
Air Blank		0.000	04:39	
Diagnostics Check		OK	04:39	

---

Data Download Date/Time: 11/14/2018 14:23



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LOUGHRAN,KEVIN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 10/19/2018 04:56:14

---

Diagnostics Check		OK	04:44	
Air Blank		0.000	04:45	
Control Test		0.079	04:45	
Air Blank		0.000	04:45	
Subject Sample	# 1	0.099	04:49	Volume Not Met
Breath Volume		0.121	04:49	
Air Blank		0.000	04:49	
Air Blank		0.000	04:51	
Subject Sample	# 2	0.086	04:54	Volume Not Met
Breath Volume		0.574	04:54	
Air Blank		0.000	04:55	
Control Test		0.079	04:55	
Air Blank		0.000	04:56	
Diagnostics Check		OK	04:56	

---

Data Download Date/Time: 11/14/2018 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: KELLIHER,MICHAEL  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: CHIRINOS

DATE/TIME OF TEST: 10/19/2018 20:52:33

---

Diagnostics Check	OK	20:46
Air Blank	0.000	20:46
Control Test	0.079	20:47
Air Blank	0.000	20:47
Subject Sample # 1	0.265	20:48
Breath Volume	1.910	20:48
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.272	20:51
Breath Volume	1.996	20:51
Air Blank	0.000	20:51
Control Test	0.080	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

---

Data Download Date/Time: 11/14/2018 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LEE,ANTHONY  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 10/20/2018 01:04:29

---

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.219	00:59
Breath Volume	2.148	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.221	01:03
Breath Volume	1.878	01:03
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

---

Data Download Date/Time: 11/14/2018 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: TRABUE,THOMAS  
ARRESTING OFFICER AGENCY: Pembroke Pines PD  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: HEITZMANN

DATE/TIME OF TEST: 10/20/2018 02:20:28

---

Diagnostics Check		OK	02:13
Air Blank		0.000	02:14
Control Test		0.080	02:14
Air Blank		0.000	02:15
Subject Sample	# 1	0.268	02:15
Breath Volume		3.203	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.267	02:18
Breath Volume		3.339	02:18
Air Blank		0.000	02:19
Control Test		0.079	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

---

Data Download Date/Time: 11/14/2018 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LOUGHRAN,KEVIN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: PETGRAVE

DATE/TIME OF TEST: 10/20/2018 04:59:31

---

Diagnostics Check		OK	04:53
Air Blank		0.000	04:53
Control Test		0.079	04:54
Air Blank		0.000	04:54
Subject Sample	# 1	0.168	04:55
Breath Volume		2.226	04:55
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.169	04:58
Breath Volume		2.445	04:58
Air Blank		0.000	04:58
Control Test		0.079	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

---

Data Download Date/Time: 11/14/2018 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: KEATING,KAYLA  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: LOREK

DATE/TIME OF TEST: 10/21/2018 00:03:14

---

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.079	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.118	23:58
Breath Volume	2.519	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.111	00:01
Breath Volume	2.058	00:01
Air Blank	0.000	00:02
Control Test	0.079	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

---

Data Download Date/Time: 11/14/2018 14:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: PACK,JEFFERY  
ARRESTING OFFICER AGENCY: Miramar PD  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: GRIFFITHS

DATE/TIME OF TEST: 10/21/2018 20:14:53

---

Diagnostics Check	OK	20:08
Air Blank	0.000	20:08
Control Test	0.079	20:08
Air Blank	0.000	20:09
Subject Sample # 1	0.174	20:10
Breath Volume	1.742	20:10
Air Blank	0.000	20:10
Air Blank	0.000	20:12
Subject Sample # 2	0.179	20:13
Breath Volume	1.808	20:13
Air Blank	0.000	20:13
Control Test	0.079	20:14
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

---

Data Download Date/Time: 11/14/2018 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: STRZALKOWSKI,SEAN  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: TRENT

DATE/TIME OF TEST: 10/21/2018 21:17:01

---

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.080	21:11
Air Blank	0.000	21:11
Subject Sample # 1	0.282	21:12
Breath Volume	1.363	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:14
Subject Sample # 2	0.286	21:15
Breath Volume	1.593	21:15
Air Blank	0.000	21:16
Control Test	0.078	21:16
Air Blank	0.000	21:16
Diagnostics Check	OK	21:17

---

Data Download Date/Time: 11/14/2018 14:23



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LOUGHRAN,KEVIN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: TRIMM

DATE/TIME OF TEST: 10/23/2018 01:56:06

---

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.000	01:51
Breath Volume	2.050	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.000	01:54
Breath Volume	1.902	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

---

Data Download Date/Time: 11/14/2018 14:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: O'DOR,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: AZAN

DATE/TIME OF TEST: 10/24/2018 23:19:37

---

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.079	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.114	23:15
Breath Volume	2.312	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.113	23:18
Breath Volume	2.179	23:18
Air Blank	0.000	23:18
Control Test	0.079	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

---

Data Download Date/Time: 11/14/2018 14:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: KELLIHER,MICHAEL  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 10/27/2018 01:36:29

---

Diagnostics Check		OK	01:25
Air Blank		0.000	01:26
Control Test		0.080	01:26
Air Blank		0.000	01:27
Subject Sample	# 1	0.223	01:27
Breath Volume		1.546	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.247	01:32
Breath Volume		2.800	01:32
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 3	0.234	01:35
Breath Volume		2.144	01:35
Air Blank		0.000	01:35
Control Test		0.079	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

---

Data Download Date/Time: 11/14/2018 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HARRIS,MILTON  
ARRESTING OFFICER AGENCY: Pembroke Pines PD  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: LAMELAS

DATE/TIME OF TEST: 10/27/2018 02:02:03

---

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.079	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.212	01:57
Breath Volume		2.167	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.217	02:00
Breath Volume		2.457	02:00
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:02

---

Data Download Date/Time: 11/14/2018 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: KEATING,KAYLA  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: FRANCISCO PEREZ

DATE/TIME OF TEST: 10/27/2018 02:13:49

---

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.079	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.208	02:09
Breath Volume	1.558	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:11
Subject Sample # 2	0.215	02:12
Breath Volume	2.839	02:12
Air Blank	0.000	02:12
Control Test	0.079	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

---

Data Download Date/Time: 11/14/2018 14:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: OLEA,ELIEZER  
ARRESTING OFFICER AGENCY: Pembroke Pines PD  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 10/28/2018 03:47:13

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.165	03:42
Breath Volume	2.187	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.165	03:45
Breath Volume	2.093	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

---

Data Download Date/Time: 11/14/2018 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: SAPP,JOSHUA  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: CHAVARRIA

DATE/TIME OF TEST: 10/31/2018 04:11:36

---

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.161	04:06
Breath Volume	2.097	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.169	04:10
Breath Volume	2.957	04:10
Air Blank	0.000	04:10
Control Test	0.079	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

---

Data Download Date/Time: 11/14/2018 14:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HURT,JONATHAN,R  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: BINGHAM,KYLE  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 10/02/2018 21:45:03

---

Diagnostics Check	OK	21:38
Air Blank	0.000	21:38
Control Test	0.080	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.000	21:40
Breath Volume	4.355	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.000	21:43
Breath Volume	4.433	21:43
Air Blank	0.000	21:44
Control Test	0.078	21:44
Air Blank	0.000	21:44
Diagnostics Check	OK	21:45

---

Data Download Date/Time: 10/16/2018 20:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: NEGERSMITH,DANIEL  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 10/05/2018 00:34:48

---

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.079	00:27
Air Blank		0.000	00:27
Subject Sample	# 1	0.246	00:29
Breath Volume		1.273	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:31
Subject Sample	# 2	0.238	00:33
Breath Volume		1.265	00:33
Air Blank		0.000	00:33
Control Test		0.078	00:34
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

---

Data Download Date/Time: 10/16/2018 20:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: SENTER,ERIC  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: ALVAREZ TORRECILLA

DATE/TIME OF TEST: 10/06/2018 01:26:41

---

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.079	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.109	01:21
Breath Volume	1.511	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.105	01:24
Breath Volume	2.121	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

---

Data Download Date/Time: 10/16/2018 20:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HURT,JONATHAN,R  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: HURT,JONATHAN  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/06/2018 03:04:02

---

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.000	02:59
Breath Volume	1.367	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.000	03:02
Breath Volume	1.882	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:04

---

Data Download Date/Time: 10/16/2018 20:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VENEZIA,STEVEN,L  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: RODRIGUES,RICHARD  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: VILLANOVA

DATE/TIME OF TEST: 10/07/2018 14:53:12

---

Diagnostics Check	OK	14:46
Air Blank	0.000	14:46
Control Test	0.079	14:46
Air Blank	0.000	14:47
Subject Sample # 1	0.295	14:48
Breath Volume	2.546	14:48
Air Blank	0.000	14:48
Air Blank	0.000	14:50
Subject Sample # 2	0.292	14:51
Breath Volume	2.109	14:51
Air Blank	0.000	14:52
Control Test	0.079	14:52
Air Blank	0.000	14:53
Diagnostics Check	OK	14:53

---

Data Download Date/Time: 10/16/2018 20:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HASTIGS,KENNETH,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: HASTINGS,KENNETH  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: ROEGNER

DATE/TIME OF TEST: 10/12/2018 02:21:32

---

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:19	
Control Test	0.079	02:19	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:20	
Control Test	0.080	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

---

Data Download Date/Time: 10/16/2018 20:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTIA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: NGO

DATE/TIME OF TEST: 10/13/2018 03:49:34

---

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.079	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.243	03:44
Breath Volume		1.929	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.251	03:47
Breath Volume		2.375	03:47
Air Blank		0.000	03:48
Control Test		0.077	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

---

Data Download Date/Time: 10/16/2018 20:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OBRIEN,DANIEL,P  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: BEAVER,MICHAEL  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 10/14/2018 03:35:12

---

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.079	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:28	No Sample Provided
Breath Volume	0.000	03:28	
Air Blank	0.000	03:28	
Air Blank	0.000	03:30	
Subject Sample	# 2	03:33	No Sample Provided
Breath Volume	0.000	03:33	
Air Blank	0.000	03:34	
Control Test	0.079	03:34	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

---

Data Download Date/Time: 10/16/2018 20:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'BRIEN,DANIEL,P  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: BEAVER,MICHAEL  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 10/14/2018 03:54:47

---

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.079	03:49
Air Blank		0.000	03:49
Subject Sample	# 1	0.164	03:50
Breath Volume		3.175	03:50
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.166	03:53
Breath Volume		2.621	03:53
Air Blank		0.000	03:53
Control Test		0.078	03:54
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

---

Data Download Date/Time: 10/16/2018 20:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EDMONDS,RICHARD,J  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: EDMONDS,RICHARD  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MCSEVNEY

DATE/TIME OF TEST: 10/15/2018 04:51:50

---

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.078	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.172	04:47
Breath Volume	1.941	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:49
Subject Sample # 2	0.170	04:50
Breath Volume	1.578	04:50
Air Blank	0.000	04:50
Control Test	0.077	04:51
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

---

Data Download Date/Time: 10/16/2018 20:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOXIE,BRIAN,P  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: KURINSKY,DAVID  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: HORNER

DATE/TIME OF TEST: 10/15/2018 22:11:04

---

Diagnostics Check	OK	22:04
Air Blank	0.000	22:05
Control Test	0.080	22:05
Air Blank	0.000	22:06
Subject Sample # 1	0.286	22:06
Breath Volume	2.183	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.288	22:09
Breath Volume	2.523	22:09
Air Blank	0.000	22:10
Control Test	0.078	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:11

---

Data Download Date/Time: 10/16/2018 20:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NUGENT,DAVID,A  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: NUGENT,DAVID  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: KRAUSS

DATE/TIME OF TEST: 10/16/2018 19:35:42

---

Diagnostics Check	OK	19:29
Air Blank	0.000	19:29
Control Test	0.079	19:29
Air Blank	0.000	19:30
Subject Sample # 1	0.176	19:31
Breath Volume	1.378	19:31
Air Blank	0.000	19:31
Air Blank	0.000	19:33
Subject Sample # 2	0.177	19:34
Breath Volume	1.824	19:34
Air Blank	0.000	19:34
Control Test	0.079	19:35
Air Blank	0.000	19:35
Diagnostics Check	OK	19:35

---

Data Download Date/Time: 10/16/2018 20:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: NEGERSMITH,DANIEL  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 10/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: CRIPE

DATE/TIME OF TEST: 10/16/2018 21:24:50

---

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.079	21:18
Air Blank	0.000	21:18
Subject Sample # 1	0.134	21:19
Breath Volume	2.589	21:19
Air Blank	0.000	21:20
Air Blank	0.000	21:22
Subject Sample # 2	0.127	21:23
Breath Volume	2.449	21:23
Air Blank	0.000	21:23
Control Test	0.077	21:24
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

---

Data Download Date/Time: 11/08/2018 19:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: BAER

DATE/TIME OF TEST: 10/18/2018 03:52:44

---

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.139	03:48
Breath Volume	2.250	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:50
Subject Sample # 2	0.131	03:51
Breath Volume	1.421	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

---

Data Download Date/Time: 11/08/2018 19:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HURT,JONATHAN,R  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: REED,CHAD  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 10/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 10/19/2018 02:46:08

---

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.000	02:41
Breath Volume	2.660	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.000	02:44
Breath Volume	2.957	02:44
Air Blank	0.000	02:45
Control Test	0.078	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

---

Data Download Date/Time: 11/08/2018 19:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: ORNELAS

DATE/TIME OF TEST: 10/19/2018 03:27:09

---

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:15	
Control Test	0.078	03:15	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:19	No Sample Provided
Breath Volume	0.000	03:19	
Air Blank	0.000	03:19	
Air Blank	0.000	03:21	
Subject Sample	# 2	0.205	03:22
Breath Volume	2.632	03:22	
Air Blank	0.000	03:23	
Air Blank	0.000	03:24	
Subject Sample	# 3	0.185	03:25
Breath Volume	2.203	03:25	
Air Blank	0.000	03:26	
Control Test	0.079	03:26	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

---

Data Download Date/Time: 11/08/2018 19:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: ORNELAS

DATE/TIME OF TEST: 10/19/2018 03:38:53

---

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.079	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.199	03:34
Breath Volume	3.042	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.186	03:37
Breath Volume	1.515	03:37
Air Blank	0.000	03:37
Control Test	0.079	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

---

Data Download Date/Time: 11/08/2018 19:00



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOXIE,BRIAN,P  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: MALANDRO,STEVE  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: DESILETS

DATE/TIME OF TEST: 10/19/2018 17:55:08

---

Diagnostics Check		OK	17:48
Air Blank		0.000	17:49
Control Test		0.079	17:49
Air Blank		0.000	17:50
Subject Sample	# 1	0.150	17:50
Breath Volume		2.203	17:50
Air Blank		0.000	17:51
Air Blank		0.000	17:53
Subject Sample	# 2	0.151	17:53
Breath Volume		2.238	17:53
Air Blank		0.000	17:54
Control Test		0.078	17:54
Air Blank		0.000	17:55
Diagnostics Check		OK	17:55

---

Data Download Date/Time: 11/08/2018 19:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HURT,JONATHAN,R  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: HURT,JONATHAN  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 10/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: NOPAL JAHUEY

DATE/TIME OF TEST: 10/19/2018 23:32:17

---

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.079	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.198	23:27
Breath Volume	2.562	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.195	23:30
Breath Volume	2.308	23:30
Air Blank	0.000	23:31
Control Test	0.079	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

---

Data Download Date/Time: 11/08/2018 19:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: EARLE

DATE/TIME OF TEST: 10/20/2018 00:43:02

---

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.079	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.114	00:38
Breath Volume	1.121	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.113	00:41
Breath Volume	1.488	00:41
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:43

---

Data Download Date/Time: 11/08/2018 19:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 10/21/2018 02:46:45

---

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.079	02:44	
Air Blank	0.000	02:44	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:45	
Control Test	0.079	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

---

Data Download Date/Time: 11/08/2018 19:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OBRIEN,DANIEL,P  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: OWENS,R  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: SHIRLEY

DATE/TIME OF TEST: 10/22/2018 00:45:27

---

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:41	
Control Test	0.080	00:41	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:44	Subject Test Refused
Breath Volume	0.000	00:44	
Air Blank	0.000	00:44	
Control Test	0.080	00:44	
Air Blank	0.000	00:45	
Diagnostics Check	OK	00:45	

---

Data Download Date/Time: 11/08/2018 19:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HASTINGS,KENNETH,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: HASTINGS,KENNETH  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: JEAN-BAPTISTE

DATE/TIME OF TEST: 10/24/2018 20:20:42

---

Diagnostics Check	OK	20:15	
Air Blank	0.000	20:15	
Control Test	0.079	20:16	
Air Blank	0.000	20:16	
Subject Sample	# 1	20:19	Subject Test Refused
Breath Volume	0.000	20:19	
Air Blank	0.000	20:19	
Control Test	0.080	20:20	
Air Blank	0.000	20:20	
Diagnostics Check	OK	20:20	

---

Data Download Date/Time: 11/08/2018 19:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HASTINGS,KENNETH,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: HASTINGS,KENNETH  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: WELSH

DATE/TIME OF TEST: 10/25/2018 05:00:22

---

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.079	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.139	04:55
Breath Volume	1.859	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:57
Subject Sample # 2	0.147	04:58
Breath Volume	1.628	04:58
Air Blank	0.000	04:59
Control Test	0.079	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

---

Data Download Date/Time: 11/08/2018 19:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: MEDINA,JOSE  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 10/16/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A2  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: NARVAEZ NARVAEZ

DATE/TIME OF TEST: 10/27/2018 04:10:01

---

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.079	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.233	04:05
Breath Volume		1.750	04:05
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.229	04:08
Breath Volume		1.863	04:08
Air Blank		0.000	04:09
Control Test		0.079	04:09
Air Blank		0.000	04:09
Diagnostics Check		OK	04:10

---

Data Download Date/Time: 11/08/2018 19:00



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNS,D,L  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: EANES,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: ESCHEN

DATE/TIME OF TEST: 10/19/2018 04:02:20

---

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.159	03:57
Breath Volume	1.566	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.164	04:00
Breath Volume	1.375	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

---

Data Download Date/Time: 11/06/2018 12:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MARRERO, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: GORE

DATE/TIME OF TEST: 10/20/2018 02:33:19

---

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.081	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.187	02:29
Breath Volume	2.035	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.180	02:31
Breath Volume	1.332	02:31
Air Blank	0.000	02:32
Control Test	0.081	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

---

Data Download Date/Time: 11/06/2018 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: TORRES, W  
ARRESTING OFFICER AGENCY: Neptune Beach PD  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: HATTEM

DATE/TIME OF TEST: 10/20/2018 04:08:21

---

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.080	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:06	Subject Test Refused
Breath Volume	0.000	04:06	
Air Blank	0.000	04:07	
Control Test	0.080	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

---

Data Download Date/Time: 11/06/2018 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KING, W  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: MESQUITA

DATE/TIME OF TEST: 10/20/2018 05:43:05

---

Diagnostics Check	OK	05:36
Air Blank	0.000	05:36
Control Test	0.081	05:37
Air Blank	0.000	05:37
Subject Sample # 1	0.068	05:38
Breath Volume	2.175	05:38
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample # 2	0.071	05:41
Breath Volume	1.929	05:41
Air Blank	0.000	05:42
Control Test	0.080	05:42
Air Blank	0.000	05:42
Diagnostics Check	OK	05:43

---

Data Download Date/Time: 11/06/2018 12:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SPLATT,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: INACIO DOS SANTOS

DATE/TIME OF TEST: 10/21/2018 00:42:42

---

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.197	00:38
Breath Volume	2.507	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.193	00:41
Breath Volume	3.207	00:41
Air Blank	0.000	00:41
Control Test	0.079	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

---

Data Download Date/Time: 11/06/2018 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BOWERS,D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 10/21/2018 06:14:15

---

Diagnostics Check	OK	06:11	
Air Blank	0.000	06:11	
Control Test	0.081	06:12	
Air Blank	0.000	06:12	
Subject Sample	# 1	06:12	Subject Test Refused
Breath Volume	0.000	06:12	
Air Blank	0.000	06:13	
Control Test	0.081	06:13	
Air Blank	0.000	06:14	
Diagnostics Check	OK	06:14	

---

Data Download Date/Time: 11/06/2018 12:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BROWN, T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: RUBINO

DATE/TIME OF TEST: 10/22/2018 00:45:05

---

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.082	00:39
Air Blank	0.000	00:40
Subject Sample # 1	0.000	00:40
Breath Volume	2.371	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.000	00:43
Breath Volume	1.121	00:43
Air Blank	0.000	00:44
Control Test	0.081	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

---

Data Download Date/Time: 11/06/2018 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNS,D,L  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MARRERO,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: RODRIGUEZ FUENTES

DATE/TIME OF TEST: 10/22/2018 02:45:04

---

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.081	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.133	02:40
Breath Volume	4.160	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:42
Subject Sample # 2	0.131	02:43
Breath Volume	4.136	02:43
Air Blank	0.000	02:44
Control Test	0.082	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:45

---

Data Download Date/Time: 11/06/2018 12:03



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BYRNE, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 10/23/2018 03:05:44

---

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.082	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.200	03:01
Breath Volume	2.800	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.201	03:04
Breath Volume	3.292	03:04
Air Blank	0.000	03:04
Control Test	0.080	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

---

Data Download Date/Time: 11/06/2018 12:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AUGUSTINE,G,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MOELLER,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: TRETICK

DATE/TIME OF TEST: 10/26/2018 01:11:23

---

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.082	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.159	01:06
Breath Volume	2.234	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.164	01:09
Breath Volume	2.121	01:09
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

---

Data Download Date/Time: 11/06/2018 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: SAENZ

DATE/TIME OF TEST: 10/26/2018 02:55:52

---

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:53	
Control Test	0.081	02:53	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:54	Subject Test Refused
Breath Volume	0.000	02:54	
Air Blank	0.000	02:54	
Control Test	0.081	02:55	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

---

Data Download Date/Time: 11/06/2018 12:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AUGUSTINE,G,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNS,B  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: BELSCHES

DATE/TIME OF TEST: 10/26/2018 05:37:44

---

Diagnostics Check	OK	05:30
Air Blank	0.000	05:31
Control Test	0.081	05:31
Air Blank	0.000	05:31
Subject Sample # 1	0.160	05:33
Breath Volume	1.894	05:33
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.151	05:36
Breath Volume	2.066	05:36
Air Blank	0.000	05:36
Control Test	0.080	05:37
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

---

Data Download Date/Time: 11/06/2018 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PAUL,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: GREAVES

DATE/TIME OF TEST: 10/27/2018 04:22:27

---

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:20	
Control Test	0.082	04:20	
Air Blank	0.000	04:20	
Subject Sample	# 1	04:21	Subject Test Refused
Breath Volume	0.000	04:21	
Air Blank	0.000	04:21	
Control Test	0.082	04:21	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:22	

---

Data Download Date/Time: 11/06/2018 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARIDA,K,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: OVERHOLSER,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 10/27/2018 19:48:02

---

Diagnostics Check	OK	19:45	
Air Blank	0.000	19:45	
Control Test	0.082	19:45	
Air Blank	0.000	19:46	
Subject Sample	# 1	19:46	Subject Test Refused
Breath Volume	0.000	19:46	
Air Blank	0.000	19:47	
Control Test	0.082	19:47	
Air Blank	0.000	19:47	
Diagnostics Check	OK	19:48	

---

Data Download Date/Time: 11/06/2018 12:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARIDA,K,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SPLATT,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: ANGELI

DATE/TIME OF TEST: 10/27/2018 22:04:19

---

Diagnostics Check	OK	21:57
Air Blank	0.000	21:58
Control Test	0.082	21:58
Air Blank	0.000	21:58
Subject Sample # 1	0.100	21:59
Breath Volume	4.199	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.099	22:02
Breath Volume	4.039	22:02
Air Blank	0.000	22:03
Control Test	0.082	22:03
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

---

Data Download Date/Time: 11/06/2018 12:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HARD,W  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: TYSON

DATE/TIME OF TEST: 10/28/2018 03:39:14

---

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.082	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.098	03:34
Breath Volume	1.578	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.096	03:37
Breath Volume	1.136	03:37
Air Blank	0.000	03:38
Control Test	0.081	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

---

Data Download Date/Time: 11/06/2018 12:03



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLS,R,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CAMACHO,K  
ARRESTING OFFICER AGENCY: Neptune Beach PD  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: TORMEY

DATE/TIME OF TEST: 10/28/2018 22:32:42

---

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.082	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.167	22:27
Breath Volume	1.570	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.161	22:31
Breath Volume	1.636	22:31
Air Blank	0.000	22:31
Control Test	0.082	22:32
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

---

Data Download Date/Time: 11/06/2018 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNS,D,L  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BYRNE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MONTIEL GARCIA

DATE/TIME OF TEST: 10/30/2018 23:47:28

---

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:44	
Control Test	0.082	23:45	
Air Blank	0.000	23:45	
Subject Sample	# 1	23:46	Subject Test Refused
Breath Volume	0.000	23:46	
Air Blank	0.000	23:46	
Control Test	0.083	23:46	
Air Blank	0.000	23:47	
Diagnostics Check	OK	23:47	

---

Data Download Date/Time: 11/06/2018 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JENSEN, D  
ARRESTING OFFICER AGENCY: Neptune Beach PD  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: RATIANARIVO

DATE/TIME OF TEST: 10/31/2018 05:09:41

---

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.083	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.145	05:04
Breath Volume		2.027	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.150	05:08
Breath Volume		2.386	05:08
Air Blank		0.000	05:08
Control Test		0.082	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

---

Data Download Date/Time: 11/06/2018 12:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRIOLAT,CASEY,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HEALY,SETH  
ARRESTING OFFICER AGENCY: Zephyrhills PD  
LAST AGENCY INSPECTION: 09/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A1  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 10/02/2018 03:05:43

---

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:54	
Control Test	0.079	02:55	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:58	No Sample Provided
Breath Volume	0.000	02:58	
Air Blank	0.000	02:59	
Air Blank	0.000	03:01	
Subject Sample	# 2	03:04	No Sample Provided
Breath Volume	0.000	03:04	
Air Blank	0.000	03:04	
Control Test	0.079	03:05	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

---

Data Download Date/Time: 11/02/2018 10:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L  
BREATH TEST OPERATOR AGENCY: Zephyrhills PD  
ARRESTING OFFICER: BAILEY,ANISA  
ARRESTING OFFICER AGENCY: Zephyrhills PD  
LAST AGENCY INSPECTION: 09/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A1  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 10/06/2018 00:19:59

---

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.078	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.323	00:14
Breath Volume	2.105	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:16
Subject Sample # 2	0.310	00:18
Breath Volume	3.058	00:18
Air Blank	0.000	00:19
Control Test	0.078	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

---

Data Download Date/Time: 11/02/2018 10:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,WILLIAM,R  
BREATH TEST OPERATOR AGENCY: Zephyrhills PD  
ARRESTING OFFICER: MALDONADO,SOLOMON  
ARRESTING OFFICER AGENCY: Zephyrhills PD  
LAST AGENCY INSPECTION: 09/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A1  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:48

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 10/19/2018 07:18:33

---

Diagnostics Check	OK	07:11
Air Blank	0.000	07:12
Control Test	0.079	07:12
Air Blank	0.000	07:13
Subject Sample # 1	0.024	07:14
Breath Volume	4.093	07:14
Air Blank	0.000	07:14
Air Blank	0.000	07:16
Subject Sample # 2	0.023	07:17
Breath Volume	4.273	07:17
Air Blank	0.000	07:17
Control Test	0.078	07:17
Air Blank	0.000	07:18
Diagnostics Check	OK	07:18

---

Data Download Date/Time: 11/02/2018 10:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: WALDO,AARON  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: SPAYSKY

DATE/TIME OF TEST: 10/06/2018 02:33:14

---

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.080	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.185	02:28
Breath Volume		1.601	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.169	02:31
Breath Volume		1.937	02:31
Air Blank		0.000	02:32
Control Test		0.078	02:32
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

---

Data Download Date/Time: 10/15/2018 08:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDO,AARON,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ZOLLOTA,VLADIMIR  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 10/07/2018 22:46:47

---

Diagnostics Check		OK	22:40
Air Blank		0.000	22:40
Control Test		0.079	22:41
Air Blank		0.000	22:41
Subject Sample	# 1	0.234	22:42
Breath Volume		2.773	22:42
Air Blank		0.000	22:43
Air Blank		0.000	22:44
Subject Sample	# 2	0.230	22:45
Breath Volume		2.449	22:45
Air Blank		0.000	22:45
Control Test		0.078	22:46
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

---

Data Download Date/Time: 10/15/2018 08:49



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: HOBART,CLAWTON  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: GILES

DATE/TIME OF TEST: 10/07/2018 23:54:27

---

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.078	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.222	23:45
Breath Volume	1.609	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.200	23:49
Breath Volume	2.445	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 3	0.207	23:52
Breath Volume	1.429	23:52
Air Blank	0.000	23:53
Control Test	0.076	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

---

Data Download Date/Time: 10/15/2018 08:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDO,AARON,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ZOLLOTA,VLADIMIR  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 10/10/2018 03:02:37

---

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:54	
Control Test	0.078	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:58	No Sample Provided
Breath Volume	0.000	02:58	
Air Blank	0.000	02:58	
Air Blank	0.000	03:00	
Subject Sample	# 2	03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:01	
Control Test	0.079	03:02	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

---

Data Download Date/Time: 10/15/2018 08:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: CRUM,ZACKARY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: POLK

DATE/TIME OF TEST: 10/28/2018 01:43:17

---

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.250	01:38
Breath Volume	2.988	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:40
Subject Sample # 2	0.252	01:41
Breath Volume	3.035	01:41
Air Blank	0.000	01:42
Control Test	0.079	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

---

Data Download Date/Time: 11/20/2018 10:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: KARAYIANES,MICHAEL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/28/2018 03:48:36

---

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.063	03:43
Breath Volume	1.507	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.063	03:47
Breath Volume	1.480	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

---

Data Download Date/Time: 11/20/2018 10:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUM,ZACKARY,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ZOLLOTA,V  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: BARGMAN

DATE/TIME OF TEST: 10/31/2018 03:44:45

---

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.106	03:40
Breath Volume	1.164	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.117	03:43
Breath Volume	2.136	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

---

Data Download Date/Time: 11/20/2018 10:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: KARAYIANES,MICHAEL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/15/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: BRANDT

DATE/TIME OF TEST: 10/31/2018 23:03:22

---

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:00	
Control Test	0.080	23:01	
Air Blank	0.000	23:01	
Subject Sample	# 1	23:02	Subject Test Refused
Breath Volume	0.000	23:02	
Air Blank	0.000	23:02	
Control Test	0.080	23:02	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

---

Data Download Date/Time: 11/20/2018 10:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALAMEDA,JORGE,L  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: MALANDRO,STEVE  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 09/06/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A3  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/01/2018 00:38:07

---

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:32	
Control Test	0.080	00:32	
Air Blank	0.000	00:33	
Subject Sample # 1	0.209	00:33	
Breath Volume	3.546	00:33	
Air Blank	0.000	00:34	
Air Blank	0.000	00:36	
Subject Sample # 2	0.206	00:36	
Breath Volume	3.597	00:36	
Air Blank	0.000	00:37	
Control Test	0.090	00:37	Interferent Detect
Air Blank	0.000	00:38	

---

Data Download Date/Time: 10/03/2018 16:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HASTINGS,KENNETH,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: HASTINGS,KENNETH  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/03/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A3  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: DIPAOLO V

DATE/TIME OF TEST: 10/04/2018 03:09:01

---

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.181	03:04
Breath Volume	3.066	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.181	03:07
Breath Volume	3.320	03:07
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:09

---

Data Download Date/Time: 10/15/2018 07:57



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HASTINGS,KENNETH,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: HASTINGS,KENNETH  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/03/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A3  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: PENNIMAN

DATE/TIME OF TEST: 10/05/2018 23:35:08

---

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.000	23:30
Breath Volume	3.550	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.000	23:33
Breath Volume	3.113	23:33
Air Blank	0.000	23:34
Control Test	0.079	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

---

Data Download Date/Time: 10/15/2018 07:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/03/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A3  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: HOLDERMAN

DATE/TIME OF TEST: 10/06/2018 03:54:18

---

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.193	03:49
Breath Volume	3.546	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.191	03:52
Breath Volume	3.941	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

---

Data Download Date/Time: 10/15/2018 07:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER, MATTHEW, D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: GOSIER, MATTHEW  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/03/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A3  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: BARGAR

DATE/TIME OF TEST: 10/06/2018 04:23:37

---

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.080	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.225	04:19
Breath Volume		2.488	04:19
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.234	04:22
Breath Volume		1.371	04:22
Air Blank		0.000	04:22
Control Test		0.079	04:23
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

---

Data Download Date/Time: 10/15/2018 07:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICE,JOHN,B  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: NORRIS,DYLAN  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/03/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A3  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: HALLETT

DATE/TIME OF TEST: 10/06/2018 16:39:08

---

Diagnostics Check	OK	16:32
Air Blank	0.000	16:32
Control Test	0.080	16:33
Air Blank	0.000	16:33
Subject Sample # 1	0.230	16:34
Breath Volume	1.761	16:34
Air Blank	0.000	16:34
Air Blank	0.000	16:36
Subject Sample # 2	0.217	16:37
Breath Volume	1.308	16:37
Air Blank	0.000	16:38
Control Test	0.080	16:38
Air Blank	0.000	16:39
Diagnostics Check	OK	16:39

---

Data Download Date/Time: 10/15/2018 07:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 10/03/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 02:22

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 10/07/2018 03:01:09

---

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.080	02:58	
Air Blank	0.000	02:58	
Subject Sample	# 1	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	03:00	
Control Test	0.080	03:00	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

---

Data Download Date/Time: 10/15/2018 07:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/03/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A3  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 10/07/2018 05:26:40

---

Diagnostics Check		OK	05:12	
Air Blank		0.000	05:12	
Control Test		0.079	05:13	
Air Blank		0.000	05:13	
Subject Sample	# 1	0.105	05:16	Volume Not Met
Breath Volume		0.648	05:16	
Air Blank		0.000	05:17	
Air Blank		0.000	05:19	
Subject Sample	# 2	0.102	05:19	
Breath Volume		1.562	05:19	
Air Blank		0.000	05:20	
Air Blank		0.000	05:22	
Subject Sample	# 3	0.101	05:25	
Breath Volume		1.457	05:25	
Air Blank		0.000	05:25	
Control Test		0.079	05:26	
Air Blank		0.000	05:26	
Diagnostics Check		OK	05:26	

---

Data Download Date/Time: 10/15/2018 07:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER, MATTHEW, D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: TAYLOR, COURTENAY

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 10/03/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A3

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 21:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: HUCKNO JR

DATE/TIME OF TEST: 10/07/2018 21:28:34

---

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:26	
Control Test	0.081	21:26	
Air Blank	0.000	21:26	
Subject Sample	# 1	21:27	Subject Test Refused
Breath Volume	0.000	21:27	
Air Blank	0.000	21:27	
Control Test	0.080	21:27	
Air Blank	0.000	21:28	
Diagnostics Check	OK	21:28	

---

Data Download Date/Time: 10/15/2018 07:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COOK,ROBERT,C  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: KWIATKOWSKA,SYLWIA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/03/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A3  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: CONKLIN

DATE/TIME OF TEST: 10/08/2018 18:19:27

---

Diagnostics Check	OK	18:16	
Air Blank	0.000	18:16	
Control Test	0.079	18:16	
Air Blank	0.000	18:17	
Subject Sample	# 1	18:18	Subject Test Refused
Breath Volume	0.000	18:18	
Air Blank	0.000	18:18	
Control Test	0.080	18:18	
Air Blank	0.000	18:19	
Diagnostics Check	OK	18:19	

---

Data Download Date/Time: 10/15/2018 07:58



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COOK,ROBERT,C  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: COOK,ROBERT  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/03/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A3  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 09:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: FERRER

DATE/TIME OF TEST: 10/09/2018 10:41:38

---

Diagnostics Check	OK	10:35
Air Blank	0.000	10:36
Control Test	0.081	10:36
Air Blank	0.000	10:37
Subject Sample # 1	0.000	10:37
Breath Volume	1.777	10:37
Air Blank	0.000	10:37
Air Blank	0.000	10:39
Subject Sample # 2	0.000	10:40
Breath Volume	1.484	10:40
Air Blank	0.000	10:40
Control Test	0.080	10:41
Air Blank	0.000	10:41
Diagnostics Check	OK	10:41

---

Data Download Date/Time: 10/15/2018 07:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NIXON,BARRY,  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: HAMMELMAM,TYLER  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A2  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: KRAUSS

DATE/TIME OF TEST: 10/26/2018 03:57:15

---

Diagnostics Check	OK	03:56	
Air Blank		03:56	Ambient Fail
Air Blank		03:57	Purge Fail

---

Data Download Date/Time: 11/26/2018 19:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NIXON,BARRY,  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: HAMMELMAN,TYLER  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17817080A2  
DRY GAS EXPIRY DATE: 08/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: KRAUSS

DATE/TIME OF TEST: 10/26/2018 04:00:18

---

Diagnostics Check	OK	03:58	
Air Blank		03:59	Ambient Fail
Air Blank	0.000	04:00	

---

Data Download Date/Time: 11/26/2018 19:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D  
BREATH TEST OPERATOR AGENCY: Plantation PD  
ARRESTING OFFICER: FARQUHARSON,KEVIN  
ARRESTING OFFICER AGENCY: Plantation PD  
LAST AGENCY INSPECTION: 09/20/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 10/06/2018 03:52:49

---

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.130	03:48
Breath Volume	1.832	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.127	03:51
Breath Volume	1.375	03:51
Air Blank	0.000	03:51
Control Test	0.078	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

---

Data Download Date/Time: 10/27/2018 14:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROSKE, OFC J,  
BREATH TEST OPERATOR AGENCY: Plantation PD  
ARRESTING OFFICER: GUE, OFC R  
ARRESTING OFFICER AGENCY: Plantation PD  
LAST AGENCY INSPECTION: 09/20/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: LOCKE

DATE/TIME OF TEST: 10/06/2018 23:46:18

---

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.079	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.058	23:42
Breath Volume		1.726	23:42
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.057	23:44
Breath Volume		1.597	23:44
Air Blank		0.000	23:45
Control Test		0.078	23:45
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

---

Data Download Date/Time: 10/27/2018 14:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOSTON,D,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BROWN,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: VARELA-SALAZAR

DATE/TIME OF TEST: 10/01/2018 01:15:25

---

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:12	
Control Test	0.081	01:13	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:14	Subject Test Refused
Breath Volume	0.000	01:14	
Air Blank	0.000	01:14	
Control Test	0.081	01:14	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

---

Data Download Date/Time: 10/17/2018 15:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOSTON,D,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BRYANT,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/01/2018 04:08:33

---

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.081	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.177	04:03
Breath Volume	1.312	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.195	04:07
Breath Volume	2.089	04:07
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

---

Data Download Date/Time: 10/17/2018 15:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JAMES,A,L  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FLORENCE,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 12:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:08

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 10/03/2018 13:32:33

---

Diagnostics Check	OK	13:29	
Air Blank	0.000	13:30	
Control Test	0.081	13:30	
Air Blank	0.000	13:30	
Subject Sample	# 1	13:31	Subject Test Refused
Breath Volume	0.000	13:31	
Air Blank	0.000	13:31	
Control Test	0.080	13:31	
Air Blank	0.000	13:32	
Diagnostics Check	OK	13:32	

---

Data Download Date/Time: 10/17/2018 15:32



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOSTON,D,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 10/05/2018 00:01:34

---

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.079	23:55
Air Blank	0.000	23:56
Subject Sample # 2	0.161	23:56
Breath Volume	1.324	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.156	00:00
Breath Volume	1.597	00:00
Air Blank	0.000	00:00
Control Test	0.080	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

---

Data Download Date/Time: 10/17/2018 15:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOSTON,D,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 10/10/2018 00:11:51

---

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:09	
Control Test	0.080	00:09	
Air Blank	0.000	00:10	
Subject Sample	# 1	00:10	Subject Test Refused
Breath Volume	0.000	00:10	
Air Blank	0.000	00:10	
Control Test	0.081	00:11	
Air Blank	0.000	00:11	
Diagnostics Check	OK	00:11	

---

Data Download Date/Time: 10/17/2018 15:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOSTON,D,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: WATCHES

DATE/TIME OF TEST: 10/11/2018 00:42:49

---

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.030	00:38
Breath Volume	2.226	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.029	00:41
Breath Volume	1.488	00:41
Air Blank	0.000	00:41
Control Test	0.081	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

---

Data Download Date/Time: 10/17/2018 15:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JAMES,A,L  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CLAYDON,P  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:08

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 10/14/2018 08:36:25

---

Diagnostics Check	OK	08:29
Air Blank	0.000	08:30
Control Test	0.080	08:30
Air Blank	0.000	08:31
Subject Sample # 1	0.087	08:31
Breath Volume	2.425	08:31
Air Blank	0.000	08:32
Air Blank	0.000	08:34
Subject Sample # 2	0.088	08:34
Breath Volume	1.687	08:34
Air Blank	0.000	08:35
Control Test	0.080	08:35
Air Blank	0.000	08:36
Diagnostics Check	OK	08:36

---

Data Download Date/Time: 10/17/2018 15:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AUGUSTINE,G,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HAWKINS,A  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: STEPNES

DATE/TIME OF TEST: 10/19/2018 04:01:01

---

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:58	
Control Test	0.081	03:58	
Air Blank	0.000	03:59	
Subject Sample	# 1	03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	04:00	
Control Test	0.080	04:00	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:01	

---

Data Download Date/Time: 11/06/2018 12:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOSTON,D,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: EANES,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: TYSINGER

DATE/TIME OF TEST: 10/20/2018 01:09:45

---

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.080	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.173	01:04
Breath Volume	2.109	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.183	01:08
Breath Volume	4.058	01:08
Air Blank	0.000	01:08
Control Test	0.079	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

---

Data Download Date/Time: 11/06/2018 12:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOSTON,D,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BROWN,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: CANNIZZARO

DATE/TIME OF TEST: 10/20/2018 02:51:53

---

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.166	02:47
Breath Volume	1.625	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.172	02:50
Breath Volume	2.441	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

---

Data Download Date/Time: 11/06/2018 12:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNS,D,L  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JENSEN,D  
ARRESTING OFFICER AGENCY: Neptune Beach PD  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: OQUENDO MEDINA

DATE/TIME OF TEST: 10/20/2018 05:44:52

---

Diagnostics Check	OK	05:38
Air Blank	0.000	05:38
Control Test	0.080	05:39
Air Blank	0.000	05:39
Subject Sample # 1	0.191	05:40
Breath Volume	2.507	05:40
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample # 2	0.192	05:43
Breath Volume	2.347	05:43
Air Blank	0.000	05:43
Control Test	0.080	05:44
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

---

Data Download Date/Time: 11/06/2018 12:16



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOSTON,D,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JESTER,B  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: MIDDLETON

DATE/TIME OF TEST: 10/22/2018 00:55:01

---

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.081	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.000	00:50
Breath Volume	2.437	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.000	00:53
Breath Volume	2.476	00:53
Air Blank	0.000	00:54
Control Test	0.081	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:55

---

Data Download Date/Time: 11/06/2018 12:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ZURN,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 10/27/2018 04:48:15

---

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.081	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.128	04:43
Breath Volume	1.417	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.131	04:46
Breath Volume	1.703	04:46
Air Blank	0.000	04:47
Control Test	0.080	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

---

Data Download Date/Time: 11/06/2018 12:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JAMES,A,L  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SPENCER,K  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A3  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 10:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:32

SUBJECT'S LAST NAME: RAYMOND

DATE/TIME OF TEST: 10/28/2018 13:59:04

---

Diagnostics Check	OK	13:52
Air Blank	0.000	13:53
Control Test	0.081	13:53
Air Blank	0.000	13:53
Subject Sample # 1	0.275	13:54
Breath Volume	2.171	13:54
Air Blank	0.000	13:55
Air Blank	0.000	13:57
Subject Sample # 2	0.290	13:57
Breath Volume	2.300	13:57
Air Blank	0.000	13:58
Control Test	0.080	13:58
Air Blank	0.000	13:58
Diagnostics Check	OK	13:59

---

Data Download Date/Time: 11/06/2018 12:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HERNANDEZ,ALAIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: TOMLIN JR

DATE/TIME OF TEST: 10/02/2018 00:44:37

---

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.078	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.036	00:39
Breath Volume	1.535	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.035	00:43
Breath Volume	1.699	00:43
Air Blank	0.000	00:43
Control Test	0.078	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

---

Data Download Date/Time: 10/29/2018 17:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HERNANDEZ,ALAIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: CRISP

DATE/TIME OF TEST: 10/02/2018 03:30:50

---

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.078	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.158	03:25
Breath Volume	1.128	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.157	03:29
Breath Volume	1.542	03:29
Air Blank	0.000	03:29
Control Test	0.077	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

---

Data Download Date/Time: 10/29/2018 17:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HERNANDEZ,ALAIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 10/06/2018 23:25:50

---

Diagnostics Check	OK	23:19
Air Blank	0.000	23:20
Control Test	0.079	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.176	23:21
Breath Volume	2.816	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:23
Subject Sample # 2	0.174	23:24
Breath Volume	2.718	23:24
Air Blank	0.000	23:24
Control Test	0.079	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

---

Data Download Date/Time: 10/29/2018 17:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HERNANDEZ,ALAIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CHAREO PEREZ

DATE/TIME OF TEST: 10/07/2018 03:13:19

---

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.078	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.209	03:09
Breath Volume	2.781	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.206	03:11
Breath Volume	2.238	03:11
Air Blank	0.000	03:12
Control Test	0.077	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

---

Data Download Date/Time: 10/29/2018 17:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALSINA, JONATHAN, D  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: BROWN, TIMOTHY  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 11:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: ALSINA

DATE/TIME OF TEST: 10/04/2018 12:15:07

---

Diagnostics Check	OK	12:11	
Air Blank	0.000	12:12	
Control Test	0.081	12:12	
Air Blank	0.000	12:13	
Subject Sample	# 1	12:13	Subject Test Refused
Breath Volume	0.000	12:13	
Air Blank	0.000	12:14	
Control Test	0.081	12:14	
Air Blank	0.000	12:15	
Diagnostics Check	OK	12:15	

---

Data Download Date/Time: 10/31/2018 01:33



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLEARE,PAMELA,F  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: SANTANA,YAMIL  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: VELASCO ARBULU

DATE/TIME OF TEST: 10/07/2018 03:19:31

---

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.081	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.153	03:15
Breath Volume	1.500	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.158	03:17
Breath Volume	2.054	03:17
Air Blank	0.000	03:18
Control Test	0.081	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

---

Data Download Date/Time: 10/31/2018 01:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LLANIO,PIERRE,  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: RIVERA,NEPHTALI  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 11:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:48

SUBJECT'S LAST NAME: SARRIA

DATE/TIME OF TEST: 10/09/2018 12:16:09

---

Diagnostics Check	OK	12:09
Air Blank	0.000	12:09
Control Test	0.081	12:10
Air Blank	0.000	12:10
Subject Sample # 1	0.201	12:11
Breath Volume	1.796	12:11
Air Blank	0.000	12:12
Air Blank	0.000	12:14
Subject Sample # 2	0.192	12:14
Breath Volume	1.945	12:14
Air Blank	0.000	12:15
Control Test	0.081	12:15
Air Blank	0.000	12:16
Diagnostics Check	OK	12:16

---

Data Download Date/Time: 10/31/2018 01:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLEARE,PAMELA,F  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: CLEARE,PAMELA  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: RIBOT

DATE/TIME OF TEST: 10/13/2018 01:36:55

---

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.082	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.156	01:30
Breath Volume	1.734	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.147	01:35
Breath Volume	1.648	01:35
Air Blank	0.000	01:35
Control Test	0.081	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

---

Data Download Date/Time: 10/31/2018 01:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLEARE,PAMELA,F  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: CARTER,MOSEL  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: CHILDERS JR

DATE/TIME OF TEST: 10/13/2018 01:52:35

---

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.000	01:48
Breath Volume	2.253	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.000	01:51
Breath Volume	2.714	01:51
Air Blank	0.000	01:51
Control Test	0.082	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

---

Data Download Date/Time: 10/31/2018 01:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MADDOX,KEOSHA,M  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: MADDOX,KEOSHA  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: MUNOZ ALVAREZ

DATE/TIME OF TEST: 10/13/2018 04:36:56

---

Diagnostics Check	OK	04:35	
Air Blank	0.000	04:35	
Control Test	0.082	04:36	
Air Blank		04:36	Improper Sample
Air Blank		04:36	Improper Sample

---

Data Download Date/Time: 10/31/2018 01:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MADDIX,KEOSHA,M  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: MADDOX,KEOSHA  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: MUNOZ ALVAREZ

DATE/TIME OF TEST: 10/13/2018 04:43:28

---

Diagnostics Check	OK	04:42	
Air Blank		04:43	Ambient Fail
Air Blank	0.000	04:43	

---

Data Download Date/Time: 10/31/2018 01:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLEARE,PAMELA,F  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: SANTANA,YUDIEL  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MASCIOPINTO

DATE/TIME OF TEST: 10/14/2018 03:33:01

---

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.082	03:32	
Air Blank		03:32	RFI Detect
Air Blank	0.000	03:33	

---

Data Download Date/Time: 10/31/2018 01:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLEARE,PAMELA,F  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: SANTANA,YUDIEL  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MASCIOPINTO

DATE/TIME OF TEST: 10/14/2018 03:41:37

---

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:39	
Control Test	0.081	03:39	
Air Blank	0.000	03:39	
Subject Sample	# 1	03:40	Subject Test Refused
Breath Volume	0.000	03:40	
Air Blank	0.000	03:40	
Control Test	0.082	03:40	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

---

Data Download Date/Time: 10/31/2018 01:33



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUTIERREZ DEL TORO,RUBEN,  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: GUTIERREZ DEL TORO,RUBEN  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: SALINAS

DATE/TIME OF TEST: 10/14/2018 17:45:24

---

Diagnostics Check	OK	17:39
Air Blank	0.000	17:39
Control Test	0.080	17:40
Air Blank	0.000	17:40
Subject Sample # 1	0.125	17:41
Breath Volume	2.496	17:41
Air Blank	0.000	17:41
Air Blank	0.000	17:43
Subject Sample # 2	0.123	17:43
Breath Volume	1.691	17:43
Air Blank	0.000	17:44
Control Test	0.079	17:44
Air Blank	0.000	17:45
Diagnostics Check	OK	17:45

---

Data Download Date/Time: 10/31/2018 01:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALBO,JOSHUA,D  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: OCHOA,GUILLERMO  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 09:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:37

SUBJECT'S LAST NAME: VILLASENOR

DATE/TIME OF TEST: 10/28/2018 10:16:21

---

Diagnostics Check	OK	10:09
Air Blank	0.000	10:10
Control Test	0.082	10:10
Air Blank	0.000	10:11
Subject Sample # 1	0.165	10:11
Breath Volume	2.796	10:11
Air Blank	0.000	10:12
Air Blank	0.000	10:14
Subject Sample # 2	0.166	10:14
Breath Volume	2.511	10:14
Air Blank	0.000	10:15
Control Test	0.082	10:15
Air Blank	0.000	10:16
Diagnostics Check	OK	10:16

---

Data Download Date/Time: 10/31/2018 01:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAGLE, TONY,  
BREATH TEST OPERATOR AGENCY: Deland PD  
ARRESTING OFFICER: CAMPBELL, CHRISTOPHER  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 10/03/2018 00:41:20

---

Diagnostics Check		OK	00:35
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.188	00:36
Breath Volume		1.718	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.194	00:39
Breath Volume		1.183	00:39
Air Blank		0.000	00:40
Control Test		0.079	00:40
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

---

Data Download Date/Time: 11/02/2018 00:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAGLE, TONY,  
BREATH TEST OPERATOR AGENCY: Deland PD  
ARRESTING OFFICER: RODRIGUEZ, HECTOR  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: GARCIA SANCHEZ

DATE/TIME OF TEST: 10/06/2018 02:11:54

---

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:08	
Control Test	0.080	02:08	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:10	Subject Test Refused
Breath Volume	0.000	02:10	
Air Blank	0.000	02:11	
Control Test	0.079	02:11	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

---

Data Download Date/Time: 11/02/2018 00:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAGLE, TONY,  
BREATH TEST OPERATOR AGENCY: Deland PD  
ARRESTING OFFICER: RODRIGUEZ, HECTOR  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: GARCIA SANCHEZ

DATE/TIME OF TEST: 10/06/2018 02:24:54

---

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:19	
Control Test	0.079	02:20	
Air Blank	0.000	02:20	
Subject Sample	# 1	02:23	Subject Test Refused
Breath Volume	0.941	02:23	
Air Blank	0.000	02:23	
Control Test	0.080	02:24	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

---

Data Download Date/Time: 11/02/2018 00:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,BRENT,M  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: JONES,BRENT  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 10/08/2018 02:41:46

---

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.150	02:36
Breath Volume	1.363	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.140	02:40
Breath Volume	1.464	02:40
Air Blank	0.000	02:40
Control Test	0.080	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

---

Data Download Date/Time: 11/02/2018 00:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALTER,CRAIG,K  
BREATH TEST OPERATOR AGENCY: Deland PD  
ARRESTING OFFICER: COCHRAN,JOHN  
ARRESTING OFFICER AGENCY: Deland PD  
LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: SHOWALTER

DATE/TIME OF TEST: 10/14/2018 00:37:56

---

Diagnostics Check		OK	00:31
Air Blank		0.000	00:32
Control Test		0.079	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.151	00:33
Breath Volume		1.519	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:35
Subject Sample	# 2	0.147	00:36
Breath Volume		2.343	00:36
Air Blank		0.000	00:37
Control Test		0.078	00:37
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

---

Data Download Date/Time: 11/02/2018 00:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CODY,RICHARD,D  
BREATH TEST OPERATOR AGENCY: Deland PD  
ARRESTING OFFICER: EDGECOMB,STEVE  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: PREVO

DATE/TIME OF TEST: 10/20/2018 20:53:55

---

Diagnostics Check	OK	20:51	
Air Blank	0.000	20:51	
Control Test	0.080	20:51	
Air Blank	0.000	20:52	
Subject Sample	# 1	20:52	Subject Test Refused
Breath Volume	0.000	20:52	
Air Blank	0.000	20:53	
Control Test	0.080	20:53	
Air Blank	0.000	20:53	
Diagnostics Check	OK	20:53	

---

Data Download Date/Time: 11/02/2018 00:18



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEMPLE,DARRYL,D  
BREATH TEST OPERATOR AGENCY: Calhoun County SO  
ARRESTING OFFICER: GOODMAN,DONALD  
ARRESTING OFFICER AGENCY: Calhoun County SO  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 10/05/2018 05:11:47

---

Diagnostics Check	OK	05:05
Air Blank	0.000	05:06
Control Test	0.079	05:06
Air Blank	0.000	05:06
Subject Sample # 1	0.189	05:07
Breath Volume	3.574	05:07
Air Blank	0.000	05:08
Air Blank	0.000	05:09
Subject Sample # 2	0.190	05:10
Breath Volume	4.000	05:10
Air Blank	0.000	05:10
Control Test	0.076	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

---

Data Download Date/Time: 10/25/2018 09:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHILLING,JOHN,R  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SCHILLING,JOHN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 06717080A6  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 11:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BREWER

DATE/TIME OF TEST: 10/20/2018 01:24:39

---

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.070	01:20
Breath Volume	2.328	01:20
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.071	01:23
Breath Volume	3.066	01:23
Air Blank	0.000	01:23
Control Test	0.078	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

---

Data Download Date/Time: 10/25/2018 09:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHIVER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Liberty County SO  
ARRESTING OFFICER: TAYLOR,CHASE  
ARRESTING OFFICER AGENCY: Liberty County SO  
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: CRUM

DATE/TIME OF TEST: 10/25/2018 17:46:31

---

Diagnostics Check	OK	17:40
Air Blank	0.000	17:40
Control Test	0.078	17:41
Air Blank	0.000	17:41
Subject Sample # 1	0.000	17:42
Breath Volume	2.015	17:42
Air Blank	0.000	17:42
Air Blank	0.000	17:44
Subject Sample # 2	0.000	17:45
Breath Volume	2.046	17:45
Air Blank	0.000	17:45
Control Test	0.078	17:45
Air Blank	0.000	17:46
Diagnostics Check	OK	17:46

---

Data Download Date/Time: 11/13/2018 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCALLISTER, WILLIAM, A  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: MELLO, ANTHONY  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 10/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 13:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 10/02/2018 14:39:55

---

Diagnostics Check	OK	14:32
Air Blank	0.000	14:32
Control Test	0.079	14:33
Air Blank	0.000	14:33
Subject Sample # 1	0.332	14:35
Breath Volume	1.187	14:35
Air Blank	0.000	14:36
Air Blank	0.000	14:37
Subject Sample # 2	0.328	14:38
Breath Volume	2.144	14:38
Air Blank	0.000	14:38
Control Test	0.077	14:39
Air Blank	0.000	14:39
Diagnostics Check	OK	14:39

---

Data Download Date/Time: 12/03/2018 02:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAY,ASHLEY,L  
BREATH TEST OPERATOR AGENCY: Eglin AFB PD  
ARRESTING OFFICER: N/A,N/A  
ARRESTING OFFICER AGENCY: Eglin AFB PD  
LAST AGENCY INSPECTION: 10/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: OTHER  
ARREST TIME: 08:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: GEFFERT

DATE/TIME OF TEST: 10/04/2018 09:11:10

---

Diagnostics Check	OK	09:05
Air Blank	0.000	09:05
Control Test	0.079	09:05
Air Blank	0.000	09:06
Subject Sample # 1	0.013	09:06
Breath Volume	2.257	09:06
Air Blank	0.000	09:07
Air Blank	0.000	09:09
Subject Sample # 2	0.013	09:09
Breath Volume	2.148	09:09
Air Blank	0.000	09:10
Control Test	0.079	09:10
Air Blank	0.000	09:11
Diagnostics Check	OK	09:11

---

Data Download Date/Time: 12/03/2018 02:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GRIFFIN,RAEFORD  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 10/02/2018 19:53:51

---

Diagnostics Check	OK	19:47
Air Blank	0.000	19:47
Control Test	0.083	19:47
Air Blank	0.000	19:48
Subject Sample # 1	0.000	19:49
Breath Volume	1.640	19:49
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample # 2	0.000	19:52
Breath Volume	1.238	19:52
Air Blank	0.000	19:52
Control Test	0.082	19:53
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

---

Data Download Date/Time: 10/10/2018 18:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BECKMAN,KEVIN,E  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: IVEY,BRIAN  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: TURMEL

DATE/TIME OF TEST: 10/04/2018 03:58:10

---

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.081	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.256	03:52
Breath Volume		2.648	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.246	03:56
Breath Volume		2.937	03:56
Air Blank		0.000	03:57
Control Test		0.080	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

---

Data Download Date/Time: 10/10/2018 18:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEQUEIRA,NATHAN,C  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SEQUEIRA,NATHAN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: RENO

DATE/TIME OF TEST: 10/06/2018 19:19:18

---

Diagnostics Check	OK	19:07	
Air Blank	0.000	19:08	
Control Test	0.082	19:08	
Air Blank	0.000	19:09	
Subject Sample	# 1	19:12	No Sample Provided
Breath Volume	0.000	19:12	
Air Blank	0.000	19:12	
Air Blank	0.000	19:14	
Subject Sample	# 2	19:17	No Sample Provided
Breath Volume	0.000	19:17	
Air Blank	0.000	19:18	
Control Test	0.082	19:18	
Air Blank	0.000	19:19	
Diagnostics Check	OK	19:19	

---

Data Download Date/Time: 10/10/2018 18:31



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GRIFFIN,RAEFORD  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: WEYNA

DATE/TIME OF TEST: 10/06/2018 19:36:06

---

Diagnostics Check	OK	19:27	
Air Blank	0.000	19:27	
Control Test	0.082	19:28	
Air Blank	0.000	19:28	
Subject Sample	# 1	19:31	No Sample Provided
Breath Volume	0.000	19:31	
Air Blank	0.000	19:32	
Air Blank	0.000	19:34	
Subject Sample	# 2	19:34	Subject Test Refused
Breath Volume	0.000	19:34	
Air Blank	0.000	19:35	
Control Test	0.082	19:35	
Air Blank	0.000	19:36	
Diagnostics Check	OK	19:36	

---

Data Download Date/Time: 10/10/2018 18:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINTERS,LANE,A  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: HAMMELMAN,TYLER  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 10/07/2018 01:07:22

---

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.082	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.173	01:02
Breath Volume		4.058	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.176	01:05
Breath Volume		2.550	01:05
Air Blank		0.000	01:06
Control Test		0.082	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

---

Data Download Date/Time: 10/10/2018 18:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRIOLAT,CASEY,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: BRIOLAT,CASEY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:34

SUBJECT'S LAST NAME: MICHEL

DATE/TIME OF TEST: 10/07/2018 07:06:01

---

Diagnostics Check	OK	06:54	
Air Blank	0.000	06:55	
Control Test	0.082	06:55	
Air Blank	0.000	06:55	
Subject Sample	# 1	06:59	No Sample Provided
Breath Volume	0.000	06:59	
Air Blank	0.000	06:59	
Air Blank	0.000	07:01	
Subject Sample	# 2	07:04	No Sample Provided
Breath Volume	0.000	07:04	
Air Blank	0.000	07:05	
Control Test	0.082	07:05	
Air Blank	0.000	07:05	
Diagnostics Check	OK	07:06	

---

Data Download Date/Time: 10/10/2018 18:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GRIFFIN,RAEFORD  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/26/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BRENNAN

DATE/TIME OF TEST: 10/08/2018 02:44:33

---

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:35	
Control Test	0.082	02:36	
Air Blank	0.000	02:36	
Subject Sample	# 1	02:39	No Sample Provided
Breath Volume	0.000	02:39	
Air Blank	0.000	02:40	
Air Blank	0.000	02:42	
Subject Sample	# 2	02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:43	
Control Test	0.083	02:43	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

---

Data Download Date/Time: 10/10/2018 18:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEQUEIRA,NATHAN,C  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SEQUEIRA,NATHAN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: VERALDI

DATE/TIME OF TEST: 10/11/2018 20:28:14

---

Diagnostics Check	OK	20:21
Air Blank	0.000	20:22
Control Test	0.081	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.239	20:23
Breath Volume	1.863	20:23
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.240	20:26
Breath Volume	1.726	20:26
Air Blank	0.000	20:27
Control Test	0.081	20:27
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

---

Data Download Date/Time: 11/26/2018 19:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NIXON,BARRY,  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: NIXON,BARRY  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: PENDRYS

DATE/TIME OF TEST: 10/11/2018 23:16:13

---

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.082	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.000	23:11
Breath Volume	1.140	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.000	23:14
Breath Volume	1.839	23:14
Air Blank	0.000	23:15
Control Test	0.082	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

---

Data Download Date/Time: 11/26/2018 19:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NIXON,BARRY,  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: NIXON,BARRY  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: THANASOULIS

DATE/TIME OF TEST: 10/13/2018 03:09:25

---

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.081	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.019	03:05
Breath Volume		1.867	03:05
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.019	03:07
Breath Volume		2.667	03:07
Air Blank		0.000	03:08
Control Test		0.082	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

---

Data Download Date/Time: 11/26/2018 19:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REAY,ALEX,M  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: REAY,ALEX  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: BUCKLEY

DATE/TIME OF TEST: 10/13/2018 07:33:45

---

Diagnostics Check	OK	07:30	
Air Blank	0.000	07:31	
Control Test	0.083	07:31	
Air Blank	0.000	07:32	
Subject Sample	# 1	07:32	Subject Test Refused
Breath Volume	0.000	07:32	
Air Blank	0.000	07:32	
Control Test	0.082	07:33	
Air Blank	0.000	07:33	
Diagnostics Check	OK	07:33	

---

Data Download Date/Time: 11/26/2018 19:04



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINTERS,LANE,A  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: MILLER,KEVIN  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 10/16/2018 03:58:50

---

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.082	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.112	03:54
Breath Volume	1.894	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.109	03:57
Breath Volume	2.808	03:57
Air Blank	0.000	03:57
Control Test	0.082	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

---

Data Download Date/Time: 11/26/2018 19:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KNORR,PETER,J  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: HAMMELMAN,TYLER  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: FARRELL

DATE/TIME OF TEST: 10/16/2018 19:31:24

---

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.082	19:25
Air Blank	0.000	19:26
Subject Sample # 1	0.000	19:26
Breath Volume	1.773	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.000	19:30
Breath Volume	2.203	19:30
Air Blank	0.000	19:30
Control Test	0.082	19:30
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

---

Data Download Date/Time: 11/26/2018 19:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORRIS, ADAM, M  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: MORRIS, ADAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/18/2018 02:21:37

---

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.081	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.127	02:17
Breath Volume	1.394	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.125	02:20
Breath Volume	2.121	02:20
Air Blank	0.000	02:20
Control Test	0.082	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

---

Data Download Date/Time: 11/26/2018 19:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NIXON,BARRY,  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: NIXON,BARRY  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: LANGEVIN

DATE/TIME OF TEST: 10/19/2018 04:14:26

---

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.082	04:09
Air Blank	0.000	04:09
Subject Sample # 1	0.000	04:10
Breath Volume	2.628	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.000	04:13
Breath Volume	2.050	04:13
Air Blank	0.000	04:13
Control Test	0.084	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

---

Data Download Date/Time: 11/26/2018 19:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEQUEIRA,NATHAN,C  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SEQUEIRA,NATHAN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:33

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/21/2018 17:59:44

---

Diagnostics Check	OK	17:53
Air Blank	0.000	17:53
Control Test	0.083	17:54
Air Blank	0.000	17:54
Subject Sample # 1	0.256	17:55
Breath Volume	1.699	17:55
Air Blank	0.000	17:55
Air Blank	0.000	17:57
Subject Sample # 2	0.251	17:58
Breath Volume	1.621	17:58
Air Blank	0.000	17:58
Control Test	0.080	17:59
Air Blank	0.000	17:59
Diagnostics Check	OK	17:59

---

Data Download Date/Time: 11/26/2018 19:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REAY,ALEX,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: REAY,ALEX  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 11:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 10/24/2018 01:29:41

---

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.083	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.000	01:25
Breath Volume	2.019	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.000	01:28
Breath Volume	1.656	01:28
Air Blank	0.000	01:28
Control Test	0.083	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

---

Data Download Date/Time: 11/26/2018 19:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNON, MATTHEW, F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HANNON, MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: BRANSTETTER

DATE/TIME OF TEST: 10/24/2018 01:35:37

---

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:33	
Control Test	0.083	01:33	
Air Blank	0.000	01:34	
Subject Sample	# 1	01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:34	
Control Test	0.083	01:35	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

---

Data Download Date/Time: 11/26/2018 19:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REAY,ALEX,M  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SMITH-LONG,JEFFERSON  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 10/24/2018 04:06:05

---

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.083	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.165	04:01
Breath Volume	1.394	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.173	04:04
Breath Volume	2.367	04:04
Air Blank	0.000	04:05
Control Test	0.082	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

---

Data Download Date/Time: 11/26/2018 19:04



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GRIFFIN,RAEFORD  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: SINGER

DATE/TIME OF TEST: 10/26/2018 02:39:15

---

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:29	
Control Test	0.082	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:33	No Sample Provided
Breath Volume	0.000	02:33	
Air Blank	0.000	02:34	
Air Blank	0.000	02:36	
Subject Sample	# 2	02:37	Subject Test Refused
Breath Volume	0.000	02:37	
Air Blank	0.000	02:38	
Control Test	0.082	02:38	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

---

Data Download Date/Time: 11/26/2018 19:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KNORR,PETER,J  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: KNORR,PETER  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: SPAIN

DATE/TIME OF TEST: 10/26/2018 21:48:39

---

Diagnostics Check	OK	21:37	
Air Blank	0.000	21:37	
Control Test	0.082	21:38	
Air Blank	0.000	21:38	
Subject Sample	# 1	21:41	No Sample Provided
Breath Volume	0.000	21:41	
Air Blank	0.000	21:42	
Air Blank	0.000	21:44	
Subject Sample	# 2	21:47	No Sample Provided
Breath Volume	0.000	21:47	
Air Blank	0.000	21:47	
Control Test	0.082	21:48	
Air Blank	0.000	21:48	
Diagnostics Check	OK	21:48	

---

Data Download Date/Time: 11/26/2018 19:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KNORR,PETER,J  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: KNORR,PETER  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SPAIN

DATE/TIME OF TEST: 10/26/2018 21:56:34

---

Diagnostics Check	OK	21:50
Air Blank	0.000	21:50
Control Test	0.081	21:50
Air Blank	0.000	21:51
Subject Sample # 1	0.154	21:52
Breath Volume	3.941	21:52
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 2	0.157	21:55
Breath Volume	3.484	21:55
Air Blank	0.000	21:55
Control Test	0.081	21:55
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

---

Data Download Date/Time: 11/26/2018 19:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAUGHERTY,D,M  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: DAUGHERTY,D  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BLAUERS

DATE/TIME OF TEST: 10/27/2018 03:38:04

---

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.082	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.000	03:33
Breath Volume	3.214	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.000	03:36
Breath Volume	3.367	03:36
Air Blank	0.000	03:37
Control Test	0.083	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:38

---

Data Download Date/Time: 11/26/2018 19:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNON, MATTHEW, F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HANNON, MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 10/27/2018 20:26:15

---

Diagnostics Check	OK	20:23	
Air Blank	0.000	20:23	
Control Test	0.083	20:24	
Air Blank	0.000	20:24	
Subject Sample	# 1	20:24	Subject Test Refused
Breath Volume	0.000	20:24	
Air Blank	0.000	20:25	
Control Test	0.082	20:25	
Air Blank	0.000	20:26	
Diagnostics Check	OK	20:26	

---

Data Download Date/Time: 11/26/2018 19:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REAY,ALEX,M  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: REAY,ALEX  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 12:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: ULRICH

DATE/TIME OF TEST: 10/28/2018 02:05:24

---

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.083	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.184	02:01
Breath Volume	2.812	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.183	02:03
Breath Volume	3.792	02:03
Air Blank	0.000	02:04
Control Test	0.082	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

---

Data Download Date/Time: 11/26/2018 19:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAUGHERTY,D,M  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: DAUGHERTY,D  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: EADY

DATE/TIME OF TEST: 10/28/2018 05:58:33

---

Diagnostics Check	OK	05:45	
Air Blank	0.000	05:45	
Control Test	0.082	05:46	
Air Blank	0.000	05:46	
Subject Sample	# 1 0.246	05:48	
Breath Volume	1.187	05:48	
Air Blank	0.000	05:48	
Air Blank	0.000	05:50	
Subject Sample	# 2 0.189	05:53	Volume Not Met
Breath Volume	0.445	05:53	
Air Blank	0.000	05:54	
Air Blank	0.000	05:56	
Subject Sample	# 3 0.239	05:57	
Breath Volume	1.511	05:57	
Air Blank	0.000	05:57	
Control Test	0.079	05:57	
Air Blank	0.000	05:58	
Diagnostics Check	OK	05:58	

---

Data Download Date/Time: 11/26/2018 19:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNON, MATTHEW, F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HANNON, MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/10/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: TONRY

DATE/TIME OF TEST: 10/29/2018 05:31:45

---

Diagnostics Check	OK	05:23	
Air Blank	0.000	05:23	
Control Test	0.082	05:24	
Air Blank	0.000	05:24	
Subject Sample	# 1	05:27	No Sample Provided
Breath Volume	0.000	05:27	
Air Blank	0.000	05:28	
Air Blank	0.000	05:30	
Subject Sample	# 2	05:30	Subject Test Refused
Breath Volume	0.000	05:30	
Air Blank	0.000	05:30	
Control Test	0.081	05:31	
Air Blank	0.000	05:31	
Diagnostics Check	OK	05:31	

---

Data Download Date/Time: 11/26/2018 19:04



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: IVEY,BRUCE,A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: CORNIER,GABRIEL  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/08/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 09:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 10/31/2018 11:49:22

---

Diagnostics Check		OK	11:43
Air Blank		0.000	11:43
Control Test		0.080	11:43
Air Blank		0.000	11:44
Subject Sample	# 1	0.305	11:44
Breath Volume		2.406	11:44
Air Blank		0.000	11:45
Air Blank		0.000	11:47
Subject Sample	# 2	0.308	11:47
Breath Volume		3.910	11:47
Air Blank		0.000	11:48
Control Test		0.078	11:48
Air Blank		0.000	11:49
Diagnostics Check		OK	11:49

---

Data Download Date/Time: 11/05/2018 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUSSIER,BRENDAN,S  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: PRATTS,ALBERT  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/01/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 437738  
DRY GAS EXPIRY DATE: 06/08/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 10/12/2018 17:51:24

---

Diagnostics Check	OK	17:44
Air Blank	0.000	17:44
Control Test	0.080	17:45
Air Blank	0.000	17:45
Subject Sample # 1	0.169	17:46
Breath Volume	1.328	17:46
Air Blank	0.000	17:47
Air Blank	0.000	17:49
Subject Sample # 2	0.154	17:49
Breath Volume	1.507	17:49
Air Blank	0.000	17:50
Control Test	0.080	17:50
Air Blank	0.000	17:51
Diagnostics Check	OK	17:51

---

Data Download Date/Time: 11/27/2018 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MADDOX,KEOSHA,M  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: MADDOX,KEOSHA  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 09/30/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: MUNOZ ALVAREZ

DATE/TIME OF TEST: 10/13/2018 04:57:31

---

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.080	04:52
Air Blank	0.000	04:52
Subject Sample # 1	0.151	04:52
Breath Volume	3.167	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.146	04:55
Breath Volume	3.277	04:55
Air Blank	0.000	04:56
Control Test	0.080	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

---

Data Download Date/Time: 10/31/2018 01:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER, ADAM, H  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: POTTER, ADAM  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 09/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: HORNSBY

DATE/TIME OF TEST: 10/16/2018 21:07:31

---

Diagnostics Check	OK	21:01
Air Blank	0.000	21:01
Control Test	0.080	21:01
Air Blank	0.000	21:02
Subject Sample # 1	0.187	21:03
Breath Volume	2.117	21:03
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.199	21:06
Breath Volume	2.562	21:06
Air Blank	0.000	21:06
Control Test	0.080	21:06
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

---

Data Download Date/Time: 10/27/2018 13:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BING, TRAE, J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: DENMARK, HAROLD  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: HALE

DATE/TIME OF TEST: 10/28/2018 02:47:26

---

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:46	
Control Test	0.079	02:46	
Air Blank		02:46	RFI Detect
Air Blank	0.000	02:47	

---

Data Download Date/Time: 11/28/2018 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BING, TRAE, J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: DENMARK, HAROLD  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: HALE

DATE/TIME OF TEST: 10/28/2018 02:56:23

---

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.080	02:54	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:55	
Control Test	0.079	02:55	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

---

Data Download Date/Time: 11/28/2018 09:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COKE,CLEVELAND,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BELCHER,ALEX  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/27/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A4  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: GATLIN

DATE/TIME OF TEST: 10/31/2018 03:38:36

---

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.165	03:34
Breath Volume	1.628	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.167	03:37
Breath Volume	2.156	03:37
Air Blank	0.000	03:37
Control Test	0.078	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

---

Data Download Date/Time: 11/28/2018 09:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PATSCH,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: PATSCH,JOSHUA  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 09/28/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 17:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: REISZ

DATE/TIME OF TEST: 10/03/2018 18:18:00

---

Diagnostics Check	OK	18:11	
Air Blank	0.000	18:12	
Control Test	0.075	18:12	
Air Blank	0.000	18:13	
Subject Sample # 1	0.000	18:13	
Breath Volume	3.214	18:13	
Air Blank	0.000	18:14	
Air Blank	0.000	18:16	
Subject Sample # 2	0.000	18:16	
Breath Volume	3.746	18:16	
Air Blank	0.000	18:17	
Control Test	0.074	18:17	Control Outside Tolerance
Air Blank	0.000	18:18	

---

Data Download Date/Time: 10/25/2018 14:01



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PATSCH,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: PATSCH,JOSHUA  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: IARIA

DATE/TIME OF TEST: 10/27/2018 17:38:35

---

Diagnostics Check	OK	17:31
Air Blank	0.000	17:32
Control Test	0.079	17:32
Air Blank	0.000	17:33
Subject Sample # 1	0.236	17:33
Breath Volume	2.355	17:33
Air Blank	0.000	17:34
Air Blank	0.000	17:36
Subject Sample # 2	0.232	17:37
Breath Volume	2.296	17:37
Air Blank	0.000	17:37
Control Test	0.079	17:37
Air Blank	0.000	17:38
Diagnostics Check	OK	17:38

---

Data Download Date/Time: 11/09/2018 12:22