

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IHNKEN,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: LICKISS,BRANT
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 11/19/2018 04:40:02

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.286	04:33
Breath Volume	1.640	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.287	04:38
Breath Volume	1.539	04:38
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:40

Data Download Date/Time: 12/28/2018 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HANSARD,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 12:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:32

SUBJECT'S LAST NAME: HARGIS

DATE/TIME OF TEST: 11/24/2018 13:02:22

Diagnostics Check	OK	12:55
Air Blank	0.000	12:56
Control Test	0.080	12:56
Air Blank	0.000	12:57
Subject Sample # 1	0.296	12:58
Breath Volume	1.710	12:58
Air Blank	0.000	12:58
Air Blank	0.000	13:00
Subject Sample # 2	0.301	13:00
Breath Volume	1.406	13:00
Air Blank	0.000	13:01
Control Test	0.080	13:01
Air Blank	0.000	13:02
Diagnostics Check	OK	13:02

Data Download Date/Time: 12/28/2018 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,BRUCE,L
BREATH TEST OPERATOR AGENCY: St. Augustine Beach
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: AHRENS

DATE/TIME OF TEST: 11/03/2018 19:35:49

Diagnostics Check	OK	19:29
Air Blank	0.000	19:30
Control Test	0.079	19:30
Air Blank	0.000	19:30
Subject Sample # 1	0.240	19:31
Breath Volume	2.828	19:31
Air Blank	0.000	19:32
Air Blank	0.000	19:33
Subject Sample # 2	0.225	19:34
Breath Volume	2.171	19:34
Air Blank	0.000	19:34
Control Test	0.079	19:35
Air Blank	0.000	19:35
Diagnostics Check	OK	19:35

Data Download Date/Time: 11/05/2018 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,BRUCE,L
BREATH TEST OPERATOR AGENCY: St. Augustine Beach
ARRESTING OFFICER: CUBBEDGE,KYLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 11/03/2018 21:55:36

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.079	21:49
Air Blank	0.000	21:49
Subject Sample # 1	0.235	21:50
Breath Volume	2.847	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:52
Subject Sample # 2	0.233	21:53
Breath Volume	1.617	21:53
Air Blank	0.000	21:54
Control Test	0.080	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 11/05/2018 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BRIGGS,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 11/04/2018 02:12:55

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:07	
Control Test	0.077	02:07	
Air Blank	0.000	02:08	
Subject Sample	# 1 0.187	02:08	
Breath Volume	2.765	02:08	
Air Blank	0.000	02:09	
Air Blank	0.000	02:10	
Subject Sample	# 2 0.189	02:11	
Breath Volume	2.648	02:11	
Air Blank	0.000	02:12	
Control Test	0.073	02:12	Control Outside Tolerance
Air Blank	0.000	02:12	

Data Download Date/Time: 11/05/2018 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,BRUCE,L
BREATH TEST OPERATOR AGENCY: St. Augustine Beach
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/05/2018 01:16:55

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.078	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.184	01:12
Breath Volume	1.878	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.170	01:15
Breath Volume	1.570	01:15
Air Blank	0.000	01:15
Control Test	0.078	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 11/05/2018 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADO,RYAN,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RADO,RYAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: SHELHART

DATE/TIME OF TEST: 11/07/2018 00:05:26

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.078	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.140	00:00
Breath Volume	1.523	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:02
Subject Sample # 2	0.140	00:03
Breath Volume	1.265	00:03
Air Blank	0.000	00:04
Control Test	0.079	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 12/09/2018 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,SCOTT,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLER,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: HALE

DATE/TIME OF TEST: 11/07/2018 03:34:02

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.161	03:29
Breath Volume	3.164	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.156	03:32
Breath Volume	2.445	03:32
Air Blank	0.000	03:33
Control Test	0.078	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:34

Data Download Date/Time: 12/09/2018 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 11/09/2018 21:20:33

Diagnostics Check	OK	21:12
Air Blank	0.000	21:13
Control Test	0.078	21:13
Air Blank	0.000	21:14
Subject Sample # 1	0.060	21:14
Breath Volume	1.324	21:14
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.062	21:19
Breath Volume	2.046	21:19
Air Blank	0.000	21:19
Control Test	0.079	21:19
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 12/09/2018 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JOHNATHAN,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BROWN,JOHNATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 11/15/2018 02:46:45

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.079	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.246	02:42
Breath Volume	1.894	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.257	02:45
Breath Volume	2.042	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 12/09/2018 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: HOLBROOK

DATE/TIME OF TEST: 11/16/2018 02:22:20

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.187	02:17
Breath Volume	1.445	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:19
Subject Sample # 2	0.192	02:20
Breath Volume	1.523	02:20
Air Blank	0.000	02:21
Control Test	0.077	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 12/09/2018 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELOIT,BRIAN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 11/18/2018 01:19:07

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.079	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.121	01:14
Breath Volume	3.355	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.116	01:17
Breath Volume	3.273	01:17
Air Blank	0.000	01:18
Control Test	0.078	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 12/09/2018 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELOIT,BRIAN,W
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 11/18/2018 01:26:17

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:23	
Control Test	0.078	01:24	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:24	Subject Test Refused
Breath Volume	0.000	01:24	
Air Blank	0.000	01:25	
Control Test	0.078	01:25	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

Data Download Date/Time: 12/09/2018 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELOIT,BRIAN,W
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: NIEVES,VICTOR
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 12:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: HILLYER

DATE/TIME OF TEST: 11/18/2018 01:57:35

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.078	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.224	01:52
Breath Volume	3.234	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.223	01:56
Breath Volume	3.324	01:56
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 12/09/2018 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELOIT,BRIAN,W
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: NELSON,ROBERT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 11/18/2018 04:02:55

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.078	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.147	03:58
Breath Volume	3.195	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:00
Subject Sample # 2	0.145	04:01
Breath Volume	2.710	04:01
Air Blank	0.000	04:02
Control Test	0.078	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 12/09/2018 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIRSHIKOVA,ALINA,A
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BEATTIE,ISALAH
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: LINK

DATE/TIME OF TEST: 11/18/2018 23:45:00

Diagnostics Check	OK	23:38
Air Blank	0.000	23:39
Control Test	0.078	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.208	23:40
Breath Volume	1.226	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.210	23:43
Breath Volume	2.531	23:43
Air Blank	0.000	23:44
Control Test	0.078	23:44
Air Blank	0.000	23:44
Diagnostics Check	OK	23:45

Data Download Date/Time: 12/09/2018 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,BRANDON,K
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: BARANOV

DATE/TIME OF TEST: 11/19/2018 16:22:18

Diagnostics Check	OK	16:16
Air Blank	0.000	16:16
Control Test	0.079	16:16
Air Blank	0.000	16:17
Subject Sample # 1	0.173	16:17
Breath Volume	2.531	16:17
Air Blank	0.000	16:18
Air Blank	0.000	16:20
Subject Sample # 2	0.166	16:20
Breath Volume	1.726	16:20
Air Blank	0.000	16:21
Control Test	0.078	16:21
Air Blank	0.000	16:22
Diagnostics Check	OK	16:22

Data Download Date/Time: 12/09/2018 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PANILAG,SAMUEL,C
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: PANILAG,SAMUEL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: VICKERY

DATE/TIME OF TEST: 11/24/2018 19:11:19

Diagnostics Check	OK	19:04
Air Blank	0.000	19:05
Control Test	0.078	19:05
Air Blank	0.000	19:05
Subject Sample # 1	0.203	19:06
Breath Volume	1.363	19:06
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample # 2	0.188	19:09
Breath Volume	1.914	19:09
Air Blank	0.000	19:10
Control Test	0.077	19:10
Air Blank	0.000	19:11
Diagnostics Check	OK	19:11

Data Download Date/Time: 12/09/2018 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADO,RYAN,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RADO,RYAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 11/26/2018 03:32:58

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.209	03:27
Breath Volume	3.160	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.191	03:31
Breath Volume	2.839	03:31
Air Blank	0.000	03:32
Control Test	0.078	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 12/09/2018 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENNESSEY,BRANDON,L
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: COOPER,NICHOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 11/26/2018 17:23:54

Diagnostics Check	OK	17:20	
Air Blank	0.000	17:20	
Control Test	0.078	17:21	
Air Blank	0.000	17:21	
Subject Sample	# 1	17:22	Subject Test Refused
Breath Volume	0.000	17:22	
Air Blank	0.000	17:22	
Control Test	0.079	17:23	
Air Blank	0.000	17:23	
Diagnostics Check	OK	17:23	

Data Download Date/Time: 12/09/2018 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 11/28/2018 21:04:13

Diagnostics Check	OK	20:54
Air Blank	0.000	20:54
Control Test	0.078	20:55
Air Blank	0.000	20:55
Subject Sample # 1	0.203	20:56
Breath Volume	3.175	20:56
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample # 2	0.179	20:59
Breath Volume	1.316	20:59
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 3	0.192	21:02
Breath Volume	2.445	21:02
Air Blank	0.000	21:03
Control Test	0.077	21:03
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

Data Download Date/Time: 12/09/2018 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALLOUN,SAMUEL,C
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BALLOUN,SAMUEL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: KLEIN

DATE/TIME OF TEST: 11/29/2018 16:48:42

Diagnostics Check	OK	16:40
Air Blank	0.000	16:41
Control Test	0.079	16:41
Air Blank	0.000	16:41
Subject Sample # 1	0.364	16:43
Breath Volume	1.742	16:43
Air Blank	0.000	16:44
Air Blank	0.000	16:45
Subject Sample # 2	0.346	16:47
Breath Volume	1.605	16:47
Air Blank	0.000	16:47
Control Test	0.078	16:48
Air Blank	0.000	16:48
Diagnostics Check	OK	16:48

Data Download Date/Time: 12/09/2018 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: KIGHT

DATE/TIME OF TEST: 11/29/2018 17:04:42

Diagnostics Check	OK	16:57
Air Blank	0.000	16:57
Control Test	0.079	16:58
Air Blank	0.000	16:58
Subject Sample # 1	0.030	17:00
Breath Volume	1.230	17:00
Air Blank	0.000	17:00
Air Blank	0.000	17:02
Subject Sample # 2	0.024	17:03
Breath Volume	1.679	17:03
Air Blank	0.000	17:03
Control Test	0.077	17:04
Air Blank	0.000	17:04
Diagnostics Check	OK	17:04

Data Download Date/Time: 12/09/2018 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMPE,CORTLAND,B
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: LAMPE,CORTLAND
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: GARRON

DATE/TIME OF TEST: 11/04/2018 00:26:39

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.221	00:21
Breath Volume	2.867	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.204	00:25
Breath Volume	1.390	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 11/27/2018 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMPE,CORTLAND,B
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: CULVER,K
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 11/04/2018 02:40:10

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.247	02:32
Breath Volume	3.304	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.226	02:35
Breath Volume	1.855	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:37
Subject Sample # 3	0.249	02:38
Breath Volume	3.425	02:38
Air Blank	0.000	02:39
Control Test	0.081	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 11/27/2018 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONLEY, SHAWNA, A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: BIONELLI, JARED
ARRESTING OFFICER AGENCY: Holly Hills PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 14:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:39

SUBJECT'S LAST NAME: ROSS JR

DATE/TIME OF TEST: 11/04/2018 16:05:55

Diagnostics Check	OK	15:59
Air Blank	0.000	16:00
Control Test	0.080	16:00
Air Blank	0.000	16:00
Subject Sample # 1	0.243	16:01
Breath Volume	3.449	16:01
Air Blank	0.000	16:02
Air Blank	0.000	16:03
Subject Sample # 2	0.231	16:04
Breath Volume	2.597	16:04
Air Blank	0.000	16:05
Control Test	0.080	16:05
Air Blank	0.000	16:05
Diagnostics Check	OK	16:05

Data Download Date/Time: 11/27/2018 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: LAWRENCE,TYLER
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: BRUNETTI

DATE/TIME OF TEST: 11/04/2018 22:32:34

Diagnostics Check	OK	22:25
Air Blank	0.000	22:25
Control Test	0.081	22:26
Air Blank	0.000	22:26
Subject Sample # 1	0.222	22:27
Breath Volume	1.906	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.212	22:30
Breath Volume	1.523	22:30
Air Blank	0.000	22:31
Control Test	0.079	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 11/27/2018 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: RIDDELL,JON
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: KRANZ

DATE/TIME OF TEST: 11/11/2018 22:41:46

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.082	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.201	22:37
Breath Volume	1.546	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.213	22:40
Breath Volume	2.257	22:40
Air Blank	0.000	22:40
Control Test	0.081	22:41
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 11/27/2018 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNNE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: LAWRENCE,TYLER
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 11/30/2018 02:07:18

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.082	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:05	Subject Test Refused
Breath Volume	0.671	02:05	
Air Blank	0.000	02:06	
Control Test	0.082	02:06	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 12/27/2018 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: FERREIRA

DATE/TIME OF TEST: 11/02/2018 22:40:09

Diagnostics Check	OK	22:32
Air Blank	0.000	22:33
Control Test	0.081	22:33
Air Blank	0.000	22:34
Subject Sample # 1	0.201	22:35
Breath Volume	1.832	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.192	22:38
Breath Volume	2.085	22:38
Air Blank	0.000	22:39
Control Test	0.081	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 11/14/2018 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JAY
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: GONZALEZ TOBAR

DATE/TIME OF TEST: 11/03/2018 00:33:14

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.082	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.208	00:28
Breath Volume	2.097	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.204	00:31
Breath Volume	2.660	00:31
Air Blank	0.000	00:32
Control Test	0.081	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 11/14/2018 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JASON
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 11/03/2018 05:31:12

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.083	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.154	05:26
Breath Volume	1.402	05:26
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.147	05:29
Breath Volume	1.394	05:29
Air Blank	0.000	05:30
Control Test	0.081	05:30
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 11/14/2018 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LIPARI,M
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:49

SUBJECT'S LAST NAME: SPATES

DATE/TIME OF TEST: 11/03/2018 08:19:39

Diagnostics Check	OK	08:13
Air Blank	0.000	08:13
Control Test	0.081	08:14
Air Blank	0.000	08:14
Subject Sample # 1	0.072	08:15
Breath Volume	2.812	08:15
Air Blank	0.000	08:15
Air Blank	0.000	08:17
Subject Sample # 2	0.073	08:18
Breath Volume	2.335	08:18
Air Blank	0.000	08:18
Control Test	0.083	08:19
Air Blank	0.000	08:19
Diagnostics Check	OK	08:19

Data Download Date/Time: 11/14/2018 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 11/04/2018 02:10:25

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.082	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.189	02:06
Breath Volume	1.582	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.186	02:08
Breath Volume	1.457	02:08
Air Blank	0.000	02:09
Control Test	0.082	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 11/14/2018 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: MCRAE

DATE/TIME OF TEST: 11/04/2018 03:11:15

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.083	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.164	03:05
Breath Volume	1.730	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.156	03:09
Breath Volume	1.386	03:09
Air Blank	0.000	03:10
Control Test	0.082	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 11/14/2018 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: ARENTZ

DATE/TIME OF TEST: 11/05/2018 00:51:24

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.082	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.190	00:46
Breath Volume	4.082	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.187	00:49
Breath Volume	3.699	00:49
Air Blank	0.000	00:50
Control Test	0.082	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 11/14/2018 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: COLLADO

DATE/TIME OF TEST: 11/05/2018 03:53:16

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.082	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.175	03:48
Breath Volume	1.617	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.170	03:51
Breath Volume	1.109	03:51
Air Blank	0.000	03:52
Control Test	0.081	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 11/14/2018 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOUGHRAN,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: TEXIDOR

DATE/TIME OF TEST: 11/10/2018 04:04:24

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.082	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.150	03:59
Breath Volume	3.152	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.149	04:02
Breath Volume	3.023	04:02
Air Blank	0.000	04:03
Control Test	0.082	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 11/14/2018 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: DILLING

DATE/TIME OF TEST: 11/10/2018 20:03:58

Diagnostics Check	OK	19:57
Air Blank	0.000	19:58
Control Test	0.082	19:58
Air Blank	0.000	19:58
Subject Sample # 1	0.199	19:59
Breath Volume	1.523	19:59
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.183	20:02
Breath Volume	1.191	20:02
Air Blank	0.000	20:03
Control Test	0.082	20:03
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

Data Download Date/Time: 11/14/2018 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: DOLISON

DATE/TIME OF TEST: 11/10/2018 23:07:30

Diagnostics Check	OK	22:56	
Air Blank	0.000	22:56	
Control Test	0.082	22:56	
Air Blank	0.000	22:57	
Subject Sample	# 1	23:00	No Sample Provided
Breath Volume	0.000	23:00	
Air Blank	0.000	23:00	
Air Blank	0.000	23:02	
Subject Sample	# 2	0.159 23:06	Volume Not Met
Breath Volume	0.019	23:06	
Air Blank	0.000	23:06	
Control Test	0.082	23:06	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

Data Download Date/Time: 11/14/2018 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SGT BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 11/11/2018 03:15:01

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.082	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.172	03:10
Breath Volume	3.773	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.171	03:13
Breath Volume	3.753	03:13
Air Blank	0.000	03:14
Control Test	0.082	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:15

Data Download Date/Time: 11/14/2018 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: MARAGH

DATE/TIME OF TEST: 11/13/2018 00:08:46

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.082	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.126	00:03
Breath Volume	1.187	00:03
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.123	00:07
Breath Volume	1.445	00:07
Air Blank	0.000	00:07
Control Test	0.083	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 11/14/2018 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BROCK,NATHAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CORDERO

DATE/TIME OF TEST: 11/13/2018 03:04:58

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.082	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.000	03:00
Breath Volume	1.550	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.000	03:03
Breath Volume	1.226	03:03
Air Blank	0.000	03:04
Control Test	0.083	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 11/14/2018 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: TAMBURINI

DATE/TIME OF TEST: 11/13/2018 23:54:18

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.082	23:48
Air Blank	0.000	23:48
Subject Sample # 1	0.181	23:49
Breath Volume	1.867	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.187	23:52
Breath Volume	2.000	23:52
Air Blank	0.000	23:53
Control Test	0.082	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 11/14/2018 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILLIAMS,ERIC
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 09:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:02

SUBJECT'S LAST NAME: AGUILAR VILLATORO

DATE/TIME OF TEST: 11/14/2018 10:34:33

Diagnostics Check	OK	10:27
Air Blank	0.000	10:28
Control Test	0.082	10:28
Air Blank	0.000	10:29
Subject Sample # 1	0.257	10:29
Breath Volume	1.992	10:29
Air Blank	0.000	10:30
Air Blank	0.000	10:32
Subject Sample # 2	0.240	10:33
Breath Volume	1.496	10:33
Air Blank	0.000	10:33
Control Test	0.082	10:33
Air Blank	0.000	10:34
Diagnostics Check	OK	10:34

Data Download Date/Time: 11/14/2018 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BAKER,MICHAEL
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 11/16/2018 03:37:11

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.148	03:32
Breath Volume	2.441	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.134	03:35
Breath Volume	1.382	03:35
Air Blank	0.000	03:36
Control Test	0.082	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 12/26/2018 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAULS,LANE
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 11/18/2018 22:19:48

Diagnostics Check	OK	22:12
Air Blank	0.000	22:13
Control Test	0.082	22:13
Air Blank	0.000	22:13
Subject Sample # 1	0.147	22:15
Breath Volume	1.421	22:15
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.143	22:18
Breath Volume	1.402	22:18
Air Blank	0.000	22:18
Control Test	0.082	22:19
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 12/26/2018 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOUGHRAN,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: COLON ALVAREZ

DATE/TIME OF TEST: 11/19/2018 02:10:49

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.082	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.082	02:06
Breath Volume	1.578	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:08
Subject Sample # 2	0.085	02:09
Breath Volume	1.714	02:09
Air Blank	0.000	02:09
Control Test	0.082	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 12/26/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BRAGA,LEONARDO
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 07:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 11/20/2018 10:33:45

Diagnostics Check	OK	10:24
Air Blank	0.000	10:24
Control Test	0.083	10:25
Air Blank	0.000	10:25
Subject Sample # 1	0.212	10:28
Breath Volume	1.195	10:28
Air Blank	0.000	10:28
Air Blank	0.000	10:30
Subject Sample # 2	0.219	10:32
Breath Volume	1.132	10:32
Air Blank	0.000	10:32
Control Test	0.082	10:33
Air Blank	0.000	10:33
Diagnostics Check	OK	10:33

Data Download Date/Time: 12/26/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERNICE,GREGORY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: PERNICE

DATE/TIME OF TEST: 11/22/2018 00:19:29

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.081	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.138	00:15
Breath Volume	2.003	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.136	00:18
Breath Volume	1.929	00:18
Air Blank	0.000	00:18
Control Test	0.082	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 12/26/2018 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BARTLETT,AMY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: SHRACK

DATE/TIME OF TEST: 11/22/2018 01:16:06

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:00	
Control Test	0.081	01:01	
Air Blank	0.000	01:01	
Subject Sample	# 1 0.234	01:04	Volume Not Met
Breath Volume	0.582	01:04	
Air Blank	0.000	01:05	
Air Blank	0.000	01:07	
Subject Sample	# 2 0.270	01:10	No .020 agreement
Breath Volume	1.539	01:10	
Air Blank	0.000	01:10	
Air Blank	0.000	01:12	
Subject Sample	# 3 0.243	01:14	No .020 agreement
Breath Volume	1.488	01:14	
Air Blank	0.000	01:15	
Control Test	0.081	01:15	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 12/26/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BARTLETT,AMY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: SHRACK

DATE/TIME OF TEST: 11/22/2018 02:35:12

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.082	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.221	02:30
Breath Volume	1.300	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.230	02:33
Breath Volume	1.781	02:33
Air Blank	0.000	02:34
Control Test	0.081	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 12/26/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: D'ALESSANDRO

DATE/TIME OF TEST: 11/22/2018 06:13:22

Diagnostics Check	OK	05:57	
Air Blank	0.000	05:57	
Control Test	0.082	05:58	
Air Blank	0.000	05:58	
Subject Sample	# 1 0.143	06:00	
Breath Volume	1.296	06:00	
Air Blank	0.000	06:01	
Air Blank	0.000	06:03	
Subject Sample	# 2	06:06	No Sample Provided
Breath Volume	0.000	06:06	
Air Blank	0.000	06:06	
Air Blank	0.000	06:08	
Subject Sample	# 3	06:12	No Sample Provided
Breath Volume	0.000	06:12	
Air Blank	0.000	06:12	
Control Test	0.082	06:12	
Air Blank	0.000	06:13	
Diagnostics Check	OK	06:13	

Data Download Date/Time: 12/26/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RODRIGUEZ,JAMES
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 11/24/2018 02:13:36

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.082	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.219	02:09
Breath Volume	2.140	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.216	02:12
Breath Volume	1.917	02:12
Air Blank	0.000	02:12
Control Test	0.082	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 12/26/2018 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: NYHUIS

DATE/TIME OF TEST: 11/24/2018 05:15:04

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.082	05:09
Air Blank	0.000	05:09
Subject Sample # 1	0.188	05:10
Breath Volume	1.468	05:10
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.190	05:13
Breath Volume	1.996	05:13
Air Blank	0.000	05:14
Control Test	0.081	05:14
Air Blank	0.000	05:14
Diagnostics Check	OK	05:15

Data Download Date/Time: 12/26/2018 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: GUTIERREZ REYES

DATE/TIME OF TEST: 11/25/2018 00:39:36

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.082	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.278	00:35
Breath Volume	1.902	00:35
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.269	00:38
Breath Volume	1.996	00:38
Air Blank	0.000	00:38
Control Test	0.081	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 12/26/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: RIVAS-ICABALZETA

DATE/TIME OF TEST: 11/25/2018 04:50:28

Diagnostics Check	OK	04:43
Air Blank	0.000	04:44
Control Test	0.083	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.263	04:45
Breath Volume	1.347	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.263	04:48
Breath Volume	2.417	04:48
Air Blank	0.000	04:49
Control Test	0.082	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 12/26/2018 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOUGHRAN,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: POZO CEDENO

DATE/TIME OF TEST: 11/25/2018 05:09:07

Diagnostics Check	OK	05:02
Air Blank	0.000	05:03
Control Test	0.082	05:03
Air Blank	0.000	05:03
Subject Sample # 1	0.030	05:04
Breath Volume	2.894	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.030	05:07
Breath Volume	3.242	05:07
Air Blank	0.000	05:08
Control Test	0.083	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 12/26/2018 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TRABUE,THOMAS
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 11/26/2018 01:09:24

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.082	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.164	01:04
Breath Volume	3.488	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.162	01:07
Breath Volume	3.078	01:07
Air Blank	0.000	01:08
Control Test	0.082	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 12/26/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PENA,RYAN
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: CARPINONE

DATE/TIME OF TEST: 11/26/2018 20:55:25

Diagnostics Check	OK	20:49
Air Blank	0.000	20:49
Control Test	0.082	20:50
Air Blank	0.000	20:50
Subject Sample # 1	0.000	20:51
Breath Volume	2.648	20:51
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.000	20:54
Breath Volume	2.246	20:54
Air Blank	0.000	20:54
Control Test	0.082	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 12/26/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/28/2018 02:56:52

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.082	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.205	02:51
Breath Volume	1.449	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.212	02:55
Breath Volume	1.628	02:55
Air Blank	0.000	02:55
Control Test	0.082	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 12/26/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MORALES JR

DATE/TIME OF TEST: 11/29/2018 01:22:04

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.083	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.000	01:17
Breath Volume	2.035	01:17
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.000	01:20
Breath Volume	1.675	01:20
Air Blank	0.000	01:21
Control Test	0.083	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:22

Data Download Date/Time: 12/26/2018 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 11/29/2018 03:47:01

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.083	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.199	03:40
Breath Volume	1.503	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.191	03:45
Breath Volume	1.343	03:45
Air Blank	0.000	03:46
Control Test	0.082	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:47

Data Download Date/Time: 12/26/2018 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SOUZA,MAYKOLL
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 11:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: MCHALE

DATE/TIME OF TEST: 11/30/2018 00:31:29

Diagnostics Check	OK	00:24
Air Blank	0.000	00:25
Control Test	0.082	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.214	00:27
Breath Volume	2.234	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.210	00:30
Breath Volume	1.503	00:30
Air Blank	0.000	00:30
Control Test	0.082	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 12/26/2018 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HARRIS,MILTON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CARDENAS

DATE/TIME OF TEST: 11/30/2018 04:33:01

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:22	
Control Test	0.082	04:22	
Air Blank	0.000	04:22	
Subject Sample	# 1	04:26	No Sample Provided
Breath Volume	0.000	04:26	
Air Blank	0.000	04:26	
Air Blank	0.000	04:28	
Subject Sample	# 2	04:31	No Sample Provided
Breath Volume	0.000	04:31	
Air Blank	0.000	04:32	
Control Test	0.083	04:32	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 12/26/2018 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE, GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE, GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: ADMIN
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 11/08/2018 13:53:06

Diagnostics Check	OK	13:47
Air Blank	0.000	13:47
Control Test	0.080	13:48
Air Blank	0.000	13:48
Subject Sample # 1	0.000	13:48
Breath Volume	1.976	13:48
Air Blank	0.000	13:49
Air Blank	0.000	13:51
Subject Sample # 2	0.000	13:51
Breath Volume	2.074	13:51
Air Blank	0.000	13:52
Control Test	0.080	13:52
Air Blank	0.000	13:53
Diagnostics Check	OK	13:53

Data Download Date/Time: 11/26/2018 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER, WILLIAM,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TUCKER, WILLIAM
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: JAIMES-VARELA

DATE/TIME OF TEST: 11/04/2018 01:33:45

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.084	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.212	01:28
Breath Volume	3.445	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.207	01:32
Breath Volume	2.253	01:32
Air Blank	0.000	01:32
Control Test	0.082	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 11/26/2018 07:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 11/05/2018 02:29:47

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.083	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.105	02:25
Breath Volume	1.597	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.117	02:28
Breath Volume	1.503	02:28
Air Blank	0.000	02:28
Control Test	0.084	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 11/26/2018 07:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 11/12/2018 02:04:55

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.085	01:57	
Air Blank	0.000	01:57	
Subject Sample	# 1	02:00	No Sample Provided
Breath Volume	0.000	02:00	
Air Blank	0.000	02:01	
Air Blank	0.000	02:03	
Subject Sample	# 2	02:03	Subject Test Refused
Breath Volume	0.000	02:03	
Air Blank	0.000	02:03	
Control Test	0.083	02:04	
Air Blank	0.000	02:04	
Diagnostics Check	OK	02:04	

Data Download Date/Time: 11/26/2018 07:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO, CLIFFORD, A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO, CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: WINSTON

DATE/TIME OF TEST: 11/14/2018 21:10:25

Diagnostics Check	OK	20:55	
Air Blank	0.000	20:55	
Control Test	0.084	20:56	
Air Blank	0.000	20:56	
Subject Sample	# 1 0.274	20:57	
Breath Volume	1.285	20:57	
Air Blank	0.000	20:58	
Air Blank	0.000	21:00	
Subject Sample	# 2 0.158	21:03	Volume Not Met
Breath Volume	0.148	21:03	
Air Blank	0.000	21:04	
Air Blank	0.000	21:05	
Subject Sample	# 3	21:09	No Sample Provided
Breath Volume	0.000	21:09	
Air Blank	0.000	21:09	
Control Test	0.083	21:09	
Air Blank	0.000	21:10	
Diagnostics Check	OK	21:10	

Data Download Date/Time: 11/26/2018 07:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: WINSTON

DATE/TIME OF TEST: 11/14/2018 21:46:38

Diagnostics Check	OK	21:31	
Air Blank	0.000	21:32	
Control Test	0.084	21:32	
Air Blank	0.000	21:33	
Subject Sample	# 1 0.266	21:35	
Breath Volume	1.460	21:35	
Air Blank	0.000	21:35	
Air Blank	0.000	21:37	
Subject Sample	# 2 0.215	21:40	Volume Not Met
Breath Volume	0.789	21:40	
Air Blank	0.000	21:41	
Air Blank	0.000	21:43	
Subject Sample	# 3 0.266	21:44	
Breath Volume	1.566	21:44	
Air Blank	0.000	21:45	
Control Test	0.082	21:45	
Air Blank	0.000	21:46	
Diagnostics Check	OK	21:46	

Data Download Date/Time: 11/26/2018 07:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER, WILLIAM,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: DELIMA, EDUARDO
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: LOPEZ-FELICIANO

DATE/TIME OF TEST: 11/17/2018 04:25:11

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.084	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.203	04:20
Breath Volume	2.312	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.198	04:23
Breath Volume	1.554	04:23
Air Blank	0.000	04:24
Control Test	0.083	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 11/26/2018 07:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER, WILLAM,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TUCKER, WILLIAM
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: TRUJILLO-MARTINEZ

DATE/TIME OF TEST: 11/22/2018 00:15:51

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.084	00:10
Air Blank	0.000	00:10
Subject Sample # 1	0.184	00:11
Breath Volume	3.285	00:11
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.182	00:14
Breath Volume	3.085	00:14
Air Blank	0.000	00:14
Control Test	0.084	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 11/26/2018 07:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SANTOS,CARLOS
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: DEORTA

DATE/TIME OF TEST: 11/24/2018 05:49:23

Diagnostics Check	OK	05:43
Air Blank	0.000	05:43
Control Test	0.084	05:44
Air Blank	0.000	05:44
Subject Sample # 1	0.027	05:44
Breath Volume	3.847	05:44
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.027	05:47
Breath Volume	3.437	05:47
Air Blank	0.000	05:48
Control Test	0.084	05:48
Air Blank	0.000	05:49
Diagnostics Check	OK	05:49

Data Download Date/Time: 11/26/2018 07:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO, CLIFFORD, A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO, LIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 11/24/2018 20:18:05

Diagnostics Check	OK	20:11
Air Blank	0.000	20:12
Control Test	0.085	20:12
Air Blank	0.000	20:12
Subject Sample # 1	0.266	20:13
Breath Volume	3.320	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:16
Subject Sample # 2	0.263	20:16
Breath Volume	3.277	20:16
Air Blank	0.000	20:17
Control Test	0.083	20:17
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

Data Download Date/Time: 11/26/2018 07:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 11/25/2018 06:30:35

Diagnostics Check	OK	06:29	
Air Blank	0.000	06:29	
Control Test	0.084	06:29	
Air Blank		06:30	RFI Detect
Air Blank	0.000	06:30	

Data Download Date/Time: 11/26/2018 07:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 11/25/2018 06:36:17

Diagnostics Check	OK	06:33	
Air Blank	0.000	06:33	
Control Test	0.084	06:33	
Air Blank	0.000	06:34	
Subject Sample	# 1	06:34	RFI Detect
Breath Volume	0.457	06:34	
Air Blank	0.000	06:35	
Control Test	0.084	06:35	
Air Blank	0.000	06:36	
Diagnostics Check	OK	06:36	

Data Download Date/Time: 11/26/2018 07:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 11/25/2018 06:45:13

Diagnostics Check	OK	06:38
Air Blank	0.000	06:39
Control Test	0.084	06:39
Air Blank	0.000	06:40
Subject Sample # 1	0.156	06:40
Breath Volume	2.968	06:40
Air Blank	0.000	06:41
Air Blank	0.000	06:43
Subject Sample # 2	0.159	06:43
Breath Volume	3.203	06:43
Air Blank	0.000	06:44
Control Test	0.080	06:44
Air Blank	0.000	06:45
Diagnostics Check	OK	06:45

Data Download Date/Time: 11/26/2018 07:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBIN,RICHARD,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: RAIHALA,JON
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: THIESS

DATE/TIME OF TEST: 11/02/2018 02:10:34

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:08	
Control Test	0.081	02:08	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:09	
Control Test	0.081	02:09	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

Data Download Date/Time: 11/27/2018 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, MICHAEL, D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: STAGE, NICOLAS
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: CAMADECA

DATE/TIME OF TEST: 11/03/2018 18:31:46

Diagnostics Check	OK	18:25
Air Blank	0.000	18:26
Control Test	0.081	18:26
Air Blank	0.000	18:27
Subject Sample # 1	0.000	18:27
Breath Volume	2.011	18:27
Air Blank	0.000	18:28
Air Blank	0.000	18:30
Subject Sample # 2	0.000	18:30
Breath Volume	1.515	18:30
Air Blank	0.000	18:30
Control Test	0.082	18:31
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

Data Download Date/Time: 11/27/2018 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEARSON,ROBERT,B
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: DEVANE,CIARA
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: CWALINA

DATE/TIME OF TEST: 11/11/2018 06:34:24

Diagnostics Check	OK	06:28
Air Blank	0.000	06:28
Control Test	0.081	06:29
Air Blank	0.000	06:29
Subject Sample # 1	0.158	06:30
Breath Volume	1.875	06:30
Air Blank	0.000	06:30
Air Blank	0.000	06:32
Subject Sample # 2	0.156	06:32
Breath Volume	2.507	06:32
Air Blank	0.000	06:33
Control Test	0.080	06:33
Air Blank	0.000	06:34
Diagnostics Check	OK	06:34

Data Download Date/Time: 11/27/2018 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARMINGTON, WILLIAM, G
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: DOUGLAS, THOMAS
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/20/2018 02:15:10

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.172	02:10
Breath Volume	1.839	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.162	02:13
Breath Volume	1.710	02:13
Air Blank	0.000	02:14
Control Test	0.081	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 11/27/2018 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBIN,RICHARD,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: DOUGLAS,THOMAS
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: LEGAULT

DATE/TIME OF TEST: 11/25/2018 04:11:15

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.081	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.158	04:06
Breath Volume	1.257	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.157	04:09
Breath Volume	1.339	04:09
Air Blank	0.000	04:10
Control Test	0.081	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 11/27/2018 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBIN,RICHARD,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: BEAULIEU,NICK
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: GRAU

DATE/TIME OF TEST: 11/26/2018 00:06:27

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.081	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.030	00:02
Breath Volume	2.636	00:02
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.030	00:05
Breath Volume	2.531	00:05
Air Blank	0.000	00:05
Control Test	0.082	00:05
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 11/27/2018 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001158

INSTRUMENT REGISTERED TO: PONCE INLET PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RYAN,WILLIAM,C

BREATH TEST OPERATOR AGENCY: Ponce Inlet PD

ARRESTING OFFICER: VALERIEN,MARK

ARRESTING OFFICER AGENCY: Ponce Inlet PD

LAST AGENCY INSPECTION: 10/01/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 814183

DRY GAS EXPIRY DATE: 07/26/2020

VIOLATION CODE: DUI

ARREST TIME: 18:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: NOVELLY

DATE/TIME OF TEST: 11/09/2018 19:38:27

Diagnostics Check	OK	19:32
Air Blank	0.000	19:32
Control Test	0.081	19:32
Air Blank	0.000	19:33
Subject Sample # 1	0.314	19:33
Breath Volume	4.523	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:36
Subject Sample # 2	0.314	19:36
Breath Volume	4.191	19:36
Air Blank	0.000	19:37
Control Test	0.081	19:37
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time: 11/15/2018 17:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRALSON,SAMUEL,E
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: MARSAN,NECOLE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: MEHLIN

DATE/TIME OF TEST: 11/09/2018 16:47:21

Diagnostics Check	OK	16:41
Air Blank	0.000	16:41
Control Test	0.081	16:42
Air Blank	0.000	16:42
Subject Sample # 1	0.000	16:43
Breath Volume	4.804	16:43
Air Blank	0.000	16:43
Air Blank	0.000	16:45
Subject Sample # 2	0.000	16:45
Breath Volume	4.660	16:45
Air Blank	0.000	16:46
Control Test	0.081	16:46
Air Blank	0.000	16:47
Diagnostics Check	OK	16:47

Data Download Date/Time: 12/04/2018 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUBAKER,KELLY,J
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: GODENZI,CAMERON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: CRESPO

DATE/TIME OF TEST: 11/23/2018 00:43:10

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:31	
Control Test	0.081	00:31	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:35	No Sample Provided
Breath Volume	0.000	00:35	
Air Blank	0.000	00:35	
Air Blank	0.000	00:37	
Subject Sample	# 2	0.203	00:38
Breath Volume	1.785	00:38	
Air Blank	0.000	00:39	
Air Blank	0.000	00:41	
Subject Sample	# 3	0.197	00:41
Breath Volume	3.136	00:41	
Air Blank	0.000	00:42	
Control Test	0.081	00:42	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 12/04/2018 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BELL,ETHAN,C

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: BELL,ETHAN

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02618080A2

DRY GAS EXPIRY DATE: 03/05/2020

VIOLATION CODE: DUI

ARREST TIME: 16:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 11/01/2018 17:45:53

Diagnostics Check	OK	17:41	
Air Blank	0.000	17:42	
Control Test	0.079	17:42	
Air Blank	0.000	17:42	
Subject Sample	# 1	17:44	Subject Test Refused
Breath Volume	0.000	17:44	
Air Blank	0.000	17:44	
Control Test	0.079	17:45	
Air Blank	0.000	17:45	
Diagnostics Check	OK	17:45	

Data Download Date/Time: 11/27/2018 21:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARPER,JIMMIE,L

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: KRUSE,KRIS

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02618080A2

DRY GAS EXPIRY DATE: 03/05/2020

VIOLATION CODE: DUI

ARREST TIME: 05:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: PERSUAD

DATE/TIME OF TEST: 11/11/2018 05:49:54

Diagnostics Check	OK	05:47	
Air Blank	0.000	05:47	
Control Test	0.079	05:47	
Air Blank	0.000	05:48	
Subject Sample	# 1	05:48	Subject Test Refused
Breath Volume	0.000	05:48	
Air Blank	0.000	05:48	
Control Test	0.079	05:49	
Air Blank	0.000	05:49	
Diagnostics Check	OK	05:49	

Data Download Date/Time: 11/27/2018 21:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: LADY LAKE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURYEA, MATTHEW,
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: DURYEA, MATTHEW
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:44

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/22/2018 16:32:43

Diagnostics Check	OK	16:17	
Air Blank	0.000	16:17	
Control Test	0.079	16:18	
Air Blank	0.000	16:18	
Subject Sample	# 1 0.136	16:20	
Breath Volume	1.691	16:20	
Air Blank	0.000	16:21	
Air Blank	0.000	16:23	
Subject Sample	# 2 0.129	16:26	Volume Not Met
Breath Volume	0.968	16:26	
Air Blank	0.000	16:27	
Air Blank	0.000	16:29	
Subject Sample	# 3 0.126	16:31	
Breath Volume	1.816	16:31	
Air Blank	0.000	16:31	
Control Test	0.074	16:32	Control Outside Tolerance
Air Blank	0.000	16:32	

Data Download Date/Time: 12/21/2018 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 279870
BREATH TEST OPERATOR: RODRIGUEZ,ROBERTO,	DRY GAS EXPIRY DATE: 02/24/2019
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)	VIOLATION CODE: BUI
ARRESTING OFFICER: HEDGEPEETH,AUSTIN	ARREST TIME: 23:18
ARRESTING OFFICER AGENCY: FFWCC (West Palm)	GENDER: Male
LAST AGENCY INSPECTION: 10/29/2018	OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: CARBRERA	DATE/TIME OF TEST: 11/23/2018 23:43:16
-------------------------------	--

Diagnostics Check	OK	23:40	
Air Blank	0.000	23:40	
Control Test	0.080	23:41	
Air Blank	0.000	23:41	
Subject Sample	# 1	23:41	Subject Test Refused
Breath Volume	0.000	23:41	
Air Blank	0.000	23:42	
Control Test	0.079	23:42	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:43	

Data Download Date/Time: 11/28/2018 18:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: CROCETTI

DATE/TIME OF TEST: 11/02/2018 23:25:04

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:22	
Control Test	0.078	23:23	
Air Blank	0.000	23:23	
Subject Sample	# 1	23:23	Subject Test Refused
Breath Volume	0.000	23:23	
Air Blank	0.000	23:24	
Control Test	0.079	23:24	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:25	

Data Download Date/Time: 11/07/2018 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: MAHON

DATE/TIME OF TEST: 11/07/2018 22:30:03

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.078	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.000	22:25
Breath Volume	1.296	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.000	22:28
Breath Volume	1.523	22:28
Air Blank	0.000	22:29
Control Test	0.078	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:30

Data Download Date/Time: 12/13/2018 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: TAMEZ

DATE/TIME OF TEST: 11/11/2018 00:18:13

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:07	
Control Test	0.078	00:07	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:11	No Sample Provided
Breath Volume	0.000	00:11	
Air Blank	0.000	00:11	
Air Blank	0.000	00:13	
Subject Sample	# 2	00:16	Subject Test Refused
Breath Volume	0.015	00:16	
Air Blank	0.000	00:17	
Control Test	0.080	00:17	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

Data Download Date/Time: 12/13/2018 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: DOMINGUEZ,JAYMES
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: CARMONA

DATE/TIME OF TEST: 11/16/2018 01:08:37

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.081	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.102	01:03
Breath Volume	1.265	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.092	01:07
Breath Volume	1.500	01:07
Air Blank	0.000	01:07
Control Test	0.078	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 12/13/2018 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: STEELE,SEAN
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: STANWAY

DATE/TIME OF TEST: 11/30/2018 03:10:33

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.171	03:06
Breath Volume	2.570	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.163	03:09
Breath Volume	2.832	03:09
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 12/13/2018 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MARTINEZ,E
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: TOHILL

DATE/TIME OF TEST: 11/29/2018 23:24:33

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.079	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.178	23:19
Breath Volume	1.488	23:19
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.173	23:23
Breath Volume	1.941	23:23
Air Blank	0.000	23:23
Control Test	0.079	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 12/04/2018 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,DAVID,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HARRIS,DAVID
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 11/05/2018 02:18:06

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.135	02:13
Breath Volume	2.601	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:15
Subject Sample # 2	0.146	02:16
Breath Volume	1.656	02:16
Air Blank	0.000	02:17
Control Test	0.078	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 12/02/2018 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,DAVID,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HERNANDEZ,LUIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 11/13/2018 19:47:05

Diagnostics Check	OK	19:35	
Air Blank	0.000	19:36	
Control Test	0.080	19:36	
Air Blank	0.000	19:36	
Subject Sample	# 1 0.127	19:40	Volume Not Met
Breath Volume	0.273	19:40	
Air Blank	0.000	19:40	
Air Blank	0.000	19:42	
Subject Sample	# 2 0.129	19:45	Volume Not Met
Breath Volume	0.015	19:45	
Air Blank	0.000	19:46	
Control Test	0.080	19:46	
Air Blank	0.000	19:47	
Diagnostics Check	OK	19:47	

Data Download Date/Time: 12/02/2018 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,DAVID,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HERNANDEZ,LUIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: STAMM

DATE/TIME OF TEST: 11/17/2018 00:22:27

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.218	00:17
Breath Volume	3.519	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.216	00:20
Breath Volume	3.402	00:20
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 12/02/2018 09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCK,AUSTIN,M
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BROCK,AUSTIN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 11:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:35

SUBJECT'S LAST NAME: PARHAM

DATE/TIME OF TEST: 11/17/2018 23:36:26

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.081	23:31	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:34	Subject Test Refused
Breath Volume	0.484	23:34	
Air Blank	0.000	23:35	
Control Test	0.081	23:35	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 12/02/2018 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAUGA,JOHN,E
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: PAUGA,JOHN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/23/2018 00:31:49

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.252	00:27
Breath Volume	1.316	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:29
Subject Sample # 2	0.260	00:30
Breath Volume	1.949	00:30
Air Blank	0.000	00:30
Control Test	0.079	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 12/02/2018 09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOHRER,M,
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BOHRER,MARK
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: CONOLEY

DATE/TIME OF TEST: 11/30/2018 03:18:01

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.081	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.180	03:13
Breath Volume	2.945	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.183	03:16
Breath Volume	2.968	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:18

Data Download Date/Time: 12/02/2018 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRIGES,JASON,J
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: HERRIGES,JASON
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: ADMIN
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: CAIN

DATE/TIME OF TEST: 11/27/2018 11:54:04

Diagnostics Check	OK	11:48
Air Blank	0.000	11:48
Control Test	0.082	11:48
Air Blank	0.000	11:49
Subject Sample # 1	0.026	11:49
Breath Volume	1.656	11:49
Air Blank	0.000	11:50
Air Blank	0.000	11:52
Subject Sample # 2	0.024	11:52
Breath Volume	1.812	11:52
Air Blank	0.000	11:53
Control Test	0.082	11:53
Air Blank	0.000	11:53
Diagnostics Check	OK	11:54

Data Download Date/Time: 12/04/2018 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,DUSTIN,D
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: HAVRANEK,ADAM
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 11/16/2018 02:43:37

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:40	
Control Test	0.080	02:40	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:42	
Control Test	0.080	02:43	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 11/21/2018 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: STREIGOLD,DAVID
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 11/12/2018 00:38:34

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.080	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.000	00:34
Breath Volume	2.433	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.000	00:37
Breath Volume	2.496	00:37
Air Blank	0.000	00:37
Control Test	0.079	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 11/29/2018 07:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: STREIGOLD,DAVID
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: ZUNUN DIAZ

DATE/TIME OF TEST: 11/18/2018 04:07:29

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.080	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.227	04:03
Breath Volume	2.304	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.232	04:05
Breath Volume	2.894	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 11/29/2018 07:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: STREIGOLD,DAVID
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: KEITH

DATE/TIME OF TEST: 11/19/2018 00:40:24

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.080	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.182	00:36
Breath Volume	2.429	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.187	00:38
Breath Volume	2.656	00:38
Air Blank	0.000	00:39
Control Test	0.079	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 11/29/2018 07:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICAH,L
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: THOMAS,MICAH
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: DOYLE

DATE/TIME OF TEST: 11/22/2018 23:24:54

Diagnostics Check	OK	23:18
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.000	23:20
Breath Volume	2.230	23:20
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.000	23:23
Breath Volume	1.152	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 11/29/2018 07:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: ANNESE,MORRIS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: WOMELSDORFF

DATE/TIME OF TEST: 11/25/2018 23:04:33

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.080	22:58
Air Blank	0.000	22:58
Subject Sample # 1	0.180	22:59
Breath Volume	1.492	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.190	23:03
Breath Volume	2.257	23:03
Air Blank	0.000	23:03
Control Test	0.080	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 11/29/2018 07:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES, STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES, STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 11/01/2018 19:54:46

Diagnostics Check	OK	19:47
Air Blank	0.000	19:48
Control Test	0.080	19:48
Air Blank	0.000	19:49
Subject Sample # 1	0.269	19:50
Breath Volume	1.246	19:50
Air Blank	0.000	19:50
Air Blank	0.000	19:52
Subject Sample # 2	0.261	19:53
Breath Volume	1.832	19:53
Air Blank	0.000	19:53
Control Test	0.079	19:54
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

Data Download Date/Time: 11/28/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WARREN,BILLY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 11/25/2018 01:30:49

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.000	01:26
Breath Volume	2.355	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.000	01:29
Breath Volume	2.441	01:29
Air Blank	0.000	01:29
Control Test	0.081	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 11/28/2018 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES, STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SCHEIDEGG, COREY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: CLIFF

DATE/TIME OF TEST: 11/30/2018 03:40:16

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.206	03:35
Breath Volume	1.507	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.206	03:38
Breath Volume	1.234	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 12/28/2018 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 11/01/2018 01:29:20

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.082	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.146	01:24
Breath Volume	2.417	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:26
Subject Sample # 2	0.141	01:27
Breath Volume	2.328	01:27
Air Blank	0.000	01:28
Control Test	0.081	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 11/15/2018 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAJEUNESSE,J,S
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: JOHNSON,M
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 10/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: GRALTON

DATE/TIME OF TEST: 11/01/2018 04:24:16

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.081	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.202	04:19
Breath Volume	3.757	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.189	04:22
Breath Volume	2.542	04:22
Air Blank	0.000	04:23
Control Test	0.081	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 11/15/2018 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAJEUNESSE,J,S
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: OSTERBERG,J
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 10/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 11/01/2018 06:00:21

Diagnostics Check	OK	05:52
Air Blank	0.000	05:52
Control Test	0.081	05:53
Air Blank	0.000	05:53
Subject Sample # 1	0.192	05:54
Breath Volume	1.304	05:54
Air Blank	0.000	05:54
Air Blank	0.000	05:56
Subject Sample # 2	0.175	05:58
Breath Volume	1.574	05:58
Air Blank	0.000	05:59
Control Test	0.081	05:59
Air Blank	0.000	06:00
Diagnostics Check	OK	06:00

Data Download Date/Time: 11/15/2018 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,TROY,P
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: DUBI,UZIEL
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 10/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: FOWLE

DATE/TIME OF TEST: 11/04/2018 04:52:05

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.081	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.171	04:47
Breath Volume	4.421	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:49
Subject Sample # 2	0.174	04:50
Breath Volume	4.085	04:50
Air Blank	0.000	04:51
Control Test	0.080	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 11/15/2018 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETTHAND,BREUNA,C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: LETTHAND,BREUNA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: DANDREA

DATE/TIME OF TEST: 11/04/2018 19:10:12

Diagnostics Check	OK	19:02
Air Blank	0.000	19:03
Control Test	0.082	19:03
Air Blank	0.000	19:04
Subject Sample # 1	0.000	19:05
Breath Volume	2.906	19:05
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample # 2	0.000	19:08
Breath Volume	1.632	19:08
Air Blank	0.000	19:09
Control Test	0.080	19:09
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

Data Download Date/Time: 11/15/2018 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,TROY,P
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: DUBI,UZIEL
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 10/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MIGUEL

DATE/TIME OF TEST: 11/05/2018 03:11:32

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.082	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.119	03:07
Breath Volume	1.347	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.120	03:10
Breath Volume	1.375	03:10
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 11/15/2018 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: SILVERIO,ALVIN
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 10/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: MYRICK

DATE/TIME OF TEST: 11/06/2018 00:47:56

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.082	00:41
Air Blank	0.000	00:42
Subject Sample # 1	0.127	00:43
Breath Volume	2.578	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.131	00:46
Breath Volume	2.632	00:46
Air Blank	0.000	00:47
Control Test	0.081	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 11/15/2018 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: FLORES,JERRY
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 10/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 11/09/2018 18:29:42

Diagnostics Check	OK	18:23
Air Blank	0.000	18:24
Control Test	0.082	18:24
Air Blank	0.000	18:24
Subject Sample # 1	0.000	18:25
Breath Volume	3.289	18:25
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 2	0.000	18:28
Breath Volume	3.316	18:28
Air Blank	0.000	18:28
Control Test	0.080	18:29
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

Data Download Date/Time: 11/15/2018 20:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: SANDBAKKEN,PAUL
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 11/17/2018 00:09:46

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:01	
Control Test	0.082	00:01	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:05	No Sample Provided
Breath Volume	0.000	00:05	
Air Blank	0.000	00:05	
Air Blank	0.000	00:07	
Subject Sample	# 2	00:08	Subject Test Refused
Breath Volume	0.000	00:08	
Air Blank	0.000	00:08	
Control Test	0.082	00:09	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:09	

Data Download Date/Time: 12/06/2018 00:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,TROY,P
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: NORGARD,BRYAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 11/17/2018 20:32:40

Diagnostics Check	OK	20:28	
Air Blank	0.000	20:29	
Control Test	0.081	20:29	
Air Blank	0.000	20:30	
Subject Sample	# 1	20:32	Sequence Aborted
Breath Volume	0.000	20:32	
Air Blank	0.000	20:32	

Data Download Date/Time: 12/06/2018 00:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,TROY,P
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: GONZALEZ,LUIS
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: IBARRA RODRIQUEZ

DATE/TIME OF TEST: 11/18/2018 00:41:40

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.081	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.233	00:37
Breath Volume	2.851	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.229	00:40
Breath Volume	3.734	00:40
Air Blank	0.000	00:40
Control Test	0.081	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 12/06/2018 00:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIJAGIC,SAMIR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ALIJAGIC,SAMIR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PETERSEN

DATE/TIME OF TEST: 11/21/2018 03:42:25

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:38	
Control Test	0.081	03:38	
Air Blank	0.000	03:39	
Subject Sample	# 1 0.193	03:40	
Breath Volume	3.246	03:40	
Air Blank		03:41	RFI Detect
Air Blank	0.000	03:41	
Control Test	0.080	03:41	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 12/06/2018 00:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: VINSON,HEATHER
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 11/22/2018 02:35:36

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:27	
Control Test	0.081	02:27	
Air Blank	0.000	02:28	
Subject Sample	# 1	02:31	No Sample Provided
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Air Blank	0.000	02:33	
Subject Sample	# 2	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:34	
Control Test	0.082	02:35	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 12/06/2018 00:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: MAHANEY,JASON
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SPEARS

DATE/TIME OF TEST: 11/22/2018 03:20:38

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.081	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.144	03:14
Breath Volume	3.535	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.146	03:19
Breath Volume	4.265	03:19
Air Blank	0.000	03:19
Control Test	0.081	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 12/06/2018 00:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: DIAZ,WILFREDO
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BLOCK

DATE/TIME OF TEST: 11/22/2018 03:56:35

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.081	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.110	03:52
Breath Volume	4.933	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.111	03:55
Breath Volume	4.144	03:55
Air Blank	0.000	03:55
Control Test	0.081	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 12/06/2018 00:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,TROY,P
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: MAHANEY,JASON
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: PRUETT

DATE/TIME OF TEST: 11/22/2018 19:30:30

Diagnostics Check	OK	19:23
Air Blank	0.000	19:23
Control Test	0.081	19:24
Air Blank	0.000	19:24
Subject Sample # 1	0.160	19:26
Breath Volume	1.484	19:26
Air Blank	0.000	19:26
Air Blank	0.000	19:28
Subject Sample # 2	0.169	19:28
Breath Volume	1.250	19:28
Air Blank	0.000	19:29
Control Test	0.082	19:29
Air Blank	0.000	19:30
Diagnostics Check	OK	19:30

Data Download Date/Time: 12/06/2018 00:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,TROY,P
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: GONZALEZ,LUIS
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: JOHANSEN

DATE/TIME OF TEST: 11/23/2018 00:04:34

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.081	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.000	00:00
Breath Volume	3.589	00:00
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 2	0.000	00:03
Breath Volume	3.808	00:03
Air Blank	0.000	00:03
Control Test	0.081	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 12/06/2018 00:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAJEUNESSE,J,S
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: JOHNSON,M
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: COWART DARLING

DATE/TIME OF TEST: 11/25/2018 02:36:33

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.082	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.000	02:32
Breath Volume	4.160	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.000	02:35
Breath Volume	3.269	02:35
Air Blank	0.000	02:35
Control Test	0.081	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 12/06/2018 00:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,TROY,P
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: CONNERS,CHRISTINA
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: DOOLEY

DATE/TIME OF TEST: 11/27/2018 20:36:12

Diagnostics Check	OK	20:24	
Air Blank	0.000	20:25	
Control Test	0.082	20:25	
Air Blank	0.000	20:25	
Subject Sample	# 1	20:28	No Sample Provided
Breath Volume	0.000	20:28	
Air Blank	0.000	20:29	
Air Blank	0.000	20:31	
Subject Sample	# 2	20:34	Volume Not Met
Breath Volume	0.019	20:34	
Air Blank	0.000	20:35	
Control Test	0.082	20:35	
Air Blank	0.000	20:36	
Diagnostics Check	OK	20:36	

Data Download Date/Time: 12/06/2018 00:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 11/28/2018 20:28:55

Diagnostics Check	OK	20:18
Air Blank	0.000	20:18
Control Test	0.081	20:19
Air Blank	0.000	20:19
Subject Sample # 1	0.148	20:21
Breath Volume	2.515	20:21
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.171	20:24
Breath Volume	2.101	20:24
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 3	0.174	20:27
Breath Volume	1.367	20:27
Air Blank	0.000	20:28
Control Test	0.082	20:28
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

Data Download Date/Time: 12/06/2018 00:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURICK,PETER,
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: BURICK,PETER
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 10/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 07:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:52

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 11/03/2018 08:19:09

Diagnostics Check	OK	08:12
Air Blank	0.000	08:13
Control Test	0.080	08:13
Air Blank	0.000	08:13
Subject Sample # 1	0.080	08:14
Breath Volume	2.917	08:14
Air Blank	0.000	08:15
Air Blank	0.000	08:17
Subject Sample # 2	0.080	08:17
Breath Volume	2.695	08:17
Air Blank	0.000	08:18
Control Test	0.080	08:18
Air Blank	0.000	08:19
Diagnostics Check	OK	08:19

Data Download Date/Time: 11/28/2018 04:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTZ,KATHRYN,N
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: WOELL,STEVEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: KNOUSE

DATE/TIME OF TEST: 11/04/2018 05:09:22

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:58	
Control Test	0.080	04:58	
Air Blank	0.000	04:59	
Subject Sample	# 1 0.167	05:02	Volume Not Met
Breath Volume	0.031	05:02	
Air Blank	0.000	05:03	
Air Blank	0.000	05:04	
Subject Sample	# 2	05:07	RFI Detect
Breath Volume	0.187	05:07	
Air Blank	0.000	05:08	
Control Test	0.079	05:08	
Air Blank	0.000	05:09	
Diagnostics Check	OK	05:09	

Data Download Date/Time: 11/28/2018 04:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: CASTELLANO,VINCENT
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 10/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 11/04/2018 21:36:30

Diagnostics Check	OK	21:29
Air Blank	0.000	21:29
Control Test	0.080	21:30
Air Blank	0.000	21:30
Subject Sample # 1	0.192	21:31
Breath Volume	2.937	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.180	21:34
Breath Volume	2.667	21:34
Air Blank	0.000	21:35
Control Test	0.078	21:35
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 11/28/2018 04:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROONEY,ZACHARY,D
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: DIEDESCH,LAURA
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 10/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: CORTEZ

DATE/TIME OF TEST: 11/11/2018 03:43:30

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.080	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.026	03:39
Breath Volume	1.410	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.023	03:41
Breath Volume	1.406	03:41
Air Blank	0.000	03:42
Control Test	0.078	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 11/28/2018 04:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: CASTELLANO,VINCENT
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 10/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: MARKLEW

DATE/TIME OF TEST: 11/13/2018 01:09:12

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.080	00:58
Air Blank	0.000	00:59
Subject Sample # 1	0.171	01:00
Breath Volume	1.167	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.149	01:04
Breath Volume	1.300	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:06
Subject Sample # 3	0.156	01:07
Breath Volume	1.894	01:07
Air Blank	0.000	01:08
Control Test	0.077	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 11/28/2018 04:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTZ,KATHRYN,N
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: GOTZ,KATHRYN
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 10/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: CALLEJAS

DATE/TIME OF TEST: 11/13/2018 03:54:33

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.152	03:49
Breath Volume	2.578	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.149	03:52
Breath Volume	2.167	03:52
Air Blank	0.000	03:53
Control Test	0.078	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 11/28/2018 04:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTZ,KATHRYN,N
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: GOTZ,KATHRYN
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 10/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: WILLMOTT

DATE/TIME OF TEST: 11/17/2018 03:10:50

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.080	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.131	03:05
Breath Volume	1.949	03:05
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.131	03:09
Breath Volume	1.234	03:09
Air Blank	0.000	03:09
Control Test	0.080	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 11/28/2018 04:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,KRISTY,L
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: BROWN,KRISTY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 10/01/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: WEEKS

DATE/TIME OF TEST: 11/18/2018 03:06:59

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.077	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.179	03:02
Breath Volume	2.652	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:04
Subject Sample # 2	0.177	03:05
Breath Volume	1.621	03:05
Air Blank	0.000	03:06
Control Test	0.078	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 11/28/2018 04:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTZ,KATHRYN,N
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: GOTZ,KATHRYN
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 10/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 11/19/2018 01:52:41

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.079	01:48
Breath Volume	1.164	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.069	01:51
Breath Volume	1.281	01:51
Air Blank	0.000	01:51
Control Test	0.079	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 11/28/2018 04:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTZ,KATHRYN,N
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: GOTZ,KATHRYN
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 10/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: STEIN

DATE/TIME OF TEST: 11/23/2018 05:14:14

Diagnostics Check	OK	05:07
Air Blank	0.000	05:07
Control Test	0.080	05:08
Air Blank	0.000	05:08
Subject Sample # 1	0.249	05:09
Breath Volume	2.101	05:09
Air Blank	0.000	05:10
Air Blank	0.000	05:11
Subject Sample # 2	0.247	05:12
Breath Volume	2.195	05:12
Air Blank	0.000	05:13
Control Test	0.079	05:13
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 11/28/2018 04:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINZ,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: BINZ,STEPHEN
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 10/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 765462
DRY GAS EXPIRY DATE: 04/13/2020
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 11/04/2018 01:44:29

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.078	01:42	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:43	
Control Test	0.078	01:43	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 11/18/2018 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GINTZ,BRIAN,W
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: HIRSCH,STEVEN
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 10/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 765462
DRY GAS EXPIRY DATE: 04/13/2020
VIOLATION CODE: DUI
ARREST TIME: 15:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:39

SUBJECT'S LAST NAME: OUTLAW

DATE/TIME OF TEST: 11/11/2018 16:09:48

Diagnostics Check	OK	15:59
Air Blank	0.000	16:00
Control Test	0.077	16:00
Air Blank	0.000	16:01
Subject Sample # 1	0.299	16:01
Breath Volume	2.367	16:01
Air Blank	0.000	16:02
Air Blank	0.000	16:04
Subject Sample # 2	0.276	16:05
Breath Volume	3.050	16:05
Air Blank	0.000	16:05
Air Blank	0.000	16:07
Subject Sample # 3	0.267	16:08
Breath Volume	3.070	16:08
Air Blank	0.000	16:08
Control Test	0.077	16:09
Air Blank	0.000	16:09
Diagnostics Check	OK	16:09

Data Download Date/Time: 11/18/2018 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINZ,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: HUDSON,ANDREW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 765462
DRY GAS EXPIRY DATE: 04/13/2020
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 11/18/2018 02:26:24

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.077	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.017	02:21
Breath Volume	1.648	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.017	02:24
Breath Volume	1.218	02:24
Air Blank	0.000	02:25
Control Test	0.077	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 11/18/2018 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: PEREZ-HERNANDEZ

DATE/TIME OF TEST: 11/10/2018 21:52:40

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.079	21:46
Air Blank	0.000	21:47
Subject Sample # 1	0.223	21:48
Breath Volume	2.964	21:48
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.215	21:51
Breath Volume	2.753	21:51
Air Blank	0.000	21:51
Control Test	0.080	21:52
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 12/10/2018 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMAN,MARIA,E
BREATH TEST OPERATOR AGENCY: Palatka PD
ARRESTING OFFICER: NUNZ,M
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: PUGH

DATE/TIME OF TEST: 11/25/2018 00:01:51

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:50	
Control Test	0.079	23:51	
Air Blank	0.000	23:51	
Subject Sample	# 2	23:54	No Sample Provided
Breath Volume	0.000	23:54	
Air Blank	0.000	23:55	
Air Blank	0.000	23:57	
Subject Sample	# 1	00:00	No Sample Provided
Breath Volume	0.000	00:00	
Air Blank	0.000	00:00	
Control Test	0.080	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 12/10/2018 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMAN,MARIA,E
BREATH TEST OPERATOR AGENCY: Palatka PD
ARRESTING OFFICER: NUNEZ,MICHAEL
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 11/25/2018 03:02:43

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.077	02:58
Breath Volume	2.769	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.077	03:01
Breath Volume	3.039	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 12/10/2018 20:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIDONATO,MIKE,J
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: DIDONATO,MIKE
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 10/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 11/18/2018 20:52:58

Diagnostics Check	OK	20:46
Air Blank	0.000	20:47
Control Test	0.082	20:47
Air Blank	0.000	20:48
Subject Sample # 1	0.146	20:48
Breath Volume	1.859	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.145	20:51
Breath Volume	1.898	20:51
Air Blank	0.000	20:52
Control Test	0.082	20:52
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 11/26/2018 07:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRITY, THOMAS, A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GARRITY, THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 11/03/2018 00:36:59

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.080	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.122	00:32
Breath Volume	2.605	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.117	00:35
Breath Volume	2.375	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 11/30/2018 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRITY, THOMAS, A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GARRITY, THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MOLNAR

DATE/TIME OF TEST: 11/04/2018 03:38:30

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.081	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.264	03:33
Breath Volume	1.378	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.265	03:37
Breath Volume	1.800	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 11/30/2018 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRATSCH,CASEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: SHAOUNI,KALEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GORDIANO

DATE/TIME OF TEST: 11/05/2018 02:56:09

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.166	02:51
Breath Volume	2.882	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.172	02:54
Breath Volume	2.214	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 11/30/2018 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRATSCH,CASEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: SHAOUNI,KALEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: WARDEN

DATE/TIME OF TEST: 11/17/2018 01:14:14

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.081	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.000	01:10
Breath Volume	2.617	01:10
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.000	01:12
Breath Volume	2.656	01:12
Air Blank	0.000	01:13
Control Test	0.081	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 11/30/2018 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAKEMAN,TAYLOR,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: SHAOUNI,KALEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: CLEMENTS

DATE/TIME OF TEST: 11/18/2018 04:34:10

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.081	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.157	04:29
Breath Volume	2.386	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.150	04:32
Breath Volume	2.386	04:32
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 11/30/2018 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARMAN, RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:21

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 11/04/2018 06:52:19

Diagnostics Check	OK	06:45
Air Blank	0.000	06:45
Control Test	0.081	06:46
Air Blank	0.000	06:46
Subject Sample # 1	0.193	06:47
Breath Volume	1.960	06:47
Air Blank	0.000	06:48
Air Blank	0.000	06:50
Subject Sample # 2	0.193	06:50
Breath Volume	2.371	06:50
Air Blank	0.000	06:51
Control Test	0.079	06:51
Air Blank	0.000	06:52
Diagnostics Check	OK	06:52

Data Download Date/Time: 11/07/2018 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ, YAISEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:46

SUBJECT'S LAST NAME: NIKITIN

DATE/TIME OF TEST: 11/04/2018 07:24:51

Diagnostics Check	OK	07:18
Air Blank	0.000	07:18
Control Test	0.079	07:19
Air Blank	0.000	07:19
Subject Sample # 1	0.208	07:20
Breath Volume	1.351	07:20
Air Blank	0.000	07:20
Air Blank	0.000	07:22
Subject Sample # 2	0.211	07:23
Breath Volume	1.261	07:23
Air Blank	0.000	07:23
Control Test	0.080	07:24
Air Blank	0.000	07:24
Diagnostics Check	OK	07:24

Data Download Date/Time: 11/07/2018 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLER,JESSICA
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: ROWLINSON

DATE/TIME OF TEST: 11/13/2018 01:09:41

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:58	
Control Test	0.080	00:59	
Air Blank	0.000	00:59	
Subject Sample	# 1	01:02	No Sample Provided
Breath Volume	0.000	01:02	
Air Blank	0.000	01:03	
Air Blank	0.000	01:05	
Subject Sample	# 2	01:08	No Sample Provided
Breath Volume	0.000	01:08	
Air Blank	0.000	01:08	
Control Test	0.080	01:09	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 11/15/2018 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: MEEKS

DATE/TIME OF TEST: 11/16/2018 03:41:00

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.165	03:36
Breath Volume	2.417	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.159	03:39
Breath Volume	1.261	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:41

Data Download Date/Time: 11/22/2018 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, CALEB, N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MENDOZA, ANGEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 11/18/2018 04:15:46

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.081	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.102	04:10
Breath Volume	1.300	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.092	04:14
Breath Volume	1.523	04:14
Air Blank	0.000	04:14
Control Test	0.079	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 11/22/2018 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PIERSALL,BRADLEY
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: TEJERA

DATE/TIME OF TEST: 11/27/2018 04:17:18

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.110	04:12
Breath Volume	1.222	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.111	04:15
Breath Volume	1.746	04:15
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 11/29/2018 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: CONTARDI,R
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: LIM

DATE/TIME OF TEST: 11/03/2018 20:00:28

Diagnostics Check	OK	19:51
Air Blank	0.000	19:51
Control Test	0.079	19:51
Air Blank	0.000	19:52
Subject Sample # 1	0.182	19:53
Breath Volume	1.261	19:53
Air Blank	0.000	19:54
Air Blank	0.000	19:55
Subject Sample # 2	0.199	19:58
Breath Volume	1.265	19:58
Air Blank	0.000	19:59
Control Test	0.077	19:59
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

Data Download Date/Time: 11/28/2018 21:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: YAKARYS,THEODOROS
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: CABRERA CANO

DATE/TIME OF TEST: 11/14/2018 05:40:41

Diagnostics Check	OK	05:37	
Air Blank	0.000	05:38	
Control Test	0.078	05:38	
Air Blank	0.000	05:39	
Subject Sample	# 1	05:39	Subject Test Refused
Breath Volume	0.000	05:39	
Air Blank	0.000	05:39	
Control Test	0.079	05:40	
Air Blank	0.000	05:40	
Diagnostics Check	OK	05:40	

Data Download Date/Time: 11/28/2018 21:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HALL,NICHOLAS,A

BREATH TEST OPERATOR AGENCY: Cocoa PD

ARRESTING OFFICER: YAKARYS,THEODOROS

ARRESTING OFFICER AGENCY: Cocoa PD

LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02618080A3

DRY GAS EXPIRY DATE: 03/05/2020

VIOLATION CODE: DUI

ARREST TIME: 20:59

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 11/29/2018 00:49:01

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:46	
Control Test	0.079	00:46	
Air Blank	0.000	00:46	
Subject Sample	# 1	00:47	Subject Test Refused
Breath Volume	0.000	00:47	
Air Blank	0.000	00:48	
Control Test	0.079	00:48	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 12/19/2018 21:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WANCA, WILLIAM, T
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: PENDER, DEREK
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 11/02/2018 19:39:42

Diagnostics Check	OK	19:37	
Air Blank	0.000	19:38	
Control Test	0.080	19:38	
Air Blank	0.000	19:38	
Subject Sample	# 1	19:39	Sequence Aborted
Breath Volume	0.000	19:39	
Air Blank	0.000	19:39	

Data Download Date/Time: 11/29/2018 05:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WANCA, WILLIAM, T
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: PENDER, DEREK
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 11/02/2018 19:45:35

Diagnostics Check	OK	19:42	
Air Blank	0.000	19:43	
Control Test	0.080	19:43	
Air Blank	0.000	19:44	
Subject Sample	# 1	19:44	Subject Test Refused
Breath Volume	0.000	19:44	
Air Blank	0.000	19:44	
Control Test	0.080	19:45	
Air Blank	0.000	19:45	
Diagnostics Check	OK	19:45	

Data Download Date/Time: 11/29/2018 05:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOMEZ,IVETTE,

BREATH TEST OPERATOR AGENCY: Rockledge PD

ARRESTING OFFICER: CLARK,JAMES

ARRESTING OFFICER AGENCY: Rockledge PD

LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 572920

DRY GAS EXPIRY DATE: 09/13/2019

VIOLATION CODE: DUI

ARREST TIME: 21:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: ESPLING

DATE/TIME OF TEST: 11/11/2018 22:35:32

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:33	
Control Test	0.080	22:33	
Air Blank	0.000	22:34	
Subject Sample	# 1	22:34	Subject Test Refused
Breath Volume	0.000	22:34	
Air Blank	0.000	22:34	
Control Test	0.079	22:34	
Air Blank	0.000	22:35	
Diagnostics Check	OK	22:35	

Data Download Date/Time: 11/29/2018 05:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CLARK,JAMES,P

BREATH TEST OPERATOR AGENCY: Rockledge PD

ARRESTING OFFICER: GARNER,TYRELL

ARRESTING OFFICER AGENCY: Rockledge PD

LAST AGENCY INSPECTION: 10/29/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 572920

DRY GAS EXPIRY DATE: 09/13/2019

VIOLATION CODE: DUI

ARREST TIME: 16:24

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: FEARRINGTON

DATE/TIME OF TEST: 11/20/2018 19:09:35

Diagnostics Check	OK	19:02
Air Blank	0.000	19:03
Control Test	0.080	19:03
Air Blank	0.000	19:03
Subject Sample # 1	0.000	19:04
Breath Volume	2.109	19:04
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample # 2	0.000	19:08
Breath Volume	2.644	19:08
Air Blank	0.000	19:08
Control Test	0.079	19:08
Air Blank	0.000	19:09
Diagnostics Check	OK	19:09

Data Download Date/Time: 11/29/2018 05:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON,EDWIN,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: KOUBEK,STEVE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 11/04/2018 01:10:47

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:05	
Control Test	0.080	01:06	
Air Blank	0.000	01:06	
Subject Sample # 1	0.250	01:09	
Breath Volume	1.136	01:09	
Air Blank		01:09	Purge Fail
Air Blank	0.000	01:10	

Data Download Date/Time: 11/27/2018 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON, EDWIN, J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: KOUBEK, STEVE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 11/04/2018 01:18:27

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:12	
Control Test	0.077	01:13	
Air Blank	0.000	01:13	
Subject Sample	# 1 0.236	01:16	Volume Not Met
Breath Volume	0.332	01:16	
Air Blank		01:17	Purge Fail
Air Blank	0.000	01:18	

Data Download Date/Time: 11/27/2018 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: KOUBEK,STEVE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 11/04/2018 01:35:51

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.078	01:30	
Air Blank	0.000	01:31	
Subject Sample	# 1 0.220	01:34	Volume Not Met
Breath Volume	0.316	01:34	
Air Blank		01:35	Purge Fail
Air Blank	0.000	01:35	

Data Download Date/Time: 11/27/2018 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON, EDWIN, J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: CRONIN, SEAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: LOPEZ AREVALO

DATE/TIME OF TEST: 11/05/2018 03:45:46

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.231	03:41
Breath Volume	2.304	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:43
Subject Sample # 2	0.232	03:44
Breath Volume	2.433	03:44
Air Blank	0.000	03:44
Control Test	0.079	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 11/27/2018 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BARNETT,LARRY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/16/2018 03:42:19

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.054	03:37
Breath Volume	2.066	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.057	03:40
Breath Volume	2.457	03:40
Air Blank	0.000	03:41
Control Test	0.078	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 11/27/2018 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON, EDWIN, J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BEARY, JONATHAN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 11/21/2018 20:51:37

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:47	
Control Test	0.079	20:48	
Air Blank	0.000	20:48	
Subject Sample	# 1 0.267	20:50	
Breath Volume	1.351	20:50	
Air Blank		20:50	Purge Fail
Air Blank	0.000	20:51	

Data Download Date/Time: 11/27/2018 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON,EDWIN,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BEARY,JONATHAN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 11/21/2018 20:58:53

Diagnostics Check	OK	20:52	
Air Blank	0.000	20:53	
Control Test	0.078	20:53	
Air Blank	0.000	20:54	
Subject Sample	# 1 0.254	20:57	Volume Not Met
Breath Volume	0.125	20:57	
Air Blank		20:58	Purge Fail
Air Blank		20:58	Purge Fail

Data Download Date/Time: 11/27/2018 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON,EDWIN,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BEARY,JONATHAN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 11/21/2018 21:14:43

Diagnostics Check	OK	21:08	
Air Blank	0.000	21:08	
Control Test	0.079	21:09	
Air Blank	0.000	21:09	
Subject Sample	# 1 0.212	21:12	Volume Not Met
Breath Volume	0.058	21:12	
Air Blank		21:13	Purge Fail
Air Blank	0.000	21:14	

Data Download Date/Time: 11/27/2018 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: LIVINGSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: ROCK

DATE/TIME OF TEST: 11/23/2018 23:40:28

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.079	23:34
Air Blank	0.000	23:34
Subject Sample # 1	0.200	23:35
Breath Volume	2.789	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.200	23:38
Breath Volume	1.921	23:38
Air Blank	0.000	23:39
Control Test	0.078	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 11/27/2018 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BARNETT,LARRY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MARTINEZ LOPEZ

DATE/TIME OF TEST: 11/25/2018 03:01:11

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.208	02:55
Breath Volume	4.027	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.191	02:59
Breath Volume	4.222	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 11/27/2018 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: TARANTINO,GRAIG
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: ELDER

DATE/TIME OF TEST: 11/29/2018 04:12:59

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.080	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.162	04:08
Breath Volume	4.222	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:10
Subject Sample # 2	0.162	04:11
Breath Volume	3.984	04:11
Air Blank	0.000	04:12
Control Test	0.081	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 12/27/2018 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JIM,E
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: SANCHEZ,CARLOS
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: KLIMKOWSKI

DATE/TIME OF TEST: 11/18/2018 00:22:49

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.080	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.216	00:18
Breath Volume	3.265	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:20
Subject Sample # 2	0.225	00:21
Breath Volume	4.187	00:21
Air Blank	0.000	00:21
Control Test	0.080	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 12/20/2018 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GRAY,CONNOR
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: LYCETTE

DATE/TIME OF TEST: 11/03/2018 04:01:44

Diagnostics Check	OK	03:54
Air Blank	0.000	03:55
Control Test	0.081	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.149	03:57
Breath Volume	3.519	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.151	04:00
Breath Volume	3.292	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 11/30/2018 15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLEASON, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MASON, RYAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: SOMERS

DATE/TIME OF TEST: 11/04/2018 06:21:23

Diagnostics Check	OK	06:14
Air Blank	0.000	06:14
Control Test	0.080	06:14
Air Blank	0.000	06:15
Subject Sample # 1	0.171	06:15
Breath Volume	1.820	06:15
Air Blank	0.000	06:16
Air Blank	0.000	06:18
Subject Sample # 2	0.172	06:19
Breath Volume	1.976	06:19
Air Blank	0.000	06:20
Control Test	0.079	06:20
Air Blank	0.000	06:21
Diagnostics Check	OK	06:21

Data Download Date/Time: 11/30/2018 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROONQUIST,DON,R
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: MCCORMICK,P
ARRESTING OFFICER AGENCY: Kennedy Space Center
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/08/2018 18:12:13

Diagnostics Check	OK	18:01
Air Blank	0.000	18:01
Control Test	0.080	18:02
Air Blank	0.000	18:02
Subject Sample # 1	0.319	18:03
Breath Volume	2.164	18:03
Air Blank	0.000	18:04
Air Blank	0.000	18:05
Subject Sample # 2	0.297	18:07
Breath Volume	2.925	18:07
Air Blank	0.000	18:08
Air Blank	0.000	18:10
Subject Sample # 3	0.312	18:10
Breath Volume	3.339	18:10
Air Blank	0.000	18:11
Control Test	0.076	18:11
Air Blank	0.000	18:12
Diagnostics Check	OK	18:12

Data Download Date/Time: 11/30/2018 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWELL,NATHANAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROWELL,NATHANAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: LAVANT

DATE/TIME OF TEST: 11/10/2018 03:34:35

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.199	03:29
Breath Volume	2.582	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.203	03:32
Breath Volume	2.484	03:32
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 11/30/2018 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,NICHOLAS,W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BAKER,D
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: IBRAGIMOV

DATE/TIME OF TEST: 11/13/2018 20:12:25

Diagnostics Check	OK	20:05
Air Blank	0.000	20:05
Control Test	0.081	20:06
Air Blank	0.000	20:06
Subject Sample # 1	0.307	20:07
Breath Volume	2.816	20:07
Air Blank	0.000	20:08
Air Blank	0.000	20:10
Subject Sample # 2	0.310	20:10
Breath Volume	1.753	20:10
Air Blank	0.000	20:11
Control Test	0.079	20:11
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 11/30/2018 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,NICHOLAS,W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: STREIFF,C
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: PEACOCK

DATE/TIME OF TEST: 11/13/2018 22:04:13

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.080	21:57
Air Blank	0.000	21:58
Subject Sample # 1	0.191	21:59
Breath Volume	1.589	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:01
Subject Sample # 2	0.186	22:02
Breath Volume	1.496	22:02
Air Blank	0.000	22:03
Control Test	0.080	22:03
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 11/30/2018 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCNEAL,DEANA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 11/14/2018 23:34:46

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.082	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.075	23:30
Breath Volume	1.640	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.067	23:33
Breath Volume	3.558	23:33
Air Blank	0.000	23:33
Control Test	0.082	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 11/30/2018 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WINSTEAD,JUSTIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 11/15/2018 23:42:37

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.080	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.153	23:34
Breath Volume	2.097	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:36
Subject Sample # 2	0.176	23:37
Breath Volume	2.179	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 3	0.151	23:41
Breath Volume	1.500	23:41
Air Blank	0.000	23:41
Control Test	0.080	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 11/30/2018 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MARCUS,H
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: SHAOUNI,KALEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: ANDERS

DATE/TIME OF TEST: 11/22/2018 04:00:15

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.081	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.195	03:55
Breath Volume	4.347	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.201	03:58
Breath Volume	4.578	03:58
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 11/30/2018 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLEASON, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GLEASON, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: FATE

DATE/TIME OF TEST: 11/22/2018 15:29:57

Diagnostics Check	OK	15:23
Air Blank	0.000	15:23
Control Test	0.081	15:24
Air Blank	0.000	15:24
Subject Sample # 1	0.197	15:25
Breath Volume	2.941	15:25
Air Blank	0.000	15:26
Air Blank	0.000	15:27
Subject Sample # 2	0.193	15:28
Breath Volume	2.582	15:28
Air Blank	0.000	15:29
Control Test	0.079	15:29
Air Blank	0.000	15:29
Diagnostics Check	OK	15:29

Data Download Date/Time: 11/30/2018 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 00:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: HODGSON

DATE/TIME OF TEST: 11/02/2018 01:13:11

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:00	
Control Test	0.081	01:01	
Air Blank	0.000	01:01	
Subject Sample	# 1	01:04	No Sample Provided
Breath Volume	0.000	01:04	
Air Blank	0.000	01:05	
Air Blank	0.000	01:07	
Subject Sample	# 2	01:08	
Breath Volume	1.656	01:08	
Air Blank	0.000	01:09	
Air Blank	0.000	01:11	
Subject Sample	# 3	01:11	
Breath Volume	1.351	01:11	
Air Blank	0.000	01:12	
Control Test	0.079	01:12	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 11/26/2018 03:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 02:21

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: MCHATTON

DATE/TIME OF TEST: 11/06/2018 02:37:48

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.081	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.115	02:33
Breath Volume	1.523	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:35
Subject Sample # 2	0.126	02:36
Breath Volume	2.351	02:36
Air Blank	0.000	02:36
Control Test	0.080	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 11/26/2018 03:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAMILTON,J,W

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HAMILTON,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 04:10

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 11/20/2018 04:43:17

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.208	04:38
Breath Volume	1.527	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.206	04:41
Breath Volume	1.515	04:41
Air Blank	0.000	04:42
Control Test	0.079	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 11/26/2018 03:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 01:32

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: SEBASTIAN

DATE/TIME OF TEST: 11/21/2018 01:49:30

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.164	01:45
Breath Volume	2.027	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.157	01:47
Breath Volume	2.039	01:47
Air Blank	0.000	01:48
Control Test	0.081	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 11/26/2018 03:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 23:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: HEFLER

DATE/TIME OF TEST: 11/22/2018 00:15:43

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:07	
Control Test	0.082	00:07	
Air Blank	0.000	00:08	
Subject Sample	# 1 0.135	00:09	
Breath Volume	1.363	00:09	
Air Blank	0.000	00:09	
Air Blank	0.000	00:11	
Subject Sample	# 2	00:14	Subject Test Refused
Breath Volume	1.867	00:14	
Air Blank	0.000	00:14	
Control Test	0.080	00:15	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

Data Download Date/Time: 11/26/2018 03:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 21:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 11/25/2018 21:44:06

Diagnostics Check	OK	21:36
Air Blank	0.000	21:36
Control Test	0.080	21:37
Air Blank	0.000	21:37
Subject Sample # 1	0.128	21:39
Breath Volume	1.574	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.135	21:42
Breath Volume	1.636	21:42
Air Blank	0.000	21:43
Control Test	0.080	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 11/26/2018 03:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROLPH,AMANDA,L

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: ROLPH,AMANDA

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 01:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: ATWOOD

DATE/TIME OF TEST: 11/05/2018 02:31:36

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.020	02:27
Breath Volume	2.449	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.019	02:30
Breath Volume	2.125	02:30
Air Blank	0.000	02:30
Control Test	0.080	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 11/26/2018 02:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SPAULDING,CODY,R

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: SPAULDING,CODY

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 17:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 11/08/2018 18:49:39

Diagnostics Check	OK	18:43
Air Blank	0.000	18:44
Control Test	0.080	18:44
Air Blank	0.000	18:44
Subject Sample # 1	0.213	18:45
Breath Volume	1.976	18:45
Air Blank	0.000	18:45
Air Blank	0.000	18:47
Subject Sample # 2	0.198	18:48
Breath Volume	1.648	18:48
Air Blank	0.000	18:48
Control Test	0.079	18:49
Air Blank	0.000	18:49
Diagnostics Check	OK	18:49

Data Download Date/Time: 11/26/2018 02:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROLPH,AMANDA,L

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: ROLPH,AMANDA

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 02:45

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 11/10/2018 03:26:08

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.114	03:21
Breath Volume	1.894	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:23
Subject Sample # 2	0.116	03:24
Breath Volume	2.074	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 11/26/2018 02:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROLPH,AMANDA,L

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: ROLPH,AMANDA

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 19:16

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 11/10/2018 20:00:56

Diagnostics Check	OK	19:54
Air Blank	0.000	19:55
Control Test	0.079	19:55
Air Blank	0.000	19:56
Subject Sample # 1	0.287	19:56
Breath Volume	1.707	19:56
Air Blank	0.000	19:57
Air Blank	0.000	19:59
Subject Sample # 2	0.303	19:59
Breath Volume	1.585	19:59
Air Blank	0.000	20:00
Control Test	0.079	20:00
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

Data Download Date/Time: 11/26/2018 02:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROLPH,AMANDA,L

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: ROLPH,AMANDA

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 20:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 11/16/2018 21:37:07

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.080	21:31
Air Blank	0.000	21:31
Subject Sample # 1	0.134	21:32
Breath Volume	1.882	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.130	21:35
Breath Volume	1.574	21:35
Air Blank	0.000	21:36
Control Test	0.079	21:36
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 11/26/2018 02:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SPAULDING,CODY,R

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: SPAULDING,CODY

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 14:46

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: WITKOWSKI

DATE/TIME OF TEST: 11/18/2018 15:26:46

Diagnostics Check	OK	15:20
Air Blank	0.000	15:20
Control Test	0.079	15:21
Air Blank	0.000	15:21
Subject Sample # 1	0.264	15:22
Breath Volume	1.820	15:22
Air Blank	0.000	15:22
Air Blank	0.000	15:24
Subject Sample # 2	0.255	15:25
Breath Volume	1.433	15:25
Air Blank	0.000	15:25
Control Test	0.078	15:26
Air Blank	0.000	15:26
Diagnostics Check	OK	15:26

Data Download Date/Time: 11/26/2018 02:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAMILTON,J,W

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HAMILTON,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 23:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: BRADBERRY

DATE/TIME OF TEST: 11/21/2018 23:56:02

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.080	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.184	23:49
Breath Volume	1.335	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.201	23:54
Breath Volume	1.550	23:54
Air Blank	0.000	23:55
Control Test	0.079	23:55
Air Blank	0.000	23:55
Diagnostics Check	OK	23:56

Data Download Date/Time: 11/26/2018 02:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SPAULDING,CODY,R

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: PENNINGTON,WILLIAM

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 11:59

GENDER: Female

OBSERVATION PERIOD BEGAN: 12:06

SUBJECT'S LAST NAME: TURLO

DATE/TIME OF TEST: 11/22/2018 12:45:45

Diagnostics Check	OK	12:34	
Air Blank	0.000	12:34	
Control Test	0.079	12:34	
Air Blank	0.000	12:35	
Subject Sample	# 1 0.199	12:38	Volume Not Met
Breath Volume	0.160	12:38	
Air Blank	0.000	12:39	
Air Blank	0.000	12:40	
Subject Sample	# 2 0.200	12:44	Volume Not Met
Breath Volume	0.152	12:44	
Air Blank	0.000	12:44	
Control Test	0.077	12:45	
Air Blank	0.000	12:45	
Diagnostics Check	OK	12:45	

Data Download Date/Time: 11/26/2018 02:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SPAULDING,CODY,R

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: PENNINGTON,WILLIAM

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 11:59

GENDER: Female

OBSERVATION PERIOD BEGAN: 12:06

SUBJECT'S LAST NAME: TURLO

DATE/TIME OF TEST: 11/22/2018 12:55:29

Diagnostics Check	OK	12:48
Air Blank	0.000	12:49
Control Test	0.077	12:49
Air Blank	0.000	12:50
Subject Sample # 1	0.195	12:50
Breath Volume	2.074	12:50
Air Blank	0.000	12:51
Air Blank	0.000	12:53
Subject Sample # 2	0.202	12:53
Breath Volume	1.222	12:53
Air Blank	0.000	12:54
Control Test	0.077	12:54
Air Blank	0.000	12:55
Diagnostics Check	OK	12:55

Data Download Date/Time: 11/26/2018 02:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,ROBERT,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LOGAN,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/24/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 00:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: COLLAZO PINTO

DATE/TIME OF TEST: 11/23/2018 01:38:09

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:34	
Control Test	0.080	01:34	
Air Blank	0.000	01:35	
Subject Sample	# 1	01:36	Subject Test Refused
Breath Volume	0.000	01:36	
Air Blank	0.000	01:37	
Control Test	0.079	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 11/26/2018 02:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,ROBERT,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: BUCKLIN,S

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 04:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: CARMEL

DATE/TIME OF TEST: 11/30/2018 05:12:19

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.080	05:05
Air Blank	0.000	05:05
Subject Sample # 1	0.153	05:06
Breath Volume	1.765	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.139	05:10
Breath Volume	1.222	05:10
Air Blank	0.000	05:11
Control Test	0.079	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 12/29/2018 21:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: IRIZARRY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 11/30/2018 03:19:58

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.080	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.241	03:15
Breath Volume	2.132	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample # 2	0.245	03:18
Breath Volume	2.105	03:18
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 12/02/2018 21:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FOSTER,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: FOEHR

DATE/TIME OF TEST: 11/01/2018 05:55:07

Diagnostics Check	OK	05:54	
Air Blank		05:54	RFI Detect
Air Blank	0.000	05:55	

Data Download Date/Time: 11/28/2018 22:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FOSTER,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: FOEHR

DATE/TIME OF TEST: 11/01/2018 06:03:25

Diagnostics Check	OK	05:57
Air Blank	0.000	05:57
Control Test	0.082	05:57
Air Blank	0.000	05:58
Subject Sample # 1	0.000	05:58
Breath Volume	2.519	05:58
Air Blank	0.000	05:59
Air Blank	0.000	06:01
Subject Sample # 2	0.000	06:02
Breath Volume	2.449	06:02
Air Blank	0.000	06:02
Control Test	0.082	06:02
Air Blank	0.000	06:03
Diagnostics Check	OK	06:03

Data Download Date/Time: 11/28/2018 22:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOD,COREY,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RAY,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: CUFF

DATE/TIME OF TEST: 11/01/2018 08:10:39

Diagnostics Check	OK	08:04
Air Blank	0.000	08:04
Control Test	0.082	08:05
Air Blank	0.000	08:05
Subject Sample # 1	0.083	08:06
Breath Volume	1.886	08:06
Air Blank	0.000	08:06
Air Blank	0.000	08:08
Subject Sample # 2	0.081	08:09
Breath Volume	1.851	08:09
Air Blank	0.000	08:09
Control Test	0.082	08:10
Air Blank	0.000	08:10
Diagnostics Check	OK	08:10

Data Download Date/Time: 11/28/2018 22:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: SERVEDIO JR

DATE/TIME OF TEST: 11/01/2018 23:48:24

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.082	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.183	23:41
Breath Volume	2.058	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.190	23:46
Breath Volume	3.183	23:46
Air Blank	0.000	23:47
Control Test	0.082	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 11/28/2018 22:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: ELLINGER

DATE/TIME OF TEST: 11/02/2018 01:46:02

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.081	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:45	
Control Test	0.081	01:45	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 11/28/2018 22:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: TIMMONS

DATE/TIME OF TEST: 11/02/2018 03:26:57

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.082	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:26	
Control Test	0.082	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 11/28/2018 22:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: ERICKSON

DATE/TIME OF TEST: 11/02/2018 21:37:39

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.081	21:31
Air Blank	0.000	21:32
Subject Sample # 1	0.000	21:32
Breath Volume	1.210	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.000	21:36
Breath Volume	1.359	21:36
Air Blank	0.000	21:36
Control Test	0.082	21:37
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 11/28/2018 22:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: MUSGRAVE

DATE/TIME OF TEST: 11/03/2018 02:57:53

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:54	
Control Test	0.082	02:55	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:56	
Control Test	0.082	02:57	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

Data Download Date/Time: 11/28/2018 22:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: CULHANE

DATE/TIME OF TEST: 11/03/2018 03:13:06

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.082	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.240	03:08
Breath Volume	2.246	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:10
Subject Sample # 2	0.224	03:11
Breath Volume	1.957	03:11
Air Blank	0.000	03:12
Control Test	0.082	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 11/28/2018 22:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: KUMATSKY

DATE/TIME OF TEST: 11/03/2018 04:25:47

Diagnostics Check	OK	04:14	
Air Blank	0.000	04:14	
Control Test	0.082	04:15	
Air Blank	0.000	04:15	
Subject Sample	# 1	04:18	No Sample Provided
Breath Volume	0.000	04:18	
Air Blank	0.000	04:19	
Air Blank	0.000	04:21	
Subject Sample	# 2	04:24	No Sample Provided
Breath Volume	0.000	04:24	
Air Blank	0.000	04:24	
Control Test	0.082	04:25	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 11/28/2018 22:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DUCKERS,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: QUESADA

DATE/TIME OF TEST: 11/03/2018 04:53:51

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.082	04:47
Air Blank	0.000	04:48
Subject Sample # 1	0.127	04:49
Breath Volume	2.667	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:51
Subject Sample # 2	0.129	04:52
Breath Volume	2.500	04:52
Air Blank	0.000	04:52
Control Test	0.081	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: AULT

DATE/TIME OF TEST: 11/03/2018 22:25:37

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.083	22:19
Air Blank	0.000	22:19
Subject Sample # 1	0.064	22:20
Breath Volume	2.117	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.067	22:24
Breath Volume	2.167	22:24
Air Blank	0.000	22:24
Control Test	0.082	22:25
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: ZUBEK

DATE/TIME OF TEST: 11/03/2018 22:57:14

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:54	
Control Test	0.083	22:55	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:55	Subject Test Refused
Breath Volume	0.000	22:55	
Air Blank	0.000	22:56	
Control Test	0.082	22:56	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,K,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE,RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: WORTH

DATE/TIME OF TEST: 11/04/2018 00:29:00

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.083	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:28	
Control Test	0.082	00:28	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:29	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE, RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE, RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: FULLAM

DATE/TIME OF TEST: 11/04/2018 01:38:14

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.083	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.195	01:33
Breath Volume	2.390	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:35
Subject Sample # 2	0.192	01:36
Breath Volume	1.914	01:36
Air Blank	0.000	01:37
Control Test	0.081	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 11/04/2018 01:11:21

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:03	
Control Test	0.082	01:03	
Air Blank	0.000	01:04	
Subject Sample	# 1 0.172	01:07	Volume Not Met
Breath Volume	0.398	01:07	
Air Blank	0.000	01:08	
Air Blank	0.000	01:09	
Subject Sample	# 2	01:09	Subject Test Refused
Breath Volume	0.000	01:09	
Air Blank	0.000	01:10	
Control Test	0.080	01:10	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: CROOK

DATE/TIME OF TEST: 11/04/2018 04:35:33

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.082	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.143	04:30
Breath Volume	1.652	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.131	04:33
Breath Volume	1.531	04:33
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FOSTER,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 11/05/2018 23:12:04

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:09	
Control Test	0.082	23:09	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:10	Subject Test Refused
Breath Volume	0.000	23:10	
Air Blank	0.000	23:11	
Control Test	0.082	23:11	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:12	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAUTHIER,ERIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 11/06/2018 02:55:50

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.082	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:54	Subject Test Refused
Breath Volume	0.000	02:54	
Air Blank	0.000	02:54	
Control Test	0.081	02:55	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 11/06/2018 05:06:54

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.081	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.061	05:02
Breath Volume	2.324	05:02
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.058	05:05
Breath Volume	2.265	05:05
Air Blank	0.000	05:05
Control Test	0.081	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOD,COREY,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CARUSO,ANTHONY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 11/06/2018 17:49:00

Diagnostics Check	OK	17:42
Air Blank	0.000	17:43
Control Test	0.083	17:43
Air Blank	0.000	17:43
Subject Sample # 1	0.134	17:44
Breath Volume	1.382	17:44
Air Blank	0.000	17:44
Air Blank	0.000	17:46
Subject Sample # 2	0.144	17:47
Breath Volume	1.714	17:47
Air Blank	0.000	17:48
Control Test	0.081	17:48
Air Blank	0.000	17:48
Diagnostics Check	OK	17:49

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBSON,JON,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DOBSON,JON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:06

SUBJECT'S LAST NAME: RUDY

DATE/TIME OF TEST: 11/07/2018 15:32:45

Diagnostics Check	OK	15:26
Air Blank	0.000	15:26
Control Test	0.082	15:27
Air Blank	0.000	15:27
Subject Sample # 1	0.000	15:28
Breath Volume	5.113	15:28
Air Blank	0.000	15:28
Air Blank	0.000	15:30
Subject Sample # 2	0.000	15:31
Breath Volume	5.351	15:31
Air Blank	0.000	15:31
Control Test	0.080	15:32
Air Blank	0.000	15:32
Diagnostics Check	OK	15:32

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: JAMIESON

DATE/TIME OF TEST: 11/07/2018 22:58:18

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.082	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.199	22:53
Breath Volume	2.144	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:55
Subject Sample # 2	0.203	22:56
Breath Volume	1.894	22:56
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/08/2018 04:05:15

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.082	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:03	Subject Test Refused
Breath Volume	0.000	04:03	
Air Blank	0.000	04:04	
Control Test	0.082	04:04	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: MILLIKEN

DATE/TIME OF TEST: 11/08/2018 05:19:05

Diagnostics Check	OK	05:13
Air Blank	0.000	05:13
Control Test	0.082	05:13
Air Blank	0.000	05:14
Subject Sample # 1	0.000	05:14
Breath Volume	2.593	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 2	0.000	05:17
Breath Volume	1.718	05:17
Air Blank	0.000	05:18
Control Test	0.082	05:18
Air Blank	0.000	05:18
Diagnostics Check	OK	05:19

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: HUSSEY

DATE/TIME OF TEST: 11/09/2018 02:43:35

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.082	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:42	
Control Test	0.081	02:42	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/09/2018 04:03:15

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:00	
Control Test	0.082	04:01	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Control Test	0.082	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VANWORP,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 11/09/2018 04:38:42

Diagnostics Check	OK	04:35	
Air Blank	0.000	04:36	
Control Test	0.082	04:36	
Air Blank	0.000	04:37	
Subject Sample	# 1	04:37	Subject Test Refused
Breath Volume	0.000	04:37	
Air Blank	0.000	04:37	
Control Test	0.081	04:38	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:38	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: KAULL JR

DATE/TIME OF TEST: 11/10/2018 00:32:53

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:30	
Control Test	0.082	00:30	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:31	
Control Test	0.082	00:32	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: LANHAM

DATE/TIME OF TEST: 11/10/2018 01:43:50

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.082	01:41	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:42	
Control Test	0.081	01:43	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: FUDGE

DATE/TIME OF TEST: 11/10/2018 02:47:53

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.082	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.132	02:43
Breath Volume	2.910	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:45
Subject Sample # 2	0.129	02:46
Breath Volume	2.386	02:46
Air Blank	0.000	02:46
Control Test	0.081	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 11/10/2018 03:21:34

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:19	
Control Test	0.082	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:20	
Control Test	0.081	03:20	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MASON,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 11/10/2018 04:02:32

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.082	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.172	03:57
Breath Volume	2.144	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.156	04:00
Breath Volume	1.507	04:00
Air Blank	0.000	04:01
Control Test	0.082	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: QUINONES-ROSA

DATE/TIME OF TEST: 11/10/2018 06:29:26

Diagnostics Check	OK	06:16	
Air Blank	0.000	06:16	
Control Test	0.082	06:16	
Air Blank	0.000	06:17	
Subject Sample	# 1 0.000	06:20	Volume Not Met
Breath Volume	0.035	06:20	
Air Blank	0.000	06:21	
Air Blank	0.000	06:22	
Subject Sample	# 2 0.134	06:24	
Breath Volume	1.644	06:24	
Air Blank	0.000	06:25	
Air Blank	0.000	06:27	
Subject Sample	# 3 0.132	06:27	
Breath Volume	2.082	06:27	
Air Blank	0.000	06:28	
Control Test	0.080	06:28	
Air Blank	0.000	06:29	
Diagnostics Check	OK	06:29	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 11/10/2018 06:37:13

Diagnostics Check	OK	06:30
Air Blank	0.000	06:31
Control Test	0.081	06:31
Air Blank	0.000	06:32
Subject Sample # 1	0.067	06:32
Breath Volume	1.925	06:32
Air Blank	0.000	06:33
Air Blank	0.000	06:35
Subject Sample # 2	0.063	06:35
Breath Volume	1.156	06:35
Air Blank	0.000	06:36
Control Test	0.081	06:36
Air Blank	0.000	06:37
Diagnostics Check	OK	06:37

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FOSTER,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: WILDER

DATE/TIME OF TEST: 11/10/2018 22:46:07

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:38	
Control Test	0.083	22:38	
Air Blank	0.000	22:38	
Subject Sample	# 1	22:42	No Sample Provided
Breath Volume	0.000	22:42	
Air Blank	0.000	22:42	
Air Blank	0.000	22:44	
Subject Sample	# 2	22:44	Subject Test Refused
Breath Volume	0.000	22:44	
Air Blank	0.000	22:45	
Control Test	0.082	22:45	
Air Blank	0.000	22:46	
Diagnostics Check	OK	22:46	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANISZEWSKI,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: PAULINO

DATE/TIME OF TEST: 11/11/2018 00:23:11

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.082	00:17
Air Blank	0.000	00:18
Subject Sample # 1	0.000	00:18
Breath Volume	1.648	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.000	00:21
Breath Volume	1.507	00:21
Air Blank	0.000	00:22
Control Test	0.082	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BOGGESS,O
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ADLEMAN

DATE/TIME OF TEST: 11/11/2018 02:47:06

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:36	
Control Test	0.082	02:36	
Air Blank	0.000	02:36	
Subject Sample	# 1	02:40	No Sample Provided
Breath Volume	0.000	02:40	
Air Blank	0.000	02:40	
Air Blank	0.000	02:42	
Subject Sample	# 2	02:45	No Sample Provided
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Control Test	0.083	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROSADO,I
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: MAHON

DATE/TIME OF TEST: 11/12/2018 00:02:43

Diagnostics Check	OK	23:59	
Air Blank	0.000	00:00	
Control Test	0.083	00:00	
Air Blank	0.000	00:01	
Subject Sample	# 1	00:01	Subject Test Refused
Breath Volume	0.000	00:01	
Air Blank	0.000	00:01	
Control Test	0.082	00:02	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:02	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMITH,ROBERT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/14/2018 05:58:01

Diagnostics Check	OK	05:54	
Air Blank	0.000	05:55	
Control Test	0.083	05:55	
Air Blank	0.000	05:55	
Subject Sample	# 1	05:56	Subject Test Refused
Breath Volume	0.000	05:56	
Air Blank	0.000	05:57	
Control Test	0.082	05:57	
Air Blank	0.000	05:57	
Diagnostics Check	OK	05:58	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMITH,KIMBERLY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:51

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/14/2018 06:17:50

Diagnostics Check	OK	06:11
Air Blank	0.000	06:11
Control Test	0.082	06:12
Air Blank	0.000	06:12
Subject Sample # 1	0.232	06:13
Breath Volume	2.054	06:13
Air Blank	0.000	06:14
Air Blank	0.000	06:15
Subject Sample # 2	0.235	06:16
Breath Volume	1.968	06:16
Air Blank	0.000	06:16
Control Test	0.081	06:17
Air Blank	0.000	06:17
Diagnostics Check	OK	06:17

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EASTTY,MARK
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: RONNEBURGER

DATE/TIME OF TEST: 11/14/2018 15:08:32

Diagnostics Check	OK	15:05	
Air Blank	0.000	15:05	
Control Test	0.082	15:06	
Air Blank	0.000	15:06	
Subject Sample	# 1	15:07	Subject Test Refused
Breath Volume	0.000	15:07	
Air Blank	0.000	15:07	
Control Test	0.081	15:07	
Air Blank	0.000	15:08	
Diagnostics Check	OK	15:08	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROSADO,I
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: WESSELS

DATE/TIME OF TEST: 11/15/2018 00:29:42

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:27	
Control Test	0.082	00:27	
Air Blank	0.000	00:28	
Subject Sample	# 1	00:28	Subject Test Refused
Breath Volume	0.000	00:28	
Air Blank	0.000	00:28	
Control Test	0.082	00:29	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOD,COREY,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOOD,COREY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: VOUTSINAS

DATE/TIME OF TEST: 11/15/2018 17:53:41

Diagnostics Check	OK	17:46
Air Blank	0.000	17:47
Control Test	0.082	17:47
Air Blank	0.000	17:48
Subject Sample # 1	0.150	17:48
Breath Volume	1.136	17:48
Air Blank	0.000	17:49
Air Blank	0.000	17:51
Subject Sample # 2	0.168	17:51
Breath Volume	3.101	17:51
Air Blank	0.000	17:52
Control Test	0.082	17:53
Air Blank	0.000	17:53
Diagnostics Check	OK	17:53

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: GANT

DATE/TIME OF TEST: 11/15/2018 22:35:20

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.082	22:29
Air Blank	0.000	22:29
Subject Sample # 1	0.179	22:30
Breath Volume	2.648	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.174	22:33
Breath Volume	2.218	22:33
Air Blank	0.000	22:34
Control Test	0.081	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: PAREDES

DATE/TIME OF TEST: 11/16/2018 01:42:15

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.082	01:39	
Air Blank	0.000	01:40	
Subject Sample	# 1	01:40	Subject Test Refused
Breath Volume	0.000	01:40	
Air Blank	0.000	01:41	
Control Test	0.083	01:41	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: CASSINI

DATE/TIME OF TEST: 11/17/2018 01:36:41

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.082	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.139	01:32
Breath Volume	1.781	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.138	01:35
Breath Volume	1.757	01:35
Air Blank	0.000	01:35
Control Test	0.082	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/17/2018 01:53:56

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.081	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:53	
Control Test	0.082	01:53	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 11/17/2018 04:12:24

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.213	04:07
Breath Volume	1.433	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample # 2	0.215	04:10
Breath Volume	1.328	04:10
Air Blank	0.000	04:11
Control Test	0.081	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: PEDROZA

DATE/TIME OF TEST: 11/17/2018 05:28:27

Diagnostics Check	OK	05:24	
Air Blank	0.000	05:25	
Control Test	0.082	05:25	
Air Blank	0.000	05:25	
Subject Sample	# 1	05:27	Subject Test Refused
Breath Volume	0.000	05:27	
Air Blank	0.000	05:27	
Control Test	0.082	05:27	
Air Blank	0.000	05:28	
Diagnostics Check	OK	05:28	

Data Download Date/Time: 11/28/2018 22:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: LA MOUREAUX

DATE/TIME OF TEST: 11/17/2018 23:05:12

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:02	
Control Test	0.082	23:02	
Air Blank	0.000	23:03	
Subject Sample	# 1	23:03	Subject Test Refused
Breath Volume	0.000	23:03	
Air Blank	0.000	23:04	
Control Test	0.082	23:04	
Air Blank	0.000	23:05	
Diagnostics Check	OK	23:05	

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: WIMBERLY

DATE/TIME OF TEST: 11/18/2018 01:47:06

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.083	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.055	01:42
Breath Volume	2.625	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.054	01:45
Breath Volume	1.941	01:45
Air Blank	0.000	01:46
Control Test	0.082	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: ENTENZA

DATE/TIME OF TEST: 11/18/2018 04:12:07

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.082	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.135	04:07
Breath Volume	1.425	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.147	04:10
Breath Volume	1.175	04:10
Air Blank	0.000	04:11
Control Test	0.082	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: PEARSON III

DATE/TIME OF TEST: 11/18/2018 04:50:20

Diagnostics Check	OK	04:47	
Air Blank	0.000	04:47	
Control Test	0.082	04:48	
Air Blank	0.000	04:48	
Subject Sample	# 1	04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:49	
Control Test	0.081	04:49	
Air Blank	0.000	04:50	
Diagnostics Check	OK	04:50	

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: GARGAN

DATE/TIME OF TEST: 11/18/2018 06:24:48

Diagnostics Check	OK	06:21	
Air Blank	0.000	06:22	
Control Test	0.082	06:22	
Air Blank	0.000	06:23	
Subject Sample	# 1	06:23	Subject Test Refused
Breath Volume	0.000	06:23	
Air Blank	0.000	06:23	
Control Test	0.082	06:24	
Air Blank	0.000	06:24	
Diagnostics Check	OK	06:24	

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBSON,JON,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DOBSON,JON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:32

SUBJECT'S LAST NAME: PORTILLO-GONZALEZ

DATE/TIME OF TEST: 11/18/2018 16:59:21

Diagnostics Check	OK	16:53
Air Blank	0.000	16:53
Control Test	0.083	16:53
Air Blank	0.000	16:54
Subject Sample # 1	0.066	16:54
Breath Volume	3.863	16:54
Air Blank	0.000	16:55
Air Blank	0.000	16:57
Subject Sample # 2	0.064	16:57
Breath Volume	3.554	16:57
Air Blank	0.000	16:58
Control Test	0.081	16:58
Air Blank	0.000	16:59
Diagnostics Check	OK	16:59

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBSON,JON,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ALBERTSON,ROBERT
ARRESTING OFFICER AGENCY: Belleair PD
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: BADWAY

DATE/TIME OF TEST: 11/18/2018 17:26:29

Diagnostics Check	OK	17:23	
Air Blank	0.000	17:24	
Control Test	0.082	17:24	
Air Blank	0.000	17:24	
Subject Sample	# 1	17:25	Subject Test Refused
Breath Volume	0.000	17:25	
Air Blank	0.000	17:25	
Control Test	0.081	17:25	
Air Blank	0.000	17:26	
Diagnostics Check	OK	17:26	

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: THAYER

DATE/TIME OF TEST: 11/18/2018 22:02:36

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.082	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.302	21:57
Breath Volume	2.386	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.290	22:00
Breath Volume	1.609	22:00
Air Blank	0.000	22:01
Control Test	0.082	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,KIOMY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CRIM,TYRONE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SCRIVEN

DATE/TIME OF TEST: 11/19/2018 03:24:14

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.083	03:22	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:23	
Control Test	0.082	03:23	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEFFINGWELL, BRADLEY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 11/19/2018 21:37:45

Diagnostics Check	OK	21:34	
Air Blank	0.000	21:35	
Control Test	0.083	21:35	
Air Blank	0.000	21:36	
Subject Sample	# 1	21:36	Subject Test Refused
Breath Volume	0.000	21:36	
Air Blank	0.000	21:36	
Control Test	0.083	21:37	
Air Blank	0.000	21:37	
Diagnostics Check	OK	21:37	

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEIB, MELANIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: WORM

DATE/TIME OF TEST: 11/20/2018 00:37:06

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.082	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.000	00:31
Breath Volume	1.132	00:31
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.000	00:35
Breath Volume	1.105	00:35
Air Blank	0.000	00:36
Control Test	0.081	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: FALZON

DATE/TIME OF TEST: 11/21/2018 02:15:33

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:13	
Control Test	0.083	02:13	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:14	Subject Test Refused
Breath Volume	0.000	02:14	
Air Blank	0.000	02:14	
Control Test	0.083	02:14	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:15	

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 11/22/2018 04:15:54

Diagnostics Check	OK	04:04	
Air Blank	0.000	04:04	
Control Test	0.083	04:05	
Air Blank	0.000	04:05	
Subject Sample	# 1	04:08	No Sample Provided
Breath Volume	0.000	04:08	
Air Blank	0.000	04:09	
Air Blank	0.000	04:11	
Subject Sample	# 2	04:14	No Sample Provided
Breath Volume	0.000	04:14	
Air Blank	0.000	04:14	
Control Test	0.083	04:15	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: QUESADA

DATE/TIME OF TEST: 11/22/2018 23:58:40

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.083	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.192	23:54
Breath Volume	1.535	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.206	23:57
Breath Volume	2.156	23:57
Air Blank	0.000	23:57
Control Test	0.082	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY, MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE, BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: SOMERS

DATE/TIME OF TEST: 11/23/2018 22:49:54

Diagnostics Check	OK	22:42
Air Blank	0.000	22:43
Control Test	0.083	22:43
Air Blank	0.000	22:44
Subject Sample # 1	0.144	22:44
Breath Volume	1.785	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.143	22:48
Breath Volume	1.832	22:48
Air Blank	0.000	22:48
Control Test	0.082	22:49
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: MENARD

DATE/TIME OF TEST: 11/23/2018 23:49:28

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:46	
Control Test	0.082	23:46	
Air Blank	0.000	23:47	
Subject Sample	# 1	23:48	Subject Test Refused
Breath Volume	0.000	23:48	
Air Blank	0.000	23:48	
Control Test	0.082	23:48	
Air Blank	0.000	23:49	
Diagnostics Check	OK	23:49	

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY, MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE, BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: MOSHTAGH

DATE/TIME OF TEST: 11/24/2018 01:22:08

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.082	01:19	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:20	Subject Test Refused
Breath Volume	0.000	01:20	
Air Blank	0.000	01:21	
Control Test	0.082	01:21	
Air Blank	0.000	01:22	
Diagnostics Check	OK	01:22	

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY, MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROSADO, IAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SAVOY

DATE/TIME OF TEST: 11/24/2018 01:31:24

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:28	
Control Test	0.081	01:29	
Air Blank	0.000	01:29	
Subject Sample	# 1	01:30	Subject Test Refused
Breath Volume	0.000	01:30	
Air Blank	0.000	01:30	
Control Test	0.081	01:30	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: SHUNK

DATE/TIME OF TEST: 11/24/2018 02:01:09

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.082	01:59	
Air Blank	0.000	01:59	
Subject Sample	# 1	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	02:00	
Control Test	0.081	02:00	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: LUTZ

DATE/TIME OF TEST: 11/24/2018 05:11:31

Diagnostics Check	OK	05:03
Air Blank	0.000	05:04
Control Test	0.083	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.000	05:06
Breath Volume	1.746	05:06
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.000	05:10
Breath Volume	1.535	05:10
Air Blank	0.000	05:10
Control Test	0.084	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,M,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: STILES,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:16

SUBJECT'S LAST NAME: WINN

DATE/TIME OF TEST: 11/24/2018 16:39:15

Diagnostics Check	OK	16:36	
Air Blank	0.000	16:36	
Control Test	0.082	16:37	
Air Blank	0.000	16:37	
Subject Sample	# 1	16:37	Subject Test Refused
Breath Volume	0.000	16:37	
Air Blank	0.000	16:38	
Control Test	0.082	16:38	
Air Blank	0.000	16:39	
Diagnostics Check	OK	16:39	

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBY,LARRY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: SIGLER

DATE/TIME OF TEST: 11/24/2018 22:26:48

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.083	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.089	22:22
Breath Volume	2.285	22:22
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.097	22:25
Breath Volume	2.992	22:25
Air Blank	0.000	22:25
Control Test	0.081	22:26
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 11/25/2018 01:54:12

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.083	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.000	01:49
Breath Volume	2.019	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.000	01:52
Breath Volume	2.230	01:52
Air Blank	0.000	01:53
Control Test	0.082	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANISZEWSKI,ANDREW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: VITULLI

DATE/TIME OF TEST: 11/26/2018 01:54:31

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.083	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.077	01:49
Breath Volume	3.292	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.080	01:52
Breath Volume	4.085	01:52
Air Blank	0.000	01:53
Control Test	0.081	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHIEBNER,JAMIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: MEHAFFEY

DATE/TIME OF TEST: 11/28/2018 15:56:54

Diagnostics Check	OK	15:50
Air Blank	0.000	15:51
Control Test	0.084	15:51
Air Blank	0.000	15:51
Subject Sample # 1	0.000	15:52
Breath Volume	2.089	15:52
Air Blank	0.000	15:52
Air Blank	0.000	15:54
Subject Sample # 2	0.000	15:55
Breath Volume	1.832	15:55
Air Blank	0.000	15:55
Control Test	0.083	15:56
Air Blank	0.000	15:56
Diagnostics Check	OK	15:56

Data Download Date/Time: 11/28/2018 22:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GARDNER,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: SOLTAU

DATE/TIME OF TEST: 11/29/2018 03:54:34

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:43	
Control Test	0.083	03:43	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:47	No Sample Provided
Breath Volume	0.000	03:47	
Air Blank	0.000	03:47	
Air Blank	0.000	03:49	
Subject Sample	# 2	03:52	Volume Not Met
Breath Volume	0.023	03:52	
Air Blank	0.000	03:53	
Control Test	0.084	03:53	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 12/26/2018 20:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GARDNER,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: SOLTAU

DATE/TIME OF TEST: 11/29/2018 04:08:30

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:57	
Control Test	0.083	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 1 0.107	04:01	Volume Not Met
Breath Volume	0.121	04:01	
Air Blank	0.000	04:01	
Air Blank	0.000	04:03	
Subject Sample	# 2 0.102	04:06	Volume Not Met
Breath Volume	0.371	04:06	
Air Blank	0.000	04:07	
Control Test	0.082	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 12/26/2018 20:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: NEELD

DATE/TIME OF TEST: 11/30/2018 00:27:52

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:25	
Control Test	0.083	00:25	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:26	Subject Test Refused
Breath Volume	0.000	00:26	
Air Blank	0.000	00:26	
Control Test	0.083	00:27	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 12/26/2018 20:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 11/30/2018 05:59:06

Diagnostics Check	OK	05:56	
Air Blank	0.000	05:56	
Control Test	0.083	05:56	
Air Blank	0.000	05:57	
Subject Sample	# 1	05:57	Subject Test Refused
Breath Volume	0.000	05:57	
Air Blank	0.000	05:58	
Control Test	0.083	05:58	
Air Blank	0.000	05:59	
Diagnostics Check	OK	05:59	

Data Download Date/Time: 12/26/2018 20:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,E,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: WHITING

DATE/TIME OF TEST: 11/30/2018 00:21:22

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:18	
Control Test	0.082	00:19	
Air Blank	0.000	00:19	
Subject Sample	# 1	00:20	Subject Test Refused
Breath Volume	0.000	00:20	
Air Blank	0.000	00:20	
Control Test	0.081	00:20	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 12/26/2018 22:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: BUCHS

DATE/TIME OF TEST: 11/30/2018 01:45:32

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.081	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.145	01:40
Breath Volume	1.839	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.140	01:44
Breath Volume	1.492	01:44
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 12/26/2018 22:15