

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,GONZALO,
BREATH TEST OPERATOR AGENCY: Medley PD
ARRESTING OFFICER: GARCIA,GONZALO
ARRESTING OFFICER AGENCY: Medley PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:27

SUBJECT'S LAST NAME: PRIMUS

DATE/TIME OF TEST: 12/12/2018 09:56:46

Diagnostics Check	OK	09:53	
Air Blank	0.000	09:54	
Control Test	0.077	09:54	
Air Blank	0.000	09:54	
Subject Sample	# 1	09:55	Subject Test Refused
Breath Volume	0.000	09:55	
Air Blank	0.000	09:55	
Control Test	0.079	09:56	
Air Blank	0.000	09:56	
Diagnostics Check	OK	09:56	

Data Download Date/Time: 01/10/2019 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUERTA,EDWIN,
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: HUERTA,EDWIN
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/30/2018 01:37:48

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:32	
Control Test	0.077	01:32	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:36	Subject Test Refused
Breath Volume	0.000	01:36	
Air Blank	0.000	01:36	
Control Test	0.078	01:37	
Air Blank	0.000	01:37	
Diagnostics Check	OK	01:37	

Data Download Date/Time: 01/10/2019 12:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROIG,DAWN,M
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: ROIG,DAWN
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 12/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: ST PHARD

DATE/TIME OF TEST: 12/16/2018 04:28:44

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.078	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.129	04:24
Breath Volume	2.023	04:24
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.124	04:27
Breath Volume	2.191	04:27
Air Blank	0.000	04:27
Control Test	0.078	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 01/11/2019 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTIS,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: CURTIS,WILLIAM
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 12/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: VILLA

DATE/TIME OF TEST: 12/16/2018 09:22:12

Diagnostics Check	OK	09:10	
Air Blank	0.000	09:11	
Control Test	0.079	09:11	
Air Blank	0.000	09:12	
Subject Sample	# 1	09:15	No Sample Provided
Breath Volume	0.000	09:15	
Air Blank	0.000	09:15	
Air Blank	0.000	09:17	
Subject Sample	# 2	09:20	No Sample Provided
Breath Volume	0.000	09:20	
Air Blank	0.000	09:21	
Control Test	0.079	09:21	
Air Blank	0.000	09:22	
Diagnostics Check	OK	09:22	

Data Download Date/Time: 01/11/2019 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELFNER,ISAAC,S
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: BURNS,JEFFERY
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 12/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/22/2018 17:07:39

Diagnostics Check	OK	17:01
Air Blank	0.000	17:01
Control Test	0.079	17:01
Air Blank	0.000	17:02
Subject Sample # 1	0.140	17:03
Breath Volume	4.214	17:03
Air Blank	0.000	17:03
Air Blank	0.000	17:05
Subject Sample # 2	0.127	17:06
Breath Volume	1.679	17:06
Air Blank	0.000	17:06
Control Test	0.078	17:07
Air Blank	0.000	17:07
Diagnostics Check	OK	17:07

Data Download Date/Time: 01/11/2019 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONETZ,JONATHAN,D
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: ESPADA,ROME
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 12/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: RAMIREZ ARTEAGA

DATE/TIME OF TEST: 12/25/2018 06:32:39

Diagnostics Check	OK	06:23	
Air Blank	0.000	06:23	
Control Test	0.078	06:24	
Air Blank	0.000	06:24	
Subject Sample	# 1	06:27	No Sample Provided
Breath Volume	0.000	06:27	
Air Blank	0.000	06:28	
Air Blank	0.000	06:30	
Subject Sample	# 2	06:31	Subject Test Refused
Breath Volume	0.000	06:31	
Air Blank	0.000	06:31	
Control Test	0.078	06:32	
Air Blank	0.000	06:32	
Diagnostics Check	OK	06:32	

Data Download Date/Time: 01/11/2019 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROIG,DAWN,M
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: ROIG,DAWN
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 12/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: MAULTSBY

DATE/TIME OF TEST: 12/26/2018 01:11:38

Diagnostics Check	OK	01:08	
Air Blank	0.000	01:09	
Control Test	0.078	01:09	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:10	Subject Test Refused
Breath Volume	0.000	01:10	
Air Blank	0.000	01:10	
Control Test	0.078	01:11	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 01/11/2019 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 12/02/2018 04:59:10

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.080	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.134	04:53
Breath Volume	1.660	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:55
Subject Sample # 2	0.132	04:57
Breath Volume	1.789	04:57
Air Blank	0.000	04:58
Control Test	0.078	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 12/13/2018 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: CACERES CORAL

DATE/TIME OF TEST: 12/03/2018 02:27:24

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.162	02:22
Breath Volume	1.996	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.161	02:25
Breath Volume	2.871	02:25
Air Blank	0.000	02:26
Control Test	0.076	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 12/13/2018 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: WALKIN

DATE/TIME OF TEST: 12/04/2018 04:55:46

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.079	04:49
Air Blank	0.000	04:49
Subject Sample # 1	0.166	04:50
Breath Volume	1.812	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.167	04:53
Breath Volume	2.320	04:53
Air Blank	0.000	04:54
Control Test	0.077	04:55
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 12/13/2018 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/06/2018 01:43:35

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.081	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.192	01:39
Breath Volume	1.945	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:41
Subject Sample # 2	0.189	01:41
Breath Volume	1.285	01:41
Air Blank	0.000	01:42
Control Test	0.078	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 12/13/2018 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,JAVIER,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/08/2018 02:28:29

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.080	02:26	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	
Control Test	0.081	02:27	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 12/13/2018 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: MARKS

DATE/TIME OF TEST: 12/08/2018 02:58:05

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.000	02:53
Breath Volume	1.105	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.000	02:56
Breath Volume	2.074	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 12/13/2018 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 12/08/2018 04:28:25

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.081	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.109	04:23
Breath Volume	1.750	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:25
Subject Sample # 2	0.100	04:26
Breath Volume	1.308	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 12/13/2018 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: FERNANDEZ,ORLANDO
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: OPORTA-CONRADO

DATE/TIME OF TEST: 12/08/2018 19:59:00

Diagnostics Check	OK	19:55	
Air Blank	0.000	19:56	
Control Test	0.080	19:56	
Air Blank	0.000	19:56	
Subject Sample	# 1	19:57	Subject Test Refused
Breath Volume	0.000	19:57	
Air Blank	0.000	19:58	
Control Test	0.080	19:58	
Air Blank	0.000	19:58	
Diagnostics Check	OK	19:59	

Data Download Date/Time: 12/13/2018 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,JAVIER,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: VEGA-IBARROLA

DATE/TIME OF TEST: 12/08/2018 22:40:38

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.080	22:33
Air Blank	0.000	22:34
Subject Sample # 1	0.140	22:35
Breath Volume	1.765	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:37
Subject Sample # 2	0.138	22:38
Breath Volume	2.503	22:38
Air Blank	0.000	22:39
Control Test	0.079	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 12/13/2018 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: GUIDO

DATE/TIME OF TEST: 12/09/2018 01:39:31

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.080	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.138	01:34
Breath Volume	1.347	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:36
Subject Sample # 2	0.142	01:37
Breath Volume	1.378	01:37
Air Blank	0.000	01:38
Control Test	0.078	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 12/13/2018 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: CORTES,OSMIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: FLORES-RODRIGUEZ

DATE/TIME OF TEST: 12/09/2018 02:27:43

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.190	02:22
Breath Volume	2.308	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.191	02:25
Breath Volume	1.156	02:25
Air Blank	0.000	02:26
Control Test	0.077	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 12/13/2018 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/09/2018 03:41:15

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.213	03:36
Breath Volume	1.519	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:38
Subject Sample # 2	0.200	03:39
Breath Volume	1.562	03:39
Air Blank	0.000	03:40
Control Test	0.077	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 12/13/2018 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: PIERRE-SAINT

DATE/TIME OF TEST: 12/09/2018 07:37:59

Diagnostics Check	OK	07:31
Air Blank	0.000	07:32
Control Test	0.079	07:32
Air Blank	0.000	07:32
Subject Sample # 1	0.200	07:33
Breath Volume	3.269	07:33
Air Blank	0.000	07:34
Air Blank	0.000	07:35
Subject Sample # 2	0.200	07:36
Breath Volume	2.468	07:36
Air Blank	0.000	07:37
Control Test	0.079	07:37
Air Blank	0.000	07:37
Diagnostics Check	OK	07:37

Data Download Date/Time: 12/13/2018 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 12/11/2018 00:31:34

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.080	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.207	00:26
Breath Volume	3.660	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.203	00:29
Breath Volume	1.492	00:29
Air Blank	0.000	00:30
Control Test	0.078	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 12/13/2018 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 12/11/2018 19:02:04

Diagnostics Check	OK	18:55
Air Blank	0.000	18:55
Control Test	0.080	18:56
Air Blank	0.000	18:56
Subject Sample # 1	0.241	18:57
Breath Volume	1.125	18:57
Air Blank	0.000	18:58
Air Blank	0.000	18:59
Subject Sample # 2	0.238	19:00
Breath Volume	1.976	19:00
Air Blank	0.000	19:01
Control Test	0.079	19:01
Air Blank	0.000	19:01
Diagnostics Check	OK	19:02

Data Download Date/Time: 12/13/2018 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: ALBELO

DATE/TIME OF TEST: 12/14/2018 20:54:14

Diagnostics Check	OK	20:47
Air Blank	0.000	20:47
Control Test	0.080	20:47
Air Blank	0.000	20:48
Subject Sample # 1	0.194	20:49
Breath Volume	1.261	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:51
Subject Sample # 2	0.192	20:52
Breath Volume	1.742	20:52
Air Blank	0.000	20:53
Control Test	0.076	20:53
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 01/04/2019 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: SERRA JR

DATE/TIME OF TEST: 12/14/2018 23:47:50

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:45	
Control Test	0.079	23:45	
Air Blank	0.000	23:46	
Subject Sample	# 1	23:46	Subject Test Refused
Breath Volume	0.000	23:46	
Air Blank	0.000	23:46	
Control Test	0.079	23:47	
Air Blank	0.000	23:47	
Diagnostics Check	OK	23:47	

Data Download Date/Time: 01/04/2019 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: RIVERA CASTRO

DATE/TIME OF TEST: 12/16/2018 00:38:52

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:28	
Control Test	0.080	00:28	
Air Blank	0.000	00:28	
Subject Sample	# 1 0.142	00:32	Volume Not Met
Breath Volume	0.042	00:32	
Air Blank	0.000	00:33	
Air Blank	0.000	00:34	
Subject Sample	# 2 0.131	00:37	Volume Not Met
Breath Volume	0.074	00:37	
Air Blank		00:38	Improper Sample
Air Blank	0.000	00:38	

Data Download Date/Time: 01/04/2019 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: RIVERA CASTRO

DATE/TIME OF TEST: 12/16/2018 00:52:50

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:42	
Control Test	0.077	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1 0.145	00:45	Volume Not Met
Breath Volume	0.285	00:45	
Air Blank	0.000	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 2 0.131	00:51	Volume Not Met
Breath Volume	0.054	00:51	
Air Blank		00:52	Purge Fail
Air Blank	0.000	00:52	

Data Download Date/Time: 01/04/2019 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: RIVERA CASTRO

DATE/TIME OF TEST: 12/16/2018 00:57:10

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:56	
Control Test	0.074	00:56	Control Outside Tolerance
Air Blank	0.000	00:57	

Data Download Date/Time: 01/04/2019 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: RIVERA CASTRO

DATE/TIME OF TEST: 12/16/2018 01:16:34

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:00	
Control Test	0.078	01:00	
Air Blank	0.000	01:01	
Subject Sample	# 1 0.134	01:03	
Breath Volume	2.285	01:03	
Air Blank	0.000	01:04	
Air Blank	0.000	01:05	
Subject Sample	# 2 0.138	01:08	Volume Not Met
Breath Volume	0.097	01:08	
Air Blank	0.000	01:10	
Air Blank	0.000	01:11	
Subject Sample	# 3 0.134	01:14	Volume Not Met
Breath Volume	0.027	01:14	
Air Blank	0.000	01:15	
Control Test	0.077	01:15	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 01/04/2019 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: RIVERA CASTRO

DATE/TIME OF TEST: 12/16/2018 01:25:00

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:19	
Control Test	0.077	01:20	
Air Blank	0.000	01:20	
Subject Sample	# 1 0.149	01:23	Volume Not Met
Breath Volume	0.000	01:23	
Air Blank		01:24	Purge Fail
Air Blank	0.000	01:25	

Data Download Date/Time: 01/04/2019 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: RIVERA CASTRO

DATE/TIME OF TEST: 12/16/2018 01:42:53

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:27	
Control Test	0.075	01:28	
Air Blank	0.000	01:28	
Subject Sample	# 1 0.110	01:31	
Breath Volume	1.578	01:31	
Air Blank	0.000	01:32	
Air Blank	0.000	01:34	
Subject Sample	# 2 0.141	01:37	Volume Not Met
Breath Volume	0.128	01:37	
Air Blank	0.000	01:38	
Air Blank	0.000	01:39	
Subject Sample	# 3 0.123	01:41	
Breath Volume	1.945	01:41	
Air Blank	0.000	01:41	
Control Test	0.075	01:42	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 01/04/2019 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVALOS,DANIEL,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DAVALOS,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: DAVILA

DATE/TIME OF TEST: 12/17/2018 04:35:27

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.153	04:27
Breath Volume	2.839	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:29
Subject Sample # 2	0.176	04:30
Breath Volume	2.523	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:32
Subject Sample # 3	0.160	04:33
Breath Volume	2.164	04:33
Air Blank	0.000	04:34
Control Test	0.076	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 01/04/2019 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYA,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: PEREZ,ANDREA
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 12/19/2018 01:02:18

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:57	
Control Test	0.080	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1 0.277	01:00	
Breath Volume	2.777	01:00	
Air Blank		01:01	Purge Fail
Air Blank	0.000	01:02	

Data Download Date/Time: 01/04/2019 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYA,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: PEREZ,ANDREA
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 12/19/2018 01:10:56

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.076	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.278	01:06
Breath Volume	2.550	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:08
Subject Sample # 2	0.274	01:09
Breath Volume	1.820	01:09
Air Blank	0.000	01:09
Control Test	0.077	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 01/04/2019 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ROLLE,ERIKA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 12/20/2018 02:31:15

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.080	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.105	02:26
Breath Volume	1.894	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.109	02:29
Breath Volume	1.269	02:29
Air Blank	0.000	02:30
Control Test	0.077	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 01/04/2019 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYA,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: GIBSON,GERONIMO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CESPEDES

DATE/TIME OF TEST: 12/21/2018 02:22:58

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.135	02:18
Breath Volume	1.527	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:20
Subject Sample # 2	0.146	02:21
Breath Volume	1.328	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 01/04/2019 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUKSA,ANDREW,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: BECKER,ALEXA
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: STENSBY

DATE/TIME OF TEST: 12/21/2018 03:48:02

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:45	
Control Test	0.080	03:45	
Air Blank	0.000	03:46	
Subject Sample	# 1	03:46	Subject Test Refused
Breath Volume	0.000	03:46	
Air Blank	0.000	03:47	
Control Test	0.080	03:47	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 01/04/2019 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUAN,RALPH,B
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: CUAN,RALPH
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: RIVERO

DATE/TIME OF TEST: 12/21/2018 19:40:17

Diagnostics Check	OK	19:37	
Air Blank	0.000	19:37	
Control Test	0.080	19:38	
Air Blank	0.000	19:38	
Subject Sample	# 1	19:38	Subject Test Refused
Breath Volume	0.000	19:38	
Air Blank	0.000	19:39	
Control Test	0.080	19:39	
Air Blank	0.000	19:40	
Diagnostics Check	OK	19:40	

Data Download Date/Time: 01/04/2019 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUAN,RALPH,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: CUAN,RALPH
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MORALES-MEDINA

DATE/TIME OF TEST: 12/22/2018 04:29:32

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:27	
Control Test	0.080	04:27	
Air Blank	0.000	04:27	
Subject Sample	# 1	04:28	Subject Test Refused
Breath Volume	0.000	04:28	
Air Blank	0.000	04:28	
Control Test	0.080	04:28	
Air Blank	0.000	04:29	
Diagnostics Check	OK	04:29	

Data Download Date/Time: 01/04/2019 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, EMILIE, M
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: GARCIA, EMILIE
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PEREZ ESCALONA

DATE/TIME OF TEST: 12/23/2018 03:40:36

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.050	03:35
Breath Volume	2.222	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.048	03:38
Breath Volume	2.281	03:38
Air Blank	0.000	03:39
Control Test	0.078	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 01/04/2019 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: RAMIREZ GERMAN

DATE/TIME OF TEST: 12/25/2018 05:16:10

Diagnostics Check	OK	05:10	
Air Blank	0.000	05:10	
Control Test	0.080	05:11	
Air Blank	0.000	05:11	
Subject Sample	# 1 0.136	05:14	Volume Not Met
Breath Volume	0.574	05:14	
Air Blank		05:15	Purge Fail
Air Blank	0.000	05:16	

Data Download Date/Time: 01/04/2019 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: RAMIREZ GERMAN

DATE/TIME OF TEST: 12/25/2018 05:26:48

Diagnostics Check	OK	05:18
Air Blank	0.000	05:18
Control Test	0.077	05:19
Air Blank	0.000	05:19
Subject Sample # 1	0.161	05:20
Breath Volume	1.777	05:20
Air Blank	0.000	05:21
Air Blank	0.000	05:23
Subject Sample # 2	0.152	05:24
Breath Volume	2.960	05:24
Air Blank	0.000	05:25
Control Test	0.078	05:26
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 01/04/2019 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: AVILA

DATE/TIME OF TEST: 12/26/2018 00:19:56

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.128	00:15
Breath Volume	1.917	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:17
Subject Sample # 2	0.128	00:18
Breath Volume	2.132	00:18
Air Blank	0.000	00:19
Control Test	0.077	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 01/04/2019 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: NESBITT

DATE/TIME OF TEST: 12/27/2018 05:57:17

Diagnostics Check	OK	05:50
Air Blank	0.000	05:50
Control Test	0.080	05:51
Air Blank	0.000	05:51
Subject Sample # 1	0.147	05:52
Breath Volume	2.898	05:52
Air Blank	0.000	05:53
Air Blank	0.000	05:55
Subject Sample # 2	0.137	05:55
Breath Volume	1.457	05:55
Air Blank	0.000	05:56
Control Test	0.078	05:56
Air Blank	0.000	05:57
Diagnostics Check	OK	05:57

Data Download Date/Time: 01/04/2019 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATTORNO,ALEXANDER,F
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GATTORNO,ALEXANDER
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 12/27/2018 16:49:13

Diagnostics Check	OK	16:46	
Air Blank	0.000	16:46	
Control Test	0.080	16:46	
Air Blank	0.000	16:47	
Subject Sample	# 1	16:47	Subject Test Refused
Breath Volume	0.000	16:47	
Air Blank	0.000	16:48	
Control Test	0.079	16:48	
Air Blank	0.000	16:49	
Diagnostics Check	OK	16:49	

Data Download Date/Time: 01/04/2019 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 12/28/2018 02:41:48

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.180	02:36
Breath Volume	1.574	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:38
Subject Sample # 2	0.182	02:39
Breath Volume	1.148	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 01/04/2019 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MALAVE GARCIA

DATE/TIME OF TEST: 12/29/2018 02:18:28

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:16	
Control Test	0.080	02:16	
Air Blank	0.000	02:16	
Subject Sample	# 1	02:17	Subject Test Refused
Breath Volume	0.000	02:17	
Air Blank	0.000	02:17	
Control Test	0.080	02:17	
Air Blank	0.000	02:18	
Diagnostics Check	OK	02:18	

Data Download Date/Time: 01/04/2019 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/29/2018 04:33:16

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.081	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.134	04:28
Breath Volume	1.652	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:30
Subject Sample # 2	0.148	04:31
Breath Volume	1.882	04:31
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 01/04/2019 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DORIA

DATE/TIME OF TEST: 12/30/2018 02:17:00

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.050	02:12
Breath Volume	1.667	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.048	02:15
Breath Volume	1.808	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:17

Data Download Date/Time: 01/04/2019 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: DORIA

DATE/TIME OF TEST: 12/30/2018 02:52:46

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.037	02:48
Breath Volume	1.355	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.038	02:51
Breath Volume	1.914	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 01/04/2019 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:47

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 12/30/2018 07:15:23

Diagnostics Check	OK	07:08
Air Blank	0.000	07:08
Control Test	0.080	07:08
Air Blank	0.000	07:09
Subject Sample # 1	0.131	07:10
Breath Volume	1.546	07:10
Air Blank	0.000	07:11
Air Blank	0.000	07:12
Subject Sample # 2	0.136	07:13
Breath Volume	1.968	07:13
Air Blank	0.000	07:14
Control Test	0.080	07:14
Air Blank	0.000	07:15
Diagnostics Check	OK	07:15

Data Download Date/Time: 01/04/2019 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PATTERSON,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: SIDDELL

DATE/TIME OF TEST: 12/01/2018 02:59:02

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.082	02:54
Breath Volume	1.507	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.081	02:57
Breath Volume	1.277	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:59

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: MATCHETT

DATE/TIME OF TEST: 12/01/2018 05:55:40

Diagnostics Check	OK	05:52	
Air Blank	0.000	05:53	
Control Test	0.080	05:53	
Air Blank	0.000	05:54	
Subject Sample	# 1	05:54	Subject Test Refused
Breath Volume	0.000	05:54	
Air Blank	0.000	05:54	
Control Test	0.080	05:55	
Air Blank	0.000	05:55	
Diagnostics Check	OK	05:55	

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 12/01/2018 22:34:19

Diagnostics Check	OK	22:27
Air Blank	0.000	22:27
Control Test	0.080	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.291	22:29
Breath Volume	1.941	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:31
Subject Sample # 2	0.311	22:32
Breath Volume	2.105	22:32
Air Blank	0.000	22:33
Control Test	0.080	22:33
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILBERGER,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 12/02/2018 17:32:31

Diagnostics Check	OK	17:26
Air Blank	0.000	17:26
Control Test	0.081	17:26
Air Blank	0.000	17:27
Subject Sample # 1	0.060	17:27
Breath Volume	2.531	17:27
Air Blank	0.000	17:28
Air Blank	0.000	17:30
Subject Sample # 2	0.055	17:31
Breath Volume	1.804	17:31
Air Blank	0.000	17:31
Control Test	0.080	17:31
Air Blank	0.000	17:32
Diagnostics Check	OK	17:32

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 12/02/2018 20:41:02

Diagnostics Check	OK	20:34
Air Blank	0.000	20:35
Control Test	0.080	20:35
Air Blank	0.000	20:36
Subject Sample # 1	0.000	20:36
Breath Volume	3.210	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.000	20:39
Breath Volume	2.234	20:39
Air Blank	0.000	20:40
Control Test	0.080	20:40
Air Blank	0.000	20:40
Diagnostics Check	OK	20:41

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: FREEBECK

DATE/TIME OF TEST: 12/03/2018 03:01:09

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.080	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.189	02:56
Breath Volume	1.328	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.202	02:59
Breath Volume	1.261	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SALISBURY

DATE/TIME OF TEST: 12/03/2018 04:01:29

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.102	03:56
Breath Volume	3.507	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.103	04:00
Breath Volume	3.484	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 12/08/2018 00:27:34

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.081	00:22
Air Blank	0.000	00:22
Subject Sample # 1	0.106	00:23
Breath Volume	1.542	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.099	00:26
Breath Volume	1.351	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: BROUGHTON

DATE/TIME OF TEST: 12/08/2018 01:47:08

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.054	01:42
Breath Volume	2.285	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.056	01:45
Breath Volume	1.136	01:45
Air Blank	0.000	01:46
Control Test	0.081	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: HAYNES

DATE/TIME OF TEST: 12/08/2018 04:34:58

Diagnostics Check	OK	04:32	
Air Blank	0.000	04:32	
Control Test	0.081	04:32	
Air Blank	0.000	04:33	
Subject Sample	# 1	04:33	Subject Test Refused
Breath Volume	0.000	04:33	
Air Blank	0.000	04:34	
Control Test	0.081	04:34	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:34	

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SANTORO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:52

SUBJECT'S LAST NAME: POCHAMPALLY

DATE/TIME OF TEST: 12/08/2018 10:19:59

Diagnostics Check	OK	10:13
Air Blank	0.000	10:13
Control Test	0.081	10:14
Air Blank	0.000	10:14
Subject Sample # 1	0.118	10:15
Breath Volume	1.187	10:15
Air Blank	0.000	10:16
Air Blank	0.000	10:17
Subject Sample # 2	0.129	10:18
Breath Volume	2.695	10:18
Air Blank	0.000	10:19
Control Test	0.080	10:19
Air Blank	0.000	10:19
Diagnostics Check	OK	10:19

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:29

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 12/09/2018 16:56:34

Diagnostics Check	OK	16:50
Air Blank	0.000	16:50
Control Test	0.080	16:51
Air Blank	0.000	16:51
Subject Sample # 1	0.161	16:52
Breath Volume	1.527	16:52
Air Blank	0.000	16:52
Air Blank	0.000	16:54
Subject Sample # 2	0.172	16:55
Breath Volume	1.792	16:55
Air Blank	0.000	16:55
Control Test	0.081	16:55
Air Blank	0.000	16:56
Diagnostics Check	OK	16:56

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: TUCK

DATE/TIME OF TEST: 12/09/2018 21:30:26

Diagnostics Check	OK	21:26	
Air Blank	0.000	21:26	
Control Test	0.080	21:27	
Air Blank	0.000	21:27	
Subject Sample	# 1	21:29	Subject Test Refused
Breath Volume	0.000	21:29	
Air Blank	0.000	21:29	
Control Test	0.081	21:29	
Air Blank	0.000	21:30	
Diagnostics Check	OK	21:30	

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: CYPHER

DATE/TIME OF TEST: 12/09/2018 22:15:09

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.081	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.161	22:10
Breath Volume	1.199	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.157	22:13
Breath Volume	2.187	22:13
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,E,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FERNANDEZ,W
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: SCHENKEL

DATE/TIME OF TEST: 12/12/2018 16:21:13

Diagnostics Check	OK	16:15
Air Blank	0.000	16:15
Control Test	0.082	16:15
Air Blank	0.000	16:16
Subject Sample # 1	0.000	16:16
Breath Volume	1.480	16:16
Air Blank	0.000	16:17
Air Blank	0.000	16:19
Subject Sample # 2	0.000	16:19
Breath Volume	1.429	16:19
Air Blank	0.000	16:20
Control Test	0.081	16:20
Air Blank	0.000	16:21
Diagnostics Check	OK	16:21

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: MASHBURN

DATE/TIME OF TEST: 12/13/2018 06:27:44

Diagnostics Check	OK	06:21
Air Blank	0.000	06:21
Control Test	0.080	06:21
Air Blank	0.000	06:22
Subject Sample # 1	0.136	06:23
Breath Volume	3.585	06:23
Air Blank	0.000	06:23
Air Blank	0.000	06:25
Subject Sample # 2	0.134	06:26
Breath Volume	3.210	06:26
Air Blank	0.000	06:26
Control Test	0.081	06:27
Air Blank	0.000	06:27
Diagnostics Check	OK	06:27

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILBERGER,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: HEMPHILL

DATE/TIME OF TEST: 12/13/2018 22:38:00

Diagnostics Check	OK	22:35	
Air Blank	0.000	22:35	
Control Test	0.080	22:35	
Air Blank	0.000	22:36	
Subject Sample	# 1	22:36	Subject Test Refused
Breath Volume	0.000	22:36	
Air Blank	0.000	22:37	
Control Test	0.080	22:37	
Air Blank	0.000	22:37	
Diagnostics Check	OK	22:38	

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DYER,W
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 12/14/2018 22:06:44

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:55	
Control Test	0.080	21:56	
Air Blank	0.000	21:56	
Subject Sample	# 1	21:59	No Sample Provided
Breath Volume	0.000	21:59	
Air Blank	0.000	22:00	
Air Blank	0.000	22:02	
Subject Sample	# 2	22:05	No Sample Provided
Breath Volume	0.000	22:05	
Air Blank	0.000	22:05	
Control Test	0.081	22:06	
Air Blank	0.000	22:06	
Diagnostics Check	OK	22:06	

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FUCCI,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:38

SUBJECT'S LAST NAME: RESURRECCION

DATE/TIME OF TEST: 12/16/2018 08:08:17

Diagnostics Check	OK	07:59
Air Blank	0.000	08:00
Control Test	0.081	08:00
Air Blank	0.000	08:00
Subject Sample # 1	0.151	08:02
Breath Volume	1.714	08:02
Air Blank	0.000	08:03
Air Blank	0.000	08:05
Subject Sample # 2	0.156	08:06
Breath Volume	1.390	08:06
Air Blank	0.000	08:07
Control Test	0.080	08:07
Air Blank	0.000	08:08
Diagnostics Check	OK	08:08

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: HUFFMAN

DATE/TIME OF TEST: 12/17/2018 22:35:29

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:33	
Control Test	0.080	22:33	
Air Blank	0.000	22:33	
Subject Sample	# 1	22:34	Subject Test Refused
Breath Volume	0.000	22:34	
Air Blank	0.000	22:34	
Control Test	0.080	22:34	
Air Blank	0.000	22:35	
Diagnostics Check	OK	22:35	

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER, JOSHUA, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMAS, J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 12/19/2018 19:25:45

Diagnostics Check	OK	19:22	
Air Blank	0.000	19:23	
Control Test	0.081	19:23	
Air Blank	0.000	19:24	
Subject Sample	# 1	19:24	Subject Test Refused
Breath Volume	0.000	19:24	
Air Blank	0.000	19:24	
Control Test	0.080	19:25	
Air Blank	0.000	19:25	
Diagnostics Check	OK	19:25	

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER, JOSHUA, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS, B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: KAHLE

DATE/TIME OF TEST: 12/19/2018 22:21:16

Diagnostics Check	OK	22:15
Air Blank	0.000	22:15
Control Test	0.080	22:16
Air Blank	0.000	22:16
Subject Sample # 1	0.000	22:16
Breath Volume	3.117	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.000	22:19
Breath Volume	1.718	22:19
Air Blank	0.000	22:20
Control Test	0.081	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 12/21/2018 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GONYEAU,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:53

SUBJECT'S LAST NAME: MERCED

DATE/TIME OF TEST: 12/21/2018 17:16:25

Diagnostics Check	OK	17:13	
Air Blank	0.000	17:14	
Control Test	0.081	17:14	
Air Blank	0.000	17:14	
Subject Sample	# 1	17:15	Subject Test Refused
Breath Volume	0.000	17:15	
Air Blank	0.000	17:15	
Control Test	0.081	17:15	
Air Blank	0.000	17:16	
Diagnostics Check	OK	17:16	

Data Download Date/Time: 01/22/2019 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: GAINNEY

DATE/TIME OF TEST: 12/21/2018 22:01:56

Diagnostics Check	OK	21:59	
Air Blank	0.000	21:59	
Control Test	0.080	21:59	
Air Blank	0.000	22:00	
Subject Sample	# 1	22:00	Subject Test Refused
Breath Volume	0.000	22:00	
Air Blank	0.000	22:01	
Control Test	0.080	22:01	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 01/22/2019 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER, JOSHUA, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BORCHART, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 12/22/2018 08:23:11

Diagnostics Check	OK	08:20	
Air Blank	0.000	08:20	
Control Test	0.081	08:21	
Air Blank	0.000	08:21	
Subject Sample	# 1	08:21	Subject Test Refused
Breath Volume	0.000	08:21	
Air Blank	0.000	08:22	
Control Test	0.079	08:22	
Air Blank	0.000	08:23	
Diagnostics Check	OK	08:23	

Data Download Date/Time: 01/22/2019 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DRAGOMIR,I
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: TURNBULL

DATE/TIME OF TEST: 12/22/2018 19:49:34

Diagnostics Check	OK	19:43
Air Blank	0.000	19:44
Control Test	0.081	19:44
Air Blank	0.000	19:44
Subject Sample # 1	0.000	19:45
Breath Volume	2.449	19:45
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample # 2	0.000	19:48
Breath Volume	2.656	19:48
Air Blank	0.000	19:48
Control Test	0.081	19:48
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

Data Download Date/Time: 01/22/2019 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SPENCER,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: LEMAY

DATE/TIME OF TEST: 12/23/2018 16:19:28

Diagnostics Check	OK	16:13
Air Blank	0.000	16:13
Control Test	0.081	16:13
Air Blank	0.000	16:14
Subject Sample # 1	0.156	16:15
Breath Volume	2.144	16:15
Air Blank	0.000	16:15
Air Blank	0.000	16:17
Subject Sample # 2	0.147	16:17
Breath Volume	2.394	16:17
Air Blank	0.000	16:18
Control Test	0.080	16:18
Air Blank	0.000	16:19
Diagnostics Check	OK	16:19

Data Download Date/Time: 01/22/2019 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: GAGNE

DATE/TIME OF TEST: 12/23/2018 22:57:31

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:40	
Control Test	0.081	22:41	
Air Blank	0.000	22:41	
Subject Sample	# 1 0.151	22:44	
Breath Volume	2.804	22:44	
Air Blank	0.000	22:45	
Air Blank	0.000	22:47	
Subject Sample	# 2	22:50	No Sample Provided
Breath Volume	0.000	22:50	
Air Blank	0.000	22:51	
Air Blank	0.000	22:53	
Subject Sample	# 3	22:56	No Sample Provided
Breath Volume	0.000	22:56	
Air Blank	0.000	22:56	
Control Test	0.080	22:56	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 01/22/2019 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/26/2018 22:39:07

Diagnostics Check	OK	22:32
Air Blank	0.000	22:32
Control Test	0.081	22:33
Air Blank	0.000	22:33
Subject Sample # 1	0.174	22:34
Breath Volume	2.507	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.162	22:37
Breath Volume	1.121	22:37
Air Blank	0.000	22:38
Control Test	0.080	22:38
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 01/22/2019 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DASHER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: STRONG

DATE/TIME OF TEST: 12/27/2018 09:22:47

Diagnostics Check	OK	09:16
Air Blank	0.000	09:17
Control Test	0.081	09:17
Air Blank	0.000	09:18
Subject Sample # 1	0.000	09:18
Breath Volume	2.164	09:18
Air Blank	0.000	09:19
Air Blank	0.000	09:21
Subject Sample # 2	0.000	09:21
Breath Volume	1.917	09:21
Air Blank	0.000	09:21
Control Test	0.081	09:22
Air Blank	0.000	09:22
Diagnostics Check	OK	09:22

Data Download Date/Time: 01/22/2019 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:33

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/29/2018 11:59:25

Diagnostics Check	OK	11:53
Air Blank	0.000	11:53
Control Test	0.080	11:54
Air Blank	0.000	11:54
Subject Sample # 1	0.190	11:55
Breath Volume	2.089	11:55
Air Blank	0.000	11:55
Air Blank	0.000	11:57
Subject Sample # 2	0.185	11:57
Breath Volume	1.898	11:57
Air Blank	0.000	11:58
Control Test	0.079	11:58
Air Blank	0.000	11:59
Diagnostics Check	OK	11:59

Data Download Date/Time: 01/22/2019 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISSETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: WALIZER

DATE/TIME OF TEST: 12/29/2018 21:29:15

Diagnostics Check	OK	21:26	
Air Blank	0.000	21:26	
Control Test	0.080	21:27	
Air Blank	0.000	21:27	
Subject Sample	# 1	21:27	Subject Test Refused
Breath Volume	0.000	21:27	
Air Blank	0.000	21:28	
Control Test	0.080	21:28	
Air Blank	0.000	21:29	
Diagnostics Check	OK	21:29	

Data Download Date/Time: 01/22/2019 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RAY,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: BROWNLEE

DATE/TIME OF TEST: 12/30/2018 08:43:43

Diagnostics Check	OK	08:40	
Air Blank	0.000	08:41	
Control Test	0.080	08:41	
Air Blank	0.000	08:42	
Subject Sample	# 1	08:42	Subject Test Refused
Breath Volume	0.000	08:42	
Air Blank	0.000	08:42	
Control Test	0.080	08:43	
Air Blank	0.000	08:43	
Diagnostics Check	OK	08:43	

Data Download Date/Time: 01/22/2019 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SORRELLS,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MCRAE

DATE/TIME OF TEST: 12/05/2018 03:29:25

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.081	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.255	03:25
Breath Volume	2.875	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.237	03:27
Breath Volume	2.449	03:27
Air Blank	0.000	03:28
Control Test	0.081	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 12/21/2018 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ALVIN

DATE/TIME OF TEST: 12/06/2018 02:43:54

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.081	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:43	
Control Test	0.081	02:43	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 12/21/2018 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/08/2018 01:06:08

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.080	00:55	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:59	No Sample Provided
Breath Volume	0.000	00:59	
Air Blank	0.000	01:00	
Air Blank	0.000	01:02	
Subject Sample	# 2	01:04	Subject Test Refused
Breath Volume	0.000	01:04	
Air Blank	0.000	01:05	
Control Test	0.080	01:05	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 12/21/2018 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BICKNELL

DATE/TIME OF TEST: 12/08/2018 23:41:08

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.081	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.000	23:36
Breath Volume	1.175	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.000	23:39
Breath Volume	1.523	23:39
Air Blank	0.000	23:40
Control Test	0.080	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 12/21/2018 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 12/09/2018 05:07:23

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:04	
Control Test	0.081	05:05	
Air Blank	0.000	05:05	
Subject Sample	# 1	05:06	Subject Test Refused
Breath Volume	0.000	05:06	
Air Blank	0.000	05:06	
Control Test	0.080	05:06	
Air Blank	0.000	05:07	
Diagnostics Check	OK	05:07	

Data Download Date/Time: 12/21/2018 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,V,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOORE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: RIVERA LOPEZ

DATE/TIME OF TEST: 12/09/2018 22:32:38

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.156	22:28
Breath Volume	1.351	22:28
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.155	22:31
Breath Volume	2.664	22:31
Air Blank	0.000	22:31
Control Test	0.078	22:32
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 12/21/2018 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SORRELLS,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: RUIZ-HERNANDEZ

DATE/TIME OF TEST: 12/10/2018 01:22:02

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:19	
Control Test	0.080	01:19	
Air Blank	0.000	01:20	
Subject Sample	# 1	01:20	Subject Test Refused
Breath Volume	0.000	01:20	
Air Blank	0.000	01:21	
Control Test	0.080	01:21	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:22	

Data Download Date/Time: 12/21/2018 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WILLIAMS,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 12/10/2018 05:21:56

Diagnostics Check	OK	05:19	
Air Blank	0.000	05:19	
Control Test	0.081	05:19	
Air Blank	0.000	05:20	
Subject Sample	# 1	05:20	Subject Test Refused
Breath Volume	0.000	05:20	
Air Blank	0.000	05:21	
Control Test	0.081	05:21	
Air Blank	0.000	05:21	
Diagnostics Check	OK	05:21	

Data Download Date/Time: 12/21/2018 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: HINKLE

DATE/TIME OF TEST: 12/12/2018 02:09:23

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.081	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.073	02:05
Breath Volume	2.402	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.071	02:07
Breath Volume	2.234	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 12/21/2018 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: EVERETT

DATE/TIME OF TEST: 12/12/2018 02:48:44

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.080	02:46	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:47	Subject Test Refused
Breath Volume	0.000	02:47	
Air Blank	0.000	02:47	
Control Test	0.080	02:48	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

Data Download Date/Time: 12/21/2018 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 12/13/2018 23:30:54

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:27	
Control Test	0.080	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:29	Subject Test Refused
Breath Volume	0.000	23:29	
Air Blank	0.000	23:29	
Control Test	0.080	23:30	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 12/21/2018 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: KANIOSKY

DATE/TIME OF TEST: 12/15/2018 01:46:15

Diagnostics Check	OK	01:39
Air Blank	0.000	01:40
Control Test	0.077	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.212	01:41
Breath Volume	2.285	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.210	01:44
Breath Volume	1.875	01:44
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 12/21/2018 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 12/16/2018 00:21:17

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.166	00:16
Breath Volume	1.867	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.161	00:19
Breath Volume	2.257	00:19
Air Blank	0.000	00:20
Control Test	0.079	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 12/21/2018 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 12/16/2018 02:32:10

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.186	02:27
Breath Volume	1.246	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.193	02:30
Breath Volume	2.058	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 12/21/2018 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SORRELLS,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: SHELAR

DATE/TIME OF TEST: 12/19/2018 03:37:59

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.081	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.201	03:30
Breath Volume	2.335	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.177	03:33
Breath Volume	2.695	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 3	0.188	03:36
Breath Volume	2.687	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 12/21/2018 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: FEWELL

DATE/TIME OF TEST: 12/20/2018 01:44:20

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.081	01:42	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:43	
Control Test	0.080	01:43	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 12/21/2018 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LANE,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/20/2018 05:43:28

Diagnostics Check	OK	05:37
Air Blank	0.000	05:38
Control Test	0.080	05:38
Air Blank	0.000	05:38
Subject Sample # 1	0.000	05:39
Breath Volume	2.082	05:39
Air Blank	0.000	05:39
Air Blank	0.000	05:41
Subject Sample # 2	0.000	05:42
Breath Volume	2.210	05:42
Air Blank	0.000	05:42
Control Test	0.080	05:42
Air Blank	0.000	05:43
Diagnostics Check	OK	05:43

Data Download Date/Time: 12/21/2018 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,V,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PATTERSON,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: DODD

DATE/TIME OF TEST: 12/20/2018 19:59:46

Diagnostics Check	OK	19:50	
Air Blank	0.000	19:51	
Control Test	0.080	19:51	
Air Blank	0.000	19:51	
Subject Sample	# 1 0.329	19:52	
Breath Volume	2.878	19:52	
Air Blank	0.000	19:53	
Air Blank	0.000	19:54	
Subject Sample	# 2 0.350	19:55	
Breath Volume	3.109	19:55	
Air Blank	0.000	19:56	
Air Blank	0.000	19:57	
Subject Sample	# 3 0.324	19:58	
Breath Volume	2.535	19:58	
Air Blank	0.000	19:58	
Control Test	0.070	19:59	Control Outside Tolerance
Air Blank	0.000	19:59	

Data Download Date/Time: 12/21/2018 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: HINKLE

DATE/TIME OF TEST: 12/22/2018 00:30:52

Diagnostics Check	OK	00:24
Air Blank	0.000	00:25
Control Test	0.080	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.114	00:26
Breath Volume	1.609	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.104	00:29
Breath Volume	2.097	00:29
Air Blank	0.000	00:29
Control Test	0.080	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 01/22/2019 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: ROSALES

DATE/TIME OF TEST: 12/22/2018 01:25:13

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:22	
Control Test	0.081	01:23	
Air Blank	0.000	01:23	
Subject Sample	# 1	01:23	Subject Test Refused
Breath Volume	0.000	01:23	
Air Blank	0.000	01:24	
Control Test	0.081	01:24	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 01/22/2019 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: MOLINA ZELEYA

DATE/TIME OF TEST: 12/22/2018 04:39:04

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:38	
		04:38	RFI Detect
Air Blank	0.000	04:39	

Data Download Date/Time: 01/22/2019 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 12/22/2018 04:48:07

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.079	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.146	04:43
Breath Volume	1.527	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.152	04:46
Breath Volume	1.472	04:46
Air Blank	0.000	04:47
Control Test	0.080	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 01/22/2019 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSEN,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: KEY

DATE/TIME OF TEST: 12/23/2018 02:40:17

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.081	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.113	02:35
Breath Volume	3.187	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.106	02:38
Breath Volume	1.316	02:38
Air Blank	0.000	02:39
Control Test	0.081	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 01/22/2019 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MURPHY,G
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/23/2018 06:00:28

Diagnostics Check	OK	05:47
Air Blank	0.000	05:47
Control Test	0.081	05:47
Air Blank	0.000	05:48
Subject Sample # 1	0.181	05:49
Breath Volume	2.019	05:49
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.159	05:53
Breath Volume	1.417	05:53
Air Blank	0.000	05:54
Air Blank	0.000	05:56
Subject Sample # 3	0.165	05:58
Breath Volume	2.039	05:58
Air Blank	0.000	05:59
Control Test	0.081	05:59
Air Blank	0.000	06:00
Diagnostics Check	OK	06:00

Data Download Date/Time: 01/22/2019 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BOWERS,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 12/25/2018 02:44:57

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:42	
Control Test	0.081	02:42	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:44	
Control Test	0.081	02:44	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 01/22/2019 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 12/25/2018 04:18:37

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.081	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.245	04:14
Breath Volume	2.355	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.254	04:17
Breath Volume	2.839	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 01/22/2019 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BOSMA

DATE/TIME OF TEST: 12/27/2018 03:36:32

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.169	03:32
Breath Volume	1.304	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.176	03:35
Breath Volume	2.957	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 01/22/2019 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BUMGARNER,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 12/29/2018 02:43:44

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.080	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:42	
Control Test	0.080	02:43	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 01/22/2019 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: TISDALE

DATE/TIME OF TEST: 12/29/2018 05:47:24

Diagnostics Check	OK	05:44	
Air Blank	0.000	05:45	
Control Test	0.080	05:45	
Air Blank	0.000	05:45	
Subject Sample	# 1	05:46	Subject Test Refused
Breath Volume	0.000	05:46	
Air Blank	0.000	05:46	
Control Test	0.080	05:46	
Air Blank	0.000	05:47	
Diagnostics Check	OK	05:47	

Data Download Date/Time: 01/22/2019 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: SCRADER

DATE/TIME OF TEST: 12/30/2018 01:39:40

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.163	01:35
Breath Volume	3.324	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.155	01:38
Breath Volume	3.089	01:38
Air Blank	0.000	01:38
Control Test	0.080	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 01/22/2019 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: GOAR

DATE/TIME OF TEST: 12/30/2018 04:33:00

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.080	04:30	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:31	Subject Test Refused
Breath Volume	0.000	04:31	
Air Blank	0.000	04:32	
Control Test	0.081	04:32	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 01/22/2019 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 12/30/2018 05:15:28

Diagnostics Check	OK	05:12	
Air Blank	0.000	05:13	
Control Test	0.080	05:13	
Air Blank	0.000	05:13	
Subject Sample	# 1	05:14	Subject Test Refused
Breath Volume	0.000	05:14	
Air Blank	0.000	05:14	
Control Test	0.080	05:14	
Air Blank	0.000	05:15	
Diagnostics Check	OK	05:15	

Data Download Date/Time: 01/22/2019 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: LIKAR

DATE/TIME OF TEST: 12/31/2018 01:42:34

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:40	
Control Test	0.080	01:40	
Air Blank	0.000	01:41	
Subject Sample	# 1	01:41	Subject Test Refused
Breath Volume	0.000	01:41	
Air Blank	0.000	01:41	
Control Test	0.081	01:41	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 01/22/2019 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: KATO

DATE/TIME OF TEST: 12/31/2018 02:41:05

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:24	
Control Test	0.080	02:24	
Air Blank	0.000	02:25	
Subject Sample	# 1 0.154	02:28	Volume Not Met
Breath Volume	0.738	02:28	
Air Blank	0.000	02:28	
Air Blank	0.000	02:30	
Subject Sample	# 2 0.155	02:34	
Breath Volume	1.683	02:34	
Air Blank	0.000	02:34	
Air Blank	0.000	02:36	
Subject Sample	# 3	02:39	Subject Test Refused
Breath Volume	0.785	02:39	
Air Blank	0.000	02:40	
Control Test	0.079	02:40	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 01/22/2019 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TURPIN,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 12/31/2018 05:18:14

Diagnostics Check	OK	05:11
Air Blank	0.000	05:12
Control Test	0.080	05:12
Air Blank	0.000	05:13
Subject Sample # 1	0.175	05:13
Breath Volume	2.367	05:13
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.174	05:16
Breath Volume	2.488	05:16
Air Blank	0.000	05:17
Control Test	0.079	05:17
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 01/22/2019 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JONES,S
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/02/2018 01:47:59

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:44	
Control Test	0.081	01:45	
Air Blank	0.000	01:45	
Subject Sample	# 1	01:46	Subject Test Refused
Breath Volume	0.000	01:46	
Air Blank	0.000	01:47	
Control Test	0.082	01:47	
Air Blank	0.000	01:47	
Diagnostics Check	OK	01:47	

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: REYNOLDS,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: LYLE

DATE/TIME OF TEST: 12/02/2018 03:34:08

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.082	03:32	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:33	
Control Test	0.081	03:33	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOLARD,RAYMOND,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: QUICK,WILFRED
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: HOLYCROSS

DATE/TIME OF TEST: 12/02/2018 18:29:02

Diagnostics Check	OK	18:22
Air Blank	0.000	18:23
Control Test	0.082	18:23
Air Blank	0.000	18:24
Subject Sample # 1	0.287	18:24
Breath Volume	3.277	18:24
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 2	0.283	18:27
Breath Volume	3.824	18:27
Air Blank	0.000	18:28
Control Test	0.080	18:28
Air Blank	0.000	18:28
Diagnostics Check	OK	18:29

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCracken,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: GODBOLD

DATE/TIME OF TEST: 12/03/2018 22:32:39

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.081	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.140	22:27
Breath Volume	1.105	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.151	22:31
Breath Volume	1.886	22:31
Air Blank	0.000	22:31
Control Test	0.080	22:32
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCracken,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: Copher,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: RAWLINS

DATE/TIME OF TEST: 12/04/2018 01:44:28

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.114	01:39
Breath Volume	1.574	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.115	01:42
Breath Volume	2.878	01:42
Air Blank	0.000	01:43
Control Test	0.081	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KELLEY,S
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: BLANKENSHIP

DATE/TIME OF TEST: 12/06/2018 18:20:50

Diagnostics Check	OK	18:14
Air Blank	0.000	18:14
Control Test	0.082	18:14
Air Blank	0.000	18:15
Subject Sample # 1	0.171	18:16
Breath Volume	2.093	18:16
Air Blank	0.000	18:16
Air Blank	0.000	18:18
Subject Sample # 2	0.158	18:19
Breath Volume	1.585	18:19
Air Blank	0.000	18:19
Control Test	0.080	18:20
Air Blank	0.000	18:20
Diagnostics Check	OK	18:20

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCracken,JD,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: LAYFIELD

DATE/TIME OF TEST: 12/08/2018 03:27:04

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.082	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.092	03:22
Breath Volume	2.214	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.103	03:25
Breath Volume	2.660	03:25
Air Blank	0.000	03:26
Control Test	0.082	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:27

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCracken,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: TERRY,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 12/08/2018 04:03:22

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.082	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.067	03:58
Breath Volume	2.730	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.061	04:01
Breath Volume	1.488	04:01
Air Blank	0.000	04:02
Control Test	0.081	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCracken,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: 2C16,KNIGHT
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BARRY

DATE/TIME OF TEST: 12/08/2018 23:18:59

Diagnostics Check	OK	23:12
Air Blank	0.000	23:13
Control Test	0.081	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.218	23:14
Breath Volume	2.113	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.233	23:17
Breath Volume	3.546	23:17
Air Blank	0.000	23:18
Control Test	0.081	23:18
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PYE

DATE/TIME OF TEST: 12/09/2018 03:39:20

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:36	
Control Test	0.081	03:37	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:38	
Control Test	0.081	03:38	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCracken,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WALTON,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: WYNN

DATE/TIME OF TEST: 12/10/2018 00:21:17

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.081	00:15
Air Blank	0.000	00:15
Subject Sample # 1	0.232	00:16
Breath Volume	1.511	00:16
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.234	00:19
Breath Volume	2.234	00:19
Air Blank	0.000	00:20
Control Test	0.080	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCracken,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: TERRY,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 12/10/2018 02:51:37

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:49	
Control Test	0.081	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:50	Subject Test Refused
Breath Volume	0.000	02:50	
Air Blank	0.000	02:50	
Control Test	0.081	02:51	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: CIPHER,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/10/2018 03:01:46

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.081	02:57	
Air Blank	0.000	02:58	
Subject Sample	# 1	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:00	
Control Test	0.081	03:01	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCracken,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 12/10/2018 03:51:53

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.178	03:47
Breath Volume	2.097	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:49
Subject Sample # 2	0.182	03:50
Breath Volume	1.906	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOLARD, RAYMOND, C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KELLY, SPENCER
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: BOGOSH

DATE/TIME OF TEST: 12/15/2018 16:45:35

Diagnostics Check	OK	16:42	
Air Blank	0.000	16:43	
Control Test	0.081	16:43	
Air Blank	0.000	16:44	
Subject Sample	# 1	16:44	Subject Test Refused
Breath Volume	0.000	16:44	
Air Blank	0.000	16:44	
Control Test	0.081	16:45	
Air Blank	0.000	16:45	
Diagnostics Check	OK	16:45	

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: REYONLDS,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: BARCARO

DATE/TIME OF TEST: 12/16/2018 21:54:46

Diagnostics Check	OK	21:46
Air Blank	0.000	21:47
Control Test	0.082	21:47
Air Blank	0.000	21:48
Subject Sample # 1	0.162	21:50
Breath Volume	2.675	21:50
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.160	21:53
Breath Volume	3.054	21:53
Air Blank	0.000	21:53
Control Test	0.080	21:54
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/18/2018 04:17:44

Diagnostics Check	OK	04:14	
Air Blank	0.000	04:15	
Control Test	0.082	04:15	
Air Blank	0.000	04:16	
Subject Sample	# 1	04:16	Subject Test Refused
Breath Volume	0.000	04:16	
Air Blank	0.000	04:16	
Control Test	0.082	04:17	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:17	

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: CHANDLER-REINSCH

DATE/TIME OF TEST: 12/18/2018 19:43:08

Diagnostics Check	OK	19:36
Air Blank	0.000	19:36
Control Test	0.082	19:37
Air Blank	0.000	19:37
Subject Sample # 1	0.012	19:38
Breath Volume	1.351	19:38
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.012	19:41
Breath Volume	1.289	19:41
Air Blank	0.000	19:42
Control Test	0.082	19:42
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,CATHERINE,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GOETHE,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 12/19/2018 23:16:27

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:03	
Control Test	0.081	23:03	
Air Blank	0.000	23:04	
Subject Sample	# 1 0.203	23:07	Volume Not Met
Breath Volume	0.203	23:07	
Air Blank	0.000	23:07	
Air Blank	0.000	23:09	
Subject Sample	# 2 0.181	23:11	
Breath Volume	2.542	23:11	
Air Blank	0.000	23:11	
Air Blank	0.000	23:13	
Subject Sample	# 3 0.179	23:14	
Breath Volume	1.621	23:14	
Air Blank	0.000	23:15	
Control Test	0.079	23:15	
Air Blank	0.000	23:16	
Diagnostics Check	OK	23:16	

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GOETHE,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: FROST

DATE/TIME OF TEST: 12/20/2018 05:38:18

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.082	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.136	05:33
Breath Volume	1.250	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.139	05:36
Breath Volume	1.164	05:36
Air Blank	0.000	05:37
Control Test	0.079	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GOETHE,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: FROST

DATE/TIME OF TEST: 12/20/2018 07:36:22

Diagnostics Check	OK	07:24	
Air Blank	0.000	07:24	
Control Test	0.080	07:24	
Air Blank	0.000	07:25	
Subject Sample	# 1 0.096	07:28	Volume Not Met
Breath Volume	0.875	07:28	
Air Blank	0.000	07:28	
Air Blank	0.000	07:30	
Subject Sample	# 2 0.104	07:31	
Breath Volume	1.242	07:31	
Air Blank	0.000	07:32	
Air Blank	0.000	07:34	
Subject Sample	# 3 0.108	07:34	
Breath Volume	1.253	07:34	
Air Blank	0.000	07:35	
Control Test	0.078	07:35	
Air Blank	0.000	07:36	
Diagnostics Check	OK	07:36	

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,B,K
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: BUNCH

DATE/TIME OF TEST: 12/20/2018 19:55:36

Diagnostics Check	OK	19:48
Air Blank	0.000	19:49
Control Test	0.081	19:49
Air Blank	0.000	19:50
Subject Sample # 1	0.174	19:51
Breath Volume	1.859	19:51
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.171	19:54
Breath Volume	1.566	19:54
Air Blank	0.000	19:54
Control Test	0.080	19:55
Air Blank	0.000	19:55
Diagnostics Check	OK	19:55

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: REYNOLDS,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: DUGGER

DATE/TIME OF TEST: 12/20/2018 22:05:39

Diagnostics Check	OK	21:59
Air Blank	0.000	22:00
Control Test	0.081	22:00
Air Blank	0.000	22:00
Subject Sample # 1	0.000	22:01
Breath Volume	1.375	22:01
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.000	22:04
Breath Volume	1.714	22:04
Air Blank	0.000	22:04
Control Test	0.081	22:05
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCracken,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 12/21/2018 21:23:32

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:11	
Control Test	0.082	21:11	
Air Blank	0.000	21:11	
Subject Sample # 1	0.000	21:13	
Breath Volume	2.644	21:13	
Air Blank	0.000	21:13	
Air Blank	0.000	21:15	
Subject Sample # 2		21:18	No Sample Provided
Breath Volume	0.000	21:18	
Air Blank	0.000	21:19	
Air Blank	0.000	21:21	
Subject Sample # 3	0.000	21:22	
Breath Volume	1.265	21:22	
Air Blank	0.000	21:22	
Control Test	0.081	21:22	
Air Blank	0.000	21:23	
Diagnostics Check	OK	21:23	

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: REYNOLDS,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/25/2018 21:21:09

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:18	
Control Test	0.082	21:18	
Air Blank	0.000	21:19	
Subject Sample	# 1	21:19	Subject Test Refused
Breath Volume	0.000	21:19	
Air Blank	0.000	21:20	
Control Test	0.081	21:20	
Air Blank	0.000	21:21	
Diagnostics Check	OK	21:21	

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCracken,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: ENNIS,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: HUENEFELD

DATE/TIME OF TEST: 12/28/2018 04:29:14

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:20	
Control Test	0.081	04:20	
Air Blank	0.000	04:20	
Subject Sample	# 1 0.129	04:24	Volume Not Met
Breath Volume	0.437	04:24	
Air Blank	0.000	04:24	
Air Blank	0.000	04:26	
Subject Sample	# 2	04:27	Subject Test Refused
Breath Volume	0.035	04:27	
Air Blank	0.000	04:28	
Control Test	0.080	04:28	
Air Blank	0.000	04:29	
Diagnostics Check	OK	04:29	

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: REYNOLDS,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 12/30/2018 01:59:40

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.176	01:53
Breath Volume	1.316	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.158	01:58
Breath Volume	1.667	01:58
Air Blank	0.000	01:58
Control Test	0.080	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 01/18/2019 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,JOSHUA,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FULTON,DARIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/08/2018 23:33:39

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:30	
Control Test	0.080	23:30	
Air Blank	0.000	23:31	
Subject Sample	# 1	23:32	RFI Detect
Breath Volume	0.000	23:32	
Air Blank	0.000	23:32	
Control Test	0.080	23:33	
Air Blank	0.000	23:33	
Diagnostics Check	OK	23:33	

Data Download Date/Time: 12/24/2018 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,JOSHUA,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FULTON,DARIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/08/2018 23:50:04

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.080	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 1 0.000	23:41	Volume Not Met
Breath Volume	0.792	23:41	
Air Blank	0.000	23:42	
Air Blank	0.000	23:44	
Subject Sample	# 2 0.000	23:44	
Breath Volume	2.195	23:44	
Air Blank	0.000	23:45	
Air Blank	0.000	23:47	
Subject Sample	# 3 0.000	23:48	
Breath Volume	1.945	23:48	
Air Blank	0.000	23:49	
Control Test	0.079	23:49	
Air Blank	0.000	23:49	
Diagnostics Check	OK	23:50	

Data Download Date/Time: 12/24/2018 16:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEAZZO, TRASEY, M
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: HOFMANN, TRAVIS
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: NIX

DATE/TIME OF TEST: 12/03/2018 11:57:14

Diagnostics Check	OK	11:50
Air Blank	0.000	11:51
Control Test	0.082	11:51
Air Blank	0.000	11:51
Subject Sample # 1	0.000	11:52
Breath Volume	1.691	11:52
Air Blank	0.000	11:52
Air Blank	0.000	11:54
Subject Sample # 2	0.000	11:55
Breath Volume	2.503	11:55
Air Blank	0.000	11:56
Control Test	0.081	11:56
Air Blank	0.000	11:57
Diagnostics Check	OK	11:57

Data Download Date/Time: 12/22/2018 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURRY, TYLER, M
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: HEATH, MARC
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: SAFFICE

DATE/TIME OF TEST: 12/11/2018 00:12:10

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.083	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.185	00:07
Breath Volume	2.714	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.179	00:10
Breath Volume	1.738	00:10
Air Blank	0.000	00:11
Control Test	0.082	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 12/22/2018 09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURRY, TYLER, M
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ONORATO, BRIANNA
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: FORBES

DATE/TIME OF TEST: 12/16/2018 03:32:40

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.082	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.000	03:28
Breath Volume	1.566	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.000	03:31
Breath Volume	2.960	03:31
Air Blank	0.000	03:31
Control Test	0.083	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 12/22/2018 09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUSHENBERY,KYLE,E
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: ANDERSON,BEN
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/19/2018 04:54:49

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:51	
Control Test	0.082	04:51	
Air Blank	0.000	04:52	
Subject Sample	# 1	04:53	Subject Test Refused
Breath Volume	0.000	04:53	
Air Blank	0.000	04:53	
Control Test	0.081	04:54	
Air Blank	0.000	04:54	
Diagnostics Check	OK	04:54	

Data Download Date/Time: 12/22/2018 09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURRY, TYLER, M
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: HEATH, MARC
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: THEOPHILE

DATE/TIME OF TEST: 12/20/2018 23:23:50

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.082	23:21	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:22	Subject Test Refused
Breath Volume	0.000	23:22	
Air Blank	0.000	23:22	
Control Test	0.081	23:23	
Air Blank	0.000	23:23	
Diagnostics Check	OK	23:23	

Data Download Date/Time: 12/22/2018 09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEAZZO, TRASEY, M
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: CREWS, JAMES
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 12/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: VELKLY

DATE/TIME OF TEST: 12/23/2018 19:37:59

Diagnostics Check	OK	19:31
Air Blank	0.000	19:32
Control Test	0.083	19:32
Air Blank	0.000	19:32
Subject Sample # 1	0.257	19:33
Breath Volume	2.789	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:35
Subject Sample # 2	0.258	19:36
Breath Volume	2.769	19:36
Air Blank	0.000	19:37
Control Test	0.082	19:37
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 01/31/2019 23:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VISMAN,CLAYTON,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: VISMAN,CLAYTON
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BUNTON

DATE/TIME OF TEST: 12/28/2018 03:37:01

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.082	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.066	03:32
Breath Volume	1.984	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.069	03:35
Breath Volume	2.511	03:35
Air Blank	0.000	03:36
Control Test	0.083	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:37

Data Download Date/Time: 01/31/2019 23:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOFMANN, TRAVIS, A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HOFMANN, TRAVIS
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 08/07/2020
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: FORTUNE

DATE/TIME OF TEST: 12/21/2018 16:06:55

Diagnostics Check	OK	16:00
Air Blank	0.000	16:01
Control Test	0.079	16:01
Air Blank	0.000	16:02
Subject Sample # 1	0.126	16:02
Breath Volume	1.890	16:02
Air Blank	0.000	16:03
Air Blank	0.000	16:05
Subject Sample # 2	0.131	16:05
Breath Volume	2.097	16:05
Air Blank	0.000	16:06
Control Test	0.079	16:06
Air Blank	0.000	16:06
Diagnostics Check	OK	16:06

Data Download Date/Time: 01/17/2019 15:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHILLER,JOSEPH,C
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: BOUKOWSKI,DAVID
ARRESTING OFFICER AGENCY: Starke PD
LAST AGENCY INSPECTION: 12/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 08:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:17

SUBJECT'S LAST NAME: MAYWALT

DATE/TIME OF TEST: 12/17/2018 09:44:03

Diagnostics Check	OK	09:37
Air Blank	0.000	09:38
Control Test	0.078	09:38
Air Blank	0.000	09:38
Subject Sample # 1	0.251	09:39
Breath Volume	3.175	09:39
Air Blank	0.000	09:40
Air Blank	0.000	09:41
Subject Sample # 2	0.249	09:42
Breath Volume	2.875	09:42
Air Blank	0.000	09:43
Control Test	0.077	09:43
Air Blank	0.000	09:43
Diagnostics Check	OK	09:44

Data Download Date/Time: 01/18/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TERRELL,G
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WILLIAMS,TERRELL
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 12/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MARQUEZ-AYALA

DATE/TIME OF TEST: 12/19/2018 03:23:43

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.079	03:15	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:19	No Sample Provided
Breath Volume	0.000	03:19	
Air Blank	0.000	03:20	
Air Blank	0.000	03:22	
Subject Sample	# 2	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:22	
Control Test	0.078	03:23	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 01/18/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWAGGERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: KROL,FRANK
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 12/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 12/29/2018 23:03:01

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.077	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.195	22:58
Breath Volume	1.597	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.193	23:01
Breath Volume	1.785	23:01
Air Blank	0.000	23:02
Control Test	0.076	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:03

Data Download Date/Time: 01/18/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWAGGERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: KROL,FRANK
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 12/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: RALFSON

DATE/TIME OF TEST: 12/30/2018 02:51:06

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.078	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.159	02:46
Breath Volume	2.734	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.155	02:49
Breath Volume	2.539	02:49
Air Blank	0.000	02:50
Control Test	0.077	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 01/18/2019 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWAGGERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: HULLENDER,ROBERT
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 12/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: WORTH

DATE/TIME OF TEST: 12/30/2018 04:20:23

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:16	
Control Test	0.078	04:16	
Air Blank	0.000	04:17	
Subject Sample	# 1	04:19	Subject Test Refused
Breath Volume	0.000	04:19	
Air Blank	0.000	04:19	
Control Test	0.078	04:19	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 01/18/2019 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMNER,DALTON,R
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: KROL,FRANK
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 11:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 12/01/2018 00:33:51

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.078	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.000	00:29
Breath Volume	2.992	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.000	00:32
Breath Volume	3.140	00:32
Air Blank	0.000	00:32
Control Test	0.079	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 12/11/2018 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL,SCOTT,M
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 12/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: ADMIN
ARREST TIME: 12:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 12/19/2018 13:06:29

Diagnostics Check	OK	13:00
Air Blank	0.000	13:01
Control Test	0.079	13:01
Air Blank	0.000	13:02
Subject Sample # 1	0.000	13:02
Breath Volume	2.187	13:02
Air Blank	0.000	13:02
Air Blank	0.000	13:04
Subject Sample # 2	0.000	13:05
Breath Volume	1.968	13:05
Air Blank	0.000	13:05
Control Test	0.079	13:05
Air Blank	0.000	13:06
Diagnostics Check	OK	13:06

Data Download Date/Time: 01/18/2019 08:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,SCOTT,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: SULLIVAN,RYAN
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: EAGER

DATE/TIME OF TEST: 12/20/2018 01:49:35

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.078	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.262	01:45
Breath Volume	1.914	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.261	01:47
Breath Volume	2.480	01:47
Air Blank	0.000	01:48
Control Test	0.078	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 12/31/2018 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/01/2018 03:40:23

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.079	03:32	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:36	No Sample Provided
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Air Blank	0.000	03:38	
Subject Sample	# 2	03:39	Subject Test Refused
Breath Volume	0.000	03:39	
Air Blank	0.000	03:39	
Control Test	0.079	03:39	
Air Blank	0.000	03:40	
Diagnostics Check	OK	03:40	

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFELT,DYLAN,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: COFFELT,DYLAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: POWERS

DATE/TIME OF TEST: 12/01/2018 21:17:45

Diagnostics Check	OK	21:11
Air Blank	0.000	21:12
Control Test	0.080	21:12
Air Blank	0.000	21:12
Subject Sample # 1	0.000	21:13
Breath Volume	2.375	21:13
Air Blank	0.000	21:14
Air Blank	0.000	21:16
Subject Sample # 2	0.000	21:16
Breath Volume	1.859	21:16
Air Blank	0.000	21:16
Control Test	0.080	21:17
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRIBB,BRIAN,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LEVIER,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 12/01/2018 23:26:36

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.079	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.136	23:22
Breath Volume	1.871	23:22
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.136	23:25
Breath Volume	2.195	23:25
Air Blank	0.000	23:25
Control Test	0.078	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAPMAN,MAVERICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 13:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:11

SUBJECT'S LAST NAME: BRANCH

DATE/TIME OF TEST: 12/02/2018 14:35:06

Diagnostics Check	OK	14:31	
Air Blank	0.000	14:31	
Control Test	0.079	14:32	
Air Blank	0.000	14:32	
Subject Sample	# 1	14:33	Subject Test Refused
Breath Volume	0.000	14:33	
Air Blank	0.000	14:34	
Control Test	0.079	14:34	
Air Blank	0.000	14:35	
Diagnostics Check	OK	14:35	

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWTON,CARSON,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: OBANNON,ANTHONY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 12/02/2018 20:19:58

Diagnostics Check	OK	20:11
Air Blank	0.000	20:11
Control Test	0.079	20:11
Air Blank	0.000	20:12
Subject Sample # 1	0.109	20:13
Breath Volume	1.828	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:16
Subject Sample # 2	0.098	20:18
Breath Volume	1.359	20:18
Air Blank	0.000	20:19
Control Test	0.079	20:19
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,WILLIAM,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ADAMS,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 12/02/2018 22:00:12

Diagnostics Check	OK	21:52
Air Blank	0.000	21:53
Control Test	0.079	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.118	21:55
Breath Volume	2.917	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:57
Subject Sample # 2	0.114	21:58
Breath Volume	2.046	21:58
Air Blank	0.000	21:59
Control Test	0.077	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRIBB,BRIAN,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LEVIER,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 12/04/2018 00:42:40

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.000	00:38
Breath Volume	2.542	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.000	00:41
Breath Volume	2.117	00:41
Air Blank	0.000	00:41
Control Test	0.079	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CLEMENTS, WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 15:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/05/2018 16:42:18

Diagnostics Check	OK	16:30	
Air Blank	0.000	16:31	
Control Test	0.079	16:31	
Air Blank	0.000	16:32	
Subject Sample	# 1	16:35	No Sample Provided
Breath Volume	0.000	16:35	
Air Blank	0.000	16:35	
Air Blank	0.000	16:37	
Subject Sample	# 2	16:40	No Sample Provided
Breath Volume	0.000	16:40	
Air Blank	0.000	16:41	
Control Test	0.080	16:41	
Air Blank	0.000	16:42	
Diagnostics Check	OK	16:42	

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CLEMENTS, WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 15:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/05/2018 17:03:35

Diagnostics Check	OK	16:56
Air Blank	0.000	16:56
Control Test	0.080	16:56
Air Blank	0.000	16:57
Subject Sample # 1	0.245	16:58
Breath Volume	2.476	16:58
Air Blank	0.000	16:59
Air Blank	0.000	17:01
Subject Sample # 2	0.250	17:02
Breath Volume	2.406	17:02
Air Blank	0.000	17:02
Control Test	0.077	17:02
Air Blank	0.000	17:03
Diagnostics Check	OK	17:03

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER, BRANDT, J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CLARK, S
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: KERR

DATE/TIME OF TEST: 12/07/2018 04:28:24

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:25	
Control Test	0.081	04:26	
Air Blank	0.000	04:26	
Subject Sample	# 1	04:27	Subject Test Refused
Breath Volume	0.000	04:27	
Air Blank	0.000	04:27	
Control Test	0.080	04:27	
Air Blank	0.000	04:28	
Diagnostics Check	OK	04:28	

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,STEPHEN,W
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: QUICK,STEPHEN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 11:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: BRAIK

DATE/TIME OF TEST: 12/08/2018 00:56:08

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.080	00:48	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:51	No Sample Provided
Breath Volume	0.000	00:51	
Air Blank	0.000	00:52	
Air Blank	0.000	00:54	
Subject Sample	# 2	00:54	Subject Test Refused
Breath Volume	0.000	00:54	
Air Blank	0.000	00:55	
Control Test	0.080	00:55	
Air Blank	0.000	00:56	
Diagnostics Check	OK	00:56	

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREW,MATTHEW,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ANDREW,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WEISS

DATE/TIME OF TEST: 12/09/2018 03:34:12

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.197	03:29
Breath Volume	3.289	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.206	03:32
Breath Volume	1.988	03:32
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWTON,CARSON,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TIFFT,DALTON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: LEE-BURNETT

DATE/TIME OF TEST: 12/09/2018 19:40:53

Diagnostics Check	OK	19:34
Air Blank	0.000	19:34
Control Test	0.078	19:34
Air Blank	0.000	19:35
Subject Sample # 1	0.197	19:35
Breath Volume	2.343	19:35
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample # 2	0.198	19:39
Breath Volume	2.171	19:39
Air Blank	0.000	19:39
Control Test	0.078	19:40
Air Blank	0.000	19:40
Diagnostics Check	OK	19:40

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRIBB,BRIAN,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CRIBB,BRIAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 12/10/2018 22:57:39

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.121	22:52
Breath Volume	2.140	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.111	22:56
Breath Volume	1.347	22:56
Air Blank	0.000	22:56
Control Test	0.079	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,GABRIEL,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROBERSON,GABRIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: BOWLEY

DATE/TIME OF TEST: 12/10/2018 23:43:53

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:37
Air Blank	0.000	23:37
Subject Sample # 1	0.000	23:38
Breath Volume	1.316	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.000	23:42
Breath Volume	2.089	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DANIEL,K
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DAVIS,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: BUNN

DATE/TIME OF TEST: 12/13/2018 01:49:19

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.079	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.153	01:44
Breath Volume	2.097	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.146	01:47
Breath Volume	2.378	01:47
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: MCMAHAN

DATE/TIME OF TEST: 12/13/2018 20:30:11

Diagnostics Check	OK	20:27	
Air Blank	0.000	20:28	
Control Test	0.079	20:28	
Air Blank	0.000	20:29	
Subject Sample	# 1	20:29	RFI Detect
Breath Volume	0.687	20:29	
Air Blank		20:29	RFI Detect
Air Blank	0.000	20:30	

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: MCMAHAN

DATE/TIME OF TEST: 12/13/2018 20:36:51

Diagnostics Check	OK	20:31	
Air Blank	0.000	20:32	
Control Test	0.079	20:32	
Air Blank	0.000	20:33	
Subject Sample	# 1 0.183	20:33	
Breath Volume	2.039	20:33	
Air Blank	0.000	20:34	
Air Blank	0.000	20:35	
Subject Sample	# 2	20:36	RFI Detect
Breath Volume	0.687	20:36	
Air Blank		20:36	RFI Detect
Air Blank	0.000	20:36	

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALABRO,ALFRED,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BUSBY,JASON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: BARRY

DATE/TIME OF TEST: 12/14/2018 21:51:12

Diagnostics Check	OK	21:42	
Air Blank	0.000	21:43	
Control Test	0.079	21:43	
Air Blank	0.000	21:44	
Subject Sample	# 1	21:47	No Sample Provided
Breath Volume	0.000	21:47	
Air Blank	0.000	21:47	
Air Blank	0.000	21:49	
Subject Sample	# 2	21:49	Subject Test Refused
Breath Volume	0.000	21:49	
Air Blank	0.000	21:50	
Control Test	0.079	21:50	
Air Blank	0.000	21:51	
Diagnostics Check	OK	21:51	

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,GABRIEL,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROBERSON,GABRIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: CEPOLSKI

DATE/TIME OF TEST: 12/14/2018 23:45:20

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.079	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.000	23:41
Breath Volume	2.839	23:41
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.000	23:43
Breath Volume	2.843	23:43
Air Blank	0.000	23:44
Control Test	0.078	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOZARD,CHRYSTAL,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BURKE,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 09:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/15/2018 11:22:15

Diagnostics Check	OK	11:15
Air Blank	0.000	11:16
Control Test	0.080	11:16
Air Blank	0.000	11:16
Subject Sample # 1	0.000	11:17
Breath Volume	2.937	11:17
Air Blank	0.000	11:18
Air Blank	0.000	11:20
Subject Sample # 2	0.000	11:20
Breath Volume	2.769	11:20
Air Blank	0.000	11:21
Control Test	0.079	11:21
Air Blank	0.000	11:22
Diagnostics Check	OK	11:22

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 12/15/2018 19:49:58

Diagnostics Check	OK	19:39
Air Blank	0.000	19:40
Control Test	0.079	19:40
Air Blank	0.000	19:41
Subject Sample # 1	0.251	19:42
Breath Volume	1.246	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.245	19:48
Breath Volume	1.160	19:48
Air Blank	0.000	19:49
Control Test	0.080	19:49
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,LLOYD,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ALEXANDER,LLOYD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 12/16/2018 02:04:41

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:01	
Control Test	0.079	02:01	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:03	Subject Test Refused
Breath Volume	0.000	02:03	
Air Blank	0.000	02:03	
Control Test	0.079	02:04	
Air Blank	0.000	02:04	
Diagnostics Check	OK	02:04	

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREW, MATTHEW, A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ANDREW, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: BECKMAN

DATE/TIME OF TEST: 12/16/2018 02:55:24

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.079	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.000	02:51
Breath Volume	2.699	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.000	02:54
Breath Volume	1.878	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEICHNER, ADAM, J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MILLER, JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 12/18/2018 00:47:57

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.248	00:43
Breath Volume	2.265	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:45
Subject Sample # 2	0.249	00:46
Breath Volume	1.636	00:46
Air Blank	0.000	00:47
Control Test	0.078	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DANIEL,K
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DAVIS,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: WALTHER

DATE/TIME OF TEST: 12/19/2018 03:52:19

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:45	
Control Test	0.080	03:45	
Air Blank	0.000	03:46	
Subject Sample	# 1 0.000	03:47	
Breath Volume	1.164	03:47	
Air Blank	0.000	03:48	
Air Blank	0.000	03:50	
Subject Sample	# 2	03:50	Subject Test Refused
Breath Volume	0.000	03:50	
Air Blank	0.000	03:51	
Control Test	0.080	03:51	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,DAVID,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DANIELS,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 14:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 12/20/2018 15:48:19

Diagnostics Check	OK	15:41
Air Blank	0.000	15:42
Control Test	0.079	15:42
Air Blank	0.000	15:43
Subject Sample # 1	0.000	15:44
Breath Volume	1.898	15:44
Air Blank	0.000	15:44
Air Blank	0.000	15:46
Subject Sample # 2	0.000	15:46
Breath Volume	1.875	15:46
Air Blank	0.000	15:47
Control Test	0.079	15:47
Air Blank	0.000	15:48
Diagnostics Check	OK	15:48

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNSFORD, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DUNSFORD, WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 12/20/2018 19:27:23

Diagnostics Check	OK	19:18	
Air Blank	0.000	19:19	
Control Test	0.079	19:19	
Air Blank	0.000	19:20	
Subject Sample	# 1	19:23	No Sample Provided
Breath Volume	0.000	19:23	
Air Blank	0.000	19:23	
Air Blank	0.000	19:25	
Subject Sample	# 2	19:26	Subject Test Refused
Breath Volume	0.000	19:26	
Air Blank	0.000	19:26	
Control Test	0.079	19:26	
Air Blank	0.000	19:27	
Diagnostics Check	OK	19:27	

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,KENNETH,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DESOUSA,CARLOS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 12/22/2018 19:58:45

Diagnostics Check	OK	19:55	
Air Blank	0.000	19:56	
Control Test	0.079	19:56	
Air Blank	0.000	19:56	
Subject Sample	# 1	19:57	Subject Test Refused
Breath Volume	0.000	19:57	
Air Blank	0.000	19:57	
Control Test	0.079	19:58	
Air Blank	0.000	19:58	
Diagnostics Check	OK	19:58	

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARGITT,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARGITT,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: SEAL

DATE/TIME OF TEST: 12/23/2018 02:16:14

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.208	02:11
Breath Volume	2.242	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.200	02:14
Breath Volume	1.875	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MUNDT,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/23/2018 05:15:20

Diagnostics Check	OK	05:12	
Air Blank	0.000	05:13	
Control Test	0.079	05:13	
Air Blank	0.000	05:13	
Subject Sample	# 1	05:13	Subject Test Refused
Breath Volume	0.000	05:13	
Air Blank	0.000	05:14	
Control Test	0.080	05:14	
Air Blank	0.000	05:15	
Diagnostics Check	OK	05:15	

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: CONE

DATE/TIME OF TEST: 12/23/2018 21:45:29

Diagnostics Check	OK	21:42	
Air Blank	0.000	21:43	
Control Test	0.079	21:43	
Air Blank	0.000	21:43	
Subject Sample	# 1	21:44	Subject Test Refused
Breath Volume	0.000	21:44	
Air Blank	0.000	21:44	
Control Test	0.079	21:44	
Air Blank	0.000	21:45	
Diagnostics Check	OK	21:45	

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARGITT,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARGITT,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: MCGAUGHY

DATE/TIME OF TEST: 12/24/2018 01:43:23

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:32	
Control Test	0.079	01:32	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:36	No Sample Provided
Breath Volume	0.000	01:36	
Air Blank	0.000	01:36	
Air Blank	0.000	01:38	
Subject Sample	# 2	01:42	No Sample Provided
Breath Volume	0.000	01:42	
Air Blank	0.000	01:42	
Control Test	0.079	01:42	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 11:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:56

SUBJECT'S LAST NAME: NANCE

DATE/TIME OF TEST: 12/24/2018 12:22:48

Diagnostics Check	OK	12:16
Air Blank	0.000	12:16
Control Test	0.080	12:17
Air Blank	0.000	12:17
Subject Sample # 1	0.064	12:18
Breath Volume	3.117	12:18
Air Blank	0.000	12:18
Air Blank	0.000	12:20
Subject Sample # 2	0.066	12:21
Breath Volume	2.710	12:21
Air Blank	0.000	12:21
Control Test	0.080	12:22
Air Blank	0.000	12:22
Diagnostics Check	OK	12:22

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: REGNA

DATE/TIME OF TEST: 12/26/2018 20:52:04

Diagnostics Check	OK	20:40	
Air Blank	0.000	20:41	
Control Test	0.079	20:41	
Air Blank	0.000	20:42	
Subject Sample	# 1	20:45	No Sample Provided
Breath Volume	0.000	20:45	
Air Blank	0.000	20:45	
Air Blank	0.000	20:47	
Subject Sample	# 2	20:50	Subject Test Refused
Breath Volume	0.000	20:50	
Air Blank	0.000	20:51	
Control Test	0.079	20:51	
Air Blank	0.000	20:51	
Diagnostics Check	OK	20:52	

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEICHNER, ADAM, J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TEICHNER, ADAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 12/28/2018 04:34:31

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:32	
Control Test	0.079	04:32	
Air Blank	0.000	04:32	
Subject Sample	# 1	04:33	Subject Test Refused
Breath Volume	0.000	04:33	
Air Blank	0.000	04:33	
Control Test	0.079	04:33	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:34	

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEINER,ZACHARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: STEINER,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: KESSLER

DATE/TIME OF TEST: 12/30/2018 05:56:07

Diagnostics Check	OK	05:49
Air Blank	0.000	05:50
Control Test	0.079	05:50
Air Blank	0.000	05:51
Subject Sample # 1	0.226	05:51
Breath Volume	4.218	05:51
Air Blank	0.000	05:52
Air Blank	0.000	05:54
Subject Sample # 2	0.213	05:54
Breath Volume	2.843	05:54
Air Blank	0.000	05:55
Control Test	0.078	05:55
Air Blank	0.000	05:56
Diagnostics Check	OK	05:56

Data Download Date/Time: 12/31/2018 20:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JORDAN,WILFORD,N

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: JORDAN,WILFORD

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 906364

DRY GAS EXPIRY DATE: 02/07/2021

VIOLATION CODE: DUI

ARREST TIME: 11:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: NIDIFFER

DATE/TIME OF TEST: 12/15/2018 00:16:32

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.079	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.294	00:12
Breath Volume	3.046	00:12
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.295	00:15
Breath Volume	2.773	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 01/31/2019 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL,DAVID,C
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: ARROYO,TRISTON
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 12/22/2018 04:27:25

Diagnostics Check	OK	04:16	
Air Blank	0.000	04:16	
Control Test	0.082	04:16	
Air Blank	0.000	04:17	
Subject Sample	# 1	04:20	No Sample Provided
Breath Volume	0.000	04:20	
Air Blank	0.000	04:20	
Air Blank	0.000	04:22	
Subject Sample	# 2	04:26	No Sample Provided
Breath Volume	0.000	04:26	
Air Blank	0.000	04:26	
Control Test	0.082	04:26	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 01/31/2019 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HOWELL,DAVID,C

BREATH TEST OPERATOR AGENCY: Marianna PD

ARRESTING OFFICER: HOWELL,DAVID

ARRESTING OFFICER AGENCY: Marianna PD

LAST AGENCY INSPECTION: 11/15/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 906364

DRY GAS EXPIRY DATE: 02/07/2021

VIOLATION CODE: DUI

ARREST TIME: 22:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: NEWSOME

DATE/TIME OF TEST: 12/25/2018 22:53:14

Diagnostics Check	OK	22:50	
Air Blank	0.000	22:50	
Control Test	0.082	22:51	
Air Blank	0.000	22:51	
Subject Sample	# 1	22:51	Subject Test Refused
Breath Volume	0.000	22:51	
Air Blank	0.000	22:52	
Control Test	0.082	22:52	
Air Blank	0.000	22:53	
Diagnostics Check	OK	22:53	

Data Download Date/Time: 01/31/2019 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUTTON,KERRY,D
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: DUTTON,KERRY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 12/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: PROFFITT

DATE/TIME OF TEST: 12/28/2018 17:28:13

Diagnostics Check	OK	17:22
Air Blank	0.000	17:22
Control Test	0.078	17:22
Air Blank	0.000	17:23
Subject Sample # 1	0.211	17:23
Breath Volume	3.921	17:23
Air Blank	0.000	17:24
Air Blank	0.000	17:26
Subject Sample # 2	0.218	17:26
Breath Volume	2.800	17:26
Air Blank	0.000	17:27
Control Test	0.077	17:27
Air Blank	0.000	17:28
Diagnostics Check	OK	17:28

Data Download Date/Time: 01/31/2019 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD, MATTHEW, E
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: MAYFIELD, MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 15:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: BUNCE

DATE/TIME OF TEST: 12/07/2018 16:44:02

Diagnostics Check	OK	16:38
Air Blank	0.000	16:38
Control Test	0.081	16:38
Air Blank	0.000	16:39
Subject Sample # 1	0.000	16:39
Breath Volume	1.355	16:39
Air Blank	0.000	16:40
Air Blank	0.000	16:42
Subject Sample # 2	0.000	16:42
Breath Volume	1.261	16:42
Air Blank	0.000	16:43
Control Test	0.081	16:43
Air Blank	0.000	16:43
Diagnostics Check	OK	16:44

Data Download Date/Time: 12/27/2018 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,BRAD,
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PIERCE,BRAD
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: RAWSON

DATE/TIME OF TEST: 12/07/2018 22:56:59

Diagnostics Check	OK	22:48
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.140	22:52
Breath Volume	2.125	22:52
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.149	22:55
Breath Volume	1.886	22:55
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 12/27/2018 11:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CHRISTY,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: ADAMS,WH
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/12/2018 02:12:39

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:10	
Control Test	0.079	02:10	
Air Blank	0.000	02:11	
Subject Sample	# 1	02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:11	
Control Test	0.079	02:12	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:12	

Data Download Date/Time: 12/27/2018 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OQUENDO,JOEL,G
BREATH TEST OPERATOR AGENCY: Springfield PD
ARRESTING OFFICER: REXFORD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Springfield PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: CLEVELAND

DATE/TIME OF TEST: 12/22/2018 04:17:53

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.080	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.090	04:13
Breath Volume	1.871	04:13
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.093	04:16
Breath Volume	1.972	04:16
Air Blank	0.000	04:16
Control Test	0.080	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 12/27/2018 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDYETTE,COREY,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BRANCH,JOSHUA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: MASTERS

DATE/TIME OF TEST: 12/25/2018 02:11:05

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.078	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.274	02:06
Breath Volume	3.093	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.270	02:09
Breath Volume	3.734	02:09
Air Blank	0.000	02:10
Control Test	0.079	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 12/27/2018 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINTOSH,ZANE,R
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: MCINTOSH,ZANE
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 171502
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 15:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:49

SUBJECT'S LAST NAME: MARCUM

DATE/TIME OF TEST: 12/04/2018 16:20:45

Diagnostics Check	OK	16:14
Air Blank	0.000	16:14
Control Test	0.082	16:15
Air Blank	0.000	16:15
Subject Sample # 1	0.206	16:16
Breath Volume	1.195	16:16
Air Blank	0.000	16:16
Air Blank	0.000	16:18
Subject Sample # 2	0.209	16:19
Breath Volume	1.445	16:19
Air Blank	0.000	16:19
Control Test	0.081	16:20
Air Blank	0.000	16:20
Diagnostics Check	OK	16:20

Data Download Date/Time: 01/04/2019 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,SETH,D
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: WILLIAMS,SETH
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 869968
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: BOHLER

DATE/TIME OF TEST: 12/19/2018 01:19:41

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.081	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.100	01:15
Breath Volume	1.187	01:15
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.093	01:18
Breath Volume	1.808	01:18
Air Blank	0.000	01:18
Control Test	0.081	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 01/04/2019 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCHULTZ,JEREMY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: GUERARA-PENA

DATE/TIME OF TEST: 12/09/2018 21:55:20

Diagnostics Check	OK	21:49
Air Blank	0.000	21:49
Control Test	0.080	21:49
Air Blank	0.000	21:50
Subject Sample # 1	0.065	21:50
Breath Volume	3.695	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.058	21:53
Breath Volume	1.687	21:53
Air Blank	0.000	21:54
Control Test	0.079	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 12/24/2018 09:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DAVID,O
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SANDERS,DAVID
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 12/13/2018 00:44:43

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.080	00:38
Air Blank	0.000	00:39
Subject Sample # 1	0.149	00:39
Breath Volume	2.734	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.148	00:43
Breath Volume	3.000	00:43
Air Blank	0.000	00:43
Control Test	0.081	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 12/24/2018 09:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AIMANOVICH,MATTHEW,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCGHEE,DARRYL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: PRIDGEON

DATE/TIME OF TEST: 12/15/2018 03:02:06

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.232	02:57
Breath Volume	2.890	02:57
Air Blank	0.000	02:58
Air Blank	0.000	02:59
Subject Sample # 2	0.238	03:00
Breath Volume	2.968	03:00
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 12/24/2018 09:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 12/16/2018 21:04:21

Diagnostics Check	OK	20:57	
Air Blank	0.000	20:58	
Control Test	0.081	20:58	
Air Blank	0.000	20:58	
Subject Sample	# 1 0.181	21:00	
Breath Volume	1.398	21:00	
Air Blank	0.000	21:00	
Air Blank	0.000	21:02	
Subject Sample	# 2 0.184	21:03	
Breath Volume	2.335	21:03	
Air Blank		21:03	RFI Detect
Air Blank	0.000	21:04	

Data Download Date/Time: 12/24/2018 09:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYDOS,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GAYDOS,JONATHAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 11:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:23

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 12/19/2018 11:56:22

Diagnostics Check	OK	11:50
Air Blank	0.000	11:50
Control Test	0.079	11:50
Air Blank	0.000	11:51
Subject Sample # 1	0.000	11:52
Breath Volume	1.128	11:52
Air Blank	0.000	11:52
Air Blank	0.000	11:54
Subject Sample # 2	0.000	11:55
Breath Volume	2.082	11:55
Air Blank	0.000	11:55
Control Test	0.080	11:55
Air Blank	0.000	11:56
Diagnostics Check	OK	11:56

Data Download Date/Time: 12/24/2018 09:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: RUIZ VARGAS

DATE/TIME OF TEST: 12/25/2018 05:18:26

Diagnostics Check	OK	05:12
Air Blank	0.000	05:12
Control Test	0.080	05:13
Air Blank	0.000	05:13
Subject Sample # 1	0.144	05:14
Breath Volume	2.175	05:14
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.145	05:16
Breath Volume	1.386	05:16
Air Blank	0.000	05:17
Control Test	0.081	05:17
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 01/23/2019 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DAVID,O
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SANDERS,DAVID
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:12

SUBJECT'S LAST NAME: LANASA

DATE/TIME OF TEST: 12/27/2018 15:48:15

Diagnostics Check	OK	15:41
Air Blank	0.000	15:42
Control Test	0.080	15:42
Air Blank	0.000	15:42
Subject Sample # 1	0.000	15:43
Breath Volume	3.390	15:43
Air Blank	0.000	15:44
Air Blank	0.000	15:46
Subject Sample # 2	0.000	15:46
Breath Volume	4.746	15:46
Air Blank	0.000	15:47
Control Test	0.080	15:47
Air Blank	0.000	15:48
Diagnostics Check	OK	15:48

Data Download Date/Time: 01/23/2019 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SEDANO

DATE/TIME OF TEST: 12/05/2018 03:10:58

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.080	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.147	03:05
Breath Volume	2.832	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.138	03:09
Breath Volume	2.101	03:09
Air Blank	0.000	03:10
Control Test	0.082	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 12/24/2018 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: COKONOUGH,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: PIZZA

DATE/TIME OF TEST: 12/07/2018 21:52:52

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:41	
Control Test	0.081	21:42	
Air Blank	0.000	21:42	
Subject Sample	# 1	21:45	No Sample Provided
Breath Volume	0.000	21:45	
Air Blank	0.000	21:46	
Air Blank	0.000	21:48	
Subject Sample	# 2	21:51	No Sample Provided
Breath Volume	0.000	21:51	
Air Blank	0.000	21:51	
Control Test	0.082	21:52	
Air Blank	0.000	21:52	
Diagnostics Check	OK	21:52	

Data Download Date/Time: 12/24/2018 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN, WILFORD, N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JORDAN, WILFORD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 12/08/2018 02:33:10

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:25	
Control Test	0.079	02:25	
Air Blank	0.000	02:25	
Subject Sample	# 1	02:29	No Sample Provided
Breath Volume	0.000	02:29	
Air Blank	0.000	02:29	
Air Blank	0.000	02:31	
Subject Sample	# 2	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:32	
Control Test	0.081	02:32	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

Data Download Date/Time: 12/24/2018 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,JODY,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GLENISTER,JOHN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: INFINGER

DATE/TIME OF TEST: 12/11/2018 23:37:47

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:35	
Control Test	0.081	23:35	
Air Blank	0.000	23:36	
Subject Sample	# 1	23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:36	
Control Test	0.081	23:37	
Air Blank	0.000	23:37	
Diagnostics Check	OK	23:37	

Data Download Date/Time: 12/24/2018 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYDOS,SAVANNAH,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TROUPE,TREY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: BAUTISTA

DATE/TIME OF TEST: 12/15/2018 23:00:12

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.080	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.236	22:55
Breath Volume	1.140	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:57
Subject Sample # 2	0.239	22:58
Breath Volume	2.437	22:58
Air Blank	0.000	22:59
Control Test	0.078	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 12/24/2018 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYDOS,SAVANNAH,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCGHEE,DARRYL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: HINOJOSAGARCIA

DATE/TIME OF TEST: 12/15/2018 23:21:28

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.080	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.207	23:16
Breath Volume	1.390	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.187	23:19
Breath Volume	1.316	23:19
Air Blank	0.000	23:20
Control Test	0.078	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 12/24/2018 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ARNOLD,ALEXIS
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 12/23/2018 20:08:20

Diagnostics Check	OK	19:55	
Air Blank	0.000	19:55	
Control Test	0.081	19:55	
Air Blank	0.000	19:56	
Subject Sample	# 1 0.077	19:59	Volume Not Met
Breath Volume	0.089	19:59	
Air Blank	0.000	20:00	
Air Blank	0.000	20:02	
Subject Sample	# 2 0.084	20:02	
Breath Volume	1.433	20:02	
Air Blank	0.000	20:03	
Air Blank	0.000	20:05	
Subject Sample	# 3 0.088	20:06	
Breath Volume	1.796	20:06	
Air Blank	0.000	20:07	
Control Test	0.079	20:07	
Air Blank	0.000	20:08	
Diagnostics Check	OK	20:08	

Data Download Date/Time: 12/24/2018 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYDOS,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GAYDOS,JONATHAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 12:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:23

SUBJECT'S LAST NAME: SCHOFIELD

DATE/TIME OF TEST: 12/24/2018 12:56:05

Diagnostics Check	OK	12:48
Air Blank	0.000	12:48
Control Test	0.082	12:48
Air Blank	0.000	12:49
Subject Sample # 1	0.000	12:50
Breath Volume	2.246	12:50
Air Blank	0.000	12:50
Air Blank	0.000	12:52
Subject Sample # 2	0.000	12:54
Breath Volume	3.445	12:54
Air Blank	0.000	12:55
Control Test	0.082	12:55
Air Blank	0.000	12:55
Diagnostics Check	OK	12:56

Data Download Date/Time: 01/23/2019 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK, STEVEN, D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK, STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: MCCANS

DATE/TIME OF TEST: 12/01/2018 03:27:33

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.218	03:22
Breath Volume	1.535	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.207	03:26
Breath Volume	1.113	03:26
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 12/05/2018 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALCAGNO, MERCEDES, T
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK, STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: GARCIALOPEZ

DATE/TIME OF TEST: 12/02/2018 02:43:07

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.077	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.149	02:38
Breath Volume	3.261	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.148	02:41
Breath Volume	3.257	02:41
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 12/05/2018 08:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: BRONSON

DATE/TIME OF TEST: 12/08/2018 21:54:04

Diagnostics Check	OK	21:47
Air Blank	0.000	21:48
Control Test	0.080	21:48
Air Blank	0.000	21:48
Subject Sample # 1	0.199	21:49
Breath Volume	3.925	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:51
Subject Sample # 2	0.215	21:52
Breath Volume	3.753	21:52
Air Blank	0.000	21:53
Control Test	0.080	21:53
Air Blank	0.000	21:53
Diagnostics Check	OK	21:54

Data Download Date/Time: 01/08/2019 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 12/15/2018 00:36:10

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.080	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.138	00:28
Breath Volume	2.414	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.163	00:31
Breath Volume	3.921	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 3	0.163	00:34
Breath Volume	3.628	00:34
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 01/08/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK, STEVEN, D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK, STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: CHAVEZ MATIAS

DATE/TIME OF TEST: 12/15/2018 23:50:20

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.093	23:45
Breath Volume	2.437	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.092	23:48
Breath Volume	2.308	23:48
Air Blank	0.000	23:49
Control Test	0.079	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 01/08/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,BRANDON,W
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VULETA,BEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 09:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: SORICELLI

DATE/TIME OF TEST: 12/21/2018 10:33:54

Diagnostics Check	OK	10:31	
Air Blank	0.000	10:31	
Control Test	0.079	10:31	
Air Blank	0.000	10:32	
Subject Sample	# 1	10:32	Subject Test Refused
Breath Volume	0.000	10:32	
Air Blank	0.000	10:33	
Control Test	0.080	10:33	
Air Blank	0.000	10:33	
Diagnostics Check	OK	10:33	

Data Download Date/Time: 01/08/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK, STEVEN, D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK, STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: BIVINS

DATE/TIME OF TEST: 12/22/2018 19:29:35

Diagnostics Check	OK	19:23
Air Blank	0.000	19:24
Control Test	0.080	19:24
Air Blank	0.000	19:24
Subject Sample # 1	0.146	19:25
Breath Volume	2.894	19:25
Air Blank	0.000	19:25
Air Blank	0.000	19:27
Subject Sample # 2	0.147	19:28
Breath Volume	2.023	19:28
Air Blank	0.000	19:28
Control Test	0.080	19:29
Air Blank	0.000	19:29
Diagnostics Check	OK	19:29

Data Download Date/Time: 01/08/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK, STEVEN, D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK, STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: JANSEN

DATE/TIME OF TEST: 12/26/2018 21:00:00

Diagnostics Check	OK	20:53
Air Blank	0.000	20:54
Control Test	0.080	20:54
Air Blank	0.000	20:54
Subject Sample # 1	0.143	20:55
Breath Volume	4.792	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.137	20:58
Breath Volume	3.480	20:58
Air Blank	0.000	20:59
Control Test	0.080	20:59
Air Blank	0.000	20:59
Diagnostics Check	OK	21:00

Data Download Date/Time: 01/08/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK, STEVEN, D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK, STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 12/28/2018 19:54:17

Diagnostics Check	OK	19:48
Air Blank	0.000	19:48
Control Test	0.080	19:48
Air Blank	0.000	19:49
Subject Sample # 1	0.065	19:49
Breath Volume	3.398	19:49
Air Blank	0.000	19:50
Air Blank	0.000	19:52
Subject Sample # 2	0.061	19:52
Breath Volume	3.937	19:52
Air Blank	0.000	19:53
Control Test	0.080	19:53
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

Data Download Date/Time: 01/08/2019 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEATON,JONATHAN,T
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: NETTLES

DATE/TIME OF TEST: 12/29/2018 00:01:45

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:59	
Control Test	0.077	23:59	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:00	
Control Test	0.079	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 01/08/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEATON,JONATHAN,T
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 12/30/2018 03:16:24

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.080	03:13	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.080	03:15	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 01/08/2019 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMERO,CESAR,
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: LIBERATO, RAMON
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 11/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 12:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 12/13/2018 01:51:03

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.078	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.192	01:45
Breath Volume	2.097	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.199	01:49
Breath Volume	2.285	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:51

Data Download Date/Time: 12/31/2018 04:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRING,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HERRING,TIMOTHY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:34

SUBJECT'S LAST NAME: BUFFORD

DATE/TIME OF TEST: 12/01/2018 17:06:24

Diagnostics Check	OK	16:54	
Air Blank	0.000	16:54	
Control Test	0.080	16:55	
Air Blank	0.000	16:55	
Subject Sample	# 1 0.117	16:58	Volume Not Met
Breath Volume	0.855	16:58	
Air Blank	0.000	16:59	
Air Blank	0.000	17:01	
Subject Sample	# 2 0.141	17:01	
Breath Volume	2.960	17:01	
Air Blank	0.000	17:02	
Air Blank	0.000	17:04	
Subject Sample	# 3 0.140	17:04	
Breath Volume	2.519	17:04	
Air Blank	0.000	17:05	
Control Test	0.076	17:05	
Air Blank	0.000	17:06	
Diagnostics Check	OK	17:06	

Data Download Date/Time: 12/28/2018 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JEREMY,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: OLIVER,MICHAEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 12/06/2018 07:05:38

Diagnostics Check	OK	06:48	
Air Blank	0.000	06:49	
Control Test	0.080	06:49	
Air Blank	0.000	06:49	
Subject Sample	# 1 0.197	06:52	
Breath Volume	1.183	06:52	
Air Blank	0.000	06:53	
Air Blank	0.000	06:55	
Subject Sample	# 2	06:58	No Sample Provided
Breath Volume	0.000	06:58	
Air Blank	0.000	06:58	
Air Blank	0.000	07:01	
Subject Sample	# 3 0.082	07:04	Volume Not Met
Breath Volume	0.027	07:04	
Air Blank	0.000	07:04	
Control Test	0.081	07:05	
Air Blank	0.000	07:05	
Diagnostics Check	OK	07:05	

Data Download Date/Time: 12/28/2018 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRING,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HERRING,TIMOTHY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: DELLA FAVE

DATE/TIME OF TEST: 12/06/2018 19:05:42

Diagnostics Check	OK	18:57
Air Blank	0.000	18:58
Control Test	0.079	18:58
Air Blank	0.000	18:58
Subject Sample # 1	0.241	19:01
Breath Volume	1.128	19:01
Air Blank	0.000	19:02
Air Blank	0.000	19:03
Subject Sample # 2	0.229	19:04
Breath Volume	1.613	19:04
Air Blank	0.000	19:04
Control Test	0.076	19:05
Air Blank	0.000	19:05
Diagnostics Check	OK	19:05

Data Download Date/Time: 12/28/2018 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,SALVADOR,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: GARCIA,SALVADOR
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:06

SUBJECT'S LAST NAME: GROTH

DATE/TIME OF TEST: 12/08/2018 13:59:38

Diagnostics Check	OK	13:55	
Air Blank	0.000	13:56	
Control Test	0.080	13:56	
Air Blank	0.000	13:57	
Subject Sample	# 1	13:58	Subject Test Refused
Breath Volume	0.000	13:58	
Air Blank	0.000	13:58	
Control Test	0.080	13:59	
Air Blank	0.000	13:59	
Diagnostics Check	OK	13:59	

Data Download Date/Time: 12/28/2018 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEEFLANG,DANIEL,B
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: GASPAR,BENJAMIN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: IMRAN

DATE/TIME OF TEST: 12/12/2018 00:31:19

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.196	00:26
Breath Volume	2.382	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.197	00:29
Breath Volume	2.457	00:29
Air Blank	0.000	00:30
Control Test	0.078	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 12/28/2018 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,BRIAN,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: REYNOLDS,MATTHEW
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: RISLEY

DATE/TIME OF TEST: 12/13/2018 23:07:40

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.079	22:59
Air Blank	0.000	23:00
Subject Sample # 1	0.136	23:01
Breath Volume	1.355	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.131	23:06
Breath Volume	1.613	23:06
Air Blank	0.000	23:06
Control Test	0.080	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 12/28/2018 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRING,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MAHLER,JOSHUA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: JACQUES

DATE/TIME OF TEST: 12/14/2018 01:28:25

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:16	
Control Test	0.079	01:17	
Air Blank	0.000	01:17	
Subject Sample	# 1 0.156	01:20	Volume Not Met
Breath Volume	0.144	01:20	
Air Blank	0.000	01:21	
Air Blank	0.000	01:23	
Subject Sample	# 2 0.166	01:23	
Breath Volume	2.691	01:23	
Air Blank	0.000	01:24	
Air Blank	0.000	01:26	
Subject Sample	# 3 0.163	01:26	
Breath Volume	1.925	01:26	
Air Blank	0.000	01:27	
Control Test	0.076	01:27	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 12/28/2018 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS',JEREMY,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: SISTARELLI,NICHOLAS
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/15/2018 03:40:31

Diagnostics Check	OK	03:39	
Air Blank		03:40	Ambient Fail
Air Blank	0.000	03:40	

Data Download Date/Time: 12/28/2018 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JEREMY,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: SISTARELLI,NICHOLAS
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/15/2018 03:49:04

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.116	03:44
Breath Volume	3.558	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.114	03:47
Breath Volume	3.535	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:49

Data Download Date/Time: 12/28/2018 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRING,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HERRING,TIMOTHY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: GALLAGHER

DATE/TIME OF TEST: 12/16/2018 03:41:39

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.000	03:37
Breath Volume	2.683	03:37
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.000	03:40
Breath Volume	2.597	03:40
Air Blank	0.000	03:40
Control Test	0.079	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 12/28/2018 18:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,BRIAN,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: THOMPSON,BRIAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:56

SUBJECT'S LAST NAME: LOBRUTTO

DATE/TIME OF TEST: 12/16/2018 08:26:14

Diagnostics Check	OK	08:19
Air Blank	0.000	08:20
Control Test	0.080	08:20
Air Blank	0.000	08:21
Subject Sample # 1	0.087	08:21
Breath Volume	2.246	08:21
Air Blank	0.000	08:22
Air Blank	0.000	08:24
Subject Sample # 2	0.085	08:24
Breath Volume	2.324	08:24
Air Blank	0.000	08:25
Control Test	0.080	08:25
Air Blank	0.000	08:26
Diagnostics Check	OK	08:26

Data Download Date/Time: 12/28/2018 18:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INSUASTY,SANDRA,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: YEBBA,RICKY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DUFF

DATE/TIME OF TEST: 12/21/2018 02:56:27

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.246	02:52
Breath Volume	1.753	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.242	02:54
Breath Volume	1.316	02:54
Air Blank	0.000	02:55
Control Test	0.078	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 12/28/2018 18:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,BRIAN,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: OCHOA,NELSON
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: HOFER

DATE/TIME OF TEST: 12/21/2018 23:11:45

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.184	23:06
Breath Volume	2.386	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.180	23:10
Breath Volume	2.546	23:10
Air Blank	0.000	23:10
Control Test	0.079	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 12/28/2018 18:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,SALVADOR,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: GARCIA,SALVADOR
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: DE ROGATIS

DATE/TIME OF TEST: 12/22/2018 20:16:16

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.080	20:11
Air Blank	0.000	20:11
Subject Sample # 1	0.000	20:12
Breath Volume	2.230	20:12
Air Blank	0.000	20:12
Air Blank	0.000	20:14
Subject Sample # 2	0.000	20:14
Breath Volume	1.703	20:14
Air Blank	0.000	20:15
Control Test	0.080	20:15
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

Data Download Date/Time: 12/28/2018 18:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRING,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HERRING,TIMOTHY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 12/22/2018 20:57:20

Diagnostics Check	OK	20:50
Air Blank	0.000	20:51
Control Test	0.080	20:51
Air Blank	0.000	20:52
Subject Sample # 1	0.126	20:52
Breath Volume	3.566	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.121	20:55
Breath Volume	2.324	20:55
Air Blank	0.000	20:56
Control Test	0.078	20:56
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 12/28/2018 18:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREVISOL,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: SISTERLLI,NICOLAS
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: BLOOD

DATE/TIME OF TEST: 12/26/2018 19:03:37

Diagnostics Check	OK	19:00	
Air Blank	0.000	19:01	
Control Test	0.079	19:01	
Air Blank	0.000	19:02	
Subject Sample	# 1	19:02	Subject Test Refused
Breath Volume	0.000	19:02	
Air Blank	0.000	19:02	
Control Test	0.080	19:03	
Air Blank	0.000	19:03	
Diagnostics Check	OK	19:03	

Data Download Date/Time: 12/28/2018 18:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOMOZA,EDUARDO,A
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SOMOZA,EDUARDO
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/01/2018 02:55:50

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:52	
Control Test	0.079	02:53	
Air Blank	0.000	02:53	
Subject Sample	# 1	02:54	Subject Test Refused
Breath Volume	0.000	02:54	
Air Blank	0.000	02:54	
Control Test	0.079	02:55	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,TERRENCE,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: NEWSOM,SGT
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 12:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:24

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/01/2018 14:04:53

Diagnostics Check	OK	13:54
Air Blank	0.000	13:55
Control Test	0.081	13:55
Air Blank	0.000	13:55
Subject Sample # 1	0.237	13:57
Breath Volume	2.218	13:57
Air Blank	0.000	13:58
Air Blank	0.000	13:59
Subject Sample # 2	0.222	14:03
Breath Volume	1.597	14:03
Air Blank	0.000	14:03
Control Test	0.080	14:04
Air Blank	0.000	14:04
Diagnostics Check	OK	14:04

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,OFFICER
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: NOBLE

DATE/TIME OF TEST: 12/01/2018 23:35:43

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:33	
Control Test	0.079	23:33	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:34	
Control Test	0.079	23:35	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:35	

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,LATONYA,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: AMUSO

DATE/TIME OF TEST: 12/03/2018 01:24:02

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:21	
Control Test	0.080	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:23	
Control Test	0.079	01:23	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRANDES,M
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: PERALA

DATE/TIME OF TEST: 12/05/2018 03:56:47

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:54	
Control Test	0.080	03:54	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Control Test	0.080	03:56	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: CAMERON

DATE/TIME OF TEST: 12/05/2018 19:48:08

Diagnostics Check	OK	19:39	
Air Blank	0.000	19:40	
Control Test	0.081	19:40	
Air Blank	0.000	19:40	
Subject Sample	# 1	19:44	No Sample Provided
Breath Volume	0.000	19:44	
Air Blank	0.000	19:44	
Air Blank	0.000	19:46	
Subject Sample	# 2	19:46	Subject Test Refused
Breath Volume	0.000	19:46	
Air Blank	0.000	19:47	
Control Test	0.081	19:47	
Air Blank	0.000	19:48	
Diagnostics Check	OK	19:48	

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: LARIOS VARGAS

DATE/TIME OF TEST: 12/07/2018 03:33:22

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.080	03:31	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.079	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: MAYNARD

DATE/TIME OF TEST: 12/07/2018 20:20:45

Diagnostics Check	OK	20:14	
Air Blank	0.000	20:14	
Control Test	0.080	20:15	
Air Blank	0.000	20:15	
Subject Sample	# 1 0.125	20:16	
Breath Volume	3.859	20:16	
Air Blank	0.000	20:16	
Air Blank	0.000	20:18	
Subject Sample	# 2	20:19	Subject Test Refused
Breath Volume	0.000	20:19	
Air Blank	0.000	20:19	
Control Test	0.079	20:20	
Air Blank	0.000	20:20	
Diagnostics Check	OK	20:20	

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SANTANA,R
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 12/07/2018 21:27:00

Diagnostics Check	OK	21:16
Air Blank	0.000	21:17
Control Test	0.080	21:17
Air Blank	0.000	21:17
Subject Sample # 1	0.237	21:20
Breath Volume	1.253	21:20
Air Blank	0.000	21:21
Air Blank	0.000	21:22
Subject Sample # 2	0.236	21:25
Breath Volume	1.656	21:25
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:26
Diagnostics Check	OK	21:27

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVIL,LUXERNE,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PREVIL,LUXERNE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: OROZCO CRUZ

DATE/TIME OF TEST: 12/08/2018 05:17:04

Diagnostics Check	OK	05:10
Air Blank	0.000	05:10
Control Test	0.080	05:11
Air Blank	0.000	05:11
Subject Sample # 1	0.055	05:12
Breath Volume	2.988	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.052	05:15
Breath Volume	2.742	05:15
Air Blank	0.000	05:16
Control Test	0.080	05:16
Air Blank	0.000	05:16
Diagnostics Check	OK	05:17

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,G
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: PLUMB,G
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: KIRBY

DATE/TIME OF TEST: 12/09/2018 03:44:45

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:33	
Control Test	0.079	03:34	
Air Blank	0.000	03:34	
Subject Sample	# 1	03:37	No Sample Provided
Breath Volume	0.000	03:37	
Air Blank	0.000	03:38	
Air Blank	0.000	03:40	
Subject Sample	# 2	03:43	No Sample Provided
Breath Volume	0.000	03:43	
Air Blank	0.000	03:43	
Control Test	0.079	03:44	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,G
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: PLUMB,G
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: KIRBY

DATE/TIME OF TEST: 12/09/2018 04:23:09

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:20	
Control Test	0.079	04:21	
Air Blank	0.000	04:21	
Subject Sample	# 1	04:21	Subject Test Refused
Breath Volume	0.000	04:21	
Air Blank	0.000	04:22	
Control Test	0.080	04:22	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:23	

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RISPOLI,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 14:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 12/10/2018 16:00:08

Diagnostics Check	OK	15:54
Air Blank	0.000	15:54
Control Test	0.080	15:54
Air Blank	0.000	15:55
Subject Sample # 1	0.184	15:55
Breath Volume	1.972	15:55
Air Blank	0.000	15:56
Air Blank	0.000	15:58
Subject Sample # 2	0.174	15:58
Breath Volume	1.910	15:58
Air Blank	0.000	15:59
Control Test	0.080	15:59
Air Blank	0.000	16:00
Diagnostics Check	OK	16:00

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVIL,LUXERNE,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PREVIL,LUXERNE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MATEO GARCIA

DATE/TIME OF TEST: 12/11/2018 02:47:19

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.169	02:42
Breath Volume	2.253	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.171	02:45
Breath Volume	2.195	02:45
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANTORSKI, MARK, A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RODRIGUEZ, ALEXANDER
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 12/12/2018 02:24:42

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:13	
Control Test	0.081	02:14	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:17	No Sample Provided
Breath Volume	0.000	02:17	
Air Blank	0.000	02:18	
Air Blank	0.000	02:20	
Subject Sample	# 2	02:23	No Sample Provided
Breath Volume	0.000	02:23	
Air Blank	0.000	02:23	
Control Test	0.081	02:24	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANTORSKI, MARK, A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RODRIGUEZ, ALEXANDER
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 12/12/2018 02:33:35

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.081	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:32	
Control Test	0.081	02:32	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HOILETT,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 12/13/2018 20:22:05

Diagnostics Check	OK	20:19	
Air Blank	0.000	20:19	
Control Test	0.080	20:20	
Air Blank	0.000	20:20	
Subject Sample	# 1	20:20	Subject Test Refused
Breath Volume	0.000	20:20	
Air Blank	0.000	20:21	
Control Test	0.080	20:21	
Air Blank	0.000	20:22	
Diagnostics Check	OK	20:22	

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE, BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CORNWELL

DATE/TIME OF TEST: 12/14/2018 22:50:06

Diagnostics Check	OK	22:43
Air Blank	0.000	22:44
Control Test	0.080	22:44
Air Blank	0.000	22:44
Subject Sample # 1	0.229	22:45
Breath Volume	1.808	22:45
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.243	22:48
Breath Volume	1.613	22:48
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MELIUS,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: DESANCTIS

DATE/TIME OF TEST: 12/18/2018 20:19:24

Diagnostics Check	OK	20:12
Air Blank	0.000	20:12
Control Test	0.081	20:13
Air Blank	0.000	20:13
Subject Sample # 1	0.160	20:15
Breath Volume	1.582	20:15
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.152	20:17
Breath Volume	1.722	20:17
Air Blank	0.000	20:18
Control Test	0.080	20:18
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 12/20/2018 22:04:30

Diagnostics Check	OK	22:01	
Air Blank	0.000	22:02	
Control Test	0.080	22:02	
Air Blank	0.000	22:02	
Subject Sample	# 1	22:03	Subject Test Refused
Breath Volume	0.000	22:03	
Air Blank	0.000	22:03	
Control Test	0.080	22:03	
Air Blank	0.000	22:04	
Diagnostics Check	OK	22:04	

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: KEGERREIS

DATE/TIME OF TEST: 12/21/2018 01:02:03

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:58	
Control Test	0.081	00:59	
Air Blank	0.000	00:59	
Subject Sample	# 1	01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:01	
Control Test	0.081	01:01	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ, OSCAR,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: DOTY, JAMES
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: WIEBELT

DATE/TIME OF TEST: 12/22/2018 03:46:19

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.137	03:40
Breath Volume	1.445	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.145	03:44
Breath Volume	1.753	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: WILKINSON,KEVIN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: FLORIAN

DATE/TIME OF TEST: 12/23/2018 03:44:08

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.080	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:43	
Control Test	0.081	03:43	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DE LEON,TEA NICOLE
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/24/2018 02:35:35

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:33	
Control Test	0.080	02:33	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:34	
Control Test	0.080	02:35	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 12/24/2018 05:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: VAUGHAN

DATE/TIME OF TEST: 12/24/2018 23:18:04

Diagnostics Check	OK	23:15	
Air Blank	0.000	23:15	
Control Test	0.080	23:15	
Air Blank	0.000	23:16	
Subject Sample	# 1	23:16	Subject Test Refused
Breath Volume	0.000	23:16	
Air Blank	0.000	23:17	
Control Test	0.080	23:17	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:18	

Data Download Date/Time: 01/04/2019 03:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,LATONYA,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: GRIGAS

DATE/TIME OF TEST: 12/25/2018 04:06:21

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.080	04:04	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:05	
Control Test	0.080	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 01/04/2019 03:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HOILETT,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: SARMIENTO

DATE/TIME OF TEST: 12/26/2018 20:09:19

Diagnostics Check	OK	20:06	
Air Blank	0.000	20:06	
Control Test	0.080	20:07	
Air Blank	0.000	20:07	
Subject Sample	# 1	20:07	Subject Test Refused
Breath Volume	0.000	20:07	
Air Blank	0.000	20:08	
Control Test	0.080	20:08	
Air Blank	0.000	20:09	
Diagnostics Check	OK	20:09	

Data Download Date/Time: 01/04/2019 03:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKEON,A
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 12/26/2018 23:18:08

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.080	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.000	23:13
Breath Volume	2.707	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.000	23:16
Breath Volume	2.078	23:16
Air Blank	0.000	23:17
Control Test	0.080	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 01/04/2019 03:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANTORSKI, MARK, A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BELL, BRIAN
ARRESTING OFFICER AGENCY: Indian River Shores
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 12/27/2018 02:38:55

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.080	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:37	Subject Test Refused
Breath Volume	0.000	02:37	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 01/04/2019 03:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CHALKER

DATE/TIME OF TEST: 12/29/2018 01:06:54

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:52	
Control Test	0.079	00:52	
Air Blank	0.000	00:53	
Subject Sample	# 1 0.040	00:56	Volume Not Met
Breath Volume	0.398	00:56	
Air Blank	0.000	00:56	
Air Blank	0.000	00:58	
Subject Sample	# 2 0.043	00:59	
Breath Volume	1.855	00:59	
Air Blank	0.000	01:00	
Air Blank	0.000	01:02	
Subject Sample	# 3 0.035	01:05	Volume Not Met
Breath Volume	0.644	01:05	
Air Blank	0.000	01:06	
Control Test	0.078	01:06	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 01/04/2019 03:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CHALKER

DATE/TIME OF TEST: 12/29/2018 01:17:08

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.032	01:12
Breath Volume	1.785	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.037	01:15
Breath Volume	1.558	01:15
Air Blank	0.000	01:16
Control Test	0.079	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 01/04/2019 03:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 12/29/2018 23:40:03

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:39	
Control Test	0.079	23:39	
Air Blank		23:39	RFI Detect
Air Blank	0.000	23:40	

Data Download Date/Time: 01/04/2019 03:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 12/29/2018 23:46:59

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:44	
Control Test	0.079	23:44	
Air Blank	0.000	23:45	
Subject Sample	# 1	23:45	Subject Test Refused
Breath Volume	0.000	23:45	
Air Blank	0.000	23:46	
Control Test	0.079	23:46	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 01/04/2019 03:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHIMENTI,RICHARD,J
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: AYALA,K
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 10:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:13

SUBJECT'S LAST NAME: GAMBARRESSE

DATE/TIME OF TEST: 12/30/2018 11:49:35

Diagnostics Check	OK	11:48	
Air Blank	0.000	11:48	
Control Test	0.079	11:48	
Air Blank		11:49	RFI Detect
Air Blank	0.000	11:49	

Data Download Date/Time: 01/04/2019 03:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHIMENTI,RICHARD,J
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: AYLALA,K
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 10:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:13

SUBJECT'S LAST NAME: GAMBARRESE

DATE/TIME OF TEST: 12/30/2018 11:57:20

Diagnostics Check	OK	11:54	
Air Blank	0.000	11:54	
Control Test	0.079	11:55	
Air Blank	0.000	11:55	
Subject Sample	# 1	11:55	Subject Test Refused
Breath Volume	0.000	11:55	
Air Blank	0.000	11:56	
Control Test	0.080	11:56	
Air Blank	0.000	11:57	
Diagnostics Check	OK	11:57	

Data Download Date/Time: 01/04/2019 03:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: DOLNICK

DATE/TIME OF TEST: 12/31/2018 01:21:26

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.080	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.060	01:16
Breath Volume	2.746	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.053	01:19
Breath Volume	1.734	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 01/04/2019 03:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YORK,JIMMY,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MACKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 12/31/2018 21:32:06

Diagnostics Check	OK	21:25
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:26
Subject Sample # 1	0.193	21:27
Breath Volume	1.628	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:29
Subject Sample # 2	0.206	21:30
Breath Volume	1.832	21:30
Air Blank	0.000	21:31
Control Test	0.079	21:31
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 01/04/2019 03:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YORK, JIMMY, A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI, KEVIN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: TERSIGNI

DATE/TIME OF TEST: 12/31/2018 22:05:35

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:03	
Control Test	0.080	22:03	
Air Blank	0.000	22:04	
Subject Sample	# 1	22:04	Subject Test Refused
Breath Volume	0.000	22:04	
Air Blank	0.000	22:04	
Control Test	0.080	22:04	
Air Blank	0.000	22:05	
Diagnostics Check	OK	22:05	

Data Download Date/Time: 01/04/2019 03:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YORK, JIMMY, A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HOILETT, DWAYNE
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 12/31/2018 22:40:21

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.079	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.215	22:35
Breath Volume	1.375	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.211	22:38
Breath Volume	1.804	22:38
Air Blank	0.000	22:39
Control Test	0.080	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 01/04/2019 03:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUMB,GREGORY,T
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: DONSIFF,MATTHEW
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 12/22/2018 21:05:21

Diagnostics Check	OK	20:57
Air Blank	0.000	20:58
Control Test	0.082	20:58
Air Blank	0.000	20:58
Subject Sample # 1	0.151	21:00
Breath Volume	1.781	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:02
Subject Sample # 2	0.166	21:03
Breath Volume	1.898	21:03
Air Blank	0.000	21:04
Control Test	0.082	21:04
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 12/24/2018 05:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASSEY,DARRELL,L
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BOGUSKY,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 12/01/2018 01:26:58

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:24	
Control Test	0.081	01:24	
Air Blank	0.000	01:25	
Subject Sample	# 1	01:25	Subject Test Refused
Breath Volume	0.000	01:25	
Air Blank	0.000	01:26	
Control Test	0.081	01:26	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

Data Download Date/Time: 12/27/2018 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JOHN,P
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: SANTANA,ELIO
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 11/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 12/22/2018 23:59:49

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.081	23:57	
Air Blank	0.000	23:58	
Subject Sample	# 1	23:58	Subject Test Refused
Breath Volume	0.000	23:58	
Air Blank	0.000	23:58	
Control Test	0.080	23:59	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 12/27/2018 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MINGS,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:03

SUBJECT'S LAST NAME: CARLIN

DATE/TIME OF TEST: 12/01/2018 17:40:50

Diagnostics Check	OK	17:33
Air Blank	0.000	17:34
Control Test	0.080	17:34
Air Blank	0.000	17:35
Subject Sample # 1	0.238	17:35
Breath Volume	1.445	17:35
Air Blank	0.000	17:36
Air Blank	0.000	17:38
Subject Sample # 2	0.240	17:39
Breath Volume	2.601	17:39
Air Blank	0.000	17:39
Control Test	0.080	17:40
Air Blank	0.000	17:40
Diagnostics Check	OK	17:40

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: CAMBRIA,STEVEN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: PECZKOWSKI

DATE/TIME OF TEST: 12/04/2018 23:44:19

Diagnostics Check	OK	23:37
Air Blank	0.000	23:38
Control Test	0.080	23:38
Air Blank	0.000	23:39
Subject Sample # 1	0.220	23:39
Breath Volume	1.281	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.217	23:42
Breath Volume	1.222	23:42
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: BREIDENBACH

DATE/TIME OF TEST: 12/06/2018 02:26:41

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.204	02:21
Breath Volume	2.371	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.198	02:24
Breath Volume	1.417	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CARDWELL,CAPTAIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: HUTSON

DATE/TIME OF TEST: 12/06/2018 22:21:15

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.080	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.115	22:16
Breath Volume	2.320	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.119	22:19
Breath Volume	2.167	22:19
Air Blank	0.000	22:20
Control Test	0.080	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,OFC,W
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: PERKINS,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: STUTZMAN

DATE/TIME OF TEST: 12/07/2018 00:02:43

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.223	23:58
Breath Volume	2.136	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.235	00:01
Breath Volume	2.410	00:01
Air Blank	0.000	00:01
Control Test	0.078	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON, HALSTON, L
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: SAYED, JEREMIAH
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: SCULLY

DATE/TIME OF TEST: 12/07/2018 02:43:51

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.169	02:39
Breath Volume	2.160	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.176	02:42
Breath Volume	4.484	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINGO,JARED,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ESTEVEZ,JOSE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: LINGO

DATE/TIME OF TEST: 12/08/2018 07:44:44

Diagnostics Check	OK	07:38
Air Blank	0.000	07:38
Control Test	0.080	07:38
Air Blank	0.000	07:39
Subject Sample # 1	0.114	07:40
Breath Volume	3.125	07:40
Air Blank	0.000	07:40
Air Blank	0.000	07:42
Subject Sample # 2	0.113	07:43
Breath Volume	3.648	07:43
Air Blank	0.000	07:43
Control Test	0.079	07:44
Air Blank	0.000	07:44
Diagnostics Check	OK	07:44

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: CAMBRIA,STEVEN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SAWYER

DATE/TIME OF TEST: 12/10/2018 01:32:37

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.249	01:27
Breath Volume	3.003	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.250	01:30
Breath Volume	2.886	01:30
Air Blank	0.000	01:31
Control Test	0.079	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BLOMGREN,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: LEACH

DATE/TIME OF TEST: 12/11/2018 01:32:21

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.080	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.075	01:27
Breath Volume	1.542	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.080	01:30
Breath Volume	1.875	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: GILSON

DATE/TIME OF TEST: 12/13/2018 00:32:11

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.143	00:27
Breath Volume	1.277	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:29
Subject Sample # 2	0.143	00:30
Breath Volume	1.914	00:30
Air Blank	0.000	00:31
Control Test	0.080	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLYNN,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: STOLL,A
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: RUSISKI

DATE/TIME OF TEST: 12/15/2018 00:06:07

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.079	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.162	00:01
Breath Volume	2.445	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:03
Subject Sample # 2	0.165	00:04
Breath Volume	2.058	00:04
Air Blank	0.000	00:05
Control Test	0.078	00:05
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCLJOSE,F
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TUCCLJOSE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: BROZA

DATE/TIME OF TEST: 12/15/2018 20:44:07

Diagnostics Check	OK	20:36
Air Blank	0.000	20:37
Control Test	0.081	20:37
Air Blank	0.000	20:38
Subject Sample # 1	0.136	20:39
Breath Volume	2.199	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.133	20:42
Breath Volume	2.347	20:42
Air Blank	0.000	20:43
Control Test	0.079	20:43
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SPENCER,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: BREEDING

DATE/TIME OF TEST: 12/16/2018 01:30:28

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.079	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.165	01:25
Breath Volume	2.027	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.170	01:28
Breath Volume	2.027	01:28
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLYNN,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CARTER,K
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 12/16/2018 02:30:46

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.000	02:26
Breath Volume	2.851	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.000	02:29
Breath Volume	3.210	02:29
Air Blank	0.000	02:29
Control Test	0.079	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTWELL,J,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CANTWELL,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 12/16/2018 05:07:36

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:02
Subject Sample # 1	0.227	05:03
Breath Volume	2.761	05:03
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.227	05:05
Breath Volume	2.597	05:05
Air Blank	0.000	05:06
Control Test	0.079	05:06
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLETON,LUCINDA,B
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: HEYMAN,LUKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:44

SUBJECT'S LAST NAME: LITTLETON

DATE/TIME OF TEST: 12/16/2018 13:11:46

Diagnostics Check	OK	13:05
Air Blank	0.000	13:05
Control Test	0.080	13:06
Air Blank	0.000	13:06
Subject Sample # 1	0.150	13:07
Breath Volume	2.773	13:07
Air Blank	0.000	13:07
Air Blank	0.000	13:09
Subject Sample # 2	0.139	13:10
Breath Volume	3.238	13:10
Air Blank	0.000	13:10
Control Test	0.079	13:11
Air Blank	0.000	13:11
Diagnostics Check	OK	13:11

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KENNY,OFFICER
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: BOUHAMID

DATE/TIME OF TEST: 12/16/2018 22:02:04

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.080	21:55
Air Blank	0.000	21:55
Subject Sample # 1	0.189	21:57
Breath Volume	1.445	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.181	22:00
Breath Volume	1.445	22:00
Air Blank	0.000	22:01
Control Test	0.077	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:02

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOLADVANDI,OFC
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: NETZEL

DATE/TIME OF TEST: 12/18/2018 02:37:02

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.079	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.188	02:32
Breath Volume	2.382	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.193	02:35
Breath Volume	2.476	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:37

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,OFC,W
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: PERKINS,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: NEWCOMBE

DATE/TIME OF TEST: 12/19/2018 22:49:25

Diagnostics Check	OK	22:42
Air Blank	0.000	22:43
Control Test	0.080	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.221	22:44
Breath Volume	1.617	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.224	22:47
Breath Volume	2.003	22:47
Air Blank	0.000	22:48
Control Test	0.079	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ROSE,W
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 12/20/2018 20:57:30

Diagnostics Check	OK	20:47
Air Blank	0.000	20:47
Control Test	0.080	20:48
Air Blank	0.000	20:48
Subject Sample # 1	0.213	20:49
Breath Volume	1.390	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.250	20:52
Breath Volume	3.453	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 3	0.235	20:55
Breath Volume	2.496	20:55
Air Blank	0.000	20:56
Control Test	0.079	20:56
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTWELL,J,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CANTWELL,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: NEWLIN

DATE/TIME OF TEST: 12/21/2018 02:32:53

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.080	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.140	02:28
Breath Volume	3.042	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:30
Subject Sample # 2	0.141	02:31
Breath Volume	3.277	02:31
Air Blank	0.000	02:31
Control Test	0.079	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: CAROTENUTO

DATE/TIME OF TEST: 12/22/2018 19:46:20

Diagnostics Check	OK	19:36
Air Blank	0.000	19:37
Control Test	0.080	19:37
Air Blank	0.000	19:37
Subject Sample # 1	0.244	19:41
Breath Volume	1.843	19:41
Air Blank	0.000	19:41
Air Blank	0.000	19:43
Subject Sample # 2	0.242	19:44
Breath Volume	1.445	19:44
Air Blank	0.000	19:45
Control Test	0.079	19:45
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEVEZ,JOSE,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ESTEVEZ,JOSE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 12/23/2018 05:22:24

Diagnostics Check	OK	05:19	
Air Blank	0.000	05:20	
Control Test	0.082	05:20	
Air Blank	0.000	05:20	
Subject Sample	# 1	05:21	Subject Test Refused
Breath Volume	0.000	05:21	
Air Blank	0.000	05:21	
Control Test	0.081	05:21	
Air Blank	0.000	05:22	
Diagnostics Check	OK	05:22	

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEVEZ,JOSE,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ESTEVEZ,JOSE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: HEFNER

DATE/TIME OF TEST: 12/23/2018 23:18:29

Diagnostics Check	OK	23:15	
Air Blank	0.000	23:16	
Control Test	0.081	23:16	
Air Blank	0.000	23:16	
Subject Sample	# 1	23:17	Subject Test Refused
Breath Volume	0.000	23:17	
Air Blank	0.000	23:17	
Control Test	0.082	23:17	
Air Blank	0.000	23:18	
Diagnostics Check	OK	23:18	

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHROTH,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: PERIN

DATE/TIME OF TEST: 12/24/2018 04:12:49

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.082	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.117	04:08
Breath Volume	1.687	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:10
Subject Sample # 2	0.112	04:11
Breath Volume	1.726	04:11
Air Blank	0.000	04:11
Control Test	0.080	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTWELL,J,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CANTWELL,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: RIVERA ROSARIO

DATE/TIME OF TEST: 12/25/2018 22:22:19

Diagnostics Check	OK	22:13
Air Blank	0.000	22:14
Control Test	0.081	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.131	22:16
Breath Volume	1.644	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.124	22:20
Breath Volume	2.378	22:20
Air Blank	0.000	22:21
Control Test	0.078	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCLJOSE,F
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TUCCLJOSE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/25/2018 23:18:04

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:12	
Control Test	0.081	23:12	
Air Blank	0.000	23:13	
Subject Sample	# 1 0.259	23:14	
Breath Volume	1.796	23:14	
Air Blank	0.000	23:15	
Air Blank	0.000	23:16	
Subject Sample	# 2 0.257	23:17	
Breath Volume	1.179	23:17	
Air Blank		23:17	RFI Detect
Air Blank	0.000	23:18	

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON, CHADDICK, R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PIERRE LOUIS, KENOL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/26/2018 12:02:29

Diagnostics Check	OK	11:55
Air Blank	0.000	11:56
Control Test	0.081	11:56
Air Blank	0.000	11:57
Subject Sample # 1	0.093	11:57
Breath Volume	2.453	11:57
Air Blank	0.000	11:58
Air Blank	0.000	12:00
Subject Sample # 2	0.087	12:00
Breath Volume	1.820	12:00
Air Blank	0.000	12:01
Control Test	0.080	12:01
Air Blank	0.000	12:02
Diagnostics Check	OK	12:02

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRITCHARD,JASON,W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUTLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: STEED

DATE/TIME OF TEST: 12/26/2018 17:15:16

Diagnostics Check	OK	17:08
Air Blank	0.000	17:09
Control Test	0.080	17:09
Air Blank	0.000	17:10
Subject Sample # 1	0.169	17:10
Breath Volume	2.625	17:10
Air Blank	0.000	17:11
Air Blank	0.000	17:13
Subject Sample # 2	0.163	17:13
Breath Volume	2.007	17:13
Air Blank	0.000	17:14
Control Test	0.077	17:14
Air Blank	0.000	17:15
Diagnostics Check	OK	17:15

Data Download Date/Time: 12/28/2018 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: HUDGINS,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 12/28/2018 23:45:29

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:34	
Control Test	0.081	23:34	
Air Blank	0.000	23:35	
Subject Sample	# 1 0.213	23:38	Volume Not Met
Breath Volume	0.015	23:38	
Air Blank	0.000	23:38	
Air Blank	0.000	23:40	
Subject Sample	# 2 0.150	23:43	Volume Not Met
Breath Volume	0.136	23:43	
Air Blank	0.000	23:44	
Control Test	0.079	23:44	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 01/30/2019 07:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,STEFANIE,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PETERSON,STEFANIE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: AMAN

DATE/TIME OF TEST: 12/29/2018 03:46:20

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.078	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.192	03:41
Breath Volume	1.156	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:43
Subject Sample # 2	0.184	03:44
Breath Volume	1.472	03:44
Air Blank	0.000	03:45
Control Test	0.077	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 01/30/2019 07:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CHRISTIE,J
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CANO

DATE/TIME OF TEST: 12/30/2018 01:21:36

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.184	01:17
Breath Volume	2.515	01:17
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.184	01:19
Breath Volume	3.070	01:19
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 01/30/2019 07:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCL,JOS,F
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TUCCL,JOSE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: GRANDE

DATE/TIME OF TEST: 12/30/2018 04:34:32

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.097	04:29
Breath Volume	3.777	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.106	04:32
Breath Volume	2.636	04:32
Air Blank	0.000	04:33
Control Test	0.078	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 01/30/2019 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MORALES, RONNIE,

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: FIEDLER, CHRISTOPHER

ARRESTING OFFICER AGENCY: Eustis PD

LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1031121

DRY GAS EXPIRY DATE: 10/12/2021

VIOLATION CODE: DUI

ARREST TIME: 04:46

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 12/08/2018 05:53:26

Diagnostics Check	OK	05:50	
Air Blank	0.000	05:51	
Control Test	0.082	05:51	
Air Blank	0.000	05:51	
Subject Sample	# 1	05:52	Subject Test Refused
Breath Volume	0.000	05:52	
Air Blank	0.000	05:52	
Control Test	0.081	05:52	
Air Blank	0.000	05:53	
Diagnostics Check	OK	05:53	

Data Download Date/Time: 01/02/2019 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MORALES, RONNIE,

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: PERRY JR, WAYNE

ARRESTING OFFICER AGENCY: Eustis PD

LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1031121

DRY GAS EXPIRY DATE: 10/12/2021

VIOLATION CODE: DUI

ARREST TIME: 03:27

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: JEWELL

DATE/TIME OF TEST: 12/09/2018 03:57:49

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:55	
Control Test	0.081	03:55	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:56	
Control Test	0.081	03:57	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 01/02/2019 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,VICTOR,E
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: CORTES,VICTOR
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: MARSALA

DATE/TIME OF TEST: 12/31/2018 03:00:20

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:52	
Control Test	0.081	02:52	
Air Blank	0.000	02:53	
Subject Sample	# 1 0.166	02:54	
Breath Volume	1.847	02:54	
Air Blank	0.000	02:54	
Air Blank	0.000	02:56	
Subject Sample	# 2 0.157	02:58	
Breath Volume	1.695	02:58	
Air Blank	0.000	02:59	
Control Test	0.080	02:59	
Air Blank		03:00	RFI Detect
Air Blank		03:00	RFI Detect

Data Download Date/Time: 01/02/2019 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,VICTOR,E
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: CORTES,VICTOR
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: MARSALA

DATE/TIME OF TEST: 12/31/2018 03:14:05

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.158	03:09
Breath Volume	1.226	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.153	03:12
Breath Volume	2.085	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 01/02/2019 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: PIPKINS

DATE/TIME OF TEST: 12/06/2018 19:48:19

Diagnostics Check	OK	19:42
Air Blank	0.000	19:42
Control Test	0.080	19:43
Air Blank	0.000	19:43
Subject Sample # 1	0.148	19:43
Breath Volume	2.195	19:43
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample # 2	0.143	19:46
Breath Volume	1.714	19:46
Air Blank	0.000	19:47
Control Test	0.079	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 12/28/2018 08:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BENEKE

DATE/TIME OF TEST: 12/07/2018 20:06:57

Diagnostics Check	OK	20:00
Air Blank	0.000	20:01
Control Test	0.079	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.164	20:02
Breath Volume	1.382	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.164	20:05
Breath Volume	1.328	20:05
Air Blank	0.000	20:06
Control Test	0.078	20:06
Air Blank	0.000	20:06
Diagnostics Check	OK	20:06

Data Download Date/Time: 12/28/2018 08:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BREIDENBACH

DATE/TIME OF TEST: 12/06/2018 01:35:40

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.079	01:30	
Air Blank	0.000	01:31	
Subject Sample	# 1	01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:34	
Control Test	0.079	01:35	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 12/28/2018 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 12/20/2018 18:45:16

Diagnostics Check	OK	18:38
Air Blank	0.000	18:38
Control Test	0.080	18:39
Air Blank	0.000	18:39
Subject Sample # 1	0.174	18:40
Breath Volume	2.976	18:40
Air Blank	0.000	18:41
Air Blank	0.000	18:43
Subject Sample # 2	0.156	18:43
Breath Volume	1.539	18:43
Air Blank	0.000	18:44
Control Test	0.078	18:44
Air Blank	0.000	18:45
Diagnostics Check	OK	18:45

Data Download Date/Time: 12/28/2018 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 12/07/2018 01:13:02

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.000	01:08
Breath Volume	2.515	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.000	01:11
Breath Volume	2.410	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:13

Data Download Date/Time: 12/28/2018 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: MAGARIS

DATE/TIME OF TEST: 12/07/2018 23:33:58

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.079	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.133	23:29
Breath Volume	1.960	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:31
Subject Sample # 2	0.133	23:32
Breath Volume	1.453	23:32
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 12/28/2018 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 12/16/2018 01:08:34

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.078	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.171	01:04
Breath Volume	3.285	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.167	01:07
Breath Volume	2.980	01:07
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 12/28/2018 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/27/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: SOMOGYI

DATE/TIME OF TEST: 12/21/2018 02:46:28

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.113	02:41
Breath Volume	2.574	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.110	02:45
Breath Volume	3.164	02:45
Air Blank	0.000	02:45
Control Test	0.078	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 12/28/2018 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: SIBERSKI-VALENTIN

DATE/TIME OF TEST: 12/02/2018 06:37:15

Diagnostics Check	OK	06:31
Air Blank	0.000	06:31
Control Test	0.082	06:32
Air Blank	0.000	06:32
Subject Sample # 1	0.134	06:32
Breath Volume	1.484	06:32
Air Blank	0.000	06:33
Air Blank	0.000	06:35
Subject Sample # 2	0.133	06:35
Breath Volume	1.855	06:35
Air Blank	0.000	06:36
Control Test	0.082	06:36
Air Blank	0.000	06:37
Diagnostics Check	OK	06:37

Data Download Date/Time: 12/17/2018 04:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANLEY,JEFFREY,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: STANLEY,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: MONTCALM

DATE/TIME OF TEST: 12/03/2018 04:31:55

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.082	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.000	04:27
Breath Volume	4.292	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.000	04:30
Breath Volume	3.984	04:30
Air Blank	0.000	04:31
Control Test	0.082	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 12/17/2018 04:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: PILLADO

DATE/TIME OF TEST: 12/09/2018 01:29:03

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.082	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.192	01:24
Breath Volume	3.996	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.200	01:27
Breath Volume	2.960	01:27
Air Blank	0.000	01:28
Control Test	0.082	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:29

Data Download Date/Time: 12/17/2018 04:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 12/14/2018 03:38:27

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.082	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.136	03:33
Breath Volume	1.503	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:35
Subject Sample # 2	0.131	03:36
Breath Volume	1.128	03:36
Air Blank	0.000	03:37
Control Test	0.081	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 12/17/2018 04:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRINSKO, THOMAS, K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON, JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 12/02/2018 03:16:35

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.079	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.178	03:12
Breath Volume	2.574	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.170	03:15
Breath Volume	2.417	03:15
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 12/18/2018 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: WOODRIDGE

DATE/TIME OF TEST: 12/03/2018 22:54:42

Diagnostics Check	OK	22:48
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.113	22:50
Breath Volume	4.750	22:50
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.112	22:53
Breath Volume	4.320	22:53
Air Blank	0.000	22:53
Control Test	0.079	22:54
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 12/18/2018 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,KENDRICK,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TURNER,KENDRICK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: VAMOSI

DATE/TIME OF TEST: 12/07/2018 19:41:55

Diagnostics Check	OK	19:33
Air Blank	0.000	19:33
Control Test	0.081	19:34
Air Blank	0.000	19:34
Subject Sample # 1	0.180	19:35
Breath Volume	1.371	19:35
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample # 2	0.168	19:40
Breath Volume	1.589	19:40
Air Blank	0.000	19:41
Control Test	0.080	19:41
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 12/18/2018 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BOSTICK,JESSE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: STICHTER

DATE/TIME OF TEST: 12/08/2018 00:52:46

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.207	00:47
Breath Volume	1.683	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.202	00:51
Breath Volume	1.742	00:51
Air Blank	0.000	00:51
Control Test	0.078	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 12/18/2018 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANKAM,DENZEL,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KANKAM,DENZEL
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/08/2018 02:46:33

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:35	
Control Test	0.080	02:35	
Air Blank	0.000	02:36	
Subject Sample	# 1	02:39	No Sample Provided
Breath Volume	0.000	02:39	
Air Blank	0.000	02:39	
Air Blank	0.000	02:41	
Subject Sample	# 2	02:44	Volume Not Met
Breath Volume	0.750	02:44	
Air Blank	0.000	02:45	
Control Test	0.080	02:45	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 12/18/2018 20:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: SCONYERS

DATE/TIME OF TEST: 12/10/2018 04:00:43

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.081	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.076	03:56
Breath Volume	4.363	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.076	03:59
Breath Volume	3.464	03:59
Air Blank	0.000	03:59
Control Test	0.081	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 12/18/2018 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWARDS,JACOB,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: ALSUBAUEALANAZI

DATE/TIME OF TEST: 12/12/2018 00:08:06

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:57	
Control Test	0.080	23:57	
Air Blank	0.000	23:57	
Subject Sample	# 1 0.133	00:00	Volume Not Met
Breath Volume	0.765	00:00	
Air Blank	0.000	00:01	
Air Blank	0.000	00:03	
Subject Sample	# 2 0.000	00:06	Volume Not Met
Breath Volume	0.015	00:06	
Air Blank	0.000	00:07	
Control Test	0.079	00:07	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 12/18/2018 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,KENDRICK,T
BREATH TEST OPERATOR AGENCY: Green Cove Springs
ARRESTING OFFICER: JOHNSON,TRAVIS
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 12/13/2018 22:33:51

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:28	
Control Test	0.080	22:28	
Air Blank	0.000	22:29	
Subject Sample	# 1	22:32	No Sample Provided
Breath Volume	0.000	22:32	
Air Blank		22:32	RFI Detect
Air Blank	0.000	22:32	
Control Test	0.080	22:33	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

Data Download Date/Time: 12/18/2018 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,KENDRICK,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: JOHNSON,TRAVIS
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 12/13/2018 22:44:16

Diagnostics Check	OK	22:35	
Air Blank	0.000	22:36	
Control Test	0.080	22:36	
Air Blank	0.000	22:36	
Subject Sample	# 1	22:40	No Sample Provided
Breath Volume	0.000	22:40	
Air Blank	0.000	22:40	
Air Blank	0.000	22:42	
Subject Sample	# 2	22:42	Subject Test Refused
Breath Volume	0.000	22:42	
Air Blank	0.000	22:43	
Control Test	0.080	22:43	
Air Blank	0.000	22:44	
Diagnostics Check	OK	22:44	

Data Download Date/Time: 12/18/2018 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANKAM,DENZEL,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KANKAM,DENZL
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 12/14/2018 01:20:26

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:09	
Control Test	0.080	01:09	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:13	No Sample Provided
Breath Volume	0.000	01:13	
Air Blank	0.000	01:13	
Air Blank	0.000	01:15	
Subject Sample	# 2	01:19	No Sample Provided
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Control Test	0.081	01:19	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 12/18/2018 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: ZEEDYK

DATE/TIME OF TEST: 12/14/2018 03:55:18

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.185	03:50
Breath Volume	2.808	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.182	03:53
Breath Volume	1.839	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 12/18/2018 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEIMIG,EMORIE,I
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: ECHEVERRIA

DATE/TIME OF TEST: 12/15/2018 02:53:14

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.115	02:48
Breath Volume	1.250	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:50
Subject Sample # 2	0.124	02:51
Breath Volume	2.199	02:51
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 12/18/2018 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: SAMS

DATE/TIME OF TEST: 12/22/2018 04:08:39

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.080	03:58	
Air Blank	0.000	03:58	
Subject Sample	# 1	04:01	No Sample Provided
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Air Blank	0.000	04:04	
Subject Sample	# 2	04:07	No Sample Provided
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Control Test	0.081	04:08	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 01/23/2019 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: BROOKE

DATE/TIME OF TEST: 12/23/2018 01:46:11

Diagnostics Check	OK	01:39
Air Blank	0.000	01:40
Control Test	0.081	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.166	01:41
Breath Volume	2.656	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.159	01:44
Breath Volume	1.753	01:44
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 01/23/2019 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,KENDRICK,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TURNER,KENDRICK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 12/23/2018 21:51:18

Diagnostics Check	OK	21:39	
Air Blank	0.000	21:40	
Control Test	0.080	21:40	
Air Blank	0.000	21:41	
Subject Sample	# 1	21:44	No Sample Provided
Breath Volume	0.000	21:44	
Air Blank	0.000	21:44	
Air Blank	0.000	21:46	
Subject Sample	# 2	21:49	No Sample Provided
Breath Volume	0.000	21:49	
Air Blank	0.000	21:50	
Control Test	0.080	21:50	
Air Blank	0.000	21:51	
Diagnostics Check	OK	21:51	

Data Download Date/Time: 01/23/2019 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 12/27/2018 03:20:40

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.107	03:16
Breath Volume	2.773	03:16
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.104	03:19
Breath Volume	2.089	03:19
Air Blank	0.000	03:19
Control Test	0.079	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 01/23/2019 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: BOKOR

DATE/TIME OF TEST: 12/28/2018 01:44:58

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.108	01:40
Breath Volume	3.964	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:42
Subject Sample # 2	0.109	01:43
Breath Volume	3.808	01:43
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 01/23/2019 21:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,STEPHAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLE,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 12/15/2018 03:38:05

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:27	
Control Test	0.080	03:27	
Air Blank	0.000	03:27	
Subject Sample	# 1	03:31	No Sample Provided
Breath Volume	0.000	03:31	
Air Blank	0.000	03:31	
Air Blank	0.000	03:33	
Subject Sample	# 2	03:36	No Sample Provided
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Control Test	0.081	03:37	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 12/26/2018 23:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,BRUCE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JOHNSON,BRUCE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: WOODARD

DATE/TIME OF TEST: 12/15/2018 05:23:41

Diagnostics Check	OK	05:22	
Air Blank	0.000	05:23	
		05:23	RFI Detect
Air Blank	0.000	05:23	

Data Download Date/Time: 12/26/2018 23:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,BRUCE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JOHNSON,BRUCE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: WOODARD

DATE/TIME OF TEST: 12/15/2018 05:38:38

Diagnostics Check	OK	05:26	
Air Blank	0.000	05:26	
Control Test	0.081	05:26	
Air Blank	0.000	05:27	
Subject Sample	# 1 0.092	05:30	Volume Not Met
Breath Volume	0.675	05:30	
Air Blank	0.000	05:31	
Air Blank	0.000	05:32	
Subject Sample	# 2 0.111	05:33	
Breath Volume	1.628	05:33	
Air Blank	0.000	05:34	
Air Blank	0.000	05:36	
Subject Sample	# 3 0.111	05:37	
Breath Volume	2.042	05:37	
Air Blank	0.000	05:37	
Control Test	0.080	05:38	
Air Blank	0.000	05:38	
Diagnostics Check	OK	05:38	

Data Download Date/Time: 12/26/2018 23:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,STEPHAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLE,STEPHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: TIEDEMANN DE WITTE

DATE/TIME OF TEST: 12/22/2018 02:38:04

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.195	02:33
Breath Volume	3.468	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.199	02:36
Breath Volume	3.527	02:36
Air Blank	0.000	02:37
Control Test	0.081	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:38

Data Download Date/Time: 12/26/2018 23:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCKENZIE,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:12

SUBJECT'S LAST NAME: FITZPATRICK

DATE/TIME OF TEST: 12/25/2018 13:39:00

Diagnostics Check	OK	13:32
Air Blank	0.000	13:33
Control Test	0.082	13:33
Air Blank	0.000	13:34
Subject Sample # 1	0.000	13:34
Breath Volume	2.246	13:34
Air Blank	0.000	13:35
Air Blank	0.000	13:37
Subject Sample # 2	0.000	13:37
Breath Volume	2.914	13:37
Air Blank	0.000	13:38
Control Test	0.081	13:38
Air Blank	0.000	13:38
Diagnostics Check	OK	13:39

Data Download Date/Time: 12/26/2018 23:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCFADDEN,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:19

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 12/28/2018 17:47:26

Diagnostics Check	OK	17:39
Air Blank	0.000	17:40
Control Test	0.081	17:40
Air Blank	0.000	17:40
Subject Sample # 1	0.254	17:42
Breath Volume	1.824	17:42
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample # 2	0.242	17:45
Breath Volume	1.679	17:45
Air Blank	0.000	17:46
Control Test	0.079	17:46
Air Blank	0.000	17:47
Diagnostics Check	OK	17:47

Data Download Date/Time: 01/30/2019 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DUNNING,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/26/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: LOIDA

DATE/TIME OF TEST: 12/28/2018 18:57:29

Diagnostics Check	OK	18:50
Air Blank	0.000	18:50
Control Test	0.080	18:51
Air Blank	0.000	18:51
Subject Sample # 1	0.160	18:52
Breath Volume	1.765	18:52
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.152	18:55
Breath Volume	1.570	18:55
Air Blank	0.000	18:56
Control Test	0.079	18:56
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 01/30/2019 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: POLLOCK,DEREK
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: DYE

DATE/TIME OF TEST: 12/02/2018 03:51:31

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.079	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.189	03:46
Breath Volume	3.203	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.190	03:50
Breath Volume	2.144	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 12/28/2018 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,DEREK,V
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: POLLOCK,DEREK
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: KLEIBER

DATE/TIME OF TEST: 12/02/2018 05:52:47

Diagnostics Check	OK	05:44
Air Blank	0.000	05:45
Control Test	0.078	05:45
Air Blank	0.000	05:46
Subject Sample # 1	0.196	05:48
Breath Volume	2.484	05:48
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample # 2	0.194	05:51
Breath Volume	3.144	05:51
Air Blank	0.000	05:51
Control Test	0.079	05:52
Air Blank	0.000	05:52
Diagnostics Check	OK	05:52

Data Download Date/Time: 12/28/2018 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,DEREK,V
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: LOPEZ,RAYMOND
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/07/2018 02:49:55

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.080	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.013	02:45
Breath Volume	2.367	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.013	02:48
Breath Volume	2.156	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 12/28/2018 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAHEEN,KEITH,B
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: BAUGHN,GUY
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: GAUTHE

DATE/TIME OF TEST: 12/09/2018 14:22:01

Diagnostics Check	OK	14:16
Air Blank	0.000	14:16
Control Test	0.079	14:16
Air Blank	0.000	14:17
Subject Sample # 1	0.093	14:17
Breath Volume	3.445	14:17
Air Blank	0.000	14:18
Air Blank	0.000	14:20
Subject Sample # 2	0.092	14:20
Breath Volume	3.457	14:20
Air Blank	0.000	14:21
Control Test	0.079	14:21
Air Blank	0.000	14:21
Diagnostics Check	OK	14:22

Data Download Date/Time: 12/28/2018 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,DEREK,V
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: TUER,DAVID
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 12/15/2018 02:36:07

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.079	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.148	02:31
Breath Volume	2.457	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.147	02:34
Breath Volume	2.570	02:34
Air Blank	0.000	02:35
Control Test	0.078	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 12/28/2018 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: ESTRADA,PAUL
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: BETTS

DATE/TIME OF TEST: 12/16/2018 02:58:18

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.132	02:53
Breath Volume	3.097	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.132	02:56
Breath Volume	4.335	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 12/28/2018 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: LOPEZ,RAYMOND
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 12/16/2018 03:13:31

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.119	03:08
Breath Volume	1.695	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.115	03:12
Breath Volume	2.917	03:12
Air Blank	0.000	03:12
Control Test	0.079	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 12/28/2018 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: LOPEZ,RAYMOND
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 12/16/2018 03:50:08

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.028	03:45
Breath Volume	2.410	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.027	03:48
Breath Volume	1.277	03:48
Air Blank	0.000	03:49
Control Test	0.079	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 12/28/2018 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,DEREK,V
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: TUER,DAVID
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 11/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 12/16/2018 19:03:11

Diagnostics Check	OK	18:56
Air Blank	0.000	18:57
Control Test	0.079	18:57
Air Blank	0.000	18:58
Subject Sample # 1	0.044	18:58
Breath Volume	2.242	18:58
Air Blank	0.000	18:59
Air Blank	0.000	19:01
Subject Sample # 2	0.041	19:01
Breath Volume	1.511	19:01
Air Blank	0.000	19:02
Control Test	0.079	19:02
Air Blank	0.000	19:03
Diagnostics Check	OK	19:03

Data Download Date/Time: 12/28/2018 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,DEREK,V
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: LOPEZ,RAYMOND
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 12/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/30/2018 04:26:50

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.079	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.239	04:22
Breath Volume	3.167	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample # 2	0.242	04:25
Breath Volume	2.375	04:25
Air Blank	0.000	04:25
Control Test	0.079	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 01/31/2019 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCFERREN, JASON
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/01/2018 02:31:36

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.081	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.133	02:26
Breath Volume	3.273	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.131	02:30
Breath Volume	1.312	02:30
Air Blank	0.000	02:30
Control Test	0.080	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 12/05/2018 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BLACKFORD, MICHAEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: HERNANDEZ COREA

DATE/TIME OF TEST: 12/03/2018 00:46:06

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.082	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.173	00:40
Breath Volume	1.855	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.172	00:44
Breath Volume	1.402	00:44
Air Blank	0.000	00:45
Control Test	0.082	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 12/05/2018 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BODE, JOSHUA
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: MARTINEZMENDEZ

DATE/TIME OF TEST: 12/03/2018 01:41:38

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.081	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.183	01:33
Breath Volume	1.109	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.204	01:37
Breath Volume	2.804	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 3	0.201	01:40
Breath Volume	2.230	01:40
Air Blank	0.000	01:40
Control Test	0.080	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 12/05/2018 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL, CRAIG
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: PANETO RIOS

DATE/TIME OF TEST: 12/05/2018 23:12:38

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.082	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.138	23:08
Breath Volume	3.664	23:08
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.140	23:11
Breath Volume	3.699	23:11
Air Blank	0.000	23:11
Control Test	0.081	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, CALEB, N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA, HERIBERTO
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: GIANNONE

DATE/TIME OF TEST: 12/05/2018 23:41:24

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.082	23:35
Air Blank	0.000	23:36
Subject Sample # 1	0.093	23:37
Breath Volume	1.152	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.094	23:39
Breath Volume	1.179	23:39
Air Blank	0.000	23:40
Control Test	0.081	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KREKEL,ALEC
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/06/2018 03:22:05

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:16	
Control Test	0.082	03:16	
Air Blank	0.000	03:17	
Subject Sample	# 1 0.262	03:17	
Breath Volume	1.578	03:17	
Air Blank	0.000	03:18	
Air Blank	0.000	03:20	
Subject Sample	# 2 0.264	03:20	
Breath Volume	1.730	03:20	
Air Blank	0.000	03:21	
Control Test	0.080	03:21	RFI Detect
Air Blank	0.000	03:22	

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KREKEL,ALEC
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/06/2018 03:30:55

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.082	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.261	03:26
Breath Volume	1.386	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample # 2	0.280	03:29
Breath Volume	1.457	03:29
Air Blank	0.000	03:30
Control Test	0.081	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOONDA, SHUAIB
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: CANDELARIO RIVERA

DATE/TIME OF TEST: 12/06/2018 21:24:33

Diagnostics Check	OK	21:18
Air Blank	0.000	21:18
Control Test	0.082	21:19
Air Blank	0.000	21:19
Subject Sample # 1	0.152	21:20
Breath Volume	2.500	21:20
Air Blank	0.000	21:20
Air Blank	0.000	21:22
Subject Sample # 2	0.155	21:23
Breath Volume	2.703	21:23
Air Blank	0.000	21:23
Control Test	0.081	21:23
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, CALEB, N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA, HERIBERTO
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: SCHALL

DATE/TIME OF TEST: 12/06/2018 22:47:15

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.082	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.182	22:42
Breath Volume	1.968	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.184	22:45
Breath Volume	1.386	22:45
Air Blank	0.000	22:46
Control Test	0.081	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAUNDERS, HUNTER
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: THIEMANN

DATE/TIME OF TEST: 12/07/2018 01:23:57

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:19	
Control Test	0.082	01:19	
Air Blank	0.000	01:20	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:23	
Control Test	0.081	01:23	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, CALEB, N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAL, RAJIV
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 12/07/2018 04:38:01

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.088	04:33
Breath Volume	1.105	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.087	04:36
Breath Volume	1.675	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:38

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONCALVES,GUSTAVO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: MATTOS

DATE/TIME OF TEST: 12/07/2018 23:06:37

Diagnostics Check	OK	23:00
Air Blank	0.000	23:00
Control Test	0.082	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.000	23:02
Breath Volume	4.273	23:02
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.000	23:05
Breath Volume	2.292	23:05
Air Blank	0.000	23:05
Control Test	0.082	23:06
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HERNANDEZ,EDUARDO
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: BONILLA SANCHEZ

DATE/TIME OF TEST: 12/08/2018 21:29:12

Diagnostics Check	OK	21:22
Air Blank	0.000	21:23
Control Test	0.082	21:23
Air Blank	0.000	21:24
Subject Sample # 1	0.187	21:24
Breath Volume	3.441	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.169	21:27
Breath Volume	1.402	21:27
Air Blank	0.000	21:28
Control Test	0.080	21:28
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VEGA,RAFAEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: HAMDAN

DATE/TIME OF TEST: 12/09/2018 03:51:51

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:40	
Control Test	0.082	03:40	
Air Blank	0.000	03:41	
Subject Sample # 1	0.168	03:41	
Breath Volume	2.550	03:41	
Air Blank	0.000	03:42	
Air Blank	0.000	03:44	
Subject Sample # 2	0.152	03:47	Slope Not Level
Breath Volume	2.761	03:47	
Air Blank	0.000	03:47	
Air Blank	0.000	03:49	
Subject Sample # 3	0.148	03:50	
Breath Volume	3.851	03:50	
Air Blank	0.000	03:50	
Control Test	0.081	03:51	
Air Blank	0.000	03:51	
Diagnostics Check	OK	03:51	

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HUDSON,CODY
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 12/09/2018 19:33:14

Diagnostics Check	OK	19:26
Air Blank	0.000	19:27
Control Test	0.081	19:27
Air Blank	0.000	19:28
Subject Sample # 1	0.185	19:28
Breath Volume	1.488	19:28
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.176	19:31
Breath Volume	1.714	19:31
Air Blank	0.000	19:32
Control Test	0.080	19:32
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LARA,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: DUROSIER

DATE/TIME OF TEST: 12/09/2018 22:08:34

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.081	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.172	22:03
Breath Volume	1.328	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.173	22:07
Breath Volume	2.648	22:07
Air Blank	0.000	22:07
Control Test	0.082	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DAVILA,SAMUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 12/10/2018 01:48:43

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:33	
Control Test	0.082	01:34	
Air Blank	0.000	01:34	
Subject Sample	# 1 0.147	01:37	Volume Not Met
Breath Volume	0.386	01:37	
Air Blank	0.000	01:38	
Air Blank	0.000	01:40	
Subject Sample	# 2 0.175	01:41	
Breath Volume	1.601	01:41	
Air Blank	0.000	01:42	
Air Blank	0.000	01:44	
Subject Sample	# 3 0.189	01:47	
Breath Volume	2.921	01:47	
Air Blank	0.000	01:47	
Control Test	0.081	01:48	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FIGUEROA,HECTOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 12/10/2018 04:14:19

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.081	04:09
Air Blank	0.000	04:09
Subject Sample # 1	0.116	04:10
Breath Volume	1.789	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.114	04:12
Breath Volume	1.316	04:12
Air Blank	0.000	04:13
Control Test	0.082	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BACON, LARRY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 12/11/2018 02:47:10

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.230	02:42
Breath Volume	1.457	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.218	02:45
Breath Volume	2.062	02:45
Air Blank	0.000	02:46
Control Test	0.081	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JACKSON, MELISSA
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 12/11/2018 21:19:18

Diagnostics Check	OK	21:04	
Air Blank	0.000	21:04	
Control Test	0.082	21:04	
Air Blank	0.000	21:05	
Subject Sample	# 1 0.182	21:08	
Breath Volume	1.117	21:08	
Air Blank	0.000	21:08	
Air Blank	0.000	21:10	
Subject Sample	# 2 0.178	21:13	Volume Not Met
Breath Volume	0.468	21:13	
Air Blank	0.000	21:14	
Air Blank	0.000	21:16	
Subject Sample	# 3 0.180	21:17	
Breath Volume	1.585	21:17	
Air Blank	0.000	21:18	
Control Test	0.080	21:18	
Air Blank	0.000	21:19	
Diagnostics Check	OK	21:19	

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAWRENCE,NICKOLAS
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 12/12/2018 01:58:02

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.082	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.240	01:52
Breath Volume	2.007	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.232	01:56
Breath Volume	2.082	01:56
Air Blank	0.000	01:57
Control Test	0.082	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:58

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KEITH, SAMUEL
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/13/2018 00:30:43

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:27	
Control Test	0.081	00:28	
Air Blank	0.000	00:28	
Subject Sample	# 1	00:29	Subject Test Refused
Breath Volume	0.000	00:29	
Air Blank	0.000	00:29	
Control Test	0.081	00:30	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PEREZ, JASMIN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: SAUPE

DATE/TIME OF TEST: 12/13/2018 02:06:12

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.081	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.160	02:01
Breath Volume	2.113	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.163	02:04
Breath Volume	3.265	02:04
Air Blank	0.000	02:05
Control Test	0.082	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HUDSON, CODY
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 12/13/2018 03:16:03

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:14	
Control Test	0.082	03:15	
Air Blank		03:15	RFI Detect
Air Blank	0.000	03:16	

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HUDSON, CODY
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 12/13/2018 03:24:04

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.081	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.194	03:19
Breath Volume	2.617	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:21
Subject Sample # 2	0.193	03:22
Breath Volume	2.585	03:22
Air Blank	0.000	03:23
Control Test	0.081	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:24

Data Download Date/Time: 12/13/2018 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: BONOMO JR

DATE/TIME OF TEST: 12/14/2018 03:35:25

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.081	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.176	03:31
Breath Volume	2.386	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.177	03:33
Breath Volume	1.925	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 12/20/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARWOOD,ARMANDO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: SHOEMAKER

DATE/TIME OF TEST: 12/14/2018 04:38:43

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.081	04:33
Air Blank	0.000	04:33
Subject Sample # 1	0.117	04:34
Breath Volume	2.101	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.121	04:37
Breath Volume	1.753	04:37
Air Blank	0.000	04:37
Control Test	0.081	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 12/20/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS, ARMAND, G
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GUEVARA, JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: COWART

DATE/TIME OF TEST: 12/14/2018 08:33:55

Diagnostics Check	OK	08:25
Air Blank	0.000	08:26
Control Test	0.081	08:26
Air Blank	0.000	08:26
Subject Sample # 1	0.168	08:28
Breath Volume	1.128	08:28
Air Blank	0.000	08:28
Air Blank	0.000	08:30
Subject Sample # 2	0.171	08:32
Breath Volume	1.632	08:32
Air Blank	0.000	08:33
Control Test	0.082	08:33
Air Blank	0.000	08:33
Diagnostics Check	OK	08:33

Data Download Date/Time: 12/20/2018 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS, ARMAND, G
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AUSTIN, ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 12/14/2018 12:44:08

Diagnostics Check	OK	12:37
Air Blank	0.000	12:38
Control Test	0.081	12:38
Air Blank	0.000	12:39
Subject Sample # 1	0.000	12:39
Breath Volume	2.777	12:39
Air Blank	0.000	12:40
Air Blank	0.000	12:42
Subject Sample # 2	0.000	12:42
Breath Volume	4.207	12:42
Air Blank	0.000	12:43
Control Test	0.080	12:43
Air Blank	0.000	12:44
Diagnostics Check	OK	12:44

Data Download Date/Time: 12/20/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, CALEB, N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KREKEL, ALEC
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: HARRIMAN

DATE/TIME OF TEST: 12/14/2018 19:53:18

Diagnostics Check	OK	19:46
Air Blank	0.000	19:47
Control Test	0.081	19:47
Air Blank	0.000	19:48
Subject Sample # 1	0.000	19:48
Breath Volume	1.730	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample # 2	0.000	19:51
Breath Volume	1.464	19:51
Air Blank	0.000	19:52
Control Test	0.081	19:52
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

Data Download Date/Time: 12/20/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: JIMENEZ VALLEJO

DATE/TIME OF TEST: 12/15/2018 05:03:58

Diagnostics Check	OK	04:47	
Air Blank	0.000	04:48	
Control Test	0.081	04:48	
Air Blank	0.000	04:49	
Subject Sample	# 1 0.136	04:52	Volume Not Met
Breath Volume	0.335	04:52	
Air Blank	0.000	04:52	
Air Blank	0.000	04:54	
Subject Sample	# 2 0.146	04:57	
Breath Volume	1.765	04:57	
Air Blank	0.000	04:58	
Air Blank	0.000	05:00	
Subject Sample	# 3 0.155	05:02	
Breath Volume	1.148	05:02	
Air Blank	0.000	05:03	
Control Test	0.079	05:03	
Air Blank	0.000	05:03	
Diagnostics Check	OK	05:03	

Data Download Date/Time: 12/20/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES-GONZALEZ,GEAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: LOPEZDELEON

DATE/TIME OF TEST: 12/15/2018 21:54:05

Diagnostics Check	OK	21:45
Air Blank	0.000	21:45
Control Test	0.082	21:46
Air Blank	0.000	21:46
Subject Sample # 1	0.182	21:49
Breath Volume	1.109	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:51
Subject Sample # 2	0.181	21:52
Breath Volume	1.226	21:52
Air Blank	0.000	21:53
Control Test	0.078	21:53
Air Blank	0.000	21:53
Diagnostics Check	OK	21:54

Data Download Date/Time: 12/20/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: EPPS,TIMOTHY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: NAJERA NAJERA

DATE/TIME OF TEST: 12/15/2018 22:39:36

Diagnostics Check	OK	22:32
Air Blank	0.000	22:33
Control Test	0.080	22:33
Air Blank	0.000	22:34
Subject Sample # 1	0.244	22:35
Breath Volume	2.167	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.244	22:38
Breath Volume	1.632	22:38
Air Blank	0.000	22:38
Control Test	0.080	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 12/20/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, CALEB, N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCHOEN, ELLIS
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 12/15/2018 23:59:25

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:44	
Control Test	0.080	23:45	
Air Blank	0.000	23:45	
Subject Sample	# 1 0.169	23:47	No .020 agreement
Breath Volume	1.949	23:47	
Air Blank	0.000	23:48	
Air Blank	0.000	23:50	
Subject Sample	# 2 0.117	23:53	Volume Not Met
Breath Volume	0.511	23:53	
Air Blank	0.000	23:54	
Air Blank	0.000	23:55	
Subject Sample	# 3 0.142	23:57	No .020 agreement
Breath Volume	1.468	23:57	
Air Blank	0.000	23:58	
Control Test	0.079	23:58	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 12/20/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, CALEB, N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCHOEN, ELLIS
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 12/16/2018 00:09:59

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.146	00:05
Breath Volume	1.859	00:05
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.138	00:08
Breath Volume	1.769	00:08
Air Blank	0.000	00:09
Control Test	0.079	00:09
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 12/20/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, CALEB, N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: COLE, KYLE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/16/2018 04:21:01

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.081	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.184	04:15
Breath Volume	1.648	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.186	04:19
Breath Volume	1.843	04:19
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:21

Data Download Date/Time: 12/20/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MENDOZA,ANGEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 12/16/2018 05:31:52

Diagnostics Check	OK	05:25
Air Blank	0.000	05:25
Control Test	0.081	05:25
Air Blank	0.000	05:26
Subject Sample # 1	0.037	05:27
Breath Volume	2.894	05:27
Air Blank	0.000	05:28
Air Blank	0.000	05:29
Subject Sample # 2	0.031	05:30
Breath Volume	1.601	05:30
Air Blank	0.000	05:30
Control Test	0.081	05:31
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 12/20/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AGOSTO, ALBA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: COCHRAN III

DATE/TIME OF TEST: 12/16/2018 21:55:16

Diagnostics Check	OK	21:48
Air Blank	0.000	21:48
Control Test	0.082	21:49
Air Blank	0.000	21:49
Subject Sample # 1	0.202	21:50
Breath Volume	1.281	21:50
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.195	21:53
Breath Volume	1.355	21:53
Air Blank	0.000	21:54
Control Test	0.081	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 12/20/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JUNCO,JONATHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: MAIGNAN

DATE/TIME OF TEST: 12/17/2018 01:14:53

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.081	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.246	01:10
Breath Volume	2.273	01:10
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.235	01:13
Breath Volume	1.312	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 12/20/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RUSSAK,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: ZELAYA

DATE/TIME OF TEST: 12/17/2018 06:56:53

Diagnostics Check	OK	06:49
Air Blank	0.000	06:50
Control Test	0.081	06:50
Air Blank	0.000	06:50
Subject Sample # 1	0.106	06:51
Breath Volume	1.308	06:51
Air Blank	0.000	06:52
Air Blank	0.000	06:54
Subject Sample # 2	0.110	06:55
Breath Volume	1.660	06:55
Air Blank	0.000	06:55
Control Test	0.082	06:56
Air Blank	0.000	06:56
Diagnostics Check	OK	06:56

Data Download Date/Time: 12/20/2018 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPEER, ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: MERRYMAN-SCOTT

DATE/TIME OF TEST: 12/18/2018 21:32:27

Diagnostics Check	OK	21:26
Air Blank	0.000	21:26
Control Test	0.081	21:27
Air Blank	0.000	21:27
Subject Sample # 1	0.127	21:28
Breath Volume	2.105	21:28
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.129	21:30
Breath Volume	3.183	21:30
Air Blank	0.000	21:31
Control Test	0.081	21:31
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 12/20/2018 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KELEMEN,JANOS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/18/2018 23:46:55

Diagnostics Check	OK	23:38
Air Blank	0.000	23:38
Control Test	0.082	23:39
Air Blank	0.000	23:39
Subject Sample # 1	0.259	23:40
Breath Volume	1.449	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.267	23:45
Breath Volume	1.960	23:45
Air Blank	0.000	23:45
Control Test	0.081	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 12/20/2018 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LABRANCHE,MINDY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: BROSSEAU

DATE/TIME OF TEST: 12/19/2018 01:59:07

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.081	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.000	01:54
Breath Volume	2.144	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.000	01:57
Breath Volume	1.226	01:57
Air Blank	0.000	01:58
Control Test	0.082	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 12/20/2018 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL,TALMON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: CHEVALIER

DATE/TIME OF TEST: 12/19/2018 04:39:08

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.081	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.161	04:34
Breath Volume	2.230	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.156	04:37
Breath Volume	2.375	04:37
Air Blank	0.000	04:38
Control Test	0.081	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 12/20/2018 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: EDMONDS,STEVEN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: RUSS

DATE/TIME OF TEST: 12/19/2018 04:51:05

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.081	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.201	04:46
Breath Volume	1.714	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.197	04:49
Breath Volume	1.230	04:49
Air Blank	0.000	04:50
Control Test	0.081	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 12/20/2018 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, CALEB, N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SMITH, JULIA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MCMAHAN

DATE/TIME OF TEST: 12/19/2018 23:44:27

Diagnostics Check	OK	23:38
Air Blank	0.000	23:38
Control Test	0.081	23:39
Air Blank	0.000	23:39
Subject Sample # 1	0.173	23:40
Breath Volume	3.367	23:40
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.176	23:42
Breath Volume	2.480	23:42
Air Blank	0.000	23:43
Control Test	0.081	23:43
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 12/20/2018 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 12/20/2018 01:08:17

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.081	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.166	01:03
Breath Volume	1.734	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.166	01:06
Breath Volume	1.968	01:06
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 12/20/2018 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, CALEB, N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS, BRIAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 12/20/2018 23:38:27

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:27	
Control Test	0.082	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:31	No Sample Provided
Breath Volume	0.000	23:31	
Air Blank	0.000	23:31	
Air Blank	0.000	23:33	
Subject Sample	# 2	23:37	No Sample Provided
Breath Volume	0.000	23:37	
Air Blank	0.000	23:37	
Control Test	0.083	23:37	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 12/27/2018 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SZESNAT,KENNY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: RIALS

DATE/TIME OF TEST: 12/21/2018 04:17:07

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.081	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.155	04:12
Breath Volume	2.410	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.158	04:15
Breath Volume	2.625	04:15
Air Blank	0.000	04:16
Control Test	0.081	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 12/27/2018 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONZALEZ,CHELSEA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 12/22/2018 02:38:47

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.082	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.204	02:34
Breath Volume	3.546	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.207	02:37
Breath Volume	3.816	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 12/27/2018 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONCALVES,GUSTAVO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 12/22/2018 21:50:47

Diagnostics Check	OK	21:44
Air Blank	0.000	21:45
Control Test	0.082	21:45
Air Blank	0.000	21:45
Subject Sample # 1	0.162	21:46
Breath Volume	2.070	21:46
Air Blank	0.000	21:47
Air Blank	0.000	21:48
Subject Sample # 2	0.164	21:49
Breath Volume	2.187	21:49
Air Blank	0.000	21:49
Control Test	0.081	21:50
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 12/27/2018 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL,TALMON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 12/23/2018 00:56:20

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.081	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.359	00:48
Breath Volume	3.113	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.336	00:51
Breath Volume	2.742	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 3	0.344	00:54
Breath Volume	3.230	00:54
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 12/27/2018 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAYLER,ROBERT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: LI

DATE/TIME OF TEST: 12/23/2018 02:57:25

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.082	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.110	02:53
Breath Volume	3.433	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.108	02:55
Breath Volume	3.527	02:55
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 12/27/2018 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DAVILA,SAMUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: PALMERPEREZ

DATE/TIME OF TEST: 12/23/2018 04:03:24

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.082	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.115	03:59
Breath Volume	2.378	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.114	04:01
Breath Volume	2.164	04:01
Air Blank	0.000	04:02
Control Test	0.081	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 12/27/2018 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAMBY,CHRISTIAN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: CRISPEN

DATE/TIME OF TEST: 12/23/2018 04:38:59

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.081	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.121	04:34
Breath Volume	3.175	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.115	04:37
Breath Volume	1.406	04:37
Air Blank	0.000	04:38
Control Test	0.081	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 12/27/2018 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DAVILA,SAMUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/24/2018 00:27:14

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:16	
Control Test	0.082	00:16	
Air Blank	0.000	00:16	
Subject Sample	# 1 0.194	00:20	Volume Not Met
Breath Volume	0.066	00:20	
Air Blank	0.000	00:20	
Air Blank	0.000	00:22	
Subject Sample	# 2 0.199	00:25	Volume Not Met
Breath Volume	0.097	00:25	
Air Blank	0.000	00:26	
Control Test	0.080	00:26	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 12/27/2018 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DAVILA,SAMUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/24/2018 00:40:20

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:28	
Control Test	0.080	00:28	
Air Blank	0.000	00:29	
Subject Sample	# 1 0.212	00:30	No .020 agreement
Breath Volume	1.316	00:30	
Air Blank	0.000	00:30	
Air Blank	0.000	00:32	
Subject Sample	# 2 0.194	00:35	Volume Not Met
Breath Volume	0.050	00:35	
Air Blank	0.000	00:36	
Air Blank	0.000	00:38	
Subject Sample	# 3 0.189	00:38	No .020 agreement
Breath Volume	1.179	00:38	
Air Blank	0.000	00:39	
Control Test	0.080	00:39	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

Data Download Date/Time: 12/27/2018 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DAVILA,SAMUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/24/2018 00:55:55

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:41	
Control Test	0.081	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1 0.184	00:45	Volume Not Met
Breath Volume	0.843	00:45	
Air Blank	0.000	00:46	
Air Blank	0.000	00:48	
Subject Sample	# 2 0.204	00:48	
Breath Volume	1.437	00:48	
Air Blank	0.000	00:49	
Air Blank	0.000	00:51	
Subject Sample	# 3 0.195	00:54	
Breath Volume	1.972	00:54	
Air Blank	0.000	00:55	
Control Test	0.080	00:55	
Air Blank	0.000	00:55	
Diagnostics Check	OK	00:55	

Data Download Date/Time: 12/27/2018 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROCCHILD,MIGUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: COLLADO

DATE/TIME OF TEST: 12/24/2018 06:30:02

Diagnostics Check	OK	06:17	
Air Blank	0.000	06:17	
Control Test	0.082	06:18	
Air Blank	0.000	06:18	
Subject Sample	# 1 0.139	06:19	
Breath Volume	1.347	06:19	
Air Blank	0.000	06:20	
Air Blank	0.000	06:22	
Subject Sample	# 2 0.129	06:25	Slope Not Level
Breath Volume	1.945	06:25	
Air Blank	0.000	06:25	
Air Blank	0.000	06:27	
Subject Sample	# 3 0.129	06:28	
Breath Volume	1.343	06:28	
Air Blank	0.000	06:29	
Control Test	0.080	06:29	
Air Blank	0.000	06:29	
Diagnostics Check	OK	06:30	

Data Download Date/Time: 12/27/2018 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,HERIBERTO
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: TERRUSA

DATE/TIME OF TEST: 12/25/2018 02:03:29

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:01	
Control Test	0.082	02:01	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Control Test	0.082	02:02	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 12/27/2018 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JUNCO,JONATHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WODDARD

DATE/TIME OF TEST: 12/25/2018 03:18:39

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.082	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.157	03:14
Breath Volume	2.105	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.152	03:17
Breath Volume	1.792	03:17
Air Blank	0.000	03:17
Control Test	0.081	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 12/27/2018 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCLEAN,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: VILACHA AYESTARAN

DATE/TIME OF TEST: 12/25/2018 17:53:04

Diagnostics Check	OK	17:45
Air Blank	0.000	17:46
Control Test	0.082	17:46
Air Blank	0.000	17:47
Subject Sample # 1	0.190	17:48
Breath Volume	1.753	17:48
Air Blank	0.000	17:48
Air Blank	0.000	17:50
Subject Sample # 2	0.179	17:51
Breath Volume	2.757	17:51
Air Blank	0.000	17:52
Control Test	0.081	17:52
Air Blank	0.000	17:52
Diagnostics Check	OK	17:53

Data Download Date/Time: 12/27/2018 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: COLLINS,MARVIN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MARTON

DATE/TIME OF TEST: 12/26/2018 02:24:01

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.082	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.100	02:19
Breath Volume	3.082	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.099	02:22
Breath Volume	3.101	02:22
Air Blank	0.000	02:23
Control Test	0.082	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:24

Data Download Date/Time: 12/27/2018 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,J
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: GODELL

DATE/TIME OF TEST: 12/27/2018 20:22:47

Diagnostics Check	OK	20:16
Air Blank	0.000	20:17
Control Test	0.081	20:17
Air Blank	0.000	20:17
Subject Sample # 1	0.167	20:18
Breath Volume	2.804	20:18
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.168	20:21
Breath Volume	2.945	20:21
Air Blank	0.000	20:21
Control Test	0.081	20:22
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 01/03/2019 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, CALEB, N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAUNDERS, HUNTER
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: ARDENE

DATE/TIME OF TEST: 12/27/2018 23:15:10

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.081	23:10
Air Blank	0.000	23:10
Subject Sample # 1	0.000	23:10
Breath Volume	1.324	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.000	23:13
Breath Volume	1.179	23:13
Air Blank	0.000	23:14
Control Test	0.081	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 01/03/2019 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 12/28/2018 04:02:55

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.082	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.164	03:58
Breath Volume	2.351	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:00
Subject Sample # 2	0.166	04:01
Breath Volume	3.542	04:01
Air Blank	0.000	04:02
Control Test	0.081	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 01/03/2019 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CHILDS

DATE/TIME OF TEST: 12/28/2018 22:48:47

Diagnostics Check	OK	22:42
Air Blank	0.000	22:43
Control Test	0.081	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.131	22:44
Breath Volume	1.183	22:44
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.131	22:47
Breath Volume	1.484	22:47
Air Blank	0.000	22:47
Control Test	0.081	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 01/03/2019 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 12/29/2018 02:46:10

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.082	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.240	02:41
Breath Volume	1.730	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.234	02:44
Breath Volume	2.453	02:44
Air Blank	0.000	02:45
Control Test	0.081	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 01/03/2019 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, CALEB, N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KUZMA, TIMOTHY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: CRAME

DATE/TIME OF TEST: 12/29/2018 05:26:00

Diagnostics Check	OK	05:17
Air Blank	0.000	05:17
Control Test	0.081	05:18
Air Blank	0.000	05:18
Subject Sample # 1	0.180	05:21
Breath Volume	1.453	05:21
Air Blank	0.000	05:21
Air Blank	0.000	05:23
Subject Sample # 2	0.192	05:24
Breath Volume	1.710	05:24
Air Blank	0.000	05:25
Control Test	0.081	05:25
Air Blank	0.000	05:25
Diagnostics Check	OK	05:26

Data Download Date/Time: 01/03/2019 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, CALEB, N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPEER, ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: PENN

DATE/TIME OF TEST: 12/30/2018 04:24:29

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.081	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.189	04:19
Breath Volume	4.039	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.189	04:22
Breath Volume	3.968	04:22
Air Blank	0.000	04:23
Control Test	0.081	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 01/03/2019 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 12/30/2018 05:04:35

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.081	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.160	04:59
Breath Volume	1.238	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.168	05:03
Breath Volume	2.152	05:03
Air Blank	0.000	05:03
Control Test	0.081	05:03
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 01/03/2019 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WOODCOX,JONATHAN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 12/30/2018 13:46:40

Diagnostics Check	OK	13:37	
Air Blank	0.000	13:38	
Control Test	0.081	13:38	
Air Blank	0.000	13:39	
Subject Sample	# 1	13:42	No Sample Provided
Breath Volume	0.000	13:42	
Air Blank	0.000	13:42	
Air Blank	0.000	13:44	
Subject Sample	# 2	13:45	Subject Test Refused
Breath Volume	0.000	13:45	
Air Blank	0.000	13:45	
Control Test	0.080	13:46	
Air Blank	0.000	13:46	
Diagnostics Check	OK	13:46	

Data Download Date/Time: 01/03/2019 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KREKEL,ALEC
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CUYLER

DATE/TIME OF TEST: 12/31/2018 03:15:00

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:12	
Control Test	0.082	03:12	
Air Blank	0.000	03:13	
Subject Sample # 1	0.000	03:13	
Breath Volume	3.402	03:13	
Air Blank		03:14	RFI Detect
Air Blank		03:14	RFI Detect
Air Blank	0.000	03:15	

Data Download Date/Time: 01/03/2019 13:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCHOEN,ELLIS
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 12/01/2018 02:21:25

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:19	
Control Test	0.081	02:19	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:20	
Control Test	0.081	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 12/05/2018 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, CALEB, N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MENDOZA, ANGEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/01/2018 04:04:22

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.100	03:59
Breath Volume	1.734	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.096	04:02
Breath Volume	2.320	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 12/05/2018 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, CALEB, N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES, NICOLE
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: COMEAU

DATE/TIME OF TEST: 12/02/2018 02:51:57

Diagnostics Check	OK	02:46	
Air Blank	0.000	02:46	
Control Test	0.081	02:47	
Air Blank	0.000	02:47	
Subject Sample # 1	0.164	02:50	
Breath Volume	1.238	02:50	
Air Blank		02:50	Improper Sample
Air Blank	0.000	02:51	
Control Test	0.082	02:51	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 12/05/2018 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, CALEB, N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES, NICOLE
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: COMEAU

DATE/TIME OF TEST: 12/02/2018 03:07:29

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:53	
Control Test	0.080	02:54	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:57	No Sample Provided
Breath Volume	0.000	02:57	
Air Blank	0.000	02:58	
Air Blank	0.000	03:00	
Subject Sample	# 2	0.169	03:01
Breath Volume	1.496	03:01	
Air Blank	0.000	03:02	
Air Blank	0.000	03:04	
Subject Sample	# 3	0.177	03:05
Breath Volume	1.871	03:05	
Air Blank	0.000	03:06	
Control Test	0.080	03:06	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 12/05/2018 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RICHARDS,ZACHARY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 12/02/2018 04:49:21

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:46	
Control Test	0.081	04:47	
Air Blank	0.000	04:47	
Subject Sample	# 1	04:47	Subject Test Refused
Breath Volume	0.000	04:47	
Air Blank	0.000	04:48	
Control Test	0.082	04:48	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time: 12/05/2018 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LISTORT,ANTHONY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 12/02/2018 08:37:13

Diagnostics Check	OK	08:28
Air Blank	0.000	08:29
Control Test	0.081	08:29
Air Blank	0.000	08:30
Subject Sample # 1	0.270	08:32
Breath Volume	2.347	08:32
Air Blank	0.000	08:32
Air Blank	0.000	08:34
Subject Sample # 2	0.257	08:35
Breath Volume	1.273	08:35
Air Blank	0.000	08:36
Control Test	0.079	08:36
Air Blank	0.000	08:37
Diagnostics Check	OK	08:37

Data Download Date/Time: 12/05/2018 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LACENTRA,DONALD
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:28

SUBJECT'S LAST NAME: WHITE-SPUNNER

DATE/TIME OF TEST: 12/02/2018 15:59:51

Diagnostics Check	OK	15:53
Air Blank	0.000	15:53
Control Test	0.082	15:54
Air Blank	0.000	15:54
Subject Sample # 1	0.250	15:55
Breath Volume	1.464	15:55
Air Blank	0.000	15:55
Air Blank	0.000	15:57
Subject Sample # 2	0.248	15:58
Breath Volume	1.246	15:58
Air Blank	0.000	15:58
Control Test	0.080	15:59
Air Blank	0.000	15:59
Diagnostics Check	OK	15:59

Data Download Date/Time: 12/05/2018 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ARNOLDY,BENJAMIN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: LUCASSANCHEZ

DATE/TIME OF TEST: 12/03/2018 00:53:42

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.082	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.194	00:49
Breath Volume	3.652	00:49
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.187	00:52
Breath Volume	4.445	00:52
Air Blank	0.000	00:52
Control Test	0.081	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 12/05/2018 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: POTTORFF

DATE/TIME OF TEST: 12/03/2018 01:54:48

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.081	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.187	01:50
Breath Volume	2.308	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:52
Subject Sample # 2	0.187	01:53
Breath Volume	2.464	01:53
Air Blank	0.000	01:53
Control Test	0.080	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 12/05/2018 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PEERS,MICHAEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: SHUKLA

DATE/TIME OF TEST: 12/04/2018 02:53:25

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.221	02:45
Breath Volume	2.066	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:47
Subject Sample # 2	0.197	02:48
Breath Volume	1.953	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 3	0.212	02:51
Breath Volume	1.222	02:51
Air Blank	0.000	02:52
Control Test	0.081	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 12/05/2018 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAWRENCE,NICKOLAS
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 11/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: DEJESUS

DATE/TIME OF TEST: 12/05/2018 04:38:46

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.217	04:34
Breath Volume	2.265	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.212	04:37
Breath Volume	2.035	04:37
Air Blank	0.000	04:37
Control Test	0.080	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 12/05/2018 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TEW, JEREMY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: RIVERA OQUENDO

DATE/TIME OF TEST: 12/07/2018 04:49:36

Diagnostics Check	OK	04:42
Air Blank	0.000	04:43
Control Test	0.082	04:43
Air Blank	0.000	04:44
Subject Sample # 1	0.127	04:44
Breath Volume	1.855	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.121	04:48
Breath Volume	1.566	04:48
Air Blank	0.000	04:48
Control Test	0.082	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 12/13/2018 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPEER, ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/08/2018 00:36:01

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.082	00:30
Air Blank	0.000	00:30
Subject Sample # 1	0.105	00:31
Breath Volume	1.675	00:31
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.111	00:34
Breath Volume	2.339	00:34
Air Blank	0.000	00:35
Control Test	0.081	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:36

Data Download Date/Time: 12/13/2018 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CADY, PETER
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: GODINEZ

DATE/TIME OF TEST: 12/08/2018 03:01:28

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.261	02:57
Breath Volume	1.894	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.264	02:59
Breath Volume	1.812	02:59
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 12/13/2018 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPEER, ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 12/09/2018 04:06:31

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.081	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.177	04:02
Breath Volume	2.074	04:02
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.173	04:04
Breath Volume	1.777	04:04
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 12/13/2018 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILKES, DAVID
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: PILE

DATE/TIME OF TEST: 12/12/2018 23:52:05

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.081	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.209	23:46
Breath Volume	1.843	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.216	23:50
Breath Volume	1.691	23:50
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 12/13/2018 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FRANCIS, RICKEY
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: KINNEY

DATE/TIME OF TEST: 12/13/2018 02:44:20

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.081	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.173	02:39
Breath Volume	1.699	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.169	02:42
Breath Volume	1.562	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 12/13/2018 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRICE, FORREST
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 12/13/2018 04:10:13

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.081	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 1 0.154	04:01	Volume Not Met
Breath Volume	0.351	04:01	
Air Blank	0.000	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 2 0.157	04:04	
Breath Volume	2.367	04:04	
Air Blank	0.000	04:05	
Air Blank	0.000	04:06	
Subject Sample	# 3 0.160	04:08	
Breath Volume	2.082	04:08	
Air Blank	0.000	04:09	
Control Test	0.080	04:09	
Air Blank	0.000	04:10	
Diagnostics Check	OK	04:10	

Data Download Date/Time: 12/13/2018 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HANSTINE, AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: WILKINSON

DATE/TIME OF TEST: 12/13/2018 23:43:45

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.081	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.127	23:39
Breath Volume	2.164	23:39
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.122	23:42
Breath Volume	2.832	23:42
Air Blank	0.000	23:42
Control Test	0.081	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 12/20/2018 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ, ERNESTO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: COLOMBO

DATE/TIME OF TEST: 12/14/2018 04:36:01

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.081	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.197	04:31
Breath Volume	2.371	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:33
Subject Sample # 2	0.197	04:34
Breath Volume	1.433	04:34
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:36

Data Download Date/Time: 12/20/2018 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KREKEL, ALEC
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: NERISPIZARRO

DATE/TIME OF TEST: 12/15/2018 03:38:02

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.081	03:26	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:29	No Sample Provided
Breath Volume	0.000	03:29	
Air Blank	0.000	03:30	
Air Blank	0.000	03:32	
Subject Sample	# 2	0.158	03:33
Breath Volume	1.660	03:33	
Air Blank	0.000	03:33	
Air Blank	0.000	03:35	
Subject Sample	# 3	0.152	03:36
Breath Volume	1.570	03:36	
Air Blank	0.000	03:37	
Control Test	0.082	03:37	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 12/20/2018 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JACKSON, MELISSA
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: PADILLACRUZ

DATE/TIME OF TEST: 12/16/2018 01:45:40

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.081	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.108	01:41
Breath Volume	2.156	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.108	01:44
Breath Volume	1.964	01:44
Air Blank	0.000	01:44
Control Test	0.081	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 12/20/2018 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MILLS, RESHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: SANCHEZ FLORES

DATE/TIME OF TEST: 12/16/2018 04:53:16

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.081	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.210	04:48
Breath Volume	2.156	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:50
Subject Sample # 2	0.215	04:51
Breath Volume	2.441	04:51
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 12/20/2018 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN, ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: MARLOWE

DATE/TIME OF TEST: 12/20/2018 03:18:29

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.081	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.184	03:13
Breath Volume	1.683	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.180	03:16
Breath Volume	3.621	03:16
Air Blank	0.000	03:17
Control Test	0.081	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 12/20/2018 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KREKEL, ALEC
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: MUNGAL

DATE/TIME OF TEST: 12/29/2018 03:18:17

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:07	
Control Test	0.081	03:07	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:11	No Sample Provided
Breath Volume	0.000	03:11	
Air Blank	0.000	03:11	
Air Blank	0.000	03:13	
Subject Sample	# 2	03:16	No Sample Provided
Breath Volume	0.000	03:16	
Air Blank	0.000	03:17	
Control Test	0.081	03:17	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 01/03/2019 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL, CRAIG
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/29/2018 04:53:19

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.081	04:47
Air Blank	0.000	04:48
Subject Sample # 1	0.153	04:48
Breath Volume	1.132	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.150	04:51
Breath Volume	1.222	04:51
Air Blank	0.000	04:52
Control Test	0.081	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 01/03/2019 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO, FABIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13917080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: KISHORE

DATE/TIME OF TEST: 12/30/2018 04:29:49

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.081	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.268	04:25
Breath Volume	1.910	04:25
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.275	04:28
Breath Volume	2.289	04:28
Air Blank	0.000	04:28
Control Test	0.080	04:29
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 01/03/2019 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,TRAVIS
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 12/01/2018 22:09:05

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.080	22:03
Air Blank	0.000	22:03
Subject Sample # 1	0.136	22:04
Breath Volume	2.312	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.136	22:07
Breath Volume	2.644	22:07
Air Blank	0.000	22:08
Control Test	0.080	22:08
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 12/20/2018 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,JERRY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 12/02/2018 18:45:56

Diagnostics Check	OK	18:36	
Air Blank	0.000	18:36	
Control Test	0.080	18:36	
Air Blank	0.000	18:37	
Subject Sample	# 1	18:40	No Sample Provided
Breath Volume	0.000	18:40	
Air Blank	0.000	18:40	
Air Blank	0.000	18:42	
Subject Sample	# 2	18:44	Subject Test Refused
Breath Volume	0.000	18:44	
Air Blank	0.000	18:45	
Control Test	0.080	18:45	
Air Blank	0.000	18:45	
Diagnostics Check	OK	18:45	

Data Download Date/Time: 12/20/2018 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,TRAVIS
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: KULT

DATE/TIME OF TEST: 12/16/2018 03:21:44

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.081	03:19	
Air Blank	0.000	03:19	
Subject Sample	# 1 0.060	03:20	
Breath Volume	3.644	03:20	
Air Blank		03:20	Improper Sample
Air Blank	0.000	03:20	
Control Test	0.081	03:21	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 12/20/2018 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD',TRAVIS
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: KULT

DATE/TIME OF TEST: 12/16/2018 03:30:32

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.058	03:26
Breath Volume	2.855	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.054	03:29
Breath Volume	3.296	03:29
Air Blank	0.000	03:29
Control Test	0.079	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 12/20/2018 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SWIM,GREG
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 12/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: CULLIPHER

DATE/TIME OF TEST: 12/20/2018 04:29:25

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.081	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.147	04:23
Breath Volume	1.296	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.167	04:27
Breath Volume	2.480	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 12/20/2018 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIESLINSKI,JOSEPH,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: ARENCIBIA

DATE/TIME OF TEST: 12/01/2018 04:30:33

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.079	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.146	04:26
Breath Volume	3.058	04:26
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.134	04:29
Breath Volume	2.199	04:29
Air Blank	0.000	04:29
Control Test	0.077	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 12/12/2018 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FESSEL,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SANCHO,BRANDON
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:29

SUBJECT'S LAST NAME: MALDONADO DE LA CRUZ

DATE/TIME OF TEST: 12/03/2018 15:55:37

Diagnostics Check	OK	15:49
Air Blank	0.000	15:50
Control Test	0.080	15:50
Air Blank	0.000	15:50
Subject Sample # 1	0.147	15:51
Breath Volume	1.992	15:51
Air Blank	0.000	15:51
Air Blank	0.000	15:53
Subject Sample # 2	0.152	15:54
Breath Volume	2.304	15:54
Air Blank	0.000	15:54
Control Test	0.080	15:55
Air Blank	0.000	15:55
Diagnostics Check	OK	15:55

Data Download Date/Time: 12/12/2018 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFY,J,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DUFFY,J
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26316080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 12/04/2018 12:16:32

Diagnostics Check	OK	12:10
Air Blank	0.000	12:10
Control Test	0.079	12:11
Air Blank	0.000	12:11
Subject Sample # 1	0.000	12:12
Breath Volume	2.453	12:12
Air Blank	0.000	12:12
Air Blank	0.000	12:14
Subject Sample # 2	0.000	12:15
Breath Volume	1.640	12:15
Air Blank	0.000	12:15
Control Test	0.079	12:15
Air Blank	0.000	12:16
Diagnostics Check	OK	12:16

Data Download Date/Time: 12/12/2018 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOOS, SAVANNAH, K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS, DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 12/04/2018 20:36:20

Diagnostics Check	OK	20:32	
Air Blank	0.000	20:33	
Control Test	0.080	20:33	
Air Blank	0.000	20:33	
Subject Sample	# 1	20:34	Subject Test Refused
Breath Volume	0.000	20:34	
Air Blank	0.000	20:35	
Control Test	0.079	20:35	
Air Blank	0.000	20:36	
Diagnostics Check	OK	20:36	

Data Download Date/Time: 12/12/2018 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOOS, SAVANNAH, K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS, DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 12/04/2018 20:44:33

Diagnostics Check	OK	20:38
Air Blank	0.000	20:38
Control Test	0.079	20:38
Air Blank	0.000	20:39
Subject Sample # 1	0.207	20:40
Breath Volume	1.496	20:40
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.195	20:43
Breath Volume	1.417	20:43
Air Blank	0.000	20:43
Control Test	0.079	20:43
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 12/12/2018 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOOS, SAVANNAH, K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SHIPPY, DYLAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 12/05/2018 02:16:50

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.079	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.201	02:12
Breath Volume	1.695	02:12
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.205	02:15
Breath Volume	2.273	02:15
Air Blank	0.000	02:15
Control Test	0.079	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 12/12/2018 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: WONG CANOVAS

DATE/TIME OF TEST: 12/06/2018 00:12:55

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.165	00:08
Breath Volume	3.019	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.163	00:11
Breath Volume	2.968	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 12/12/2018 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUSKEVICH,CURT,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS,D
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: FREDERICK

DATE/TIME OF TEST: 12/09/2018 00:59:18

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.227	00:54
Breath Volume	2.320	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.226	00:57
Breath Volume	2.558	00:57
Air Blank	0.000	00:58
Control Test	0.079	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 12/12/2018 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOOS, SAVANNAH, K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS, DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/10/2018 00:22:30

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:18	
Control Test	0.080	00:19	
Air Blank	0.000	00:19	
Subject Sample	# 1	00:21	Subject Test Refused
Breath Volume	0.000	00:21	
Air Blank	0.000	00:21	
Control Test	0.081	00:21	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 12/12/2018 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOOS, SAVANNAH, K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SCARLATO, MICHAEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HAYDEN

DATE/TIME OF TEST: 12/10/2018 03:16:14

Diagnostics Check	OK	03:15	
Air Blank		03:15	RFI Detect
Air Blank	0.000	03:16	

Data Download Date/Time: 12/12/2018 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOOS, SAVANNAH, K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SCARLATO, MICHAEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HAYDEN

DATE/TIME OF TEST: 12/10/2018 03:30:37

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.079	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:23	No Sample Provided
Breath Volume	0.000	03:23	
Air Blank	0.000	03:23	
Air Blank	0.000	03:25	
Subject Sample	# 2	03:29	Volume Not Met
Breath Volume	1.394	03:29	
Air Blank	0.000	03:29	
Control Test	0.080	03:30	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 12/12/2018 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOOS, SAVANNAH, K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SCARLATO, MICHAEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HAYDEN

DATE/TIME OF TEST: 12/10/2018 03:40:02

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.127	03:35
Breath Volume	1.601	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.127	03:38
Breath Volume	1.800	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:40

Data Download Date/Time: 12/12/2018 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BARKLEY,J
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: SYS CHECK
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BARKLEY

DATE/TIME OF TEST: 12/13/2018 22:48:09

Diagnostics Check	OK	22:42
Air Blank	0.000	22:42
Control Test	0.079	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.000	22:44
Breath Volume	3.046	22:44
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.000	22:46
Breath Volume	3.207	22:46
Air Blank	0.000	22:47
Control Test	0.078	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 01/09/2019 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GALLO,ANDREW
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:36

SUBJECT'S LAST NAME: VANPOELVOORDE

DATE/TIME OF TEST: 12/15/2018 13:07:11

Diagnostics Check	OK	13:00
Air Blank	0.000	13:01
Control Test	0.080	13:01
Air Blank	0.000	13:02
Subject Sample # 1	0.228	13:02
Breath Volume	3.988	13:02
Air Blank	0.000	13:03
Air Blank	0.000	13:05
Subject Sample # 2	0.220	13:05
Breath Volume	2.980	13:05
Air Blank	0.000	13:06
Control Test	0.078	13:06
Air Blank	0.000	13:07
Diagnostics Check	OK	13:07

Data Download Date/Time: 01/09/2019 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOOS,SAVANNAH,K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 12/18/2018 01:07:05

Diagnostics Check	OK	01:06	
Air Blank		01:06	RFI Detect
Air Blank	0.000	01:07	

Data Download Date/Time: 01/09/2019 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOOS,SAVANNAH,K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 12/18/2018 01:15:34

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.038	01:11
Breath Volume	2.847	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.036	01:14
Breath Volume	2.964	01:14
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 01/09/2019 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,CHARLES,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MILLS,CHARLES
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: WHARTON

DATE/TIME OF TEST: 12/22/2018 03:41:08

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.109	03:36
Breath Volume	3.335	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.110	03:39
Breath Volume	3.476	03:39
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 01/09/2019 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BARKLEY,JEFFREY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 12/23/2018 02:14:47

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.163	02:07
Breath Volume	1.523	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.198	02:10
Breath Volume	3.750	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 3	0.199	02:13
Breath Volume	3.277	02:13
Air Blank	0.000	02:13
Control Test	0.078	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 01/09/2019 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIESLINSKI,JOSEPH,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: KOMOLL

DATE/TIME OF TEST: 12/23/2018 03:14:42

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.079	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.189	03:10
Breath Volume	2.503	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.183	03:13
Breath Volume	2.414	03:13
Air Blank	0.000	03:13
Control Test	0.078	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 01/09/2019 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: ALBRIGHT

DATE/TIME OF TEST: 12/28/2018 22:40:06

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.308	22:35
Breath Volume	2.410	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.308	22:38
Breath Volume	2.464	22:38
Air Blank	0.000	22:39
Control Test	0.079	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 01/09/2019 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ENGH,WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/30/2018 03:59:55

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.124	03:54
Breath Volume	1.656	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.128	03:58
Breath Volume	2.406	03:58
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 01/09/2019 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEARN,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FILLAK,MICHAEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 12/30/2018 19:36:48

Diagnostics Check	OK	19:30
Air Blank	0.000	19:31
Control Test	0.079	19:31
Air Blank	0.000	19:32
Subject Sample # 1	0.194	19:32
Breath Volume	2.691	19:32
Air Blank	0.000	19:33
Air Blank	0.000	19:34
Subject Sample # 2	0.184	19:35
Breath Volume	3.289	19:35
Air Blank	0.000	19:35
Control Test	0.079	19:36
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

Data Download Date/Time: 01/09/2019 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOOS, SAVANNAH, K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SUPER, M
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: ZINDERMAN

DATE/TIME OF TEST: 12/31/2018 12:02:22

Diagnostics Check	OK	11:56
Air Blank	0.000	11:56
Control Test	0.079	11:57
Air Blank	0.000	11:57
Subject Sample # 1	0.000	11:57
Breath Volume	1.421	11:57
Air Blank	0.000	11:58
Air Blank	0.000	12:00
Subject Sample # 2	0.000	12:01
Breath Volume	1.175	12:01
Air Blank	0.000	12:01
Control Test	0.079	12:01
Air Blank	0.000	12:02
Diagnostics Check	OK	12:02

Data Download Date/Time: 01/09/2019 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADLEY,KELLY,L
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: KOZUB

DATE/TIME OF TEST: 12/09/2018 01:27:01

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.076	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.184	01:22
Breath Volume	2.304	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.170	01:25
Breath Volume	1.582	01:25
Air Blank	0.000	01:26
Control Test	0.077	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:27

Data Download Date/Time: 01/02/2019 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADLEY,KELLY,L
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GARCIA-PAZ,SAMUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: GATZIARIS

DATE/TIME OF TEST: 12/09/2018 02:00:07

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.077	01:49	
Air Blank	0.000	01:50	
Subject Sample	# 1 0.156	01:53	Volume Not Met
Breath Volume	0.050	01:53	
Air Blank	0.000	01:54	
Air Blank	0.000	01:55	
Subject Sample	# 2	01:58	Subject Test Refused
Breath Volume	0.171	01:58	
Air Blank	0.000	01:59	
Control Test	0.077	01:59	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 01/02/2019 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADLEY,KELLY,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STUDZINSKI,KATINA
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: PETROPOULOS

DATE/TIME OF TEST: 12/09/2018 02:07:43

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:04	
Control Test	0.077	02:05	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Control Test	0.077	02:07	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 01/02/2019 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 12/25/2018 22:27:06

Diagnostics Check	OK	22:26	
Air Blank		22:26	RFI Detect
Air Blank	0.000	22:27	

Data Download Date/Time: 01/15/2019 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 12/25/2018 22:39:03

Diagnostics Check	OK	22:28	
Air Blank	0.000	22:29	
Control Test	0.080	22:29	
Air Blank	0.000	22:29	
Subject Sample	# 1	22:33	No Sample Provided
Breath Volume	0.000	22:33	
Air Blank	0.000	22:33	
Air Blank	0.000	22:35	
Subject Sample	# 2	22:37	Subject Test Refused
Breath Volume	0.000	22:37	
Air Blank	0.000	22:38	
Control Test	0.079	22:38	
Air Blank	0.000	22:38	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 01/15/2019 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BRIGGS,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: CANOVA

DATE/TIME OF TEST: 12/01/2018 00:23:02

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.081	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.195	00:18
Breath Volume	1.988	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:20
Subject Sample # 2	0.191	00:21
Breath Volume	2.425	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:23

Data Download Date/Time: 12/09/2018 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CAPALLIA

DATE/TIME OF TEST: 12/01/2018 03:31:20

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.082	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.132	03:24
Breath Volume	1.214	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:26
Subject Sample # 2	0.135	03:29
Breath Volume	1.746	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 12/09/2018 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: VALES

DATE/TIME OF TEST: 12/02/2018 01:57:34

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.248	01:52
Breath Volume	3.320	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.228	01:55
Breath Volume	1.976	01:55
Air Blank	0.000	01:56
Control Test	0.081	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 12/09/2018 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: THOMPSON,MSON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: JACALONE

DATE/TIME OF TEST: 12/02/2018 03:19:32

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.082	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.093	03:14
Breath Volume	2.132	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.093	03:17
Breath Volume	2.414	03:17
Air Blank	0.000	03:18
Control Test	0.082	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 12/09/2018 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: JENSEN,DAVID
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ESPARZA NAVARRO

DATE/TIME OF TEST: 12/04/2018 01:49:11

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.238	01:43
Breath Volume	1.550	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.239	01:47
Breath Volume	1.894	01:47
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 12/09/2018 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GROW

DATE/TIME OF TEST: 12/06/2018 01:47:07

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.090	01:42
Breath Volume	1.960	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:44
Subject Sample # 2	0.094	01:45
Breath Volume	2.851	01:45
Air Blank	0.000	01:46
Control Test	0.082	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 12/09/2018 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASH,JUSTIN,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: EDDINS,BRETT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: CHILDERS

DATE/TIME OF TEST: 12/07/2018 03:20:02

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:11	
Control Test	0.082	03:11	
Air Blank	0.000	03:12	
Subject Sample	# 1 0.218	03:15	
Breath Volume	2.222	03:15	
Air Blank	0.000	03:16	
Air Blank	0.000	03:18	
Subject Sample	# 2	03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 12/09/2018 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASH,JUSTIN,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 12/07/2018 03:40:22

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.081	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.171	03:35
Breath Volume	1.820	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.173	03:38
Breath Volume	1.839	03:38
Air Blank	0.000	03:39
Control Test	0.081	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 12/09/2018 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MILLER,CHRIS
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: SHANKS

DATE/TIME OF TEST: 12/08/2018 22:32:40

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.081	22:27
Air Blank	0.000	22:27
Subject Sample # 1	0.185	22:27
Breath Volume	2.144	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.176	22:31
Breath Volume	2.050	22:31
Air Blank	0.000	22:31
Control Test	0.080	22:32
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 12/09/2018 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

INSTRUMENT REGISTERED TO: MICCOSUKEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MISTY,L
BREATH TEST OPERATOR AGENCY: Miccosukee PD
ARRESTING OFFICER: GONZALEZ,DANIEL
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 12/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: CHARRON

DATE/TIME OF TEST: 12/14/2018 05:26:33

Diagnostics Check	OK	05:19
Air Blank	0.000	05:19
Control Test	0.078	05:20
Air Blank	0.000	05:20
Subject Sample # 1	0.190	05:21
Breath Volume	1.128	05:21
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.199	05:24
Breath Volume	2.171	05:24
Air Blank	0.000	05:25
Control Test	0.076	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 01/10/2019 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 12/02/2018 04:07:55

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.079	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1 0.250	04:00	Volume Not Met
Breath Volume	0.164	04:00	
Air Blank	0.000	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 2 0.217	04:05	Volume Not Met
Breath Volume	0.152	04:05	
Air Blank	0.000	04:06	
Control Test	0.078	04:07	
Air Blank	0.000	04:07	
Diagnostics Check	OK	04:07	

Data Download Date/Time: 12/03/2018 03:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 12/02/2018 04:17:08

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:10	
Control Test	0.077	04:11	
Air Blank	0.000	04:11	
Subject Sample	# 1 0.230	04:12	
Breath Volume	1.496	04:12	
Air Blank	0.000	04:13	
Air Blank	0.000	04:14	
Subject Sample	# 2 0.251	04:15	
Breath Volume	1.449	04:15	
Air Blank		04:16	Purge Fail
Air Blank	0.000	04:17	

Data Download Date/Time: 12/03/2018 03:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 12/02/2018 04:26:54

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.077	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.237	04:21
Breath Volume	2.097	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.239	04:25
Breath Volume	1.890	04:25
Air Blank	0.000	04:25
Control Test	0.078	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 12/03/2018 03:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: RODRIGUEZ SALGADO

DATE/TIME OF TEST: 12/26/2018 00:24:13

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.080	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.137	00:18
Breath Volume	1.710	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.136	00:22
Breath Volume	2.179	00:22
Air Blank	0.000	00:23
Control Test	0.078	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 01/07/2019 00:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,L
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONSS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 12/30/2018 03:14:14

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:01	
Control Test	0.080	03:02	
Air Blank	0.000	03:02	
Subject Sample	# 1 0.095	03:05	Volume Not Met
Breath Volume	0.605	03:05	
Air Blank	0.000	03:06	
Air Blank	0.000	03:08	
Subject Sample	# 2 0.105	03:09	
Breath Volume	1.699	03:09	
Air Blank	0.000	03:10	
Air Blank	0.000	03:12	
Subject Sample	# 3 0.109	03:12	
Breath Volume	2.191	03:12	
Air Blank	0.000	03:13	
Control Test	0.077	03:13	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 01/07/2019 00:42