

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DADDARIO,ASHLEY,A  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: GROLEAU,L  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 09/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CALLAWAY

DATE/TIME OF TEST: 10/03/2019 03:14:38

---

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.078	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.129	03:09
Breath Volume	3.417	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.138	03:13
Breath Volume	4.863	03:13
Air Blank	0.000	03:13
Control Test	0.078	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

---

Data Download Date/Time: 11/06/2019 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,DEVIN,A

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: HOWARD,RICHARD

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 09/27/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 18:57

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: WILKES

DATE/TIME OF TEST: 10/04/2019 19:37:29

---

Diagnostics Check	OK	19:34	
Air Blank	0.000	19:35	
Control Test	0.078	19:35	
Air Blank	0.000	19:35	
Subject Sample	# 1	19:36	Subject Test Refused
Breath Volume	0.000	19:36	
Air Blank	0.000	19:36	
Control Test	0.077	19:36	
Air Blank	0.000	19:37	
Diagnostics Check	OK	19:37	

---

Data Download Date/Time: 11/06/2019 11:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DADDARIO,ASHLEY,A  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: RAMOS,J  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 09/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: WATERS

DATE/TIME OF TEST: 10/05/2019 23:27:36

---

Diagnostics Check	OK	23:19	
Air Blank	0.000	23:19	
Control Test	0.077	23:19	
Air Blank	0.000	23:20	
Subject Sample	# 1 0.000	23:23	Volume Not Met
Breath Volume	0.015	23:23	
Air Blank	0.000	23:24	
Air Blank	0.000	23:25	
Subject Sample	# 2	23:26	Subject Test Refused
Breath Volume	0.000	23:26	
Air Blank	0.000	23:26	
Control Test	0.076	23:27	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:27	

---

Data Download Date/Time: 11/06/2019 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TURO,JAMES,M

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: BASS,J

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 09/27/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 23:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: DICKERSON II

DATE/TIME OF TEST: 10/08/2019 00:15:56

---

Diagnostics Check		OK	00:08
Air Blank		0.000	00:08
Control Test		0.078	00:09
Air Blank		0.000	00:09
Subject Sample	# 1	0.264	00:11
Breath Volume		4.085	00:11
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.270	00:14
Breath Volume		4.386	00:14
Air Blank		0.000	00:15
Control Test		0.077	00:15
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

---

Data Download Date/Time: 11/06/2019 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CASON,CHRISTOPHER,N

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: WIGGLSWORTH,C

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 09/27/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 02:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: BOYNTON

DATE/TIME OF TEST: 10/11/2019 03:53:43

---

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.077	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.165	03:45
Breath Volume		1.238	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.186	03:48
Breath Volume		2.507	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 3	0.189	03:52
Breath Volume		1.300	03:52
Air Blank		0.000	03:52
Control Test		0.077	03:53
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

---

Data Download Date/Time: 11/06/2019 11:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOWERMASTER,MARY,E  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: HOWARD,RUSS  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 09/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 10:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:25

SUBJECT'S LAST NAME: LOOMIS

DATE/TIME OF TEST: 10/14/2019 11:56:49

---

Diagnostics Check	OK	11:45
Air Blank	0.000	11:45
Control Test	0.077	11:46
Air Blank	0.000	11:46
Subject Sample # 1	0.329	11:48
Breath Volume	3.113	11:48
Air Blank	0.000	11:49
Air Blank	0.000	11:50
Subject Sample # 2	0.297	11:51
Breath Volume	2.187	11:51
Air Blank	0.000	11:52
Air Blank	0.000	11:53
Subject Sample # 3	0.301	11:55
Breath Volume	2.023	11:55
Air Blank	0.000	11:55
Control Test	0.077	11:56
Air Blank	0.000	11:56
Diagnostics Check	OK	11:56

---

Data Download Date/Time: 11/06/2019 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: GRICE,W  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 09/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: FEASTER

DATE/TIME OF TEST: 10/17/2019 20:46:24

---

Diagnostics Check	OK	20:42	
Air Blank	0.000	20:43	
Control Test	0.078	20:43	
Air Blank	0.000	20:43	
Subject Sample	# 1	20:45	Subject Test Refused
Breath Volume	0.000	20:45	
Air Blank	0.000	20:45	
Control Test	0.077	20:45	
Air Blank	0.000	20:46	
Diagnostics Check	OK	20:46	

---

Data Download Date/Time: 11/06/2019 11:29





**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,DEVIN,A

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: WIGGLESWORTH,C

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 09/27/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 20:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 10/20/2019 21:13:12

---

Diagnostics Check		OK	21:06
Air Blank		0.000	21:06
Control Test		0.076	21:06
Air Blank		0.000	21:07
Subject Sample	# 1	0.230	21:08
Breath Volume		1.246	21:08
Air Blank		0.000	21:08
Air Blank		0.000	21:10
Subject Sample	# 2	0.211	21:11
Breath Volume		1.808	21:11
Air Blank		0.000	21:12
Control Test		0.077	21:12
Air Blank		0.000	21:13
Diagnostics Check		OK	21:13

---

Data Download Date/Time: 11/06/2019 11:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: WIGGLESWORTH,C  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 09/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: WILLOUGHBY

DATE/TIME OF TEST: 10/21/2019 04:03:46

---

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.076	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.155	03:59
Breath Volume	2.515	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.157	04:02
Breath Volume	2.140	04:02
Air Blank	0.000	04:02
Control Test	0.076	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

---

Data Download Date/Time: 11/06/2019 11:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVENPORT, JOSHUA, J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: DAVENPORT, JOSHUA  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 10/22/2019 23:10:26

---

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:07	
Control Test	0.077	23:07	
Air Blank	0.000	23:08	
Subject Sample	# 1	23:09	Subject Test Refused
Breath Volume	0.000	23:09	
Air Blank	0.000	23:09	
Control Test	0.077	23:09	
Air Blank	0.000	23:10	
Diagnostics Check	OK	23:10	

---

Data Download Date/Time: 11/06/2019 11:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROY,GARY,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: ROY,GARY  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: FUHRMAN

DATE/TIME OF TEST: 10/01/2019 23:15:37

---

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.079	23:10
Air Blank	0.000	23:10
Subject Sample # 1	0.119	23:11
Breath Volume	2.035	23:11
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.124	23:14
Breath Volume	2.132	23:14
Air Blank	0.000	23:14
Control Test	0.079	23:15
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

---

Data Download Date/Time: 10/03/2019 13:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,JEFFREY,A  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: THOMPSON,JEFFREY  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: QUINTANA

DATE/TIME OF TEST: 10/02/2019 00:40:33

---

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.078	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.271	00:36
Breath Volume	3.539	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.258	00:39
Breath Volume	2.476	00:39
Air Blank	0.000	00:39
Control Test	0.079	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

---

Data Download Date/Time: 10/03/2019 13:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,JEFFREY,A  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: COLLINS,TIMOTHY  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 10/02/2019 01:09:15

---

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:01	
Control Test		0.078	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1	0.041	01:05	Volume Not Met
Breath Volume		0.109	01:05	
Air Blank		0.000	01:05	
Air Blank		0.000	01:07	
Subject Sample	# 2		01:07	Subject Test Refused
Breath Volume		0.000	01:07	
Air Blank		0.000	01:08	
Control Test		0.078	01:08	
Air Blank		0.000	01:09	
Diagnostics Check		OK	01:09	

---

Data Download Date/Time: 10/03/2019 13:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,JEFFREY,A  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: THOMPSON,JEFFREY  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/03/2019 00:29:00

---

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.162	00:23
Breath Volume	2.210	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.163	00:27
Breath Volume	2.250	00:27
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:29

---

Data Download Date/Time: 10/03/2019 13:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: NORTHWAY,GEORGIA  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 10/03/2019 03:37:45

---

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:29	
Control Test	0.081	03:30	
Air Blank	0.000	03:30	
Subject Sample	# 1	03:33	No Sample Provided
Breath Volume	0.000	03:33	
Air Blank	0.000	03:34	
Air Blank	0.000	03:36	
Subject Sample	# 2	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Control Test	0.080	03:37	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

---

Data Download Date/Time: 10/03/2019 13:28



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 181692
BREATH TEST OPERATOR: SIRMANS, ABNER, K	DRY GAS EXPIRY DATE: 02/05/2021
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD	VIOLATION CODE: DUI
ARRESTING OFFICER: SIRMANS, ABNER	ARREST TIME: 21:27
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD	GENDER: Male
LAST AGENCY INSPECTION: 09/12/2019	OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: PABLO-HERNANDEZ      DATE/TIME OF TEST: 10/13/2019      22:20:23

---

Diagnostics Check	OK	22:12
Air Blank	0.000	22:13
Control Test	0.081	22:13
Air Blank	0.000	22:14
Subject Sample      # 1	0.198	22:14
Breath Volume	2.519	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample      # 2	0.197	22:18
Breath Volume	2.027	22:18
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

---

Data Download Date/Time: 11/13/2019 14:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOLL, MICHAEL,  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: WILSON, MICHAEL  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 09/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 10/14/2019 01:21:48

---

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.081	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.091	01:15
Breath Volume	1.117	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.084	01:20
Breath Volume	1.949	01:20
Air Blank	0.000	01:20
Control Test	0.081	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

---

Data Download Date/Time: 11/13/2019 14:06







# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOLL, MICHAEL,  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: WOLL, MICHAEL  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 09/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 10/23/2019 23:35:33

---

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:32	
Control Test	0.082	23:33	
Air Blank	0.000	23:33	
Subject Sample	# 1	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:34	
Control Test	0.080	23:34	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:35	

---

Data Download Date/Time: 11/13/2019 14:06









**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLIS,TERENCE,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: HARM,C  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 10/05/2019 03:33:07

---

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.145	03:28
Breath Volume	2.152	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.136	03:31
Breath Volume	1.648	03:31
Air Blank	0.000	03:32
Control Test	0.078	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

---

Data Download Date/Time: 10/30/2019 11:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HOLLIS,TERENCE,L

BREATH TEST OPERATOR AGENCY: Sumter County SO

ARRESTING OFFICER: HARM,C

ARRESTING OFFICER AGENCY: Sumter County SO

LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A2

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 11:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: GOBLE

DATE/TIME OF TEST: 10/10/2019 12:23:37

---

Diagnostics Check	OK	12:20	
Air Blank	0.000	12:21	
Control Test	0.079	12:21	
Air Blank	0.000	12:22	
Subject Sample	# 1	12:22	Subject Test Refused
Breath Volume	0.000	12:22	
Air Blank	0.000	12:22	
Control Test	0.079	12:23	
Air Blank	0.000	12:23	
Diagnostics Check	OK	12:23	

---

Data Download Date/Time: 10/30/2019 11:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUCHNICKI,ZACHARIAH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop K  
ARRESTING OFFICER: REDWAN,ALFREDO  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: MUCHNICKI

DATE/TIME OF TEST: 10/11/2019 20:38:59

---

Diagnostics Check	OK	20:32
Air Blank	0.000	20:33
Control Test	0.078	20:33
Air Blank	0.000	20:34
Subject Sample # 1	0.000	20:34
Breath Volume	2.070	20:34
Air Blank	0.000	20:35
Air Blank	0.000	20:37
Subject Sample # 2	0.000	20:37
Breath Volume	2.437	20:37
Air Blank	0.000	20:38
Control Test	0.078	20:38
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

---

Data Download Date/Time: 10/30/2019 11:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: SLAY,ADRIAN  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 10/12/2019 02:13:20

---

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:10	
Control Test	0.078	02:11	
Air Blank	0.000	02:11	
Subject Sample	# 1	02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:12	
Control Test	0.078	02:12	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

---

Data Download Date/Time: 10/30/2019 11:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUTTON,EDWIN,C  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: ORTIZ,B  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: STEINBERG

DATE/TIME OF TEST: 10/12/2019 05:26:21

---

Diagnostics Check	OK	05:19
Air Blank	0.000	05:19
Control Test	0.079	05:20
Air Blank	0.000	05:20
Subject Sample # 1	0.150	05:21
Breath Volume	1.226	05:21
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.168	05:24
Breath Volume	1.113	05:24
Air Blank	0.000	05:25
Control Test	0.078	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

---

Data Download Date/Time: 10/30/2019 11:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WELLS,AUSTIN,D

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: WELLS,AUSTIN

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A2

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 04:08

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: CODY

DATE/TIME OF TEST: 10/20/2019 04:56:16

---

Diagnostics Check	OK	04:53	
Air Blank	0.000	04:53	
Control Test	0.077	04:54	
Air Blank	0.000	04:54	
Subject Sample	# 1	04:54	Subject Test Refused
Breath Volume	0.000	04:54	
Air Blank	0.000	04:55	
Control Test	0.079	04:55	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:56	

---

Data Download Date/Time: 10/30/2019 11:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: CASIMIRO,BRANDON  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 10/21/2019 20:29:29

---

Diagnostics Check	OK	20:20
Air Blank	0.000	20:21
Control Test	0.078	20:21
Air Blank	0.000	20:21
Subject Sample # 1	0.148	20:24
Breath Volume	1.601	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 2	0.143	20:27
Breath Volume	1.328	20:27
Air Blank	0.000	20:28
Control Test	0.078	20:28
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

---

Data Download Date/Time: 10/30/2019 11:26



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,WYATT,M  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: HARM,CONNOR  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 10/24/2019 18:27:57

---

Diagnostics Check	OK	18:25	
Air Blank	0.000	18:25	
Control Test	0.078	18:25	
Air Blank	0.000	18:26	
Subject Sample	# 1	18:26	Subject Test Refused
Breath Volume	0.000	18:26	
Air Blank	0.000	18:27	
Control Test	0.079	18:27	
Air Blank	0.000	18:27	
Diagnostics Check	OK	18:27	

---

Data Download Date/Time: 10/30/2019 11:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,WYATT,M  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: PERDIKES,PAUL  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: MAYS

DATE/TIME OF TEST: 10/24/2019 22:45:46

---

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.078	22:39
Air Blank	0.000	22:39
Subject Sample # 1	0.000	22:41
Breath Volume	1.500	22:41
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.000	22:44
Breath Volume	1.347	22:44
Air Blank	0.000	22:44
Control Test	0.078	22:45
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

---

Data Download Date/Time: 10/30/2019 11:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LITZKOW,CARLTON,D  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: LITZKOW,CARLTON  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: NORDSTROM

DATE/TIME OF TEST: 10/05/2019 04:10:14

---

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.079	04:05	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:08	RFI Detect
Breath Volume	0.000	04:08	
Air Blank	0.000	04:09	
Control Test	0.080	04:09	
Air Blank	0.000	04:10	
Diagnostics Check	OK	04:10	

---

Data Download Date/Time: 10/23/2019 14:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LITZKOW,CARLTON,D  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: LITZKOW,CARLTON  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: NORDSTROM

DATE/TIME OF TEST: 10/05/2019 04:23:23

---

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:12	
Control Test	0.080	04:12	
Air Blank	0.000	04:13	
Subject Sample	# 1	04:16	No Sample Provided
Breath Volume	0.000	04:16	
Air Blank	0.000	04:16	
Air Blank	0.000	04:18	
Subject Sample	# 2	04:21	No Sample Provided
Breath Volume	0.000	04:21	
Air Blank	0.000	04:22	
Control Test	0.080	04:22	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:23	

---

Data Download Date/Time: 10/23/2019 14:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVAREZ,ROBERTO,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: ALVAREZ,ROBERTO  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: ODOM

DATE/TIME OF TEST: 10/05/2019 19:24:55

---

Diagnostics Check	OK	19:22	
Air Blank	0.000	19:22	
Control Test	0.080	19:22	
Air Blank	0.000	19:23	
Subject Sample	# 1	19:23	Subject Test Refused
Breath Volume	0.000	19:23	
Air Blank	0.000	19:24	
Control Test	0.081	19:24	
Air Blank	0.000	19:24	
Diagnostics Check	OK	19:24	

---

Data Download Date/Time: 10/23/2019 14:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LITZKOW,CARLTON,D  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: LITZKOW,CARLTON  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: DILLENBECK

DATE/TIME OF TEST: 10/06/2019 22:56:45

---

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:45	
Control Test	0.080	22:46	
Air Blank	0.000	22:46	
Subject Sample	# 1	22:49	No Sample Provided
Breath Volume	0.000	22:49	
Air Blank	0.000	22:50	
Air Blank	0.000	22:52	
Subject Sample	# 2	22:55	No Sample Provided
Breath Volume	0.000	22:55	
Air Blank	0.000	22:55	
Control Test	0.081	22:56	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

---

Data Download Date/Time: 10/23/2019 14:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DILLON,JUSTIN,P  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: WORTH,GARRETT  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: AMERSON

DATE/TIME OF TEST: 10/08/2019 06:22:06

---

Diagnostics Check	OK	06:15
Air Blank	0.000	06:15
Control Test	0.080	06:15
Air Blank	0.000	06:16
Subject Sample # 1	0.132	06:16
Breath Volume	1.226	06:16
Air Blank	0.000	06:17
Air Blank	0.000	06:19
Subject Sample # 2	0.143	06:20
Breath Volume	3.382	06:20
Air Blank	0.000	06:21
Control Test	0.080	06:21
Air Blank	0.000	06:22
Diagnostics Check	OK	06:22

---

Data Download Date/Time: 10/23/2019 14:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA,BECKY,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BROWN,RICK  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: VOLKMANN

DATE/TIME OF TEST: 10/10/2019 18:16:19

---

Diagnostics Check	OK	18:13	
Air Blank	0.000	18:13	
Control Test	0.080	18:14	
Air Blank	0.000	18:14	
Subject Sample	# 1	18:14	Subject Test Refused
Breath Volume	0.000	18:14	
Air Blank	0.000	18:15	
Control Test	0.080	18:15	
Air Blank	0.000	18:16	
Diagnostics Check	OK	18:16	

---

Data Download Date/Time: 10/23/2019 14:43



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LITZKOW,CLAYTON,P  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: LITZKOW,CLAYTON  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: WILDER

DATE/TIME OF TEST: 10/15/2019 03:18:57

---

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.081	03:08	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:11	No Sample Provided
Breath Volume	0.000	03:11	
Air Blank	0.000	03:12	
Air Blank	0.000	03:14	
Subject Sample	# 2	03:17	No Sample Provided
Breath Volume	0.000	03:17	
Air Blank	0.000	03:18	
Control Test	0.082	03:18	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

---

Data Download Date/Time: 10/23/2019 14:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CABRAL,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BEDENBAUGH,JUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 10/19/2019 22:32:34

---

Diagnostics Check	OK	22:24
Air Blank	0.000	22:25
Control Test	0.080	22:25
Air Blank	0.000	22:26
Subject Sample # 1	0.159	22:28
Breath Volume	1.644	22:28
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.168	22:31
Breath Volume	1.328	22:31
Air Blank	0.000	22:31
Control Test	0.079	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

---

Data Download Date/Time: 10/23/2019 14:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CABRAL,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: CABRAL,JONATHAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 10/20/2019 03:29:55

---

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.081	03:19	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:22	No Sample Provided
Breath Volume	0.000	03:22	
Air Blank	0.000	03:23	
Air Blank	0.000	03:25	
Subject Sample	# 2	03:28	No Sample Provided
Breath Volume	0.000	03:28	
Air Blank	0.000	03:28	
Control Test	0.080	03:29	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:29	

---

Data Download Date/Time: 10/23/2019 14:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LITZKOW,CLAYTON,P  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: LITZKOW,CLAYTON  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: LOZADA

DATE/TIME OF TEST: 10/20/2019 03:58:13

---

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.081	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.224	03:53
Breath Volume		2.535	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.219	03:56
Breath Volume		2.503	03:56
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

---

Data Download Date/Time: 10/23/2019 14:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TURRO,NICHOLAS,J  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: TURRO,NICHOLAS  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 10:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: SALES

DATE/TIME OF TEST: 10/20/2019 13:02:25

---

Diagnostics Check	OK	12:56
Air Blank	0.000	12:56
Control Test	0.080	12:56
Air Blank	0.000	12:57
Subject Sample # 1	0.091	12:57
Breath Volume	3.441	12:57
Air Blank	0.000	12:58
Air Blank	0.000	13:00
Subject Sample # 2	0.078	13:00
Breath Volume	1.765	13:00
Air Blank	0.000	13:01
Control Test	0.080	13:01
Air Blank	0.000	13:02
Diagnostics Check	OK	13:02

---

Data Download Date/Time: 10/23/2019 14:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURINGTON,PATRICK,A  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: PURINGTON,PATRICK  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 10/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: WAYLEN

DATE/TIME OF TEST: 10/27/2019 04:24:25

---

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.080	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.189	04:20
Breath Volume	4.148	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.177	04:22
Breath Volume	2.984	04:22
Air Blank	0.000	04:23
Control Test	0.077	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

---

Data Download Date/Time: 11/19/2019 11:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRICKER,JEFFREY,S  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: FRICKER,JEFFREY  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 10/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 07:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 10/29/2019 08:57:17

---

Diagnostics Check	OK	08:49
Air Blank	0.000	08:49
Control Test	0.081	08:49
Air Blank	0.000	08:50
Subject Sample # 1	0.092	08:52
Breath Volume	2.558	08:52
Air Blank	0.000	08:53
Air Blank	0.000	08:55
Subject Sample # 2	0.082	08:55
Breath Volume	2.031	08:55
Air Blank	0.000	08:56
Control Test	0.076	08:56
Air Blank	0.000	08:57
Diagnostics Check	OK	08:57

---

Data Download Date/Time: 11/19/2019 11:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURINGTON,PATRICK,A  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: PURINGTON,PATRICK  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 10/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: STRAUGHN

DATE/TIME OF TEST: 10/31/2019 23:20:49

---

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:09	
Control Test	0.079	23:10	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:13	No Sample Provided
Breath Volume	0.000	23:13	
Air Blank	0.000	23:14	
Air Blank	0.000	23:16	
Subject Sample	# 2	23:19	No Sample Provided
Breath Volume	0.000	23:19	
Air Blank	0.000	23:19	
Control Test	0.081	23:20	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

---

Data Download Date/Time: 11/19/2019 11:14



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMILTON, RAY, C  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: HAMILTON, RAY  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 09/06/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:06

SUBJECT'S LAST NAME: SWAILS

DATE/TIME OF TEST: 10/06/2019 07:35:00

---

Diagnostics Check	OK	07:28
Air Blank	0.000	07:28
Control Test	0.081	07:29
Air Blank	0.000	07:29
Subject Sample # 1	0.138	07:30
Breath Volume	4.246	07:30
Air Blank	0.000	07:30
Air Blank	0.000	07:32
Subject Sample # 2	0.138	07:33
Breath Volume	4.089	07:33
Air Blank	0.000	07:34
Control Test	0.080	07:34
Air Blank	0.000	07:34
Diagnostics Check	OK	07:35

---

Data Download Date/Time: 10/10/2019 09:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALKER, LANDON, L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SMITH, CHAD  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 09/06/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: HORNUNG

DATE/TIME OF TEST: 10/06/2019 23:25:00

---

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:22	
Control Test	0.080	23:22	
Air Blank	0.000	23:23	
Subject Sample	# 1	23:23	Subject Test Refused
Breath Volume	0.000	23:23	
Air Blank	0.000	23:24	
Control Test	0.080	23:24	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:25	

---

Data Download Date/Time: 10/10/2019 09:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FITCHETT, ADAM, R  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SANCHEZ, FRANCISCO  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 09/06/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 10/09/2019 23:00:33

---

Diagnostics Check	OK	22:53
Air Blank	0.000	22:54
Control Test	0.081	22:54
Air Blank	0.000	22:55
Subject Sample # 1	0.072	22:56
Breath Volume	1.808	22:56
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.072	22:58
Breath Volume	2.363	22:58
Air Blank	0.000	22:59
Control Test	0.078	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

---

Data Download Date/Time: 10/10/2019 09:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALKER, LANDON, L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WALKER, LANDON  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 09/06/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 12:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ZAPATA

DATE/TIME OF TEST: 10/10/2019 01:25:23

---

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.080	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:23	Subject Test Refused
Breath Volume	0.101	01:23	
Air Blank	0.000	01:24	
Control Test	0.080	01:24	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

---

Data Download Date/Time: 10/10/2019 09:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMALL,BRANNON,A  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SMALL,BRANNON  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: SUAZO CACERES

DATE/TIME OF TEST: 10/13/2019 00:30:27

---

Diagnostics Check		OK	00:23
Air Blank		0.000	00:23
Control Test		0.080	00:24
Air Blank		0.000	00:24
Subject Sample	# 1	0.207	00:25
Breath Volume		2.285	00:25
Air Blank		0.000	00:26
Air Blank		0.000	00:27
Subject Sample	# 2	0.209	00:28
Breath Volume		2.042	00:28
Air Blank		0.000	00:29
Control Test		0.078	00:29
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

---

Data Download Date/Time: 11/07/2019 10:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMALL,BRANNON,A  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: MARSHALL,ALEXANDRIA  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: LUNDY

DATE/TIME OF TEST: 10/13/2019 23:48:31

---

Diagnostics Check	OK	23:40	
Air Blank	0.000	23:40	
Control Test	0.079	23:40	
Air Blank	0.000	23:41	
Subject Sample	# 1	23:44	No Sample Provided
Breath Volume	0.000	23:44	
Air Blank	0.000	23:44	
Air Blank	0.000	23:46	
Subject Sample	# 2	23:47	Subject Test Refused
Breath Volume	0.000	23:47	
Air Blank	0.000	23:47	
Control Test	0.080	23:47	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

---

Data Download Date/Time: 11/07/2019 10:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMALL,BRANNON,A  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SMALL,BRANNON  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 10/18/2019 03:47:01

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.217	03:42
Breath Volume	1.570	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:44
Subject Sample # 2	0.209	03:45
Breath Volume	1.730	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:47

---

Data Download Date/Time: 11/07/2019 10:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALKER, LONDON, L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SANCHEZ, FRANCISCO  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: LIZARDO

DATE/TIME OF TEST: 10/20/2019 02:47:43

---

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:45	
Control Test	0.081	02:45	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:46	
Control Test	0.080	02:47	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

---

Data Download Date/Time: 11/07/2019 10:49



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALKER, LANDON, L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SIMPKINS, BRANDON  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: THEDE

DATE/TIME OF TEST: 10/20/2019 05:39:36

---

Diagnostics Check	OK	05:33
Air Blank	0.000	05:33
Control Test	0.080	05:33
Air Blank	0.000	05:34
Subject Sample # 1	0.000	05:35
Breath Volume	3.222	05:35
Air Blank	0.000	05:35
Air Blank	0.000	05:37
Subject Sample # 2	0.000	05:38
Breath Volume	1.472	05:38
Air Blank	0.000	05:38
Control Test	0.079	05:38
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

---

Data Download Date/Time: 11/07/2019 10:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALKER, LANDON, L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SMITH, CHAD  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: MAFORTE

DATE/TIME OF TEST: 10/20/2019 05:53:41

---

Diagnostics Check	OK	05:50	
Air Blank	0.000	05:51	
Control Test	0.080	05:51	
Air Blank	0.000	05:52	
Subject Sample	# 1	05:52	Subject Test Refused
Breath Volume	0.000	05:52	
Air Blank	0.000	05:52	
Control Test	0.080	05:53	
Air Blank	0.000	05:53	
Diagnostics Check	OK	05:53	

---

Data Download Date/Time: 11/07/2019 10:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALMATEER,TRACY,D  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: PALMATEER,TRACY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: LOCICERO

DATE/TIME OF TEST: 10/21/2019 04:37:14

---

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.176	04:32
Breath Volume	1.230	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.158	04:35
Breath Volume	1.867	04:35
Air Blank	0.000	04:36
Control Test	0.079	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

---

Data Download Date/Time: 11/07/2019 10:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIFER,JUSTIN,S  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: PIFER,JUSTIN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: BERRIOS

DATE/TIME OF TEST: 10/23/2019 19:12:13

---

Diagnostics Check	OK	19:08	
Air Blank	0.000	19:09	
Control Test	0.081	19:09	
Air Blank	0.000	19:10	
Subject Sample	# 1	19:10	Subject Test Refused
Breath Volume	0.000	19:10	
Air Blank	0.000	19:11	
Control Test	0.081	19:11	
Air Blank	0.000	19:12	
Diagnostics Check	OK	19:12	

---

Data Download Date/Time: 11/07/2019 10:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALKER, LANDON, L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WALKER, LANDON  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: BAUTISTA

DATE/TIME OF TEST: 10/24/2019 02:06:33

---

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.171	02:00
Breath Volume	1.328	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.171	02:04
Breath Volume	1.437	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

---

Data Download Date/Time: 11/07/2019 10:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMALL,BRANNON,A  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: JONES,CARLOS  
ARRESTING OFFICER AGENCY: Shalimar PD  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 10/25/2019 22:26:41

---

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.080	22:19
Air Blank	0.000	22:19
Subject Sample # 1	0.210	22:21
Breath Volume	1.707	22:21
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.193	22:24
Breath Volume	1.164	22:24
Air Blank	0.000	22:25
Control Test	0.078	22:26
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

---

Data Download Date/Time: 11/07/2019 10:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMALL,BRANNON,A  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SMALL,BRANNON  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: MARSH

DATE/TIME OF TEST: 10/27/2019 00:53:36

---

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.080	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.183	00:48
Breath Volume	1.738	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.171	00:52
Breath Volume	2.164	00:52
Air Blank	0.000	00:52
Control Test	0.077	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

---

Data Download Date/Time: 11/07/2019 10:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMALL,BRANNON,A  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: MARSHALL,ALEXANDRIA  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: MAKAR

DATE/TIME OF TEST: 10/28/2019 03:24:03

---

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.121	03:18
Breath Volume	1.199	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.122	03:22
Breath Volume	1.132	03:22
Air Blank	0.000	03:23
Control Test	0.078	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:24

---

Data Download Date/Time: 11/07/2019 10:49



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRANE,RICHARD,B  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: DRANE,RICHARD  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 11:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: SCOVILL

DATE/TIME OF TEST: 10/31/2019 00:57:44

---

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.075	00:53
Breath Volume	1.734	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.073	00:56
Breath Volume	1.335	00:56
Air Blank	0.000	00:56
Control Test	0.080	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

---

Data Download Date/Time: 11/07/2019 10:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: CRAWFORD,BENJAMIN  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 10/31/2019 03:01:40

---

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.196	02:57
Breath Volume	3.460	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.201	03:00
Breath Volume	3.550	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

---

Data Download Date/Time: 11/07/2019 10:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: IVEY,BRUCE,A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: IVEY,BRUCE  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: OLDROYD

DATE/TIME OF TEST: 10/06/2019 22:37:54

---

Diagnostics Check	OK	22:31
Air Blank	0.000	22:31
Control Test	0.082	22:32
Air Blank	0.000	22:32
Subject Sample # 1	0.162	22:33
Breath Volume	4.566	22:33
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.162	22:36
Breath Volume	3.976	22:36
Air Blank	0.000	22:36
Control Test	0.080	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

---

Data Download Date/Time: 11/08/2019 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURINGTON,PATRICK,A  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: PURINGTON,PATRICK  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: NEUHAUS

DATE/TIME OF TEST: 10/22/2019 03:27:23

---

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.184	03:22
Breath Volume	2.464	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample # 2	0.184	03:25
Breath Volume	2.601	03:25
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

---

Data Download Date/Time: 10/23/2019 14:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOGGAN,MARLON,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: BOGGAN,MARLON  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 09/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 10/07/2019 00:17:24

---

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:06	
Control Test	0.081	00:06	
Air Blank	0.000	00:07	
Subject Sample	# 1	00:10	No Sample Provided
Breath Volume	0.000	00:10	
Air Blank	0.000	00:10	
Air Blank	0.000	00:12	
Subject Sample	# 2	00:16	No Sample Provided
Breath Volume	0.000	00:16	
Air Blank	0.000	00:16	
Control Test	0.081	00:16	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

---

Data Download Date/Time: 11/06/2019 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALKER, LANDON, L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WALKER, LANDON  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 09/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BLAND

DATE/TIME OF TEST: 10/15/2019 02:16:10

---

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:13	
Control Test	0.080	02:14	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:14	Subject Test Refused
Breath Volume	0.000	02:14	
Air Blank	0.000	02:15	
Control Test	0.081	02:15	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

---

Data Download Date/Time: 11/06/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: MORENO,JACOB  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 09/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: SETTLES

DATE/TIME OF TEST: 10/20/2019 01:30:35

---

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.079	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.186	01:25
Breath Volume	1.492	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.183	01:28
Breath Volume	2.371	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

---

Data Download Date/Time: 11/06/2019 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOWNING,CADEN,T  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: CRAIG,ANYCE  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 09/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:32

SUBJECT'S LAST NAME: FREDRICK

DATE/TIME OF TEST: 10/25/2019 01:05:55

---

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:51	
Control Test	0.080	00:52	
Air Blank	0.000	00:52	
Subject Sample	# 1 0.072	00:54	
Breath Volume	1.410	00:54	
Air Blank	0.000	00:55	
Air Blank	0.000	00:57	
Subject Sample	# 2 0.075	01:00	Volume Not Met
Breath Volume	0.500	01:00	
Air Blank	0.000	01:01	
Air Blank	0.000	01:03	
Subject Sample	# 3 0.070	01:04	
Breath Volume	1.351	01:04	
Air Blank	0.000	01:04	
Control Test	0.077	01:05	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

---

Data Download Date/Time: 11/06/2019 10:37



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURGESS,MASON,W  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: BURGESS,MASON  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 09/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SOUTHARD STOKES

DATE/TIME OF TEST: 10/26/2019 03:36:32

---

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.156	03:32
Breath Volume	1.820	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.144	03:34
Breath Volume	1.882	03:34
Air Blank	0.000	03:35
Control Test	0.078	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

---

Data Download Date/Time: 11/06/2019 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: KIMBRIL,JOSHUA  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 09/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GOLDMAN

DATE/TIME OF TEST: 10/27/2019 04:32:14

---

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:21	
Control Test	0.079	04:21	
Air Blank	0.000	04:22	
Subject Sample	# 1	04:25	No Sample Provided
Breath Volume	0.000	04:25	
Air Blank	0.000	04:25	
Air Blank	0.000	04:27	
Subject Sample	# 2	04:30	No Sample Provided
Breath Volume	0.000	04:30	
Air Blank	0.000	04:31	
Control Test	0.080	04:31	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

---

Data Download Date/Time: 11/06/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STEINER,ZACHARY,A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: STEINER,ZACHARY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 10/02/2019 18:33:30

---

Diagnostics Check	OK	18:24
Air Blank	0.000	18:25
Control Test	0.083	18:25
Air Blank	0.000	18:26
Subject Sample # 1	0.140	18:28
Breath Volume	1.632	18:28
Air Blank	0.000	18:29
Air Blank	0.000	18:31
Subject Sample # 2	0.142	18:31
Breath Volume	1.519	18:31
Air Blank	0.000	18:32
Control Test	0.082	18:32
Air Blank	0.000	18:33
Diagnostics Check	OK	18:33

---

Data Download Date/Time: 10/24/2019 21:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERRILL,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MERRILL,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:59

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 10/02/2019 18:48:38

---

Diagnostics Check	OK	18:45	
Air Blank	0.000	18:45	
Control Test	0.082	18:46	
Air Blank	0.000	18:46	
Subject Sample	# 1 0.140	18:47	
Breath Volume	2.601	18:47	
Air Blank		18:48	RFI Detect
Air Blank		18:48	RFI Detect
Air Blank	0.000	18:48	

---

Data Download Date/Time: 10/24/2019 21:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERRILL,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MERRILL,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:59

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 10/02/2019 19:00:49

---

Diagnostics Check	OK	18:53
Air Blank	0.000	18:54
Control Test	0.083	18:54
Air Blank	0.000	18:54
Subject Sample # 1	0.139	18:55
Breath Volume	2.398	18:55
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample # 2	0.140	18:59
Breath Volume	2.675	18:59
Air Blank	0.000	18:59
Control Test	0.083	19:00
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

---

Data Download Date/Time: 10/24/2019 21:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOGGAN,MARLON,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: BOGGAN,MARLON  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: RAYMOND

DATE/TIME OF TEST: 10/06/2019 01:08:02

---

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.083	00:58
Air Blank	0.000	00:59
Subject Sample # 1	0.147	01:02
Breath Volume	1.187	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.136	01:06
Breath Volume	1.257	01:06
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:08

---

Data Download Date/Time: 10/24/2019 21:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE,JAMES,I  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MOORE,JAMES  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 10/06/2019 05:18:54

---

Diagnostics Check	OK	05:12
Air Blank	0.000	05:12
Control Test	0.084	05:13
Air Blank	0.000	05:13
Subject Sample # 1	0.000	05:14
Breath Volume	1.410	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 2	0.000	05:17
Breath Volume	3.707	05:17
Air Blank	0.000	05:18
Control Test	0.083	05:18
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

---

Data Download Date/Time: 10/24/2019 21:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND II,JACK,W  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: HOLLAND II,JACK  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: DUPUY

DATE/TIME OF TEST: 10/10/2019 19:34:27

---

Diagnostics Check	OK	19:27
Air Blank	0.000	19:28
Control Test	0.083	19:28
Air Blank	0.000	19:29
Subject Sample # 1	0.078	19:29
Breath Volume	2.652	19:29
Air Blank	0.000	19:30
Air Blank	0.000	19:32
Subject Sample # 2	0.080	19:32
Breath Volume	3.191	19:32
Air Blank	0.000	19:33
Control Test	0.083	19:33
Air Blank	0.000	19:34
Diagnostics Check	OK	19:34

---

Data Download Date/Time: 10/24/2019 21:12



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOGGAN,MARLON,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: BOGGAN,MARLON  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 10/11/2019 02:57:22

---

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.084	02:56	
Air Blank		02:56	RFI Detect
Air Blank	0.000	02:57	

---

Data Download Date/Time: 10/24/2019 21:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOGGAN,MARLON,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: BOGGAN,MARLON  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 10/11/2019 03:06:17

---

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.084	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.176	03:01
Breath Volume	2.121	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:03
Subject Sample # 2	0.180	03:04
Breath Volume	1.238	03:04
Air Blank	0.000	03:05
Control Test	0.082	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

---

Data Download Date/Time: 10/24/2019 21:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STEINER,ZACHARY,A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: STEINER,ZACHARY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: EARLE

DATE/TIME OF TEST: 10/11/2019 23:21:53

---

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.084	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.110	23:17
Breath Volume	2.367	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:19
Subject Sample # 2	0.113	23:20
Breath Volume	2.691	23:20
Air Blank	0.000	23:20
Control Test	0.083	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

---

Data Download Date/Time: 10/24/2019 21:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOLDEN,RYAN,L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: HARTBARGER,WESLEY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: FLAHERTY

DATE/TIME OF TEST: 10/12/2019 19:12:58

---

Diagnostics Check		OK	19:01	
Air Blank		0.000	19:01	
Control Test		0.083	19:01	
Air Blank		0.000	19:02	
Subject Sample	# 1	0.214	19:05	Volume Not Met
Breath Volume		0.152	19:05	
Air Blank		0.000	19:06	
Air Blank		0.000	19:08	
Subject Sample	# 2	0.176	19:11	Volume Not Met
Breath Volume		0.039	19:11	
Air Blank		0.000	19:12	
Control Test		0.081	19:12	
Air Blank		0.000	19:12	
Diagnostics Check		OK	19:12	

---

Data Download Date/Time: 10/24/2019 21:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOGGAN,MARLON,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: BOGGAN,MARLON  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: HUNTINGTON

DATE/TIME OF TEST: 10/14/2019 05:46:47

---

Diagnostics Check	OK	05:40
Air Blank	0.000	05:41
Control Test	0.083	05:41
Air Blank	0.000	05:42
Subject Sample # 1	0.063	05:42
Breath Volume	2.511	05:42
Air Blank	0.000	05:43
Air Blank	0.000	05:44
Subject Sample # 2	0.062	05:45
Breath Volume	1.824	05:45
Air Blank	0.000	05:45
Control Test	0.083	05:46
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

---

Data Download Date/Time: 10/24/2019 21:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLEMENTS,WILIAM,J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CLEMENTS,WILIAM  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 10/15/2019 15:52:40

---

Diagnostics Check	OK	15:45
Air Blank	0.000	15:46
Control Test	0.083	15:46
Air Blank	0.000	15:46
Subject Sample # 1	0.143	15:48
Breath Volume	1.996	15:48
Air Blank	0.000	15:48
Air Blank	0.000	15:50
Subject Sample # 2	0.141	15:51
Breath Volume	2.035	15:51
Air Blank	0.000	15:51
Control Test	0.082	15:52
Air Blank	0.000	15:52
Diagnostics Check	OK	15:52

---

Data Download Date/Time: 10/24/2019 21:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURKE,MATTHEW,S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: BURKE,MATTHEW  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: HOBBS

DATE/TIME OF TEST: 10/16/2019 00:31:59

---

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:25	
Control Test	0.083	00:25	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:29	No Sample Provided
Breath Volume	0.000	00:29	
Air Blank	0.000	00:29	
Air Blank		00:31	Sequence Aborted
Air Blank	0.000	00:31	

---

Data Download Date/Time: 10/24/2019 21:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND II,JACK,W  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: HOLLAND II,JACK  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:46

SUBJECT'S LAST NAME: NAQUIN III

DATE/TIME OF TEST: 10/17/2019 16:15:18

---

Diagnostics Check	OK	16:11	
Air Blank	0.000	16:12	
Control Test	0.083	16:12	
Air Blank	0.000	16:12	
Subject Sample	# 1	16:13	Subject Test Refused
Breath Volume	0.000	16:13	
Air Blank	0.000	16:14	
Control Test	0.083	16:14	
Air Blank	0.000	16:15	
Diagnostics Check	OK	16:15	

---

Data Download Date/Time: 10/24/2019 21:12



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOGGAN,MARLON,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: BOGGAN,MARLON  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 10/18/2019 04:54:06

---

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:51	
Control Test	0.084	04:52	
Air Blank	0.000	04:52	
Subject Sample	# 1	04:52	Subject Test Refused
Breath Volume	0.000	04:52	
Air Blank	0.000	04:53	
Control Test	0.083	04:53	
Air Blank	0.000	04:54	
Diagnostics Check	OK	04:54	

---

Data Download Date/Time: 10/24/2019 21:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DESOUSA,CARLOS,J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: DESOUSA,CARLOS  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: FITE

DATE/TIME OF TEST: 10/19/2019 02:13:38

---

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.083	02:10	
Air Blank	0.000	02:10	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:12	
Control Test	0.084	02:13	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

---

Data Download Date/Time: 10/24/2019 21:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOGGAN,ARLON,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: BOGGAN,MARLON  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 10/19/2019 04:51:41

---

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.083	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.243	04:47
Breath Volume	1.882	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:49
Subject Sample # 2	0.229	04:50
Breath Volume	2.332	04:50
Air Blank	0.000	04:50
Control Test	0.081	04:51
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

---

Data Download Date/Time: 10/24/2019 21:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOGGAN,MARLON,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: BOGGAN,MARLON  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 10/20/2019 03:46:38

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.082	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.178	03:41
Breath Volume	1.167	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.172	03:45
Breath Volume	1.855	03:45
Air Blank	0.000	03:45
Control Test	0.081	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

---

Data Download Date/Time: 10/24/2019 21:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOYD,J,J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: BOYD,J  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: LAPP

DATE/TIME OF TEST: 10/20/2019 04:28:30

---

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:17	
Control Test	0.082	04:17	
Air Blank	0.000	04:18	
Subject Sample	# 1	04:21	No Sample Provided
Breath Volume	0.000	04:21	
Air Blank	0.000	04:21	
Air Blank	0.000	04:24	
Subject Sample	# 2	04:27	No Sample Provided
Breath Volume	0.000	04:27	
Air Blank	0.000	04:27	
Control Test	0.083	04:27	
Air Blank	0.000	04:28	
Diagnostics Check	OK	04:28	

---

Data Download Date/Time: 10/24/2019 21:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOYD,J,J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: BURGETT,D  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 10/20/2019 05:06:04

---

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.083	05:00
Air Blank	0.000	05:01
Subject Sample # 1	0.118	05:01
Breath Volume	3.003	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.119	05:04
Breath Volume	3.457	05:04
Air Blank	0.000	05:05
Control Test	0.083	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:06

---

Data Download Date/Time: 10/24/2019 21:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CHAPMAN,MAVERICK  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 10/21/2019 17:38:58

---

Diagnostics Check	OK	17:31
Air Blank	0.000	17:31
Control Test	0.082	17:31
Air Blank	0.000	17:32
Subject Sample # 1	0.263	17:33
Breath Volume	1.316	17:33
Air Blank	0.000	17:33
Air Blank	0.000	17:35
Subject Sample # 2	0.246	17:37
Breath Volume	1.281	17:37
Air Blank	0.000	17:38
Control Test	0.082	17:38
Air Blank	0.000	17:38
Diagnostics Check	OK	17:38

---

Data Download Date/Time: 10/24/2019 21:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND II,JACK,W  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: HOLLAND II,JACK  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: TISDALE

DATE/TIME OF TEST: 10/23/2019 22:53:06

---

Diagnostics Check	OK	22:50	
Air Blank	0.000	22:50	
Control Test	0.083	22:50	
Air Blank	0.000	22:51	
Subject Sample	# 1	22:51	Subject Test Refused
Breath Volume	0.000	22:51	
Air Blank	0.000	22:52	
Control Test	0.084	22:52	
Air Blank	0.000	22:53	
Diagnostics Check	OK	22:53	

---

Data Download Date/Time: 10/24/2019 21:13



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBINSON,STUART,P  
BREATH TEST OPERATOR AGENCY: Lake City PD  
ARRESTING OFFICER: MILLIGAN,CONNIE  
ARRESTING OFFICER AGENCY: Lake City PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: HAMPTON

DATE/TIME OF TEST: 10/05/2019 02:24:05

---

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:15	
Control Test	0.079	02:15	
Air Blank	0.000	02:16	
Subject Sample	# 1 0.299	02:19	Volume Not Met
Breath Volume	0.082	02:19	
Air Blank	0.000	02:20	
Air Blank	0.000	02:22	
Subject Sample	# 2	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:23	
Control Test	0.077	02:23	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

---

Data Download Date/Time: 11/12/2019 07:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,BRETT,L  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: ELLIOTT,NIGEL  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: WITT

DATE/TIME OF TEST: 10/06/2019 02:30:10

---

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:27	
Control Test	0.080	02:27	
Air Blank	0.000	02:28	
Subject Sample # 1	0.198	02:28	
Breath Volume	2.589	02:28	
Air Blank		02:28	RFI Detect
Air Blank	0.000	02:29	
Control Test	0.075	02:29	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

---

Data Download Date/Time: 11/12/2019 07:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,BRETT,L  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: ELLIOTT,NIGEL  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: WITT

DATE/TIME OF TEST: 10/06/2019 02:37:47

---

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.079	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.200	02:33
Breath Volume	2.023	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.198	02:36
Breath Volume	2.074	02:36
Air Blank	0.000	02:36
Control Test	0.080	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

---

Data Download Date/Time: 11/12/2019 07:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,DANIEL,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MARTIN,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: FONTENOT

DATE/TIME OF TEST: 10/20/2019 18:38:40

---

Diagnostics Check	OK	18:31
Air Blank	0.000	18:31
Control Test	0.079	18:32
Air Blank	0.000	18:32
Subject Sample # 1	0.014	18:33
Breath Volume	1.789	18:33
Air Blank	0.000	18:34
Air Blank	0.000	18:36
Subject Sample # 2	0.012	18:37
Breath Volume	2.195	18:37
Air Blank	0.000	18:37
Control Test	0.077	18:38
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

---

Data Download Date/Time: 11/12/2019 07:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DURR,KYLE,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BAURES,BRANDYN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 10/21/2019 02:03:15

---

Diagnostics Check		OK	01:56
Air Blank		0.000	01:57
Control Test		0.080	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.208	01:58
Breath Volume		2.382	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.221	02:01
Breath Volume		1.835	02:01
Air Blank		0.000	02:02
Control Test		0.079	02:02
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

---

Data Download Date/Time: 11/12/2019 07:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CREWS,C,L  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: ELLIOT,N  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: DICAPNA

DATE/TIME OF TEST: 10/23/2019 22:19:47

---

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:15	
Control Test	0.079	22:16	
Air Blank	0.000	22:16	
Subject Sample	# 1	22:18	Subject Test Refused
Breath Volume	0.000	22:18	
Air Blank	0.000	22:18	
Control Test	0.078	22:19	
Air Blank	0.000	22:19	
Diagnostics Check	OK	22:19	

---

Data Download Date/Time: 11/12/2019 07:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBINSON,STUART,P  
BREATH TEST OPERATOR AGENCY: Lake City PD  
ARRESTING OFFICER: ROBINSON,STUART  
ARRESTING OFFICER AGENCY: Lake City PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: SNEED

DATE/TIME OF TEST: 10/23/2019 23:37:06

---

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.193	23:32
Breath Volume	1.519	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:34
Subject Sample # 2	0.186	23:35
Breath Volume	1.476	23:35
Air Blank	0.000	23:36
Control Test	0.077	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

---

Data Download Date/Time: 11/12/2019 07:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARVEY,DAVID,A  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: HARVEY,DAVID  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: SANTANGELO

DATE/TIME OF TEST: 10/27/2019 02:53:26

---

Diagnostics Check	OK	02:46	
Air Blank	0.000	02:47	
Control Test	0.080	02:47	
Air Blank	0.000	02:47	
Subject Sample	# 1 0.162	02:48	
Breath Volume	2.054	02:48	
Air Blank	0.000	02:49	
Air Blank	0.000	02:51	
Subject Sample	# 2 0.158	02:51	
Breath Volume	3.230	02:51	
Air Blank	0.000	02:52	
Control Test	0.077	02:52	
Air Blank		02:52	RFI Detect
Air Blank	0.000	02:53	

---

Data Download Date/Time: 11/12/2019 07:13



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BALDREE,TASHAHNDA,J  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BALDREE,TASHAHNDA  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 10/29/2019 03:51:30

---

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.196	03:46
Breath Volume	2.355	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.182	03:49
Breath Volume	3.511	03:49
Air Blank	0.000	03:50
Control Test	0.078	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

---

Data Download Date/Time: 11/12/2019 07:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CREWS,MICHAEL,B  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: CREWS,MICHAEL  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: HOBBY

DATE/TIME OF TEST: 10/30/2019 22:45:46

---

Diagnostics Check	OK	22:39
Air Blank	0.000	22:39
Control Test	0.079	22:40
Air Blank	0.000	22:40
Subject Sample # 1	0.166	22:41
Breath Volume	2.039	22:41
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.166	22:44
Breath Volume	1.964	22:44
Air Blank	0.000	22:44
Control Test	0.079	22:45
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

---

Data Download Date/Time: 11/12/2019 07:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CREWS,MICHAEL,B  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: MONTEMURNO,MATT  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: RHODEN

DATE/TIME OF TEST: 10/31/2019 04:05:55

---

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.080	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.269	04:01
Breath Volume	2.437	04:01
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.261	04:04
Breath Volume	2.238	04:04
Air Blank	0.000	04:04
Control Test	0.079	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

---

Data Download Date/Time: 11/12/2019 07:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUSGROVE,DUSTIN,K  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MUSGROVE,DUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 09/08/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: GILLYARD

DATE/TIME OF TEST: 10/12/2019 22:49:36

---

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.082	22:44
Air Blank	0.000	22:44
Subject Sample # 1	0.230	22:45
Breath Volume	3.425	22:45
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.227	22:47
Breath Volume	2.855	22:47
Air Blank	0.000	22:48
Control Test	0.080	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

---

Data Download Date/Time: 11/04/2019 13:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BATES,BRIAN,C  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BATES,BRIAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 09/08/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: CHANEY

DATE/TIME OF TEST: 10/19/2019 16:36:07

---

Diagnostics Check	OK	16:33	
Air Blank	0.000	16:33	
Control Test	0.081	16:33	
Air Blank	0.000	16:34	
Subject Sample	# 1	16:34	Subject Test Refused
Breath Volume	0.000	16:34	
Air Blank	0.000	16:35	
Control Test	0.081	16:35	
Air Blank	0.000	16:36	
Diagnostics Check	OK	16:36	

---

Data Download Date/Time: 11/04/2019 13:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MACFADDEN,RYAN,E  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MUSGROVE,DUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 09/08/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 10/22/2019 23:10:07

---

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.082	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.000	23:05
Breath Volume	3.242	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.000	23:08
Breath Volume	2.078	23:08
Air Blank	0.000	23:09
Control Test	0.081	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

---

Data Download Date/Time: 11/04/2019 13:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAURES, BRANDYN, K  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BAURES, BRANDYN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 09/08/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: MACDONALD

DATE/TIME OF TEST: 10/23/2019 23:57:24

---

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.082	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.071	23:52
Breath Volume	2.472	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.069	23:55
Breath Volume	2.121	23:55
Air Blank	0.000	23:56
Control Test	0.083	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

---

Data Download Date/Time: 11/04/2019 13:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FISCHER, MICHAEL, A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: FISCHER, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 09/08/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ROLLING

DATE/TIME OF TEST: 10/26/2019 00:17:17

---

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.082	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.262	00:12
Breath Volume		3.093	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.261	00:15
Breath Volume		2.820	00:15
Air Blank		0.000	00:16
Control Test		0.080	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

---

Data Download Date/Time: 11/04/2019 13:35



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUSGROVE,DUSTIN,K  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MUSGROVE,DUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 09/08/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: CECIL

DATE/TIME OF TEST: 10/27/2019 19:13:16

---

Diagnostics Check	OK	19:06
Air Blank	0.000	19:07
Control Test	0.083	19:07
Air Blank	0.000	19:08
Subject Sample # 1	0.087	19:08
Breath Volume	1.960	19:08
Air Blank	0.000	19:09
Air Blank	0.000	19:11
Subject Sample # 2	0.083	19:11
Breath Volume	1.812	19:11
Air Blank	0.000	19:12
Control Test	0.080	19:12
Air Blank	0.000	19:13
Diagnostics Check	OK	19:13

---

Data Download Date/Time: 11/04/2019 13:35





**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAWYER,DARRYL,B  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SAWYER,DARRYL  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/07/2020  
VIOLATION CODE: DUI  
ARREST TIME: 10:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:28

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 10/06/2019 11:56:10

---

Diagnostics Check	OK	11:49
Air Blank	0.000	11:49
Control Test	0.083	11:50
Air Blank	0.000	11:50
Subject Sample # 1	0.204	11:51
Breath Volume	3.269	11:51
Air Blank	0.000	11:52
Air Blank	0.000	11:53
Subject Sample # 2	0.201	11:54
Breath Volume	3.273	11:54
Air Blank	0.000	11:55
Control Test	0.080	11:55
Air Blank	0.000	11:56
Diagnostics Check	OK	11:56

---

Data Download Date/Time: 11/05/2019 23:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SISK,CARLOS,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SISK,CARLOS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/07/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: CIBOROWSKI

DATE/TIME OF TEST: 10/08/2019 21:42:11

---

Diagnostics Check		OK	21:27	
Air Blank		0.000	21:28	
Control Test		0.083	21:28	
Air Blank		0.000	21:29	
Subject Sample	# 1	0.203	21:32	Volume Not Met
Breath Volume		0.097	21:32	
Air Blank		0.000	21:33	
Air Blank		0.000	21:34	
Subject Sample	# 2	0.191	21:35	No .020 agreement
Breath Volume		1.375	21:35	
Air Blank		0.000	21:36	
Air Blank		0.000	21:38	
Subject Sample	# 3	0.162	21:40	No .020 agreement
Breath Volume		1.566	21:40	
Air Blank		0.000	21:41	
Control Test		0.075	21:41	
Air Blank		0.000	21:42	
Diagnostics Check		OK	21:42	

---

Data Download Date/Time: 11/05/2019 23:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SISK,CARLOS,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SISK,CARLOS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/07/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: NOLAN

DATE/TIME OF TEST: 10/13/2019 16:08:37

---

Diagnostics Check	OK	16:01
Air Blank	0.000	16:02
Control Test	0.083	16:02
Air Blank	0.000	16:02
Subject Sample # 1	0.167	16:03
Breath Volume	4.250	16:03
Air Blank	0.000	16:04
Air Blank	0.000	16:05
Subject Sample # 2	0.177	16:06
Breath Volume	4.613	16:06
Air Blank	0.000	16:07
Control Test	0.080	16:07
Air Blank	0.000	16:08
Diagnostics Check	OK	16:08

---

Data Download Date/Time: 11/05/2019 23:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YASUDA,SHOHEI,  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: BROWN,DAVE  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/01/2019

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/07/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:08

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 10/17/2019 15:32:31

---

Diagnostics Check	OK	15:29	
Air Blank	0.000	15:30	
Control Test	0.083	15:30	
Air Blank	0.000	15:30	
Subject Sample	# 1	15:31	Subject Test Refused
Breath Volume	0.000	15:31	
Air Blank	0.000	15:31	
Control Test	0.083	15:31	
Air Blank	0.000	15:32	
Diagnostics Check	OK	15:32	

---

Data Download Date/Time: 11/05/2019 23:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAGNER,BRETT,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: WAGNER,BRETT  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/07/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 10/17/2019 22:10:07

---

Diagnostics Check	OK	22:07	
Air Blank	0.000	22:07	
Control Test	0.083	22:07	
Air Blank	0.000	22:08	
Subject Sample	# 1	22:08	RFI Detect
Breath Volume	0.000	22:08	
Air Blank	0.000	22:09	
Control Test	0.084	22:09	
Air Blank	0.000	22:10	
Diagnostics Check	OK	22:10	

---

Data Download Date/Time: 11/05/2019 23:53



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAGNER,BRETT,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: WAGNER,BRETT  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/07/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 10/17/2019 22:30:42

---

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:21	
Control Test	0.083	22:21	
Air Blank	0.000	22:22	
Subject Sample # 1	0.162	22:23	
Breath Volume	1.484	22:23	
Air Blank	0.000	22:24	
Air Blank	0.000	22:26	
Subject Sample # 2	0.146	22:29	Volume Not Met
Breath Volume	0.050	22:29	
Air Blank		22:30	Purge Fail
Air Blank	0.000	22:30	

---

Data Download Date/Time: 11/05/2019 23:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAGNER,BRETT,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: WAGNER,BRETT  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/07/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 10/17/2019 22:41:21

---

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.076	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.151	22:36
Breath Volume	1.644	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.147	22:39
Breath Volume	1.855	22:39
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

---

Data Download Date/Time: 11/05/2019 23:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RADCLIFF,MICHAEL,R  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: RADCLIFF,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/07/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: DIETZ

DATE/TIME OF TEST: 10/20/2019 20:59:39

---

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:46	
Control Test	0.082	20:47	
Air Blank	0.000	20:47	
Subject Sample	# 1	20:50	No Sample Provided
Breath Volume	0.000	20:50	
Air Blank	0.000	20:51	
Air Blank	0.000	20:53	
Subject Sample	# 2	20:54	
Breath Volume	1.457	20:54	
Air Blank	0.000	20:55	
Air Blank	0.000	20:57	
Subject Sample	# 3	20:58	
Breath Volume	1.347	20:58	
Air Blank		20:59	Purge Fail
Air Blank	0.000	20:59	

---

Data Download Date/Time: 11/05/2019 23:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RADCLIFF,MICHAEL,R  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: RADCLIFF,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/07/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: DIETZ

DATE/TIME OF TEST: 10/20/2019 21:12:32

---

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.076	21:06
Air Blank	0.000	21:06
Subject Sample # 1	0.206	21:07
Breath Volume	1.898	21:07
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 2	0.214	21:10
Breath Volume	1.812	21:10
Air Blank	0.000	21:11
Control Test	0.078	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

---

Data Download Date/Time: 11/05/2019 23:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEIERS,BLAINE,T  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: WEIERS,BLAINE  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 09/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: STEWARD

DATE/TIME OF TEST: 10/02/2019 03:17:17

---

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:06	
Control Test		0.082	03:06	
Air Blank		0.000	03:06	
Subject Sample	# 1	0.171	03:10	Volume Not Met
Breath Volume		0.117	03:10	
Air Blank		0.000	03:10	
Air Blank		0.000	03:12	
Subject Sample	# 2	0.192	03:15	Volume Not Met
Breath Volume		0.191	03:15	
Air Blank		0.000	03:16	
Control Test		0.081	03:16	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:17	

---

Data Download Date/Time: 10/19/2019 03:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMS,KYLE,N  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: WHITE,TRAVIS  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 09/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/03/2019 03:22:01

---

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.081	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.257	03:17
Breath Volume		3.613	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:19
Subject Sample	# 2	0.259	03:20
Breath Volume		3.910	03:20
Air Blank		0.000	03:21
Control Test		0.081	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:22

---

Data Download Date/Time: 10/19/2019 03:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCAPIN,BRANDON,L  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: SCAPIN,BRANDON  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 09/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: BLUE

DATE/TIME OF TEST: 10/04/2019 03:02:36

---

Diagnostics Check	OK	02:59	
Air Blank	0.000	03:00	
Control Test	0.081	03:00	
Air Blank	0.000	03:01	
Subject Sample	# 1	03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:01	
Control Test	0.081	03:02	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

---

Data Download Date/Time: 10/19/2019 03:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEIERS,BLAINE,T  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: WEIERS,BLAINE  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 09/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BUCHALTER

DATE/TIME OF TEST: 10/04/2019 22:26:56

---

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.080	22:21
Air Blank	0.000	22:21
Subject Sample # 1	0.156	22:22
Breath Volume	2.023	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:24
Subject Sample # 2	0.150	22:25
Breath Volume	2.089	22:25
Air Blank	0.000	22:26
Control Test	0.081	22:26
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

---

Data Download Date/Time: 10/19/2019 03:30



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHURCHILL,JON,W  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: HUNTER,JOHN  
ARRESTING OFFICER AGENCY: OTHER Fed  
LAST AGENCY INSPECTION: 09/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: SHORT

DATE/TIME OF TEST: 10/06/2019 01:36:29

---

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.082	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.154	01:31
Breath Volume	4.523	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.151	01:34
Breath Volume	3.984	01:34
Air Blank	0.000	01:35
Control Test	0.081	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

---

Data Download Date/Time: 10/19/2019 03:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEIERS,BLAINE,T  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: LLOYD,GEORGE  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 09/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: RACKLEY

DATE/TIME OF TEST: 10/10/2019 18:17:16

---

Diagnostics Check	OK	18:10
Air Blank	0.000	18:11
Control Test	0.080	18:11
Air Blank	0.000	18:12
Subject Sample # 1	0.171	18:12
Breath Volume	3.316	18:12
Air Blank	0.000	18:13
Air Blank	0.000	18:15
Subject Sample # 2	0.174	18:15
Breath Volume	3.550	18:15
Air Blank	0.000	18:16
Control Test	0.081	18:16
Air Blank	0.000	18:17
Diagnostics Check	OK	18:17

---

Data Download Date/Time: 10/19/2019 03:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMS,KYLE,N  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: SIMMS,KYLE  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 09/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: WOODY

DATE/TIME OF TEST: 10/12/2019 03:26:28

---

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.081	03:24	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Control Test	0.081	03:25	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

---

Data Download Date/Time: 10/19/2019 03:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMS,KYLE,N  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: SIMMS,KYLE  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 09/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: SANDI

DATE/TIME OF TEST: 10/13/2019 03:10:02

---

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.081	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.148	03:05
Breath Volume	1.644	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.155	03:08
Breath Volume	2.492	03:08
Air Blank	0.000	03:09
Control Test	0.081	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:10

---

Data Download Date/Time: 10/19/2019 03:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERSON,THOMAS,C  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: WEIERS,BLAINE  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 10/19/2019

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: BLYTHE

DATE/TIME OF TEST: 10/20/2019 03:58:52

---

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.081	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.240	03:54
Breath Volume	3.777	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:56
Subject Sample # 2	0.235	03:57
Breath Volume	2.742	03:57
Air Blank	0.000	03:57
Control Test	0.079	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

---

Data Download Date/Time: 11/22/2019 17:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BANKS,SEAN,C  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: BANKS,SEAN  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 10/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:01

SUBJECT'S LAST NAME: REEVER

DATE/TIME OF TEST: 10/25/2019 13:28:03

---

Diagnostics Check	OK	13:21
Air Blank	0.000	13:22
Control Test	0.080	13:22
Air Blank	0.000	13:22
Subject Sample # 1	0.000	13:23
Breath Volume	1.351	13:23
Air Blank	0.000	13:24
Air Blank	0.000	13:26
Subject Sample # 2	0.000	13:26
Breath Volume	2.941	13:26
Air Blank	0.000	13:27
Control Test	0.081	13:27
Air Blank	0.000	13:27
Diagnostics Check	OK	13:28

---

Data Download Date/Time: 11/22/2019 17:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITE,TRAVIS,D  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: WHITE,TRAVIS  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 10/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: LOVING

DATE/TIME OF TEST: 10/26/2019 23:59:26

---

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.253	23:54
Breath Volume	4.539	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.271	23:57
Breath Volume	4.441	23:57
Air Blank	0.000	23:58
Control Test	0.080	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

---

Data Download Date/Time: 11/22/2019 17:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHURCHILL,JON,W  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: CHURCHILL,JON  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 10/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 10/29/2019 21:26:59

---

Diagnostics Check	OK	21:20
Air Blank	0.000	21:21
Control Test	0.081	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.040	21:22
Breath Volume	3.339	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.038	21:25
Breath Volume	2.632	21:25
Air Blank	0.000	21:26
Control Test	0.081	21:26
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

---

Data Download Date/Time: 11/22/2019 17:45



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 10/02/2019 01:39:34

---

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.000	01:35
Breath Volume	4.000	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.000	01:38
Breath Volume	3.886	01:38
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

---

Data Download Date/Time: 10/22/2019 15:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: GRAZIANO,STEVEN  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/03/2019 01:54:08

---

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.152	01:47
Breath Volume	1.410	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.156	01:52
Breath Volume	1.265	01:52
Air Blank	0.000	01:53
Control Test	0.078	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

---

Data Download Date/Time: 10/22/2019 15:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: GRAZIANO,STEVEN  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 10/04/2019 05:15:45

---

Diagnostics Check	OK	05:08
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.092	05:11
Breath Volume	3.246	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.096	05:14
Breath Volume	2.320	05:14
Air Blank	0.000	05:14
Control Test	0.079	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

---

Data Download Date/Time: 10/22/2019 15:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HAGANEY

DATE/TIME OF TEST: 10/05/2019 00:34:16

---

Diagnostics Check		OK	00:19	
Air Blank		0.000	00:20	
Control Test		0.080	00:20	
Air Blank		0.000	00:21	
Subject Sample	# 1	0.104	00:24	Volume Not Met
Breath Volume		0.992	00:24	
Air Blank		0.000	00:24	
Air Blank		0.000	00:26	
Subject Sample	# 2	0.123	00:28	
Breath Volume		2.000	00:28	
Air Blank		0.000	00:29	
Air Blank		0.000	00:31	
Subject Sample	# 3	0.119	00:32	
Breath Volume		1.425	00:32	
Air Blank		0.000	00:33	
Control Test		0.079	00:33	
Air Blank		0.000	00:34	
Diagnostics Check		OK	00:34	

---

Data Download Date/Time: 10/22/2019 15:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: GASPAR,BENJAMIN  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 10/05/2019 04:02:47

---

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.080	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.237	03:58
Breath Volume		2.417	03:58
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.244	04:01
Breath Volume		2.609	04:01
Air Blank		0.000	04:01
Control Test		0.080	04:02
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

---

Data Download Date/Time: 10/22/2019 15:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: RUIZ CARMONA

DATE/TIME OF TEST: 10/06/2019 19:00:01

---

Diagnostics Check	OK	18:53
Air Blank	0.000	18:53
Control Test	0.080	18:54
Air Blank	0.000	18:54
Subject Sample # 1	0.145	18:55
Breath Volume	2.800	18:55
Air Blank	0.000	18:56
Air Blank	0.000	18:57
Subject Sample # 2	0.145	18:58
Breath Volume	2.902	18:58
Air Blank	0.000	18:59
Control Test	0.079	18:59
Air Blank	0.000	18:59
Diagnostics Check	OK	19:00

---

Data Download Date/Time: 10/22/2019 15:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: MCINERNEY,JESSE  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: GLENNON

DATE/TIME OF TEST: 10/06/2019 19:48:23

---

Diagnostics Check	OK	19:42
Air Blank	0.000	19:42
Control Test	0.079	19:42
Air Blank	0.000	19:43
Subject Sample # 1	0.262	19:43
Breath Volume	2.320	19:43
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample # 2	0.258	19:46
Breath Volume	2.335	19:46
Air Blank	0.000	19:47
Control Test	0.079	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

---

Data Download Date/Time: 10/22/2019 15:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: ALVARADO MARTINEZ

DATE/TIME OF TEST: 10/07/2019 03:03:21

---

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.241	02:58
Breath Volume	2.492	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.227	03:01
Breath Volume	1.824	03:01
Air Blank	0.000	03:02
Control Test	0.078	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

---

Data Download Date/Time: 10/22/2019 15:09



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: FOSTER,SAMANTHA  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 10/07/2019 20:59:29

---

Diagnostics Check	OK	20:45	
Air Blank	0.000	20:45	
Control Test	0.081	20:45	
Air Blank	0.000	20:46	
Subject Sample # 1	0.121	20:48	
Breath Volume	1.992	20:48	
Air Blank	0.000	20:49	
Air Blank	0.000	20:50	
Subject Sample # 2	0.109	20:54	Volume Not Met
Breath Volume	0.644	20:54	
Air Blank	0.000	20:54	
Air Blank	0.000	20:56	
Subject Sample # 3	0.121	20:57	
Breath Volume	2.531	20:57	
Air Blank	0.000	20:58	
Control Test	0.080	20:58	
Air Blank	0.000	20:59	
Diagnostics Check	OK	20:59	

---

Data Download Date/Time: 10/22/2019 15:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/10/2019 00:29:00

---

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.080	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.239	00:24
Breath Volume	2.648	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:26
Subject Sample # 2	0.240	00:27
Breath Volume	2.593	00:27
Air Blank	0.000	00:28
Control Test	0.079	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:29

---

Data Download Date/Time: 10/22/2019 15:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PREVIL,LUXERNE,  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: PREVIL,LUXERNE  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 10/11/2019 01:50:42

---

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.218	01:46
Breath Volume	1.824	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:48
Subject Sample # 2	0.209	01:49
Breath Volume	1.609	01:49
Air Blank	0.000	01:49
Control Test	0.079	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

---

Data Download Date/Time: 10/22/2019 15:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CASTRILLON

DATE/TIME OF TEST: 10/11/2019 02:11:42

---

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.079	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.066	02:07
Breath Volume	2.914	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.066	02:10
Breath Volume	3.078	02:10
Air Blank	0.000	02:10
Control Test	0.079	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

---

Data Download Date/Time: 10/22/2019 15:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: COLSON

DATE/TIME OF TEST: 10/16/2019 02:56:17

---

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.000	02:51
Breath Volume	2.964	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.000	02:54
Breath Volume	2.953	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

---

Data Download Date/Time: 10/22/2019 15:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SELPH,JASON,H  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: JONES,CORTLAND  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:32

SUBJECT'S LAST NAME: LASALA

DATE/TIME OF TEST: 10/16/2019 10:58:06

---

Diagnostics Check	OK	10:52
Air Blank	0.000	10:52
Control Test	0.079	10:52
Air Blank	0.000	10:53
Subject Sample # 1	0.000	10:53
Breath Volume	2.476	10:53
Air Blank	0.000	10:54
Air Blank	0.000	10:56
Subject Sample # 2	0.000	10:56
Breath Volume	1.554	10:56
Air Blank	0.000	10:57
Control Test	0.080	10:57
Air Blank	0.000	10:58
Diagnostics Check	OK	10:58

---

Data Download Date/Time: 10/22/2019 15:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: RODRIGUEZ RUBIO

DATE/TIME OF TEST: 10/18/2019 22:38:37

---

Diagnostics Check	OK	22:32
Air Blank	0.000	22:32
Control Test	0.079	22:33
Air Blank	0.000	22:33
Subject Sample # 1	0.224	22:34
Breath Volume	1.304	22:34
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.218	22:37
Breath Volume	1.785	22:37
Air Blank	0.000	22:37
Control Test	0.079	22:38
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

---

Data Download Date/Time: 10/22/2019 15:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 10/19/2019 03:23:03

---

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.296	03:18
Breath Volume		2.273	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:20
Subject Sample	# 2	0.296	03:21
Breath Volume		2.328	03:21
Air Blank		0.000	03:22
Control Test		0.080	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:23

---

Data Download Date/Time: 10/22/2019 15:09



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 10/20/2019 21:07:59

---

Diagnostics Check	OK	21:01
Air Blank	0.000	21:02
Control Test	0.080	21:02
Air Blank	0.000	21:02
Subject Sample # 1	0.178	21:03
Breath Volume	3.019	21:03
Air Blank	0.000	21:04
Air Blank	0.000	21:05
Subject Sample # 2	0.179	21:06
Breath Volume	3.160	21:06
Air Blank	0.000	21:07
Control Test	0.081	21:07
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

---

Data Download Date/Time: 10/22/2019 15:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MIANO,MELISSA,A  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: MIANO,MELISSA  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 20:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 10/20/2019 21:45:11

---

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.079	21:39
Air Blank	0.000	21:40
Subject Sample # 1	0.152	21:40
Breath Volume	3.199	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.152	21:43
Breath Volume	2.640	21:43
Air Blank	0.000	21:44
Control Test	0.079	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

---

Data Download Date/Time: 10/22/2019 15:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: LAMPLOUGH

DATE/TIME OF TEST: 10/23/2019 22:16:05

---

Diagnostics Check	OK	22:09
Air Blank	0.000	22:10
Control Test	0.080	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.052	22:11
Breath Volume	2.636	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:13
Subject Sample # 2	0.050	22:14
Breath Volume	2.847	22:14
Air Blank	0.000	22:15
Control Test	0.080	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:16

---

Data Download Date/Time: 11/20/2019 14:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOODRUFF,MARK,D  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: FRASIER,MICHAEL  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 10/22/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: GATES

DATE/TIME OF TEST: 10/24/2019 23:12:34

---

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.081	23:07
Air Blank	0.000	23:07
Subject Sample # 1	0.122	23:08
Breath Volume	2.058	23:08
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.121	23:11
Breath Volume	2.039	23:11
Air Blank	0.000	23:11
Control Test	0.080	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

---

Data Download Date/Time: 11/20/2019 14:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUCKLES,SCHIEFER,S  
BREATH TEST OPERATOR AGENCY: FL Dept of AG  
ARRESTING OFFICER: BUCKLES,SCHIEFER  
ARRESTING OFFICER AGENCY: FL Dept of AG  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: HYATT

DATE/TIME OF TEST: 10/08/2019 00:49:48

---

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.257	00:45
Breath Volume	1.156	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:47
Subject Sample # 2	0.257	00:48
Breath Volume	1.109	00:48
Air Blank	0.000	00:48
Control Test	0.080	00:49
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

---

Data Download Date/Time: 11/27/2019 11:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HALL,L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: LASH

DATE/TIME OF TEST: 10/01/2019 21:11:05

---

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:01	
Control Test	0.079	21:01	
Air Blank	0.000	21:02	
Subject Sample	# 1	21:05	No Sample Provided
Breath Volume	0.000	21:05	
Air Blank	0.000	21:05	
Air Blank	0.000	21:07	
Subject Sample	# 2	21:09	Subject Test Refused
Breath Volume	0.000	21:09	
Air Blank	0.000	21:10	
Control Test	0.079	21:10	
Air Blank	0.000	21:11	
Diagnostics Check	OK	21:11	

---

Data Download Date/Time: 10/10/2019 14:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DE JESUS,V  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: KOUCHEKIAN

DATE/TIME OF TEST: 10/03/2019 03:33:02

---

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:22	
Control Test	0.079	03:22	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:26	No Sample Provided
Breath Volume	0.000	03:26	
Air Blank	0.000	03:26	
Air Blank	0.000	03:28	
Subject Sample	# 2	03:31	No Sample Provided
Breath Volume	0.000	03:31	
Air Blank	0.000	03:32	
Control Test	0.079	03:32	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:33	

---

Data Download Date/Time: 10/10/2019 14:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LYON,M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 10/04/2019 03:21:07

---

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.079	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.132	03:16
Breath Volume	1.296	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.139	03:19
Breath Volume	1.734	03:19
Air Blank	0.000	03:20
Control Test	0.079	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

---

Data Download Date/Time: 10/10/2019 14:25



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GARRELL,Q  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: BEERS

DATE/TIME OF TEST: 10/04/2019 19:01:48

---

Diagnostics Check	OK	18:55
Air Blank	0.000	18:55
Control Test	0.081	18:56
Air Blank	0.000	18:56
Subject Sample # 1	0.000	18:57
Breath Volume	2.195	18:57
Air Blank	0.000	18:57
Air Blank	0.000	18:59
Subject Sample # 2	0.000	19:00
Breath Volume	1.792	19:00
Air Blank	0.000	19:00
Control Test	0.079	19:01
Air Blank	0.000	19:01
Diagnostics Check	OK	19:01

---

Data Download Date/Time: 10/10/2019 14:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HALL,L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: WOODCOCK

DATE/TIME OF TEST: 10/05/2019 03:06:41

---

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.181	03:01
Breath Volume	1.527	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.178	03:05
Breath Volume	1.605	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

---

Data Download Date/Time: 10/10/2019 14:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SCHAEFER,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: COMINGS GARCIA

DATE/TIME OF TEST: 10/05/2019 04:44:11

---

Diagnostics Check		OK	04:32	
Air Blank		0.000	04:33	
Control Test		0.079	04:33	
Air Blank		0.000	04:33	
Subject Sample	# 1	0.054	04:37	Volume Not Met
Breath Volume		0.250	04:37	
Air Blank		0.000	04:37	
Air Blank		0.000	04:39	
Subject Sample	# 2	0.062	04:42	Volume Not Met
Breath Volume		0.449	04:42	
Air Blank		0.000	04:43	
Control Test		0.078	04:43	
Air Blank		0.000	04:44	
Diagnostics Check		OK	04:44	

---

Data Download Date/Time: 10/10/2019 14:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SCHAEFER,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: COMINGS GARCIA

DATE/TIME OF TEST: 10/05/2019 05:06:56

---

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.079	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.065	05:00
Breath Volume	2.253	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.064	05:05
Breath Volume	2.066	05:05
Air Blank	0.000	05:06
Control Test	0.080	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

---

Data Download Date/Time: 10/10/2019 14:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN TREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: LOPEZ-MARTINEZ

DATE/TIME OF TEST: 10/06/2019 03:25:33

---

Diagnostics Check		OK	03:17
Air Blank		0.000	03:17
Control Test		0.079	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.151	03:20
Breath Volume		1.687	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:22
Subject Sample	# 2	0.148	03:23
Breath Volume		1.234	03:23
Air Blank		0.000	03:24
Control Test		0.077	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

---

Data Download Date/Time: 10/10/2019 14:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WASHINGTON,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 10/06/2019 04:44:02

---

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.080	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.185	04:39
Breath Volume	1.613	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:41
Subject Sample # 2	0.189	04:42
Breath Volume	1.773	04:42
Air Blank	0.000	04:43
Control Test	0.078	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:44

---

Data Download Date/Time: 10/10/2019 14:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DEJSUS,V  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 10/08/2019 03:43:14

---

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.079	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.224	03:38
Breath Volume	2.285	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.204	03:41
Breath Volume	1.355	03:41
Air Blank	0.000	03:42
Control Test	0.078	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

---

Data Download Date/Time: 10/10/2019 14:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: REPLOGLE

DATE/TIME OF TEST: 10/08/2019 04:16:21

---

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:11	
Control Test	0.079	04:11	
Air Blank	0.000	04:11	
Subject Sample	# 1	04:14	Subject Test Refused
Breath Volume	0.000	04:14	
Air Blank	0.000	04:15	
Control Test	0.079	04:15	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

---

Data Download Date/Time: 10/10/2019 14:25



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: WALSH,SHANNON  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: BEEDLE

DATE/TIME OF TEST: 10/17/2019 22:52:47

---

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.080	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.070	22:47
Breath Volume	1.914	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.070	22:51
Breath Volume	1.980	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

---

Data Download Date/Time: 10/23/2019 23:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: PITCHER,DARRYL  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 10/26/2019 02:21:30

---

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.078	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.082	02:17
Breath Volume	2.167	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.082	02:20
Breath Volume	2.859	02:20
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

---

Data Download Date/Time: 11/09/2019 01:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,P  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,BRONSON  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 10/27/2019 02:00:33

---

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.079	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:55	No Sample Provided
Breath Volume	0.000	01:55	
Air Blank	0.000	01:56	
Air Blank	0.000	01:58	
Subject Sample	# 2	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	01:59	
Control Test	0.079	01:59	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

---

Data Download Date/Time: 11/09/2019 01:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,P  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANDI  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: KEY

DATE/TIME OF TEST: 10/27/2019 03:26:10

---

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:25	
Control Test	0.077	03:25	Interferent Detect
Air Blank	0.000	03:26	

---

Data Download Date/Time: 11/09/2019 01:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,P  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANDI  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: KEY

DATE/TIME OF TEST: 10/27/2019 03:39:11

---

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:28	
Control Test	0.077	03:28	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:32	No Sample Provided
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Air Blank	0.000	03:34	
Subject Sample	# 2	03:37	No Sample Provided
Breath Volume	0.000	03:37	
Air Blank	0.000	03:38	
Control Test	0.078	03:38	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

---

Data Download Date/Time: 11/09/2019 01:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,ETHAN,C  
BREATH TEST OPERATOR AGENCY: Clermont PD  
ARRESTING OFFICER: BELL,ETHAN  
ARRESTING OFFICER AGENCY: Clermont PD  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1099238  
DRY GAS EXPIRY DATE: 03/01/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: DESMOND

DATE/TIME OF TEST: 10/02/2019 23:24:49

---

Diagnostics Check	OK	23:18
Air Blank	0.000	23:19
Control Test	0.077	23:19
Air Blank	0.000	23:19
Subject Sample # 1	0.047	23:20
Breath Volume	1.355	23:20
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.048	23:23
Breath Volume	1.832	23:23
Air Blank	0.000	23:23
Control Test	0.077	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

---

Data Download Date/Time: 10/23/2019 23:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,BRONSON  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1099238  
DRY GAS EXPIRY DATE: 03/01/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 10/10/2019 01:02:54

---

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.077	00:57
Air Blank	0.000	00:57
Subject Sample # 1	0.151	00:58
Breath Volume	1.355	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.159	01:01
Breath Volume	1.492	01:01
Air Blank	0.000	01:01
Control Test	0.076	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

---

Data Download Date/Time: 10/23/2019 23:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANDRA  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1099238  
DRY GAS EXPIRY DATE: 03/01/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: BOROVY

DATE/TIME OF TEST: 10/17/2019 00:13:40

---

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:09	
Control Test	0.076	00:09	
Air Blank	0.000	00:10	
Subject Sample	# 1	00:12	Subject Test Refused
Breath Volume	0.000	00:12	
Air Blank	0.000	00:12	
Control Test	0.077	00:13	
Air Blank	0.000	00:13	
Diagnostics Check	OK	00:13	

---

Data Download Date/Time: 10/23/2019 23:18



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,BRONSON  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1099238  
DRY GAS EXPIRY DATE: 03/01/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SIKES

DATE/TIME OF TEST: 10/17/2019 03:42:18

---

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:30	
Control Test	0.077	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1 0.171	03:34	Volume Not Met
Breath Volume	0.253	03:34	
Air Blank	0.000	03:35	
Air Blank	0.000	03:36	
Subject Sample	# 2 0.195	03:37	
Breath Volume	1.718	03:37	
Air Blank	0.000	03:38	
Air Blank	0.000	03:40	
Subject Sample	# 3 0.194	03:40	
Breath Volume	2.062	03:40	
Air Blank	0.000	03:41	
Control Test	0.074	03:41	Control Outside Tolerance
Air Blank	0.000	03:42	

---

Data Download Date/Time: 10/23/2019 23:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,BRONSON  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1099238  
DRY GAS EXPIRY DATE: 03/01/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SIKES

DATE/TIME OF TEST: 10/17/2019 03:50:24

---

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.075	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.186	03:45
Breath Volume	1.496	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.182	03:48
Breath Volume	1.367	03:48
Air Blank	0.000	03:49
Control Test	0.076	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

---

Data Download Date/Time: 10/23/2019 23:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANDRA  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1099238  
DRY GAS EXPIRY DATE: 03/01/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: PEEPLES

DATE/TIME OF TEST: 10/17/2019 23:22:31

---

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:11	
Control Test	0.078	23:11	
Air Blank	0.000	23:11	
Subject Sample	# 1	23:15	No Sample Provided
Breath Volume	0.000	23:15	
Air Blank	0.000	23:15	
Air Blank	0.000	23:17	
Subject Sample	# 2	0.140	23:17
Breath Volume	3.140	23:17	
Air Blank	0.000	23:18	
Air Blank	0.000	23:20	
Subject Sample	# 3	0.141	23:20
Breath Volume	3.152	23:20	
Air Blank	0.000	23:21	
Control Test	0.077	23:21	
Air Blank	0.000	23:22	
Diagnostics Check	OK	23:22	

---

Data Download Date/Time: 10/23/2019 23:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: PITCHER,DARRYL  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1099238  
DRY GAS EXPIRY DATE: 03/01/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 10/18/2019 01:34:55

---

Diagnostics Check		OK	01:23	
Air Blank		0.000	01:23	
Control Test		0.078	01:23	
Air Blank		0.000	01:24	
Subject Sample	# 1	0.121	01:27	Volume Not Met
Breath Volume		0.074	01:27	
Air Blank		0.000	01:28	
Air Blank		0.000	01:30	
Subject Sample	# 2	0.177	01:33	Volume Not Met
Breath Volume		0.089	01:33	
Air Blank		0.000	01:33	
Control Test		0.077	01:34	
Air Blank		0.000	01:34	
Diagnostics Check		OK	01:34	

---

Data Download Date/Time: 10/23/2019 23:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: PITCHER,DARRYL  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1099238  
DRY GAS EXPIRY DATE: 03/01/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 10/18/2019 01:54:33

---

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:38	
Control Test		0.076	01:38	
Air Blank		0.000	01:39	
Subject Sample	# 1	0.159	01:42	Volume Not Met
Breath Volume		0.246	01:42	
Air Blank		0.000	01:43	
Air Blank		0.000	01:45	
Subject Sample	# 2	0.186	01:47	
Breath Volume		1.718	01:47	
Air Blank		0.000	01:48	
Air Blank		0.000	01:49	
Subject Sample	# 3	0.150	01:52	Volume Not Met
Breath Volume		0.246	01:52	
Air Blank		0.000	01:53	
Control Test		0.076	01:53	
Air Blank		0.000	01:54	
Diagnostics Check		OK	01:54	

---

Data Download Date/Time: 10/23/2019 23:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: PITCHER,DARRYL  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1099238  
DRY GAS EXPIRY DATE: 03/01/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 10/18/2019 02:02:29

---

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:01	
Control Test	0.074	02:01	Control Outside Tolerance
Air Blank	0.000	02:02	

---

Data Download Date/Time: 10/23/2019 23:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: PITCHER,DARRYL  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1099238  
DRY GAS EXPIRY DATE: 03/01/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 10/18/2019 02:14:02

---

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.076	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.168	02:07
Breath Volume	1.191	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:09
Subject Sample # 2	0.182	02:12
Breath Volume	1.785	02:12
Air Blank	0.000	02:13
Control Test	0.077	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:14

---

Data Download Date/Time: 10/23/2019 23:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHESSHER,SANDRA,W  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,B  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1099238  
DRY GAS EXPIRY DATE: 03/01/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 10/24/2019 03:08:26

---

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.077	03:03	
Air Blank	0.000	03:04	
Subject Sample # 1	0.172	03:04	
Breath Volume	3.542	03:04	
Air Blank	0.000	03:05	
Air Blank		03:07	RFI Detect
Air Blank	0.000	03:07	
Control Test	0.078	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

---

Data Download Date/Time: 11/09/2019 01:58



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHESSHER,SANDRA,W  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,B  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1099238  
DRY GAS EXPIRY DATE: 03/01/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 10/24/2019 03:16:24

---

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.077	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.168	03:11
Breath Volume	2.847	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.163	03:14
Breath Volume	3.253	03:14
Air Blank	0.000	03:15
Control Test	0.076	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

---

Data Download Date/Time: 11/09/2019 01:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOMMERSDORF,JOHN,G  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: REBANDO,BRIAN  
ARRESTING OFFICER AGENCY: Clermont PD  
LAST AGENCY INSPECTION: 09/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A2  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: AVILES

DATE/TIME OF TEST: 10/03/2019 02:05:46

---

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:03	
Control Test	0.080	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Control Test	0.080	02:05	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

---

Data Download Date/Time: 10/17/2019 04:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON,MARK,S  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: WHITE,DONALD  
ARRESTING OFFICER AGENCY: Mount Dora PD  
LAST AGENCY INSPECTION: 09/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A2  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 10/04/2019 02:18:32

---

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.076	02:11	
Air Blank	0.000	02:12	
Subject Sample # 1	0.179	02:13	
Breath Volume	2.144	02:13	
Air Blank	0.000	02:14	
Air Blank	0.000	02:16	
Subject Sample # 2	0.175	02:17	
Breath Volume	2.175	02:17	
Air Blank	0.000	02:17	
Control Test	0.072	02:17	Control Outside Tolerance
Air Blank	0.000	02:18	

---

Data Download Date/Time: 10/17/2019 04:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALSH,SHANNON,L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: WALSH,S

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 09/16/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 34917080A2

DRY GAS EXPIRY DATE: 02/05/2020

VIOLATION CODE: DUI

ARREST TIME: 01:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: GONZALEZ PEREZ

DATE/TIME OF TEST: 10/15/2019 02:21:32

---

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.160	02:16
Breath Volume	1.406	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.160	02:20
Breath Volume	1.906	02:20
Air Blank	0.000	02:20
Control Test	0.078	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

---

Data Download Date/Time: 10/17/2019 04:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORDOVA IZAGUIRRE,AMADO,  
BREATH TEST OPERATOR AGENCY: Fruitland Park PD  
ARRESTING OFFICER: WALSH,S  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 09/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A2  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 00:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: CORDOVA IZAGUIRRE

DATE/TIME OF TEST: 10/16/2019 01:24:34

---

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.142	01:19
Breath Volume	1.562	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.139	01:23
Breath Volume	1.496	01:23
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

---

Data Download Date/Time: 10/17/2019 04:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREEN,ERIC,

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: TYCE,DUSTIN

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 10/17/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 34917080A2

DRY GAS EXPIRY DATE: 02/05/2020

VIOLATION CODE: DUI

ARREST TIME: 14:28

GENDER: Female

OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: LAWS

DATE/TIME OF TEST: 10/31/2019 15:27:13

---

Diagnostics Check	OK	15:23	
Air Blank	0.000	15:24	
Control Test	0.079	15:24	
Air Blank	0.000	15:25	
Subject Sample	# 1 0.099	15:25	
Breath Volume	2.003	15:25	
Air Blank		15:25	Improper Sample
Air Blank	0.000	15:26	
Control Test	0.080	15:26	
Air Blank	0.000	15:27	
Diagnostics Check	OK	15:27	

---

Data Download Date/Time: 11/23/2019 09:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREEN,ERIC,  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: TYCE,DUSTIN  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 10/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A2  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: LAWS

DATE/TIME OF TEST: 10/31/2019 15:41:15

---

Diagnostics Check	OK	15:34
Air Blank	0.000	15:35
Control Test	0.079	15:35
Air Blank	0.000	15:36
Subject Sample # 1	0.095	15:36
Breath Volume	1.867	15:36
Air Blank	0.000	15:37
Air Blank	0.000	15:39
Subject Sample # 2	0.092	15:39
Breath Volume	1.878	15:39
Air Blank	0.000	15:40
Control Test	0.079	15:40
Air Blank	0.000	15:41
Diagnostics Check	OK	15:41

---

Data Download Date/Time: 11/23/2019 09:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALSH,SHANNON,L  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: WALSH,SHANNON  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 10/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A2  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: WOODRUFF

DATE/TIME OF TEST: 10/31/2019 21:36:30

---

Diagnostics Check	OK	21:30
Air Blank	0.000	21:30
Control Test	0.079	21:31
Air Blank	0.000	21:31
Subject Sample # 1	0.000	21:32
Breath Volume	2.265	21:32
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.000	21:35
Breath Volume	1.242	21:35
Air Blank	0.000	21:35
Control Test	0.079	21:35
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

---

Data Download Date/Time: 11/23/2019 09:13



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIVERS,DAWN,T  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: WARE,GLENN  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 10/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 09:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:37

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/26/2019 11:03:50

---

Diagnostics Check	OK	10:57
Air Blank	0.000	10:57
Control Test	0.077	10:57
Air Blank	0.000	10:58
Subject Sample # 1	0.149	10:59
Breath Volume	1.378	10:59
Air Blank	0.000	10:59
Air Blank	0.000	11:01
Subject Sample # 2	0.136	11:02
Breath Volume	1.296	11:02
Air Blank	0.000	11:02
Control Test	0.077	11:03
Air Blank	0.000	11:03
Diagnostics Check	OK	11:03

---

Data Download Date/Time: 10/30/2019 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: WARE,GLENN  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 10/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: ROWLAND

DATE/TIME OF TEST: 10/26/2019 22:41:13

---

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.077	22:33
Air Blank	0.000	22:34
Subject Sample # 1	0.223	22:36
Breath Volume	1.453	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.223	22:39
Breath Volume	1.750	22:39
Air Blank	0.000	22:40
Control Test	0.076	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

---

Data Download Date/Time: 10/30/2019 10:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: HOLLOWAY,JACOB  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 10/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: SLINKARD

DATE/TIME OF TEST: 10/27/2019 18:41:26

---

Diagnostics Check	OK	18:34
Air Blank	0.000	18:34
Control Test	0.077	18:35
Air Blank	0.000	18:35
Subject Sample # 1	0.234	18:36
Breath Volume	1.546	18:36
Air Blank	0.000	18:37
Air Blank	0.000	18:39
Subject Sample # 2	0.239	18:39
Breath Volume	1.394	18:39
Air Blank	0.000	18:40
Control Test	0.076	18:40
Air Blank	0.000	18:41
Diagnostics Check	OK	18:41

---

Data Download Date/Time: 10/30/2019 10:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: LOMBARDI,ANTHONY  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 10/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: BUTTS

DATE/TIME OF TEST: 10/27/2019 19:12:49

---

Diagnostics Check	OK	19:09	
Air Blank	0.000	19:10	
Control Test	0.076	19:10	
Air Blank	0.000	19:11	
Subject Sample	# 1	19:11	Subject Test Refused
Breath Volume	0.000	19:11	
Air Blank	0.000	19:11	
Control Test	0.077	19:12	
Air Blank	0.000	19:12	
Diagnostics Check	OK	19:12	

---

Data Download Date/Time: 10/30/2019 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: WOLSTENHOLM,A  
ARRESTING OFFICER AGENCY: FHP TROOP C  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: ROZEBOOM

DATE/TIME OF TEST: 10/03/2019 01:29:26

---

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.171	01:24
Breath Volume	1.230	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.164	01:27
Breath Volume	1.378	01:27
Air Blank	0.000	01:28
Control Test	0.078	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

---

Data Download Date/Time: 10/09/2019 19:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: VANTREESE,S  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 10/03/2019 03:27:23

---

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.079	03:22
Air Blank		0.000	03:22
Subject Sample	# 1	0.157	03:23
Breath Volume		2.425	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.162	03:25
Breath Volume		2.460	03:25
Air Blank		0.000	03:26
Control Test		0.079	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

---

Data Download Date/Time: 10/09/2019 19:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: ROSALES,F  
ARRESTING OFFICER AGENCY: PLANT CITY PD  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: ELWELL III

DATE/TIME OF TEST: 10/04/2019 01:07:58

---

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.079	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.283	01:03
Breath Volume	1.847	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.286	01:06
Breath Volume	2.484	01:06
Air Blank	0.000	01:07
Control Test	0.078	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

---

Data Download Date/Time: 10/09/2019 19:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: HOWARD,D  
ARRESTING OFFICER AGENCY: FHP TROOP C  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: CANALES DERAS

DATE/TIME OF TEST: 10/04/2019 05:29:37

---

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.079	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.099	05:25
Breath Volume	1.976	05:25
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.106	05:28
Breath Volume	3.261	05:28
Air Blank	0.000	05:28
Control Test	0.078	05:29
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

---

Data Download Date/Time: 10/09/2019 19:09



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: ANDREWS,S  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 10/04/2019 23:34:31

---

Diagnostics Check		OK	23:22	
Air Blank		0.000	23:22	
Control Test		0.080	23:23	
Air Blank		0.000	23:23	
Subject Sample	# 1	0.108	23:26	Volume Not Met
Breath Volume		0.046	23:26	
Air Blank		0.000	23:27	
Air Blank		0.000	23:29	
Subject Sample	# 2	0.136	23:29	
Breath Volume		2.058	23:29	
Air Blank		0.000	23:30	
Air Blank		0.000	23:32	
Subject Sample	# 3	0.132	23:33	
Breath Volume		1.828	23:33	
Air Blank		0.000	23:33	
Control Test		0.077	23:33	
Air Blank		0.000	23:34	
Diagnostics Check		OK	23:34	

---

Data Download Date/Time: 10/09/2019 19:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: ARCHAMBEAU,M  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: PETTRY

DATE/TIME OF TEST: 10/05/2019 02:57:00

---

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.078	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.139	02:51
Breath Volume	1.804	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.149	02:55
Breath Volume	2.210	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:57

---

Data Download Date/Time: 10/09/2019 19:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: VANTREESE,S  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: KENDALL

DATE/TIME OF TEST: 10/05/2019 03:38:31

---

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:27	
Control Test	0.079	03:27	
Air Blank	0.000	03:28	
Subject Sample	# 1	03:31	No Sample Provided
Breath Volume	0.000	03:31	
Air Blank	0.000	03:32	
Air Blank	0.000	03:34	
Subject Sample	# 2	03:37	No Sample Provided
Breath Volume	0.000	03:37	
Air Blank	0.000	03:37	
Control Test	0.079	03:37	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

---

Data Download Date/Time: 10/09/2019 19:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: VAN TREESE,S  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: KENDALL

DATE/TIME OF TEST: 10/05/2019 03:56:46

---

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.127	03:52
Breath Volume	1.984	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.128	03:55
Breath Volume	2.277	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

---

Data Download Date/Time: 10/09/2019 19:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: DRISCOLL,B  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: TOLEDO GARCIA

DATE/TIME OF TEST: 10/06/2019 00:20:48

---

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.079	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.130	00:14
Breath Volume	1.730	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:16
Subject Sample # 2	0.128	00:19
Breath Volume	2.109	00:19
Air Blank	0.000	00:19
Control Test	0.080	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

---

Data Download Date/Time: 10/09/2019 19:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: LYON,M  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: RIVAS HERNANDEZ

DATE/TIME OF TEST: 10/06/2019 04:30:45

---

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.079	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.178	04:26
Breath Volume	1.605	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:28
Subject Sample # 2	0.164	04:29
Breath Volume	2.179	04:29
Air Blank	0.000	04:29
Control Test	0.079	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

---

Data Download Date/Time: 10/09/2019 19:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: ARCHAMBEAU,M  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: LOGAN

DATE/TIME OF TEST: 10/06/2019 23:39:47

---

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:34
Subject Sample # 1	0.190	23:35
Breath Volume	3.480	23:35
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.198	23:38
Breath Volume	3.824	23:38
Air Blank	0.000	23:38
Control Test	0.078	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

---

Data Download Date/Time: 10/09/2019 19:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: BUGATTI

DATE/TIME OF TEST: 10/07/2019 22:27:36

---

Diagnostics Check	OK	22:23	
Air Blank	0.000	22:23	
Control Test	0.079	22:24	
Air Blank	0.000	22:24	
Subject Sample	# 1	22:26	Subject Test Refused
Breath Volume	0.000	22:26	
Air Blank	0.000	22:26	
Control Test	0.079	22:27	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:27	

---

Data Download Date/Time: 10/09/2019 19:09



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 10/08/2019 00:20:46

---

Diagnostics Check		OK	00:14
Air Blank		0.000	00:15
Control Test		0.080	00:15
Air Blank		0.000	00:16
Subject Sample	# 1	0.176	00:16
Breath Volume		1.996	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:18
Subject Sample	# 2	0.175	00:19
Breath Volume		2.355	00:19
Air Blank		0.000	00:19
Control Test		0.079	00:20
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

---

Data Download Date/Time: 10/09/2019 19:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: FISHER,T  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: UTSLER

DATE/TIME OF TEST: 10/08/2019 02:49:15

---

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.079	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.225	02:44
Breath Volume	1.281	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.239	02:47
Breath Volume	2.574	02:47
Air Blank	0.000	02:48
Control Test	0.078	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

---

Data Download Date/Time: 10/09/2019 19:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAY,ERICK,J  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: DAY,ERICK  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 09/08/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 10/01/2019 18:19:03

---

Diagnostics Check	OK	18:09
Air Blank	0.000	18:10
Control Test	0.079	18:10
Air Blank	0.000	18:10
Subject Sample # 1	0.130	18:13
Breath Volume	1.101	18:13
Air Blank	0.000	18:14
Air Blank	0.000	18:15
Subject Sample # 2	0.120	18:17
Breath Volume	1.246	18:17
Air Blank	0.000	18:18
Control Test	0.078	18:18
Air Blank	0.000	18:18
Diagnostics Check	OK	18:19

---

Data Download Date/Time: 10/04/2019 17:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JAQUES,J,C  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: JAQUES,J  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 09/08/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:44

SUBJECT'S LAST NAME: FRANCHI

DATE/TIME OF TEST: 10/02/2019 16:12:48

---

Diagnostics Check	OK	16:06
Air Blank	0.000	16:06
Control Test	0.079	16:07
Air Blank	0.000	16:07
Subject Sample # 1	0.185	16:08
Breath Volume	1.664	16:08
Air Blank	0.000	16:08
Air Blank	0.000	16:10
Subject Sample # 2	0.192	16:11
Breath Volume	1.554	16:11
Air Blank	0.000	16:11
Control Test	0.078	16:12
Air Blank	0.000	16:12
Diagnostics Check	OK	16:12

---

Data Download Date/Time: 10/04/2019 17:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAMPIASI,JAKE,L  
BREATH TEST OPERATOR AGENCY: FHP Troop K  
ARRESTING OFFICER: LAMPIASI,JAKE  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 09/08/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: GORDAN

DATE/TIME OF TEST: 10/03/2019 00:39:15

---

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.079	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.141	00:33
Breath Volume	3.316	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.127	00:37
Breath Volume	2.488	00:37
Air Blank	0.000	00:38
Control Test	0.077	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

---

Data Download Date/Time: 10/04/2019 17:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,JOSE,G  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: DELGADO,JOSE  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: SON CHIGUIL

DATE/TIME OF TEST: 10/05/2019 01:49:21

---

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.079	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.194	01:44
Breath Volume	2.253	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.188	01:47
Breath Volume	2.859	01:47
Air Blank	0.000	01:48
Control Test	0.078	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

---

Data Download Date/Time: 11/03/2019 22:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEER,BRADLEY,J  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: MEER,BRADLEY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:23

SUBJECT'S LAST NAME: HANIFL

DATE/TIME OF TEST: 10/09/2019 15:50:32

---

Diagnostics Check	OK	15:44
Air Blank	0.000	15:44
Control Test	0.079	15:45
Air Blank	0.000	15:45
Subject Sample # 1	0.000	15:46
Breath Volume	1.625	15:46
Air Blank	0.000	15:46
Air Blank	0.000	15:48
Subject Sample # 2	0.000	15:49
Breath Volume	2.863	15:49
Air Blank	0.000	15:49
Control Test	0.078	15:49
Air Blank	0.000	15:50
Diagnostics Check	OK	15:50

---

Data Download Date/Time: 11/03/2019 22:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKES,JOHN,D  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: WILKES,JOHN  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: HEGGERICK

DATE/TIME OF TEST: 10/12/2019 19:00:10

---

Diagnostics Check	OK	18:53
Air Blank	0.000	18:53
Control Test	0.079	18:54
Air Blank	0.000	18:54
Subject Sample # 1	0.178	18:55
Breath Volume	3.121	18:55
Air Blank	0.000	18:56
Air Blank	0.000	18:57
Subject Sample # 2	0.171	18:58
Breath Volume	2.539	18:58
Air Blank	0.000	18:59
Control Test	0.079	18:59
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

---

Data Download Date/Time: 11/03/2019 22:40



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLASZYK,JACOB,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BLASZYK,JACOB  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: LOPEZ PABLO

DATE/TIME OF TEST: 10/12/2019 20:43:50

---

Diagnostics Check	OK	20:37
Air Blank	0.000	20:37
Control Test	0.079	20:37
Air Blank	0.000	20:38
Subject Sample # 1	0.335	20:39
Breath Volume	1.925	20:39
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.343	20:42
Breath Volume	1.851	20:42
Air Blank	0.000	20:42
Control Test	0.079	20:43
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

---

Data Download Date/Time: 11/03/2019 22:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRADY,REBECCA,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BRADY,REBECCA  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:23

SUBJECT'S LAST NAME: CORLISS

DATE/TIME OF TEST: 10/13/2019 06:53:23

---

Diagnostics Check	OK	06:45
Air Blank	0.000	06:46
Control Test	0.079	06:46
Air Blank	0.000	06:47
Subject Sample # 1	0.000	06:49
Breath Volume	2.425	06:49
Air Blank	0.000	06:49
Air Blank	0.000	06:51
Subject Sample # 2	0.000	06:52
Breath Volume	2.000	06:52
Air Blank	0.000	06:52
Control Test	0.079	06:52
Air Blank	0.000	06:53
Diagnostics Check	OK	06:53

---

Data Download Date/Time: 11/03/2019 22:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: ANGERVIL,CLAUDY  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: KODNER

DATE/TIME OF TEST: 10/15/2019 03:32:30

---

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.079	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.212	03:27
Breath Volume	1.667	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:29
Subject Sample # 2	0.227	03:30
Breath Volume	1.402	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

---

Data Download Date/Time: 11/03/2019 22:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARROLL, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: CARROLL, MICHAEL  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: LOPEZ ORTIZ

DATE/TIME OF TEST: 10/15/2019 22:32:13

---

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:26
Subject Sample # 1	0.112	22:27
Breath Volume	2.628	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.112	22:30
Breath Volume	2.445	22:30
Air Blank	0.000	22:31
Control Test	0.079	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

---

Data Download Date/Time: 11/03/2019 22:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: CAPLAN,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: SALAS

DATE/TIME OF TEST: 10/21/2019 01:10:04

---

Diagnostics Check		OK	00:56	
Air Blank		0.000	00:56	
Control Test		0.078	00:56	
Air Blank		0.000	00:57	
Subject Sample	# 1	0.087	01:00	Volume Not Met
Breath Volume		0.000	01:00	
Air Blank		0.000	01:01	
Air Blank		0.000	01:03	
Subject Sample	# 2	0.087	01:03	
Breath Volume		2.242	01:03	
Air Blank		0.000	01:04	
Air Blank		0.000	01:06	
Subject Sample	# 3	0.090	01:08	
Breath Volume		2.632	01:08	
Air Blank		0.000	01:09	
Control Test		0.079	01:09	
Air Blank		0.000	01:09	
Diagnostics Check		OK	01:10	

---

Data Download Date/Time: 11/03/2019 22:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSON,DAVID,E  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: JACOBSON,DAVID  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: CANEVA

DATE/TIME OF TEST: 10/21/2019 18:42:10

---

Diagnostics Check	OK	18:35
Air Blank	0.000	18:36
Control Test	0.079	18:36
Air Blank	0.000	18:36
Subject Sample # 1	0.247	18:37
Breath Volume	2.699	18:37
Air Blank	0.000	18:37
Air Blank	0.000	18:39
Subject Sample # 2	0.253	18:40
Breath Volume	1.847	18:40
Air Blank	0.000	18:41
Control Test	0.079	18:41
Air Blank	0.000	18:42
Diagnostics Check	OK	18:42

---

Data Download Date/Time: 11/03/2019 22:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSON, DAVID, E  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: ROGEL-CARINO, JANNETH  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: KUTNER

DATE/TIME OF TEST: 10/22/2019 16:21:34

---

Diagnostics Check	OK	16:14
Air Blank	0.000	16:14
Control Test	0.079	16:14
Air Blank	0.000	16:15
Subject Sample # 1	0.129	16:17
Breath Volume	2.855	16:17
Air Blank	0.000	16:17
Air Blank	0.000	16:19
Subject Sample # 2	0.133	16:20
Breath Volume	2.773	16:20
Air Blank	0.000	16:20
Control Test	0.078	16:20
Air Blank	0.000	16:21
Diagnostics Check	OK	16:21

---

Data Download Date/Time: 11/03/2019 22:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CERNUTO, MATTHEW, D  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: CERNUTO, MATTHEW  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 10/23/2019 15:28:13

---

Diagnostics Check		OK	15:13	
Air Blank		0.000	15:14	
Control Test		0.079	15:14	
Air Blank		0.000	15:14	
Subject Sample	# 1	0.171	15:18	Volume Not Met
Breath Volume		0.156	15:18	
Air Blank		0.000	15:18	
Air Blank		0.000	15:20	
Subject Sample	# 2	0.203	15:20	
Breath Volume		1.496	15:20	
Air Blank		0.000	15:21	
Air Blank		0.000	15:23	
Subject Sample	# 3	0.153	15:26	Volume Not Met
Breath Volume		0.164	15:26	
Air Blank		0.000	15:27	
Control Test		0.078	15:27	
Air Blank		0.000	15:28	
Diagnostics Check		OK	15:28	

---

Data Download Date/Time: 11/03/2019 22:40



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEER,BRADLEY,J  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: MEER,BRADLEY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: ROSILES

DATE/TIME OF TEST: 10/23/2019 17:43:27

---

Diagnostics Check	OK	17:37
Air Blank	0.000	17:37
Control Test	0.079	17:37
Air Blank	0.000	17:38
Subject Sample # 1	0.000	17:39
Breath Volume	3.871	17:39
Air Blank	0.000	17:39
Air Blank	0.000	17:41
Subject Sample # 2	0.000	17:42
Breath Volume	2.468	17:42
Air Blank	0.000	17:42
Control Test	0.078	17:42
Air Blank	0.000	17:43
Diagnostics Check	OK	17:43

---

Data Download Date/Time: 11/03/2019 22:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENOVIL,GEDEON,  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BRENOVIL,GEDEON  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 13:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:43

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 10/24/2019 15:10:17

---

Diagnostics Check	OK	15:04
Air Blank	0.000	15:04
Control Test	0.080	15:04
Air Blank	0.000	15:05
Subject Sample # 1	0.161	15:05
Breath Volume	2.167	15:05
Air Blank	0.000	15:06
Air Blank	0.000	15:08
Subject Sample # 2	0.150	15:08
Breath Volume	1.609	15:08
Air Blank	0.000	15:09
Control Test	0.077	15:09
Air Blank	0.000	15:10
Diagnostics Check	OK	15:10

---

Data Download Date/Time: 11/03/2019 22:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKES,JOHN,D  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: WILKES,JOHN  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:21

SUBJECT'S LAST NAME: LINDSLEY

DATE/TIME OF TEST: 10/25/2019 18:55:47

---

Diagnostics Check	OK	18:48
Air Blank	0.000	18:49
Control Test	0.079	18:49
Air Blank	0.000	18:49
Subject Sample # 1	0.222	18:50
Breath Volume	2.691	18:50
Air Blank	0.000	18:51
Air Blank	0.000	18:53
Subject Sample # 2	0.223	18:54
Breath Volume	3.078	18:54
Air Blank	0.000	18:54
Control Test	0.078	18:55
Air Blank	0.000	18:55
Diagnostics Check	OK	18:55

---

Data Download Date/Time: 11/03/2019 22:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: THOMAS,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 10/25/2019 21:02:15

---

Diagnostics Check	OK	20:55
Air Blank	0.000	20:56
Control Test	0.079	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.000	20:57
Breath Volume	1.738	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.000	21:00
Breath Volume	1.707	21:00
Air Blank	0.000	21:01
Control Test	0.079	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

---

Data Download Date/Time: 11/03/2019 22:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ISHAM, MATTHEW, H  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: ISHAM, MATTHEW  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: TREVER

DATE/TIME OF TEST: 10/26/2019 22:31:41

---

Diagnostics Check	OK	22:25
Air Blank	0.000	22:25
Control Test	0.079	22:26
Air Blank	0.000	22:26
Subject Sample # 1	0.148	22:27
Breath Volume	1.953	22:27
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.148	22:30
Breath Volume	2.234	22:30
Air Blank	0.000	22:30
Control Test	0.078	22:31
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

---

Data Download Date/Time: 11/03/2019 22:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKES,JOHN,D  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: WILKES,JOHN  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:53

SUBJECT'S LAST NAME: SCHAEFER

DATE/TIME OF TEST: 10/27/2019 17:28:44

---

Diagnostics Check	OK	17:19
Air Blank	0.000	17:20
Control Test	0.079	17:20
Air Blank	0.000	17:21
Subject Sample # 1	0.334	17:23
Breath Volume	2.003	17:23
Air Blank	0.000	17:24
Air Blank	0.000	17:25
Subject Sample # 2	0.336	17:27
Breath Volume	2.585	17:27
Air Blank	0.000	17:27
Control Test	0.079	17:28
Air Blank	0.000	17:28
Diagnostics Check	OK	17:28

---

Data Download Date/Time: 11/03/2019 22:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: THOMAS,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: ROSE-BUTT

DATE/TIME OF TEST: 10/27/2019 22:32:15

---

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.079	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.267	22:25
Breath Volume	1.214	22:25
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.274	22:30
Breath Volume	1.242	22:30
Air Blank	0.000	22:31
Control Test	0.079	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

---

Data Download Date/Time: 11/03/2019 22:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAY,ERICK,J  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: DAY,ERICK  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 07:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:59

SUBJECT'S LAST NAME: MERKEL

DATE/TIME OF TEST: 10/29/2019 08:47:23

---

Diagnostics Check	OK	08:40
Air Blank	0.000	08:41
Control Test	0.079	08:41
Air Blank	0.000	08:41
Subject Sample # 1	0.138	08:42
Breath Volume	1.300	08:42
Air Blank	0.000	08:43
Air Blank	0.000	08:44
Subject Sample # 2	0.131	08:45
Breath Volume	1.117	08:45
Air Blank	0.000	08:46
Control Test	0.078	08:46
Air Blank	0.000	08:47
Diagnostics Check	OK	08:47

---

Data Download Date/Time: 11/03/2019 22:40



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLASZYK,JACOB,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BLASZYK,JACOB  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: KLECZEK

DATE/TIME OF TEST: 10/31/2019 00:46:35

---

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.000	00:42
Breath Volume	3.152	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.000	00:45
Breath Volume	3.484	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

---

Data Download Date/Time: 11/03/2019 22:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ANDREWS,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: FONOTI

DATE/TIME OF TEST: 10/02/2019 04:57:00

---

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.080	04:49
Air Blank	0.000	04:49
Subject Sample # 1	0.182	04:51
Breath Volume	2.132	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.170	04:55
Breath Volume	1.500	04:55
Air Blank	0.000	04:56
Control Test	0.078	04:56
Air Blank	0.000	04:56
Diagnostics Check	OK	04:57

---

Data Download Date/Time: 10/10/2019 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SCHAEFER,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/04/2019 21:47:27

---

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.080	21:42
Air Blank	0.000	21:42
Subject Sample # 1	0.000	21:42
Breath Volume	1.195	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.000	21:46
Breath Volume	1.441	21:46
Air Blank	0.000	21:46
Control Test	0.080	21:46
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

---

Data Download Date/Time: 10/10/2019 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LYON,M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: BEASLEY

DATE/TIME OF TEST: 10/04/2019 22:46:49

---

Diagnostics Check		OK	22:34	
Air Blank		0.000	22:34	
Control Test		0.080	22:34	
Air Blank		0.000	22:35	
Subject Sample	# 1	0.130	22:38	Volume Not Met
Breath Volume		0.917	22:38	
Air Blank		0.000	22:39	
Air Blank		0.000	22:41	
Subject Sample	# 2	0.135	22:41	
Breath Volume		1.304	22:41	
Air Blank		0.000	22:42	
Air Blank		0.000	22:44	
Subject Sample	# 3	0.133	22:45	
Breath Volume		1.386	22:45	
Air Blank		0.000	22:45	
Control Test		0.078	22:46	
Air Blank		0.000	22:46	
Diagnostics Check		OK	22:46	

---

Data Download Date/Time: 10/10/2019 14:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SULAIMAN,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: TALFORD IV

DATE/TIME OF TEST: 10/05/2019 02:37:01

---

Diagnostics Check		OK	02:23	
Air Blank		0.000	02:24	
Control Test		0.080	02:24	
Air Blank		0.000	02:25	
Subject Sample	# 1	0.123	02:28	Volume Not Met
Breath Volume		0.000	02:28	
Air Blank		0.000	02:29	
Air Blank		0.000	02:30	
Subject Sample	# 2	0.124	02:32	
Breath Volume		1.675	02:32	
Air Blank		0.000	02:33	
Air Blank		0.000	02:34	
Subject Sample	# 3	0.126	02:35	
Breath Volume		1.558	02:35	
Air Blank		0.000	02:36	
Control Test		0.078	02:36	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:37	

---

Data Download Date/Time: 10/10/2019 14:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ANDREWS,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 10/05/2019 05:25:28

---

Diagnostics Check	OK	05:18
Air Blank	0.000	05:18
Control Test	0.080	05:18
Air Blank	0.000	05:19
Subject Sample # 1	0.172	05:20
Breath Volume	1.910	05:20
Air Blank	0.000	05:21
Air Blank	0.000	05:23
Subject Sample # 2	0.178	05:23
Breath Volume	1.972	05:23
Air Blank	0.000	05:24
Control Test	0.079	05:24
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

---

Data Download Date/Time: 10/10/2019 14:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HALL,L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/06/2019 00:37:51

---

Diagnostics Check	OK	00:22	
Air Blank	0.000	00:22	
Control Test	0.079	00:23	
Air Blank	0.000	00:23	
Subject Sample	# 1 0.185	00:26	
Breath Volume	1.382	00:26	
Air Blank	0.000	00:27	
Air Blank	0.000	00:29	
Subject Sample	# 2 0.160	00:32	Volume Not Met
Breath Volume	0.765	00:32	
Air Blank	0.000	00:33	
Air Blank	0.000	00:35	
Subject Sample	# 3 0.173	00:36	
Breath Volume	1.550	00:36	
Air Blank	0.000	00:36	
Control Test	0.076	00:37	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

---

Data Download Date/Time: 10/10/2019 14:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SULAIMAN,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 10/06/2019 04:29:36

---

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.172	04:24
Breath Volume	3.804	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.176	04:27
Breath Volume	2.781	04:27
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

---

Data Download Date/Time: 10/10/2019 14:35



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: COSTA,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: BEIGHTEL

DATE/TIME OF TEST: 10/06/2019 05:15:20

---

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.079	05:07
Air Blank	0.000	05:07
Subject Sample # 1	0.049	05:09
Breath Volume	1.218	05:09
Air Blank	0.000	05:10
Air Blank	0.000	05:11
Subject Sample # 2	0.045	05:13
Breath Volume	1.628	05:13
Air Blank	0.000	05:14
Control Test	0.080	05:14
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

---

Data Download Date/Time: 10/10/2019 14:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/09/2019 04:19:40

---

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.163	04:14
Breath Volume	2.042	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:16
Subject Sample # 2	0.151	04:18
Breath Volume	2.289	04:18
Air Blank	0.000	04:18
Control Test	0.079	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

---

Data Download Date/Time: 10/10/2019 14:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: PITCHER,DARRYL  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 10/24/2019 22:44:15

---

Diagnostics Check		OK	22:29	
Air Blank		0.000	22:29	
Control Test		0.079	22:30	
Air Blank		0.000	22:30	
Subject Sample	# 1	0.132	22:33	Volume Not Met
Breath Volume		0.082	22:33	
Air Blank		0.000	22:34	
Air Blank		0.000	22:36	
Subject Sample	# 2	0.146	22:38	
Breath Volume		1.457	22:38	
Air Blank		0.000	22:39	
Air Blank		0.000	22:40	
Subject Sample	# 3	0.136	22:42	
Breath Volume		1.210	22:42	
Air Blank		0.000	22:43	
Control Test		0.079	22:43	
Air Blank		0.000	22:44	
Diagnostics Check		OK	22:44	

---

Data Download Date/Time: 11/09/2019 01:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,P  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,BRONSON  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/25/2019 21:39:34

---

Diagnostics Check	OK	21:36	
Air Blank	0.000	21:36	
Control Test	0.075	21:37	
Air Blank	0.000	21:37	
Subject Sample	# 1	21:38	Subject Test Refused
Breath Volume	0.000	21:38	
Air Blank	0.000	21:38	
Control Test	0.075	21:38	
Air Blank	0.000	21:39	
Diagnostics Check	OK	21:39	

---

Data Download Date/Time: 11/09/2019 01:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSS,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 10/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 11:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:26

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/03/2019 12:59:58

---

Diagnostics Check	OK	12:52
Air Blank	0.000	12:53
Control Test	0.081	12:53
Air Blank	0.000	12:54
Subject Sample # 1	0.036	12:55
Breath Volume	2.597	12:55
Air Blank	0.000	12:55
Air Blank	0.000	12:57
Subject Sample # 2	0.034	12:58
Breath Volume	2.089	12:58
Air Blank	0.000	12:59
Control Test	0.081	12:59
Air Blank	0.000	12:59
Diagnostics Check	OK	12:59

---

Data Download Date/Time: 11/03/2019 19:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHERI,DIANE,M  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: BARBOUR,DANIEL  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 10/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: ARMATO

DATE/TIME OF TEST: 10/03/2019 23:02:53

---

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.080	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.206	22:58
Breath Volume	1.984	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.215	23:01
Breath Volume	1.574	23:01
Air Blank	0.000	23:01
Control Test	0.080	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

---

Data Download Date/Time: 11/03/2019 19:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WURST,JOSHUA,D  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: WURST,JOSHUA  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 10/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: RUSIECKI

DATE/TIME OF TEST: 10/06/2019 16:03:17

---

Diagnostics Check	OK	15:56
Air Blank	0.000	15:57
Control Test	0.081	15:57
Air Blank	0.000	15:58
Subject Sample # 1	0.123	15:58
Breath Volume	1.542	15:58
Air Blank	0.000	15:59
Air Blank	0.000	16:01
Subject Sample # 2	0.124	16:01
Breath Volume	2.519	16:01
Air Blank	0.000	16:02
Control Test	0.079	16:02
Air Blank	0.000	16:03
Diagnostics Check	OK	16:03

---

Data Download Date/Time: 11/03/2019 19:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANCE,CHARLES,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: LANCE,CHARLES  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 10/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: AGUDELO

DATE/TIME OF TEST: 10/12/2019 02:08:29

---

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.080	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.215	02:04
Breath Volume	3.023	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.214	02:07
Breath Volume	2.488	02:07
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

---

Data Download Date/Time: 11/03/2019 19:38



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PELFREY,PAUL,D  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: BARBOUR,DAN  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 10/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: WILSON BURTZ

DATE/TIME OF TEST: 10/17/2019 02:09:45

---

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.185	02:04
Breath Volume		2.515	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:06
Subject Sample	# 2	0.176	02:08
Breath Volume		2.476	02:08
Air Blank		0.000	02:08
Control Test		0.081	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

---

Data Download Date/Time: 11/03/2019 19:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PELFREY,PAUL,D  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: POPPELREITER,DERICK  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 10/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/26/2019 02:32:50

---

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.080	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.197	02:28
Breath Volume	2.355	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.199	02:31
Breath Volume	2.113	02:31
Air Blank	0.000	02:31
Control Test	0.079	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

---

Data Download Date/Time: 11/03/2019 19:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,STEPHANIE,N  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: ALLEN,FORREST  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 10/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: WYSONG

DATE/TIME OF TEST: 10/27/2019 00:21:47

---

Diagnostics Check		OK	00:15
Air Blank		0.000	00:15
Control Test		0.080	00:16
Air Blank		0.000	00:16
Subject Sample	# 1	0.119	00:17
Breath Volume		1.222	00:17
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.125	00:20
Breath Volume		1.136	00:20
Air Blank		0.000	00:20
Control Test		0.078	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

---

Data Download Date/Time: 11/03/2019 19:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUEHLER,DEP,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BUEHLER,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: POPP

DATE/TIME OF TEST: 10/03/2019 03:04:22

---

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.178	02:59
Breath Volume	2.140	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.172	03:02
Breath Volume	2.304	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

---

Data Download Date/Time: 10/15/2019 10:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUEHLER,DEP,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: ECKERT,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DLUTOWSKI

DATE/TIME OF TEST: 10/22/2019 02:19:34

---

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.078	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.240	02:15
Breath Volume	1.164	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.249	02:18
Breath Volume	1.421	02:18
Air Blank	0.000	02:18
Control Test	0.078	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

---

Data Download Date/Time: 11/26/2019 09:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUEHLER,DEP,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: ECKERT,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: COGHLAN

DATE/TIME OF TEST: 10/23/2019 01:46:53

---

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.079	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.158	01:42
Breath Volume		3.003	01:42
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.163	01:45
Breath Volume		3.300	01:45
Air Blank		0.000	01:45
Control Test		0.078	01:46
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

---

Data Download Date/Time: 11/26/2019 09:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUEHLER,DEP,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: ECKERT,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/15/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: JOHN

DATE/TIME OF TEST: 10/26/2019 20:52:08

---

Diagnostics Check	OK	20:45
Air Blank	0.000	20:45
Control Test	0.079	20:46
Air Blank	0.000	20:46
Subject Sample # 1	0.213	20:47
Breath Volume	1.953	20:47
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample # 2	0.214	20:50
Breath Volume	2.554	20:50
Air Blank	0.000	20:51
Control Test	0.078	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

---

Data Download Date/Time: 11/26/2019 09:12