

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,RYAN,M  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: GIBSON,GERONIMO  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 620051  
DRY GAS EXPIRY DATE: 11/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: IRIZARRY

DATE/TIME OF TEST: 10/05/2019 01:21:16

---

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.083	01:19	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:20	
Control Test	0.082	01:20	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:21	

---

Data Download Date/Time: 10/24/2019 11:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUJAN,ARTURO,J  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: GARCIA,DASHIELL  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 620051  
DRY GAS EXPIRY DATE: 11/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 10/06/2019 18:17:47

---

Diagnostics Check	OK	18:11
Air Blank	0.000	18:11
Control Test	0.082	18:12
Air Blank	0.000	18:12
Subject Sample # 1	0.211	18:13
Breath Volume	1.484	18:13
Air Blank	0.000	18:13
Air Blank	0.000	18:15
Subject Sample # 2	0.218	18:16
Breath Volume	3.093	18:16
Air Blank	0.000	18:16
Control Test	0.081	18:17
Air Blank	0.000	18:17
Diagnostics Check	OK	18:17

---

Data Download Date/Time: 10/24/2019 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NAYA,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: NAYA,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 620051  
DRY GAS EXPIRY DATE: 11/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: VILLARREAL

DATE/TIME OF TEST: 10/11/2019 04:29:20

---

Diagnostics Check		OK	04:18
Air Blank		0.000	04:18
Control Test		0.081	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.216	04:21
Breath Volume		1.785	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:23
Subject Sample	# 2	0.189	04:24
Breath Volume		2.062	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 3	0.194	04:27
Breath Volume		1.910	04:27
Air Blank		0.000	04:28
Control Test		0.082	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

---

Data Download Date/Time: 10/24/2019 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA QUINONES,IGNACIO,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: ROSAS,ANYELINNA  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 620051  
DRY GAS EXPIRY DATE: 11/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: VERGARA TOBON

DATE/TIME OF TEST: 10/19/2019 20:59:37

---

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.082	20:53
Air Blank	0.000	20:54
Subject Sample # 1	0.325	20:54
Breath Volume	2.085	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.321	20:58
Breath Volume	1.402	20:58
Air Blank	0.000	20:58
Control Test	0.082	20:59
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

---

Data Download Date/Time: 10/24/2019 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,RYAN,M  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: SISNEY,WHITNEY  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 620051  
DRY GAS EXPIRY DATE: 11/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: GALARRETA

DATE/TIME OF TEST: 10/20/2019 01:35:46

---

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:33	
Control Test	0.082	01:33	
Air Blank	0.000	01:34	
Subject Sample	# 1	01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:34	
Control Test	0.082	01:35	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

---

Data Download Date/Time: 10/24/2019 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STEELE,ANDREW,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: CONTRERAS,ORLANDO  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 620051  
DRY GAS EXPIRY DATE: 11/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: DOS PASSOS PINTO

DATE/TIME OF TEST: 10/22/2019 03:26:18

---

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.082	03:24	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:25	
Control Test	0.082	03:25	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

---

Data Download Date/Time: 10/24/2019 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITE,AVERY,D  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: WHITE,AVERY  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CASAMAYOR

DATE/TIME OF TEST: 10/12/2019 01:52:30

---

Diagnostics Check	OK	01:51	
Air Blank		01:52	Ambient Fail
Air Blank	0.000	01:52	

---

Data Download Date/Time: 11/04/2019 07:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITE,AVERY,D  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: WHITE,AVERY  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CASAMAYOR

DATE/TIME OF TEST: 10/12/2019 02:02:15

---

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.112	01:57
Breath Volume	1.531	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.115	02:00
Breath Volume	1.101	02:00
Air Blank	0.000	02:01
Control Test	0.079	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

---

Data Download Date/Time: 11/04/2019 07:19



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELENDEZ,DANIEL,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: POLANCO,JESSICA  
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 11:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 10/03/2019 00:55:21

---

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.082	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.126	00:49
Breath Volume	1.804	00:49
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.127	00:53
Breath Volume	1.859	00:53
Air Blank	0.000	00:54
Control Test	0.082	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

---

Data Download Date/Time: 10/30/2019 11:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUMPHRY,KERRY,R  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: AZICRI,LEON  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: PUNAY RAYMUNDO

DATE/TIME OF TEST: 10/03/2019 03:46:39

---

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.082	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.231	03:41
Breath Volume	1.503	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.232	03:45
Breath Volume	2.093	03:45
Air Blank	0.000	03:45
Control Test	0.082	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

---

Data Download Date/Time: 10/30/2019 11:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CONFESSORE,STEVE,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: ACEVEDO,KATHLEEN  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/05/2019 00:32:15

---

Diagnostics Check		OK	00:18	
Air Blank		0.000	00:19	
Control Test		0.082	00:19	
Air Blank		0.000	00:20	
Subject Sample	# 1	0.204	00:23	Volume Not Met
Breath Volume		0.359	00:23	
Air Blank		0.000	00:23	
Air Blank		0.000	00:25	
Subject Sample	# 2	0.206	00:26	
Breath Volume		1.464	00:26	
Air Blank		0.000	00:27	
Air Blank		0.000	00:29	
Subject Sample	# 3	0.208	00:30	
Breath Volume		1.902	00:30	
Air Blank		0.000	00:31	
Control Test		0.080	00:31	
Air Blank		0.000	00:32	
Diagnostics Check		OK	00:32	

---

Data Download Date/Time: 10/30/2019 11:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAIGE,LEON,I  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: PAIGE,LEON  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: TAPIA MALARIN

DATE/TIME OF TEST: 10/06/2019 23:31:57

---

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.082	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.161	23:27
Breath Volume	2.480	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.154	23:30
Breath Volume	2.828	23:30
Air Blank	0.000	23:31
Control Test	0.081	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

---

Data Download Date/Time: 10/30/2019 11:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAIGE,LEON,I  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: PAIGE,LEON  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/30/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: BLAIZE

DATE/TIME OF TEST: 10/11/2019 02:39:01

---

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:34	
Control Test	0.082	02:34	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:37	Subject Test Refused
Breath Volume	0.000	02:37	
Air Blank	0.000	02:38	
Control Test	0.083	02:38	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:39	

---

Data Download Date/Time: 10/30/2019 11:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAIGE,LEON,I  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: PAIGE,LEON  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/30/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: HERNANDEZ MORALES

DATE/TIME OF TEST: 10/18/2019 00:39:12

---

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.082	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.068	00:34
Breath Volume	3.203	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.076	00:37
Breath Volume	4.398	00:37
Air Blank	0.000	00:38
Control Test	0.082	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

---

Data Download Date/Time: 10/30/2019 11:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELENDEZ,DANIEL,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: MELENDEZ,DANIEL  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 10/18/2019 06:17:00

---

Diagnostics Check		OK	06:10
Air Blank		0.000	06:11
Control Test		0.082	06:11
Air Blank		0.000	06:11
Subject Sample	# 1	0.171	06:12
Breath Volume		1.843	06:12
Air Blank		0.000	06:13
Air Blank		0.000	06:15
Subject Sample	# 2	0.173	06:15
Breath Volume		1.894	06:15
Air Blank		0.000	06:16
Control Test		0.081	06:16
Air Blank		0.000	06:16
Diagnostics Check		OK	06:17

---

Data Download Date/Time: 10/30/2019 11:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELENDEZ,DANIEL,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: MELENDEZ,DANIEL  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: BOS-LUN

DATE/TIME OF TEST: 10/19/2019 06:30:16

---

Diagnostics Check	OK	06:23
Air Blank	0.000	06:23
Control Test	0.082	06:24
Air Blank	0.000	06:24
Subject Sample # 1	0.200	06:25
Breath Volume	3.117	06:25
Air Blank	0.000	06:26
Air Blank	0.000	06:28
Subject Sample # 2	0.199	06:28
Breath Volume	1.464	06:28
Air Blank	0.000	06:29
Control Test	0.082	06:29
Air Blank	0.000	06:30
Diagnostics Check	OK	06:30

---

Data Download Date/Time: 10/30/2019 11:42



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, JONATHAN,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: RODRIGUEZ, RICHARD  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175968  
DRY GAS EXPIRY DATE: 07/30/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: ALLI-BALUKOFF

DATE/TIME OF TEST: 10/19/2019 22:24:21

---

Diagnostics Check		OK	22:11	
Air Blank		0.000	22:11	
Control Test		0.082	22:11	
Air Blank		0.000	22:12	
Subject Sample	# 1	0.171	22:15	Volume Not Met
Breath Volume		0.734	22:15	
Air Blank		0.000	22:16	
Air Blank		0.000	22:17	
Subject Sample	# 2	0.166	22:19	
Breath Volume		1.601	22:19	
Air Blank		0.000	22:20	
Air Blank		0.000	22:22	
Subject Sample	# 3	0.183	22:22	
Breath Volume		1.777	22:22	
Air Blank		0.000	22:23	
Control Test		0.081	22:23	
Air Blank		0.000	22:24	
Diagnostics Check		OK	22:24	

---

Data Download Date/Time: 10/30/2019 11:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELENDEZ,DANIEL,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: MELENDEZ,DANIEL  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/20/2019 01:39:45

---

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.082	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.196	01:31
Breath Volume	2.441	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.175	01:35
Breath Volume	1.761	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 3	0.188	01:38
Breath Volume	2.160	01:38
Air Blank	0.000	01:38
Control Test	0.080	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

---

Data Download Date/Time: 10/30/2019 11:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELENDEZ,DANIEL,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: MELENDEZ,DANIEL  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: DUARTE

DATE/TIME OF TEST: 10/20/2019 03:45:52

---

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.082	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.211	03:41
Breath Volume	2.414	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.203	03:44
Breath Volume	3.953	03:44
Air Blank	0.000	03:44
Control Test	0.077	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

---

Data Download Date/Time: 10/30/2019 11:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAIGE,LEON,I  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: PAIGE,LEON  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/20/2019 23:13:33

---

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.080	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.139	23:08
Breath Volume	2.781	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:10
Subject Sample # 2	0.142	23:12
Breath Volume	3.335	23:12
Air Blank	0.000	23:12
Control Test	0.081	23:12
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

---

Data Download Date/Time: 10/30/2019 11:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAIGE,LEON,I  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: PAIGE,LEON  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: AGUILAR HERNANDEZ

DATE/TIME OF TEST: 10/21/2019 03:01:35

---

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:51	
Control Test		0.082	02:51	
Air Blank		0.000	02:52	
Subject Sample	# 1	0.194	02:55	Volume Not Met
Breath Volume		0.707	02:55	
Air Blank		0.000	02:55	
Air Blank		0.000	02:57	
Subject Sample	# 2		03:00	Subject Test Refused
Breath Volume		0.238	03:00	
Air Blank		0.000	03:00	
Control Test		0.081	03:00	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

---

Data Download Date/Time: 10/30/2019 11:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HORTON,REGINALD,B  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: HORTON,REGINALD  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: FLECHA

DATE/TIME OF TEST: 10/24/2019 05:05:06

---

Diagnostics Check	OK	04:58
Air Blank	0.000	04:59
Control Test	0.081	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.145	05:00
Breath Volume	1.664	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.162	05:03
Breath Volume	1.417	05:03
Air Blank	0.000	05:04
Control Test	0.081	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

---

Data Download Date/Time: 10/30/2019 11:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAIGE,LEON,I  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: PAIGE,LEON  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PETRARCA

DATE/TIME OF TEST: 10/26/2019 03:57:29

---

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.082	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.126	03:53
Breath Volume	2.562	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.125	03:56
Breath Volume	2.527	03:56
Air Blank	0.000	03:56
Control Test	0.082	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

---

Data Download Date/Time: 10/30/2019 11:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOMINGUEZ,FREDERICK,A  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: RODRIGUEZ,ELIZABETH  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175968  
DRY GAS EXPIRY DATE: 07/30/2022  
VIOLATION CODE: DUI  
ARREST TIME: 09:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:39

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 10/26/2019 10:06:49

---

Diagnostics Check	OK	09:59
Air Blank	0.000	10:00
Control Test	0.083	10:00
Air Blank	0.000	10:01
Subject Sample # 1	0.169	10:01
Breath Volume	2.882	10:01
Air Blank	0.000	10:02
Air Blank	0.000	10:04
Subject Sample # 2	0.172	10:05
Breath Volume	3.273	10:05
Air Blank	0.000	10:05
Control Test	0.081	10:06
Air Blank	0.000	10:06
Diagnostics Check	OK	10:06

---

Data Download Date/Time: 10/30/2019 11:43



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAIGE,LEON,I  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: PAIGE,LEON  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/27/2019 04:45:08

---

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:42	
Control Test	0.082	04:43	
Air Blank	0.000	04:43	
Subject Sample	# 1	04:43	Subject Test Refused
Breath Volume	0.000	04:43	
Air Blank	0.000	04:44	
Control Test	0.082	04:44	
Air Blank	0.000	04:45	
Diagnostics Check	OK	04:45	

---

Data Download Date/Time: 10/30/2019 11:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELENDEZ,DANIEL,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: MELENDEZ,DANIEL  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:31

SUBJECT'S LAST NAME: ENCALADA

DATE/TIME OF TEST: 10/27/2019 06:52:27

---

Diagnostics Check	OK	06:51	
Air Blank		06:51	RFI Detect
Air Blank	0.000	06:52	

---

Data Download Date/Time: 10/30/2019 11:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELENDEZ,DANIEL,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: MELENDEZ,DANIEL  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:31

SUBJECT'S LAST NAME: ENCALADA

DATE/TIME OF TEST: 10/27/2019 07:01:43

---

Diagnostics Check	OK	06:55
Air Blank	0.000	06:55
Control Test	0.082	06:56
Air Blank	0.000	06:56
Subject Sample # 1	0.162	06:57
Breath Volume	3.316	06:57
Air Blank	0.000	06:57
Air Blank	0.000	06:59
Subject Sample # 2	0.162	07:00
Breath Volume	3.414	07:00
Air Blank	0.000	07:00
Control Test	0.082	07:01
Air Blank	0.000	07:01
Diagnostics Check	OK	07:01

---

Data Download Date/Time: 10/30/2019 11:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELENDEZ,DANIEL,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: HACKWORTH,EARL  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: ALEAGA

DATE/TIME OF TEST: 10/28/2019 03:44:53

---

Diagnostics Check		OK	03:38
Air Blank		0.000	03:38
Control Test		0.081	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.183	03:40
Breath Volume		2.500	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:42
Subject Sample	# 2	0.175	03:43
Breath Volume		2.121	03:43
Air Blank		0.000	03:43
Control Test		0.081	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

---

Data Download Date/Time: 10/30/2019 11:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOMINGUEZ,ERIC,  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: THOMAS,WAYNE  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814181  
DRY GAS EXPIRY DATE: 07/24/2020  
VIOLATION CODE: DUI  
ARREST TIME: 09:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:07

SUBJECT'S LAST NAME: DIAZ SIERRA

DATE/TIME OF TEST: 10/30/2019 10:31:00

---

Diagnostics Check	OK	10:27	
Air Blank	0.000	10:28	
Control Test	0.082	10:28	
Air Blank	0.000	10:29	
Subject Sample	# 1	10:29	Subject Test Refused
Breath Volume	0.000	10:29	
Air Blank	0.000	10:30	
Control Test	0.081	10:30	
Air Blank	0.000	10:30	
Diagnostics Check	OK	10:31	

---

Data Download Date/Time: 10/30/2019 11:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOMINGUEZ,FREDERICK,A  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: DOMINGUEZ,FREDERICK  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 10/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175968  
DRY GAS EXPIRY DATE: 07/30/2022  
VIOLATION CODE: DUI  
ARREST TIME: 08:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:44

SUBJECT'S LAST NAME: QUISPE PENA

DATE/TIME OF TEST: 10/31/2019 09:11:18

---

Diagnostics Check	OK	09:08	
Air Blank	0.000	09:08	
Control Test	0.079	09:09	
Air Blank	0.000	09:09	
Subject Sample	# 1	09:09	Subject Test Refused
Breath Volume	0.000	09:09	
Air Blank	0.000	09:10	
Control Test	0.080	09:10	
Air Blank	0.000	09:11	
Diagnostics Check	OK	09:11	

---

Data Download Date/Time: 11/15/2019 11:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO,GEORGE,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PACHECO,GEORGE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 10/27/2019 07:34:49

---

Diagnostics Check	OK	07:31	
Air Blank	0.000	07:32	
Control Test	0.078	07:32	
Air Blank	0.000	07:33	
Subject Sample	# 1	07:33	Subject Test Refused
Breath Volume	0.000	07:33	
Air Blank	0.000	07:33	
Control Test	0.078	07:34	
Air Blank	0.000	07:34	
Diagnostics Check	OK	07:34	

---

Data Download Date/Time: 10/31/2019 02:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,STEVEN,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: ADAMS,STEVEN  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 09/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: FONSECA

DATE/TIME OF TEST: 10/09/2019 00:02:23

---

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.081	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.128	23:57
Breath Volume	3.035	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.134	00:00
Breath Volume	2.898	00:00
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

---

Data Download Date/Time: 10/31/2019 02:36



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,STEVEN,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: ADAMS,STEVEN  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 09/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: COSTA-ALVAREZ

DATE/TIME OF TEST: 10/26/2019 22:42:50

---

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.081	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.198	22:37
Breath Volume	1.718	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.193	22:41
Breath Volume	1.742	22:41
Air Blank	0.000	22:41
Control Test	0.081	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

---

Data Download Date/Time: 10/31/2019 02:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GATTORNO,ALEXANDER,F  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: GATTORNO,ALEXANDER  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 09/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 07:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:57

SUBJECT'S LAST NAME: RANDOLPH

DATE/TIME OF TEST: 10/27/2019 09:01:12

---

Diagnostics Check	OK	08:54
Air Blank	0.000	08:54
Control Test	0.080	08:54
Air Blank	0.000	08:55
Subject Sample # 1	0.191	08:56
Breath Volume	2.738	08:56
Air Blank	0.000	08:56
Air Blank	0.000	08:58
Subject Sample # 2	0.191	08:59
Breath Volume	2.503	08:59
Air Blank	0.000	09:00
Control Test	0.081	09:00
Air Blank	0.000	09:01
Diagnostics Check	OK	09:01

---

Data Download Date/Time: 10/31/2019 02:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTILLO,JOSUET,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: CASTILLO,JOSUET  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 09/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BELTRAN

DATE/TIME OF TEST: 10/07/2019 04:02:21

---

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.066	03:57
Breath Volume	3.089	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.067	04:00
Breath Volume	2.976	04:00
Air Blank	0.000	04:01
Control Test	0.081	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

---

Data Download Date/Time: 10/31/2019 01:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VIVEIROS,KYLE,M  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: VIVEIROS,KYLE  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 09/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:38

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 10/25/2019 12:24:46

---

Diagnostics Check	OK	12:23	
Air Blank	0.000	12:23	
Control Test	0.079	12:24	
Air Blank		12:24	RFI Detect
Air Blank	0.000	12:24	

---

Data Download Date/Time: 10/31/2019 01:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VIVEIROS,KYLE,M  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: VIVEIROS,KYLE  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 09/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:38

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 10/25/2019 12:29:58

---

Diagnostics Check	OK	12:27	
Air Blank	0.000	12:27	
Control Test	0.079	12:27	
Air Blank	0.000	12:28	
Subject Sample	# 1	12:28	Subject Test Refused
Breath Volume	0.000	12:28	
Air Blank	0.000	12:29	
Control Test	0.080	12:29	
Air Blank	0.000	12:29	
Diagnostics Check	OK	12:29	

---

Data Download Date/Time: 10/31/2019 01:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTILLO,JOSUET,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: CASTILLO,JOSUET  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 09/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: QUINTANILLAGUTIERREZ

DATE/TIME OF TEST: 10/26/2019 06:01:11

---

Diagnostics Check	OK	05:58	
Air Blank	0.000	05:58	
Control Test	0.080	05:59	
Air Blank	0.000	05:59	
Subject Sample	# 1	05:59	Subject Test Refused
Breath Volume	0.000	05:59	
Air Blank	0.000	06:00	
Control Test	0.080	06:00	
Air Blank	0.000	06:01	
Diagnostics Check	OK	06:01	

---

Data Download Date/Time: 10/31/2019 01:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VARGAS,D,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: VARGAS,D  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 10/03/2019 23:02:01

---

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:53	
Control Test	0.080	22:53	
Air Blank	0.000	22:53	
Subject Sample	# 1 0.209	22:57	Volume Not Met
Breath Volume	0.402	22:57	
Air Blank	0.000	22:57	
Air Blank	0.000	22:59	
Subject Sample	# 2	23:00	Subject Test Refused
Breath Volume	0.035	23:00	
Air Blank	0.000	23:01	
Control Test	0.079	23:01	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:02	

---

Data Download Date/Time: 10/31/2019 01:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANDERSON,S,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: WEDDERBURN,A  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 12:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: CERVANTES

DATE/TIME OF TEST: 10/01/2019 12:43:05

---

Diagnostics Check	OK	12:37
Air Blank	0.000	12:37
Control Test	0.079	12:37
Air Blank	0.000	12:38
Subject Sample # 1	0.000	12:38
Breath Volume	2.453	12:38
Air Blank	0.000	12:39
Air Blank	0.000	12:41
Subject Sample # 2	0.000	12:41
Breath Volume	1.847	12:41
Air Blank	0.000	12:42
Control Test	0.078	12:42
Air Blank	0.000	12:43
Diagnostics Check	OK	12:43

---

Data Download Date/Time: 10/28/2019 06:40



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLANUEVA, MAURICIO,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: ALADIN, KENDRICK  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 10/19/2019 01:52:46

---

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:47	
Control Test	0.079	01:47	
Air Blank	0.000	01:48	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:51	
Control Test	0.078	01:52	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

---

Data Download Date/Time: 10/28/2019 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLANUEVA, MAURICIO,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: ALADIN, KENDRICK  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: IGBINO GHENE

DATE/TIME OF TEST: 10/27/2019 00:13:50

---

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:06	
Control Test	0.079	00:06	
Air Blank	0.000	00:07	
Subject Sample	# 1	00:10	No Sample Provided
Breath Volume	0.000	00:10	
Air Blank	0.000	00:10	
Air Blank		00:12	RFI Detect
Air Blank	0.000	00:12	
Control Test	0.079	00:13	
Air Blank	0.000	00:13	
Diagnostics Check	OK	00:13	

---

Data Download Date/Time: 10/28/2019 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLANUEVA, MAURICIO,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: ALADIN, KENDRICK  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 09/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: IGBINO GHENE

DATE/TIME OF TEST: 10/27/2019 00:20:33

---

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:16	
Control Test	0.079	00:17	
Air Blank	0.000	00:17	
Subject Sample	# 1	00:19	Subject Test Refused
Breath Volume	0.000	00:19	
Air Blank	0.000	00:19	
Control Test	0.079	00:19	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

---

Data Download Date/Time: 10/28/2019 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BASFORD,NICOLE,M  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: BASFORD,NICOLE  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 09/13/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/21/2021  
VIOLATION CODE: BUI  
ARREST TIME: 16:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: STONER

DATE/TIME OF TEST: 10/06/2019 17:10:32

---

Diagnostics Check	OK	17:07	
Air Blank	0.000	17:07	
Control Test	0.080	17:08	
Air Blank	0.000	17:08	
Subject Sample	# 1	17:09	Subject Test Refused
Breath Volume	0.000	17:09	
Air Blank	0.000	17:09	
Control Test	0.081	17:09	
Air Blank	0.000	17:10	
Diagnostics Check	OK	17:10	

---

Data Download Date/Time: 10/17/2019 14:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: PULASKI,DAMON  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 18:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 10/13/2019 20:25:03

---

Diagnostics Check	OK	20:18
Air Blank	0.000	20:19
Control Test	0.079	20:19
Air Blank	0.000	20:20
Subject Sample # 1	0.000	20:20
Breath Volume	1.773	20:20
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.000	20:23
Breath Volume	1.925	20:23
Air Blank	0.000	20:24
Control Test	0.077	20:24
Air Blank	0.000	20:24
Diagnostics Check	OK	20:25

---

Data Download Date/Time: 11/15/2019 20:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROTHCHILD,FOREST,W  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: CANAMERO,ALBERT  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 17:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 10/18/2019 18:29:39

---

Diagnostics Check	OK	18:26	
Air Blank	0.000	18:27	
Control Test	0.077	18:27	
Air Blank	0.000	18:28	
Subject Sample	# 1	18:28	Subject Test Refused
Breath Volume	0.000	18:28	
Air Blank	0.000	18:28	
Control Test	0.078	18:29	
Air Blank	0.000	18:29	
Diagnostics Check	OK	18:29	

---

Data Download Date/Time: 11/15/2019 20:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROTHCHILD,FOREST,W  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: ROTHCHILD,FOREST  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 19:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: OPOLSKI

DATE/TIME OF TEST: 10/27/2019 19:35:38

---

Diagnostics Check	OK	19:29
Air Blank	0.000	19:29
Control Test	0.078	19:29
Air Blank	0.000	19:30
Subject Sample # 1	0.073	19:31
Breath Volume	2.949	19:31
Air Blank	0.000	19:31
Air Blank	0.000	19:33
Subject Sample # 2	0.072	19:34
Breath Volume	3.121	19:34
Air Blank	0.000	19:34
Control Test	0.078	19:35
Air Blank	0.000	19:35
Diagnostics Check	OK	19:35

---

Data Download Date/Time: 11/15/2019 20:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,MALACHI,C  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: MCCORMACK,EAMON  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 22:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: MAGLICH

DATE/TIME OF TEST: 10/05/2019 23:57:50

---

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:56	
Control Test	0.000	23:57	Control Outside Tolerance
Air Blank	0.000	23:57	

---

Data Download Date/Time: 11/15/2019 19:18



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,MALACHI,C  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: MCCORMACK,EAMON  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 22:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: MAGLICH

DATE/TIME OF TEST: 10/06/2019 00:05:44

---

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:01	
Control Test	0.077	00:02	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:03	Subject Test Refused
Breath Volume	0.000	00:03	
Air Blank	0.000	00:04	
Control Test	0.079	00:04	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

---

Data Download Date/Time: 11/15/2019 19:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,MALACHI,C  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: MCCORMACK,EAMON  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 17:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:37

SUBJECT'S LAST NAME: CARY

DATE/TIME OF TEST: 10/12/2019 18:23:51

---

Diagnostics Check	OK	18:19	
Air Blank	0.000	18:20	
Control Test	0.079	18:20	
Air Blank	0.000	18:21	
Subject Sample	# 1	18:21	Subject Test Refused
Breath Volume	0.000	18:21	
Air Blank	0.000	18:22	
Control Test	0.078	18:22	
Air Blank	0.000	18:23	
Diagnostics Check	OK	18:23	

---

Data Download Date/Time: 11/15/2019 19:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,MALACHI,C  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: PASS,JUSTIN  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 20:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: DOTSON

DATE/TIME OF TEST: 10/12/2019 21:23:23

---

Diagnostics Check	OK	21:16
Air Blank	0.000	21:17
Control Test	0.080	21:17
Air Blank	0.000	21:18
Subject Sample # 1	0.067	21:18
Breath Volume	3.074	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.069	21:21
Breath Volume	3.203	21:21
Air Blank	0.000	21:22
Control Test	0.079	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

---

Data Download Date/Time: 11/15/2019 19:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,MALACHI,C  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: BERG,TIM  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 20:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: NAEDER

DATE/TIME OF TEST: 10/13/2019 20:54:40

---

Diagnostics Check	OK	20:50	
Air Blank	0.000	20:51	
Control Test	0.079	20:51	
Air Blank	0.000	20:52	
Subject Sample	# 1	20:52	Subject Test Refused
Breath Volume	0.000	20:52	
Air Blank	0.000	20:53	
Control Test	0.079	20:53	
Air Blank	0.000	20:54	
Diagnostics Check	OK	20:54	

---

Data Download Date/Time: 11/15/2019 19:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: OLVERA

DATE/TIME OF TEST: 10/02/2019 23:57:33

---

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.156	23:53
Breath Volume	1.371	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.149	23:56
Breath Volume	1.425	23:56
Air Blank	0.000	23:56
Control Test	0.078	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

---

Data Download Date/Time: 10/22/2019 15:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/03/2019 22:17:01

---

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.079	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.187	22:10
Breath Volume	2.117	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.182	22:15
Breath Volume	1.562	22:15
Air Blank	0.000	22:16
Control Test	0.078	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:17

---

Data Download Date/Time: 10/22/2019 15:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: JUAN

DATE/TIME OF TEST: 10/13/2019 00:41:21

---

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.079	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.211	00:34
Breath Volume	1.148	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.223	00:39
Breath Volume	1.484	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

---

Data Download Date/Time: 10/22/2019 15:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: DATTOLO

DATE/TIME OF TEST: 10/13/2019 21:32:04

---

Diagnostics Check	OK	21:25
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:27
Subject Sample # 1	0.180	21:27
Breath Volume	1.988	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.177	21:30
Breath Volume	1.425	21:30
Air Blank	0.000	21:31
Control Test	0.079	21:31
Air Blank	0.000	21:31
Diagnostics Check	OK	21:32

---

Data Download Date/Time: 10/22/2019 15:01



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: CARIAS MORALES

DATE/TIME OF TEST: 10/14/2019 00:50:21

---

Diagnostics Check	OK	00:42
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:44
Subject Sample # 1	0.253	00:45
Breath Volume	2.105	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.263	00:48
Breath Volume	1.679	00:48
Air Blank	0.000	00:49
Control Test	0.078	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

---

Data Download Date/Time: 10/22/2019 15:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: URBANCIK

DATE/TIME OF TEST: 10/17/2019 02:57:17

---

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.138	02:52
Breath Volume	1.406	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.149	02:55
Breath Volume	1.472	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

---

Data Download Date/Time: 10/22/2019 15:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: CAMMILLERI

DATE/TIME OF TEST: 10/18/2019 02:03:24

---

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.157	01:59
Breath Volume	1.578	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.158	02:01
Breath Volume	1.546	02:01
Air Blank	0.000	02:02
Control Test	0.080	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

---

Data Download Date/Time: 10/22/2019 15:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MEMBRENO CASTILLO

DATE/TIME OF TEST: 10/26/2019 01:22:15

---

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.078	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.175	01:17
Breath Volume	2.535	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.179	01:20
Breath Volume	2.343	01:20
Air Blank	0.000	01:21
Control Test	0.077	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

---

Data Download Date/Time: 11/18/2019 08:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: RAMIREZ OLVERA

DATE/TIME OF TEST: 10/26/2019 23:10:43

---

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.079	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.182	23:06
Breath Volume	1.867	23:06
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.193	23:09
Breath Volume	2.250	23:09
Air Blank	0.000	23:09
Control Test	0.078	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

---

Data Download Date/Time: 11/18/2019 08:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: NORTHE

DATE/TIME OF TEST: 10/27/2019 23:12:45

---

Diagnostics Check	OK	23:06
Air Blank	0.000	23:07
Control Test	0.078	23:07
Air Blank	0.000	23:07
Subject Sample # 1	0.174	23:08
Breath Volume	2.480	23:08
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.171	23:11
Breath Volume	1.500	23:11
Air Blank	0.000	23:11
Control Test	0.077	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

---

Data Download Date/Time: 11/18/2019 08:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: NAVA NAVA

DATE/TIME OF TEST: 10/28/2019 03:09:27

---

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.187	03:05
Breath Volume	2.250	03:05
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.190	03:08
Breath Volume	1.785	03:08
Air Blank	0.000	03:08
Control Test	0.078	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

---

Data Download Date/Time: 11/18/2019 08:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINS,KYLE,L  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: LUNA,MICHAEL  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: VARGAS GRESS

DATE/TIME OF TEST: 10/02/2019 03:05:58

---

Diagnostics Check		OK	02:59
Air Blank		0.000	03:00
Control Test		0.080	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.223	03:01
Breath Volume		2.335	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:03
Subject Sample	# 2	0.209	03:04
Breath Volume		3.160	03:04
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

---

Data Download Date/Time: 10/09/2019 12:33



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SOMMERS 13120,E  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: CARMONA-ALVARENGA

DATE/TIME OF TEST: 10/04/2019 00:08:22

---

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.079	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.221	00:02
Breath Volume	2.433	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.217	00:06
Breath Volume	1.796	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

---

Data Download Date/Time: 10/09/2019 12:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SORTORE 17101,D  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: HINDS

DATE/TIME OF TEST: 10/04/2019 01:13:31

---

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.121	01:08
Breath Volume	1.363	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.130	01:12
Breath Volume	1.265	01:12
Air Blank	0.000	01:12
Control Test	0.078	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

---

Data Download Date/Time: 10/09/2019 12:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALL,CHRISTINE,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HALL,CHRISTINE  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: LOCKMAN

DATE/TIME OF TEST: 10/05/2019 01:09:23

---

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.188	01:04
Breath Volume		2.449	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.190	01:07
Breath Volume		2.335	01:07
Air Blank		0.000	01:08
Control Test		0.078	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

---

Data Download Date/Time: 10/09/2019 12:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,JOSIAH,R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SMITH,JOSIAH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: WECHSEL

DATE/TIME OF TEST: 10/05/2019 02:41:27

---

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.079	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.188	02:36
Breath Volume	1.613	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.191	02:39
Breath Volume	1.746	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

---

Data Download Date/Time: 10/09/2019 12:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALL,CHRISTINE,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HALL,CHRISTINE  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 10/05/2019 18:22:59

---

Diagnostics Check	OK	18:16
Air Blank	0.000	18:17
Control Test	0.079	18:17
Air Blank	0.000	18:18
Subject Sample # 1	0.155	18:18
Breath Volume	3.820	18:18
Air Blank	0.000	18:19
Air Blank	0.000	18:21
Subject Sample # 2	0.147	18:21
Breath Volume	3.832	18:21
Air Blank	0.000	18:22
Control Test	0.080	18:22
Air Blank	0.000	18:22
Diagnostics Check	OK	18:22

---

Data Download Date/Time: 10/09/2019 12:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, JOSIAH, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: KUPKA, R  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: HORBACH

DATE/TIME OF TEST: 10/06/2019 00:25:49

---

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.079	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.186	00:21
Breath Volume	1.246	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:23
Subject Sample # 2	0.185	00:24
Breath Volume	1.222	00:24
Air Blank	0.000	00:24
Control Test	0.079	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

---

Data Download Date/Time: 10/09/2019 12:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CAMEJO,JAZMIN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 10/06/2019 03:38:11

---

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.156	03:33
Breath Volume	2.816	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.153	03:36
Breath Volume	2.753	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

---

Data Download Date/Time: 10/09/2019 12:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SMITH,JOSIAH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: HERNANDEZ ROMERO

DATE/TIME OF TEST: 10/06/2019 03:59:00

---

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.238	03:54
Breath Volume	1.621	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.228	03:57
Breath Volume	2.082	03:57
Air Blank	0.000	03:58
Control Test	0.078	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:59

---

Data Download Date/Time: 10/09/2019 12:33



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, JOSIAH, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HIDEG, JARED  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: PAGE

DATE/TIME OF TEST: 10/06/2019 04:42:07

---

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.080	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.214	04:37
Breath Volume	2.285	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.212	04:40
Breath Volume	2.195	04:40
Air Blank	0.000	04:41
Control Test	0.079	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

---

Data Download Date/Time: 10/09/2019 12:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINS,KYLE,L  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: TELUS,JEAN  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: GALVAN

DATE/TIME OF TEST: 10/06/2019 06:41:04

---

Diagnostics Check	OK	06:35	
Air Blank	0.000	06:35	
Control Test	0.079	06:35	
Air Blank	0.000	06:36	
Subject Sample	# 1 0.106	06:37	
Breath Volume	4.617	06:37	
Air Blank	0.000	06:37	
Air Blank	0.000	06:39	
Subject Sample	# 2	06:39	Subject Test Refused
Breath Volume	0.000	06:39	
Air Blank	0.000	06:40	
Control Test	0.078	06:40	
Air Blank	0.000	06:40	
Diagnostics Check	OK	06:41	

---

Data Download Date/Time: 10/09/2019 12:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHAEFER, KENDALL, A  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: WALSH, C  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: LARA

DATE/TIME OF TEST: 10/09/2019 18:35:17

---

Diagnostics Check	OK	18:28
Air Blank	0.000	18:29
Control Test	0.080	18:29
Air Blank	0.000	18:30
Subject Sample # 1	0.163	18:30
Breath Volume	2.609	18:30
Air Blank	0.000	18:31
Air Blank	0.000	18:33
Subject Sample # 2	0.166	18:33
Breath Volume	1.964	18:33
Air Blank	0.000	18:34
Control Test	0.080	18:34
Air Blank	0.000	18:35
Diagnostics Check	OK	18:35

---

Data Download Date/Time: 11/20/2019 10:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEINER,RYAN,T  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: BEINER,RYAN  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: REY RODRIGUEZ

DATE/TIME OF TEST: 10/10/2019 04:13:12

---

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:08	
Control Test	0.079	04:08	
Air Blank	0.000	04:08	
Subject Sample # 1	0.146	04:09	
Breath Volume	2.402	04:09	
Air Blank	0.000	04:10	
Air Blank		04:11	RFI Detect
Air Blank	0.000	04:12	
Control Test	0.079	04:12	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

---

Data Download Date/Time: 11/20/2019 10:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEINER,RYAN,T  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: BEINER,RYAN  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: REY RODRIGUEZ

DATE/TIME OF TEST: 10/10/2019 06:00:07

---

Diagnostics Check	OK	05:53
Air Blank	0.000	05:54
Control Test	0.080	05:54
Air Blank	0.000	05:55
Subject Sample # 1	0.113	05:55
Breath Volume	1.523	05:55
Air Blank	0.000	05:56
Air Blank	0.000	05:58
Subject Sample # 2	0.098	05:58
Breath Volume	1.835	05:58
Air Blank	0.000	05:59
Control Test	0.079	05:59
Air Blank	0.000	06:00
Diagnostics Check	OK	06:00

---

Data Download Date/Time: 11/20/2019 10:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: WINCELOWICZ,LARRY  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: COMSTOCK

DATE/TIME OF TEST: 10/11/2019 03:27:41

---

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.081	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.234	03:19
Breath Volume		1.105	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.212	03:23
Breath Volume		2.238	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 3	0.214	03:26
Breath Volume		2.453	03:26
Air Blank		0.000	03:26
Control Test		0.078	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

---

Data Download Date/Time: 11/20/2019 10:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,EARTHEN,A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BROWN,EARTHEN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 10/12/2019 05:19:11

---

Diagnostics Check	OK	05:07
Air Blank	0.000	05:08
Control Test	0.079	05:08
Air Blank	0.000	05:09
Subject Sample # 1	0.177	05:09
Breath Volume	2.062	05:09
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.156	05:14
Breath Volume	1.453	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 3	0.168	05:17
Breath Volume	1.570	05:17
Air Blank	0.000	05:18
Control Test	0.079	05:18
Air Blank	0.000	05:19
Diagnostics Check	OK	05:19

---

Data Download Date/Time: 11/20/2019 10:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HIDEG 17018,J  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: CALVO-DIAZ

DATE/TIME OF TEST: 10/12/2019 23:30:20

---

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.079	23:24
Air Blank	0.000	23:24
Subject Sample # 1	0.156	23:25
Breath Volume	1.417	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.144	23:28
Breath Volume	1.250	23:28
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

---

Data Download Date/Time: 11/20/2019 10:41



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDIN, JEREMY, L  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SOTO, MICHAEL  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 10:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:19

SUBJECT'S LAST NAME: SALINAS JR

DATE/TIME OF TEST: 10/13/2019 11:49:09

---

Diagnostics Check		OK	11:42
Air Blank		0.000	11:43
Control Test		0.080	11:43
Air Blank		0.000	11:44
Subject Sample	# 1	0.257	11:44
Breath Volume		1.460	11:44
Air Blank		0.000	11:45
Air Blank		0.000	11:47
Subject Sample	# 2	0.256	11:47
Breath Volume		1.363	11:47
Air Blank		0.000	11:48
Control Test		0.077	11:48
Air Blank		0.000	11:49
Diagnostics Check		OK	11:49

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BRADFORD 577,H  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 10/14/2019 03:51:03

---

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.221	03:46
Breath Volume	1.953	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.206	03:49
Breath Volume	1.136	03:49
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:51

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINS,KYLE,L  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: BROWN,TERRELL  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: KOZLOWSKI

DATE/TIME OF TEST: 10/15/2019 02:46:19

---

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.200	02:41
Breath Volume	1.914	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.202	02:44
Breath Volume	2.296	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

---

Data Download Date/Time: 11/20/2019 10:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,EARTHEN,A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BROWN,EARTHEN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: RUSS

DATE/TIME OF TEST: 10/17/2019 21:36:48

---

Diagnostics Check	OK	21:30
Air Blank	0.000	21:30
Control Test	0.079	21:31
Air Blank	0.000	21:31
Subject Sample # 1	0.161	21:32
Breath Volume	1.648	21:32
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.168	21:35
Breath Volume	1.472	21:35
Air Blank	0.000	21:35
Control Test	0.080	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, JOSIAH, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SMITH, JOSIAH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: LATORRE

DATE/TIME OF TEST: 10/18/2019 22:38:38

---

Diagnostics Check	OK	22:32
Air Blank	0.000	22:33
Control Test	0.080	22:33
Air Blank	0.000	22:34
Subject Sample # 1	0.000	22:34
Breath Volume	2.199	22:34
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.000	22:37
Breath Volume	2.636	22:37
Air Blank	0.000	22:37
Control Test	0.080	22:38
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINS,KYLE,L  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: VICIOSO,YEURIS  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: LAZARTE ZURITA

DATE/TIME OF TEST: 10/19/2019 05:40:46

---

Diagnostics Check		OK	05:34
Air Blank		0.000	05:34
Control Test		0.079	05:34
Air Blank		0.000	05:35
Subject Sample	# 1	0.208	05:36
Breath Volume		1.492	05:36
Air Blank		0.000	05:36
Air Blank		0.000	05:38
Subject Sample	# 2	0.228	05:39
Breath Volume		2.554	05:39
Air Blank		0.000	05:39
Control Test		0.080	05:40
Air Blank		0.000	05:40
Diagnostics Check		OK	05:40

---

Data Download Date/Time: 11/20/2019 10:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALL,CHRISTINE,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HALL,CHRISTINE  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: COSLET

DATE/TIME OF TEST: 10/19/2019 21:43:36

---

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.077	21:38
Breath Volume	1.269	21:38
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.080	21:42
Breath Volume	1.375	21:42
Air Blank	0.000	21:42
Control Test	0.080	21:43
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HIDEG 17018,J  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: LANG

DATE/TIME OF TEST: 10/22/2019 03:32:46

---

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.169	03:28
Breath Volume	3.292	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.172	03:31
Breath Volume	3.195	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

---

Data Download Date/Time: 11/20/2019 10:42



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDIN, JEREMY, L  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SOTO, MICHAEL  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:28

SUBJECT'S LAST NAME: APARECIDO DE SOUZA

DATE/TIME OF TEST: 10/22/2019 15:03:21

---

Diagnostics Check	OK	14:56
Air Blank	0.000	14:57
Control Test	0.079	14:57
Air Blank	0.000	14:58
Subject Sample # 1	0.000	14:58
Breath Volume	1.160	14:58
Air Blank	0.000	14:58
Air Blank	0.000	15:00
Subject Sample # 2	0.000	15:01
Breath Volume	1.570	15:01
Air Blank	0.000	15:02
Control Test	0.079	15:02
Air Blank	0.000	15:03
Diagnostics Check	OK	15:03

---

Data Download Date/Time: 11/20/2019 10:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: JENSEN,BRADLEY  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: SWIFT

DATE/TIME OF TEST: 10/23/2019 22:15:10

---

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.080	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.000	22:10
Breath Volume	1.789	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.000	22:13
Breath Volume	1.421	22:13
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GONZALES,DAVID  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: PIETERSE

DATE/TIME OF TEST: 10/23/2019 23:05:18

---

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.080	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.162	22:58
Breath Volume	1.226	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.173	23:03
Breath Volume	2.523	23:03
Air Blank	0.000	23:04
Control Test	0.079	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: NORMAN,CHARLES  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 10/25/2019 00:41:49

---

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.181	00:37
Breath Volume	1.308	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.199	00:40
Breath Volume	1.375	00:40
Air Blank	0.000	00:40
Control Test	0.080	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

---

Data Download Date/Time: 11/20/2019 10:42



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDIN,JEEMY,L  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SUTPHIN,STEPHEN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: BHARATH

DATE/TIME OF TEST: 10/26/2019 05:50:28

---

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.079	05:45
Air Blank	0.000	05:45
Subject Sample # 1	0.190	05:46
Breath Volume	1.703	05:46
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.181	05:48
Breath Volume	1.242	05:48
Air Blank	0.000	05:49
Control Test	0.079	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

---

Data Download Date/Time: 11/20/2019 10:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,JOSIAH,R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SMITH,JOSIAH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MANCHESTER

DATE/TIME OF TEST: 10/28/2019 18:30:32

---

Diagnostics Check	OK	18:23
Air Blank	0.000	18:24
Control Test	0.079	18:24
Air Blank	0.000	18:24
Subject Sample # 1	0.171	18:25
Breath Volume	1.773	18:25
Air Blank	0.000	18:26
Air Blank	0.000	18:28
Subject Sample # 2	0.178	18:29
Breath Volume	1.750	18:29
Air Blank	0.000	18:29
Control Test	0.079	18:29
Air Blank	0.000	18:30
Diagnostics Check	OK	18:30

---

Data Download Date/Time: 11/20/2019 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, JOSIAH, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: TOMPKINA, KENDALL  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 10/30/2019 03:20:30

---

Diagnostics Check		OK	03:14
Air Blank		0.000	03:14
Control Test		0.080	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.066	03:16
Breath Volume		2.089	03:16
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.063	03:19
Breath Volume		2.160	03:19
Air Blank		0.000	03:19
Control Test		0.078	03:19
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

---

Data Download Date/Time: 11/20/2019 10:42



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BRADFORD #577,H  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: FERNANDEZ-PAREDES

DATE/TIME OF TEST: 10/31/2019 03:56:31

---

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.228	03:51
Breath Volume		1.628	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.227	03:55
Breath Volume		1.417	03:55
Air Blank		0.000	03:55
Control Test		0.079	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

---

Data Download Date/Time: 11/20/2019 10:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOGES,KEVIN,R  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: ELLISON,DAVID  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 11:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: LOGES

DATE/TIME OF TEST: 10/31/2019 05:41:06

---

Diagnostics Check	OK	05:40	
Air Blank		05:40	RFI Detect
Air Blank		05:41	RFI Detect

---

Data Download Date/Time: 11/20/2019 10:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELLISON,DAVID,J  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: ELLISON,DAVID  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: LOGES

DATE/TIME OF TEST: 10/31/2019 05:46:47

---

Diagnostics Check	OK	05:43	
Air Blank	0.000	05:44	
Control Test	0.080	05:44	
Air Blank	0.000	05:45	
Subject Sample	# 1	05:45	Subject Test Refused
Breath Volume	0.000	05:45	
Air Blank	0.000	05:45	
Control Test	0.079	05:46	
Air Blank		05:46	RFI Detect
Air Blank	0.000	05:46	

---

Data Download Date/Time: 11/20/2019 10:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BONSALL, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BONSALL, M  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/31/2019 20:32:01

---

Diagnostics Check	OK	20:25
Air Blank	0.000	20:25
Control Test	0.079	20:25
Air Blank	0.000	20:26
Subject Sample # 1	0.083	20:27
Breath Volume	2.000	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.085	20:30
Breath Volume	1.898	20:30
Air Blank	0.000	20:31
Control Test	0.079	20:31
Air Blank	0.000	20:31
Diagnostics Check	OK	20:32

---

Data Download Date/Time: 11/20/2019 10:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINDHOLM,JOHN,T  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LINDHOLM,JOHN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CASH RAMIREZ

DATE/TIME OF TEST: 10/14/2019 04:01:51

---

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.081	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.194	03:57
Breath Volume	1.769	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.197	04:00
Breath Volume	2.003	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

---

Data Download Date/Time: 11/20/2019 10:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STROBEL, RAYMOND, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: STROBEL, RAYMOND  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 10/15/2019 15:57:01

---

Diagnostics Check	OK	15:53	
Air Blank	0.000	15:54	
Control Test	0.080	15:54	
Air Blank	0.000	15:55	
Subject Sample	# 1	15:55	Subject Test Refused
Breath Volume	0.000	15:55	
Air Blank	0.000	15:56	
Control Test	0.079	15:56	
Air Blank	0.000	15:56	
Diagnostics Check	OK	15:57	

---

Data Download Date/Time: 11/20/2019 10:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STROBEL, RAYMOND, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: STROBEL, RAYMOND  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 10/15/2019 16:09:18

---

Diagnostics Check	OK	16:02
Air Blank	0.000	16:02
Control Test	0.080	16:02
Air Blank	0.000	16:03
Subject Sample # 1	0.211	16:04
Breath Volume	2.195	16:04
Air Blank	0.000	16:04
Air Blank	0.000	16:06
Subject Sample # 2	0.200	16:07
Breath Volume	2.035	16:07
Air Blank	0.000	16:08
Control Test	0.080	16:08
Air Blank	0.000	16:09
Diagnostics Check	OK	16:09

---

Data Download Date/Time: 11/20/2019 10:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STROBEL, RAYMOND, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CHERNOFF, R  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 10/15/2019 17:24:56

---

Diagnostics Check	OK	17:17
Air Blank	0.000	17:18
Control Test	0.080	17:18
Air Blank	0.000	17:18
Subject Sample # 1	0.122	17:19
Breath Volume	1.558	17:19
Air Blank	0.000	17:20
Air Blank	0.000	17:22
Subject Sample # 2	0.120	17:23
Breath Volume	1.136	17:23
Air Blank	0.000	17:24
Control Test	0.080	17:24
Air Blank	0.000	17:24
Diagnostics Check	OK	17:24

---

Data Download Date/Time: 11/20/2019 10:29



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOULD,LEONARD,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GOULD,LEONARD  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:11

SUBJECT'S LAST NAME: LA MARCH

DATE/TIME OF TEST: 10/16/2019 15:37:32

---

Diagnostics Check	OK	15:31
Air Blank	0.000	15:31
Control Test	0.081	15:32
Air Blank	0.000	15:32
Subject Sample # 1	0.051	15:33
Breath Volume	1.996	15:33
Air Blank	0.000	15:33
Air Blank	0.000	15:35
Subject Sample # 2	0.051	15:35
Breath Volume	2.085	15:35
Air Blank	0.000	15:36
Control Test	0.080	15:36
Air Blank	0.000	15:37
Diagnostics Check	OK	15:37

---

Data Download Date/Time: 11/20/2019 10:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINDHOLM,JOHN,T  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LINDHOLM,JOHN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: DUMAY-GARDINER

DATE/TIME OF TEST: 10/25/2019 21:25:38

---

Diagnostics Check		OK	21:15
Air Blank		0.000	21:16
Control Test		0.080	21:16
Air Blank		0.000	21:17
Subject Sample	# 1	0.188	21:17
Breath Volume		2.382	21:17
Air Blank		0.000	21:18
Air Blank		0.000	21:20
Subject Sample	# 2	0.213	21:20
Breath Volume		2.257	21:20
Air Blank		0.000	21:21
Air Blank		0.000	21:23
Subject Sample	# 3	0.213	21:24
Breath Volume		1.988	21:24
Air Blank		0.000	21:24
Control Test		0.078	21:25
Air Blank		0.000	21:25
Diagnostics Check		OK	21:25

---

Data Download Date/Time: 11/20/2019 10:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: DAVIS,DANIEL  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BRUCE

DATE/TIME OF TEST: 10/25/2019 21:46:53

---

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:40
Subject Sample # 1	0.279	21:41
Breath Volume	1.210	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:43
Subject Sample # 2	0.269	21:45
Breath Volume	2.250	21:45
Air Blank	0.000	21:45
Control Test	0.079	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

---

Data Download Date/Time: 11/20/2019 10:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINDHOLM,JOHN,T  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: ROSE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 10/25/2019 22:02:09

---

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.078	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.000	21:57
Breath Volume	2.039	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.000	22:00
Breath Volume	1.324	22:00
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

---

Data Download Date/Time: 11/20/2019 10:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BONSALL, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BONSALL, M  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SISONGKHAM

DATE/TIME OF TEST: 10/25/2019 22:12:01

---

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.079	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.188	22:07
Breath Volume	2.089	22:07
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.175	22:10
Breath Volume	2.085	22:10
Air Blank	0.000	22:11
Control Test	0.079	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:12

---

Data Download Date/Time: 11/20/2019 10:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOULD, LEONARD,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HIDEG, JARED  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: BOXENBAUM

DATE/TIME OF TEST: 10/26/2019 22:15:59

---

Diagnostics Check	OK	22:09
Air Blank	0.000	22:10
Control Test	0.080	22:10
Air Blank	0.000	22:10
Subject Sample # 1	0.180	22:11
Breath Volume	2.074	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:13
Subject Sample # 2	0.181	22:14
Breath Volume	1.441	22:14
Air Blank	0.000	22:15
Control Test	0.079	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

---

Data Download Date/Time: 11/20/2019 10:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BONSALL, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: ECKERT, K  
ARRESTING OFFICER AGENCY: Sanibel PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ARANHA

DATE/TIME OF TEST: 10/27/2019 01:32:45

---

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.079	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.127	01:27
Breath Volume	1.855	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.125	01:31
Breath Volume	1.671	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

---

Data Download Date/Time: 11/20/2019 10:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSS,CHEYANNE,K  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: OHERN,ANDREW  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: LOSACANO

DATE/TIME OF TEST: 10/02/2019 00:19:43

---

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:15	
Control Test	0.080	00:15	
Air Blank	0.000	00:16	
Subject Sample	# 1	00:18	Subject Test Refused
Breath Volume	0.000	00:18	
Air Blank	0.000	00:18	
Control Test	0.081	00:19	
Air Blank	0.000	00:19	
Diagnostics Check	OK	00:19	

---

Data Download Date/Time: 10/09/2019 08:43



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: ANDERSEN

DATE/TIME OF TEST: 10/03/2019 01:44:35

---

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.080	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.200	01:37
Breath Volume		1.214	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:39
Subject Sample	# 2	0.205	01:42
Breath Volume		2.031	01:42
Air Blank		0.000	01:43
Control Test		0.080	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

---

Data Download Date/Time: 10/09/2019 08:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: VOORHEES

DATE/TIME OF TEST: 10/04/2019 00:40:02

---

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:35	
Control Test	0.080	00:35	
Air Blank	0.000	00:35	
Subject Sample # 1	0.000	00:36	
Breath Volume	3.066	00:36	
Air Blank	0.000	00:36	
Air Blank	0.000	00:38	
Subject Sample # 2	0.000	00:39	
Breath Volume	3.238	00:39	
Air Blank		00:39	RFI Detect
Air Blank		00:40	RFI Detect

---

Data Download Date/Time: 10/09/2019 08:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSH,DANIEL,C  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: RUSH,D  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: GARNETT

DATE/TIME OF TEST: 10/04/2019 03:12:13

---

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.131	03:07
Breath Volume	1.425	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:09
Subject Sample # 2	0.143	03:10
Breath Volume	2.226	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

---

Data Download Date/Time: 10/09/2019 08:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINI,RYAN,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MARTINI,RYAN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DAUGHTREY

DATE/TIME OF TEST: 10/05/2019 02:29:33

---

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.081	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.000	02:25
Breath Volume	2.421	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.000	02:28
Breath Volume	2.039	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

---

Data Download Date/Time: 10/09/2019 08:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: BARKLEY,JEFFREY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 10/06/2019 04:30:41

---

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:26	
Control Test	0.080	04:26	
Air Blank	0.000	04:27	
Subject Sample	# 1	04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:29	
Control Test	0.080	04:30	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

---

Data Download Date/Time: 10/09/2019 08:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FUGIEL,ADRIAN,M  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: LOETHEN,KEITH  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 09/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: LAUER

DATE/TIME OF TEST: 10/07/2019 00:58:55

---

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.000	00:54
Breath Volume	1.714	00:54
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.000	00:57
Breath Volume	1.644	00:57
Air Blank	0.000	00:57
Control Test	0.080	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

---

Data Download Date/Time: 10/09/2019 08:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: BARKLEY,JEFFREY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: OTHER  
ARREST TIME: 07:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 10/09/2019 09:29:57

---

Diagnostics Check	OK	09:24
Air Blank	0.000	09:24
Control Test	0.081	09:24
Air Blank	0.000	09:25
Subject Sample # 1	0.000	09:25
Breath Volume	3.347	09:25
Air Blank	0.000	09:26
Air Blank	0.000	09:28
Subject Sample # 2	0.000	09:28
Breath Volume	3.296	09:28
Air Blank	0.000	09:29
Control Test	0.080	09:29
Air Blank	0.000	09:29
Diagnostics Check	OK	09:29

---

Data Download Date/Time: 11/20/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: DELUCIA

DATE/TIME OF TEST: 10/11/2019 22:45:57

---

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.081	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.144	22:41
Breath Volume	3.601	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:43
Subject Sample # 2	0.139	22:44
Breath Volume	3.085	22:44
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

---

Data Download Date/Time: 11/20/2019 10:37



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/12/2019 02:40:52

---

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:29	
Control Test	0.082	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:33	No Sample Provided
Breath Volume	0.000	02:33	
Air Blank	0.000	02:34	
Air Blank	0.000	02:36	
Subject Sample	# 2	02:39	No Sample Provided
Breath Volume	0.000	02:39	
Air Blank	0.000	02:39	
Control Test	0.080	02:40	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

---

Data Download Date/Time: 11/20/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,ALFONSO,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: DURHAM,KARI  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: RIVERI

DATE/TIME OF TEST: 10/12/2019 19:07:25

---

Diagnostics Check	OK	18:59
Air Blank	0.000	19:00
Control Test	0.080	19:00
Air Blank	0.000	19:01
Subject Sample # 1	0.100	19:02
Breath Volume	1.292	19:02
Air Blank	0.000	19:03
Air Blank	0.000	19:04
Subject Sample # 2	0.101	19:05
Breath Volume	1.546	19:05
Air Blank	0.000	19:06
Control Test	0.079	19:06
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

---

Data Download Date/Time: 11/20/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: PALUMBO

DATE/TIME OF TEST: 10/13/2019 02:19:10

---

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.126	02:14
Breath Volume	2.585	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:16
Subject Sample # 2	0.128	02:17
Breath Volume	3.015	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

---

Data Download Date/Time: 11/20/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: SCARLATO,FRANK  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: LOWERY

DATE/TIME OF TEST: 10/13/2019 02:39:33

---

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.079	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.137	02:34
Breath Volume	3.050	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.143	02:37
Breath Volume	3.937	02:37
Air Blank	0.000	02:38
Control Test	0.079	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

---

Data Download Date/Time: 11/20/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZARRILLO,N,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: DOTSON,W  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 10/14/2019 03:28:02

---

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.080	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.175	03:23
Breath Volume		2.203	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.163	03:26
Breath Volume		2.500	03:26
Air Blank		0.000	03:27
Control Test		0.079	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:28

---

Data Download Date/Time: 11/20/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINI,RYAN,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MARTINI,RYAN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: BABCOCK

DATE/TIME OF TEST: 10/16/2019 06:29:07

---

Diagnostics Check	OK	06:22
Air Blank	0.000	06:23
Control Test	0.081	06:23
Air Blank	0.000	06:24
Subject Sample # 1	0.000	06:24
Breath Volume	1.812	06:24
Air Blank	0.000	06:25
Air Blank	0.000	06:27
Subject Sample # 2	0.000	06:27
Breath Volume	1.761	06:27
Air Blank	0.000	06:28
Control Test	0.081	06:28
Air Blank	0.000	06:29
Diagnostics Check	OK	06:29

---

Data Download Date/Time: 11/20/2019 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATO,MARK,P  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: CABRAL,MATTHEW  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: BODENMANN

DATE/TIME OF TEST: 10/16/2019 13:38:23

---

Diagnostics Check	OK	13:31
Air Blank	0.000	13:32
Control Test	0.080	13:32
Air Blank	0.000	13:33
Subject Sample # 1	0.146	13:33
Breath Volume	2.195	13:33
Air Blank	0.000	13:34
Air Blank	0.000	13:36
Subject Sample # 2	0.136	13:36
Breath Volume	2.214	13:36
Air Blank	0.000	13:37
Control Test	0.080	13:37
Air Blank	0.000	13:38
Diagnostics Check	OK	13:38

---

Data Download Date/Time: 11/20/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUFFY,J,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: COLWELL,C  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: DALBORA

DATE/TIME OF TEST: 10/17/2019 02:54:38

---

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.177	02:49
Breath Volume	1.328	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.180	02:52
Breath Volume	1.613	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

---

Data Download Date/Time: 11/20/2019 10:37



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUFFY,J,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: THOMAS,D  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: BELLMAN

DATE/TIME OF TEST: 10/17/2019 03:38:08

---

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.215	03:32
Breath Volume	1.144	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.219	03:36
Breath Volume	1.542	03:36
Air Blank	0.000	03:37
Control Test	0.078	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

---

Data Download Date/Time: 11/20/2019 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,SAMANTHA,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: CARVALHO,ASHLEY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SHOEMAKER

DATE/TIME OF TEST: 10/20/2019 02:48:24

---

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:36	
Control Test	0.080	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:40	No Sample Provided
Breath Volume	0.000	02:40	
Air Blank	0.000	02:40	
Air Blank	0.000	02:42	
Subject Sample	# 2	0.221	02:43
Breath Volume	1.535	02:43	
Air Blank	0.000	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 3	0.221	02:46
Breath Volume	2.394	02:46	
Air Blank	0.000	02:47	
Control Test	0.080	02:47	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

---

Data Download Date/Time: 11/20/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUFFY,J,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: BALL,P  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: YEHUDA

DATE/TIME OF TEST: 10/22/2019 23:42:26

---

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.080	23:31	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:35	No Sample Provided
Breath Volume	0.000	23:35	
Air Blank	0.000	23:35	
Air Blank	0.000	23:37	
Subject Sample	# 2	23:41	No Sample Provided
Breath Volume	0.000	23:41	
Air Blank	0.000	23:41	
Control Test	0.080	23:41	
Air Blank	0.000	23:42	
Diagnostics Check	OK	23:42	

---

Data Download Date/Time: 11/20/2019 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,SAMANTHA,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: CARVALHO,ASHLEY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 10/23/2019 23:59:04

---

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:46	
Control Test	0.080	23:47	
Air Blank	0.000	23:47	
Subject Sample	# 1 0.186	23:49	
Breath Volume	1.136	23:49	
Air Blank	0.000	23:49	
Air Blank	0.000	23:51	
Subject Sample	# 2 0.168	23:54	Slope Not Level
Breath Volume	1.218	23:54	
Air Blank	0.000	23:55	
Air Blank	0.000	23:57	
Subject Sample	# 3	23:57	Subject Test Refused
Breath Volume	0.000	23:57	
Air Blank	0.000	23:58	
Control Test	0.079	23:58	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:59	

---

Data Download Date/Time: 11/20/2019 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLS, MATTHEW, K  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: ROSALES, PATRICK  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: GIANCOLA

DATE/TIME OF TEST: 10/25/2019 01:15:12

---

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:12	
Control Test	0.080	01:13	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:13	Subject Test Refused
Breath Volume	0.000	01:13	
Air Blank	0.000	01:14	
Control Test	0.080	01:14	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

---

Data Download Date/Time: 11/20/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: FORYST,JADE  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 10/26/2019 02:17:03

---

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.080	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.090	02:12
Breath Volume	1.738	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.086	02:15
Breath Volume	1.343	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:17

---

Data Download Date/Time: 11/20/2019 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUFFY,J,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: LOWE,A  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: PEPLINSKI

DATE/TIME OF TEST: 10/27/2019 20:48:27

---

Diagnostics Check	OK	20:41
Air Blank	0.000	20:42
Control Test	0.080	20:42
Air Blank	0.000	20:43
Subject Sample # 1	0.120	20:43
Breath Volume	3.246	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.116	20:46
Breath Volume	3.035	20:46
Air Blank	0.000	20:47
Control Test	0.082	20:47
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

---

Data Download Date/Time: 11/20/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,SAMANTHA,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: SONG,GUANG  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: KUNTZ

DATE/TIME OF TEST: 10/29/2019 23:05:28

---

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.080	22:59
Air Blank	0.000	23:00
Subject Sample # 1	0.158	23:00
Breath Volume	1.710	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.159	23:03
Breath Volume	2.113	23:03
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

---

Data Download Date/Time: 11/20/2019 10:37



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: DUPUY

DATE/TIME OF TEST: 10/31/2019 03:02:26

---

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.143	02:57
Breath Volume	2.851	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.142	03:00
Breath Volume	2.957	03:00
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

---

Data Download Date/Time: 11/20/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: SCARLATO,MICHAEL  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: CORAL MORALES

DATE/TIME OF TEST: 10/31/2019 03:56:09

---

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:45	
Control Test	0.080	03:45	
Air Blank	0.000	03:45	
Subject Sample	# 1	03:49	No Sample Provided
Breath Volume	0.000	03:49	
Air Blank	0.000	03:49	
Air Blank	0.000	03:51	
Subject Sample	# 2	03:54	No Sample Provided
Breath Volume	0.000	03:54	
Air Blank	0.000	03:55	
Control Test	0.080	03:55	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

---

Data Download Date/Time: 11/20/2019 10:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, ALFONSO,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: COLWELL, CHASE  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/31/2019 04:29:47

---

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.080	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.050	04:24
Breath Volume		2.027	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.043	04:28
Breath Volume		1.164	04:28
Air Blank		0.000	04:28
Control Test		0.080	04:29
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

---

Data Download Date/Time: 11/20/2019 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,DANNY,W  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: OCHS,DANIEL  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: ESTEVA

DATE/TIME OF TEST: 10/31/2019 19:50:11

---

Diagnostics Check	OK	19:43
Air Blank	0.000	19:44
Control Test	0.080	19:44
Air Blank	0.000	19:44
Subject Sample # 1	0.000	19:45
Breath Volume	4.160	19:45
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 2	0.000	19:48
Breath Volume	4.375	19:48
Air Blank	0.000	19:49
Control Test	0.079	19:49
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

---

Data Download Date/Time: 11/20/2019 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPEZ,JOSE,D  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: LOPEZ,JOSE  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 10/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/08/2019 03:30:39

---

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.282	03:25
Breath Volume	3.187	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.276	03:28
Breath Volume	1.660	03:28
Air Blank	0.000	03:29
Control Test	0.079	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

---

Data Download Date/Time: 11/04/2019 17:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYKHASKIV,MICHAEL,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: VASQUEZ,DANIEL  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 10/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: HINOJOSA

DATE/TIME OF TEST: 10/31/2019 22:45:28

---

Diagnostics Check	OK	22:39
Air Blank	0.000	22:39
Control Test	0.079	22:40
Air Blank	0.000	22:40
Subject Sample # 1	0.040	22:41
Breath Volume	2.394	22:41
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.033	22:43
Breath Volume	1.351	22:43
Air Blank	0.000	22:44
Control Test	0.078	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

---

Data Download Date/Time: 11/04/2019 17:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUNN,CDFC,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: DAVIDSON,M  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 09/25/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: DREW

DATE/TIME OF TEST: 10/01/2019 03:56:19

---

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.078	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.183	03:51
Breath Volume	1.980	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:53
Subject Sample # 2	0.178	03:54
Breath Volume	1.890	03:54
Air Blank	0.000	03:55
Control Test	0.078	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

---

Data Download Date/Time: 12/05/2019 14:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EPPS,QUINTELLA,  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: CLENDING,JAMES  
ARRESTING OFFICER AGENCY: Glades County SO  
LAST AGENCY INSPECTION: 10/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A2  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 10/24/2019 00:55:16

---

Diagnostics Check		OK	00:46
Air Blank		0.000	00:47
Control Test		0.075	00:47
Air Blank		0.000	00:47
Subject Sample	# 1	0.136	00:50
Breath Volume		2.203	00:50
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.132	00:53
Breath Volume		1.113	00:53
Air Blank		0.000	00:54
Control Test		0.075	00:54
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

---

Data Download Date/Time: 10/26/2019 12:11



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KELLER, DAVID,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: KELLER, DAVID  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/03/2019 02:38:07

---

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.079	02:32	
Air Blank	0.000	02:33	
Subject Sample	# 1 0.286	02:34	
Breath Volume	1.226	02:34	
Air Blank	0.000	02:34	
Air Blank	0.000	02:36	
Subject Sample	# 2	02:36	RFI Detect
Breath Volume	0.000	02:36	
Air Blank	0.000	02:37	
Control Test	0.079	02:37	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

---

Data Download Date/Time: 10/30/2019 21:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KELLER,DAVID,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: KELLER,DAVID  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 10/04/2019 23:11:19

---

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:08	
Control Test	0.079	23:08	
Air Blank	0.000	23:09	
Subject Sample	# 1	23:09	Subject Test Refused
Breath Volume	0.000	23:09	
Air Blank	0.000	23:10	
Control Test	0.079	23:10	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

---

Data Download Date/Time: 10/30/2019 21:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KELLER,DAVID,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: KELLER,DAVID  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ELLISON

DATE/TIME OF TEST: 10/05/2019 04:06:24

---

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.079	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.175	04:02
Breath Volume		1.261	04:02
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.181	04:04
Breath Volume		1.164	04:04
Air Blank		0.000	04:05
Control Test		0.079	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

---

Data Download Date/Time: 10/30/2019 21:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KELLER,DAVID,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: KELLER,DAVID  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: STREET

DATE/TIME OF TEST: 10/10/2019 04:52:25

---

Diagnostics Check	OK	04:51	
Air Blank		04:52	RFI Detect
Air Blank		04:52	RFI Detect

---

Data Download Date/Time: 10/30/2019 21:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KELLER,DAVID,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: KELLER,DAVID  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: STREET

DATE/TIME OF TEST: 10/10/2019 04:59:08

---

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:57	
Control Test	0.080	04:58	
Air Blank	0.000	04:58	
		04:58	RFI Detect
Air Blank	0.000	04:59	

---

Data Download Date/Time: 10/30/2019 21:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KELLER,DAVID,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: KELLER,DAVID  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 10/11/2019 02:36:38

---

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:35	
Control Test	0.079	02:35	
Air Blank		02:36	RFI Detect
Air Blank	0.000	02:36	

---

Data Download Date/Time: 10/30/2019 21:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KELLER, DAVID,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: 1L34, KELLER  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 10/11/2019 02:43:38

---

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:40	
Control Test	0.079	02:40	
Air Blank	0.000	02:41	
Subject Sample # 1	0.000	02:41	
Breath Volume	2.148	02:41	
Air Blank		02:41	RFI Detect
Air Blank	0.000	02:42	
Control Test	0.080	02:42	
Air Blank	0.000	02:43	
Diagnostics Check	Fail	02:43	Voltage/Current Test
Air Blank	0.000	02:43	

---

Data Download Date/Time: 10/30/2019 21:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KELLER,DAVID,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: KELLER,DAVID  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 10/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: LY

DATE/TIME OF TEST: 10/31/2019 03:06:03

---

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:00	
Control Test	0.078	03:00	
Air Blank	0.000	03:01	
Subject Sample	# 1 0.191	03:01	
Breath Volume	1.363	03:01	
Air Blank	0.000	03:02	
Air Blank	0.000	03:04	
Subject Sample	# 2 0.197	03:05	
Breath Volume	1.898	03:05	
Air Blank		03:05	RFI Detect
Air Blank		03:06	RFI Detect

---

Data Download Date/Time: 11/25/2019 09:42



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NELSON, THOMAS, E  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: SWINDELL, SEAN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 09/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: BUI  
ARREST TIME: 19:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 10/13/2019 20:52:18

---

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:47	
Control Test	0.077	20:47	
Air Blank	0.000	20:48	
Subject Sample	# 1	20:50	Subject Test Refused
Breath Volume	0.000	20:50	
Air Blank	0.000	20:51	
Control Test	0.078	20:51	
Air Blank		20:51	RFI Detect
Air Blank	0.000	20:52	

---

Data Download Date/Time: 10/14/2019 11:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NELSON, THOMAS, E  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: SWINDELL, SEAN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 09/17/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: BUI  
ARREST TIME: 19:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 10/13/2019 20:59:32

---

Diagnostics Check	OK	20:56	
Air Blank	0.000	20:57	
Control Test	0.077	20:57	
Air Blank	0.000	20:57	
Subject Sample	# 1	20:58	Subject Test Refused
Breath Volume	0.000	20:58	
Air Blank	0.000	20:58	
Control Test	0.077	20:58	
Air Blank	0.000	20:59	
Diagnostics Check	OK	20:59	

---

Data Download Date/Time: 10/14/2019 11:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARREN,BRIAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: ADAMCZYK,ANDREW  
ARRESTING OFFICER AGENCY: Jefferson County SO  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 10/12/2019 23:43:48

---

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.078	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.077	23:39
Breath Volume	1.492	23:39
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.076	23:42
Breath Volume	1.863	23:42
Air Blank	0.000	23:42
Control Test	0.078	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

---

Data Download Date/Time: 11/19/2019 10:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FOLSOM, CHRISTOPHER, M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: ADAMCZYK, ANDY  
ARRESTING OFFICER AGENCY: Jefferson County SO  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: GEATHERS

DATE/TIME OF TEST: 10/27/2019 00:57:22

---

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.077	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.104	00:52
Breath Volume		2.449	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.107	00:55
Breath Volume		2.687	00:55
Air Blank		0.000	00:56
Control Test		0.078	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

---

Data Download Date/Time: 11/19/2019 10:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FOLSOM, CHRISTOPHER, M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: ADAMCZYK, ANDY  
ARRESTING OFFICER AGENCY: Jefferson County SO  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: GEATHERS

DATE/TIME OF TEST: 10/27/2019 01:15:29

---

Diagnostics Check		OK	01:07
Air Blank		0.000	01:08
Control Test		0.078	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.103	01:09
Breath Volume		1.390	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.101	01:13
Breath Volume		3.179	01:13
Air Blank		0.000	01:14
Control Test		0.078	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

---

Data Download Date/Time: 11/19/2019 10:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FOLSOM, CHRISTOPHER, M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: ADAMCZYK, ANDY  
ARRESTING OFFICER AGENCY: Jefferson County SO  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 10/27/2019 04:23:02

---

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.078	04:16
Air Blank		0.000	04:16
Subject Sample	# 1	0.208	04:17
Breath Volume		1.363	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.212	04:21
Breath Volume		2.296	04:21
Air Blank		0.000	04:22
Control Test		0.078	04:22
Air Blank		0.000	04:22
Diagnostics Check		OK	04:23

---

Data Download Date/Time: 11/19/2019 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARREN,BRIAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: PEACOCK,JAMES  
ARRESTING OFFICER AGENCY: Jefferson County SO  
LAST AGENCY INSPECTION: 10/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 10/31/2019 19:24:28

---

Diagnostics Check	OK	19:17
Air Blank	0.000	19:18
Control Test	0.078	19:18
Air Blank	0.000	19:19
Subject Sample # 1	0.000	19:20
Breath Volume	2.808	19:20
Air Blank	0.000	19:20
Air Blank	0.000	19:22
Subject Sample # 2	0.000	19:23
Breath Volume	2.023	19:23
Air Blank	0.000	19:23
Control Test	0.078	19:23
Air Blank	0.000	19:24
Diagnostics Check	OK	19:24

---

Data Download Date/Time: 11/19/2019 10:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STROHECKER,KYNDAL,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: STROHECKER,KYNDAL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 09/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GOWDY

DATE/TIME OF TEST: 10/07/2019 02:01:02

---

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.083	01:56	
Air Blank	0.000	01:57	
Subject Sample	# 1 0.165	01:59	
Breath Volume	1.347	01:59	
Air Blank		02:00	Purge Fail
Air Blank	0.000	02:01	

---

Data Download Date/Time: 10/15/2019 13:07



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STROHECKER,KYNDAL,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: STROHECKER,KYNDAL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 09/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GOWDY

DATE/TIME OF TEST: 10/07/2019 02:09:51

---

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:03	
Control Test	0.081	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1 0.172	02:04	
Breath Volume	1.285	02:04	
Air Blank	0.000	02:05	
Air Blank	0.000	02:07	
Subject Sample	# 2	02:08	RFI Detect
Breath Volume	1.261	02:08	
Air Blank	0.000	02:08	
Control Test	0.081	02:09	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

---

Data Download Date/Time: 10/15/2019 13:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STROHECKER,KYNDAL,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: STROHECKER,KYNDAL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 09/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GOWDY

DATE/TIME OF TEST: 10/07/2019 02:12:14

---

Diagnostics Check	OK	02:11	
Air Blank		02:11	RFI Detect
Air Blank	0.000	02:12	

---

Data Download Date/Time: 10/15/2019 13:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STROHECKER,KYNDAL,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: STROHECKER,KYNDAL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 09/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GOWDY

DATE/TIME OF TEST: 10/07/2019 02:29:47

---

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:15	
Control Test	0.081	02:15	
Air Blank	0.000	02:15	
Subject Sample # 1	0.169	02:16	
Breath Volume	1.285	02:16	
Air Blank	0.000	02:17	
Air Blank	0.000	02:19	
Subject Sample # 2	0.155	02:22	Volume Not Met
Breath Volume	0.781	02:22	
Air Blank	0.000	02:23	
Air Blank	0.000	02:25	
Subject Sample # 3	0.158	02:28	Slope Not Level
Breath Volume	1.117	02:28	
Air Blank	0.000	02:28	
Control Test	0.080	02:29	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

---

Data Download Date/Time: 10/15/2019 13:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STROHECKER,KYNDAL,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: STROHECKER,KYNDAL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 09/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: GOWDY

DATE/TIME OF TEST: 10/07/2019 03:22:50

---

Diagnostics Check		OK	03:11	
Air Blank		0.000	03:11	
Control Test		0.082	03:12	
Air Blank		0.000	03:12	
Subject Sample	# 1	0.146	03:15	Volume Not Met
Breath Volume		1.019	03:15	
Air Blank		0.000	03:16	
Air Blank		0.000	03:18	
Subject Sample	# 2	0.153	03:21	Slope Not Level
Breath Volume		1.304	03:21	
Air Blank		0.000	03:21	
Control Test		0.081	03:22	
Air Blank		0.000	03:22	
Diagnostics Check		OK	03:22	

---

Data Download Date/Time: 10/15/2019 13:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STROHECKER,KYNDAL,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: STROHECKER,KYNDAL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 09/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: GOWDY

DATE/TIME OF TEST: 10/07/2019 03:32:16

---

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.155	03:26
Breath Volume	1.402	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.149	03:30
Breath Volume	1.316	03:30
Air Blank	0.000	03:31
Control Test	0.083	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

---

Data Download Date/Time: 10/15/2019 13:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLS,ARMAND,G  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SANTTI,ROBERT  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: ADMIN  
ARREST TIME: 13:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:37

SUBJECT'S LAST NAME: WAESCO

DATE/TIME OF TEST: 10/02/2019 15:06:10

---

Diagnostics Check	OK	15:00
Air Blank	0.000	15:00
Control Test	0.082	15:00
Air Blank	0.000	15:01
Subject Sample # 1	0.041	15:01
Breath Volume	3.414	15:01
Air Blank	0.000	15:02
Air Blank	0.000	15:04
Subject Sample # 2	0.041	15:04
Breath Volume	3.691	15:04
Air Blank	0.000	15:05
Control Test	0.080	15:05
Air Blank	0.000	15:06
Diagnostics Check	OK	15:06

---

Data Download Date/Time: 10/03/2019 16:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: TRUONG,DAT  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: MORALESVAZQUEZ

DATE/TIME OF TEST: 10/04/2019 00:12:37

---

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.082	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.229	00:08
Breath Volume	1.550	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.234	00:11
Breath Volume	1.125	00:11
Air Blank	0.000	00:11
Control Test	0.081	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

---

Data Download Date/Time: 10/24/2019 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SIMPSON,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: GUERRERO-VIDAL

DATE/TIME OF TEST: 10/04/2019 22:14:45

---

Diagnostics Check		OK	22:03	
Air Blank		0.000	22:03	
Control Test		0.083	22:03	
Air Blank		0.000	22:04	
Subject Sample	# 1	0.090	22:07	Volume Not Met
Breath Volume		0.589	22:07	
Air Blank		0.000	22:08	
Air Blank		0.000	22:09	
Subject Sample	# 2	0.077	22:10	
Breath Volume		1.925	22:10	
Air Blank		0.000	22:10	
Air Blank		0.000	22:12	
Subject Sample	# 3	0.082	22:13	
Breath Volume		1.957	22:13	
Air Blank		0.000	22:13	
Control Test		0.081	22:14	
Air Blank		0.000	22:14	
Diagnostics Check		OK	22:14	

---

Data Download Date/Time: 10/24/2019 14:41



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DUCOS, CHRISTIAN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: BIEN AIME

DATE/TIME OF TEST: 10/05/2019 00:20:22

---

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.082	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.046	00:16
Breath Volume		2.296	00:16
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.042	00:18
Breath Volume		1.351	00:18
Air Blank		0.000	00:19
Control Test		0.081	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

---

Data Download Date/Time: 10/24/2019 14:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HOLIDAY,LUCAS  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 10/05/2019 02:44:32

---

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.082	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.000	02:40
Breath Volume	1.644	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.000	02:43
Breath Volume	1.343	02:43
Air Blank	0.000	02:43
Control Test	0.082	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

---

Data Download Date/Time: 10/24/2019 14:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HARMAN, RICHARD  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: CASELLAU

DATE/TIME OF TEST: 10/05/2019 15:56:05

---

Diagnostics Check	OK	15:49
Air Blank	0.000	15:50
Control Test	0.081	15:50
Air Blank	0.000	15:50
Subject Sample # 1	0.302	15:51
Breath Volume	1.886	15:51
Air Blank	0.000	15:51
Air Blank	0.000	15:53
Subject Sample # 2	0.292	15:54
Breath Volume	1.824	15:54
Air Blank	0.000	15:55
Control Test	0.081	15:55
Air Blank	0.000	15:55
Diagnostics Check	OK	15:56

---

Data Download Date/Time: 10/24/2019 14:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WISE,IAN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: PRIKRYL

DATE/TIME OF TEST: 10/05/2019 21:03:08

---

Diagnostics Check	OK	20:51	
Air Blank	0.000	20:52	
Control Test	0.082	20:52	
Air Blank	0.000	20:53	
Subject Sample	# 1	20:56	No Sample Provided
Breath Volume	0.000	20:56	
Air Blank	0.000	20:56	
Air Blank	0.000	20:58	
Subject Sample	# 2	21:01	No Sample Provided
Breath Volume	0.000	21:01	
Air Blank	0.000	21:02	
Control Test	0.082	21:02	
Air Blank	0.000	21:03	
Diagnostics Check	OK	21:03	

---

Data Download Date/Time: 10/24/2019 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MARCELLON,WILBENS  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: LIMEHOUSE

DATE/TIME OF TEST: 10/06/2019 05:06:37

---

Diagnostics Check	OK	05:00
Air Blank	0.000	05:00
Control Test	0.082	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.101	05:02
Breath Volume	1.191	05:02
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.102	05:05
Breath Volume	2.328	05:05
Air Blank	0.000	05:05
Control Test	0.081	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

---

Data Download Date/Time: 10/24/2019 14:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: FIGUEROA,GENESIS  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: BAEZFEBO

DATE/TIME OF TEST: 10/07/2019 01:59:45

---

Diagnostics Check		OK	01:52
Air Blank		0.000	01:53
Control Test		0.082	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.237	01:54
Breath Volume		2.703	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.231	01:58
Breath Volume		2.300	01:58
Air Blank		0.000	01:58
Control Test		0.081	01:59
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

---

Data Download Date/Time: 10/24/2019 14:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KRAUSS,JOHN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: LOVETTE

DATE/TIME OF TEST: 10/07/2019 23:33:32

---

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.082	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.195	23:29
Breath Volume	2.578	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.194	23:32
Breath Volume	2.921	23:32
Air Blank	0.000	23:32
Control Test	0.081	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

---

Data Download Date/Time: 10/24/2019 14:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ABREU,JUAN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 10/09/2019 23:51:50

---

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.082	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.162	23:46
Breath Volume	1.207	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.163	23:50
Breath Volume	1.359	23:50
Air Blank	0.000	23:50
Control Test	0.081	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

---

Data Download Date/Time: 10/24/2019 14:41



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: FIGUEROA,GENESIS  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: YU

DATE/TIME OF TEST: 10/10/2019 23:04:20

---

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.082	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.000	23:00
Breath Volume	3.019	23:00
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.000	23:02
Breath Volume	3.640	23:02
Air Blank	0.000	23:03
Control Test	0.082	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

---

Data Download Date/Time: 10/24/2019 14:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MCFERREN,JASON  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: WALKES

DATE/TIME OF TEST: 10/11/2019 01:18:25

---

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.083	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.028	01:14
Breath Volume	2.539	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.026	01:16
Breath Volume	2.093	01:16
Air Blank	0.000	01:17
Control Test	0.082	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

---

Data Download Date/Time: 10/24/2019 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HARWOOD,ARMANDO  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 10/12/2019 02:29:36

---

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.082	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.130	02:25
Breath Volume	1.726	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.125	02:28
Breath Volume	1.250	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

---

Data Download Date/Time: 10/24/2019 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VELAZQUEZ,JAMES  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 07:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: FIGUEROARODRIGUEZ

DATE/TIME OF TEST: 10/12/2019 08:42:00

---

Diagnostics Check	OK	08:39	
Air Blank	0.000	08:39	
Control Test	0.081	08:39	
Air Blank	0.000	08:40	
Subject Sample	# 1	08:40	Subject Test Refused
Breath Volume	0.000	08:40	
Air Blank	0.000	08:41	
Control Test	0.081	08:41	
Air Blank	0.000	08:41	
Diagnostics Check	OK	08:42	

---

Data Download Date/Time: 10/24/2019 14:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MCLEAN, NICHOLAS  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 10:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: DOMINGODOMINGO

DATE/TIME OF TEST: 10/12/2019 12:16:39

---

Diagnostics Check	OK	12:09
Air Blank	0.000	12:09
Control Test	0.082	12:10
Air Blank	0.000	12:10
Subject Sample # 1	0.126	12:12
Breath Volume	2.203	12:12
Air Blank	0.000	12:12
Air Blank	0.000	12:14
Subject Sample # 2	0.120	12:15
Breath Volume	2.531	12:15
Air Blank	0.000	12:15
Control Test	0.081	12:16
Air Blank	0.000	12:16
Diagnostics Check	OK	12:16

---

Data Download Date/Time: 10/24/2019 14:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BRIGGS,DERIENNE  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: RIVERA DIAZ

DATE/TIME OF TEST: 10/12/2019 19:46:39

---

Diagnostics Check	OK	19:40
Air Blank	0.000	19:41
Control Test	0.082	19:41
Air Blank	0.000	19:41
Subject Sample # 1	0.000	19:42
Breath Volume	2.515	19:42
Air Blank	0.000	19:42
Air Blank	0.000	19:44
Subject Sample # 2	0.000	19:45
Breath Volume	4.667	19:45
Air Blank	0.000	19:45
Control Test	0.081	19:46
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

---

Data Download Date/Time: 10/24/2019 14:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GARVIN,MAJOR  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: BURTON

DATE/TIME OF TEST: 10/13/2019 01:53:19

---

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.082	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.195	01:49
Breath Volume	2.742	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.200	01:51
Breath Volume	2.824	01:51
Air Blank	0.000	01:52
Control Test	0.081	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

---

Data Download Date/Time: 10/24/2019 14:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SANCHEZ,JOSE  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: REISS

DATE/TIME OF TEST: 10/13/2019 03:17:49

---

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.164	03:13
Breath Volume		2.171	03:13
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.167	03:16
Breath Volume		1.753	03:16
Air Blank		0.000	03:16
Control Test		0.079	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

---

Data Download Date/Time: 10/24/2019 14:41



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BINET,MARLON  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: ONOFREHERNANDEZ

DATE/TIME OF TEST: 10/13/2019 20:29:49

---

Diagnostics Check	OK	20:22
Air Blank	0.000	20:23
Control Test	0.082	20:23
Air Blank	0.000	20:24
Subject Sample # 1	0.190	20:25
Breath Volume	2.796	20:25
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 2	0.193	20:28
Breath Volume	3.050	20:28
Air Blank	0.000	20:28
Control Test	0.081	20:29
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

---

Data Download Date/Time: 10/24/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANER, OSVALDO,  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: FACEY, ANDRE  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: BAYONNE

DATE/TIME OF TEST: 10/14/2019 04:18:49

---

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.082	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.167	04:14
Breath Volume	1.562	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.151	04:17
Breath Volume	1.109	04:17
Air Blank	0.000	04:17
Control Test	0.082	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

---

Data Download Date/Time: 10/24/2019 14:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KAVAJA,ARIJETA  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: QUINTANILLA

DATE/TIME OF TEST: 10/14/2019 18:15:33

---

Diagnostics Check	OK	18:09
Air Blank	0.000	18:09
Control Test	0.082	18:10
Air Blank	0.000	18:10
Subject Sample # 1	0.000	18:11
Breath Volume	2.847	18:11
Air Blank	0.000	18:11
Air Blank	0.000	18:13
Subject Sample # 2	0.000	18:14
Breath Volume	3.285	18:14
Air Blank	0.000	18:14
Control Test	0.081	18:14
Air Blank	0.000	18:15
Diagnostics Check	OK	18:15

---

Data Download Date/Time: 10/24/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS, TABITHA, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WISE, IAN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: CONCEPCIONHERNANDEZ

DATE/TIME OF TEST: 10/15/2019 03:41:34

---

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.082	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.172	03:36
Breath Volume	1.167	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.172	03:40
Breath Volume	1.875	03:40
Air Blank	0.000	03:40
Control Test	0.081	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

---

Data Download Date/Time: 10/24/2019 14:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HARRISON,ROBBIE  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: AGRONT

DATE/TIME OF TEST: 10/16/2019 03:09:41

---

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.082	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.054	03:05
Breath Volume	2.042	03:05
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.055	03:08
Breath Volume	2.074	03:08
Air Blank	0.000	03:08
Control Test	0.081	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

---

Data Download Date/Time: 10/24/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DOMINGUEZ, JAYMES  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 10/16/2019 23:04:32

---

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.083	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.098	22:59
Breath Volume	1.300	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.088	23:03
Breath Volume	1.570	23:03
Air Blank	0.000	23:03
Control Test	0.082	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

---

Data Download Date/Time: 10/24/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SHEIMAN,RICHARD  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: PEREZVERDUGO

DATE/TIME OF TEST: 10/17/2019 00:54:02

---

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.081	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.199	00:49
Breath Volume	1.531	00:49
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.210	00:52
Breath Volume	1.929	00:52
Air Blank	0.000	00:53
Control Test	0.081	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:54

---

Data Download Date/Time: 10/24/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PARRISH,KAYLA  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 10/17/2019 02:38:57

---

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.081	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.220	02:34
Breath Volume	4.015	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample # 2	0.213	02:37
Breath Volume	3.347	02:37
Air Blank	0.000	02:38
Control Test	0.081	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

---

Data Download Date/Time: 10/24/2019 14:42



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PEERS,MICHAEL  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: BLUNT

DATE/TIME OF TEST: 10/17/2019 03:49:54

---

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.149	03:43
Breath Volume	1.273	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.157	03:48
Breath Volume	2.722	03:48
Air Blank	0.000	03:48
Control Test	0.080	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

---

Data Download Date/Time: 10/24/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLS,ARMAND,G  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GUEVARA,JOSHUA  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 09:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:39

SUBJECT'S LAST NAME: KORNILOV

DATE/TIME OF TEST: 10/17/2019 11:12:08

---

Diagnostics Check	OK	11:02
Air Blank	0.000	11:02
Control Test	0.081	11:02
Air Blank	0.000	11:03
Subject Sample # 1	0.160	11:06
Breath Volume	1.218	11:06
Air Blank	0.000	11:07
Air Blank	0.000	11:08
Subject Sample # 2	0.164	11:10
Breath Volume	1.957	11:10
Air Blank	0.000	11:11
Control Test	0.080	11:11
Air Blank	0.000	11:12
Diagnostics Check	OK	11:12

---

Data Download Date/Time: 10/24/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: FACEY,ANDRE  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: DIXSON

DATE/TIME OF TEST: 10/18/2019 00:00:17

---

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:46	
Control Test		0.082	23:46	
Air Blank		0.000	23:47	
Subject Sample	# 1	0.131	23:50	Volume Not Met
Breath Volume		0.117	23:50	
Air Blank		0.000	23:50	
Air Blank		0.000	23:52	
Subject Sample	# 2	0.130	23:54	
Breath Volume		1.453	23:54	
Air Blank		0.000	23:55	
Air Blank		0.000	23:57	
Subject Sample	# 3	0.139	23:58	
Breath Volume		2.738	23:58	
Air Blank		0.000	23:59	
Control Test		0.081	23:59	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:00	

---

Data Download Date/Time: 10/24/2019 14:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SPEER,ANDREW  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: ABDUALINOV

DATE/TIME OF TEST: 10/18/2019 00:46:17

---

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.133	00:41
Breath Volume	2.636	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.127	00:44
Breath Volume	3.375	00:44
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

---

Data Download Date/Time: 10/24/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MARTINEZ,ERNESTO  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 10/18/2019 22:04:44

---

Diagnostics Check		OK	21:58
Air Blank		0.000	21:58
Control Test		0.081	21:58
Air Blank		0.000	21:59
Subject Sample	# 1	0.257	22:00
Breath Volume		1.445	22:00
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.243	22:03
Breath Volume		1.808	22:03
Air Blank		0.000	22:03
Control Test		0.080	22:04
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

---

Data Download Date/Time: 10/24/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HOLLEY,DEANDRE  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ASCANIO GARCIA

DATE/TIME OF TEST: 10/19/2019 01:46:40

---

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.082	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.189	01:41
Breath Volume	3.289	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.187	01:45
Breath Volume	3.257	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

---

Data Download Date/Time: 10/24/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HYATT,MARCUS  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: GARCIA-ZAPATA

DATE/TIME OF TEST: 10/19/2019 02:52:55

---

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.082	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.154	02:47
Breath Volume	1.164	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.169	02:51
Breath Volume	1.832	02:51
Air Blank	0.000	02:52
Control Test	0.082	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

---

Data Download Date/Time: 10/24/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ABREU,JUAN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 10/19/2019 04:42:20

---

Diagnostics Check		OK	04:32
Air Blank		0.000	04:33
Control Test		0.081	04:33
Air Blank		0.000	04:33
Subject Sample	# 1	0.255	04:37
Breath Volume		1.269	04:37
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.258	04:40
Breath Volume		1.792	04:40
Air Blank		0.000	04:41
Control Test		0.079	04:41
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

---

Data Download Date/Time: 10/24/2019 14:42



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HARRISON,ROBBIE  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 10/19/2019 22:43:31

---

Diagnostics Check	OK	22:36
Air Blank	0.000	22:37
Control Test	0.081	22:37
Air Blank	0.000	22:38
Subject Sample # 1	0.092	22:38
Breath Volume	2.890	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.088	22:42
Breath Volume	2.964	22:42
Air Blank	0.000	22:42
Control Test	0.081	22:42
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

---

Data Download Date/Time: 10/24/2019 14:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SPEER,ANDREW  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CURREY

DATE/TIME OF TEST: 10/20/2019 03:23:05

---

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:19	
Control Test	0.082	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:22	Sequence Aborted
Breath Volume	0.000	03:22	
Air Blank	0.000	03:23	

---

Data Download Date/Time: 10/24/2019 14:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: NDOUMBE,FREDERIC  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: GOMEZHERNANDEZ

DATE/TIME OF TEST: 10/20/2019 04:06:49

---

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.082	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.134	04:02
Breath Volume	3.679	04:02
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.130	04:05
Breath Volume	3.695	04:05
Air Blank	0.000	04:05
Control Test	0.081	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

---

Data Download Date/Time: 10/24/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS, TABITHA, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WISE, IAN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: OQUENDO

DATE/TIME OF TEST: 10/20/2019 23:18:03

---

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.082	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.096	23:13
Breath Volume	1.628	23:13
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.104	23:16
Breath Volume	5.136	23:16
Air Blank	0.000	23:17
Control Test	0.082	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:18

---

Data Download Date/Time: 10/24/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HOLLEY,DEANDRE  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MATOS RIOS

DATE/TIME OF TEST: 10/21/2019 03:35:53

---

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.082	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.193	03:31
Breath Volume	2.824	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:33
Subject Sample # 2	0.173	03:34
Breath Volume	1.351	03:34
Air Blank	0.000	03:34
Control Test	0.081	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

---

Data Download Date/Time: 10/24/2019 14:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS, TABITHA, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PATINO, NICK  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: VISSER

DATE/TIME OF TEST: 10/21/2019 05:40:26

---

Diagnostics Check	OK	05:34
Air Blank	0.000	05:34
Control Test	0.082	05:34
Air Blank	0.000	05:35
Subject Sample # 1	0.192	05:35
Breath Volume	1.765	05:35
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.186	05:38
Breath Volume	1.875	05:38
Air Blank	0.000	05:39
Control Test	0.081	05:39
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

---

Data Download Date/Time: 10/24/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HERNANDEZ, JEFFRIE  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: RAMOSROMAN

DATE/TIME OF TEST: 10/21/2019 16:39:33

---

Diagnostics Check	OK	16:33
Air Blank	0.000	16:33
Control Test	0.082	16:33
Air Blank	0.000	16:34
Subject Sample # 1	0.000	16:35
Breath Volume	3.617	16:35
Air Blank	0.000	16:35
Air Blank	0.000	16:37
Subject Sample # 2	0.000	16:38
Breath Volume	1.167	16:38
Air Blank	0.000	16:38
Control Test	0.082	16:38
Air Blank	0.000	16:39
Diagnostics Check	OK	16:39

---

Data Download Date/Time: 10/24/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PARRISH,KAYLA  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 10/21/2019 22:43:31

---

Diagnostics Check		OK	22:36
Air Blank		0.000	22:36
Control Test		0.083	22:36
Air Blank		0.000	22:37
Subject Sample	# 1	0.204	22:38
Breath Volume		1.730	22:38
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.195	22:42
Breath Volume		1.910	22:42
Air Blank		0.000	22:42
Control Test		0.081	22:42
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

---

Data Download Date/Time: 10/24/2019 14:42



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS, TABITHA, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DOMINGUEZ, JAYMES  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: CHAVEZROJO

DATE/TIME OF TEST: 10/22/2019 20:57:08

---

Diagnostics Check	OK	20:50
Air Blank	0.000	20:51
Control Test	0.082	20:51
Air Blank	0.000	20:52
Subject Sample # 1	0.048	20:52
Breath Volume	3.871	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.047	20:55
Breath Volume	4.121	20:55
Air Blank	0.000	20:56
Control Test	0.081	20:56
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

---

Data Download Date/Time: 10/24/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MCFERREN,JASON  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: BERRETTA

DATE/TIME OF TEST: 10/22/2019 23:42:40

---

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.082	23:32	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:35	No Sample Provided
Breath Volume	0.000	23:35	
Air Blank	0.000	23:36	
Air Blank	0.000	23:38	
Subject Sample	# 2	23:41	No Sample Provided
Breath Volume	0.000	23:41	
Air Blank	0.000	23:41	
Control Test	0.083	23:42	
Air Blank	0.000	23:42	
Diagnostics Check	OK	23:42	

---

Data Download Date/Time: 10/24/2019 14:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ROCCHILD,MIGUEL  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BRANNIN

DATE/TIME OF TEST: 10/23/2019 01:21:21

---

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.082	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.049	01:16
Breath Volume	1.960	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.049	01:19
Breath Volume	2.378	01:19
Air Blank	0.000	01:20
Control Test	0.082	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

---

Data Download Date/Time: 10/24/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RAMIREZ,MATTHEW  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: VENTURAJIMENEZ

DATE/TIME OF TEST: 10/23/2019 02:14:22

---

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.082	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:07	No Sample Provided
Breath Volume	0.000	02:07	
Air Blank	0.000	02:07	
Air Blank	0.000	02:09	
Subject Sample	# 2	02:12	No Sample Provided
Breath Volume	0.000	02:12	
Air Blank	0.000	02:13	
Control Test	0.081	02:13	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:14	

---

Data Download Date/Time: 10/24/2019 14:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS, TABITHA, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MCPHEE, WILLIS  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 10/23/2019 02:53:39

---

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.081	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.000	02:49
Breath Volume	1.425	02:49
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.000	02:52
Breath Volume	1.781	02:52
Air Blank	0.000	02:52
Control Test	0.082	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

---

Data Download Date/Time: 10/24/2019 14:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MOLNE,LUIS  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: MOREJON VILLAFANA

DATE/TIME OF TEST: 10/24/2019 19:10:56

---

Diagnostics Check	OK	19:04
Air Blank	0.000	19:05
Control Test	0.081	19:05
Air Blank	0.000	19:06
Subject Sample # 1	0.000	19:06
Breath Volume	3.605	19:06
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample # 2	0.000	19:09
Breath Volume	1.667	19:09
Air Blank	0.000	19:10
Control Test	0.080	19:10
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

---

Data Download Date/Time: 10/31/2019 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WISE,IAN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 10/25/2019 02:33:52

---

Diagnostics Check		OK	02:19	
Air Blank		0.000	02:19	
Control Test		0.081	02:19	
Air Blank		0.000	02:20	
Subject Sample	# 1	0.210	02:21	
Breath Volume		1.402	02:21	
Air Blank		0.000	02:21	
Air Blank		0.000	02:23	
Subject Sample	# 2	0.188	02:26	Volume Not Met
Breath Volume		0.140	02:26	
Air Blank		0.000	02:27	
Air Blank		0.000	02:29	
Subject Sample	# 3	0.159	02:32	Volume Not Met
Breath Volume		0.121	02:32	
Air Blank		0.000	02:32	
Control Test		0.080	02:33	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

---

Data Download Date/Time: 10/31/2019 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAKMAN,JASON  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: BOONE

DATE/TIME OF TEST: 10/26/2019 01:29:32

---

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.082	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.211	01:24
Breath Volume		1.792	01:24
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.212	01:28
Breath Volume		2.199	01:28
Air Blank		0.000	01:28
Control Test		0.081	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

---

Data Download Date/Time: 10/31/2019 15:21



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HOPKINS,JEFFREY  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: HERNANDEZ FLORES

DATE/TIME OF TEST: 10/26/2019 03:08:48

---

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.082	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.169	03:04
Breath Volume	4.171	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.177	03:07
Breath Volume	5.210	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

---

Data Download Date/Time: 10/31/2019 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BRIGGS, DERIENNE  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 08:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 10/26/2019 09:55:44

---

Diagnostics Check	OK	09:49
Air Blank	0.000	09:49
Control Test	0.081	09:50
Air Blank	0.000	09:50
Subject Sample # 1	0.170	09:51
Breath Volume	1.902	09:51
Air Blank	0.000	09:51
Air Blank	0.000	09:53
Subject Sample # 2	0.161	09:54
Breath Volume	1.144	09:54
Air Blank	0.000	09:54
Control Test	0.082	09:55
Air Blank	0.000	09:55
Diagnostics Check	OK	09:55

---

Data Download Date/Time: 10/31/2019 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CORREA,ASHLEY  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: LE

DATE/TIME OF TEST: 10/26/2019 21:06:11

---

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:50	
Control Test	0.082	20:50	
Air Blank	0.000	20:50	
Subject Sample	# 1 0.243	20:53	
Breath Volume	1.132	20:53	
Air Blank	0.000	20:53	
Air Blank	0.000	20:55	
Subject Sample	# 2 0.251	20:58	Volume Not Met
Breath Volume	1.042	20:58	
Air Blank	0.000	20:59	
Air Blank	0.000	21:01	
Subject Sample	# 3 0.248	21:04	
Breath Volume	1.230	21:04	
Air Blank	0.000	21:05	
Control Test	0.079	21:05	
Air Blank	0.000	21:06	
Diagnostics Check	OK	21:06	

---

Data Download Date/Time: 10/31/2019 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DAVIS, DWAYNE  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: MATTOX

DATE/TIME OF TEST: 10/27/2019 01:05:39

---

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.081	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.234	01:01
Breath Volume	3.511	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.231	01:04
Breath Volume	3.199	01:04
Air Blank	0.000	01:04
Control Test	0.081	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

---

Data Download Date/Time: 10/31/2019 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: LAWRENCE,NICKOLAS  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: KENAN

DATE/TIME OF TEST: 10/27/2019 03:31:41

---

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.082	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.106	03:27
Breath Volume	1.281	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.105	03:30
Breath Volume	2.187	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

---

Data Download Date/Time: 10/31/2019 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PRICE, FORREST  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 10/27/2019 17:11:08

---

Diagnostics Check	OK	17:05
Air Blank	0.000	17:05
Control Test	0.082	17:05
Air Blank	0.000	17:06
Subject Sample # 1	0.287	17:06
Breath Volume	1.925	17:06
Air Blank	0.000	17:07
Air Blank	0.000	17:09
Subject Sample # 2	0.289	17:09
Breath Volume	2.675	17:09
Air Blank	0.000	17:10
Control Test	0.080	17:10
Air Blank	0.000	17:11
Diagnostics Check	OK	17:11

---

Data Download Date/Time: 10/31/2019 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HEBEL,GRIFFIN  
ARRESTING OFFICER AGENCY: Windermere PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: PAULSON

DATE/TIME OF TEST: 10/27/2019 23:56:27

---

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:45	
Control Test	0.082	23:45	
Air Blank	0.000	23:46	
Subject Sample	# 1 0.265	23:49	Volume Not Met
Breath Volume	1.097	23:49	
Air Blank	0.000	23:49	
Air Blank	0.000	23:51	
Subject Sample	# 2 0.274	23:54	Volume Not Met
Breath Volume	0.031	23:54	
Air Blank	0.000	23:55	
Control Test	0.080	23:55	
Air Blank	0.000	23:56	
Diagnostics Check	OK	23:56	

---

Data Download Date/Time: 10/31/2019 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HEBEL,GRIFFIN  
ARRESTING OFFICER AGENCY: Windermere PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: PAULSON

DATE/TIME OF TEST: 10/28/2019 00:06:28

---

Diagnostics Check		OK	23:58
Air Blank		0.000	23:59
Control Test		0.080	23:59
Air Blank		0.000	23:59
Subject Sample	# 1	0.294	00:01
Breath Volume		2.253	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.283	00:04
Breath Volume		1.886	00:04
Air Blank		0.000	00:05
Control Test		0.079	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

---

Data Download Date/Time: 10/31/2019 15:21



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELVILLE,KELLY,M  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ANDREWS,GARY  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:53

SUBJECT'S LAST NAME: RIVERAROBLES

DATE/TIME OF TEST: 10/29/2019 10:24:15

---

Diagnostics Check	OK	10:17
Air Blank	0.000	10:17
Control Test	0.081	10:18
Air Blank	0.000	10:18
Subject Sample # 1	0.172	10:19
Breath Volume	1.285	10:19
Air Blank	0.000	10:20
Air Blank	0.000	10:22
Subject Sample # 2	0.162	10:22
Breath Volume	1.457	10:22
Air Blank	0.000	10:23
Control Test	0.080	10:23
Air Blank	0.000	10:24
Diagnostics Check	OK	10:24

---

Data Download Date/Time: 10/31/2019 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HOLLEY,DEANDRE  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HEMMINGS

DATE/TIME OF TEST: 10/30/2019 00:36:36

---

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.081	00:30
Air Blank	0.000	00:30
Subject Sample # 1	0.109	00:31
Breath Volume	1.421	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.113	00:35
Breath Volume	1.628	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

---

Data Download Date/Time: 10/31/2019 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WISE,IAN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: ROTELLA

DATE/TIME OF TEST: 10/30/2019 01:13:14

---

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.082	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.000	01:08
Breath Volume	3.078	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.000	01:11
Breath Volume	4.011	01:11
Air Blank	0.000	01:12
Control Test	0.081	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

---

Data Download Date/Time: 10/31/2019 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GIBSON,JAMES  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MOSHER

DATE/TIME OF TEST: 10/30/2019 03:05:49

---

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.082	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.277	03:01
Breath Volume	1.390	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:03
Subject Sample # 2	0.269	03:04
Breath Volume	1.792	03:04
Air Blank	0.000	03:04
Control Test	0.080	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

---

Data Download Date/Time: 10/31/2019 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KERR,BRANDON  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 10/30/2019 23:50:47

---

Diagnostics Check	OK	23:44
Air Blank	0.000	23:45
Control Test	0.082	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.167	23:46
Breath Volume	1.781	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:48
Subject Sample # 2	0.169	23:49
Breath Volume	1.335	23:49
Air Blank	0.000	23:49
Control Test	0.082	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

---

Data Download Date/Time: 10/31/2019 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIAMSON,DAWN,P  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MARTELL,SIONE  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 10/31/2019 04:39:07

---

Diagnostics Check		OK	04:32
Air Blank		0.000	04:33
Control Test		0.082	04:33
Air Blank		0.000	04:34
Subject Sample	# 1	0.210	04:34
Breath Volume		2.023	04:34
Air Blank		0.000	04:35
Air Blank		0.000	04:37
Subject Sample	# 2	0.198	04:37
Breath Volume		1.957	04:37
Air Blank		0.000	04:38
Control Test		0.082	04:38
Air Blank		0.000	04:39
Diagnostics Check		OK	04:39

---

Data Download Date/Time: 10/31/2019 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELVILLE,KELLY,M  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HOLIDAY,LUCAS  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:56

SUBJECT'S LAST NAME: BEHAR

DATE/TIME OF TEST: 10/31/2019 10:26:54

---

Diagnostics Check	OK	10:20
Air Blank	0.000	10:21
Control Test	0.082	10:21
Air Blank	0.000	10:21
Subject Sample # 1	0.000	10:22
Breath Volume	1.371	10:22
Air Blank	0.000	10:23
Air Blank	0.000	10:25
Subject Sample # 2	0.000	10:25
Breath Volume	2.179	10:25
Air Blank	0.000	10:25
Control Test	0.081	10:26
Air Blank	0.000	10:26
Diagnostics Check	OK	10:26

---

Data Download Date/Time: 10/31/2019 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WISE,IAN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: PERSAD

DATE/TIME OF TEST: 10/11/2019 01:24:04

---

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.080	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.000	01:19
Breath Volume	2.621	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.000	01:22
Breath Volume	3.003	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:24

---

Data Download Date/Time: 10/24/2019 14:47



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SPEER,ANDREW  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MARION

DATE/TIME OF TEST: 10/12/2019 03:50:32

---

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.080	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.066	03:46
Breath Volume	1.312	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.065	03:49
Breath Volume	2.234	03:49
Air Blank	0.000	03:49
Control Test	0.079	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

---

Data Download Date/Time: 10/24/2019 14:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAINES,BRANT  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: ODWIN

DATE/TIME OF TEST: 10/13/2019 03:20:19

---

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.109	03:15
Breath Volume	3.070	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.104	03:18
Breath Volume	2.597	03:18
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

---

Data Download Date/Time: 10/24/2019 14:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RIVERA,VICTOR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 10/17/2019 02:41:00

---

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.147	02:36
Breath Volume	1.109	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.158	02:39
Breath Volume	1.613	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:41

---

Data Download Date/Time: 10/24/2019 14:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BODE,JOSHUA  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 10/19/2019 02:15:40

---

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.080	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.170	02:11
Breath Volume	2.132	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.161	02:14
Breath Volume	2.226	02:14
Air Blank	0.000	02:14
Control Test	0.078	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

---

Data Download Date/Time: 10/24/2019 14:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RIDEAUX,ADAM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: MCFALL

DATE/TIME OF TEST: 10/24/2019 01:54:09

---

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.079	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.208	01:49
Breath Volume	3.132	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:51
Subject Sample # 2	0.208	01:52
Breath Volume	3.117	01:52
Air Blank	0.000	01:53
Control Test	0.078	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

---

Data Download Date/Time: 10/24/2019 14:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COPELAND,MICHAEL,C  
BREATH TEST OPERATOR AGENCY: Quincy PD  
ARRESTING OFFICER: MATLOCK,LOGAN  
ARRESTING OFFICER AGENCY: Quincy PD  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MEJIA DELGADO

DATE/TIME OF TEST: 10/05/2019 01:26:31

---

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:21	
Control Test	0.077	01:22	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:24	Subject Test Refused
Breath Volume	0.453	01:24	
Air Blank	0.000	01:25	
Control Test	0.079	01:25	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

---

Data Download Date/Time: 10/09/2019 11:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGGS,JOSHUA,A  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MATLOCK,LOGAN  
ARRESTING OFFICER AGENCY: Quincy PD  
LAST AGENCY INSPECTION: 09/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 10/06/2019 01:34:22

---

Diagnostics Check		OK	01:27
Air Blank		0.000	01:28
Control Test		0.077	01:28
Air Blank		0.000	01:29
Subject Sample	# 1	0.137	01:29
Breath Volume		2.992	01:29
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.131	01:32
Breath Volume		2.953	01:32
Air Blank		0.000	01:33
Control Test		0.078	01:33
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

---

Data Download Date/Time: 10/09/2019 11:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,ASHLEY,J  
BREATH TEST OPERATOR AGENCY: Quincy PD  
ARRESTING OFFICER: CLARK,ASHLEY  
ARRESTING OFFICER AGENCY: Quincy PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: CYL 011  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/12/2019 01:57:13

---

Diagnostics Check		OK	01:50
Air Blank		0.000	01:51
Control Test		0.078	01:51
Air Blank		0.000	01:52
Subject Sample	# 1	0.144	01:52
Breath Volume		2.097	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.138	01:55
Breath Volume		1.882	01:55
Air Blank		0.000	01:56
Control Test		0.077	01:56
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

---

Data Download Date/Time: 11/06/2019 14:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,ASHLEY,J  
BREATH TEST OPERATOR AGENCY: Quincy PD  
ARRESTING OFFICER: CLARK,ASHLEY  
ARRESTING OFFICER AGENCY: Quincy PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 10/13/2019 01:37:45

---

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.000	01:33
Breath Volume	1.863	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.000	01:36
Breath Volume	1.335	01:36
Air Blank	0.000	01:36
Control Test	0.077	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

---

Data Download Date/Time: 11/06/2019 14:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: THOMPSON,RONDRICKA,N

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: THOMPSON,RONDRICKA

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29017080A3

DRY GAS EXPIRY DATE: 12/05/2019

VIOLATION CODE: DUI

ARREST TIME: 19:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: ROYCE

DATE/TIME OF TEST: 10/23/2019 20:21:50

---

Diagnostics Check	OK	20:21	
Air Blank		20:21	RFI Detect
Air Blank	0.000	20:21	

---

Data Download Date/Time: 11/06/2019 14:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,RONDRICKA,N  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: THOMPSON,RONDRICKA  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: ROYCE

DATE/TIME OF TEST: 10/23/2019 20:31:00

---

Diagnostics Check	OK	20:24
Air Blank	0.000	20:24
Control Test	0.079	20:25
Air Blank	0.000	20:25
Subject Sample # 1	0.124	20:26
Breath Volume	1.414	20:26
Air Blank	0.000	20:26
Air Blank	0.000	20:28
Subject Sample # 2	0.138	20:29
Breath Volume	3.765	20:29
Air Blank	0.000	20:30
Control Test	0.079	20:30
Air Blank	0.000	20:30
Diagnostics Check	OK	20:31

---

Data Download Date/Time: 11/06/2019 14:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGGS,JOSHUA,A  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MATLOCK,LOGAN  
ARRESTING OFFICER AGENCY: Quincy PD  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: BOLAN

DATE/TIME OF TEST: 10/24/2019 00:31:48

---

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.078	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.164	00:27
Breath Volume	3.269	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.164	00:30
Breath Volume	3.542	00:30
Air Blank	0.000	00:30
Control Test	0.079	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

---

Data Download Date/Time: 11/06/2019 14:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGGS,JOSHUA,A  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: BIGGS,JOSHUA  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 10/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 10/28/2019 00:42:38

---

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.077	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.179	00:38
Breath Volume	2.781	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.196	00:41
Breath Volume	3.500	00:41
Air Blank	0.000	00:41
Control Test	0.079	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

---

Data Download Date/Time: 11/06/2019 14:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SAYERS,MICHELLE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 09/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: HILLMAN

DATE/TIME OF TEST: 10/01/2019 23:47:13

---

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.247	23:42
Breath Volume	3.312	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.240	23:45
Breath Volume	2.976	23:45
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

---

Data Download Date/Time: 10/03/2019 23:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 09/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 10/02/2019 00:18:51

---

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.078	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.000	00:14
Breath Volume	1.652	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.000	00:17
Breath Volume	1.632	00:17
Air Blank	0.000	00:17
Control Test	0.078	00:18
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

---

Data Download Date/Time: 10/03/2019 23:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SAYERS,MICHELLE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 10/05/2019 02:28:32

---

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.078	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.085	02:24
Breath Volume	1.738	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.082	02:27
Breath Volume	1.335	02:27
Air Blank	0.000	02:27
Control Test	0.077	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

---

Data Download Date/Time: 11/26/2019 14:36



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AKINS,JUSTIN,M  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: AKINS,JUSTIN  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BORGELLA

DATE/TIME OF TEST: 10/06/2019 03:17:04

---

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.011	03:12
Breath Volume	1.394	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.009	03:15
Breath Volume	1.324	03:15
Air Blank	0.000	03:16
Control Test	0.077	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:17

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: HARRIS,D  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ECHEVERRY-ARROYAVE

DATE/TIME OF TEST: 10/06/2019 03:52:24

---

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.078	03:46
Air Blank		0.000	03:47
Subject Sample	# 1	0.128	03:47
Breath Volume		3.144	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.128	03:50
Breath Volume		2.910	03:50
Air Blank		0.000	03:51
Control Test		0.077	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PADILLA,H  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: ESTIVERNE

DATE/TIME OF TEST: 10/06/2019 21:02:08

---

Diagnostics Check	OK	20:51
Air Blank	0.000	20:52
Control Test	0.079	20:52
Air Blank	0.000	20:52
Subject Sample # 1	0.152	20:55
Breath Volume	1.144	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:57
Subject Sample # 2	0.160	21:00
Breath Volume	1.871	21:00
Air Blank	0.000	21:01
Control Test	0.078	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCONNELL,CHARLES,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: CRAWFORD,DAVID  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:59

SUBJECT'S LAST NAME: BARONE

DATE/TIME OF TEST: 10/08/2019 18:25:41

---

Diagnostics Check	OK	18:19
Air Blank	0.000	18:19
Control Test	0.077	18:20
Air Blank	0.000	18:20
Subject Sample # 1	0.000	18:21
Breath Volume	3.054	18:21
Air Blank	0.000	18:21
Air Blank	0.000	18:23
Subject Sample # 2	0.000	18:24
Breath Volume	1.941	18:24
Air Blank	0.000	18:24
Control Test	0.076	18:25
Air Blank	0.000	18:25
Diagnostics Check	OK	18:25

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AKINS,JUSTIN,M  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: CAMPOS,BARBARA  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: MENSHAWY-BAKR

DATE/TIME OF TEST: 10/11/2019 00:30:27

---

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.078	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.000	00:26
Breath Volume	3.398	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.000	00:29
Breath Volume	1.949	00:29
Air Blank	0.000	00:29
Control Test	0.079	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYONS,DANIEL,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: ACEVEDO,ELIZABETH  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BAUTISTA-JIMENEZ

DATE/TIME OF TEST: 10/14/2019 01:18:28

---

Diagnostics Check		OK	01:05	
Air Blank		0.000	01:06	
Control Test		0.079	01:06	
Air Blank		0.000	01:07	
Subject Sample	# 1	0.154	01:10	Volume Not Met
Breath Volume		0.718	01:10	
Air Blank		0.000	01:11	
Air Blank		0.000	01:12	
Subject Sample	# 2	0.173	01:13	No .020 agreement
Breath Volume		2.039	01:13	
Air Blank		0.000	01:14	
Air Blank		0.000	01:15	
Subject Sample	# 3	0.152	01:16	No .020 agreement
Breath Volume		1.570	01:16	
Air Blank		0.000	01:17	
Control Test		0.077	01:17	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYONS,DANIEL,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: AVEVEDO,ELIZABETH  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BAUTISTA-JIMENEZ

DATE/TIME OF TEST: 10/14/2019 01:33:55

---

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.079	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.166	01:29
Breath Volume	1.890	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:31
Subject Sample # 2	0.150	01:32
Breath Volume	1.777	01:32
Air Blank	0.000	01:33
Control Test	0.076	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAEZ,BAYRON,L  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: GUZMAN,V  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 10/14/2019 13:41:42

---

Diagnostics Check	OK	13:35
Air Blank	0.000	13:35
Control Test	0.079	13:36
Air Blank	0.000	13:36
Subject Sample # 1	0.226	13:37
Breath Volume	2.058	13:37
Air Blank	0.000	13:37
Air Blank	0.000	13:39
Subject Sample # 2	0.214	13:40
Breath Volume	1.578	13:40
Air Blank	0.000	13:40
Control Test	0.078	13:41
Air Blank	0.000	13:41
Diagnostics Check	OK	13:41

---

Data Download Date/Time: 11/26/2019 14:36



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOUVRES,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: DOUVRES,MICHAEL  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: HICKOK

DATE/TIME OF TEST: 10/14/2019 22:47:28

---

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.080	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.077	22:42
Breath Volume	2.683	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.083	22:45
Breath Volume	2.914	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

---

Data Download Date/Time: 11/26/2019 14:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SAYERS,MICHELLE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WISSMANN

DATE/TIME OF TEST: 10/15/2019 04:27:15

---

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.076	04:25	
Air Blank	0.000	04:25	
Subject Sample	# 1	04:25	Subject Test Refused
Breath Volume	0.000	04:25	
Air Blank	0.000	04:26	
Control Test	0.078	04:26	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEKLE,JAMES,D  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: MARTINEZ,CHARLES  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:11

SUBJECT'S LAST NAME: PALMIERI

DATE/TIME OF TEST: 10/18/2019 13:38:35

---

Diagnostics Check	OK	13:32
Air Blank	0.000	13:32
Control Test	0.079	13:33
Air Blank	0.000	13:33
Subject Sample # 1	0.064	13:34
Breath Volume	2.207	13:34
Air Blank	0.000	13:34
Air Blank	0.000	13:36
Subject Sample # 2	0.060	13:37
Breath Volume	2.980	13:37
Air Blank	0.000	13:37
Control Test	0.079	13:37
Air Blank	0.000	13:38
Diagnostics Check	OK	13:38

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAEZ,BAYRON,L  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: FILALI,N  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 09:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: LAMPLEY

DATE/TIME OF TEST: 10/19/2019 11:12:46

---

Diagnostics Check	OK	11:06
Air Blank	0.000	11:06
Control Test	0.080	11:07
Air Blank	0.000	11:07
Subject Sample # 1	0.000	11:08
Breath Volume	1.703	11:08
Air Blank	0.000	11:08
Air Blank	0.000	11:10
Subject Sample # 2	0.000	11:11
Breath Volume	1.703	11:11
Air Blank	0.000	11:11
Control Test	0.078	11:12
Air Blank	0.000	11:12
Diagnostics Check	OK	11:12

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: ALFARAJ

DATE/TIME OF TEST: 10/20/2019 07:22:40

---

Diagnostics Check	OK	07:16
Air Blank	0.000	07:16
Control Test	0.079	07:16
Air Blank	0.000	07:17
Subject Sample # 1	0.115	07:18
Breath Volume	1.949	07:18
Air Blank	0.000	07:18
Air Blank	0.000	07:20
Subject Sample # 2	0.113	07:21
Breath Volume	1.843	07:21
Air Blank	0.000	07:21
Control Test	0.078	07:22
Air Blank	0.000	07:22
Diagnostics Check	OK	07:22

---

Data Download Date/Time: 11/26/2019 14:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOUVRES,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: DOUVRES,MICHAEL  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: GERENA MARTINEZ

DATE/TIME OF TEST: 10/20/2019 20:45:19

---

Diagnostics Check	OK	20:38
Air Blank	0.000	20:38
Control Test	0.079	20:38
Air Blank	0.000	20:39
Subject Sample # 1	0.164	20:40
Breath Volume	1.875	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.158	20:43
Breath Volume	1.453	20:43
Air Blank	0.000	20:44
Control Test	0.078	20:44
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

---

Data Download Date/Time: 11/26/2019 14:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: CAMACHO

DATE/TIME OF TEST: 10/20/2019 22:14:30

---

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.078	22:08
Air Blank	0.000	22:09
Subject Sample # 1	0.000	22:09
Breath Volume	1.371	22:09
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample # 2	0.000	22:13
Breath Volume	1.472	22:13
Air Blank	0.000	22:13
Control Test	0.077	22:13
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

---

Data Download Date/Time: 11/26/2019 14:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,DAVID,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: CRAWFORD,DAVID  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: MELENDEZ HUERTAS

DATE/TIME OF TEST: 10/22/2019 03:23:36

---

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.024	03:19
Breath Volume	3.285	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.023	03:22
Breath Volume	2.941	03:22
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

---

Data Download Date/Time: 11/26/2019 14:36



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,ALBERT,S  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: RODRIGUEZ,ALBERT  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: GONZALEZ ARVELO

DATE/TIME OF TEST: 10/23/2019 01:19:55

---

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.079	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.000	01:15
Breath Volume	1.542	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.000	01:18
Breath Volume	2.335	01:18
Air Blank	0.000	01:19
Control Test	0.079	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

---

Data Download Date/Time: 11/26/2019 14:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: OMALLEY

DATE/TIME OF TEST: 10/23/2019 20:47:49

---

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.079	20:42
Air Blank	0.000	20:42
Subject Sample # 1	0.238	20:43
Breath Volume	2.058	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:45
Subject Sample # 2	0.236	20:46
Breath Volume	1.781	20:46
Air Blank	0.000	20:46
Control Test	0.079	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUNN,COLLIN,J  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: GYSLER,ALEX  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 10/24/2019 02:39:55

---

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.076	02:32	
Air Blank	0.000	02:33	
Subject Sample # 1	0.181	02:34	
Breath Volume	2.824	02:34	
Air Blank	0.000	02:35	
Air Blank	0.000	02:37	
Subject Sample # 2	0.196	02:38	
Breath Volume	1.683	02:38	
Air Blank	0.000	02:39	
Control Test	0.074	02:39	Control Outside Tolerance
Air Blank	0.000	02:39	

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,MICHAEL,K  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: FILALI FIKRI,N  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 10/24/2019 17:26:28

---

Diagnostics Check	OK	17:14	
Air Blank	0.000	17:14	
Control Test	0.076	17:14	
Air Blank	0.000	17:15	
Subject Sample	# 1 0.330	17:16	
Breath Volume	1.359	17:16	
Air Blank	0.000	17:17	
Air Blank	0.000	17:19	
Subject Sample	# 2 0.353	17:20	
Breath Volume	2.183	17:20	
Air Blank	0.000	17:21	
Air Blank	0.000	17:23	
Subject Sample	# 3	17:25	Subject Test Refused
Breath Volume	0.000	17:25	
Air Blank	0.000	17:25	
Control Test	0.072	17:25	Control Outside Tolerance
Air Blank	0.000	17:26	

---

Data Download Date/Time: 11/26/2019 14:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: CLEEK

DATE/TIME OF TEST: 10/29/2019 01:18:25

---

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.079	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.000	01:14
Breath Volume	2.527	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.000	01:17
Breath Volume	2.843	01:17
Air Blank	0.000	01:17
Control Test	0.076	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

---

Data Download Date/Time: 11/26/2019 14:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AKINS,JUSTIN,M  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: AKINS,JUSTIN  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: MALDONADO-RVIERA

DATE/TIME OF TEST: 10/30/2019 01:48:21

---

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.079	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.120	01:41
Breath Volume	1.191	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:43
Subject Sample # 2	0.120	01:46
Breath Volume	1.640	01:46
Air Blank	0.000	01:47
Control Test	0.078	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

---

Data Download Date/Time: 11/26/2019 14:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PYE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: REYES-MELENDZ

DATE/TIME OF TEST: 10/30/2019 03:49:00

---

Diagnostics Check		OK	03:35	
Air Blank		0.000	03:36	
Control Test		0.077	03:36	
Air Blank		0.000	03:37	
Subject Sample	# 1	0.102	03:40	Volume Not Met
Breath Volume		0.398	03:40	
Air Blank		0.000	03:40	
Air Blank		0.000	03:42	
Subject Sample	# 2	0.139	03:43	
Breath Volume		1.898	03:43	
Air Blank		0.000	03:43	
Air Blank		0.000	03:45	
Subject Sample	# 3	0.120	03:47	
Breath Volume		1.210	03:47	
Air Blank		0.000	03:48	
Control Test		0.075	03:48	
Air Blank		0.000	03:48	
Diagnostics Check		OK	03:49	

---

Data Download Date/Time: 11/26/2019 14:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESCOBAR,RIGOBERTO,  
BREATH TEST OPERATOR AGENCY: FHP Troop K  
ARRESTING OFFICER: ESCOBAR,RIGOBERTO  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646645  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: ABELL

DATE/TIME OF TEST: 10/30/2019 23:37:00

---

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.079	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.234	23:32
Breath Volume	2.128	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.236	23:35
Breath Volume	1.531	23:35
Air Blank	0.000	23:36
Control Test	0.077	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:37

---

Data Download Date/Time: 11/26/2019 14:36



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONDELLO, ANTHONY, J  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: DODD, MARCUS  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: SIRRELL

DATE/TIME OF TEST: 10/04/2019 03:10:50

---

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.221	03:06
Breath Volume	1.566	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.210	03:09
Breath Volume	2.128	03:09
Air Blank	0.000	03:09
Control Test	0.079	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

---

Data Download Date/Time: 11/26/2019 15:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUNN,COLLIN,J  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: HELLER,M  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: FINES

DATE/TIME OF TEST: 10/05/2019 00:57:44

---

Diagnostics Check		OK	00:49
Air Blank		0.000	00:50
Control Test		0.079	00:50
Air Blank		0.000	00:51
Subject Sample	# 1	0.081	00:52
Breath Volume		1.601	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:54
Subject Sample	# 2	0.077	00:56
Breath Volume		1.335	00:56
Air Blank		0.000	00:56
Control Test		0.079	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

---

Data Download Date/Time: 11/26/2019 15:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ SUERO,JOSE,M  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: PEREZ SUERO,JOSE  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: AVRAMICH

DATE/TIME OF TEST: 10/21/2019 22:26:05

---

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.169	22:19
Breath Volume	1.390	22:19
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.158	22:24
Breath Volume	1.570	22:24
Air Blank	0.000	22:25
Control Test	0.081	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

---

Data Download Date/Time: 11/26/2019 15:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,ALBERT,S  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: RODRIGUEZ,ALBERT  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: ASTACIO

DATE/TIME OF TEST: 10/26/2019 04:14:09

---

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.000	04:09
Breath Volume	2.925	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.000	04:12
Breath Volume	2.425	04:12
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

---

Data Download Date/Time: 11/26/2019 15:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYONS,DANIEL,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: DOYEN,ALEXIA  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 10/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: DASILVA

DATE/TIME OF TEST: 10/26/2019 19:53:47

---

Diagnostics Check	OK	19:47
Air Blank	0.000	19:47
Control Test	0.079	19:47
Air Blank	0.000	19:48
Subject Sample # 1	0.186	19:49
Breath Volume	1.785	19:49
Air Blank	0.000	19:50
Air Blank	0.000	19:51
Subject Sample # 2	0.188	19:52
Breath Volume	1.558	19:52
Air Blank	0.000	19:52
Control Test	0.078	19:53
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

---

Data Download Date/Time: 11/26/2019 15:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESCHBACH,DANIEL,M  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: TEAT,MATTHEW  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 09/06/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: DICKERSON

DATE/TIME OF TEST: 10/11/2019 23:05:59

---

Diagnostics Check	OK	22:59
Air Blank	0.000	23:00
Control Test	0.080	23:00
Air Blank	0.000	23:01
Subject Sample # 1	0.193	23:01
Breath Volume	4.093	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.188	23:04
Breath Volume	4.167	23:04
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

---

Data Download Date/Time: 10/30/2019 15:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEIDNER,CODY,M  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: GIVANS,GARRETH  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 09/06/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: AULT

DATE/TIME OF TEST: 10/17/2019 02:22:56

---

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.242	02:18
Breath Volume	2.648	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.246	02:21
Breath Volume	1.906	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

---

Data Download Date/Time: 10/30/2019 15:10

