

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CAPLAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BALTAZAR

DATE/TIME OF TEST: 02/10/2019 04:17:30

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.196	04:12
Breath Volume		2.472	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.198	04:15
Breath Volume		2.507	04:15
Air Blank		0.000	04:16
Control Test		0.079	04:16
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 03/28/2019 23:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,O
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: COALE

DATE/TIME OF TEST: 02/11/2019 04:17:59

Diagnostics Check		OK	04:08
Air Blank		0.000	04:09
Control Test		0.079	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.235	04:10
Breath Volume		2.093	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:12
Subject Sample	# 2	0.214	04:13
Breath Volume		1.230	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:15
Subject Sample	# 3	0.233	04:16
Breath Volume		2.718	04:16
Air Blank		0.000	04:17
Control Test		0.078	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 03/28/2019 23:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: KRZYWICKI

DATE/TIME OF TEST: 02/12/2019 21:38:25

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.079	21:32
Air Blank	0.000	21:33
Subject Sample # 1	0.178	21:33
Breath Volume	3.535	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.188	21:36
Breath Volume	3.839	21:36
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 03/28/2019 23:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENOVIL,GEDEON,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRENOVIL,GEDEON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:46

SUBJECT'S LAST NAME: MYRIE

DATE/TIME OF TEST: 02/13/2019 12:18:39

Diagnostics Check	OK	12:10
Air Blank	0.000	12:10
Control Test	0.080	12:11
Air Blank	0.000	12:11
Subject Sample # 1	0.000	12:14
Breath Volume	3.394	12:14
Air Blank	0.000	12:14
Air Blank	0.000	12:16
Subject Sample # 2	0.000	12:17
Breath Volume	2.316	12:17
Air Blank	0.000	12:17
Control Test	0.079	12:18
Air Blank	0.000	12:18
Diagnostics Check	OK	12:18

Data Download Date/Time: 03/28/2019 23:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CAPLAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: LORENZO PEDRO

DATE/TIME OF TEST: 02/13/2019 23:35:42

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.079	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.265	23:30
Breath Volume	2.136	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.256	23:34
Breath Volume	1.710	23:34
Air Blank	0.000	23:34
Control Test	0.079	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 03/28/2019 23:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 02/14/2019 00:15:02

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:06	
Control Test		0.079	00:06	
Air Blank		0.000	00:07	
Subject Sample	# 1	0.173	00:10	Volume Not Met
Breath Volume		0.015	00:10	
Air Blank		0.000	00:10	
Air Blank		0.000	00:12	
Subject Sample	# 2		00:13	Subject Test Refused
Breath Volume		0.000	00:13	
Air Blank		0.000	00:14	
Control Test		0.079	00:14	
Air Blank		0.000	00:14	
Diagnostics Check		OK	00:15	

Data Download Date/Time: 03/28/2019 23:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTY,J,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CUTY,J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: LEROY

DATE/TIME OF TEST: 02/16/2019 19:28:00

Diagnostics Check	OK	19:21
Air Blank	0.000	19:21
Control Test	0.079	19:22
Air Blank	0.000	19:22
Subject Sample # 1	0.258	19:23
Breath Volume	1.996	19:23
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample # 2	0.258	19:26
Breath Volume	1.558	19:26
Air Blank	0.000	19:27
Control Test	0.079	19:27
Air Blank	0.000	19:27
Diagnostics Check	OK	19:28

Data Download Date/Time: 03/28/2019 23:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: VENTURA-GARCIA

DATE/TIME OF TEST: 02/16/2019 23:08:13

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.079	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.206	23:03
Breath Volume	2.511	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.206	23:06
Breath Volume	2.890	23:06
Air Blank	0.000	23:07
Control Test	0.079	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 03/28/2019 23:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 02/17/2019 21:04:58

Diagnostics Check		OK	20:57
Air Blank		0.000	20:57
Control Test		0.080	20:58
Air Blank		0.000	20:58
Subject Sample	# 1	0.198	21:00
Breath Volume		1.550	21:00
Air Blank		0.000	21:00
Air Blank		0.000	21:02
Subject Sample	# 2	0.203	21:03
Breath Volume		1.601	21:03
Air Blank		0.000	21:04
Control Test		0.080	21:04
Air Blank		0.000	21:04
Diagnostics Check		OK	21:04

Data Download Date/Time: 03/28/2019 23:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETSCHA,GEORGE,K
BREATH TEST OPERATOR AGENCY: Sewall's Point PD
ARRESTING OFFICER: BETSCHA,GEORGE
ARRESTING OFFICER AGENCY: Sewall's Point PD
LAST AGENCY INSPECTION: 02/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: O'DONNELL

DATE/TIME OF TEST: 02/18/2019 20:24:45

Diagnostics Check	OK	20:20	
Air Blank	0.000	20:20	
Control Test	0.079	20:21	
Air Blank	0.000	20:21	
Subject Sample	# 1	20:23	Subject Test Refused
Breath Volume	0.000	20:23	
Air Blank	0.000	20:23	
Control Test	0.080	20:24	
Air Blank	0.000	20:24	
Diagnostics Check	OK	20:24	

Data Download Date/Time: 03/28/2019 23:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEEDHAM,K,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: LEEDHAM,K
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:35

SUBJECT'S LAST NAME: STABB

DATE/TIME OF TEST: 02/20/2019 14:04:23

Diagnostics Check	OK	13:57
Air Blank	0.000	13:58
Control Test	0.080	13:58
Air Blank	0.000	13:58
Subject Sample # 1	0.194	13:59
Breath Volume	2.371	13:59
Air Blank	0.000	14:00
Air Blank	0.000	14:01
Subject Sample # 2	0.185	14:02
Breath Volume	1.363	14:02
Air Blank	0.000	14:03
Control Test	0.080	14:03
Air Blank	0.000	14:04
Diagnostics Check	OK	14:04

Data Download Date/Time: 03/28/2019 23:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISHAM, MATTHEW, H
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ISHAM, MATTHEW
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: ORR

DATE/TIME OF TEST: 02/21/2019 21:35:43

Diagnostics Check		OK	21:24	
Air Blank		0.000	21:24	
Control Test		0.079	21:24	
Air Blank		0.000	21:25	
Subject Sample	# 1	0.040	21:28	Volume Not Met
Breath Volume		0.226	21:28	
Air Blank		0.000	21:29	
Air Blank		0.000	21:31	
Subject Sample	# 2	0.031	21:34	Volume Not Met
Breath Volume		0.460	21:34	
Air Blank		0.000	21:34	
Control Test		0.079	21:35	
Air Blank		0.000	21:35	
Diagnostics Check		OK	21:35	

Data Download Date/Time: 03/28/2019 23:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISHAM, MATTHEW, H
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ISHAM, MATTHEW
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: ORR

DATE/TIME OF TEST: 02/21/2019 21:49:49

Diagnostics Check		OK	21:38	
Air Blank		0.000	21:38	
Control Test		0.079	21:39	
Air Blank		0.000	21:39	
Subject Sample	# 1	0.030	21:42	Volume Not Met
Breath Volume		0.316	21:42	
Air Blank		0.000	21:43	
Air Blank		0.000	21:45	
Subject Sample	# 2	0.027	21:48	Volume Not Met
Breath Volume		0.855	21:48	
Air Blank		0.000	21:48	
Control Test		0.078	21:49	
Air Blank		0.000	21:49	
Diagnostics Check		OK	21:49	

Data Download Date/Time: 03/28/2019 23:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREGORY,AMANDA,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: KASPROW,MARK
ARRESTING OFFICER AGENCY: Jupiter Island
LAST AGENCY INSPECTION: 02/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 02/22/2019 02:01:42

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.051	01:57
Breath Volume	1.808	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.051	02:00
Breath Volume	2.039	02:00
Air Blank	0.000	02:00
Control Test	0.079	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 03/28/2019 23:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/22/2019 21:16:30

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.079	21:11
Air Blank	0.000	21:11
Subject Sample # 1	0.000	21:12
Breath Volume	1.859	21:12
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.000	21:15
Breath Volume	1.636	21:15
Air Blank	0.000	21:15
Control Test	0.079	21:15
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 03/28/2019 23:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,JOSE,G
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DELGADO,JOSE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/09/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: HAMMOND

DATE/TIME OF TEST: 02/24/2019 22:43:51

Diagnostics Check		OK	22:33
Air Blank		0.000	22:33
Control Test		0.079	22:34
Air Blank		0.000	22:34
Subject Sample	# 1	0.268	22:35
Breath Volume		1.464	22:35
Air Blank		0.000	22:36
Air Blank		0.000	22:38
Subject Sample	# 2	0.232	22:39
Breath Volume		1.363	22:39
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 3	0.258	22:42
Breath Volume		1.621	22:42
Air Blank		0.000	22:42
Control Test		0.078	22:43
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 03/28/2019 23:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: HIATT

DATE/TIME OF TEST: 02/25/2019 20:42:33

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.079	20:37
Air Blank	0.000	20:37
Subject Sample # 1	0.161	20:38
Breath Volume	2.828	20:38
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.151	20:41
Breath Volume	2.664	20:41
Air Blank	0.000	20:41
Control Test	0.079	20:41
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

Data Download Date/Time: 03/28/2019 23:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: LIPNICK

DATE/TIME OF TEST: 02/15/2019 02:36:32

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.082	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.060	02:31
Breath Volume	2.441	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.062	02:34
Breath Volume	3.285	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 03/06/2019 20:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 02/16/2019 01:35:37

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.082	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.181	01:28
Breath Volume	1.191	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:30
Subject Sample # 2	0.180	01:33
Breath Volume	1.539	01:33
Air Blank	0.000	01:34
Control Test	0.083	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 03/06/2019 20:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: KRUSE,KRIS
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 02/17/2019 00:59:43

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:51	
Control Test	0.082	00:51	
Air Blank	0.000	00:52	
Subject Sample	# 1	00:55	No Sample Provided
Breath Volume	0.000	00:55	
Air Blank	0.000	00:55	
Air Blank	0.000	00:57	
Subject Sample	# 2	00:58	Subject Test Refused
Breath Volume	0.000	00:58	
Air Blank	0.000	00:58	
Control Test	0.082	00:59	
Air Blank	0.000	00:59	
Diagnostics Check	OK	00:59	

Data Download Date/Time: 03/06/2019 20:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: PETTIS

DATE/TIME OF TEST: 02/22/2019 02:26:44

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.081	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.076	02:22
Breath Volume	3.671	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.071	02:25
Breath Volume	3.066	02:25
Air Blank	0.000	02:25
Control Test	0.079	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 03/06/2019 20:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: ZIMMER

DATE/TIME OF TEST: 02/22/2019 19:04:17

Diagnostics Check	OK	18:51	
Air Blank	0.000	18:51	
Control Test	0.081	18:52	
Air Blank	0.000	18:52	
Subject Sample	# 1 0.102	18:53	
Breath Volume	1.441	18:53	
Air Blank	0.000	18:54	
Air Blank	0.000	18:55	
Subject Sample	# 2 0.074	18:59	Volume Not Met
Breath Volume	0.789	18:59	
Air Blank	0.000	18:59	
Air Blank	0.000	19:01	
Subject Sample	# 3 0.093	19:02	
Breath Volume	1.792	19:02	
Air Blank	0.000	19:03	
Control Test	0.079	19:03	
Air Blank	0.000	19:04	
Diagnostics Check	OK	19:04	

Data Download Date/Time: 03/06/2019 20:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS,HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: ATNIP

DATE/TIME OF TEST: 02/02/2019 05:47:36

Diagnostics Check	OK	05:40
Air Blank	0.000	05:40
Control Test	0.077	05:41
Air Blank	0.000	05:41
Subject Sample # 1	0.171	05:42
Breath Volume	1.484	05:42
Air Blank	0.000	05:42
Air Blank	0.000	05:44
Subject Sample # 2	0.161	05:45
Breath Volume	1.351	05:45
Air Blank	0.000	05:46
Control Test	0.078	05:46
Air Blank	0.000	05:47
Diagnostics Check	OK	05:47

Data Download Date/Time: 02/28/2019 18:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS,HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/04/2019 00:55:16

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.079	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.000	00:50
Breath Volume	3.167	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.000	00:53
Breath Volume	1.867	00:53
Air Blank	0.000	00:54
Control Test	0.078	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 02/28/2019 18:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRITY, THOMAS, A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GARRITY, THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 02/09/2019 01:55:27

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.078	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.116	01:50
Breath Volume	1.343	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.123	01:53
Breath Volume	1.363	01:53
Air Blank	0.000	01:54
Control Test	0.077	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 02/28/2019 18:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBIN,GASTON,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: QUIMBAYO,MONICA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 02/11/2019 00:54:51

Diagnostics Check		OK	00:48
Air Blank		0.000	00:48
Control Test		0.076	00:48
Air Blank		0.000	00:49
Subject Sample	# 1	0.244	00:50
Breath Volume		1.523	00:50
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.231	00:53
Breath Volume		1.777	00:53
Air Blank		0.000	00:53
Control Test		0.077	00:54
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 02/28/2019 18:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBIN,GASTON,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: SHAOUNI,KALEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A3
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: JONAS

DATE/TIME OF TEST: 02/23/2019 04:15:50

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.078	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.157	04:11
Breath Volume	1.496	04:11
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.152	04:14
Breath Volume	1.472	04:14
Air Blank	0.000	04:14
Control Test	0.078	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 02/28/2019 18:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: HESTER,JAMES
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 01/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 02/03/2019 01:29:17

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.121	01:24
Breath Volume	1.562	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.119	01:27
Breath Volume	1.496	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 02/22/2019 00:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANO,MAIRA,H
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: KEARNS,HUNTER
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 02/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 02/23/2019 05:34:47

Diagnostics Check	OK	05:29	
Air Blank	0.000	05:30	
Control Test	0.081	05:30	
Air Blank	0.000	05:31	
Subject Sample	# 1	05:33	Subject Test Refused
Breath Volume	0.097	05:33	
Air Blank	0.000	05:33	
Control Test	0.082	05:34	
Air Blank	0.000	05:34	
Diagnostics Check	OK	05:34	

Data Download Date/Time: 03/12/2019 22:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: EDGE

DATE/TIME OF TEST: 02/07/2019 02:54:23

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.079	02:39	
Air Blank	0.000	02:39	
Subject Sample	# 1 0.141	02:41	
Breath Volume	1.277	02:41	
Air Blank	0.000	02:42	
Air Blank	0.000	02:44	
Subject Sample	# 2 0.118	02:47	Volume Not Met
Breath Volume	0.218	02:47	
Air Blank	0.000	02:48	
Air Blank	0.000	02:49	
Subject Sample	# 3 0.139	02:52	
Breath Volume	1.195	02:52	
Air Blank	0.000	02:53	
Control Test	0.079	02:53	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 02/26/2019 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: PINEDA GARCIA

DATE/TIME OF TEST: 02/16/2019 03:00:15

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.079	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.168	02:55
Breath Volume	1.855	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.171	02:58
Breath Volume	1.601	02:58
Air Blank	0.000	02:59
Control Test	0.079	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 02/26/2019 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: GONZALEZ ZAMORA

DATE/TIME OF TEST: 02/18/2019 03:13:30

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.079	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.180	03:09
Breath Volume	1.902	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.184	03:11
Breath Volume	2.449	03:11
Air Blank	0.000	03:12
Control Test	0.078	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 02/26/2019 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 02/28/2019 23:17:53

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.080	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.000	23:13
Breath Volume	1.832	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.000	23:16
Breath Volume	1.773	23:16
Air Blank	0.000	23:16
Control Test	0.078	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 03/25/2019 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BRACK,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: DENT

DATE/TIME OF TEST: 02/01/2019 02:10:34

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.080	02:05
Air Blank		0.000	02:05
Subject Sample	# 1	0.202	02:06
Breath Volume		1.175	02:06
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.199	02:08
Breath Volume		1.820	02:08
Air Blank		0.000	02:09
Control Test		0.080	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 02/08/2019 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZEITZ,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 02/04/2019 00:19:27

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.179	00:14
Breath Volume	1.941	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.174	00:17
Breath Volume	1.753	00:17
Air Blank	0.000	00:18
Control Test	0.079	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 02/08/2019 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ALDOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MUNROE

DATE/TIME OF TEST: 02/04/2019 01:19:43

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.093	01:15
Breath Volume	2.121	01:15
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.094	01:18
Breath Volume	1.535	01:18
Air Blank	0.000	01:18
Control Test	0.080	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 02/08/2019 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 02/04/2019 04:16:37

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.159	04:11
Breath Volume	1.492	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.159	04:15
Breath Volume	2.050	04:15
Air Blank	0.000	04:15
Control Test	0.079	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 02/08/2019 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARBER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: CRAIG

DATE/TIME OF TEST: 02/04/2019 22:28:11

Diagnostics Check	OK	22:18	
Air Blank	0.000	22:19	
Control Test	0.081	22:19	
Air Blank	0.000	22:20	
Subject Sample	# 1	22:23	No Sample Provided
Breath Volume	0.000	22:23	
Air Blank	0.000	22:23	
Air Blank	0.000	22:25	
Subject Sample	# 2	22:26	Subject Test Refused
Breath Volume	0.000	22:26	
Air Blank	0.000	22:27	
Control Test	0.081	22:27	
Air Blank	0.000	22:28	
Diagnostics Check	OK	22:28	

Data Download Date/Time: 02/08/2019 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JON
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BHAGWANDEEN

DATE/TIME OF TEST: 02/05/2019 01:10:49

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.202	01:02
Breath Volume	2.171	01:02
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.181	01:06
Breath Volume	1.644	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 3	0.199	01:09
Breath Volume	2.359	01:09
Air Blank	0.000	01:09
Control Test	0.077	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 02/08/2019 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DALTON,RYAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: DONIS ALVAREZ

DATE/TIME OF TEST: 02/05/2019 22:39:53

Diagnostics Check	OK	22:32
Air Blank	0.000	22:32
Control Test	0.080	22:33
Air Blank	0.000	22:33
Subject Sample # 1	0.151	22:35
Breath Volume	3.429	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.154	22:38
Breath Volume	3.144	22:38
Air Blank	0.000	22:38
Control Test	0.080	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 02/08/2019 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCDONOUGH,MARK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 02/07/2019 00:46:01

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.080	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.170	00:41
Breath Volume		2.406	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:43
Subject Sample	# 2	0.154	00:44
Breath Volume		1.117	00:44
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:45
Diagnostics Check		OK	00:46

Data Download Date/Time: 02/08/2019 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MOHAMMED,JOSHUA
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 02/09/2019 03:05:10

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.126	03:00
Breath Volume	1.335	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.134	03:03
Breath Volume	2.003	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: REYES YANEZ

DATE/TIME OF TEST: 02/09/2019 05:08:58

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.079	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.155	05:04
Breath Volume	1.933	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:06
Subject Sample # 2	0.145	05:07
Breath Volume	1.292	05:07
Air Blank	0.000	05:08
Control Test	0.079	05:08
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CANTELAR,JOAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: LINARES

DATE/TIME OF TEST: 02/09/2019 17:44:24

Diagnostics Check	OK	17:37
Air Blank	0.000	17:37
Control Test	0.080	17:38
Air Blank	0.000	17:38
Subject Sample # 1	0.194	17:39
Breath Volume	1.671	17:39
Air Blank	0.000	17:40
Air Blank	0.000	17:42
Subject Sample # 2	0.194	17:42
Breath Volume	1.996	17:42
Air Blank	0.000	17:43
Control Test	0.079	17:43
Air Blank	0.000	17:44
Diagnostics Check	OK	17:44

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/09/2019 20:43:06

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.080	20:37
Air Blank	0.000	20:37
Subject Sample # 1	0.202	20:38
Breath Volume	2.027	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.190	20:41
Breath Volume	1.531	20:41
Air Blank	0.000	20:42
Control Test	0.079	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: JESUS MENTEJO

DATE/TIME OF TEST: 02/10/2019 04:58:43

Diagnostics Check	OK	04:52
Air Blank	0.000	04:52
Control Test	0.079	04:53
Air Blank	0.000	04:53
Subject Sample # 1	0.167	04:54
Breath Volume	3.593	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:56
Subject Sample # 2	0.173	04:57
Breath Volume	3.527	04:57
Air Blank	0.000	04:57
Control Test	0.079	04:58
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COX,TYLER
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 02/10/2019 12:21:04

Diagnostics Check	OK	12:12
Air Blank	0.000	12:12
Control Test	0.079	12:13
Air Blank	0.000	12:13
Subject Sample # 1	0.181	12:16
Breath Volume	1.769	12:16
Air Blank	0.000	12:17
Air Blank	0.000	12:18
Subject Sample # 2	0.177	12:19
Breath Volume	1.601	12:19
Air Blank	0.000	12:20
Control Test	0.077	12:20
Air Blank	0.000	12:20
Diagnostics Check	OK	12:21

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCCOY,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: LARA ESCALERA

DATE/TIME OF TEST: 02/11/2019 05:19:22

Diagnostics Check	OK	05:13
Air Blank	0.000	05:13
Control Test	0.081	05:13
Air Blank	0.000	05:14
Subject Sample # 1	0.138	05:14
Breath Volume	1.746	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 2	0.128	05:17
Breath Volume	1.210	05:17
Air Blank	0.000	05:18
Control Test	0.081	05:18
Air Blank	0.000	05:19
Diagnostics Check	OK	05:19

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STEPHAN,ROBERT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MERCADO

DATE/TIME OF TEST: 02/13/2019 23:46:39

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.080	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.113	23:42
Breath Volume	3.359	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.121	23:45
Breath Volume	4.609	23:45
Air Blank	0.000	23:45
Control Test	0.079	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYNIHAN,KERIANE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: BEZARES

DATE/TIME OF TEST: 02/16/2019 05:01:56

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.080	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.059	04:57
Breath Volume	1.175	04:57
Air Blank	0.000	04:58
Air Blank	0.000	04:59
Subject Sample # 2	0.054	05:00
Breath Volume	1.207	05:00
Air Blank	0.000	05:01
Control Test	0.079	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HALPERN,ZACHARY
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:09

SUBJECT'S LAST NAME: RANDALL

DATE/TIME OF TEST: 02/16/2019 06:46:22

Diagnostics Check	OK	06:40
Air Blank	0.000	06:40
Control Test	0.079	06:40
Air Blank	0.000	06:41
Subject Sample # 1	0.143	06:41
Breath Volume	2.347	06:41
Air Blank	0.000	06:42
Air Blank	0.000	06:44
Subject Sample # 2	0.147	06:44
Breath Volume	2.296	06:44
Air Blank	0.000	06:45
Control Test	0.080	06:45
Air Blank	0.000	06:46
Diagnostics Check	OK	06:46

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: IV,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: SAGONAS

DATE/TIME OF TEST: 02/16/2019 18:25:12

Diagnostics Check	OK	18:18
Air Blank	0.000	18:19
Control Test	0.081	18:19
Air Blank	0.000	18:20
Subject Sample # 1	0.000	18:20
Breath Volume	2.539	18:20
Air Blank	0.000	18:21
Air Blank	0.000	18:23
Subject Sample # 2	0.000	18:23
Breath Volume	3.054	18:23
Air Blank	0.000	18:24
Control Test	0.080	18:24
Air Blank	0.000	18:25
Diagnostics Check	OK	18:25

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 02/17/2019 03:25:07

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.221	03:20
Breath Volume		2.492	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.215	03:23
Breath Volume		2.238	03:23
Air Blank		0.000	03:24
Control Test		0.078	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: ORTIZ SALES

DATE/TIME OF TEST: 02/17/2019 07:33:57

Diagnostics Check		OK	07:20
Air Blank		0.000	07:20
Control Test		0.079	07:21
Air Blank		0.000	07:21
Subject Sample	# 1	0.259	07:24
Breath Volume		1.386	07:24
Air Blank		0.000	07:25
Air Blank		0.000	07:26
Subject Sample	# 2	0.231	07:28
Breath Volume		1.132	07:28
Air Blank		0.000	07:29
Air Blank		0.000	07:31
Subject Sample	# 3	0.235	07:32
Breath Volume		1.101	07:32
Air Blank		0.000	07:33
Control Test		0.076	07:33
Air Blank		0.000	07:33
Diagnostics Check		OK	07:33

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: ARROCHA

DATE/TIME OF TEST: 02/18/2019 23:25:37

Diagnostics Check		OK	23:09	
Air Blank		0.000	23:09	
Control Test		0.080	23:10	
Air Blank		0.000	23:10	
Subject Sample	# 1	0.056	23:13	Volume Not Met
Breath Volume		0.218	23:13	
Air Blank		0.000	23:14	
Air Blank		0.000	23:16	
Subject Sample	# 2	0.054	23:18	
Breath Volume		1.714	23:18	
Air Blank		0.000	23:19	
Air Blank		0.000	23:20	
Subject Sample	# 3	0.052	23:24	Volume Not Met
Breath Volume		0.019	23:24	
Air Blank		0.000	23:24	
Control Test		0.078	23:25	
Air Blank		0.000	23:25	
Diagnostics Check		OK	23:25	

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: CHILEL-OCHA

DATE/TIME OF TEST: 02/19/2019 04:22:22

Diagnostics Check		OK	04:15
Air Blank		0.000	04:16
Control Test		0.080	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.208	04:17
Breath Volume		2.242	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.202	04:20
Breath Volume		1.988	04:20
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LESLIE,OMAR
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: SOLOMON

DATE/TIME OF TEST: 02/20/2019 21:53:11

Diagnostics Check	OK	21:47
Air Blank	0.000	21:47
Control Test	0.080	21:48
Air Blank	0.000	21:48
Subject Sample # 1	0.000	21:48
Breath Volume	1.894	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.000	21:51
Breath Volume	2.265	21:51
Air Blank	0.000	21:52
Control Test	0.080	21:52
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: PABLO VELASQUEZ

DATE/TIME OF TEST: 02/22/2019 21:19:52

Diagnostics Check	OK	21:11
Air Blank	0.000	21:12
Control Test	0.080	21:12
Air Blank	0.000	21:13
Subject Sample # 1	0.138	21:14
Breath Volume	2.414	21:14
Air Blank	0.000	21:14
Air Blank	0.000	21:16
Subject Sample # 2	0.137	21:18
Breath Volume	2.554	21:18
Air Blank	0.000	21:18
Control Test	0.080	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOLOWAY,ALAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: VOLUNGIS

DATE/TIME OF TEST: 02/22/2019 22:27:36

Diagnostics Check	OK	22:20
Air Blank	0.000	22:21
Control Test	0.080	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.176	22:22
Breath Volume	1.683	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.183	22:25
Breath Volume	2.082	22:25
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: HEAVENER

DATE/TIME OF TEST: 02/23/2019 02:16:57

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.080	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.199	02:12
Breath Volume	1.117	02:12
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.202	02:15
Breath Volume	1.457	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREG
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: VIVIAN

DATE/TIME OF TEST: 02/23/2019 21:26:27

Diagnostics Check		OK	21:11	
Air Blank		0.000	21:12	
Control Test		0.080	21:12	
Air Blank		0.000	21:13	
Subject Sample	# 1	0.078	21:16	Volume Not Met
Breath Volume		0.035	21:16	
Air Blank		0.000	21:16	
Air Blank		0.000	21:18	
Subject Sample	# 2	0.087	21:21	
Breath Volume		1.226	21:21	
Air Blank		0.000	21:22	
Air Blank		0.000	21:23	
Subject Sample	# 3	0.086	21:24	
Breath Volume		1.128	21:24	
Air Blank		0.000	21:25	
Control Test		0.079	21:25	
Air Blank		0.000	21:26	
Diagnostics Check		OK	21:26	

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JENNINGS,DAVON
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: TIERNEY

DATE/TIME OF TEST: 02/24/2019 01:53:47

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.080	01:48	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:52	
Control Test	0.080	01:53	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JENNINGS,DAVON
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: TIERNEY

DATE/TIME OF TEST: 02/24/2019 02:24:28

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
Control Test	0.073	02:23	Control Outside Tolerance
Air Blank	0.000	02:24	

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: GONZALEZ DORANTE

DATE/TIME OF TEST: 02/24/2019 23:58:24

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.080	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.140	23:53
Breath Volume	2.035	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.145	23:56
Breath Volume	1.917	23:56
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 02/25/2019 02:29:37

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.080	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.169	02:24
Breath Volume		1.695	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.168	02:28
Breath Volume		1.804	02:28
Air Blank		0.000	02:28
Control Test		0.079	02:29
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-
RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: NARINE

DATE/TIME OF TEST: 02/27/2019 00:50:46

Diagnostics Check		OK	00:44
Air Blank		0.000	00:44
Control Test		0.080	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.298	00:46
Breath Volume		2.781	00:46
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.291	00:49
Breath Volume		2.894	00:49
Air Blank		0.000	00:49
Control Test		0.079	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 03/15/2019 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BLACKBURN

DATE/TIME OF TEST: 02/17/2019 03:10:15

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.081	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.122	03:05
Breath Volume	2.855	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.111	03:08
Breath Volume	1.593	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 03/01/2019 17:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: VANDERVELDE

DATE/TIME OF TEST: 02/17/2019 20:52:25

Diagnostics Check	OK	20:46
Air Blank	0.000	20:46
Control Test	0.080	20:46
Air Blank	0.000	20:47
Subject Sample # 1	0.052	20:47
Breath Volume	2.238	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.052	20:50
Breath Volume	3.742	20:50
Air Blank	0.000	20:51
Control Test	0.081	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 03/01/2019 17:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MYERS,MATTHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: REMELIUS

DATE/TIME OF TEST: 02/03/2019 18:57:20

Diagnostics Check	OK	18:51
Air Blank	0.000	18:51
Control Test	0.080	18:52
Air Blank	0.000	18:52
Subject Sample # 1	0.000	18:53
Breath Volume	1.308	18:53
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.000	18:55
Breath Volume	1.363	18:55
Air Blank	0.000	18:56
Control Test	0.080	18:56
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 02/18/2019 12:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRYAN,STEVEN,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HODGES,MATHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/06/2019 00:49:51

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.081	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.000	00:45
Breath Volume	2.132	00:45
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.000	00:48
Breath Volume	2.062	00:48
Air Blank	0.000	00:48
Control Test	0.081	00:49
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 02/18/2019 12:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HANSARD,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: MARCHESE

DATE/TIME OF TEST: 02/07/2019 03:19:26

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.081	03:13
Air Blank		0.000	03:14
Subject Sample	# 1	0.194	03:14
Breath Volume		4.453	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.193	03:17
Breath Volume		3.214	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 02/18/2019 12:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: KARWAT

DATE/TIME OF TEST: 02/07/2019 23:40:18

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.081	23:33
Air Blank	0.000	23:34
Subject Sample # 1	0.000	23:35
Breath Volume	1.148	23:35
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.000	23:38
Breath Volume	1.324	23:38
Air Blank	0.000	23:39
Control Test	0.080	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 02/18/2019 12:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 02/01/2019 01:00:37

Diagnostics Check		OK	00:54
Air Blank		0.000	00:54
Control Test		0.082	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.172	00:55
Breath Volume		1.796	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.170	00:58
Breath Volume		1.507	00:58
Air Blank		0.000	00:59
Control Test		0.081	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 02/04/2019 19:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 02/01/2019 04:28:45

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.081	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.119	04:24
Breath Volume	2.332	04:24
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.123	04:27
Breath Volume	3.117	04:27
Air Blank	0.000	04:27
Control Test	0.080	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 02/04/2019 19:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: ALEMAN LOPEZ

DATE/TIME OF TEST: 02/01/2019 23:02:27

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.081	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.078	22:57
Breath Volume	1.722	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.084	23:00
Breath Volume	2.761	23:00
Air Blank	0.000	23:01
Control Test	0.081	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 02/04/2019 19:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 02/02/2019 01:34:03

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.081	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.099	01:29
Breath Volume	2.445	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:31
Subject Sample # 2	0.096	01:32
Breath Volume	2.710	01:32
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:34

Data Download Date/Time: 02/04/2019 19:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: DIAZ GOMEZ

DATE/TIME OF TEST: 02/02/2019 03:23:02

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.080	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.130	03:17
Breath Volume	1.488	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.136	03:21
Breath Volume	1.882	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:23

Data Download Date/Time: 02/04/2019 19:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: CROWTHER

DATE/TIME OF TEST: 02/02/2019 04:25:12

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:12	
Control Test		0.081	04:12	
Air Blank		0.000	04:13	
Subject Sample	# 1	0.095	04:16	Volume Not Met
Breath Volume		0.171	04:16	
Air Blank		0.000	04:17	
Air Blank		0.000	04:18	
Subject Sample	# 2	0.098	04:20	
Breath Volume		2.171	04:20	
Air Blank		0.000	04:21	
Air Blank		0.000	04:22	
Subject Sample	# 3	0.097	04:23	
Breath Volume		1.632	04:23	
Air Blank		0.000	04:24	
Control Test		0.079	04:24	
Air Blank		0.000	04:25	
Diagnostics Check		OK	04:25	

Data Download Date/Time: 02/04/2019 19:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BELEN,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:37

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 02/02/2019 10:04:15

Diagnostics Check	OK	09:58
Air Blank	0.000	09:58
Control Test	0.081	09:58
Air Blank	0.000	09:59
Subject Sample # 1	0.000	09:59
Breath Volume	2.246	09:59
Air Blank	0.000	10:00
Air Blank	0.000	10:02
Subject Sample # 2	0.000	10:02
Breath Volume	2.234	10:02
Air Blank	0.000	10:03
Control Test	0.081	10:03
Air Blank	0.000	10:04
Diagnostics Check	OK	10:04

Data Download Date/Time: 02/04/2019 19:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROSALES,F
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 01/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: ALVAREZ SALGADO

DATE/TIME OF TEST: 02/02/2019 22:54:38

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.081	22:43
Air Blank	0.000	22:44
Subject Sample # 1	0.267	22:45
Breath Volume	2.074	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.229	22:49
Breath Volume	1.207	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 3	0.249	22:52
Breath Volume	1.601	22:52
Air Blank	0.000	22:53
Control Test	0.077	22:54
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 02/04/2019 19:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER, T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: OHNSTAD

DATE/TIME OF TEST: 02/03/2019 01:00:51

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.080	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.212	00:56
Breath Volume	3.179	00:56
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.211	00:59
Breath Volume	3.824	00:59
Air Blank	0.000	00:59
Control Test	0.080	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 02/04/2019 19:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: LARSEN

DATE/TIME OF TEST: 02/03/2019 04:51:21

Diagnostics Check	OK	04:43
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.168	04:46
Breath Volume	1.437	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.157	04:49
Breath Volume	1.187	04:49
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 02/04/2019 19:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: RAD

DATE/TIME OF TEST: 02/03/2019 23:44:00

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.081	23:38
Air Blank	0.000	23:38
Subject Sample # 1	0.196	23:39
Breath Volume	2.769	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:41
Subject Sample # 2	0.197	23:42
Breath Volume	1.996	23:42
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:44

Data Download Date/Time: 02/04/2019 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 02/04/2019 03:48:05

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.177	03:43
Breath Volume	1.816	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.177	03:46
Breath Volume	1.746	03:46
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 02/04/2019 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 02/04/2019 05:23:29

Diagnostics Check		OK	05:17
Air Blank		0.000	05:17
Control Test		0.080	05:18
Air Blank		0.000	05:18
Subject Sample	# 1	0.154	05:18
Breath Volume		1.300	05:18
Air Blank		0.000	05:19
Air Blank		0.000	05:21
Subject Sample	# 2	0.155	05:21
Breath Volume		1.675	05:21
Air Blank		0.000	05:22
Control Test		0.080	05:22
Air Blank		0.000	05:23
Diagnostics Check		OK	05:23

Data Download Date/Time: 02/04/2019 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/19/2019 02:11:39

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.081	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.070	02:06
Breath Volume	1.886	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.067	02:10
Breath Volume	1.617	02:10
Air Blank	0.000	02:10
Control Test	0.080	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ATABAY

DATE/TIME OF TEST: 02/19/2019 21:57:15

Diagnostics Check		OK	21:50
Air Blank		0.000	21:51
Control Test		0.081	21:51
Air Blank		0.000	21:52
Subject Sample	# 1	0.220	21:52
Breath Volume		2.378	21:52
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.218	21:55
Breath Volume		1.449	21:55
Air Blank		0.000	21:56
Control Test		0.080	21:56
Air Blank		0.000	21:57
Diagnostics Check		OK	21:57

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/20/2019 05:23:18

Diagnostics Check	OK	05:17
Air Blank	0.000	05:17
Control Test	0.081	05:18
Air Blank	0.000	05:18
Subject Sample # 1	0.000	05:19
Breath Volume	3.191	05:19
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 2	0.000	05:21
Breath Volume	3.175	05:21
Air Blank	0.000	05:22
Control Test	0.081	05:22
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FOSTER,E 5143
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:27

SUBJECT'S LAST NAME: CELIS

DATE/TIME OF TEST: 02/20/2019 14:54:10

Diagnostics Check	OK	14:48
Air Blank	0.000	14:48
Control Test	0.081	14:48
Air Blank	0.000	14:49
Subject Sample # 1	0.000	14:49
Breath Volume	2.582	14:49
Air Blank	0.000	14:50
Air Blank	0.000	14:52
Subject Sample # 2	0.000	14:52
Breath Volume	3.085	14:52
Air Blank	0.000	14:53
Control Test	0.080	14:53
Air Blank	0.000	14:54
Diagnostics Check	OK	14:54

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: REDKEY

DATE/TIME OF TEST: 02/21/2019 04:33:06

Diagnostics Check		OK	04:26
Air Blank		0.000	04:27
Control Test		0.081	04:27
Air Blank		0.000	04:28
Subject Sample	# 1	0.201	04:28
Breath Volume		2.253	04:28
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.199	04:31
Breath Volume		1.820	04:31
Air Blank		0.000	04:32
Control Test		0.079	04:32
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND, M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: PANTELIS

DATE/TIME OF TEST: 02/22/2019 04:51:34

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.081	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.157	04:46
Breath Volume	1.539	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.173	04:49
Breath Volume	2.242	04:49
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LOPEZ,C
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 02/22/2019 17:31:36

Diagnostics Check	OK	17:21
Air Blank	0.000	17:22
Control Test	0.080	17:22
Air Blank	0.000	17:22
Subject Sample # 1	0.184	17:23
Breath Volume	1.847	17:23
Air Blank	0.000	17:24
Air Blank	0.000	17:25
Subject Sample # 2	0.206	17:26
Breath Volume	2.871	17:26
Air Blank	0.000	17:27
Air Blank	0.000	17:29
Subject Sample # 3	0.200	17:29
Breath Volume	2.621	17:29
Air Blank	0.000	17:30
Control Test	0.079	17:30
Air Blank	0.000	17:31
Diagnostics Check	OK	17:31

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: JAIMES

DATE/TIME OF TEST: 02/22/2019 22:06:51

Diagnostics Check		OK	21:53	
Air Blank		0.000	21:53	
Control Test		0.080	21:54	
Air Blank		0.000	21:54	
Subject Sample	# 1	0.056	21:57	Volume Not Met
Breath Volume		1.085	21:57	
Air Blank		0.000	21:58	
Air Blank		0.000	22:00	
Subject Sample	# 2	0.059	22:01	
Breath Volume		1.476	22:01	
Air Blank		0.000	22:02	
Air Blank		0.000	22:04	
Subject Sample	# 3	0.060	22:05	
Breath Volume		1.574	22:05	
Air Blank		0.000	22:05	
Control Test		0.079	22:06	
Air Blank		0.000	22:06	
Diagnostics Check		OK	22:06	

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NONE,NA
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 02/22/2019 22:47:37

Diagnostics Check	OK	22:41	
Air Blank	0.000	22:42	
Control Test	0.081	22:42	
Air Blank	0.000	22:43	
Subject Sample	# 1 0.121	22:43	
Breath Volume	2.773	22:43	
Air Blank	0.000	22:44	
Air Blank	0.000	22:46	
Subject Sample	# 2	22:46	Subject Test Refused
Breath Volume	0.000	22:46	
Air Blank	0.000	22:46	
Control Test	0.081	22:47	
Air Blank	0.000	22:47	
Diagnostics Check	OK	22:47	

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: TETREAULT

DATE/TIME OF TEST: 02/23/2019 02:54:55

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.081	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.113	02:49
Breath Volume	1.562	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.112	02:53
Breath Volume	2.988	02:53
Air Blank	0.000	02:53
Control Test	0.080	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: GRESHAM

DATE/TIME OF TEST: 02/23/2019 03:22:00

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.081	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.194	03:16
Breath Volume	1.585	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.196	03:20
Breath Volume	1.757	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:22

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VASALLO, Y
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 02/23/2019 04:51:03

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:49	
Control Test	0.081	04:50	
Air Blank		04:50	RFI Detect
Air Blank	0.000	04:51	

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VASALLO, Y
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 02/23/2019 04:59:46

Diagnostics Check	OK	04:52
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:53
Subject Sample # 1	0.184	04:54
Breath Volume	3.421	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample # 2	0.170	04:58
Breath Volume	1.609	04:58
Air Blank	0.000	04:58
Control Test	0.079	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBERU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: MARTIN GODINEZ

DATE/TIME OF TEST: 02/24/2019 01:31:02

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.080	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.148	01:26
Breath Volume	1.707	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:28
Subject Sample # 2	0.156	01:29
Breath Volume	2.648	01:29
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:31

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PALFY,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: RAMIREZ LORENZO

DATE/TIME OF TEST: 02/24/2019 02:12:09

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.080	02:05
Air Blank		0.000	02:06
Subject Sample	# 1	0.201	02:07
Breath Volume		1.269	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.216	02:10
Breath Volume		1.476	02:10
Air Blank		0.000	02:11
Control Test		0.079	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: RAMOS CRESPO

DATE/TIME OF TEST: 02/24/2019 05:27:06

Diagnostics Check		OK	05:17
Air Blank		0.000	05:18
Control Test		0.080	05:18
Air Blank		0.000	05:18
Subject Sample	# 1	0.139	05:22
Breath Volume		1.878	05:22
Air Blank		0.000	05:22
Air Blank		0.000	05:24
Subject Sample	# 2	0.146	05:25
Breath Volume		1.671	05:25
Air Blank		0.000	05:26
Control Test		0.077	05:26
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 02/25/2019 01:37:09

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.081	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.189	01:32
Breath Volume		2.089	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.189	01:35
Breath Volume		2.031	01:35
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: BODERICK

DATE/TIME OF TEST: 02/25/2019 22:21:13

Diagnostics Check	OK	22:15
Air Blank	0.000	22:15
Control Test	0.081	22:16
Air Blank	0.000	22:16
Subject Sample # 1	0.229	22:16
Breath Volume	2.140	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.235	22:19
Breath Volume	1.148	22:19
Air Blank	0.000	22:20
Control Test	0.079	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: LYGARKOS

DATE/TIME OF TEST: 02/26/2019 03:02:26

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.080	02:55
Air Blank		0.000	02:55
Subject Sample	# 1	0.066	02:57
Breath Volume		2.003	02:57
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.057	03:00
Breath Volume		1.234	03:00
Air Blank		0.000	03:01
Control Test		0.079	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 03/13/2019 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: SOPER

DATE/TIME OF TEST: 02/26/2019 05:03:49

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.080	04:58
Air Blank	0.000	04:58
Subject Sample # 1	0.156	04:59
Breath Volume	2.714	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:01
Subject Sample # 2	0.148	05:02
Breath Volume	2.277	05:02
Air Blank	0.000	05:02
Control Test	0.080	05:03
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 03/13/2019 18:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: COZIK

DATE/TIME OF TEST: 02/27/2019 01:15:58

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.081	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.304	01:11
Breath Volume	3.179	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:13
Subject Sample # 2	0.291	01:14
Breath Volume	2.878	01:14
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 03/13/2019 18:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: ALEXANDER JR

DATE/TIME OF TEST: 02/28/2019 00:11:09

Diagnostics Check		OK	00:03
Air Blank		0.000	00:04
Control Test		0.080	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.136	00:06
Breath Volume		1.601	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:08
Subject Sample	# 2	0.141	00:09
Breath Volume		3.214	00:09
Air Blank		0.000	00:10
Control Test		0.080	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 03/13/2019 18:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHAEFER,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: BAUTISTA HERNANDEZ

DATE/TIME OF TEST: 02/28/2019 02:01:12

Diagnostics Check		OK	01:53
Air Blank		0.000	01:53
Control Test		0.081	01:54
Air Blank		0.000	01:54
Subject Sample	# 1	0.156	01:56
Breath Volume		1.171	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:58
Subject Sample	# 2	0.151	01:59
Breath Volume		2.691	01:59
Air Blank		0.000	02:00
Control Test		0.079	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 03/13/2019 18:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PARLIER,STEVEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 02/01/2019 02:26:21

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.083	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.105	02:21
Breath Volume	2.898	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.103	02:24
Breath Volume	3.390	02:24
Air Blank	0.000	02:25
Control Test	0.082	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 02/13/2019 07:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESTIFO, CHARLES, T
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LANDRY, KATHERINE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 02/01/2019 03:54:53

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.083	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.000	03:49
Breath Volume	1.339	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.000	03:53
Breath Volume	1.167	03:53
Air Blank	0.000	03:53
Control Test	0.080	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 02/13/2019 07:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEWTON,JAMES,H
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: BROESCH,MICHAEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/02/2019 04:36:51

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:31
Subject Sample # 1	0.164	04:32
Breath Volume	1.984	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:34
Subject Sample # 2	0.161	04:35
Breath Volume	1.812	04:35
Air Blank	0.000	04:35
Control Test	0.080	04:36
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 02/13/2019 07:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAMUEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: WILKEY

DATE/TIME OF TEST: 02/03/2019 03:32:38

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.082	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.080	03:28
Breath Volume	4.042	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.080	03:31
Breath Volume	3.855	03:31
Air Blank	0.000	03:31
Control Test	0.081	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 02/13/2019 07:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/25/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: SERVICE

DATE/TIME OF TEST: 02/05/2019 01:23:47

Diagnostics Check		OK	01:13
Air Blank		0.000	01:14
Control Test		0.082	01:14
Air Blank		0.000	01:15
Subject Sample	# 1	0.209	01:15
Breath Volume		1.683	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.174	01:19
Breath Volume		1.761	01:19
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 3	0.195	01:22
Breath Volume		2.492	01:22
Air Blank		0.000	01:22
Control Test		0.081	01:23
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 02/13/2019 07:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: FINKLEA

DATE/TIME OF TEST: 02/05/2019 04:36:55

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:27	
Control Test	0.081	04:28	
Air Blank	0.000	04:28	
Subject Sample	# 1	04:31	No Sample Provided
Breath Volume	0.000	04:31	
Air Blank	0.000	04:32	
Air Blank	0.000	04:34	
Subject Sample	# 2	04:35	Subject Test Refused
Breath Volume	0.000	04:35	
Air Blank	0.000	04:35	
Control Test	0.080	04:36	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

Data Download Date/Time: 02/13/2019 07:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAFKJEN,THOMAS,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: SOWRY,SHANNON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 02/07/2019 15:08:18

Diagnostics Check	OK	15:02
Air Blank	0.000	15:02
Control Test	0.082	15:03
Air Blank	0.000	15:03
Subject Sample # 1	0.000	15:04
Breath Volume	2.351	15:04
Air Blank	0.000	15:04
Air Blank	0.000	15:06
Subject Sample # 2	0.000	15:06
Breath Volume	2.105	15:06
Air Blank	0.000	15:07
Control Test	0.080	15:07
Air Blank	0.000	15:08
Diagnostics Check	OK	15:08

Data Download Date/Time: 02/13/2019 07:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARTINEZ,HENDRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/08/2019 03:37:08

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.081	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.166	03:32
Breath Volume		1.199	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.161	03:35
Breath Volume		2.015	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 02/13/2019 07:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAMUEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 02/09/2019 22:31:55

Diagnostics Check	OK	22:25
Air Blank	0.000	22:25
Control Test	0.083	22:26
Air Blank	0.000	22:26
Subject Sample # 1	0.015	22:27
Breath Volume	2.980	22:27
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.014	22:30
Breath Volume	1.261	22:30
Air Blank	0.000	22:31
Control Test	0.083	22:31
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 02/13/2019 07:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/10/2019 00:07:25

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.082	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.091	00:02
Breath Volume	2.414	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.091	00:05
Breath Volume	1.808	00:05
Air Blank	0.000	00:06
Control Test	0.082	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 02/13/2019 07:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BROWN,BRENDAN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 02/10/2019 02:31:31

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.084	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.186	02:26
Breath Volume	3.085	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.183	02:29
Breath Volume	2.687	02:29
Air Blank	0.000	02:30
Control Test	0.082	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 02/13/2019 07:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS, JEREMIAH, S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS, JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: CROWE

DATE/TIME OF TEST: 02/10/2019 03:36:18

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.083	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.160	03:31
Breath Volume	3.699	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.152	03:34
Breath Volume	3.316	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 02/13/2019 07:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PARLIER,STEVEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/10/2019 21:41:07

Diagnostics Check	OK	21:34
Air Blank	0.000	21:35
Control Test	0.082	21:35
Air Blank	0.000	21:36
Subject Sample # 1	0.133	21:36
Breath Volume	2.242	21:36
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.121	21:39
Breath Volume	1.800	21:39
Air Blank	0.000	21:40
Control Test	0.081	21:40
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 02/13/2019 07:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENE,ROBERT,R
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: BURNETT,J
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: RUTTER

DATE/TIME OF TEST: 02/11/2019 19:28:28

Diagnostics Check	OK	19:22
Air Blank	0.000	19:22
Control Test	0.081	19:23
Air Blank	0.000	19:23
Subject Sample # 1	0.202	19:24
Breath Volume	3.113	19:24
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample # 2	0.199	19:26
Breath Volume	3.218	19:26
Air Blank	0.000	19:27
Control Test	0.081	19:27
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

Data Download Date/Time: 02/13/2019 07:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: BRAZWELL

DATE/TIME OF TEST: 02/12/2019 02:04:05

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:01	
Control Test	0.082	02:02	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:03	
Control Test	0.082	02:03	
Air Blank	0.000	02:04	
Diagnostics Check	OK	02:04	

Data Download Date/Time: 02/13/2019 07:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAVES,TIMOTHY,F
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: GRAVES,TIMOTHY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 08:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:53

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 02/12/2019 09:29:14

Diagnostics Check	OK	09:22
Air Blank	0.000	09:22
Control Test	0.082	09:22
Air Blank	0.000	09:23
Subject Sample # 1	0.000	09:24
Breath Volume	1.437	09:24
Air Blank	0.000	09:25
Air Blank	0.000	09:27
Subject Sample # 2	0.000	09:27
Breath Volume	1.410	09:27
Air Blank	0.000	09:28
Control Test	0.082	09:28
Air Blank	0.000	09:29
Diagnostics Check	OK	09:29

Data Download Date/Time: 02/13/2019 07:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARTINEZ,HENDRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ROTE

DATE/TIME OF TEST: 02/13/2019 01:32:00

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.081	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.107	01:27
Breath Volume	3.589	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:29
Subject Sample # 2	0.109	01:30
Breath Volume	3.164	01:30
Air Blank	0.000	01:31
Control Test	0.082	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:32

Data Download Date/Time: 02/13/2019 07:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TASCHETTA,JOSEPH,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PARLIER,STEVEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: REMATA

DATE/TIME OF TEST: 02/15/2019 02:11:59

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.077	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.221	02:06
Breath Volume	1.664	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.225	02:10
Breath Volume	3.605	02:10
Air Blank	0.000	02:11
Control Test	0.076	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 03/21/2019 07:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDRY,KATHERINE,E
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LANDRY,KATHERINE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/15/2019 02:27:15

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.076	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.192	02:21
Breath Volume	1.429	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.183	02:25
Breath Volume	1.656	02:25
Air Blank	0.000	02:26
Control Test	0.077	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 03/21/2019 07:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS, JEREMIAH, S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS, JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: HAZZARD

DATE/TIME OF TEST: 02/16/2019 02:32:30

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.079	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.238	02:28
Breath Volume		1.312	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.246	02:30
Breath Volume		1.652	02:30
Air Blank		0.000	02:31
Control Test		0.075	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 03/21/2019 07:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VO,SEBASTIAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: HEARD

DATE/TIME OF TEST: 02/16/2019 19:38:05

Diagnostics Check	OK	19:31
Air Blank	0.000	19:32
Control Test	0.079	19:32
Air Blank	0.000	19:33
Subject Sample # 1	0.078	19:33
Breath Volume	1.566	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:36
Subject Sample # 2	0.072	19:36
Breath Volume	1.304	19:36
Air Blank	0.000	19:37
Control Test	0.076	19:37
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time: 03/21/2019 07:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALON,DENNIS,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: SALON,DENNIS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: DEAKTOR

DATE/TIME OF TEST: 02/17/2019 21:57:26

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.080	21:51
Air Blank	0.000	21:52
Subject Sample # 1	0.074	21:52
Breath Volume	1.453	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.070	21:55
Breath Volume	1.917	21:55
Air Blank	0.000	21:56
Control Test	0.077	21:56
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 03/21/2019 07:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KAYLAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: TIMMONS,ALLAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 13:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 02/19/2019 14:44:40

Diagnostics Check		OK	14:35
Air Blank		0.000	14:35
Control Test		0.079	14:35
Air Blank		0.000	14:36
Subject Sample	# 1	0.285	14:37
Breath Volume		2.285	14:37
Air Blank		0.000	14:37
Air Blank		0.000	14:39
Subject Sample	# 2	0.306	14:39
Breath Volume		2.105	14:39
Air Blank		0.000	14:40
Air Blank		0.000	14:42
Subject Sample	# 3	0.283	14:43
Breath Volume		1.636	14:43
Air Blank		0.000	14:43
Control Test		0.078	14:44
Air Blank		0.000	14:44
Diagnostics Check		OK	14:44

Data Download Date/Time: 03/21/2019 07:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANNING,CARL,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MANNING,CARL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: GANDY

DATE/TIME OF TEST: 02/20/2019 19:54:23

Diagnostics Check	OK	19:48
Air Blank	0.000	19:48
Control Test	0.080	19:48
Air Blank	0.000	19:49
Subject Sample # 1	0.000	19:50
Breath Volume	5.074	19:50
Air Blank	0.000	19:50
Air Blank	0.000	19:52
Subject Sample # 2	0.000	19:53
Breath Volume	2.105	19:53
Air Blank	0.000	19:53
Control Test	0.075	19:53
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

Data Download Date/Time: 03/21/2019 07:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,JAMES,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BROWN,BRENDAN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 02/21/2019 22:54:30

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.079	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.000	22:49
Breath Volume	2.285	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.000	22:53
Breath Volume	2.273	22:53
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 03/21/2019 07:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEITE,STEVEN,G
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEITE,STEVEN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: BENTLEY

DATE/TIME OF TEST: 02/22/2019 01:56:32

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:54	
Control Test	0.079	01:54	
Air Blank	0.000	01:54	
Subject Sample	# 1	01:55	Subject Test Refused
Breath Volume	0.000	01:55	
Air Blank	0.000	01:55	
Control Test	0.080	01:55	
Air Blank	0.000	01:56	
Diagnostics Check	OK	01:56	

Data Download Date/Time: 03/21/2019 07:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VO,SEBASTIAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 02/24/2019 00:41:15

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.079	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.164	00:36
Breath Volume	1.585	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:38
Subject Sample # 2	0.152	00:39
Breath Volume	1.378	00:39
Air Blank	0.000	00:40
Control Test	0.075	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 03/21/2019 07:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGALADO,JORGE,I
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: REGALADO,JORGE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: TACUBA SANCHEZ

DATE/TIME OF TEST: 02/24/2019 21:41:55

Diagnostics Check	OK	21:35
Air Blank	0.000	21:36
Control Test	0.078	21:36
Air Blank	0.000	21:36
Subject Sample # 1	0.183	21:37
Breath Volume	1.488	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:39
Subject Sample # 2	0.177	21:40
Breath Volume	2.281	21:40
Air Blank	0.000	21:40
Control Test	0.076	21:41
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 03/21/2019 07:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VO,SEBASTIAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 02/26/2019 01:39:00

Diagnostics Check		OK	01:32
Air Blank		0.000	01:33
Control Test		0.080	01:33
Air Blank		0.000	01:33
Subject Sample	# 1	0.242	01:34
Breath Volume		2.417	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:36
Subject Sample	# 2	0.249	01:37
Breath Volume		2.718	01:37
Air Blank		0.000	01:38
Control Test		0.077	01:38
Air Blank		0.000	01:38
Diagnostics Check		OK	01:39

Data Download Date/Time: 03/21/2019 07:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KAYLAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DAVIS,KAYLAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: MEDEIROS

DATE/TIME OF TEST: 02/27/2019 06:39:47

Diagnostics Check	OK	06:33
Air Blank	0.000	06:33
Control Test	0.079	06:34
Air Blank	0.000	06:34
Subject Sample # 1	0.171	06:35
Breath Volume	2.007	06:35
Air Blank	0.000	06:35
Air Blank	0.000	06:37
Subject Sample # 2	0.176	06:38
Breath Volume	1.539	06:38
Air Blank	0.000	06:38
Control Test	0.079	06:39
Air Blank	0.000	06:39
Diagnostics Check	OK	06:39

Data Download Date/Time: 03/21/2019 07:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ,EDUARDO,M

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: RODRIGUEZ,EDUARDO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 01/29/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 640976

DRY GAS EXPIRY DATE: 11/21/2019

VIOLATION CODE: DUI

ARREST TIME: 23:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: RUBIO

DATE/TIME OF TEST: 02/01/2019 00:27:30

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:16	
Control Test	0.080	00:16	
Air Blank	0.000	00:17	
Subject Sample	# 1	00:20	No Sample Provided
Breath Volume	0.000	00:20	
Air Blank	0.000	00:20	
Air Blank	0.000	00:22	
Subject Sample	# 2	00:26	No Sample Provided
Breath Volume	0.000	00:26	
Air Blank	0.000	00:26	
Control Test	0.079	00:26	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 02/27/2019 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,EDUARDO,M
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUEZ,EDUARDO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: RUBIO

DATE/TIME OF TEST: 02/01/2019 00:48:07

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:32	
Control Test	0.079	00:32	
Air Blank	0.000	00:33	
Subject Sample	# 1 0.177	00:35	
Breath Volume	1.406	00:35	
Air Blank	0.000	00:35	
Air Blank	0.000	00:37	
Subject Sample	# 2 0.139	00:40	Volume Not Met
Breath Volume	0.335	00:40	
Air Blank	0.000	00:41	
Air Blank	0.000	00:43	
Subject Sample	# 3 0.116	00:46	Volume Not Met
Breath Volume	0.097	00:46	
Air Blank	0.000	00:47	
Control Test	0.077	00:47	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:48	

Data Download Date/Time: 02/27/2019 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,EDUARDO,M
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUEZ,EDUARDO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: RUBIO

DATE/TIME OF TEST: 02/01/2019 01:11:00

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:07	
Control Test	0.078	01:07	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:09	Subject Test Refused
Breath Volume	0.054	01:09	
Air Blank	0.000	01:10	
Control Test	0.079	01:10	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 02/27/2019 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALVEZ,JOSE,R
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PEREIRA,L
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: BATISTA

DATE/TIME OF TEST: 02/03/2019 00:23:37

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:18	
Control Test	0.080	00:18	
Air Blank	0.000	00:19	
Subject Sample	# 1	00:22	Subject Test Refused
Breath Volume	0.578	00:22	
Air Blank	0.000	00:22	
Control Test	0.081	00:23	
Air Blank	0.000	00:23	
Diagnostics Check	OK	00:23	

Data Download Date/Time: 02/27/2019 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALVEZ,JOSE,R
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PEREIRA,LUCAS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: BATISTA-BATISTA

DATE/TIME OF TEST: 02/03/2019 00:43:33

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:40	
Control Test	0.078	00:41	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:42	Subject Test Refused
Breath Volume	0.000	00:42	
Air Blank	0.000	00:42	
Control Test	0.079	00:42	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 02/27/2019 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ESPINOZA, ANAVELKYS,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: RODRIGUEZ, ARNIEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 01/29/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 640976

DRY GAS EXPIRY DATE: 11/21/2019

VIOLATION CODE: DUI

ARREST TIME: 21:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 02/15/2019 22:25:10

Diagnostics Check	OK	22:22	
Air Blank	0.000	22:22	
Control Test	0.079	22:22	
Air Blank	0.000	22:23	
Subject Sample	# 1	22:23	Subject Test Refused
Breath Volume	0.000	22:23	
Air Blank	0.000	22:24	
Control Test	0.079	22:24	
Air Blank	0.000	22:25	
Diagnostics Check	OK	22:25	

Data Download Date/Time: 02/27/2019 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: HUDSON,ANDREW
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 02/01/2019 20:19:05

Diagnostics Check	OK	20:16	
Air Blank	0.000	20:16	
Control Test	0.077	20:16	
Air Blank	0.000	20:17	
Subject Sample	# 1	20:17	Subject Test Refused
Breath Volume	0.000	20:17	
Air Blank	0.000	20:18	
Control Test	0.078	20:18	
Air Blank	0.000	20:18	
Diagnostics Check	OK	20:19	

Data Download Date/Time: 02/20/2019 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: HUDSON,ANDREW
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 02/02/2019 02:30:11

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.213	02:25
Breath Volume	2.375	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.213	02:28
Breath Volume	2.199	02:28
Air Blank	0.000	02:29
Control Test	0.076	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 02/20/2019 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: MARCONI

DATE/TIME OF TEST: 02/20/2019 03:18:54

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.078	03:16	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:17	
Control Test	0.078	03:18	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 02/20/2019 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 02/03/2019 02:31:41

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.078	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.143	02:23
Breath Volume	3.152	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.112	02:26
Breath Volume	1.355	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 3	0.134	02:29
Breath Volume	3.382	02:29
Air Blank	0.000	02:30
Control Test	0.075	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 02/20/2019 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: HUDSON,ANDREW
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 02/03/2019 04:57:42

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.078	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.122	04:53
Breath Volume	1.925	04:53
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.129	04:56
Breath Volume	2.464	04:56
Air Blank	0.000	04:56
Control Test	0.076	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 02/20/2019 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: HUDSON,ANDREW
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: SUSS

DATE/TIME OF TEST: 02/04/2019 00:38:12

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:35	
Control Test	0.078	00:35	
Air Blank	0.000	00:36	
Subject Sample	# 1	00:36	Subject Test Refused
Breath Volume	0.000	00:36	
Air Blank	0.000	00:37	
Control Test	0.077	00:37	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

Data Download Date/Time: 02/20/2019 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/06/2019 23:10:37

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.077	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.172	23:05
Breath Volume	2.378	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.172	23:08
Breath Volume	2.996	23:08
Air Blank	0.000	23:09
Control Test	0.078	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 02/20/2019 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: HUDSON,ANDREW
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: TUFO

DATE/TIME OF TEST: 02/07/2019 04:06:41

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.078	04:03	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:05	Subject Test Refused
Breath Volume	0.000	04:05	
Air Blank	0.000	04:05	
Control Test	0.077	04:06	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 02/20/2019 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ALEXANDRA,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODDY,AARON
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:47

SUBJECT'S LAST NAME: MENSE

DATE/TIME OF TEST: 02/12/2019 09:18:10

Diagnostics Check	OK	09:11
Air Blank	0.000	09:12
Control Test	0.078	09:12
Air Blank	0.000	09:12
Subject Sample # 1	0.000	09:13
Breath Volume	2.273	09:13
Air Blank	0.000	09:14
Air Blank	0.000	09:16
Subject Sample # 2	0.000	09:16
Breath Volume	1.738	09:16
Air Blank	0.000	09:17
Control Test	0.078	09:17
Air Blank	0.000	09:18
Diagnostics Check	OK	09:18

Data Download Date/Time: 02/20/2019 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MARTIN-REYES,JUAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 02/12/2019 20:37:02

Diagnostics Check	OK	20:30
Air Blank	0.000	20:30
Control Test	0.077	20:31
Air Blank	0.000	20:31
Subject Sample # 1	0.215	20:32
Breath Volume	2.617	20:32
Air Blank	0.000	20:33
Air Blank	0.000	20:34
Subject Sample # 2	0.219	20:35
Breath Volume	2.882	20:35
Air Blank	0.000	20:36
Control Test	0.077	20:36
Air Blank	0.000	20:36
Diagnostics Check	OK	20:37

Data Download Date/Time: 02/20/2019 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLIER, JEREMEY, J
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: TELLIER, JEREMEY
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: BROCKNER

DATE/TIME OF TEST: 02/18/2019 00:13:02

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:01	
Control Test	0.078	00:01	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:05	No Sample Provided
Breath Volume	0.000	00:05	
Air Blank	0.000	00:06	
Air Blank	0.000	00:08	
Subject Sample	# 2	0.163	00:08
Breath Volume	1.574	00:08	
Air Blank	0.000	00:09	
Air Blank	0.000	00:10	
Subject Sample	# 3	0.164	00:11
Breath Volume	1.632	00:11	
Air Blank	0.000	00:12	
Control Test	0.076	00:12	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:13	

Data Download Date/Time: 02/20/2019 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: BARROSO

DATE/TIME OF TEST: 02/18/2019 19:38:33

Diagnostics Check	OK	19:35	
Air Blank	0.000	19:36	
Control Test	0.077	19:36	
Air Blank	0.000	19:36	
Subject Sample	# 1	19:37	Subject Test Refused
Breath Volume	0.000	19:37	
Air Blank	0.000	19:37	
Control Test	0.077	19:37	
Air Blank	0.000	19:38	
Diagnostics Check	OK	19:38	

Data Download Date/Time: 02/20/2019 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TYLER,BRYAN,D
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: TYLER,BRYAN
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: SCHIANO DI COLA

DATE/TIME OF TEST: 02/19/2019 21:07:33

Diagnostics Check	OK	21:04	
Air Blank	0.000	21:05	
Control Test	0.078	21:05	
Air Blank	0.000	21:06	
Subject Sample	# 1	21:06	Subject Test Refused
Breath Volume	0.000	21:06	
Air Blank	0.000	21:06	
Control Test	0.079	21:06	
Air Blank	0.000	21:07	
Diagnostics Check	OK	21:07	

Data Download Date/Time: 02/20/2019 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TYLER,BRYAN,D
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: TYLER,BRYAN
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 01/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 02/20/2019 03:17:50

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:15	
Control Test	0.079	03:15	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:16	Subject Test Refused
Breath Volume	0.000	03:16	
Air Blank	0.000	03:16	
Control Test	0.077	03:17	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 02/20/2019 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 02/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 02/21/2019 21:51:20

Diagnostics Check	OK	21:48	
Air Blank	0.000	21:48	
Control Test	0.078	21:49	
Air Blank	0.000	21:49	
Subject Sample	# 1	21:49	Subject Test Refused
Breath Volume	0.000	21:49	
Air Blank	0.000	21:50	
Control Test	0.079	21:50	
Air Blank	0.000	21:51	
Diagnostics Check	OK	21:51	

Data Download Date/Time: 03/20/2019 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRIGA,JUAN,
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: CABRERA,CARLOS
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 01/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/08/2019 18:36:49

Diagnostics Check	OK	18:23
Air Blank	0.000	18:23
Control Test	0.077	18:24
Air Blank	0.000	18:24
Subject Sample # 1	0.218	18:25
Breath Volume	2.527	18:25
Air Blank	0.000	18:26
Air Blank	0.000	18:28
Subject Sample # 2	0.196	18:30
Breath Volume	1.656	18:30
Air Blank	0.000	18:31
Air Blank	0.000	18:33
Subject Sample # 3	0.183	18:35
Breath Volume	1.347	18:35
Air Blank	0.000	18:35
Control Test	0.075	18:36
Air Blank	0.000	18:36
Diagnostics Check	OK	18:36

Data Download Date/Time: 02/15/2019 07:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAIKEL,H
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: CRUZ,EDISON
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 01/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: FIQYINA

DATE/TIME OF TEST: 02/11/2019 11:59:40

Diagnostics Check		OK	11:50
Air Blank		0.000	11:50
Control Test		0.077	11:51
Air Blank		0.000	11:51
Subject Sample	# 1	0.262	11:52
Breath Volume		1.242	11:52
Air Blank		0.000	11:53
Air Blank		0.000	11:55
Subject Sample	# 2	0.254	11:57
Breath Volume		2.074	11:57
Air Blank		0.000	11:58
Control Test		0.076	11:58
Air Blank		0.000	11:59
Diagnostics Check		OK	11:59

Data Download Date/Time: 02/15/2019 07:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,BRIENNA,M
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: BROWN,BRIENNA
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 01/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/13/2019 11:05:57

Diagnostics Check	OK	10:54	
Air Blank	0.000	10:55	
Control Test	0.076	10:55	
Air Blank	0.000	10:55	
Subject Sample	# 1 0.000	10:59	Volume Not Met
Breath Volume	0.464	10:59	
Air Blank	0.000	10:59	
Air Blank	0.000	11:01	
Subject Sample	# 2 0.000	11:01	
Breath Volume	1.816	11:01	
Air Blank	0.000	11:02	
Air Blank	0.000	11:04	
Subject Sample	# 3	11:04	Subject Test Refused
Breath Volume	0.000	11:04	
Air Blank	0.000	11:05	
Control Test	0.076	11:05	
Air Blank	0.000	11:05	
Diagnostics Check	OK	11:05	

Data Download Date/Time: 02/15/2019 07:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOLOWAY,ALAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: BOUCHARD

DATE/TIME OF TEST: 02/08/2019 23:48:36

Diagnostics Check	OK	23:38
Air Blank	0.000	23:39
Control Test	0.080	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.155	23:41
Breath Volume	1.980	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.140	23:46
Breath Volume	1.421	23:46
Air Blank	0.000	23:47
Control Test	0.079	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ALDOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 02/10/2019 00:51:16

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.188	00:46
Breath Volume	3.210	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.182	00:49
Breath Volume	2.171	00:49
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOLOWAY,ALAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/10/2019 18:56:15

Diagnostics Check	OK	18:49
Air Blank	0.000	18:49
Control Test	0.080	18:50
Air Blank	0.000	18:50
Subject Sample # 1	0.199	18:51
Breath Volume	1.265	18:51
Air Blank	0.000	18:52
Air Blank	0.000	18:53
Subject Sample # 2	0.184	18:54
Breath Volume	1.910	18:54
Air Blank	0.000	18:55
Control Test	0.080	18:55
Air Blank	0.000	18:56
Diagnostics Check	OK	18:56

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: GRENZ

DATE/TIME OF TEST: 02/10/2019 19:10:41

Diagnostics Check	OK	19:04
Air Blank	0.000	19:04
Control Test	0.079	19:04
Air Blank	0.000	19:05
Subject Sample # 1	0.154	19:06
Breath Volume	3.562	19:06
Air Blank	0.000	19:06
Air Blank	0.000	19:08
Subject Sample # 2	0.150	19:09
Breath Volume	3.152	19:09
Air Blank	0.000	19:09
Control Test	0.080	19:10
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BIDO,GISELLE
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: ARAYA CORDERO

DATE/TIME OF TEST: 02/10/2019 21:38:37

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.081	21:31
Air Blank	0.000	21:32
Subject Sample # 1	0.233	21:33
Breath Volume	1.687	21:33
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.240	21:36
Breath Volume	4.066	21:36
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: QUIROS,MARCIEL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: SILVESTRE DOMINGO

DATE/TIME OF TEST: 02/10/2019 21:58:48

Diagnostics Check	OK	21:51
Air Blank	0.000	21:52
Control Test	0.080	21:52
Air Blank	0.000	21:53
Subject Sample # 1	0.221	21:54
Breath Volume	2.949	21:54
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.220	21:57
Breath Volume	3.484	21:57
Air Blank	0.000	21:57
Control Test	0.080	21:58
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLOWERS,GARY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: KAHN

DATE/TIME OF TEST: 02/11/2019 21:32:38

Diagnostics Check	OK	21:25
Air Blank	0.000	21:25
Control Test	0.080	21:26
Air Blank	0.000	21:26
Subject Sample # 1	0.170	21:27
Breath Volume	1.671	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:29
Subject Sample # 2	0.169	21:30
Breath Volume	1.921	21:30
Air Blank	0.000	21:31
Control Test	0.080	21:32
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: GASPAR LEON

DATE/TIME OF TEST: 02/12/2019 00:35:03

Diagnostics Check		OK	00:27
Air Blank		0.000	00:28
Control Test		0.080	00:28
Air Blank		0.000	00:29
Subject Sample	# 1	0.227	00:30
Breath Volume		2.078	00:30
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.217	00:33
Breath Volume		1.441	00:33
Air Blank		0.000	00:34
Control Test		0.079	00:34
Air Blank		0.000	00:34
Diagnostics Check		OK	00:35

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: PLATT

DATE/TIME OF TEST: 02/12/2019 02:42:36

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.080	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.190	02:38
Breath Volume	2.289	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.184	02:40
Breath Volume	1.886	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DALTON,RYAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: STOFFT

DATE/TIME OF TEST: 02/13/2019 21:35:30

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.080	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.246	21:30
Breath Volume	1.707	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample # 2	0.228	21:33
Breath Volume	1.558	21:33
Air Blank	0.000	21:34
Control Test	0.079	21:34
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: ANTRIM

DATE/TIME OF TEST: 02/13/2019 23:41:33

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.080	23:35
Air Blank	0.000	23:36
Subject Sample # 1	0.103	23:37
Breath Volume	2.726	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.102	23:39
Breath Volume	2.359	23:39
Air Blank	0.000	23:40
Control Test	0.080	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROMERO,JUAN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: ROULEAU

DATE/TIME OF TEST: 02/14/2019 06:34:03

Diagnostics Check		OK	06:27
Air Blank		0.000	06:27
Control Test		0.081	06:27
Air Blank		0.000	06:28
Subject Sample	# 1	0.224	06:29
Breath Volume		1.250	06:29
Air Blank		0.000	06:30
Air Blank		0.000	06:31
Subject Sample	# 2	0.235	06:32
Breath Volume		1.457	06:32
Air Blank		0.000	06:33
Control Test		0.078	06:33
Air Blank		0.000	06:33
Diagnostics Check		OK	06:34

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CADE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:35

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 02/15/2019 14:08:19

Diagnostics Check	OK	14:01
Air Blank	0.000	14:02
Control Test	0.080	14:02
Air Blank	0.000	14:02
Subject Sample # 1	0.188	14:03
Breath Volume	2.054	14:03
Air Blank	0.000	14:04
Air Blank	0.000	14:06
Subject Sample # 2	0.178	14:06
Breath Volume	1.238	14:06
Air Blank	0.000	14:07
Control Test	0.079	14:07
Air Blank	0.000	14:08
Diagnostics Check	OK	14:08

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: KERCADO

DATE/TIME OF TEST: 02/16/2019 04:38:08

Diagnostics Check		OK	04:24	
Air Blank		0.000	04:24	
Control Test		0.081	04:25	
Air Blank		0.000	04:25	
Subject Sample	# 1	0.185	04:27	No .020 agreement
Breath Volume		2.398	04:27	
Air Blank		0.000	04:28	
Air Blank		0.000	04:30	
Subject Sample	# 2	0.154	04:33	Volume Not Met
Breath Volume		0.000	04:33	
Air Blank		0.000	04:34	
Air Blank		0.000	04:35	
Subject Sample	# 3	0.162	04:36	No .020 agreement
Breath Volume		1.832	04:36	
Air Blank		0.000	04:37	
Control Test		0.078	04:37	
Air Blank		0.000	04:38	
Diagnostics Check		OK	04:38	

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: KERCADO

DATE/TIME OF TEST: 02/16/2019 04:51:46

Diagnostics Check		OK	04:39	
Air Blank		0.000	04:40	
Control Test		0.078	04:40	
Air Blank		0.000	04:41	
Subject Sample	# 1	0.152	04:44	Volume Not Met
Breath Volume		0.054	04:44	
Air Blank		0.000	04:44	
Air Blank		0.000	04:46	
Subject Sample	# 2	0.148	04:47	
Breath Volume		1.843	04:47	
Air Blank		0.000	04:48	
Air Blank		0.000	04:50	
Subject Sample	# 3		04:50	Subject Test Refused
Breath Volume		0.000	04:50	
Air Blank		0.000	04:50	
Control Test		0.078	04:51	
Air Blank		0.000	04:51	
Diagnostics Check		OK	04:51	

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 02/16/2019 11:42:21

Diagnostics Check	OK	11:35
Air Blank	0.000	11:35
Control Test	0.080	11:36
Air Blank	0.000	11:36
Subject Sample # 1	0.131	11:37
Breath Volume	1.585	11:37
Air Blank	0.000	11:38
Air Blank	0.000	11:39
Subject Sample # 2	0.125	11:40
Breath Volume	1.593	11:40
Air Blank	0.000	11:41
Control Test	0.080	11:41
Air Blank	0.000	11:42
Diagnostics Check	OK	11:42

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DENIS,ISMARY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/17/2019 04:20:03

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.167	04:15
Breath Volume	2.492	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:17
Subject Sample # 2	0.171	04:18
Breath Volume	1.730	04:18
Air Blank	0.000	04:19
Control Test	0.078	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:20

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LASSO,JAVIER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: RAMIREZ MIRANDA

DATE/TIME OF TEST: 02/17/2019 10:39:19

Diagnostics Check	OK	10:32
Air Blank	0.000	10:33
Control Test	0.080	10:33
Air Blank	0.000	10:34
Subject Sample # 1	0.150	10:34
Breath Volume	3.378	10:34
Air Blank	0.000	10:35
Air Blank	0.000	10:37
Subject Sample # 2	0.154	10:37
Breath Volume	2.179	10:37
Air Blank	0.000	10:38
Control Test	0.080	10:38
Air Blank	0.000	10:39
Diagnostics Check	OK	10:39

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: MOLINA-LOPEZ

DATE/TIME OF TEST: 02/17/2019 20:50:45

Diagnostics Check	OK	20:44
Air Blank	0.000	20:44
Control Test	0.080	20:44
Air Blank	0.000	20:45
Subject Sample # 1	0.199	20:45
Breath Volume	4.007	20:45
Air Blank	0.000	20:46
Air Blank	0.000	20:48
Subject Sample # 2	0.190	20:49
Breath Volume	3.640	20:49
Air Blank	0.000	20:49
Control Test	0.079	20:50
Air Blank	0.000	20:50
Diagnostics Check	OK	20:50

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: IV,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: MCDONOUGH

DATE/TIME OF TEST: 02/17/2019 21:07:01

Diagnostics Check	OK	21:00
Air Blank	0.000	21:00
Control Test	0.080	21:01
Air Blank	0.000	21:01
Subject Sample # 1	0.000	21:02
Breath Volume	4.250	21:02
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample # 2	0.000	21:05
Breath Volume	4.238	21:05
Air Blank	0.000	21:06
Control Test	0.080	21:06
Air Blank	0.000	21:06
Diagnostics Check	OK	21:07

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: RICHIE

DATE/TIME OF TEST: 02/17/2019 23:34:35

Diagnostics Check	OK	23:19	
Air Blank	0.000	23:20	
Control Test	0.080	23:20	
Air Blank	0.000	23:20	
Subject Sample # 1	0.251	23:22	
Breath Volume	2.316	23:22	
Air Blank	0.000	23:23	
Air Blank	0.000	23:25	
Subject Sample # 2	0.227	23:28	Volume Not Met
Breath Volume	0.105	23:28	
Air Blank	0.000	23:28	
Air Blank	0.000	23:30	
Subject Sample # 3		23:33	Subject Test Refused
Breath Volume	0.000	23:33	
Air Blank	0.000	23:33	
Control Test	0.077	23:33	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: VILLALOBOS NOLASCO

DATE/TIME OF TEST: 02/18/2019 02:08:52

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.080	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.164	02:04
Breath Volume		1.609	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:06
Subject Sample	# 2	0.164	02:07
Breath Volume		2.238	02:07
Air Blank		0.000	02:07
Control Test		0.079	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/18/2019 22:59:57

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.081	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.259	22:48
Breath Volume	1.601	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.234	22:53
Breath Volume	1.984	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 3	0.240	22:58
Breath Volume	1.816	22:58
Air Blank	0.000	22:58
Control Test	0.077	22:59
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 02/19/2019 01:25:30

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.080	01:19
Air Blank	0.000	01:20
Subject Sample # 1	0.125	01:20
Breath Volume	1.113	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.120	01:23
Breath Volume	1.667	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/19/2019 04:24:22

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.162	04:19
Breath Volume	2.015	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:21
Subject Sample # 2	0.166	04:22
Breath Volume	1.878	04:22
Air Blank	0.000	04:23
Control Test	0.079	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGRIFF,CORNELIUS
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: DEOSARAN

DATE/TIME OF TEST: 02/19/2019 17:21:43

Diagnostics Check	OK	17:14
Air Blank	0.000	17:15
Control Test	0.080	17:15
Air Blank	0.000	17:16
Subject Sample # 1	0.199	17:16
Breath Volume	1.921	17:16
Air Blank	0.000	17:17
Air Blank	0.000	17:19
Subject Sample # 2	0.199	17:20
Breath Volume	1.734	17:20
Air Blank	0.000	17:20
Control Test	0.080	17:21
Air Blank	0.000	17:21
Diagnostics Check	OK	17:21

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BRODBECK,JAMES
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: TESCHKE

DATE/TIME OF TEST: 02/20/2019 20:22:00

Diagnostics Check	OK	20:15
Air Blank	0.000	20:15
Control Test	0.081	20:15
Air Blank	0.000	20:16
Subject Sample # 1	0.117	20:17
Breath Volume	2.167	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:19
Subject Sample # 2	0.118	20:20
Breath Volume	2.453	20:20
Air Blank	0.000	20:21
Control Test	0.080	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:22

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GARCIA,KIMBERLY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: ST-PAUL

DATE/TIME OF TEST: 02/23/2019 21:56:15

Diagnostics Check	OK	21:49
Air Blank	0.000	21:50
Control Test	0.081	21:50
Air Blank	0.000	21:51
Subject Sample # 1	0.152	21:51
Breath Volume	1.648	21:51
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 2	0.149	21:54
Breath Volume	1.335	21:54
Air Blank	0.000	21:55
Control Test	0.080	21:55
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SANDERS,JOEY
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 02/24/2019 01:34:57

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.185	01:30
Breath Volume	2.804	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.182	01:33
Breath Volume	3.273	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 02/24/2019 02:25:08

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.165	02:20
Breath Volume	2.152	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample # 2	0.164	02:23
Breath Volume	2.738	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MARTINEZ HERNANDEZ

DATE/TIME OF TEST: 02/24/2019 02:46:34

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.080	02:40
Air Blank		0.000	02:41
Subject Sample	# 1	0.207	02:42
Breath Volume		1.312	02:42
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.203	02:44
Breath Volume		1.636	02:44
Air Blank		0.000	02:45
Control Test		0.079	02:45
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 03/15/2019 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DE LEON MARTINEZ

DATE/TIME OF TEST: 02/24/2019 03:32:34

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.181	03:28
Breath Volume	2.472	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.182	03:30
Breath Volume	2.164	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 03/15/2019 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BELCHER,PAUL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:38

SUBJECT'S LAST NAME: GIRON

DATE/TIME OF TEST: 02/24/2019 09:34:31

Diagnostics Check	OK	09:28
Air Blank	0.000	09:28
Control Test	0.081	09:28
Air Blank	0.000	09:29
Subject Sample # 1	0.185	09:29
Breath Volume	3.335	09:29
Air Blank	0.000	09:30
Air Blank	0.000	09:32
Subject Sample # 2	0.187	09:32
Breath Volume	3.648	09:32
Air Blank	0.000	09:33
Control Test	0.080	09:33
Air Blank	0.000	09:34
Diagnostics Check	OK	09:34

Data Download Date/Time: 03/15/2019 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASTILLO,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: TIMPSON

DATE/TIME OF TEST: 02/24/2019 16:43:53

Diagnostics Check	OK	16:37
Air Blank	0.000	16:37
Control Test	0.081	16:38
Air Blank	0.000	16:38
Subject Sample # 1	0.000	16:39
Breath Volume	3.292	16:39
Air Blank	0.000	16:39
Air Blank	0.000	16:41
Subject Sample # 2	0.000	16:42
Breath Volume	1.593	16:42
Air Blank	0.000	16:42
Control Test	0.081	16:43
Air Blank	0.000	16:43
Diagnostics Check	OK	16:43

Data Download Date/Time: 03/15/2019 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MC DERMOTT

DATE/TIME OF TEST: 02/25/2019 00:39:03

Diagnostics Check		OK	00:32
Air Blank		0.000	00:33
Control Test		0.080	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.164	00:34
Breath Volume		1.351	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.160	00:37
Breath Volume		1.507	00:37
Air Blank		0.000	00:38
Control Test		0.082	00:38
Air Blank		0.000	00:38
Diagnostics Check		OK	00:39

Data Download Date/Time: 03/15/2019 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WERTMAN,GREGORY
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: KOSTKA

DATE/TIME OF TEST: 02/25/2019 03:20:38

Diagnostics Check		OK	03:14
Air Blank		0.000	03:14
Control Test		0.081	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.229	03:16
Breath Volume		3.230	03:16
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.224	03:18
Breath Volume		3.246	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:20
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 03/15/2019 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: RESENDIZ

DATE/TIME OF TEST: 02/25/2019 03:54:30

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.195	03:49
Breath Volume	3.199	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.195	03:52
Breath Volume	3.378	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 03/15/2019 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JOHNSON,CHARLES
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:57

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 02/25/2019 14:32:33

Diagnostics Check	OK	14:25
Air Blank	0.000	14:25
Control Test	0.081	14:26
Air Blank	0.000	14:26
Subject Sample # 1	0.189	14:27
Breath Volume	1.359	14:27
Air Blank	0.000	14:28
Air Blank	0.000	14:30
Subject Sample # 2	0.184	14:30
Breath Volume	1.292	14:30
Air Blank	0.000	14:31
Control Test	0.079	14:31
Air Blank	0.000	14:32
Diagnostics Check	OK	14:32

Data Download Date/Time: 03/15/2019 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREG
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/25/2019 22:13:40

Diagnostics Check	OK	22:09	
Air Blank	0.000	22:09	
Control Test	0.080	22:10	
Air Blank	0.000	22:10	
Subject Sample	# 1	22:12	Subject Test Refused
Breath Volume	0.000	22:12	
Air Blank	0.000	22:12	
Control Test	0.082	22:13	
Air Blank	0.000	22:13	
Diagnostics Check	OK	22:13	

Data Download Date/Time: 03/15/2019 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: RICCARDI

DATE/TIME OF TEST: 02/26/2019 21:17:12

Diagnostics Check		OK	21:10
Air Blank		0.000	21:10
Control Test		0.080	21:11
Air Blank		0.000	21:11
Subject Sample	# 1	0.173	21:12
Breath Volume		2.437	21:12
Air Blank		0.000	21:13
Air Blank		0.000	21:14
Subject Sample	# 2	0.177	21:15
Breath Volume		2.652	21:15
Air Blank		0.000	21:16
Control Test		0.080	21:16
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

Data Download Date/Time: 03/15/2019 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YACINTHE,WESLEY
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: GIPNER

DATE/TIME OF TEST: 02/27/2019 00:15:33

Diagnostics Check		OK	00:09
Air Blank		0.000	00:09
Control Test		0.080	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.165	00:11
Breath Volume		1.265	00:11
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.181	00:13
Breath Volume		1.265	00:13
Air Blank		0.000	00:14
Control Test		0.080	00:14
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 03/15/2019 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PHILIPPE,BENDJY
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: ORLANDO

DATE/TIME OF TEST: 02/27/2019 01:49:34

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.081	01:43
Air Blank		0.000	01:43
Subject Sample	# 1	0.195	01:44
Breath Volume		2.601	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.175	01:47
Breath Volume		1.730	01:47
Air Blank		0.000	01:48
Control Test		0.080	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 03/15/2019 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 02/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: HERSOM

DATE/TIME OF TEST: 02/28/2019 00:21:34

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.080	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.181	00:16
Breath Volume		2.300	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:18
Subject Sample	# 2	0.162	00:19
Breath Volume		1.597	00:19
Air Blank		0.000	00:20
Control Test		0.080	00:20
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 03/15/2019 14:01