Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUERTA,EDWIN,
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD
ARRESTING OFFICER: HUERTA,EDWIN
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD
LAST AGENCY INSPECTION: 02/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 03/02/2019 04:42:01

Diagnostics Check		OK	04:32	
Air Blank		0.000	04:32	
Control Test		0.078	04:33	
Air Blank		0.000	04:33	
Subject Sample	# 1		04:36	No Sample Provided
Breath Volume		0.000	04:36	
Air Blank		0.000	04:37	
Air Blank		0.000	04:39	
Subject Sample	# 2		04:40	Subject Test Refused
Breath Volume		0.000	04:40	
Air Blank		0.000	04:41	
Control Test		0.079	04:41	
Air Blank		0.000	04:41	
Diagnostics Check		OK	04:42	

Data Download Date/Time: 03/27/2019 10:13

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUERTA,EDWIN,
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD
ARRESTING OFFICER: HUERTA,EDWIN
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD
LAST AGENCY INSPECTION: 02/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/02/2019 23:28:13

Diagnostics Check		OK	23:25	
Air Blank		0.000	23:25	
Control Test		0.077	23:26	
Air Blank		0.000	23:26	
Subject Sample	# 1		23:26	Subject Test Refused
Breath Volume		0.000	23:26	
Air Blank		0.000	23:27	
Control Test		0.077	23:27	
Air Blank		0.000	23:28	
Diagnostics Check		OK	23:28	

Data Download Date/Time: 03/27/2019 10:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVARRIA,DEGLYS,E
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: CHAVARRIA,DEGLYS
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 02/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 03/18/2019 23:48:42

Diagnostics Check		OK	23:42
Air Blank		0.000	23:43
Control Test		0.077	23:43
Air Blank		0.000	23:43
Subject Sample	# 1	0.046	23:44
Breath Volume		1.652	23:44
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.044	23:47
Breath Volume		1.222	23:47
Air Blank		0.000	23:47
Control Test		0.077	23:48
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 03/27/2019 10:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORISMOND,WISNER,
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: DORISMOND,WISNER
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 02/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: ABRAHAM

DATE/TIME OF TEST: 03/23/2019 05:13:26

Diagnostics Check		OK	05:10	
Air Blank		0.000	05:10	
Control Test		0.077	05:11	
Air Blank		0.000	05:11	
Subject Sample	# 1		05:12	Subject Test Refused
Breath Volume		0.000	05:12	
Air Blank		0.000	05:12	
Control Test		0.079	05:12	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 03/27/2019 10:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: RAWLUK,S
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: LEAK

DATE/TIME OF TEST: 03/03/2019 03:12:43

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:10	
Control Test		0.078	03:10	
Air Blank		0.000	03:11	
Subject Sample	# 1		03:11	Subject Test Refused
Breath Volume		0.000	03:11	
Air Blank		0.000	03:11	
Control Test		0.079	03:12	
Air Blank		0.000	03:12	
Diagnostics Check		OK	03:12	

Data Download Date/Time: 03/22/2019 10:12

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINLAND,JUSTIN,D
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MARQUEZ-GOMEZ,LYDIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: PROANO

DATE/TIME OF TEST: 03/04/2019 12:39:21

Diagnostics Check		OK	12:35	
Air Blank		0.000	12:35	
Control Test		0.078	12:36	
Air Blank		0.000	12:36	
Subject Sample	# 1		12:38	Subject Test Refused
Breath Volume		0.000	12:38	
Air Blank		0.000	12:38	
Control Test		0.078	12:38	
Air Blank		0.000	12:39	
Diagnostics Check		OK	12:39	

Data Download Date/Time: 03/22/2019 10:12

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA',JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: ANZARDO REYNALDO

DATE/TIME OF TEST: 03/07/2019 01:48:12

Diagnostics Check		OK	01:38
Air Blank		0.000	01:39
Control Test		0.080	01:39
Air Blank		0.000	01:39
Subject Sample	# 1	0.175	01:42
Breath Volume		1.761	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.177	01:46
Breath Volume		1.667	01:46
Air Blank		0.000	01:47
Control Test		0.080	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: GARAY-CHACON

DATE/TIME OF TEST: 03/07/2019 05:43:12

Diagnostics Check		OK	05:37
Air Blank		0.000	05:37
Control Test		0.080	05:37
Air Blank		0.000	05:38
Subject Sample	# 1	0.000	05:38
Breath Volume		1.296	05:38
Air Blank		0.000	05:39
Air Blank		0.000	05:41
Subject Sample	# 2	0.000	05:41
Breath Volume		1.195	05:41
Air Blank		0.000	05:42
Control Test		0.080	05:42
Air Blank		0.000	05:43
Diagnostics Check		OK	05:43

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: RUEDA

DATE/TIME OF TEST: 03/08/2019 03:43:53

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.000	03:39
Breath Volume		1.691	03:39
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.000	03:42
Breath Volume		1.816	03:42
Air Blank		0.000	03:42
Control Test		0.079	03:43
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,R
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: CARRICABURU

DATE/TIME OF TEST: 03/08/2019 05:27:01

Diagnostics Check		OK	05:24	
Air Blank		0.000	05:24	
Control Test		0.080	05:24	
Air Blank		0.000	05:25	
Subject Sample	# 1		05:25	Subject Test Refused
Breath Volume		0.000	05:25	
Air Blank		0.000	05:26	
Control Test		0.080	05:26	
Air Blank		0.000	05:26	
Diagnostics Check		OK	05:27	

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: ABURTO

DATE/TIME OF TEST: 03/09/2019 02:50:52

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.080	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.073	02:46
Breath Volume		2.066	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:48
Subject Sample	# 2	0.069	02:49
Breath Volume		1.718	02:49
Air Blank		0.000	02:49
Control Test		0.078	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GUASP MIRANDA

DATE/TIME OF TEST: 03/09/2019 03:41:45

Diagnostics Check		OK	03:34
Air Blank		0.000	03:35
Control Test		0.080	03:35
Air Blank		0.000	03:36
Subject Sample	# 1	0.158	03:36
Breath Volume		1.148	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.159	03:40
Breath Volume		1.617	03:40
Air Blank		0.000	03:40
Control Test		0.080	03:41
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: CARDOSA CARDOSA

DATE/TIME OF TEST: 03/09/2019 04:29:18

Diagnostics Check		OK	04:26	
Air Blank		0.000	04:26	
Control Test		0.079	04:27	
Air Blank		0.000	04:27	
Subject Sample	# 1		04:27	Subject Test Refused
Breath Volume		0.000	04:27	
Air Blank		0.000	04:28	
Control Test		0.079	04:28	
Air Blank		0.000	04:29	
Diagnostics Check		OK	04:29	

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: LOPEZ-LOPEZ

DATE/TIME OF TEST: 03/09/2019 04:42:00

Diagnostics Check		OK	04:35
Air Blank		0.000	04:35
Control Test		0.080	04:35
Air Blank		0.000	04:36
Subject Sample	# 1	0.152	04:36
Breath Volume		1.171	04:36
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.155	04:40
Breath Volume		2.101	04:40
Air Blank		0.000	04:41
Control Test		0.079	04:41
Air Blank		0.000	04:41
Diagnostics Check		OK	04:42

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: GAYVORONSKAYA

DATE/TIME OF TEST: 03/09/2019 05:04:46

Diagnostics Check		OK	04:57
Air Blank		0.000	04:58
Control Test		0.079	04:58
Air Blank		0.000	04:59
Subject Sample	# 1	0.186	05:00
Breath Volume		2.031	05:00
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.187	05:03
Breath Volume		1.585	05:03
Air Blank		0.000	05:03
Control Test		0.078	05:04
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: CORTES,OSMIN
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 03/10/2019 00:12:31

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.079	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.170	00:07
Breath Volume		1.781	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.169	00:10
Breath Volume		1.898	00:10
Air Blank		0.000	00:11
Control Test		0.078	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 03/10/2019 01:28:38

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.079	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.180	01:23
Breath Volume		2.722	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.182	01:26
Breath Volume		2.246	01:26
Air Blank		0.000	01:27
Control Test		0.079	01:27
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 03/10/2019 03:21:51

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:19	
Control Test		0.080	03:19	
Air Blank		0.000	03:20	
Subject Sample	# 1		03:20	Subject Test Refused
Breath Volume		0.000	03:20	
Air Blank		0.000	03:20	
Control Test		0.081	03:21	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAPLACELIERE,LEONARDO,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LAPLACELIERE,LEONARDO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 03/10/2019 16:41:00

Diagnostics Check		OK	16:31	
Air Blank		0.000	16:31	
Control Test		0.079	16:31	
Air Blank		0.000	16:32	
Subject Sample	# 1	0.278	16:35	
Breath Volume		1.625	16:35	
Air Blank		0.000	16:36	
Air Blank		0.000	16:37	
Subject Sample	# 2	0.247	16:39	
Breath Volume		1.957	16:39	
Air Blank			16:39	Improper Sample
Air Blank		0.000	16:40	
Control Test		0.075	16:40	
Air Blank		0.000	16:40	
Diagnostics Check		OK	16:41	

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAPLACELIERE,LEONARDO,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LAPLACELIERE,LEONARDO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 03/10/2019 16:46:28

Diagnostics Check		OK	16:43	
Air Blank		0.000	16:44	
Control Test		0.077	16:44	
Air Blank		0.000	16:44	
Subject Sample	# 1		16:45	Subject Test Refused
Breath Volume		0.000	16:45	
Air Blank		0.000	16:45	
Control Test		0.078	16:45	
Air Blank		0.000	16:46	
Diagnostics Check		OK	16:46	

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MORALES,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 03/13/2019 04:14:58

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.080	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.201	04:08
Breath Volume		1.242	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.197	04:13
Breath Volume		1.921	04:13
Air Blank		0.000	04:14
Control Test		0.079	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: ESTEBAN

DATE/TIME OF TEST: 03/13/2019 19:49:20

Diagnostics Check		OK	19:46	
Air Blank		0.000	19:46	
Control Test		0.079	19:47	
Air Blank		0.000	19:47	
Subject Sample	# 1		19:47	Subject Test Refused
Breath Volume		0.000	19:47	
Air Blank		0.000	19:48	
Control Test		0.080	19:48	
Air Blank		0.000	19:49	
Diagnostics Check		OK	19:49	

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: GONZALEZ PENA

DATE/TIME OF TEST: 03/15/2019 03:58:24

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.137	03:53
Breath Volume		1.394	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:55
Subject Sample	# 2	0.139	03:56
Breath Volume		3.226	03:56
Air Blank		0.000	03:57
Control Test		0.078	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: LORENZO

DATE/TIME OF TEST: 03/17/2019 01:07:24

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:04	
Control Test		0.080	01:04	
Air Blank		0.000	01:05	
Subject Sample	# 1		01:06	Subject Test Refused
Breath Volume		0.000	01:06	
Air Blank		0.000	01:06	
Control Test		0.080	01:06	
Air Blank		0.000	01:07	
Diagnostics Check		OK	01:07	

Data Download Date/Time: 04/18/2019 09:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: DE LEON MORALES

DATE/TIME OF TEST: 03/17/2019 02:49:18

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.079	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.148	02:44
Breath Volume		2.320	02:44
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.151	02:47
Breath Volume		1.976	02:47
Air Blank		0.000	02:48
Control Test		0.078	02:48
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORGES,SEAN,P
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: BORGES,SEAN
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:52

SUBJECT'S LAST NAME: BELISARIO

DATE/TIME OF TEST: 03/17/2019 07:17:43

Diagnostics Check		OK	07:14	
Air Blank		0.000	07:15	
Control Test		0.080	07:15	
Air Blank		0.000	07:16	
Subject Sample	# 1		07:16	Subject Test Refused
Breath Volume		0.000	07:16	
Air Blank		0.000	07:16	
Control Test		0.079	07:17	
Air Blank		0.000	07:17	
Diagnostics Check		OK	07:17	

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: NAPOLES SILVA

DATE/TIME OF TEST: 03/18/2019 00:23:07

Diagnostics Check		OK	00:15
Air Blank		0.000	00:16
Control Test		0.080	00:16
Air Blank		0.000	00:16
Subject Sample	# 1	0.102	00:18
Breath Volume		2.589	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:20
Subject Sample	# 2	0.100	00:21
Breath Volume		1.363	00:21
Air Blank		0.000	00:22
Control Test		0.079	00:22
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 03/18/2019 01:11:24

Diagnostics Check		OK	01:08	
Air Blank		0.000	01:09	
Control Test		0.079	01:09	
Air Blank		0.000	01:09	
Subject Sample	# 1		01:10	Subject Test Refused
Breath Volume		0.000	01:10	
Air Blank		0.000	01:10	
Control Test		0.079	01:10	
Air Blank		0.000	01:11	
Diagnostics Check		OK	01:11	

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CANELLIZ

DATE/TIME OF TEST: 03/18/2019 01:46:42

Diagnostics Check		OK	01:43	
Air Blank		0.000	01:44	
Control Test		0.080	01:44	
Air Blank		0.000	01:44	
Subject Sample	# 1		01:45	Subject Test Refused
Breath Volume		0.000	01:45	
Air Blank		0.000	01:45	
Control Test		0.079	01:46	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 03/22/2019 03:15:47

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:10	
Control Test		0.080	03:11	
Air Blank		0.000	03:11	
Subject Sample	# 1		03:14	Subject Test Refused
Breath Volume		0.496	03:14	
Air Blank		0.000	03:14	
Control Test		0.080	03:15	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:15	

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 03/22/2019 03:23:33

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:18	
Control Test		0.079	03:19	
Air Blank		0.000	03:19	
Subject Sample	# 1		03:21	Subject Test Refused
Breath Volume		0.160	03:21	
Air Blank		0.000	03:22	
Control Test		0.079	03:22	
Air Blank		0.000	03:23	
Diagnostics Check		OK	03:23	

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: JUNCO HORMIGO

DATE/TIME OF TEST: 03/22/2019 04:04:23

Diagnostics Check		OK	03:55
Air Blank		0.000	03:55
Control Test		0.079	03:56
Air Blank		0.000	03:56
Subject Sample	# 1	0.154	03:59
Breath Volume		1.238	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.151	04:02
Breath Volume		2.113	04:02
Air Blank		0.000	04:03
Control Test		0.078	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BETANCUR

DATE/TIME OF TEST: 03/23/2019 01:41:06

Diagnostics Check		OK	01:37	
Air Blank		0.000	01:37	
Control Test		0.080	01:38	
Air Blank		0.000	01:38	
Subject Sample	# 1		01:39	Subject Test Refused
Breath Volume		0.000	01:39	
Air Blank		0.000	01:40	
Control Test		0.080	01:40	
Air Blank		0.000	01:41	
Diagnostics Check		OK	01:41	

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 03/23/2019 03:42:20

Diagnostics Check		OK	03:39	
Air Blank		0.000	03:39	
Control Test		0.080	03:40	
Air Blank		0.000	03:40	
Subject Sample	# 1		03:40	Subject Test Refused
Breath Volume		0.000	03:40	
Air Blank		0.000	03:41	
Control Test		0.080	03:41	
Air Blank		0.000	03:42	
Diagnostics Check		OK	03:42	

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: FERNANDEZ,ORLANDO
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: ORDONEZ

DATE/TIME OF TEST: 03/23/2019 21:20:48

Diagnostics Check		OK	21:14	
Air Blank		0.000	21:15	
Control Test		0.080	21:15	
Air Blank		0.000	21:15	
Subject Sample	# 1		21:19	Subject Test Refused
Breath Volume		0.101	21:19	
Air Blank		0.000	21:19	
Control Test		0.079	21:20	
Air Blank		0.000	21:20	
Diagnostics Check		OK	21:20	

Data Download Date/Time: 04/18/2019 09:21

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUBELA,CHANTEL,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: CUBELA,CHANTEL
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BETANCUR BASULTO

DATE/TIME OF TEST: 03/27/2019 04:08:11

Diagnostics Check		OK	04:05	
Air Blank		0.000	04:05	
Control Test		0.080	04:06	
Air Blank		0.000	04:06	
Subject Sample	# 1		04:06	Subject Test Refused
Breath Volume		0.000	04:06	
Air Blank		0.000	04:07	
Control Test		0.081	04:07	
Air Blank		0.000	04:08	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 04/18/2019 09:22

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUBELA,CHANTEL,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: CUBELA,CHANTEL
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: BETANCUR BASULTO

DATE/TIME OF TEST: 03/27/2019 04:28:50

Diagnostics Check		OK	04:25	
Air Blank		0.000	04:26	
Control Test		0.081	04:26	
Air Blank		0.000	04:27	
Subject Sample	# 1		04:27	Subject Test Refused
Breath Volume		0.000	04:27	
Air Blank		0.000	04:27	
Control Test		0.080	04:28	
Air Blank		0.000	04:28	
Diagnostics Check		OK	04:28	

Data Download Date/Time: 04/18/2019 09:22

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,JAVIER,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: DIAZ,JAVIER
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:38

SUBJECT'S LAST NAME: CORTINA

DATE/TIME OF TEST: 03/30/2019 08:01:33

Diagnostics Check		OK	07:58	
Air Blank		0.000	07:59	
Control Test		0.080	07:59	
Air Blank		0.000	07:59	
Subject Sample	# 1		08:00	Subject Test Refused
Breath Volume		0.000	08:00	
Air Blank		0.000	08:00	
Control Test		0.080	08:00	
Air Blank		0.000	08:01	
Diagnostics Check		OK	08:01	

Data Download Date/Time: 04/18/2019 09:22

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: FARNSTROM

DATE/TIME OF TEST: 03/31/2019 05:22:31

Diagnostics Check		OK	05:10	
Air Blank		0.000	05:11	
Control Test		0.079	05:11	
Air Blank		0.000	05:12	
Subject Sample	# 1	0.210	05:15	Volume Not Met
Breath Volume		0.332	05:15	
Air Blank		0.000	05:16	
Air Blank		0.000	05:17	
Subject Sample	# 2	0.227	05:20	Volume Not Met
Breath Volume		0.078	05:20	
Air Blank		0.000	05:21	
Control Test		0.078	05:21	
Air Blank		0.000	05:22	
Diagnostics Check		OK	05:22	

Data Download Date/Time: 04/18/2019 09:22

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDO,RAAJU,S
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: SIDDO,RAAJU
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: FARNSTROM

DATE/TIME OF TEST: 03/31/2019 05:28:54

Diagnostics Check		OK	05:25	
Air Blank		0.000	05:26	
Control Test		0.078	05:26	
Air Blank		0.000	05:27	
Subject Sample	# 1		05:27	Subject Test Refused
Breath Volume		0.000	05:27	
Air Blank		0.000	05:27	
Control Test		0.079	05:28	
Air Blank		0.000	05:28	
Diagnostics Check		OK	05:28	

Data Download Date/Time: 04/18/2019 09:22

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: KHNAIZIR

DATE/TIME OF TEST: 03/01/2019 05:09:03

Diagnostics Check		OK	05:02
Air Blank		0.000	05:02
Control Test		0.081	05:03
Air Blank		0.000	05:03
Subject Sample	# 1	0.140	05:04
Breath Volume		1.503	05:04
Air Blank		0.000	05:04
Air Blank		0.000	05:06
Subject Sample	# 2	0.142	05:07
Breath Volume		1.187	05:07
Air Blank		0.000	05:08
Control Test		0.080	05:08
Air Blank		0.000	05:08
Diagnostics Check		OK	05:09

Data Download Date/Time: 03/04/2019 13:18

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,A,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 03/01/2019 21:17:49

Diagnostics Check		OK	21:11
Air Blank		0.000	21:11
Control Test		0.080	21:11
Air Blank		0.000	21:12
Subject Sample	# 1	0.110	21:13
Breath Volume		1.804	21:13
Air Blank		0.000	21:14
Air Blank		0.000	21:15
Subject Sample	# 2	0.106	21:16
Breath Volume		1.953	21:16
Air Blank		0.000	21:16
Control Test		0.080	21:17
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

Data Download Date/Time: 03/04/2019 13:18

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DANIELS,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: SANTOS-RODRIGUEZ

DATE/TIME OF TEST: 03/02/2019 18:32:25

Diagnostics Check		OK	18:25
Air Blank		0.000	18:25
Control Test		0.080	18:26
Air Blank		0.000	18:26
Subject Sample	# 1	0.174	18:27
Breath Volume		2.285	18:27
Air Blank		0.000	18:28
Air Blank		0.000	18:30
Subject Sample	# 2	0.167	18:30
Breath Volume		1.761	18:30
Air Blank		0.000	18:31
Control Test		0.079	18:31
Air Blank		0.000	18:32
Diagnostics Check		OK	18:32

Data Download Date/Time: 03/04/2019 13:18

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 03/03/2019 22:16:13

Diagnostics Check		OK	22:13	
Air Blank		0.000	22:13	
Control Test		0.080	22:13	
Air Blank		0.000	22:14	
Subject Sample	# 1		22:14	Subject Test Refused
Breath Volume		0.000	22:14	
Air Blank		0.000	22:15	
Control Test		0.080	22:15	
Air Blank		0.000	22:16	
Diagnostics Check		OK	22:16	

Data Download Date/Time: 03/04/2019 13:18

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SNYDER,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/04/2019 16:24:09

Diagnostics Check		OK	16:15
Air Blank		0.000	16:16
Control Test		0.081	16:16
Air Blank		0.000	16:17
Subject Sample	# 1	0.161	16:19
Breath Volume		1.140	16:19
Air Blank		0.000	16:19
Air Blank		0.000	16:21
Subject Sample	# 2	0.173	16:22
Breath Volume		1.152	16:22
Air Blank		0.000	16:23
Control Test		0.080	16:23
Air Blank		0.000	16:24
Diagnostics Check		OK	16:24

Data Download Date/Time: 03/15/2019 11:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 03/04/2019 21:02:09

Diagnostics Check		OK	20:59	
Air Blank		0.000	20:59	
Control Test		0.080	21:00	
Air Blank		0.000	21:00	
Subject Sample	# 1		21:00	Subject Test Refused
Breath Volume		0.000	21:00	
Air Blank		0.000	21:01	
Control Test		0.081	21:01	
Air Blank		0.000	21:02	
Diagnostics Check		OK	21:02	

Data Download Date/Time: 03/15/2019 11:50

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,A,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: BARELA

DATE/TIME OF TEST: 03/05/2019 19:17:07

Diagnostics Check		OK	19:14	
Air Blank		0.000	19:14	
Control Test		0.082	19:15	
Air Blank		0.000	19:15	
Subject Sample	# 1		19:15	Subject Test Refused
Breath Volume		0.000	19:15	
Air Blank		0.000	19:16	
Control Test		0.081	19:16	
Air Blank		0.000	19:17	
Diagnostics Check		OK	19:17	

Data Download Date/Time: 03/15/2019 11:50

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 03/06/2019 15:57:04

Diagnostics Check		OK	15:51
Air Blank		0.000	15:51
Control Test		0.082	15:51
Air Blank		0.000	15:52
Subject Sample	# 1	0.000	15:52
Breath Volume		3.042	15:52
Air Blank		0.000	15:53
Air Blank		0.000	15:55
Subject Sample	# 2	0.000	15:55
Breath Volume		1.910	15:55
Air Blank		0.000	15:56
Control Test		0.082	15:56
Air Blank		0.000	15:56
Diagnostics Check		OK	15:57

Data Download Date/Time: 03/15/2019 11:50

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARDWELL,M
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: DRILL

DATE/TIME OF TEST: 03/06/2019 18:47:21

Diagnostics Check		OK	18:40
Air Blank		0.000	18:41
Control Test		0.082	18:41
Air Blank		0.000	18:41
Subject Sample	# 1	0.090	18:42
Breath Volume		2.449	18:42
Air Blank		0.000	18:43
Air Blank		0.000	18:44
Subject Sample	# 2	0.089	18:45
Breath Volume		1.535	18:45
Air Blank		0.000	18:46
Control Test		0.082	18:46
Air Blank		0.000	18:47
Diagnostics Check		OK	18:47

Data Download Date/Time: 03/15/2019 11:50

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: CREWS

DATE/TIME OF TEST: 03/06/2019 20:51:23

Diagnostics Check		OK	20:45
Air Blank		0.000	20:45
Control Test		0.081	20:45
Air Blank		0.000	20:46
Subject Sample	# 1	0.095	20:46
Breath Volume		1.480	20:46
Air Blank		0.000	20:47
Air Blank		0.000	20:49
Subject Sample	# 2	0.096	20:49
Breath Volume		2.164	20:49
Air Blank		0.000	20:50
Control Test		0.081	20:50
Air Blank		0.000	20:51
Diagnostics Check		OK	20:51

Data Download Date/Time: 03/15/2019 11:50

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: AL-DAHARAA

DATE/TIME OF TEST: 03/06/2019 22:11:52

Diagnostics Check		OK	22:05
Air Blank		0.000	22:05
Control Test		0.081	22:06
Air Blank		0.000	22:06
Subject Sample	# 1	0.156	22:07
Breath Volume		2.207	22:07
Air Blank		0.000	22:08
Air Blank		0.000	22:09
Subject Sample	# 2	0.142	22:10
Breath Volume		1.785	22:10
Air Blank		0.000	22:10
Control Test		0.080	22:11
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 03/15/2019 11:50

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,M,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 03/09/2019 19:21:35

Diagnostics Check		OK	19:18	
Air Blank		0.000	19:19	
Control Test		0.081	19:19	
Air Blank		0.000	19:19	
Subject Sample	# 1		19:20	Subject Test Refused
Breath Volume		0.000	19:20	
Air Blank		0.000	19:20	
Control Test		0.080	19:21	
Air Blank		0.000	19:21	
Diagnostics Check		OK	19:21	

Data Download Date/Time: 03/15/2019 11:50

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,M,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: LEMUS

DATE/TIME OF TEST: 03/09/2019 22:25:20

Diagnostics Check		OK	22:22	
Air Blank		0.000	22:23	
Control Test		0.080	22:23	
Air Blank		0.000	22:23	
Subject Sample	# 1		22:23	Subject Test Refused
Breath Volume		0.000	22:23	
Air Blank		0.000	22:24	
Control Test		0.080	22:24	
Air Blank		0.000	22:25	
Diagnostics Check		OK	22:25	

Data Download Date/Time: 03/15/2019 11:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 03/10/2019 18:27:46

Diagnostics Check		OK	18:21
Air Blank		0.000	18:21
Control Test		0.080	18:22
Air Blank		0.000	18:22
Subject Sample	# 1	0.211	18:23
Breath Volume		3.843	18:23
Air Blank		0.000	18:23
Air Blank		0.000	18:25
Subject Sample	# 2	0.206	18:26
Breath Volume		3.785	18:26
Air Blank		0.000	18:26
Control Test		0.080	18:27
Air Blank		0.000	18:27
Diagnostics Check		OK	18:27

Data Download Date/Time: 03/15/2019 11:50

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: GOLOVIN

DATE/TIME OF TEST: 03/11/2019 21:07:36

Diagnostics Check		OK	21:01
Air Blank		0.000	21:01
Control Test		0.078	21:01
Air Blank		0.000	21:02
Subject Sample	# 1	0.051	21:03
Breath Volume		3.828	21:03
Air Blank		0.000	21:03
Air Blank		0.000	21:05
Subject Sample	# 2	0.050	21:06
Breath Volume		3.753	21:06
Air Blank		0.000	21:06
Control Test		0.077	21:07
Air Blank		0.000	21:07
Diagnostics Check		OK	21:07

Data Download Date/Time: 03/15/2019 11:50

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMAS,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: DYKES

DATE/TIME OF TEST: 03/13/2019 05:27:19

Diagnostics Check		OK	05:20
Air Blank		0.000	05:21
Control Test		0.080	05:21
Air Blank		0.000	05:22
Subject Sample	# 1	0.118	05:22
Breath Volume		2.152	05:22
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.119	05:25
Breath Volume		1.769	05:25
Air Blank		0.000	05:26
Control Test		0.080	05:26
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 03/15/2019 11:50

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONROE,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 03/17/2019 00:01:11

Diagnostics Check		OK	23:58	
Air Blank		0.000	23:58	
Control Test		0.081	23:59	
Air Blank		0.000	23:59	
Subject Sample	# 1		23:59	Subject Test Refused
Breath Volume		0.000	23:59	
Air Blank		0.000	00:00	
Control Test		0.080	00:00	
Air Blank		0.000	00:01	
Diagnostics Check		OK	00:01	

Data Download Date/Time: 04/16/2019 15:19

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DRAGOMIR RAMIREZ,I
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: STRINGER

DATE/TIME OF TEST: 03/17/2019 03:16:30

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.081	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.165	03:12
Breath Volume		3.148	03:12
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.165	03:15
Breath Volume		3.218	03:15
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 04/16/2019 15:19

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,A,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: LOVERIDGE

DATE/TIME OF TEST: 03/17/2019 20:28:41

Diagnostics Check		OK	20:22
Air Blank		0.000	20:22
Control Test		0.082	20:22
Air Blank		0.000	20:23
Subject Sample	# 1	0.190	20:23
Breath Volume		2.253	20:23
Air Blank		0.000	20:24
Air Blank		0.000	20:26
Subject Sample	# 2	0.196	20:27
Breath Volume		2.960	20:27
Air Blank		0.000	20:27
Control Test		0.080	20:28
Air Blank		0.000	20:28
Diagnostics Check		OK	20:28

Data Download Date/Time: 04/16/2019 15:19

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,A,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: CASADY

DATE/TIME OF TEST: 03/17/2019 21:59:09

Diagnostics Check		OK	21:52
Air Blank		0.000	21:53
Control Test		0.080	21:53
Air Blank		0.000	21:54
Subject Sample	# 1	0.220	21:54
Breath Volume		2.722	21:54
Air Blank		0.000	21:55
Air Blank		0.000	21:57
Subject Sample	# 2	0.224	21:57
Breath Volume		3.734	21:57
Air Blank		0.000	21:58
Control Test		0.080	21:58
Air Blank		0.000	21:59
Diagnostics Check		OK	21:59

Data Download Date/Time: 04/16/2019 15:19

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALDREP,B
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: WATERS

DATE/TIME OF TEST: 03/24/2019 02:53:53

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:51	
Control Test		0.082	02:51	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:52	Subject Test Refused
Breath Volume		0.000	02:52	
Air Blank		0.000	02:53	
Control Test		0.081	02:53	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 04/16/2019 15:19

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARMENTER,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: OLSCHEWSKE

DATE/TIME OF TEST: 03/01/2019 04:48:51

Diagnostics Check		OK	04:42
Air Blank		0.000	04:43
Control Test		0.081	04:43
Air Blank		0.000	04:44
Subject Sample	# 1	0.120	04:44
Breath Volume		1.804	04:44
Air Blank		0.000	04:45
Air Blank		0.000	04:46
Subject Sample	# 2	0.118	04:47
Breath Volume		1.671	04:47
Air Blank		0.000	04:47
Control Test		0.081	04:48
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 03/04/2019 13:26

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/01/2019 06:12:26

Diagnostics Check		OK	06:06
Air Blank		0.000	06:06
Control Test		0.081	06:07
Air Blank		0.000	06:07
Subject Sample	# 1	0.158	06:08
Breath Volume		3.089	06:08
Air Blank		0.000	06:08
Air Blank		0.000	06:10
Subject Sample	# 2	0.150	06:10
Breath Volume		2.308	06:10
Air Blank		0.000	06:11
Control Test		0.079	06:11
Air Blank		0.000	06:12
Diagnostics Check		OK	06:12

Data Download Date/Time: 03/04/2019 13:26

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: SMART

DATE/TIME OF TEST: 03/02/2019 00:29:25

Diagnostics Check		OK	00:23
Air Blank		0.000	00:23
Control Test		0.080	00:24
Air Blank		0.000	00:24
Subject Sample	# 1	0.000	00:25
Breath Volume		2.890	00:25
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.000	00:28
Breath Volume		2.324	00:28
Air Blank		0.000	00:28
Control Test		0.081	00:28
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 03/04/2019 13:26

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HICKSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: MASTERS

DATE/TIME OF TEST: 03/02/2019 02:32:37

Diagnostics Check		OK	02:29	
Air Blank		0.000	02:30	
Control Test		0.081	02:30	
Air Blank		0.000	02:30	
Subject Sample	# 1		02:31	Subject Test Refused
Breath Volume		0.000	02:31	
Air Blank		0.000	02:31	
Control Test		0.080	02:32	
Air Blank		0.000	02:32	
Diagnostics Check		OK	02:32	

Data Download Date/Time: 03/04/2019 13:26

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/02/2019 04:30:21

Diagnostics Check		OK	04:27	
Air Blank		0.000	04:27	
Control Test		0.081	04:28	
Air Blank		0.000	04:28	
Subject Sample	# 1		04:28	Subject Test Refused
Breath Volume		0.000	04:28	
Air Blank		0.000	04:29	
Control Test		0.081	04:29	
Air Blank		0.000	04:30	
Diagnostics Check		OK	04:30	

Data Download Date/Time: 03/04/2019 13:26

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,F,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: ABDULLAH

DATE/TIME OF TEST: 03/02/2019 07:28:26

Diagnostics Check		OK	07:21
Air Blank		0.000	07:22
Control Test		0.081	07:22
Air Blank		0.000	07:23
Subject Sample	# 1	0.229	07:23
Breath Volume		2.281	07:23
Air Blank		0.000	07:24
Air Blank		0.000	07:26
Subject Sample	# 2	0.234	07:26
Breath Volume		2.250	07:26
Air Blank		0.000	07:27
Control Test		0.080	07:27
Air Blank		0.000	07:28
Diagnostics Check		OK	07:28

Data Download Date/Time: 03/04/2019 13:26

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,F,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KELLY,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:47

SUBJECT'S LAST NAME: FREDRICK

DATE/TIME OF TEST: 03/02/2019 09:14:34

Diagnostics Check		OK	09:07
Air Blank		0.000	09:08
Control Test		0.081	09:08
Air Blank		0.000	09:09
Subject Sample	# 1	0.120	09:10
Breath Volume		2.617	09:10
Air Blank		0.000	09:10
Air Blank		0.000	09:12
Subject Sample	# 2	0.121	09:13
Breath Volume		2.886	09:13
Air Blank		0.000	09:13
Control Test		0.080	09:13
Air Blank		0.000	09:14
Diagnostics Check		OK	09:14

Data Download Date/Time: 03/04/2019 13:27

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: GARZA

DATE/TIME OF TEST: 03/02/2019 23:38:04

Diagnostics Check		OK	23:31
Air Blank		0.000	23:32
Control Test		0.080	23:32
Air Blank		0.000	23:33
Subject Sample	# 1	0.214	23:33
Breath Volume		3.320	23:33
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.207	23:36
Breath Volume		2.992	23:36
Air Blank		0.000	23:37
Control Test		0.080	23:37
Air Blank		0.000	23:37
Diagnostics Check		OK	23:38

Data Download Date/Time: 03/04/2019 13:27

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 03/03/2019 02:23:47

Diagnostics Check		OK	02:20	
Air Blank		0.000	02:21	
Control Test		0.081	02:21	
Air Blank		0.000	02:22	
Subject Sample	# 1		02:22	Subject Test Refused
Breath Volume		0.000	02:22	
Air Blank		0.000	02:22	
Control Test		0.081	02:23	
Air Blank		0.000	02:23	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 03/04/2019 13:27

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: PROM

DATE/TIME OF TEST: 03/03/2019 02:59:52

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:57	
Control Test		0.080	02:57	
Air Blank		0.000	02:58	
Subject Sample	# 1		02:58	Subject Test Refused
Breath Volume		0.000	02:58	
Air Blank		0.000	02:58	
Control Test		0.080	02:59	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 03/04/2019 13:27

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MCFARLAND,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: PECK

DATE/TIME OF TEST: 03/03/2019 05:10:10

Diagnostics Check		OK	05:04
Air Blank		0.000	05:04
Control Test		0.080	05:04
Air Blank		0.000	05:05
Subject Sample	# 1	0.183	05:05
Breath Volume		3.730	05:05
Air Blank		0.000	05:06
Air Blank		0.000	05:08
Subject Sample	# 2	0.183	05:08
Breath Volume		2.449	05:08
Air Blank		0.000	05:09
Control Test		0.080	05:09
Air Blank		0.000	05:10
Diagnostics Check		OK	05:10

Data Download Date/Time: 03/04/2019 13:27

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: LOCKWOOD

DATE/TIME OF TEST: 03/03/2019 23:43:59

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.080	23:39
Air Blank		0.000	23:39
Subject Sample	# 1	0.000	23:39
Breath Volume		1.738	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.000	23:42
Breath Volume		1.632	23:42
Air Blank		0.000	23:43
Control Test		0.080	23:43
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 03/04/2019 13:27

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MALLOY

DATE/TIME OF TEST: 03/04/2019 02:41:51

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.080	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.109	02:37
Breath Volume		1.664	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:39
Subject Sample	# 2	0.100	02:40
Breath Volume		1.535	02:40
Air Blank		0.000	02:40
Control Test		0.081	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 03/04/2019 13:27

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: ADKINS

DATE/TIME OF TEST: 03/05/2019 00:52:03

Diagnostics Check		OK	00:46
Air Blank		0.000	00:46
Control Test		0.081	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.000	00:47
Breath Volume		2.726	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.000	00:50
Breath Volume		2.914	00:50
Air Blank		0.000	00:51
Control Test		0.081	00:51
Air Blank		0.000	00:51
Diagnostics Check		OK	00:52

Data Download Date/Time: 03/15/2019 11:58

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BUMGARNER,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 03/05/2019 01:46:34

Diagnostics Check		OK	01:35	
Air Blank		0.000	01:35	
Control Test		0.080	01:35	
Air Blank		0.000	01:36	
Subject Sample	# 1		01:39	No Sample Provided
Breath Volume		0.000	01:39	
Air Blank		0.000	01:40	
Air Blank		0.000	01:42	
Subject Sample	# 2		01:45	No Sample Provided
Breath Volume		0.000	01:45	
Air Blank		0.000	01:45	
Control Test		0.081	01:45	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 03/15/2019 11:58

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/05/2019 03:55:29

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:53	
Control Test		0.080	03:53	
Air Blank		0.000	03:53	
Subject Sample	# 1		03:54	Subject Test Refused
Breath Volume		0.000	03:54	
Air Blank		0.000	03:54	
Control Test		0.081	03:54	
Air Blank		0.000	03:55	
Diagnostics Check		OK	03:55	

Data Download Date/Time: 03/15/2019 11:58

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: LYDAY

DATE/TIME OF TEST: 03/07/2019 01:35:08

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:32	
Control Test		0.081	01:33	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:33	Subject Test Refused
Breath Volume		0.000	01:33	
Air Blank		0.000	01:34	
Control Test		0.081	01:34	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 03/15/2019 11:58

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,SAVANNAH,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: RUTLEDGE

DATE/TIME OF TEST: 03/07/2019 17:42:16

Diagnostics Check		OK	17:35
Air Blank		0.000	17:36
Control Test		0.081	17:36
Air Blank		0.000	17:37
Subject Sample	# 1	0.000	17:37
Breath Volume		1.433	17:37
Air Blank		0.000	17:38
Air Blank		0.000	17:40
Subject Sample	# 2	0.000	17:40
Breath Volume		2.972	17:40
Air Blank		0.000	17:41
Control Test		0.081	17:41
Air Blank		0.000	17:42
Diagnostics Check		OK	17:42

Data Download Date/Time: 03/15/2019 11:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,SAVANNAH,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/07/2019 18:56:29

Diagnostics Check		OK	18:50
Air Blank		0.000	18:50
Control Test		0.082	18:51
Air Blank		0.000	18:51
Subject Sample	# 1	0.149	18:52
Breath Volume		1.832	18:52
Air Blank		0.000	18:52
Air Blank		0.000	18:54
Subject Sample	# 2	0.147	18:54
Breath Volume		3.324	18:54
Air Blank		0.000	18:55
Control Test		0.081	18:55
Air Blank		0.000	18:56
Diagnostics Check		OK	18:56

Data Download Date/Time: 03/15/2019 11:58

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,SAVANNAH,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: HALMAJAN

DATE/TIME OF TEST: 03/07/2019 22:07:33

Diagnostics Check		OK	21:59
Air Blank		0.000	22:00
Control Test		0.082	22:00
Air Blank		0.000	22:00
Subject Sample	# 1	0.106	22:03
Breath Volume		2.261	22:03
Air Blank		0.000	22:03
Air Blank		0.000	22:05
Subject Sample	# 2	0.106	22:06
Breath Volume		2.386	22:06
Air Blank		0.000	22:06
Control Test		0.080	22:06
Air Blank		0.000	22:07
Diagnostics Check		OK	22:07

Data Download Date/Time: 03/15/2019 11:58

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARPER,T
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CALLAHAN

DATE/TIME OF TEST: 03/08/2019 02:26:55

Diagnostics Check		OK	02:24	
Air Blank		0.000	02:24	
Control Test		0.081	02:24	
Air Blank		0.000	02:25	
Subject Sample	# 1		02:25	Subject Test Refused
Breath Volume		0.000	02:25	
Air Blank		0.000	02:26	
Control Test		0.082	02:26	
Air Blank		0.000	02:26	
Diagnostics Check		OK	02:26	

Data Download Date/Time: 03/15/2019 11:59

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: BEARD

DATE/TIME OF TEST: 03/08/2019 03:07:10

Diagnostics Check		OK	03:04	
Air Blank		0.000	03:04	
Control Test		0.081	03:05	
Air Blank		0.000	03:05	
Subject Sample	# 1		03:05	Subject Test Refused
Breath Volume		0.000	03:05	
Air Blank		0.000	03:06	
Control Test		0.081	03:06	
Air Blank		0.000	03:07	
Diagnostics Check		OK	03:07	

Data Download Date/Time: 03/15/2019 11:59

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MUNGER,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: HILL-TRULL

DATE/TIME OF TEST: 03/08/2019 04:56:21

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.081	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.208	04:51
Breath Volume		3.777	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.204	04:54
Breath Volume		3.105	04:54
Air Blank		0.000	04:55
Control Test		0.080	04:55
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 03/15/2019 11:59

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUTZE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:43

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 03/08/2019 06:06:02

Diagnostics Check		OK	06:03	
Air Blank		0.000	06:03	
Control Test		0.081	06:03	
Air Blank		0.000	06:04	
Subject Sample	# 1		06:04	Subject Test Refused
Breath Volume		0.000	06:04	
Air Blank		0.000	06:05	
Control Test		0.082	06:05	
Air Blank		0.000	06:05	
Diagnostics Check		OK	06:06	

Data Download Date/Time: 03/15/2019 11:59

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASS,H,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: GARCIA MORALES

DATE/TIME OF TEST: 03/09/2019 22:54:10

Diagnostics Check		OK	22:47	
Air Blank		0.000	22:48	
Control Test		0.081	22:48	
Air Blank		0.000	22:49	
Subject Sample	# 1	0.164	22:49	
Breath Volume		1.492	22:49	
Air Blank		0.000	22:50	
Air Blank		0.000	22:51	
Subject Sample	# 2	0.170	22:52	
Breath Volume		2.230	22:52	
Air Blank		0.000	22:52	
Control Test		0.080	22:53	
Air Blank			22:53	RFI Detect
Air Blank		0.000	22:54	

Data Download Date/Time: 03/15/2019 11:59

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASS,H,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: GARCIA MORALES

DATE/TIME OF TEST: 03/09/2019 23:20:05

Diagnostics Check		OK	23:13
Air Blank		0.000	23:14
Control Test		0.081	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.153	23:15
Breath Volume		2.332	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.149	23:18
Breath Volume		1.765	23:18
Air Blank		0.000	23:19
Control Test		0.081	23:19
Air Blank		0.000	23:19
Diagnostics Check		OK	23:20

Data Download Date/Time: 03/15/2019 11:59

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: GUEST

DATE/TIME OF TEST: 03/12/2019 22:07:32

Diagnostics Check		OK	22:00
Air Blank		0.000	22:00
Control Test		0.081	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.154	22:02
Breath Volume		1.699	22:02
Air Blank		0.000	22:03
Air Blank		0.000	22:05
Subject Sample	# 2	0.154	22:06
Breath Volume		2.542	22:06
Air Blank		0.000	22:06
Control Test		0.081	22:06
Air Blank		0.000	22:07
Diagnostics Check		OK	22:07

Data Download Date/Time: 03/15/2019 11:59

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,I
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:49

SUBJECT'S LAST NAME: PEAVEY

DATE/TIME OF TEST: 03/13/2019 15:14:31

Diagnostics Check		OK	15:11	
Air Blank		0.000	15:12	
Control Test		0.081	15:12	
Air Blank		0.000	15:12	
Subject Sample	# 1		15:13	Subject Test Refused
Breath Volume		0.000	15:13	
Air Blank		0.000	15:13	
Control Test		0.080	15:13	
Air Blank		0.000	15:14	
Diagnostics Check		OK	15:14	

Data Download Date/Time: 03/15/2019 11:59

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/13/2019 19:17:50

Diagnostics Check		OK	19:14	
Air Blank		0.000	19:15	
Control Test		0.081	19:15	
Air Blank		0.000	19:16	
Subject Sample	# 1		19:16	Subject Test Refused
Breath Volume		0.000	19:16	
Air Blank		0.000	19:16	
Control Test		0.080	19:17	
Air Blank		0.000	19:17	
Diagnostics Check		OK	19:17	

Data Download Date/Time: 03/15/2019 11:59

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: BERGMAN

DATE/TIME OF TEST: 03/14/2019 19:03:44

Diagnostics Check		OK	19:00	
Air Blank		0.000	19:01	
Control Test		0.080	19:01	
Air Blank		0.000	19:02	
Subject Sample	# 1		19:02	Subject Test Refused
Breath Volume		0.000	19:02	
Air Blank		0.000	19:02	
Control Test		0.080	19:03	
Air Blank		0.000	19:03	
Diagnostics Check		OK	19:03	

Data Download Date/Time: 03/15/2019 11:59

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TIRRI,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: TRIFONOVA

DATE/TIME OF TEST: 03/15/2019 13:47:08

Diagnostics Check		OK	13:40
Air Blank		0.000	13:41
Control Test		0.080	13:41
Air Blank		0.000	13:42
Subject Sample	# 1	0.000	13:42
Breath Volume		2.199	13:42
Air Blank		0.000	13:43
Air Blank		0.000	13:45
Subject Sample	# 2	0.000	13:45
Breath Volume		1.804	13:45
Air Blank		0.000	13:46
Control Test		0.080	13:46
Air Blank		0.000	13:47
Diagnostics Check		OK	13:47

Data Download Date/Time: 04/16/2019 15:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: MADDOX

DATE/TIME OF TEST: 03/15/2019 22:32:23

Diagnostics Check		OK	22:28	
Air Blank		0.000	22:29	
Control Test		0.081	22:29	
Air Blank		0.000	22:30	
Subject Sample	# 1		22:31	Subject Test Refused
Breath Volume		0.000	22:31	
Air Blank		0.000	22:31	
Control Test		0.081	22:31	
Air Blank		0.000	22:32	
Diagnostics Check		OK	22:32	

Data Download Date/Time: 04/16/2019 15:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: WEYAND

DATE/TIME OF TEST: 03/17/2019 23:18:43

Diagnostics Check		OK	23:15	
Air Blank		0.000	23:16	
Control Test		0.081	23:16	
Air Blank		0.000	23:16	
Subject Sample	# 1		23:17	Subject Test Refused
Breath Volume		0.000	23:17	
Air Blank		0.000	23:17	
Control Test		0.082	23:18	
Air Blank		0.000	23:18	
Diagnostics Check		OK	23:18	

Data Download Date/Time: 04/16/2019 15:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: CONGER

DATE/TIME OF TEST: 03/18/2019 02:48:29

Diagnostics Check		OK	02:45	
Air Blank		0.000	02:46	
Control Test		0.082	02:46	
Air Blank		0.000	02:46	
Subject Sample	# 1		02:47	Subject Test Refused
Breath Volume		0.000	02:47	
Air Blank		0.000	02:47	
Control Test		0.081	02:47	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:48	

Data Download Date/Time: 04/16/2019 15:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SOSA,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: LOPEZ NOLASCO

DATE/TIME OF TEST: 03/18/2019 05:47:35

Diagnostics Check		OK	05:44	
Air Blank		0.000	05:45	
Control Test		0.081	05:45	
Air Blank		0.000	05:46	
Subject Sample	# 1		05:46	Subject Test Refused
Breath Volume		0.000	05:46	
Air Blank		0.000	05:46	
Control Test		0.081	05:47	
Air Blank		0.000	05:47	
Diagnostics Check		OK	05:47	

Data Download Date/Time: 04/16/2019 15:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOUMARD,MARCEL,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J.R.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: BRIDGES

DATE/TIME OF TEST: 03/18/2019 21:04:43

Diagnostics Check		OK	20:58
Air Blank		0.000	20:58
Control Test		0.081	20:58
Air Blank		0.000	20:59
Subject Sample	# 1	0.159	20:59
Breath Volume		3.589	20:59
Air Blank		0.000	21:00
Air Blank		0.000	21:02
Subject Sample	# 2	0.155	21:03
Breath Volume		3.117	21:03
Air Blank		0.000	21:03
Control Test		0.081	21:04
Air Blank		0.000	21:04
Diagnostics Check		OK	21:04

Data Download Date/Time: 04/16/2019 15:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASS,H,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: ABEL

DATE/TIME OF TEST: 03/22/2019 00:00:40

Diagnostics Check		OK	23:57	
Air Blank		0.000	23:58	
Control Test		0.081	23:58	
Air Blank		0.000	23:59	
Subject Sample	# 1		23:59	Subject Test Refused
Breath Volume		0.000	23:59	
Air Blank		0.000	23:59	
Control Test		0.081	00:00	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:00	

Data Download Date/Time: 04/16/2019 15:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASS,H,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEAN,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: TRIPLETT

DATE/TIME OF TEST: 03/22/2019 03:26:09

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:23	
Control Test		0.081	03:24	
Air Blank		0.000	03:24	
Subject Sample	# 1		03:24	Subject Test Refused
Breath Volume		0.000	03:24	
Air Blank		0.000	03:25	
Control Test		0.082	03:25	
Air Blank		0.000	03:26	
Diagnostics Check		OK	03:26	

Data Download Date/Time: 04/16/2019 15:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASS,H,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/23/2019 03:36:24

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.081	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.110	03:29
Breath Volume		1.117	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.103	03:34
Breath Volume		1.187	03:34
Air Blank		0.000	03:35
Control Test		0.081	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 04/16/2019 15:31

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASS,H,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALDREP,B
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: DOZIER

DATE/TIME OF TEST: 03/23/2019 04:33:17

Diagnostics Check		OK	04:21	
Air Blank		0.000	04:21	
Control Test		0.081	04:21	
Air Blank		0.000	04:22	
Subject Sample	# 1	0.100	04:25	Volume Not Met
Breath Volume		0.199	04:25	
Air Blank		0.000	04:26	
Air Blank		0.000	04:28	
Subject Sample	# 2	0.096	04:28	
Breath Volume		1.710	04:28	
Air Blank		0.000	04:29	
Air Blank		0.000	04:31	
Subject Sample	# 3	0.083	04:31	
Breath Volume		2.429	04:31	
Air Blank		0.000	04:32	
Control Test		0.081	04:32	
Air Blank		0.000	04:33	
Diagnostics Check		OK	04:33	

Data Download Date/Time: 04/16/2019 15:31

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASS,H,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: VO

DATE/TIME OF TEST: 03/26/2019 01:00:09

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.081	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.046	00:54
Breath Volume		1.148	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.044	00:58
Breath Volume		1.183	00:58
Air Blank		0.000	00:59
Control Test		0.081	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 04/16/2019 15:31

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: HOOKS

DATE/TIME OF TEST: 03/30/2019 05:27:25

Diagnostics Check		OK	05:20
Air Blank		0.000	05:20
Control Test		0.080	05:20
Air Blank		0.000	05:21
Subject Sample	# 1	0.206	05:22
Breath Volume		3.460	05:22
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.196	05:25
Breath Volume		2.343	05:25
Air Blank		0.000	05:26
Control Test		0.080	05:26
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 04/16/2019 15:31

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASS,H,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: MANIER

DATE/TIME OF TEST: 03/30/2019 23:20:49

Diagnostics Check		OK	23:13
Air Blank		0.000	23:14
Control Test		0.079	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.140	23:16
Breath Volume		1.957	23:16
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.153	23:19
Breath Volume		2.363	23:19
Air Blank		0.000	23:19
Control Test		0.080	23:20
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 04/16/2019 15:31

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: PADILLA HERNANDEZ

DATE/TIME OF TEST: 03/02/2019 01:21:04

Diagnostics Check		OK	01:14
Air Blank		0.000	01:15
Control Test		0.079	01:15
Air Blank		0.000	01:16
Subject Sample	# 1	0.126	01:16
Breath Volume		2.324	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.120	01:19
Breath Volume		1.144	01:19
Air Blank		0.000	01:20
Control Test		0.079	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:21

Data Download Date/Time: 03/04/2019 13:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: SHOLLY

DATE/TIME OF TEST: 03/02/2019 03:30:36

Diagnostics Check		OK	03:27	
Air Blank		0.000	03:28	
Control Test		0.079	03:28	
Air Blank		0.000	03:29	
Subject Sample	# 1		03:29	Subject Test Refused
Breath Volume		0.000	03:29	
Air Blank		0.000	03:29	
Control Test		0.079	03:30	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:30	

Data Download Date/Time: 03/04/2019 13:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: HEINECK

DATE/TIME OF TEST: 03/02/2019 22:02:06

Diagnostics Check		OK	21:55
Air Blank		0.000	21:56
Control Test		0.079	21:56
Air Blank		0.000	21:57
Subject Sample	# 1	0.255	21:57
Breath Volume		2.390	21:57
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.252	22:00
Breath Volume		2.140	22:00
Air Blank		0.000	22:01
Control Test		0.079	22:01
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 03/04/2019 13:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: PEREZ LOPEZ

DATE/TIME OF TEST: 03/03/2019 02:11:19

Diagnostics Check		OK	02:04
Air Blank		0.000	02:05
Control Test		0.078	02:05
Air Blank		0.000	02:06
Subject Sample	# 1	0.202	02:06
Breath Volume		2.644	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.195	02:09
Breath Volume		1.843	02:09
Air Blank		0.000	02:10
Control Test		0.079	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 03/04/2019 13:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: FORREST

DATE/TIME OF TEST: 03/03/2019 03:07:54

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.079	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.105	03:03
Breath Volume		1.359	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.108	03:06
Breath Volume		1.816	03:06
Air Blank		0.000	03:06
Control Test		0.078	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 03/04/2019 13:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: ROBLES LOPEZ

DATE/TIME OF TEST: 03/03/2019 19:13:42

Diagnostics Check		OK	19:10	
Air Blank		0.000	19:11	
Control Test		0.080	19:11	
Air Blank		0.000	19:12	
Subject Sample	# 1		19:12	Subject Test Refused
Breath Volume		0.000	19:12	
Air Blank		0.000	19:12	
Control Test		0.079	19:13	
Air Blank		0.000	19:13	
Diagnostics Check		OK	19:13	

Data Download Date/Time: 03/04/2019 13:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,A,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 03/03/2019 22:39:25

Diagnostics Check		OK	22:33
Air Blank		0.000	22:33
Control Test		0.080	22:34
Air Blank		0.000	22:34
Subject Sample	# 1	0.063	22:34
Breath Volume		1.382	22:34
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.062	22:37
Breath Volume		1.250	22:37
Air Blank		0.000	22:38
Control Test		0.080	22:38
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 03/04/2019 13:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: SEARCY

DATE/TIME OF TEST: 03/04/2019 02:20:11

Diagnostics Check		OK	02:13
Air Blank		0.000	02:14
Control Test		0.079	02:14
Air Blank		0.000	02:15
Subject Sample	# 1	0.139	02:15
Breath Volume		1.566	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.128	02:18
Breath Volume		1.488	02:18
Air Blank		0.000	02:19
Control Test		0.079	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 03/04/2019 13:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: KOLODICH

DATE/TIME OF TEST: 03/05/2019 00:44:25

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:42	
Control Test		0.080	00:42	
Air Blank		0.000	00:42	
Subject Sample	# 1		00:43	Subject Test Refused
Breath Volume		0.000	00:43	
Air Blank		0.000	00:43	
Control Test		0.080	00:43	
Air Blank		0.000	00:44	
Diagnostics Check		OK	00:44	

Data Download Date/Time: 03/15/2019 11:47

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: BAIRD

DATE/TIME OF TEST: 03/05/2019 03:25:09

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:22	
Control Test		0.080	03:23	
Air Blank		0.000	03:23	
Subject Sample	# 1		03:23	Subject Test Refused
Breath Volume		0.000	03:23	
Air Blank		0.000	03:24	
Control Test		0.080	03:24	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 03/15/2019 11:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: CHAPNICK

DATE/TIME OF TEST: 03/05/2019 23:04:43

Diagnostics Check		OK	23:01	
Air Blank		0.000	23:02	
Control Test		0.081	23:02	
Air Blank		0.000	23:03	
Subject Sample	# 1		23:03	Subject Test Refused
Breath Volume		0.000	23:03	
Air Blank		0.000	23:03	
Control Test		0.081	23:04	
Air Blank		0.000	23:04	
Diagnostics Check		OK	23:04	

Data Download Date/Time: 03/15/2019 11:48

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: MONROE

DATE/TIME OF TEST: 03/05/2019 23:15:26

Diagnostics Check		OK	23:08
Air Blank		0.000	23:09
Control Test		0.080	23:09
Air Blank		0.000	23:10
Subject Sample	# 1	0.007	23:10
Breath Volume		3.558	23:10
Air Blank		0.000	23:11
Air Blank		0.000	23:13
Subject Sample	# 2	0.000	23:14
Breath Volume		1.871	23:14
Air Blank		0.000	23:14
Control Test		0.081	23:14
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 03/15/2019 11:48

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: DEPENZA

DATE/TIME OF TEST: 03/06/2019 00:25:05

Diagnostics Check		OK	00:19
Air Blank		0.000	00:19
Control Test		0.080	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.134	00:20
Breath Volume		2.093	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.134	00:23
Breath Volume		2.605	00:23
Air Blank		0.000	00:24
Control Test		0.080	00:24
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 03/15/2019 11:48

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GATES

DATE/TIME OF TEST: 03/07/2019 04:07:18

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.104	04:02
Breath Volume		3.917	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.096	04:05
Breath Volume		4.046	04:05
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 03/15/2019 11:48

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HERNANDEZ ROMAN

DATE/TIME OF TEST: 03/09/2019 02:07:51

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:56	
Control Test		0.080	01:56	
Air Blank		0.000	01:56	
Subject Sample	# 1	0.164	01:57	
Breath Volume		2.562	01:57	
Air Blank		0.000	01:58	
Air Blank		0.000	02:00	
Subject Sample	# 2	0.134	02:03	Slope Not Level
Breath Volume		1.226	02:03	
Air Blank		0.000	02:03	
Air Blank		0.000	02:05	
Subject Sample	# 3	0.172	02:06	
Breath Volume		2.152	02:06	
Air Blank		0.000	02:06	
Control Test		0.078	02:07	
Air Blank		0.000	02:07	
Diagnostics Check		OK	02:07	

Data Download Date/Time: 03/15/2019 11:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SOSA,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: KEARLEY

DATE/TIME OF TEST: 03/09/2019 05:31:41

Diagnostics Check		OK	05:28	
Air Blank		0.000	05:29	
Control Test		0.080	05:29	
Air Blank		0.000	05:30	
Subject Sample	# 1		05:30	Subject Test Refused
Breath Volume		0.000	05:30	
Air Blank		0.000	05:30	
Control Test		0.080	05:31	
Air Blank		0.000	05:31	
Diagnostics Check		OK	05:31	

Data Download Date/Time: 03/15/2019 11:48

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: BLOODWORTH

DATE/TIME OF TEST: 03/10/2019 01:24:48

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.079	01:18
Air Blank		0.000	01:19
Subject Sample	# 1	0.226	01:20
Breath Volume		2.199	01:20
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.237	01:23
Breath Volume		2.914	01:23
Air Blank		0.000	01:23
Control Test		0.079	01:24
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 03/15/2019 11:48

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 03/10/2019 05:44:01

Diagnostics Check		OK	05:41	
Air Blank		0.000	05:41	
Control Test		0.079	05:41	
Air Blank		0.000	05:42	
Subject Sample	# 1		05:42	Subject Test Refused
Breath Volume		0.000	05:42	
Air Blank		0.000	05:43	
Control Test		0.080	05:43	
Air Blank		0.000	05:43	
Diagnostics Check		OK	05:44	

Data Download Date/Time: 03/15/2019 11:48

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: SIKES

DATE/TIME OF TEST: 03/10/2019 20:33:59

Diagnostics Check		OK	20:31	
Air Blank		0.000	20:31	
Control Test		0.080	20:31	
Air Blank		0.000	20:32	
Subject Sample	# 1		20:32	Subject Test Refused
Breath Volume		0.000	20:32	
Air Blank		0.000	20:33	
Control Test		0.079	20:33	
Air Blank		0.000	20:33	
Diagnostics Check		OK	20:33	

Data Download Date/Time: 03/15/2019 11:48

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: YU

DATE/TIME OF TEST: 03/10/2019 21:02:44

Diagnostics Check		OK	20:55
Air Blank		0.000	20:56
Control Test		0.080	20:56
Air Blank		0.000	20:56
Subject Sample	# 1	0.118	20:57
Breath Volume		2.003	20:57
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.115	21:01
Breath Volume		1.988	21:01
Air Blank		0.000	21:01
Control Test		0.078	21:02
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 03/15/2019 11:48

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: RUSS

DATE/TIME OF TEST: 03/11/2019 01:31:39

Diagnostics Check		OK	01:28	
Air Blank		0.000	01:29	
Control Test		0.080	01:29	
Air Blank		0.000	01:30	
Subject Sample	# 1		01:30	Subject Test Refused
Breath Volume		0.000	01:30	
Air Blank		0.000	01:30	
Control Test		0.080	01:31	
Air Blank		0.000	01:31	
Diagnostics Check		OK	01:31	

Data Download Date/Time: 03/15/2019 11:48

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HANLON

DATE/TIME OF TEST: 03/12/2019 03:07:06

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.077	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.134	03:02
Breath Volume		2.464	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:04
Subject Sample	# 2	0.136	03:05
Breath Volume		3.550	03:05
Air Blank		0.000	03:06
Control Test		0.077	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 03/15/2019 11:48

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: REGISTER

DATE/TIME OF TEST: 03/14/2019 04:51:18

Diagnostics Check		OK	04:44
Air Blank		0.000	04:45
Control Test		0.080	04:45
Air Blank		0.000	04:46
Subject Sample	# 1	0.000	04:47
Breath Volume		1.285	04:47
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.000	04:49
Breath Volume		1.878	04:49
Air Blank		0.000	04:50
Control Test		0.079	04:50
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 03/15/2019 11:48

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 03/16/2019 00:57:49

Diagnostics Check		OK	00:51
Air Blank		0.000	00:52
Control Test		0.080	00:52
Air Blank		0.000	00:52
Subject Sample	# 1	0.199	00:53
Breath Volume		2.265	00:53
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.200	00:56
Breath Volume		1.769	00:56
Air Blank		0.000	00:56
Control Test		0.079	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 03/16/2019 02:07:15

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.115	02:02
Breath Volume		3.187	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.119	02:05
Breath Volume		3.121	02:05
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SORRELLS,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: CANDY

DATE/TIME OF TEST: 03/16/2019 06:45:47

Diagnostics Check		OK	06:42	
Air Blank		0.000	06:43	
Control Test		0.080	06:43	
Air Blank		0.000	06:44	
Subject Sample	# 1		06:44	Subject Test Refused
Breath Volume		0.000	06:44	
Air Blank		0.000	06:44	
Control Test		0.080	06:45	
Air Blank		0.000	06:45	
Diagnostics Check		OK	06:45	

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: MOSBY

DATE/TIME OF TEST: 03/17/2019 02:03:01

Diagnostics Check		OK	01:56
Air Blank		0.000	01:57
Control Test		0.080	01:57
Air Blank		0.000	01:58
Subject Sample	# 1	0.185	01:58
Breath Volume		2.347	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:00
Subject Sample	# 2	0.194	02:01
Breath Volume		2.574	02:01
Air Blank		0.000	02:02
Control Test		0.080	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:03

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BODENSTEIN

DATE/TIME OF TEST: 03/17/2019 02:42:35

Diagnostics Check		OK	02:36
Air Blank		0.000	02:36
Control Test		0.080	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.155	02:37
Breath Volume		4.437	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.150	02:41
Breath Volume		4.367	02:41
Air Blank		0.000	02:41
Control Test		0.080	02:41
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: PULASKI

DATE/TIME OF TEST: 03/17/2019 04:17:21

Diagnostics Check		OK	04:14	
Air Blank		0.000	04:15	
Control Test		0.080	04:15	
Air Blank		0.000	04:15	
Subject Sample	# 1		04:15	Subject Test Refused
Breath Volume		0.000	04:15	
Air Blank		0.000	04:16	
Control Test		0.080	04:16	
Air Blank		0.000	04:17	
Diagnostics Check		OK	04:17	

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 03/18/2019 01:45:33

Diagnostics Check		OK	01:42	
Air Blank		0.000	01:43	
Control Test		0.081	01:43	
Air Blank		0.000	01:44	
Subject Sample	# 1		01:44	Subject Test Refused
Breath Volume		0.000	01:44	
Air Blank		0.000	01:44	
Control Test		0.080	01:44	
Air Blank		0.000	01:45	
Diagnostics Check		OK	01:45	

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 03/18/2019 04:26:04

Diagnostics Check		OK	04:23	
Air Blank		0.000	04:23	
Control Test		0.080	04:24	
Air Blank		0.000	04:24	
Subject Sample	# 1		04:24	Subject Test Refused
Breath Volume		0.000	04:24	
Air Blank		0.000	04:25	
Control Test		0.081	04:25	
Air Blank		0.000	04:25	
Diagnostics Check		OK	04:26	

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DALESSIO,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: FLOYD

DATE/TIME OF TEST: 03/18/2019 06:19:43

Diagnostics Check		OK	06:13
Air Blank		0.000	06:13
Control Test		0.080	06:14
Air Blank		0.000	06:14
Subject Sample	# 1	0.138	06:15
Breath Volume		2.214	06:15
Air Blank		0.000	06:15
Air Blank		0.000	06:17
Subject Sample	# 2	0.136	06:18
Breath Volume		2.707	06:18
Air Blank		0.000	06:18
Control Test		0.080	06:19
Air Blank		0.000	06:19
Diagnostics Check		OK	06:19

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: SALGADO

DATE/TIME OF TEST: 03/20/2019 04:31:25

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.080	04:25
Air Blank		0.000	04:26
Subject Sample	# 1	0.155	04:26
Breath Volume		3.289	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.154	04:29
Breath Volume		3.578	04:29
Air Blank		0.000	04:30
Control Test		0.080	04:30
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: MOHLER

DATE/TIME OF TEST: 03/21/2019 03:06:30

Diagnostics Check		OK	03:03	
Air Blank		0.000	03:04	
Control Test		0.080	03:04	
Air Blank		0.000	03:04	
Subject Sample	# 1		03:05	Subject Test Refused
Breath Volume		0.000	03:05	
Air Blank		0.000	03:05	
Control Test		0.080	03:05	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:06	

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: SANTOS ACOSTA

DATE/TIME OF TEST: 03/22/2019 02:50:53

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.080	02:45
Air Blank		0.000	02:45
Subject Sample	# 1	0.171	02:46
Breath Volume		1.273	02:46
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.161	02:49
Breath Volume		1.511	02:49
Air Blank		0.000	02:49
Control Test		0.080	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CHRISTIAN

DATE/TIME OF TEST: 03/22/2019 03:45:30

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:43	
Control Test		0.080	03:43	
Air Blank		0.000	03:43	
Subject Sample	# 1		03:44	Subject Test Refused
Breath Volume		0.000	03:44	
Air Blank		0.000	03:44	
Control Test		0.080	03:44	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALDREP,B
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 03/22/2019 04:40:44

Diagnostics Check		OK	04:34
Air Blank		0.000	04:34
Control Test		0.080	04:35
Air Blank		0.000	04:35
Subject Sample	# 1	0.094	04:36
Breath Volume		3.570	04:36
Air Blank		0.000	04:36
Air Blank		0.000	04:38
Subject Sample	# 2	0.088	04:39
Breath Volume		1.890	04:39
Air Blank		0.000	04:39
Control Test		0.080	04:40
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: HOGENCAMP

DATE/TIME OF TEST: 03/23/2019 02:56:18

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.081	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.069	02:51
Breath Volume		2.855	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.070	02:54
Breath Volume		3.597	02:54
Air Blank		0.000	02:55
Control Test		0.080	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SORRELS,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: BOWMAN

DATE/TIME OF TEST: 03/23/2019 03:46:05

Diagnostics Check		OK	03:43	
Air Blank		0.000	03:43	
Control Test		0.080	03:44	
Air Blank		0.000	03:44	
Subject Sample	# 1		03:44	Subject Test Refused
Breath Volume		0.000	03:44	
Air Blank		0.000	03:45	
Control Test		0.080	03:45	
Air Blank		0.000	03:46	
Diagnostics Check		OK	03:46	

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SOSA,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 03/23/2019 04:26:00

Diagnostics Check		OK	04:19
Air Blank		0.000	04:20
Control Test		0.080	04:20
Air Blank		0.000	04:21
Subject Sample	# 1	0.186	04:21
Breath Volume		1.187	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:23
Subject Sample	# 2	0.182	04:24
Breath Volume		1.726	04:24
Air Blank		0.000	04:25
Control Test		0.080	04:25
Air Blank		0.000	04:25
Diagnostics Check		OK	04:26

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: HOGUE

DATE/TIME OF TEST: 03/24/2019 01:11:25

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.079	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.274	01:06
Breath Volume		3.085	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.276	01:09
Breath Volume		3.027	01:09
Air Blank		0.000	01:10
Control Test		0.079	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: ANDERSEN

DATE/TIME OF TEST: 03/24/2019 02:41:22

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.080	02:35
Air Blank		0.000	02:36
Subject Sample	# 1	0.113	02:36
Breath Volume		3.117	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.102	02:39
Breath Volume		3.570	02:39
Air Blank		0.000	02:40
Control Test		0.080	02:40
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 03/24/2019 05:34:06

Diagnostics Check		OK	05:31	
Air Blank		0.000	05:31	
Control Test		0.080	05:32	
Air Blank		0.000	05:32	
Subject Sample	# 1		05:32	Subject Test Refused
Breath Volume		0.000	05:32	
Air Blank		0.000	05:33	
Control Test		0.080	05:33	
Air Blank		0.000	05:34	
Diagnostics Check		OK	05:34	

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 03/29/2019 01:51:59

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.080	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.144	01:47
Breath Volume		2.394	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.145	01:50
Breath Volume		2.367	01:50
Air Blank		0.000	01:51
Control Test		0.080	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: ENSLEY

DATE/TIME OF TEST: 03/29/2019 03:21:44

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:19	
Control Test		0.080	03:19	
Air Blank		0.000	03:20	
Subject Sample	# 1		03:20	Subject Test Refused
Breath Volume		0.000	03:20	
Air Blank		0.000	03:20	
Control Test		0.080	03:21	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BEALL,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SAPP

DATE/TIME OF TEST: 03/30/2019 02:46:50

Diagnostics Check		OK	02:44	
Air Blank		0.000	02:44	
Control Test		0.079	02:44	
Air Blank		0.000	02:45	
Subject Sample	# 1		02:45	Subject Test Refused
Breath Volume		0.000	02:45	
Air Blank		0.000	02:45	
Control Test		0.080	02:46	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:46	

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORRO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: BATTS

DATE/TIME OF TEST: 03/30/2019 04:35:04

Diagnostics Check		OK	04:28
Air Blank		0.000	04:29
Control Test		0.080	04:29
Air Blank		0.000	04:29
Subject Sample	# 1	0.214	04:30
Breath Volume		1.917	04:30
Air Blank		0.000	04:31
Air Blank		0.000	04:32
Subject Sample	# 2	0.213	04:33
Breath Volume		1.710	04:33
Air Blank		0.000	04:34
Control Test		0.080	04:34
Air Blank		0.000	04:34
Diagnostics Check		OK	04:35

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/30/2019 23:12:52

Diagnostics Check		OK	23:06
Air Blank		0.000	23:07
Control Test		0.080	23:07
Air Blank		0.000	23:07
Subject Sample	# 1	0.176	23:08
Breath Volume		2.570	23:08
Air Blank		0.000	23:09
Air Blank		0.000	23:10
Subject Sample	# 2	0.170	23:11
Breath Volume		1.468	23:11
Air Blank		0.000	23:11
Control Test		0.079	23:12
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: GAINES

DATE/TIME OF TEST: 03/31/2019 01:32:43

Diagnostics Check		OK	01:26
Air Blank		0.000	01:27
Control Test		0.080	01:27
Air Blank		0.000	01:27
Subject Sample	# 1	0.000	01:28
Breath Volume		4.355	01:28
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.000	01:31
Breath Volume		4.378	01:31
Air Blank		0.000	01:31
Control Test		0.080	01:32
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: ALEXI GOMEZ

DATE/TIME OF TEST: 03/31/2019 02:41:27

Diagnostics Check		OK	02:38	
Air Blank		0.000	02:39	
Control Test		0.079	02:39	
Air Blank		0.000	02:39	
Subject Sample	# 1		02:40	Subject Test Refused
Breath Volume		0.000	02:40	
Air Blank		0.000	02:40	
Control Test		0.080	02:40	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PAUL,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 03/31/2019 04:52:39

Diagnostics Check		OK	04:39	
Air Blank		0.000	04:40	
Control Test		0.079	04:40	
Air Blank		0.000	04:41	
Subject Sample	# 1	0.100	04:44	Volume Not Met
Breath Volume		0.601	04:44	
Air Blank		0.000	04:45	
Air Blank		0.000	04:46	
Subject Sample	# 2	0.112	04:47	
Breath Volume		1.273	04:47	
Air Blank		0.000	04:48	
Air Blank		0.000	04:50	
Subject Sample	# 3	0.114	04:51	
Breath Volume		1.613	04:51	
Air Blank		0.000	04:51	
Control Test		0.078	04:52	
Air Blank		0.000	04:52	
Diagnostics Check		OK	04:52	

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: PRZEPASNIAK

DATE/TIME OF TEST: 03/31/2019 22:40:46

Diagnostics Check		OK	22:32
Air Blank		0.000	22:32
Control Test		0.080	22:33
Air Blank		0.000	22:33
Subject Sample	# 1	0.189	22:35
Breath Volume		1.308	22:35
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.180	22:39
Breath Volume		1.171	22:39
Air Blank		0.000	22:39
Control Test		0.078	22:40
Air Blank		0.000	22:40
Diagnostics Check		OK	22:40

Data Download Date/Time: 04/16/2019 15:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDDERLY,G,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SOLTERO,H
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: BAIN

DATE/TIME OF TEST: 03/02/2019 17:37:28

Diagnostics Check		OK	17:33	
Air Blank		0.000	17:33	
Control Test		0.077	17:34	
Air Blank		0.000	17:34	
Subject Sample	# 1		17:36	Subject Test Refused
Breath Volume		0.000	17:36	
Air Blank		0.000	17:36	
Control Test		0.078	17:36	
Air Blank		0.000	17:37	
Diagnostics Check		OK	17:37	

Data Download Date/Time: 04/01/2019 16:29

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WALTON,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: LITTLETON

DATE/TIME OF TEST: 03/02/2019 22:40:23

Diagnostics Check		OK	22:33
Air Blank		0.000	22:34
Control Test		0.077	22:34
Air Blank		0.000	22:35
Subject Sample	# 1	0.000	22:35
Breath Volume		1.988	22:35
Air Blank		0.000	22:36
Air Blank		0.000	22:38
Subject Sample	# 2	0.000	22:39
Breath Volume		2.261	22:39
Air Blank		0.000	22:39
Control Test		0.078	22:39
Air Blank		0.000	22:40
Diagnostics Check		OK	22:40

Data Download Date/Time: 04/01/2019 16:29

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,J,
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GALLOWAY,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: JADOO

DATE/TIME OF TEST: 03/03/2019 15:14:34

Diagnostics Check		OK	15:08
Air Blank		0.000	15:08
Control Test		0.077	15:09
Air Blank		0.000	15:09
Subject Sample	# 1	0.290	15:10
Breath Volume		2.335	15:10
Air Blank		0.000	15:10
Air Blank		0.000	15:12
Subject Sample	# 2	0.278	15:13
Breath Volume		2.449	15:13
Air Blank		0.000	15:13
Control Test		0.076	15:13
Air Blank		0.000	15:14
Diagnostics Check		OK	15:14

Data Download Date/Time: 04/01/2019 16:29

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SHOOK,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 03/08/2019 01:48:02

Diagnostics Check		OK	01:41
Air Blank		0.000	01:42
Control Test		0.078	01:42
Air Blank		0.000	01:42
Subject Sample	# 1	0.196	01:43
Breath Volume		2.273	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:45
Subject Sample	# 2	0.202	01:46
Breath Volume		2.125	01:46
Air Blank		0.000	01:47
Control Test		0.078	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:48

Data Download Date/Time: 04/01/2019 16:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 03/09/2019 00:37:14

Diagnostics Check		OK	00:30
Air Blank		0.000	00:31
Control Test		0.077	00:31
Air Blank		0.000	00:32
Subject Sample	# 1	0.000	00:32
Breath Volume		2.207	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.000	00:35
Breath Volume		1.765	00:35
Air Blank		0.000	00:36
Control Test		0.077	00:36
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 04/01/2019 16:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRACKEN,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUZE,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MEEKS

DATE/TIME OF TEST: 03/12/2019 00:32:23

Diagnostics Check		OK	00:25
Air Blank		0.000	00:26
Control Test		0.075	00:26
Air Blank		0.000	00:26
Subject Sample	# 1	0.164	00:27
Breath Volume		2.597	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.166	00:30
Breath Volume		2.605	00:30
Air Blank		0.000	00:31
Control Test		0.078	00:31
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 04/01/2019 16:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: KLAAREN

DATE/TIME OF TEST: 03/12/2019 01:33:15

Diagnostics Check		OK	01:27
Air Blank		0.000	01:27
Control Test		0.077	01:27
Air Blank		0.000	01:28
Subject Sample	# 1	0.000	01:28
Breath Volume		2.812	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.000	01:31
Breath Volume		2.957	01:31
Air Blank		0.000	01:32
Control Test		0.077	01:32
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 04/01/2019 16:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: COATS

DATE/TIME OF TEST: 03/17/2019 03:03:35

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.077	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.200	02:59
Breath Volume		1.281	02:59
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.190	03:02
Breath Volume		1.890	03:02
Air Blank		0.000	03:02
Control Test		0.078	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 04/01/2019 16:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRACKEN,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: 2C16,HUNTER
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 03/17/2019 22:10:16

Diagnostics Check		OK	22:00
Air Blank		0.000	22:01
Control Test		0.078	22:01
Air Blank		0.000	22:02
Subject Sample	# 1	0.228	22:05
Breath Volume		2.632	22:05
Air Blank		0.000	22:05
Air Blank		0.000	22:07
Subject Sample	# 2	0.217	22:08
Breath Volume		2.757	22:08
Air Blank		0.000	22:09
Control Test		0.077	22:09
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 04/01/2019 16:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SCHADOWSKY

DATE/TIME OF TEST: 03/18/2019 03:57:52

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.077	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.119	03:53
Breath Volume		2.105	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:55
Subject Sample	# 2	0.115	03:56
Breath Volume		2.394	03:56
Air Blank		0.000	03:56
Control Test		0.078	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 04/01/2019 16:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOLARD,RAYMOND,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JACKSON,ROBERT
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:52

SUBJECT'S LAST NAME: SPILLIS

DATE/TIME OF TEST: 03/18/2019 16:19:35

Diagnostics Check		OK	16:13
Air Blank		0.000	16:14
Control Test		0.078	16:14
Air Blank		0.000	16:14
Subject Sample	# 1	0.000	16:15
Breath Volume		2.531	16:15
Air Blank		0.000	16:15
Air Blank		0.000	16:17
Subject Sample	# 2	0.000	16:18
Breath Volume		3.917	16:18
Air Blank		0.000	16:18
Control Test		0.077	16:18
Air Blank		0.000	16:19
Diagnostics Check		OK	16:19

Data Download Date/Time: 04/01/2019 16:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDDERLY,G,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: NEWLON,J
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:28

SUBJECT'S LAST NAME: CLEGHORN

DATE/TIME OF TEST: 03/21/2019 14:55:01

Diagnostics Check		OK	14:49	
Air Blank		0.000	14:50	
Control Test		0.077	14:50	
Air Blank		0.000	14:51	
Subject Sample	# 1		14:53	Subject Test Refused
Breath Volume		0.000	14:53	
Air Blank		0.000	14:54	
Control Test		0.077	14:54	
Air Blank		0.000	14:54	
Diagnostics Check		OK	14:55	

Data Download Date/Time: 04/01/2019 16:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BURSE,E
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: CHUMIOQUE

DATE/TIME OF TEST: 03/23/2019 04:37:02

Diagnostics Check		OK	04:34	
Air Blank		0.000	04:34	
Control Test		0.078	04:34	
Air Blank		0.000	04:35	
Subject Sample	# 1		04:35	Subject Test Refused
Breath Volume		0.000	04:35	
Air Blank		0.000	04:36	
Control Test		0.079	04:36	
Air Blank		0.000	04:36	
Diagnostics Check		OK	04:37	

Data Download Date/Time: 04/01/2019 16:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOLARD,RAYMOND,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: PINKSTON,IVAN
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:31

SUBJECT'S LAST NAME: USHER

DATE/TIME OF TEST: 03/27/2019 14:07:44

Diagnostics Check	OK	14:06	
Air Blank	0.000	14:06	
Control Test	0.074	14:07	Control Outside Tolerance
Air Blank	0.000	14:07	

Data Download Date/Time: 04/01/2019 16:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOLARD,RAYMOND,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: PINKSTON,IVAN
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:31

SUBJECT'S LAST NAME: USHER

DATE/TIME OF TEST: 03/27/2019 14:24:52

Diagnostics Check		OK	14:12	
Air Blank		0.000	14:13	
Control Test		0.076	14:13	
Air Blank		0.000	14:14	
Subject Sample	# 1	0.348	14:17	Volume Not Met
Breath Volume		1.078	14:17	
Air Blank		0.000	14:18	
Air Blank		0.000	14:19	
Subject Sample	# 2	0.327	14:22	Volume Not Met
Breath Volume		0.683	14:22	
Air Blank		0.000	14:23	
Control Test		0.073	14:24	Control Outside Tolerance
Air Blank		0.000	14:24	

Data Download Date/Time: 04/01/2019 16:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SCHMIDT,RYAN
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: LUGG

DATE/TIME OF TEST: 03/27/2019 16:49:02

Diagnostics Check		OK	16:41
Air Blank		0.000	16:41
Control Test		0.077	16:41
Air Blank		0.000	16:42
Subject Sample	# 1	0.149	16:44
Breath Volume		1.332	16:44
Air Blank		0.000	16:44
Air Blank		0.000	16:46
Subject Sample	# 2	0.151	16:47
Breath Volume		2.625	16:47
Air Blank		0.000	16:48
Control Test		0.075	16:48
Air Blank		0.000	16:48
Diagnostics Check		OK	16:49

Data Download Date/Time: 04/01/2019 16:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,J,
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: MOO YOUNG

DATE/TIME OF TEST: 03/30/2019 02:18:04

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:15	
Control Test		0.077	02:16	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:16	Subject Test Refused
Breath Volume		0.000	02:16	
Air Blank		0.000	02:17	
Control Test		0.078	02:17	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 04/01/2019 16:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: TERRY,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: CASSADA

DATE/TIME OF TEST: 03/31/2019 00:49:08

Diagnostics Check		OK	00:46	
Air Blank		0.000	00:46	
Control Test		0.078	00:47	
Air Blank		0.000	00:47	
Subject Sample	# 1		00:47	Subject Test Refused
Breath Volume		0.000	00:47	
Air Blank		0.000	00:48	
Control Test		0.077	00:48	
Air Blank		0.000	00:49	
Diagnostics Check		OK	00:49	

Data Download Date/Time: 04/01/2019 16:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: COPHER,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: MELTON

DATE/TIME OF TEST: 03/31/2019 01:39:36

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.078	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.000	01:35
Breath Volume		2.089	01:35
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.000	01:38
Breath Volume		2.210	01:38
Air Blank		0.000	01:38
Control Test		0.077	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 04/01/2019 16:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,J,
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BORLAND

DATE/TIME OF TEST: 03/31/2019 03:30:12

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.077	03:24
Air Blank		0.000	03:25
Subject Sample	# 1	0.120	03:25
Breath Volume		2.449	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.128	03:28
Breath Volume		4.429	03:28
Air Blank		0.000	03:29
Control Test		0.077	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 04/01/2019 16:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRACKEN,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 03/31/2019 19:44:13

Diagnostics Check		OK	19:36
Air Blank		0.000	19:36
Control Test		0.078	19:36
Air Blank		0.000	19:37
Subject Sample	# 1	0.081	19:39
Breath Volume		1.246	19:39
Air Blank		0.000	19:39
Air Blank		0.000	19:41
Subject Sample	# 2	0.086	19:42
Breath Volume		1.480	19:42
Air Blank		0.000	19:43
Control Test		0.077	19:43
Air Blank		0.000	19:44
Diagnostics Check		OK	19:44

Data Download Date/Time: 04/01/2019 16:30

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YAKUBSIN,JAMES,R
BREATH TEST OPERATOR AGENCY: High Springs PD
ARRESTING OFFICER: FRICKE,JON
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 03/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 03/24/2019 23:43:57

Diagnostics Check		OK	23:40	
Air Blank		0.000	23:41	
Control Test		0.081	23:41	
Air Blank		0.000	23:42	
Subject Sample	# 1		23:42	Subject Test Refused
Breath Volume		0.000	23:42	
Air Blank		0.000	23:43	
Control Test		0.080	23:43	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:43	

Data Download Date/Time: 03/28/2019 02:48

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,CALEB,A
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: HEATH,MARC
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: RUMSEY

DATE/TIME OF TEST: 03/04/2019 23:17:26

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.082	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.205	23:13
Breath Volume		2.332	23:13
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.207	23:15
Breath Volume		2.269	23:15
Air Blank		0.000	23:16
Control Test		0.083	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 04/17/2019 10:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURRY,TYLER,M
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: HIGGINBOTHAM,JEFFREY
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SHAMOEL

DATE/TIME OF TEST: 03/20/2019 03:43:18

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.084	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.099	03:38
Breath Volume		2.960	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.105	03:41
Breath Volume		2.812	03:41
Air Blank		0.000	03:42
Control Test		0.082	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 04/17/2019 10:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: RAULERSON

DATE/TIME OF TEST: 03/23/2019 23:08:18

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.084	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.000	23:03
Breath Volume		1.367	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.000	23:06
Breath Volume		1.351	23:06
Air Blank		0.000	23:07
Control Test		0.082	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 04/17/2019 10:17

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KRIS,M
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: MALLARD,J
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 03/29/2019 23:42:15

Diagnostics Check		OK	23:30	
Air Blank		0.000	23:31	
Control Test		0.082	23:31	
Air Blank		0.000	23:32	
Subject Sample	# 1		23:35	No Sample Provided
Breath Volume		0.000	23:35	
Air Blank		0.000	23:35	
Air Blank		0.000	23:37	
Subject Sample	# 2		23:40	No Sample Provided
Breath Volume		0.000	23:40	
Air Blank		0.000	23:41	
Control Test		0.083	23:41	
Air Blank		0.000	23:42	
Diagnostics Check		OK	23:42	

Data Download Date/Time: 04/17/2019 10:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAEYAERT,JEFF,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: NAEYAERT,JEFF
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 08/07/2020
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 03/04/2019 21:57:14

Diagnostics Check	OK	21:56	
Air Blank		21:56	RFI Detect
Air Blank		21:57	RFI Detect

Data Download Date/Time: 03/05/2019 16:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAEYAERT,JEFF,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: NAEYAERT,JEFF
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 08/07/2020
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 03/04/2019 22:03:44

Diagnostics Check		OK	22:00	
Air Blank		0.000	22:00	
Control Test		0.080	22:01	
Air Blank		0.000	22:01	
Subject Sample	# 1	0.037	22:02	
Breath Volume		1.273	22:02	
Air Blank			22:02	RFI Detect
Air Blank		0.000	22:02	
Control Test		0.079	22:03	
Air Blank		0.000	22:03	
Diagnostics Check		OK	22:03	

Data Download Date/Time: 03/05/2019 16:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAEYAERT,JEFF,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: NAEYAERT,JEFF
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 08/07/2020
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 03/04/2019 22:07:52

Diagnostics Check	OK	22:06	
Air Blank	0.000	22:06	
Control Test	0.079	22:06	
Air Blank	0.000	22:07	
		22:07	RFI Detect
Air Blank	0.000	22:07	

Data Download Date/Time: 03/05/2019 16:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAEYAERT,JEFF,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: NAEYAERT,JEFF
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 08/07/2020
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 03/04/2019 22:15:44

Diagnostics Check		OK	22:09
Air Blank		0.000	22:10
Control Test		0.079	22:10
Air Blank		0.000	22:11
Subject Sample	# 1	0.034	22:11
Breath Volume		1.695	22:11
Air Blank		0.000	22:11
Air Blank		0.000	22:13
Subject Sample	# 2	0.032	22:14
Breath Volume		1.476	22:14
Air Blank		0.000	22:14
Control Test		0.078	22:15
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 03/05/2019 16:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOFMANN,TRAVIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HOFMANN,TRAVIS
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823252
DRY GAS EXPIRY DATE: 08/07/2020
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/15/2019 21:31:43

Diagnostics Check		OK	21:28	
Air Blank		0.000	21:29	
Control Test		0.080	21:29	
Air Blank		0.000	21:30	
Subject Sample	# 1		21:30	Subject Test Refused
Breath Volume		0.000	21:30	
Air Blank		0.000	21:30	
Control Test		0.080	21:31	
Air Blank		0.000	21:31	
Diagnostics Check		OK	21:31	

Data Download Date/Time: 03/28/2019 14:27

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TERRELL,G
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: ROONEY,MICHAEL
ARRESTING OFFICER AGENCY: Starke PD
LAST AGENCY INSPECTION: 03/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: ALTMAN

DATE/TIME OF TEST: 03/16/2019 03:10:11

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.078	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.252	03:05
Breath Volume		2.558	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.233	03:08
Breath Volume		2.386	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 04/30/2019 13:58

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,WILLIAM,D
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: LONG,KATHLEEN
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 03/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 09:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:32

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 03/07/2019 11:06:44

Diagnostics Check		OK	11:00
Air Blank		0.000	11:00
Control Test		0.078	11:01
Air Blank		0.000	11:01
Subject Sample	# 1	0.208	11:02
Breath Volume		1.207	11:02
Air Blank		0.000	11:02
Air Blank		0.000	11:04
Subject Sample	# 2	0.206	11:05
Breath Volume		2.722	11:05
Air Blank		0.000	11:05
Control Test		0.078	11:06
Air Blank		0.000	11:06
Diagnostics Check		OK	11:06

Data Download Date/Time: 04/29/2019 18:24

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESHER,TIFFANY,A
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: VALENCIA,J
ARRESTING OFFICER AGENCY: Gilchrist County SO
LAST AGENCY INSPECTION: 03/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HERNANDEZ DIAZ

DATE/TIME OF TEST: 03/23/2019 23:21:07

Diagnostics Check		OK	23:14
Air Blank		0.000	23:14
Control Test		0.079	23:15
Air Blank		0.000	23:15
Subject Sample	# 1	0.223	23:16
Breath Volume		1.511	23:16
Air Blank		0.000	23:17
Air Blank		0.000	23:18
Subject Sample	# 2	0.237	23:19
Breath Volume		1.789	23:19
Air Blank		0.000	23:20
Control Test		0.078	23:20
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 04/29/2019 18:24

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESHER,TIFFANY,A
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: PAGAN,M
ARRESTING OFFICER AGENCY: Cedar Keys PD
LAST AGENCY INSPECTION: 03/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 03/24/2019 00:20:10

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.079	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.114	00:15
Breath Volume		1.832	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.120	00:18
Breath Volume		1.714	00:18
Air Blank		0.000	00:19
Control Test		0.079	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 04/29/2019 18:24

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASH,EARL,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DOWNING,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/24/2019 21:14:51

Diagnostics Check		OK	21:12	
Air Blank		0.000	21:12	
Control Test		0.078	21:12	
Air Blank		0.000	21:13	
Subject Sample	# 1		21:13	Subject Test Refused
Breath Volume		0.000	21:13	
Air Blank		0.000	21:13	
Control Test		0.078	21:14	
Air Blank		0.000	21:14	
Diagnostics Check		OK	21:14	

Data Download Date/Time: 04/29/2019 18:24

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKER,DEREK,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: LONG,KATHY
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 03/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: PASCUAL MARTINEZ

DATE/TIME OF TEST: 03/29/2019 18:06:49

Diagnostics Check		OK	17:55	
Air Blank		0.000	17:55	
Control Test		0.079	17:55	
Air Blank		0.000	17:56	
Subject Sample	# 1	0.146	17:59	Volume Not Met
Breath Volume		0.050	17:59	
Air Blank		0.000	18:00	
Air Blank		0.000	18:01	
Subject Sample	# 2	0.165	18:02	
Breath Volume		2.613	18:02	
Air Blank		0.000	18:03	
Air Blank		0.000	18:04	
Subject Sample	# 3	0.172	18:05	
Breath Volume		1.515	18:05	
Air Blank		0.000	18:05	
Control Test		0.077	18:06	
Air Blank		0.000	18:06	
Diagnostics Check		OK	18:06	

Data Download Date/Time: 04/29/2019 18:24

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWLING,DAKOTA,A
BREATH TEST OPERATOR AGENCY: Cedar Keys PD
ARRESTING OFFICER: HAMILTON,LUCAS
ARRESTING OFFICER AGENCY: Cedar Keys PD
LAST AGENCY INSPECTION: 03/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/30/2019 20:40:17

Diagnostics Check		OK	20:37	
Air Blank		0.000	20:37	
Control Test		0.079	20:38	
Air Blank		0.000	20:38	
Subject Sample	# 1		20:38	Subject Test Refused
Breath Volume		0.000	20:38	
Air Blank		0.000	20:39	
Control Test		0.079	20:39	
Air Blank		0.000	20:40	
Diagnostics Check		OK	20:40	

Data Download Date/Time: 04/29/2019 18:24

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,JEREMIAH,J
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: FINKELL,JOHN
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 03/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805701
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 03/31/2019 19:23:14

Diagnostics Check		OK	19:16
Air Blank		0.000	19:17
Control Test		0.078	19:17
Air Blank		0.000	19:18
Subject Sample	# 1	0.185	19:18
Breath Volume		1.519	19:18
Air Blank		0.000	19:19
Air Blank		0.000	19:21
Subject Sample	# 2	0.185	19:21
Breath Volume		2.179	19:21
Air Blank		0.000	19:22
Control Test		0.077	19:22
Air Blank		0.000	19:23
Diagnostics Check		OK	19:23

Data Download Date/Time: 04/29/2019 18:24

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDIS,BRYAN,S
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: LANDIS,BRYAN
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 02/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 937240
DRY GAS EXPIRY DATE: 04/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 03/02/2019 01:29:01

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.080	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.095	01:24
Breath Volume		2.210	01:24
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.100	01:27
Breath Volume		2.855	01:27
Air Blank		0.000	01:28
Control Test		0.079	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:29

Data Download Date/Time: 03/18/2019 13:54

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 16:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: PAGE

DATE/TIME OF TEST: 03/29/2019 18:34:08

Diagnostics Check		OK	18:31	
Air Blank		0.000	18:31	
Control Test		0.078	18:32	
Air Blank		0.000	18:32	
Subject Sample	# 1		18:32	Subject Test Refused
Breath Volume		0.000	18:32	
Air Blank		0.000	18:33	
Control Test		0.078	18:33	
Air Blank		0.000	18:34	
Diagnostics Check		OK	18:34	

Data Download Date/Time: 03/29/2019 19:43

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGGAN,MARLON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOGGAN,MARLON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG805702
DRY GAS EXPIRY DATE: 02/26/2020
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: WAITS

DATE/TIME OF TEST: 03/30/2019 00:44:11

Diagnostics Check		OK	00:38
Air Blank		0.000	00:38
Control Test		0.077	00:38
Air Blank		0.000	00:39
Subject Sample	# 1	0.203	00:39
Breath Volume		2.035	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.208	00:42
Breath Volume		1.691	00:42
Air Blank		0.000	00:43
Control Test		0.077	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 04/30/2019 15:26

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOHN,F
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: HURST

DATE/TIME OF TEST: 03/02/2019 00:47:41

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.079	00:42
Air Blank		0.000	00:42
Subject Sample	# 1	0.238	00:43
Breath Volume		1.445	00:43
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.224	00:46
Breath Volume		2.089	00:46
Air Blank		0.000	00:46
Control Test		0.079	00:47
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 03/29/2019 19:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVIER,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LEVIER,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 03/02/2019 03:29:59

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.079	03:24
Air Blank		0.000	03:25
Subject Sample	# 1	0.098	03:25
Breath Volume		1.332	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.100	03:28
Breath Volume		2.625	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 03/29/2019 19:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRIBB,BRIAN,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CRIBB,BRIAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 07:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/02/2019 08:47:27

Diagnostics Check		OK	08:41
Air Blank		0.000	08:41
Control Test		0.080	08:41
Air Blank		0.000	08:42
Subject Sample	# 1	0.153	08:43
Breath Volume		2.113	08:43
Air Blank		0.000	08:43
Air Blank		0.000	08:45
Subject Sample	# 2	0.137	08:45
Breath Volume		1.378	08:45
Air Blank		0.000	08:46
Control Test		0.078	08:46
Air Blank		0.000	08:47
Diagnostics Check		OK	08:47

Data Download Date/Time: 03/29/2019 19:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,WILLIAM,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ADAMS,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 03/02/2019 17:36:36

Diagnostics Check		OK	17:33	
Air Blank		0.000	17:34	
Control Test		0.079	17:34	
Air Blank		0.000	17:34	
Subject Sample	# 1		17:35	Subject Test Refused
Breath Volume		0.000	17:35	
Air Blank		0.000	17:35	
Control Test		0.079	17:36	
Air Blank		0.000	17:36	
Diagnostics Check		OK	17:36	

Data Download Date/Time: 03/29/2019 19:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: JOINER

DATE/TIME OF TEST: 03/03/2019 05:40:32

Diagnostics Check		OK	05:34
Air Blank		0.000	05:34
Control Test		0.080	05:34
Air Blank		0.000	05:35
Subject Sample	# 1	0.109	05:36
Breath Volume		3.421	05:36
Air Blank		0.000	05:36
Air Blank		0.000	05:38
Subject Sample	# 2	0.112	05:39
Breath Volume		3.589	05:39
Air Blank		0.000	05:39
Control Test		0.080	05:39
Air Blank		0.000	05:40
Diagnostics Check		OK	05:40

Data Download Date/Time: 03/29/2019 19:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 03/03/2019 22:00:35

Diagnostics Check		OK	21:52
Air Blank		0.000	21:52
Control Test		0.079	21:53
Air Blank		0.000	21:53
Subject Sample	# 1	0.144	21:56
Breath Volume		3.382	21:56
Air Blank		0.000	21:56
Air Blank		0.000	21:58
Subject Sample	# 2	0.140	21:59
Breath Volume		3.218	21:59
Air Blank		0.000	21:59
Control Test		0.080	21:59
Air Blank		0.000	22:00
Diagnostics Check		OK	22:00

Data Download Date/Time: 03/29/2019 19:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DANIEL,K
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DAVIS,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: BOUTWELL

DATE/TIME OF TEST: 03/06/2019 00:39:52

Diagnostics Check		OK	00:33
Air Blank		0.000	00:34
Control Test		0.081	00:34
Air Blank		0.000	00:35
Subject Sample	# 1	0.000	00:35
Breath Volume		1.968	00:35
Air Blank		0.000	00:35
Air Blank		0.000	00:38
Subject Sample	# 2	0.000	00:38
Breath Volume		1.878	00:38
Air Blank		0.000	00:38
Control Test		0.080	00:39
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 03/29/2019 19:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUCE,HEATHER,N
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BRUCE,HEATHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 03/07/2019 03:14:39

Diagnostics Check		OK	03:11	
Air Blank		0.000	03:12	
Control Test		0.081	03:12	
Air Blank		0.000	03:13	
Subject Sample	# 1		03:13	Subject Test Refused
Breath Volume		0.000	03:13	
Air Blank		0.000	03:13	
Control Test		0.081	03:14	
Air Blank		0.000	03:14	
Diagnostics Check		OK	03:14	

Data Download Date/Time: 03/29/2019 19:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESOUSA,CARLOS,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DESOUSA,CARLOS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 03/09/2019 00:50:50

Diagnostics Check		OK	00:47	
Air Blank		0.000	00:48	
Control Test		0.081	00:48	
Air Blank		0.000	00:49	
Subject Sample	# 1		00:49	Subject Test Refused
Breath Volume		0.000	00:49	
Air Blank		0.000	00:49	
Control Test		0.080	00:50	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 03/29/2019 19:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOUGH,AARON,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GOUGH,AARON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 03/09/2019 02:01:05

Diagnostics Check		OK	01:49	
Air Blank		0.000	01:49	
Control Test		0.080	01:50	
Air Blank		0.000	01:50	
Subject Sample	# 1	0.165	01:53	Volume Not Met
Breath Volume		0.863	01:53	
Air Blank		0.000	01:54	
Air Blank		0.000	01:56	
Subject Sample	# 2	0.126	01:59	Volume Not Met
Breath Volume		0.609	01:59	
Air Blank		0.000	02:00	
Control Test		0.080	02:00	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 03/29/2019 19:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOUGH,AARON,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GOUGH,AARON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 03/09/2019 02:09:49

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.161	02:04
Breath Volume		1.816	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.157	02:08
Breath Volume		1.824	02:08
Air Blank		0.000	02:08
Control Test		0.079	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 03/29/2019 19:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,MICHAEL,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: RAMIREZ,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 11:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:10

SUBJECT'S LAST NAME: KOENIG

DATE/TIME OF TEST: 03/09/2019 12:38:20

Diagnostics Check		OK	12:31
Air Blank		0.000	12:32
Control Test		0.079	12:32
Air Blank		0.000	12:33
Subject Sample	# 1	0.274	12:33
Breath Volume		3.210	12:33
Air Blank		0.000	12:34
Air Blank		0.000	12:36
Subject Sample	# 2	0.275	12:36
Breath Volume		3.183	12:36
Air Blank		0.000	12:37
Control Test		0.078	12:37
Air Blank		0.000	12:38
Diagnostics Check		OK	12:38

Data Download Date/Time: 03/29/2019 19:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 03/09/2019 20:09:27

Diagnostics Check		OK	20:02
Air Blank		0.000	20:03
Control Test		0.080	20:03
Air Blank		0.000	20:04
Subject Sample	# 1	0.157	20:04
Breath Volume		1.347	20:04
Air Blank		0.000	20:05
Air Blank		0.000	20:07
Subject Sample	# 2	0.145	20:07
Breath Volume		1.203	20:07
Air Blank		0.000	20:08
Control Test		0.080	20:08
Air Blank		0.000	20:09
Diagnostics Check		OK	20:09

Data Download Date/Time: 03/29/2019 19:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUCE,HEATHER,N
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BRUCE,HEATHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: STARLING

DATE/TIME OF TEST: 03/10/2019 21:32:27

Diagnostics Check		OK	21:25
Air Blank		0.000	21:26
Control Test		0.080	21:26
Air Blank		0.000	21:26
Subject Sample	# 1	0.073	21:27
Breath Volume		1.683	21:27
Air Blank		0.000	21:28
Air Blank		0.000	21:29
Subject Sample	# 2	0.071	21:30
Breath Volume		1.476	21:30
Air Blank		0.000	21:31
Control Test		0.079	21:31
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 03/29/2019 19:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: PROB/PAROL
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: MONTERO

DATE/TIME OF TEST: 03/11/2019 01:03:15

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.137	00:58
Breath Volume		3.449	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.135	01:01
Breath Volume		3.105	01:01
Air Blank		0.000	01:02
Control Test		0.078	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 03/29/2019 19:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVOTH,DANIEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARVOTH,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 03/11/2019 01:52:30

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.079	01:46
Air Blank		0.000	01:46
Subject Sample	# 1	0.197	01:48
Breath Volume		2.898	01:48
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.194	01:51
Breath Volume		3.347	01:51
Air Blank		0.000	01:51
Control Test		0.079	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 03/29/2019 19:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOGGETTE,MASON,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DOGGETTE,MASON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: CARMICAL

DATE/TIME OF TEST: 03/13/2019 17:38:30

Diagnostics Check		OK	17:30
Air Blank		0.000	17:31
Control Test		0.079	17:31
Air Blank		0.000	17:32
Subject Sample	# 1	0.207	17:34
Breath Volume		2.398	17:34
Air Blank		0.000	17:34
Air Blank		0.000	17:36
Subject Sample	# 2	0.203	17:36
Breath Volume		1.542	17:36
Air Blank		0.000	17:37
Control Test		0.078	17:37
Air Blank		0.000	17:38
Diagnostics Check		OK	17:38

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAPMAN,MAVERICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: KUKLO

DATE/TIME OF TEST: 03/15/2019 00:20:09

Diagnostics Check		OK	00:13
Air Blank		0.000	00:13
Control Test		0.080	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.222	00:15
Breath Volume		2.843	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.224	00:18
Breath Volume		2.175	00:18
Air Blank		0.000	00:19
Control Test		0.078	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,COREY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BERRY,COREY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/15/2019 22:50:49

Diagnostics Check		OK	22:44
Air Blank		0.000	22:44
Control Test		0.080	22:45
Air Blank		0.000	22:45
Subject Sample	# 1	0.160	22:46
Breath Volume		2.035	22:46
Air Blank		0.000	22:47
Air Blank		0.000	22:48
Subject Sample	# 2	0.158	22:49
Breath Volume		1.625	22:49
Air Blank		0.000	22:49
Control Test		0.077	22:50
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: PERREAULT

DATE/TIME OF TEST: 03/15/2019 23:48:36

Diagnostics Check		OK	23:42
Air Blank		0.000	23:43
Control Test		0.080	23:43
Air Blank		0.000	23:43
Subject Sample	# 1	0.000	23:44
Breath Volume		3.886	23:44
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.000	23:47
Breath Volume		4.011	23:47
Air Blank		0.000	23:47
Control Test		0.079	23:48
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVIER,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LEVIER,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: BOHANNON

DATE/TIME OF TEST: 03/16/2019 02:00:33

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.080	01:55
Air Blank		0.000	01:55
Subject Sample	# 1	0.097	01:56
Breath Volume		3.863	01:56
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.097	01:59
Breath Volume		3.792	01:59
Air Blank		0.000	01:59
Control Test		0.079	01:59
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAPMAN,MAVERICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: FRANCO

DATE/TIME OF TEST: 03/17/2019 00:57:46

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.081	00:52
Air Blank		0.000	00:52
Subject Sample	# 1	0.116	00:53
Breath Volume		3.105	00:53
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.110	00:56
Breath Volume		2.285	00:56
Air Blank		0.000	00:56
Control Test		0.079	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUCHANAN,ASHLEY,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPEZ,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BUCHANAN

DATE/TIME OF TEST: 03/17/2019 01:41:52

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.079	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.196	01:37
Breath Volume		3.156	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:39
Subject Sample	# 2	0.199	01:40
Breath Volume		2.968	01:40
Air Blank		0.000	01:40
Control Test		0.078	01:41
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: BOUYOUKOS

DATE/TIME OF TEST: 03/17/2019 22:39:34

Diagnostics Check		OK	22:33
Air Blank		0.000	22:33
Control Test		0.081	22:34
Air Blank		0.000	22:34
Subject Sample	# 1	0.077	22:35
Breath Volume		1.527	22:35
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.071	22:38
Breath Volume		2.035	22:38
Air Blank		0.000	22:38
Control Test		0.079	22:38
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: REGIMBAL

DATE/TIME OF TEST: 03/18/2019 00:49:28

Diagnostics Check		OK	00:45	
Air Blank		0.000	00:46	
Control Test		0.081	00:46	
Air Blank		0.000	00:47	
Subject Sample	# 1		00:48	Subject Test Refused
Breath Volume		0.000	00:48	
Air Blank		0.000	00:48	
Control Test		0.080	00:48	
Air Blank		0.000	00:49	
Diagnostics Check		OK	00:49	

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUCE,HEATHER,N
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BURGETT,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 03/18/2019 03:14:37

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.000	03:10
Breath Volume		2.468	03:10
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.000	03:13
Breath Volume		1.320	03:13
Air Blank		0.000	03:13
Control Test		0.080	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEINER,ZACHARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: STEINER,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 03/21/2019 18:43:13

Diagnostics Check		OK	18:36
Air Blank		0.000	18:36
Control Test		0.081	18:37
Air Blank		0.000	18:37
Subject Sample	# 1	0.239	18:38
Breath Volume		1.828	18:38
Air Blank		0.000	18:38
Air Blank		0.000	18:40
Subject Sample	# 2	0.245	18:41
Breath Volume		2.757	18:41
Air Blank		0.000	18:42
Control Test		0.081	18:42
Air Blank		0.000	18:43
Diagnostics Check		OK	18:43

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVOTH,DANIEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARVOTH,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: HARDING

DATE/TIME OF TEST: 03/21/2019 21:37:45

Diagnostics Check		OK	21:30
Air Blank		0.000	21:30
Control Test		0.081	21:30
Air Blank		0.000	21:31
Subject Sample	# 1	0.213	21:33
Breath Volume		3.539	21:33
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 2	0.214	21:36
Breath Volume		4.191	21:36
Air Blank		0.000	21:36
Control Test		0.080	21:37
Air Blank		0.000	21:37
Diagnostics Check		OK	21:37

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPLING,JEFF,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: APPLING,JEFF
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: MIRRA

DATE/TIME OF TEST: 03/22/2019 23:22:54

Diagnostics Check		OK	23:14	
Air Blank		0.000	23:14	
Control Test		0.081	23:14	
Air Blank		0.000	23:15	
Subject Sample	# 1		23:18	No Sample Provided
Breath Volume		0.000	23:18	
Air Blank		0.000	23:18	
Air Blank		0.000	23:20	
Subject Sample	# 2		23:21	Subject Test Refused
Breath Volume		0.000	23:21	
Air Blank		0.000	23:21	
Control Test		0.080	23:22	
Air Blank		0.000	23:22	
Diagnostics Check		OK	23:22	

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,COREY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WEAVER,COREY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ARRIERO

DATE/TIME OF TEST: 03/23/2019 01:27:02

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.174	01:22
Breath Volume		1.417	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.175	01:25
Breath Volume		1.367	01:25
Air Blank		0.000	01:26
Control Test		0.080	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:27

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAPMAN,MAVERICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: STEINER

DATE/TIME OF TEST: 03/23/2019 01:58:18

Diagnostics Check		OK	01:49	
Air Blank		0.000	01:50	
Control Test		0.081	01:50	
Air Blank		0.000	01:51	
Subject Sample	# 1		01:54	No Sample Provided
Breath Volume		0.000	01:54	
Air Blank		0.000	01:54	
Air Blank		0.000	01:56	
Subject Sample	# 2		01:56	Subject Test Refused
Breath Volume		0.000	01:56	
Air Blank		0.000	01:57	
Control Test		0.081	01:57	
Air Blank		0.000	01:58	
Diagnostics Check		OK	01:58	

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUCE,HEATHER,N
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BRUCE,HEATHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: ADCOX

DATE/TIME OF TEST: 03/24/2019 00:08:20

Diagnostics Check		OK	00:01
Air Blank		0.000	00:02
Control Test		0.080	00:02
Air Blank		0.000	00:02
Subject Sample	# 1	0.185	00:03
Breath Volume		1.500	00:03
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.183	00:06
Breath Volume		2.488	00:06
Air Blank		0.000	00:07
Control Test		0.079	00:07
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: GOLSBY

DATE/TIME OF TEST: 03/24/2019 01:37:20

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:33	
Control Test		0.080	01:33	
Air Blank		0.000	01:34	
Subject Sample	# 1		01:35	RFI Detect
Breath Volume		0.000	01:35	
Air Blank		0.000	01:36	
Control Test		0.080	01:36	
Air Blank		0.000	01:37	
Diagnostics Check		OK	01:37	

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: GOLSBY

DATE/TIME OF TEST: 03/24/2019 01:48:13

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.080	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.149	01:43
Breath Volume		1.238	01:43
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.158	01:46
Breath Volume		1.554	01:46
Air Blank		0.000	01:47
Control Test		0.079	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DANNY,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MILLER,DANNY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: ASBLE

DATE/TIME OF TEST: 03/25/2019 00:10:50

Diagnostics Check		OK	00:07	
Air Blank		0.000	00:08	
Control Test		0.080	00:08	
Air Blank		0.000	00:09	
Subject Sample	# 1		00:09	Subject Test Refused
Breath Volume		0.000	00:09	
Air Blank		0.000	00:09	
Control Test		0.080	00:10	
Air Blank		0.000	00:10	
Diagnostics Check		OK	00:10	

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYD,JACOB,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOYD,JACOB
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: JERNIGAN

DATE/TIME OF TEST: 03/26/2019 21:29:38

Diagnostics Check		OK	21:22
Air Blank		0.000	21:23
Control Test		0.080	21:23
Air Blank		0.000	21:24
Subject Sample	# 1	0.050	21:24
Breath Volume		1.734	21:24
Air Blank		0.000	21:25
Air Blank		0.000	21:27
Subject Sample	# 2	0.045	21:28
Breath Volume		1.671	21:28
Air Blank		0.000	21:28
Control Test		0.080	21:29
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,PAUL,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: RODRIGUEZ,PAUL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/27/2019 02:13:36

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.080	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.110	02:09
Breath Volume		2.449	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.111	02:12
Breath Volume		2.304	02:12
Air Blank		0.000	02:12
Control Test		0.079	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 03/29/2019 19:16

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFITH,DALTON,M
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: GRIFFITH,DALTON
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/02/2019 23:07:47

Diagnostics Check		OK	23:04	
Air Blank		0.000	23:05	
Control Test		0.082	23:05	
Air Blank		0.000	23:05	
Subject Sample	# 1		23:06	Subject Test Refused
Breath Volume		0.000	23:06	
Air Blank		0.000	23:06	
Control Test		0.082	23:07	
Air Blank		0.000	23:07	
Diagnostics Check		OK	23:07	

Data Download Date/Time: 03/07/2019 12:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARTER,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: EDWARDS,CHEREE
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 10:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:04

SUBJECT'S LAST NAME: JENKINS JR

DATE/TIME OF TEST: 03/28/2019 11:40:31

Diagnostics Check		OK	11:37	
Air Blank		0.000	11:38	
Control Test		0.083	11:38	
Air Blank		0.000	11:38	
Subject Sample	# 1		11:39	Subject Test Refused
Breath Volume		0.000	11:39	
Air Blank		0.000	11:39	
Control Test		0.082	11:39	
Air Blank		0.000	11:40	
Diagnostics Check		OK	11:40	

Data Download Date/Time: 04/12/2019 11:20

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDYETTE,COREY,L
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: MIDYETTE,COREY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 13:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: PRESTON

DATE/TIME OF TEST: 03/28/2019 16:59:01

Diagnostics Check		OK	16:52
Air Blank		0.000	16:53
Control Test		0.080	16:53
Air Blank		0.000	16:53
Subject Sample	# 1	0.258	16:54
Breath Volume		3.042	16:54
Air Blank		0.000	16:55
Air Blank		0.000	16:56
Subject Sample	# 2	0.259	16:57
Breath Volume		3.183	16:57
Air Blank		0.000	16:58
Control Test		0.079	16:58
Air Blank		0.000	16:58
Diagnostics Check		OK	16:59

Data Download Date/Time: 05/01/2019 09:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,ERIC,D
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BROWN,ERIC
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 03/28/2019 21:52:05

Diagnostics Check		OK	21:45
Air Blank		0.000	21:46
Control Test		0.080	21:46
Air Blank		0.000	21:46
Subject Sample	# 1	0.144	21:47
Breath Volume		2.882	21:47
Air Blank		0.000	21:48
Air Blank		0.000	21:50
Subject Sample	# 2	0.143	21:50
Breath Volume		2.128	21:50
Air Blank		0.000	21:51
Control Test		0.081	21:51
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 05/01/2019 09:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: STEADMAN

DATE/TIME OF TEST: 03/29/2019 03:58:52

Diagnostics Check		OK	03:55	
Air Blank		0.000	03:56	
Control Test		0.082	03:56	
Air Blank		0.000	03:56	
Subject Sample	# 1		03:57	Subject Test Refused
Breath Volume		0.000	03:57	
Air Blank		0.000	03:57	
Control Test		0.080	03:58	
Air Blank		0.000	03:58	
Diagnostics Check		OK	03:58	

Data Download Date/Time: 05/01/2019 09:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG722402
DRY GAS EXPIRY DATE: 08/12/2019
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:44

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/31/2019 13:07:09

Diagnostics Check		OK	13:04	
Air Blank		0.000	13:04	
Control Test		0.080	13:04	
Air Blank		0.000	13:05	
Subject Sample	# 1		13:05	Subject Test Refused
Breath Volume		0.000	13:05	
Air Blank		0.000	13:06	
Control Test		0.080	13:06	
Air Blank		0.000	13:07	
Diagnostics Check		OK	13:07	

Data Download Date/Time: 05/01/2019 09:15

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL,CHAD,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MCGOWIN,JAKE
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: MAYNOR

DATE/TIME OF TEST: 03/01/2019 07:18:25

Diagnostics Check		OK	07:15	
Air Blank		0.000	07:16	
Control Test		0.079	07:16	
Air Blank		0.000	07:16	
Subject Sample	# 1		07:17	Subject Test Refused
Breath Volume		0.000	07:17	
Air Blank		0.000	07:17	
Control Test		0.080	07:17	
Air Blank		0.000	07:18	
Diagnostics Check		OK	07:18	

Data Download Date/Time: 03/27/2019 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: SNODDY

DATE/TIME OF TEST: 03/02/2019 03:23:34

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.080	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.148	03:18
Breath Volume		1.609	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.140	03:22
Breath Volume		2.691	03:22
Air Blank		0.000	03:22
Control Test		0.080	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 03/27/2019 11:45

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: URDANETA PEREZ

DATE/TIME OF TEST: 03/03/2019 03:59:13

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.079	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.144	03:54
Breath Volume		1.687	03:54
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.152	03:57
Breath Volume		1.929	03:57
Air Blank		0.000	03:58
Control Test		0.079	03:58
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 03/27/2019 11:45

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: PATISHTAN ARGUELLO

DATE/TIME OF TEST: 03/04/2019 02:41:34

Diagnostics Check		OK	02:34
Air Blank		0.000	02:34
Control Test		0.080	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.162	02:36
Breath Volume		1.949	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.175	02:40
Breath Volume		1.531	02:40
Air Blank		0.000	02:40
Control Test		0.079	02:40
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 03/27/2019 11:45

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTTER
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 03/04/2019 23:06:27

Diagnostics Check		OK	23:00
Air Blank		0.000	23:00
Control Test		0.081	23:00
Air Blank		0.000	23:01
Subject Sample	# 1	0.000	23:01
Breath Volume		3.128	23:01
Air Blank		0.000	23:02
Air Blank		0.000	23:04
Subject Sample	# 2	0.000	23:05
Breath Volume		3.234	23:05
Air Blank		0.000	23:05
Control Test		0.081	23:05
Air Blank		0.000	23:06
Diagnostics Check		OK	23:06

Data Download Date/Time: 03/27/2019 11:45

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAWYER,DARRYL,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAWYER,DARRYL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 03/05/2019 22:01:57

Diagnostics Check		OK	21:55
Air Blank		0.000	21:56
Control Test		0.081	21:56
Air Blank		0.000	21:57
Subject Sample	# 1	0.112	21:57
Breath Volume		2.968	21:57
Air Blank		0.000	21:58
Air Blank		0.000	21:59
Subject Sample	# 2	0.117	22:00
Breath Volume		2.816	22:00
Air Blank		0.000	22:01
Control Test		0.079	22:01
Air Blank		0.000	22:01
Diagnostics Check		OK	22:01

Data Download Date/Time: 03/27/2019 11:45

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSIUS,STANLEY,W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: JONES,D
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 03/06/2019 18:37:24

Diagnostics Check		OK	18:34	
Air Blank		0.000	18:35	
Control Test		0.081	18:35	
Air Blank		0.000	18:35	
Subject Sample	# 1		18:36	Subject Test Refused
Breath Volume		0.000	18:36	
Air Blank		0.000	18:36	
Control Test		0.081	18:36	
Air Blank		0.000	18:37	
Diagnostics Check		OK	18:37	

Data Download Date/Time: 03/27/2019 11:45

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CHRISTY,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: NELSON,RYAN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:56

SUBJECT'S LAST NAME: GENTRY

DATE/TIME OF TEST: 03/08/2019 12:22:52

Diagnostics Check		OK	12:16	
Air Blank		0.000	12:17	
Control Test		0.080	12:17	
Air Blank		0.000	12:18	
Subject Sample	# 1	0.239	12:18	
Breath Volume		2.453	12:18	
Air Blank		0.000	12:19	
Air Blank		0.000	12:21	
Subject Sample	# 2		12:21	Subject Test Refused
Breath Volume		0.000	12:21	
Air Blank		0.000	12:21	
Control Test		0.078	12:22	
Air Blank		0.000	12:22	
Diagnostics Check		OK	12:22	

Data Download Date/Time: 03/27/2019 11:45

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: PEREZ-HERNANDEZ

DATE/TIME OF TEST: 03/09/2019 01:27:20

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.157	01:22
Breath Volume		2.039	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.157	01:25
Breath Volume		2.441	01:25
Air Blank		0.000	01:26
Control Test		0.080	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 03/27/2019 11:45

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 03/09/2019 21:20:26

Diagnostics Check		OK	21:11	
Air Blank		0.000	21:12	
Control Test		0.080	21:12	
Air Blank		0.000	21:13	
Subject Sample	# 1		21:16	No Sample Provided
Breath Volume		0.000	21:16	
Air Blank		0.000	21:16	
Air Blank		0.000	21:18	
Subject Sample	# 2		21:19	Subject Test Refused
Breath Volume		0.000	21:19	
Air Blank		0.000	21:19	
Control Test		0.080	21:19	
Air Blank		0.000	21:20	
Diagnostics Check		OK	21:20	

Data Download Date/Time: 03/27/2019 11:45

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 03/10/2019 23:11:00

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:56	
Control Test		0.079	22:56	
Air Blank		0.000	22:57	
Subject Sample	# 1	0.171	22:58	
Breath Volume		1.566	22:58	
Air Blank		0.000	22:58	
Air Blank		0.000	23:00	
Subject Sample	# 2		23:03	No Sample Provided
Breath Volume		0.000	23:03	
Air Blank		0.000	23:04	
Air Blank		0.000	23:06	
Subject Sample	# 3		23:09	No Sample Provided
Breath Volume		0.000	23:09	
Air Blank		0.000	23:10	
Control Test		0.080	23:10	
Air Blank		0.000	23:10	
Diagnostics Check		OK	23:11	

Data Download Date/Time: 03/27/2019 11:45

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 03/10/2019 23:36:50

Diagnostics Check		OK	23:30
Air Blank		0.000	23:30
Control Test		0.080	23:31
Air Blank		0.000	23:31
Subject Sample	# 1	0.157	23:32
Breath Volume		1.312	23:32
Air Blank		0.000	23:32
Air Blank		0.000	23:34
Subject Sample	# 2	0.149	23:35
Breath Volume		1.160	23:35
Air Blank		0.000	23:35
Control Test		0.080	23:36
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 03/27/2019 11:45

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,JAMES,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: WALKER,JAMES
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MCNABB

DATE/TIME OF TEST: 03/14/2019 01:42:35

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:32	
Control Test		0.080	01:33	
Air Blank		0.000	01:33	
Subject Sample	# 1	0.192	01:36	Volume Not Met
Breath Volume		0.167	01:36	
Air Blank		0.000	01:37	
Air Blank		0.000	01:39	
Subject Sample	# 2		01:41	Subject Test Refused
Breath Volume		0.113	01:41	
Air Blank		0.000	01:41	
Control Test		0.079	01:41	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:42	

Data Download Date/Time: 03/27/2019 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: LOCKE

DATE/TIME OF TEST: 03/17/2019 01:26:52

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:15	
Control Test		0.081	01:16	
Air Blank		0.000	01:16	
Subject Sample	# 1	0.075	01:19	Volume Not Met
Breath Volume		0.625	01:19	
Air Blank		0.000	01:20	
Air Blank		0.000	01:22	
Subject Sample	# 2	0.074	01:25	Volume Not Met
Breath Volume		0.722	01:25	
Air Blank		0.000	01:25	
Control Test		0.080	01:26	
Air Blank		0.000	01:26	
Diagnostics Check		OK	01:26	

Data Download Date/Time: 03/27/2019 11:45

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FURMAN,FREDDIE,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 05:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: TOMASIEWICZ

DATE/TIME OF TEST: 03/17/2019 08:12:56

Diagnostics Check		OK	08:06
Air Blank		0.000	08:06
Control Test		0.080	08:06
Air Blank		0.000	08:07
Subject Sample	# 1	0.126	08:08
Breath Volume		2.304	08:08
Air Blank		0.000	08:08
Air Blank		0.000	08:10
Subject Sample	# 2	0.119	08:11
Breath Volume		1.593	08:11
Air Blank		0.000	08:12
Control Test		0.077	08:12
Air Blank		0.000	08:12
Diagnostics Check		OK	08:12

Data Download Date/Time: 03/27/2019 11:45

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD,MATTHEW,E
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: MAYFIELD,MATHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 10:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:25

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 03/20/2019 11:52:02

Diagnostics Check		OK	11:45
Air Blank		0.000	11:46
Control Test		0.081	11:46
Air Blank		0.000	11:47
Subject Sample	# 1	0.038	11:47
Breath Volume		2.957	11:47
Air Blank		0.000	11:48
Air Blank		0.000	11:50
Subject Sample	# 2	0.034	11:50
Breath Volume		1.734	11:50
Air Blank		0.000	11:51
Control Test		0.080	11:51
Air Blank		0.000	11:51
Diagnostics Check		OK	11:52

Data Download Date/Time: 03/27/2019 11:45

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,RYAN,D
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HART,HEATHER
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: CARMONA

DATE/TIME OF TEST: 03/24/2019 00:16:43

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.078	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.168	00:12
Breath Volume		1.414	00:12
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.175	00:15
Breath Volume		1.707	00:15
Air Blank		0.000	00:15
Control Test		0.079	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 03/27/2019 11:45

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRITT,DAKOTA,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MERRITT,DAKOTA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CARELLA

DATE/TIME OF TEST: 03/24/2019 00:51:27

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.081	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.234	00:46
Breath Volume		2.058	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.245	00:49
Breath Volume		2.757	00:49
Air Blank		0.000	00:50
Control Test		0.080	00:50
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 03/27/2019 11:45

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORDHAM,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: FORDHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 869968
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 14:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:12

SUBJECT'S LAST NAME: CLAYTON

DATE/TIME OF TEST: 03/15/2019 15:35:56

Diagnostics Check		OK	15:33	
Air Blank		0.000	15:33	
Control Test		0.080	15:33	
Air Blank		0.000	15:34	
Subject Sample	# 1		15:34	Subject Test Refused
Breath Volume		0.000	15:34	
Air Blank		0.000	15:35	
Control Test		0.080	15:35	
Air Blank		0.000	15:35	
Diagnostics Check		OK	15:35	

Data Download Date/Time: 03/31/2019 15:01

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINTOSH,ZANE,R
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: MCINTOSH,ZANE
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 171502
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/16/2019 23:30:12

Diagnostics Check		OK	23:27	
Air Blank		0.000	23:27	
Control Test		0.081	23:28	
Air Blank		0.000	23:28	
Subject Sample	# 1		23:28	Subject Test Refused
Breath Volume		0.000	23:28	
Air Blank		0.000	23:29	
Control Test		0.080	23:29	
Air Blank		0.000	23:30	
Diagnostics Check		OK	23:30	

Data Download Date/Time: 03/31/2019 15:01

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DAVID,O
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: WHITAKER,JOHNNY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 14:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:23

SUBJECT'S LAST NAME: GREMILLION

DATE/TIME OF TEST: 03/03/2019 14:58:22

Diagnostics Check		OK	14:51
Air Blank		0.000	14:52
Control Test		0.080	14:52
Air Blank		0.000	14:53
Subject Sample	# 1	0.201	14:53
Breath Volume		3.050	14:53
Air Blank		0.000	14:54
Air Blank		0.000	14:56
Subject Sample	# 2	0.198	14:56
Breath Volume		1.941	14:56
Air Blank		0.000	14:57
Control Test		0.079	14:57
Air Blank		0.000	14:58
Diagnostics Check		OK	14:58

Data Download Date/Time: 03/23/2019 10:46

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOBLES,DARIUS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NOBLES,DARIUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BOENNING

DATE/TIME OF TEST: 03/03/2019 23:29:10

Diagnostics Check		OK	23:23
Air Blank		0.000	23:23
Control Test		0.080	23:23
Air Blank		0.000	23:24
Subject Sample	# 1	0.023	23:24
Breath Volume		3.773	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.023	23:27
Breath Volume		3.800	23:27
Air Blank		0.000	23:28
Control Test		0.080	23:28
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 03/23/2019 10:46

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUNCEY,KATHLEEN,R
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: SCHOFIELD

DATE/TIME OF TEST: 03/10/2019 15:37:28

Diagnostics Check		OK	15:31
Air Blank		0.000	15:31
Control Test		0.080	15:32
Air Blank		0.000	15:32
Subject Sample	# 1	0.000	15:33
Breath Volume		3.722	15:33
Air Blank		0.000	15:33
Air Blank		0.000	15:35
Subject Sample	# 2	0.000	15:36
Breath Volume		3.832	15:36
Air Blank		0.000	15:36
Control Test		0.079	15:36
Air Blank		0.000	15:37
Diagnostics Check		OK	15:37

Data Download Date/Time: 03/23/2019 10:46

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON,STEVEN,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BOSCHERT,RYAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 03/22/2019 22:49:31

Diagnostics Check		OK	22:43
Air Blank		0.000	22:43
Control Test		0.080	22:44
Air Blank		0.000	22:44
Subject Sample	# 1	0.228	22:45
Breath Volume		2.839	22:45
Air Blank		0.000	22:45
Air Blank		0.000	22:47
Subject Sample	# 2	0.224	22:47
Breath Volume		2.437	22:47
Air Blank		0.000	22:48
Control Test		0.080	22:48
Air Blank		0.000	22:49
Diagnostics Check		OK	22:49

Data Download Date/Time: 03/23/2019 10:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCHULTZ,JEREMY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: YAHRAUS

DATE/TIME OF TEST: 03/07/2019 21:26:11

Diagnostics Check		OK	21:19
Air Blank		0.000	21:20
Control Test		0.080	21:20
Air Blank		0.000	21:20
Subject Sample	# 1	0.326	21:21
Breath Volume		1.546	21:21
Air Blank		0.000	21:22
Air Blank		0.000	21:24
Subject Sample	# 2	0.319	21:24
Breath Volume		1.468	21:24
Air Blank		0.000	21:25
Control Test		0.080	21:25
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 03/23/2019 10:40

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERY,JOHN,W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 03/10/2019 01:22:01

Diagnostics Check		OK	01:19	
Air Blank		0.000	01:19	
Control Test		0.080	01:19	
Air Blank		0.000	01:20	
Subject Sample	# 1		01:20	Subject Test Refused
Breath Volume		0.000	01:20	
Air Blank		0.000	01:21	
Control Test		0.081	01:21	
Air Blank		0.000	01:21	
Diagnostics Check		OK	01:22	

Data Download Date/Time: 03/23/2019 10:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: RICHTER

DATE/TIME OF TEST: 03/17/2019 22:56:53

Diagnostics Check		OK	22:39	
Air Blank		0.000	22:40	
Control Test		0.081	22:40	
Air Blank		0.000	22:41	
Subject Sample	# 1		22:44	No Sample Provided
Breath Volume		0.000	22:44	
Air Blank		0.000	22:44	
Air Blank		0.000	22:46	
Subject Sample	# 2	0.318	22:49	
Breath Volume		1.101	22:49	
Air Blank		0.000	22:50	
Air Blank		0.000	22:52	
Subject Sample	# 3		22:55	No Sample Provided
Breath Volume		0.000	22:55	
Air Blank		0.000	22:55	
Control Test		0.077	22:56	
Air Blank		0.000	22:56	
Diagnostics Check		OK	22:56	

Data Download Date/Time: 03/23/2019 10:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: COKONOUGHER,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 03/18/2019 00:39:06

Diagnostics Check		OK	00:27	
Air Blank		0.000	00:28	
Control Test		0.081	00:28	
Air Blank		0.000	00:28	
Subject Sample	# 1		00:32	No Sample Provided
Breath Volume		0.000	00:32	
Air Blank		0.000	00:32	
Air Blank		0.000	00:34	
Subject Sample	# 2		00:37	No Sample Provided
Breath Volume		0.000	00:37	
Air Blank		0.000	00:38	
Control Test		0.082	00:38	
Air Blank		0.000	00:39	
Diagnostics Check		OK	00:39	

Data Download Date/Time: 03/23/2019 10:40

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,CONNOR,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: VENTRY,GENE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 03/24/2019 21:57:03

Diagnostics Check		OK	21:51	
Air Blank		0.000	21:52	
Control Test		0.081	21:52	
Air Blank		0.000	21:53	
Subject Sample	# 1		21:55	Subject Test Refused
Breath Volume		0.000	21:55	
Air Blank		0.000	21:56	
Control Test		0.082	21:56	
Air Blank		0.000	21:56	
Diagnostics Check		OK	21:57	

Data Download Date/Time: 04/05/2019 10:18

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,B,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: COLEMAN,T
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 11:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:57

SUBJECT'S LAST NAME: TANN

DATE/TIME OF TEST: 03/25/2019 12:27:11

Diagnostics Check		OK	12:20
Air Blank		0.000	12:20
Control Test		0.080	12:21
Air Blank		0.000	12:21
Subject Sample	# 1	0.209	12:22
Breath Volume		1.503	12:22
Air Blank		0.000	12:22
Air Blank		0.000	12:24
Subject Sample	# 2	0.218	12:25
Breath Volume		1.843	12:25
Air Blank		0.000	12:26
Control Test		0.079	12:26
Air Blank		0.000	12:27
Diagnostics Check		OK	12:27

Data Download Date/Time: 04/05/2019 10:18

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: SCHUDDE

DATE/TIME OF TEST: 03/29/2019 21:14:31

Diagnostics Check		OK	21:00	
Air Blank		0.000	21:01	
Control Test		0.081	21:01	
Air Blank		0.000	21:02	
Subject Sample	# 1	0.207	21:05	Volume Not Met
Breath Volume		0.164	21:05	
Air Blank		0.000	21:05	
Air Blank		0.000	21:07	
Subject Sample	# 2	0.247	21:08	
Breath Volume		1.890	21:08	
Air Blank		0.000	21:09	
Air Blank		0.000	21:11	
Subject Sample	# 3	0.266	21:12	
Breath Volume		2.078	21:12	
Air Blank		0.000	21:13	
Control Test		0.079	21:13	
Air Blank		0.000	21:14	
Diagnostics Check		OK	21:14	

Data Download Date/Time: 04/05/2019 10:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,BRANDON,W
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VULETA,BEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 03/02/2019 00:30:55

Diagnostics Check		OK	00:24
Air Blank		0.000	00:25
Control Test		0.079	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.237	00:26
Breath Volume		1.796	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:28
Subject Sample	# 2	0.246	00:29
Breath Volume		2.789	00:29
Air Blank		0.000	00:30
Control Test		0.079	00:30
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 03/06/2019 10:35

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,BRANDON,W
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VULETA,BEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 03/02/2019 03:21:46

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.150	03:17
Breath Volume		1.281	03:17
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.142	03:20
Breath Volume		1.597	03:20
Air Blank		0.000	03:20
Control Test		0.079	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 03/06/2019 10:35

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: KINSER

DATE/TIME OF TEST: 03/07/2019 22:54:04

Diagnostics Check		OK	22:50	
Air Blank		0.000	22:50	
Control Test		0.080	22:51	
Air Blank		0.000	22:51	
Subject Sample	# 1		22:52	Subject Test Refused
Breath Volume		0.000	22:52	
Air Blank		0.000	22:53	
Control Test		0.081	22:53	
Air Blank		0.000	22:53	
Diagnostics Check		OK	22:54	

Data Download Date/Time: 04/15/2019 14:58

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,S
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BERGQUIST

DATE/TIME OF TEST: 03/09/2019 03:28:32

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:26	
Control Test		0.079	03:26	
Air Blank		0.000	03:27	
Subject Sample	# 1		03:27	Subject Test Refused
Breath Volume		0.000	03:27	
Air Blank		0.000	03:27	
Control Test		0.080	03:27	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 04/15/2019 14:58

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 03/16/2019 00:31:55

Diagnostics Check		OK	00:25
Air Blank		0.000	00:26
Control Test		0.079	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.195	00:27
Breath Volume		2.957	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.200	00:30
Breath Volume		3.031	00:30
Air Blank		0.000	00:31
Control Test		0.079	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 04/15/2019 14:58

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BILLIE

DATE/TIME OF TEST: 03/17/2019 02:59:45

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:57	
Control Test		0.080	02:57	
Air Blank		0.000	02:58	
Subject Sample	# 1		02:58	Subject Test Refused
Breath Volume		0.000	02:58	
Air Blank		0.000	02:58	
Control Test		0.079	02:59	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 04/15/2019 14:58

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CARL,C
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: ARMENTA

DATE/TIME OF TEST: 03/23/2019 03:09:14

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.080	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.259	03:04
Breath Volume		2.410	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.273	03:07
Breath Volume		3.792	03:07
Air Blank		0.000	03:08
Control Test		0.080	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 04/15/2019 14:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 03/23/2019 18:21:12

Diagnostics Check		OK	18:14
Air Blank		0.000	18:14
Control Test		0.080	18:15
Air Blank		0.000	18:15
Subject Sample	# 1	0.172	18:16
Breath Volume		2.351	18:16
Air Blank		0.000	18:17
Air Blank		0.000	18:19
Subject Sample	# 2	0.181	18:19
Breath Volume		2.496	18:19
Air Blank		0.000	18:20
Control Test		0.080	18:20
Air Blank		0.000	18:21
Diagnostics Check		OK	18:21

Data Download Date/Time: 04/15/2019 14:58

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: KINDLUND

DATE/TIME OF TEST: 03/23/2019 22:47:42

Diagnostics Check		OK	22:44	
Air Blank		0.000	22:45	
Control Test		0.080	22:45	
Air Blank		0.000	22:46	
Subject Sample	# 1		22:46	Subject Test Refused
Breath Volume		0.000	22:46	
Air Blank		0.000	22:46	
Control Test		0.080	22:47	
Air Blank		0.000	22:47	
Diagnostics Check		OK	22:47	

Data Download Date/Time: 04/15/2019 14:58

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DAVID,E
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: ROBLES-BRUNO

DATE/TIME OF TEST: 03/30/2019 03:06:16

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.079	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.235	03:01
Breath Volume		2.875	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.221	03:04
Breath Volume		1.160	03:04
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 04/15/2019 14:58

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMERO,CESAR,
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: ROMERO,CESAR
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 03/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: WELKER

DATE/TIME OF TEST: 03/06/2019 21:20:10

Diagnostics Check		OK	21:13
Air Blank		0.000	21:14
Control Test		0.078	21:14
Air Blank		0.000	21:15
Subject Sample	# 1	0.241	21:15
Breath Volume		4.320	21:15
Air Blank		0.000	21:16
Air Blank		0.000	21:18
Subject Sample	# 2	0.252	21:18
Breath Volume		2.156	21:18
Air Blank		0.000	21:19
Control Test		0.078	21:19
Air Blank		0.000	21:20
Diagnostics Check		OK	21:20

Data Download Date/Time: 04/16/2019 18:37

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: WILLIAMS,MICHAEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: POGGI

DATE/TIME OF TEST: 03/02/2019 22:13:15

Diagnostics Check		OK	22:05
Air Blank		0.000	22:05
Control Test		0.079	22:06
Air Blank		0.000	22:06
Subject Sample	# 1	0.238	22:07
Breath Volume		1.539	22:07
Air Blank		0.000	22:08
Air Blank		0.000	22:10
Subject Sample	# 2	0.234	22:11
Breath Volume		1.714	22:11
Air Blank		0.000	22:12
Control Test		0.078	22:12
Air Blank		0.000	22:13
Diagnostics Check		OK	22:13

Data Download Date/Time: 03/28/2019 18:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OJEDA,NELSON,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MCINERNEY,JESSE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 03/06/2019 02:47:52

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.080	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.186	02:42
Breath Volume		3.187	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.188	02:46
Breath Volume		3.117	02:46
Air Blank		0.000	02:46
Control Test		0.077	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 03/28/2019 18:50

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,IAN,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HARRIS,IAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:02

SUBJECT'S LAST NAME: SOMAL

DATE/TIME OF TEST: 03/09/2019 10:32:26

Diagnostics Check		OK	10:26
Air Blank		0.000	10:26
Control Test		0.081	10:27
Air Blank		0.000	10:27
Subject Sample	# 1	0.044	10:28
Breath Volume		1.648	10:28
Air Blank		0.000	10:28
Air Blank		0.000	10:30
Subject Sample	# 2	0.043	10:30
Breath Volume		1.945	10:30
Air Blank		0.000	10:31
Control Test		0.080	10:31
Air Blank		0.000	10:32
Diagnostics Check		OK	10:32

Data Download Date/Time: 03/28/2019 18:50

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEEFLANG,DANIEL,B
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MARREEL,SCOTT
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:01

SUBJECT'S LAST NAME: HYLTON

DATE/TIME OF TEST: 03/10/2019 08:31:23

Diagnostics Check		OK	08:25
Air Blank		0.000	08:25
Control Test		0.080	08:26
Air Blank		0.000	08:26
Subject Sample	# 1	0.249	08:26
Breath Volume		3.476	08:26
Air Blank		0.000	08:27
Air Blank		0.000	08:29
Subject Sample	# 2	0.235	08:29
Breath Volume		3.101	08:29
Air Blank		0.000	08:30
Control Test		0.077	08:30
Air Blank		0.000	08:31
Diagnostics Check		OK	08:31

Data Download Date/Time: 03/28/2019 18:50

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRING,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MCINERNEY,JESSE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 03/16/2019 23:55:31

Diagnostics Check		OK	23:49
Air Blank		0.000	23:49
Control Test		0.080	23:49
Air Blank		0.000	23:50
Subject Sample	# 1	0.106	23:51
Breath Volume		1.300	23:51
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.107	23:53
Breath Volume		1.578	23:53
Air Blank		0.000	23:54
Control Test		0.078	23:54
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 03/28/2019 18:50

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEEFLANG,DANIEL,B
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: GASPAR,BENJAMIN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: TOPPIN

DATE/TIME OF TEST: 03/17/2019 03:13:30

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.080	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.146	03:09
Breath Volume		2.609	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.141	03:11
Breath Volume		2.929	03:11
Air Blank		0.000	03:12
Control Test		0.078	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 03/28/2019 18:50

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREVISOL,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TREVISOL,JOSEPH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 03/20/2019 17:59:53

Diagnostics Check		OK	17:47	
Air Blank		0.000	17:48	
Control Test		0.080	17:48	
Air Blank		0.000	17:49	
Subject Sample	# 1	0.357	17:52	Volume Not Met
Breath Volume		0.054	17:52	
Air Blank		0.000	17:53	
Air Blank		0.000	17:54	
Subject Sample	# 2	0.381	17:55	
Breath Volume		1.773	17:55	
Air Blank		0.000	17:56	
Air Blank		0.000	17:57	
Subject Sample	# 3	0.388	17:58	
Breath Volume		1.484	17:58	
Air Blank		0.000	17:58	
Control Test		0.077	17:59	
Air Blank		0.000	17:59	
Diagnostics Check		OK	17:59	

Data Download Date/Time: 03/28/2019 18:50

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,BRIAN,P
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: GREEN,BRIAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: FISHMAN

DATE/TIME OF TEST: 03/05/2019 20:22:29

Diagnostics Check		OK	20:15
Air Blank		0.000	20:16
Control Test		0.080	20:16
Air Blank		0.000	20:17
Subject Sample	# 1	0.118	20:17
Breath Volume		1.128	20:17
Air Blank		0.000	20:18
Air Blank		0.000	20:19
Subject Sample	# 2	0.132	20:20
Breath Volume		2.359	20:20
Air Blank		0.000	20:21
Control Test		0.080	20:21
Air Blank		0.000	20:22
Diagnostics Check		OK	20:22

Data Download Date/Time: 04/12/2019 06:05

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 03/01/2019 21:19:20

Diagnostics Check		OK	21:08	
Air Blank		0.000	21:08	
Control Test		0.081	21:08	
Air Blank		0.000	21:09	
Subject Sample	# 1		21:12	No Sample Provided
Breath Volume		0.000	21:12	
Air Blank		0.000	21:12	
Air Blank		0.000	21:14	
Subject Sample	# 2		21:17	Subject Test Refused
Breath Volume		0.000	21:17	
Air Blank		0.000	21:18	
Control Test		0.083	21:18	
Air Blank		0.000	21:19	
Diagnostics Check		OK	21:19	

Data Download Date/Time: 03/04/2019 03:07

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDELL,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 05:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 03/02/2019 06:28:38

Diagnostics Check		OK	06:21
Air Blank		0.000	06:22
Control Test		0.082	06:22
Air Blank		0.000	06:22
Subject Sample	# 1	0.104	06:23
Breath Volume		3.113	06:23
Air Blank		0.000	06:24
Air Blank		0.000	06:26
Subject Sample	# 2	0.104	06:27
Breath Volume		2.949	06:27
Air Blank		0.000	06:27
Control Test		0.082	06:28
Air Blank		0.000	06:28
Diagnostics Check		OK	06:28

Data Download Date/Time: 03/04/2019 03:07

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FENTON,WILLIAM,B
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FENTON,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 14:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:46

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 03/02/2019 15:20:08

Diagnostics Check		OK	15:14
Air Blank		0.000	15:14
Control Test		0.081	15:14
Air Blank		0.000	15:15
Subject Sample	# 1	0.113	15:15
Breath Volume		1.894	15:15
Air Blank		0.000	15:16
Air Blank		0.000	15:18
Subject Sample	# 2	0.114	15:18
Breath Volume		2.761	15:18
Air Blank		0.000	15:19
Control Test		0.081	15:19
Air Blank		0.000	15:20
Diagnostics Check		OK	15:20

Data Download Date/Time: 03/04/2019 03:07

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,DAVID
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: HALACY

DATE/TIME OF TEST: 03/04/2019 23:07:47

Diagnostics Check		OK	22:58	
Air Blank		0.000	22:59	
Control Test		0.081	22:59	
Air Blank		0.000	23:00	
Subject Sample	# 1	0.285	23:01	
Breath Volume		1.121	23:01	
Air Blank		0.000	23:02	
Air Blank		0.000	23:04	
Subject Sample	# 2		23:06	Subject Test Refused
Breath Volume		0.285	23:06	
Air Blank		0.000	23:06	
Control Test		0.081	23:07	
Air Blank		0.000	23:07	
Diagnostics Check		OK	23:07	

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HOILETT,DWAYNE
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: LEADBETTER

DATE/TIME OF TEST: 03/06/2019 20:10:19

Diagnostics Check		OK	20:03
Air Blank		0.000	20:03
Control Test		0.083	20:03
Air Blank		0.000	20:04
Subject Sample	# 1	0.110	20:05
Breath Volume		2.855	20:05
Air Blank		0.000	20:06
Air Blank		0.000	20:08
Subject Sample	# 2	0.094	20:08
Breath Volume		2.707	20:08
Air Blank		0.000	20:09
Control Test		0.082	20:09
Air Blank		0.000	20:10
Diagnostics Check		OK	20:10

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUEHN,KYLE,B
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SCROGGIN,JOHN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 03/08/2019 00:34:19

Diagnostics Check		OK	00:28
Air Blank		0.000	00:28
Control Test		0.083	00:29
Air Blank		0.000	00:29
Subject Sample	# 1	0.000	00:30
Breath Volume		2.472	00:30
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.000	00:32
Breath Volume		1.765	00:32
Air Blank		0.000	00:33
Control Test		0.083	00:33
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 03/09/2019 03:13:32

Diagnostics Check		OK	03:06
Air Blank		0.000	03:07
Control Test		0.081	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.046	03:08
Breath Volume		1.199	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.051	03:12
Breath Volume		1.621	03:12
Air Blank		0.000	03:12
Control Test		0.082	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: QUINN,THOMAS
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: GODFREY

DATE/TIME OF TEST: 03/09/2019 20:34:16

Diagnostics Check		OK	20:31	
Air Blank		0.000	20:31	
Control Test		0.083	20:32	
Air Blank		0.000	20:32	
Subject Sample	# 1		20:32	Subject Test Refused
Breath Volume		0.000	20:32	
Air Blank		0.000	20:33	
Control Test		0.082	20:33	
Air Blank		0.000	20:34	
Diagnostics Check		OK	20:34	

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,OFC
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 03/09/2019 21:35:50

Diagnostics Check		OK	21:33	
Air Blank		0.000	21:33	
Control Test		0.083	21:33	
Air Blank		0.000	21:34	
Subject Sample	# 1		21:34	Subject Test Refused
Breath Volume		0.000	21:34	
Air Blank		0.000	21:34	
Control Test		0.083	21:35	
Air Blank		0.000	21:35	
Diagnostics Check		OK	21:35	

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BARKWELL,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: ANDERON

DATE/TIME OF TEST: 03/10/2019 07:55:50

Diagnostics Check		OK	07:41	
Air Blank		0.000	07:42	
Control Test		0.082	07:42	
Air Blank		0.000	07:43	
Subject Sample	# 1	0.150	07:46	Volume Not Met
Breath Volume		0.085	07:46	
Air Blank		0.000	07:46	
Air Blank		0.000	07:48	
Subject Sample	# 2	0.178	07:49	
Breath Volume		1.414	07:49	
Air Blank		0.000	07:50	
Air Blank		0.000	07:52	
Subject Sample	# 3	0.170	07:54	
Breath Volume		1.847	07:54	
Air Blank		0.000	07:54	
Control Test		0.080	07:55	
Air Blank		0.000	07:55	
Diagnostics Check		OK	07:55	

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FENTON,WILLIAM,B
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FENTON,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 13:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 03/11/2019 14:48:49

Diagnostics Check		OK	14:39
Air Blank		0.000	14:40
Control Test		0.082	14:40
Air Blank		0.000	14:41
Subject Sample	# 1	0.046	14:42
Breath Volume		1.136	14:42
Air Blank		0.000	14:43
Air Blank		0.000	14:44
Subject Sample	# 2	0.042	14:47
Breath Volume		1.414	14:47
Air Blank		0.000	14:47
Control Test		0.082	14:48
Air Blank		0.000	14:48
Diagnostics Check		OK	14:48

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,ALFRED,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: WHITTINGTON,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: PREDUE

DATE/TIME OF TEST: 03/11/2019 19:52:30

Diagnostics Check		OK	19:46
Air Blank		0.000	19:46
Control Test		0.082	19:46
Air Blank		0.000	19:47
Subject Sample	# 1	0.151	19:48
Breath Volume		2.050	19:48
Air Blank		0.000	19:48
Air Blank		0.000	19:50
Subject Sample	# 2	0.149	19:50
Breath Volume		1.351	19:50
Air Blank		0.000	19:51
Control Test		0.081	19:51
Air Blank		0.000	19:52
Diagnostics Check		OK	19:52

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FARLESS,G
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BOOKER

DATE/TIME OF TEST: 03/12/2019 01:21:04

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.082	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.224	01:16
Breath Volume		1.652	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:18
Subject Sample	# 2	0.218	01:19
Breath Volume		2.281	01:19
Air Blank		0.000	01:20
Control Test		0.080	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:21

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: JUMPING EAGLE

DATE/TIME OF TEST: 03/13/2019 23:09:12

Diagnostics Check		OK	23:06	
Air Blank		0.000	23:06	
Control Test		0.083	23:07	
Air Blank		0.000	23:07	
Subject Sample	# 1		23:07	Subject Test Refused
Breath Volume		0.000	23:07	
Air Blank		0.000	23:08	
Control Test		0.083	23:08	
Air Blank		0.000	23:09	
Diagnostics Check		OK	23:09	

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BARCARO,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: AGNEW

DATE/TIME OF TEST: 03/15/2019 00:56:38

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:52	
Control Test		0.081	00:53	
Air Blank		0.000	00:53	
Subject Sample	# 1		00:55	Subject Test Refused
Breath Volume		0.000	00:55	
Air Blank		0.000	00:55	
Control Test		0.082	00:56	
Air Blank		0.000	00:56	
Diagnostics Check		OK	00:56	

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKEON,A
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 03/17/2019 20:48:56

Diagnostics Check		OK	20:46	
Air Blank		0.000	20:46	
Control Test		0.082	20:46	
Air Blank		0.000	20:47	
Subject Sample	# 1		20:47	Subject Test Refused
Breath Volume		0.000	20:47	
Air Blank		0.000	20:48	
Control Test		0.082	20:48	
Air Blank		0.000	20:48	
Diagnostics Check		OK	20:48	

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEREZ,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: ORBEN

DATE/TIME OF TEST: 03/17/2019 22:15:47

Diagnostics Check		OK	22:09
Air Blank		0.000	22:10
Control Test		0.082	22:10
Air Blank		0.000	22:10
Subject Sample	# 1	0.177	22:11
Breath Volume		2.425	22:11
Air Blank		0.000	22:12
Air Blank		0.000	22:13
Subject Sample	# 2	0.170	22:14
Breath Volume		1.859	22:14
Air Blank		0.000	22:14
Control Test		0.081	22:15
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FARLESS,GARY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ALTO

DATE/TIME OF TEST: 03/19/2019 00:21:40

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.081	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.140	00:13
Breath Volume		1.769	00:13
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.165	00:16
Breath Volume		1.179	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 3	0.141	00:20
Breath Volume		2.167	00:20
Air Blank		0.000	00:20
Control Test		0.081	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: VALLAS,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BENTON

DATE/TIME OF TEST: 03/21/2019 02:44:31

Diagnostics Check		OK	02:41	
Air Blank		0.000	02:42	
Control Test		0.082	02:42	
Air Blank		0.000	02:42	
Subject Sample	# 1		02:43	Subject Test Refused
Breath Volume		0.000	02:43	
Air Blank		0.000	02:43	
Control Test		0.081	02:43	
Air Blank		0.000	02:44	
Diagnostics Check		OK	02:44	

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUEHN,KYLE,B
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: DOTY,JAMES
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: YORKER

DATE/TIME OF TEST: 03/21/2019 21:35:38

Diagnostics Check		OK	21:28
Air Blank		0.000	21:28
Control Test		0.082	21:29
Air Blank		0.000	21:29
Subject Sample	# 1	0.113	21:30
Breath Volume		1.296	21:30
Air Blank		0.000	21:31
Air Blank		0.000	21:32
Subject Sample	# 2	0.119	21:34
Breath Volume		1.640	21:34
Air Blank		0.000	21:34
Control Test		0.083	21:35
Air Blank		0.000	21:35
Diagnostics Check		OK	21:35

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: DYAL

DATE/TIME OF TEST: 03/22/2019 19:31:15

Diagnostics Check		OK	19:19	
Air Blank		0.000	19:20	
Control Test		0.082	19:20	
Air Blank		0.000	19:20	
Subject Sample	# 1	0.167	19:24	Volume Not Met
Breath Volume		0.957	19:24	
Air Blank		0.000	19:24	
Air Blank		0.000	19:26	
Subject Sample	# 2	0.171	19:29	Volume Not Met
Breath Volume		0.246	19:29	
Air Blank		0.000	19:30	
Control Test		0.080	19:30	
Air Blank		0.000	19:31	
Diagnostics Check		OK	19:31	

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONSECA,GIOVANNY,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: FONSECA,GIOVANNY
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 03/23/2019 00:45:20

Diagnostics Check		OK	00:36
Air Blank		0.000	00:37
Control Test		0.082	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.112	00:40
Breath Volume		2.035	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:42
Subject Sample	# 2	0.101	00:43
Breath Volume		1.484	00:43
Air Blank		0.000	00:44
Control Test		0.080	00:44
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: VALUS,ALEXIS
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: EDALATKHAH

DATE/TIME OF TEST: 03/23/2019 02:39:33

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:37	
Control Test		0.082	02:37	
Air Blank		0.000	02:38	
Subject Sample	# 1		02:38	Subject Test Refused
Breath Volume		0.000	02:38	
Air Blank		0.000	02:38	
Control Test		0.082	02:38	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGGERS,RYAN,K
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EGGERS,RYAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: STRIBLING

DATE/TIME OF TEST: 03/23/2019 05:00:58

Diagnostics Check		OK	04:57	
Air Blank		0.000	04:58	
Control Test		0.083	04:58	
Air Blank		0.000	04:59	
Subject Sample	# 1		04:59	Subject Test Refused
Breath Volume		0.000	04:59	
Air Blank		0.000	05:00	
Control Test		0.083	05:00	
Air Blank		0.000	05:00	
Diagnostics Check		OK	05:00	

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:04

SUBJECT'S LAST NAME: LAMOTTE

DATE/TIME OF TEST: 03/23/2019 06:41:15

Diagnostics Check		OK	06:33
Air Blank		0.000	06:33
Control Test		0.083	06:33
Air Blank		0.000	06:34
Subject Sample	# 1	0.217	06:36
Breath Volume		1.734	06:36
Air Blank		0.000	06:36
Air Blank		0.000	06:38
Subject Sample	# 2	0.206	06:39
Breath Volume		1.371	06:39
Air Blank		0.000	06:40
Control Test		0.082	06:40
Air Blank		0.000	06:41
Diagnostics Check		OK	06:41

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: VALUS,ALEXIS
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: DOSIO

DATE/TIME OF TEST: 03/24/2019 03:13:15

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:10	
Control Test		0.082	03:11	
Air Blank		0.000	03:11	
Subject Sample	# 1		03:11	Subject Test Refused
Breath Volume		0.000	03:11	
Air Blank		0.000	03:12	
Control Test		0.082	03:12	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HATT,CODY,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HATT,CODY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GARCIA LIRANZO

DATE/TIME OF TEST: 03/24/2019 05:08:28

Diagnostics Check		OK	05:02
Air Blank		0.000	05:02
Control Test		0.082	05:02
Air Blank		0.000	05:03
Subject Sample	# 1	0.129	05:03
Breath Volume		1.375	05:03
Air Blank		0.000	05:04
Air Blank		0.000	05:06
Subject Sample	# 2	0.130	05:06
Breath Volume		1.531	05:06
Air Blank		0.000	05:07
Control Test		0.082	05:07
Air Blank		0.000	05:08
Diagnostics Check		OK	05:08

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: VILLIER,F
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 03/24/2019 23:13:00

Diagnostics Check		OK	23:06
Air Blank		0.000	23:06
Control Test		0.082	23:07
Air Blank		0.000	23:07
Subject Sample	# 1	0.195	23:08
Breath Volume		1.187	23:08
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.186	23:11
Breath Volume		1.625	23:11
Air Blank		0.000	23:12
Control Test		0.082	23:12
Air Blank		0.000	23:12
Diagnostics Check		OK	23:13

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: KLECKNER

DATE/TIME OF TEST: 03/25/2019 04:06:17

Diagnostics Check		OK	03:58
Air Blank		0.000	03:59
Control Test		0.082	03:59
Air Blank		0.000	04:00
Subject Sample	# 1	0.186	04:01
Breath Volume		1.906	04:01
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.189	04:04
Breath Volume		1.753	04:04
Air Blank		0.000	04:05
Control Test		0.082	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: KALISZ

DATE/TIME OF TEST: 03/29/2019 04:05:38

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:01	
Control Test		0.082	04:02	
Air Blank		0.000	04:02	
Subject Sample	# 1		04:04	Subject Test Refused
Breath Volume		0.000	04:04	
Air Blank		0.000	04:04	
Control Test		0.083	04:05	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:05	

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: WHITTINGTON,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: KREISEL

DATE/TIME OF TEST: 03/30/2019 00:38:36

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.082	00:31
Air Blank		0.000	00:32
Subject Sample	# 1	0.243	00:33
Breath Volume		1.660	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.229	00:37
Breath Volume		1.519	00:37
Air Blank		0.000	00:37
Control Test		0.081	00:38
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: STOREY

DATE/TIME OF TEST: 03/30/2019 04:01:48

Diagnostics Check		OK	03:58	
Air Blank		0.000	03:58	
Control Test		0.082	03:58	
Air Blank		0.000	03:59	
Subject Sample	# 1		04:00	Subject Test Refused
Breath Volume		0.000	04:00	
Air Blank		0.000	04:00	
Control Test		0.082	04:01	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRANDES,M
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: CORLETT

DATE/TIME OF TEST: 03/30/2019 22:30:23

Diagnostics Check		OK	22:27	
Air Blank		0.000	22:28	
Control Test		0.081	22:28	
Air Blank		0.000	22:28	
Subject Sample	# 1		22:29	Subject Test Refused
Breath Volume		0.000	22:29	
Air Blank		0.000	22:29	
Control Test		0.082	22:29	
Air Blank		0.000	22:30	
Diagnostics Check		OK	22:30	

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOMOZA,EDUARDO,A
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SOMOZA,EDUARDO
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: TANNER

DATE/TIME OF TEST: 03/30/2019 23:08:14

Diagnostics Check		OK	23:02	
Air Blank		0.000	23:03	
Control Test		0.081	23:03	
Air Blank		0.000	23:03	
Subject Sample	# 1		23:07	No Sample Provided
Breath Volume		0.000	23:07	
Air Blank		0.000	23:07	
Air Blank		0.000	23:08	

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOMOZA,EDUARDO,A
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SOMOZA,EDUARDO
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: GAY

DATE/TIME OF TEST: 03/30/2019 23:18:59

Diagnostics Check		OK	23:15	
Air Blank		0.000	23:15	
Control Test		0.082	23:16	
Air Blank		0.000	23:16	
Subject Sample	# 1		23:17	Subject Test Refused
Breath Volume		0.000	23:17	
Air Blank		0.000	23:18	
Control Test		0.082	23:18	
Air Blank		0.000	23:18	
Diagnostics Check		OK	23:18	

Data Download Date/Time: 04/12/2019 06:02

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JOHN,P
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: COX,RILEY
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 02/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: MEZA-GONZALEZ

DATE/TIME OF TEST: 03/03/2019 03:53:13

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:50	
Control Test		0.081	03:51	
Air Blank		0.000	03:51	
Subject Sample	# 1		03:51	Subject Test Refused
Breath Volume		0.000	03:51	
Air Blank		0.000	03:52	
Control Test		0.082	03:52	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 03/09/2019 22:53

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,JOSIE,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WELLS,JOSIE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 03/07/2019 03:13:50

Diagnostics Check		OK	03:04	
Air Blank		0.000	03:05	
Control Test		0.081	03:05	
Air Blank		0.000	03:06	
Subject Sample	# 1	0.102	03:09	Volume Not Met
Breath Volume		0.066	03:09	
Air Blank		0.000	03:09	
Air Blank		0.000	03:11	
Subject Sample	# 2		03:12	Subject Test Refused
Breath Volume		0.097	03:12	
Air Blank		0.000	03:12	
Control Test		0.081	03:13	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 03/09/2019 22:53

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JOHN,P
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: PAEPLOW,KYLE
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 03/09/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: LATTIMORE

DATE/TIME OF TEST: 03/31/2019 00:32:15

Diagnostics Check		OK	00:25
Air Blank		0.000	00:26
Control Test		0.081	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.115	00:27
Breath Volume		1.644	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.114	00:30
Breath Volume		1.558	00:30
Air Blank		0.000	00:31
Control Test		0.081	00:31
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 04/23/2019 15:18

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDOREK,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: OSTROSKI,OFC
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 03/01/2019 22:15:37

Diagnostics Check		OK	22:12	
Air Blank		0.000	22:13	
Control Test		0.079	22:13	
Air Blank		0.000	22:14	
Subject Sample	# 1		22:14	Subject Test Refused
Breath Volume		0.000	22:14	
Air Blank		0.000	22:14	
Control Test		0.079	22:15	
Air Blank		0.000	22:15	
Diagnostics Check		OK	22:15	

Data Download Date/Time: 03/27/2019 08:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDOREK,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: YOCK,GEORGE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: WITKOWICKI

DATE/TIME OF TEST: 03/02/2019 01:00:02

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.079	00:54
Air Blank		0.000	00:54
Subject Sample	# 1	0.119	00:55
Breath Volume		1.988	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:57
Subject Sample	# 2	0.124	00:58
Breath Volume		2.148	00:58
Air Blank		0.000	00:59
Control Test		0.078	00:59
Air Blank		0.000	00:59
Diagnostics Check		OK	01:00

Data Download Date/Time: 03/27/2019 08:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDOREK,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PARIS,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: HORAN

DATE/TIME OF TEST: 03/02/2019 03:13:42

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.143	03:08
Breath Volume		2.089	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.134	03:12
Breath Volume		1.261	03:12
Air Blank		0.000	03:12
Control Test		0.078	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 03/27/2019 08:13

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUBRANO,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: COKER,OFC
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 03/02/2019 06:20:48

Diagnostics Check		OK	06:14
Air Blank		0.000	06:14
Control Test		0.079	06:14
Air Blank		0.000	06:15
Subject Sample	# 1	0.095	06:16
Breath Volume		3.222	06:16
Air Blank		0.000	06:16
Air Blank		0.000	06:18
Subject Sample	# 2	0.083	06:19
Breath Volume		1.339	06:19
Air Blank		0.000	06:19
Control Test		0.078	06:20
Air Blank		0.000	06:20
Diagnostics Check		OK	06:20

Data Download Date/Time: 03/27/2019 08:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRITCHARD,DEP,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: VLECK,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 03/03/2019 16:47:05

Diagnostics Check		OK	16:40
Air Blank		0.000	16:41
Control Test		0.079	16:41
Air Blank		0.000	16:42
Subject Sample	# 1	0.203	16:42
Breath Volume		2.218	16:42
Air Blank		0.000	16:43
Air Blank		0.000	16:45
Subject Sample	# 2	0.200	16:45
Breath Volume		1.421	16:45
Air Blank		0.000	16:46
Control Test		0.078	16:46
Air Blank		0.000	16:46
Diagnostics Check		OK	16:47

Data Download Date/Time: 03/27/2019 08:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERMILLION,OFFICER,B
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: VERMILLION,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: UNKNOWN

DATE/TIME OF TEST: 03/03/2019 23:00:41

Diagnostics Check		OK	22:52
Air Blank		0.000	22:52
Control Test		0.080	22:53
Air Blank		0.000	22:53
Subject Sample	# 1	0.229	22:55
Breath Volume		1.621	22:55
Air Blank		0.000	22:56
Air Blank		0.000	22:57
Subject Sample	# 2	0.242	22:58
Breath Volume		1.695	22:58
Air Blank		0.000	22:59
Control Test		0.075	23:00
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 03/27/2019 08:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDOREK,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: THIESEN,FRED
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: BUFFINGTON

DATE/TIME OF TEST: 03/04/2019 01:52:54

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.079	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.209	01:47
Breath Volume		1.949	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.217	01:51
Breath Volume		1.601	01:51
Air Blank		0.000	01:51
Control Test		0.077	01:52
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 03/27/2019 08:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,OFC,W
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: PERKINS,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: FORSON

DATE/TIME OF TEST: 03/05/2019 23:06:12

Diagnostics Check		OK	22:59
Air Blank		0.000	22:59
Control Test		0.080	22:59
Air Blank		0.000	23:00
Subject Sample	# 1	0.238	23:01
Breath Volume		2.144	23:01
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.247	23:04
Breath Volume		2.183	23:04
Air Blank		0.000	23:05
Control Test		0.079	23:05
Air Blank		0.000	23:06
Diagnostics Check		OK	23:06

Data Download Date/Time: 03/27/2019 08:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLYNN,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SHORT,S
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 03/06/2019 00:47:41

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.080	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.068	00:41
Breath Volume		1.808	00:41
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.072	00:45
Breath Volume		1.777	00:45
Air Blank		0.000	00:46
Control Test		0.078	00:47
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 03/27/2019 08:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: REVATTA HUAMAN

DATE/TIME OF TEST: 03/06/2019 02:47:22

Diagnostics Check		OK	02:41	
Air Blank		0.000	02:42	
Control Test		0.078	02:42	
Air Blank		0.000	02:43	
Subject Sample	# 1		02:46	Subject Test Refused
Breath Volume		0.000	02:46	
Air Blank		0.000	02:46	
Control Test		0.079	02:46	
Air Blank		0.000	02:47	
Diagnostics Check		OK	02:47	

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FELTOVIC,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FELTOVIC,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: CECIL

DATE/TIME OF TEST: 03/07/2019 12:19:52

Diagnostics Check		OK	12:13
Air Blank		0.000	12:14
Control Test		0.079	12:14
Air Blank		0.000	12:15
Subject Sample	# 1	0.219	12:15
Breath Volume		3.101	12:15
Air Blank		0.000	12:16
Air Blank		0.000	12:17
Subject Sample	# 2	0.222	12:18
Breath Volume		3.070	12:18
Air Blank		0.000	12:18
Control Test		0.079	12:19
Air Blank		0.000	12:19
Diagnostics Check		OK	12:19

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:21

SUBJECT'S LAST NAME: JIMENEZ SOTO

DATE/TIME OF TEST: 03/08/2019 18:51:40

Diagnostics Check		OK	18:43
Air Blank		0.000	18:44
Control Test		0.078	18:44
Air Blank		0.000	18:45
Subject Sample	# 1	0.224	18:46
Breath Volume		2.730	18:46
Air Blank		0.000	18:47
Air Blank		0.000	18:49
Subject Sample	# 2	0.226	18:49
Breath Volume		1.980	18:49
Air Blank		0.000	18:50
Control Test		0.076	18:51
Air Blank		0.000	18:51
Diagnostics Check		OK	18:51

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CHRISTIE,J
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: JENSEN

DATE/TIME OF TEST: 03/08/2019 21:42:59

Diagnostics Check		OK	21:35
Air Blank		0.000	21:36
Control Test		0.079	21:36
Air Blank		0.000	21:37
Subject Sample	# 1	0.254	21:37
Breath Volume		1.632	21:37
Air Blank		0.000	21:38
Air Blank		0.000	21:40
Subject Sample	# 2	0.245	21:41
Breath Volume		1.203	21:41
Air Blank		0.000	21:42
Control Test		0.077	21:42
Air Blank		0.000	21:42
Diagnostics Check		OK	21:42

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORD,JOSHUA,N
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LORD,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: HALE

DATE/TIME OF TEST: 03/09/2019 03:33:19

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.079	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.135	03:28
Breath Volume		2.269	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.133	03:31
Breath Volume		2.414	03:31
Air Blank		0.000	03:32
Control Test		0.078	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,STEFANIE,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PETERSON,STEFANIE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: CASHMORE

DATE/TIME OF TEST: 03/11/2019 00:38:05

Diagnostics Check		OK	00:31
Air Blank		0.000	00:32
Control Test		0.080	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.259	00:33
Breath Volume		2.675	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:35
Subject Sample	# 2	0.257	00:36
Breath Volume		2.125	00:36
Air Blank		0.000	00:37
Control Test		0.078	00:37
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SHORT,OFFICER
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: GUENTHER

DATE/TIME OF TEST: 03/11/2019 01:04:03

Diagnostics Check		OK	00:57
Air Blank		0.000	00:58
Control Test		0.079	00:58
Air Blank		0.000	00:58
Subject Sample	# 1	0.203	00:59
Breath Volume		2.386	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:01
Subject Sample	# 2	0.198	01:02
Breath Volume		2.421	01:02
Air Blank		0.000	01:03
Control Test		0.078	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:04

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERR,CALEB,W
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: KERR,CALEB
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: CURLES

DATE/TIME OF TEST: 03/11/2019 23:38:40

Diagnostics Check		OK	23:32
Air Blank		0.000	23:32
Control Test		0.080	23:33
Air Blank		0.000	23:33
Subject Sample	# 1	0.103	23:34
Breath Volume		1.847	23:34
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.104	23:37
Breath Volume		1.519	23:37
Air Blank		0.000	23:37
Control Test		0.079	23:38
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALLUCCA,MARIO,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WATSON,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: MARQUEZ CRUZ

DATE/TIME OF TEST: 03/14/2019 21:08:16

Diagnostics Check		OK	21:00
Air Blank		0.000	21:01
Control Test		0.079	21:01
Air Blank		0.000	21:02
Subject Sample	# 1	0.106	21:03
Breath Volume		2.621	21:03
Air Blank		0.000	21:04
Air Blank		0.000	21:06
Subject Sample	# 2	0.098	21:06
Breath Volume		2.515	21:06
Air Blank		0.000	21:07
Control Test		0.077	21:07
Air Blank		0.000	21:08
Diagnostics Check		OK	21:08

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,T
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 03/16/2019 01:21:36

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.080	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.145	01:16
Breath Volume		4.585	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.146	01:19
Breath Volume		4.347	01:19
Air Blank		0.000	01:20
Control Test		0.079	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERR,CALEB,W
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: KERR,CALEB
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CONNORS

DATE/TIME OF TEST: 03/16/2019 03:32:10

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.078	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.128	03:27
Breath Volume		2.656	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:29
Subject Sample	# 2	0.133	03:30
Breath Volume		3.222	03:30
Air Blank		0.000	03:31
Control Test		0.078	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLOESCU,A
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: MENDOZA-SALAZAR

DATE/TIME OF TEST: 03/16/2019 05:04:37

Diagnostics Check		OK	04:57
Air Blank		0.000	04:58
Control Test		0.079	04:58
Air Blank		0.000	04:59
Subject Sample	# 1	0.106	04:59
Breath Volume		3.257	04:59
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.107	05:03
Breath Volume		2.800	05:03
Air Blank		0.000	05:03
Control Test		0.078	05:04
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,SCOTT,H
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,AARON
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BRACKINS

DATE/TIME OF TEST: 03/17/2019 00:11:20

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.079	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.000	00:07
Breath Volume		2.902	00:07
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.000	00:09
Breath Volume		2.945	00:09
Air Blank		0.000	00:10
Control Test		0.078	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICK,A.,R
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: 6L13,6L13
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: NICK

DATE/TIME OF TEST: 03/18/2019 00:02:05

Diagnostics Check		OK	23:55
Air Blank		0.000	23:56
Control Test		0.079	23:56
Air Blank		0.000	23:57
Subject Sample	# 2	0.000	23:57
Breath Volume		4.367	23:57
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.000	00:00
Breath Volume		2.488	00:00
Air Blank		0.000	00:01
Control Test		0.078	00:01
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRELSFORD,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: WILKE

DATE/TIME OF TEST: 03/18/2019 16:21:42

Diagnostics Check		OK	16:15
Air Blank		0.000	16:15
Control Test		0.079	16:16
Air Blank		0.000	16:16
Subject Sample	# 1	0.254	16:17
Breath Volume		2.050	16:17
Air Blank		0.000	16:17
Air Blank		0.000	16:19
Subject Sample	# 2	0.246	16:19
Breath Volume		1.937	16:19
Air Blank		0.000	16:20
Control Test		0.078	16:21
Air Blank		0.000	16:21
Diagnostics Check		OK	16:21

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTWELL,J,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CANTWELL,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GREGOIRE

DATE/TIME OF TEST: 03/20/2019 04:15:30

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:12	
Control Test		0.079	04:12	
Air Blank		0.000	04:13	
Subject Sample	# 1	0.237	04:13	
Breath Volume		3.976	04:13	
Air Blank			04:13	RFI Detect
Air Blank		0.000	04:14	
Control Test		0.079	04:14	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:15	

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTWELL,J,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CANTWELL,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GREGOIRE

DATE/TIME OF TEST: 03/20/2019 04:24:35

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.079	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.230	04:19
Breath Volume		4.691	04:19
Air Blank		0.000	04:20
Air Blank		0.000	04:22
Subject Sample	# 2	0.241	04:22
Breath Volume		4.242	04:22
Air Blank		0.000	04:23
Control Test		0.077	04:23
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICK,A,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,A.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 03/20/2019 18:12:39

Diagnostics Check		OK	18:06
Air Blank		0.000	18:06
Control Test		0.079	18:07
Air Blank		0.000	18:07
Subject Sample	# 1	0.000	18:08
Breath Volume		4.832	18:08
Air Blank		0.000	18:08
Air Blank		0.000	18:10
Subject Sample	# 2	0.000	18:11
Breath Volume		4.332	18:11
Air Blank		0.000	18:11
Control Test		0.077	18:12
Air Blank		0.000	18:12
Diagnostics Check		OK	18:12

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,L
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 03/23/2019 04:07:36

Diagnostics Check		OK	04:04	
Air Blank		0.000	04:05	
Control Test		0.080	04:05	
Air Blank		0.000	04:06	
Subject Sample	# 1		04:06	Subject Test Refused
Breath Volume		0.000	04:06	
Air Blank		0.000	04:06	
Control Test		0.079	04:07	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: TIRADO

DATE/TIME OF TEST: 03/24/2019 02:48:46

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:34	
Control Test		0.080	02:35	
Air Blank		0.000	02:35	
Subject Sample	# 1	0.202	02:37	
Breath Volume		1.664	02:37	
Air Blank		0.000	02:38	
Air Blank		0.000	02:39	
Subject Sample	# 2	0.178	02:43	Volume Not Met
Breath Volume		0.128	02:43	
Air Blank		0.000	02:43	
Air Blank		0.000	02:45	
Subject Sample	# 3	0.189	02:47	
Breath Volume		1.464	02:47	
Air Blank		0.000	02:47	
Control Test		0.075	02:48	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:48	

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DREZNIN,M
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: CHARETTE

DATE/TIME OF TEST: 03/25/2019 20:43:45

Diagnostics Check		OK	20:37
Air Blank		0.000	20:37
Control Test		0.079	20:38
Air Blank		0.000	20:38
Subject Sample	# 1	0.000	20:39
Breath Volume		2.300	20:39
Air Blank		0.000	20:39
Air Blank		0.000	20:41
Subject Sample	# 2	0.000	20:42
Breath Volume		2.765	20:42
Air Blank		0.000	20:42
Control Test		0.079	20:43
Air Blank		0.000	20:43
Diagnostics Check		OK	20:43

Data Download Date/Time: 03/27/2019 08:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: JERRELLS

DATE/TIME OF TEST: 03/26/2019 21:06:31

Diagnostics Check		OK	21:00
Air Blank		0.000	21:00
Control Test		0.079	21:00
Air Blank		0.000	21:01
Subject Sample	# 1	0.108	21:01
Breath Volume		2.445	21:01
Air Blank		0.000	21:02
Air Blank		0.000	21:04
Subject Sample	# 2	0.109	21:04
Breath Volume		1.742	21:04
Air Blank		0.000	21:05
Control Test		0.078	21:05
Air Blank		0.000	21:06
Diagnostics Check		OK	21:06

Data Download Date/Time: 03/27/2019 08:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEYMAN,LUKE,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: HEYMAN,LUKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:04

SUBJECT'S LAST NAME: GIULIANI

DATE/TIME OF TEST: 03/28/2019 15:31:50

Diagnostics Check		OK	15:24
Air Blank		0.000	15:25
Control Test		0.079	15:25
Air Blank		0.000	15:26
Subject Sample	# 1	0.197	15:27
Breath Volume		1.367	15:27
Air Blank		0.000	15:28
Air Blank		0.000	15:29
Subject Sample	# 2	0.197	15:30
Breath Volume		1.468	15:30
Air Blank		0.000	15:30
Control Test		0.076	15:31
Air Blank		0.000	15:31
Diagnostics Check		OK	15:31

Data Download Date/Time: 04/16/2019 08:09

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: KNEBL

DATE/TIME OF TEST: 03/30/2019 02:30:58

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.181	02:26
Breath Volume		1.273	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:28
Subject Sample	# 2	0.183	02:29
Breath Volume		1.691	02:29
Air Blank		0.000	02:30
Control Test		0.077	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 04/16/2019 08:09

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GUZMANNERI

DATE/TIME OF TEST: 03/30/2019 20:12:08

Diagnostics Check		OK	20:05	
Air Blank		0.000	20:06	
Control Test		0.079	20:06	
Air Blank		0.000	20:07	
Subject Sample	# 1	0.284	20:08	
Breath Volume		1.585	20:08	
Air Blank		0.000	20:08	
Air Blank		0.000	20:10	
Subject Sample	# 2		20:10	RFI Detect
Breath Volume		0.000	20:10	
Air Blank		0.000	20:11	
Control Test		0.077	20:11	
Air Blank		0.000	20:12	
Diagnostics Check		OK	20:12	

Data Download Date/Time: 04/16/2019 08:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GUZMANNERI

DATE/TIME OF TEST: 03/30/2019 20:22:53

Diagnostics Check		OK	20:14	
Air Blank		0.000	20:14	
Control Test		0.078	20:15	
Air Blank		0.000	20:15	
Subject Sample	# 1	0.279	20:17	
Breath Volume		2.652	20:17	
Air Blank		0.000	20:18	
Air Blank		0.000	20:19	
Subject Sample	# 2	0.290	20:21	
Breath Volume		1.449	20:21	
Air Blank		0.000	20:21	
Control Test		0.074	20:22	Control Outside Tolerance
Air Blank		0.000	20:22	

Data Download Date/Time: 04/16/2019 08:09

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A2
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GUZMANNERI

DATE/TIME OF TEST: 03/30/2019 20:27:20

Diagnostics Check	OK	20:25	
Air Blank	0.000	20:26	
Control Test	0.072	20:26	Control Outside Tolerance
Air Blank	0.000	20:27	

Data Download Date/Time: 04/16/2019 08:09

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GUZMAN NERI

DATE/TIME OF TEST: 03/30/2019 20:56:30

Diagnostics Check		OK	20:49
Air Blank		0.000	20:49
Control Test		0.080	20:49
Air Blank		0.000	20:50
Subject Sample	# 1	0.275	20:51
Breath Volume		2.292	20:51
Air Blank		0.000	20:52
Air Blank		0.000	20:53
Subject Sample	# 2	0.281	20:54
Breath Volume		2.183	20:54
Air Blank		0.000	20:55
Control Test		0.077	20:55
Air Blank		0.000	20:56
Diagnostics Check		OK	20:56

Data Download Date/Time: 04/16/2019 08:09

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: COLLINS,R
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: PETTIES

DATE/TIME OF TEST: 03/31/2019 03:58:58

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.080	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.215	03:54
Breath Volume		2.257	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:56
Subject Sample	# 2	0.200	03:57
Breath Volume		1.414	03:57
Air Blank		0.000	03:58
Control Test		0.077	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 04/16/2019 08:09

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: DAWKINS

DATE/TIME OF TEST: 03/31/2019 04:53:08

Diagnostics Check		OK	04:40	
Air Blank		0.000	04:40	
Control Test		0.079	04:41	
Air Blank		0.000	04:41	
Subject Sample	# 1	0.156	04:44	Volume Not Met
Breath Volume		0.183	04:44	
Air Blank		0.000	04:45	
Air Blank		0.000	04:47	
Subject Sample	# 2	0.152	04:48	
Breath Volume		2.148	04:48	
Air Blank		0.000	04:49	
Air Blank		0.000	04:50	
Subject Sample	# 3	0.157	04:51	
Breath Volume		2.910	04:51	
Air Blank		0.000	04:52	
Control Test		0.078	04:52	
Air Blank		0.000	04:53	
Diagnostics Check		OK	04:53	

Data Download Date/Time: 04/16/2019 08:09

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: STEELE,ZACHARY
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: RODRIGUEZ CASTILLO

DATE/TIME OF TEST: 03/09/2019 00:43:26

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.079	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.195	00:39
Breath Volume		2.000	00:39
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.199	00:41
Breath Volume		2.777	00:41
Air Blank		0.000	00:42
Control Test		0.077	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 04/01/2019 12:24

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PARKER,STEPHEN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: VALDEZ

DATE/TIME OF TEST: 03/11/2019 00:35:47

Diagnostics Check		OK	00:29
Air Blank		0.000	00:30
Control Test		0.079	00:30
Air Blank		0.000	00:30
Subject Sample	# 1	0.204	00:31
Breath Volume		3.562	00:31
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.205	00:34
Breath Volume		4.128	00:34
Air Blank		0.000	00:34
Control Test		0.079	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 04/01/2019 12:24

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: FIEDLER,CHRIS
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 03/13/2019 23:53:34

Diagnostics Check		OK	23:47
Air Blank		0.000	23:47
Control Test		0.081	23:48
Air Blank		0.000	23:48
Subject Sample	# 1	0.149	23:49
Breath Volume		3.082	23:49
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.131	23:52
Breath Volume		1.703	23:52
Air Blank		0.000	23:52
Control Test		0.080	23:52
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 04/01/2019 12:24

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: STERLING,WILLIAM
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/14/2019 00:21:00

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.080	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.000	00:16
Breath Volume		3.429	00:16
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.000	00:19
Breath Volume		2.214	00:19
Air Blank		0.000	00:20
Control Test		0.079	00:20
Air Blank		0.000	00:20
Diagnostics Check		OK	00:21

Data Download Date/Time: 04/01/2019 12:24

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: STEELE,ZACHARY
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: CUSTARD

DATE/TIME OF TEST: 03/15/2019 02:04:04

Diagnostics Check		OK	01:57
Air Blank		0.000	01:57
Control Test		0.080	01:57
Air Blank		0.000	01:58
Subject Sample	# 1	0.162	01:59
Breath Volume		2.390	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:01
Subject Sample	# 2	0.143	02:02
Breath Volume		1.503	02:02
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:04

Data Download Date/Time: 04/01/2019 12:24

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ALLRED,JAMES
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: JAGGERS

DATE/TIME OF TEST: 03/15/2019 04:40:51

Diagnostics Check		OK	04:38	
Air Blank		0.000	04:38	
Control Test		0.081	04:38	
Air Blank		0.000	04:39	
Subject Sample	# 1		04:39	Subject Test Refused
Breath Volume		0.000	04:39	
Air Blank		0.000	04:39	
Control Test		0.081	04:40	
Air Blank		0.000	04:40	
Diagnostics Check		OK	04:40	

Data Download Date/Time: 04/01/2019 12:25

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,SEAN,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ALLRED,JAMES
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: STAPELTON

DATE/TIME OF TEST: 03/17/2019 00:46:50

Diagnostics Check		OK	00:37	
Air Blank		0.000	00:38	
Control Test		0.080	00:38	
Air Blank		0.000	00:39	
Subject Sample	# 1		00:42	No Sample Provided
Breath Volume		0.000	00:42	
Air Blank		0.000	00:42	
Air Blank		0.000	00:44	
Subject Sample	# 2		00:45	Subject Test Refused
Breath Volume		0.000	00:45	
Air Blank		0.000	00:45	
Control Test		0.081	00:46	
Air Blank		0.000	00:46	
Diagnostics Check		OK	00:46	

Data Download Date/Time: 04/01/2019 12:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: STERLING,WILLIAM
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: LIZZANO

DATE/TIME OF TEST: 03/23/2019 22:14:56

Diagnostics Check		OK	22:11	
Air Blank		0.000	22:12	
Control Test		0.081	22:12	
Air Blank		0.000	22:13	
Subject Sample	# 1		22:13	Subject Test Refused
Breath Volume		0.000	22:13	
Air Blank		0.000	22:14	
Control Test		0.081	22:14	
Air Blank		0.000	22:14	
Diagnostics Check		OK	22:14	

Data Download Date/Time: 04/01/2019 12:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: WADE,CODY
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 03/25/2019 01:18:01

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:15	
Control Test		0.080	01:15	
Air Blank		0.000	01:16	
Subject Sample	# 1		01:16	Subject Test Refused
Breath Volume		0.000	01:16	
Air Blank		0.000	01:17	
Control Test		0.081	01:17	
Air Blank		0.000	01:17	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 04/01/2019 12:25

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 03/06/2019 22:36:55

Diagnostics Check		OK	22:30
Air Blank		0.000	22:30
Control Test		0.081	22:31
Air Blank		0.000	22:31
Subject Sample	# 1	0.216	22:32
Breath Volume		1.929	22:32
Air Blank		0.000	22:33
Air Blank		0.000	22:34
Subject Sample	# 2	0.204	22:35
Breath Volume		1.902	22:35
Air Blank		0.000	22:36
Control Test		0.081	22:36
Air Blank		0.000	22:36
Diagnostics Check		OK	22:36

Data Download Date/Time: 03/30/2019 07:31

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BAYER

DATE/TIME OF TEST: 03/21/2019 02:12:10

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.080	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.176	02:07
Breath Volume		2.722	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.173	02:10
Breath Volume		2.437	02:10
Air Blank		0.000	02:11
Control Test		0.080	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 03/30/2019 07:31

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 03/23/2019 00:40:20

Diagnostics Check		OK	00:33
Air Blank		0.000	00:34
Control Test		0.080	00:34
Air Blank		0.000	00:34
Subject Sample	# 1	0.107	00:35
Breath Volume		1.785	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.113	00:38
Breath Volume		2.394	00:38
Air Blank		0.000	00:39
Control Test		0.080	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 03/30/2019 07:31

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/03/2019 00:18:19

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.082	00:10
Air Blank		0.000	00:11
Subject Sample	# 1	0.208	00:13
Breath Volume		1.195	00:13
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.218	00:16
Breath Volume		1.195	00:16
Air Blank		0.000	00:17
Control Test		0.080	00:17
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 03/31/2019 22:35

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MARC,E
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SMITH,MARC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: CLEMENTS

DATE/TIME OF TEST: 03/04/2019 00:44:47

Diagnostics Check		OK	00:38
Air Blank		0.000	00:38
Control Test		0.081	00:39
Air Blank		0.000	00:39
Subject Sample	# 1	0.209	00:40
Breath Volume		2.968	00:40
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.198	00:43
Breath Volume		2.703	00:43
Air Blank		0.000	00:43
Control Test		0.082	00:44
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 03/31/2019 22:35

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,CHANNING,L
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: LEATHEROW,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 07:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 03/12/2019 09:49:42

Diagnostics Check		OK	09:43
Air Blank		0.000	09:44
Control Test		0.083	09:44
Air Blank		0.000	09:44
Subject Sample	# 1	0.000	09:45
Breath Volume		3.578	09:45
Air Blank		0.000	09:45
Air Blank		0.000	09:47
Subject Sample	# 2	0.000	09:48
Breath Volume		3.578	09:48
Air Blank		0.000	09:48
Control Test		0.082	09:49
Air Blank		0.000	09:49
Diagnostics Check		OK	09:49

Data Download Date/Time: 03/31/2019 22:35

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MARC,E
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SMITH,MARC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 12:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 03/17/2019 01:48:26

Diagnostics Check		OK	01:41
Air Blank		0.000	01:42
Control Test		0.082	01:42
Air Blank		0.000	01:42
Subject Sample	# 1	0.115	01:43
Breath Volume		1.953	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.116	01:46
Breath Volume		1.820	01:46
Air Blank		0.000	01:47
Control Test		0.082	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 03/31/2019 22:35

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEVES,VIRGILIO,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ESTEVES,VIRGILIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: WOHLBOLD

DATE/TIME OF TEST: 03/18/2019 15:48:03

Diagnostics Check		OK	15:40
Air Blank		0.000	15:40
Control Test		0.082	15:41
Air Blank		0.000	15:41
Subject Sample	# 1	0.196	15:43
Breath Volume		3.152	15:43
Air Blank		0.000	15:44
Air Blank		0.000	15:45
Subject Sample	# 2	0.191	15:46
Breath Volume		2.453	15:46
Air Blank		0.000	15:47
Control Test		0.081	15:47
Air Blank		0.000	15:47
Diagnostics Check		OK	15:48

Data Download Date/Time: 03/31/2019 22:35

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 03/22/2019 23:55:35

Diagnostics Check		OK	23:49
Air Blank		0.000	23:50
Control Test		0.083	23:50
Air Blank		0.000	23:50
Subject Sample	# 1	0.116	23:51
Breath Volume		1.925	23:51
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.103	23:54
Breath Volume		1.496	23:54
Air Blank		0.000	23:54
Control Test		0.083	23:54
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 03/31/2019 22:35

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JATEFF,TIFFANY,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: JATEFF,TIFFANY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: HIERS

DATE/TIME OF TEST: 03/24/2019 22:51:47

Diagnostics Check		OK	22:45
Air Blank		0.000	22:46
Control Test		0.082	22:46
Air Blank		0.000	22:46
Subject Sample	# 1	0.107	22:47
Breath Volume		3.355	22:47
Air Blank		0.000	22:48
Air Blank		0.000	22:49
Subject Sample	# 2	0.106	22:50
Breath Volume		3.371	22:50
Air Blank		0.000	22:50
Control Test		0.082	22:51
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 03/31/2019 22:35

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/26/2019 01:07:04

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.082	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.000	01:02
Breath Volume		2.117	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.000	01:05
Breath Volume		2.277	01:05
Air Blank		0.000	01:06
Control Test		0.083	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:07

Data Download Date/Time: 03/31/2019 22:35

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,KENDRICK,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TURNER,KENDRICK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: RYON

DATE/TIME OF TEST: 03/02/2019 21:16:26

Diagnostics Check		OK	21:09
Air Blank		0.000	21:09
Control Test		0.080	21:10
Air Blank		0.000	21:10
Subject Sample	# 1	0.199	21:11
Breath Volume		2.738	21:11
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 2	0.199	21:14
Breath Volume		2.738	21:14
Air Blank		0.000	21:15
Control Test		0.078	21:15
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

Data Download Date/Time: 03/21/2019 17:43

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENDT,HENDRIK,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BERNARD,KRISTOPHER
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 03/03/2019 03:25:07

Diagnostics Check		OK	03:13	
Air Blank		0.000	03:14	
Control Test		0.079	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:18	No Sample Provided
Breath Volume		0.000	03:18	
Air Blank		0.000	03:18	
Air Blank		0.000	03:20	
Subject Sample	# 2		03:23	No Sample Provided
Breath Volume		0.000	03:23	
Air Blank		0.000	03:24	
Control Test		0.080	03:24	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 03/21/2019 17:43

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,JASON,S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: ROWE,JASON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:07

SUBJECT'S LAST NAME: BRISTOL

DATE/TIME OF TEST: 03/03/2019 11:32:35

Diagnostics Check	OK	11:31	
Air Blank		11:32	RFI Detect
Air Blank	0.000	11:32	

Data Download Date/Time: 03/21/2019 17:43

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,JASON,S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: ROWE,JASON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:07

SUBJECT'S LAST NAME: BRISTOL

DATE/TIME OF TEST: 03/03/2019 11:41:41

Diagnostics Check		OK	11:35
Air Blank		0.000	11:35
Control Test		0.079	11:35
Air Blank		0.000	11:36
Subject Sample	# 1	0.275	11:37
Breath Volume		1.679	11:37
Air Blank		0.000	11:38
Air Blank		0.000	11:39
Subject Sample	# 2	0.272	11:40
Breath Volume		1.882	11:40
Air Blank		0.000	11:40
Control Test		0.077	11:41
Air Blank		0.000	11:41
Diagnostics Check		OK	11:41

Data Download Date/Time: 03/21/2019 17:43

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POSTLE,SHELLEY,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BRICE,DAVID
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:53

SUBJECT'S LAST NAME: SIMMS

DATE/TIME OF TEST: 03/05/2019 11:18:58

Diagnostics Check		OK	11:16	
Air Blank		0.000	11:16	
Control Test		0.080	11:16	
Air Blank		0.000	11:17	
Subject Sample	# 1		11:17	Subject Test Refused
Breath Volume		0.000	11:17	
Air Blank		0.000	11:18	
Control Test		0.080	11:18	
Air Blank		0.000	11:18	
Diagnostics Check		OK	11:18	

Data Download Date/Time: 03/21/2019 17:43

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/09/2019 04:57:57

Diagnostics Check		OK	04:50
Air Blank		0.000	04:51
Control Test		0.079	04:51
Air Blank		0.000	04:52
Subject Sample	# 1	0.178	04:53
Breath Volume		1.578	04:53
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.171	04:56
Breath Volume		2.160	04:56
Air Blank		0.000	04:57
Control Test		0.078	04:57
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 03/21/2019 17:43

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: WOODY

DATE/TIME OF TEST: 03/09/2019 22:09:26

Diagnostics Check		OK	21:55	
Air Blank		0.000	21:56	
Control Test		0.080	21:56	
Air Blank		0.000	21:57	
Subject Sample	# 1	0.117	22:00	Volume Not Met
Breath Volume		0.308	22:00	
Air Blank		0.000	22:01	
Air Blank		0.000	22:02	
Subject Sample	# 2	0.144	22:04	
Breath Volume		1.472	22:04	
Air Blank		0.000	22:04	
Air Blank		0.000	22:06	
Subject Sample	# 3	0.145	22:07	
Breath Volume		1.285	22:07	
Air Blank		0.000	22:08	
Control Test		0.078	22:08	
Air Blank		0.000	22:09	
Diagnostics Check		OK	22:09	

Data Download Date/Time: 03/21/2019 17:43

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: ABDULKARIM

DATE/TIME OF TEST: 03/15/2019 03:35:25

Diagnostics Check		OK	03:24	
Air Blank		0.000	03:24	
Control Test		0.080	03:24	
Air Blank		0.000	03:25	
Subject Sample	# 1		03:28	No Sample Provided
Breath Volume		0.000	03:28	
Air Blank		0.000	03:28	
Air Blank		0.000	03:30	
Subject Sample	# 2		03:34	No Sample Provided
Breath Volume		0.000	03:34	
Air Blank		0.000	03:34	
Control Test		0.080	03:34	
Air Blank		0.000	03:35	
Diagnostics Check		OK	03:35	

Data Download Date/Time: 03/21/2019 17:43

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/20/2019 03:26:59

Diagnostics Check		OK	03:15	
Air Blank		0.000	03:16	
Control Test		0.081	03:16	
Air Blank		0.000	03:16	
Subject Sample	# 1		03:20	No Sample Provided
Breath Volume		0.000	03:20	
Air Blank		0.000	03:20	
Air Blank		0.000	03:22	
Subject Sample	# 2		03:25	No Sample Provided
Breath Volume		0.000	03:25	
Air Blank		0.000	03:26	
Control Test		0.081	03:26	
Air Blank		0.000	03:26	
Diagnostics Check		OK	03:26	

Data Download Date/Time: 03/21/2019 17:43

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,KENDRICK,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: JOHNSON,TRAVIS
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: REGISTER

DATE/TIME OF TEST: 03/22/2019 04:51:00

Diagnostics Check		OK	04:43
Air Blank		0.000	04:43
Control Test		0.081	04:43
Air Blank		0.000	04:44
Subject Sample	# 1	0.155	04:46
Breath Volume		1.878	04:46
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.157	04:49
Breath Volume		1.277	04:49
Air Blank		0.000	04:50
Control Test		0.079	04:50
Air Blank		0.000	04:50
Diagnostics Check		OK	04:51

Data Download Date/Time: 04/24/2019 00:54

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 03/23/2019 23:05:25

Diagnostics Check		OK	22:57
Air Blank		0.000	22:58
Control Test		0.080	22:58
Air Blank		0.000	22:58
Subject Sample	# 1	0.204	23:00
Breath Volume		3.324	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.204	23:03
Breath Volume		3.925	23:03
Air Blank		0.000	23:04
Control Test		0.080	23:04
Air Blank		0.000	23:05
Diagnostics Check		OK	23:05

Data Download Date/Time: 04/24/2019 00:54

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLOSKA,EDWARD,J
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: RIOS,MARGARITO
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 03/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MAYS

DATE/TIME OF TEST: 03/30/2019 02:36:46

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:31	
Control Test		0.080	02:32	
Air Blank		0.000	02:32	
Subject Sample	# 1	0.000	02:32	
Breath Volume		2.250	02:32	
Air Blank		0.000	02:33	
Air Blank		0.000	02:35	
Subject Sample	# 2	0.000	02:35	
Breath Volume		4.433	02:35	
Air Blank			02:36	Improper Sample
Air Blank		0.000	02:36	

Data Download Date/Time: 04/17/2019 11:27

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLOSKA,EDWARD,J
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: RIOS,MARGARITO
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 03/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MAYS

DATE/TIME OF TEST: 03/30/2019 02:44:50

Diagnostics Check		OK	02:41	
Air Blank		0.000	02:42	
Control Test		0.079	02:42	
Air Blank		0.000	02:42	
Subject Sample	# 1	0.000	02:43	
Breath Volume		5.187	02:43	
Air Blank			02:43	Improper Sample
Air Blank		0.000	02:43	
Control Test		0.079	02:44	
Air Blank		0.000	02:44	
Diagnostics Check		OK	02:44	

Data Download Date/Time: 04/17/2019 11:27

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLOSKA,EDWARD,J
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: RIOS,MARGARITO
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 03/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MAYS

DATE/TIME OF TEST: 03/30/2019 02:50:48

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:48	
Control Test		0.079	02:48	
Air Blank		0.000	02:48	
Subject Sample	# 1	0.000	02:49	
Breath Volume		1.703	02:49	
Air Blank			02:49	Improper Sample
Air Blank		0.000	02:49	
Control Test		0.080	02:50	
Air Blank		0.000	02:50	
Diagnostics Check		OK	02:50	

Data Download Date/Time: 04/17/2019 11:27

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLOSKA,EDWARD,J
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: RIOS,MARGARITO
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 03/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MAYS

DATE/TIME OF TEST: 03/30/2019 04:36:00

Diagnostics Check	OK	04:35	
Air Blank		04:35	RFI Detect
Air Blank	0.000	04:36	

Data Download Date/Time: 04/17/2019 11:27

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLOSKA,EDWARD,J
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: RIOS,MARGARITO
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 03/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MAYS

DATE/TIME OF TEST: 03/30/2019 04:44:38

Diagnostics Check		OK	04:38
Air Blank		0.000	04:39
Control Test		0.079	04:39
Air Blank		0.000	04:39
Subject Sample	# 1	0.000	04:40
Breath Volume		4.187	04:40
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.000	04:43
Breath Volume		4.472	04:43
Air Blank		0.000	04:43
Control Test		0.079	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 04/17/2019 11:27

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASLINE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: ORTIZ TIRADO

DATE/TIME OF TEST: 03/01/2019 13:44:45

Diagnostics Check		OK	13:40	
Air Blank		0.000	13:40	
Control Test		0.081	13:41	
Air Blank		0.000	13:41	
Subject Sample	# 1		13:43	RFI Detect
Breath Volume		1.570	13:43	
Air Blank		0.000	13:43	
Control Test		0.081	13:44	
Air Blank		0.000	13:44	
Diagnostics Check		OK	13:44	

Data Download Date/Time: 03/25/2019 12:40

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SWIM,GREG
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 02/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: MCGAULEY

DATE/TIME OF TEST: 03/05/2019 00:00:34

Diagnostics Check		OK	23:54
Air Blank		0.000	23:54
Control Test		0.081	23:54
Air Blank		0.000	23:55
Subject Sample	# 1	0.196	23:56
Breath Volume		2.343	23:56
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 2	0.180	23:59
Breath Volume		1.875	23:59
Air Blank		0.000	23:59
Control Test		0.080	23:59
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 03/25/2019 12:40

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ALEXANDER,DAVID
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: DIONNE

DATE/TIME OF TEST: 03/23/2019 02:44:30

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.081	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.197	02:40
Breath Volume		3.652	02:40
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.197	02:42
Breath Volume		2.328	02:42
Air Blank		0.000	02:43
Control Test		0.080	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 03/25/2019 12:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,TRAVIS
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: RAGER

DATE/TIME OF TEST: 03/28/2019 21:10:48

Diagnostics Check		OK	21:04
Air Blank		0.000	21:04
Control Test		0.082	21:04
Air Blank		0.000	21:05
Subject Sample	# 1	0.330	21:05
Breath Volume		1.503	21:05
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.339	21:09
Breath Volume		1.382	21:09
Air Blank		0.000	21:09
Control Test		0.081	21:10
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 04/21/2019 08:43

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAU,J,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,J
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ERICKSON

DATE/TIME OF TEST: 03/02/2019 04:27:00

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.079	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.101	04:22
Breath Volume		3.292	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.104	04:25
Breath Volume		2.871	04:25
Air Blank		0.000	04:26
Control Test		0.079	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:27

Data Download Date/Time: 03/06/2019 12:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,SARAH,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SCARLATO,FRANK
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: BERNOT

DATE/TIME OF TEST: 03/02/2019 06:10:44

Diagnostics Check		OK	06:04
Air Blank		0.000	06:05
Control Test		0.079	06:05
Air Blank		0.000	06:05
Subject Sample	# 1	0.043	06:06
Breath Volume		1.542	06:06
Air Blank		0.000	06:07
Air Blank		0.000	06:08
Subject Sample	# 2	0.042	06:09
Breath Volume		2.281	06:09
Air Blank		0.000	06:09
Control Test		0.079	06:10
Air Blank		0.000	06:10
Diagnostics Check		OK	06:10

Data Download Date/Time: 03/06/2019 12:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIESLINSKI,JOSEPH,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SHIPPY,D
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 03/02/2019 23:55:52

Diagnostics Check		OK	23:52	
Air Blank		0.000	23:52	
Control Test		0.080	23:53	
Air Blank		0.000	23:53	
Subject Sample	# 1		23:54	Subject Test Refused
Breath Volume		0.000	23:54	
Air Blank		0.000	23:54	
Control Test		0.081	23:55	
Air Blank		0.000	23:55	
Diagnostics Check		OK	23:55	

Data Download Date/Time: 03/06/2019 12:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOBLER,AMANDA,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HARRISON,WILLY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: STALNAKER

DATE/TIME OF TEST: 03/03/2019 21:18:38

Diagnostics Check		OK	21:12
Air Blank		0.000	21:12
Control Test		0.080	21:12
Air Blank		0.000	21:13
Subject Sample	# 1	0.190	21:13
Breath Volume		1.375	21:13
Air Blank		0.000	21:14
Air Blank		0.000	21:16
Subject Sample	# 2	0.182	21:17
Breath Volume		1.593	21:17
Air Blank		0.000	21:17
Control Test		0.080	21:18
Air Blank		0.000	21:18
Diagnostics Check		OK	21:18

Data Download Date/Time: 03/06/2019 12:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILLAK,MICHAEL,O
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SCARLATO,FRANK
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: KEEFE

DATE/TIME OF TEST: 03/04/2019 21:28:08

Diagnostics Check		OK	21:20
Air Blank		0.000	21:21
Control Test		0.080	21:21
Air Blank		0.000	21:22
Subject Sample	# 1	0.076	21:23
Breath Volume		1.390	21:23
Air Blank		0.000	21:23
Air Blank		0.000	21:25
Subject Sample	# 2	0.073	21:26
Breath Volume		1.152	21:26
Air Blank		0.000	21:27
Control Test		0.079	21:27
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 03/06/2019 12:14

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAU,J,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: WOODY

DATE/TIME OF TEST: 03/07/2019 23:50:40

Diagnostics Check	OK	23:49	
Air Blank		23:50	RFI Detect
Air Blank	0.000	23:50	

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAU,J,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: WOODY

DATE/TIME OF TEST: 03/08/2019 00:00:55

Diagnostics Check		OK	23:53
Air Blank		0.000	23:53
Control Test		0.080	23:54
Air Blank		0.000	23:54
Subject Sample	# 1	0.060	23:55
Breath Volume		1.128	23:55
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.061	23:59
Breath Volume		2.160	23:59
Air Blank		0.000	00:00
Control Test		0.080	00:00
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 03/09/2019 02:32:21

Diagnostics Check		OK	02:19	
Air Blank		0.000	02:19	
Control Test		0.079	02:19	
Air Blank		0.000	02:20	
Subject Sample	# 1	0.240	02:23	Volume Not Met
Breath Volume		0.121	02:23	
Air Blank		0.000	02:24	
Air Blank		0.000	02:25	
Subject Sample	# 2	0.272	02:27	
Breath Volume		1.574	02:27	
Air Blank		0.000	02:28	
Air Blank		0.000	02:30	
Subject Sample	# 3	0.256	02:30	
Breath Volume		1.550	02:30	
Air Blank		0.000	02:31	
Control Test		0.078	02:31	
Air Blank		0.000	02:32	
Diagnostics Check		OK	02:32	

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SATTERLEE,ANDREW,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SATTERLEE,ANDREW
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:27

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 03/10/2019 16:00:31

Diagnostics Check		OK	15:53
Air Blank		0.000	15:54
Control Test		0.080	15:54
Air Blank		0.000	15:54
Subject Sample	# 1	0.264	15:55
Breath Volume		3.312	15:55
Air Blank		0.000	15:56
Air Blank		0.000	15:58
Subject Sample	# 2	0.246	15:59
Breath Volume		3.273	15:59
Air Blank		0.000	15:59
Control Test		0.079	15:59
Air Blank		0.000	16:00
Diagnostics Check		OK	16:00

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: ALICEA JR

DATE/TIME OF TEST: 03/11/2019 03:04:49

Diagnostics Check		OK	02:58
Air Blank		0.000	02:59
Control Test		0.080	02:59
Air Blank		0.000	02:59
Subject Sample	# 1	0.218	03:00
Breath Volume		2.781	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:02
Subject Sample	# 2	0.206	03:03
Breath Volume		2.308	03:03
Air Blank		0.000	03:03
Control Test		0.079	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIESLINSKI,JOSEPH,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: MCCRORK

DATE/TIME OF TEST: 03/12/2019 22:01:39

Diagnostics Check		OK	21:55
Air Blank		0.000	21:55
Control Test		0.080	21:56
Air Blank		0.000	21:56
Subject Sample	# 1	0.339	21:57
Breath Volume		3.808	21:57
Air Blank		0.000	21:58
Air Blank		0.000	21:59
Subject Sample	# 2	0.330	22:00
Breath Volume		3.699	22:00
Air Blank		0.000	22:00
Control Test		0.077	22:01
Air Blank		0.000	22:01
Diagnostics Check		OK	22:01

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,MATTHEW,K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MILLS,MATTHEW
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HUGHEY

DATE/TIME OF TEST: 03/14/2019 00:08:04

Diagnostics Check		OK	00:00
Air Blank		0.000	00:01
Control Test		0.080	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.211	00:03
Breath Volume		2.628	00:03
Air Blank		0.000	00:04
Air Blank		0.000	00:05
Subject Sample	# 2	0.218	00:06
Breath Volume		1.738	00:06
Air Blank		0.000	00:07
Control Test		0.079	00:07
Air Blank		0.000	00:07
Diagnostics Check		OK	00:08

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 03/14/2019 02:23:00

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.080	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.099	02:18
Breath Volume		1.792	02:18
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.102	02:21
Breath Volume		2.429	02:21
Air Blank		0.000	02:22
Control Test		0.080	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:23

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KULKO,JON,D
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ESTERLE,HUGH
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:12

SUBJECT'S LAST NAME: DENZ

DATE/TIME OF TEST: 03/14/2019 16:03:21

Diagnostics Check		OK	15:51	
Air Blank		0.000	15:51	
Control Test		0.080	15:52	
Air Blank		0.000	15:52	
Subject Sample	# 1	0.191	15:55	Volume Not Met
Breath Volume		0.101	15:55	
Air Blank		0.000	15:56	
Air Blank		0.000	15:58	
Subject Sample	# 2	0.206	15:58	No .020 agreement
Breath Volume		3.000	15:58	
Air Blank		0.000	15:59	
Air Blank		0.000	16:01	
Subject Sample	# 3	0.182	16:01	No .020 agreement
Breath Volume		1.468	16:01	
Air Blank		0.000	16:02	
Control Test		0.079	16:02	
Air Blank		0.000	16:03	
Diagnostics Check		OK	16:03	

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIESLINSKI,JOSEPH,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: BONDI

DATE/TIME OF TEST: 03/15/2019 21:48:53

Diagnostics Check	OK	21:47	
Air Blank	0.000	21:47	
Control Test	0.080	21:48	
Air Blank		21:48	RFI Detect
Air Blank	0.000	21:48	

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIESLINSKI,JOSEPH,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: BONDI

DATE/TIME OF TEST: 03/15/2019 21:57:39

Diagnostics Check		OK	21:51
Air Blank		0.000	21:51
Control Test		0.080	21:52
Air Blank		0.000	21:52
Subject Sample	# 1	0.097	21:53
Breath Volume		1.800	21:53
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.099	21:56
Breath Volume		1.855	21:56
Air Blank		0.000	21:56
Control Test		0.078	21:57
Air Blank		0.000	21:57
Diagnostics Check		OK	21:57

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIESLINSKI,JOSEPH,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SCARLATO,FRANK
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 03/16/2019 02:13:48

Diagnostics Check		OK	02:07
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.024	02:09
Breath Volume		3.929	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.022	02:12
Breath Volume		3.019	02:12
Air Blank		0.000	02:12
Control Test		0.078	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIESLINSKI,JOSEPH,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: MONAHAN

DATE/TIME OF TEST: 03/16/2019 02:41:49

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.079	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.099	02:37
Breath Volume		3.128	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:39
Subject Sample	# 2	0.098	02:40
Breath Volume		3.160	02:40
Air Blank		0.000	02:40
Control Test		0.079	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUSKEVICH,CURT,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,J
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: SHEERAN-QUIROZ

DATE/TIME OF TEST: 03/18/2019 03:05:23

Diagnostics Check		OK	02:58
Air Blank		0.000	02:58
Control Test		0.080	02:58
Air Blank		0.000	02:59
Subject Sample	# 1	0.196	03:00
Breath Volume		2.566	03:00
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.184	03:03
Breath Volume		1.445	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAU,J,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: KONDAS,S
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: OQUENDO

DATE/TIME OF TEST: 03/18/2019 21:57:36

Diagnostics Check		OK	21:51
Air Blank		0.000	21:51
Control Test		0.080	21:52
Air Blank		0.000	21:52
Subject Sample	# 1	0.279	21:53
Breath Volume		3.171	21:53
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.266	21:56
Breath Volume		1.921	21:56
Air Blank		0.000	21:56
Control Test		0.078	21:57
Air Blank		0.000	21:57
Diagnostics Check		OK	21:57

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIESLINSKI,JOSEPH,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: OTTINGER

DATE/TIME OF TEST: 03/20/2019 21:03:50

Diagnostics Check		OK	20:57
Air Blank		0.000	20:58
Control Test		0.080	20:58
Air Blank		0.000	20:59
Subject Sample	# 1	0.000	20:59
Breath Volume		1.417	20:59
Air Blank		0.000	21:00
Air Blank		0.000	21:02
Subject Sample	# 2	0.000	21:02
Breath Volume		1.195	21:02
Air Blank		0.000	21:02
Control Test		0.079	21:03
Air Blank		0.000	21:03
Diagnostics Check		OK	21:03

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIESLINSKI,JOSEPH,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BROWN,SAMANTHA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: BEGBIE JR

DATE/TIME OF TEST: 03/22/2019 00:18:15

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:15	
Control Test		0.080	00:16	
Air Blank		0.000	00:16	
Subject Sample	# 1		00:16	Subject Test Refused
Breath Volume		0.000	00:16	
Air Blank		0.000	00:17	
Control Test		0.080	00:17	
Air Blank		0.000	00:18	
Diagnostics Check		OK	00:18	

Data Download Date/Time: 04/17/2019 10:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: LUCKFELT

DATE/TIME OF TEST: 03/23/2019 04:28:44

Diagnostics Check		OK	04:21	
Air Blank		0.000	04:22	
Control Test		0.080	04:22	
Air Blank		0.000	04:23	
Subject Sample	# 1	0.000	04:24	
Breath Volume		1.515	04:24	
Air Blank		0.000	04:24	
Air Blank		0.000	04:26	
Subject Sample	# 2	0.000	04:27	
Breath Volume		2.000	04:27	
Air Blank		0.000	04:27	
Control Test		0.079	04:28	
Air Blank			04:28	RFI Detect
Air Blank		0.000	04:28	

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/24/2019 03:38:55

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.080	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.000	03:34
Breath Volume		2.347	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.000	03:37
Breath Volume		2.003	03:37
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 04/17/2019 10:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSH,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOWE,ALEXANDER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: KING HEWETT

DATE/TIME OF TEST: 03/26/2019 03:22:41

Diagnostics Check		OK	03:13
Air Blank		0.000	03:14
Control Test		0.080	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.153	03:15
Breath Volume		1.687	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.170	03:21
Breath Volume		4.218	03:21
Air Blank		0.000	03:21
Control Test		0.079	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIESLINSKI,JOSEPH,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SCARLATO,FRANK
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 03/27/2019 04:25:25

Diagnostics Check		OK	04:19
Air Blank		0.000	04:19
Control Test		0.080	04:20
Air Blank		0.000	04:20
Subject Sample	# 1	0.225	04:21
Breath Volume		2.906	04:21
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.214	04:23
Breath Volume		2.531	04:23
Air Blank		0.000	04:24
Control Test		0.076	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUSER,SGT,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: OLIVER,R
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/28/2019 10:02:51

Diagnostics Check		OK	09:56
Air Blank		0.000	09:57
Control Test		0.080	09:57
Air Blank		0.000	09:58
Subject Sample	# 1	0.000	09:58
Breath Volume		1.546	09:58
Air Blank		0.000	09:59
Air Blank		0.000	10:01
Subject Sample	# 2	0.000	10:01
Breath Volume		1.578	10:01
Air Blank		0.000	10:01
Control Test		0.080	10:02
Air Blank		0.000	10:02
Diagnostics Check		OK	10:02

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIESLINSKI,JOSEPH,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: SHREVE

DATE/TIME OF TEST: 03/31/2019 01:18:34

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.080	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.112	01:14
Breath Volume		1.398	01:14
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.108	01:17
Breath Volume		1.429	01:17
Air Blank		0.000	01:17
Control Test		0.079	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 04/17/2019 10:42

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JOSE,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: LOPEZ,JOSE
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: DE CAMPS

DATE/TIME OF TEST: 03/10/2019 21:47:44

Diagnostics Check		OK	21:40	
Air Blank		0.000	21:40	
Control Test		0.079	21:40	
Air Blank		0.000	21:41	
Subject Sample	# 1	0.174	21:42	
Breath Volume		2.371	21:42	
Air Blank		0.000	21:43	
Air Blank		0.000	21:44	
Subject Sample	# 2		21:46	Subject Test Refused
Breath Volume		1.082	21:46	
Air Blank		0.000	21:46	
Control Test		0.079	21:47	
Air Blank		0.000	21:47	
Diagnostics Check		OK	21:47	

Data Download Date/Time: 04/02/2019 11:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JOSE,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: LOPEZ,JOSE
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: VILLAGOMEZ CUEVAS

DATE/TIME OF TEST: 03/23/2019 04:10:10

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:04
Subject Sample	# 1	0.184	04:05
Breath Volume		2.175	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:07
Subject Sample	# 2	0.186	04:08
Breath Volume		1.859	04:08
Air Blank		0.000	04:09
Control Test		0.079	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 04/02/2019 11:13

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,SEAN,C
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TAYLOR,BRAY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: ALENDAREZ

DATE/TIME OF TEST: 03/04/2019 01:38:29

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.080	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.182	01:34
Breath Volume		2.210	01:34
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.187	01:36
Breath Volume		2.078	01:36
Air Blank		0.000	01:37
Control Test		0.081	01:37
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 03/26/2019 11:47

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASH,JUSTIN,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CASH,JUSTIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: GALVIN

DATE/TIME OF TEST: 03/05/2019 21:39:14

Diagnostics Check		OK	21:33
Air Blank		0.000	21:33
Control Test		0.082	21:33
Air Blank		0.000	21:34
Subject Sample	# 1	0.294	21:34
Breath Volume		2.906	21:34
Air Blank		0.000	21:35
Air Blank		0.000	21:37
Subject Sample	# 2	0.288	21:37
Breath Volume		2.285	21:37
Air Blank		0.000	21:38
Control Test		0.081	21:38
Air Blank		0.000	21:39
Diagnostics Check		OK	21:39

Data Download Date/Time: 03/26/2019 11:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: TAM

DATE/TIME OF TEST: 03/09/2019 01:46:59

Diagnostics Check		OK	01:41
Air Blank		0.000	01:41
Control Test		0.081	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.081	01:42
Breath Volume		2.414	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.082	01:45
Breath Volume		2.621	01:45
Air Blank		0.000	01:46
Control Test		0.081	01:46
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 03/26/2019 11:47

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASH,JUSTIN,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: THOMPSON,MASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 03/09/2019 22:45:58

Diagnostics Check		OK	22:39
Air Blank		0.000	22:39
Control Test		0.081	22:39
Air Blank		0.000	22:40
Subject Sample	# 1	0.279	22:41
Breath Volume		2.640	22:41
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.281	22:44
Breath Volume		1.750	22:44
Air Blank		0.000	22:45
Control Test		0.081	22:45
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 03/26/2019 11:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASH,JUSTIN,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MANGOLD,KENNETH
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: PRAKASH

DATE/TIME OF TEST: 03/11/2019 01:53:41

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.082	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.146	01:49
Breath Volume		2.125	01:49
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.129	01:52
Breath Volume		1.191	01:52
Air Blank		0.000	01:52
Control Test		0.080	01:53
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 03/26/2019 11:47

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASH,JUSTIN,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SASTRE,STACCI
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: PAPAS

DATE/TIME OF TEST: 03/11/2019 02:43:52

Diagnostics Check		OK	02:39	
Air Blank		0.000	02:40	
Control Test		0.081	02:40	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:42	Subject Test Refused
Breath Volume		0.000	02:42	
Air Blank		0.000	02:42	
Control Test		0.082	02:43	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 03/26/2019 11:47

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASH,JUSTIN,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: FORSYTHE

DATE/TIME OF TEST: 03/11/2019 03:19:57

Diagnostics Check		OK	03:13
Air Blank		0.000	03:14
Control Test		0.081	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.149	03:15
Breath Volume		1.777	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.144	03:18
Breath Volume		1.257	03:18
Air Blank		0.000	03:19
Control Test		0.082	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 03/26/2019 11:47

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASH,JUSTIN,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: WEBSTER

DATE/TIME OF TEST: 03/11/2019 03:49:57

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.080	03:42
Air Blank		0.000	03:42
Subject Sample	# 1	0.202	03:44
Breath Volume		1.312	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.203	03:48
Breath Volume		1.773	03:48
Air Blank		0.000	03:49
Control Test		0.081	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 03/26/2019 11:47

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASH,JUSTIN,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HETZLER,BRANDON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: GUY

DATE/TIME OF TEST: 03/23/2019 05:16:24

Diagnostics Check	OK	05:15	
Air Blank	0.000	05:15	
Control Test	0.081	05:15	RFI Detect
Air Blank	0.000	05:16	

Data Download Date/Time: 03/26/2019 11:47

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASH,JUSTIN,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: THOMPSON,MASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: BELLAMY

DATE/TIME OF TEST: 03/24/2019 19:17:14

Diagnostics Check		OK	19:11
Air Blank		0.000	19:11
Control Test		0.081	19:11
Air Blank		0.000	19:12
Subject Sample	# 1	0.349	19:12
Breath Volume		1.539	19:12
Air Blank		0.000	19:13
Air Blank		0.000	19:15
Subject Sample	# 2	0.350	19:15
Breath Volume		1.230	19:15
Air Blank		0.000	19:16
Control Test		0.081	19:16
Air Blank		0.000	19:17
Diagnostics Check		OK	19:17

Data Download Date/Time: 03/26/2019 11:47

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINSKY,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: LINSKY,MICHAEL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: YELTON

DATE/TIME OF TEST: 03/25/2019 17:12:39

Diagnostics Check		OK	17:00
Air Blank		0.000	17:01
Control Test		0.081	17:01
Air Blank		0.000	17:02
Subject Sample	# 1	0.318	17:02
Breath Volume		1.726	17:02
Air Blank		0.000	17:03
Air Blank		0.000	17:05
Subject Sample	# 2	0.290	17:08
Breath Volume		1.554	17:08
Air Blank		0.000	17:08
Air Blank		0.000	17:10
Subject Sample	# 3	0.297	17:11
Breath Volume		1.425	17:11
Air Blank		0.000	17:11
Control Test		0.080	17:12
Air Blank		0.000	17:12
Diagnostics Check		OK	17:12

Data Download Date/Time: 03/26/2019 11:47

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASH,JUSTIN,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MANGOLD,KENNETH
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: SALVATORE

DATE/TIME OF TEST: 03/28/2019 00:42:48

Diagnostics Check		OK	00:38	
Air Blank		0.000	00:39	
Control Test		0.082	00:39	
Air Blank		0.000	00:39	
Subject Sample	# 1		00:41	Subject Test Refused
Breath Volume		0.000	00:41	
Air Blank		0.000	00:41	
Control Test		0.082	00:42	
Air Blank		0.000	00:42	
Diagnostics Check		OK	00:42	

Data Download Date/Time: 04/18/2019 14:51

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,THOMAS,H
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: THOMPSON,MASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 03/29/2019 01:09:51

Diagnostics Check		OK	01:03
Air Blank		0.000	01:04
Control Test		0.081	01:04
Air Blank		0.000	01:04
Subject Sample	# 1	0.082	01:05
Breath Volume		1.738	01:05
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.078	01:08
Breath Volume		2.316	01:08
Air Blank		0.000	01:08
Control Test		0.080	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 04/18/2019 14:51

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: NORTH MIAMI PD
ARRESTING OFFICER: PINEDA,MIGUEL
ARRESTING OFFICER AGENCY: NORTH MIAMI PD
LAST AGENCY INSPECTION: 03/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34418080A3
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: JARRETT

DATE/TIME OF TEST: 03/13/2019 05:38:20

Diagnostics Check		OK	05:31
Air Blank		0.000	05:31
Control Test		0.081	05:32
Air Blank		0.000	05:32
Subject Sample	# 1	0.244	05:33
Breath Volume		1.839	05:33
Air Blank		0.000	05:34
Air Blank		0.000	05:35
Subject Sample	# 2	0.242	05:36
Breath Volume		2.480	05:36
Air Blank		0.000	05:37
Control Test		0.080	05:37
Air Blank		0.000	05:38
Diagnostics Check		OK	05:38

Data Download Date/Time: 04/23/2019 15:50

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: NORTH MIAMI PD
ARRESTING OFFICER: CARTER,ERICKA
ARRESTING OFFICER AGENCY: NORTH MIAMI PD
LAST AGENCY INSPECTION: 03/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34418080A3
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 03/26/2019 00:10:37

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.082	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.000	00:06
Breath Volume		3.203	00:06
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.000	00:09
Breath Volume		4.070	00:09
Air Blank		0.000	00:09
Control Test		0.082	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 04/23/2019 15:50

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

INSTRUMENT REGISTERED TO: MICCOSUKEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,RENE,
BREATH TEST OPERATOR AGENCY: MICCOSUKEE PD
ARRESTING OFFICER: BAKER,M
ARRESTING OFFICER AGENCY: MICCOSUKEE PD
LAST AGENCY INSPECTION: 02/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 03/14/2019 02:46:42

Diagnostics Check		OK	02:43	
Air Blank		0.000	02:44	
Control Test		0.078	02:44	
Air Blank		0.000	02:45	
Subject Sample	# 1		02:45	Subject Test Refused
Breath Volume		0.000	02:45	
Air Blank		0.000	02:45	
Control Test		0.078	02:46	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:46	

Data Download Date/Time: 03/27/2019 09:56

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

INSTRUMENT REGISTERED TO: MICCOSUKEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MISTY,L
BREATH TEST OPERATOR AGENCY: MICCOSUKEE PD
ARRESTING OFFICER: DOMINGUEZ,HORACIO
ARRESTING OFFICER AGENCY: MICCOSUKEE PD
LAST AGENCY INSPECTION: 02/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 03/15/2019 03:12:48

Diagnostics Check		OK	03:00	
Air Blank		0.000	03:01	
Control Test		0.077	03:01	
Air Blank		0.000	03:02	
Subject Sample	# 1	0.107	03:05	Volume Not Met
Breath Volume		0.277	03:05	
Air Blank		0.000	03:06	
Air Blank		0.000	03:07	
Subject Sample	# 2	0.106	03:10	Volume Not Met
Breath Volume		0.207	03:10	
Air Blank		0.000	03:11	
Control Test		0.076	03:12	
Air Blank		0.000	03:12	
Diagnostics Check		OK	03:12	

Data Download Date/Time: 03/27/2019 09:56

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001714

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEFFIELD,KHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: WRZESINSKI,TIMOTHY
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: ALCANTARA LUNA

DATE/TIME OF TEST: 03/02/2019 04:20:38

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:19	
Control Test	0.077	04:19	
Air Blank		04:20	Improper Sample
Air Blank	0.000	04:20	

Data Download Date/Time: 03/29/2019 14:09

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001714

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEFFIELD,KHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: WRZESINSKI,TIMOTHY
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 02/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: ALACANTARA LUNA

DATE/TIME OF TEST: 03/02/2019 04:29:54

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.076	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.179	04:25
Breath Volume		3.238	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:27
Subject Sample	# 2	0.175	04:28
Breath Volume		2.582	04:28
Air Blank		0.000	04:28
Control Test		0.076	04:29
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 03/29/2019 14:09

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 03/10/2019 01:03:39

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.079	00:58
Air Blank		0.000	00:58
Subject Sample	# 1	0.049	00:59
Breath Volume		1.730	00:59
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.047	01:02
Breath Volume		1.261	01:02
Air Blank		0.000	01:02
Control Test		0.078	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 03/18/2019 22:41

Printed: 21 June 2019

COBRA ©1997-2014 CMI, Inc.