

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEGGETT,JARED,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: LEGGETT,JARED
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CORTEZ BOLANOS

DATE/TIME OF TEST: 04/01/2019 02:18:06

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.080	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.208	02:13
Breath Volume		1.515	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:15
Subject Sample	# 2	0.210	02:16
Breath Volume		2.250	02:16
Air Blank		0.000	02:17
Control Test		0.079	02:17
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 04/23/2019 16:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPURLOCK,ROBERT,V
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPURLOCK,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: ADMIN
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: SPURLOCK

DATE/TIME OF TEST: 04/03/2019 19:38:26

Diagnostics Check	OK	19:32
Air Blank	0.000	19:33
Control Test	0.079	19:33
Air Blank	0.000	19:34
Subject Sample # 1	0.000	19:34
Breath Volume	1.589	19:34
Air Blank	0.000	19:34
Air Blank	0.000	19:36
Subject Sample # 2	0.000	19:37
Breath Volume	1.921	19:37
Air Blank	0.000	19:37
Control Test	0.080	19:37
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time: 04/23/2019 16:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,RYAN,C
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: VAUGHN,RYAN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: WIEGMAN

DATE/TIME OF TEST: 04/06/2019 04:03:07

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.000	03:58
Breath Volume	1.890	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.000	04:01
Breath Volume	2.003	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 04/23/2019 16:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,RYAN,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: VAUGHN,RYAN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: RHODES

DATE/TIME OF TEST: 04/12/2019 06:00:07

Diagnostics Check	OK	05:57	
Air Blank	0.000	05:57	
Control Test	0.081	05:58	
Air Blank	0.000	05:58	
Subject Sample	# 1	05:58	Subject Test Refused
Breath Volume	0.000	05:58	
Air Blank	0.000	05:59	
Control Test	0.081	05:59	
Air Blank	0.000	06:00	
Diagnostics Check	OK	06:00	

Data Download Date/Time: 04/23/2019 16:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VREEKEN,RANDEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: VREEKEN,RANDEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:52

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 04/13/2019 14:18:51

Diagnostics Check	OK	14:12
Air Blank	0.000	14:13
Control Test	0.081	14:13
Air Blank	0.000	14:14
Subject Sample # 1	0.000	14:14
Breath Volume	3.199	14:14
Air Blank	0.000	14:15
Air Blank	0.000	14:17
Subject Sample # 2	0.000	14:17
Breath Volume	3.089	14:17
Air Blank	0.000	14:17
Control Test	0.078	14:18
Air Blank	0.000	14:18
Diagnostics Check	OK	14:18

Data Download Date/Time: 04/23/2019 16:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,RYAN,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: VAUGHN,RYAN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 04/14/2019 02:55:52

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.203	02:51
Breath Volume	1.941	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.205	02:54
Breath Volume	2.312	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 04/23/2019 16:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWEN,RYAN,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BOWEN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: JOSE-MANUEL

DATE/TIME OF TEST: 04/15/2019 02:42:24

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.080	02:41	
Air Blank	0.000	02:42	
		02:42	RFI Detect
Air Blank		02:42	RFI Detect

Data Download Date/Time: 04/23/2019 16:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWEN,RYAN,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BOWEN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: JOSE-MANUEL

DATE/TIME OF TEST: 04/15/2019 02:46:55

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:45	
Control Test	0.080	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 04/23/2019 16:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: SKAGGS

DATE/TIME OF TEST: 04/03/2019 00:58:35

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.082	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.096	00:53
Breath Volume	3.687	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.099	00:56
Breath Volume	4.074	00:56
Air Blank	0.000	00:57
Control Test	0.082	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 04/22/2019 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPURLOCK,ROBERT,V
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPURLOCK,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: BORJAS-GAITAN

DATE/TIME OF TEST: 04/04/2019 20:19:51

Diagnostics Check		OK	20:13
Air Blank		0.000	20:13
Control Test		0.082	20:13
Air Blank		0.000	20:14
Subject Sample	# 1	0.250	20:15
Breath Volume		2.519	20:15
Air Blank		0.000	20:16
Air Blank		0.000	20:17
Subject Sample	# 2	0.235	20:18
Breath Volume		2.339	20:18
Air Blank		0.000	20:18
Control Test		0.080	20:19
Air Blank		0.000	20:19
Diagnostics Check		OK	20:19

Data Download Date/Time: 04/22/2019 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 04/06/2019 00:38:49

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.082	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.143	00:34
Breath Volume	2.425	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.140	00:37
Breath Volume	2.730	00:37
Air Blank	0.000	00:37
Control Test	0.081	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 04/22/2019 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MANESS

DATE/TIME OF TEST: 04/07/2019 00:28:39

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.082	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.152	00:24
Breath Volume	1.863	00:24
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.152	00:26
Breath Volume	2.621	00:26
Air Blank	0.000	00:27
Control Test	0.082	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 04/22/2019 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: ARIAS-TORRES

DATE/TIME OF TEST: 04/08/2019 02:28:43

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.082	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.194	02:23
Breath Volume	1.910	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:25
Subject Sample # 2	0.196	02:27
Breath Volume	1.968	02:27
Air Blank	0.000	02:27
Control Test	0.081	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 04/22/2019 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: MAULBECK

DATE/TIME OF TEST: 04/10/2019 21:49:53

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:47	
Control Test	0.081	21:47	
Air Blank	0.000	21:48	
Subject Sample	# 1	21:48	Subject Test Refused
Breath Volume	0.000	21:48	
Air Blank	0.000	21:48	
Control Test	0.081	21:49	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:49	

Data Download Date/Time: 04/22/2019 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: MAULBECK

DATE/TIME OF TEST: 04/10/2019 22:12:28

Diagnostics Check	OK	22:05
Air Blank	0.000	22:06
Control Test	0.081	22:06
Air Blank	0.000	22:07
Subject Sample # 1	0.172	22:07
Breath Volume	2.117	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.167	22:10
Breath Volume	1.617	22:10
Air Blank	0.000	22:11
Control Test	0.080	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 04/22/2019 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: MAYS

DATE/TIME OF TEST: 04/11/2019 21:29:19

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.081	21:23
Air Blank	0.000	21:24
Subject Sample # 1	0.000	21:25
Breath Volume	1.953	21:25
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.000	21:27
Breath Volume	1.550	21:27
Air Blank	0.000	21:28
Control Test	0.081	21:28
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 04/22/2019 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 04/11/2019 23:19:06

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.081	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.000	23:14
Breath Volume	2.546	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.000	23:17
Breath Volume	1.781	23:17
Air Blank	0.000	23:18
Control Test	0.079	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 04/22/2019 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: TANIS

DATE/TIME OF TEST: 04/13/2019 01:13:51

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.082	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.106	01:08
Breath Volume	1.582	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.099	01:12
Breath Volume	1.617	01:12
Air Blank	0.000	01:12
Control Test	0.080	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 04/22/2019 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/20/2019 02:57:41

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.082	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.185	02:53
Breath Volume	2.167	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.175	02:56
Breath Volume	1.269	02:56
Air Blank	0.000	02:56
Control Test	0.082	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 04/22/2019 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: BURKETT

DATE/TIME OF TEST: 04/23/2019 23:50:42

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.082	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.080	23:44
Breath Volume	1.210	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.080	23:49
Breath Volume	1.652	23:49
Air Blank	0.000	23:49
Control Test	0.082	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 05/20/2019 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: SUAZO GONZALEZ

DATE/TIME OF TEST: 04/24/2019 23:05:56

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.081	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.202	23:00
Breath Volume	2.023	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.198	23:04
Breath Volume	1.613	23:04
Air Blank	0.000	23:04
Control Test	0.080	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 05/20/2019 13:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: DEBAUCHE

DATE/TIME OF TEST: 04/30/2019 22:18:52

Diagnostics Check		OK	22:06	
Air Blank		0.000	22:07	
Control Test		0.082	22:07	
Air Blank		0.000	22:07	
Subject Sample	# 1	0.198	22:11	Volume Not Met
Breath Volume		0.937	22:11	
Air Blank		0.000	22:11	
Air Blank		0.000	22:13	
Subject Sample	# 2	0.192	22:14	
Breath Volume		1.250	22:14	
Air Blank		0.000	22:14	
Air Blank		0.000	22:16	
Subject Sample	# 3	0.197	22:17	
Breath Volume		1.523	22:17	
Air Blank		0.000	22:17	
Control Test		0.080	22:18	
Air Blank		0.000	22:18	
Diagnostics Check		OK	22:18	

Data Download Date/Time: 05/20/2019 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHAEFER,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: GEEDING

DATE/TIME OF TEST: 04/01/2019 21:56:17

Diagnostics Check	OK	21:48
Air Blank	0.000	21:48
Control Test	0.080	21:49
Air Blank	0.000	21:49
Subject Sample # 1	0.100	21:50
Breath Volume	3.195	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.106	21:54
Breath Volume	2.058	21:54
Air Blank	0.000	21:55
Control Test	0.081	21:55
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 04/04/2019 17:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHERMAN,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 04/02/2019 16:53:25

Diagnostics Check		OK	16:41	
Air Blank		0.000	16:42	
Control Test		0.080	16:42	
Air Blank		0.000	16:42	
Subject Sample	# 1	0.226	16:46	Volume Not Met
Breath Volume		0.328	16:46	
Air Blank		0.000	16:46	
Air Blank		0.000	16:48	
Subject Sample	# 2	0.236	16:51	Volume Not Met
Breath Volume		0.257	16:51	
Air Blank		0.000	16:52	
Control Test		0.078	16:52	
Air Blank		0.000	16:53	
Diagnostics Check		OK	16:53	

Data Download Date/Time: 04/04/2019 17:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 04/03/2019 04:04:23

Diagnostics Check		OK	03:57
Air Blank		0.000	03:58
Control Test		0.080	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.145	03:59
Breath Volume		2.210	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:01
Subject Sample	# 2	0.150	04:02
Breath Volume		2.789	04:02
Air Blank		0.000	04:03
Control Test		0.081	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 04/04/2019 17:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 04/04/2019 03:44:42

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.000	03:40
Breath Volume	1.617	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.000	03:43
Breath Volume	1.132	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 04/04/2019 17:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLORES,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 04/04/2019 05:48:36

Diagnostics Check		OK	05:41
Air Blank		0.000	05:42
Control Test		0.080	05:42
Air Blank		0.000	05:43
Subject Sample	# 1	0.183	05:43
Breath Volume		2.687	05:43
Air Blank		0.000	05:44
Air Blank		0.000	05:46
Subject Sample	# 2	0.186	05:46
Breath Volume		1.902	05:46
Air Blank		0.000	05:47
Control Test		0.080	05:47
Air Blank		0.000	05:48
Diagnostics Check		OK	05:48

Data Download Date/Time: 04/04/2019 17:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAMPBELL,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: MORSE

DATE/TIME OF TEST: 04/12/2019 20:18:47

Diagnostics Check	OK	20:11
Air Blank	0.000	20:11
Control Test	0.080	20:12
Air Blank	0.000	20:12
Subject Sample # 1	0.157	20:13
Breath Volume	3.277	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:16
Subject Sample # 2	0.145	20:17
Breath Volume	1.628	20:17
Air Blank	0.000	20:17
Control Test	0.079	20:18
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

Data Download Date/Time: 05/06/2019 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: PANKRATH

DATE/TIME OF TEST: 04/12/2019 22:19:51

Diagnostics Check		OK	22:03	
Air Blank		0.000	22:04	
Control Test		0.079	22:04	
Air Blank		0.000	22:05	
Subject Sample	# 1	0.207	22:08	Volume Not Met
Breath Volume		0.226	22:08	
Air Blank		0.000	22:09	
Air Blank		0.000	22:10	
Subject Sample	# 2	0.219	22:13	
Breath Volume		1.601	22:13	
Air Blank		0.000	22:14	
Air Blank		0.000	22:15	
Subject Sample	# 3	0.228	22:18	
Breath Volume		1.511	22:18	
Air Blank		0.000	22:18	
Control Test		0.075	22:19	
Air Blank		0.000	22:19	
Diagnostics Check		OK	22:19	

Data Download Date/Time: 05/06/2019 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: SHUTTS

DATE/TIME OF TEST: 04/12/2019 22:40:42

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.078	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.033	22:36
Breath Volume	1.949	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.031	22:39
Breath Volume	2.203	22:39
Air Blank	0.000	22:39
Control Test	0.078	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 05/06/2019 14:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: RUSSO

DATE/TIME OF TEST: 04/12/2019 22:58:40

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.079	22:49
Air Blank	0.000	22:49
Subject Sample # 1	0.210	22:50
Breath Volume	1.750	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.231	22:54
Breath Volume	1.496	22:54
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 3	0.212	22:57
Breath Volume	1.226	22:57
Air Blank	0.000	22:57
Control Test	0.077	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 05/06/2019 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: PYLES

DATE/TIME OF TEST: 04/13/2019 02:06:38

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.080	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.139	02:01
Breath Volume	1.468	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.140	02:05
Breath Volume	1.398	02:05
Air Blank	0.000	02:05
Control Test	0.079	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 05/06/2019 14:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: STRACK,ROBERT
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: WETHERELL

DATE/TIME OF TEST: 04/06/2019 23:01:10

Diagnostics Check	OK	22:53
Air Blank	0.000	22:54
Control Test	0.080	22:54
Air Blank	0.000	22:55
Subject Sample # 1	0.000	22:56
Breath Volume	3.828	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.000	22:59
Breath Volume	3.789	22:59
Air Blank	0.000	23:00
Control Test	0.080	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 04/30/2019 06:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,D,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: ROSADO,S
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: GIBBONS

DATE/TIME OF TEST: 04/21/2019 13:07:41

Diagnostics Check	OK	13:01
Air Blank	0.000	13:01
Control Test	0.080	13:01
Air Blank	0.000	13:02
Subject Sample # 1	0.229	13:02
Breath Volume	1.746	13:02
Air Blank	0.000	13:03
Air Blank	0.000	13:05
Subject Sample # 2	0.226	13:06
Breath Volume	1.875	13:06
Air Blank	0.000	13:06
Control Test	0.080	13:07
Air Blank	0.000	13:07
Diagnostics Check	OK	13:07

Data Download Date/Time: 04/30/2019 06:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: RENTERIA,BARTOLO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: PRAZERES

DATE/TIME OF TEST: 04/05/2019 23:58:22

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.078	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.112	23:53
Breath Volume	2.691	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.096	23:56
Breath Volume	2.113	23:56
Air Blank	0.000	23:57
Control Test	0.078	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 04/19/2019 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STABLER,JULIE,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VANCAMP,JEFFREY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:06

SUBJECT'S LAST NAME: STABLER

DATE/TIME OF TEST: 04/11/2019 15:46:10

Diagnostics Check	OK	15:39
Air Blank	0.000	15:40
Control Test	0.079	15:40
Air Blank	0.000	15:41
Subject Sample # 1	0.000	15:41
Breath Volume	1.683	15:41
Air Blank	0.000	15:42
Air Blank	0.000	15:44
Subject Sample # 2	0.000	15:44
Breath Volume	1.582	15:44
Air Blank	0.000	15:45
Control Test	0.077	15:45
Air Blank	0.000	15:46
Diagnostics Check	OK	15:46

Data Download Date/Time: 04/19/2019 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,BRITTANIE,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ERMOVICK,ADAM
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: FIERSTEIN

DATE/TIME OF TEST: 04/12/2019 00:38:08

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:35	
Control Test	0.078	00:35	
Air Blank	0.000	00:36	
Subject Sample	# 1	00:36	Subject Test Refused
Breath Volume	0.000	00:36	
Air Blank	0.000	00:37	
Control Test	0.079	00:37	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

Data Download Date/Time: 04/19/2019 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTERIA,BARTOLO,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BURNETTE,ASHLEY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: BERMAN

DATE/TIME OF TEST: 04/13/2019 04:00:03

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.079	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.145	03:55
Breath Volume	1.960	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:57
Subject Sample # 2	0.134	03:58
Breath Volume	1.734	03:58
Air Blank	0.000	03:59
Control Test	0.078	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	04:00

Data Download Date/Time: 04/19/2019 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: RENTERIA,BARTOLO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: KONIGSBURG

DATE/TIME OF TEST: 04/19/2019 02:04:41

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:02	
Control Test	0.079	02:02	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:03	Subject Test Refused
Breath Volume	0.000	02:03	
Air Blank	0.000	02:03	
Control Test	0.078	02:04	
Air Blank	0.000	02:04	
Diagnostics Check	OK	02:04	

Data Download Date/Time: 04/19/2019 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTERIA,BARTOLO,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: PAYNE,JONATHAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: AMSDEN

DATE/TIME OF TEST: 04/19/2019 06:19:09

Diagnostics Check	OK	06:16	
Air Blank	0.000	06:16	
Control Test	0.079	06:17	
Air Blank	0.000	06:17	
Subject Sample	# 1	06:17	Subject Test Refused
Breath Volume	0.000	06:17	
Air Blank	0.000	06:18	
Control Test	0.079	06:18	
Air Blank	0.000	06:19	
Diagnostics Check	OK	06:19	

Data Download Date/Time: 04/19/2019 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTERIA,BARTOLO,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SAAVEDRA,ALONSO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 04/24/2019 01:23:33

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.078	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:22	
Control Test	0.079	01:22	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 05/24/2019 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: BAUDUIN

DATE/TIME OF TEST: 04/24/2019 05:15:44

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.079	05:10
Air Blank	0.000	05:10
Subject Sample # 1	0.157	05:11
Breath Volume	4.164	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.145	05:14
Breath Volume	2.343	05:14
Air Blank	0.000	05:14
Control Test	0.079	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 05/14/2019 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCA,LUIS,C
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: GIZ,JOAQUIN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: SQUIRE

DATE/TIME OF TEST: 04/21/2019 23:07:51

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.081	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.268	23:03
Breath Volume	1.820	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:05
Subject Sample # 2	0.278	23:06
Breath Volume	1.335	23:06
Air Blank	0.000	23:06
Control Test	0.078	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 05/01/2019 01:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCA,LUIS,C
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: HERNANDEZ,JORGE
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 03/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 04/26/2019 03:58:29

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:54	
Control Test	0.082	03:54	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Control Test	0.081	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 05/01/2019 01:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,TABARIE,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,TABARIE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 972301
DRY GAS EXPIRY DATE: 06/15/2021
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: PURVIS

DATE/TIME OF TEST: 04/12/2019 23:14:59

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.166	23:10
Breath Volume	2.085	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:12
Subject Sample # 2	0.155	23:13
Breath Volume	1.562	23:13
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 05/02/2019 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: THAKKAR

DATE/TIME OF TEST: 04/05/2019 19:17:49

Diagnostics Check	OK	19:10
Air Blank	0.000	19:11
Control Test	0.080	19:11
Air Blank	0.000	19:11
Subject Sample # 1	0.193	19:13
Breath Volume	2.089	19:13
Air Blank	0.000	19:13
Air Blank	0.000	19:15
Subject Sample # 2	0.199	19:16
Breath Volume	2.054	19:16
Air Blank	0.000	19:16
Control Test	0.079	19:17
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

Data Download Date/Time: 05/02/2019 17:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/20/2019 01:14:51

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.081	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.169	01:09
Breath Volume	2.441	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.168	01:13
Breath Volume	2.527	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 05/02/2019 17:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PROKOP

DATE/TIME OF TEST: 04/25/2019 02:47:13

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.078	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.105	02:42
Breath Volume	4.792	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:44
Subject Sample # 2	0.111	02:45
Breath Volume	4.046	02:45
Air Blank	0.000	02:46
Control Test	0.078	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 05/20/2019 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: TOSCANO

DATE/TIME OF TEST: 04/25/2019 05:01:04

Diagnostics Check	OK	04:54
Air Blank	0.000	04:54
Control Test	0.077	04:55
Air Blank	0.000	04:55
Subject Sample # 1	0.218	04:56
Breath Volume	2.312	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:58
Subject Sample # 2	0.215	04:59
Breath Volume	1.824	04:59
Air Blank	0.000	05:00
Control Test	0.077	05:00
Air Blank	0.000	05:00
Diagnostics Check	OK	05:01

Data Download Date/Time: 05/20/2019 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 04/26/2019 02:20:21

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.079	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.185	02:15
Breath Volume	5.328	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.177	02:18
Breath Volume	4.414	02:18
Air Blank	0.000	02:19
Control Test	0.077	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 05/20/2019 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VONADOR,JERARD,E
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: VONADOR,JERARD
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: OCHOA

DATE/TIME OF TEST: 04/28/2019 04:22:25

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:20	
Control Test	0.079	04:20	
Air Blank	0.000	04:20	
Subject Sample	# 1	04:21	Subject Test Refused
Breath Volume	0.000	04:21	
Air Blank	0.000	04:21	
Control Test	0.078	04:21	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:22	

Data Download Date/Time: 05/20/2019 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: GOOD

DATE/TIME OF TEST: 04/29/2019 19:51:31

Diagnostics Check		OK	19:39	
Air Blank		0.000	19:39	
Control Test		0.078	19:40	
Air Blank		0.000	19:40	
Subject Sample	# 1	0.191	19:43	Volume Not Met
Breath Volume		0.000	19:43	
Air Blank		0.000	19:44	
Air Blank		0.000	19:46	
Subject Sample	# 2	0.165	19:46	
Breath Volume		1.964	19:46	
Air Blank		0.000	19:47	
Air Blank		0.000	19:49	
Subject Sample	# 3	0.162	19:49	
Breath Volume		1.906	19:49	
Air Blank		0.000	19:50	
Control Test		0.076	19:50	
Air Blank		0.000	19:51	
Diagnostics Check		OK	19:51	

Data Download Date/Time: 05/20/2019 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 04/30/2019 03:54:39

Diagnostics Check		OK	03:47
Air Blank		0.000	03:48
Control Test		0.078	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.211	03:49
Breath Volume		1.957	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.209	03:52
Breath Volume		2.007	03:52
Air Blank		0.000	03:53
Control Test		0.077	03:54
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 05/20/2019 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 04/30/2019 22:40:34

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.079	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.152	22:35
Breath Volume	2.312	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.144	22:38
Breath Volume	1.640	22:38
Air Blank	0.000	22:39
Control Test	0.078	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 05/20/2019 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: HAYS

DATE/TIME OF TEST: 04/05/2019 00:55:32

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:51	
Control Test	0.076	00:51	
Air Blank	0.000	00:52	
Subject Sample	# 1	00:54	Subject Test Refused
Breath Volume	0.000	00:54	
Air Blank	0.000	00:54	
Control Test	0.077	00:54	
Air Blank	0.000	00:55	
Diagnostics Check	OK	00:55	

Data Download Date/Time: 04/19/2019 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: RIVERA REYES

DATE/TIME OF TEST: 04/05/2019 01:32:38

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.078	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.151	01:27
Breath Volume	1.843	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.149	01:31
Breath Volume	1.878	01:31
Air Blank	0.000	01:31
Control Test	0.076	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 04/19/2019 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: ROOPNARINE

DATE/TIME OF TEST: 04/06/2019 03:25:56

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.077	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.191	03:20
Breath Volume	1.488	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.197	03:24
Breath Volume	1.511	03:24
Air Blank	0.000	03:24
Control Test	0.076	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 04/19/2019 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: HILLWAKEFIELD

DATE/TIME OF TEST: 04/11/2019 19:09:35

Diagnostics Check	OK	19:02
Air Blank	0.000	19:03
Control Test	0.078	19:03
Air Blank	0.000	19:04
Subject Sample # 1	0.000	19:05
Breath Volume	1.640	19:05
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample # 2	0.000	19:08
Breath Volume	2.105	19:08
Air Blank	0.000	19:08
Control Test	0.077	19:08
Air Blank	0.000	19:09
Diagnostics Check	OK	19:09

Data Download Date/Time: 04/19/2019 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: ELSEN

DATE/TIME OF TEST: 04/12/2019 02:29:05

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:14	
Control Test		0.078	02:15	
Air Blank		0.000	02:15	
Subject Sample	# 1	0.144	02:18	
Breath Volume		1.390	02:18	
Air Blank		0.000	02:19	
Air Blank		0.000	02:21	
Subject Sample	# 2	0.155	02:24	Volume Not Met
Breath Volume		0.066	02:24	
Air Blank		0.000	02:25	
Air Blank		0.000	02:26	
Subject Sample	# 3	0.140	02:27	
Breath Volume		1.792	02:27	
Air Blank		0.000	02:28	
Control Test		0.075	02:28	
Air Blank		0.000	02:29	
Diagnostics Check		OK	02:29	

Data Download Date/Time: 04/19/2019 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: STACHELSKI

DATE/TIME OF TEST: 04/14/2019 00:41:32

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:39	
Control Test	0.078	00:39	
Air Blank	0.000	00:39	
Subject Sample	# 1	00:40	Subject Test Refused
Breath Volume	0.000	00:40	
Air Blank	0.000	00:40	
Control Test	0.077	00:40	
Air Blank	0.000	00:41	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 04/19/2019 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: ARROYO SANTIAGO

DATE/TIME OF TEST: 04/20/2019 03:06:47

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.078	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.073	03:01
Breath Volume	1.566	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.075	03:05
Breath Volume	2.800	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 05/14/2019 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBROAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: LABARDY

DATE/TIME OF TEST: 04/21/2019 02:19:29

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.198	02:14
Breath Volume	2.425	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.185	02:17
Breath Volume	1.984	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 05/14/2019 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BELL,ETHAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: HENSHAW

DATE/TIME OF TEST: 04/26/2019 23:38:44

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:27	
Control Test	0.078	23:28	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:31	No Sample Provided
Breath Volume	0.000	23:31	
Air Blank	0.000	23:32	
Air Blank	0.000	23:34	
Subject Sample	# 2	23:37	No Sample Provided
Breath Volume	0.000	23:37	
Air Blank	0.000	23:37	
Control Test	0.078	23:38	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 05/14/2019 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WALSH,S
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A5
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SUMMERLIN

DATE/TIME OF TEST: 04/28/2019 03:10:56

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.078	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.171	03:06
Breath Volume	2.640	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.168	03:09
Breath Volume	3.242	03:09
Air Blank	0.000	03:09
Control Test	0.079	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 05/14/2019 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEVES,VIRGILIO,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ESTEVES,VIRGILIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: VERMILLION

DATE/TIME OF TEST: 04/20/2019 23:04:22

Diagnostics Check	OK	22:59	
Air Blank	0.000	22:59	
Control Test	0.081	23:00	
Air Blank	0.000	23:00	
Subject Sample # 1	0.115	23:01	
Breath Volume	1.292	23:01	
Air Blank	0.000	23:02	
Air Blank		23:03	Ambient Fail
Air Blank	0.000	23:04	

Data Download Date/Time: 04/30/2019 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEVES,VIRGILIO,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ESTEVES,VIRGILIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: VERMILLION

DATE/TIME OF TEST: 04/20/2019 23:14:34

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.114	23:09
Breath Volume	3.949	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.108	23:12
Breath Volume	3.687	23:12
Air Blank	0.000	23:13
Control Test	0.080	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 04/30/2019 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: CARCAMO ALMENDAREZ

DATE/TIME OF TEST: 04/07/2019 02:33:25

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.078	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.155	02:27
Breath Volume	1.542	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.153	02:31
Breath Volume	1.527	02:31
Air Blank	0.000	02:32
Control Test	0.076	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 05/02/2019 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: MIXER

DATE/TIME OF TEST: 04/08/2019 03:26:17

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.077	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.226	03:21
Breath Volume		2.609	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.231	03:24
Breath Volume		2.320	03:24
Air Blank		0.000	03:25
Control Test		0.077	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 05/02/2019 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: POTTS

DATE/TIME OF TEST: 04/13/2019 04:46:24

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.078	04:40
Air Blank	0.000	04:41
Subject Sample # 1	0.171	04:41
Breath Volume	1.761	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.188	04:44
Breath Volume	2.277	04:44
Air Blank	0.000	04:45
Control Test	0.077	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 05/02/2019 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: GALARZA JR

DATE/TIME OF TEST: 04/20/2019 02:21:01

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.114	02:16
Breath Volume	1.687	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:18
Subject Sample # 2	0.119	02:19
Breath Volume	2.195	02:19
Air Blank	0.000	02:19
Control Test	0.079	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:21

Data Download Date/Time: 05/02/2019 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: SETTLEMYRE

DATE/TIME OF TEST: 04/21/2019 02:39:35

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.134	02:34
Breath Volume	2.351	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.138	02:37
Breath Volume	1.808	02:37
Air Blank	0.000	02:38
Control Test	0.081	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 05/02/2019 16:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: SILVA SANCHEZ

DATE/TIME OF TEST: 04/21/2019 20:10:37

Diagnostics Check	OK	19:55	
Air Blank	0.000	19:56	
Control Test	0.080	19:56	
Air Blank	0.000	19:57	
Subject Sample	# 1 0.201	19:59	
Breath Volume	1.898	19:59	
Air Blank	0.000	20:00	
Air Blank	0.000	20:02	
Subject Sample	# 2 0.158	20:05	Volume Not Met
Breath Volume	0.562	20:05	
Air Blank	0.000	20:06	
Air Blank	0.000	20:07	
Subject Sample	# 3 0.199	20:08	
Breath Volume	1.433	20:08	
Air Blank	0.000	20:09	
Control Test	0.079	20:09	
Air Blank	0.000	20:10	
Diagnostics Check	OK	20:10	

Data Download Date/Time: 05/02/2019 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENYAN,M,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KENYAN,MA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: ENGLER

DATE/TIME OF TEST: 04/01/2019 23:18:55

Diagnostics Check	OK	23:12
Air Blank	0.000	23:13
Control Test	0.079	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.014	23:14
Breath Volume	2.617	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:16
Subject Sample # 2	0.014	23:17
Breath Volume	1.574	23:17
Air Blank	0.000	23:18
Control Test	0.079	23:18
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 04/22/2019 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: VANHORN,MICHAEL
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: ONEILL

DATE/TIME OF TEST: 04/11/2019 21:48:17

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.079	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.041	21:42
Breath Volume	1.355	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.038	21:46
Breath Volume	1.496	21:46
Air Blank	0.000	21:47
Control Test	0.078	21:47
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 04/22/2019 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: ROEHSLER

DATE/TIME OF TEST: 04/02/2019 21:59:24

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:50	
Control Test	0.078	21:51	
Air Blank	0.000	21:51	
Subject Sample # 1	0.204	21:52	
Breath Volume	1.433	21:52	
Air Blank	0.000	21:53	
Air Blank	0.000	21:55	
Subject Sample # 2		21:58	Subject Test Refused
Breath Volume	0.000	21:58	
Air Blank	0.000	21:58	
Control Test	0.078	21:58	
Air Blank	0.000	21:59	
Diagnostics Check	OK	21:59	

Data Download Date/Time: 04/04/2019 07:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 04/06/2019 02:02:34

Diagnostics Check	OK	01:59	
Air Blank	0.000	02:00	
Control Test	0.079	02:00	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:01	
Control Test	0.079	02:01	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 05/02/2019 09:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SWOGGER, EDWARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: LONGTINE

DATE/TIME OF TEST: 04/06/2019 18:02:10

Diagnostics Check	OK	17:55
Air Blank	0.000	17:56
Control Test	0.078	17:56
Air Blank	0.000	17:57
Subject Sample # 1	0.200	17:57
Breath Volume	2.425	17:57
Air Blank	0.000	17:58
Air Blank	0.000	18:00
Subject Sample # 2	0.211	18:00
Breath Volume	2.515	18:00
Air Blank	0.000	18:01
Control Test	0.077	18:01
Air Blank	0.000	18:02
Diagnostics Check	OK	18:02

Data Download Date/Time: 05/02/2019 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ JR,RAUL,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: GONZALEZ JR,RAUL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: FORTHMAN

DATE/TIME OF TEST: 04/06/2019 23:06:34

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:55	
Control Test		0.078	22:55	
Air Blank		0.000	22:56	
Subject Sample	# 1	0.000	22:59	Volume Not Met
Breath Volume		0.175	22:59	
Air Blank		0.000	23:00	
Air Blank		0.000	23:02	
Subject Sample	# 2	0.000	23:05	Volume Not Met
Breath Volume		0.050	23:05	
Air Blank		0.000	23:05	
Control Test		0.076	23:05	
Air Blank		0.000	23:06	
Diagnostics Check		OK	23:06	

Data Download Date/Time: 05/02/2019 09:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SWOGGER, EDWARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: RIVERA COLON

DATE/TIME OF TEST: 04/09/2019 22:39:27

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:36	
Control Test	0.078	22:37	
Air Blank	0.000	22:37	
Subject Sample	# 1	22:38	RFI Detect
Breath Volume	0.000	22:38	
Air Blank	0.000	22:38	
Control Test	0.078	22:38	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 05/02/2019 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SWOGGER, EDWARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: RIVERA COLON

DATE/TIME OF TEST: 04/09/2019 22:46:43

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.079	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.076	22:42
Breath Volume	1.945	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.074	22:45
Breath Volume	2.171	22:45
Air Blank	0.000	22:45
Control Test	0.079	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 05/02/2019 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: TURTON

DATE/TIME OF TEST: 04/10/2019 22:15:01

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:12	
Control Test	0.078	22:12	
Air Blank	0.000	22:12	
Subject Sample	# 1	22:14	Subject Test Refused
Breath Volume	0.000	22:14	
Air Blank		22:14	RFI Detect
Air Blank	0.000	22:15	

Data Download Date/Time: 05/02/2019 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: JEAN BAPTISTE

DATE/TIME OF TEST: 04/11/2019 00:07:46

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.078	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.131	00:03
Breath Volume	1.957	00:03
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.126	00:06
Breath Volume	1.656	00:06
Air Blank	0.000	00:06
Control Test	0.078	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 05/02/2019 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/12/2019 02:22:22

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:19	
Control Test	0.079	02:19	
Air Blank	0.000	02:20	
Subject Sample	# 1 0.202	02:20	
Breath Volume	1.691	02:20	
Air Blank		02:20	RFI Detect
Air Blank	0.000	02:21	
Control Test	0.078	02:21	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 05/02/2019 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/12/2019 02:31:25

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.078	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.193	02:26
Breath Volume	1.871	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:28
Subject Sample # 2	0.212	02:29
Breath Volume	2.414	02:29
Air Blank	0.000	02:30
Control Test	0.078	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 05/02/2019 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HANNA,WARDELL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: SCHLEMMER

DATE/TIME OF TEST: 04/16/2019 22:21:16

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:16	
Control Test	0.079	22:16	
Air Blank	0.000	22:17	
Subject Sample	# 1	22:19	Subject Test Refused
Breath Volume	0.000	22:19	
Air Blank	0.000	22:20	
Control Test	0.078	22:20	
Air Blank	0.000	22:21	
Diagnostics Check	OK	22:21	

Data Download Date/Time: 05/02/2019 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,L
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:33

SUBJECT'S LAST NAME: HERNANDEZ CAMPOS

DATE/TIME OF TEST: 04/20/2019 15:59:20

Diagnostics Check	OK	15:54	
Air Blank	0.000	15:55	
Control Test	0.078	15:55	
Air Blank	0.000	15:56	
Subject Sample	# 1	15:57	Subject Test Refused
Breath Volume	0.000	15:57	
Air Blank	0.000	15:58	
Control Test	0.078	15:58	
Air Blank	0.000	15:59	
Diagnostics Check	OK	15:59	

Data Download Date/Time: 05/02/2019 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLINA,IGNACIO,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MOLINA,IGNACIO
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 04/22/2019 23:47:25

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.097	23:42
Breath Volume	4.859	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.097	23:45
Breath Volume	4.828	23:45
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 05/02/2019 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLINA,IGNACIO,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MOLINA,IGNACIO
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: LAGO

DATE/TIME OF TEST: 04/23/2019 21:02:46

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.079	20:57
Air Blank	0.000	20:57
Subject Sample # 1	0.017	20:58
Breath Volume	3.578	20:58
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.017	21:01
Breath Volume	3.457	21:01
Air Blank	0.000	21:01
Control Test	0.078	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 05/02/2019 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: MERITHEW

DATE/TIME OF TEST: 04/24/2019 21:12:03

Diagnostics Check		OK	21:05
Air Blank		0.000	21:06
Control Test		0.078	21:06
Air Blank		0.000	21:06
Subject Sample	# 1	0.154	21:07
Breath Volume		2.671	21:07
Air Blank		0.000	21:08
Air Blank		0.000	21:09
Subject Sample	# 2	0.154	21:10
Breath Volume		3.968	21:10
Air Blank		0.000	21:11
Control Test		0.078	21:11
Air Blank		0.000	21:11
Diagnostics Check		OK	21:12

Data Download Date/Time: 05/02/2019 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/25/2019 02:29:08

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.000	02:24
Breath Volume	3.304	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.000	02:27
Breath Volume	2.824	02:27
Air Blank	0.000	02:28
Control Test	0.078	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 05/02/2019 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLINA,IGNACIO,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MOLINA,IGNACIO
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/27/2019 03:48:23

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:45	
Control Test	0.078	03:46	
Air Blank	0.000	03:46	
Subject Sample	# 1	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:47	
Control Test	0.079	03:47	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 05/02/2019 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLINA,IGNACIO,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: JACKSON,CODY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/28/2019 01:19:00

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:16	
Control Test	0.078	01:16	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:17	Subject Test Refused
Breath Volume	0.000	01:17	
Air Blank	0.000	01:18	
Control Test	0.078	01:18	
Air Blank	0.000	01:18	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 05/02/2019 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,JOEL,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RIOS,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: SCHREIBER

DATE/TIME OF TEST: 04/28/2019 21:54:28

Diagnostics Check	OK	21:51	
Air Blank	0.000	21:52	
Control Test	0.078	21:52	
Air Blank	0.000	21:52	
Subject Sample	# 1	21:53	Subject Test Refused
Breath Volume	0.000	21:53	
Air Blank	0.000	21:53	
Control Test	0.078	21:53	
Air Blank	0.000	21:54	
Diagnostics Check	OK	21:54	

Data Download Date/Time: 05/02/2019 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,L
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:41

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 04/29/2019 12:04:13

Diagnostics Check	OK	12:01	
Air Blank	0.000	12:01	
Control Test	0.078	12:02	
Air Blank	0.000	12:02	
Subject Sample	# 1	12:02	Subject Test Refused
Breath Volume	0.000	12:02	
Air Blank	0.000	12:03	
Control Test	0.079	12:03	
Air Blank	0.000	12:04	
Diagnostics Check	OK	12:04	

Data Download Date/Time: 05/02/2019 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,AARON,K
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: DALTON,JASON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: BUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: SHORE

DATE/TIME OF TEST: 04/27/2019 19:20:30

Diagnostics Check	OK	19:09	
Air Blank	0.000	19:09	
Control Test	0.079	19:09	
Air Blank	0.000	19:10	
Subject Sample	# 1	19:13	No Sample Provided
Breath Volume	0.000	19:13	
Air Blank	0.000	19:13	
Air Blank	0.000	19:15	
Subject Sample	# 2	19:19	No Sample Provided
Breath Volume	0.000	19:19	
Air Blank	0.000	19:19	
Control Test	0.079	19:19	
Air Blank	0.000	19:20	
Diagnostics Check	OK	19:20	

Data Download Date/Time: 05/20/2019 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORD,JOSHUA,N
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LORD,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: DEVANE

DATE/TIME OF TEST: 04/28/2019 03:17:22

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.143	03:12
Breath Volume	2.257	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.136	03:15
Breath Volume	2.187	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 05/20/2019 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIDSON,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: KLOBUCHAR,RICHARD
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: BUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: LEHMANN

DATE/TIME OF TEST: 04/28/2019 20:46:20

Diagnostics Check	OK	20:34	
Air Blank	0.000	20:35	
Control Test	0.079	20:35	
Air Blank	0.000	20:36	
Subject Sample	# 1	20:39	No Sample Provided
Breath Volume	0.000	20:39	
Air Blank	0.000	20:39	
Air Blank	0.000	20:41	
Subject Sample	# 2	20:44	No Sample Provided
Breath Volume	0.000	20:44	
Air Blank	0.000	20:45	
Control Test	0.079	20:45	
Air Blank	0.000	20:46	
Diagnostics Check	OK	20:46	

Data Download Date/Time: 05/20/2019 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUIZ,JONATHAN,O
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: RUIZ,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: WEBER

DATE/TIME OF TEST: 04/29/2019 22:02:42

Diagnostics Check	OK	21:53
Air Blank	0.000	21:54
Control Test	0.080	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.033	21:57
Breath Volume	1.226	21:57
Air Blank	0.000	21:58
Air Blank	0.000	21:59
Subject Sample # 2	0.030	22:01
Breath Volume	1.316	22:01
Air Blank	0.000	22:01
Control Test	0.079	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 05/20/2019 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIGAN,WILLIAM,B
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: CARR,DEVON
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 04/22/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 04/30/2019 21:13:59

Diagnostics Check	OK	21:11	
Air Blank	0.000	21:11	
Control Test	0.080	21:11	
Air Blank	0.000	21:12	
Subject Sample	# 1	21:12	Subject Test Refused
Breath Volume	0.000	21:12	
Air Blank	0.000	21:13	
Control Test	0.079	21:13	
Air Blank	0.000	21:13	
Diagnostics Check	OK	21:13	

Data Download Date/Time: 05/20/2019 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: CRUZ MURELLO

DATE/TIME OF TEST: 04/06/2019 01:39:57

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:34	
Control Test	0.082	01:35	
Air Blank	0.000	01:35	
Subject Sample # 1	0.200	01:38	Volume Not Met
Breath Volume	0.054	01:38	
Air Blank		01:39	Purge Fail
Air Blank	0.000	01:39	

Data Download Date/Time: 04/22/2019 09:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: CRUZ MURELLO

DATE/TIME OF TEST: 04/06/2019 01:46:15

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:42	
Control Test	0.081	01:42	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:44	Subject Test Refused
Breath Volume	0.035	01:44	
Air Blank	0.000	01:45	
Control Test	0.079	01:45	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 04/22/2019 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: VILLALBA ROIG

DATE/TIME OF TEST: 04/20/2019 22:55:43

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.080	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.127	22:50
Breath Volume	3.484	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.126	22:54
Breath Volume	1.757	22:54
Air Blank	0.000	22:54
Control Test	0.079	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 04/22/2019 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: CARMONA ALVA

DATE/TIME OF TEST: 04/21/2019 02:55:08

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.082	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.159	02:50
Breath Volume	3.031	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.155	02:53
Breath Volume	2.546	02:53
Air Blank	0.000	02:54
Control Test	0.083	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 04/22/2019 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MEJIA MATIAS

DATE/TIME OF TEST: 04/21/2019 23:17:03

Diagnostics Check	OK	23:09
Air Blank	0.000	23:10
Control Test	0.083	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.208	23:12
Breath Volume	2.550	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.213	23:15
Breath Volume	2.277	23:15
Air Blank	0.000	23:16
Control Test	0.083	23:16
Air Blank	0.000	23:16
Diagnostics Check	OK	23:17

Data Download Date/Time: 04/22/2019 09:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

INSTRUMENT REGISTERED TO: FHP TROOP L

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CLARKE,ROMAINE,S

BREATH TEST OPERATOR AGENCY: FHP Troop L

ARRESTING OFFICER: CLARKE,ROMAINE

ARRESTING OFFICER AGENCY: FHP Troop L

LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 915310

DRY GAS EXPIRY DATE: 04/27/2021

VIOLATION CODE: DUI

ARREST TIME: 15:00

GENDER: Female

OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 04/04/2019 15:23:57

Diagnostics Check		OK	15:17
Air Blank		0.000	15:17
Control Test		0.077	15:18
Air Blank		0.000	15:18
Subject Sample	# 1	0.235	15:19
Breath Volume		2.316	15:19
Air Blank		0.000	15:19
Air Blank		0.000	15:21
Subject Sample	# 2	0.239	15:22
Breath Volume		2.433	15:22
Air Blank		0.000	15:23
Control Test		0.078	15:23
Air Blank		0.000	15:23
Diagnostics Check		OK	15:23

Data Download Date/Time: 04/19/2019 21:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

INSTRUMENT REGISTERED TO: FHP TROOP L

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HATT,CODY,L

BREATH TEST OPERATOR AGENCY: FHP Troop L

ARRESTING OFFICER: MEER,BRADLEY

ARRESTING OFFICER AGENCY: FHP Troop L

LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 915310

DRY GAS EXPIRY DATE: 04/27/2021

VIOLATION CODE: DUI

ARREST TIME: 18:44

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: OLSEN

DATE/TIME OF TEST: 04/05/2019 19:48:42

Diagnostics Check	OK	19:41
Air Blank	0.000	19:42
Control Test	0.077	19:42
Air Blank	0.000	19:43
Subject Sample # 1	0.000	19:44
Breath Volume	1.828	19:44
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample # 2	0.000	19:47
Breath Volume	1.144	19:47
Air Blank	0.000	19:47
Control Test	0.076	19:48
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 04/19/2019 21:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

INSTRUMENT REGISTERED TO: FHP TROOP L

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HATT,CODY,L

BREATH TEST OPERATOR AGENCY: FHP Troop L

ARRESTING OFFICER: HATT,CODY

ARRESTING OFFICER AGENCY: FHP Troop L

LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 915310

DRY GAS EXPIRY DATE: 04/27/2021

VIOLATION CODE: DUI

ARREST TIME: 23:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: LOFTON

DATE/TIME OF TEST: 04/20/2019 23:53:47

Diagnostics Check		OK	23:40	
Air Blank		0.000	23:40	
Control Test		0.079	23:40	
Air Blank		0.000	23:41	
Subject Sample	# 1	0.068	23:44	Volume Not Met
Breath Volume		0.054	23:44	
Air Blank		0.000	23:45	
Air Blank		0.000	23:47	
Subject Sample	# 2	0.090	23:48	
Breath Volume		1.917	23:48	
Air Blank		0.000	23:49	
Air Blank		0.000	23:50	
Subject Sample	# 3	0.091	23:52	
Breath Volume		3.359	23:52	
Air Blank		0.000	23:52	
Control Test		0.077	23:53	
Air Blank		0.000	23:53	
Diagnostics Check		OK	23:53	

Data Download Date/Time: 05/23/2019 21:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

INSTRUMENT REGISTERED TO: FHP TROOP L

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HATT,CODY,L

BREATH TEST OPERATOR AGENCY: FHP Troop L

ARRESTING OFFICER: HATT,CODY

ARRESTING OFFICER AGENCY: FHP Troop L

LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 915310

DRY GAS EXPIRY DATE: 04/27/2021

VIOLATION CODE: DUI

ARREST TIME: 01:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 04/22/2019 02:23:25

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:21	
Control Test	0.078	02:21	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Control Test	0.077	02:22	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 05/23/2019 21:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MONGEON

DATE/TIME OF TEST: 04/25/2019 00:03:03

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.078	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.251	23:57
Breath Volume	2.164	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.253	00:01
Breath Volume	2.394	00:01
Air Blank	0.000	00:02
Control Test	0.078	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:03

Data Download Date/Time: 05/27/2019 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: ORTIZ ALBAREZ

DATE/TIME OF TEST: 04/30/2019 20:07:39

Diagnostics Check	OK	20:00
Air Blank	0.000	20:00
Control Test	0.078	20:01
Air Blank	0.000	20:01
Subject Sample # 1	0.203	20:03
Breath Volume	1.585	20:03
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.199	20:06
Breath Volume	1.398	20:06
Air Blank	0.000	20:06
Control Test	0.077	20:07
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 05/27/2019 18:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HEALY,GREGORY,B

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: HEALY,GREGORY

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 906364

DRY GAS EXPIRY DATE: 02/07/2021

VIOLATION CODE: DUI

ARREST TIME: 02:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: HIGGINBOTHAM

DATE/TIME OF TEST: 04/07/2019 02:46:04

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.081	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.158	02:41
Breath Volume	1.976	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:43
Subject Sample # 2	0.160	02:44
Breath Volume	2.445	02:44
Air Blank	0.000	02:45
Control Test	0.081	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:46

Data Download Date/Time: 04/24/2019 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ENNIS,JAMES
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 04/14/2019 23:26:43

Diagnostics Check	OK	23:19
Air Blank	0.000	23:20
Control Test	0.080	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.157	23:22
Breath Volume	1.449	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:24
Subject Sample # 2	0.157	23:25
Breath Volume	1.570	23:25
Air Blank	0.000	23:25
Control Test	0.079	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 04/24/2019 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: BALL
DATE/TIME OF TEST: 04/30/2019 18:46:26

Diagnostics Check	OK	18:34	
Air Blank	0.000	18:35	
Control Test	0.079	18:35	
Air Blank	0.000	18:36	
Subject Sample	# 1	18:39	No Sample Provided
Breath Volume	0.000	18:39	
Air Blank	0.000	18:39	
Air Blank	0.000	18:41	
Subject Sample	# 2	18:45	No Sample Provided
Breath Volume	0.000	18:45	
Air Blank	0.000	18:45	
Control Test	0.079	18:45	
Air Blank	0.000	18:46	
Diagnostics Check	OK	18:46	

Data Download Date/Time: 05/23/2019 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 04/30/2019 19:16:17

Diagnostics Check	OK	19:09
Air Blank	0.000	19:10
Control Test	0.079	19:10
Air Blank	0.000	19:11
Subject Sample # 1	0.182	19:11
Breath Volume	3.105	19:11
Air Blank	0.000	19:12
Air Blank	0.000	19:14
Subject Sample # 2	0.174	19:14
Breath Volume	2.972	19:14
Air Blank	0.000	19:15
Control Test	0.079	19:15
Air Blank	0.000	19:16
Diagnostics Check	OK	19:16

Data Download Date/Time: 05/23/2019 14:27

**Florida Department of Law Enforcement
Alcohol Testing Program**

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006769

INSTRUMENT REGISTERED TO: FHP TROOP L
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TODD,ZION,D
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GUNSAULLUS

DATE/TIME OF TEST: 04/19/2019 23:37:54

Diagnostics Check	OK	23:31
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.125	23:33
Breath Volume	2.468	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:35
Subject Sample # 2	0.122	23:36
Breath Volume	2.507	23:36
Air Blank	0.000	23:36
Control Test	0.081	23:37
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 04/24/2019 01:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006769

INSTRUMENT REGISTERED TO: FHP TROOP L	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 02618080A3
BREATH TEST OPERATOR: TODD,ZION,D	DRY GAS EXPIRY DATE: 03/05/2020
BREATH TEST OPERATOR AGENCY: FHP Troop L	VIOLATION CODE: DUI
ARRESTING OFFICER: TODD,ZION	ARREST TIME: 03:30
ARRESTING OFFICER AGENCY: FHP Troop L	GENDER: Male
LAST AGENCY INSPECTION: 04/19/2019	OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DOMINGUEZ PEREZ DATE/TIME OF TEST: 04/20/2019 04:05:28

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.081	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.253	04:00
Breath Volume	2.089	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.266	04:03
Breath Volume	2.464	04:03
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 04/24/2019 01:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006769

INSTRUMENT REGISTERED TO: FHP TROOP L

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TODD,ZION,D

BREATH TEST OPERATOR AGENCY: FHP Troop L

ARRESTING OFFICER: TODD,ZION

ARRESTING OFFICER AGENCY: FHP Troop L

LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02618080A3

DRY GAS EXPIRY DATE: 03/05/2020

VIOLATION CODE: DUI

ARREST TIME: 02:25

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: COMPER

DATE/TIME OF TEST: 04/24/2019 03:14:05

Diagnostics Check		OK	03:04
Air Blank		0.000	03:05
Control Test		0.081	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.165	03:06
Breath Volume		1.855	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.192	03:09
Breath Volume		1.921	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 3	0.169	03:12
Breath Volume		1.671	03:12
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 06/10/2019 19:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: HERRERIAS

DATE/TIME OF TEST: 04/14/2019 04:04:02

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.000	03:59
Breath Volume	1.894	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.000	04:02
Breath Volume	1.164	04:02
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:04

Data Download Date/Time: 05/08/2019 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATTORNO,ALEXANDER,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GATTORNO,ALEXANDER
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:54

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/30/2019 13:53:13

Diagnostics Check	OK	13:46
Air Blank	0.000	13:47
Control Test	0.079	13:47
Air Blank	0.000	13:48
Subject Sample # 1	0.314	13:48
Breath Volume	2.281	13:48
Air Blank	0.000	13:49
Air Blank	0.000	13:51
Subject Sample # 2	0.295	13:51
Breath Volume	1.835	13:51
Air Blank	0.000	13:52
Control Test	0.080	13:52
Air Blank	0.000	13:53
Diagnostics Check	OK	13:53

Data Download Date/Time: 05/08/2019 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CRAWFORD,BENJAMIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: LUJANO

DATE/TIME OF TEST: 04/04/2019 01:05:08

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:02	
Control Test	0.082	01:02	
Air Blank	0.000	01:03	
Subject Sample	# 1	01:03	Subject Test Refused
Breath Volume	0.000	01:03	
Air Blank	0.000	01:04	
Control Test	0.082	01:04	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 05/03/2019 05:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CRAWFORD,BENJAMIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: BARRON

DATE/TIME OF TEST: 04/06/2019 04:46:59

Diagnostics Check		OK	04:34	
Air Blank		0.000	04:34	
Control Test		0.080	04:35	
Air Blank		0.000	04:35	
Subject Sample	# 1	0.144	04:38	Volume Not Met
Breath Volume		0.511	04:38	
Air Blank		0.000	04:39	
Air Blank		0.000	04:41	
Subject Sample	# 2	0.183	04:42	
Breath Volume		1.929	04:42	
Air Blank		0.000	04:43	
Air Blank		0.000	04:44	
Subject Sample	# 3	0.170	04:45	
Breath Volume		1.363	04:45	
Air Blank		0.000	04:46	
Control Test		0.077	04:46	
Air Blank		0.000	04:46	
Diagnostics Check		OK	04:46	

Data Download Date/Time: 05/03/2019 05:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CRAWFORD,BEJAMIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: DOUGHERTY

DATE/TIME OF TEST: 04/07/2019 20:44:17

Diagnostics Check	OK	20:41	
Air Blank	0.000	20:41	
Control Test	0.080	20:42	
Air Blank	0.000	20:42	
Subject Sample	# 1	20:42	Subject Test Refused
Breath Volume	0.000	20:42	
Air Blank	0.000	20:43	
Control Test	0.080	20:43	
Air Blank	0.000	20:44	
Diagnostics Check	OK	20:44	

Data Download Date/Time: 05/03/2019 05:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CRAWFORD,BEJAMIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/17/2019 22:06:32

Diagnostics Check	OK	22:03	
Air Blank	0.000	22:04	
Control Test	0.080	22:04	
Air Blank	0.000	22:04	
Subject Sample	# 1	22:05	Subject Test Refused
Breath Volume	0.000	22:05	
Air Blank	0.000	22:05	
Control Test	0.079	22:05	
Air Blank	0.000	22:06	
Diagnostics Check	OK	22:06	

Data Download Date/Time: 05/03/2019 05:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,EDGARDO,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CARTER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: BRUNS

DATE/TIME OF TEST: 04/20/2019 04:14:04

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:11	
Control Test	0.081	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:12	Subject Test Refused
Breath Volume	0.000	04:12	
Air Blank	0.000	04:13	
Control Test	0.081	04:13	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:14	

Data Download Date/Time: 05/03/2019 05:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,DONALD,H
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: GREEN,DONALD
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 04/21/2019 04:54:11

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:51	
Control Test	0.082	04:52	
Air Blank	0.000	04:52	
Subject Sample	# 1	04:53	Sequence Aborted
Breath Volume	0.000	04:53	
Air Blank	0.000	04:54	

Data Download Date/Time: 05/03/2019 05:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,DONALD,H
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: GREEN,DONALD
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 04/21/2019 05:04:50

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.079	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.134	05:00
Breath Volume	1.800	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:02
Subject Sample # 2	0.137	05:03
Breath Volume	1.554	05:03
Air Blank	0.000	05:03
Control Test	0.077	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 05/03/2019 05:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CRAWFORD,BENJAMIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: SMITHEY

DATE/TIME OF TEST: 04/22/2019 22:42:22

Diagnostics Check		OK	22:34
Air Blank		0.000	22:34
Control Test		0.081	22:34
Air Blank		0.000	22:35
Subject Sample	# 1	0.156	22:36
Breath Volume		1.460	22:36
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.152	22:40
Breath Volume		1.636	22:40
Air Blank		0.000	22:41
Control Test		0.080	22:41
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 05/03/2019 05:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CRAWFORD,BENJAMIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 03/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: C230113764600

DATE/TIME OF TEST: 04/25/2019 02:12:18

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.081	02:10	
Air Blank	0.000	02:10	
Subject Sample	# 1	02:10	Subject Test Refused
Breath Volume	0.000	02:10	
Air Blank	0.000	02:11	
Control Test	0.081	02:11	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:12	

Data Download Date/Time: 05/03/2019 05:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CRAWFORD,BENJAMIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: CARCAMO

DATE/TIME OF TEST: 04/27/2019 02:49:40

Diagnostics Check	OK	02:46	
Air Blank	0.000	02:47	
Control Test	0.081	02:47	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:48	Subject Test Refused
Breath Volume	0.000	02:48	
Air Blank	0.000	02:48	
Control Test	0.081	02:49	
Air Blank	0.000	02:49	
Diagnostics Check	OK	02:49	

Data Download Date/Time: 05/03/2019 05:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARRO,CASIMIRO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: HERNANDEZ GONZALEZ

DATE/TIME OF TEST: 04/05/2019 21:25:59

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.077	21:17
Air Blank	0.000	21:18
Subject Sample # 1	0.191	21:21
Breath Volume	1.324	21:21
Air Blank	0.000	21:21
Air Blank	0.000	21:23
Subject Sample # 2	0.182	21:24
Breath Volume	1.332	21:24
Air Blank	0.000	21:25
Control Test	0.076	21:25
Air Blank	0.000	21:25
Diagnostics Check	OK	21:25

Data Download Date/Time: 04/18/2019 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONSEES,D,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ELLIS,T
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 04/01/2019 21:17:59

Diagnostics Check	OK	21:11
Air Blank	0.000	21:12
Control Test	0.078	21:12
Air Blank	0.000	21:13
Subject Sample # 1	0.000	21:13
Breath Volume	4.234	21:13
Air Blank	0.000	21:14
Air Blank	0.000	21:16
Subject Sample # 2	0.000	21:16
Breath Volume	2.949	21:16
Air Blank	0.000	21:17
Control Test	0.078	21:17
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,J,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ARNOLD,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 04/04/2019 00:03:38

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.078	23:58
Air Blank	0.000	23:58
Subject Sample # 2	0.218	23:59
Breath Volume	3.429	23:59
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.235	00:02
Breath Volume	2.429	00:02
Air Blank	0.000	00:02
Control Test	0.078	00:03
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARR,L,J
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BARBUTO,B
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: MAZARRIEGOS

DATE/TIME OF TEST: 04/04/2019 23:02:00

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.078	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.066	22:57
Breath Volume	2.554	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.067	23:00
Breath Volume	2.996	23:00
Air Blank	0.000	23:01
Control Test	0.078	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:02

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,E,W
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: DERRIG,J
ARRESTING OFFICER AGENCY: MARCO ISLAND PD
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: VELARDE

DATE/TIME OF TEST: 04/06/2019 02:03:06

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.079	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.154	01:58
Breath Volume	2.691	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.150	02:01
Breath Volume	3.007	02:01
Air Blank	0.000	02:02
Control Test	0.078	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRATT,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PRATT,N
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: SILHAVA

DATE/TIME OF TEST: 04/07/2019 02:30:44

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.079	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.258	02:26
Breath Volume	1.972	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.242	02:29
Breath Volume	1.898	02:29
Air Blank	0.000	02:29
Control Test	0.079	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: RUSSELL,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ANADELL

DATE/TIME OF TEST: 04/08/2019 23:47:54

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.078	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.252	23:43
Breath Volume	1.437	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.260	23:46
Breath Volume	1.199	23:46
Air Blank	0.000	23:46
Control Test	0.078	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GASH,CHASE
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: RIVES

DATE/TIME OF TEST: 04/11/2019 01:03:48

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.079	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.080	00:59
Breath Volume	2.425	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:01
Subject Sample # 2	0.079	01:02
Breath Volume	1.777	01:02
Air Blank	0.000	01:02
Control Test	0.078	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: DIORIO II

DATE/TIME OF TEST: 04/11/2019 20:29:04

Diagnostics Check	OK	20:20
Air Blank	0.000	20:21
Control Test	0.079	20:21
Air Blank	0.000	20:21
Subject Sample # 1	0.185	20:23
Breath Volume	1.792	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.188	20:27
Breath Volume	1.800	20:27
Air Blank	0.000	20:28
Control Test	0.077	20:28
Air Blank	0.000	20:28
Diagnostics Check	OK	20:29

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLMSTEAD,J,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GASH,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: RHODES

DATE/TIME OF TEST: 04/12/2019 01:11:26

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.077	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.047	01:06
Breath Volume	3.468	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.046	01:09
Breath Volume	2.687	01:09
Air Blank	0.000	01:10
Control Test	0.078	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,D,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: LEONI,S
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: IRIBAR

DATE/TIME OF TEST: 04/12/2019 01:29:28

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.077	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.175	01:24
Breath Volume	3.855	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.178	01:27
Breath Volume	3.781	01:27
Air Blank	0.000	01:28
Control Test	0.077	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOEL,Y,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: VALLE,M
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: JOSE BAUTISTA

DATE/TIME OF TEST: 04/14/2019 05:10:11

Diagnostics Check		OK	05:02
Air Blank		0.000	05:02
Control Test		0.081	05:03
Air Blank		0.000	05:03
Subject Sample	# 1	0.083	05:05
Breath Volume		1.675	05:05
Air Blank		0.000	05:06
Air Blank		0.000	05:08
Subject Sample	# 2	0.081	05:08
Breath Volume		1.656	05:08
Air Blank		0.000	05:09
Control Test		0.077	05:09
Air Blank		0.000	05:10
Diagnostics Check		OK	05:10

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: STECKMAN,N
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: GLYNN

DATE/TIME OF TEST: 04/15/2019 02:29:30

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.077	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.132	02:24
Breath Volume	1.410	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.141	02:27
Breath Volume	1.488	02:27
Air Blank	0.000	02:28
Control Test	0.077	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PEPITONE,MICHAEL
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: HIDALGO ORTEGA

DATE/TIME OF TEST: 04/16/2019 00:43:23

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.078	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.106	00:38
Breath Volume	1.324	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.103	00:41
Breath Volume	1.460	00:41
Air Blank	0.000	00:42
Control Test	0.077	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ELLIS,T
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: AJQUI CHINGO

DATE/TIME OF TEST: 04/17/2019 00:43:15

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.078	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.160	00:38
Breath Volume	2.097	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:40
Subject Sample # 2	0.158	00:41
Breath Volume	1.632	00:41
Air Blank	0.000	00:42
Control Test	0.078	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS, MATTHEW, L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MCCOY, J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 04/17/2019 04:31:45

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.078	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.000	04:27
Breath Volume	3.078	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.000	04:30
Breath Volume	2.152	04:30
Air Blank	0.000	04:30
Control Test	0.078	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARP,RYAN,M
BREATH TEST OPERATOR AGENCY: NAPLES PD
ARRESTING OFFICER: HARP,RYAN
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: PROYCE

DATE/TIME OF TEST: 04/18/2019 16:13:03

Diagnostics Check	OK	16:06
Air Blank	0.000	16:07
Control Test	0.078	16:07
Air Blank	0.000	16:07
Subject Sample # 1	0.317	16:08
Breath Volume	1.890	16:08
Air Blank	0.000	16:09
Air Blank	0.000	16:10
Subject Sample # 2	0.309	16:11
Breath Volume	1.718	16:11
Air Blank	0.000	16:12
Control Test	0.076	16:12
Air Blank	0.000	16:12
Diagnostics Check	OK	16:13

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOTO,SAMUEL,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: TRAVIS,NPD
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:44

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/20/2019 15:13:08

Diagnostics Check	OK	15:07
Air Blank	0.000	15:07
Control Test	0.078	15:07
Air Blank	0.000	15:08
Subject Sample # 1	0.109	15:08
Breath Volume	1.945	15:08
Air Blank	0.000	15:09
Air Blank	0.000	15:11
Subject Sample # 2	0.109	15:11
Breath Volume	1.917	15:11
Air Blank	0.000	15:12
Control Test	0.078	15:12
Air Blank	0.000	15:13
Diagnostics Check	OK	15:13

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNORAT,ADELY,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GASH,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: CULLEN

DATE/TIME OF TEST: 04/21/2019 03:06:26

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.078	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.171	03:01
Breath Volume	1.281	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.168	03:04
Breath Volume	2.070	03:04
Air Blank	0.000	03:05
Control Test	0.078	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ESPINOSA GOMEZ

DATE/TIME OF TEST: 04/21/2019 04:03:28

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.078	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.189	03:58
Breath Volume	3.035	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.185	04:01
Breath Volume	3.195	04:01
Air Blank	0.000	04:02
Control Test	0.077	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERTILUS,FREDELYNE,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: FARRO,L.
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: KASAP

DATE/TIME OF TEST: 04/21/2019 21:08:53

Diagnostics Check	OK	21:02
Air Blank	0.000	21:02
Control Test	0.079	21:03
Air Blank	0.000	21:03
Subject Sample # 1	0.156	21:04
Breath Volume	1.164	21:04
Air Blank	0.000	21:05
Air Blank	0.000	21:06
Subject Sample # 2	0.151	21:07
Breath Volume	1.429	21:07
Air Blank	0.000	21:07
Control Test	0.077	21:08
Air Blank	0.000	21:08
Diagnostics Check	OK	21:08

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAFFEE,C,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MCCOY,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 04/21/2019 22:07:11

Diagnostics Check		OK	21:57
Air Blank		0.000	21:57
Control Test		0.079	21:58
Air Blank		0.000	21:58
Subject Sample	# 1	0.304	21:59
Breath Volume		1.390	21:59
Air Blank		0.000	22:00
Air Blank		0.000	22:01
Subject Sample	# 2	0.280	22:02
Breath Volume		1.316	22:02
Air Blank		0.000	22:03
Air Blank		0.000	22:04
Subject Sample	# 3	0.288	22:05
Breath Volume		2.125	22:05
Air Blank		0.000	22:06
Control Test		0.077	22:06
Air Blank		0.000	22:07
Diagnostics Check		OK	22:07

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,K,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GROSSKLAS,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 04/23/2019 03:46:47

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.114	03:42
Breath Volume	3.492	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:44
Subject Sample # 2	0.112	03:45
Breath Volume	3.542	03:45
Air Blank	0.000	03:45
Control Test	0.076	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,S,C
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: REED,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: MULIAR

DATE/TIME OF TEST: 04/23/2019 20:28:07

Diagnostics Check	OK	20:21
Air Blank	0.000	20:22
Control Test	0.078	20:22
Air Blank	0.000	20:23
Subject Sample # 1	0.000	20:23
Breath Volume	1.453	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.000	20:26
Breath Volume	1.281	20:26
Air Blank	0.000	20:27
Control Test	0.079	20:27
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CARPENTER,CHRISTOPHER
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: FERRARA

DATE/TIME OF TEST: 04/26/2019 23:03:42

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.078	22:58
Air Blank	0.000	22:58
Subject Sample # 1	0.178	22:59
Breath Volume	2.097	22:59
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.173	23:02
Breath Volume	2.125	23:02
Air Blank	0.000	23:02
Control Test	0.078	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLINGS,EMANUEL,R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: STECKMAN,N
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: BOBOWSKI

DATE/TIME OF TEST: 04/27/2019 23:42:34

Diagnostics Check		OK	23:29	
Air Blank		0.000	23:30	
Control Test		0.079	23:30	
Air Blank		0.000	23:31	
Subject Sample	# 1	0.044	23:34	Volume Not Met
Breath Volume		0.105	23:34	
Air Blank		0.000	23:34	
Air Blank		0.000	23:36	
Subject Sample	# 2	0.041	23:37	
Breath Volume		2.031	23:37	
Air Blank		0.000	23:38	
Air Blank		0.000	23:40	
Subject Sample	# 3	0.039	23:41	
Breath Volume		1.792	23:41	
Air Blank		0.000	23:41	
Control Test		0.077	23:41	
Air Blank		0.000	23:42	
Diagnostics Check		OK	23:42	

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVAS,W,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: REED,R
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: CARCINI-MIGUEL

DATE/TIME OF TEST: 04/28/2019 06:07:32

Diagnostics Check	OK	06:00
Air Blank	0.000	06:01
Control Test	0.078	06:01
Air Blank	0.000	06:01
Subject Sample # 1	0.172	06:02
Breath Volume	1.554	06:02
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 2	0.185	06:05
Breath Volume	1.835	06:05
Air Blank	0.000	06:06
Control Test	0.077	06:06
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,Y.,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SHERAN,,
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/28/2019 23:43:41

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.079	23:38
Air Blank	0.000	23:38
Subject Sample # 1	0.000	23:39
Breath Volume	1.398	23:39
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.000	23:42
Breath Volume	2.414	23:42
Air Blank	0.000	23:42
Control Test	0.079	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 05/02/2019 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODI,DOMENIC,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ASHBY,N
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 03/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/01/2019 02:13:25

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.080	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.243	02:08
Breath Volume		1.949	02:08
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.258	02:11
Breath Volume		1.406	02:11
Air Blank		0.000	02:12
Control Test		0.079	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 04/01/2019 14:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JOSE,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: LOPEZ,JOSE
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: HERNANDEZ AGUILAR

DATE/TIME OF TEST: 04/14/2019 00:33:40

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.081	00:27
Air Blank		0.000	00:28
Subject Sample	# 1	0.205	00:28
Breath Volume		1.355	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.209	00:31
Breath Volume		1.574	00:31
Air Blank		0.000	00:32
Control Test		0.080	00:33
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 05/03/2019 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAFFARAS,CELINA,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HOLLINGSWORTH,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: LORT

DATE/TIME OF TEST: 04/01/2019 03:23:37

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.079	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.203	03:18
Breath Volume	4.328	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.194	03:21
Breath Volume	3.378	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 04/01/2019 06:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LISTER,ROBERT
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 04/03/2019 22:04:08

Diagnostics Check	OK	21:56
Air Blank	0.000	21:57
Control Test	0.079	21:57
Air Blank	0.000	21:58
Subject Sample # 1	0.223	21:59
Breath Volume	3.257	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:01
Subject Sample # 2	0.224	22:02
Breath Volume	3.878	22:02
Air Blank	0.000	22:03
Control Test	0.080	22:03
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 05/03/2019 21:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,BRADLEY,W
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCOTT,BRADLEY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: VILLELLA

DATE/TIME OF TEST: 04/04/2019 10:00:51

Diagnostics Check	OK	09:54
Air Blank	0.000	09:55
Control Test	0.080	09:55
Air Blank	0.000	09:56
Subject Sample # 1	0.000	09:56
Breath Volume	3.476	09:56
Air Blank	0.000	09:56
Air Blank	0.000	09:58
Subject Sample # 2	0.000	09:59
Breath Volume	3.691	09:59
Air Blank	0.000	09:59
Control Test	0.081	10:00
Air Blank	0.000	10:00
Diagnostics Check	OK	10:00

Data Download Date/Time: 05/03/2019 21:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,D,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: NELSON,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: VELASCO LOPEZ

DATE/TIME OF TEST: 04/07/2019 02:33:14

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.174	02:28
Breath Volume	2.582	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:30
Subject Sample # 2	0.172	02:31
Breath Volume	3.183	02:31
Air Blank	0.000	02:32
Control Test	0.081	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 05/03/2019 21:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,D,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PATTEN,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: DE LEON GOMEZ

DATE/TIME OF TEST: 04/07/2019 03:25:44

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.078	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.205	03:20
Breath Volume	2.687	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.218	03:24
Breath Volume	4.000	03:24
Air Blank	0.000	03:24
Control Test	0.082	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 05/03/2019 21:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,HENRY,E
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CREWS,HENRY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: VILLELLA

DATE/TIME OF TEST: 04/08/2019 10:41:14

Diagnostics Check	OK	10:35
Air Blank	0.000	10:35
Control Test	0.079	10:35
Air Blank	0.000	10:36
Subject Sample # 1	0.000	10:36
Breath Volume	3.609	10:36
Air Blank	0.000	10:37
Air Blank	0.000	10:39
Subject Sample # 2	0.000	10:39
Breath Volume	4.492	10:39
Air Blank	0.000	10:40
Control Test	0.080	10:40
Air Blank	0.000	10:41
Diagnostics Check	OK	10:41

Data Download Date/Time: 05/03/2019 21:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER,CHRIS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: FENTON

DATE/TIME OF TEST: 04/09/2019 01:51:26

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.079	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.151	01:46
Breath Volume	2.074	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.148	01:49
Breath Volume	2.261	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 05/03/2019 21:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KATSARELAS,KEVIN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: ALFARO-PEREZ

DATE/TIME OF TEST: 04/11/2019 23:17:56

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.198	23:13
Breath Volume	1.875	23:13
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.197	23:16
Breath Volume	2.375	23:16
Air Blank	0.000	23:16
Control Test	0.080	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 05/03/2019 21:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,DONNA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TAYLOR,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: VENER

DATE/TIME OF TEST: 04/15/2019 18:59:47

Diagnostics Check	OK	18:52
Air Blank	0.000	18:52
Control Test	0.080	18:53
Air Blank	0.000	18:53
Subject Sample # 1	0.213	18:54
Breath Volume	4.187	18:54
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.215	18:58
Breath Volume	2.550	18:58
Air Blank	0.000	18:58
Control Test	0.080	18:59
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

Data Download Date/Time: 05/03/2019 21:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ESTRADA,KEVIN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: DOUBLEDAY

DATE/TIME OF TEST: 04/16/2019 20:55:38

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.079	20:49
Air Blank	0.000	20:49
Subject Sample # 1	0.091	20:50
Breath Volume	1.867	20:50
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.086	20:54
Breath Volume	1.484	20:54
Air Blank	0.000	20:54
Control Test	0.081	20:55
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 05/03/2019 21:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BALTIMORE,KEVIN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/17/2019 01:37:16

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.134	01:32
Breath Volume	1.859	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:34
Subject Sample # 2	0.141	01:35
Breath Volume	1.398	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 05/03/2019 21:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORDSTROM,DEPUTY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TAMERIS,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: RUFF

DATE/TIME OF TEST: 04/20/2019 07:21:56

Diagnostics Check	OK	07:15
Air Blank	0.000	07:16
Control Test	0.080	07:16
Air Blank	0.000	07:16
Subject Sample # 1	0.079	07:17
Breath Volume	1.808	07:17
Air Blank	0.000	07:18
Air Blank	0.000	07:19
Subject Sample # 2	0.072	07:20
Breath Volume	1.335	07:20
Air Blank	0.000	07:20
Control Test	0.078	07:21
Air Blank	0.000	07:21
Diagnostics Check	OK	07:21

Data Download Date/Time: 05/03/2019 21:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,D,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LAROSA,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 04/21/2019 05:37:44

Diagnostics Check	OK	05:29
Air Blank	0.000	05:30
Control Test	0.080	05:30
Air Blank	0.000	05:30
Subject Sample # 1	0.000	05:32
Breath Volume	2.285	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.000	05:36
Breath Volume	2.222	05:36
Air Blank	0.000	05:36
Control Test	0.080	05:37
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 05/03/2019 21:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUCKER,DEPUTY,S
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JONES,JASON
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:06

SUBJECT'S LAST NAME: BYNUM

DATE/TIME OF TEST: 04/21/2019 07:36:35

Diagnostics Check	OK	07:30
Air Blank	0.000	07:30
Control Test	0.080	07:30
Air Blank	0.000	07:31
Subject Sample # 1	0.133	07:31
Breath Volume	4.039	07:31
Air Blank	0.000	07:32
Air Blank	0.000	07:34
Subject Sample # 2	0.117	07:34
Breath Volume	3.937	07:34
Air Blank	0.000	07:35
Control Test	0.079	07:35
Air Blank	0.000	07:36
Diagnostics Check	OK	07:36

Data Download Date/Time: 05/03/2019 21:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSZE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HOLLINGSWORTH,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: NICOLOSI

DATE/TIME OF TEST: 04/23/2019 20:18:54

Diagnostics Check	OK	20:12
Air Blank	0.000	20:12
Control Test	0.079	20:12
Air Blank	0.000	20:13
Subject Sample # 1	0.357	20:14
Breath Volume	3.902	20:14
Air Blank	0.000	20:15
Air Blank	0.000	20:16
Subject Sample # 2	0.368	20:17
Breath Volume	3.917	20:17
Air Blank	0.000	20:17
Control Test	0.078	20:18
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

Data Download Date/Time: 05/03/2019 21:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROSPER,BENSON,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: REID,N
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: YONAKA

DATE/TIME OF TEST: 04/23/2019 21:02:09

Diagnostics Check	OK	20:55
Air Blank	0.000	20:56
Control Test	0.079	20:56
Air Blank	0.000	20:56
Subject Sample # 1	0.181	20:57
Breath Volume	2.140	20:57
Air Blank	0.000	20:58
Air Blank	0.000	20:59
Subject Sample # 2	0.174	21:00
Breath Volume	1.742	21:00
Air Blank	0.000	21:01
Control Test	0.078	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 05/03/2019 21:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSZE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORAGHAN,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 04/23/2019 21:57:02

Diagnostics Check	OK	21:50
Air Blank	0.000	21:51
Control Test	0.079	21:51
Air Blank	0.000	21:52
Subject Sample # 1	0.000	21:52
Breath Volume	3.132	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.000	21:55
Breath Volume	3.238	21:55
Air Blank	0.000	21:56
Control Test	0.081	21:56
Air Blank	0.000	21:56
Diagnostics Check	OK	21:57

Data Download Date/Time: 05/03/2019 21:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FIORINO,R
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: BODE

DATE/TIME OF TEST: 04/25/2019 23:54:53

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.079	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.142	23:50
Breath Volume	2.148	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.152	23:53
Breath Volume	2.863	23:53
Air Blank	0.000	23:53
Control Test	0.079	23:54
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 05/03/2019 21:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSZE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: EARNER,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: SENNI

DATE/TIME OF TEST: 04/26/2019 22:25:05

Diagnostics Check	OK	22:15
Air Blank	0.000	22:15
Control Test	0.080	22:16
Air Blank	0.000	22:16
Subject Sample # 1	0.186	22:19
Breath Volume	2.050	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:21
Subject Sample # 2	0.180	22:23
Breath Volume	1.707	22:23
Air Blank	0.000	22:24
Control Test	0.079	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 05/03/2019 21:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MILLER,KAYLEE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 04/30/2019 04:25:52

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.159	04:21
Breath Volume		2.339	04:21
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.144	04:24
Breath Volume		1.714	04:24
Air Blank		0.000	04:24
Control Test		0.078	04:25
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 05/03/2019 21:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: SGRO,MICHAEL
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: VOZZOLA

DATE/TIME OF TEST: 04/19/2019 02:33:42

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.081	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.182	02:29
Breath Volume	1.714	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.193	02:32
Breath Volume	2.714	02:32
Air Blank	0.000	02:32
Control Test	0.081	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 05/09/2019 17:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: HART,STEVEN
ARRESTING OFFICER AGENCY: OTHER STATE
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: LOCK II

DATE/TIME OF TEST: 04/20/2019 03:39:27

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.081	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.057	03:34
Breath Volume	3.019	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.057	03:37
Breath Volume	3.113	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 05/09/2019 17:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: FORT LAUDERDALE PD
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: RUBIANO RAMOS

DATE/TIME OF TEST: 04/21/2019 03:04:09

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.081	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.111	02:59
Breath Volume	1.417	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:01
Subject Sample # 2	0.109	03:02
Breath Volume	1.980	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 05/09/2019 17:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: BERMAN,MARC
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: RODDY

DATE/TIME OF TEST: 04/22/2019 23:25:26

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.081	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.163	23:18
Breath Volume	1.222	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.163	23:23
Breath Volume	1.355	23:23
Air Blank	0.000	23:24
Control Test	0.081	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 05/09/2019 17:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: FORT LAUDERDALE PD
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: DELEON

DATE/TIME OF TEST: 04/25/2019 00:36:37

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.080	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.182	00:31
Breath Volume	2.476	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.185	00:35
Breath Volume	3.187	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 05/09/2019 17:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: ZA VALETA DIAZ

DATE/TIME OF TEST: 04/25/2019 02:32:01

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.228	02:27
Breath Volume	2.656	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:29
Subject Sample # 2	0.229	02:30
Breath Volume	3.281	02:30
Air Blank	0.000	02:31
Control Test	0.079	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:32

Data Download Date/Time: 05/09/2019 17:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: FORT LAUDERDALE PD
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BORN

DATE/TIME OF TEST: 04/26/2019 22:28:08

Diagnostics Check	OK	22:21
Air Blank	0.000	22:22
Control Test	0.081	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.190	22:23
Breath Volume	2.753	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:25
Subject Sample # 2	0.174	22:26
Breath Volume	1.140	22:26
Air Blank	0.000	22:27
Control Test	0.081	22:27
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 05/09/2019 17:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: IRELAN,ANDREW
ARRESTING OFFICER AGENCY: MARGATE PD
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 04/28/2019 04:14:09

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.205	04:09
Breath Volume		1.480	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:11
Subject Sample	# 2	0.203	04:12
Breath Volume		1.738	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 05/09/2019 17:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: FORT LAUDERDALE PD
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: MARTINEZ PEREZ

DATE/TIME OF TEST: 04/28/2019 04:44:03

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.080	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.108	04:39
Breath Volume	2.453	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.107	04:42
Breath Volume	2.417	04:42
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:44

Data Download Date/Time: 05/09/2019 17:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER STATE
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: CAJUSTE

DATE/TIME OF TEST: 04/28/2019 05:05:35

Diagnostics Check	OK	04:58
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	05:00
Subject Sample # 1	0.076	05:00
Breath Volume	1.843	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.076	05:04
Breath Volume	2.000	05:04
Air Blank	0.000	05:04
Control Test	0.080	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 05/09/2019 17:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: SGT BOHM,JONATHAN
ARRESTING OFFICER AGENCY: FORT LAUDERDALE PD
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: CALIXTO

DATE/TIME OF TEST: 04/28/2019 05:14:37

Diagnostics Check	OK	05:10	
Air Blank	0.000	05:10	
Control Test	0.080	05:10	
Air Blank	0.000	05:11	
Subject Sample	# 1	05:13	Slope Not Met
Breath Volume	0.472	05:13	
Air Blank	0.000	05:13	
Control Test	0.080	05:14	
Air Blank	0.000	05:14	
Diagnostics Check	OK	05:14	

Data Download Date/Time: 05/09/2019 17:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: SGT BOHM,JONATHAN
ARRESTING OFFICER AGENCY: FORT LAUDERDALE PD
LAST AGENCY INSPECTION: 04/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: CALIXTO

DATE/TIME OF TEST: 04/28/2019 05:44:45

Diagnostics Check	OK	05:38
Air Blank	0.000	05:38
Control Test	0.080	05:39
Air Blank	0.000	05:39
Subject Sample # 1	0.135	05:40
Breath Volume	1.753	05:40
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample # 2	0.140	05:43
Breath Volume	2.207	05:43
Air Blank	0.000	05:43
Control Test	0.079	05:44
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

Data Download Date/Time: 05/09/2019 17:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALAS,LUIS,E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SALAS,LUIS
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: STIFFLER

DATE/TIME OF TEST: 04/04/2019 00:07:47

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:56	
Control Test	0.078	23:56	
Air Blank	0.000	23:57	
Subject Sample	# 1	00:00	No Sample Provided
Breath Volume	0.000	00:00	
Air Blank	0.000	00:01	
Air Blank	0.000	00:03	
Subject Sample	# 2	00:06	No Sample Provided
Breath Volume	0.000	00:06	
Air Blank	0.000	00:06	
Control Test	0.080	00:07	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 04/05/2019 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALAS,LUIS,E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PARSONS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: SHUKRI

DATE/TIME OF TEST: 04/04/2019 03:05:44

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:56	
Control Test	0.078	02:57	
Air Blank	0.000	02:57	
Subject Sample	# 1 0.197	02:59	
Breath Volume	1.976	02:59	
Air Blank	0.000	03:00	
Air Blank	0.000	03:02	
Subject Sample	# 2 0.189	03:03	
Breath Volume	1.687	03:03	
Air Blank	0.000	03:04	
Control Test	0.077	03:04	
Air Blank	0.000	03:05	
Diagnostics Check	Fail	03:05	Voltage/Current Test
Air Blank	0.000	03:05	

Data Download Date/Time: 04/05/2019 09:57