

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAYS,DARRYL,A  
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD  
ARRESTING OFFICER: MAYS,DARRYL  
ARRESTING OFFICER AGENCY: HOMESTEAD PD  
LAST AGENCY INSPECTION: 04/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 937240  
DRY GAS EXPIRY DATE: 04/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: MARTIN CARRILLO

DATE/TIME OF TEST: 05/08/2019 01:02:53

---

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.203	00:58
Breath Volume		1.367	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:00
Subject Sample	# 2	0.196	01:01
Breath Volume		2.726	01:01
Air Blank		0.000	01:01
Control Test		0.078	01:02
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

---

Data Download Date/Time: 05/28/2019 16:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,P  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: WALSH,S  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1099238  
DRY GAS EXPIRY DATE: 03/01/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: HERNADEZ-HERNANDEZ

DATE/TIME OF TEST: 05/02/2019 20:38:39

---

Diagnostics Check	OK	20:29
Air Blank	0.000	20:29
Control Test	0.080	20:30
Air Blank	0.000	20:30
Subject Sample # 1	0.204	20:33
Breath Volume	1.617	20:33
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.190	20:36
Breath Volume	1.269	20:36
Air Blank	0.000	20:37
Control Test	0.079	20:37
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

---

Data Download Date/Time: 05/16/2019 16:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,P  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANDI  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1099238  
DRY GAS EXPIRY DATE: 03/01/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 05/12/2019 01:45:08

---

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.100	01:39
Breath Volume	1.816	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:41
Subject Sample # 2	0.098	01:43
Breath Volume	1.414	01:43
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

---

Data Download Date/Time: 05/16/2019 16:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIVINS,HEATHER,C  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: INGERSOLL,NICK  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MACKKEY

DATE/TIME OF TEST: 05/12/2019 02:52:04

---

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.076	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.174	02:47
Breath Volume	1.308	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.174	02:50
Breath Volume	1.484	02:50
Air Blank	0.000	02:51
Control Test	0.075	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:52

---

Data Download Date/Time: 05/27/2019 16:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIVINS, HEATHER, C  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: APPLGATE, A  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: DA COSTA

DATE/TIME OF TEST: 05/12/2019 23:08:28

---

Diagnostics Check	OK	22:56		
Air Blank	0.000	22:57		
Control Test	0.077	22:57		
Air Blank	0.000	22:58		
Subject Sample	# 1	23:01	No Sample Provided	
Breath Volume	0.000	23:01		
Air Blank	0.000	23:01		
Air Blank	0.000	23:03		
Subject Sample	# 2	0.064	23:06	Volume Not Met
Breath Volume	0.015	23:06		
Air Blank	0.000	23:07		
Control Test	0.078	23:07		
Air Blank	0.000	23:08		
Diagnostics Check	OK	23:08		

---

Data Download Date/Time: 05/27/2019 16:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIVINS,HEATHER,C  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: APPLGATE,A  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: DA COSTA

DATE/TIME OF TEST: 05/12/2019 23:20:31

---

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.076	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.069	23:15
Breath Volume	1.453	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.078	23:18
Breath Volume	1.843	23:18
Air Blank	0.000	23:19
Control Test	0.076	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

---

Data Download Date/Time: 05/27/2019 16:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHAOUNI,K,  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: QUIMBAYO,M  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: WITHERSPOON

DATE/TIME OF TEST: 05/18/2019 03:37:06

---

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:33	
Control Test	0.077	03:33	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:35	Subject Test Refused
Breath Volume	0.000	03:35	
Air Blank	0.000	03:36	
Control Test	0.078	03:36	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

---

Data Download Date/Time: 05/27/2019 16:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PILL,JEREMY,L  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: OUELLETTE,SARAH  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: KRUSE

DATE/TIME OF TEST: 05/19/2019 02:45:17

---

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.077	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.127	02:40
Breath Volume	2.652	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.122	02:43
Breath Volume	3.285	02:43
Air Blank	0.000	02:44
Control Test	0.075	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

---

Data Download Date/Time: 05/27/2019 16:21



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIVINS, HEATHER, C  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: WEIDNER, CODY  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: VOSNICK

DATE/TIME OF TEST: 05/21/2019 03:00:48

---

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.078	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1 0.174	02:55	
Breath Volume	1.660	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:58	
Subject Sample	# 2 0.155	02:59	
Breath Volume	1.457	02:59	
Air Blank	0.000	02:59	
Control Test	0.074	03:00	Control Outside Tolerance
Air Blank	0.000	03:00	

---

Data Download Date/Time: 05/27/2019 16:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIVINS, HEATHER, C  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: WEIDNER, CODY  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: VOSNICK

DATE/TIME OF TEST: 05/21/2019 03:09:50

---

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.076	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.170	03:04
Breath Volume		1.355	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.169	03:08
Breath Volume		1.500	03:08
Air Blank		0.000	03:08
Control Test		0.076	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

---

Data Download Date/Time: 05/27/2019 16:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIVINS,HEATHER,C  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: BIVINS,HEATHER  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/25/2019 03:33:28

---

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.161	03:28
Breath Volume	1.433	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.150	03:31
Breath Volume	1.261	03:31
Air Blank	0.000	03:32
Control Test	0.076	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

---

Data Download Date/Time: 05/27/2019 16:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIVINS, HEATHER, C  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: BROOKS, NEILL  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: LAWS

DATE/TIME OF TEST: 05/26/2019 05:24:12

---

Diagnostics Check	OK	05:17
Air Blank	0.000	05:18
Control Test	0.078	05:18
Air Blank	0.000	05:19
Subject Sample # 1	0.129	05:19
Breath Volume	2.183	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.126	05:22
Breath Volume	1.406	05:22
Air Blank	0.000	05:23
Control Test	0.076	05:23
Air Blank	0.000	05:24
Diagnostics Check	OK	05:24

---

Data Download Date/Time: 05/27/2019 16:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KEARNEY, BRANDON, G  
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)  
ARRESTING OFFICER: LIGHTSEY, DUSTIN  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: BUI  
ARREST TIME: 17:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 05/26/2019 17:46:30

---

Diagnostics Check	OK	17:39	
Air Blank	0.000	17:40	
Control Test	0.077	17:40	
Air Blank	0.000	17:41	
Subject Sample # 1	0.064	17:41	
Breath Volume	1.136	17:41	
Air Blank	0.000	17:42	
Air Blank	0.000	17:44	
Subject Sample # 2	0.068	17:44	
Breath Volume	1.519	17:44	
Air Blank	0.000	17:45	
Control Test	0.074	17:45	Control Outside Tolerance
Air Blank	0.000	17:46	

---

Data Download Date/Time: 05/27/2019 16:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KEARNEY,BRANDON,G  
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)  
ARRESTING OFFICER: LIGHTSEY,DUSTIN  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A3  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: BUI  
ARREST TIME: 17:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 05/26/2019 17:50:14

---

Diagnostics Check	OK	17:48	
Air Blank	0.000	17:49	
Control Test	0.072	17:49	Control Outside Tolerance
Air Blank	0.000	17:50	

---

Data Download Date/Time: 05/27/2019 16:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COREY,GAVIN,T  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: SCELFO,ZACHARY  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 437738  
DRY GAS EXPIRY DATE: 06/08/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: STORY

DATE/TIME OF TEST: 05/12/2019 02:47:03

---

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.172	02:42
Breath Volume	2.140	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:44
Subject Sample # 2	0.168	02:45
Breath Volume	2.085	02:45
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:47

---

Data Download Date/Time: 05/28/2019 22:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COREY,GAVIN,T  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: SCELFO,ZACHARY  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 437738  
DRY GAS EXPIRY DATE: 06/08/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: SAUCIER

DATE/TIME OF TEST: 05/14/2019 01:10:41

---

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.082	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.118	01:05
Breath Volume	2.574	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.122	01:09
Breath Volume	3.136	01:09
Air Blank	0.000	01:09
Control Test	0.081	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

---

Data Download Date/Time: 05/28/2019 22:43



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HESTER,JAMES,R  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: MORALES,R  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 437738  
DRY GAS EXPIRY DATE: 06/08/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MENDOZA-GALVAN

DATE/TIME OF TEST: 05/26/2019 03:14:21

---

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.081	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.171	03:09
Breath Volume		1.460	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.174	03:12
Breath Volume		1.300	03:12
Air Blank		0.000	03:13
Control Test		0.081	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

---

Data Download Date/Time: 05/28/2019 22:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HESTER,JAMES,R  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: ANDERSON,K  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 437738  
DRY GAS EXPIRY DATE: 06/08/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: PEREZ GOMEZ

DATE/TIME OF TEST: 05/26/2019 03:30:22

---

Diagnostics Check		OK	03:24
Air Blank		0.000	03:24
Control Test		0.081	03:25
Air Blank		0.000	03:25
Subject Sample	# 1	0.186	03:25
Breath Volume		1.460	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.188	03:28
Breath Volume		1.468	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

---

Data Download Date/Time: 05/28/2019 22:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCNEIL,DARRIN,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: COLON,RAYMOND  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 437738  
DRY GAS EXPIRY DATE: 06/08/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: VARGAS RODRIGUEZ

DATE/TIME OF TEST: 05/28/2019 20:18:42

---

Diagnostics Check		OK	20:02	
Air Blank		0.000	20:02	
Control Test		0.083	20:03	
Air Blank		0.000	20:03	
Subject Sample	# 1	0.292	20:06	Volume Not Met
Breath Volume		0.359	20:06	
Air Blank		0.000	20:07	
Air Blank		0.000	20:09	
Subject Sample	# 2	0.278	20:11	
Breath Volume		1.902	20:11	
Air Blank		0.000	20:12	
Air Blank		0.000	20:13	
Subject Sample	# 3	0.279	20:16	Volume Not Met
Breath Volume		0.136	20:16	
Air Blank		0.000	20:17	
Control Test		0.078	20:18	
Air Blank		0.000	20:18	
Diagnostics Check		OK	20:18	

---

Data Download Date/Time: 05/28/2019 22:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: GESELL

DATE/TIME OF TEST: 05/07/2019 01:40:31

---

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.078	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.214	01:34
Breath Volume		2.093	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.209	01:38
Breath Volume		1.617	01:38
Air Blank		0.000	01:39
Control Test		0.079	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

---

Data Download Date/Time: 05/20/2019 09:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: PRESSWOOD

DATE/TIME OF TEST: 05/11/2019 02:45:57

---

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.078	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.141	02:41
Breath Volume	1.679	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:43
Subject Sample # 2	0.136	02:44
Breath Volume	1.625	02:44
Air Blank	0.000	02:45
Control Test	0.077	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

---

Data Download Date/Time: 05/20/2019 09:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: BAUTISTA MARTINEZ

DATE/TIME OF TEST: 05/12/2019 01:57:22

---

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.079	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.195	01:52
Breath Volume	1.140	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.193	01:55
Breath Volume	1.816	01:55
Air Blank	0.000	01:56
Control Test	0.077	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

---

Data Download Date/Time: 05/20/2019 09:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: MAYO

DATE/TIME OF TEST: 05/12/2019 22:10:29

---

Diagnostics Check	OK	22:03
Air Blank	0.000	22:04
Control Test	0.078	22:04
Air Blank	0.000	22:04
Subject Sample # 1	0.306	22:05
Breath Volume	1.464	22:05
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.294	22:08
Breath Volume	1.753	22:08
Air Blank	0.000	22:09
Control Test	0.076	22:09
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

---

Data Download Date/Time: 05/20/2019 09:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: MOROSALES

DATE/TIME OF TEST: 05/15/2019 23:51:12

---

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.127	23:46
Breath Volume	1.679	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.122	23:49
Breath Volume	1.593	23:49
Air Blank	0.000	23:50
Control Test	0.077	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

---

Data Download Date/Time: 05/20/2019 09:28



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GWODZ,OFC,R  
BREATH TEST OPERATOR AGENCY: Bradenton PD  
ARRESTING OFFICER: PELLEGRINE,OFC  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 05/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DAVENPORT

DATE/TIME OF TEST: 05/21/2019 02:16:49

---

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:04	
Control Test	0.078	02:05	
Air Blank	0.000	02:05	
Subject Sample # 1	0.231	02:06	
Breath Volume	1.742	02:06	
Air Blank	0.000	02:06	
Air Blank	0.000	02:08	
Subject Sample # 2	0.221	02:11	Slope Not Level
Breath Volume	1.476	02:11	
Air Blank	0.000	02:12	
Air Blank	0.000	02:14	
Subject Sample # 3	0.223	02:15	
Breath Volume	1.187	02:15	
Air Blank	0.000	02:15	
Control Test	0.074	02:16	Control Outside Tolerance
Air Blank	0.000	02:16	

---

Data Download Date/Time: 06/25/2019 09:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 05/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 05/25/2019 01:41:26

---

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.078	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.176	01:36
Breath Volume	2.191	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.177	01:39
Breath Volume	2.179	01:39
Air Blank	0.000	01:40
Control Test	0.078	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

---

Data Download Date/Time: 06/25/2019 09:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 05/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: UTOSH

DATE/TIME OF TEST: 05/25/2019 22:42:46

---

Diagnostics Check	OK	22:35
Air Blank	0.000	22:36
Control Test	0.077	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.173	22:37
Breath Volume	2.085	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.181	22:41
Breath Volume	1.781	22:41
Air Blank	0.000	22:41
Control Test	0.077	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

---

Data Download Date/Time: 06/25/2019 09:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMPSON,STEVEN,M  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SIMPSON,STEVEN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 05/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:06

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 05/27/2019 15:30:34

---

Diagnostics Check	OK	15:27	
Air Blank	0.000	15:27	
Control Test	0.078	15:27	
Air Blank	0.000	15:28	
Subject Sample	# 1	15:28	Interferent Detect
Breath Volume	2.636	15:28	
Air Blank	0.000	15:29	
Control Test	0.078	15:29	
Air Blank	0.000	15:30	
Diagnostics Check	OK	15:30	

---

Data Download Date/Time: 06/25/2019 09:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMPSON,STEVEN,M  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SIMPSON,STEVEN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 05/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:06

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 05/27/2019 15:40:14

---

Diagnostics Check	OK	15:36	
Air Blank	0.000	15:37	
Control Test	0.077	15:37	
Air Blank	0.000	15:37	
Subject Sample	# 1	15:38	Interferent Detect
Breath Volume	2.796	15:38	
Air Blank	0.000	15:39	
Control Test	0.077	15:39	
Air Blank	0.000	15:40	
Diagnostics Check	OK	15:40	

---

Data Download Date/Time: 06/25/2019 09:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: MONTEMAGNO,CARLO  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 05/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: LOPES

DATE/TIME OF TEST: 05/29/2019 02:44:39

---

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.078	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.075	02:39
Breath Volume	2.992	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.074	02:43
Breath Volume	3.750	02:43
Air Blank	0.000	02:43
Control Test	0.078	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

---

Data Download Date/Time: 06/25/2019 09:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUICK,BENJAMIN,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: QUICK,BENJAMIN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 05/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: LAMDA

DATE/TIME OF TEST: 05/31/2019 21:24:39

---

Diagnostics Check	OK	21:18
Air Blank	0.000	21:18
Control Test	0.079	21:19
Air Blank	0.000	21:19
Subject Sample # 1	0.000	21:20
Breath Volume	3.046	21:20
Air Blank	0.000	21:20
Air Blank	0.000	21:22
Subject Sample # 2	0.000	21:23
Breath Volume	3.218	21:23
Air Blank	0.000	21:23
Control Test	0.076	21:24
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

---

Data Download Date/Time: 06/25/2019 09:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANTORO,CAMERON,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 04/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/02/2019 00:43:27

---

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:40	
Control Test	0.081	00:41	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:42	Subject Test Refused
Breath Volume	0.000	00:42	
Air Blank	0.000	00:42	
Control Test	0.080	00:42	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

---

Data Download Date/Time: 05/24/2019 16:18



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANTORO,CAMERON,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 04/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 05/02/2019 04:02:19

---

Diagnostics Check		OK	03:49	
Air Blank		0.000	03:49	
Control Test		0.081	03:50	
Air Blank		0.000	03:50	
Subject Sample	# 1	0.125	03:53	Volume Not Met
Breath Volume		0.531	03:53	
Air Blank		0.000	03:54	
Air Blank		0.000	03:56	
Subject Sample	# 2	0.160	03:57	
Breath Volume		1.152	03:57	
Air Blank		0.000	03:57	
Air Blank		0.000	03:59	
Subject Sample	# 3	0.162	04:00	
Breath Volume		1.117	04:00	
Air Blank		0.000	04:01	
Control Test		0.079	04:01	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

---

Data Download Date/Time: 05/24/2019 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARCHER, JOSHUA, J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: BAUMGARTNER, NATHANIAL  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 04/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 13:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:58

SUBJECT'S LAST NAME: NADAL

DATE/TIME OF TEST: 05/02/2019 15:32:40

---

Diagnostics Check	OK	15:23
Air Blank	0.000	15:24
Control Test	0.080	15:24
Air Blank	0.000	15:25
Subject Sample # 1	0.235	15:26
Breath Volume	1.617	15:26
Air Blank	0.000	15:27
Air Blank	0.000	15:29
Subject Sample # 2	0.235	15:31
Breath Volume	1.335	15:31
Air Blank	0.000	15:31
Control Test	0.079	15:32
Air Blank	0.000	15:32
Diagnostics Check	OK	15:32

---

Data Download Date/Time: 05/24/2019 16:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARCHER, JOSHUA, J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY, RYAN  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 04/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: RUSTIA

DATE/TIME OF TEST: 05/07/2019 03:02:41

---

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.277	02:54
Breath Volume	4.097	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:56
Subject Sample # 2	0.211	02:57
Breath Volume	1.117	02:57
Air Blank	0.000	02:58
Air Blank	0.000	02:59
Subject Sample # 3	0.263	03:00
Breath Volume	2.437	03:00
Air Blank	0.000	03:01
Control Test	0.077	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

---

Data Download Date/Time: 05/24/2019 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALTSTATT, MATTHEW,  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: BRIGGS, G  
ARRESTING OFFICER AGENCY: Orange Park PD  
LAST AGENCY INSPECTION: 04/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: DOHERTY

DATE/TIME OF TEST: 05/11/2019 04:37:16

---

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:24	
Control Test	0.080	04:24	
Air Blank	0.000	04:25	
Subject Sample # 1	0.160	04:26	
Breath Volume	1.335	04:26	
Air Blank	0.000	04:27	
Air Blank	0.000	04:29	
Subject Sample # 2		04:32	No Sample Provided
Breath Volume	0.000	04:32	
Air Blank	0.000	04:32	
Air Blank	0.000	04:34	
Subject Sample # 3		04:35	Subject Test Refused
Breath Volume	0.000	04:35	
Air Blank	0.000	04:36	
Control Test	0.080	04:36	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

---

Data Download Date/Time: 05/24/2019 16:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINSON,C,A  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: BAUMGARTNER,N  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 04/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 07:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 05/11/2019 09:48:22

---

Diagnostics Check	OK	09:33	
Air Blank	0.000	09:33	
Control Test	0.080	09:34	
Air Blank	0.000	09:34	
Subject Sample	# 1 0.122	09:35	
Breath Volume	2.027	09:35	
Air Blank	0.000	09:36	
Air Blank	0.000	09:38	
Subject Sample	# 2	09:41	No Sample Provided
Breath Volume	0.000	09:41	
Air Blank	0.000	09:41	
Air Blank	0.000	09:43	
Subject Sample	# 3	09:47	No Sample Provided
Breath Volume	0.000	09:47	
Air Blank	0.000	09:47	
Control Test	0.080	09:47	
Air Blank	0.000	09:48	
Diagnostics Check	OK	09:48	

---

Data Download Date/Time: 05/24/2019 16:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINSON,C,A  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: BAUMGARTNER,N  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 04/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 07:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 05/11/2019 10:25:57

---

Diagnostics Check	OK	10:11	
Air Blank	0.000	10:11	
Control Test	0.080	10:12	
Air Blank	0.000	10:12	
Subject Sample # 1	0.097	10:13	
Breath Volume	1.496	10:13	
Air Blank	0.000	10:13	
Air Blank	0.000	10:15	
Subject Sample # 2		10:18	No Sample Provided
Breath Volume	0.000	10:18	
Air Blank	0.000	10:19	
Air Blank	0.000	10:21	
Subject Sample # 3		10:24	No Sample Provided
Breath Volume	0.000	10:24	
Air Blank	0.000	10:25	
Control Test	0.080	10:25	
Air Blank	0.000	10:25	
Diagnostics Check	OK	10:25	

---

Data Download Date/Time: 05/24/2019 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINSON,C,A  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: BAUMGARTNER,N  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 04/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 07:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 05/11/2019 10:58:06

---

Diagnostics Check	OK	10:51
Air Blank	0.000	10:52
Control Test	0.080	10:52
Air Blank	0.000	10:53
Subject Sample # 1	0.093	10:53
Breath Volume	1.449	10:53
Air Blank	0.000	10:54
Air Blank	0.000	10:56
Subject Sample # 2	0.094	10:56
Breath Volume	1.640	10:56
Air Blank	0.000	10:57
Control Test	0.079	10:57
Air Blank	0.000	10:58
Diagnostics Check	OK	10:58

---

Data Download Date/Time: 05/24/2019 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASLINE, T'NEISHA, S  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY, R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 04/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: CORBETT

DATE/TIME OF TEST: 05/16/2019 19:01:56

---

Diagnostics Check	OK	18:53	
Air Blank	0.000	18:53	
Control Test	0.080	18:54	
Air Blank	0.000	18:54	
Subject Sample	# 1	18:57	No Sample Provided
Breath Volume	0.000	18:57	
Air Blank	0.000	18:58	
Air Blank	0.000	19:00	
Subject Sample	# 2	19:00	Subject Test Refused
Breath Volume	0.000	19:00	
Air Blank	0.000	19:01	
Control Test	0.080	19:01	
Air Blank	0.000	19:01	
Diagnostics Check	OK	19:01	

---

Data Download Date/Time: 05/24/2019 16:18



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RENTZ,JARROD,C  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: FALL,J  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 04/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: YARBROUGH

DATE/TIME OF TEST: 05/19/2019 01:58:58

---

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.080	01:52
Air Blank	0.000	01:53
Subject Sample # 1	0.151	01:53
Breath Volume	1.800	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.158	01:57
Breath Volume	2.183	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

---

Data Download Date/Time: 05/24/2019 16:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RENTZ,JARROD,C  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: MITCHELL,C  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 04/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SAKRASKA

DATE/TIME OF TEST: 05/19/2019 02:55:53

---

Diagnostics Check	OK	02:55	
Air Blank		02:55	RFI Detect
Air Blank	0.000	02:55	

---

Data Download Date/Time: 05/24/2019 16:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARLOW,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: LOVETT,KEITH  
ARRESTING OFFICER AGENCY: Orange Park PD  
LAST AGENCY INSPECTION: 04/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: APPLE

DATE/TIME OF TEST: 05/20/2019 03:24:59

---

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:16	
Control Test	0.080	03:16	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:20	No Sample Provided
Breath Volume	0.000	03:20	
Air Blank	0.000	03:20	
Air Blank	0.000	03:22	
Subject Sample	# 2	03:23	Subject Test Refused
Breath Volume	0.000	03:23	
Air Blank	0.000	03:24	
Control Test	0.080	03:24	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

---

Data Download Date/Time: 05/24/2019 16:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANTORO,CAMERON,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 04/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/21/2019 00:57:51

---

Diagnostics Check	OK	00:56	
Air Blank	0.000	00:56	
Control Test	0.080	00:57	
Air Blank		00:57	RFI Detect
Air Blank	0.000	00:57	

---

Data Download Date/Time: 05/24/2019 16:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANTORO,CAMERON,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 04/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/21/2019 01:06:43

---

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.203	01:01
Breath Volume	3.109	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.202	01:05
Breath Volume	3.179	01:05
Air Blank	0.000	01:05
Control Test	0.079	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

---

Data Download Date/Time: 05/24/2019 16:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANTORO,CAMERON,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 04/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: STACKMAN

DATE/TIME OF TEST: 05/21/2019 21:07:17

---

Diagnostics Check	OK	21:04	
Air Blank	0.000	21:05	
Control Test	0.080	21:05	
Air Blank	0.000	21:05	
Subject Sample	# 1	21:06	RFI Detect
Breath Volume	2.179	21:06	
Air Blank		21:06	RFI Detect
Air Blank	0.000	21:07	

---

Data Download Date/Time: 05/24/2019 16:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANTORO,CAMERON,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 04/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: STACKMAN

DATE/TIME OF TEST: 05/21/2019 21:16:20

---

Diagnostics Check	OK	21:09
Air Blank	0.000	21:10
Control Test	0.079	21:10
Air Blank	0.000	21:11
Subject Sample # 1	0.137	21:11
Breath Volume	2.082	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.142	21:14
Breath Volume	1.871	21:14
Air Blank	0.000	21:15
Control Test	0.079	21:15
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

---

Data Download Date/Time: 05/24/2019 16:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RENTZ,JARROD,C  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: PRITCHARD,T  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 04/21/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/22/2019 22:32:43

---

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.079	22:26
Air Blank	0.000	22:26
Subject Sample # 1	0.160	22:27
Breath Volume	2.437	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.156	22:31
Breath Volume	2.050	22:31
Air Blank	0.000	22:31
Control Test	0.079	22:32
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

---

Data Download Date/Time: 05/24/2019 16:20



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALTSTATT, MATTHEW, B  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: BARBER, E  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 19:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: SCHNESKE

DATE/TIME OF TEST: 05/25/2019 21:46:28

---

Diagnostics Check		OK	21:34	
Air Blank		0.000	21:35	
Control Test		0.080	21:35	
Air Blank		0.000	21:35	
Subject Sample	# 1	0.057	21:39	Volume Not Met
Breath Volume		0.796	21:39	
Air Blank		0.000	21:39	
Air Blank		0.000	21:41	
Subject Sample	# 2	0.073	21:41	
Breath Volume		1.867	21:41	
Air Blank		0.000	21:42	
Air Blank		0.000	21:44	
Subject Sample	# 3	0.073	21:44	
Breath Volume		1.644	21:44	
Air Blank		0.000	21:45	
Control Test		0.079	21:45	
Air Blank		0.000	21:46	
Diagnostics Check		OK	21:46	

---

Data Download Date/Time: 06/25/2019 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARCHER, JOSHUA, J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY, RYAN  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: DAMRON

DATE/TIME OF TEST: 05/26/2019 01:54:50

---

Diagnostics Check		OK	01:41	
Air Blank		0.000	01:42	
Control Test		0.081	01:42	
Air Blank		0.000	01:43	
Subject Sample	# 1		01:46	No Sample Provided
Breath Volume		0.000	01:46	
Air Blank		0.000	01:46	
Air Blank		0.000	01:48	
Subject Sample	# 2	0.233	01:49	
Breath Volume		2.472	01:49	
Air Blank		0.000	01:50	
Air Blank		0.000	01:52	
Subject Sample	# 3	0.214	01:53	
Breath Volume		1.570	01:53	
Air Blank		0.000	01:53	
Control Test		0.079	01:54	
Air Blank		0.000	01:54	
Diagnostics Check		OK	01:54	

---

Data Download Date/Time: 06/25/2019 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARCHER,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: PRITCHARD,T  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: COSTANZO

DATE/TIME OF TEST: 05/26/2019 03:13:48

---

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.080	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:09	No Sample Provided
Breath Volume	0.000	03:09	
Air Blank	0.000	03:10	
Air Blank	0.000	03:12	
Subject Sample	# 2	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:12	
Control Test	0.080	03:13	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

---

Data Download Date/Time: 06/25/2019 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARCHER,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: HANEY,M  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 16:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: JUSTUS

DATE/TIME OF TEST: 05/26/2019 18:04:50

---

Diagnostics Check	OK	18:04	
Air Blank		18:04	RFI Detect
Air Blank	0.000	18:04	

---

Data Download Date/Time: 06/25/2019 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARCHER, JOSHUA, J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: HANEY, MICHAEL  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 16:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: JUSTUS

DATE/TIME OF TEST: 05/26/2019 18:17:01

---

Diagnostics Check	OK	18:08	
Air Blank	0.000	18:08	
Control Test	0.080	18:09	
Air Blank	0.000	18:09	
Subject Sample	# 1	18:12	No Sample Provided
Breath Volume	0.000	18:12	
Air Blank	0.000	18:13	
Air Blank	0.000	18:15	
Subject Sample	# 2	18:15	Subject Test Refused
Breath Volume	0.000	18:15	
Air Blank	0.000	18:16	
Control Test	0.080	18:16	
Air Blank	0.000	18:16	
Diagnostics Check	OK	18:17	

---

Data Download Date/Time: 06/25/2019 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARCHER, JOSHUA, J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: BARBER, JAMES  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 20:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: BRINKLEY

DATE/TIME OF TEST: 05/26/2019 21:41:12

---

Diagnostics Check	OK	21:31	
Air Blank	0.000	21:31	
Control Test	0.080	21:32	
Air Blank	0.000	21:32	
Subject Sample	# 1	21:35	No Sample Provided
Breath Volume	0.000	21:35	
Air Blank	0.000	21:36	
Air Blank	0.000	21:38	
Subject Sample	# 2	21:39	Subject Test Refused
Breath Volume	0.000	21:39	
Air Blank	0.000	21:40	
Control Test	0.079	21:40	
Air Blank	0.000	21:41	
Diagnostics Check	OK	21:41	

---

Data Download Date/Time: 06/25/2019 08:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASLINE, T'NEISHA, S  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY, R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: CHAUDURY

DATE/TIME OF TEST: 05/27/2019 01:16:12

---

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.249	01:11
Breath Volume		1.890	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.234	01:14
Breath Volume		1.152	01:14
Air Blank		0.000	01:15
Control Test		0.079	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

---

Data Download Date/Time: 06/25/2019 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARCHER,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: BRIGGS,G  
ARRESTING OFFICER AGENCY: Orange Park PD  
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: TISON

DATE/TIME OF TEST: 05/27/2019 03:59:18

---

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.136	03:54
Breath Volume	3.160	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.131	03:57
Breath Volume	2.054	03:57
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

---

Data Download Date/Time: 06/25/2019 08:12



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANTIAGO,DEPUTY,  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: MILLER,TROOPER  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 13:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 05/27/2019 14:39:46

---

Diagnostics Check	OK	14:31
Air Blank	0.000	14:32
Control Test	0.080	14:32
Air Blank	0.000	14:33
Subject Sample # 1	0.190	14:35
Breath Volume	1.894	14:35
Air Blank	0.000	14:35
Air Blank	0.000	14:37
Subject Sample # 2	0.191	14:38
Breath Volume	1.777	14:38
Air Blank	0.000	14:38
Control Test	0.078	14:39
Air Blank	0.000	14:39
Diagnostics Check	OK	14:39

---

Data Download Date/Time: 06/25/2019 08:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANTIAGO,DEPUTY,  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RAMOS,DEPUTY  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: AYLESBURY

DATE/TIME OF TEST: 05/27/2019 16:10:20

---

Diagnostics Check	OK	16:03
Air Blank	0.000	16:04
Control Test	0.079	16:04
Air Blank	0.000	16:05
Subject Sample # 1	0.048	16:05
Breath Volume	1.781	16:05
Air Blank	0.000	16:06
Air Blank	0.000	16:08
Subject Sample # 2	0.048	16:08
Breath Volume	1.613	16:08
Air Blank	0.000	16:09
Control Test	0.079	16:09
Air Blank	0.000	16:10
Diagnostics Check	OK	16:10

---

Data Download Date/Time: 06/25/2019 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, MATTHEW, P  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: LOVETT, K  
ARRESTING OFFICER AGENCY: Orange Park PD  
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 05/28/2019 03:33:32

---

Diagnostics Check	OK	03:32	
Air Blank		03:33	RFI Detect
Air Blank	0.000	03:33	

---

Data Download Date/Time: 06/25/2019 08:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, MATTHEW, P  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: LOVETT, K  
ARRESTING OFFICER AGENCY: Orange Park PD  
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 05/28/2019 03:41:49

---

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.079	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.185	03:37
Breath Volume	2.316	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:39
Subject Sample # 2	0.183	03:40
Breath Volume	2.632	03:40
Air Blank	0.000	03:40
Control Test	0.080	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

---

Data Download Date/Time: 06/25/2019 08:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARCHER, JOSHUA, J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: BRIGGS, G  
ARRESTING OFFICER AGENCY: Orange Park PD  
LAST AGENCY INSPECTION: 05/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A1  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/31/2019 04:24:09

---

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:12	
Control Test		0.079	04:12	
Air Blank		0.000	04:13	
Subject Sample	# 1	0.200	04:14	
Breath Volume		1.167	04:14	
Air Blank		0.000	04:14	
Air Blank		0.000	04:16	
Subject Sample	# 2	0.216	04:19	Volume Not Met
Breath Volume		0.160	04:19	
Air Blank		0.000	04:20	
Air Blank		0.000	04:22	
Subject Sample	# 3		04:22	Subject Test Refused
Breath Volume		0.000	04:22	
Air Blank		0.000	04:23	
Control Test		0.076	04:23	
Air Blank		0.000	04:24	
Diagnostics Check		OK	04:24	

---

Data Download Date/Time: 06/25/2019 08:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALMER,SAMANTHA,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SORIANO,REYNALDO  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: HERARD

DATE/TIME OF TEST: 05/04/2019 02:57:49

---

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.119	02:53
Breath Volume	1.425	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.119	02:56
Breath Volume	2.210	02:56
Air Blank	0.000	02:56
Control Test	0.079	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

---

Data Download Date/Time: 06/13/2019 15:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FREY,JACOB  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: GARCIA MORALES

DATE/TIME OF TEST: 05/05/2019 01:06:41

---

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.079	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.186	01:02
Breath Volume	3.359	01:02
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.190	01:05
Breath Volume	2.636	01:05
Air Blank	0.000	01:05
Control Test	0.079	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

---

Data Download Date/Time: 06/13/2019 15:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: BONET,LUIS  
ARRESTING OFFICER AGENCY: Delray Beach PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: DICKSON

DATE/TIME OF TEST: 05/05/2019 02:00:56

---

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.079	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.220	01:54
Breath Volume	1.984	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.208	01:59
Breath Volume	2.277	01:59
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

---

Data Download Date/Time: 06/13/2019 15:38



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEA,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FREY,JACOB  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:06

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 05/05/2019 07:41:47

---

Diagnostics Check	OK	07:30
Air Blank	0.000	07:31
Control Test	0.080	07:31
Air Blank	0.000	07:32
Subject Sample # 1	0.157	07:34
Breath Volume	1.136	07:34
Air Blank	0.000	07:35
Air Blank	0.000	07:37
Subject Sample # 2	0.150	07:40
Breath Volume	1.628	07:40
Air Blank	0.000	07:40
Control Test	0.078	07:41
Air Blank	0.000	07:41
Diagnostics Check	OK	07:41

---

Data Download Date/Time: 06/13/2019 15:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHAFFER,NANCY,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WARD,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: SLY

DATE/TIME OF TEST: 05/05/2019 22:00:21

---

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.080	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.181	21:55
Breath Volume	2.039	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.189	21:58
Breath Volume	2.269	21:58
Air Blank	0.000	21:59
Control Test	0.079	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

---

Data Download Date/Time: 06/13/2019 15:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: DALTON,RYAN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: RIIHIAHO

DATE/TIME OF TEST: 05/05/2019 23:54:51

---

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.080	23:43	
Air Blank	0.000	23:44	
Subject Sample	# 1 0.000	23:47	Volume Not Met
Breath Volume	0.109	23:47	
Air Blank	0.000	23:48	
Air Blank	0.000	23:50	
Subject Sample	# 2 0.000	23:50	
Breath Volume	2.117	23:50	
Air Blank	0.000	23:51	
Air Blank	0.000	23:53	
Subject Sample	# 3 0.000	23:53	
Breath Volume	2.296	23:53	
Air Blank	0.000	23:53	
Control Test	0.079	23:54	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

---

Data Download Date/Time: 06/13/2019 15:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: YOCHUM,CRAIG  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: KLEIN

DATE/TIME OF TEST: 05/06/2019 01:40:37

---

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:36	
Control Test	0.079	01:37	
Air Blank	0.000	01:37	
Subject Sample	# 1	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:39	
Control Test	0.079	01:40	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

---

Data Download Date/Time: 06/13/2019 15:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: JENNINGS,DAVON  
ARRESTING OFFICER AGENCY: Boynton Beach PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HATHHORN

DATE/TIME OF TEST: 05/06/2019 02:22:40

---

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.219	02:18
Breath Volume	1.828	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.218	02:21
Breath Volume	1.851	02:21
Air Blank	0.000	02:21
Control Test	0.078	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

---

Data Download Date/Time: 06/13/2019 15:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ROCHA,LUIS  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: JUAN

DATE/TIME OF TEST: 05/06/2019 03:10:58

---

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.204	03:06
Breath Volume	1.828	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.187	03:09
Breath Volume	1.160	03:09
Air Blank	0.000	03:10
Control Test	0.077	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

---

Data Download Date/Time: 06/13/2019 15:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHAFFER,NANCY,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: LYNCH,GREGORY  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 05/07/2019 21:32:49

---

Diagnostics Check	OK	21:25
Air Blank	0.000	21:26
Control Test	0.080	21:26
Air Blank	0.000	21:27
Subject Sample # 1	0.125	21:28
Breath Volume	1.652	21:28
Air Blank	0.000	21:29
Air Blank	0.000	21:30
Subject Sample # 2	0.118	21:31
Breath Volume	1.183	21:31
Air Blank	0.000	21:31
Control Test	0.080	21:32
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

---

Data Download Date/Time: 06/13/2019 15:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FREY,JACOB  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: ALEXANDRE

DATE/TIME OF TEST: 05/07/2019 23:33:44

---

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.244	23:28
Breath Volume	2.375	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.242	23:32
Breath Volume	2.277	23:32
Air Blank	0.000	23:32
Control Test	0.079	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

---

Data Download Date/Time: 06/13/2019 15:38



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: PARKS,MATTHEW  
ARRESTING OFFICER AGENCY: Lantana PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 05/08/2019 21:09:45

---

Diagnostics Check	OK	21:02
Air Blank	0.000	21:03
Control Test	0.079	21:03
Air Blank	0.000	21:04
Subject Sample # 1	0.262	21:05
Breath Volume	2.417	21:05
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.264	21:08
Breath Volume	2.531	21:08
Air Blank	0.000	21:08
Control Test	0.079	21:09
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

---

Data Download Date/Time: 06/13/2019 15:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: LYNCH,GREGORY  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: ADAIR

DATE/TIME OF TEST: 05/09/2019 02:27:59

---

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.227	02:23
Breath Volume	1.480	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:25
Subject Sample # 2	0.211	02:26
Breath Volume	1.414	02:26
Air Blank	0.000	02:27
Control Test	0.078	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

---

Data Download Date/Time: 06/13/2019 15:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: OWEN,MATTHEW  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: LIMANI

DATE/TIME OF TEST: 05/09/2019 19:07:07

---

Diagnostics Check	OK	18:58
Air Blank	0.000	18:59
Control Test	0.080	18:59
Air Blank	0.000	19:00
Subject Sample # 1	0.176	19:02
Breath Volume	1.382	19:02
Air Blank	0.000	19:02
Air Blank	0.000	19:04
Subject Sample # 2	0.179	19:05
Breath Volume	1.929	19:05
Air Blank	0.000	19:06
Control Test	0.079	19:06
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

---

Data Download Date/Time: 06/13/2019 15:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FERNANDEZ,ALAIN  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 09:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 05/10/2019 11:44:51

---

Diagnostics Check		OK	11:37
Air Blank		0.000	11:38
Control Test		0.080	11:38
Air Blank		0.000	11:39
Subject Sample	# 1	0.229	11:39
Breath Volume		1.953	11:39
Air Blank		0.000	11:40
Air Blank		0.000	11:42
Subject Sample	# 2	0.231	11:43
Breath Volume		2.320	11:43
Air Blank		0.000	11:43
Control Test		0.079	11:44
Air Blank		0.000	11:44
Diagnostics Check		OK	11:44

---

Data Download Date/Time: 06/13/2019 15:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SCHAEFER,JOHN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: SKEEN

DATE/TIME OF TEST: 05/10/2019 22:10:45

---

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.080	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.146	22:06
Breath Volume	2.156	22:06
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.137	22:09
Breath Volume	1.687	22:09
Air Blank	0.000	22:09
Control Test	0.080	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

---

Data Download Date/Time: 06/13/2019 15:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ZEITZ,PATRICK  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GAMA PEREZ

DATE/TIME OF TEST: 05/11/2019 00:53:10

---

Diagnostics Check		OK	00:46
Air Blank		0.000	00:47
Control Test		0.080	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.204	00:48
Breath Volume		2.218	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.198	00:51
Breath Volume		1.832	00:51
Air Blank		0.000	00:52
Control Test		0.079	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

---

Data Download Date/Time: 06/13/2019 15:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: HARVEY,ANTHONY  
ARRESTING OFFICER AGENCY: Lantana PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 05/11/2019 02:22:28

---

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.169	02:17
Breath Volume	1.464	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.159	02:20
Breath Volume	1.175	02:20
Air Blank	0.000	02:21
Control Test	0.077	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

---

Data Download Date/Time: 06/13/2019 15:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FREY,JACOB  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: YUKENAVITCH

DATE/TIME OF TEST: 05/11/2019 02:50:24

---

Diagnostics Check		OK	02:43
Air Blank		0.000	02:44
Control Test		0.079	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.157	02:45
Breath Volume		2.417	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.167	02:48
Breath Volume		2.398	02:48
Air Blank		0.000	02:49
Control Test		0.078	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

---

Data Download Date/Time: 06/13/2019 15:39



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MARCH,THOMAS  
ARRESTING OFFICER AGENCY: Palm Beach PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BATISTA

DATE/TIME OF TEST: 05/11/2019 03:34:23

---

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.079	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.196	03:29
Breath Volume	1.789	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.199	03:32
Breath Volume	1.882	03:32
Air Blank	0.000	03:33
Control Test	0.077	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

---

Data Download Date/Time: 06/13/2019 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: REECE,COREY  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:09

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 05/11/2019 15:41:31

---

Diagnostics Check	OK	15:34
Air Blank	0.000	15:35
Control Test	0.079	15:35
Air Blank	0.000	15:36
Subject Sample # 1	0.000	15:36
Breath Volume	4.089	15:36
Air Blank	0.000	15:37
Air Blank	0.000	15:39
Subject Sample # 2	0.000	15:40
Breath Volume	3.046	15:40
Air Blank	0.000	15:40
Control Test	0.079	15:40
Air Blank	0.000	15:41
Diagnostics Check	OK	15:41

---

Data Download Date/Time: 06/13/2019 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WARD,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: ARCHILA SANTOS

DATE/TIME OF TEST: 05/11/2019 23:36:46

---

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.079	23:30
Air Blank	0.000	23:31
Subject Sample # 1	0.217	23:31
Breath Volume	2.234	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.212	23:35
Breath Volume	1.859	23:35
Air Blank	0.000	23:35
Control Test	0.078	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

---

Data Download Date/Time: 06/13/2019 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: STREIGOLD,DAVID  
ARRESTING OFFICER AGENCY: West Palm Beach PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 09:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:34

SUBJECT'S LAST NAME: CHAGNON

DATE/TIME OF TEST: 05/12/2019 11:34:00

---

Diagnostics Check	OK	11:27
Air Blank	0.000	11:28
Control Test	0.079	11:28
Air Blank	0.000	11:29
Subject Sample # 1	0.000	11:29
Breath Volume	1.316	11:29
Air Blank	0.000	11:30
Air Blank	0.000	11:32
Subject Sample # 2	0.000	11:32
Breath Volume	1.250	11:32
Air Blank	0.000	11:33
Control Test	0.079	11:33
Air Blank	0.000	11:33
Diagnostics Check	OK	11:34

---

Data Download Date/Time: 06/13/2019 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: TUANG,NGIN  
ARRESTING OFFICER AGENCY: Lantana PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: GUTIERREZ-DE LA CRUZ

DATE/TIME OF TEST: 05/13/2019 02:00:34

---

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:47	
Control Test		0.080	01:47	
Air Blank		0.000	01:47	
Subject Sample	# 1	0.169	01:51	Volume Not Met
Breath Volume		0.519	01:51	
Air Blank		0.000	01:51	
Air Blank		0.000	01:53	
Subject Sample	# 2	0.192	01:54	
Breath Volume		1.105	01:54	
Air Blank		0.000	01:55	
Air Blank		0.000	01:57	
Subject Sample	# 3	0.193	01:58	
Breath Volume		1.664	01:58	
Air Blank		0.000	01:59	
Control Test		0.076	01:59	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:00	

---

Data Download Date/Time: 06/13/2019 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: IZZARONE,TINA  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:08

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 05/13/2019 07:40:26

---

Diagnostics Check	OK	07:34
Air Blank	0.000	07:34
Control Test	0.080	07:34
Air Blank	0.000	07:35
Subject Sample # 1	0.173	07:35
Breath Volume	1.964	07:35
Air Blank	0.000	07:36
Air Blank	0.000	07:38
Subject Sample # 2	0.173	07:38
Breath Volume	2.414	07:38
Air Blank	0.000	07:39
Control Test	0.079	07:39
Air Blank	0.000	07:40
Diagnostics Check	OK	07:40

---

Data Download Date/Time: 06/13/2019 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: IZZARONE,TINA  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 05/14/2019 16:01:18

---

Diagnostics Check	OK	15:49	
Air Blank	0.000	15:50	
Control Test	0.080	15:50	
Air Blank	0.000	15:50	
Subject Sample	# 1 0.182	15:54	Slope Not Level
Breath Volume	1.125	15:54	
Air Blank	0.000	15:54	
Air Blank	0.000	15:56	
Subject Sample	# 2 0.180	15:59	Volume Not Met
Breath Volume	0.972	15:59	
Air Blank	0.000	16:00	
Control Test	0.077	16:00	
Air Blank	0.000	16:01	
Diagnostics Check	OK	16:01	

---

Data Download Date/Time: 06/13/2019 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: IZZARONE,TINA  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 05/14/2019 16:12:17

---

Diagnostics Check	OK	16:03
Air Blank	0.000	16:03
Control Test	0.077	16:04
Air Blank	0.000	16:04
Subject Sample # 1	0.181	16:05
Breath Volume	1.101	16:05
Air Blank	0.000	16:06
Air Blank	0.000	16:08
Subject Sample # 2	0.183	16:10
Breath Volume	1.320	16:10
Air Blank	0.000	16:11
Control Test	0.078	16:11
Air Blank	0.000	16:12
Diagnostics Check	OK	16:12

---

Data Download Date/Time: 06/13/2019 15:39



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHAFFER,NANCY,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FRIEDMAN,R.  
ARRESTING OFFICER AGENCY: Palm Beach PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/14/2019 21:55:47

---

Diagnostics Check	OK	21:48
Air Blank	0.000	21:48
Control Test	0.078	21:49
Air Blank	0.000	21:49
Subject Sample # 1	0.034	21:51
Breath Volume	2.117	21:51
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.030	21:54
Breath Volume	1.386	21:54
Air Blank	0.000	21:54
Control Test	0.078	21:55
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

---

Data Download Date/Time: 06/13/2019 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SMITH,RYAN  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: O'BRIEN

DATE/TIME OF TEST: 05/14/2019 23:58:38

---

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.079	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.260	23:54
Breath Volume	2.691	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.253	23:57
Breath Volume	1.996	23:57
Air Blank	0.000	23:57
Control Test	0.078	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

---

Data Download Date/Time: 06/13/2019 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALMER,SAMANTHA,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MARTINEZ,JACOB  
ARRESTING OFFICER AGENCY: West Palm Beach PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: LOVERDE

DATE/TIME OF TEST: 05/15/2019 23:28:07

---

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:17	
Control Test	0.080	23:17	
Air Blank	0.000	23:18	
Subject Sample	# 1 0.000	23:21	Volume Not Met
Breath Volume	0.000	23:21	
Air Blank	0.000	23:21	
Air Blank	0.000	23:23	
Subject Sample	# 2 0.000	23:26	Volume Not Met
Breath Volume	0.531	23:26	
Air Blank	0.000	23:27	
Control Test	0.080	23:27	
Air Blank	0.000	23:28	
Diagnostics Check	OK	23:28	

---

Data Download Date/Time: 06/13/2019 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SORIANO,REYNALDO  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: OCHOA MUNGUIA

DATE/TIME OF TEST: 05/16/2019 00:52:22

---

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.165	00:47
Breath Volume	1.484	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.161	00:50
Breath Volume	1.121	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

---

Data Download Date/Time: 06/13/2019 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FREY,JACOB  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 05/16/2019 22:34:18

---

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.080	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.106	22:29
Breath Volume	2.269	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.109	22:32
Breath Volume	2.242	22:32
Air Blank	0.000	22:33
Control Test	0.079	22:33
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

---

Data Download Date/Time: 06/13/2019 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: TORRES,ANDREW  
ARRESTING OFFICER AGENCY: FL Atlantic UPD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: CARRINGTON

DATE/TIME OF TEST: 05/17/2019 20:17:15

---

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.080	20:11
Air Blank	0.000	20:11
Subject Sample # 1	0.380	20:12
Breath Volume	2.238	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:14
Subject Sample # 2	0.391	20:15
Breath Volume	4.121	20:15
Air Blank	0.000	20:16
Control Test	0.078	20:16
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

---

Data Download Date/Time: 06/13/2019 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ETIENNE,PIERRE  
ARRESTING OFFICER AGENCY: West Palm Beach PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: FONTANA

DATE/TIME OF TEST: 05/18/2019 00:26:02

---

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.080	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.192	00:21
Breath Volume	2.664	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.177	00:24
Breath Volume	1.800	00:24
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:26

---

Data Download Date/Time: 06/13/2019 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ERIKSSON,CIRSTEN  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: MENA

DATE/TIME OF TEST: 05/18/2019 01:42:18

---

Diagnostics Check		OK	01:30	
Air Blank		0.000	01:31	
Control Test		0.079	01:31	
Air Blank		0.000	01:31	
Subject Sample	# 1	0.151	01:35	Volume Not Met
Breath Volume		0.167	01:35	
Air Blank		0.000	01:35	
Air Blank		0.000	01:37	
Subject Sample	# 2	0.175	01:40	Volume Not Met
Breath Volume		0.242	01:40	
Air Blank		0.000	01:41	
Control Test		0.078	01:41	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:42	

---

Data Download Date/Time: 06/13/2019 15:39



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ZESUT,BRANDON  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 05/18/2019 02:57:42

---

Diagnostics Check		OK	02:51
Air Blank		0.000	02:51
Control Test		0.080	02:52
Air Blank		0.000	02:52
Subject Sample	# 1	0.206	02:53
Breath Volume		1.660	02:53
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.204	02:56
Breath Volume		1.542	02:56
Air Blank		0.000	02:56
Control Test		0.079	02:57
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

---

Data Download Date/Time: 06/13/2019 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SCHAEFER,JOHN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: COWEN

DATE/TIME OF TEST: 05/18/2019 21:52:17

---

Diagnostics Check	OK	21:42
Air Blank	0.000	21:43
Control Test	0.080	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.158	21:46
Breath Volume	1.753	21:46
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.163	21:50
Breath Volume	1.246	21:50
Air Blank	0.000	21:51
Control Test	0.079	21:51
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

---

Data Download Date/Time: 06/13/2019 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: QUIROS,MARCIEL  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CINTO GUZMAN

DATE/TIME OF TEST: 05/19/2019 02:27:55

---

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.174	02:23
Breath Volume	2.125	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:25
Subject Sample # 2	0.166	02:26
Breath Volume	1.996	02:26
Air Blank	0.000	02:27
Control Test	0.078	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

---

Data Download Date/Time: 06/13/2019 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ESCARAN,LESTER  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:36

SUBJECT'S LAST NAME: HERNANDEZ PEREZ

DATE/TIME OF TEST: 05/19/2019 17:10:17

---

Diagnostics Check	OK	17:03
Air Blank	0.000	17:04
Control Test	0.080	17:04
Air Blank	0.000	17:04
Subject Sample # 1	0.263	17:05
Breath Volume	2.695	17:05
Air Blank	0.000	17:06
Air Blank	0.000	17:08
Subject Sample # 2	0.257	17:08
Breath Volume	3.113	17:08
Air Blank	0.000	17:09
Control Test	0.078	17:09
Air Blank	0.000	17:10
Diagnostics Check	OK	17:10

---

Data Download Date/Time: 06/13/2019 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHAFFER,NANCY,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CRUZ,GARDENIA  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MEJIA VENTURA

DATE/TIME OF TEST: 05/19/2019 19:20:02

---

Diagnostics Check	OK	19:13
Air Blank	0.000	19:13
Control Test	0.079	19:13
Air Blank	0.000	19:14
Subject Sample # 1	0.167	19:15
Breath Volume	1.769	19:15
Air Blank	0.000	19:16
Air Blank	0.000	19:17
Subject Sample # 2	0.171	19:18
Breath Volume	2.035	19:18
Air Blank	0.000	19:19
Control Test	0.079	19:19
Air Blank	0.000	19:19
Diagnostics Check	OK	19:20

---

Data Download Date/Time: 06/13/2019 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: YOCHUM,CRAIG  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: JEFFERSON

DATE/TIME OF TEST: 05/20/2019 00:51:27

---

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.112	00:46
Breath Volume	1.816	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.120	00:49
Breath Volume	2.328	00:49
Air Blank	0.000	00:50
Control Test	0.079	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

---

Data Download Date/Time: 06/13/2019 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ROCHA,LUIS  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: NIGRO

DATE/TIME OF TEST: 05/20/2019 04:17:38

---

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.000	04:13
Breath Volume	2.996	04:13
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.000	04:16
Breath Volume	3.011	04:16
Air Blank	0.000	04:16
Control Test	0.080	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

---

Data Download Date/Time: 06/13/2019 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ROMERO,JUAN  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: JAIMEN

DATE/TIME OF TEST: 05/20/2019 06:29:23

---

Diagnostics Check	OK	06:24	
Air Blank	0.000	06:24	
Control Test	0.080	06:24	
Air Blank	0.000	06:25	
Subject Sample	# 1 0.140	06:25	
Breath Volume	2.007	06:25	
Air Blank	0.000	06:26	
Air Blank	0.000	06:28	
Subject Sample	# 2	06:28	RFI Detect
Breath Volume	0.195	06:28	
Air Blank		06:28	RFI Detect
Air Blank	0.000	06:29	

---

Data Download Date/Time: 06/13/2019 15:39



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ROMERO,JUAN  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: JAIMEN

DATE/TIME OF TEST: 05/20/2019 06:37:04

---

Diagnostics Check	OK	06:30
Air Blank	0.000	06:31
Control Test	0.080	06:31
Air Blank	0.000	06:32
Subject Sample # 1	0.135	06:32
Breath Volume	2.058	06:32
Air Blank	0.000	06:33
Air Blank	0.000	06:35
Subject Sample # 2	0.135	06:35
Breath Volume	1.441	06:35
Air Blank	0.000	06:36
Control Test	0.079	06:36
Air Blank	0.000	06:36
Diagnostics Check	OK	06:37

---

Data Download Date/Time: 06/13/2019 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: LAWRENCE,TIMOTHY  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:42

SUBJECT'S LAST NAME: RAMIREZ Y RAMIREZ

DATE/TIME OF TEST: 05/20/2019 16:14:40

---

Diagnostics Check	OK	16:08
Air Blank	0.000	16:08
Control Test	0.080	16:09
Air Blank	0.000	16:09
Subject Sample # 1	0.224	16:10
Breath Volume	2.250	16:10
Air Blank	0.000	16:10
Air Blank	0.000	16:12
Subject Sample # 2	0.222	16:13
Breath Volume	1.558	16:13
Air Blank	0.000	16:13
Control Test	0.079	16:14
Air Blank	0.000	16:14
Diagnostics Check	OK	16:14

---

Data Download Date/Time: 06/13/2019 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ROMERO,JUAN  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 05/21/2019 04:12:20

---

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.079	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.357	04:07
Breath Volume		2.011	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.373	04:10
Breath Volume		1.281	04:10
Air Blank		0.000	04:11
Control Test		0.077	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

---

Data Download Date/Time: 06/13/2019 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHAFFER,NANCY,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WHITE,E.K.  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 05/21/2019 21:41:19

---

Diagnostics Check	OK	21:33
Air Blank	0.000	21:34
Control Test	0.079	21:34
Air Blank	0.000	21:35
Subject Sample # 1	0.283	21:36
Breath Volume	2.015	21:36
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.296	21:39
Breath Volume	2.261	21:39
Air Blank	0.000	21:40
Control Test	0.079	21:40
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

---

Data Download Date/Time: 06/13/2019 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WHITE,EDWARD  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 05/23/2019 03:20:15

---

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.178	03:14
Breath Volume	1.246	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:16
Subject Sample # 2	0.181	03:18
Breath Volume	1.550	03:18
Air Blank	0.000	03:19
Control Test	0.078	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

---

Data Download Date/Time: 06/13/2019 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: LYNCH,GREGORY  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: DELEON

DATE/TIME OF TEST: 05/23/2019 23:37:02

---

Diagnostics Check		OK	23:30
Air Blank		0.000	23:30
Control Test		0.080	23:30
Air Blank		0.000	23:31
Subject Sample	# 1	0.257	23:32
Breath Volume		3.445	23:32
Air Blank		0.000	23:33
Air Blank		0.000	23:34
Subject Sample	# 2	0.246	23:35
Breath Volume		3.320	23:35
Air Blank		0.000	23:36
Control Test		0.079	23:36
Air Blank		0.000	23:36
Diagnostics Check		OK	23:37

---

Data Download Date/Time: 06/13/2019 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: LOPEZ,NIKOL  
ARRESTING OFFICER AGENCY: West Palm Beach PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: HENDRIX

DATE/TIME OF TEST: 05/24/2019 19:31:03

---

Diagnostics Check	OK	19:26	
Air Blank	0.000	19:27	
Control Test	0.080	19:27	
Air Blank	0.000	19:28	
Subject Sample	# 1	19:29	Subject Test Refused
Breath Volume	0.000	19:29	
Air Blank	0.000	19:30	
Control Test	0.080	19:30	
Air Blank	0.000	19:30	
Diagnostics Check	OK	19:31	

---

Data Download Date/Time: 06/13/2019 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALMER,SAMANTHA,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CARVER,CAMERON  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BUDD

DATE/TIME OF TEST: 05/25/2019 01:18:13

---

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.081	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.138	01:13
Breath Volume	1.964	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.140	01:16
Breath Volume	2.339	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

---

Data Download Date/Time: 06/13/2019 15:39



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALMER,SAMANTHA,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ZEITZ,PATRICK  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 05/25/2019 01:52:28

---

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.127	01:48
Breath Volume	2.355	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.133	01:50
Breath Volume	2.371	01:50
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

---

Data Download Date/Time: 06/13/2019 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALMER,SAMANTHA,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: BUNDRICK,MATHEW  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/25/2019 03:03:23

---

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.094	02:58
Breath Volume	1.527	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.092	03:01
Breath Volume	1.402	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

---

Data Download Date/Time: 06/13/2019 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALMER,SAMANTHA,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ESCARAN,LESTER  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GARZON RODRIGUEZ

DATE/TIME OF TEST: 05/25/2019 04:12:25

---

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.132	04:07
Breath Volume	1.843	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample # 2	0.127	04:10
Breath Volume	1.433	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

---

Data Download Date/Time: 06/13/2019 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHAFFER,NANCY,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: AGAUGLU,Y  
ARRESTING OFFICER AGENCY: Palm Beach PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: MONDRO

DATE/TIME OF TEST: 05/25/2019 19:32:01

---

Diagnostics Check		OK	19:20	
Air Blank		0.000	19:20	
Control Test		0.080	19:21	
Air Blank		0.000	19:21	
Subject Sample	# 1	0.181	19:24	Volume Not Met
Breath Volume		0.894	19:24	
Air Blank		0.000	19:25	
Air Blank		0.000	19:27	
Subject Sample	# 2	0.144	19:30	Volume Not Met
Breath Volume		0.507	19:30	
Air Blank		0.000	19:31	
Control Test		0.078	19:31	
Air Blank		0.000	19:31	
Diagnostics Check		OK	19:32	

---

Data Download Date/Time: 06/13/2019 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHAFFER,NANCY,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: AGAAGLU,Y  
ARRESTING OFFICER AGENCY: Palm Beach PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: MONDRO

DATE/TIME OF TEST: 05/25/2019 19:46:35

---

Diagnostics Check		OK	19:34	
Air Blank		0.000	19:35	
Control Test		0.079	19:35	
Air Blank		0.000	19:36	
Subject Sample	# 1	0.155	19:39	Volume Not Met
Breath Volume		0.421	19:39	
Air Blank		0.000	19:40	
Air Blank		0.000	19:41	
Subject Sample	# 2	0.139	19:44	Volume Not Met
Breath Volume		0.417	19:44	
Air Blank		0.000	19:45	
Control Test		0.078	19:45	
Air Blank		0.000	19:46	
Diagnostics Check		OK	19:46	

---

Data Download Date/Time: 06/13/2019 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHAFFER,NANCY,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: AGAUGLU,Y  
ARRESTING OFFICER AGENCY: Palm Beach PD  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: MONDRO

DATE/TIME OF TEST: 05/25/2019 19:56:14

---

Diagnostics Check	OK	19:48
Air Blank	0.000	19:49
Control Test	0.078	19:49
Air Blank	0.000	19:50
Subject Sample # 1	0.184	19:51
Breath Volume	1.585	19:51
Air Blank	0.000	19:52
Air Blank	0.000	19:53
Subject Sample # 2	0.189	19:54
Breath Volume	1.320	19:54
Air Blank	0.000	19:55
Control Test	0.077	19:55
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

---

Data Download Date/Time: 06/13/2019 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHAFFER,NANCY,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CISSON,J  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 05/25/2019 21:10:31

---

Diagnostics Check	OK	21:03
Air Blank	0.000	21:03
Control Test	0.079	21:04
Air Blank	0.000	21:04
Subject Sample # 1	0.048	21:06
Breath Volume	1.761	21:06
Air Blank	0.000	21:06
Air Blank	0.000	21:08
Subject Sample # 2	0.048	21:08
Breath Volume	2.503	21:08
Air Blank	0.000	21:09
Control Test	0.079	21:09
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

---

Data Download Date/Time: 06/13/2019 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHAFFER,NANCY,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: BOYD,ADAM  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: MCROBERTS

DATE/TIME OF TEST: 05/25/2019 22:12:20

---

Diagnostics Check	OK	22:03
Air Blank	0.000	22:04
Control Test	0.080	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.141	22:07
Breath Volume	2.003	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:09
Subject Sample # 2	0.139	22:10
Breath Volume	2.398	22:10
Air Blank	0.000	22:11
Control Test	0.079	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

---

Data Download Date/Time: 06/13/2019 15:40



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WHITE,KEITH  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: MENDEZ PASCUAL

DATE/TIME OF TEST: 05/26/2019 02:48:55

---

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.211	02:44
Breath Volume	1.953	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:46
Subject Sample # 2	0.207	02:47
Breath Volume	1.507	02:47
Air Blank	0.000	02:47
Control Test	0.080	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

---

Data Download Date/Time: 06/13/2019 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: TRUDEAU,RYAN  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: GRIMARD

DATE/TIME OF TEST: 05/26/2019 23:23:58

---

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.177	23:19
Breath Volume	1.609	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.176	23:22
Breath Volume	1.921	23:22
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

---

Data Download Date/Time: 06/13/2019 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SCHNEIDER,DAVID  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 05/27/2019 00:10:58

---

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.080	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.140	00:06
Breath Volume	3.402	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:08
Subject Sample # 2	0.141	00:09
Breath Volume	3.476	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

---

Data Download Date/Time: 06/13/2019 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SMITH,MARK  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: DE ARMAS RIVERO

DATE/TIME OF TEST: 05/27/2019 00:55:29

---

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.080	00:50
Air Blank	0.000	00:50
Subject Sample # 1	0.141	00:51
Breath Volume	1.421	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.138	00:53
Breath Volume	1.433	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

---

Data Download Date/Time: 06/13/2019 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: POINTU,PATRICK  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: KAMEN

DATE/TIME OF TEST: 05/27/2019 01:40:21

---

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.079	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.000	01:36
Breath Volume	3.515	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.000	01:38
Breath Volume	3.257	01:38
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

---

Data Download Date/Time: 06/13/2019 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: D'ANDREA,ADAM  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 07:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:43

SUBJECT'S LAST NAME: MARTINEZ-VASQUEZ

DATE/TIME OF TEST: 05/27/2019 08:15:42

---

Diagnostics Check		OK	08:08
Air Blank		0.000	08:08
Control Test		0.081	08:08
Air Blank		0.000	08:09
Subject Sample	# 1	0.248	08:10
Breath Volume		2.003	08:10
Air Blank		0.000	08:11
Air Blank		0.000	08:13
Subject Sample	# 2	0.247	08:14
Breath Volume		2.039	08:14
Air Blank		0.000	08:14
Control Test		0.078	08:15
Air Blank		0.000	08:15
Diagnostics Check		OK	08:15

---

Data Download Date/Time: 06/13/2019 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: D'ANDREA,ADAM  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 13:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:38

SUBJECT'S LAST NAME: ESPINOZA-SALAZA

DATE/TIME OF TEST: 05/27/2019 14:11:23

---

Diagnostics Check	OK	14:08	
Air Blank	0.000	14:08	
Control Test	0.080	14:08	
Air Blank	0.000	14:09	
Subject Sample	# 1	14:10	Subject Test Refused
Breath Volume	0.000	14:10	
Air Blank	0.000	14:10	
Control Test	0.080	14:10	
Air Blank	0.000	14:11	
Diagnostics Check	OK	14:11	

---

Data Download Date/Time: 06/13/2019 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ZEITZ,PATRICK  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: RODRIGUEZ HERNANADEZ

DATE/TIME OF TEST: 05/27/2019 23:41:48

---

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.080	23:35
Air Blank	0.000	23:36
Subject Sample # 1	0.185	23:36
Breath Volume	1.417	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.172	23:40
Breath Volume	1.753	23:40
Air Blank	0.000	23:40
Control Test	0.080	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

---

Data Download Date/Time: 06/13/2019 15:40



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: POINTU,PATRICK  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: ABUDEYE GUERRA

DATE/TIME OF TEST: 05/28/2019 03:29:59

---

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.169	03:24
Breath Volume	1.617	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.166	03:28
Breath Volume	1.902	03:28
Air Blank	0.000	03:29
Control Test	0.079	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

---

Data Download Date/Time: 06/13/2019 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FERNANDEZ,ALAIN  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MORRELL

DATE/TIME OF TEST: 05/28/2019 04:00:39

---

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.079	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.150	03:56
Breath Volume	1.871	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.149	03:59
Breath Volume	1.160	03:59
Air Blank	0.000	03:59
Control Test	0.079	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

---

Data Download Date/Time: 06/13/2019 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WEHRLE,ERIC  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:04

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 05/29/2019 08:45:51

---

Diagnostics Check	OK	08:39
Air Blank	0.000	08:39
Control Test	0.081	08:39
Air Blank	0.000	08:40
Subject Sample # 1	0.202	08:40
Breath Volume	1.828	08:40
Air Blank	0.000	08:41
Air Blank	0.000	08:43
Subject Sample # 2	0.197	08:44
Breath Volume	2.281	08:44
Air Blank	0.000	08:44
Control Test	0.079	08:45
Air Blank	0.000	08:45
Diagnostics Check	OK	08:45

---

Data Download Date/Time: 06/13/2019 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CARVER,CAMERON  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: FILOON

DATE/TIME OF TEST: 05/30/2019 22:53:18

---

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.081	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.028	22:48
Breath Volume	1.519	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.028	22:51
Breath Volume	1.515	22:51
Air Blank	0.000	22:52
Control Test	0.080	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

---

Data Download Date/Time: 06/13/2019 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: LYNCH,GREGORY  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 05/03/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: REITER

DATE/TIME OF TEST: 05/31/2019 22:34:02

---

Diagnostics Check	OK	22:27
Air Blank	0.000	22:27
Control Test	0.080	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.000	22:29
Breath Volume	5.089	22:29
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.000	22:32
Breath Volume	4.500	22:32
Air Blank	0.000	22:33
Control Test	0.080	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:34

---

Data Download Date/Time: 06/13/2019 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,CHRIS,J  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MILLER,CHRIS  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: TAPIA

DATE/TIME OF TEST: 05/04/2019 23:54:27

---

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Subject Sample # 1	0.114	23:49
Breath Volume	3.460	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.109	23:52
Breath Volume	3.054	23:52
Air Blank	0.000	23:53
Control Test	0.077	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

---

Data Download Date/Time: 05/27/2019 20:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLEY,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: JUBERT

DATE/TIME OF TEST: 05/05/2019 00:52:01

---

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:48	
Control Test	0.080	00:48	
Air Blank	0.000	00:49	
Subject Sample	# 1	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:51	
Control Test	0.080	00:51	
Air Blank	0.000	00:51	
Diagnostics Check	OK	00:52	

---

Data Download Date/Time: 05/27/2019 20:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLEY,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: JUBERT

DATE/TIME OF TEST: 05/05/2019 01:09:55

---

Diagnostics Check		OK	00:57	
Air Blank		0.000	00:57	
Control Test		0.080	00:57	
Air Blank		0.000	00:58	
Subject Sample	# 1	0.093	01:01	Volume Not Met
Breath Volume		0.398	01:01	
Air Blank		0.000	01:02	
Air Blank		0.000	01:04	
Subject Sample	# 2	0.108	01:05	
Breath Volume		1.390	01:05	
Air Blank		0.000	01:06	
Air Blank		0.000	01:07	
Subject Sample	# 3	0.115	01:08	
Breath Volume		1.472	01:08	
Air Blank		0.000	01:08	
Control Test		0.079	01:09	
Air Blank		0.000	01:09	
Diagnostics Check		OK	01:09	

---

Data Download Date/Time: 05/27/2019 20:58



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLEY,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: ALSHAMRANI

DATE/TIME OF TEST: 05/05/2019 04:04:56

---

Diagnostics Check		OK	03:58
Air Blank		0.000	03:59
Control Test		0.079	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.123	04:00
Breath Volume		2.210	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:02
Subject Sample	# 2	0.118	04:03
Breath Volume		1.992	04:03
Air Blank		0.000	04:04
Control Test		0.078	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

---

Data Download Date/Time: 05/27/2019 20:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLEY,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 05/10/2019 02:51:57

---

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.165	02:45
Breath Volume	2.070	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.166	02:50
Breath Volume	1.808	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

---

Data Download Date/Time: 05/27/2019 20:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLEY,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: GRIER

DATE/TIME OF TEST: 05/11/2019 01:45:15

---

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.075	01:40
Breath Volume	3.144	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.076	01:43
Breath Volume	2.496	01:43
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

---

Data Download Date/Time: 05/27/2019 20:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLEY,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 05/12/2019 01:56:30

---

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.079	01:50	
Air Blank	0.000	01:50	
Subject Sample # 1	0.158	01:51	
Breath Volume	3.355	01:51	
Air Blank	0.000	01:52	
Air Blank	0.000	01:53	
Subject Sample # 2	0.158	01:54	
Breath Volume	2.050	01:54	
Air Blank	0.000	01:55	
Control Test	0.081	01:55	
Air Blank	0.000	01:55	
Diagnostics Check	Fail	01:56	Analytical Stability Test
Air Blank	0.000	01:56	

---

Data Download Date/Time: 05/27/2019 20:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLEY,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 05/12/2019 02:05:27

---

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.079	01:59
Air Blank		0.000	01:59
Subject Sample	# 1	0.146	02:00
Breath Volume		1.515	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.163	02:03
Breath Volume		3.574	02:03
Air Blank		0.000	02:04
Control Test		0.079	02:04
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

---

Data Download Date/Time: 05/27/2019 20:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLEY,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: NOLAN-WILLIAMS

DATE/TIME OF TEST: 05/24/2019 01:02:32

---

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.081	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.275	00:57
Breath Volume		1.500	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.279	01:00
Breath Volume		1.355	01:00
Air Blank		0.000	01:01
Control Test		0.081	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

---

Data Download Date/Time: 05/27/2019 20:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLEY,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: YANNAM

DATE/TIME OF TEST: 05/24/2019 23:40:39

---

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.080	23:33
Air Blank	0.000	23:34
Subject Sample # 1	0.133	23:36
Breath Volume	1.984	23:36
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.132	23:39
Breath Volume	1.187	23:39
Air Blank	0.000	23:39
Control Test	0.079	23:40
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

---

Data Download Date/Time: 05/27/2019 20:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLEY,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: PINEDA LAGUNAS

DATE/TIME OF TEST: 05/25/2019 02:52:14

---

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.175	02:47
Breath Volume	1.128	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:49
Subject Sample # 2	0.180	02:50
Breath Volume	3.945	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

---

Data Download Date/Time: 05/27/2019 20:58



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLEY,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/24/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: MYERS 3RD

DATE/TIME OF TEST: 05/25/2019 22:53:31

---

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.080	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.207	22:49
Breath Volume	3.300	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.202	22:51
Breath Volume	1.964	22:51
Air Blank	0.000	22:52
Control Test	0.080	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

---

Data Download Date/Time: 05/27/2019 20:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GODENZI,CAMERON,M  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: BROWN,TREVOR  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 04/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: STOCKMAN

DATE/TIME OF TEST: 05/11/2019 19:55:44

---

Diagnostics Check	OK	19:46	
Air Blank	0.000	19:47	
Control Test	0.079	19:47	
Air Blank	0.000	19:47	
Subject Sample # 1	0.175	19:49	
Breath Volume	1.246	19:49	
Air Blank	0.000	19:50	
Air Blank	0.000	19:51	
Subject Sample # 2	0.166	19:54	
Breath Volume	1.156	19:54	
Air Blank		19:55	Purge Fail
Air Blank	0.000	19:55	

---

Data Download Date/Time: 05/20/2019 11:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GODENZI,CAMERON,M  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: BROWN,TREVOR  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 04/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: STOCKMAN

DATE/TIME OF TEST: 05/11/2019 20:17:58

---

Diagnostics Check	OK	20:03	
Air Blank	0.000	20:03	
Control Test	0.076	20:04	
Air Blank	0.000	20:04	
Subject Sample # 1	0.166	20:07	
Breath Volume	1.101	20:07	
Air Blank	0.000	20:08	
Air Blank	0.000	20:09	
Subject Sample # 2	0.151	20:12	Slope Not Level
Breath Volume	1.367	20:12	
Air Blank	0.000	20:13	
Air Blank	0.000	20:15	
Subject Sample # 3	0.162	20:16	
Breath Volume	1.203	20:16	
Air Blank	0.000	20:17	
Control Test	0.074	20:17	Control Outside Tolerance
Air Blank	0.000	20:17	

---

Data Download Date/Time: 05/20/2019 11:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GODENZI,CAMERON,M  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: BROWN,TREVOR  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 04/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: STOCKMAN

DATE/TIME OF TEST: 05/11/2019 21:09:14

---

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:01	
Control Test	0.079	21:01	
Air Blank	0.000	21:02	
Subject Sample	# 1 0.144	21:02	
Breath Volume	1.359	21:02	
Air Blank	0.000	21:03	
Air Blank	0.000	21:05	
Subject Sample	# 2	21:07	Slope Not Met
Breath Volume	0.472	21:07	
Air Blank	0.000	21:08	
Control Test	0.076	21:08	
Air Blank	0.000	21:09	
Diagnostics Check	OK	21:09	

---

Data Download Date/Time: 05/20/2019 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GODENZI,CAMERON,M  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: BROWN,TREVOR  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 04/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 644476  
DRY GAS EXPIRY DATE: 12/01/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: STOCKMAN

DATE/TIME OF TEST: 05/11/2019 21:44:55

---

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:38	
Control Test	0.078	21:39	
Air Blank	0.000	21:39	
Subject Sample # 1	0.121	21:40	
Breath Volume	1.695	21:40	
Air Blank	0.000	21:40	
Air Blank	0.000	21:42	
Subject Sample # 2	0.120	21:43	
Breath Volume	1.464	21:43	
Air Blank	0.000	21:44	
Control Test	0.074	21:44	Control Outside Tolerance
Air Blank	0.000	21:44	

---

Data Download Date/Time: 05/20/2019 11:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: JAMES, ROYCE  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 05/20/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1048227  
DRY GAS EXPIRY DATE: 11/15/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: KRUG

DATE/TIME OF TEST: 05/21/2019 13:32:35

---

Diagnostics Check	OK	13:25
Air Blank	0.000	13:25
Control Test	0.079	13:26
Air Blank	0.000	13:26
Subject Sample # 1	0.066	13:27
Breath Volume	1.656	13:27
Air Blank	0.000	13:28
Air Blank	0.000	13:30
Subject Sample # 2	0.064	13:31
Breath Volume	1.617	13:31
Air Blank	0.000	13:31
Control Test	0.079	13:31
Air Blank	0.000	13:32
Diagnostics Check	OK	13:32

---

Data Download Date/Time: 06/17/2019 12:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KLENOFSKY,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: HARRIGAN

DATE/TIME OF TEST: 05/03/2019 04:54:34

---

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.080	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.145	04:49
Breath Volume	4.011	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:51
Subject Sample # 2	0.153	04:52
Breath Volume	3.378	04:52
Air Blank	0.000	04:53
Control Test	0.079	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

---

Data Download Date/Time: 05/06/2019 14:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WEG, L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: SRIKANTH

DATE/TIME OF TEST: 05/03/2019 05:31:06

---

Diagnostics Check	OK	05:24
Air Blank	0.000	05:25
Control Test	0.080	05:25
Air Blank	0.000	05:26
Subject Sample # 1	0.016	05:26
Breath Volume	3.531	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.016	05:29
Breath Volume	3.578	05:29
Air Blank	0.000	05:30
Control Test	0.080	05:30
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

---

Data Download Date/Time: 05/06/2019 14:09



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SCHAEFER,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 05/03/2019 22:46:19

---

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.081	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.169	22:41
Breath Volume	3.695	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.172	22:44
Breath Volume	3.605	22:44
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

---

Data Download Date/Time: 05/06/2019 14:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: COSTA, A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 05/04/2019 03:28:19

---

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.082	03:23
Breath Volume	1.449	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.087	03:26
Breath Volume	2.082	03:26
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

---

Data Download Date/Time: 05/06/2019 14:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DOBSON,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 05/04/2019 23:03:49

---

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.080	22:58
Air Blank	0.000	22:59
Subject Sample # 1	0.058	22:59
Breath Volume	2.687	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:01
Subject Sample # 2	0.060	23:02
Breath Volume	2.773	23:02
Air Blank	0.000	23:02
Control Test	0.081	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

---

Data Download Date/Time: 05/06/2019 14:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SCHAEFER,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/05/2019 04:18:41

---

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.081	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.170	04:14
Breath Volume	3.695	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.170	04:17
Breath Volume	2.585	04:17
Air Blank	0.000	04:17
Control Test	0.079	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

---

Data Download Date/Time: 05/06/2019 14:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VASALLO,Y  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: GATICA RAMIREZ

DATE/TIME OF TEST: 05/05/2019 04:52:39

---

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.081	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.115	04:48
Breath Volume	2.652	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.112	04:51
Breath Volume	2.925	04:51
Air Blank	0.000	04:51
Control Test	0.079	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

---

Data Download Date/Time: 05/06/2019 14:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ANDERSON,C  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 05/05/2019 05:24:47

---

Diagnostics Check		OK	05:18
Air Blank		0.000	05:18
Control Test		0.080	05:18
Air Blank		0.000	05:19
Subject Sample	# 1	0.200	05:20
Breath Volume		1.796	05:20
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.189	05:23
Breath Volume		2.058	05:23
Air Blank		0.000	05:23
Control Test		0.078	05:24
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

---

Data Download Date/Time: 05/06/2019 14:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DRISCOLL,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: ALASCAGUA CASTANEDA

DATE/TIME OF TEST: 05/06/2019 03:29:24

---

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.081	03:24	
Air Blank	0.000	03:24	
Subject Sample	# 1 0.214	03:27	Slope Not Level
Breath Volume	1.847	03:27	
Air Blank		03:28	Purge Fail
Air Blank	0.000	03:29	

---

Data Download Date/Time: 05/06/2019 14:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DRISCOLL,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: ALASCAGUA CASTANEDA

DATE/TIME OF TEST: 05/06/2019 03:41:53

---

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.077	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.219	03:35
Breath Volume	2.449	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.225	03:40
Breath Volume	1.851	03:40
Air Blank	0.000	03:40
Control Test	0.078	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

---

Data Download Date/Time: 05/06/2019 14:10



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 05/08/2019 05:15:47

---

Diagnostics Check		OK	05:08
Air Blank		0.000	05:08
Control Test		0.080	05:09
Air Blank		0.000	05:09
Subject Sample	# 1	0.135	05:10
Breath Volume		2.437	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.137	05:14
Breath Volume		2.929	05:14
Air Blank		0.000	05:14
Control Test		0.079	05:15
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

---

Data Download Date/Time: 06/11/2019 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: GALLO

DATE/TIME OF TEST: 05/08/2019 21:46:32

---

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:40
Subject Sample # 1	0.175	21:42
Breath Volume	2.332	21:42
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.173	21:44
Breath Volume	1.351	21:44
Air Blank	0.000	21:45
Control Test	0.079	21:45
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

---

Data Download Date/Time: 06/11/2019 08:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: COSTA,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/09/2019 03:22:50

---

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.081	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.000	03:17
Breath Volume	3.203	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.000	03:21
Breath Volume	2.750	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

---

Data Download Date/Time: 06/11/2019 08:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: HILLS JR

DATE/TIME OF TEST: 05/10/2019 20:41:53

---

Diagnostics Check	OK	20:38	
Air Blank	0.000	20:39	
Control Test	0.080	20:39	
Air Blank	0.000	20:40	
Subject Sample	# 1	20:40	Subject Test Refused
Breath Volume	0.000	20:40	
Air Blank	0.000	20:40	
Control Test	0.081	20:41	
Air Blank	0.000	20:41	
Diagnostics Check	OK	20:41	

---

Data Download Date/Time: 06/11/2019 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: HORNSBY

DATE/TIME OF TEST: 05/11/2019 04:39:39

---

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.081	04:33
Air Blank	0.000	04:33
Subject Sample # 1	0.155	04:35
Breath Volume	1.246	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:37
Subject Sample # 2	0.152	04:38
Breath Volume	1.125	04:38
Air Blank	0.000	04:38
Control Test	0.078	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

---

Data Download Date/Time: 06/11/2019 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: BENITEZ DOMINGUEZ

DATE/TIME OF TEST: 05/11/2019 22:16:58

---

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.080	22:11
Air Blank	0.000	22:11
Subject Sample # 1	0.011	22:12
Breath Volume	3.757	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.008	22:15
Breath Volume	2.101	22:15
Air Blank	0.000	22:16
Control Test	0.079	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

---

Data Download Date/Time: 06/11/2019 08:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CAMPBELL,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 05/12/2019 02:24:18

---

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:16	
Control Test	0.080	02:16	
Air Blank	0.000	02:16	
Subject Sample	# 1	02:20	No Sample Provided
Breath Volume	0.000	02:20	
Air Blank	0.000	02:20	
Air Blank	0.000	02:22	
Subject Sample	# 2	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:23	
Control Test	0.081	02:23	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

---

Data Download Date/Time: 06/11/2019 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CAMPBELL,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 05/12/2019 02:33:08

---

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:29	
Control Test	0.080	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:32	
Control Test	0.080	02:32	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

---

Data Download Date/Time: 06/11/2019 08:12



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DEJESUS,V  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/12/2019 20:28:23

---

Diagnostics Check	OK	20:22
Air Blank	0.000	20:22
Control Test	0.081	20:22
Air Blank	0.000	20:23
Subject Sample # 1	0.059	20:23
Breath Volume	2.984	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.054	20:26
Breath Volume	3.480	20:26
Air Blank	0.000	20:27
Control Test	0.080	20:27
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

---

Data Download Date/Time: 06/11/2019 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: COSTA,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: BALLENTINE

DATE/TIME OF TEST: 05/13/2019 22:19:50

---

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.082	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.000	22:15
Breath Volume	1.847	22:15
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.000	22:18
Breath Volume	2.410	22:18
Air Blank	0.000	22:18
Control Test	0.081	22:19
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

---

Data Download Date/Time: 06/11/2019 08:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KLENOFSKY,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: PAULEY

DATE/TIME OF TEST: 05/16/2019 04:23:07

---

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.081	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.087	04:18
Breath Volume		2.675	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.073	04:21
Breath Volume		1.207	04:21
Air Blank		0.000	04:22
Control Test		0.080	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

---

Data Download Date/Time: 06/11/2019 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DRISCOLL,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/18/2019 03:28:02

---

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.081	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.149	03:23
Breath Volume	1.769	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.160	03:26
Breath Volume	1.175	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:28

---

Data Download Date/Time: 06/11/2019 08:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PALFY,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: NICHOLS VACHUNEK

DATE/TIME OF TEST: 05/19/2019 01:19:58

---

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.080	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.028	01:15
Breath Volume	1.179	01:15
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.027	01:18
Breath Volume	1.378	01:18
Air Blank	0.000	01:19
Control Test	0.080	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

---

Data Download Date/Time: 06/11/2019 08:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: COSTA,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: NEDLEY

DATE/TIME OF TEST: 05/19/2019 05:48:47

---

Diagnostics Check		OK	05:38
Air Blank		0.000	05:38
Control Test		0.081	05:38
Air Blank		0.000	05:39
Subject Sample	# 1	0.197	05:42
Breath Volume		1.496	05:42
Air Blank		0.000	05:43
Air Blank		0.000	05:44
Subject Sample	# 2	0.190	05:47
Breath Volume		1.250	05:47
Air Blank		0.000	05:47
Control Test		0.079	05:48
Air Blank		0.000	05:48
Diagnostics Check		OK	05:48

---

Data Download Date/Time: 06/11/2019 08:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN TREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: RADCLIFFE

DATE/TIME OF TEST: 05/19/2019 23:09:43

---

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.080	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.186	23:04
Breath Volume	1.632	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:06
Subject Sample # 2	0.181	23:08
Breath Volume	1.433	23:08
Air Blank	0.000	23:08
Control Test	0.079	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

---

Data Download Date/Time: 06/11/2019 08:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FISHER,T  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 05/21/2019 02:59:59

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.081	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.130	02:54
Breath Volume	2.371	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.132	02:58
Breath Volume	2.382	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

---

Data Download Date/Time: 06/11/2019 08:13



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BRABAND, M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: FRANSEN

DATE/TIME OF TEST: 05/21/2019 23:32:18

---

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.081	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.119	23:27
Breath Volume	2.113	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.119	23:30
Breath Volume	2.019	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

---

Data Download Date/Time: 06/11/2019 08:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RODRIGUEZ,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: CRAVOTTA

DATE/TIME OF TEST: 05/23/2019 02:52:16

---

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.081	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.123	02:47
Breath Volume	3.253	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.122	02:50
Breath Volume	3.414	02:50
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

---

Data Download Date/Time: 06/11/2019 08:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VISSER,A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: HINKIN

DATE/TIME OF TEST: 05/25/2019 04:07:43

---

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.081	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.148	04:03
Breath Volume	3.398	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.149	04:06
Breath Volume	3.492	04:06
Air Blank	0.000	04:06
Control Test	0.081	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

---

Data Download Date/Time: 06/11/2019 08:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WOLSTENHOLM,A  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: KOCHANOWSKI-KNOWLES

DATE/TIME OF TEST: 05/25/2019 04:56:51

---

Diagnostics Check	OK	04:49
Air Blank	0.000	04:50
Control Test	0.081	04:50
Air Blank	0.000	04:51
Subject Sample # 1	0.225	04:52
Breath Volume	2.617	04:52
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.211	04:55
Breath Volume	2.792	04:55
Air Blank	0.000	04:55
Control Test	0.079	04:56
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

---

Data Download Date/Time: 06/11/2019 08:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VISSER,A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 05/25/2019 20:03:20

---

Diagnostics Check	OK	19:56
Air Blank	0.000	19:56
Control Test	0.082	19:57
Air Blank	0.000	19:57
Subject Sample # 1	0.289	19:58
Breath Volume	3.402	19:58
Air Blank	0.000	19:59
Air Blank	0.000	20:00
Subject Sample # 2	0.275	20:01
Breath Volume	2.460	20:01
Air Blank	0.000	20:02
Control Test	0.080	20:02
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

---

Data Download Date/Time: 06/11/2019 08:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KLENOFSKY,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31517080A1  
DRY GAS EXPIRY DATE: 01/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 05/25/2019 22:49:08

---

Diagnostics Check	OK	22:42
Air Blank	0.000	22:42
Control Test	0.081	22:42
Air Blank	0.000	22:43
Subject Sample # 1	0.000	22:44
Breath Volume	2.714	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.000	22:47
Breath Volume	1.492	22:47
Air Blank	0.000	22:48
Control Test	0.081	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

---

Data Download Date/Time: 06/11/2019 08:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RIVERA,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: MOONEY

DATE/TIME OF TEST: 05/26/2019 04:45:02

---

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.081	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.151	04:40
Breath Volume	1.683	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:42
Subject Sample # 2	0.156	04:43
Breath Volume	2.226	04:43
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:44
Diagnostics Check	OK	04:45

---

Data Download Date/Time: 06/11/2019 08:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/27/2019 01:51:14

---

Diagnostics Check		OK	01:41
Air Blank		0.000	01:42
Control Test		0.080	01:42
Air Blank		0.000	01:42
Subject Sample	# 1	0.164	01:44
Breath Volume		2.097	01:44
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.163	01:49
Breath Volume		2.265	01:49
Air Blank		0.000	01:50
Control Test		0.081	01:50
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

---

Data Download Date/Time: 06/11/2019 08:13



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BORRENBERGEN

DATE/TIME OF TEST: 05/28/2019 03:24:28

---

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:11	
Control Test		0.081	03:11	
Air Blank		0.000	03:12	
Subject Sample	# 1	0.143	03:12	
Breath Volume		1.625	03:12	
Air Blank		0.000	03:13	
Air Blank		0.000	03:15	
Subject Sample	# 2	0.125	03:18	Volume Not Met
Breath Volume		0.195	03:18	
Air Blank		0.000	03:19	
Air Blank		0.000	03:21	
Subject Sample	# 3	0.131	03:22	
Breath Volume		1.238	03:22	
Air Blank		0.000	03:23	
Control Test		0.077	03:23	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:24	

---

Data Download Date/Time: 06/11/2019 08:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 05/30/2019 00:41:42

---

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.131	00:37
Breath Volume	2.679	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.131	00:40
Breath Volume	2.613	00:40
Air Blank	0.000	00:40
Control Test	0.079	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

---

Data Download Date/Time: 06/11/2019 08:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: THORNE, S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: PAREDES

DATE/TIME OF TEST: 05/31/2019 04:47:06

---

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.081	04:40
Air Blank	0.000	04:41
Subject Sample # 1	0.177	04:42
Breath Volume	2.289	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.166	04:45
Breath Volume	1.320	04:45
Air Blank	0.000	04:46
Control Test	0.078	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

---

Data Download Date/Time: 06/11/2019 08:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SULAIMAN,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/02/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: LAZARO

DATE/TIME OF TEST: 05/31/2019 19:59:34

---

Diagnostics Check	OK	19:51
Air Blank	0.000	19:51
Control Test	0.081	19:52
Air Blank	0.000	19:52
Subject Sample # 1	0.185	19:54
Breath Volume	1.875	19:54
Air Blank	0.000	19:55
Air Blank	0.000	19:56
Subject Sample # 2	0.176	19:57
Breath Volume	1.492	19:57
Air Blank	0.000	19:58
Control Test	0.077	19:58
Air Blank	0.000	19:59
Diagnostics Check	OK	19:59

---

Data Download Date/Time: 06/11/2019 08:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENLEY, WILLIAM, C  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: SHORETTE, BRIAN  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 05/06/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 792976  
DRY GAS EXPIRY DATE: 06/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: MENZIE

DATE/TIME OF TEST: 05/06/2019 21:26:30

---

Diagnostics Check	OK	21:19
Air Blank	0.000	21:19
Control Test	0.080	21:20
Air Blank	0.000	21:20
Subject Sample # 1	0.159	21:21
Breath Volume	3.097	21:21
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.162	21:24
Breath Volume	3.261	21:24
Air Blank	0.000	21:25
Control Test	0.083	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

---

Data Download Date/Time: 06/12/2019 20:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: MCGUIRE,HOWARD  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 05/06/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 792976  
DRY GAS EXPIRY DATE: 06/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 05/23/2019 21:42:06

---

Diagnostics Check	OK	21:30	
Air Blank	0.000	21:31	
Control Test	0.081	21:31	
Air Blank	0.000	21:31	
Subject Sample	# 1	21:35	No Sample Provided
Breath Volume	0.000	21:35	
Air Blank	0.000	21:35	
Air Blank	0.000	21:37	
Subject Sample	# 2	21:40	No Sample Provided
Breath Volume	0.000	21:40	
Air Blank	0.000	21:41	
Control Test	0.081	21:41	
Air Blank	0.000	21:42	
Diagnostics Check	OK	21:42	

---

Data Download Date/Time: 06/12/2019 20:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENLEY, WILLIAM, C  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: HENLEY, WILLIAM  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 05/06/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 792976  
DRY GAS EXPIRY DATE: 06/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 05/27/2019 21:40:18

---

Diagnostics Check	OK	21:32
Air Blank	0.000	21:33
Control Test	0.082	21:33
Air Blank	0.000	21:34
Subject Sample # 1	0.213	21:35
Breath Volume	2.480	21:35
Air Blank	0.000	21:36
Air Blank	0.000	21:37
Subject Sample # 2	0.199	21:38
Breath Volume	1.542	21:38
Air Blank	0.000	21:39
Control Test	0.081	21:39
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

---

Data Download Date/Time: 06/12/2019 20:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORELLANA,ADALBERTO,L  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ORELLANA,ADALBERTO  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: AICARDI

DATE/TIME OF TEST: 05/02/2019 07:51:08

---

Diagnostics Check	OK	07:37	
Air Blank	0.000	07:37	
Control Test	0.079	07:38	
Air Blank	0.000	07:38	
Subject Sample	# 1 0.157	07:41	Volume Not Met
Breath Volume	0.691	07:41	
Air Blank	0.000	07:42	
Air Blank	0.000	07:44	
Subject Sample	# 2 0.217	07:45	
Breath Volume	3.621	07:45	
Air Blank	0.000	07:45	
Air Blank	0.000	07:47	
Subject Sample	# 3 0.215	07:49	
Breath Volume	3.128	07:49	
Air Blank	0.000	07:50	
Control Test	0.073	07:50	Control Outside Tolerance
Air Blank	0.000	07:51	

---

Data Download Date/Time: 05/06/2019 21:28



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,KAYLAN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: DAVIS,KAYLAN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/10/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 09:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:28

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 05/04/2019 11:02:31

---

Diagnostics Check	OK	10:55
Air Blank	0.000	10:55
Control Test	0.080	10:56
Air Blank	0.000	10:56
Subject Sample # 1	0.000	10:57
Breath Volume	1.886	10:57
Air Blank	0.000	10:57
Air Blank	0.000	10:59
Subject Sample # 2	0.000	11:01
Breath Volume	2.000	11:01
Air Blank	0.000	11:01
Control Test	0.080	11:01
Air Blank	0.000	11:02
Diagnostics Check	OK	11:02

---

Data Download Date/Time: 05/06/2019 21:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,NELSON,Y  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: HERNANDEZ,NELSON  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: BOYCE

DATE/TIME OF TEST: 05/04/2019 03:04:20

---

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:00	
Control Test	0.079	03:01	
Air Blank	0.000	03:01	
Subject Sample	# 1	03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:03	
Control Test	0.079	03:03	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

---

Data Download Date/Time: 05/31/2019 12:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,MARQUISE,A  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: WILLIAMS,MARQUISE  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: STEPHENSON

DATE/TIME OF TEST: 05/09/2019 02:26:41

---

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:19	
Control Test	0.079	02:20	
Air Blank	0.000	02:20	
Subject Sample	# 1 0.172	02:22	
Breath Volume	2.359	02:22	
Air Blank	0.000	02:23	
Air Blank	0.000	02:24	
Subject Sample	# 2 0.166	02:25	
Breath Volume	1.691	02:25	
Air Blank		02:25	Improper Sample
Air Blank	0.000	02:26	

---

Data Download Date/Time: 05/31/2019 12:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,MARQUISE,A  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: WILLIAMS,MARQUISE  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: STEPHENSON

DATE/TIME OF TEST: 05/09/2019 02:43:43

---

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.078	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.182	02:38
Breath Volume	1.710	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.181	02:42
Breath Volume	2.660	02:42
Air Blank	0.000	02:42
Control Test	0.079	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

---

Data Download Date/Time: 05/31/2019 12:58



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MIGUEL,A  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: MEDINA,MIGUEL  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 05/11/2019 21:20:20

---

Diagnostics Check		OK	21:10
Air Blank		0.000	21:10
Control Test		0.075	21:10
Air Blank		0.000	21:11
Subject Sample	# 1	0.214	21:14
Breath Volume		1.191	21:14
Air Blank		0.000	21:15
Air Blank		0.000	21:16
Subject Sample	# 2	0.217	21:18
Breath Volume		1.636	21:18
Air Blank		0.000	21:19
Control Test		0.077	21:19
Air Blank		0.000	21:20
Diagnostics Check		OK	21:20

---

Data Download Date/Time: 05/31/2019 12:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA,BRYAN,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: GARCIA,BRYAN  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MAIZAS

DATE/TIME OF TEST: 05/14/2019 02:40:26

---

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:38	
Subject Sample	# 1	02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:39	
Control Test	0.079	02:39	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

---

Data Download Date/Time: 05/31/2019 12:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORENO,PEDRO,J  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: MORENO,PEDRO  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: PANEQUE GARCIA

DATE/TIME OF TEST: 05/14/2019 21:16:41

---

Diagnostics Check	OK	21:13	
Air Blank	0.000	21:13	
Control Test	0.080	21:13	
Air Blank	0.000	21:14	
Subject Sample	# 1	21:15	Subject Test Refused
Breath Volume	0.000	21:15	
Air Blank	0.000	21:15	
Control Test	0.079	21:16	
Air Blank	0.000	21:16	
Diagnostics Check	OK	21:16	

---

Data Download Date/Time: 05/31/2019 12:58



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ACUNA,C,A  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: MUNNE,L  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 14:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:28

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 05/15/2019 15:00:33

---

Diagnostics Check		OK	14:48	
Air Blank		0.000	14:49	
Control Test		0.079	14:49	
Air Blank		0.000	14:49	
Subject Sample	# 1	0.248	14:53	Volume Not Met
Breath Volume		0.156	14:53	
Air Blank		0.000	14:53	
Air Blank		0.000	14:55	
Subject Sample	# 2	0.255	14:58	Volume Not Met
Breath Volume		0.875	14:58	
Air Blank		0.000	14:59	
Control Test		0.077	14:59	
Air Blank		0.000	15:00	
Diagnostics Check		OK	15:00	

---

Data Download Date/Time: 05/31/2019 12:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HERNANDEZ,JASIEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: PINIO,BRYAN

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 640976

DRY GAS EXPIRY DATE: 11/21/2019

VIOLATION CODE: DUI

ARREST TIME: 00:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SILVESTIRI

DATE/TIME OF TEST: 05/16/2019 01:33:17

---

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:30	
Control Test	0.078	01:30	
Air Blank	0.000	01:30	
Subject Sample	# 1	01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:32	
Control Test	0.078	01:32	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

---

Data Download Date/Time: 05/31/2019 12:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUINTANA,JORGE,F  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO,GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: PINEDA LEON

DATE/TIME OF TEST: 05/18/2019 03:09:02

---

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.079	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:08	
Control Test	0.080	03:08	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:09	

---

Data Download Date/Time: 05/31/2019 12:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,K,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: GONZALEZ,K  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 05/23/2019 01:31:07

---

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:27	
Control Test	0.080	01:28	
Air Blank	0.000	01:28	
Subject Sample	# 1	01:29	Subject Test Refused
Breath Volume	0.000	01:29	
Air Blank	0.000	01:30	
Control Test	0.080	01:30	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

---

Data Download Date/Time: 05/31/2019 12:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WASHINGTON,R,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: WASHINGTON,R  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 05/25/2019 01:10:50

---

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.081	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.163	01:06
Breath Volume	1.890	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:08
Subject Sample # 2	0.158	01:09
Breath Volume	1.394	01:09
Air Blank	0.000	01:09
Control Test	0.079	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

---

Data Download Date/Time: 05/31/2019 12:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANCO,GREGORY,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO,GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: SALABARRIA

DATE/TIME OF TEST: 05/25/2019 21:53:53

---

Diagnostics Check	OK	21:47
Air Blank	0.000	21:48
Control Test	0.080	21:48
Air Blank	0.000	21:48
Subject Sample # 1	0.295	21:49
Breath Volume	2.253	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:51
Subject Sample # 2	0.308	21:52
Breath Volume	2.187	21:52
Air Blank	0.000	21:52
Control Test	0.081	21:53
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

---

Data Download Date/Time: 05/31/2019 12:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WASHINGTON,R,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: ARREOLA,M  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: GARCIA BARRERA

DATE/TIME OF TEST: 05/26/2019 00:39:03

---

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:36	
Control Test	0.080	00:36	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:37	Subject Test Refused
Breath Volume	0.000	00:37	
Air Blank	0.000	00:38	
Control Test	0.080	00:38	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:39	

---

Data Download Date/Time: 05/31/2019 12:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 640976
BREATH TEST OPERATOR: AREVALO-RODRIGUEZ,NILSON,G	DRY GAS EXPIRY DATE: 11/21/2019
BREATH TEST OPERATOR AGENCY: Miami PD	VIOLATION CODE: DUI
ARRESTING OFFICER: CABRERA,C.	ARREST TIME: 08:40
ARRESTING OFFICER AGENCY: Miami PD	GENDER: Male
LAST AGENCY INSPECTION: 04/29/2019	OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: AREVALO-RODRIGUEZ      DATE/TIME OF TEST: 05/26/2019      08:29:10

---

Diagnostics Check	OK	08:26	
Air Blank	0.000	08:26	
Control Test	0.081	08:27	
Air Blank	0.000	08:27	
Subject Sample	# 1	08:27	Subject Test Refused
Breath Volume	0.000	08:27	
Air Blank	0.000	08:28	
Control Test	0.080	08:28	
Air Blank	0.000	08:29	
Diagnostics Check	OK	08:29	

---

Data Download Date/Time: 05/31/2019 12:59



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEROSIER,JOANNE,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: DEROSIER,JOANNE  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: MORA

DATE/TIME OF TEST: 05/26/2019 17:44:03

---

Diagnostics Check	OK	17:38	
Air Blank	0.000	17:39	
Control Test	0.080	17:39	
Air Blank	0.000	17:40	
Subject Sample	# 1	17:43	No Sample Provided
Breath Volume	0.000	17:43	
Air Blank		17:43	Sequence Aborted
Air Blank	0.000	17:44	

---

Data Download Date/Time: 05/31/2019 12:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DEROSIER,JOANNE,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: DEROSIER,JOANNE

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 640976

DRY GAS EXPIRY DATE: 11/21/2019

VIOLATION CODE: DUI

ARREST TIME: 16:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: MORA

DATE/TIME OF TEST: 05/26/2019 17:53:58

---

Diagnostics Check	OK	17:51	
Air Blank	0.000	17:51	
Control Test	0.080	17:51	
Air Blank	0.000	17:52	
Subject Sample	# 1	17:52	Subject Test Refused
Breath Volume	0.000	17:52	
Air Blank	0.000	17:53	
Control Test	0.081	17:53	
Air Blank	0.000	17:53	
Diagnostics Check	OK	17:53	

---

Data Download Date/Time: 05/31/2019 12:59



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASTOR,JOSE,F  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: FRANCO,G  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 04/29/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 640976  
DRY GAS EXPIRY DATE: 11/21/2019  
VIOLATION CODE: DUI  
ARREST TIME: 05:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:16

SUBJECT'S LAST NAME: TORRES-IRIAS

DATE/TIME OF TEST: 05/31/2019 06:40:58

---

Diagnostics Check	OK	06:37	
Air Blank	0.000	06:38	
Control Test	0.080	06:38	
Air Blank	0.000	06:38	
Subject Sample	# 1	06:39	Subject Test Refused
Breath Volume	0.000	06:39	
Air Blank	0.000	06:40	
Control Test	0.080	06:40	
Air Blank	0.000	06:40	
Diagnostics Check	OK	06:40	

---

Data Download Date/Time: 05/31/2019 12:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: CHLEBANOWSKI,RYAN  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 05/02/2019 20:23:13

---

Diagnostics Check	OK	20:20	
Air Blank	0.000	20:20	
Control Test	0.078	20:20	
Air Blank	0.000	20:21	
Subject Sample	# 1	20:21	Subject Test Refused
Breath Volume	0.000	20:21	
Air Blank	0.000	20:22	
Control Test	0.078	20:22	
Air Blank	0.000	20:23	
Diagnostics Check	OK	20:23	

---

Data Download Date/Time: 06/05/2019 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TYLER,BRYAN,D  
BREATH TEST OPERATOR AGENCY: KEY WEST PD  
ARRESTING OFFICER: TYLER,BRYAN  
ARRESTING OFFICER AGENCY: KEY WEST PD  
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: ESCOBAR SOSA

DATE/TIME OF TEST: 05/05/2019 01:19:00

---

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.078	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.251	01:14
Breath Volume		1.566	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:16
Subject Sample	# 2	0.244	01:17
Breath Volume		1.917	01:17
Air Blank		0.000	01:18
Control Test		0.077	01:18
Air Blank		0.000	01:18
Diagnostics Check		OK	01:19

---

Data Download Date/Time: 06/05/2019 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TYLER,BRYAN,D  
BREATH TEST OPERATOR AGENCY: KEY WEST PD  
ARRESTING OFFICER: WAITE,KENNETH  
ARRESTING OFFICER AGENCY: KEY WEST PD  
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 05/06/2019 02:20:27

---

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.079	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.190	02:16
Breath Volume	1.324	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.181	02:18
Breath Volume	1.109	02:18
Air Blank	0.000	02:19
Control Test	0.076	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

---

Data Download Date/Time: 06/05/2019 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOGERLEY,NEIL,F  
BREATH TEST OPERATOR AGENCY: KEY WEST PD  
ARRESTING OFFICER: CHESNEY,KYLE  
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)  
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: LANCASTER

DATE/TIME OF TEST: 05/10/2019 23:23:27

---

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.079	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.115	23:18
Breath Volume	1.578	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.111	23:21
Breath Volume	1.757	23:21
Air Blank	0.000	23:22
Control Test	0.078	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

---

Data Download Date/Time: 06/05/2019 11:04



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,ANDREW,J  
BREATH TEST OPERATOR AGENCY: KEY WEST PD  
ARRESTING OFFICER: HUDSON,ANDREW  
ARRESTING OFFICER AGENCY: KEY WEST PD  
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: HESTER

DATE/TIME OF TEST: 05/12/2019 00:59:22

---

Diagnostics Check		OK	00:52
Air Blank		0.000	00:53
Control Test		0.079	00:53
Air Blank		0.000	00:54
Subject Sample	# 1	0.216	00:54
Breath Volume		1.667	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.219	00:57
Breath Volume		2.242	00:57
Air Blank		0.000	00:58
Control Test		0.077	00:58
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

---

Data Download Date/Time: 06/05/2019 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,ANDREW,J  
BREATH TEST OPERATOR AGENCY: KEY WEST PD  
ARRESTING OFFICER: CUNEO,EDWARD  
ARRESTING OFFICER AGENCY: KEY WEST PD  
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: ABREU

DATE/TIME OF TEST: 05/12/2019 02:18:53

---

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:16	
Control Test	0.079	02:16	
Air Blank	0.000	02:17	
Subject Sample	# 1	02:17	Subject Test Refused
Breath Volume	0.000	02:17	
Air Blank	0.000	02:17	
Control Test	0.079	02:18	
Air Blank	0.000	02:18	
Diagnostics Check	OK	02:18	

---

Data Download Date/Time: 06/05/2019 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLEN,JOHN,M  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: RODDY,AARON  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: SEAY

DATE/TIME OF TEST: 05/12/2019 17:51:30

---

Diagnostics Check	OK	17:44
Air Blank	0.000	17:45
Control Test	0.079	17:45
Air Blank	0.000	17:46
Subject Sample # 1	0.048	17:46
Breath Volume	1.414	17:46
Air Blank	0.000	17:47
Air Blank	0.000	17:49
Subject Sample # 2	0.047	17:49
Breath Volume	1.503	17:49
Air Blank	0.000	17:50
Control Test	0.078	17:50
Air Blank	0.000	17:51
Diagnostics Check	OK	17:51

---

Data Download Date/Time: 06/05/2019 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,ANDREW,J  
BREATH TEST OPERATOR AGENCY: KEY WEST PD  
ARRESTING OFFICER: CHESNEY,KYLE  
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)  
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: ALVAREZ SOTOMAYOR

DATE/TIME OF TEST: 05/12/2019 20:20:50

---

Diagnostics Check		OK	20:12	
Air Blank		0.000	20:12	
Control Test		0.079	20:13	
Air Blank		0.000	20:13	
Subject Sample	# 1	0.132	20:16	Volume Not Met
Breath Volume		0.585	20:16	
Air Blank		0.000	20:17	
Air Blank		0.000	20:19	
Subject Sample	# 2		20:19	Subject Test Refused
Breath Volume		0.000	20:19	
Air Blank		0.000	20:19	
Control Test		0.076	20:20	
Air Blank		0.000	20:20	
Diagnostics Check		OK	20:20	

---

Data Download Date/Time: 06/05/2019 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,ANDREW,J  
BREATH TEST OPERATOR AGENCY: KEY WEST PD  
ARRESTING OFFICER: HUDSON,ANDREW  
ARRESTING OFFICER AGENCY: KEY WEST PD  
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/17/2019 01:28:35

---

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.078	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.185	01:23
Breath Volume		1.625	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.183	01:26
Breath Volume		1.648	01:26
Air Blank		0.000	01:27
Control Test		0.078	01:27
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

---

Data Download Date/Time: 06/05/2019 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TYLER,BRYAN,D  
BREATH TEST OPERATOR AGENCY: KEY WEST PD  
ARRESTING OFFICER: OGGERI,TIM  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: TRUE

DATE/TIME OF TEST: 05/18/2019 02:04:08

---

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.078	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.194	01:59
Breath Volume	2.183	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.196	02:02
Breath Volume	1.757	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

---

Data Download Date/Time: 06/05/2019 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KEOHANE,KUNIKO,  
BREATH TEST OPERATOR AGENCY: KEY WEST PD  
ARRESTING OFFICER: KEOHANE,KUNIKO  
ARRESTING OFFICER AGENCY: KEY WEST PD  
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 10:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:44

SUBJECT'S LAST NAME: ALCINDOR

DATE/TIME OF TEST: 05/26/2019 11:29:01

---

Diagnostics Check	OK	11:20
Air Blank	0.000	11:20
Control Test	0.079	11:21
Air Blank	0.000	11:21
Subject Sample # 1	0.208	11:23
Breath Volume	1.375	11:23
Air Blank	0.000	11:23
Air Blank	0.000	11:25
Subject Sample # 2	0.211	11:27
Breath Volume	1.441	11:27
Air Blank	0.000	11:28
Control Test	0.077	11:28
Air Blank	0.000	11:28
Diagnostics Check	OK	11:29

---

Data Download Date/Time: 06/05/2019 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,ANDREW,J  
BREATH TEST OPERATOR AGENCY: KEY WEST PD  
ARRESTING OFFICER: HUDSON,ANDREW  
ARRESTING OFFICER AGENCY: KEY WEST PD  
LAST AGENCY INSPECTION: 05/01/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SZEGLIN

DATE/TIME OF TEST: 05/27/2019 02:40:08

---

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.078	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.146	02:35
Breath Volume	2.984	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:37
Subject Sample # 2	0.144	02:38
Breath Volume	2.656	02:38
Air Blank	0.000	02:39
Control Test	0.078	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

---

Data Download Date/Time: 06/05/2019 11:04



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BERMAN,MARC  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 05/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A1  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: HERNANDEZ CASTELLANO

DATE/TIME OF TEST: 05/17/2019 23:58:10

---

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.078	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.178	23:53
Breath Volume	1.695	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.163	23:56
Breath Volume	1.910	23:56
Air Blank	0.000	23:57
Control Test	0.078	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

---

Data Download Date/Time: 06/13/2019 13:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: O'DOR,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 05/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19817080A1  
DRY GAS EXPIRY DATE: 09/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: ASTORGA

DATE/TIME OF TEST: 05/18/2019 00:25:10

---

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.078	00:17
Air Blank	0.000	00:18
Subject Sample # 1	0.000	00:20
Breath Volume	1.492	00:20
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.000	00:23
Breath Volume	1.125	00:23
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

---

Data Download Date/Time: 06/13/2019 13:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STRALKOWSKI,SEAN,D  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: STRZALKOWSKI,SEAN  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 05/09/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: DE MOURA

DATE/TIME OF TEST: 05/24/2019 05:24:07

---

Diagnostics Check		OK	05:16
Air Blank		0.000	05:16
Control Test		0.078	05:16
Air Blank		0.000	05:17
Subject Sample	# 1	0.132	05:18
Breath Volume		2.148	05:18
Air Blank		0.000	05:18
Air Blank		0.000	05:20
Subject Sample	# 2	0.121	05:22
Breath Volume		1.109	05:22
Air Blank		0.000	05:23
Control Test		0.079	05:23
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

---

Data Download Date/Time: 06/13/2019 13:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALMER,SAMANTHA,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: GONZALEZ,ROBERTO  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: POLANCO

DATE/TIME OF TEST: 05/01/2019 00:24:57

---

Diagnostics Check	OK	00:14
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.128	00:17
Breath Volume	1.375	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.130	00:23
Breath Volume	1.449	00:23
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

---

Data Download Date/Time: 05/03/2019 13:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALMER,SAMANTHA,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SOLOWAY,ALAN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: BLANCHARD

DATE/TIME OF TEST: 05/02/2019 01:01:01

---

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.080	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.044	00:56
Breath Volume	4.121	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:58
Subject Sample # 2	0.044	00:59
Breath Volume	4.132	00:59
Air Blank	0.000	01:00
Control Test	0.079	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:01

---

Data Download Date/Time: 05/03/2019 13:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALMER,SAMANTHA,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: BORROWS,ANDREW  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 05/02/2019 03:18:44

---

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:07	
Control Test		0.080	03:07	
Air Blank		0.000	03:08	
Subject Sample	# 1	0.080	03:11	Volume Not Met
Breath Volume		0.128	03:11	
Air Blank		0.000	03:12	
Air Blank		0.000	03:14	
Subject Sample	# 2	0.075	03:17	Volume Not Met
Breath Volume		0.085	03:17	
Air Blank		0.000	03:17	
Control Test		0.079	03:18	
Air Blank		0.000	03:18	
Diagnostics Check		OK	03:18	

---

Data Download Date/Time: 05/03/2019 13:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALMER,SAMANTHA,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: BORROWS,ANDREW  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 05/02/2019 03:37:58

---

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.078	03:21	
Air Blank	0.000	03:22	
Subject Sample	# 1 0.084	03:25	
Breath Volume	1.429	03:25	
Air Blank	0.000	03:25	
Air Blank	0.000	03:27	
Subject Sample	# 2 0.073	03:30	Volume Not Met
Breath Volume	0.164	03:30	
Air Blank	0.000	03:31	
Air Blank	0.000	03:33	
Subject Sample	# 3 0.070	03:36	Volume Not Met
Breath Volume	0.015	03:36	
Air Blank	0.000	03:37	
Control Test	0.079	03:37	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

---

Data Download Date/Time: 05/03/2019 13:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALMER,SAMANTHA,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MACK,CHEYENNE  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: ARMBRUSTER

DATE/TIME OF TEST: 05/02/2019 05:15:14

---

Diagnostics Check		OK	05:00	
Air Blank		0.000	05:00	
Control Test		0.080	05:01	
Air Blank		0.000	05:01	
Subject Sample	# 1	0.117	05:03	
Breath Volume		1.257	05:03	
Air Blank		0.000	05:04	
Air Blank		0.000	05:06	
Subject Sample	# 2	0.126	05:09	Volume Not Met
Breath Volume		0.085	05:09	
Air Blank		0.000	05:10	
Air Blank		0.000	05:11	
Subject Sample	# 3	0.127	05:13	
Breath Volume		1.343	05:13	
Air Blank		0.000	05:14	
Control Test		0.077	05:14	
Air Blank		0.000	05:15	
Diagnostics Check		OK	05:15	

---

Data Download Date/Time: 05/03/2019 13:19



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CASTROS,DENNIS  
ARRESTING OFFICER AGENCY: Boynton Beach PD  
LAST AGENCY INSPECTION: 04/19/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 05/02/2019 23:54:03

---

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.080	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.000	23:48
Breath Volume	1.246	23:48
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.000	23:52
Breath Volume	1.753	23:52
Air Blank	0.000	23:53
Control Test	0.081	23:53
Air Blank	0.000	23:53
Diagnostics Check	OK	23:54

---

Data Download Date/Time: 05/03/2019 13:19