Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/04/2019 18:28:11

Diagnostics Check		OK	18:26	
Air Blank		0.000	18:26	
Control Test		0.080	18:26	
Air Blank		0.000	18:27	
Subject Sample	# 1		18:27	Sequence Aborted
Breath Volume		0.000	18:27	
Air Blank		0.000	18:28	

Data Download Date/Time: 07/17/2019 12:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/04/2019 18:54:16

Diagnostics Check		OK	18:48
Air Blank		0.000	18:48
Control Test		0.081	18:49
Air Blank		0.000	18:49
Subject Sample	# 1	0.000	18:50
Breath Volume		2.507	18:50
Air Blank		0.000	18:50
Air Blank		0.000	18:52
Subject Sample	# 2	0.000	18:52
Breath Volume		2.726	18:52
Air Blank		0.000	18:53
Control Test		0.080	18:53
Air Blank		0.000	18:54
Diagnostics Check		OK	18:54

Data Download Date/Time: 07/17/2019 12:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODWARD,KYLER,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WODWARD,KYLER
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: VANCE

DATE/TIME OF TEST: 07/04/2019 19:58:31

Diagnostics Check		OK	19:52	
Air Blank		0.000	19:53	
Control Test		0.078	19:53	
Air Blank		0.000	19:53	
Subject Sample	# 1		19:56	Subject Test Refused
Breath Volume		0.125	19:56	
Air Blank		0.000	19:57	
Control Test		0.078	19:57	
Air Blank		0.000	19:58	
Diagnostics Check		OK	19:58	

Data Download Date/Time: 07/17/2019 12:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,RHETT,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARVEY,RHETT
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 07/05/2019 00:14:15

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.080	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.172	00:09
Breath Volume		1.203	00:09
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 2	0.159	00:12
Breath Volume		2.000	00:12
Air Blank		0.000	00:13
Control Test		0.079	00:13
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 07/17/2019 12:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE,CHARLES,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 07/05/2019 19:11:26

Diagnostics Check		OK	19:05
Air Blank		0.000	19:05
Control Test		0.080	19:06
Air Blank		0.000	19:06
Subject Sample	# 1	0.000	19:07
Breath Volume		1.281	19:07
Air Blank		0.000	19:07
Air Blank		0.000	19:09
Subject Sample	# 2	0.000	19:10
Breath Volume		2.738	19:10
Air Blank		0.000	19:10
Control Test		0.080	19:10
Air Blank		0.000	19:11
Diagnostics Check		OK	19:11

Data Download Date/Time: 07/17/2019 12:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAS,BRANDYN,M
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: DEAS,BRANDYN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: KJELLEVOLD

DATE/TIME OF TEST: 07/05/2019 21:21:50

Diagnostics Check		OK	21:15
Air Blank		0.000	21:16
Control Test		0.079	21:16
Air Blank		0.000	21:16
Subject Sample	# 1	0.144	21:17
Breath Volume		1.878	21:17
Air Blank		0.000	21:18
Air Blank		0.000	21:19
Subject Sample	# 2	0.146	21:20
Breath Volume		1.703	21:20
Air Blank		0.000	21:20
Control Test		0.079	21:21
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 07/17/2019 12:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: GERMANY

DATE/TIME OF TEST: 07/06/2019 05:15:00

Diagnostics Check		OK	05:08
Air Blank		0.000	05:08
Control Test		0.080	05:09
Air Blank		0.000	05:09
Subject Sample	# 1	0.105	05:10
Breath Volume		1.136	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:12
Subject Sample	# 2	0.112	05:13
Breath Volume		1.363	05:13
Air Blank		0.000	05:14
Control Test		0.079	05:14
Air Blank		0.000	05:14
Diagnostics Check		OK	05:15

Data Download Date/Time: 07/17/2019 12:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,CLIFFORD,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CROUCH,CLIFFORD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:59

SUBJECT'S LAST NAME: SAENZ

DATE/TIME OF TEST: 07/06/2019 07:35:20

Diagnostics Check		OK	07:26
Air Blank		0.000	07:27
Control Test		0.080	07:27
Air Blank		0.000	07:28
Subject Sample	# 1	0.189	07:29
Breath Volume		2.402	07:29
Air Blank		0.000	07:29
Air Blank		0.000	07:31
Subject Sample	# 2	0.188	07:33
Breath Volume		2.671	07:33
Air Blank		0.000	07:34
Control Test		0.080	07:34
Air Blank		0.000	07:35
Diagnostics Check		OK	07:35

Data Download Date/Time: 07/17/2019 12:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: THOMPSON,JEFFREY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: SHAIKH

DATE/TIME OF TEST: 07/06/2019 22:43:30

Diagnostics Check		OK	22:37
Air Blank		0.000	22:37
Control Test		0.081	22:38
Air Blank		0.000	22:38
Subject Sample	# 1	0.131	22:39
Breath Volume		3.304	22:39
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.133	22:42
Breath Volume		3.554	22:42
Air Blank		0.000	22:42
Control Test		0.080	22:42
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 07/17/2019 12:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,ROBERT,A
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WESTMAN,PATRICK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: THOMAS-BRYANT

DATE/TIME OF TEST: 07/07/2019 03:22:56

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.079	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.167	03:18
Breath Volume		1.902	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:20
Subject Sample	# 2	0.167	03:21
Breath Volume		1.750	03:21
Air Blank		0.000	03:22
Control Test		0.079	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 07/17/2019 12:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:43

SUBJECT'S LAST NAME: DOSS

DATE/TIME OF TEST: 07/12/2019 06:14:55

Diagnostics Check		OK	06:08
Air Blank		0.000	06:09
Control Test		0.080	06:09
Air Blank		0.000	06:10
Subject Sample	# 1	0.187	06:10
Breath Volume		2.398	06:10
Air Blank		0.000	06:11
Air Blank		0.000	06:12
Subject Sample	# 2	0.187	06:13
Breath Volume		2.257	06:13
Air Blank		0.000	06:14
Control Test		0.080	06:14
Air Blank		0.000	06:14
Diagnostics Check		OK	06:14

Data Download Date/Time: 07/17/2019 12:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE,CHARLES,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 07/13/2019 21:21:21

Diagnostics Check		OK	21:15
Air Blank		0.000	21:15
Control Test		0.080	21:16
Air Blank		0.000	21:16
Subject Sample	# 1	0.011	21:17
Breath Volume		2.808	21:17
Air Blank		0.000	21:17
Air Blank		0.000	21:19
Subject Sample	# 2	0.011	21:20
Breath Volume		2.085	21:20
Air Blank		0.000	21:20
Control Test		0.080	21:20
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 07/17/2019 12:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISCHER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: FISCHER,ANDREW
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: WHITE SPRINGER

DATE/TIME OF TEST: 07/14/2019 03:59:12

Diagnostics Check		OK	03:52
Air Blank		0.000	03:52
Control Test		0.080	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.185	03:54
Breath Volume		1.410	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:56
Subject Sample	# 2	0.186	03:57
Breath Volume		1.535	03:57
Air Blank		0.000	03:58
Control Test		0.079	03:58
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 07/17/2019 12:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: WAID

DATE/TIME OF TEST: 07/14/2019 05:45:20

Diagnostics Check		OK	05:38
Air Blank		0.000	05:38
Control Test		0.080	05:39
Air Blank		0.000	05:39
Subject Sample	# 1	0.185	05:40
Breath Volume		1.695	05:40
Air Blank		0.000	05:41
Air Blank		0.000	05:43
Subject Sample	# 2	0.195	05:43
Breath Volume		1.437	05:43
Air Blank		0.000	05:44
Control Test		0.079	05:44
Air Blank		0.000	05:45
Diagnostics Check		OK	05:45

Data Download Date/Time: 07/17/2019 12:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 07/01/2019 05:56:29

Diagnostics Check		OK	05:50
Air Blank		0.000	05:50
Control Test		0.080	05:50
Air Blank		0.000	05:51
Subject Sample	# 1	0.138	05:52
Breath Volume		1.703	05:52
Air Blank		0.000	05:52
Air Blank		0.000	05:54
Subject Sample	# 2	0.138	05:55
Breath Volume		1.421	05:55
Air Blank		0.000	05:55
Control Test		0.079	05:55
Air Blank		0.000	05:56
Diagnostics Check		OK	05:56

Data Download Date/Time: 07/17/2019 11:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEFFIELD,BRANDON,S
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: BAUMANN,BRANDON
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: JANNEH

DATE/TIME OF TEST: 07/06/2019 01:30:28

Diagnostics Check		OK	01:23
Air Blank		0.000	01:24
Control Test		0.080	01:24
Air Blank		0.000	01:25
Subject Sample	# 1	0.000	01:26
Breath Volume		1.132	01:26
Air Blank		0.000	01:26
Air Blank		0.000	01:28
Subject Sample	# 2	0.000	01:29
Breath Volume		2.519	01:29
Air Blank		0.000	01:29
Control Test		0.079	01:29
Air Blank		0.000	01:30
Diagnostics Check		OK	01:30

Data Download Date/Time: 07/17/2019 11:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: DEMING,W
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: VONTRAPP

DATE/TIME OF TEST: 07/17/2019 01:57:11

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:53	
Control Test		0.080	01:53	
Air Blank		0.000	01:53	
Subject Sample	# 1		01:56	Subject Test Refused
Breath Volume		0.906	01:56	
Air Blank			01:56	RFI Detect
Air Blank		0.000	01:57	

Data Download Date/Time: 07/17/2019 11:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 07/17/2019 23:39:08

Diagnostics Check		OK	23:33
Air Blank		0.000	23:33
Control Test		0.080	23:33
Air Blank		0.000	23:34
Subject Sample	# 1	0.000	23:34
Breath Volume		2.066	23:34
Air Blank		0.000	23:35
Air Blank		0.000	23:37
Subject Sample	# 2	0.000	23:37
Breath Volume		3.222	23:37
Air Blank		0.000	23:38
Control Test		0.080	23:38
Air Blank		0.000	23:39
Diagnostics Check		OK	23:39

Data Download Date/Time: 08/26/2019 10:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE,CHARLES,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/17/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: REZA

DATE/TIME OF TEST: 07/18/2019 23:21:46

Diagnostics Check		OK	23:15
Air Blank		0.000	23:16
Control Test		0.079	23:16
Air Blank		0.000	23:16
Subject Sample	# 1	0.000	23:17
Breath Volume		3.238	23:17
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.000	23:20
Breath Volume		3.035	23:20
Air Blank		0.000	23:20
Control Test		0.080	23:21
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 08/26/2019 10:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNON,C,
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: BROWN,J
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 07/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: MCWHORTER

DATE/TIME OF TEST: 07/20/2019 03:53:03

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.080	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.148	03:45
Breath Volume		4.281	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.124	03:48
Breath Volume		1.722	03:48
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 3	0.121	03:51
Breath Volume		1.832	03:51
Air Blank		0.000	03:52
Control Test		0.079	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:53

Data Download Date/Time: 08/26/2019 10:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITAKER,H,H
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: WHITAKER,H
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: MEADE

DATE/TIME OF TEST: 07/20/2019 22:48:35

Diagnostics Check		OK	22:41
Air Blank		0.000	22:42
Control Test		0.080	22:42
Air Blank		0.000	22:43
Subject Sample	# 1	0.217	22:44
Breath Volume		1.734	22:44
Air Blank		0.000	22:44
Air Blank		0.000	22:46
Subject Sample	# 2	0.226	22:47
Breath Volume		2.453	22:47
Air Blank		0.000	22:47
Control Test		0.079	22:48
Air Blank		0.000	22:48
Diagnostics Check		OK	22:48

Data Download Date/Time: 08/26/2019 10:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE,CHARLES,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 07/21/2019 23:47:29

Diagnostics Check		OK	23:44	
Air Blank		0.000	23:45	
Control Test		0.080	23:45	
Air Blank		0.000	23:45	
Subject Sample	# 1		23:46	Subject Test Refused
Breath Volume		0.000	23:46	
Air Blank		0.000	23:46	
Control Test		0.080	23:46	
Air Blank		0.000	23:47	
Diagnostics Check		OK	23:47	

Data Download Date/Time: 08/26/2019 10:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 07/22/2019 23:21:54

Diagnostics Check		OK	23:15
Air Blank		0.000	23:16
Control Test		0.079	23:16
Air Blank		0.000	23:17
Subject Sample	# 1	0.105	23:17
Breath Volume		3.117	23:17
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.110	23:20
Breath Volume		3.378	23:20
Air Blank		0.000	23:21
Control Test		0.079	23:21
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 08/26/2019 10:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,RAYMOND,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WHITE,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: GORTON

DATE/TIME OF TEST: 07/24/2019 01:19:20

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.080	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.000	01:14
Breath Volume		4.113	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.000	01:17
Breath Volume		4.316	01:17
Air Blank		0.000	01:18
Control Test		0.080	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 08/26/2019 10:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD,LEE,M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: VARNER,TYLER
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 07/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/25/2019 03:08:41

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.080	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.000	03:04
Breath Volume		2.808	03:04
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.000	03:07
Breath Volume		4.554	03:07
Air Blank		0.000	03:07
Control Test		0.081	03:08
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 08/26/2019 10:31

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: HERNANDEZ RODRIGUEZ

DATE/TIME OF TEST: 07/27/2019 03:22:29

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.213	03:18
Breath Volume		1.261	03:18
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.207	03:21
Breath Volume		1.746	03:21
Air Blank		0.000	03:21
Control Test		0.079	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 08/26/2019 10:31

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 07/28/2019 04:00:38

Diagnostics Check		OK	03:54
Air Blank		0.000	03:54
Control Test		0.079	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.202	03:56
Breath Volume		2.296	03:56
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.191	03:59
Breath Volume		2.402	03:59
Air Blank		0.000	03:59
Control Test		0.080	04:00
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 08/26/2019 10:31

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PURINGTON,PATRICK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 07/02/2019 03:31:54

Diagnostics Check		OK	03:29	
Air Blank		0.000	03:29	
Control Test		0.080	03:29	
Air Blank		0.000	03:30	
Subject Sample	# 1		03:30	Subject Test Refused
Breath Volume		0.000	03:30	
Air Blank		0.000	03:30	
Control Test		0.080	03:31	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 07/12/2019 15:23

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SABO,JESSICA,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SABO,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/05/2019 05:38:56

Diagnostics Check		OK	05:32
Air Blank		0.000	05:33
Control Test		0.081	05:33
Air Blank		0.000	05:33
Subject Sample	# 1	0.066	05:34
Breath Volume		4.039	05:34
Air Blank		0.000	05:35
Air Blank		0.000	05:36
Subject Sample	# 2	0.063	05:37
Breath Volume		3.050	05:37
Air Blank		0.000	05:38
Control Test		0.079	05:38
Air Blank		0.000	05:38
Diagnostics Check		OK	05:38

Data Download Date/Time: 07/12/2019 15:23

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PURINGTON,PATRICK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 07/11/2019 01:33:32

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:22	
Control Test		0.080	01:22	
Air Blank		0.000	01:23	
Subject Sample	# 1	0.120	01:26	Volume Not Met
Breath Volume		0.195	01:26	
Air Blank		0.000	01:27	
Air Blank		0.000	01:28	
Subject Sample	# 2	0.118	01:29	
Breath Volume		1.996	01:29	
Air Blank		0.000	01:29	
Air Blank		0.000	01:31	
Subject Sample	# 3	0.121	01:32	
Breath Volume		1.425	01:32	
Air Blank		0.000	01:32	
Control Test		0.076	01:32	
Air Blank		0.000	01:33	
Diagnostics Check		OK	01:33	

Data Download Date/Time: 07/12/2019 15:23

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HOWARD,ANTHONY
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 07/13/2019 04:31:01

Diagnostics Check		OK	04:19	
Air Blank		0.000	04:20	
Control Test		0.080	04:20	
Air Blank		0.000	04:20	
Subject Sample	# 1		04:24	No Sample Provided
Breath Volume		0.000	04:24	
Air Blank		0.000	04:24	
Air Blank		0.000	04:26	
Subject Sample	# 2		04:29	No Sample Provided
Breath Volume		0.000	04:29	
Air Blank		0.000	04:30	
Control Test		0.081	04:30	
Air Blank		0.000	04:30	
Diagnostics Check		OK	04:31	

Data Download Date/Time: 08/27/2019 16:32

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITZKOW,CLAYTON,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUGHES,DAVID
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/14/2019 01:14:46

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.079	01:08
Air Blank		0.000	01:08
Subject Sample	# 1	0.130	01:09
Breath Volume		1.750	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.140	01:13
Breath Volume		3.324	01:13
Air Blank		0.000	01:13
Control Test		0.078	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 08/27/2019 16:32

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PARKS,DONALD
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: KLEMA

DATE/TIME OF TEST: 07/14/2019 03:30:23

Diagnostics Check		OK	03:27	
Air Blank		0.000	03:27	
Control Test		0.080	03:28	
Air Blank		0.000	03:28	
Subject Sample	# 1		03:29	Subject Test Refused
Breath Volume		0.000	03:29	
Air Blank		0.000	03:29	
Control Test		0.080	03:29	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:30	

Data Download Date/Time: 08/27/2019 16:32

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCEACHIN,SCOTT,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MURPHY,JAVON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: ORONA

DATE/TIME OF TEST: 07/21/2019 04:20:16

Diagnostics Check		OK	04:14
Air Blank		0.000	04:14
Control Test		0.081	04:14
Air Blank		0.000	04:15
Subject Sample	# 1	0.167	04:15
Breath Volume		1.437	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.174	04:18
Breath Volume		1.535	04:18
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 08/27/2019 16:32

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITZKOW,CARLTON,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LITZKOW,CARLTON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 07/23/2019 05:52:58

Diagnostics Check		OK	05:47	
Air Blank		0.000	05:47	
Control Test		0.079	05:48	
Air Blank		0.000	05:48	
Subject Sample	# 1		05:51	Subject Test Refused
Breath Volume		0.000	05:51	
Air Blank		0.000	05:52	
Control Test		0.081	05:52	
Air Blank		0.000	05:52	
Diagnostics Check		OK	05:52	

Data Download Date/Time: 08/27/2019 16:32

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLON,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MOORE,GARY
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 07/24/2019 21:17:34

Diagnostics Check		OK	21:12	
Air Blank		0.000	21:13	
Control Test		0.079	21:13	
Air Blank		0.000	21:14	
Subject Sample	# 1		21:17	Sequence Aborted
Breath Volume		0.000	21:17	
Air Blank		0.000	21:17	

Data Download Date/Time: 08/27/2019 16:32

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITZKOW,CARLTON,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LITZKOW,CARLTON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 07/27/2019 07:00:07

Diagnostics Check		OK	06:52
Air Blank		0.000	06:53
Control Test		0.080	06:53
Air Blank		0.000	06:54
Subject Sample	# 1	0.126	06:54
Breath Volume		1.472	06:54
Air Blank		0.000	06:55
Air Blank		0.000	06:57
Subject Sample	# 2	0.136	06:58
Breath Volume		3.261	06:58
Air Blank		0.000	06:59
Control Test		0.079	06:59
Air Blank		0.000	07:00
Diagnostics Check		OK	07:00

Data Download Date/Time: 08/27/2019 16:32

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: PINO

DATE/TIME OF TEST: 07/27/2019 07:18:40

Diagnostics Check		OK	07:11
Air Blank		0.000	07:11
Control Test		0.079	07:12
Air Blank		0.000	07:12
Subject Sample	# 1	0.151	07:14
Breath Volume		1.449	07:14
Air Blank		0.000	07:14
Air Blank		0.000	07:16
Subject Sample	# 2	0.148	07:17
Breath Volume		1.269	07:17
Air Blank		0.000	07:17
Control Test		0.079	07:18
Air Blank		0.000	07:18
Diagnostics Check		OK	07:18

Data Download Date/Time: 08/27/2019 16:32

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/27/2019 21:23:55

Diagnostics Check		OK	21:12	
Air Blank		0.000	21:12	
Control Test		0.080	21:13	
Air Blank		0.000	21:13	
Subject Sample	# 1	0.152	21:16	Volume Not Met
Breath Volume		0.269	21:16	
Air Blank		0.000	21:17	
Air Blank		0.000	21:19	
Subject Sample	# 2	0.125	21:22	Volume Not Met
Breath Volume		0.707	21:22	
Air Blank		0.000	21:22	
Control Test		0.080	21:23	
Air Blank		0.000	21:23	
Diagnostics Check		OK	21:23	

Data Download Date/Time: 08/27/2019 16:32

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ROBERTO,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ALVAREZ,ROBERTO
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:07

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 07/28/2019 14:41:14

Diagnostics Check		OK	14:38	
Air Blank		0.000	14:38	
Control Test		0.079	14:39	
Air Blank		0.000	14:39	
Subject Sample	# 1		14:39	Subject Test Refused
Breath Volume		0.000	14:39	
Air Blank		0.000	14:40	
Control Test		0.079	14:40	
Air Blank		0.000	14:41	
Diagnostics Check		OK	14:41	

Data Download Date/Time: 08/27/2019 16:32

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUNDENBURG,TRACY,C
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BOSTICK,ANTONIO
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: VISE

DATE/TIME OF TEST: 07/28/2019 16:17:19

Diagnostics Check		OK	16:10
Air Blank		0.000	16:10
Control Test		0.080	16:11
Air Blank		0.000	16:11
Subject Sample	# 1	0.175	16:12
Breath Volume		1.820	16:12
Air Blank		0.000	16:13
Air Blank		0.000	16:15
Subject Sample	# 2	0.191	16:15
Breath Volume		2.328	16:15
Air Blank		0.000	16:16
Control Test		0.076	16:16
Air Blank		0.000	16:17
Diagnostics Check		OK	16:17

Data Download Date/Time: 08/27/2019 16:32

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTSCHER,REBECCA,L
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: BUTSCHER,REBECCA
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/30/2019 03:32:00

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:29	
Control Test		0.080	03:29	
Air Blank		0.000	03:30	
Subject Sample	# 1		03:30	Subject Test Refused
Breath Volume		0.000	03:30	
Air Blank		0.000	03:31	
Control Test		0.080	03:31	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:32	

Data Download Date/Time: 08/27/2019 16:32

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,WENDELL,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPENCE,JUSTIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 07/01/2019 22:45:01

Diagnostics Check		OK	22:38
Air Blank		0.000	22:38
Control Test		0.080	22:39
Air Blank		0.000	22:39
Subject Sample	# 1	0.128	22:40
Breath Volume		2.683	22:40
Air Blank		0.000	22:41
Air Blank		0.000	22:42
Subject Sample	# 2	0.118	22:43
Breath Volume		1.707	22:43
Air Blank		0.000	22:44
Control Test		0.079	22:44
Air Blank		0.000	22:44
Diagnostics Check		OK	22:45

Data Download Date/Time: 07/16/2019 09:07

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,ROYCE,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPENCE,JUSTIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: B650716884070

DATE/TIME OF TEST: 07/02/2019 23:52:04

Diagnostics Check		OK	23:43	
Air Blank		0.000	23:44	
Control Test		0.080	23:44	
Air Blank		0.000	23:44	
Subject Sample	# 1		23:47	No Sample Provided
Breath Volume		0.000	23:47	
Air Blank		0.000	23:48	
Air Blank		0.000	23:50	
Subject Sample	# 2		23:50	Subject Test Refused
Breath Volume		0.000	23:50	
Air Blank		0.000	23:51	
Control Test		0.080	23:51	
Air Blank		0.000	23:51	
Diagnostics Check		OK	23:52	

Data Download Date/Time: 07/16/2019 09:07

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILHELM,TONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: TORANZO

DATE/TIME OF TEST: 07/06/2019 23:46:17

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.081	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.138	23:41
Breath Volume		1.734	23:41
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.134	23:44
Breath Volume		1.359	23:44
Air Blank		0.000	23:45
Control Test		0.084	23:45
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 07/16/2019 09:07

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPENCE,JUSTIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MCGEE

DATE/TIME OF TEST: 07/07/2019 00:11:45

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.080	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.269	00:07
Breath Volume		3.175	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:09
Subject Sample	# 2	0.260	00:10
Breath Volume		2.812	00:10
Air Blank		0.000	00:10
Control Test		0.077	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 07/16/2019 09:07

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLEOD,JOEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCLEOD,JOEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/07/2019 04:29:26

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.080	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.141	04:24
Breath Volume		2.058	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.144	04:27
Breath Volume		3.054	04:27
Air Blank		0.000	04:28
Control Test		0.082	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 07/16/2019 09:07

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,ROYCE,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPENCE,JUSTIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: SEGUIAS

DATE/TIME OF TEST: 07/07/2019 05:27:00

Diagnostics Check		OK	05:19
Air Blank		0.000	05:20
Control Test		0.081	05:20
Air Blank		0.000	05:20
Subject Sample	# 1	0.109	05:21
Breath Volume		1.148	05:21
Air Blank		0.000	05:22
Air Blank		0.000	05:23
Subject Sample	# 2	0.099	05:25
Breath Volume		1.656	05:25
Air Blank		0.000	05:26
Control Test		0.080	05:26
Air Blank		0.000	05:26
Diagnostics Check		OK	05:27

Data Download Date/Time: 07/16/2019 09:07

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: TRINH

DATE/TIME OF TEST: 07/08/2019 03:48:15

Diagnostics Check		OK	03:36	
Air Blank		0.000	03:37	
Control Test		0.081	03:37	
Air Blank		0.000	03:38	
Subject Sample	# 1		03:41	No Sample Provided
Breath Volume		0.000	03:41	
Air Blank		0.000	03:41	
Air Blank		0.000	03:43	
Subject Sample	# 2		03:46	No Sample Provided
Breath Volume		0.000	03:46	
Air Blank		0.000	03:47	
Control Test		0.080	03:47	
Air Blank		0.000	03:48	
Diagnostics Check		OK	03:48	

Data Download Date/Time: 07/16/2019 09:07

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MARTIN,JAMES
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GARCIA SIFUENTES

DATE/TIME OF TEST: 07/11/2019 01:52:59

Diagnostics Check		OK	01:50	
Air Blank		0.000	01:50	
Control Test		0.081	01:50	
Air Blank		0.000	01:51	
Subject Sample	# 1		01:51	Subject Test Refused
Breath Volume		0.000	01:51	
Air Blank		0.000	01:52	
Control Test		0.081	01:52	
Air Blank		0.000	01:52	
Diagnostics Check		OK	01:52	

Data Download Date/Time: 07/16/2019 09:07

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,RAY,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAMILTON,RAY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/13/2019 21:07:56

Diagnostics Check		OK	21:04	
Air Blank		0.000	21:04	
Control Test		0.079	21:05	
Air Blank		0.000	21:05	
Subject Sample	# 1		21:06	Subject Test Refused
Breath Volume		0.000	21:06	
Air Blank		0.000	21:07	
Control Test		0.079	21:07	
Air Blank		0.000	21:07	
Diagnostics Check		OK	21:07	

Data Download Date/Time: 07/16/2019 09:07

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOBLES,DARIUS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NOBLES,DARIUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 07/14/2019 00:30:24

Diagnostics Check		OK	00:27	
Air Blank		0.000	00:28	
Control Test		0.080	00:28	
Air Blank		0.000	00:28	
Subject Sample	# 1		00:29	Subject Test Refused
Breath Volume		0.000	00:29	
Air Blank		0.000	00:29	
Control Test		0.081	00:29	
Air Blank		0.000	00:30	
Diagnostics Check		OK	00:30	

Data Download Date/Time: 07/16/2019 09:07

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUS,CHRISTINA,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DUS,CHRISTINA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 06:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:17

SUBJECT'S LAST NAME: REDMON

DATE/TIME OF TEST: 07/14/2019 07:42:46

Diagnostics Check		OK	07:39	
Air Blank		0.000	07:40	
Control Test		0.080	07:40	
Air Blank		0.000	07:41	
Subject Sample	# 1		07:41	Subject Test Refused
Breath Volume		0.000	07:41	
Air Blank		0.000	07:41	
Control Test		0.080	07:42	
Air Blank		0.000	07:42	
Diagnostics Check		OK	07:42	

Data Download Date/Time: 07/16/2019 09:07

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEAU,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NEAU,MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: SANTIZ LOPEZ

DATE/TIME OF TEST: 07/14/2019 22:10:55

Diagnostics Check		OK	22:04
Air Blank		0.000	22:05
Control Test		0.080	22:05
Air Blank		0.000	22:05
Subject Sample	# 1	0.196	22:06
Breath Volume		3.031	22:06
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.192	22:09
Breath Volume		3.535	22:09
Air Blank		0.000	22:09
Control Test		0.078	22:10
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 07/16/2019 09:07

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNHAM,J,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURNHAM,J
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: SANFORD

DATE/TIME OF TEST: 07/15/2019 01:45:04

Diagnostics Check		OK	01:42	
Air Blank		0.000	01:42	
Control Test		0.080	01:42	
Air Blank		0.000	01:43	
Subject Sample	# 1		01:43	Subject Test Refused
Breath Volume		0.000	01:43	
Air Blank		0.000	01:44	
Control Test		0.080	01:44	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:45	

Data Download Date/Time: 07/16/2019 09:07

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMALL,BRANNON,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MOONEY,PATRICK
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 08:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 07/15/2019 09:47:16

Diagnostics Check		OK	09:40
Air Blank		0.000	09:41
Control Test		0.080	09:41
Air Blank		0.000	09:42
Subject Sample	# 1	0.142	09:42
Breath Volume		2.453	09:42
Air Blank		0.000	09:43
Air Blank		0.000	09:45
Subject Sample	# 2	0.144	09:45
Breath Volume		3.406	09:45
Air Blank		0.000	09:46
Control Test		0.078	09:46
Air Blank		0.000	09:47
Diagnostics Check		OK	09:47

Data Download Date/Time: 07/16/2019 09:18

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,LANDON,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WALKER,LANDON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 07/17/2019 18:00:46

Diagnostics Check		OK	17:56	
Air Blank		0.000	17:57	
Control Test		0.078	17:57	
Air Blank		0.000	17:58	
Subject Sample	# 1		17:59	Subject Test Refused
Breath Volume		0.250	17:59	
Air Blank		0.000	17:59	
Control Test		0.080	18:00	
Air Blank		0.000	18:00	
Diagnostics Check		OK	18:00	

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: JOHNSON,KARRY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 07/17/2019 21:11:45

Diagnostics Check		OK	21:08	
Air Blank		0.000	21:09	
Control Test		0.081	21:09	
Air Blank		0.000	21:10	
Subject Sample	# 1		21:10	Subject Test Refused
Breath Volume		0.000	21:10	
Air Blank		0.000	21:10	
Control Test		0.080	21:11	
Air Blank		0.000	21:11	
Diagnostics Check		OK	21:11	

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORTON,JAMES,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LINCOLN,AUSTIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 07/18/2019 02:14:22

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.080	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.000	02:09
Breath Volume		1.910	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.000	02:12
Breath Volume		1.351	02:12
Air Blank		0.000	02:13
Control Test		0.080	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CRAWFORD,BEJAMIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: LOVELACE

DATE/TIME OF TEST: 07/18/2019 06:11:57

Diagnostics Check		OK	06:01
Air Blank		0.000	06:02
Control Test		0.080	06:02
Air Blank		0.000	06:02
Subject Sample	# 1	0.098	06:05
Breath Volume		1.238	06:05
Air Blank		0.000	06:06
Air Blank		0.000	06:07
Subject Sample	# 2	0.096	06:10
Breath Volume		1.226	06:10
Air Blank		0.000	06:11
Control Test		0.080	06:11
Air Blank		0.000	06:11
Diagnostics Check		OK	06:11

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONARCH,BRIAN,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MONARCH,BRIAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: CISNEROS

DATE/TIME OF TEST: 07/18/2019 20:58:20

Diagnostics Check		OK	20:52
Air Blank		0.000	20:52
Control Test		0.081	20:53
Air Blank		0.000	20:53
Subject Sample	# 1	0.000	20:54
Breath Volume		2.851	20:54
Air Blank		0.000	20:54
Air Blank		0.000	20:56
Subject Sample	# 2	0.000	20:56
Breath Volume		3.753	20:56
Air Blank		0.000	20:57
Control Test		0.078	20:57
Air Blank		0.000	20:58
Diagnostics Check		OK	20:58

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPENCE,JUSTIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 07/19/2019 22:37:16

Diagnostics Check		OK	22:30
Air Blank		0.000	22:30
Control Test		0.081	22:30
Air Blank		0.000	22:31
Subject Sample	# 1	0.140	22:31
Breath Volume		1.406	22:31
Air Blank		0.000	22:32
Air Blank		0.000	22:34
Subject Sample	# 2	0.127	22:35
Breath Volume		1.250	22:35
Air Blank		0.000	22:36
Control Test		0.080	22:36
Air Blank		0.000	22:37
Diagnostics Check		OK	22:37

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOBLES,DARIUS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NOBLES,DARIUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: RENFRO

DATE/TIME OF TEST: 07/20/2019 05:03:57

Diagnostics Check		OK	04:56
Air Blank		0.000	04:57
Control Test		0.080	04:57
Air Blank		0.000	04:58
Subject Sample	# 1	0.138	04:59
Breath Volume		2.226	04:59
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.137	05:02
Breath Volume		1.992	05:02
Air Blank		0.000	05:03
Control Test		0.080	05:03
Air Blank		0.000	05:03
Diagnostics Check		OK	05:03

Data Download Date/Time: 08/08/2019 09:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILHELM,TONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GONSALVES

DATE/TIME OF TEST: 07/21/2019 02:58:38

Diagnostics Check		OK	02:55	
Air Blank		0.000	02:56	
Control Test		0.080	02:56	
Air Blank		0.000	02:57	
Subject Sample	# 1		02:57	Subject Test Refused
Breath Volume		0.000	02:57	
Air Blank		0.000	02:57	
Control Test		0.080	02:58	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONARCH,BRIAN,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MONARCH,BRIAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GASKIN

DATE/TIME OF TEST: 07/23/2019 01:24:33

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:20	
Control Test		0.080	01:21	
Air Blank		0.000	01:21	
Subject Sample	# 1		01:22	Subject Test Refused
Breath Volume		0.070	01:22	
Air Blank		0.000	01:23	
Control Test		0.081	01:23	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNHAM,J,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURNHAM,J
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: SHIRAH

DATE/TIME OF TEST: 07/23/2019 22:53:21

Diagnostics Check		OK	22:45
Air Blank		0.000	22:45
Control Test		0.083	22:46
Air Blank		0.000	22:46
Subject Sample	# 1	0.179	22:48
Breath Volume		1.316	22:48
Air Blank		0.000	22:49
Air Blank		0.000	22:51
Subject Sample	# 2	0.179	22:51
Breath Volume		1.714	22:51
Air Blank		0.000	22:52
Control Test		0.079	22:52
Air Blank		0.000	22:53
Diagnostics Check		OK	22:53

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,RYAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 13:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: PRETLOW TUCKER

DATE/TIME OF TEST: 07/24/2019 14:23:07

Diagnostics Check	OK	14:21	
Air Blank	0.000	14:22	
Control Test	0.088	14:22	Control Outside Tolerance
Air Blank	0.000	14:23	

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,RYAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 13:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: PRETLOW TUCKER

DATE/TIME OF TEST: 07/24/2019 14:28:50

Diagnostics Check		OK	14:26	
Air Blank		0.000	14:26	
Control Test		0.080	14:26	
Air Blank		0.000	14:27	
Subject Sample	# 1		14:27	Subject Test Refused
Breath Volume		0.000	14:27	
Air Blank		0.000	14:27	
Control Test		0.079	14:28	
Air Blank		0.000	14:28	
Diagnostics Check		OK	14:28	

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: GLANCEY

DATE/TIME OF TEST: 07/25/2019 00:36:26

Diagnostics Check		OK	00:33	
Air Blank		0.000	00:34	
Control Test		0.080	00:34	
Air Blank		0.000	00:34	
Subject Sample	# 1		00:35	Subject Test Refused
Breath Volume		0.000	00:35	
Air Blank		0.000	00:35	
Control Test		0.080	00:35	
Air Blank		0.000	00:36	
Diagnostics Check		OK	00:36	

Data Download Date/Time: 08/08/2019 09:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: GLANCEY

DATE/TIME OF TEST: 07/25/2019 00:54:23

Diagnostics Check		OK	00:45	
Air Blank		0.000	00:46	
Control Test		0.080	00:46	
Air Blank		0.000	00:47	
Subject Sample	# 1		00:50	No Sample Provided
Breath Volume		0.000	00:50	
Air Blank		0.000	00:50	
Air Blank		0.000	00:52	
Subject Sample	# 2		00:53	Subject Test Refused
Breath Volume		0.000	00:53	
Air Blank		0.000	00:53	
Control Test		0.080	00:53	
Air Blank		0.000	00:54	
Diagnostics Check		OK	00:54	

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KIMBRIL,JOSHUA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 06:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: LINTON

DATE/TIME OF TEST: 07/25/2019 07:02:56

Diagnostics Check		OK	06:56	
Air Blank		0.000	06:57	
Control Test		0.080	06:57	
Air Blank		0.000	06:57	
Subject Sample	# 1	0.149	06:58	
Breath Volume		2.285	06:58	
Air Blank		0.000	06:59	
Air Blank		0.000	07:00	
Subject Sample	# 2	0.151	07:01	
Breath Volume		2.105	07:01	
Air Blank		0.000	07:02	
Control Test		0.079	07:02	Interferent Detect
Air Blank		0.000	07:02	

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KIMBRIL,JOSHUA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 06:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: LINTON

DATE/TIME OF TEST: 07/25/2019 07:12:29

Diagnostics Check		OK	07:05
Air Blank		0.000	07:06
Control Test		0.080	07:06
Air Blank		0.000	07:06
Subject Sample	# 1	0.158	07:07
Breath Volume		3.039	07:07
Air Blank		0.000	07:08
Air Blank		0.000	07:10
Subject Sample	# 2	0.155	07:10
Breath Volume		2.070	07:10
Air Blank		0.000	07:11
Control Test		0.079	07:11
Air Blank		0.000	07:12
Diagnostics Check		OK	07:12

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGGAN,MARLON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOGGAN,MARLON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 15:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:04

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 07/25/2019 16:32:17

Diagnostics Check		OK	16:25
Air Blank		0.000	16:26
Control Test		0.080	16:26
Air Blank		0.000	16:27
Subject Sample	# 1	0.000	16:27
Breath Volume		3.460	16:27
Air Blank		0.000	16:28
Air Blank		0.000	16:30
Subject Sample	# 2	0.000	16:30
Breath Volume		3.273	16:30
Air Blank		0.000	16:31
Control Test		0.080	16:31
Air Blank		0.000	16:32
Diagnostics Check		OK	16:32

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEENEY,ERIC,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FEENEY,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 07/25/2019 22:22:19

Diagnostics Check		OK	22:15
Air Blank		0.000	22:16
Control Test		0.081	22:16
Air Blank		0.000	22:17
Subject Sample	# 1	0.237	22:17
Breath Volume		4.070	22:17
Air Blank		0.000	22:18
Air Blank		0.000	22:20
Subject Sample	# 2	0.223	22:20
Breath Volume		2.304	22:20
Air Blank		0.000	22:21
Control Test		0.079	22:21
Air Blank		0.000	22:22
Diagnostics Check		OK	22:22

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILDE,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: FORTHE

DATE/TIME OF TEST: 07/26/2019 01:25:35

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:21	
Control Test		0.080	01:22	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:24	Subject Test Refused
Breath Volume		0.000	01:24	
Air Blank		0.000	01:24	
Control Test		0.080	01:24	
Air Blank		0.000	01:25	
Diagnostics Check		OK	01:25	

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORTON,JAMES,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MORENO,WESTON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: LEWANDA

DATE/TIME OF TEST: 07/26/2019 21:25:35

Diagnostics Check		OK	21:19
Air Blank		0.000	21:20
Control Test		0.080	21:20
Air Blank		0.000	21:20
Subject Sample	# 1	0.000	21:21
Breath Volume		3.085	21:21
Air Blank		0.000	21:21
Air Blank		0.000	21:23
Subject Sample	# 2	0.000	21:24
Breath Volume		2.777	21:24
Air Blank		0.000	21:24
Control Test		0.081	21:24
Air Blank		0.000	21:25
Diagnostics Check		OK	21:25

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORTON,JAMES,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DUH,CASEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: HIGHSMITH

DATE/TIME OF TEST: 07/26/2019 22:09:05

Diagnostics Check		OK	21:53	
Air Blank		0.000	21:53	
Control Test		0.081	21:54	
Air Blank		0.000	21:54	
Subject Sample	# 1	0.146	21:57	
Breath Volume		2.332	21:57	
Air Blank		0.000	21:58	
Air Blank		0.000	22:00	
Subject Sample	# 2	0.128	22:03	Volume Not Met
Breath Volume		0.097	22:03	
Air Blank		0.000	22:04	
Air Blank		0.000	22:06	
Subject Sample	# 3	0.146	22:07	
Breath Volume		2.964	22:07	
Air Blank		0.000	22:08	
Control Test		0.078	22:08	
Air Blank		0.000	22:09	
Diagnostics Check		OK	22:09	

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNHAM,J,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURNHAM,J
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: CHE AKE

DATE/TIME OF TEST: 07/26/2019 22:35:16

Diagnostics Check		OK	22:28
Air Blank		0.000	22:29
Control Test		0.079	22:29
Air Blank		0.000	22:29
Subject Sample	# 1	0.233	22:30
Breath Volume		1.988	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.233	22:33
Breath Volume		1.851	22:33
Air Blank		0.000	22:34
Control Test		0.080	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORTON,JAMES,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: WEATHERSBY

DATE/TIME OF TEST: 07/27/2019 00:32:01

Diagnostics Check		OK	00:25
Air Blank		0.000	00:26
Control Test		0.080	00:26
Air Blank		0.000	00:26
Subject Sample	# 1	0.098	00:27
Breath Volume		1.292	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.084	00:30
Breath Volume		1.453	00:30
Air Blank		0.000	00:31
Control Test		0.080	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:32

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORTON,JAMES,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WEAVER-RASSA

DATE/TIME OF TEST: 07/27/2019 03:38:04

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.080	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.186	03:33
Breath Volume		2.589	03:33
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.182	03:36
Breath Volume		2.835	03:36
Air Blank		0.000	03:37
Control Test		0.079	03:37
Air Blank		0.000	03:37
Diagnostics Check		OK	03:38

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,RAY',C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAMILTON,RAY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: CAPELLAN PAREDES

DATE/TIME OF TEST: 07/27/2019 05:43:11

Diagnostics Check		OK	05:38	
Air Blank		0.000	05:39	
Control Test		0.079	05:39	
Air Blank		0.000	05:40	
Subject Sample	# 1		05:41	Subject Test Refused
Breath Volume		0.000	05:41	
Air Blank		0.000	05:42	
Control Test		0.080	05:42	
Air Blank		0.000	05:43	
Diagnostics Check		OK	05:43	

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARQUIS,NORMAND,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARQUIS,NORMAND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 16:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: HAJDUK

DATE/TIME OF TEST: 07/27/2019 17:11:51

Diagnostics Check		OK	17:04
Air Blank		0.000	17:05
Control Test		0.080	17:05
Air Blank		0.000	17:05
Subject Sample	# 1	0.288	17:06
Breath Volume		1.816	17:06
Air Blank		0.000	17:07
Air Blank		0.000	17:09
Subject Sample	# 2	0.290	17:10
Breath Volume		2.617	17:10
Air Blank		0.000	17:10
Control Test		0.080	17:11
Air Blank		0.000	17:11
Diagnostics Check		OK	17:11

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOBLES,DARIUS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NOBLES,DARIUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 07/30/2019 05:45:52

Diagnostics Check		OK	05:37	
Air Blank		0.000	05:37	
Control Test		0.081	05:38	
Air Blank		0.000	05:38	
Subject Sample	# 1		05:41	No Sample Provided
Breath Volume		0.000	05:41	
Air Blank		0.000	05:42	
Air Blank		0.000	05:44	
Subject Sample	# 2		05:44	Subject Test Refused
Breath Volume		0.000	05:44	
Air Blank		0.000	05:44	
Control Test		0.081	05:45	
Air Blank		0.000	05:45	
Diagnostics Check		OK	05:45	

Data Download Date/Time: 08/08/2019 09:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,ROYCE,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: DOWNEY

DATE/TIME OF TEST: 07/30/2019 19:58:11

Diagnostics Check		OK	19:51
Air Blank		0.000	19:51
Control Test		0.081	19:51
Air Blank		0.000	19:52
Subject Sample	# 1	0.106	19:53
Breath Volume		1.230	19:53
Air Blank		0.000	19:54
Air Blank		0.000	19:56
Subject Sample	# 2	0.112	19:56
Breath Volume		1.167	19:56
Air Blank		0.000	19:57
Control Test		0.080	19:57
Air Blank		0.000	19:58
Diagnostics Check		OK	19:58

Data Download Date/Time: 08/08/2019 09:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLL,MICHAEL,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WOLL,MICHAEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BRENNAN

DATE/TIME OF TEST: 07/05/2019 04:01:36

Diagnostics Check		OK	03:55
Air Blank		0.000	03:56
Control Test		0.080	03:56
Air Blank		0.000	03:56
Subject Sample	# 1	0.186	03:57
Breath Volume		3.558	03:57
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.182	04:00
Breath Volume		3.007	04:00
Air Blank		0.000	04:00
Control Test		0.085	04:01
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 07/26/2019 11:59

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,EDGARDO,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WOLL,MICHAEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: RECILLAS

DATE/TIME OF TEST: 07/13/2019 04:59:18

Diagnostics Check		OK	04:54	
Air Blank		0.000	04:55	
Control Test		0.080	04:55	
Air Blank		0.000	04:55	
Subject Sample	# 1		04:57	Subject Test Refused
Breath Volume		0.000	04:57	
Air Blank		0.000	04:58	
Control Test		0.081	04:58	
Air Blank		0.000	04:59	
Diagnostics Check		OK	04:59	

Data Download Date/Time: 07/26/2019 11:59

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,JOSHUA,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: TUCKER,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: CASTON

DATE/TIME OF TEST: 07/11/2019 21:25:23

Diagnostics Check		OK	21:18
Air Blank		0.000	21:18
Control Test		0.080	21:18
Air Blank		0.000	21:19
Subject Sample	# 1	0.166	21:20
Breath Volume		1.585	21:20
Air Blank		0.000	21:21
Air Blank		0.000	21:22
Subject Sample	# 2	0.162	21:23
Breath Volume		4.066	21:23
Air Blank		0.000	21:24
Control Test		0.080	21:24
Air Blank		0.000	21:25
Diagnostics Check		OK	21:25

Data Download Date/Time: 07/19/2019 23:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNCH,CHAD,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LYNCH,CHAD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: HASBROUER

DATE/TIME OF TEST: 07/12/2019 04:51:56

Diagnostics Check		OK	04:44
Air Blank		0.000	04:45
Control Test		0.079	04:45
Air Blank		0.000	04:46
Subject Sample	# 1	0.198	04:47
Breath Volume		2.980	04:47
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.179	04:50
Breath Volume		1.605	04:50
Air Blank		0.000	04:51
Control Test		0.080	04:51
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 07/19/2019 23:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGALADO,JORGE,I
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: REGALADO,JORGE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 07/13/2019 15:58:49

Diagnostics Check		OK	15:52
Air Blank		0.000	15:53
Control Test		0.079	15:53
Air Blank		0.000	15:54
Subject Sample	# 1	0.235	15:54
Breath Volume		3.093	15:54
Air Blank		0.000	15:55
Air Blank		0.000	15:56
Subject Sample	# 2	0.229	15:57
Breath Volume		2.429	15:57
Air Blank		0.000	15:57
Control Test		0.078	15:58
Air Blank		0.000	15:58
Diagnostics Check		OK	15:58

Data Download Date/Time: 07/19/2019 23:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEPBURN,ANDREW,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 12:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:06

SUBJECT'S LAST NAME: CARMICHAEL

DATE/TIME OF TEST: 07/16/2019 13:01:51

Diagnostics Check		OK	12:54
Air Blank		0.000	12:54
Control Test		0.080	12:55
Air Blank		0.000	12:55
Subject Sample	# 1	0.140	12:56
Breath Volume		1.257	12:56
Air Blank		0.000	12:57
Air Blank		0.000	12:58
Subject Sample	# 2	0.139	13:00
Breath Volume		1.695	13:00
Air Blank		0.000	13:00
Control Test		0.078	13:01
Air Blank		0.000	13:01
Diagnostics Check		OK	13:01

Data Download Date/Time: 07/19/2019 23:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA TRAINING
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY,WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: HENLEY

DATE/TIME OF TEST: 07/04/2019 15:38:51

Diagnostics Check		OK	15:29	
Air Blank		0.000	15:30	
Control Test		0.079	15:30	
Air Blank		0.000	15:31	
Subject Sample	# 1	0.000	15:34	Volume Not Met
Breath Volume		0.136	15:34	
Air Blank		0.000	15:34	
Air Blank		0.000	15:36	
Subject Sample	# 2		15:37	Slope Not Met
Breath Volume		1.007	15:37	
Air Blank		0.000	15:37	
Control Test		0.077	15:38	
Air Blank		0.000	15:38	
Diagnostics Check		OK	15:38	

Data Download Date/Time: 07/12/2019 11:00

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA TRAINING
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY,WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: HENLEY

DATE/TIME OF TEST: 07/04/2019 15:47:35

Diagnostics Check		OK	15:42	
Air Blank		0.000	15:43	
Control Test		0.080	15:43	
Air Blank		0.000	15:44	
Subject Sample	# 1		15:46	Slope Not Met
Breath Volume		1.328	15:46	
Air Blank		0.000	15:46	
Control Test		0.080	15:47	
Air Blank		0.000	15:47	
Diagnostics Check		OK	15:47	

Data Download Date/Time: 07/12/2019 10:43

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA TRAINING
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAMUEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: BLANK

DATE/TIME OF TEST: 07/04/2019 16:26:42

Diagnostics Check	OK	16:24	
Air Blank	0.000	16:25	
Control Test	0.079	16:25	
Air Blank	0.000	16:26	
		16:26	RFI Detect
Air Blank	0.000	16:26	

Data Download Date/Time: 07/12/2019 10:43

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA TRAINING
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: BLANK

DATE/TIME OF TEST: 07/04/2019 16:36:23

Diagnostics Check		OK	16:29	
Air Blank		0.000	16:29	
Control Test		0.080	16:30	
Air Blank		0.000	16:30	
Subject Sample	# 1	0.000	16:32	
Breath Volume		2.472	16:32	
Air Blank		0.000	16:32	
Air Blank		0.000	16:34	
Subject Sample	# 2		16:35	RFI Detect
Breath Volume		0.000	16:35	
Air Blank		0.000	16:35	
Control Test		0.079	16:35	
Air Blank		0.000	16:36	
Diagnostics Check		OK	16:36	

Data Download Date/Time: 07/12/2019 10:43

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUMPHREY,MATTHEW,D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: SCOTT,TRAVIS
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 183393
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: BUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: LIPFORD

DATE/TIME OF TEST: 07/04/2019 17:55:22

Diagnostics Check		OK	17:49
Air Blank		0.000	17:49
Control Test		0.082	17:49
Air Blank		0.000	17:50
Subject Sample	# 1	0.136	17:50
Breath Volume		3.734	17:50
Air Blank		0.000	17:51
Air Blank		0.000	17:53
Subject Sample	# 2	0.135	17:53
Breath Volume		3.437	17:53
Air Blank		0.000	17:54
Control Test		0.080	17:54
Air Blank		0.000	17:55
Diagnostics Check		OK	17:55

Data Download Date/Time: 07/26/2019 03:16

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONEY,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MALONEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 10:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: DONALD

DATE/TIME OF TEST: 07/27/2019 12:07:19

Diagnostics Check		OK	12:00
Air Blank		0.000	12:01
Control Test		0.081	12:01
Air Blank		0.000	12:02
Subject Sample	# 1	0.245	12:02
Breath Volume		1.347	12:02
Air Blank		0.000	12:03
Air Blank		0.000	12:05
Subject Sample	# 2	0.251	12:05
Breath Volume		1.617	12:05
Air Blank		0.000	12:06
Control Test		0.081	12:06
Air Blank		0.000	12:07
Diagnostics Check		OK	12:07

Data Download Date/Time: 08/05/2019 15:10

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLNAR,JARROD,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BARTLETT,SHELTON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: JARCZYK

DATE/TIME OF TEST: 07/01/2019 20:57:37

Diagnostics Check		OK	20:50
Air Blank		0.000	20:51
Control Test		0.079	20:51
Air Blank		0.000	20:52
Subject Sample	# 1	0.243	20:52
Breath Volume		1.777	20:52
Air Blank		0.000	20:53
Air Blank		0.000	20:55
Subject Sample	# 2	0.239	20:56
Breath Volume		1.960	20:56
Air Blank		0.000	20:56
Control Test		0.077	20:57
Air Blank		0.000	20:57
Diagnostics Check		OK	20:57

Data Download Date/Time: 07/02/2019 14:39

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,RAY,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAMILTON,RAY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: PRESTARRI

DATE/TIME OF TEST: 07/05/2019 05:46:17

Diagnostics Check		OK	05:42	
Air Blank		0.000	05:42	
Control Test		0.080	05:42	
Air Blank		0.000	05:43	
Subject Sample	# 1		05:44	Subject Test Refused
Breath Volume		0.000	05:44	
Air Blank		0.000	05:45	
Control Test		0.079	05:45	
Air Blank		0.000	05:46	
Diagnostics Check		OK	05:46	

Data Download Date/Time: 07/31/2019 09:40

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,DUSTAN,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RICE,DUSTAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 07/05/2019 21:23:57

Diagnostics Check		OK	21:20	
Air Blank		0.000	21:21	
Control Test		0.078	21:21	
Air Blank		0.000	21:22	
Subject Sample	# 1		21:22	Subject Test Refused
Breath Volume		0.000	21:22	
Air Blank		0.000	21:23	
Control Test		0.078	21:23	
Air Blank		0.000	21:23	
Diagnostics Check		OK	21:23	

Data Download Date/Time: 07/31/2019 09:40

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMATEER,TRACY,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VALENTINE,RONALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: HUME

DATE/TIME OF TEST: 07/09/2019 22:46:45

Diagnostics Check		OK	22:39
Air Blank		0.000	22:40
Control Test		0.080	22:40
Air Blank		0.000	22:40
Subject Sample	# 1	0.143	22:42
Breath Volume		2.832	22:42
Air Blank		0.000	22:43
Air Blank		0.000	22:44
Subject Sample	# 2	0.139	22:45
Breath Volume		1.199	22:45
Air Blank		0.000	22:45
Control Test		0.076	22:46
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 07/31/2019 09:40

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ANGIOLA,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: BRANNON

DATE/TIME OF TEST: 07/18/2019 17:10:56

Diagnostics Check		OK	17:03
Air Blank		0.000	17:03
Control Test		0.079	17:04
Air Blank		0.000	17:04
Subject Sample	# 1	0.212	17:06
Breath Volume		4.371	17:06
Air Blank		0.000	17:06
Air Blank		0.000	17:08
Subject Sample	# 2	0.227	17:09
Breath Volume		4.289	17:09
Air Blank		0.000	17:10
Control Test		0.079	17:10
Air Blank		0.000	17:10
Diagnostics Check		OK	17:10

Data Download Date/Time: 07/31/2019 09:40

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLINNIX,JERAD,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DUS,CHRISTINA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 07:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: DOUD

DATE/TIME OF TEST: 07/19/2019 08:31:14

Diagnostics Check		OK	08:19	
Air Blank		0.000	08:20	
Control Test		0.079	08:20	
Air Blank		0.000	08:21	
Subject Sample	# 1		08:24	No Sample Provided
Breath Volume		0.000	08:24	
Air Blank		0.000	08:24	
Air Blank		0.000	08:26	
Subject Sample	# 2		08:29	No Sample Provided
Breath Volume		0.000	08:29	
Air Blank		0.000	08:30	
Control Test		0.079	08:30	
Air Blank		0.000	08:31	
Diagnostics Check		OK	08:31	

Data Download Date/Time: 07/31/2019 09:40

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERTENS,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LEWIS,C
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 07/20/2019 05:06:10

Diagnostics Check		OK	04:58
Air Blank		0.000	04:59
Control Test		0.079	04:59
Air Blank		0.000	05:00
Subject Sample	# 1	0.251	05:01
Breath Volume		2.593	05:01
Air Blank		0.000	05:01
Air Blank		0.000	05:03
Subject Sample	# 2	0.239	05:04
Breath Volume		3.437	05:04
Air Blank		0.000	05:05
Control Test		0.079	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 07/31/2019 09:40

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLINNIX,JERAD,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ANGILLETTA,SALVATOR
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 08:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:53

SUBJECT'S LAST NAME: A436441882590

DATE/TIME OF TEST: 07/20/2019 09:24:06

Diagnostics Check		OK	09:17
Air Blank		0.000	09:18
Control Test		0.078	09:18
Air Blank		0.000	09:18
Subject Sample	# 1	0.000	09:19
Breath Volume		4.757	09:19
Air Blank		0.000	09:20
Air Blank		0.000	09:22
Subject Sample	# 2	0.000	09:22
Breath Volume		4.722	09:22
Air Blank		0.000	09:23
Control Test		0.078	09:23
Air Blank		0.000	09:24
Diagnostics Check		OK	09:24

Data Download Date/Time: 07/31/2019 09:40

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MELTON

DATE/TIME OF TEST: 07/24/2019 00:01:03

Diagnostics Check		OK	23:52	
Air Blank		0.000	23:52	
Control Test		0.080	23:53	
Air Blank		0.000	23:53	
Subject Sample	# 1		23:56	No Sample Provided
Breath Volume		0.000	23:56	
Air Blank		0.000	23:57	
Air Blank		0.000	23:59	
Subject Sample	# 2		23:59	Subject Test Refused
Breath Volume		0.000	23:59	
Air Blank		0.000	00:00	
Control Test		0.080	00:00	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:01	

Data Download Date/Time: 07/31/2019 09:40

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 07/24/2019 02:17:41

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.080	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.147	02:13
Breath Volume		2.804	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.136	02:16
Breath Volume		2.593	02:16
Air Blank		0.000	02:16
Control Test		0.078	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 07/31/2019 09:40

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEARS,TODD,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SEARS,TODD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: WATERSON

DATE/TIME OF TEST: 07/30/2019 00:36:09

Diagnostics Check		OK	00:28
Air Blank		0.000	00:29
Control Test		0.079	00:29
Air Blank		0.000	00:29
Subject Sample	# 1	0.348	00:31
Breath Volume		1.632	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:33
Subject Sample	# 2	0.355	00:34
Breath Volume		2.113	00:34
Air Blank		0.000	00:35
Control Test		0.078	00:35
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 07/31/2019 09:40

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIMMERMANN,JOHN,C
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WATSON,CLAYTON
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 07/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: YOTT

DATE/TIME OF TEST: 07/28/2019 04:20:37

Diagnostics Check		OK	04:17	
Air Blank		0.000	04:18	
Control Test		0.078	04:18	
Air Blank		0.000	04:19	
Subject Sample	# 1		04:19	Subject Test Refused
Breath Volume		0.000	04:19	
Air Blank		0.000	04:19	
Control Test		0.078	04:20	
Air Blank		0.000	04:20	
Diagnostics Check		OK	04:20	

Data Download Date/Time: 07/31/2019 23:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLIE,JOHN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TILLIE,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: HALE

DATE/TIME OF TEST: 07/29/2019 01:18:59

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:18	
Control Test	0.089	01:18	Control Outside Tolerance
Air Blank	0.000	01:18	

Data Download Date/Time: 07/31/2019 23:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLIE,JOHN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TILLIE,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: HALE

DATE/TIME OF TEST: 07/29/2019 01:29:40

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.078	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.211	01:24
Breath Volume		1.703	01:24
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.213	01:27
Breath Volume		1.894	01:27
Air Blank		0.000	01:28
Control Test		0.076	01:29
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 07/31/2019 23:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRIBB,BRIAN,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CRIBB,BRIAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 07:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 07/01/2019 08:56:00

Diagnostics Check		OK	08:50
Air Blank		0.000	08:50
Control Test		0.083	08:50
Air Blank		0.000	08:51
Subject Sample	# 1	0.000	08:51
Breath Volume		1.847	08:51
Air Blank		0.000	08:52
Air Blank		0.000	08:54
Subject Sample	# 2	0.000	08:54
Breath Volume		1.453	08:54
Air Blank		0.000	08:55
Control Test		0.083	08:55
Air Blank		0.000	08:55
Diagnostics Check		OK	08:56

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFELT,DYLAN,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: COFFELT,DYLAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: MADDOX

DATE/TIME OF TEST: 07/04/2019 20:05:17

Diagnostics Check		OK	19:59
Air Blank		0.000	19:59
Control Test		0.084	20:00
Air Blank		0.000	20:00
Subject Sample	# 1	0.000	20:01
Breath Volume		2.171	20:01
Air Blank		0.000	20:01
Air Blank		0.000	20:03
Subject Sample	# 2	0.000	20:03
Breath Volume		1.882	20:03
Air Blank		0.000	20:04
Control Test		0.084	20:04
Air Blank		0.000	20:05
Diagnostics Check		OK	20:05

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,LOUIE,S
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ROBERSON,LOUIE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: BUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: TOUCHSTONE

DATE/TIME OF TEST: 07/04/2019 23:40:30

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:35	
Control Test		0.083	23:36	
Air Blank		0.000	23:36	
Subject Sample	# 1		23:39	Subject Test Refused
Breath Volume		0.000	23:39	
Air Blank		0.000	23:39	
Control Test		0.083	23:39	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POSEY,BENJAMIN,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: POSEY,BENJAMIN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 07/06/2019 02:16:36

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:08	
Control Test		0.084	02:08	
Air Blank		0.000	02:09	
Subject Sample	# 1	0.035	02:12	Volume Not Met
Breath Volume		0.687	02:12	
Air Blank		0.000	02:12	
Air Blank		0.000	02:14	
Subject Sample	# 2		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Control Test		0.082	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: BRADSHAW

DATE/TIME OF TEST: 07/06/2019 04:55:48

Diagnostics Check		OK	04:53	
Air Blank		0.000	04:53	
Control Test		0.084	04:53	
Air Blank		0.000	04:54	
Subject Sample	# 1		04:54	Subject Test Refused
Breath Volume		0.000	04:54	
Air Blank		0.000	04:54	
Control Test		0.084	04:55	
Air Blank		0.000	04:55	
Diagnostics Check		OK	04:55	

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUCE,HEATHER,N
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BRUCE,HEATHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: BABB

DATE/TIME OF TEST: 07/07/2019 00:53:18

Diagnostics Check		OK	00:46
Air Blank		0.000	00:47
Control Test		0.083	00:47
Air Blank		0.000	00:47
Subject Sample	# 1	0.187	00:48
Breath Volume		1.632	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:50
Subject Sample	# 2	0.170	00:51
Breath Volume		2.105	00:51
Air Blank		0.000	00:52
Control Test		0.082	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVOTH,DANIEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARVOTH,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: WILLHAM

DATE/TIME OF TEST: 07/07/2019 02:20:49

Diagnostics Check		OK	02:12	
Air Blank		0.000	02:12	
Control Test		0.083	02:13	
Air Blank		0.000	02:13	
Subject Sample	# 1		02:16	No Sample Provided
Breath Volume		0.000	02:16	
Air Blank		0.000	02:17	
Air Blank		0.000	02:19	
Subject Sample	# 2		02:19	Subject Test Refused
Breath Volume		0.000	02:19	
Air Blank		0.000	02:19	
Control Test		0.083	02:20	
Air Blank		0.000	02:20	
Diagnostics Check		OK	02:20	

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWTON,CARSON,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HOWTON,CARSON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: MCCRANIE

DATE/TIME OF TEST: 07/10/2019 19:42:29

Diagnostics Check		OK	19:34	
Air Blank		0.000	19:34	
Control Test		0.083	19:34	
Air Blank		0.000	19:35	
Subject Sample	# 1		19:38	No Sample Provided
Breath Volume		0.000	19:38	
Air Blank		0.000	19:38	
Air Blank		0.000	19:40	
Subject Sample	# 2		19:41	Subject Test Refused
Breath Volume		0.000	19:41	
Air Blank		0.000	19:41	
Control Test		0.083	19:41	
Air Blank		0.000	19:42	
Diagnostics Check		OK	19:42	

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOHNER,KYLER,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOHNER,KYLER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: ONEAL

DATE/TIME OF TEST: 07/11/2019 22:29:13

Diagnostics Check		OK	22:22
Air Blank		0.000	22:23
Control Test		0.082	22:23
Air Blank		0.000	22:24
Subject Sample	# 1	0.144	22:24
Breath Volume		4.648	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.147	22:27
Breath Volume		4.742	22:27
Air Blank		0.000	22:28
Control Test		0.082	22:28
Air Blank		0.000	22:29
Diagnostics Check		OK	22:29

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: PETTIGREW

DATE/TIME OF TEST: 07/14/2019 06:02:42

Diagnostics Check		OK	05:58	
Air Blank		0.000	05:59	
Control Test		0.083	05:59	
Air Blank		0.000	06:00	
Subject Sample	# 1		06:01	Subject Test Refused
Breath Volume		0.000	06:01	
Air Blank		0.000	06:01	
Control Test		0.084	06:02	
Air Blank		0.000	06:02	
Diagnostics Check		OK	06:02	

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOGGETTE,MASON,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DOGGETTE,MASON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/14/2019 21:49:35

Diagnostics Check		OK	21:43
Air Blank		0.000	21:43
Control Test		0.083	21:44
Air Blank		0.000	21:44
Subject Sample	# 1	0.189	21:45
Breath Volume		1.792	21:45
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.173	21:48
Breath Volume		1.199	21:48
Air Blank		0.000	21:48
Control Test		0.083	21:48
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DANNY,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: AVERY,BAYLEE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 07/15/2019 22:18:45

Diagnostics Check		OK	22:12
Air Blank		0.000	22:13
Control Test		0.084	22:13
Air Blank		0.000	22:13
Subject Sample	# 1	0.000	22:14
Breath Volume		1.660	22:14
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.000	22:17
Breath Volume		2.281	22:17
Air Blank		0.000	22:17
Control Test		0.084	22:18
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,DAVID,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DANIELS,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: DUEITT

DATE/TIME OF TEST: 07/17/2019 01:24:00

Diagnostics Check		OK	01:17
Air Blank		0.000	01:18
Control Test		0.083	01:18
Air Blank		0.000	01:19
Subject Sample	# 1	0.171	01:19
Breath Volume		2.949	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.187	01:22
Breath Volume		3.667	01:22
Air Blank		0.000	01:23
Control Test		0.082	01:23
Air Blank		0.000	01:23
Diagnostics Check		OK	01:24

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,COREY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WEAVER,COREY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: IRWIN

DATE/TIME OF TEST: 07/20/2019 01:38:08

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.084	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.202	01:33
Breath Volume		3.910	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:35
Subject Sample	# 2	0.188	01:36
Breath Volume		1.187	01:36
Air Blank		0.000	01:37
Control Test		0.081	01:37
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDEN,RYAN,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GOLDEN,RYAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 07/23/2019 23:48:43

Diagnostics Check		OK	23:41
Air Blank		0.000	23:42
Control Test		0.083	23:42
Air Blank		0.000	23:43
Subject Sample	# 1	0.168	23:44
Breath Volume		2.703	23:44
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.170	23:47
Breath Volume		2.875	23:47
Air Blank		0.000	23:47
Control Test		0.082	23:48
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,JOSHUA,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: TUCKER,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 12:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: PATRICK

DATE/TIME OF TEST: 07/24/2019 01:59:05

Diagnostics Check		OK	01:51
Air Blank		0.000	01:52
Control Test		0.083	01:52
Air Blank		0.000	01:52
Subject Sample	# 1	0.000	01:53
Breath Volume		2.929	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.000	01:57
Breath Volume		2.789	01:57
Air Blank		0.000	01:58
Control Test		0.083	01:58
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: WHIDDON

DATE/TIME OF TEST: 07/27/2019 21:35:53

Diagnostics Check		OK	21:29
Air Blank		0.000	21:29
Control Test		0.082	21:30
Air Blank		0.000	21:30
Subject Sample	# 1	0.233	21:31
Breath Volume		3.960	21:31
Air Blank		0.000	21:32
Air Blank		0.000	21:33
Subject Sample	# 2	0.239	21:34
Breath Volume		3.691	21:34
Air Blank		0.000	21:34
Control Test		0.081	21:35
Air Blank		0.000	21:35
Diagnostics Check		OK	21:35

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: FARLEY

DATE/TIME OF TEST: 07/28/2019 01:47:42

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.083	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.166	01:42
Breath Volume		3.875	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.172	01:46
Breath Volume		4.308	01:46
Air Blank		0.000	01:46
Control Test		0.082	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERHUNE,JEREMY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TERHUNE,JEREMY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 09:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:46

SUBJECT'S LAST NAME: HOLLEY

DATE/TIME OF TEST: 07/28/2019 10:22:37

Diagnostics Check		OK	10:16
Air Blank		0.000	10:16
Control Test		0.084	10:16
Air Blank		0.000	10:17
Subject Sample	# 1	0.000	10:18
Breath Volume		2.433	10:18
Air Blank		0.000	10:18
Air Blank		0.000	10:20
Subject Sample	# 2	0.000	10:21
Breath Volume		1.902	10:21
Air Blank		0.000	10:21
Control Test		0.083	10:22
Air Blank		0.000	10:22
Diagnostics Check		OK	10:22

Data Download Date/Time: 07/30/2019 21:30

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOHNER,KYLER,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOHNER,KYLER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: HOVERMALE

DATE/TIME OF TEST: 07/31/2019 22:35:10

Diagnostics Check		OK	22:26	
Air Blank		0.000	22:27	
Control Test		0.083	22:27	
Air Blank		0.000	22:27	
Subject Sample	# 1		22:31	No Sample Provided
Breath Volume		0.000	22:31	
Air Blank		0.000	22:31	
Air Blank		0.000	22:33	
Subject Sample	# 2		22:33	Subject Test Refused
Breath Volume		0.000	22:33	
Air Blank		0.000	22:34	
Control Test		0.084	22:34	
Air Blank		0.000	22:35	
Diagnostics Check		OK	22:35	

Data Download Date/Time: 08/23/2019 19:26

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAS,ALFRED,A
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: DEAS,ALFRED
ARRESTING OFFICER AGENCY: Lafayette County SO
LAST AGENCY INSPECTION: 07/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: CUBBAGE

DATE/TIME OF TEST: 07/07/2019 23:03:42

Diagnostics Check		OK	22:57
Air Blank		0.000	22:57
Control Test		0.078	22:57
Air Blank		0.000	22:58
Subject Sample	# 1	0.180	22:58
Breath Volume		1.535	22:58
Air Blank		0.000	22:59
Air Blank		0.000	23:01
Subject Sample	# 2	0.178	23:02
Breath Volume		1.894	23:02
Air Blank		0.000	23:02
Control Test		0.077	23:03
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 09/04/2019 17:09

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MONTEMURNO,MATT
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: MILEWSKI

DATE/TIME OF TEST: 07/01/2019 22:46:32

Diagnostics Check		OK	22:40
Air Blank		0.000	22:40
Control Test		0.079	22:41
Air Blank		0.000	22:41
Subject Sample	# 1	0.232	22:42
Breath Volume		3.019	22:42
Air Blank		0.000	22:42
Air Blank		0.000	22:44
Subject Sample	# 2	0.236	22:45
Breath Volume		2.875	22:45
Air Blank		0.000	22:45
Control Test		0.079	22:45
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 07/18/2019 06:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/04/2019 01:02:45

Diagnostics Check		OK	00:59	
Air Blank		0.000	01:00	
Control Test		0.080	01:00	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:01	Subject Test Refused
Breath Volume		0.000	01:01	
Air Blank		0.000	01:01	
Control Test		0.080	01:02	
Air Blank		0.000	01:02	
Diagnostics Check		OK	01:02	

Data Download Date/Time: 07/18/2019 06:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MCFARLAND

DATE/TIME OF TEST: 07/04/2019 22:41:59

Diagnostics Check		OK	22:35
Air Blank		0.000	22:36
Control Test		0.079	22:36
Air Blank		0.000	22:37
Subject Sample	# 1	0.311	22:37
Breath Volume		3.277	22:37
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.302	22:40
Breath Volume		3.378	22:40
Air Blank		0.000	22:41
Control Test		0.079	22:41
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 07/18/2019 06:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CALHOUN

DATE/TIME OF TEST: 07/05/2019 22:47:36

Diagnostics Check		OK	22:40
Air Blank		0.000	22:41
Control Test		0.080	22:41
Air Blank		0.000	22:42
Subject Sample	# 1	0.121	22:42
Breath Volume		1.398	22:42
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 2	0.120	22:46
Breath Volume		1.628	22:46
Air Blank		0.000	22:46
Control Test		0.080	22:47
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 07/18/2019 06:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: CREWS,MICHAEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/06/2019 01:42:05

Diagnostics Check		OK	01:35	
Air Blank		0.000	01:36	
Control Test		0.079	01:36	
Air Blank		0.000	01:37	
Subject Sample	# 1	0.063	01:37	
Breath Volume		2.660	01:37	
Air Blank		0.000	01:38	
Air Blank		0.000	01:40	
Subject Sample	# 2		01:41	Subject Test Refused
Breath Volume		0.000	01:41	
Air Blank		0.000	01:41	
Control Test		0.079	01:41	
Air Blank			01:41	RFI Detect
Air Blank			01:42	RFI Detect

Data Download Date/Time: 07/18/2019 06:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY,MATHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY,MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 07/06/2019 03:00:52

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.080	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.142	02:56
Breath Volume		1.175	02:56
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.146	02:59
Breath Volume		1.824	02:59
Air Blank		0.000	02:59
Control Test		0.080	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 07/18/2019 06:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: GEIGER

DATE/TIME OF TEST: 07/06/2019 03:35:14

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:30
Subject Sample	# 1	0.148	03:30
Breath Volume		2.464	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.147	03:33
Breath Volume		1.933	03:33
Air Blank		0.000	03:34
Control Test		0.078	03:34
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 07/18/2019 06:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: CREWS,MICHAEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: ARTEAGA

DATE/TIME OF TEST: 07/07/2019 03:02:09

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.167	02:57
Breath Volume		1.746	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.180	03:00
Breath Volume		1.578	03:00
Air Blank		0.000	03:01
Control Test		0.079	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 07/18/2019 06:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: NASH,JAMES
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: ERIKSON

DATE/TIME OF TEST: 07/07/2019 05:40:40

Diagnostics Check		OK	05:34
Air Blank		0.000	05:34
Control Test		0.080	05:35
Air Blank		0.000	05:35
Subject Sample	# 1	0.188	05:36
Breath Volume		3.515	05:36
Air Blank		0.000	05:36
Air Blank		0.000	05:38
Subject Sample	# 2	0.186	05:39
Breath Volume		2.601	05:39
Air Blank		0.000	05:39
Control Test		0.080	05:40
Air Blank		0.000	05:40
Diagnostics Check		OK	05:40

Data Download Date/Time: 07/18/2019 06:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MONTEMURNO,MATT
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: GERHARDT

DATE/TIME OF TEST: 07/12/2019 03:10:49

Diagnostics Check		OK	03:04
Air Blank		0.000	03:05
Control Test		0.080	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.127	03:06
Breath Volume		2.511	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:08
Subject Sample	# 2	0.122	03:09
Breath Volume		2.914	03:09
Air Blank		0.000	03:09
Control Test		0.079	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 07/18/2019 06:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: WYCHE

DATE/TIME OF TEST: 07/13/2019 01:53:37

Diagnostics Check		OK	01:47
Air Blank		0.000	01:48
Control Test		0.081	01:48
Air Blank		0.000	01:48
Subject Sample	# 1	0.149	01:49
Breath Volume		1.132	01:49
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.143	01:52
Breath Volume		1.714	01:52
Air Blank		0.000	01:52
Control Test		0.079	01:53
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 07/18/2019 06:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,STUART,P
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: MILLIGAN,CONNIE
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: LESLIE

DATE/TIME OF TEST: 07/14/2019 03:28:58

Diagnostics Check		OK	03:26	
Air Blank		0.000	03:26	
Control Test		0.080	03:26	
Air Blank		0.000	03:27	
Subject Sample	# 1		03:27	Subject Test Refused
Breath Volume		0.000	03:27	
Air Blank		0.000	03:28	
Control Test		0.080	03:28	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 07/18/2019 06:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAY,DARREL,E
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MONTEMURNO,MATTHEW
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: REGISTER

DATE/TIME OF TEST: 07/17/2019 01:09:38

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:03
Subject Sample	# 1	0.213	01:04
Breath Volume		1.816	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.210	01:08
Breath Volume		1.808	01:08
Air Blank		0.000	01:08
Control Test		0.076	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 07/18/2019 06:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: CREWS,MICHAEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GARY

DATE/TIME OF TEST: 07/17/2019 03:24:30

Diagnostics Check		OK	03:15	
Air Blank		0.000	03:15	
Control Test		0.080	03:16	
Air Blank		0.000	03:16	
Subject Sample	# 1	0.175	03:19	Volume Not Met
Breath Volume		0.195	03:19	
Air Blank		0.000	03:20	
Air Blank		0.000	03:22	
Subject Sample	# 2		03:22	Subject Test Refused
Breath Volume		0.132	03:22	
Air Blank		0.000	03:23	
Control Test		0.079	03:23	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:24	

Data Download Date/Time: 07/18/2019 06:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: CREWS,MICHAEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: ADMIN
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: C

DATE/TIME OF TEST: 07/17/2019 03:31:18

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:26	
Control Test		0.079	03:26	
Air Blank		0.000	03:27	
Subject Sample	# 1	0.000	03:27	
Breath Volume		2.003	03:27	
Air Blank		0.000	03:27	
Air Blank		0.000	03:29	
Subject Sample	# 2		03:29	Subject Test Refused
Breath Volume		0.000	03:29	
Air Blank		0.000	03:30	
Control Test		0.079	03:30	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 07/18/2019 06:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: CREWS,MICHAEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GARY

DATE/TIME OF TEST: 07/17/2019 03:39:34

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:33	
Control Test		0.079	03:33	
Air Blank		0.000	03:34	
Subject Sample	# 1	0.183	03:34	
Breath Volume		1.179	03:34	
Air Blank		0.000	03:35	
Air Blank		0.000	03:37	
Subject Sample	# 2		03:38	Subject Test Refused
Breath Volume		0.136	03:38	
Air Blank		0.000	03:38	
Control Test		0.079	03:38	
Air Blank		0.000	03:39	
Diagnostics Check		OK	03:39	

Data Download Date/Time: 07/18/2019 06:22

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: CREWS,MICHAEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 07/20/2019 02:54:11

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.078	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.086	02:49
Breath Volume		1.925	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.085	02:52
Breath Volume		1.792	02:52
Air Blank		0.000	02:53
Control Test		0.079	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 08/22/2019 17:19

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MONTEMURNO,MATT
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 07/20/2019 04:03:00

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.080	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.147	03:58
Breath Volume		4.519	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.147	04:01
Breath Volume		4.312	04:01
Air Blank		0.000	04:02
Control Test		0.078	04:02
Air Blank		0.000	04:02
Diagnostics Check		OK	04:03

Data Download Date/Time: 08/22/2019 17:19

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: MCNEILL

DATE/TIME OF TEST: 07/20/2019 20:51:05

Diagnostics Check	OK	20:50	
Air Blank		20:50	RFI Detect
Air Blank	0.000	20:51	

Data Download Date/Time: 08/22/2019 17:19

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: MCNEILL

DATE/TIME OF TEST: 07/20/2019 20:52:48

Diagnostics Check	OK	20:52	
Air Blank		20:52	RFI Detect
Air Blank		20:52	RFI Detect

Data Download Date/Time: 08/22/2019 17:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: MCNEILL

DATE/TIME OF TEST: 07/20/2019 21:03:05

Diagnostics Check		OK	20:56
Air Blank		0.000	20:57
Control Test		0.079	20:57
Air Blank		0.000	20:58
Subject Sample	# 1	0.028	20:58
Breath Volume		3.019	20:58
Air Blank		0.000	20:59
Air Blank		0.000	21:01
Subject Sample	# 2	0.027	21:01
Breath Volume		3.558	21:01
Air Blank		0.000	21:02
Control Test		0.078	21:02
Air Blank		0.000	21:03
Diagnostics Check		OK	21:03

Data Download Date/Time: 08/22/2019 17:19

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDREE,TASHAHNDA,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MONTEMURNO,MATTHEW
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: HARVIN

DATE/TIME OF TEST: 07/21/2019 02:47:02

Diagnostics Check		OK	02:43	
Air Blank		0.000	02:43	
Control Test		0.080	02:43	
Air Blank		0.000	02:44	
Subject Sample	# 1		02:45	Subject Test Refused
Breath Volume		0.000	02:45	
Air Blank		0.000	02:46	
Control Test		0.080	02:46	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:47	

Data Download Date/Time: 08/22/2019 17:19

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDREE,TASHAHNDA,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALDREE,TASHAHNDA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/21/2019 21:52:11

Diagnostics Check		OK	21:45
Air Blank		0.000	21:46
Control Test		0.079	21:46
Air Blank		0.000	21:47
Subject Sample	# 1	0.192	21:47
Breath Volume		1.976	21:47
Air Blank		0.000	21:48
Air Blank		0.000	21:50
Subject Sample	# 2	0.191	21:50
Breath Volume		1.718	21:50
Air Blank		0.000	21:51
Control Test		0.079	21:51
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 08/22/2019 17:19

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MONTEMURNO,MATT
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: KITCHINGS

DATE/TIME OF TEST: 07/26/2019 02:32:42

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.081	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.000	02:28
Breath Volume		2.316	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.000	02:31
Breath Volume		1.445	02:31
Air Blank		0.000	02:31
Control Test		0.080	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 08/22/2019 17:19

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: MAHONE

DATE/TIME OF TEST: 07/27/2019 18:59:47

Diagnostics Check		OK	18:51	
Air Blank		0.000	18:51	
Control Test		0.079	18:51	
Air Blank		0.000	18:52	
Subject Sample	# 1	0.000	18:55	Volume Not Met
Breath Volume		0.000	18:55	
Air Blank		0.000	18:56	
Air Blank		0.000	18:58	
Subject Sample	# 2		18:58	Subject Test Refused
Breath Volume		0.000	18:58	
Air Blank		0.000	18:58	
Control Test		0.079	18:59	
Air Blank		0.000	18:59	
Diagnostics Check		OK	18:59	

Data Download Date/Time: 08/22/2019 17:19

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,STUART,P
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 07/28/2019 02:42:28

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.080	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.233	02:35
Breath Volume		3.234	02:35
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.194	02:37
Breath Volume		1.781	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 3	0.227	02:40
Breath Volume		3.406	02:40
Air Blank		0.000	02:41
Control Test		0.077	02:41
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 08/22/2019 17:19

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MARTELL,NICOLE
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 07/30/2019 00:48:46

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.080	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.340	00:44
Breath Volume		1.949	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:46
Subject Sample	# 2	0.326	00:47
Breath Volume		1.968	00:47
Air Blank		0.000	00:47
Control Test		0.080	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 08/22/2019 17:19

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MONTEMURNO,MATT
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: HOOEY

DATE/TIME OF TEST: 07/31/2019 03:05:09

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.079	02:58
Air Blank		0.000	02:58
Subject Sample	# 1	0.270	02:59
Breath Volume		2.175	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.253	03:03
Breath Volume		1.734	03:03
Air Blank		0.000	03:04
Control Test		0.080	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 08/22/2019 17:19

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISCHER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FISCHER,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 07/03/2019 22:55:57

Diagnostics Check		OK	22:53	
Air Blank		0.000	22:53	
Control Test		0.082	22:53	
Air Blank		0.000	22:54	
Subject Sample	# 1		22:54	Subject Test Refused
Breath Volume		0.000	22:54	
Air Blank		0.000	22:55	
Control Test		0.082	22:55	
Air Blank		0.000	22:55	
Diagnostics Check		OK	22:55	

Data Download Date/Time: 07/24/2019 11:45

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORE,MATTHEW,A
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HOMAN,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: BUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: SHIRLEY

DATE/TIME OF TEST: 07/04/2019 19:06:10

Diagnostics Check		OK	18:59
Air Blank		0.000	18:59
Control Test		0.080	19:00
Air Blank		0.000	19:00
Subject Sample	# 1	0.136	19:01
Breath Volume		1.859	19:01
Air Blank		0.000	19:02
Air Blank		0.000	19:04
Subject Sample	# 2	0.139	19:04
Breath Volume		1.683	19:04
Air Blank		0.000	19:05
Control Test		0.079	19:05
Air Blank		0.000	19:06
Diagnostics Check		OK	19:06

Data Download Date/Time: 07/26/2019 03:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LERNER,MATTHEW,H
BREATH TEST OPERATOR AGENCY: Washington County SO
ARRESTING OFFICER: LERNER,MATTHEW
ARRESTING OFFICER AGENCY: Washington County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BOSSERT

DATE/TIME OF TEST: 07/06/2019 23:27:01

Diagnostics Check		OK	23:15	
Air Blank		0.000	23:16	
Control Test		0.080	23:16	
Air Blank		0.000	23:16	
Subject Sample	# 1	0.155	23:20	Volume Not Met
Breath Volume		0.945	23:20	
Air Blank		0.000	23:20	
Air Blank		0.000	23:22	
Subject Sample	# 2		23:25	No Sample Provided
Breath Volume		0.000	23:25	
Air Blank		0.000	23:26	
Control Test		0.078	23:26	
Air Blank		0.000	23:26	
Diagnostics Check		OK	23:27	

Data Download Date/Time: 07/26/2019 03:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONEY,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MALONEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: LOCKE

DATE/TIME OF TEST: 07/26/2019 18:56:21

Diagnostics Check	OK	18:55	
Air Blank		18:55	RFI Detect
Air Blank	0.000	18:56	

Data Download Date/Time: 08/05/2019 14:38

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONEY,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MALONEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: LOCKE

DATE/TIME OF TEST: 07/26/2019 19:07:35

Diagnostics Check		OK	19:00
Air Blank		0.000	19:00
Control Test		0.079	19:01
Air Blank		0.000	19:01
Subject Sample	# 1	0.185	19:02
Breath Volume		2.605	19:02
Air Blank		0.000	19:03
Air Blank		0.000	19:04
Subject Sample	# 2	0.183	19:06
Breath Volume		3.671	19:06
Air Blank		0.000	19:06
Control Test		0.079	19:06
Air Blank		0.000	19:07
Diagnostics Check		OK	19:07

Data Download Date/Time: 08/05/2019 14:38

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 07/05/2019 21:23:50

Diagnostics Check		OK	21:20	
Air Blank		0.000	21:20	
Control Test		0.082	21:21	
Air Blank		0.000	21:21	
Subject Sample	# 1		21:22	Subject Test Refused
Breath Volume		0.000	21:22	
Air Blank		0.000	21:22	
Control Test		0.082	21:23	
Air Blank		0.000	21:23	
Diagnostics Check		OK	21:23	

Data Download Date/Time: 07/26/2019 03:23

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: EMUNDSON

DATE/TIME OF TEST: 07/06/2019 04:45:21

Diagnostics Check		OK	04:42	
Air Blank		0.000	04:42	
Control Test		0.082	04:43	
Air Blank		0.000	04:43	
Subject Sample	# 1		04:43	Subject Test Refused
Breath Volume		0.000	04:43	
Air Blank		0.000	04:44	
Control Test		0.081	04:44	
Air Blank		0.000	04:45	
Diagnostics Check		OK	04:45	

Data Download Date/Time: 07/26/2019 03:23

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: KELLER

DATE/TIME OF TEST: 07/23/2019 23:05:21

Diagnostics Check		OK	22:56	
Air Blank		0.000	22:57	
Control Test		0.082	22:57	
Air Blank		0.000	22:58	
Subject Sample	# 1		23:01	No Sample Provided
Breath Volume		0.000	23:01	
Air Blank		0.000	23:01	
Air Blank		0.000	23:03	
Subject Sample	# 2		23:03	Subject Test Refused
Breath Volume		0.000	23:03	
Air Blank		0.000	23:04	
Control Test		0.082	23:04	
Air Blank		0.000	23:05	
Diagnostics Check		OK	23:05	

Data Download Date/Time: 07/26/2019 03:23

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: WEEKS

DATE/TIME OF TEST: 07/27/2019 21:28:53

Diagnostics Check		OK	21:21
Air Blank		0.000	21:22
Control Test		0.083	21:22
Air Blank		0.000	21:23
Subject Sample	# 1	0.145	21:24
Breath Volume		1.847	21:24
Air Blank		0.000	21:24
Air Blank		0.000	21:26
Subject Sample	# 2	0.149	21:27
Breath Volume		1.179	21:27
Air Blank		0.000	21:27
Control Test		0.080	21:28
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 08/05/2019 15:33

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,WILFORD,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HURST,CARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 14:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: BLAKNEY

DATE/TIME OF TEST: 07/05/2019 15:24:06

Diagnostics Check		OK	15:21	
Air Blank		0.000	15:21	
Control Test		0.081	15:22	
Air Blank		0.000	15:22	
Subject Sample	# 1		15:22	Subject Test Refused
Breath Volume		0.000	15:22	
Air Blank		0.000	15:23	
Control Test		0.080	15:23	
Air Blank		0.000	15:24	
Diagnostics Check		OK	15:24	

Data Download Date/Time: 07/26/2019 02:40

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEFF,STEPHEN,
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: CANILLAS

DATE/TIME OF TEST: 07/04/2019 20:38:58

Diagnostics Check		OK	20:33	
Air Blank		0.000	20:33	
Control Test		0.081	20:33	
Air Blank		0.000	20:34	
Subject Sample	# 1	0.164	20:37	Volume Not Met
Breath Volume		0.238	20:37	
Air Blank			20:38	Purge Fail
Air Blank		0.000	20:38	

Data Download Date/Time: 07/14/2019 10:24

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEFF,STEPHEN,
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: CANILLAS

DATE/TIME OF TEST: 07/04/2019 20:46:24

Diagnostics Check		OK	20:40	
Air Blank		0.000	20:41	
Control Test		0.080	20:41	
Air Blank		0.000	20:42	
Subject Sample	# 1	0.171	20:44	
Breath Volume		2.171	20:44	
Air Blank			20:45	Purge Fail
Air Blank		0.000	20:46	

Data Download Date/Time: 07/14/2019 10:24

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEFF,STEPHEN,
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: CANILLAS

DATE/TIME OF TEST: 07/04/2019 20:58:37

Diagnostics Check		OK	20:48	
Air Blank		0.000	20:49	
Control Test		0.080	20:49	
Air Blank		0.000	20:49	
Subject Sample	# 1	0.177	20:50	
Breath Volume		1.683	20:50	
Air Blank		0.000	20:51	
Air Blank		0.000	20:53	
Subject Sample	# 2	0.153	20:54	
Breath Volume		1.343	20:54	
Air Blank		0.000	20:55	
Air Blank			20:57	RFI Detect
Air Blank		0.000	20:57	
Control Test		0.080	20:58	
Air Blank		0.000	20:58	
Diagnostics Check		OK	20:58	

Data Download Date/Time: 07/14/2019 10:24

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEFF,STEPHEN,
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: CANILLAS

DATE/TIME OF TEST: 07/04/2019 21:05:47

Diagnostics Check		OK	21:00	
Air Blank		0.000	21:00	
Control Test		0.080	21:01	
Air Blank		0.000	21:01	
Subject Sample	# 1	0.160	21:04	
Breath Volume		1.472	21:04	
Air Blank			21:05	Purge Fail
Air Blank		0.000	21:05	

Data Download Date/Time: 07/14/2019 10:24

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: WELTZ

DATE/TIME OF TEST: 07/13/2019 03:18:42

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:13
Subject Sample	# 1	0.145	03:14
Breath Volume		1.636	03:14
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.141	03:17
Breath Volume		1.867	03:17
Air Blank		0.000	03:17
Control Test		0.080	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 07/14/2019 10:24

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUZBEE,HEATHER,N
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: CHECHAK,LANA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 07/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 07/18/2019 02:43:09

Diagnostics Check		OK	02:34
Air Blank		0.000	02:34
Control Test		0.081	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.153	02:36
Breath Volume		1.753	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.147	02:41
Breath Volume		1.414	02:41
Air Blank		0.000	02:42
Control Test		0.083	02:42
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 08/15/2019 11:42

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: CRUMBLEY

DATE/TIME OF TEST: 07/19/2019 04:39:36

Diagnostics Check		OK	04:33
Air Blank		0.000	04:33
Control Test		0.081	04:34
Air Blank		0.000	04:34
Subject Sample	# 1	0.138	04:35
Breath Volume		1.750	04:35
Air Blank		0.000	04:35
Air Blank		0.000	04:37
Subject Sample	# 2	0.142	04:38
Breath Volume		2.105	04:38
Air Blank		0.000	04:38
Control Test		0.082	04:39
Air Blank		0.000	04:39
Diagnostics Check		OK	04:39

Data Download Date/Time: 08/15/2019 11:42

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDRY,KATHERINE,E
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LANDRY,KATHERINE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 07/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: RELIHAN

DATE/TIME OF TEST: 07/27/2019 05:16:55

Diagnostics Check		OK	05:10
Air Blank		0.000	05:10
Control Test		0.081	05:11
Air Blank		0.000	05:11
Subject Sample	# 1	0.080	05:12
Breath Volume		3.613	05:12
Air Blank		0.000	05:13
Air Blank		0.000	05:14
Subject Sample	# 2	0.080	05:15
Breath Volume		1.773	05:15
Air Blank		0.000	05:16
Control Test		0.081	05:16
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 08/15/2019 11:42

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKENSON,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: LUGG,CHRIS
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 07/14/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: BUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: STURGEON

DATE/TIME OF TEST: 07/27/2019 21:30:55

Diagnostics Check		OK	21:24
Air Blank		0.000	21:25
Control Test		0.081	21:25
Air Blank		0.000	21:25
Subject Sample	# 1	0.133	21:26
Breath Volume		2.992	21:26
Air Blank		0.000	21:27
Air Blank		0.000	21:28
Subject Sample	# 2	0.116	21:29
Breath Volume		2.281	21:29
Air Blank		0.000	21:30
Control Test		0.082	21:30
Air Blank		0.000	21:30
Diagnostics Check		OK	21:30

Data Download Date/Time: 08/15/2019 11:42

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: CEASER

DATE/TIME OF TEST: 07/01/2019 00:33:22

Diagnostics Check		OK	00:21	
Air Blank		0.000	00:22	
Control Test		0.081	00:22	
Air Blank		0.000	00:22	
Subject Sample	# 1	0.170	00:26	Volume Not Met
Breath Volume		0.035	00:26	
Air Blank		0.000	00:26	
Air Blank		0.000	00:28	
Subject Sample	# 2	0.160	00:31	Volume Not Met
Breath Volume		0.167	00:31	
Air Blank		0.000	00:32	
Control Test		0.077	00:32	
Air Blank		0.000	00:33	
Diagnostics Check		OK	00:33	

Data Download Date/Time: 07/24/2019 08:41

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: PUZA

DATE/TIME OF TEST: 07/04/2019 00:23:48

Diagnostics Check		OK	00:17
Air Blank		0.000	00:17
Control Test		0.080	00:18
Air Blank		0.000	00:18
Subject Sample	# 1	0.171	00:19
Breath Volume		2.406	00:19
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.156	00:22
Breath Volume		1.554	00:22
Air Blank		0.000	00:22
Control Test		0.080	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 07/24/2019 08:41

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: DELANCY

DATE/TIME OF TEST: 07/04/2019 20:36:20

Diagnostics Check		OK	20:30
Air Blank		0.000	20:30
Control Test		0.081	20:30
Air Blank		0.000	20:31
Subject Sample	# 1	0.086	20:31
Breath Volume		1.664	20:31
Air Blank		0.000	20:32
Air Blank		0.000	20:34
Subject Sample	# 2	0.087	20:34
Breath Volume		1.519	20:34
Air Blank		0.000	20:35
Control Test		0.080	20:35
Air Blank		0.000	20:36
Diagnostics Check		OK	20:36

Data Download Date/Time: 07/24/2019 08:41

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PEREZ,RICKY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: CAMARA

DATE/TIME OF TEST: 07/04/2019 22:28:42

Diagnostics Check		OK	22:24	
Air Blank		0.000	22:24	
Control Test		0.080	22:25	
Air Blank		0.000	22:25	
Subject Sample	# 1		22:27	Subject Test Refused
Breath Volume		0.000	22:27	
Air Blank		0.000	22:27	
Control Test		0.081	22:28	
Air Blank		0.000	22:28	
Diagnostics Check		OK	22:28	

Data Download Date/Time: 07/24/2019 08:41

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 07/05/2019 00:05:48

Diagnostics Check		OK	23:59
Air Blank		0.000	00:00
Control Test		0.080	00:00
Air Blank		0.000	00:00
Subject Sample	# 1	0.174	00:01
Breath Volume		2.671	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:03
Subject Sample	# 2	0.182	00:04
Breath Volume		2.347	00:04
Air Blank		0.000	00:04
Control Test		0.081	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 07/24/2019 08:41

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ANTON,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WETZEL

DATE/TIME OF TEST: 07/06/2019 03:35:55

Diagnostics Check		OK	03:26
Air Blank		0.000	03:27
Control Test		0.080	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.234	03:28
Breath Volume		1.464	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.265	03:31
Breath Volume		1.824	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:33
Subject Sample	# 3	0.248	03:34
Breath Volume		1.910	03:34
Air Blank		0.000	03:35
Control Test		0.080	03:35
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 07/24/2019 08:41

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 07/09/2019 04:02:20

Diagnostics Check		OK	03:55
Air Blank		0.000	03:56
Control Test		0.081	03:56
Air Blank		0.000	03:56
Subject Sample	# 1	0.206	03:57
Breath Volume		3.152	03:57
Air Blank		0.000	03:58
Air Blank		0.000	03:59
Subject Sample	# 2	0.201	04:00
Breath Volume		2.085	04:00
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 07/24/2019 08:41

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/10/2019 02:43:35

Diagnostics Check		OK	02:36
Air Blank		0.000	02:36
Control Test		0.080	02:37
Air Blank		0.000	02:37
Subject Sample	# 1	0.022	02:38
Breath Volume		2.191	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.020	02:42
Breath Volume		1.136	02:42
Air Blank		0.000	02:42
Control Test		0.080	02:42
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 07/24/2019 08:41

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVIL,LUXERNE,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PREVIL,LUXERNE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 07/10/2019 19:39:29

Diagnostics Check		OK	19:33
Air Blank		0.000	19:33
Control Test		0.080	19:33
Air Blank		0.000	19:34
Subject Sample	# 1	0.146	19:34
Breath Volume		3.273	19:34
Air Blank		0.000	19:35
Air Blank		0.000	19:37
Subject Sample	# 2	0.146	19:37
Breath Volume		2.722	19:37
Air Blank		0.000	19:38
Control Test		0.080	19:38
Air Blank		0.000	19:39
Diagnostics Check		OK	19:39

Data Download Date/Time: 07/24/2019 08:41

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PEREZ,RICKY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: HANFMANN

DATE/TIME OF TEST: 07/14/2019 05:22:20

Diagnostics Check		OK	05:12
Air Blank		0.000	05:13
Control Test		0.080	05:13
Air Blank		0.000	05:14
Subject Sample	# 1	0.208	05:14
Breath Volume		2.335	05:14
Air Blank		0.000	05:15
Air Blank		0.000	05:17
Subject Sample	# 2	0.184	05:17
Breath Volume		1.500	05:17
Air Blank		0.000	05:18
Air Blank		0.000	05:19
Subject Sample	# 3	0.205	05:20
Breath Volume		2.503	05:20
Air Blank		0.000	05:21
Control Test		0.079	05:21
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 07/24/2019 08:41

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 07/14/2019 21:31:43

Diagnostics Check		OK	21:24
Air Blank		0.000	21:24
Control Test		0.082	21:25
Air Blank		0.000	21:25
Subject Sample	# 1	0.019	21:26
Breath Volume		1.570	21:26
Air Blank		0.000	21:26
Air Blank		0.000	21:28
Subject Sample	# 2	0.018	21:30
Breath Volume		1.550	21:30
Air Blank		0.000	21:30
Control Test		0.081	21:31
Air Blank		0.000	21:31
Diagnostics Check		OK	21:31

Data Download Date/Time: 07/24/2019 08:41

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAX,ERIC,C
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MONDRAGON,LIONEL
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: ESSLINGER TIMLIN

DATE/TIME OF TEST: 07/17/2019 00:00:12

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.081	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.241	23:55
Breath Volume		1.210	23:55
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 2	0.223	23:58
Breath Volume		1.503	23:58
Air Blank		0.000	23:59
Control Test		0.080	23:59
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 07/24/2019 08:41

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAX,ERIC,C
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WAX,ERIC
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 07/17/2019 01:23:56

Diagnostics Check		OK	01:17
Air Blank		0.000	01:18
Control Test		0.081	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.071	01:19
Breath Volume		2.328	01:19
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.069	01:22
Breath Volume		3.101	01:22
Air Blank		0.000	01:23
Control Test		0.080	01:23
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 07/24/2019 08:41

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAX,ERIC,C
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WAX,ERIC
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:48

SUBJECT'S LAST NAME: DENYKO

DATE/TIME OF TEST: 07/21/2019 13:43:31

Diagnostics Check		OK	13:36
Air Blank		0.000	13:37
Control Test		0.080	13:37
Air Blank		0.000	13:38
Subject Sample	# 1	0.256	13:38
Breath Volume		2.582	13:38
Air Blank		0.000	13:39
Air Blank		0.000	13:41
Subject Sample	# 2	0.260	13:41
Breath Volume		2.585	13:41
Air Blank		0.000	13:42
Control Test		0.080	13:42
Air Blank		0.000	13:43
Diagnostics Check		OK	13:43

Data Download Date/Time: 07/24/2019 08:41

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: YULFO

DATE/TIME OF TEST: 07/26/2019 21:41:20

Diagnostics Check		OK	21:34
Air Blank		0.000	21:34
Control Test		0.081	21:34
Air Blank		0.000	21:35
Subject Sample	# 1	0.023	21:36
Breath Volume		1.121	21:36
Air Blank		0.000	21:37
Air Blank		0.000	21:39
Subject Sample	# 2	0.020	21:39
Breath Volume		1.187	21:39
Air Blank		0.000	21:40
Control Test		0.081	21:40
Air Blank		0.000	21:41
Diagnostics Check		OK	21:41

Data Download Date/Time: 08/27/2019 10:58

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SHADLE

DATE/TIME OF TEST: 07/27/2019 02:22:13

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.081	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.193	02:17
Breath Volume		1.796	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.188	02:20
Breath Volume		2.007	02:20
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 08/27/2019 10:59

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: OSTEEN,JONATHAN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BAUMGARDNER

DATE/TIME OF TEST: 07/27/2019 03:53:24

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.081	03:46
Air Blank		0.000	03:47
Subject Sample	# 1	0.118	03:48
Breath Volume		1.292	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:50
Subject Sample	# 2	0.111	03:51
Breath Volume		1.160	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 08/27/2019 10:59

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 07/27/2019 20:39:54

Diagnostics Check		OK	20:31
Air Blank		0.000	20:32
Control Test		0.080	20:32
Air Blank		0.000	20:32
Subject Sample	# 1	0.134	20:34
Breath Volume		2.058	20:34
Air Blank		0.000	20:35
Air Blank		0.000	20:37
Subject Sample	# 2	0.125	20:38
Breath Volume		1.773	20:38
Air Blank		0.000	20:38
Control Test		0.079	20:39
Air Blank		0.000	20:39
Diagnostics Check		OK	20:39

Data Download Date/Time: 08/27/2019 10:59

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: KIRCHNER,MATTHEW
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/28/2019 03:01:51

Diagnostics Check		OK	02:52	
Air Blank		0.000	02:52	
Control Test		0.080	02:53	
Air Blank		0.000	02:53	
Subject Sample	# 1	0.208	02:55	
Breath Volume		1.390	02:55	
Air Blank		0.000	02:56	
Air Blank		0.000	02:58	
Subject Sample	# 2		03:00	Subject Test Refused
Breath Volume		0.148	03:00	
Air Blank		0.000	03:00	
Control Test		0.079	03:01	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

Data Download Date/Time: 08/27/2019 10:59

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHNELLBACHER,ALBERT,J
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: SCHNELLBACHER,ALBERT
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/28/2019 19:50:43

Diagnostics Check		OK	19:45	
Air Blank		0.000	19:46	
Control Test		0.079	19:46	
Air Blank		0.000	19:46	
Subject Sample	# 1	0.159	19:47	
Breath Volume		2.027	19:47	
Air Blank		0.000	19:47	
Air Blank		0.000	19:49	
Subject Sample	# 2	0.157	19:50	
Breath Volume		2.828	19:50	
Air Blank			19:50	RFI Detect
Air Blank			19:50	RFI Detect

Data Download Date/Time: 08/27/2019 10:59

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: KIRCHNER,MATTHEW
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: NICK

DATE/TIME OF TEST: 07/28/2019 20:31:01

Diagnostics Check		OK	20:23
Air Blank		0.000	20:23
Control Test		0.080	20:24
Air Blank		0.000	20:24
Subject Sample	# 1	0.258	20:26
Breath Volume		1.636	20:26
Air Blank		0.000	20:27
Air Blank		0.000	20:28
Subject Sample	# 2	0.254	20:29
Breath Volume		2.335	20:29
Air Blank		0.000	20:30
Control Test		0.075	20:30
Air Blank		0.000	20:30
Diagnostics Check		OK	20:31

Data Download Date/Time: 08/27/2019 10:59

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,WILFORD,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JORDAN,WILFORD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: SHEARER

DATE/TIME OF TEST: 07/15/2019 11:53:46

Diagnostics Check		OK	11:47
Air Blank		0.000	11:47
Control Test		0.084	11:48
Air Blank		0.000	11:48
Subject Sample	# 1	0.105	11:49
Breath Volume		2.222	11:49
Air Blank		0.000	11:49
Air Blank		0.000	11:51
Subject Sample	# 2	0.105	11:52
Breath Volume		2.269	11:52
Air Blank		0.000	11:52
Control Test		0.084	11:53
Air Blank		0.000	11:53
Diagnostics Check		OK	11:53

Data Download Date/Time: 07/28/2019 12:21

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH JR,LORNE,J
BREATH TEST OPERATOR AGENCY: Hamilton County SO
ARRESTING OFFICER: SMITH JR,LORNE
ARRESTING OFFICER AGENCY: Hamilton County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: DEBIEN

DATE/TIME OF TEST: 07/12/2019 19:56:58

Diagnostics Check		OK	19:53	
Air Blank		0.000	19:54	
Control Test		0.080	19:54	
Air Blank		0.000	19:54	
Subject Sample	# 1		19:55	Subject Test Refused
Breath Volume		0.000	19:55	
Air Blank		0.000	19:56	
Control Test		0.081	19:56	
Air Blank		0.000	19:56	
Diagnostics Check		OK	19:56	

Data Download Date/Time: 08/01/2019 01:20

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELAN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PHELAN,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 07/27/2019 11:06:00

Diagnostics Check		OK	10:59
Air Blank		0.000	11:00
Control Test		0.081	11:00
Air Blank		0.000	11:01
Subject Sample	# 1	0.000	11:01
Breath Volume		2.230	11:01
Air Blank		0.000	11:02
Air Blank		0.000	11:04
Subject Sample	# 2	0.000	11:04
Breath Volume		1.800	11:04
Air Blank		0.000	11:05
Control Test		0.080	11:05
Air Blank		0.000	11:05
Diagnostics Check		OK	11:06

Data Download Date/Time: 08/01/2019 01:20

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 07/28/2019 22:49:00

Diagnostics Check		OK	22:39
Air Blank		0.000	22:40
Control Test		0.081	22:40
Air Blank		0.000	22:40
Subject Sample	# 1	0.253	22:41
Breath Volume		1.246	22:41
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.232	22:44
Breath Volume		1.332	22:44
Air Blank		0.000	22:45
Air Blank		0.000	22:46
Subject Sample	# 3	0.233	22:47
Breath Volume		1.945	22:47
Air Blank		0.000	22:48
Control Test		0.079	22:48
Air Blank		0.000	22:48
Diagnostics Check		OK	22:49

Data Download Date/Time: 08/01/2019 01:20

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 07/10/2019 01:00:28

Diagnostics Check		OK	00:54
Air Blank		0.000	00:55
Control Test		0.080	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.000	00:56
Breath Volume		1.507	00:56
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.000	00:59
Breath Volume		1.824	00:59
Air Blank		0.000	00:59
Control Test		0.080	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 07/10/2019 13:16

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FERGUSON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: IBARRA

DATE/TIME OF TEST: 07/12/2019 03:35:26

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:32	
Control Test		0.080	03:32	
Air Blank		0.000	03:33	
Subject Sample	# 1		03:34	Subject Test Refused
Breath Volume		0.000	03:34	
Air Blank		0.000	03:34	
Control Test		0.080	03:34	
Air Blank		0.000	03:35	
Diagnostics Check		OK	03:35	

Data Download Date/Time: 08/06/2019 14:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: WALKINGSTICK

DATE/TIME OF TEST: 07/13/2019 02:57:08

Diagnostics Check		OK	02:49
Air Blank		0.000	02:49
Control Test		0.081	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.225	02:51
Breath Volume		1.601	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:53
Subject Sample	# 2	0.209	02:55
Breath Volume		1.183	02:55
Air Blank		0.000	02:56
Control Test		0.079	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 08/06/2019 14:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: MAGIDA

DATE/TIME OF TEST: 07/13/2019 03:14:38

Diagnostics Check		OK	03:06
Air Blank		0.000	03:07
Control Test		0.079	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.119	03:09
Breath Volume		1.664	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.119	03:13
Breath Volume		1.476	03:13
Air Blank		0.000	03:13
Control Test		0.079	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 08/06/2019 14:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 07/13/2019 05:31:35

Diagnostics Check		OK	05:24
Air Blank		0.000	05:25
Control Test		0.078	05:25
Air Blank		0.000	05:26
Subject Sample	# 1	0.177	05:27
Breath Volume		2.617	05:27
Air Blank		0.000	05:27
Air Blank		0.000	05:29
Subject Sample	# 2	0.178	05:30
Breath Volume		2.644	05:30
Air Blank		0.000	05:30
Control Test		0.080	05:30
Air Blank		0.000	05:31
Diagnostics Check		OK	05:31

Data Download Date/Time: 08/06/2019 14:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: DIAZ SOTO

DATE/TIME OF TEST: 07/13/2019 22:22:49

Diagnostics Check		OK	22:11	
Air Blank		0.000	22:11	
Control Test		0.080	22:12	
Air Blank		0.000	22:12	
Subject Sample	# 1		22:15	No Sample Provided
Breath Volume		0.000	22:15	
Air Blank		0.000	22:16	
Air Blank		0.000	22:18	
Subject Sample	# 2		22:21	No Sample Provided
Breath Volume		0.000	22:21	
Air Blank		0.000	22:21	
Control Test		0.080	22:22	
Air Blank		0.000	22:22	
Diagnostics Check		OK	22:22	

Data Download Date/Time: 08/06/2019 14:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: GALLEGO MARCOS

DATE/TIME OF TEST: 07/13/2019 23:55:55

Diagnostics Check		OK	23:49
Air Blank		0.000	23:49
Control Test		0.079	23:50
Air Blank		0.000	23:50
Subject Sample	# 1	0.259	23:51
Breath Volume		1.878	23:51
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.254	23:54
Breath Volume		2.085	23:54
Air Blank		0.000	23:54
Control Test		0.080	23:55
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 08/06/2019 14:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 07/14/2019 02:08:55

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.079	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.135	02:04
Breath Volume		2.382	02:04
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.136	02:07
Breath Volume		2.542	02:07
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 08/06/2019 14:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: RIANO GARCIA

DATE/TIME OF TEST: 07/14/2019 19:37:25

Diagnostics Check		OK	19:28
Air Blank		0.000	19:28
Control Test		0.080	19:29
Air Blank		0.000	19:29
Subject Sample	# 1	0.203	19:32
Breath Volume		2.246	19:32
Air Blank		0.000	19:33
Air Blank		0.000	19:35
Subject Sample	# 2	0.183	19:35
Breath Volume		2.378	19:35
Air Blank		0.000	19:36
Control Test		0.078	19:36
Air Blank		0.000	19:37
Diagnostics Check		OK	19:37

Data Download Date/Time: 08/06/2019 14:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: TANNER

DATE/TIME OF TEST: 07/15/2019 00:13:41

Diagnostics Check		OK	00:06
Air Blank		0.000	00:06
Control Test		0.080	00:06
Air Blank		0.000	00:07
Subject Sample	# 1	0.154	00:08
Breath Volume		2.437	00:08
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 2	0.160	00:12
Breath Volume		3.312	00:12
Air Blank		0.000	00:12
Control Test		0.077	00:13
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 08/06/2019 14:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: ACKER

DATE/TIME OF TEST: 07/16/2019 00:32:08

Diagnostics Check		OK	00:25
Air Blank		0.000	00:26
Control Test		0.080	00:26
Air Blank		0.000	00:26
Subject Sample	# 1	0.049	00:27
Breath Volume		1.894	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.049	00:30
Breath Volume		1.605	00:30
Air Blank		0.000	00:31
Control Test		0.079	00:31
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 08/06/2019 14:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: MATTHAY

DATE/TIME OF TEST: 07/17/2019 02:21:03

Diagnostics Check		OK	02:13
Air Blank		0.000	02:14
Control Test		0.080	02:14
Air Blank		0.000	02:15
Subject Sample	# 1	0.110	02:15
Breath Volume		2.570	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.112	02:19
Breath Volume		2.546	02:19
Air Blank		0.000	02:20
Control Test		0.080	02:20
Air Blank		0.000	02:20
Diagnostics Check		OK	02:21

Data Download Date/Time: 08/06/2019 14:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: WORVIE

DATE/TIME OF TEST: 07/18/2019 05:38:25

Diagnostics Check		OK	05:31
Air Blank		0.000	05:32
Control Test		0.080	05:32
Air Blank		0.000	05:32
Subject Sample	# 1	0.072	05:33
Breath Volume		2.933	05:33
Air Blank		0.000	05:34
Air Blank		0.000	05:36
Subject Sample	# 2	0.070	05:36
Breath Volume		3.144	05:36
Air Blank		0.000	05:37
Control Test		0.079	05:37
Air Blank		0.000	05:38
Diagnostics Check		OK	05:38

Data Download Date/Time: 08/06/2019 14:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 07/19/2019 23:36:36

Diagnostics Check		OK	23:29
Air Blank		0.000	23:29
Control Test		0.079	23:29
Air Blank		0.000	23:30
Subject Sample	# 1	0.292	23:31
Breath Volume		1.613	23:31
Air Blank		0.000	23:32
Air Blank		0.000	23:34
Subject Sample	# 2	0.287	23:34
Breath Volume		1.308	23:34
Air Blank		0.000	23:35
Control Test		0.078	23:35
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 08/06/2019 14:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WEG,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: MCCLENDON

DATE/TIME OF TEST: 07/20/2019 00:46:19

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.080	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.162	00:41
Breath Volume		1.652	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:43
Subject Sample	# 2	0.163	00:44
Breath Volume		1.832	00:44
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 08/06/2019 14:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: MAHAMUD

DATE/TIME OF TEST: 07/20/2019 02:40:31

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:38	
Control Test		0.079	02:38	
Air Blank		0.000	02:38	
Subject Sample	# 1		02:39	Subject Test Refused
Breath Volume		0.000	02:39	
Air Blank		0.000	02:39	
Control Test		0.081	02:39	
Air Blank		0.000	02:40	
Diagnostics Check		OK	02:40	

Data Download Date/Time: 08/06/2019 14:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: CREQUE

DATE/TIME OF TEST: 07/21/2019 02:11:15

Diagnostics Check		OK	02:01
Air Blank		0.000	02:02
Control Test		0.080	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.168	02:06
Breath Volume		1.960	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:08
Subject Sample	# 2	0.175	02:09
Breath Volume		1.699	02:09
Air Blank		0.000	02:10
Control Test		0.078	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 08/06/2019 14:03

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 07/21/2019 03:26:56

Diagnostics Check		OK	03:19
Air Blank		0.000	03:20
Control Test		0.079	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.190	03:22
Breath Volume		1.378	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:24
Subject Sample	# 2	0.200	03:25
Breath Volume		1.867	03:25
Air Blank		0.000	03:25
Control Test		0.077	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 08/06/2019 14:04

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FERGUSON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: KONITZER

DATE/TIME OF TEST: 07/21/2019 04:19:11

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.079	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.191	04:14
Breath Volume		2.355	04:14
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.183	04:17
Breath Volume		1.718	04:17
Air Blank		0.000	04:18
Control Test		0.077	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 08/06/2019 14:04

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WEG,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: QUINTERO URBANEJA

DATE/TIME OF TEST: 07/21/2019 23:50:22

Diagnostics Check		OK	23:43
Air Blank		0.000	23:43
Control Test		0.080	23:44
Air Blank		0.000	23:44
Subject Sample	# 1	0.165	23:45
Breath Volume		1.566	23:45
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.169	23:48
Breath Volume		2.902	23:48
Air Blank		0.000	23:49
Control Test		0.079	23:49
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 08/06/2019 14:04

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGQUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/24/2019 05:18:50

Diagnostics Check		OK	05:12
Air Blank		0.000	05:12
Control Test		0.081	05:13
Air Blank		0.000	05:13
Subject Sample	# 1	0.123	05:14
Breath Volume		3.242	05:14
Air Blank		0.000	05:14
Air Blank		0.000	05:16
Subject Sample	# 2	0.118	05:17
Breath Volume		2.636	05:17
Air Blank		0.000	05:17
Control Test		0.080	05:18
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 08/06/2019 14:04

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/25/2019 01:09:28

Diagnostics Check		OK	01:03
Air Blank		0.000	01:03
Control Test		0.079	01:04
Air Blank		0.000	01:04
Subject Sample	# 1	0.258	01:04
Breath Volume		1.746	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.257	01:07
Breath Volume		1.320	01:07
Air Blank		0.000	01:08
Control Test		0.079	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 08/06/2019 14:04

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KLENOFSKY,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: WALLS

DATE/TIME OF TEST: 07/25/2019 04:31:17

Diagnostics Check		OK	04:24
Air Blank		0.000	04:25
Control Test		0.080	04:25
Air Blank		0.000	04:26
Subject Sample	# 1	0.174	04:26
Breath Volume		3.433	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.176	04:29
Breath Volume		2.632	04:29
Air Blank		0.000	04:30
Control Test		0.079	04:30
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 08/06/2019 14:04

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: SHELLENBERGER

DATE/TIME OF TEST: 07/27/2019 03:13:23

Diagnostics Check		OK	03:06
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.164	03:08
Breath Volume		4.667	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.163	03:11
Breath Volume		4.421	03:11
Air Blank		0.000	03:12
Control Test		0.080	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 08/06/2019 14:04

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: MEDINA DIAZ

DATE/TIME OF TEST: 07/28/2019 03:57:57

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.196	03:53
Breath Volume		2.230	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.185	03:56
Breath Volume		1.304	03:56
Air Blank		0.000	03:57
Control Test		0.078	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 08/06/2019 14:04

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: HERNANDEZ BUENO

DATE/TIME OF TEST: 07/28/2019 05:33:22

Diagnostics Check		OK	05:27
Air Blank		0.000	05:27
Control Test		0.080	05:28
Air Blank		0.000	05:28
Subject Sample	# 1	0.189	05:28
Breath Volume		3.542	05:28
Air Blank		0.000	05:29
Air Blank		0.000	05:31
Subject Sample	# 2	0.186	05:31
Breath Volume		3.566	05:31
Air Blank		0.000	05:32
Control Test		0.079	05:32
Air Blank		0.000	05:33
Diagnostics Check		OK	05:33

Data Download Date/Time: 08/06/2019 14:04

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: DURAN LOYO

DATE/TIME OF TEST: 07/29/2019 04:28:46

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:27	
Control Test	0.081	04:27	Interferent Detect
Air Blank		04:28	Purge Fail

Data Download Date/Time: 08/06/2019 14:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: HOOD

DATE/TIME OF TEST: 07/29/2019 05:45:35

Diagnostics Check		OK	05:42	
Air Blank		0.000	05:42	
Control Test		0.079	05:43	
Air Blank		0.000	05:43	
Subject Sample	# 1		05:44	Interferent Detect
Breath Volume		2.359	05:44	
Air Blank		0.000	05:44	
Control Test		0.078	05:45	Interferent Detect
Air Blank		0.000	05:45	

Data Download Date/Time: 08/06/2019 14:04

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES,JONATHAN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: NETTLES,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1761808A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 07/05/2019 00:24:35

Diagnostics Check		OK	00:17
Air Blank		0.000	00:18
Control Test		0.081	00:18
Air Blank		0.000	00:19
Subject Sample	# 1	0.102	00:19
Breath Volume		2.101	00:19
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.097	00:23
Breath Volume		1.957	00:23
Air Blank		0.000	00:23
Control Test		0.081	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 07/28/2019 21:26

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,ANTHONY,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ESQUIVEL,ANTHONY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 07/23/2019 02:54:42

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:51	
Control Test		0.081	02:52	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:53	Subject Test Refused
Breath Volume		0.000	02:53	
Air Blank		0.000	02:53	
Control Test		0.082	02:54	
Air Blank		0.000	02:54	
Diagnostics Check		OK	02:54	

Data Download Date/Time: 08/15/2019 11:25

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARNEST,WILLIAM,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: EARNEST,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: FELICIANO

DATE/TIME OF TEST: 07/24/2019 22:52:06

Diagnostics Check		OK	22:45
Air Blank		0.000	22:45
Control Test		0.081	22:45
Air Blank		0.000	22:46
Subject Sample	# 1	0.280	22:47
Breath Volume		3.152	22:47
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.261	22:50
Breath Volume		2.937	22:50
Air Blank		0.000	22:51
Control Test		0.081	22:51
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

Data Download Date/Time: 08/15/2019 11:25

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: OVERSTREET

DATE/TIME OF TEST: 07/25/2019 20:43:57

Diagnostics Check		OK	20:36
Air Blank		0.000	20:37
Control Test		0.081	20:37
Air Blank		0.000	20:38
Subject Sample	# 1	0.082	20:38
Breath Volume		1.101	20:38
Air Blank		0.000	20:39
Air Blank		0.000	20:40
Subject Sample	# 2	0.077	20:42
Breath Volume		1.171	20:42
Air Blank		0.000	20:43
Control Test		0.081	20:43
Air Blank		0.000	20:43
Diagnostics Check		OK	20:43

Data Download Date/Time: 08/15/2019 11:25

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINTON,BLAKE,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MINTON,BLAKE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 07/27/2019 03:45:23

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.082	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.159	03:40
Breath Volume		1.457	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.161	03:43
Breath Volume		2.328	03:43
Air Blank		0.000	03:44
Control Test		0.082	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 08/15/2019 11:25

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTIGLIONE,S.M.,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CASTIGLIONE,S.M
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: LARKIN

DATE/TIME OF TEST: 07/28/2019 05:18:07

Diagnostics Check		OK	05:12
Air Blank		0.000	05:12
Control Test		0.081	05:12
Air Blank		0.000	05:13
Subject Sample	# 1	0.137	05:13
Breath Volume		3.386	05:13
Air Blank		0.000	05:14
Air Blank		0.000	05:16
Subject Sample	# 2	0.128	05:16
Breath Volume		3.007	05:16
Air Blank		0.000	05:17
Control Test		0.081	05:17
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 08/15/2019 11:25

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,COLLEEN,E
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KENT,NATHAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 07/04/2019 19:49:32

Diagnostics Check		OK	19:43
Air Blank		0.000	19:43
Control Test		0.081	19:43
Air Blank		0.000	19:44
Subject Sample	# 1	0.137	19:44
Breath Volume		1.972	19:44
Air Blank		0.000	19:45
Air Blank		0.000	19:47
Subject Sample	# 2	0.152	19:47
Breath Volume		1.164	19:47
Air Blank		0.000	19:48
Control Test		0.080	19:48
Air Blank		0.000	19:49
Diagnostics Check		OK	19:49

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELM,DEREK,W
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HELM,DEREK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: DEFRANCISCO

DATE/TIME OF TEST: 07/05/2019 03:57:03

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.081	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.130	03:52
Breath Volume		1.292	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:54
Subject Sample	# 2	0.134	03:55
Breath Volume		1.929	03:55
Air Blank		0.000	03:56
Control Test		0.081	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:57

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,PATRICK,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KIBBEY,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: SEYBOLD

DATE/TIME OF TEST: 07/05/2019 04:29:51

Diagnostics Check		OK	04:23
Air Blank		0.000	04:24
Control Test		0.081	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.172	04:25
Breath Volume		3.062	04:25
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.168	04:28
Breath Volume		2.945	04:28
Air Blank		0.000	04:28
Control Test		0.082	04:29
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATIF,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SMITH,BRIAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:47

SUBJECT'S LAST NAME: BURELA

DATE/TIME OF TEST: 07/09/2019 18:18:57

Diagnostics Check		OK	18:07	
Air Blank		0.000	18:07	
Control Test		0.082	18:08	
Air Blank		0.000	18:08	
Subject Sample	# 1	0.222	18:11	Volume Not Met
Breath Volume		0.250	18:11	
Air Blank		0.000	18:12	
Air Blank		0.000	18:14	
Subject Sample	# 2	0.212	18:17	Volume Not Met
Breath Volume		0.000	18:17	
Air Blank		0.000	18:17	
Control Test		0.080	18:18	
Air Blank		0.000	18:18	
Diagnostics Check		OK	18:18	

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MITRO,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 07/11/2019 03:10:10

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:07	
Control Test		0.082	03:07	
Air Blank		0.000	03:08	
Subject Sample	# 1		03:08	Subject Test Refused
Breath Volume		0.000	03:08	
Air Blank		0.000	03:09	
Control Test		0.082	03:09	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: LECLERC

DATE/TIME OF TEST: 07/12/2019 02:37:08

Diagnostics Check		OK	02:29
Air Blank		0.000	02:30
Control Test		0.083	02:30
Air Blank		0.000	02:31
Subject Sample	# 1	0.215	02:32
Breath Volume		2.265	02:32
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.208	02:35
Breath Volume		2.207	02:35
Air Blank		0.000	02:36
Control Test		0.082	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MCKINLEY

DATE/TIME OF TEST: 07/12/2019 03:12:48

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.082	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.137	03:08
Breath Volume		2.324	03:08
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.141	03:11
Breath Volume		2.324	03:11
Air Blank		0.000	03:11
Control Test		0.083	03:12
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JONATHON,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: VARVARO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: WILLEMIN

DATE/TIME OF TEST: 07/12/2019 03:52:59

Diagnostics Check		OK	03:46
Air Blank		0.000	03:47
Control Test		0.083	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.059	03:48
Breath Volume		4.125	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:50
Subject Sample	# 2	0.058	03:51
Breath Volume		4.183	03:51
Air Blank		0.000	03:52
Control Test		0.082	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: ALVAREZ ARMBRUSTER

DATE/TIME OF TEST: 07/12/2019 04:52:25

Diagnostics Check		OK	04:49	
Air Blank		0.000	04:49	
Control Test		0.082	04:49	
Air Blank		0.000	04:50	
Subject Sample	# 1		04:50	RFI Detect
Breath Volume		0.386	04:50	
Air Blank		0.000	04:51	
Control Test		0.082	04:51	
Air Blank		0.000	04:52	
Diagnostics Check		OK	04:52	

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: ALVAREZ ARMBRUSTER

DATE/TIME OF TEST: 07/12/2019 05:03:03

Diagnostics Check		OK	04:54
Air Blank		0.000	04:55
Control Test		0.082	04:55
Air Blank		0.000	04:56
Subject Sample	# 1	0.163	04:57
Breath Volume		2.234	04:57
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.162	05:01
Breath Volume		1.882	05:01
Air Blank		0.000	05:02
Control Test		0.080	05:02
Air Blank		0.000	05:02
Diagnostics Check		OK	05:03

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELM,DEREK,W
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIBBEY,BRAD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: CRAMER

DATE/TIME OF TEST: 07/13/2019 06:05:59

Diagnostics Check		OK	05:59
Air Blank		0.000	05:59
Control Test		0.082	05:59
Air Blank		0.000	06:00
Subject Sample	# 1	0.179	06:00
Breath Volume		1.101	06:00
Air Blank		0.000	06:01
Air Blank		0.000	06:03
Subject Sample	# 2	0.172	06:04
Breath Volume		2.714	06:04
Air Blank		0.000	06:05
Control Test		0.082	06:05
Air Blank		0.000	06:05
Diagnostics Check		OK	06:05

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEIRCE,HAROLD,F
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SCHWARZ,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:49

SUBJECT'S LAST NAME: GILLIS

DATE/TIME OF TEST: 07/13/2019 14:18:37

Diagnostics Check		OK	14:09
Air Blank		0.000	14:10
Control Test		0.082	14:10
Air Blank		0.000	14:11
Subject Sample	# 1	0.062	14:13
Breath Volume		1.425	14:13
Air Blank		0.000	14:14
Air Blank		0.000	14:16
Subject Sample	# 2	0.064	14:17
Breath Volume		1.812	14:17
Air Blank		0.000	14:17
Control Test		0.082	14:18
Air Blank		0.000	14:18
Diagnostics Check		OK	14:18

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATIF,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RAMOS,MICHAEL
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 07/13/2019 23:21:12

Diagnostics Check		OK	23:14
Air Blank		0.000	23:14
Control Test		0.082	23:15
Air Blank		0.000	23:15
Subject Sample	# 1	0.232	23:16
Breath Volume		1.761	23:16
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.237	23:19
Breath Volume		1.156	23:19
Air Blank		0.000	23:20
Control Test		0.082	23:20
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATIF,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WILSON,KYLE
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: LIBELL

DATE/TIME OF TEST: 07/14/2019 21:22:26

Diagnostics Check		OK	21:16
Air Blank		0.000	21:16
Control Test		0.081	21:16
Air Blank		0.000	21:17
Subject Sample	# 1	0.188	21:18
Breath Volume		2.136	21:18
Air Blank		0.000	21:18
Air Blank		0.000	21:20
Subject Sample	# 2	0.173	21:20
Breath Volume		1.863	21:20
Air Blank		0.000	21:21
Control Test		0.081	21:21
Air Blank		0.000	21:22
Diagnostics Check		OK	21:22

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,JONATHAN,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MATHIS,JONATHAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: ABRAHAM

DATE/TIME OF TEST: 07/15/2019 18:04:29

Diagnostics Check		OK	17:52	
Air Blank		0.000	17:53	
Control Test		0.082	17:53	
Air Blank		0.000	17:54	
Subject Sample	# 1	0.142	17:57	Volume Not Met
Breath Volume		0.230	17:57	
Air Blank		0.000	17:57	
Air Blank		0.000	17:59	
Subject Sample	# 2	0.111	18:02	Volume Not Met
Breath Volume		0.289	18:02	
Air Blank		0.000	18:03	
Control Test		0.081	18:03	
Air Blank		0.000	18:04	
Diagnostics Check		OK	18:04	

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: FRYAR

DATE/TIME OF TEST: 07/15/2019 19:21:16

Diagnostics Check		OK	19:08	
Air Blank		0.000	19:08	
Control Test		0.082	19:08	
Air Blank		0.000	19:09	
Subject Sample	# 1	0.160	19:12	Volume Not Met
Breath Volume		0.546	19:12	
Air Blank		0.000	19:13	
Air Blank		0.000	19:15	
Subject Sample	# 2	0.167	19:15	
Breath Volume		1.308	19:15	
Air Blank		0.000	19:16	
Air Blank		0.000	19:18	
Subject Sample	# 3	0.179	19:19	
Breath Volume		2.070	19:19	
Air Blank		0.000	19:20	
Control Test		0.079	19:20	
Air Blank		0.000	19:21	
Diagnostics Check		OK	19:21	

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELM,DEREK,W
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HELM,DEREK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/19/2019 00:13:24

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.083	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.041	00:09
Breath Volume		2.097	00:09
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 2	0.040	00:11
Breath Volume		2.164	00:11
Air Blank		0.000	00:12
Control Test		0.082	00:12
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: BONDERSON

DATE/TIME OF TEST: 07/21/2019 01:55:47

Diagnostics Check		OK	01:48
Air Blank		0.000	01:49
Control Test		0.082	01:49
Air Blank		0.000	01:49
Subject Sample	# 1	0.147	01:50
Breath Volume		1.929	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.145	01:54
Breath Volume		1.660	01:54
Air Blank		0.000	01:54
Control Test		0.082	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/20/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: MECCARIELLO

DATE/TIME OF TEST: 07/21/2019 02:38:37

Diagnostics Check		OK	02:33	
Air Blank		0.000	02:33	
Control Test		0.080	02:34	
Air Blank		0.000	02:34	
Subject Sample	# 1		02:37	Subject Test Refused
Breath Volume		0.000	02:37	
Air Blank		0.000	02:37	
Control Test		0.080	02:38	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:38	

Data Download Date/Time: 07/21/2019 13:37

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,JONATHAN,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MATHIS,JONATHAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:07

SUBJECT'S LAST NAME: ATKINSON

DATE/TIME OF TEST: 07/21/2019 18:34:13

Diagnostics Check		OK	18:27
Air Blank		0.000	18:27
Control Test		0.079	18:28
Air Blank		0.000	18:28
Subject Sample	# 1	0.316	18:29
Breath Volume		2.519	18:29
Air Blank		0.000	18:30
Air Blank		0.000	18:31
Subject Sample	# 2	0.308	18:32
Breath Volume		1.851	18:32
Air Blank		0.000	18:33
Control Test		0.080	18:33
Air Blank		0.000	18:34
Diagnostics Check		OK	18:34

Data Download Date/Time: 08/25/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: ZIMMER

DATE/TIME OF TEST: 07/25/2019 00:21:46

Diagnostics Check		OK	00:17	
Air Blank		0.000	00:17	
Control Test		0.082	00:18	
Air Blank		0.000	00:18	
Subject Sample	# 1		00:20	Subject Test Refused
Breath Volume		0.000	00:20	
Air Blank		0.000	00:20	
Control Test		0.082	00:21	
Air Blank		0.000	00:21	
Diagnostics Check		OK	00:21	

Data Download Date/Time: 08/25/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIMPSON,RACHEL,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SCHMIDT,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 07/25/2019 07:37:07

Diagnostics Check		OK	07:30
Air Blank		0.000	07:31
Control Test		0.083	07:31
Air Blank		0.000	07:31
Subject Sample	# 1	0.131	07:32
Breath Volume		1.695	07:32
Air Blank		0.000	07:33
Air Blank		0.000	07:34
Subject Sample	# 2	0.144	07:35
Breath Volume		2.226	07:35
Air Blank		0.000	07:36
Control Test		0.082	07:36
Air Blank		0.000	07:37
Diagnostics Check		OK	07:37

Data Download Date/Time: 08/25/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,BILLY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ORTIZ,BILLY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/25/2019 20:40:06

Diagnostics Check		OK	20:25	
Air Blank		0.000	20:25	
Control Test		0.082	20:26	
Air Blank		0.000	20:26	
Subject Sample	# 1	0.000	20:27	
Breath Volume		1.644	20:27	
Air Blank		0.000	20:27	
Air Blank		0.000	20:29	
Subject Sample	# 2	0.000	20:33	Volume Not Met
Breath Volume		0.656	20:33	
Air Blank		0.000	20:33	
Air Blank		0.000	20:35	
Subject Sample	# 3	0.000	20:38	Volume Not Met
Breath Volume		0.117	20:38	
Air Blank		0.000	20:39	
Control Test		0.082	20:39	
Air Blank		0.000	20:40	
Diagnostics Check		OK	20:40	

Data Download Date/Time: 08/25/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,BILLY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ORTIZ,BILLY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/25/2019 20:54:56

Diagnostics Check		OK	20:43	
Air Blank		0.000	20:43	
Control Test		0.081	20:44	
Air Blank		0.000	20:44	
Subject Sample	# 1	0.000	20:44	
Breath Volume		2.097	20:44	
Air Blank		0.000	20:45	
Air Blank		0.000	20:47	
Subject Sample	# 2	0.000	20:50	Volume Not Met
Breath Volume		0.070	20:50	
Air Blank		0.000	20:51	
Air Blank		0.000	20:53	
Subject Sample	# 3	0.000	20:53	
Breath Volume		2.031	20:53	
Air Blank		0.000	20:54	
Control Test		0.081	20:54	
Air Blank		0.000	20:54	
Diagnostics Check		OK	20:54	

Data Download Date/Time: 08/25/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 07/26/2019 01:09:37

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:05	
Control Test		0.082	01:05	
Air Blank		0.000	01:06	
Subject Sample	# 1		01:08	Subject Test Refused
Breath Volume		0.000	01:08	
Air Blank		0.000	01:08	
Control Test		0.083	01:09	
Air Blank		0.000	01:09	
Diagnostics Check		OK	01:09	

Data Download Date/Time: 08/25/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEIRCE,HAROLD,F
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SMITH,BRIAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:02

SUBJECT'S LAST NAME: CORREIA

DATE/TIME OF TEST: 07/26/2019 16:31:45

Diagnostics Check		OK	16:22
Air Blank		0.000	16:23
Control Test		0.081	16:23
Air Blank		0.000	16:24
Subject Sample	# 1	0.311	16:25
Breath Volume		1.253	16:25
Air Blank		0.000	16:26
Air Blank		0.000	16:27
Subject Sample	# 2	0.312	16:30
Breath Volume		1.921	16:30
Air Blank		0.000	16:30
Control Test		0.079	16:31
Air Blank		0.000	16:31
Diagnostics Check		OK	16:31

Data Download Date/Time: 08/25/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELM,DEREK,W
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HELM,DEREK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 07/29/2019 04:13:58

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.081	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.102	04:09
Breath Volume		1.527	04:09
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.098	04:12
Breath Volume		1.230	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 08/25/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: AKIEME

DATE/TIME OF TEST: 07/30/2019 20:36:44

Diagnostics Check		OK	20:24	
Air Blank		0.000	20:25	
Control Test		0.081	20:25	
Air Blank		0.000	20:25	
Subject Sample	# 1	0.000	20:28	Volume Not Met
Breath Volume		0.085	20:28	
Air Blank		0.000	20:29	
Air Blank		0.000	20:31	
Subject Sample	# 2	0.194	20:32	
Breath Volume		1.179	20:32	
Air Blank		0.000	20:32	
Air Blank		0.000	20:34	
Subject Sample	# 3	0.192	20:35	
Breath Volume		1.804	20:35	
Air Blank		0.000	20:35	
Control Test		0.080	20:36	
Air Blank		0.000	20:36	
Diagnostics Check		OK	20:36	

Data Download Date/Time: 08/25/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETROVICH,DARREN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PETROVICH,DARREN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/21/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:19

SUBJECT'S LAST NAME: GRZETICH

DATE/TIME OF TEST: 07/31/2019 14:56:31

Diagnostics Check		OK	14:42	
Air Blank		0.000	14:42	
Control Test		0.079	14:43	
Air Blank		0.000	14:43	
Subject Sample	# 1	0.000	14:43	
Breath Volume		2.121	14:43	
Air Blank		0.000	14:44	
Air Blank		0.000	14:46	
Subject Sample	# 2	0.000	14:49	Volume Not Met
Breath Volume		0.722	14:49	
Air Blank		0.000	14:50	
Air Blank		0.000	14:52	
Subject Sample	# 3	0.000	14:55	Volume Not Met
Breath Volume		0.613	14:55	
Air Blank		0.000	14:55	
Control Test		0.081	14:55	
Air Blank		0.000	14:56	
Diagnostics Check		OK	14:56	

Data Download Date/Time: 08/25/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HALL,JESSE
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:23

SUBJECT'S LAST NAME: BENITEZ

DATE/TIME OF TEST: 07/04/2019 08:51:10

Diagnostics Check		OK	08:44
Air Blank		0.000	08:45
Control Test		0.080	08:45
Air Blank		0.000	08:46
Subject Sample	# 1	0.000	08:46
Breath Volume		1.144	08:46
Air Blank		0.000	08:47
Air Blank		0.000	08:49
Subject Sample	# 2	0.000	08:49
Breath Volume		1.328	08:49
Air Blank		0.000	08:50
Control Test		0.079	08:50
Air Blank		0.000	08:51
Diagnostics Check		OK	08:51

Data Download Date/Time: 07/31/2019 16:38

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: DELAINO

DATE/TIME OF TEST: 07/04/2019 20:07:35

Diagnostics Check		OK	20:01
Air Blank		0.000	20:01
Control Test		0.080	20:02
Air Blank		0.000	20:02
Subject Sample	# 1	0.000	20:03
Breath Volume		3.167	20:03
Air Blank		0.000	20:03
Air Blank		0.000	20:05
Subject Sample	# 2	0.000	20:06
Breath Volume		5.203	20:06
Air Blank		0.000	20:06
Control Test		0.078	20:06
Air Blank		0.000	20:07
Diagnostics Check		OK	20:07

Data Download Date/Time: 07/31/2019 16:38

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: POLAK

DATE/TIME OF TEST: 07/05/2019 00:16:12

Diagnostics Check		OK	00:04	
Air Blank		0.000	00:05	
Control Test		0.078	00:05	
Air Blank		0.000	00:06	
Subject Sample	# 1	0.000	00:09	Volume Not Met
Breath Volume		0.031	00:09	
Air Blank		0.000	00:09	
Air Blank		0.000	00:11	
Subject Sample	# 2	0.000	00:14	Volume Not Met
Breath Volume		0.714	00:14	
Air Blank		0.000	00:15	
Control Test		0.078	00:15	
Air Blank		0.000	00:16	
Diagnostics Check		OK	00:16	

Data Download Date/Time: 07/31/2019 16:38

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JARROD,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: FLYNN,PATRICK
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: WEILER

DATE/TIME OF TEST: 07/08/2019 16:41:35

Diagnostics Check		OK	16:35
Air Blank		0.000	16:36
Control Test		0.080	16:36
Air Blank		0.000	16:36
Subject Sample	# 1	0.000	16:37
Breath Volume		2.613	16:37
Air Blank		0.000	16:37
Air Blank		0.000	16:39
Subject Sample	# 2	0.000	16:40
Breath Volume		1.980	16:40
Air Blank		0.000	16:40
Control Test		0.079	16:40
Air Blank		0.000	16:41
Diagnostics Check		OK	16:41

Data Download Date/Time: 07/31/2019 16:38

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: AKINS

DATE/TIME OF TEST: 07/09/2019 00:55:51

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.078	00:50
Air Blank		0.000	00:50
Subject Sample	# 1	0.210	00:51
Breath Volume		4.257	00:51
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.207	00:54
Breath Volume		5.414	00:54
Air Blank		0.000	00:54
Control Test		0.077	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 07/31/2019 16:38

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: RUTHERFORD

DATE/TIME OF TEST: 07/09/2019 22:52:25

Diagnostics Check		OK	22:39	
Air Blank		0.000	22:40	
Control Test		0.078	22:40	
Air Blank		0.000	22:41	
Subject Sample	# 1	0.081	22:44	Volume Not Met
Breath Volume		0.964	22:44	
Air Blank		0.000	22:44	
Air Blank		0.000	22:46	
Subject Sample	# 2	0.080	22:47	
Breath Volume		1.109	22:47	
Air Blank		0.000	22:48	
Air Blank		0.000	22:50	
Subject Sample	# 3	0.078	22:50	
Breath Volume		1.382	22:50	
Air Blank		0.000	22:51	
Control Test		0.078	22:51	
Air Blank		0.000	22:52	
Diagnostics Check		OK	22:52	

Data Download Date/Time: 07/31/2019 16:39

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: KINDER,ARRON
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: ELWELL

DATE/TIME OF TEST: 07/12/2019 22:01:22

Diagnostics Check		OK	21:57	
Air Blank		0.000	21:58	
Control Test		0.079	21:58	
Air Blank		0.000	21:59	
Subject Sample	# 1		21:59	Subject Test Refused
Breath Volume		0.000	21:59	
Air Blank		0.000	22:00	
Control Test		0.079	22:00	
Air Blank		0.000	22:01	
Diagnostics Check		OK	22:01	

Data Download Date/Time: 07/31/2019 16:39

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ROESEL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/13/2019 04:26:35

Diagnostics Check		OK	04:14
Air Blank		0.000	04:15
Control Test		0.079	04:15
Air Blank		0.000	04:15
Subject Sample	# 1	0.184	04:17
Breath Volume		2.050	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:19
Subject Sample	# 2	0.151	04:21
Breath Volume		1.218	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:23
Subject Sample	# 3	0.157	04:24
Breath Volume		1.535	04:24
Air Blank		0.000	04:25
Control Test		0.078	04:25
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 07/31/2019 16:39

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SULLIVAN,BRANDON
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: HUBBERT

DATE/TIME OF TEST: 07/15/2019 00:16:55

Diagnostics Check		OK	00:13	
Air Blank		0.000	00:14	
Control Test		0.079	00:14	
Air Blank		0.000	00:14	
Subject Sample	# 1		00:15	Subject Test Refused
Breath Volume		0.000	00:15	
Air Blank		0.000	00:15	
Control Test		0.078	00:16	
Air Blank		0.000	00:16	
Diagnostics Check		OK	00:16	

Data Download Date/Time: 07/31/2019 16:39

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP-METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WHITMORE,JAMES
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: SIERRA JR

DATE/TIME OF TEST: 07/16/2019 17:40:42

Diagnostics Check		OK	17:37	
Air Blank		0.000	17:38	
Control Test		0.080	17:38	
Air Blank		0.000	17:39	
Subject Sample	# 1		17:39	Subject Test Refused
Breath Volume		0.000	17:39	
Air Blank		0.000	17:39	
Control Test		0.079	17:40	
Air Blank		0.000	17:40	
Diagnostics Check		OK	17:40	

Data Download Date/Time: 07/31/2019 16:39

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTTON,EDWIN,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MIDDLETON,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:46

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 07/19/2019 09:17:33

Diagnostics Check		OK	09:10
Air Blank		0.000	09:11
Control Test		0.079	09:11
Air Blank		0.000	09:11
Subject Sample	# 1	0.000	09:13
Breath Volume		1.414	09:13
Air Blank		0.000	09:13
Air Blank		0.000	09:15
Subject Sample	# 2	0.000	09:16
Breath Volume		1.574	09:16
Air Blank		0.000	09:16
Control Test		0.079	09:16
Air Blank		0.000	09:17
Diagnostics Check		OK	09:17

Data Download Date/Time: 07/31/2019 16:39

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,ERIC,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ORTIZ,BILLY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 07/20/2019 04:22:30

Diagnostics Check		OK	04:19	
Air Blank		0.000	04:19	
Control Test		0.079	04:20	
Air Blank		0.000	04:20	
Subject Sample	# 1		04:21	Subject Test Refused
Breath Volume		0.000	04:21	
Air Blank		0.000	04:21	
Control Test		0.079	04:21	
Air Blank		0.000	04:22	
Diagnostics Check		OK	04:22	

Data Download Date/Time: 07/31/2019 16:39

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVERS,DAWN,T
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: DOBO,ANTHONEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: DOBSON

DATE/TIME OF TEST: 07/21/2019 02:18:16

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:15	
Control Test		0.079	02:16	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:16	Subject Test Refused
Breath Volume		0.000	02:16	
Air Blank		0.000	02:17	
Control Test		0.078	02:17	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 07/31/2019 16:39

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: STEIN

DATE/TIME OF TEST: 07/23/2019 03:06:10

Diagnostics Check		OK	03:03	
Air Blank		0.000	03:03	
Control Test		0.079	03:04	
Air Blank		0.000	03:04	
Subject Sample	# 1		03:04	Subject Test Refused
Breath Volume		0.000	03:04	
Air Blank		0.000	03:05	
Control Test		0.079	03:05	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:06	

Data Download Date/Time: 07/31/2019 16:39

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JARROD,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NOTTINGHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: MELTON

DATE/TIME OF TEST: 07/23/2019 13:47:00

Diagnostics Check		OK	13:40
Air Blank		0.000	13:41
Control Test		0.080	13:41
Air Blank		0.000	13:42
Subject Sample	# 1	0.011	13:42
Breath Volume		2.070	13:42
Air Blank		0.000	13:43
Air Blank		0.000	13:45
Subject Sample	# 2	0.011	13:45
Breath Volume		2.140	13:45
Air Blank		0.000	13:46
Control Test		0.080	13:46
Air Blank		0.000	13:46
Diagnostics Check		OK	13:47

Data Download Date/Time: 07/31/2019 16:39

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,ERIC,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SCHMIDT,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: TWITO

DATE/TIME OF TEST: 07/25/2019 00:07:13

Diagnostics Check		OK	00:00
Air Blank		0.000	00:01
Control Test		0.079	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.144	00:02
Breath Volume		2.222	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:04
Subject Sample	# 2	0.140	00:05
Breath Volume		2.253	00:05
Air Blank		0.000	00:06
Control Test		0.078	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 07/31/2019 16:39

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 07/27/2019 21:14:06

Diagnostics Check		OK	21:11	
Air Blank		0.000	21:11	
Control Test		0.078	21:12	
Air Blank		0.000	21:12	
Subject Sample	# 1		21:12	Subject Test Refused
Breath Volume		0.000	21:12	
Air Blank		0.000	21:13	
Control Test		0.079	21:13	
Air Blank		0.000	21:14	
Diagnostics Check		OK	21:14	

Data Download Date/Time: 07/31/2019 16:39

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: OTERO,HECTOR
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 07/27/2019 23:45:24

Diagnostics Check		OK	23:42	
Air Blank		0.000	23:42	
Control Test		0.078	23:43	
Air Blank		0.000	23:43	
Subject Sample	# 1		23:44	Subject Test Refused
Breath Volume		0.000	23:44	
Air Blank		0.000	23:44	
Control Test		0.078	23:44	
Air Blank		0.000	23:45	
Diagnostics Check		OK	23:45	

Data Download Date/Time: 07/31/2019 16:39

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: DEVONISH

DATE/TIME OF TEST: 07/28/2019 05:02:24

Diagnostics Check		OK	04:59	
Air Blank		0.000	05:00	
Control Test		0.080	05:00	
Air Blank		0.000	05:00	
Subject Sample	# 1		05:01	Subject Test Refused
Breath Volume		0.000	05:01	
Air Blank		0.000	05:01	
Control Test		0.079	05:01	
Air Blank		0.000	05:02	
Diagnostics Check		OK	05:02	

Data Download Date/Time: 07/31/2019 16:39

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HUTCHINS,JENNIFER
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 06/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BLANKINSHIP

DATE/TIME OF TEST: 07/29/2019 00:42:02

Diagnostics Check		OK	00:35
Air Blank		0.000	00:36
Control Test		0.079	00:36
Air Blank		0.000	00:37
Subject Sample	# 1	0.089	00:37
Breath Volume		2.062	00:37
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.090	00:40
Breath Volume		2.210	00:40
Air Blank		0.000	00:41
Control Test		0.079	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:42

Data Download Date/Time: 07/31/2019 16:39

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORINASH,GARY,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SCHMIDT,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: SLAVER

DATE/TIME OF TEST: 07/30/2019 21:41:59

Diagnostics Check		OK	21:34
Air Blank		0.000	21:35
Control Test		0.080	21:35
Air Blank		0.000	21:36
Subject Sample	# 1	0.000	21:37
Breath Volume		1.511	21:37
Air Blank		0.000	21:37
Air Blank		0.000	21:39
Subject Sample	# 2	0.000	21:40
Breath Volume		1.812	21:40
Air Blank		0.000	21:41
Control Test		0.077	21:41
Air Blank		0.000	21:41
Diagnostics Check		OK	21:41

Data Download Date/Time: 07/31/2019 16:39

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCAZZIO,SEAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOLTZHOUSE,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: CHRISTIANSEN

DATE/TIME OF TEST: 07/01/2019 23:12:27

Diagnostics Check		OK	23:09	
Air Blank		0.000	23:10	
Control Test		0.077	23:10	
Air Blank		0.000	23:10	
Subject Sample	# 1		23:11	Subject Test Refused
Breath Volume		0.000	23:11	
Air Blank		0.000	23:11	
Control Test		0.077	23:11	
Air Blank		0.000	23:12	
Diagnostics Check		OK	23:12	

Data Download Date/Time: 07/06/2019 19:19

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DADDARIO,ASHLEY,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GRIFFIS,M
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: HARVESTER

DATE/TIME OF TEST: 07/06/2019 19:53:37

Diagnostics Check		OK	19:50	
Air Blank		0.000	19:51	
Control Test		0.077	19:51	
Air Blank		0.000	19:52	
Subject Sample	# 1		19:52	Subject Test Refused
Breath Volume		0.000	19:52	
Air Blank		0.000	19:52	
Control Test		0.077	19:53	
Air Blank		0.000	19:53	
Diagnostics Check		OK	19:53	

Data Download Date/Time: 07/31/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DADDARIO,ASHLEY,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: SAAYMAN,F
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: CORNE

DATE/TIME OF TEST: 07/11/2019 04:22:31

Diagnostics Check		OK	04:15
Air Blank		0.000	04:16
Control Test		0.076	04:16
Air Blank		0.000	04:16
Subject Sample	# 1	0.218	04:17
Breath Volume		3.796	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.225	04:20
Breath Volume		3.765	04:20
Air Blank		0.000	04:21
Control Test		0.077	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 07/31/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DADDARIO,ASHLEY,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HANEY,M
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: HEJNY

DATE/TIME OF TEST: 07/11/2019 22:51:43

Diagnostics Check		OK	22:48	
Air Blank		0.000	22:49	
Control Test		0.076	22:49	
Air Blank		0.000	22:49	
Subject Sample	# 1		22:50	Subject Test Refused
Breath Volume		0.000	22:50	
Air Blank		0.000	22:50	
Control Test		0.077	22:51	
Air Blank		0.000	22:51	
Diagnostics Check		OK	22:51	

Data Download Date/Time: 07/31/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DADDARIO,ASHLEY,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BOSWELL,A
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: WEILER

DATE/TIME OF TEST: 07/12/2019 02:41:05

Diagnostics Check		OK	02:34
Air Blank		0.000	02:35
Control Test		0.077	02:35
Air Blank		0.000	02:36
Subject Sample	# 1	0.000	02:36
Breath Volume		1.796	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.000	02:39
Breath Volume		2.328	02:39
Air Blank		0.000	02:40
Control Test		0.078	02:40
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 07/31/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,STEVEN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WARD,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 07/12/2019 23:19:36

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.077	23:12
Air Blank		0.000	23:12
Subject Sample	# 1	0.268	23:14
Breath Volume		1.753	23:14
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.255	23:17
Breath Volume		2.121	23:17
Air Blank		0.000	23:18
Control Test		0.076	23:18
Air Blank		0.000	23:19
Diagnostics Check		OK	23:19

Data Download Date/Time: 07/31/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DADDARIO,ASHLEY,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BOSTIC,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 07/14/2019 04:19:59

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.077	04:14
Air Blank		0.000	04:14
Subject Sample	# 1	0.117	04:15
Breath Volume		3.519	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.116	04:18
Breath Volume		3.699	04:18
Air Blank		0.000	04:19
Control Test		0.077	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 07/31/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERMASTER,MARY,E
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOWARD,RUSS
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 07/17/2019 16:03:54

Diagnostics Check		OK	15:57
Air Blank		0.000	15:57
Control Test		0.077	15:58
Air Blank		0.000	15:58
Subject Sample	# 1	0.197	15:59
Breath Volume		1.734	15:59
Air Blank		0.000	15:59
Air Blank		0.000	16:01
Subject Sample	# 2	0.195	16:02
Breath Volume		1.839	16:02
Air Blank		0.000	16:02
Control Test		0.077	16:03
Air Blank		0.000	16:03
Diagnostics Check		OK	16:03

Data Download Date/Time: 07/31/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WIGGLESWORTH,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: COSTELLO

DATE/TIME OF TEST: 07/18/2019 01:12:09

Diagnostics Check		OK	01:04
Air Blank		0.000	01:05
Control Test		0.077	01:05
Air Blank		0.000	01:05
Subject Sample	# 1	0.192	01:07
Breath Volume		2.996	01:07
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.188	01:10
Breath Volume		2.289	01:10
Air Blank		0.000	01:11
Control Test		0.076	01:11
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 07/31/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEVORAK,MATTHEW,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: JOHNSON,D
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: TOTTEN-DELUCCI

DATE/TIME OF TEST: 07/20/2019 22:31:11

Diagnostics Check		OK	22:28	
Air Blank		0.000	22:28	
Control Test		0.077	22:28	
Air Blank		0.000	22:29	
Subject Sample	# 1		22:29	Subject Test Refused
Breath Volume		0.000	22:29	
Air Blank		0.000	22:30	
Control Test		0.076	22:30	
Air Blank		0.000	22:31	
Diagnostics Check		OK	22:31	

Data Download Date/Time: 07/31/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,CURTIS,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CLOUD,TOD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: LUIZ

DATE/TIME OF TEST: 07/21/2019 08:21:57

Diagnostics Check		OK	08:15
Air Blank		0.000	08:16
Control Test		0.077	08:16
Air Blank		0.000	08:16
Subject Sample	# 1	0.216	08:17
Breath Volume		3.460	08:17
Air Blank		0.000	08:18
Air Blank		0.000	08:19
Subject Sample	# 2	0.215	08:20
Breath Volume		3.718	08:20
Air Blank		0.000	08:21
Control Test		0.076	08:21
Air Blank		0.000	08:21
Diagnostics Check		OK	08:21

Data Download Date/Time: 07/31/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,STEVEN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WARD,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: LAUGHLIN

DATE/TIME OF TEST: 07/22/2019 23:17:27

Diagnostics Check		OK	23:10
Air Blank		0.000	23:10
Control Test		0.077	23:11
Air Blank		0.000	23:11
Subject Sample	# 1	0.116	23:12
Breath Volume		2.824	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.121	23:15
Breath Volume		2.195	23:15
Air Blank		0.000	23:16
Control Test		0.077	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 07/31/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WIGGLESWORTH,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: SCHLENK

DATE/TIME OF TEST: 07/27/2019 01:24:58

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.076	01:19
Air Blank		0.000	01:19
Subject Sample	# 1	0.205	01:20
Breath Volume		2.253	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:22
Subject Sample	# 2	0.207	01:23
Breath Volume		2.437	01:23
Air Blank		0.000	01:24
Control Test		0.077	01:24
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 07/31/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEVORAK,M,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,KYLE
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: CORRIVEAU

DATE/TIME OF TEST: 07/30/2019 01:35:26

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:32	
Control Test		0.076	01:33	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:34	Subject Test Refused
Breath Volume		0.000	01:34	
Air Blank		0.000	01:34	
Control Test		0.077	01:34	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 07/31/2019 12:06

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: DEAS

DATE/TIME OF TEST: 07/04/2019 21:52:56

Diagnostics Check		OK	21:49	
Air Blank		0.000	21:50	
Control Test		0.076	21:50	
Air Blank		0.000	21:51	
Subject Sample	# 1		21:51	Subject Test Refused
Breath Volume		0.000	21:51	
Air Blank		0.000	21:52	
Control Test		0.076	21:52	
Air Blank		0.000	21:52	
Diagnostics Check		OK	21:52	

Data Download Date/Time: 08/23/2019 15:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 07/04/2019 22:56:42

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.076	22:51
Air Blank		0.000	22:51
Subject Sample	# 1	0.128	22:52
Breath Volume		2.820	22:52
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.126	22:55
Breath Volume		2.726	22:55
Air Blank		0.000	22:55
Control Test		0.076	22:56
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 08/23/2019 15:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CLARK,KURTIS
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/05/2019 01:36:09

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:35	
Control Test	0.074	01:35	Control Outside Tolerance
Air Blank	0.000	01:36	

Data Download Date/Time: 08/23/2019 15:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CLARK,KURTIS
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/05/2019 01:45:13

Diagnostics Check		OK	01:38
Air Blank		0.000	01:38
Control Test		0.076	01:38
Air Blank		0.000	01:39
Subject Sample	# 1	0.096	01:40
Breath Volume		1.941	01:40
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.095	01:43
Breath Volume		1.167	01:43
Air Blank		0.000	01:44
Control Test		0.076	01:44
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 08/23/2019 15:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PITTS

DATE/TIME OF TEST: 07/05/2019 02:47:13

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.076	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.166	02:42
Breath Volume		2.351	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.153	02:45
Breath Volume		1.238	02:45
Air Blank		0.000	02:46
Control Test		0.076	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 08/23/2019 15:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: STURM

DATE/TIME OF TEST: 07/06/2019 00:24:44

Diagnostics Check		OK	00:18
Air Blank		0.000	00:18
Control Test		0.077	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.086	00:20
Breath Volume		1.906	00:20
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.091	00:23
Breath Volume		2.546	00:23
Air Blank		0.000	00:23
Control Test		0.076	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 08/23/2019 15:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 07/07/2019 01:55:49

Diagnostics Check		OK	01:50	
Air Blank		0.000	01:50	
Control Test		0.076	01:51	
Air Blank		0.000	01:51	
Subject Sample	# 1	0.175	01:52	
Breath Volume		1.355	01:52	
Air Blank		0.000	01:52	
Air Blank		0.000	01:54	
Subject Sample	# 2	0.195	01:55	
Breath Volume		1.773	01:55	
Air Blank			01:55	RFI Detect
Air Blank		0.000	01:55	

Data Download Date/Time: 08/23/2019 15:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 07/07/2019 02:11:47

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.077	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.181	02:07
Breath Volume		1.750	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.192	02:10
Breath Volume		2.210	02:10
Air Blank		0.000	02:10
Control Test		0.076	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 08/23/2019 15:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 14:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: HOOPER

DATE/TIME OF TEST: 07/20/2019 15:04:09

Diagnostics Check		OK	14:57
Air Blank		0.000	14:57
Control Test		0.077	14:58
Air Blank		0.000	14:58
Subject Sample	# 1	0.235	14:59
Breath Volume		2.175	14:59
Air Blank		0.000	14:59
Air Blank		0.000	15:01
Subject Sample	# 2	0.229	15:02
Breath Volume		1.343	15:02
Air Blank		0.000	15:03
Control Test		0.076	15:03
Air Blank		0.000	15:04
Diagnostics Check		OK	15:04

Data Download Date/Time: 08/23/2019 15:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: VANLOON

DATE/TIME OF TEST: 07/23/2019 22:47:21

Diagnostics Check		OK	22:40
Air Blank		0.000	22:40
Control Test		0.075	22:41
Air Blank		0.000	22:41
Subject Sample	# 1	0.127	22:42
Breath Volume		1.296	22:42
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 2	0.135	22:45
Breath Volume		1.531	22:45
Air Blank		0.000	22:46
Control Test		0.075	22:46
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 08/23/2019 15:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WALSH,SHANNON
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 06/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 07/02/2019 04:06:41

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.080	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.137	04:02
Breath Volume		1.476	04:02
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.155	04:05
Breath Volume		2.007	04:05
Air Blank		0.000	04:05
Control Test		0.080	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 07/26/2019 16:31

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YODER,JACQUELYN,S
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: HEUSER,JOSEPH
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 06/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: YODER

DATE/TIME OF TEST: 07/11/2019 23:20:26

Diagnostics Check		OK	23:16	
Air Blank		0.000	23:16	
Control Test		0.080	23:17	
Air Blank		0.000	23:17	
Subject Sample	# 1		23:19	Subject Test Refused
Breath Volume		0.000	23:19	
Air Blank		0.000	23:19	
Control Test		0.080	23:19	
Air Blank		0.000	23:20	
Diagnostics Check		OK	23:20	

Data Download Date/Time: 07/26/2019 16:31

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WALSH,SHANNON
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 06/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: PEREZ RUIZ

DATE/TIME OF TEST: 07/21/2019 03:25:36

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.081	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.142	03:21
Breath Volume		1.667	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.142	03:24
Breath Volume		3.054	03:24
Air Blank		0.000	03:24
Control Test		0.080	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 07/26/2019 16:31

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,KEVIN,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WALSH,SHANNON
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 06/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A2
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MARTINEZ-LOPEZ

DATE/TIME OF TEST: 07/22/2019 22:42:16

Diagnostics Check		OK	22:35
Air Blank		0.000	22:36
Control Test		0.080	22:36
Air Blank		0.000	22:37
Subject Sample	# 1	0.129	22:37
Breath Volume		2.125	22:37
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.124	22:40
Breath Volume		2.480	22:40
Air Blank		0.000	22:41
Control Test		0.080	22:41
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 07/26/2019 16:31

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 07/04/2019 05:33:59

Diagnostics Check		OK	05:25
Air Blank		0.000	05:25
Control Test		0.079	05:26
Air Blank		0.000	05:26
Subject Sample	# 1	0.183	05:29
Breath Volume		2.160	05:29
Air Blank		0.000	05:30
Air Blank		0.000	05:31
Subject Sample	# 2	0.185	05:32
Breath Volume		1.886	05:32
Air Blank		0.000	05:33
Control Test		0.078	05:33
Air Blank		0.000	05:33
Diagnostics Check		OK	05:33

Data Download Date/Time: 07/10/2019 14:09

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: DACEY

DATE/TIME OF TEST: 07/05/2019 01:25:15

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.079	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.135	01:20
Breath Volume		1.914	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.130	01:23
Breath Volume		2.359	01:23
Air Blank		0.000	01:24
Control Test		0.078	01:24
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 07/10/2019 14:09

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: TORRES SANTIAGO

DATE/TIME OF TEST: 07/05/2019 03:53:24

Diagnostics Check		OK	03:46
Air Blank		0.000	03:47
Control Test		0.078	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.110	03:48
Breath Volume		2.140	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.100	03:51
Breath Volume		1.136	03:51
Air Blank		0.000	03:52
Control Test		0.079	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 07/10/2019 14:09

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: GIBAS

DATE/TIME OF TEST: 07/05/2019 04:37:24

Diagnostics Check		OK	04:25	
Air Blank		0.000	04:26	
Control Test		0.079	04:26	
Air Blank		0.000	04:27	
Subject Sample	# 1	0.213	04:30	Volume Not Met
Breath Volume		0.195	04:30	
Air Blank		0.000	04:30	
Air Blank		0.000	04:32	
Subject Sample	# 2	0.188	04:35	Volume Not Met
Breath Volume		0.304	04:35	
Air Blank		0.000	04:36	
Control Test		0.078	04:36	
Air Blank		0.000	04:37	
Diagnostics Check		OK	04:37	

Data Download Date/Time: 07/10/2019 14:09

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: GIBAS

DATE/TIME OF TEST: 07/05/2019 04:51:12

Diagnostics Check		OK	04:39	
Air Blank		0.000	04:40	
Control Test		0.077	04:40	
Air Blank		0.000	04:40	
Subject Sample	# 1	0.218	04:44	Volume Not Met
Breath Volume		0.121	04:44	
Air Blank		0.000	04:44	
Air Blank		0.000	04:46	
Subject Sample	# 2	0.198	04:49	Volume Not Met
Breath Volume		0.152	04:49	
Air Blank		0.000	04:50	
Control Test		0.076	04:50	
Air Blank		0.000	04:51	
Diagnostics Check		OK	04:51	

Data Download Date/Time: 07/10/2019 14:09

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/06/2019 04:36:13

Diagnostics Check		OK	04:30
Air Blank		0.000	04:30
Control Test		0.079	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.139	04:31
Breath Volume		2.562	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.141	04:34
Breath Volume		2.621	04:34
Air Blank		0.000	04:35
Control Test		0.079	04:35
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 07/10/2019 14:09

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: QUINTERO

DATE/TIME OF TEST: 07/06/2019 20:06:08

Diagnostics Check		OK	19:58
Air Blank		0.000	19:59
Control Test		0.079	19:59
Air Blank		0.000	20:00
Subject Sample	# 1	0.047	20:01
Breath Volume		2.261	20:01
Air Blank		0.000	20:02
Air Blank		0.000	20:04
Subject Sample	# 2	0.043	20:04
Breath Volume		1.562	20:04
Air Blank		0.000	20:05
Control Test		0.079	20:05
Air Blank		0.000	20:06
Diagnostics Check		OK	20:06

Data Download Date/Time: 07/10/2019 14:10

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 07/07/2019 01:24:08

Diagnostics Check		OK	01:17
Air Blank		0.000	01:18
Control Test		0.079	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.152	01:19
Breath Volume		2.359	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.152	01:22
Breath Volume		2.121	01:22
Air Blank		0.000	01:23
Control Test		0.079	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 07/10/2019 14:10

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 07/07/2019 05:10:00

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.079	05:04
Air Blank		0.000	05:04
Subject Sample	# 1	0.132	05:05
Breath Volume		1.468	05:05
Air Blank		0.000	05:06
Air Blank		0.000	05:07
Subject Sample	# 2	0.131	05:08
Breath Volume		2.007	05:08
Air Blank		0.000	05:09
Control Test		0.079	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:10

Data Download Date/Time: 07/10/2019 14:10

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 06/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 07/08/2019 04:20:34

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.080	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.120	04:15
Breath Volume		1.480	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.120	04:19
Breath Volume		1.203	04:19
Air Blank		0.000	04:19
Control Test		0.079	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 07/10/2019 14:10

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERAVIL,JEAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: DERAVIL,JEAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 07/06/2019 21:45:34

Diagnostics Check		OK	21:39	
Air Blank		0.000	21:39	
Control Test		0.076	21:40	
Air Blank		0.000	21:40	
Subject Sample	# 1	0.170	21:41	
Breath Volume		2.195	21:41	
Air Blank		0.000	21:41	
Air Blank		0.000	21:43	
Subject Sample	# 2	0.173	21:43	
Breath Volume		2.480	21:43	
Air Blank		0.000	21:44	
Control Test		0.073	21:44	Control Outside Tolerance
Air Blank		0.000	21:45	

Data Download Date/Time: 07/30/2019 21:45

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVIL,LUXERNE,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PREVIL,LUXERNE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: ISAACS

DATE/TIME OF TEST: 07/17/2019 02:12:53

Diagnostics Check		OK	02:06
Air Blank		0.000	02:07
Control Test		0.078	02:07
Air Blank		0.000	02:07
Subject Sample	# 1	0.290	02:08
Breath Volume		1.226	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:10
Subject Sample	# 2	0.271	02:11
Breath Volume		1.308	02:11
Air Blank		0.000	02:11
Control Test		0.076	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 07/30/2019 21:45

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVIL,LUXERNE,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PREVIL,LUXERNE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PASTRANA

DATE/TIME OF TEST: 07/21/2019 04:17:17

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.077	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.185	04:12
Breath Volume		1.937	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.185	04:15
Breath Volume		1.750	04:15
Air Blank		0.000	04:16
Control Test		0.077	04:16
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 07/30/2019 21:45

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,KYLE,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PATTERSON,KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: BUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: PAOLETTI

DATE/TIME OF TEST: 07/04/2019 20:31:49

Diagnostics Check		OK	20:24
Air Blank		0.000	20:24
Control Test		0.080	20:24
Air Blank		0.000	20:25
Subject Sample	# 1	0.170	20:26
Breath Volume		1.285	20:26
Air Blank		0.000	20:26
Air Blank		0.000	20:28
Subject Sample	# 2	0.173	20:30
Breath Volume		1.789	20:30
Air Blank		0.000	20:30
Control Test		0.080	20:31
Air Blank		0.000	20:31
Diagnostics Check		OK	20:31

Data Download Date/Time: 07/30/2019 17:46

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORROW,ELISHA,D
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: MORROW,ELISHA
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: BUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 07/04/2019 21:47:42

Diagnostics Check		OK	21:41
Air Blank		0.000	21:41
Control Test		0.079	21:42
Air Blank		0.000	21:42
Subject Sample	# 1	0.062	21:43
Breath Volume		1.863	21:43
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.056	21:46
Breath Volume		2.039	21:46
Air Blank		0.000	21:46
Control Test		0.078	21:47
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 07/30/2019 17:46

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BREVIK,LONNIE,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BREVIK,LONNIE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: BUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: GABREY

DATE/TIME OF TEST: 07/04/2019 22:40:30

Diagnostics Check		OK	22:37	
Air Blank		0.000	22:38	
Control Test		0.079	22:38	
Air Blank		0.000	22:38	
Subject Sample	# 1		22:39	Subject Test Refused
Breath Volume		0.000	22:39	
Air Blank		0.000	22:39	
Control Test		0.079	22:39	
Air Blank		0.000	22:40	
Diagnostics Check		OK	22:40	

Data Download Date/Time: 07/30/2019 17:46

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,JOSE,G
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CAPLAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: MAURY GOYO

DATE/TIME OF TEST: 07/05/2019 05:12:45

Diagnostics Check		OK	05:05
Air Blank		0.000	05:05
Control Test		0.079	05:05
Air Blank		0.000	05:06
Subject Sample	# 1	0.191	05:08
Breath Volume		2.199	05:08
Air Blank		0.000	05:08
Air Blank		0.000	05:10
Subject Sample	# 2	0.191	05:11
Breath Volume		3.667	05:11
Air Blank		0.000	05:11
Control Test		0.078	05:12
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 07/30/2019 17:46

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: RICE,MICHAEL
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 07/07/2019 08:03:08

Diagnostics Check		OK	07:56
Air Blank		0.000	07:56
Control Test		0.079	07:57
Air Blank		0.000	07:57
Subject Sample	# 1	0.038	07:58
Breath Volume		3.246	07:58
Air Blank		0.000	07:59
Air Blank		0.000	08:00
Subject Sample	# 2	0.038	08:01
Breath Volume		1.671	08:01
Air Blank		0.000	08:02
Control Test		0.080	08:02
Air Blank		0.000	08:03
Diagnostics Check		OK	08:03

Data Download Date/Time: 07/30/2019 17:46

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CARROLL,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: BUI
ARREST TIME: 15:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: AYRES

DATE/TIME OF TEST: 07/07/2019 16:58:09

Diagnostics Check		OK	16:55	
Air Blank		0.000	16:55	
Control Test		0.079	16:55	
Air Blank		0.000	16:56	
Subject Sample	# 1		16:56	Subject Test Refused
Breath Volume		0.000	16:56	
Air Blank		0.000	16:57	
Control Test		0.079	16:57	
Air Blank		0.000	16:58	
Diagnostics Check		OK	16:58	

Data Download Date/Time: 07/30/2019 17:46

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,CODY,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BROWN,CODY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 07/08/2019 00:08:46

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:05	
Control Test		0.079	00:05	
Air Blank		0.000	00:06	
Subject Sample	# 1		00:08	Sequence Aborted
Breath Volume		0.132	00:08	
Air Blank		0.000	00:08	

Data Download Date/Time: 07/30/2019 17:46

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,CODY,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BROWN,CODY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: JILLSON

DATE/TIME OF TEST: 07/08/2019 21:09:02

Diagnostics Check		OK	21:02
Air Blank		0.000	21:03
Control Test		0.079	21:03
Air Blank		0.000	21:04
Subject Sample	# 1	0.142	21:04
Breath Volume		2.074	21:04
Air Blank		0.000	21:05
Air Blank		0.000	21:07
Subject Sample	# 2	0.144	21:07
Breath Volume		2.289	21:07
Air Blank		0.000	21:08
Control Test		0.076	21:08
Air Blank		0.000	21:08
Diagnostics Check		OK	21:09

Data Download Date/Time: 07/30/2019 17:46

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CAPLAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: COWARDIN

DATE/TIME OF TEST: 07/09/2019 00:45:47

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.079	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.258	00:41
Breath Volume		3.171	00:41
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.263	00:44
Breath Volume		3.121	00:44
Air Blank		0.000	00:44
Control Test		0.078	00:45
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 07/30/2019 17:46

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CAPLAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/09/2019 03:44:35

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:33	
Control Test		0.079	03:33	
Air Blank		0.000	03:34	
Subject Sample	# 1	0.151	03:37	Volume Not Met
Breath Volume		0.015	03:37	
Air Blank		0.000	03:38	
Air Blank		0.000	03:39	
Subject Sample	# 2	0.151	03:43	Volume Not Met
Breath Volume		0.167	03:43	
Air Blank		0.000	03:43	
Control Test		0.078	03:43	
Air Blank		0.000	03:44	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 07/30/2019 17:46

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,CODY,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BROWN,CODY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: IZZO

DATE/TIME OF TEST: 07/12/2019 01:39:55

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.079	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.179	01:35
Breath Volume		4.316	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:37
Subject Sample	# 2	0.180	01:38
Breath Volume		4.714	01:38
Air Blank		0.000	01:38
Control Test		0.078	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 07/30/2019 17:46

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,ALBERT,M
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: COLLINS,ALBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: POKRANDT

DATE/TIME OF TEST: 07/12/2019 03:20:08

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:09	
Control Test		0.079	03:09	
Air Blank		0.000	03:10	
Subject Sample	# 1		03:13	No Sample Provided
Breath Volume		0.000	03:13	
Air Blank		0.000	03:13	
Air Blank		0.000	03:15	
Subject Sample	# 2		03:18	No Sample Provided
Breath Volume		0.000	03:18	
Air Blank		0.000	03:19	
Control Test		0.079	03:19	
Air Blank		0.000	03:20	
Diagnostics Check		OK	03:20	

Data Download Date/Time: 07/30/2019 17:46

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CAPLAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: ENYEART

DATE/TIME OF TEST: 07/13/2019 03:30:03

Diagnostics Check		OK	03:22
Air Blank		0.000	03:23
Control Test		0.079	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.202	03:25
Breath Volume		1.855	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.197	03:28
Breath Volume		1.644	03:28
Air Blank		0.000	03:29
Control Test		0.078	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:30

Data Download Date/Time: 07/30/2019 17:46

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,DILLAN,R
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: HUDSON,DILLAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: BUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: MULCAHY

DATE/TIME OF TEST: 07/13/2019 20:03:16

Diagnostics Check		OK	19:56
Air Blank		0.000	19:56
Control Test		0.080	19:56
Air Blank		0.000	19:57
Subject Sample	# 1	0.061	19:58
Breath Volume		2.355	19:58
Air Blank		0.000	19:58
Air Blank		0.000	20:00
Subject Sample	# 2	0.064	20:01
Breath Volume		2.769	20:01
Air Blank		0.000	20:02
Control Test		0.079	20:02
Air Blank		0.000	20:03
Diagnostics Check		OK	20:03

Data Download Date/Time: 07/30/2019 17:46

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: HUDSON,DILLAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: BUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: HEFNER

DATE/TIME OF TEST: 07/14/2019 19:11:43

Diagnostics Check		OK	19:08	
Air Blank		0.000	19:09	
Control Test		0.079	19:09	
Air Blank		0.000	19:09	
Subject Sample	# 1		19:10	Subject Test Refused
Breath Volume		0.000	19:10	
Air Blank		0.000	19:10	
Control Test		0.079	19:11	
Air Blank		0.000	19:11	
Diagnostics Check		OK	19:11	

Data Download Date/Time: 07/30/2019 17:46

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,DAVID,E
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: JACOBSON,DAVID
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: TEIXEIRA

DATE/TIME OF TEST: 07/15/2019 00:58:46

Diagnostics Check		OK	00:49
Air Blank		0.000	00:50
Control Test		0.078	00:50
Air Blank		0.000	00:51
Subject Sample	# 1	0.224	00:52
Breath Volume		1.261	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.213	00:57
Breath Volume		1.316	00:57
Air Blank		0.000	00:57
Control Test		0.078	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 07/30/2019 17:46

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGGARD,RICKY,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HAGGARD,RICKY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: KOSAR

DATE/TIME OF TEST: 07/17/2019 22:02:46

Diagnostics Check		OK	21:53	
Air Blank		0.000	21:53	
Control Test		0.079	21:53	
Air Blank		0.000	21:54	
Subject Sample	# 1		21:57	No Sample Provided
Breath Volume		0.000	21:57	
Air Blank		0.000	21:58	
Air Blank		0.000	22:00	
Subject Sample	# 2		22:01	Subject Test Refused
Breath Volume		0.000	22:01	
Air Blank		0.000	22:01	
Control Test		0.079	22:02	
Air Blank		0.000	22:02	
Diagnostics Check		OK	22:02	

Data Download Date/Time: 07/30/2019 17:47

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREGORY,AMANDA,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GREGORY,AMANDA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/21/2019 03:51:30

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.080	03:45
Air Blank		0.000	03:46
Subject Sample	# 1	0.149	03:46
Breath Volume		1.722	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.146	03:49
Breath Volume		1.800	03:49
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 07/30/2019 17:47

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,ERICK,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DAY,ERICK
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 07/22/2019 10:33:39

Diagnostics Check		OK	10:26
Air Blank		0.000	10:27
Control Test		0.079	10:27
Air Blank		0.000	10:27
Subject Sample	# 1	0.201	10:28
Breath Volume		1.210	10:28
Air Blank		0.000	10:28
Air Blank		0.000	10:30
Subject Sample	# 2	0.186	10:32
Breath Volume		1.175	10:32
Air Blank		0.000	10:32
Control Test		0.079	10:33
Air Blank		0.000	10:33
Diagnostics Check		OK	10:33

Data Download Date/Time: 07/30/2019 17:47

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: GAROFALO

DATE/TIME OF TEST: 07/22/2019 20:32:16

Diagnostics Check		OK	20:25
Air Blank		0.000	20:26
Control Test		0.080	20:26
Air Blank		0.000	20:27
Subject Sample	# 1	0.000	20:27
Breath Volume		3.628	20:27
Air Blank		0.000	20:28
Air Blank		0.000	20:30
Subject Sample	# 2	0.000	20:30
Breath Volume		2.078	20:30
Air Blank		0.000	20:31
Control Test		0.079	20:31
Air Blank		0.000	20:32
Diagnostics Check		OK	20:32

Data Download Date/Time: 07/30/2019 17:47

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEEDHAM,K,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: LEEDHAM,K
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:48

SUBJECT'S LAST NAME: PEREZ-PEREZ

DATE/TIME OF TEST: 07/26/2019 16:15:36

Diagnostics Check		OK	16:09
Air Blank		0.000	16:09
Control Test		0.079	16:10
Air Blank		0.000	16:10
Subject Sample	# 1	0.231	16:11
Breath Volume		2.460	16:11
Air Blank		0.000	16:11
Air Blank		0.000	16:13
Subject Sample	# 2	0.231	16:13
Breath Volume		1.949	16:13
Air Blank		0.000	16:14
Control Test		0.079	16:14
Air Blank		0.000	16:15
Diagnostics Check		OK	16:15

Data Download Date/Time: 07/30/2019 17:47

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOPF,GRACE-MARIE,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: ZOPF,GRACE-MARIE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/27/2019 02:40:29

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.079	02:33
Air Blank		0.000	02:34
Subject Sample	# 1	0.091	02:36
Breath Volume		1.183	02:36
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.084	02:38
Breath Volume		1.484	02:38
Air Blank		0.000	02:39
Control Test		0.078	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 07/30/2019 17:47

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,CODY,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BROWN,CODY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: CECERE

DATE/TIME OF TEST: 07/27/2019 05:36:19

Diagnostics Check		OK	05:30
Air Blank		0.000	05:30
Control Test		0.079	05:30
Air Blank		0.000	05:31
Subject Sample	# 1	0.000	05:31
Breath Volume		4.964	05:31
Air Blank		0.000	05:32
Air Blank		0.000	05:34
Subject Sample	# 2	0.000	05:34
Breath Volume		5.042	05:34
Air Blank		0.000	05:35
Control Test		0.077	05:35
Air Blank		0.000	05:36
Diagnostics Check		OK	05:36

Data Download Date/Time: 07/30/2019 17:47

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOPF,GRACE-MARIE,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: ZOPF,GRACE-MARIE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: TEETOR

DATE/TIME OF TEST: 07/27/2019 20:42:20

Diagnostics Check		OK	20:35	
Air Blank		0.000	20:35	
Control Test		0.079	20:35	
Air Blank		0.000	20:36	
Subject Sample	# 1	0.153	20:37	
Breath Volume		1.511	20:37	
Air Blank		0.000	20:38	
Air Blank		0.000	20:40	
Subject Sample	# 2	0.152	20:40	
Breath Volume		1.843	20:40	
Air Blank		0.000	20:41	
Control Test		0.071	20:41	Control Outside Tolerance
Air Blank		0.000	20:42	

Data Download Date/Time: 07/30/2019 17:47

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEEDHAM,K,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: LEEDHAM,K
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:49

SUBJECT'S LAST NAME: TATE

DATE/TIME OF TEST: 07/28/2019 14:16:20

Diagnostics Check		OK	14:10
Air Blank		0.000	14:10
Control Test		0.079	14:11
Air Blank		0.000	14:11
Subject Sample	# 1	0.046	14:11
Breath Volume		1.324	14:11
Air Blank		0.000	14:12
Air Blank		0.000	14:14
Subject Sample	# 2	0.046	14:14
Breath Volume		1.382	14:14
Air Blank		0.000	14:15
Control Test		0.079	14:15
Air Blank		0.000	14:16
Diagnostics Check		OK	14:16

Data Download Date/Time: 07/30/2019 17:47

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: ANDRES FRANCISCO

DATE/TIME OF TEST: 07/28/2019 21:45:07

Diagnostics Check		OK	21:38
Air Blank		0.000	21:38
Control Test		0.079	21:38
Air Blank		0.000	21:39
Subject Sample	# 1	0.215	21:40
Breath Volume		2.035	21:40
Air Blank		0.000	21:41
Air Blank		0.000	21:42
Subject Sample	# 2	0.214	21:43
Breath Volume		3.070	21:43
Air Blank		0.000	21:44
Control Test		0.079	21:44
Air Blank		0.000	21:45
Diagnostics Check		OK	21:45

Data Download Date/Time: 07/30/2019 17:47

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDERSON,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 07/09/2019 22:33:13

Diagnostics Check		OK	22:26
Air Blank		0.000	22:26
Control Test		0.081	22:26
Air Blank		0.000	22:27
Subject Sample	# 1	0.198	22:28
Breath Volume		1.671	22:28
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.199	22:31
Breath Volume		1.921	22:31
Air Blank		0.000	22:32
Control Test		0.077	22:32
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 07/10/2019 09:29

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: DINEHART

DATE/TIME OF TEST: 07/10/2019 03:43:06

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.080	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.131	03:38
Breath Volume		1.664	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.134	03:41
Breath Volume		2.070	03:41
Air Blank		0.000	03:42
Control Test		0.079	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 07/10/2019 09:29

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: GARCIA MENDEZ

DATE/TIME OF TEST: 07/11/2019 21:05:25

Diagnostics Check		OK	20:57
Air Blank		0.000	20:58
Control Test		0.080	20:58
Air Blank		0.000	20:59
Subject Sample	# 1	0.161	21:00
Breath Volume		1.136	21:00
Air Blank		0.000	21:01
Air Blank		0.000	21:02
Subject Sample	# 2	0.165	21:03
Breath Volume		1.285	21:03
Air Blank		0.000	21:04
Control Test		0.079	21:04
Air Blank		0.000	21:05
Diagnostics Check		OK	21:05

Data Download Date/Time: 08/06/2019 14:08

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: SHOUP

DATE/TIME OF TEST: 07/12/2019 04:11:57

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.191	04:07
Breath Volume		1.953	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:09
Subject Sample	# 2	0.192	04:10
Breath Volume		2.476	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 08/06/2019 14:08

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WEG,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: PLAGGEMEYER

DATE/TIME OF TEST: 07/12/2019 04:43:23

Diagnostics Check		OK	04:34
Air Blank		0.000	04:35
Control Test		0.080	04:35
Air Blank		0.000	04:35
Subject Sample	# 1	0.000	04:37
Breath Volume		1.355	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.000	04:42
Breath Volume		1.925	04:42
Air Blank		0.000	04:42
Control Test		0.081	04:42
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 08/06/2019 14:08

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: SILBERMAN

DATE/TIME OF TEST: 07/12/2019 19:37:26

Diagnostics Check		OK	19:30
Air Blank		0.000	19:30
Control Test		0.080	19:31
Air Blank		0.000	19:31
Subject Sample	# 1	0.243	19:32
Breath Volume		1.562	19:32
Air Blank		0.000	19:33
Air Blank		0.000	19:34
Subject Sample	# 2	0.241	19:35
Breath Volume		1.339	19:35
Air Blank		0.000	19:36
Control Test		0.079	19:36
Air Blank		0.000	19:37
Diagnostics Check		OK	19:37

Data Download Date/Time: 08/06/2019 14:08

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: FABIAN

DATE/TIME OF TEST: 07/12/2019 21:49:10

Diagnostics Check		OK	21:42
Air Blank		0.000	21:43
Control Test		0.080	21:43
Air Blank		0.000	21:44
Subject Sample	# 1	0.161	21:44
Breath Volume		1.335	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.161	21:47
Breath Volume		1.511	21:47
Air Blank		0.000	21:48
Control Test		0.079	21:48
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 08/06/2019 14:08

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VASALLO,Y
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: RUIZ-CANTU

DATE/TIME OF TEST: 07/13/2019 04:51:16

Diagnostics Check		OK	04:44
Air Blank		0.000	04:44
Control Test		0.080	04:44
Air Blank		0.000	04:45
Subject Sample	# 1	0.180	04:46
Breath Volume		1.906	04:46
Air Blank		0.000	04:47
Air Blank		0.000	04:48
Subject Sample	# 2	0.183	04:49
Breath Volume		2.144	04:49
Air Blank		0.000	04:50
Control Test		0.079	04:50
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 08/06/2019 14:08

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: PEREZ GONZALEZ

DATE/TIME OF TEST: 07/14/2019 02:39:32

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.080	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.192	02:35
Breath Volume		1.718	02:35
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.178	02:37
Breath Volume		1.449	02:37
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 08/06/2019 14:08

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: PARKS

DATE/TIME OF TEST: 07/14/2019 03:13:33

Diagnostics Check		OK	03:04	
Air Blank		0.000	03:04	
Control Test		0.080	03:05	
Air Blank		0.000	03:05	
Subject Sample	# 1	0.155	03:08	Volume Not Met
Breath Volume		0.000	03:08	
Air Blank		0.000	03:09	
Air Blank		0.000	03:11	
Subject Sample	# 2		03:12	Subject Test Refused
Breath Volume		0.000	03:12	
Air Blank		0.000	03:12	
Control Test		0.078	03:12	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 08/06/2019 14:08

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 07/14/2019 22:27:40

Diagnostics Check		OK	22:20
Air Blank		0.000	22:20
Control Test		0.080	22:21
Air Blank		0.000	22:21
Subject Sample	# 1	0.165	22:22
Breath Volume		2.464	22:22
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.160	22:26
Breath Volume		1.773	22:26
Air Blank		0.000	22:26
Control Test		0.079	22:27
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 08/06/2019 14:08

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: RAMIREZ HERRERA

DATE/TIME OF TEST: 07/18/2019 03:38:28

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.080	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.120	03:33
Breath Volume		1.312	03:33
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.129	03:36
Breath Volume		2.605	03:36
Air Blank		0.000	03:37
Control Test		0.079	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 08/06/2019 14:08

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: NAJERA

DATE/TIME OF TEST: 07/20/2019 02:28:30

Diagnostics Check		OK	02:21
Air Blank		0.000	02:22
Control Test		0.080	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.180	02:23
Breath Volume		2.648	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.181	02:26
Breath Volume		2.812	02:26
Air Blank		0.000	02:27
Control Test		0.080	02:27
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 08/06/2019 14:08

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GARCIA MARTINEZ

DATE/TIME OF TEST: 07/20/2019 03:06:39

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:51	
Control Test		0.080	02:51	
Air Blank		0.000	02:52	
Subject Sample	# 1	0.161	02:53	
Breath Volume		1.523	02:53	
Air Blank		0.000	02:54	
Air Blank		0.000	02:56	
Subject Sample	# 2	0.142	02:59	Volume Not Met
Breath Volume		0.035	02:59	
Air Blank		0.000	03:00	
Air Blank		0.000	03:01	
Subject Sample	# 3	0.146	03:04	Volume Not Met
Breath Volume		0.289	03:04	
Air Blank		0.000	03:05	
Control Test		0.077	03:06	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:06	

Data Download Date/Time: 08/06/2019 14:08

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 07/20/2019 04:22:47

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.079	04:16
Air Blank		0.000	04:16
Subject Sample	# 1	0.174	04:18
Breath Volume		2.164	04:18
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.162	04:21
Breath Volume		1.507	04:21
Air Blank		0.000	04:21
Control Test		0.079	04:22
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 08/06/2019 14:08

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: GAY

DATE/TIME OF TEST: 07/20/2019 22:51:03

Diagnostics Check		OK	22:44
Air Blank		0.000	22:45
Control Test		0.080	22:45
Air Blank		0.000	22:46
Subject Sample	# 1	0.057	22:46
Breath Volume		1.816	22:46
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.056	22:49
Breath Volume		2.242	22:49
Air Blank		0.000	22:50
Control Test		0.080	22:50
Air Blank		0.000	22:50
Diagnostics Check		OK	22:51

Data Download Date/Time: 08/06/2019 14:08

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 07/21/2019 00:42:52

Diagnostics Check		OK	00:28	
Air Blank		0.000	00:28	
Control Test		0.079	00:28	
Air Blank		0.000	00:29	
Subject Sample	# 1	0.170	00:29	
Breath Volume		1.320	00:29	
Air Blank		0.000	00:30	
Air Blank		0.000	00:32	
Subject Sample	# 2	0.179	00:35	Slope Not Level
Breath Volume		1.652	00:35	
Air Blank		0.000	00:36	
Air Blank		0.000	00:37	
Subject Sample	# 3	0.177	00:41	Volume Not Met
Breath Volume		0.925	00:41	
Air Blank		0.000	00:41	
Control Test		0.078	00:42	
Air Blank		0.000	00:42	
Diagnostics Check		OK	00:42	

Data Download Date/Time: 08/06/2019 14:08

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: OLMOS MENDIOLA

DATE/TIME OF TEST: 07/21/2019 02:56:06

Diagnostics Check		OK	02:43	
Air Blank		0.000	02:44	
Control Test		0.080	02:44	
Air Blank		0.000	02:45	
Subject Sample	# 1	0.201	02:48	Volume Not Met
Breath Volume		0.601	02:48	
Air Blank		0.000	02:49	
Air Blank		0.000	02:50	
Subject Sample	# 2	0.271	02:51	
Breath Volume		2.359	02:51	
Air Blank		0.000	02:52	
Air Blank		0.000	02:53	
Subject Sample	# 3	0.266	02:54	
Breath Volume		2.621	02:54	
Air Blank		0.000	02:55	
Control Test		0.078	02:55	
Air Blank		0.000	02:56	
Diagnostics Check		OK	02:56	

Data Download Date/Time: 08/06/2019 14:08

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULLIVAN,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/21/2019 03:39:09

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.078	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.063	03:34
Breath Volume		2.167	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:36
Subject Sample	# 2	0.060	03:37
Breath Volume		1.414	03:37
Air Blank		0.000	03:38
Control Test		0.079	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 08/06/2019 14:08

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: KAYIAN

DATE/TIME OF TEST: 07/25/2019 01:41:30

Diagnostics Check		OK	01:29	
Air Blank		0.000	01:30	
Control Test		0.081	01:30	
Air Blank		0.000	01:30	
Subject Sample	# 1	0.095	01:31	
Breath Volume		1.769	01:31	
Air Blank		0.000	01:32	
Air Blank		0.000	01:33	
Subject Sample	# 2	0.084	01:37	Volume Not Met
Breath Volume		0.968	01:37	
Air Blank		0.000	01:37	
Air Blank		0.000	01:39	
Subject Sample	# 3	0.092	01:39	
Breath Volume		1.589	01:39	
Air Blank		0.000	01:40	
Control Test		0.078	01:40	
Air Blank		0.000	01:41	
Diagnostics Check		OK	01:41	

Data Download Date/Time: 08/06/2019 14:08

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 07/27/2019 05:06:29

Diagnostics Check		OK	04:59
Air Blank		0.000	05:00
Control Test		0.080	05:00
Air Blank		0.000	05:00
Subject Sample	# 1	0.159	05:01
Breath Volume		1.757	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.152	05:04
Breath Volume		1.593	05:04
Air Blank		0.000	05:05
Control Test		0.079	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 08/06/2019 14:09

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 07/28/2019 04:43:41

Diagnostics Check		OK	04:36
Air Blank		0.000	04:37
Control Test		0.080	04:37
Air Blank		0.000	04:38
Subject Sample	# 1	0.104	04:39
Breath Volume		1.644	04:39
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.100	04:42
Breath Volume		1.417	04:42
Air Blank		0.000	04:42
Control Test		0.080	04:43
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 08/06/2019 14:09

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: DURAN LOYO

DATE/TIME OF TEST: 07/29/2019 04:40:21

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.079	04:33
Air Blank		0.000	04:33
Subject Sample	# 1	0.113	04:35
Breath Volume		1.800	04:35
Air Blank		0.000	04:36
Air Blank		0.000	04:37
Subject Sample	# 2	0.110	04:38
Breath Volume		2.156	04:38
Air Blank		0.000	04:39
Control Test		0.078	04:39
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 08/06/2019 14:09

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MATHEIS JR

DATE/TIME OF TEST: 07/30/2019 01:12:48

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.000	01:08
Breath Volume		1.691	01:08
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.000	01:11
Breath Volume		1.242	01:11
Air Blank		0.000	01:11
Control Test		0.079	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 08/06/2019 14:09

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FERGUSON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/30/2019 04:48:08

Diagnostics Check		OK	04:39	
Air Blank		0.000	04:40	
Control Test		0.080	04:40	
Air Blank		0.000	04:40	
Subject Sample	# 1		04:44	No Sample Provided
Breath Volume		0.000	04:44	
Air Blank		0.000	04:44	
Air Blank		0.000	04:46	
Subject Sample	# 2		04:46	Subject Test Refused
Breath Volume		0.000	04:46	
Air Blank		0.000	04:47	
Control Test		0.080	04:47	
Air Blank		0.000	04:48	
Diagnostics Check		OK	04:48	

Data Download Date/Time: 08/06/2019 14:09

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: ZAMBRANO VACA

DATE/TIME OF TEST: 07/30/2019 23:41:11

Diagnostics Check		OK	23:37	
Air Blank		0.000	23:38	
Control Test		0.080	23:38	
Air Blank		0.000	23:39	
Subject Sample	# 1		23:39	Subject Test Refused
Breath Volume		0.000	23:39	
Air Blank		0.000	23:40	
Control Test		0.080	23:40	
Air Blank		0.000	23:41	
Diagnostics Check		OK	23:41	

Data Download Date/Time: 08/06/2019 14:09

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BURNSIDE,CURTIS
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: POIO

DATE/TIME OF TEST: 07/03/2019 23:11:53

Diagnostics Check		OK	23:06
Air Blank		0.000	23:06
Control Test		0.079	23:06
Air Blank		0.000	23:07
Subject Sample	# 1	0.000	23:07
Breath Volume		2.238	23:07
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.000	23:10
Breath Volume		1.156	23:10
Air Blank		0.000	23:10
Control Test		0.079	23:11
Air Blank		0.000	23:11
Diagnostics Check		OK	23:11

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,JOSE,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PARISH,CLINT
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 06:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: RONDON

DATE/TIME OF TEST: 07/04/2019 07:45:14

Diagnostics Check		OK	07:42	
Air Blank		0.000	07:42	
Control Test		0.079	07:43	
Air Blank		0.000	07:43	
Subject Sample	# 1		07:43	Subject Test Refused
Breath Volume		0.000	07:43	
Air Blank		0.000	07:44	
Control Test		0.078	07:44	
Air Blank		0.000	07:45	
Diagnostics Check		OK	07:45	

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PARISH,C
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 07/04/2019 16:05:26

Diagnostics Check		OK	16:02	
Air Blank		0.000	16:03	
Control Test		0.078	16:03	
Air Blank		0.000	16:03	
Subject Sample	# 1		16:04	Subject Test Refused
Breath Volume		0.000	16:04	
Air Blank		0.000	16:04	
Control Test		0.078	16:04	
Air Blank		0.000	16:05	
Diagnostics Check		OK	16:05	

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,JOHN,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: FAHNESTOCK,ZACH
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: BUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: SPRAY

DATE/TIME OF TEST: 07/04/2019 20:25:06

Diagnostics Check		OK	20:18
Air Blank		0.000	20:19
Control Test		0.078	20:19
Air Blank		0.000	20:19
Subject Sample	# 1	0.199	20:20
Breath Volume		2.250	20:20
Air Blank		0.000	20:21
Air Blank		0.000	20:22
Subject Sample	# 2	0.200	20:23
Breath Volume		3.144	20:23
Air Blank		0.000	20:24
Control Test		0.079	20:24
Air Blank		0.000	20:25
Diagnostics Check		OK	20:25

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIJAGIC,SAMIR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ALIJAGIC,SAMIR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: DOMINGUEZ VILLARREAL

DATE/TIME OF TEST: 07/05/2019 01:45:14

Diagnostics Check		OK	01:38
Air Blank		0.000	01:38
Control Test		0.079	01:39
Air Blank		0.000	01:39
Subject Sample	# 1	0.138	01:40
Breath Volume		1.527	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:42
Subject Sample	# 2	0.140	01:43
Breath Volume		2.605	01:43
Air Blank		0.000	01:44
Control Test		0.078	01:44
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,JOSHUA,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: APONTE,JOSHUA
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 07/05/2019 02:05:34

Diagnostics Check		OK	02:03	
Air Blank		0.000	02:03	
Control Test		0.078	02:04	
Air Blank		0.000	02:04	
Subject Sample	# 1		02:04	Subject Test Refused
Breath Volume		0.000	02:04	
Air Blank			02:05	RFI Detect
Air Blank		0.000	02:05	

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,JOSHUA,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: APONTE,JOSHUA
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 07/05/2019 02:10:47

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:08	
Control Test		0.079	02:08	
Air Blank		0.000	02:08	
Subject Sample	# 1		02:09	Subject Test Refused
Breath Volume		0.000	02:09	
Air Blank		0.000	02:09	
Control Test		0.079	02:10	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOLLERY,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: DOLLERY,MICHAEL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: WIELICKI

DATE/TIME OF TEST: 07/05/2019 17:24:25

Diagnostics Check		OK	17:17
Air Blank		0.000	17:18
Control Test		0.079	17:18
Air Blank		0.000	17:19
Subject Sample	# 1	0.151	17:19
Breath Volume		3.851	17:19
Air Blank		0.000	17:20
Air Blank		0.000	17:22
Subject Sample	# 2	0.161	17:22
Breath Volume		4.031	17:22
Air Blank		0.000	17:23
Control Test		0.079	17:23
Air Blank		0.000	17:24
Diagnostics Check		OK	17:24

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,V
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: WILLIAMS,Z
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MELLEN

DATE/TIME OF TEST: 07/05/2019 22:09:12

Diagnostics Check		OK	22:01
Air Blank		0.000	22:01
Control Test		0.079	22:02
Air Blank		0.000	22:02
Subject Sample	# 1	0.130	22:03
Breath Volume		1.218	22:03
Air Blank		0.000	22:04
Air Blank		0.000	22:05
Subject Sample	# 2	0.131	22:07
Breath Volume		1.191	22:07
Air Blank		0.000	22:08
Control Test		0.081	22:08
Air Blank		0.000	22:09
Diagnostics Check		OK	22:09

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE,KRISTOPHER
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: ELLERBE

DATE/TIME OF TEST: 07/06/2019 01:35:44

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:33	
Control Test		0.079	01:33	
Air Blank		0.000	01:34	
Subject Sample	# 1		01:34	Subject Test Refused
Breath Volume		0.000	01:34	
Air Blank		0.000	01:34	
Control Test		0.079	01:35	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,NICHOLAS,L
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: PEREZ,NICHOLAS
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 07/06/2019 02:12:36

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.079	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.074	02:07
Breath Volume		1.183	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.077	02:11
Breath Volume		1.925	02:11
Air Blank		0.000	02:11
Control Test		0.079	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,JOSHUA,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: CLARK,KURTIS
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SANDOVAL

DATE/TIME OF TEST: 07/06/2019 23:49:02

Diagnostics Check		OK	23:46	
Air Blank		0.000	23:46	
Control Test		0.079	23:46	
Air Blank		0.000	23:47	
Subject Sample	# 1		23:47	Subject Test Refused
Breath Volume		0.000	23:47	
Air Blank		0.000	23:48	
Control Test		0.078	23:48	
Air Blank		0.000	23:48	
Diagnostics Check		OK	23:49	

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMMERS,HAROLD,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 11:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: SUMMERS

DATE/TIME OF TEST: 07/07/2019 00:50:56

Diagnostics Check		OK	00:48	
Air Blank		0.000	00:48	
Control Test		0.079	00:48	
Air Blank		0.000	00:49	
Subject Sample	# 1		00:49	Subject Test Refused
Breath Volume		0.000	00:49	
Air Blank		0.000	00:50	
Control Test		0.078	00:50	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: TAVERAS

DATE/TIME OF TEST: 07/07/2019 04:45:06

Diagnostics Check		OK	04:38
Air Blank		0.000	04:38
Control Test		0.078	04:38
Air Blank		0.000	04:39
Subject Sample	# 1	0.157	04:40
Breath Volume		3.167	04:40
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.158	04:43
Breath Volume		2.082	04:43
Air Blank		0.000	04:44
Control Test		0.079	04:44
Air Blank		0.000	04:45
Diagnostics Check		OK	04:45

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,WILLIAM,S
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MCKENZIE,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: LADY

DATE/TIME OF TEST: 07/07/2019 17:54:04

Diagnostics Check		OK	17:46
Air Blank		0.000	17:46
Control Test		0.078	17:46
Air Blank		0.000	17:47
Subject Sample	# 1	0.084	17:48
Breath Volume		1.179	17:48
Air Blank		0.000	17:49
Air Blank		0.000	17:51
Subject Sample	# 2	0.086	17:52
Breath Volume		2.085	17:52
Air Blank		0.000	17:53
Control Test		0.079	17:53
Air Blank		0.000	17:53
Diagnostics Check		OK	17:54

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIJAGIC,SAMIR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ALIJAGIC,SAMIR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 07/10/2019 04:33:53

Diagnostics Check		OK	04:27
Air Blank		0.000	04:27
Control Test		0.079	04:28
Air Blank		0.000	04:28
Subject Sample	# 1	0.163	04:29
Breath Volume		3.894	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:31
Subject Sample	# 2	0.156	04:32
Breath Volume		3.312	04:32
Air Blank		0.000	04:32
Control Test		0.077	04:33
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILLESPIE,ANDREW,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 07/12/2019 21:52:20

Diagnostics Check		OK	21:45
Air Blank		0.000	21:46
Control Test		0.078	21:46
Air Blank		0.000	21:46
Subject Sample	# 1	0.142	21:47
Breath Volume		2.125	21:47
Air Blank		0.000	21:48
Air Blank		0.000	21:50
Subject Sample	# 2	0.147	21:50
Breath Volume		2.210	21:50
Air Blank		0.000	21:51
Control Test		0.080	21:51
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WALSH,SHANNON
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: PRESCOTT

DATE/TIME OF TEST: 07/14/2019 03:43:42

Diagnostics Check		OK	03:36
Air Blank		0.000	03:37
Control Test		0.079	03:37
Air Blank		0.000	03:38
Subject Sample	# 1	0.186	03:38
Breath Volume		2.093	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.181	03:42
Breath Volume		2.011	03:42
Air Blank		0.000	03:42
Control Test		0.079	03:43
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAJEUNESSE,J,
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: WRAY,S
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 07/15/2019 20:45:20

Diagnostics Check		OK	20:41	
Air Blank		0.000	20:42	
Control Test		0.079	20:42	
Air Blank		0.000	20:42	
Subject Sample	# 1		20:43	Subject Test Refused
Breath Volume		0.000	20:43	
Air Blank		0.000	20:44	
Control Test		0.078	20:44	
Air Blank		0.000	20:45	
Diagnostics Check		OK	20:45	

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADE,CODY,L
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: WADE,CODY
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 06:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/17/2019 07:42:13

Diagnostics Check		OK	07:39	
Air Blank		0.000	07:39	
Control Test		0.079	07:40	
Air Blank		0.000	07:40	
Subject Sample	# 1		07:40	Subject Test Refused
Breath Volume		0.000	07:40	
Air Blank		0.000	07:41	
Control Test		0.079	07:41	
Air Blank		0.000	07:42	
Diagnostics Check		OK	07:42	

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: JOHNS,BILLY
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:27

SUBJECT'S LAST NAME: ALI

DATE/TIME OF TEST: 07/18/2019 17:14:32

Diagnostics Check		OK	17:05	
Air Blank		0.000	17:06	
Control Test		0.079	17:06	
Air Blank		0.000	17:07	
Subject Sample	# 1	0.256	17:10	Volume Not Met
Breath Volume		0.035	17:10	
Air Blank		0.000	17:10	
Air Blank		0.000	17:12	
Subject Sample	# 2		17:13	Subject Test Refused
Breath Volume		0.000	17:13	
Air Blank		0.000	17:13	
Control Test		0.077	17:13	
Air Blank		0.000	17:14	
Diagnostics Check		OK	17:14	

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPRAUN,ERICK,L
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: DUBI,YZIEL
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 06:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: ABRAHAM

DATE/TIME OF TEST: 07/20/2019 07:40:31

Diagnostics Check		OK	07:33
Air Blank		0.000	07:33
Control Test		0.079	07:33
Air Blank		0.000	07:34
Subject Sample	# 1	0.221	07:35
Breath Volume		2.433	07:35
Air Blank		0.000	07:36
Air Blank		0.000	07:38
Subject Sample	# 2	0.223	07:38
Breath Volume		1.871	07:38
Air Blank		0.000	07:39
Control Test		0.079	07:39
Air Blank		0.000	07:40
Diagnostics Check		OK	07:40

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIJAGIC,SAMIR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ALIJAGIC,SAMIR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 06:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 07/20/2019 08:07:10

Diagnostics Check		OK	08:00	
Air Blank		0.000	08:01	
Control Test		0.078	08:01	
Air Blank		0.000	08:01	
Subject Sample	# 1	0.222	08:05	Volume Not Met
Breath Volume		0.140	08:05	
Air Blank			08:05	RFI Detect
Air Blank		0.000	08:06	
Control Test		0.080	08:06	
Air Blank		0.000	08:07	
Diagnostics Check		OK	08:07	

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIJAGIC,SAMIR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ALIJAGIC,SAMIR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 06:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 07/20/2019 08:11:59

Diagnostics Check	OK	08:10	
Air Blank	0.000	08:11	
Control Test	0.026	08:11	RFI Detect
Air Blank	0.000	08:11	

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIJAGIC,SAMIR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ALIJAGIC,SAMIR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 06:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 07/20/2019 08:24:23

Diagnostics Check		OK	08:13	
Air Blank		0.000	08:13	
Control Test		0.078	08:14	
Air Blank		0.000	08:14	
Subject Sample	# 1	0.200	08:17	Volume Not Met
Breath Volume		0.062	08:17	
Air Blank		0.000	08:18	
Air Blank		0.000	08:20	
Subject Sample	# 2		08:23	No Sample Provided
Breath Volume		0.000	08:23	
Air Blank			08:23	Sequence Aborted
Air Blank		0.000	08:24	

Data Download Date/Time: 08/23/2019 16:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,JOHN,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: GONZALEZ,LUIS
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 07/24/2019 02:24:39

Diagnostics Check		OK	02:11	
Air Blank		0.000	02:12	
Control Test		0.078	02:12	
Air Blank		0.000	02:13	
Subject Sample	# 1	0.167	02:14	
Breath Volume		1.804	02:14	
Air Blank		0.000	02:14	
Air Blank		0.000	02:16	
Subject Sample	# 2	0.166	02:19	Volume Not Met
Breath Volume		0.035	02:19	
Air Blank		0.000	02:20	
Air Blank		0.000	02:22	
Subject Sample	# 3	0.171	02:23	
Breath Volume		1.261	02:23	
Air Blank		0.000	02:23	
Control Test		0.077	02:24	
Air Blank		0.000	02:24	
Diagnostics Check		OK	02:24	

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,JOSHUA,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: APONTE,JOSHUA
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 14:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: DENNISON

DATE/TIME OF TEST: 07/27/2019 15:17:37

Diagnostics Check		OK	15:15	
Air Blank		0.000	15:15	
Control Test		0.079	15:15	
Air Blank		0.000	15:16	
Subject Sample	# 1	0.000	15:16	
Breath Volume		1.472	15:16	
Air Blank			15:17	RFI Detect
Air Blank			15:17	RFI Detect
Air Blank		0.000	15:17	

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,JOSHUA,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: APONTE,JOSHUA
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 14:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: DENNISON

DATE/TIME OF TEST: 07/27/2019 15:26:00

Diagnostics Check		OK	15:19
Air Blank		0.000	15:20
Control Test		0.079	15:20
Air Blank		0.000	15:21
Subject Sample	# 1	0.000	15:21
Breath Volume		2.207	15:21
Air Blank		0.000	15:22
Air Blank		0.000	15:24
Subject Sample	# 2	0.000	15:24
Breath Volume		1.738	15:24
Air Blank		0.000	15:25
Control Test		0.079	15:25
Air Blank		0.000	15:25
Diagnostics Check		OK	15:26

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,JOHN,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: LOPEZ,BRITTANY
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 07/27/2019 22:46:39

Diagnostics Check		OK	22:39
Air Blank		0.000	22:40
Control Test		0.079	22:40
Air Blank		0.000	22:41
Subject Sample	# 1	0.179	22:42
Breath Volume		3.328	22:42
Air Blank		0.000	22:42
Air Blank		0.000	22:44
Subject Sample	# 2	0.179	22:45
Breath Volume		3.539	22:45
Air Blank		0.000	22:45
Control Test		0.079	22:46
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,JOHN,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: WESTENDORF,JORDAN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 07/28/2019 04:51:43

Diagnostics Check		OK	04:45
Air Blank		0.000	04:45
Control Test		0.078	04:46
Air Blank		0.000	04:46
Subject Sample	# 1	0.197	04:47
Breath Volume		2.386	04:47
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.209	04:50
Breath Volume		3.097	04:50
Air Blank		0.000	04:50
Control Test		0.080	04:51
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,JOHN,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: GONZALEZ,LUIS
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: MARCY

DATE/TIME OF TEST: 07/29/2019 01:21:49

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.078	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.282	01:16
Breath Volume		1.566	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.271	01:20
Breath Volume		2.035	01:20
Air Blank		0.000	01:20
Control Test		0.079	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: GONZALEZ,LUIS
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 07/31/2019 21:58:14

Diagnostics Check		OK	21:45	
Air Blank		0.000	21:46	
Control Test		0.079	21:46	
Air Blank		0.000	21:47	
Subject Sample	# 1	0.237	21:50	Volume Not Met
Breath Volume		0.367	21:50	
Air Blank		0.000	21:51	
Air Blank		0.000	21:52	
Subject Sample	# 2	0.251	21:53	
Breath Volume		1.859	21:53	
Air Blank		0.000	21:54	
Air Blank		0.000	21:55	
Subject Sample	# 3	0.239	21:56	
Breath Volume		1.843	21:56	
Air Blank		0.000	21:57	
Control Test		0.078	21:57	
Air Blank		0.000	21:58	
Diagnostics Check		OK	21:58	

Data Download Date/Time: 08/23/2019 16:48

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WURST,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WURST,JOSHUA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: MCDONOUGH

DATE/TIME OF TEST: 07/04/2019 05:55:25

Diagnostics Check		OK	05:48
Air Blank		0.000	05:49
Control Test		0.080	05:49
Air Blank		0.000	05:50
Subject Sample	# 1	0.088	05:50
Breath Volume		1.816	05:50
Air Blank		0.000	05:51
Air Blank		0.000	05:53
Subject Sample	# 2	0.082	05:53
Breath Volume		1.449	05:53
Air Blank		0.000	05:54
Control Test		0.079	05:54
Air Blank		0.000	05:55
Diagnostics Check		OK	05:55

Data Download Date/Time: 08/04/2019 21:00

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 07/04/2019 20:45:11

Diagnostics Check		OK	20:38
Air Blank		0.000	20:39
Control Test		0.079	20:39
Air Blank		0.000	20:40
Subject Sample	# 1	0.206	20:40
Breath Volume		1.632	20:40
Air Blank		0.000	20:41
Air Blank		0.000	20:43
Subject Sample	# 2	0.221	20:43
Breath Volume		1.507	20:43
Air Blank		0.000	20:44
Control Test		0.079	20:44
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 08/04/2019 21:00

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WURST,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WURST,JOSHUA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: WILKINSON

DATE/TIME OF TEST: 07/09/2019 06:02:46

Diagnostics Check		OK	05:56
Air Blank		0.000	05:56
Control Test		0.080	05:57
Air Blank		0.000	05:57
Subject Sample	# 1	0.179	05:58
Breath Volume		1.281	05:58
Air Blank		0.000	05:58
Air Blank		0.000	06:00
Subject Sample	# 2	0.181	06:01
Breath Volume		1.550	06:01
Air Blank		0.000	06:01
Control Test		0.079	06:02
Air Blank		0.000	06:02
Diagnostics Check		OK	06:02

Data Download Date/Time: 08/04/2019 21:00

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MONTANO

DATE/TIME OF TEST: 07/14/2019 22:32:28

Diagnostics Check		OK	22:25
Air Blank		0.000	22:25
Control Test		0.080	22:26
Air Blank		0.000	22:26
Subject Sample	# 1	0.178	22:27
Breath Volume		1.339	22:27
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.178	22:31
Breath Volume		1.687	22:31
Air Blank		0.000	22:31
Control Test		0.080	22:31
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 08/04/2019 21:00

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/15/2019 01:59:14

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.078	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.265	01:52
Breath Volume		1.121	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.267	01:57
Breath Volume		2.082	01:57
Air Blank		0.000	01:58
Control Test		0.079	01:58
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 08/04/2019 21:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: DUPREE

DATE/TIME OF TEST: 07/22/2019 01:49:47

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.081	01:44
Air Blank		0.000	01:44
Subject Sample	# 1	0.135	01:45
Breath Volume		3.441	01:45
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.134	01:48
Breath Volume		3.117	01:48
Air Blank		0.000	01:48
Control Test		0.080	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 08/04/2019 21:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: SHARP

DATE/TIME OF TEST: 07/26/2019 23:24:28

Diagnostics Check		OK	23:18
Air Blank		0.000	23:18
Control Test		0.080	23:18
Air Blank		0.000	23:19
Subject Sample	# 1	0.157	23:20
Breath Volume		2.031	23:20
Air Blank		0.000	23:20
Air Blank		0.000	23:22
Subject Sample	# 2	0.161	23:23
Breath Volume		1.726	23:23
Air Blank		0.000	23:23
Control Test		0.080	23:23
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 08/04/2019 21:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: STAPLES

DATE/TIME OF TEST: 07/28/2019 03:47:24

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.080	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.160	03:42
Breath Volume		1.941	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.178	03:45
Breath Volume		2.148	03:45
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 08/04/2019 21:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/28/2019 23:17:14

Diagnostics Check		OK	23:09
Air Blank		0.000	23:10
Control Test		0.080	23:10
Air Blank		0.000	23:10
Subject Sample	# 1	0.185	23:12
Breath Volume		1.613	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.188	23:15
Breath Volume		1.175	23:15
Air Blank		0.000	23:16
Control Test		0.079	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 08/04/2019 21:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGOWAN,SAVANNAH,K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DURHAM,K
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:47

SUBJECT'S LAST NAME: CASTELLANOS

DATE/TIME OF TEST: 07/02/2019 12:20:23

Diagnostics Check		OK	12:14
Air Blank		0.000	12:14
Control Test		0.079	12:15
Air Blank		0.000	12:15
Subject Sample	# 1	0.000	12:16
Breath Volume		1.417	12:16
Air Blank		0.000	12:16
Air Blank		0.000	12:18
Subject Sample	# 2	0.000	12:19
Breath Volume		2.539	12:19
Air Blank		0.000	12:19
Control Test		0.080	12:19
Air Blank		0.000	12:20
Diagnostics Check		OK	12:20

Data Download Date/Time: 07/19/2019 10:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFY,J,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOWE,A
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/02/2019 22:41:49

Diagnostics Check		OK	22:35
Air Blank		0.000	22:36
Control Test		0.080	22:36
Air Blank		0.000	22:36
Subject Sample	# 1	0.135	22:37
Breath Volume		2.519	22:37
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.142	22:40
Breath Volume		2.457	22:40
Air Blank		0.000	22:40
Control Test		0.079	22:41
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 07/19/2019 10:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILLAK,MICHAEL,O
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GUGLIOTTA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: LAZO

DATE/TIME OF TEST: 07/04/2019 00:33:00

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.080	00:26
Air Blank		0.000	00:26
Subject Sample	# 1	0.226	00:28
Breath Volume		2.117	00:28
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.213	00:31
Breath Volume		1.750	00:31
Air Blank		0.000	00:32
Control Test		0.079	00:32
Air Blank		0.000	00:32
Diagnostics Check		OK	00:33

Data Download Date/Time: 07/19/2019 10:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,CHEYANNE,K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ROSS,CHEYANNE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 07/04/2019 03:33:03

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.080	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.175	03:28
Breath Volume		2.925	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:30
Subject Sample	# 2	0.173	03:31
Breath Volume		2.527	03:31
Air Blank		0.000	03:32
Control Test		0.079	03:32
Air Blank		0.000	03:32
Diagnostics Check		OK	03:33

Data Download Date/Time: 07/19/2019 10:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SHIPPY,DYLAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: FLOWER

DATE/TIME OF TEST: 07/06/2019 01:04:31

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.079	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.193	01:00
Breath Volume		2.496	01:00
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.195	01:03
Breath Volume		1.750	01:03
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 07/19/2019 10:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MEDINA-DAMICO

DATE/TIME OF TEST: 07/06/2019 03:16:13

Diagnostics Check		OK	03:09
Air Blank		0.000	03:10
Control Test		0.080	03:10
Air Blank		0.000	03:10
Subject Sample	# 1	0.181	03:11
Breath Volume		2.777	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.171	03:14
Breath Volume		1.960	03:14
Air Blank		0.000	03:15
Control Test		0.078	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 07/19/2019 10:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FESSEL,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: MAYERS

DATE/TIME OF TEST: 07/06/2019 22:03:11

Diagnostics Check		OK	21:56
Air Blank		0.000	21:57
Control Test		0.078	21:57
Air Blank		0.000	21:57
Subject Sample	# 1	0.182	21:58
Breath Volume		1.281	21:58
Air Blank		0.000	21:59
Air Blank		0.000	22:00
Subject Sample	# 2	0.174	22:01
Breath Volume		2.402	22:01
Air Blank		0.000	22:02
Control Test		0.079	22:02
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 07/19/2019 10:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FESSEL,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MADURO,VICTOR
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: BAUTISTA

DATE/TIME OF TEST: 07/07/2019 00:41:40

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:35	
Control Test		0.078	00:35	
Air Blank		0.000	00:36	
Subject Sample	# 1	0.000	00:36	
Breath Volume		2.746	00:36	
Air Blank		0.000	00:37	
Air Blank		0.000	00:39	
Subject Sample	# 2	0.000	00:39	
Breath Volume		2.570	00:39	
Air Blank		0.000	00:40	
Control Test		0.078	00:40	
Air Blank		0.000	00:41	
Diagnostics Check		Fail	00:41	Temperature Regulation Test
Air Blank		0.000	00:41	

Data Download Date/Time: 07/19/2019 10:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FESSEL,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MADURO,VICTOR
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: BAUTISTA

DATE/TIME OF TEST: 07/07/2019 01:13:14

Diagnostics Check		OK	01:06
Air Blank		0.000	01:07
Control Test		0.078	01:07
Air Blank		0.000	01:07
Subject Sample	# 1	0.000	01:09
Breath Volume		1.390	01:09
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.000	01:11
Breath Volume		2.042	01:11
Air Blank		0.000	01:12
Control Test		0.078	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 07/19/2019 10:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOWE,ALEXANDER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: VADNEY

DATE/TIME OF TEST: 07/07/2019 01:49:00

Diagnostics Check		OK	01:44	
Air Blank		0.000	01:45	
Control Test		0.078	01:45	
Air Blank		0.000	01:46	
Subject Sample	# 1		01:47	Subject Test Refused
Breath Volume		0.000	01:47	
Air Blank		0.000	01:48	
Control Test		0.079	01:48	
Air Blank		0.000	01:48	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 07/19/2019 10:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FESSEL,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ROODY,HEATH
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 07/07/2019 02:04:41

Diagnostics Check		OK	01:51	
Air Blank		0.000	01:51	
Control Test		0.079	01:52	
Air Blank		0.000	01:52	
Subject Sample	# 1	0.220	01:54	
Breath Volume		1.812	01:54	
Air Blank		0.000	01:54	
Air Blank		0.000	01:56	
Subject Sample	# 2	0.212	01:59	Volume Not Met
Breath Volume		0.097	01:59	
Air Blank		0.000	02:00	
Air Blank		0.000	02:02	
Subject Sample	# 3	0.221	02:03	
Breath Volume		2.460	02:03	
Air Blank		0.000	02:03	
Control Test		0.077	02:04	
Air Blank		0.000	02:04	
Diagnostics Check		OK	02:04	

Data Download Date/Time: 07/19/2019 10:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FESSEL,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: NERONI,KRISTI
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/07/2019 02:52:31

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.078	02:47
Air Blank		0.000	02:47
Subject Sample	# 1	0.035	02:48
Breath Volume		1.589	02:48
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.037	02:51
Breath Volume		2.023	02:51
Air Blank		0.000	02:51
Control Test		0.077	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 07/19/2019 10:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUGIEL,ADRIAN,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SUPER,MICHAEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: LIMATOLA

DATE/TIME OF TEST: 07/12/2019 21:21:27

Diagnostics Check		OK	21:14
Air Blank		0.000	21:14
Control Test		0.079	21:15
Air Blank		0.000	21:15
Subject Sample	# 1	0.000	21:16
Breath Volume		1.660	21:16
Air Blank		0.000	21:17
Air Blank		0.000	21:19
Subject Sample	# 2	0.000	21:20
Breath Volume		2.179	21:20
Air Blank		0.000	21:20
Control Test		0.079	21:20
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 07/19/2019 10:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGH,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CHESNUT,J
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: AGUILAR RIVERA

DATE/TIME OF TEST: 07/12/2019 23:10:21

Diagnostics Check		OK	23:03
Air Blank		0.000	23:04
Control Test		0.079	23:04
Air Blank		0.000	23:05
Subject Sample	# 1	0.193	23:05
Breath Volume		1.625	23:05
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.195	23:08
Breath Volume		1.503	23:08
Air Blank		0.000	23:09
Control Test		0.079	23:09
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 07/19/2019 10:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGH,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: STEMPEL,T
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: KARCZEWSKI

DATE/TIME OF TEST: 07/15/2019 00:48:44

Diagnostics Check		OK	00:40	
Air Blank		0.000	00:40	
Control Test		0.079	00:41	
Air Blank		0.000	00:41	
Subject Sample	# 1		00:44	No Sample Provided
Breath Volume		0.000	00:44	
Air Blank		0.000	00:45	
Air Blank		0.000	00:47	
Subject Sample	# 2		00:47	Subject Test Refused
Breath Volume		0.000	00:47	
Air Blank		0.000	00:47	
Control Test		0.079	00:48	
Air Blank		0.000	00:48	
Diagnostics Check		OK	00:48	

Data Download Date/Time: 07/19/2019 10:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOWE,ALEXANDER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: HEPNER

DATE/TIME OF TEST: 07/16/2019 23:42:20

Diagnostics Check		OK	23:35
Air Blank		0.000	23:36
Control Test		0.081	23:36
Air Blank		0.000	23:37
Subject Sample	# 1	0.055	23:37
Breath Volume		2.984	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.054	23:40
Breath Volume		2.941	23:40
Air Blank		0.000	23:41
Control Test		0.081	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 07/19/2019 10:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SHIPPY,DYLAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: WELTER

DATE/TIME OF TEST: 07/17/2019 01:01:48

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.199	00:57
Breath Volume		2.527	00:57
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.200	01:00
Breath Volume		2.527	01:00
Air Blank		0.000	01:00
Control Test		0.080	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 07/19/2019 10:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: FORTIN

DATE/TIME OF TEST: 07/18/2019 03:44:55

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.080	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.178	03:40
Breath Volume		1.964	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.179	03:43
Breath Volume		2.027	03:43
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 07/19/2019 10:01

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOWE,ALEXANDER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: MAC

DATE/TIME OF TEST: 07/19/2019 22:53:50

Diagnostics Check		OK	22:47
Air Blank		0.000	22:48
Control Test		0.080	22:48
Air Blank		0.000	22:48
Subject Sample	# 1	0.000	22:49
Breath Volume		3.273	22:49
Air Blank		0.000	22:49
Air Blank		0.000	22:51
Subject Sample	# 2	0.000	22:52
Breath Volume		3.875	22:52
Air Blank		0.000	22:52
Control Test		0.080	22:53
Air Blank		0.000	22:53
Diagnostics Check		OK	22:53

Data Download Date/Time: 11/16/2019 12:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GRAY,DANNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: AMEZAGA

DATE/TIME OF TEST: 07/20/2019 03:32:40

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:26	
Control Test		0.080	03:26	
Air Blank		0.000	03:26	
Subject Sample	# 1	0.086	03:27	
Breath Volume		1.441	03:27	
Air Blank		0.000	03:27	
Air Blank		0.000	03:29	
Subject Sample	# 2	0.085	03:30	
Breath Volume		1.125	03:30	
Air Blank		0.000	03:31	
Control Test		0.079	03:31	
Air Blank		0.000	03:32	
Diagnostics Check		Fail	03:32	Temperature Regulation Test
Air Blank		0.000	03:32	

Data Download Date/Time: 11/16/2019 12:57

Printed: 08 January 2020

COBRA ©1997-2014 CMI, Inc.