

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: HANDY

DATE/TIME OF TEST: 01/11/2020 03:28:58

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.079	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.169	03:23
Breath Volume		1.265	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.157	03:27
Breath Volume		1.203	03:27
Air Blank		0.000	03:28
Control Test		0.078	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 02/07/2020 19:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: BERNAL PEREZ

DATE/TIME OF TEST: 01/29/2020 21:58:20

Diagnostics Check	OK	21:49	
Air Blank	0.000	21:50	
Control Test	0.080	21:50	
Air Blank	0.000	21:51	
Subject Sample	# 1	21:54	No Sample Provided
Breath Volume	0.000	21:54	
Air Blank	0.000	21:54	
Air Blank	0.000	21:56	
Subject Sample	# 2	21:56	Subject Test Refused
Breath Volume	0.000	21:56	
Air Blank	0.000	21:57	
Control Test	0.081	21:57	
Air Blank	0.000	21:58	
Diagnostics Check	OK	21:58	

Data Download Date/Time: 01/30/2020 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 12/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: GRAVES

DATE/TIME OF TEST: 01/25/2020 01:28:34

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.078	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.075	01:24
Breath Volume	2.792	01:24
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.073	01:27
Breath Volume	2.945	01:27
Air Blank	0.000	01:27
Control Test	0.078	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 01/28/2020 22:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 12/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: GRANISON

DATE/TIME OF TEST: 01/25/2020 04:01:29

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.078	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.125	03:56
Breath Volume	2.671	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.119	03:59
Breath Volume	1.484	03:59
Air Blank	0.000	04:00
Control Test	0.078	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 01/28/2020 22:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ISAKSON

DATE/TIME OF TEST: 01/29/2020 01:53:41

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.079	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.215	01:49
Breath Volume	1.980	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.214	01:52
Breath Volume	1.390	01:52
Air Blank	0.000	01:52
Control Test	0.079	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 03/29/2020 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAOUNI,K,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: SHEHAN,K
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 01/25/2020 02:13:05

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.081	02:07
Air Blank		0.000	02:07
Subject Sample	# 1	0.247	02:08
Breath Volume		1.390	02:08
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.234	02:11
Breath Volume		1.691	02:11
Air Blank		0.000	02:12
Control Test		0.079	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:13

Data Download Date/Time: 01/29/2020 08:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: ANDERSON,K
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: GERMANY

DATE/TIME OF TEST: 01/01/2020 03:38:15

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:23	
Control Test		0.081	03:23	
Air Blank		0.000	03:24	
Subject Sample	# 1	0.148	03:26	
Breath Volume		1.191	03:26	
Air Blank		0.000	03:27	
Air Blank		0.000	03:29	
Subject Sample	# 2	0.139	03:32	Volume Not Met
Breath Volume		0.152	03:32	
Air Blank		0.000	03:33	
Air Blank		0.000	03:35	
Subject Sample	# 3	0.142	03:36	
Breath Volume		1.128	03:36	
Air Blank		0.000	03:37	
Control Test		0.078	03:37	
Air Blank		0.000	03:38	
Diagnostics Check		OK	03:38	

Data Download Date/Time: 01/14/2020 18:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COREY,GAVIN,T
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CORONILLA,DANIEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 01/03/2020 03:45:38

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:43	
Control Test	0.080	03:43	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:44	
Control Test	0.080	03:45	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 01/14/2020 18:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COREY,GAVIN,T
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CORONILLA,DANIEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 01/11/2020 04:22:29

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:13	
Control Test	0.081	04:13	
Air Blank	0.000	04:14	
Subject Sample # 1	0.248	04:15	
Breath Volume	1.207	04:15	
Air Blank	0.000	04:16	
Air Blank	0.000	04:18	
Subject Sample # 2	0.228	04:21	Volume Not Met
Breath Volume	0.015	04:21	
Air Blank		04:22	Purge Fail
Air Blank	0.000	04:22	

Data Download Date/Time: 01/14/2020 18:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COREY,GAVIN,T
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CORONILLA,DANIEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 01/11/2020 04:54:03

Diagnostics Check		OK	04:45
Air Blank		0.000	04:46
Control Test		0.078	04:46
Air Blank		0.000	04:46
Subject Sample	# 1	0.244	04:49
Breath Volume		1.222	04:49
Air Blank		0.000	04:50
Air Blank		0.000	04:51
Subject Sample	# 2	0.232	04:52
Breath Volume		1.472	04:52
Air Blank		0.000	04:53
Control Test		0.077	04:53
Air Blank		0.000	04:53
Diagnostics Check		OK	04:54

Data Download Date/Time: 01/14/2020 18:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AFONSO, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CORONILLA, DANIEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 01/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: LEON TREJO

DATE/TIME OF TEST: 01/16/2020 00:24:15

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:21	
Control Test	0.081	00:22	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:22	Subject Test Refused
Breath Volume	0.000	00:22	
Air Blank	0.000	00:23	
Control Test	0.080	00:23	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 01/29/2020 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AFONSO,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: AFONSO,MICHAEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 01/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: VASQUEZ-MENDEZ

DATE/TIME OF TEST: 01/18/2020 03:03:38

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.081	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.178	02:58
Breath Volume	1.902	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.174	03:01
Breath Volume	2.015	03:01
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 01/29/2020 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COREY,GAVIN,T
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CORONILLA,DANIEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 01/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 01/25/2020 01:12:44

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:10	
Control Test	0.080	01:10	
Air Blank	0.000	01:11	
Subject Sample	# 1	01:11	Subject Test Refused
Breath Volume	0.000	01:11	
Air Blank	0.000	01:11	
Control Test	0.080	01:12	
Air Blank	0.000	01:12	
Diagnostics Check	OK	01:12	

Data Download Date/Time: 01/29/2020 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORD,JOSHUA,N
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LORD,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 01/18/2020 02:38:56

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:25	
Control Test	0.079	02:25	
Air Blank	0.000	02:26	
Subject Sample	# 1 0.200	02:27	
Breath Volume	1.644	02:27	
Air Blank	0.000	02:28	
Air Blank	0.000	02:30	
Subject Sample	# 2 0.171	02:33	Volume Not Met
Breath Volume	0.660	02:33	
Air Blank	0.000	02:34	
Air Blank	0.000	02:35	
Subject Sample	# 3 0.192	02:37	
Breath Volume	1.511	02:37	
Air Blank	0.000	02:37	
Control Test	0.076	02:38	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 02/15/2020 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2020
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 01/19/2020 05:36:47

Diagnostics Check	OK	05:27
Air Blank	0.000	05:28
Control Test	0.079	05:28
Air Blank	0.000	05:29
Subject Sample # 1	0.179	05:29
Breath Volume	2.183	05:29
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.177	05:35
Breath Volume	1.585	05:35
Air Blank	0.000	05:35
Control Test	0.079	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 02/15/2020 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ZAVALA BORJAS

DATE/TIME OF TEST: 01/20/2020 02:19:21

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.161	02:14
Breath Volume	2.050	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:16
Subject Sample # 2	0.158	02:17
Breath Volume	2.242	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 02/15/2020 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2020
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 01/21/2020 04:44:27

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:41	
Control Test	0.080	04:41	
Air Blank	0.000	04:42	
Subject Sample	# 1	04:43	Subject Test Refused
Breath Volume	0.000	04:43	
Air Blank	0.000	04:43	
Control Test	0.080	04:43	
Air Blank	0.000	04:44	
Diagnostics Check	OK	04:44	

Data Download Date/Time: 02/15/2020 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2020
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GROWNEY

DATE/TIME OF TEST: 01/23/2020 01:02:35

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:58	
Control Test	0.080	00:58	
Air Blank	0.000	00:59	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:01	
Control Test	0.081	01:01	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 02/15/2020 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURCHETT,JASON,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: WENTLAND,ALEX
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2020
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BARRON

DATE/TIME OF TEST: 01/24/2020 00:32:28

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.161	00:27
Breath Volume	2.199	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.153	00:30
Breath Volume	1.242	00:30
Air Blank	0.000	00:31
Control Test	0.077	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 02/15/2020 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2020
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/25/2020 03:36:01

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.154	03:29
Breath Volume	5.320	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.148	03:34
Breath Volume	4.019	03:34
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:36

Data Download Date/Time: 02/15/2020 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMBERLIN,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TOMBERLIN,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2020
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: WOOTEN

DATE/TIME OF TEST: 01/26/2020 04:47:03

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.080	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.129	04:42
Breath Volume	1.625	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:44
Subject Sample # 2	0.131	04:45
Breath Volume	2.453	04:45
Air Blank	0.000	04:46
Control Test	0.077	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:47

Data Download Date/Time: 02/15/2020 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: METZLER

DATE/TIME OF TEST: 01/26/2020 05:59:47

Diagnostics Check	OK	05:52
Air Blank	0.000	05:53
Control Test	0.080	05:53
Air Blank	0.000	05:54
Subject Sample # 1	0.132	05:54
Breath Volume	2.101	05:54
Air Blank	0.000	05:55
Air Blank	0.000	05:57
Subject Sample # 2	0.133	05:58
Breath Volume	2.636	05:58
Air Blank	0.000	05:58
Control Test	0.080	05:59
Air Blank	0.000	05:59
Diagnostics Check	OK	05:59

Data Download Date/Time: 02/15/2020 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMBERLIN,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TOMBERLIN,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: ESCOBEDO GOMEZ

DATE/TIME OF TEST: 01/27/2020 02:09:18

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:06	
Control Test	0.081	02:07	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:07	Subject Test Refused
Breath Volume	0.000	02:07	
Air Blank	0.000	02:08	
Control Test	0.080	02:08	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

Data Download Date/Time: 02/15/2020 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2020
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GIL TOVAR

DATE/TIME OF TEST: 01/27/2020 03:37:55

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.128	03:33
Breath Volume	3.968	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:35
Subject Sample # 2	0.125	03:36
Breath Volume	4.230	03:36
Air Blank	0.000	03:36
Control Test	0.080	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 02/15/2020 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURCHETT,JASON,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: WENTLAND,ALEX
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2020
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: GAVIA

DATE/TIME OF TEST: 01/28/2020 00:31:11

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:28	
Control Test	0.079	00:29	
Air Blank	0.000	00:29	
Subject Sample	# 1	00:29	Subject Test Refused
Breath Volume	0.000	00:29	
Air Blank	0.000	00:30	
Control Test	0.080	00:30	
Air Blank	0.000	00:31	
Diagnostics Check	OK	00:31	

Data Download Date/Time: 02/15/2020 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,HAROLD,
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: BAEZ,HAROLD
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2020
VIOLATION CODE: DUI
ARREST TIME: 09:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:29

SUBJECT'S LAST NAME: NEWBECK

DATE/TIME OF TEST: 01/28/2020 11:05:04

Diagnostics Check	OK	10:53	
Air Blank	0.000	10:54	
Control Test	0.080	10:54	
Air Blank	0.000	10:54	
Subject Sample	# 1	10:58	No Sample Provided
Breath Volume	0.000	10:58	
Air Blank	0.000	10:58	
Air Blank	0.000	11:00	
Subject Sample	# 2	11:03	No Sample Provided
Breath Volume	0.000	11:03	
Air Blank	0.000	11:04	
Control Test	0.080	11:04	
Air Blank	0.000	11:04	
Diagnostics Check	OK	11:05	

Data Download Date/Time: 02/15/2020 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,HAROLD,
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: BAEZ,HAROLD
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2020
VIOLATION CODE: DUI
ARREST TIME: 09:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:48

SUBJECT'S LAST NAME: NEWBECK

DATE/TIME OF TEST: 01/28/2020 11:17:17

Diagnostics Check	OK	11:09
Air Blank	0.000	11:09
Control Test	0.080	11:09
Air Blank	0.000	11:10
Subject Sample # 1	0.000	11:11
Breath Volume	1.578	11:11
Air Blank	0.000	11:12
Air Blank	0.000	11:14
Subject Sample # 2	0.000	11:15
Breath Volume	1.628	11:15
Air Blank	0.000	11:16
Control Test	0.076	11:16
Air Blank	0.000	11:17
Diagnostics Check	OK	11:17

Data Download Date/Time: 02/15/2020 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MARTIN PEREZ

DATE/TIME OF TEST: 01/30/2020 02:36:04

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:33	
Control Test	0.080	02:33	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:35	
Control Test	0.080	02:35	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 02/15/2020 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2020
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: GARCIA ALVAREZ

DATE/TIME OF TEST: 01/31/2020 03:58:23

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.080	03:56	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Control Test	0.080	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 02/15/2020 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMBERLIN,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TOMBERLIN,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2020
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: VILLAMAR VEGA

DATE/TIME OF TEST: 01/31/2020 05:02:04

Diagnostics Check	OK	04:59	
Air Blank	0.000	04:59	
Control Test	0.079	04:59	
Air Blank	0.000	05:00	
Subject Sample	# 1	05:00	Subject Test Refused
Breath Volume	0.000	05:00	
Air Blank	0.000	05:01	
Control Test	0.080	05:01	
Air Blank	0.000	05:01	
Diagnostics Check	OK	05:02	

Data Download Date/Time: 02/15/2020 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 01/04/2020 02:59:21

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:55	
Control Test	0.079	02:55	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:58	
Control Test	0.080	02:58	
Air Blank	0.000	02:59	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 01/27/2020 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASLINE,T'NEISHA,S
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TOLLIK,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: REYNOSO

DATE/TIME OF TEST: 01/05/2020 06:33:59

Diagnostics Check	OK	06:24
Air Blank	0.000	06:25
Control Test	0.080	06:25
Air Blank	0.000	06:26
Subject Sample # 1	0.203	06:29
Breath Volume	1.171	06:29
Air Blank	0.000	06:29
Air Blank	0.000	06:31
Subject Sample # 2	0.208	06:32
Breath Volume	1.433	06:32
Air Blank	0.000	06:33
Control Test	0.077	06:33
Air Blank	0.000	06:33
Diagnostics Check	OK	06:33

Data Download Date/Time: 01/27/2020 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: RANKIN

DATE/TIME OF TEST: 01/10/2020 01:35:25

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:27	
Control Test	0.080	01:27	
Air Blank	0.000	01:28	
Subject Sample	# 1	01:31	No Sample Provided
Breath Volume	0.000	01:31	
Air Blank	0.000	01:31	
Air Blank	0.000	01:33	
Subject Sample	# 2	01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:34	
Control Test	0.080	01:34	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 01/27/2020 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: EURE

DATE/TIME OF TEST: 01/18/2020 04:36:35

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.157	04:31
Breath Volume	1.292	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:33
Subject Sample # 2	0.147	04:34
Breath Volume	1.156	04:34
Air Blank	0.000	04:35
Control Test	0.075	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 01/27/2020 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: MINARD

DATE/TIME OF TEST: 01/18/2020 23:38:18

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.080	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.192	23:33
Breath Volume	3.386	23:33
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.186	23:36
Breath Volume	2.617	23:36
Air Blank	0.000	23:37
Control Test	0.076	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 01/27/2020 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: POWERS,L
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 01/19/2020 01:27:31

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:24	
Control Test	0.080	01:24	
Air Blank	0.000	01:25	
Subject Sample	# 1	01:26	Subject Test Refused
Breath Volume	0.000	01:26	
Air Blank	0.000	01:26	
Control Test	0.080	01:26	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:27	

Data Download Date/Time: 01/27/2020 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: CARDASCIA

DATE/TIME OF TEST: 01/19/2020 22:03:44

Diagnostics Check	OK	21:58	
Air Blank	0.000	21:59	
Control Test	0.080	21:59	
Air Blank	0.000	21:59	
Subject Sample	# 1	22:02	Subject Test Refused
Breath Volume	0.000	22:02	
Air Blank	0.000	22:02	
Control Test	0.080	22:03	
Air Blank	0.000	22:03	
Diagnostics Check	OK	22:03	

Data Download Date/Time: 01/27/2020 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PYNE

DATE/TIME OF TEST: 01/20/2020 01:36:22

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.070	01:31
Breath Volume	2.042	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:33
Subject Sample # 2	0.063	01:34
Breath Volume	1.281	01:34
Air Blank	0.000	01:35
Control Test	0.077	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 01/27/2020 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD, CHARLES, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD, T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: YERBY MOORE

DATE/TIME OF TEST: 01/21/2020 00:15:36

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.081	00:10
Air Blank	0.000	00:10
Subject Sample # 1	0.000	00:11
Breath Volume	1.636	00:11
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.000	00:14
Breath Volume	1.835	00:14
Air Blank	0.000	00:14
Control Test	0.077	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 01/27/2020 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASLINE, T'NEISHA, S
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ROOT, K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 01/23/2020 17:29:35

Diagnostics Check	OK	17:23
Air Blank	0.000	17:23
Control Test	0.080	17:23
Air Blank	0.000	17:24
Subject Sample # 1	0.000	17:25
Breath Volume	1.269	17:25
Air Blank	0.000	17:25
Air Blank	0.000	17:27
Subject Sample # 2	0.000	17:28
Breath Volume	1.582	17:28
Air Blank	0.000	17:28
Control Test	0.077	17:28
Air Blank	0.000	17:29
Diagnostics Check	OK	17:29

Data Download Date/Time: 01/27/2020 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD, CHARLES, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD, T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: RIGBY

DATE/TIME OF TEST: 01/25/2020 03:50:18

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:39	
Control Test		0.080	03:39	
Air Blank		0.000	03:40	
Subject Sample	# 1	0.130	03:43	Volume Not Met
Breath Volume		0.882	03:43	
Air Blank		0.000	03:44	
Air Blank		0.000	03:45	
Subject Sample	# 2		03:48	No Sample Provided
Breath Volume		0.000	03:48	
Air Blank		0.000	03:49	
Control Test		0.077	03:49	
Air Blank		0.000	03:50	
Diagnostics Check		OK	03:50	

Data Download Date/Time: 01/27/2020 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD,CHARLES,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: RIGBY

DATE/TIME OF TEST: 01/25/2020 04:10:37

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.134	04:06
Breath Volume	1.578	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.139	04:09
Breath Volume	1.476	04:09
Air Blank	0.000	04:09
Control Test	0.078	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 01/27/2020 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD, CHARLES, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FAIRCLOTH, D
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/25/2020 04:45:20

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.132	04:40
Breath Volume	1.457	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.130	04:43
Breath Volume	1.632	04:43
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 01/27/2020 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD, CHARLES, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITE, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: TEW

DATE/TIME OF TEST: 01/25/2020 05:13:17

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.079	05:07
Air Blank	0.000	05:07
Subject Sample # 1	0.153	05:08
Breath Volume	1.289	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:10
Subject Sample # 2	0.155	05:11
Breath Volume	1.730	05:11
Air Blank	0.000	05:12
Control Test	0.079	05:12
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 01/27/2020 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LOWRY,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: FLORIBERTO

DATE/TIME OF TEST: 01/26/2020 06:51:24

Diagnostics Check	OK	06:44
Air Blank	0.000	06:44
Control Test	0.081	06:45
Air Blank	0.000	06:45
Subject Sample # 1	0.134	06:46
Breath Volume	1.125	06:46
Air Blank	0.000	06:47
Air Blank	0.000	06:48
Subject Sample # 2	0.133	06:49
Breath Volume	1.640	06:49
Air Blank	0.000	06:50
Control Test	0.081	06:50
Air Blank	0.000	06:51
Diagnostics Check	OK	06:51

Data Download Date/Time: 01/27/2020 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON, MATTHEW, P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD, T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: TARVIT

DATE/TIME OF TEST: 01/26/2020 19:10:45

Diagnostics Check	OK	19:04
Air Blank	0.000	19:04
Control Test	0.081	19:05
Air Blank	0.000	19:05
Subject Sample # 1	0.153	19:06
Breath Volume	2.207	19:06
Air Blank	0.000	19:06
Air Blank	0.000	19:08
Subject Sample # 2	0.162	19:09
Breath Volume	2.312	19:09
Air Blank	0.000	19:09
Control Test	0.080	19:10
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

Data Download Date/Time: 01/27/2020 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DALTON,RYAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 01/01/2020 05:05:06

Diagnostics Check	OK	04:58
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.191	05:00
Breath Volume	2.300	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.192	05:03
Breath Volume	1.960	05:03
Air Blank	0.000	05:04
Control Test	0.079	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 01/17/2020 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MASTERSON,MONIQUE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: TESSENHOLTZ

DATE/TIME OF TEST: 01/01/2020 07:56:49

Diagnostics Check	OK	07:50
Air Blank	0.000	07:50
Control Test	0.080	07:50
Air Blank	0.000	07:51
Subject Sample # 1	0.206	07:51
Breath Volume	2.359	07:51
Air Blank	0.000	07:52
Air Blank	0.000	07:54
Subject Sample # 2	0.204	07:55
Breath Volume	2.117	07:55
Air Blank	0.000	07:55
Control Test	0.080	07:56
Air Blank	0.000	07:56
Diagnostics Check	OK	07:56

Data Download Date/Time: 01/17/2020 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ,MARQUIS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:09

SUBJECT'S LAST NAME: CANO

DATE/TIME OF TEST: 01/01/2020 16:45:11

Diagnostics Check	OK	16:38
Air Blank	0.000	16:38
Control Test	0.080	16:38
Air Blank	0.000	16:39
Subject Sample # 1	0.176	16:40
Breath Volume	2.359	16:40
Air Blank	0.000	16:41
Air Blank	0.000	16:43
Subject Sample # 2	0.179	16:43
Breath Volume	2.285	16:43
Air Blank	0.000	16:44
Control Test	0.080	16:44
Air Blank	0.000	16:45
Diagnostics Check	OK	16:45

Data Download Date/Time: 01/17/2020 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 01/03/2020 01:04:04

Diagnostics Check	OK	00:56
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.092	00:59
Breath Volume	2.843	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.087	01:02
Breath Volume	1.589	01:02
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:04

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ALIEV

DATE/TIME OF TEST: 01/03/2020 03:57:16

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.122	03:52
Breath Volume	1.640	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:54
Subject Sample # 2	0.124	03:55
Breath Volume	1.734	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: ESTEVEZ

DATE/TIME OF TEST: 01/04/2020 01:26:34

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.051	01:21
Breath Volume	2.562	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.053	01:25
Breath Volume	3.453	01:25
Air Blank	0.000	01:25
Control Test	0.080	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TATE,COREY
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: FONSECA

DATE/TIME OF TEST: 01/04/2020 02:52:03

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.059	02:47
Breath Volume	1.402	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.060	02:50
Breath Volume	1.699	02:50
Air Blank	0.000	02:51
Control Test	0.081	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:52

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TUANG,NGIN
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: FOLSOM

DATE/TIME OF TEST: 01/04/2020 05:17:09

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.092	05:12
Breath Volume	1.531	05:12
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.086	05:15
Breath Volume	1.257	05:15
Air Blank	0.000	05:16
Control Test	0.080	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PETTL,JENNA
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:13

SUBJECT'S LAST NAME: GUMAER

DATE/TIME OF TEST: 01/04/2020 08:45:18

Diagnostics Check		OK	08:39
Air Blank		0.000	08:39
Control Test		0.080	08:39
Air Blank		0.000	08:40
Subject Sample	# 1	0.000	08:40
Breath Volume		2.007	08:40
Air Blank		0.000	08:41
Air Blank		0.000	08:43
Subject Sample	# 2	0.000	08:43
Breath Volume		1.597	08:43
Air Blank		0.000	08:44
Control Test		0.080	08:44
Air Blank		0.000	08:45
Diagnostics Check		OK	08:45

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LUNDGREN,KYLE
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: ALICE

DATE/TIME OF TEST: 01/05/2020 01:54:15

Diagnostics Check		OK	01:42	
Air Blank		0.000	01:43	
Control Test		0.080	01:43	
Air Blank		0.000	01:43	
Subject Sample	# 1	0.131	01:47	Volume Not Met
Breath Volume		0.058	01:47	
Air Blank		0.000	01:47	
Air Blank		0.000	01:49	
Subject Sample	# 2	0.123	01:52	Volume Not Met
Breath Volume		0.000	01:52	
Air Blank		0.000	01:53	
Control Test		0.079	01:53	
Air Blank		0.000	01:54	
Diagnostics Check		OK	01:54	

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LUNDGREN,KYLE
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: ALICE

DATE/TIME OF TEST: 01/05/2020 02:02:39

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.078	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.136	01:58
Breath Volume	1.781	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.130	02:01
Breath Volume	2.554	02:01
Air Blank	0.000	02:01
Control Test	0.078	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KYSOR,MITCHELL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: BELLO

DATE/TIME OF TEST: 01/05/2020 04:56:24

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.081	04:50
Air Blank	0.000	04:51
Subject Sample # 1	0.133	04:51
Breath Volume	1.492	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.131	04:54
Breath Volume	1.902	04:54
Air Blank	0.000	04:55
Control Test	0.081	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MARCH,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: ISAIN

DATE/TIME OF TEST: 01/05/2020 21:38:28

Diagnostics Check	OK	21:27	
Air Blank	0.000	21:27	
Control Test	0.080	21:27	
Air Blank	0.000	21:28	
Subject Sample	# 1	21:31	No Sample Provided
Breath Volume	0.000	21:31	
Air Blank	0.000	21:31	
Air Blank	0.000	21:33	
Subject Sample	# 2	21:37	No Sample Provided
Breath Volume	0.000	21:37	
Air Blank	0.000	21:37	
Control Test	0.080	21:37	
Air Blank	0.000	21:38	
Diagnostics Check	OK	21:38	

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: COLIN

DATE/TIME OF TEST: 01/06/2020 01:47:05

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.129	01:42
Breath Volume	1.535	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.129	01:45
Breath Volume	1.171	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JOHNSON,CARL
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: HOBBS JR

DATE/TIME OF TEST: 01/07/2020 03:49:29

Diagnostics Check		OK	03:42
Air Blank		0.000	03:43
Control Test		0.081	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.161	03:44
Breath Volume		1.527	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.162	03:47
Breath Volume		1.250	03:47
Air Blank		0.000	03:48
Control Test		0.080	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ORTEGA,RICARDO
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:06

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 01/07/2020 16:03:16

Diagnostics Check	OK	15:56
Air Blank	0.000	15:56
Control Test	0.081	15:56
Air Blank	0.000	15:57
Subject Sample # 1	0.259	15:58
Breath Volume	1.675	15:58
Air Blank	0.000	15:59
Air Blank	0.000	16:01
Subject Sample # 2	0.267	16:01
Breath Volume	1.949	16:01
Air Blank	0.000	16:02
Control Test	0.080	16:02
Air Blank	0.000	16:03
Diagnostics Check	OK	16:03

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,O
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESPOSITO,DAVID
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: LASDON

DATE/TIME OF TEST: 01/07/2020 18:13:45

Diagnostics Check	OK	18:07
Air Blank	0.000	18:07
Control Test	0.081	18:08
Air Blank	0.000	18:08
Subject Sample # 1	0.117	18:09
Breath Volume	1.722	18:09
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.121	18:12
Breath Volume	1.835	18:12
Air Blank	0.000	18:12
Control Test	0.080	18:13
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 01/08/2020 04:05:33

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.131	04:00
Breath Volume	2.250	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.132	04:04
Breath Volume	2.746	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: ROSEN

DATE/TIME OF TEST: 01/08/2020 21:40:58

Diagnostics Check	OK	21:34
Air Blank	0.000	21:34
Control Test	0.081	21:34
Air Blank	0.000	21:35
Subject Sample # 1	0.141	21:36
Breath Volume	1.644	21:36
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.137	21:39
Breath Volume	1.851	21:39
Air Blank	0.000	21:40
Control Test	0.081	21:40
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 01/09/2020 23:57:45

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.090	23:52
Breath Volume	1.335	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.092	23:56
Breath Volume	3.046	23:56
Air Blank	0.000	23:56
Control Test	0.081	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 01/10/2020 01:27:56

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.126	01:23
Breath Volume	1.902	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:25
Subject Sample # 2	0.125	01:26
Breath Volume	1.976	01:26
Air Blank	0.000	01:27
Control Test	0.078	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASTRO,D
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: QUE

DATE/TIME OF TEST: 01/10/2020 02:54:02

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.161	02:49
Breath Volume	1.789	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.157	02:52
Breath Volume	2.000	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:54

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: LAPRE

DATE/TIME OF TEST: 01/11/2020 01:58:26

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.080	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.147	01:53
Breath Volume	1.609	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.145	01:56
Breath Volume	3.347	01:56
Air Blank	0.000	01:57
Control Test	0.078	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOLOWAY,ALAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SNACHEZ

DATE/TIME OF TEST: 01/12/2020 02:48:55

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.080	02:43
Air Blank		0.000	02:43
Subject Sample	# 1	0.255	02:44
Breath Volume		2.382	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:46
Subject Sample	# 2	0.258	02:47
Breath Volume		2.390	02:47
Air Blank		0.000	02:48
Control Test		0.079	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOLOWAY,ALAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 01/13/2020 01:07:16

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.160	01:02
Breath Volume	1.855	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.157	01:05
Breath Volume	2.335	01:05
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: JEZOWSKI

DATE/TIME OF TEST: 01/14/2020 00:33:15

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.080	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.165	00:28
Breath Volume	4.382	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.161	00:31
Breath Volume	2.781	00:31
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 01/14/2020 01:28:29

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.080	01:19
Air Blank		0.000	01:19
Subject Sample	# 1	0.213	01:20
Breath Volume		1.449	01:20
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.236	01:23
Breath Volume		2.207	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 3	0.221	01:26
Breath Volume		1.585	01:26
Air Blank		0.000	01:27
Control Test		0.077	01:27
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: BRIGGS

DATE/TIME OF TEST: 01/14/2020 22:20:31

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.105	22:16
Breath Volume	2.046	22:16
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.105	22:18
Breath Volume	1.445	22:18
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 01/16/2020 04:24:48

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.080	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.188	04:19
Breath Volume		2.035	04:19
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.170	04:23
Breath Volume		1.472	04:23
Air Blank		0.000	04:23
Control Test		0.080	04:24
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MOREA,DEAN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 01/16/2020 22:19:18

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.080	22:13
Air Blank	0.000	22:14
Subject Sample # 1	0.154	22:14
Breath Volume	1.519	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.155	22:17
Breath Volume	1.425	22:17
Air Blank	0.000	22:18
Control Test	0.080	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 01/17/2020 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROCHA,LUIS
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: JOSE ALONZO

DATE/TIME OF TEST: 01/18/2020 02:27:39

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.162	02:23
Breath Volume	3.746	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.157	02:26
Breath Volume	3.312	02:26
Air Blank	0.000	02:26
Control Test	0.079	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 02/14/2020 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TUANG,NGIN
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: CHILD

DATE/TIME OF TEST: 01/18/2020 04:30:59

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.118	04:26
Breath Volume	2.250	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:28
Subject Sample # 2	0.118	04:29
Breath Volume	3.410	04:29
Air Blank	0.000	04:30
Control Test	0.079	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 02/14/2020 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MENDEZ,ERIC
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: HEINTZ-CATCHPOLE

DATE/TIME OF TEST: 01/19/2020 01:01:35

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.081	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.182	00:57
Breath Volume	1.414	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.184	01:00
Breath Volume	1.382	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 02/14/2020 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: ALBAN

DATE/TIME OF TEST: 01/19/2020 02:45:09

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.127	02:40
Breath Volume	1.753	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.128	02:43
Breath Volume	1.886	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 02/14/2020 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DALTON,RYAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 01/19/2020 04:33:10

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.137	04:28
Breath Volume	1.484	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.134	04:31
Breath Volume	1.652	04:31
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 02/14/2020 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MENDEZ, ERIC
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 01/19/2020 23:58:51

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.080	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.231	23:54
Breath Volume	2.492	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.230	23:57
Breath Volume	2.800	23:57
Air Blank	0.000	23:57
Control Test	0.079	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 02/14/2020 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGILLICUDDY, STEVEN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MACINTOSH III

DATE/TIME OF TEST: 01/20/2020 02:08:03

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.151	02:03
Breath Volume	3.054	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:05
Subject Sample # 2	0.150	02:06
Breath Volume	3.242	02:06
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:08

Data Download Date/Time: 02/14/2020 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: REECE,COREY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: CUCURELLA

DATE/TIME OF TEST: 01/22/2020 09:06:06

Diagnostics Check	OK	08:59
Air Blank	0.000	09:00
Control Test	0.081	09:00
Air Blank	0.000	09:00
Subject Sample # 1	0.000	09:01
Breath Volume	1.246	09:01
Air Blank	0.000	09:02
Air Blank	0.000	09:04
Subject Sample # 2	0.000	09:04
Breath Volume	1.589	09:04
Air Blank	0.000	09:05
Control Test	0.081	09:05
Air Blank	0.000	09:06
Diagnostics Check	OK	09:06

Data Download Date/Time: 02/14/2020 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: REECE,COREY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:24

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 01/23/2020 10:47:01

Diagnostics Check	OK	10:40
Air Blank	0.000	10:41
Control Test	0.080	10:41
Air Blank	0.000	10:41
Subject Sample # 1	0.030	10:42
Breath Volume	1.996	10:42
Air Blank	0.000	10:43
Air Blank	0.000	10:45
Subject Sample # 2	0.029	10:45
Breath Volume	1.503	10:45
Air Blank	0.000	10:46
Control Test	0.081	10:46
Air Blank	0.000	10:46
Diagnostics Check	OK	10:47

Data Download Date/Time: 02/14/2020 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HUYETT,MARK
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:43

SUBJECT'S LAST NAME: NUTTALL

DATE/TIME OF TEST: 01/23/2020 12:18:23

Diagnostics Check	OK	12:11
Air Blank	0.000	12:12
Control Test	0.080	12:12
Air Blank	0.000	12:13
Subject Sample # 1	0.000	12:13
Breath Volume	3.089	12:13
Air Blank	0.000	12:14
Air Blank	0.000	12:16
Subject Sample # 2	0.000	12:17
Breath Volume	2.558	12:17
Air Blank	0.000	12:17
Control Test	0.080	12:17
Air Blank	0.000	12:18
Diagnostics Check	OK	12:18

Data Download Date/Time: 02/14/2020 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: REINER

DATE/TIME OF TEST: 01/23/2020 22:22:29

Diagnostics Check	OK	22:15
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.000	22:17
Breath Volume	3.480	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.000	22:21
Breath Volume	3.316	22:21
Air Blank	0.000	22:21
Control Test	0.080	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASTRO,D
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: BOZIEUX

DATE/TIME OF TEST: 01/24/2020 01:40:13

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.080	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.244	01:35
Breath Volume	1.988	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:37
Subject Sample # 2	0.241	01:38
Breath Volume	1.218	01:38
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGILLICUDDY,STEVEN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: BENNION

DATE/TIME OF TEST: 01/24/2020 02:23:35

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.079	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.242	02:18
Breath Volume	2.636	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.255	02:21
Breath Volume	2.523	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BONET,L
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/24/2020 02:50:15

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.307	02:45
Breath Volume	2.500	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.308	02:48
Breath Volume	1.910	02:48
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,DAVID
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: WEISS

DATE/TIME OF TEST: 01/25/2020 00:46:01

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.133	00:41
Breath Volume	1.437	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.130	00:44
Breath Volume	1.441	00:44
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:46

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DIAZ MORALES

DATE/TIME OF TEST: 01/25/2020 03:51:08

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.081	03:42
Air Blank		0.000	03:42
Subject Sample	# 1	0.189	03:43
Breath Volume		1.992	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:45
Subject Sample	# 2	0.212	03:46
Breath Volume		1.507	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 3	0.198	03:49
Breath Volume		1.265	03:49
Air Blank		0.000	03:50
Control Test		0.078	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LOPEZ, KAROL
ARRESTING OFFICER AGENCY: Tequesta PD
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: SALAS

DATE/TIME OF TEST: 01/25/2020 05:16:24

Diagnostics Check	OK	05:09
Air Blank	0.000	05:10
Control Test	0.080	05:10
Air Blank	0.000	05:10
Subject Sample # 1	0.140	05:11
Breath Volume	2.937	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.133	05:14
Breath Volume	1.296	05:14
Air Blank	0.000	05:15
Control Test	0.079	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH, GREG
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: PEREZ CHILEL

DATE/TIME OF TEST: 01/26/2020 04:07:47

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.072	04:03
Breath Volume	1.687	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.073	04:06
Breath Volume	1.261	04:06
Air Blank	0.000	04:06
Control Test	0.080	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MARCH, THOMAS
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ALEXIS

DATE/TIME OF TEST: 01/27/2020 03:18:16

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.081	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.211	03:13
Breath Volume	1.546	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:15
Subject Sample # 2	0.217	03:16
Breath Volume	1.285	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MCWILLIAMS

DATE/TIME OF TEST: 01/27/2020 23:50:27

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.045	23:45
Breath Volume	2.667	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.045	23:48
Breath Volume	2.671	23:48
Air Blank	0.000	23:49
Control Test	0.080	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS, ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: STAVOLA

DATE/TIME OF TEST: 01/28/2020 22:47:52

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.080	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.124	22:43
Breath Volume	1.804	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:45
Subject Sample # 2	0.122	22:46
Breath Volume	1.871	22:46
Air Blank	0.000	22:46
Control Test	0.079	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MILLER, KARLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: EDMOND

DATE/TIME OF TEST: 01/29/2020 01:34:07

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.158	01:29
Breath Volume	1.195	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.165	01:32
Breath Volume	1.359	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DE YOUNG

DATE/TIME OF TEST: 01/29/2020 03:31:23

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.160	03:26
Breath Volume	3.183	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample # 2	0.162	03:29
Breath Volume	2.597	03:29
Air Blank	0.000	03:30
Control Test	0.078	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BUDISH,HARRISON
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: NAVAMUEL

DATE/TIME OF TEST: 01/29/2020 23:47:40

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.243	23:43
Breath Volume	4.785	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.233	23:46
Breath Volume	4.312	23:46
Air Blank	0.000	23:46
Control Test	0.079	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: HERRERA ORDAZ

DATE/TIME OF TEST: 01/30/2020 00:59:43

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.080	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.169	00:55
Breath Volume	2.171	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.175	00:58
Breath Volume	1.898	00:58
Air Blank	0.000	00:58
Control Test	0.079	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,DAVID
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: HONEYCUTT

DATE/TIME OF TEST: 01/31/2020 00:06:04

Diagnostics Check		OK	23:59
Air Blank		0.000	00:00
Control Test		0.080	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.214	00:01
Breath Volume		1.402	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:03
Subject Sample	# 2	0.199	00:04
Breath Volume		2.324	00:04
Air Blank		0.000	00:05
Control Test		0.080	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:06

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: WEINSTEIN

DATE/TIME OF TEST: 01/31/2020 00:54:33

Diagnostics Check		OK	00:40	
Air Blank		0.000	00:41	
Control Test		0.079	00:41	
Air Blank		0.000	00:42	
Subject Sample	# 1	0.148	00:45	Volume Not Met
Breath Volume		1.031	00:45	
Air Blank		0.000	00:46	
Air Blank		0.000	00:47	
Subject Sample	# 2	0.153	00:49	
Breath Volume		1.273	00:49	
Air Blank		0.000	00:50	
Air Blank		0.000	00:51	
Subject Sample	# 3	0.151	00:52	
Breath Volume		1.191	00:52	
Air Blank		0.000	00:53	
Control Test		0.077	00:53	
Air Blank		0.000	00:54	
Diagnostics Check		OK	00:54	

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: STIGALL

DATE/TIME OF TEST: 01/31/2020 01:12:35

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.079	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.131	01:08
Breath Volume	1.871	01:08
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.139	01:11
Breath Volume	2.144	01:11
Air Blank	0.000	01:11
Control Test	0.078	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:34

SUBJECT'S LAST NAME: TAMBRIZ AJPOP

DATE/TIME OF TEST: 01/31/2020 18:11:12

Diagnostics Check	OK	18:04
Air Blank	0.000	18:04
Control Test	0.080	18:05
Air Blank	0.000	18:05
Subject Sample # 1	0.284	18:06
Breath Volume	1.101	18:06
Air Blank	0.000	18:07
Air Blank	0.000	18:09
Subject Sample # 2	0.272	18:09
Breath Volume	1.664	18:09
Air Blank	0.000	18:10
Control Test	0.080	18:10
Air Blank	0.000	18:11
Diagnostics Check	OK	18:11

Data Download Date/Time: 02/14/2020 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: GRIFFITHS

DATE/TIME OF TEST: 01/01/2020 02:08:52

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.083	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.067	02:04
Breath Volume	1.136	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.067	02:07
Breath Volume	1.507	02:07
Air Blank	0.000	02:07
Control Test	0.081	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 01/16/2020 18:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: CRANFORD

DATE/TIME OF TEST: 01/02/2020 21:31:48

Diagnostics Check	OK	21:24
Air Blank	0.000	21:24
Control Test	0.082	21:24
Air Blank	0.000	21:25
Subject Sample # 1	0.140	21:27
Breath Volume	2.292	21:27
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample # 2	0.136	21:29
Breath Volume	2.828	21:29
Air Blank	0.000	21:30
Control Test	0.078	21:31
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time: 01/16/2020 18:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 01/17/2020 00:48:26

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.082	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.141	00:43
Breath Volume	3.089	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.138	00:46
Breath Volume	4.355	00:46
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 02/12/2020 19:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 01/18/2020 02:45:49

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.083	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.150	02:40
Breath Volume	2.808	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.147	02:44
Breath Volume	3.710	02:44
Air Blank	0.000	02:44
Control Test	0.079	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 02/12/2020 19:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: CARVELLI

DATE/TIME OF TEST: 01/19/2020 04:53:52

Diagnostics Check	OK	04:47
Air Blank	0.000	04:47
Control Test	0.082	04:48
Air Blank	0.000	04:48
Subject Sample # 1	0.144	04:49
Breath Volume	1.683	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:51
Subject Sample # 2	0.140	04:52
Breath Volume	2.621	04:52
Air Blank	0.000	04:52
Control Test	0.080	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 02/12/2020 19:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARLA
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ONORATO,BRIANA
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: 19907209

DATE/TIME OF TEST: 01/25/2020 20:00:31

Diagnostics Check	OK	19:53	
Air Blank	0.000	19:54	
Control Test	0.083	19:54	
Air Blank	0.000	19:54	
Subject Sample # 1	0.207	19:55	
Breath Volume	3.574	19:55	
Air Blank	0.000	19:56	
Air Blank	0.000	19:58	
Subject Sample # 2	0.190	19:59	
Breath Volume	1.582	19:59	
Air Blank		20:00	Purge Fail
Air Blank	0.000	20:00	

Data Download Date/Time: 02/12/2020 19:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARLA
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ONORATO,BRIANA
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: MCHALE

DATE/TIME OF TEST: 01/25/2020 20:12:26

Diagnostics Check	OK	20:05	
Air Blank	0.000	20:06	
Control Test	0.080	20:06	
Air Blank	0.000	20:07	
Subject Sample # 1	0.180	20:07	
Breath Volume	1.562	20:07	
Air Blank	0.000	20:08	
Air Blank	0.000	20:10	
Subject Sample # 2	0.191	20:10	
Breath Volume	3.156	20:10	
Air Blank		20:11	Purge Fail
Air Blank	0.000	20:12	

Data Download Date/Time: 02/12/2020 19:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 01/26/2020 04:20:35

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.083	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.117	04:15
Breath Volume	2.105	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.116	04:18
Breath Volume	1.640	04:18
Air Blank	0.000	04:19
Control Test	0.081	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 02/12/2020 19:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTENOUR,COREY,T
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: RITTENOUR,COREY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: ALVARADO MORELOS

DATE/TIME OF TEST: 01/11/2020 23:48:18

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.081	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.221	23:43
Breath Volume	4.187	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:45
Subject Sample # 2	0.225	23:46
Breath Volume	3.523	23:46
Air Blank	0.000	23:47
Control Test	0.079	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 01/27/2020 12:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: ANDREESSEN,SAMUEL
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 01/29/2020 03:01:37

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.081	02:47	
Air Blank	0.000	02:48	
Subject Sample	# 1 0.135	02:51	Volume Not Met
Breath Volume	0.222	02:51	
Air Blank	0.000	02:52	
Air Blank	0.000	02:54	
Subject Sample	# 2 0.121	02:56	
Breath Volume	1.792	02:56	
Air Blank	0.000	02:57	
Air Blank	0.000	02:58	
Subject Sample	# 3 0.121	02:59	
Breath Volume	1.433	02:59	
Air Blank	0.000	03:00	
Control Test	0.077	03:00	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 02/24/2020 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMS,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: KLEIN

DATE/TIME OF TEST: 01/01/2020 00:13:48

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.081	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.179	00:09
Breath Volume	1.437	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.180	00:12
Breath Volume	1.480	00:12
Air Blank	0.000	00:12
Control Test	0.080	00:13
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 01/06/2020 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KLENOFSKY,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 01/01/2020 00:37:00

Diagnostics Check	OK	00:29
Air Blank	0.000	00:29
Control Test	0.081	00:29
Air Blank	0.000	00:30
Subject Sample # 1	0.339	00:32
Breath Volume	3.183	00:32
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.332	00:35
Breath Volume	2.929	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:37

Data Download Date/Time: 01/06/2020 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROSALES,F
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: CASTANEDA SANTIZO

DATE/TIME OF TEST: 01/01/2020 02:36:31

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.290	02:31
Breath Volume	1.726	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.301	02:34
Breath Volume	2.191	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 01/06/2020 18:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:43

SUBJECT'S LAST NAME: MARUTHOOPARAMPIL

DATE/TIME OF TEST: 01/01/2020 06:10:04

Diagnostics Check	OK	06:03
Air Blank	0.000	06:04
Control Test	0.080	06:04
Air Blank	0.000	06:04
Subject Sample # 1	0.139	06:05
Breath Volume	1.117	06:05
Air Blank	0.000	06:06
Air Blank	0.000	06:08
Subject Sample # 2	0.153	06:08
Breath Volume	1.523	06:08
Air Blank	0.000	06:09
Control Test	0.080	06:09
Air Blank	0.000	06:09
Diagnostics Check	OK	06:10

Data Download Date/Time: 01/06/2020 18:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: CARONONGAN JR

DATE/TIME OF TEST: 01/01/2020 22:15:16

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.081	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.000	22:10
Breath Volume	1.113	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.000	22:13
Breath Volume	3.625	22:13
Air Blank	0.000	22:14
Control Test	0.081	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 01/06/2020 18:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 01/03/2020 02:35:54

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.080	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.058	02:31
Breath Volume		1.437	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.059	02:34
Breath Volume		1.632	02:34
Air Blank		0.000	02:34
Control Test		0.080	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 01/06/2020 18:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: OGLESBY

DATE/TIME OF TEST: 01/03/2020 04:26:35

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.081	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.129	04:21
Breath Volume	1.707	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.135	04:24
Breath Volume	1.785	04:24
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 01/06/2020 18:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: BEDOYA

DATE/TIME OF TEST: 01/04/2020 05:26:32

Diagnostics Check		OK	05:19
Air Blank		0.000	05:19
Control Test		0.080	05:19
Air Blank		0.000	05:20
Subject Sample	# 1	0.125	05:21
Breath Volume		1.500	05:21
Air Blank		0.000	05:22
Air Blank		0.000	05:23
Subject Sample	# 2	0.125	05:24
Breath Volume		1.390	05:24
Air Blank		0.000	05:25
Control Test		0.079	05:25
Air Blank		0.000	05:26
Diagnostics Check		OK	05:26

Data Download Date/Time: 01/06/2020 18:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 01/05/2020 02:29:50

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.012	02:25
Breath Volume	1.761	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.012	02:28
Breath Volume	1.140	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 01/06/2020 18:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: EBANKS

DATE/TIME OF TEST: 01/05/2020 22:01:34

Diagnostics Check		OK	21:55
Air Blank		0.000	21:55
Control Test		0.081	21:56
Air Blank		0.000	21:56
Subject Sample	# 1	0.245	21:56
Breath Volume		4.140	21:56
Air Blank		0.000	21:57
Air Blank		0.000	21:59
Subject Sample	# 2	0.245	21:59
Breath Volume		4.414	21:59
Air Blank		0.000	22:00
Control Test		0.081	22:00
Air Blank		0.000	22:01
Diagnostics Check		OK	22:01

Data Download Date/Time: 01/06/2020 18:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: CATALANI

DATE/TIME OF TEST: 01/06/2020 22:19:35

Diagnostics Check	OK	22:12
Air Blank	0.000	22:13
Control Test	0.082	22:13
Air Blank	0.000	22:13
Subject Sample # 1	0.265	22:14
Breath Volume	1.863	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.264	22:17
Breath Volume	1.925	22:17
Air Blank	0.000	22:18
Control Test	0.081	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RUIZ,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: VALLADARES

DATE/TIME OF TEST: 01/08/2020 04:46:20

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:41	
Control Test	0.081	04:41	
Air Blank	0.000	04:42	
Subject Sample	# 1	04:44	Subject Test Refused
Breath Volume	0.000	04:44	
Air Blank	0.000	04:45	
Control Test	0.081	04:45	
Air Blank	0.000	04:46	
Diagnostics Check	OK	04:46	

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: REYES MARCOS

DATE/TIME OF TEST: 01/11/2020 00:41:15

Diagnostics Check		OK	00:34
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:35
Subject Sample	# 1	0.175	00:36
Breath Volume		2.683	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.173	00:39
Breath Volume		1.726	00:39
Air Blank		0.000	00:40
Control Test		0.080	00:40
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: FETHERLIN

DATE/TIME OF TEST: 01/11/2020 01:25:41

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.080	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.323	01:20
Breath Volume	3.894	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.326	01:23
Breath Volume	3.406	01:23
Air Blank	0.000	01:24
Control Test	0.079	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: KESSLER JR

DATE/TIME OF TEST: 01/11/2020 04:05:10

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.081	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.178	04:00
Breath Volume	3.089	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.176	04:03
Breath Volume	2.789	04:03
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: NAVA NAVARRETE

DATE/TIME OF TEST: 01/11/2020 23:10:35

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.199	23:05
Breath Volume	1.484	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.205	23:08
Breath Volume	1.363	23:08
Air Blank	0.000	23:09
Control Test	0.079	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: CHILDRESS JR

DATE/TIME OF TEST: 01/12/2020 04:22:25

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.163	04:15
Breath Volume		1.812	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.146	04:20
Breath Volume		1.832	04:20
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 01/12/2020 05:13:07

Diagnostics Check	OK	05:07
Air Blank	0.000	05:07
Control Test	0.079	05:07
Air Blank	0.000	05:08
Subject Sample # 1	0.124	05:08
Breath Volume	2.867	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.122	05:11
Breath Volume	2.425	05:11
Air Blank	0.000	05:12
Control Test	0.078	05:12
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: RUMICK

DATE/TIME OF TEST: 01/12/2020 05:40:23

Diagnostics Check	OK	05:33
Air Blank	0.000	05:34
Control Test	0.079	05:34
Air Blank	0.000	05:34
Subject Sample # 1	0.120	05:35
Breath Volume	1.343	05:35
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.127	05:38
Breath Volume	1.578	05:38
Air Blank	0.000	05:39
Control Test	0.079	05:39
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: OJEDA

DATE/TIME OF TEST: 01/13/2020 01:02:41

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.080	00:57
Air Blank	0.000	00:57
Subject Sample # 1	0.133	00:58
Breath Volume	1.722	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.148	01:01
Breath Volume	2.289	01:01
Air Blank	0.000	01:01
Control Test	0.080	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 01/13/2020 03:21:35

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.080	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.168	03:17
Breath Volume		2.085	03:17
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.166	03:19
Breath Volume		1.113	03:19
Air Blank		0.000	03:20
Control Test		0.079	03:20
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: MCKEE

DATE/TIME OF TEST: 01/13/2020 05:22:06

Diagnostics Check	OK	05:15
Air Blank	0.000	05:16
Control Test	0.080	05:16
Air Blank	0.000	05:17
Subject Sample # 1	0.164	05:17
Breath Volume	2.636	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:19
Subject Sample # 2	0.163	05:20
Breath Volume	2.761	05:20
Air Blank	0.000	05:21
Control Test	0.080	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KLENOFSKY,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: ALLUMS

DATE/TIME OF TEST: 01/14/2020 02:56:07

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.000	02:51
Breath Volume	2.582	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.000	02:54
Breath Volume	1.667	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FERGUSON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: HERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 01/15/2020 03:03:23

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.120	02:58
Breath Volume	2.839	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.117	03:01
Breath Volume	2.074	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/16/2020 03:02:52

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:48	
Control Test	0.080	02:49	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:52	No Sample Provided
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Air Blank	0.000	02:55	
Subject Sample	# 2	0.205	02:57
Breath Volume	1.519	02:57	
Air Blank	0.000	02:58	
Air Blank	0.000	03:00	
Subject Sample	# 3	0.189	03:01
Breath Volume	1.242	03:01	
Air Blank	0.000	03:01	
Control Test	0.078	03:02	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: ROLAND

DATE/TIME OF TEST: 01/17/2020 02:41:28

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.115	02:36
Breath Volume	3.339	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.112	02:39
Breath Volume	3.835	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: RECCHIO

DATE/TIME OF TEST: 01/17/2020 04:33:58

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.080	04:27
Air Blank		0.000	04:27
Subject Sample	# 1	0.188	04:29
Breath Volume		2.402	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:31
Subject Sample	# 2	0.186	04:32
Breath Volume		2.015	04:32
Air Blank		0.000	04:33
Control Test		0.079	04:33
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: KUC

DATE/TIME OF TEST: 01/17/2020 23:25:21

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.200	23:20
Breath Volume	2.531	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.204	23:23
Breath Volume	2.542	23:23
Air Blank	0.000	23:24
Control Test	0.079	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BETT

DATE/TIME OF TEST: 01/18/2020 01:11:52

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.081	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.170	01:07
Breath Volume	3.152	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:09
Subject Sample # 2	0.178	01:10
Breath Volume	2.617	01:10
Air Blank	0.000	01:10
Control Test	0.078	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ZENTENO

DATE/TIME OF TEST: 01/18/2020 03:19:23

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.081	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.136	03:14
Breath Volume		2.152	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:16
Subject Sample	# 2	0.127	03:17
Breath Volume		2.195	03:17
Air Blank		0.000	03:18
Control Test		0.079	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 02/05/2020 18:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: COLLAZO SANTANA

DATE/TIME OF TEST: 01/18/2020 04:32:59

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.080	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.135	04:28
Breath Volume	1.574	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:30
Subject Sample # 2	0.131	04:31
Breath Volume	1.632	04:31
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: DONOVAN

DATE/TIME OF TEST: 01/18/2020 04:50:08

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.080	04:42
Air Blank	0.000	04:43
Subject Sample # 1	0.199	04:44
Breath Volume	1.695	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.211	04:48
Breath Volume	1.968	04:48
Air Blank	0.000	04:49
Control Test	0.080	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: PEDRO

DATE/TIME OF TEST: 01/19/2020 00:20:10

Diagnostics Check		OK	00:13
Air Blank		0.000	00:13
Control Test		0.080	00:14
Air Blank		0.000	00:14
Subject Sample	# 1	0.231	00:15
Breath Volume		2.031	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.221	00:18
Breath Volume		2.097	00:18
Air Blank		0.000	00:19
Control Test		0.079	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: YEARWOOD

DATE/TIME OF TEST: 01/19/2020 01:17:25

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.081	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.000	01:12
Breath Volume	2.882	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.000	01:16
Breath Volume	2.660	01:16
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FERGUSON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: POUERIE

DATE/TIME OF TEST: 01/19/2020 03:38:06

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.162	03:33
Breath Volume	1.574	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:35
Subject Sample # 2	0.168	03:36
Breath Volume	1.683	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: MENDIETA

DATE/TIME OF TEST: 01/19/2020 05:47:31

Diagnostics Check		OK	05:40
Air Blank		0.000	05:40
Control Test		0.081	05:41
Air Blank		0.000	05:41
Subject Sample	# 1	0.157	05:42
Breath Volume		3.933	05:42
Air Blank		0.000	05:43
Air Blank		0.000	05:45
Subject Sample	# 2	0.155	05:45
Breath Volume		3.296	05:45
Air Blank		0.000	05:46
Control Test		0.080	05:46
Air Blank		0.000	05:47
Diagnostics Check		OK	05:47

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TORRES JR

DATE/TIME OF TEST: 01/20/2020 02:58:48

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.120	02:53
Breath Volume	1.464	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.123	02:57
Breath Volume	1.578	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILCHER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: ORTIZ NEGRON

DATE/TIME OF TEST: 01/21/2020 19:51:35

Diagnostics Check		OK	19:44
Air Blank		0.000	19:44
Control Test		0.081	19:45
Air Blank		0.000	19:45
Subject Sample	# 1	0.208	19:46
Breath Volume		2.527	19:46
Air Blank		0.000	19:47
Air Blank		0.000	19:49
Subject Sample	# 2	0.204	19:49
Breath Volume		2.015	19:49
Air Blank		0.000	19:50
Control Test		0.079	19:50
Air Blank		0.000	19:51
Diagnostics Check		OK	19:51

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: MCPHEE III

DATE/TIME OF TEST: 01/24/2020 22:40:54

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.081	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.141	22:36
Breath Volume	2.441	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.126	22:39
Breath Volume	1.335	22:39
Air Blank	0.000	22:39
Control Test	0.080	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: KETNOR

DATE/TIME OF TEST: 01/25/2020 03:41:23

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.081	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.117	03:35
Breath Volume	1.753	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.112	03:39
Breath Volume	1.441	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: LEMUS SARDINAS

DATE/TIME OF TEST: 01/25/2020 23:13:49

Diagnostics Check	OK	23:02		
Air Blank	0.000	23:02		
Control Test	0.082	23:02		
Air Blank	0.000	23:03		
Subject Sample	# 1	23:06	No Sample Provided	
Breath Volume	0.000	23:06		
Air Blank	0.000	23:06		
Air Blank	0.000	23:08		
Subject Sample	# 2	0.202	23:12	Volume Not Met
Breath Volume	0.363	23:12		
Air Blank	0.000	23:12		
Control Test	0.082	23:13		
Air Blank	0.000	23:13		
Diagnostics Check	OK	23:13		

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: LEMUS SARDINAS

DATE/TIME OF TEST: 01/25/2020 23:17:54

Diagnostics Check	OK	23:15	
Air Blank	0.000	23:15	
Control Test	0.080	23:15	
Air Blank	0.000	23:16	
Subject Sample	# 1	23:16	Subject Test Refused
Breath Volume	0.000	23:16	
Air Blank	0.000	23:16	
Control Test	0.081	23:17	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:17	

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/26/2020 00:36:57

Diagnostics Check		OK	00:29
Air Blank		0.000	00:30
Control Test		0.081	00:30
Air Blank		0.000	00:30
Subject Sample	# 1	0.228	00:32
Breath Volume		1.253	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:34
Subject Sample	# 2	0.235	00:35
Breath Volume		1.453	00:35
Air Blank		0.000	00:36
Control Test		0.079	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: MCDANIELS

DATE/TIME OF TEST: 01/26/2020 02:11:46

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.081	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.127	02:07
Breath Volume	2.441	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.126	02:10
Breath Volume	1.445	02:10
Air Blank	0.000	02:10
Control Test	0.080	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: JARAMILLO

DATE/TIME OF TEST: 01/26/2020 04:46:54

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.080	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.086	04:42
Breath Volume	2.156	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:44
Subject Sample # 2	0.080	04:45
Breath Volume	1.250	04:45
Air Blank	0.000	04:45
Control Test	0.081	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: WEISS

DATE/TIME OF TEST: 01/27/2020 20:21:11

Diagnostics Check	OK	20:15
Air Blank	0.000	20:15
Control Test	0.081	20:15
Air Blank	0.000	20:16
Subject Sample # 1	0.000	20:16
Breath Volume	1.429	20:16
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample # 2	0.000	20:19
Breath Volume	3.234	20:19
Air Blank	0.000	20:20
Control Test	0.081	20:20
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: MAROTO CHACON

DATE/TIME OF TEST: 01/29/2020 05:43:28

Diagnostics Check	OK	05:36
Air Blank	0.000	05:37
Control Test	0.081	05:37
Air Blank	0.000	05:38
Subject Sample # 1	0.102	05:38
Breath Volume	1.109	05:38
Air Blank	0.000	05:39
Air Blank	0.000	05:41
Subject Sample # 2	0.114	05:41
Breath Volume	3.179	05:41
Air Blank	0.000	05:42
Control Test	0.080	05:42
Air Blank	0.000	05:43
Diagnostics Check	OK	05:43

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/30/2020 01:04:15

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.081	00:58
Air Blank		0.000	00:58
Subject Sample	# 1	0.134	00:59
Breath Volume		1.808	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.136	01:02
Breath Volume		3.136	01:02
Air Blank		0.000	01:03
Control Test		0.080	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: NEEDHAM

DATE/TIME OF TEST: 01/31/2020 01:37:55

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.081	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.095	01:33
Breath Volume	2.179	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.099	01:36
Breath Volume	3.457	01:36
Air Blank	0.000	01:36
Control Test	0.080	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 02/05/2020 18:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,KALEIGH,C
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: SCELFO,ZACHARY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 01/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: BILLIE

DATE/TIME OF TEST: 01/27/2020 22:46:50

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.088	22:42
Breath Volume	2.421	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.095	22:45
Breath Volume	3.187	22:45
Air Blank	0.000	22:45
Control Test	0.079	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 02/03/2020 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COREY,GAVIN,T
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: MARQUEZ,MICHAEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 01/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BRENNEMAN

DATE/TIME OF TEST: 01/30/2020 21:23:48

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.081	21:17
Air Blank	0.000	21:18
Subject Sample # 1	0.248	21:19
Breath Volume	3.140	21:19
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.242	21:22
Breath Volume	2.976	21:22
Air Blank	0.000	21:22
Control Test	0.081	21:23
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

Data Download Date/Time: 02/03/2020 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HOUSEHOLDER

DATE/TIME OF TEST: 01/15/2020 02:15:37

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.080	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.186	02:10
Breath Volume		2.242	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.194	02:14
Breath Volume		1.496	02:14
Air Blank		0.000	02:14
Control Test		0.078	02:15
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 02/14/2020 21:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTE,JACOB,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: GRIFFIN,DONALD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 01/25/2020 05:11:29

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:06
Subject Sample # 1	0.109	05:06
Breath Volume	1.855	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.112	05:09
Breath Volume	2.007	05:09
Air Blank	0.000	05:10
Control Test	0.078	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 02/14/2020 21:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,KEITH,C
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: HERNANDEZ,KEITH
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: BUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 01/26/2020 22:28:53

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:26	
Control Test	0.080	22:26	
Air Blank	0.000	22:27	
Subject Sample	# 1	22:27	Subject Test Refused
Breath Volume	0.000	22:27	
Air Blank	0.000	22:27	
Control Test	0.080	22:28	
Air Blank	0.000	22:28	
Diagnostics Check	OK	22:28	

Data Download Date/Time: 02/28/2020 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,AVERY,D
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GILBERT,KATRINA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 01/05/2020 22:51:27

Diagnostics Check		OK	22:39	
Air Blank		0.000	22:40	
Control Test		0.079	22:40	
Air Blank		0.000	22:41	
Subject Sample	# 1	0.240	22:44	Volume Not Met
Breath Volume		0.000	22:44	
Air Blank		0.000	22:44	
Air Blank		0.000	22:46	
Subject Sample	# 2	0.221	22:49	Volume Not Met
Breath Volume		0.113	22:49	
Air Blank		0.000	22:50	
Control Test		0.077	22:50	
Air Blank		0.000	22:51	
Diagnostics Check		OK	22:51	

Data Download Date/Time: 01/21/2020 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,AVERY,D
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GILBERT,KATRINA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 01/05/2020 23:22:44

Diagnostics Check		OK	23:13	
Air Blank		0.000	23:13	
Control Test		0.078	23:13	
Air Blank		0.000	23:14	
Subject Sample	# 1	0.222	23:17	Volume Not Met
Breath Volume		0.777	23:17	
Air Blank		0.000	23:18	
Air Blank		0.000	23:20	
Subject Sample	# 2		23:21	Subject Test Refused
Breath Volume		0.210	23:21	
Air Blank		0.000	23:21	
Control Test		0.077	23:22	
Air Blank		0.000	23:22	
Diagnostics Check		OK	23:22	

Data Download Date/Time: 01/21/2020 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CASTELLANO,ALAIN,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CASTELLANO,ALAIN

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 18:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: REINOSO

DATE/TIME OF TEST: 01/07/2020 18:31:53

Diagnostics Check	OK	18:29	
Air Blank	0.000	18:29	
Control Test	0.079	18:29	
Air Blank	0.000	18:30	
Subject Sample	# 1	18:30	Subject Test Refused
Breath Volume	0.000	18:30	
Air Blank	0.000	18:30	
Control Test	0.079	18:31	
Air Blank	0.000	18:31	
Diagnostics Check	OK	18:31	

Data Download Date/Time: 01/21/2020 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,BRANDON,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,M
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 01/10/2020 12:01:57

Diagnostics Check	OK	11:55
Air Blank	0.000	11:56
Control Test	0.080	11:56
Air Blank	0.000	11:57
Subject Sample # 1	0.000	11:57
Breath Volume	2.679	11:57
Air Blank	0.000	11:58
Air Blank	0.000	12:00
Subject Sample # 2	0.000	12:00
Breath Volume	4.570	12:00
Air Blank	0.000	12:01
Control Test	0.077	12:01
Air Blank	0.000	12:01
Diagnostics Check	OK	12:01

Data Download Date/Time: 01/21/2020 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,DAVID,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BRAVO,JIM
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: SOZA

DATE/TIME OF TEST: 01/12/2020 04:16:45

Diagnostics Check		OK	04:09
Air Blank		0.000	04:10
Control Test		0.080	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.038	04:12
Breath Volume		2.941	04:12
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.037	04:15
Breath Volume		2.398	04:15
Air Blank		0.000	04:15
Control Test		0.078	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 01/21/2020 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GARCIA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 08:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: MENDEZ SEVILLA

DATE/TIME OF TEST: 01/12/2020 09:21:57

Diagnostics Check	OK	09:15	
Air Blank	0.000	09:16	
Control Test	0.079	09:16	
Air Blank	0.000	09:17	
Subject Sample # 1	0.160	09:18	
Breath Volume	1.527	09:18	
Air Blank	0.000	09:19	
Air Blank		09:20	Improper Sample
Air Blank	0.000	09:21	
Control Test	0.080	09:21	
Air Blank	0.000	09:21	
Diagnostics Check	OK	09:21	

Data Download Date/Time: 01/21/2020 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GARCIA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 08:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: MENDEZ SEVILLA

DATE/TIME OF TEST: 01/12/2020 09:31:33

Diagnostics Check	OK	09:23
Air Blank	0.000	09:24
Control Test	0.079	09:24
Air Blank	0.000	09:25
Subject Sample # 1	0.158	09:26
Breath Volume	2.304	09:26
Air Blank	0.000	09:26
Air Blank	0.000	09:28
Subject Sample # 2	0.142	09:29
Breath Volume	1.523	09:29
Air Blank	0.000	09:30
Control Test	0.076	09:30
Air Blank	0.000	09:31
Diagnostics Check	OK	09:31

Data Download Date/Time: 01/21/2020 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPAZ,TAMMY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DE MELLO,KATHERINE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 12:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 01/14/2020 02:02:45

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.289	01:57
Breath Volume	3.917	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.285	02:01
Breath Volume	2.609	02:01
Air Blank	0.000	02:01
Control Test	0.076	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 01/21/2020 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASTRA,SUNEY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: LASTRA,SUNEY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 08:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:42

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 01/15/2020 09:11:18

Diagnostics Check		OK	09:03
Air Blank		0.000	09:03
Control Test		0.079	09:04
Air Blank		0.000	09:04
Subject Sample	# 1	0.190	09:06
Breath Volume		1.292	09:06
Air Blank		0.000	09:07
Air Blank		0.000	09:09
Subject Sample	# 2	0.195	09:09
Breath Volume		2.625	09:09
Air Blank		0.000	09:10
Control Test		0.077	09:10
Air Blank		0.000	09:11
Diagnostics Check		OK	09:11

Data Download Date/Time: 01/21/2020 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARGIA,FRANK,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TARGIA,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 10:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: FERREL

DATE/TIME OF TEST: 01/16/2020 11:42:55

Diagnostics Check	OK	11:35
Air Blank	0.000	11:36
Control Test	0.079	11:36
Air Blank	0.000	11:36
Subject Sample # 1	0.000	11:38
Breath Volume	5.398	11:38
Air Blank	0.000	11:39
Air Blank	0.000	11:41
Subject Sample # 2	0.000	11:41
Breath Volume	3.433	11:41
Air Blank	0.000	11:42
Control Test	0.079	11:42
Air Blank	0.000	11:42
Diagnostics Check	OK	11:42

Data Download Date/Time: 01/21/2020 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNNDEZ,NELSON,Y
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,NELSON
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ESTRELLA

DATE/TIME OF TEST: 01/18/2020 04:19:47

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:16	
Control Test	0.079	04:16	
Air Blank	0.000	04:17	
Subject Sample	# 1	04:18	Subject Test Refused
Breath Volume	0.000	04:18	
Air Blank	0.000	04:18	
Control Test	0.079	04:19	
Air Blank	0.000	04:19	
Diagnostics Check	OK	04:19	

Data Download Date/Time: 01/21/2020 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBUERNE,ALIAN,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DIAZ,DANELIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 07:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:54

SUBJECT'S LAST NAME: LOPEZ SANTOS

DATE/TIME OF TEST: 01/19/2020 08:36:27

Diagnostics Check	OK	08:27	
Air Blank	0.000	08:28	
Control Test	0.080	08:28	
Air Blank	0.000	08:29	
Subject Sample	# 1 0.228	08:29	
Breath Volume	1.562	08:29	
Air Blank	0.000	08:30	
Air Blank	0.000	08:32	
Subject Sample	# 2	08:34	RFI Detect
Breath Volume	0.382	08:34	
Air Blank	0.000	08:35	
Control Test	0.078	08:35	
Air Blank	0.000	08:36	
Diagnostics Check	OK	08:36	

Data Download Date/Time: 01/21/2020 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBUERNE,ALIAN,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DIAZ,DANELIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 07:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:54

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 01/19/2020 08:48:59

Diagnostics Check	OK	08:40	
Air Blank	0.000	08:41	
Control Test	0.077	08:41	
Air Blank	0.000	08:41	
Subject Sample # 1	0.201	08:44	
Breath Volume	1.304	08:44	
Air Blank	0.000	08:45	
Air Blank	0.000	08:46	
Subject Sample # 2	0.228	08:47	
Breath Volume	2.921	08:47	
Air Blank		08:48	Purge Fail
Air Blank	0.000	08:48	

Data Download Date/Time: 01/21/2020 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NARDONI,J,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: LOPEZ,L
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 01/19/2020 19:51:01

Diagnostics Check	OK	19:44	
Air Blank	0.000	19:45	
Control Test	0.078	19:45	
Air Blank	0.000	19:45	
Subject Sample	# 1 0.280	19:46	
Breath Volume	3.828	19:46	
Air Blank	0.000	19:47	
Air Blank	0.000	19:48	
Subject Sample	# 2 0.271	19:49	
Breath Volume	3.457	19:49	
Air Blank	0.000	19:50	
Control Test	0.069	19:50	Control Outside Tolerance
Air Blank	0.000	19:51	

Data Download Date/Time: 01/21/2020 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANTILLA,JUAN,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BRAVO,JIM
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: MANTILLA

DATE/TIME OF TEST: 01/21/2020 19:33:37

Diagnostics Check	OK	19:29	
Air Blank	0.000	19:30	
Control Test	0.080	19:30	
Air Blank	0.000	19:30	
Subject Sample	# 1	19:32	Subject Test Refused
Breath Volume	0.000	19:32	
Air Blank	0.000	19:32	
Control Test	0.080	19:33	
Air Blank	0.000	19:33	
Diagnostics Check	OK	19:33	

Data Download Date/Time: 02/28/2020 13:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANTILLA,JUAN,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BRAVO,JIM
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: M534432772460

DATE/TIME OF TEST: 01/21/2020 20:20:14

Diagnostics Check	OK	20:17	
Air Blank	0.000	20:17	
Control Test	0.080	20:18	
Air Blank	0.000	20:18	
Subject Sample	# 1	20:18	Subject Test Refused
Breath Volume	0.000	20:18	
Air Blank	0.000	20:19	
Control Test	0.080	20:19	
Air Blank	0.000	20:20	
Diagnostics Check	OK	20:20	

Data Download Date/Time: 02/28/2020 13:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERNANDEZ,ANGEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: FERNANDEZ,ANGEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 21:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CHICAS

DATE/TIME OF TEST: 01/23/2020 23:32:32

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:23	
Control Test	0.079	23:24	
Air Blank	0.000	23:24	
Subject Sample	# 1	23:27	No Sample Provided
Breath Volume	0.000	23:27	
Air Blank	0.000	23:28	
Air Blank	0.000	23:30	
Subject Sample	# 2	23:31	Subject Test Refused
Breath Volume	0.000	23:31	
Air Blank	0.000	23:31	
Control Test	0.078	23:31	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

Data Download Date/Time: 02/28/2020 13:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONZALEZ,ANTHONY,H

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: GONZALEZ,A

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 20:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: CAMARGO ZULUAGA

DATE/TIME OF TEST: 01/24/2020 21:50:34

Diagnostics Check	OK	21:49	
Air Blank	0.000	21:49	
Control Test	0.072	21:49	Control Outside Tolerance
Air Blank	0.000	21:50	

Data Download Date/Time: 02/28/2020 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,A,H
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,A
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: CAMARGO ZULUAGA

DATE/TIME OF TEST: 01/24/2020 21:59:36

Diagnostics Check	OK	21:53	
Air Blank	0.000	21:53	
Control Test	0.079	21:54	
Air Blank	0.000	21:54	
Subject Sample	# 1 0.176	21:55	
Breath Volume	3.300	21:55	
Air Blank	0.000	21:55	
Air Blank	0.000	21:57	
Subject Sample	# 2 0.169	21:57	
Breath Volume	3.324	21:57	
Air Blank	0.000	21:58	
Control Test	0.073	21:59	Control Outside Tolerance
Air Blank	0.000	21:59	

Data Download Date/Time: 02/28/2020 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,A,H
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,A
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: CAMARGO ZULUAGA

DATE/TIME OF TEST: 01/24/2020 22:09:26

Diagnostics Check	OK	22:03
Air Blank	0.000	22:03
Control Test	0.077	22:03
Air Blank	0.000	22:04
Subject Sample # 1	0.176	22:04
Breath Volume	3.226	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.172	22:07
Breath Volume	2.937	22:07
Air Blank	0.000	22:08
Control Test	0.075	22:08
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 02/28/2020 13:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMERO,JOHN,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ROMERO,J
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 09:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 01/25/2020 10:44:37

Diagnostics Check	OK	10:41	
Air Blank	0.000	10:41	
Control Test	0.080	10:42	
Air Blank	0.000	10:42	
Subject Sample	# 1	10:43	Subject Test Refused
Breath Volume	0.000	10:43	
Air Blank	0.000	10:43	
Control Test	0.079	10:44	
Air Blank	0.000	10:44	
Diagnostics Check	OK	10:44	

Data Download Date/Time: 02/28/2020 13:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,R,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: WASHINGTON,R
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: HARMON

DATE/TIME OF TEST: 01/27/2020 01:54:30

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.079	01:52	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:53	Subject Test Refused
Breath Volume	0.000	01:53	
Air Blank	0.000	01:53	
Control Test	0.079	01:53	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 02/28/2020 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CABRERA,JOSE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 01/31/2020 04:36:31

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.078	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.090	04:31
Breath Volume	2.097	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.084	04:34
Breath Volume	1.742	04:34
Air Blank	0.000	04:35
Control Test	0.077	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 02/28/2020 13:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: BACK

DATE/TIME OF TEST: 01/02/2020 21:22:08

Diagnostics Check		OK	21:15
Air Blank		0.000	21:16
Control Test		0.079	21:16
Air Blank		0.000	21:16
Subject Sample	# 1	0.112	21:17
Breath Volume		1.777	21:17
Air Blank		0.000	21:18
Air Blank		0.000	21:19
Subject Sample	# 2	0.121	21:20
Breath Volume		1.125	21:20
Air Blank		0.000	21:21
Control Test		0.079	21:21
Air Blank		0.000	21:22
Diagnostics Check		OK	21:22

Data Download Date/Time: 01/16/2020 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLINA,IGNACIO,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MOLINA,IGNACIO
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 01/03/2020 19:15:56

Diagnostics Check	OK	19:06	
Air Blank	0.000	19:06	
Control Test	0.080	19:06	
Air Blank	0.000	19:07	
Subject Sample	# 1 0.000	19:10	Volume Not Met
Breath Volume	0.296	19:10	
Air Blank	0.000	19:10	
Air Blank	0.000	19:12	
Subject Sample	# 2	19:14	Subject Test Refused
Breath Volume	0.109	19:14	
Air Blank	0.000	19:15	
Control Test	0.080	19:15	
Air Blank	0.000	19:15	
Diagnostics Check	OK	19:15	

Data Download Date/Time: 01/16/2020 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLINA,IGNACIO,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MOLINA,IGNACIO
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BRINDLE

DATE/TIME OF TEST: 01/05/2020 00:58:19

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.079	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.179	00:53
Breath Volume	1.839	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.194	00:56
Breath Volume	2.304	00:56
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 01/16/2020 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ DEL TORO,RUBEN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: GUTIERREZ DEL TORO,RUBEN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: BARAGONA

DATE/TIME OF TEST: 01/05/2020 18:13:59

Diagnostics Check	OK	18:12	
Air Blank	0.000	18:13	
Control Test	0.080	18:13	RFI Detect
Air Blank	0.000	18:13	

Data Download Date/Time: 01/16/2020 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ DEL TORO,RUBEN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: GUTIERREZ DEL TORO,RUBEN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: BARAGONA

DATE/TIME OF TEST: 01/05/2020 18:18:57

Diagnostics Check	OK	18:15	
Air Blank	0.000	18:16	
Control Test	0.080	18:16	
Air Blank	0.000	18:17	
Subject Sample	# 1	18:17	Subject Test Refused
Breath Volume	0.000	18:17	
Air Blank	0.000	18:18	
Control Test	0.080	18:18	
Air Blank	0.000	18:18	
Diagnostics Check	OK	18:18	

Data Download Date/Time: 01/16/2020 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 01/05/2020 21:36:41

Diagnostics Check	OK	21:30
Air Blank	0.000	21:30
Control Test	0.080	21:31
Air Blank	0.000	21:31
Subject Sample # 1	0.185	21:32
Breath Volume	2.406	21:32
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.182	21:35
Breath Volume	2.085	21:35
Air Blank	0.000	21:35
Control Test	0.080	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 01/16/2020 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,R
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: SOARES

DATE/TIME OF TEST: 01/06/2020 22:28:53

Diagnostics Check	OK	22:22
Air Blank	0.000	22:23
Control Test	0.079	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.000	22:24
Breath Volume	3.859	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.000	22:27
Breath Volume	3.859	22:27
Air Blank	0.000	22:27
Control Test	0.080	22:28
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 01/16/2020 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPPE,ZACHARY,D
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: WEICHHAN,DYLAN
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 01/07/2020 20:02:18

Diagnostics Check	OK	19:55
Air Blank	0.000	19:55
Control Test	0.080	19:56
Air Blank	0.000	19:56
Subject Sample # 1	0.123	19:57
Breath Volume	1.699	19:57
Air Blank	0.000	19:57
Air Blank	0.000	19:59
Subject Sample # 2	0.127	20:00
Breath Volume	2.343	20:00
Air Blank	0.000	20:01
Control Test	0.078	20:01
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

Data Download Date/Time: 01/16/2020 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUGAN,DANIEL,T
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: COELO,ANA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: DUGAN

DATE/TIME OF TEST: 01/10/2020 02:38:21

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:35	
Control Test	0.079	02:36	
Air Blank	0.000	02:36	
Subject Sample	# 1	02:36	Subject Test Refused
Breath Volume	0.000	02:36	
Air Blank	0.000	02:37	
Control Test	0.079	02:37	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 01/16/2020 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLINA,IGNACIO,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: COELO,ANA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: DUGAN

DATE/TIME OF TEST: 01/10/2020 02:45:24

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:43	
Control Test	0.079	02:43	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:44	Subject Test Refused
Breath Volume	0.000	02:44	
Air Blank	0.000	02:44	
Control Test	0.078	02:44	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 01/16/2020 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: TUCKUS

DATE/TIME OF TEST: 01/11/2020 03:52:39

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.079	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.071	03:48
Breath Volume	1.246	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.074	03:51
Breath Volume	2.757	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 01/16/2020 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: LOJA QUIZHPE

DATE/TIME OF TEST: 01/12/2020 21:41:20

Diagnostics Check		OK	21:33
Air Blank		0.000	21:34
Control Test		0.079	21:34
Air Blank		0.000	21:35
Subject Sample	# 1	0.169	21:36
Breath Volume		2.402	21:36
Air Blank		0.000	21:36
Air Blank		0.000	21:38
Subject Sample	# 2	0.170	21:39
Breath Volume		1.503	21:39
Air Blank		0.000	21:40
Control Test		0.078	21:40
Air Blank		0.000	21:41
Diagnostics Check		OK	21:41

Data Download Date/Time: 01/16/2020 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RIVAS,JUSTIN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: NEJI

DATE/TIME OF TEST: 01/17/2020 04:27:23

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.079	04:24	
Air Blank	0.000	04:25	
Subject Sample	# 1	04:26	Subject Test Refused
Breath Volume	0.000	04:26	
Air Blank	0.000	04:26	
Control Test	0.079	04:26	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 02/04/2020 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLINA,IGNACIO,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MOLINA,IGNACIO
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 01/18/2020 19:05:29

Diagnostics Check	OK	19:01	
Air Blank	0.000	19:02	
Control Test	0.079	19:02	
Air Blank	0.000	19:03	
Subject Sample	# 1	19:04	Subject Test Refused
Breath Volume	0.000	19:04	
Air Blank	0.000	19:04	
Control Test	0.079	19:04	
Air Blank	0.000	19:05	
Diagnostics Check	OK	19:05	

Data Download Date/Time: 02/04/2020 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,JOEL,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: JACOBS,GARRETT
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: STRAHSBURG

DATE/TIME OF TEST: 01/18/2020 23:46:05

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.078	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.164	23:41
Breath Volume	2.820	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:43
Subject Sample # 2	0.167	23:44
Breath Volume	1.480	23:44
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 02/04/2020 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MILLER,JAIME
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: MOLLINEDA DIAZ

DATE/TIME OF TEST: 01/21/2020 04:18:49

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.079	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.176	04:13
Breath Volume		1.308	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.182	04:17
Breath Volume		1.101	04:17
Air Blank		0.000	04:17
Control Test		0.078	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 02/04/2020 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: FRANTA

DATE/TIME OF TEST: 01/22/2020 02:26:16

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
Control Test	0.079	02:23	
Air Blank	0.000	02:24	
Subject Sample	# 1	02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:25	
Control Test	0.078	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 02/04/2020 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: ZIMMERMAN

DATE/TIME OF TEST: 01/25/2020 00:45:47

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.078	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.154	00:41
Breath Volume	1.328	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.159	00:44
Breath Volume	1.351	00:44
Air Blank	0.000	00:44
Control Test	0.077	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 02/04/2020 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MILLER,JAIME
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: HEDMAN PERDOMO

DATE/TIME OF TEST: 01/25/2020 01:58:00

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:55	
Control Test	0.078	01:55	
Air Blank	0.000	01:56	
Subject Sample	# 1	01:56	Subject Test Refused
Breath Volume	0.000	01:56	
Air Blank	0.000	01:57	
Control Test	0.078	01:57	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:58	

Data Download Date/Time: 02/04/2020 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 01/26/2020 05:04:26

Diagnostics Check		OK	04:52	
Air Blank		0.000	04:52	
Control Test		0.079	04:53	
Air Blank		0.000	04:53	
Subject Sample	# 1	0.090	04:56	Volume Not Met
Breath Volume		0.093	04:56	
Air Blank		0.000	04:57	
Air Blank		0.000	04:59	
Subject Sample	# 2	0.153	04:59	
Breath Volume		1.937	04:59	
Air Blank		0.000	05:00	
Air Blank		0.000	05:02	
Subject Sample	# 3	0.149	05:02	
Breath Volume		2.019	05:02	
Air Blank		0.000	05:03	
Control Test		0.078	05:03	
Air Blank		0.000	05:04	
Diagnostics Check		OK	05:04	

Data Download Date/Time: 02/04/2020 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 01/27/2020 04:55:39

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:52	
Control Test	0.078	04:53	
Air Blank	0.000	04:53	
Subject Sample	# 1	04:54	Subject Test Refused
Breath Volume	0.000	04:54	
Air Blank	0.000	04:54	
Control Test	0.079	04:55	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:55	

Data Download Date/Time: 02/04/2020 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,JOEL,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RIOS,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: SONDAG

DATE/TIME OF TEST: 01/30/2020 01:23:04

Diagnostics Check		OK	01:16
Air Blank		0.000	01:17
Control Test		0.079	01:17
Air Blank		0.000	01:17
Subject Sample	# 1	0.289	01:18
Breath Volume		2.308	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:20
Subject Sample	# 2	0.295	01:21
Breath Volume		2.582	01:21
Air Blank		0.000	01:22
Control Test		0.079	01:22
Air Blank		0.000	01:22
Diagnostics Check		OK	01:23

Data Download Date/Time: 02/04/2020 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 01/30/2020 03:18:29

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.078	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.156	03:13
Breath Volume		1.187	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.156	03:16
Breath Volume		1.324	03:16
Air Blank		0.000	03:17
Control Test		0.079	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 02/04/2020 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SWOGGER, EDWARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 01/30/2020 19:23:52

Diagnostics Check	OK	19:20	
Air Blank	0.000	19:21	
Control Test	0.078	19:21	
Air Blank	0.000	19:22	
Subject Sample	# 1	19:22	RFI Detect
Breath Volume	0.000	19:22	
Air Blank	0.000	19:22	
Control Test	0.078	19:23	
Air Blank	0.000	19:23	
Diagnostics Check	OK	19:23	

Data Download Date/Time: 02/04/2020 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SWOGGER, EDWARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 01/30/2020 19:32:10

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.078	19:26
Air Blank	0.000	19:26
Subject Sample # 1	0.134	19:27
Breath Volume	3.179	19:27
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.115	19:30
Breath Volume	2.816	19:30
Air Blank	0.000	19:31
Control Test	0.079	19:31
Air Blank	0.000	19:32
Diagnostics Check	OK	19:32

Data Download Date/Time: 02/04/2020 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SWOGGER, EDWARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 01/30/2020 19:44:39

Diagnostics Check	OK	19:38
Air Blank	0.000	19:38
Control Test	0.078	19:39
Air Blank	0.000	19:39
Subject Sample # 1	0.133	19:40
Breath Volume	2.851	19:40
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.129	19:43
Breath Volume	3.273	19:43
Air Blank	0.000	19:43
Control Test	0.078	19:44
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

Data Download Date/Time: 02/04/2020 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,JOEL,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: FARR,JASON
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CARRERAS DE ARA

DATE/TIME OF TEST: 01/31/2020 03:10:12

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:58	
Control Test		0.079	02:58	
Air Blank		0.000	02:58	
Subject Sample	# 1	0.123	03:02	Volume Not Met
Breath Volume		0.000	03:02	
Air Blank		0.000	03:02	
Air Blank		0.000	03:04	
Subject Sample	# 2	0.146	03:05	
Breath Volume		1.496	03:05	
Air Blank		0.000	03:06	
Air Blank		0.000	03:08	
Subject Sample	# 3		03:08	Subject Test Refused
Breath Volume		0.000	03:08	
Air Blank		0.000	03:09	
Control Test		0.077	03:09	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 02/04/2020 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,JOEL,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: FARR,JASON
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: CARRERAS DE ARA

DATE/TIME OF TEST: 01/31/2020 03:57:26

Diagnostics Check		OK	03:44	
Air Blank		0.000	03:44	
Control Test		0.078	03:45	
Air Blank		0.000	03:45	
Subject Sample	# 1	0.123	03:48	Volume Not Met
Breath Volume		0.796	03:48	
Air Blank		0.000	03:49	
Air Blank		0.000	03:51	
Subject Sample	# 2	0.132	03:51	
Breath Volume		1.714	03:51	
Air Blank		0.000	03:52	
Air Blank		0.000	03:54	
Subject Sample	# 3		03:55	Subject Test Refused
Breath Volume		0.054	03:55	
Air Blank		0.000	03:56	
Control Test		0.078	03:56	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 02/04/2020 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STEELE,MICHAEL
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 07:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: HATTAWAY

DATE/TIME OF TEST: 01/01/2020 09:16:52

Diagnostics Check	OK	09:09
Air Blank	0.000	09:10
Control Test	0.080	09:10
Air Blank	0.000	09:11
Subject Sample # 1	0.180	09:12
Breath Volume	3.492	09:12
Air Blank	0.000	09:12
Air Blank	0.000	09:14
Subject Sample # 2	0.183	09:15
Breath Volume	3.500	09:15
Air Blank	0.000	09:15
Control Test	0.078	09:16
Air Blank	0.000	09:16
Diagnostics Check	OK	09:16

Data Download Date/Time: 01/16/2020 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: CAFFI

DATE/TIME OF TEST: 01/04/2020 05:22:21

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.079	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.123	05:17
Breath Volume	1.406	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.126	05:20
Breath Volume	1.332	05:20
Air Blank	0.000	05:21
Control Test	0.078	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 01/16/2020 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PRESUME,JEHUD
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: NORWOOD

DATE/TIME OF TEST: 01/05/2020 20:55:32

Diagnostics Check	OK	20:49
Air Blank	0.000	20:49
Control Test	0.079	20:50
Air Blank	0.000	20:50
Subject Sample # 1	0.224	20:50
Breath Volume	2.273	20:50
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.234	20:53
Breath Volume	1.820	20:53
Air Blank	0.000	20:54
Control Test	0.080	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 01/16/2020 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ANTOINE,KAHLYL
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: GENET

DATE/TIME OF TEST: 01/05/2020 21:40:19

Diagnostics Check		OK	21:27	
Air Blank		0.000	21:27	
Control Test		0.079	21:28	
Air Blank		0.000	21:28	
Subject Sample	# 1	0.184	21:31	Volume Not Met
Breath Volume		0.917	21:31	
Air Blank		0.000	21:32	
Air Blank		0.000	21:34	
Subject Sample	# 2	0.215	21:34	
Breath Volume		1.691	21:34	
Air Blank		0.000	21:35	
Air Blank		0.000	21:37	
Subject Sample	# 3	0.223	21:38	
Breath Volume		2.640	21:38	
Air Blank		0.000	21:39	
Control Test		0.077	21:39	
Air Blank		0.000	21:40	
Diagnostics Check		OK	21:40	

Data Download Date/Time: 01/16/2020 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SHIN

DATE/TIME OF TEST: 01/06/2020 03:37:05

Diagnostics Check		OK	03:24	
Air Blank		0.000	03:25	
Control Test		0.079	03:25	
Air Blank		0.000	03:26	
Subject Sample	# 1	0.139	03:29	Volume Not Met
Breath Volume		0.000	03:29	
Air Blank		0.000	03:30	
Air Blank		0.000	03:31	
Subject Sample	# 2	0.147	03:32	
Breath Volume		1.910	03:32	
Air Blank		0.000	03:33	
Air Blank		0.000	03:34	
Subject Sample	# 3	0.163	03:35	
Breath Volume		2.898	03:35	
Air Blank		0.000	03:36	
Control Test		0.078	03:36	
Air Blank		0.000	03:36	
Diagnostics Check		OK	03:37	

Data Download Date/Time: 01/16/2020 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: MCGILL

DATE/TIME OF TEST: 01/07/2020 00:55:31

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.199	00:51
Breath Volume	3.222	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.183	00:53
Breath Volume	1.636	00:53
Air Blank	0.000	00:54
Control Test	0.078	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 01/16/2020 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'DOR,ROBERT,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: BEASLEY

DATE/TIME OF TEST: 01/07/2020 23:22:52

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.079	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.146	23:18
Breath Volume	2.332	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.147	23:21
Breath Volume	2.332	23:21
Air Blank	0.000	23:21
Control Test	0.080	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 01/16/2020 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/08/2020 02:37:24

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.184	02:32
Breath Volume	2.804	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.190	02:35
Breath Volume	2.785	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 01/16/2020 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 01/10/2020 04:33:36

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.079	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.146	04:29
Breath Volume	1.718	04:29
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.145	04:32
Breath Volume	1.746	04:32
Air Blank	0.000	04:32
Control Test	0.078	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 01/16/2020 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERNANDEZ,ERNESTO
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: PASCO CABEZA

DATE/TIME OF TEST: 01/10/2020 21:01:15

Diagnostics Check	OK	20:54
Air Blank	0.000	20:55
Control Test	0.080	20:55
Air Blank	0.000	20:56
Subject Sample # 1	0.115	20:56
Breath Volume	1.757	20:56
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample # 2	0.124	20:59
Breath Volume	2.730	20:59
Air Blank	0.000	21:00
Control Test	0.080	21:00
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time: 01/16/2020 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: GARWOOD

DATE/TIME OF TEST: 01/11/2020 04:05:49

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.149	04:00
Breath Volume	1.921	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.142	04:04
Breath Volume	1.847	04:04
Air Blank	0.000	04:04
Control Test	0.079	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 01/16/2020 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WASHINGTON,YVETTE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: MONCADA

DATE/TIME OF TEST: 01/11/2020 04:23:57

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.198	04:19
Breath Volume	2.234	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.190	04:22
Breath Volume	2.175	04:22
Air Blank	0.000	04:23
Control Test	0.078	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 01/16/2020 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,MARK,D
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 01/12/2020 23:04:41

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.078	22:58
Air Blank	0.000	22:59
Subject Sample # 1	0.137	23:00
Breath Volume	2.953	23:00
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.132	23:03
Breath Volume	2.929	23:03
Air Blank	0.000	23:03
Control Test	0.079	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 01/16/2020 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CABRERA,JOSEPH
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/13/2020 23:29:18

Diagnostics Check		OK	23:22
Air Blank		0.000	23:23
Control Test		0.079	23:23
Air Blank		0.000	23:24
Subject Sample	# 1	0.201	23:24
Breath Volume		2.308	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.183	23:27
Breath Volume		1.503	23:27
Air Blank		0.000	23:28
Control Test		0.079	23:28
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 01/16/2020 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: STINE

DATE/TIME OF TEST: 01/16/2020 00:13:44

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.080	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.290	00:08
Breath Volume	1.214	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:10
Subject Sample # 2	0.283	00:12
Breath Volume	1.167	00:12
Air Blank	0.000	00:12
Control Test	0.079	00:13
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 01/16/2020 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, YVETTE, D
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HARRIS, DELUCA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 08:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:52

SUBJECT'S LAST NAME: DAMPMAN

DATE/TIME OF TEST: 01/16/2020 09:28:58

Diagnostics Check	OK	09:22
Air Blank	0.000	09:22
Control Test	0.080	09:22
Air Blank	0.000	09:23
Subject Sample # 1	0.000	09:24
Breath Volume	3.121	09:24
Air Blank	0.000	09:24
Air Blank	0.000	09:26
Subject Sample # 2	0.000	09:27
Breath Volume	1.906	09:27
Air Blank	0.000	09:28
Control Test	0.081	09:28
Air Blank	0.000	09:28
Diagnostics Check	OK	09:28

Data Download Date/Time: 01/16/2020 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'DOR,ROBERT,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: CLEARY

DATE/TIME OF TEST: 01/20/2020 20:50:50

Diagnostics Check	OK	20:44
Air Blank	0.000	20:44
Control Test	0.081	20:44
Air Blank	0.000	20:45
Subject Sample # 1	0.224	20:46
Breath Volume	3.070	20:46
Air Blank	0.000	20:46
Air Blank	0.000	20:48
Subject Sample # 2	0.224	20:49
Breath Volume	4.273	20:49
Air Blank	0.000	20:49
Control Test	0.079	20:50
Air Blank	0.000	20:50
Diagnostics Check	OK	20:50

Data Download Date/Time: 02/21/2020 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: DESHOMMES

DATE/TIME OF TEST: 01/21/2020 23:09:00

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.080	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.199	23:04
Breath Volume	2.128	23:04
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.214	23:07
Breath Volume	1.324	23:07
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:09

Data Download Date/Time: 02/21/2020 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,B,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,B
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BOTZEK

DATE/TIME OF TEST: 01/29/2020 01:40:15

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.055	01:34
Breath Volume	1.593	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.054	01:38
Breath Volume	1.453	01:38
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 02/21/2020 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: DACOSTA

DATE/TIME OF TEST: 01/17/2020 21:28:54

Diagnostics Check	OK	21:22
Air Blank	0.000	21:23
Control Test	0.080	21:23
Air Blank	0.000	21:23
Subject Sample # 1	0.193	21:24
Breath Volume	1.343	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:26
Subject Sample # 2	0.188	21:27
Breath Volume	1.359	21:27
Air Blank	0.000	21:27
Control Test	0.080	21:28
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

Data Download Date/Time: 02/14/2020 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BOYD,ADAM
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CEDENO TUAREZ

DATE/TIME OF TEST: 01/19/2020 01:38:52

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.081	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.181	01:34
Breath Volume	2.343	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.182	01:37
Breath Volume	2.890	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 02/14/2020 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: SUSTACHE

DATE/TIME OF TEST: 01/19/2020 03:48:12

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:33	
Control Test		0.080	03:33	
Air Blank		0.000	03:34	
Subject Sample	# 1	0.159	03:37	Volume Not Met
Breath Volume		0.949	03:37	
Air Blank		0.000	03:38	
Air Blank		0.000	03:39	
Subject Sample	# 2	0.153	03:40	
Breath Volume		1.707	03:40	
Air Blank		0.000	03:41	
Air Blank		0.000	03:43	
Subject Sample	# 3	0.158	03:46	Volume Not Met
Breath Volume		0.074	03:46	
Air Blank		0.000	03:47	
Control Test		0.079	03:47	
Air Blank		0.000	03:48	
Diagnostics Check		OK	03:48	

Data Download Date/Time: 02/14/2020 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOHN,MARK
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 01/20/2020 01:57:35

Diagnostics Check		OK	01:45	
Air Blank		0.000	01:46	
Control Test		0.080	01:46	
Air Blank		0.000	01:47	
Subject Sample	# 1	0.200	01:50	Volume Not Met
Breath Volume		0.933	01:50	
Air Blank		0.000	01:51	
Air Blank		0.000	01:52	
Subject Sample	# 2	0.213	01:55	Volume Not Met
Breath Volume		0.714	01:55	
Air Blank		0.000	01:56	
Control Test		0.076	01:56	
Air Blank		0.000	01:57	
Diagnostics Check		OK	01:57	

Data Download Date/Time: 02/14/2020 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: LUGO

DATE/TIME OF TEST: 01/20/2020 04:19:55

Diagnostics Check		OK	04:13
Air Blank		0.000	04:14
Control Test		0.080	04:14
Air Blank		0.000	04:14
Subject Sample	# 1	0.200	04:15
Breath Volume		2.636	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:17
Subject Sample	# 2	0.187	04:18
Breath Volume		2.347	04:18
Air Blank		0.000	04:18
Control Test		0.080	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 02/14/2020 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: TICE

DATE/TIME OF TEST: 01/20/2020 23:13:48

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.145	23:09
Breath Volume	1.890	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.144	23:12
Breath Volume	2.097	23:12
Air Blank	0.000	23:12
Control Test	0.080	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 02/14/2020 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HETTCHEN,JOHN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: DETOMA

DATE/TIME OF TEST: 01/22/2020 03:54:04

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.090	03:49
Breath Volume	3.339	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:51
Subject Sample # 2	0.087	03:52
Breath Volume	1.917	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:54

Data Download Date/Time: 02/14/2020 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,DAVID
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: RENSTROM

DATE/TIME OF TEST: 01/24/2020 02:16:37

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.081	02:10
Air Blank		0.000	02:11
Subject Sample	# 1	0.132	02:11
Breath Volume		1.441	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.125	02:14
Breath Volume		1.132	02:14
Air Blank		0.000	02:15
Control Test		0.081	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 02/14/2020 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: NUTTLE

DATE/TIME OF TEST: 01/25/2020 20:43:59

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.081	20:37
Air Blank	0.000	20:37
Subject Sample # 1	0.193	20:39
Breath Volume	1.382	20:39
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.200	20:42
Breath Volume	1.363	20:42
Air Blank	0.000	20:43
Control Test	0.080	20:43
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 02/14/2020 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: PEREZ-DIAZ

DATE/TIME OF TEST: 01/25/2020 21:29:56

Diagnostics Check	OK	21:22
Air Blank	0.000	21:23
Control Test	0.080	21:23
Air Blank	0.000	21:23
Subject Sample # 1	0.223	21:25
Breath Volume	2.078	21:25
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.234	21:28
Breath Volume	2.406	21:28
Air Blank	0.000	21:28
Control Test	0.079	21:29
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 02/14/2020 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 01/26/2020 02:34:17

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.136	02:27
Breath Volume	1.285	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:29
Subject Sample # 2	0.129	02:32
Breath Volume	1.304	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 02/14/2020 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: WHEELS

DATE/TIME OF TEST: 01/26/2020 23:19:28

Diagnostics Check	OK	23:12
Air Blank	0.000	23:13
Control Test	0.081	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.137	23:14
Breath Volume	2.054	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.118	23:17
Breath Volume	1.222	23:17
Air Blank	0.000	23:18
Control Test	0.079	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 02/14/2020 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 01/27/2020 23:34:55

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.081	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.282	23:29
Breath Volume	1.843	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.293	23:33
Breath Volume	1.964	23:33
Air Blank	0.000	23:33
Control Test	0.079	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 02/14/2020 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 01/28/2020 01:00:38

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.080	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.242	00:55
Breath Volume		1.640	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.231	00:58
Breath Volume		1.953	00:58
Air Blank		0.000	00:59
Control Test		0.077	01:00
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 02/14/2020 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: HATFIELD

DATE/TIME OF TEST: 01/28/2020 04:09:21

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.199	04:03
Breath Volume	1.601	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.198	04:07
Breath Volume	2.199	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 02/14/2020 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJEDA, ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: FRANKE

DATE/TIME OF TEST: 01/28/2020 09:22:52

Diagnostics Check	OK	09:16
Air Blank	0.000	09:16
Control Test	0.080	09:17
Air Blank	0.000	09:17
Subject Sample # 1	0.369	09:18
Breath Volume	3.531	09:18
Air Blank	0.000	09:19
Air Blank	0.000	09:20
Subject Sample # 2	0.364	09:21
Breath Volume	3.390	09:21
Air Blank	0.000	09:21
Control Test	0.079	09:22
Air Blank	0.000	09:22
Diagnostics Check	OK	09:22

Data Download Date/Time: 02/14/2020 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAWRENCE,TIMOTHY
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:41

SUBJECT'S LAST NAME: KERN-HAYES

DATE/TIME OF TEST: 01/28/2020 16:38:32

Diagnostics Check	OK	16:31
Air Blank	0.000	16:31
Control Test	0.080	16:32
Air Blank	0.000	16:32
Subject Sample # 1	0.063	16:33
Breath Volume	1.367	16:33
Air Blank	0.000	16:34
Air Blank	0.000	16:35
Subject Sample # 2	0.061	16:36
Breath Volume	1.472	16:36
Air Blank	0.000	16:37
Control Test	0.080	16:37
Air Blank	0.000	16:38
Diagnostics Check	OK	16:38

Data Download Date/Time: 02/14/2020 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PRADO,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 01/28/2020 21:28:34

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.081	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.255	21:23
Breath Volume	1.351	21:23
Air Blank	0.000	21:24
Air Blank	0.000	21:26
Subject Sample # 2	0.261	21:26
Breath Volume	1.472	21:26
Air Blank	0.000	21:27
Control Test	0.080	21:27
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

Data Download Date/Time: 02/14/2020 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: SARAVIA

DATE/TIME OF TEST: 01/29/2020 00:11:33

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.080	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.058	00:06
Breath Volume	1.984	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.054	00:09
Breath Volume	1.195	00:09
Air Blank	0.000	00:10
Control Test	0.079	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 02/14/2020 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOLOWAY,ALAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 01/29/2020 01:13:14

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.183	01:08
Breath Volume		3.367	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.180	01:11
Breath Volume		3.421	01:11
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 02/14/2020 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROMERO,JUAN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/29/2020 02:31:55

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.176	02:27
Breath Volume	1.535	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.185	02:30
Breath Volume	2.179	02:30
Air Blank	0.000	02:30
Control Test	0.080	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 02/14/2020 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: LUIS GARCIA

DATE/TIME OF TEST: 01/29/2020 02:57:02

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.222	02:52
Breath Volume	3.398	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.228	02:55
Breath Volume	3.429	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:57

Data Download Date/Time: 02/14/2020 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,DAVID
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: ANTONUCCI

DATE/TIME OF TEST: 01/29/2020 21:44:12

Diagnostics Check	OK	21:37
Air Blank	0.000	21:38
Control Test	0.081	21:38
Air Blank	0.000	21:39
Subject Sample # 1	0.000	21:39
Breath Volume	2.750	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.000	21:42
Breath Volume	3.207	21:42
Air Blank	0.000	21:43
Control Test	0.081	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 02/14/2020 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: RODRIGUEZ MONTENEGRO

DATE/TIME OF TEST: 01/30/2020 03:34:44

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.081	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.139	03:29
Breath Volume	2.453	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.136	03:33
Breath Volume	1.960	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 02/14/2020 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DELANEY,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:38

SUBJECT'S LAST NAME: CONTRERAS CRUZ

DATE/TIME OF TEST: 01/30/2020 10:11:54

Diagnostics Check	OK	10:05
Air Blank	0.000	10:05
Control Test	0.081	10:06
Air Blank	0.000	10:06
Subject Sample # 1	0.125	10:07
Breath Volume	2.980	10:07
Air Blank	0.000	10:08
Air Blank	0.000	10:09
Subject Sample # 2	0.125	10:10
Breath Volume	2.976	10:10
Air Blank	0.000	10:10
Control Test	0.081	10:11
Air Blank	0.000	10:11
Diagnostics Check	OK	10:11

Data Download Date/Time: 02/14/2020 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 01/31/2020 04:01:04

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.140	03:56
Breath Volume	1.304	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:58
Subject Sample # 2	0.142	03:59
Breath Volume	1.484	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:01

Data Download Date/Time: 02/14/2020 13:18