

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUICK,BENJAMIN,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: QUICK,BENJAMIN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 01/10/2020 21:12:46

---

Diagnostics Check	OK	21:06
Air Blank	0.000	21:06
Control Test	0.080	21:07
Air Blank	0.000	21:07
Subject Sample # 1	0.164	21:08
Breath Volume	1.910	21:08
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 2	0.162	21:11
Breath Volume	1.601	21:11
Air Blank	0.000	21:11
Control Test	0.080	21:12
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

---

Data Download Date/Time: 02/15/2020 13:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/11/2020 02:13:38

---

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:10	
Control Test	0.080	02:11	
Air Blank	0.000	02:11	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:12	
Control Test	0.080	02:13	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

---

Data Download Date/Time: 02/15/2020 13:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: OGLINE

DATE/TIME OF TEST: 01/11/2020 02:26:48

---

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.080	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.169	02:21
Breath Volume	4.195	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.172	02:25
Breath Volume	3.710	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

---

Data Download Date/Time: 02/15/2020 13:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PODHYSKI,PHILIP,A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: MASON,SAMUELL  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 01/11/2020 03:56:31

---

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.176	03:51
Breath Volume	3.375	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.176	03:54
Breath Volume	3.406	03:54
Air Blank	0.000	03:55
Control Test	0.078	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

---

Data Download Date/Time: 02/15/2020 13:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: OLARTE

DATE/TIME OF TEST: 01/12/2020 03:01:46

---

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.217	02:57
Breath Volume	1.414	02:57
Air Blank	0.000	02:58
Air Blank	0.000	02:59
Subject Sample # 2	0.216	03:00
Breath Volume	1.835	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

---

Data Download Date/Time: 02/15/2020 13:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PODHYSKI,PHILIP,A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: RIVERA GUTIERREZ

DATE/TIME OF TEST: 01/13/2020 03:37:33

---

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:35	
Control Test	0.080	03:35	
Air Blank	0.000	03:35	
Subject Sample	# 1	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Control Test	0.080	03:36	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

---

Data Download Date/Time: 02/15/2020 13:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GWODZ,OFC,R  
BREATH TEST OPERATOR AGENCY: Bradenton PD  
ARRESTING OFFICER: WENTLAND,ALEX  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 01/14/2020 04:21:52

---

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:19	
Control Test	0.080	04:19	
Air Blank	0.000	04:20	
Subject Sample	# 1 0.174	04:20	
Breath Volume	1.593	04:20	
Air Blank		04:21	Purge Fail
Air Blank	0.000	04:21	

---

Data Download Date/Time: 02/15/2020 13:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GWODZ,OFC,R  
BREATH TEST OPERATOR AGENCY: Bradenton PD  
ARRESTING OFFICER: WENTLAND,ALEX  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 01/14/2020 04:30:19

---

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.183	04:25
Breath Volume	1.773	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.183	04:28
Breath Volume	1.511	04:28
Air Blank	0.000	04:29
Control Test	0.078	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

---

Data Download Date/Time: 02/15/2020 13:54



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 01/17/2020 01:27:48

---

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:22	
Control Test	0.081	01:22	
Air Blank	0.000	01:23	
Subject Sample	# 1	01:26	Subject Test Refused
Breath Volume	0.000	01:26	
Air Blank	0.000	01:26	
Control Test	0.080	01:27	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:27	

---

Data Download Date/Time: 02/15/2020 13:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRETT,CODY,R  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BARRETT,CODY  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 01/17/2020 02:48:17

---

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.080	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.192	02:43
Breath Volume	3.222	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:45
Subject Sample # 2	0.186	02:46
Breath Volume	2.925	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

---

Data Download Date/Time: 02/15/2020 13:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUMMINGS,GARY,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: CUMMINGS,GARY  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 01/17/2020 06:17:19

---

Diagnostics Check	OK	06:14	
Air Blank	0.000	06:14	
Control Test	0.080	06:14	
Air Blank	0.000	06:15	
Subject Sample	# 1	06:15	Subject Test Refused
Breath Volume	0.000	06:15	
Air Blank	0.000	06:16	
Control Test	0.081	06:16	
Air Blank	0.000	06:17	
Diagnostics Check	OK	06:17	

---

Data Download Date/Time: 02/15/2020 13:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/18/2020 22:26:07

---

Diagnostics Check	OK	22:19
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:20
Subject Sample # 1	0.151	22:21
Breath Volume	1.515	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:23
Subject Sample # 2	0.154	22:24
Breath Volume	1.511	22:24
Air Blank	0.000	22:25
Control Test	0.080	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

---

Data Download Date/Time: 02/15/2020 13:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: FORESTANDI

DATE/TIME OF TEST: 01/22/2020 21:01:34

---

Diagnostics Check	OK	20:55
Air Blank	0.000	20:55
Control Test	0.080	20:56
Air Blank	0.000	20:56
Subject Sample # 1	0.354	20:57
Breath Volume	2.882	20:57
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample # 2	0.338	20:59
Breath Volume	2.265	20:59
Air Blank	0.000	21:00
Control Test	0.080	21:00
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

---

Data Download Date/Time: 02/15/2020 13:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: MONTEMAGNO,CARLO  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: SCHALLER

DATE/TIME OF TEST: 01/22/2020 22:35:43

---

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.080	22:30
Air Blank	0.000	22:30
Subject Sample # 1	0.071	22:31
Breath Volume	1.816	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:33
Subject Sample # 2	0.072	22:34
Breath Volume	1.378	22:34
Air Blank	0.000	22:34
Control Test	0.080	22:35
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

---

Data Download Date/Time: 02/15/2020 13:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/25/2020 02:44:51

---

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.081	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.104	02:40
Breath Volume	1.777	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:42
Subject Sample # 2	0.107	02:43
Breath Volume	2.160	02:43
Air Blank	0.000	02:43
Control Test	0.079	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

---

Data Download Date/Time: 02/15/2020 13:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRETT,CODY,R  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BARRETT,CODY  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: LOPEZ MAZARIEGOS

DATE/TIME OF TEST: 01/25/2020 22:47:45

---

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.080	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.122	22:42
Breath Volume	2.687	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.122	22:46
Breath Volume	2.097	22:46
Air Blank	0.000	22:46
Control Test	0.080	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

---

Data Download Date/Time: 02/15/2020 13:54



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: MARTINEZ-SOLORZANO

DATE/TIME OF TEST: 01/26/2020 03:21:08

---

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.081	03:19	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:20	
Control Test	0.080	03:20	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

---

Data Download Date/Time: 02/15/2020 13:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VONADOR, JERARD, E  
BREATH TEST OPERATOR AGENCY: Bradenton PD  
ARRESTING OFFICER: MOORE, ADAM  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: CORONA

DATE/TIME OF TEST: 01/31/2020 22:50:33

---

Diagnostics Check	OK	22:47	
Air Blank	0.000	22:48	
Control Test	0.081	22:48	
Air Blank	0.000	22:49	
Subject Sample	# 1	22:49	Subject Test Refused
Breath Volume	0.000	22:49	
Air Blank	0.000	22:49	
Control Test	0.081	22:49	
Air Blank	0.000	22:50	
Diagnostics Check	OK	22:50	

---

Data Download Date/Time: 02/15/2020 13:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 01/01/2020 03:27:19

---

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.081	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.119	03:22
Breath Volume	3.332	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample # 2	0.119	03:25
Breath Volume	3.472	03:25
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

---

Data Download Date/Time: 01/21/2020 07:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PODHYSKI,PHILIP,A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: CASSIDY

DATE/TIME OF TEST: 01/01/2020 23:14:16

---

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:09	
Control Test	0.081	23:10	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:12	Subject Test Refused
Breath Volume	0.000	23:12	
Air Blank	0.000	23:13	
Control Test	0.081	23:13	
Air Blank	0.000	23:14	
Diagnostics Check	OK	23:14	

---

Data Download Date/Time: 01/21/2020 07:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ;6360101350092163387,WILLIAM,A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: HAEUSSLER,BRETT  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 09:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: MIHM

DATE/TIME OF TEST: 01/03/2020 11:12:23

---

Diagnostics Check	OK	11:11	
Air Blank		11:11	RFI Detect
Air Blank	0.000	11:12	

---

Data Download Date/Time: 01/21/2020 07:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLE,CALVIN,K  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: HAEUSSLER,BRETT  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 09:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: MIHM

DATE/TIME OF TEST: 01/03/2020 11:21:21

---

Diagnostics Check	OK	11:14
Air Blank	0.000	11:15
Control Test	0.081	11:15
Air Blank	0.000	11:16
Subject Sample # 1	0.000	11:16
Breath Volume	4.773	11:16
Air Blank	0.000	11:17
Air Blank	0.000	11:19
Subject Sample # 2	0.000	11:19
Breath Volume	5.121	11:19
Air Blank	0.000	11:20
Control Test	0.080	11:20
Air Blank	0.000	11:21
Diagnostics Check	OK	11:21

---

Data Download Date/Time: 01/21/2020 07:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VAUGHN,RYAN,  
BREATH TEST OPERATOR AGENCY: Bradenton PD  
ARRESTING OFFICER: MOORE,ADAM  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: BUSSELL

DATE/TIME OF TEST: 01/03/2020 21:35:57

---

Diagnostics Check	OK	21:32	
Air Blank	0.000	21:32	
Control Test	0.081	21:32	
Air Blank	0.000	21:33	
Subject Sample	# 1	21:34	Subject Test Refused
Breath Volume	0.000	21:34	
Air Blank	0.000	21:35	
Control Test	0.080	21:35	
Air Blank	0.000	21:35	
Diagnostics Check	OK	21:35	

---

Data Download Date/Time: 01/21/2020 07:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GWODZ,OFC,R  
BREATH TEST OPERATOR AGENCY: Bradenton PD  
ARRESTING OFFICER: MOORE,ADAM  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/04/2020 05:08:36

---

Diagnostics Check	OK	05:05	
Air Blank	0.000	05:06	
Control Test	0.081	05:06	
Air Blank	0.000	05:06	
Subject Sample	# 1	05:07	Subject Test Refused
Breath Volume	0.000	05:07	
Air Blank	0.000	05:07	
Control Test	0.080	05:07	
Air Blank	0.000	05:08	
Diagnostics Check	OK	05:08	

---

Data Download Date/Time: 01/21/2020 07:24



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: SANCHEZ TORRES

DATE/TIME OF TEST: 01/06/2020 02:02:35

---

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.081	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.122	01:57
Breath Volume	2.679	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.130	02:00
Breath Volume	3.652	02:00
Air Blank	0.000	02:01
Control Test	0.080	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

---

Data Download Date/Time: 01/21/2020 07:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: FRIDAY

DATE/TIME OF TEST: 01/07/2020 05:08:01

---

Diagnostics Check	OK	05:07	
Air Blank		05:07	RFI Detect
Air Blank	0.000	05:08	

---

Data Download Date/Time: 01/21/2020 07:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: FRIDAY

DATE/TIME OF TEST: 01/07/2020 05:13:03

---

Diagnostics Check	OK	05:09	
Air Blank	0.000	05:10	
Control Test	0.082	05:10	
Air Blank	0.000	05:11	
Subject Sample	# 1	05:11	Subject Test Refused
Breath Volume	0.000	05:11	
Air Blank	0.000	05:12	
Control Test	0.082	05:12	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:13	

---

Data Download Date/Time: 01/21/2020 07:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: KIPPENBERG

DATE/TIME OF TEST: 01/08/2020 02:09:20

---

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:02	
Control Test	0.081	02:02	
Air Blank	0.000	02:03	
Subject Sample	# 1 0.197	02:04	
Breath Volume	1.765	02:04	
Air Blank	0.000	02:05	
Air Blank	0.000	02:07	
Subject Sample	# 2	02:07	Subject Test Refused
Breath Volume	0.000	02:07	
Air Blank	0.000	02:08	
Control Test	0.080	02:08	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

---

Data Download Date/Time: 01/21/2020 07:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RODRIGUEZ,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 01/01/2020 02:57:53

---

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.081	02:51
Air Blank		0.000	02:52
Subject Sample	# 1	0.155	02:53
Breath Volume		1.429	02:53
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.147	02:56
Breath Volume		1.679	02:56
Air Blank		0.000	02:56
Control Test		0.080	02:57
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

---

Data Download Date/Time: 01/06/2020 18:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: VIVES

DATE/TIME OF TEST: 01/01/2020 05:57:11

---

Diagnostics Check		OK	05:49
Air Blank		0.000	05:49
Control Test		0.080	05:50
Air Blank		0.000	05:50
Subject Sample	# 1	0.166	05:52
Breath Volume		1.183	05:52
Air Blank		0.000	05:53
Air Blank		0.000	05:54
Subject Sample	# 2	0.167	05:55
Breath Volume		1.292	05:55
Air Blank		0.000	05:56
Control Test		0.078	05:56
Air Blank		0.000	05:57
Diagnostics Check		OK	05:57

---

Data Download Date/Time: 01/06/2020 18:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WASHINGTON,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:56

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 01/01/2020 06:23:40

---

Diagnostics Check	OK	06:16
Air Blank	0.000	06:16
Control Test	0.079	06:17
Air Blank	0.000	06:17
Subject Sample # 1	0.090	06:18
Breath Volume	2.300	06:18
Air Blank	0.000	06:19
Air Blank	0.000	06:21
Subject Sample # 2	0.089	06:22
Breath Volume	1.851	06:22
Air Blank	0.000	06:22
Control Test	0.080	06:23
Air Blank	0.000	06:23
Diagnostics Check	OK	06:23

---

Data Download Date/Time: 01/06/2020 18:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RODRIGUEZ,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 01/01/2020 22:59:10

---

Diagnostics Check	OK	22:53	
Air Blank	0.000	22:53	
Control Test	0.080	22:54	
Air Blank	0.000	22:54	
Subject Sample	# 1	22:57	Subject Test Refused
Breath Volume	0.000	22:57	
Air Blank	0.000	22:58	
Control Test	0.081	22:58	
Air Blank	0.000	22:59	
Diagnostics Check	OK	22:59	

---

Data Download Date/Time: 01/06/2020 18:22



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HALL,L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: BREEDEN

DATE/TIME OF TEST: 01/02/2020 21:40:23

---

Diagnostics Check	OK	21:33
Air Blank	0.000	21:34
Control Test	0.080	21:34
Air Blank	0.000	21:34
Subject Sample # 1	0.228	21:35
Breath Volume	1.121	21:35
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.229	21:38
Breath Volume	1.511	21:38
Air Blank	0.000	21:39
Control Test	0.078	21:39
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

---

Data Download Date/Time: 01/06/2020 18:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WHITE,M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: MILETY

DATE/TIME OF TEST: 01/03/2020 01:15:20

---

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.079	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.067	01:10
Breath Volume		2.289	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.065	01:13
Breath Volume		1.996	01:13
Air Blank		0.000	01:14
Control Test		0.079	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

---

Data Download Date/Time: 01/06/2020 18:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ANDREWS,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: FREEMAN JR

DATE/TIME OF TEST: 01/03/2020 03:39:24

---

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.135	03:34
Breath Volume	1.664	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.132	03:37
Breath Volume	1.636	03:37
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

---

Data Download Date/Time: 01/06/2020 18:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ANDREWS,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: MCCORD

DATE/TIME OF TEST: 01/03/2020 06:11:22

---

Diagnostics Check	OK	06:04
Air Blank	0.000	06:05
Control Test	0.080	06:05
Air Blank	0.000	06:06
Subject Sample # 1	0.000	06:07
Breath Volume	1.398	06:07
Air Blank	0.000	06:07
Air Blank	0.000	06:09
Subject Sample # 2	0.000	06:09
Breath Volume	1.734	06:09
Air Blank	0.000	06:10
Control Test	0.079	06:10
Air Blank	0.000	06:11
Diagnostics Check	OK	06:11

---

Data Download Date/Time: 01/06/2020 18:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: KIRK

DATE/TIME OF TEST: 01/04/2020 03:48:03

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.113	03:43
Breath Volume	1.507	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.118	03:46
Breath Volume	3.332	03:46
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:48

---

Data Download Date/Time: 01/06/2020 18:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKIPPER,ROGER,G  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD,J 287  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: APATA

DATE/TIME OF TEST: 01/04/2020 04:47:52

---

Diagnostics Check		OK	04:41
Air Blank		0.000	04:41
Control Test		0.080	04:42
Air Blank		0.000	04:42
Subject Sample	# 1	0.079	04:43
Breath Volume		1.160	04:43
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.085	04:46
Breath Volume		1.738	04:46
Air Blank		0.000	04:46
Control Test		0.080	04:47
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

---

Data Download Date/Time: 01/06/2020 18:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKIPPER,ROGER,G  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KLENOFSKY,A 255482  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 01/04/2020 05:44:14

---

Diagnostics Check	OK	05:37
Air Blank	0.000	05:37
Control Test	0.079	05:38
Air Blank	0.000	05:38
Subject Sample # 1	0.138	05:39
Breath Volume	2.769	05:39
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample # 2	0.145	05:42
Breath Volume	3.570	05:42
Air Blank	0.000	05:43
Control Test	0.079	05:43
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

---

Data Download Date/Time: 01/06/2020 18:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: BOYKINS

DATE/TIME OF TEST: 01/05/2020 00:43:35

---

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.105	00:38
Breath Volume	1.136	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.101	00:42
Breath Volume	1.496	00:42
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

---

Data Download Date/Time: 01/06/2020 18:22



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: MANZA

DATE/TIME OF TEST: 01/05/2020 03:21:18

---

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.000	03:17
Breath Volume	1.765	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.000	03:19
Breath Volume	2.472	03:19
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

---

Data Download Date/Time: 01/06/2020 18:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: LEAVITT

DATE/TIME OF TEST: 01/05/2020 05:02:36

---

Diagnostics Check		OK	04:55
Air Blank		0.000	04:55
Control Test		0.080	04:56
Air Blank		0.000	04:56
Subject Sample	# 1	0.130	04:57
Breath Volume		1.109	04:57
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.141	05:01
Breath Volume		2.429	05:01
Air Blank		0.000	05:01
Control Test		0.080	05:01
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

---

Data Download Date/Time: 01/06/2020 18:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/06/2020 01:25:54

---

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.131	01:21
Breath Volume	1.292	01:21
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.134	01:24
Breath Volume	1.882	01:24
Air Blank	0.000	01:24
Control Test	0.079	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

---

Data Download Date/Time: 01/06/2020 18:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONG,JOSHUA,A  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PATEL,JASON  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: MEADE

DATE/TIME OF TEST: 01/01/2020 21:27:02

---

Diagnostics Check	OK	21:20
Air Blank	0.000	21:20
Control Test	0.080	21:21
Air Blank	0.000	21:21
Subject Sample # 1	0.195	21:22
Breath Volume	1.863	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:24
Subject Sample # 2	0.190	21:25
Breath Volume	1.953	21:25
Air Blank	0.000	21:26
Control Test	0.078	21:26
Air Blank	0.000	21:26
Diagnostics Check	OK	21:27

---

Data Download Date/Time: 01/28/2020 17:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMA,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: LIMA,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: HULING

DATE/TIME OF TEST: 01/04/2020 19:44:07

---

Diagnostics Check	OK	19:43	
Air Blank		19:43	RFI Detect
Air Blank	0.000	19:44	

---

Data Download Date/Time: 01/28/2020 17:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMA, MICHAEL, A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: LIMA, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: HULING

DATE/TIME OF TEST: 01/04/2020 19:51:24

---

Diagnostics Check	OK	19:45
Air Blank	0.000	19:45
Control Test	0.080	19:46
Air Blank	0.000	19:46
Subject Sample # 1	0.000	19:47
Breath Volume	2.203	19:47
Air Blank	0.000	19:47
Air Blank	0.000	19:49
Subject Sample # 2	0.000	19:50
Breath Volume	2.199	19:50
Air Blank	0.000	19:50
Control Test	0.079	19:50
Air Blank	0.000	19:51
Diagnostics Check	OK	19:51

---

Data Download Date/Time: 01/28/2020 17:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CROUCH,J,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PIERRE,S  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: SCHWARTZ

DATE/TIME OF TEST: 01/05/2020 23:06:50

---

Diagnostics Check	OK	22:55	
Air Blank	0.000	22:56	
Control Test	0.081	22:56	
Air Blank	0.000	22:56	
Subject Sample	# 1 0.278	22:59	
Breath Volume	1.343	22:59	
Air Blank	0.000	23:01	
Air Blank	0.000	23:02	
Subject Sample	# 2 0.287	23:05	Volume Not Met
Breath Volume	0.113	23:05	
Air Blank		23:06	Purge Fail
Air Blank	0.000	23:06	

---

Data Download Date/Time: 01/28/2020 17:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CROUCH,J,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PIERRE,S  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: SCHWARTZ

DATE/TIME OF TEST: 01/05/2020 23:47:24

---

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.078	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.247	23:41
Breath Volume	1.199	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:43
Subject Sample # 2	0.247	23:45
Breath Volume	1.972	23:45
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

---

Data Download Date/Time: 01/28/2020 17:14



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCARTHY, EVAN, V  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: MCCARTHY, EVAN  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 01/08/2020 02:21:55

---

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.081	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.295	02:17
Breath Volume	3.128	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:19
Subject Sample # 2	0.285	02:20
Breath Volume	3.324	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

---

Data Download Date/Time: 01/28/2020 17:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PARKER,CARL  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: COTE

DATE/TIME OF TEST: 01/11/2020 21:46:02

---

Diagnostics Check	OK	21:42	
Air Blank	0.000	21:42	
Control Test	0.080	21:43	
Air Blank	0.000	21:43	
Subject Sample	# 1	21:44	Subject Test Refused
Breath Volume	0.000	21:44	
Air Blank	0.000	21:45	
Control Test	0.080	21:45	
Air Blank	0.000	21:45	
Diagnostics Check	OK	21:46	

---

Data Download Date/Time: 01/28/2020 17:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: BEANE,BRANDON  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 12:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:04

SUBJECT'S LAST NAME: CHOUN

DATE/TIME OF TEST: 01/12/2020 13:28:52

---

Diagnostics Check	OK	13:25	
Air Blank	0.000	13:26	
Control Test	0.080	13:26	
Air Blank	0.000	13:27	
Subject Sample	# 1	13:27	Subject Test Refused
Breath Volume	0.000	13:27	
Air Blank	0.000	13:27	
Control Test	0.079	13:28	
Air Blank	0.000	13:28	
Diagnostics Check	OK	13:28	

---

Data Download Date/Time: 01/28/2020 17:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONG,JOSHUA,A  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PATEL,JASON  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: HEANG

DATE/TIME OF TEST: 01/15/2020 20:51:25

---

Diagnostics Check	OK	20:44
Air Blank	0.000	20:44
Control Test	0.080	20:45
Air Blank	0.000	20:45
Subject Sample # 1	0.217	20:46
Breath Volume	2.148	20:46
Air Blank	0.000	20:47
Air Blank	0.000	20:48
Subject Sample # 2	0.207	20:49
Breath Volume	2.199	20:49
Air Blank	0.000	20:50
Control Test	0.077	20:50
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

---

Data Download Date/Time: 01/28/2020 17:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONG,JOSHUA,A  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: COZZONE,GAETANO  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 01/16/2020 21:36:27

---

Diagnostics Check	OK	21:33	
Air Blank	0.000	21:33	
Control Test	0.080	21:34	
Air Blank	0.000	21:34	
Subject Sample	# 1	21:35	Subject Test Refused
Breath Volume	0.000	21:35	
Air Blank	0.000	21:35	
Control Test	0.080	21:35	
Air Blank	0.000	21:36	
Diagnostics Check	OK	21:36	

---

Data Download Date/Time: 01/28/2020 17:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCARTHY, EVAN, V  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: MCCARTHY, EVAN  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: CENTOFANTI

DATE/TIME OF TEST: 01/17/2020 04:32:05

---

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.196	04:26
Breath Volume	1.375	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.178	04:30
Breath Volume	2.058	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

---

Data Download Date/Time: 01/28/2020 17:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STOGDON,BRAD,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PATEL,JASON  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: THOMSON

DATE/TIME OF TEST: 01/25/2020 21:16:05

---

Diagnostics Check	OK	21:13	
Air Blank	0.000	21:13	
Control Test	0.080	21:13	
Air Blank	0.000	21:14	
Subject Sample	# 1	21:14	Subject Test Refused
Breath Volume	0.000	21:14	
Air Blank	0.000	21:15	
Control Test	0.080	21:15	
Air Blank	0.000	21:16	
Diagnostics Check	OK	21:16	

---

Data Download Date/Time: 01/28/2020 17:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAINES,KAYLA,M  
BREATH TEST OPERATOR AGENCY: Flagler Beach PD  
ARRESTING OFFICER: RAINES,KAYLA  
ARRESTING OFFICER AGENCY: Flagler Beach PD  
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 01/26/2020 03:16:12

---

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:10
Subject Sample # 1	0.334	03:11
Breath Volume	1.589	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:13
Subject Sample # 2	0.324	03:14
Breath Volume	1.750	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

---

Data Download Date/Time: 01/28/2020 17:14



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEAVER,D,  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: STRACK,R  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: DIEDEL

DATE/TIME OF TEST: 01/26/2020 03:41:31

---

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.079	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.248	03:36
Breath Volume		1.109	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.262	03:39
Breath Volume		1.707	03:39
Air Blank		0.000	03:40
Control Test		0.079	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

---

Data Download Date/Time: 01/28/2020 17:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PARKER,CARL  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: STILLEY

DATE/TIME OF TEST: 01/26/2020 07:46:48

---

Diagnostics Check	OK	07:38
Air Blank	0.000	07:38
Control Test	0.080	07:38
Air Blank	0.000	07:39
Subject Sample # 1	0.126	07:41
Breath Volume	1.175	07:41
Air Blank	0.000	07:42
Air Blank	0.000	07:43
Subject Sample # 2	0.116	07:45
Breath Volume	2.039	07:45
Air Blank	0.000	07:45
Control Test	0.080	07:46
Air Blank	0.000	07:46
Diagnostics Check	OK	07:46

---

Data Download Date/Time: 01/28/2020 17:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHMID,ERNEST,J  
BREATH TEST OPERATOR AGENCY: Kennedy Space Center  
ARRESTING OFFICER: HOOAND,AUBREY  
ARRESTING OFFICER AGENCY: Kennedy Space Center  
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 10:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:33

SUBJECT'S LAST NAME: SPRINGSTROH

DATE/TIME OF TEST: 01/08/2020 11:05:15

---

Diagnostics Check	OK	10:54	
Air Blank	0.000	10:55	
Control Test	0.078	10:55	
Air Blank	0.000	10:56	
Subject Sample	# 1 0.196	10:59	
Breath Volume	2.382	10:59	
Air Blank	0.000	11:00	
Air Blank	0.000	11:01	
Subject Sample	# 2 0.201	11:03	
Breath Volume	1.570	11:03	
Air Blank	0.000	11:04	
Control Test	0.074	11:04	Control Outside Tolerance
Air Blank	0.000	11:05	

---

Data Download Date/Time: 01/12/2020 15:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAUCKNER,CHRISTOPHER,A  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: CALHOUN,KEVIN  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 01/03/2020 01:05:05

---

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.079	00:59
Air Blank		0.000	00:59
Subject Sample	# 1	0.014	01:00
Breath Volume		1.390	01:00
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.013	01:03
Breath Volume		1.968	01:03
Air Blank		0.000	01:04
Control Test		0.079	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:05

---

Data Download Date/Time: 01/31/2020 12:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,HALEY,M  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: CASAS,JAVIER  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: BURSTEIN

DATE/TIME OF TEST: 01/05/2020 00:26:45

---

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.079	00:20
Air Blank		0.000	00:21
Subject Sample	# 1	0.234	00:22
Breath Volume		2.136	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:24
Subject Sample	# 2	0.232	00:24
Breath Volume		1.210	00:24
Air Blank		0.000	00:25
Control Test		0.079	00:26
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

---

Data Download Date/Time: 01/31/2020 12:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COON,REBECCA,L  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: CALHOUN,KEVIN  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 01/17/2020 02:48:34

---

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.079	02:45	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:47	Subject Test Refused
Breath Volume	0.000	02:47	
Air Blank	0.000	02:47	
Control Test	0.080	02:47	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

---

Data Download Date/Time: 01/31/2020 12:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COON,REBECCA,L  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: RENTERIA,BARTOLO  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ALLISON

DATE/TIME OF TEST: 01/17/2020 23:00:15

---

Diagnostics Check	OK	22:57	
Air Blank	0.000	22:57	
Control Test	0.079	22:58	
Air Blank	0.000	22:58	
Subject Sample	# 1	22:58	Subject Test Refused
Breath Volume	0.000	22:58	
Air Blank	0.000	22:59	
Control Test	0.080	22:59	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:00	

---

Data Download Date/Time: 01/31/2020 12:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,A,  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: PAPIA,A  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: VAN VOLKENBURG

DATE/TIME OF TEST: 01/24/2020 17:05:07

---

Diagnostics Check	OK	16:57
Air Blank	0.000	16:58
Control Test	0.079	16:58
Air Blank	0.000	16:58
Subject Sample # 1	0.126	16:59
Breath Volume	2.792	16:59
Air Blank	0.000	17:00
Air Blank	0.000	17:02
Subject Sample # 2	0.127	17:03
Breath Volume	1.707	17:03
Air Blank	0.000	17:04
Control Test	0.079	17:04
Air Blank	0.000	17:05
Diagnostics Check	OK	17:05

---

Data Download Date/Time: 01/31/2020 12:09



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RENTERIA,BARTOLO,  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: COON,REBECCA  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: MURDOCH

DATE/TIME OF TEST: 01/30/2020 05:10:17

---

Diagnostics Check	OK	05:07	
Air Blank	0.000	05:07	
Control Test	0.080	05:08	
Air Blank	0.000	05:08	
Subject Sample	# 1	05:08	Subject Test Refused
Breath Volume	0.000	05:08	
Air Blank	0.000	05:09	
Control Test	0.080	05:09	
Air Blank	0.000	05:10	
Diagnostics Check	OK	05:10	

---

Data Download Date/Time: 01/31/2020 12:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,HALEY,M  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: CARECCIA,GREGORY  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CONO

DATE/TIME OF TEST: 01/31/2020 03:16:00

---

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.080	03:14	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank		03:15	RFI Detect
Air Blank	0.000	03:16	

---

Data Download Date/Time: 01/31/2020 12:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,HALEY,M  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: CARECCIA,GREGORY  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CONO

DATE/TIME OF TEST: 01/31/2020 03:21:49

---

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:20	
Control Test	0.080	03:21	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

---

Data Download Date/Time: 01/31/2020 12:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RENTERIA,BARTOLO,  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: VRABEL,RYAN  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:21

SUBJECT'S LAST NAME: TREJO RAMIREZ

DATE/TIME OF TEST: 01/31/2020 06:54:13

---

Diagnostics Check	OK	06:47
Air Blank	0.000	06:47
Control Test	0.080	06:48
Air Blank	0.000	06:48
Subject Sample # 1	0.246	06:49
Breath Volume	2.644	06:49
Air Blank	0.000	06:50
Air Blank	0.000	06:51
Subject Sample # 2	0.229	06:52
Breath Volume	2.492	06:52
Air Blank	0.000	06:53
Control Test	0.079	06:53
Air Blank	0.000	06:54
Diagnostics Check	OK	06:54

---

Data Download Date/Time: 01/31/2020 12:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANTELAR,JOAN,  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: CANTELAR,JOAN  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 12/30/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: ACANDA LAZO

DATE/TIME OF TEST: 01/12/2020 02:06:02

---

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.080	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.096	02:01
Breath Volume	3.382	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:03
Subject Sample # 2	0.092	02:04
Breath Volume	2.730	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:06

---

Data Download Date/Time: 01/27/2020 15:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTILLO,CARLOS,M  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: CASTILLO,CARLOS  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: RUSH

DATE/TIME OF TEST: 01/31/2020 02:15:00

---

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:05	
Control Test		0.080	02:06	
Air Blank		0.000	02:06	
Subject Sample	# 1	0.308	02:09	Volume Not Met
Breath Volume		0.636	02:09	
Air Blank		0.000	02:10	
Air Blank		0.000	02:12	
Subject Sample	# 2		02:13	Subject Test Refused
Breath Volume		0.000	02:13	
Air Blank		0.000	02:14	
Control Test		0.071	02:14	Control Outside Tolerance
Air Blank		0.000	02:15	

---

Data Download Date/Time: 02/29/2020 22:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROJAS,GUILLERMO,A  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: ROJAS,GUILLERMO  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 796576  
DRY GAS EXPIRY DATE: 06/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: BULLARD

DATE/TIME OF TEST: 01/31/2020 23:08:50

---

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:06	
Control Test	0.081	23:06	
Air Blank	0.000	23:07	
Subject Sample	# 1	23:07	Subject Test Refused
Breath Volume	0.000	23:07	
Air Blank	0.000	23:07	
Control Test	0.080	23:08	
Air Blank	0.000	23:08	
Diagnostics Check	OK	23:08	

---

Data Download Date/Time: 02/29/2020 22:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA,VICTOR,J  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: RIVERA,VICTOR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: ROUSSEAU

DATE/TIME OF TEST: 01/05/2020 03:46:14

---

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.120	03:41
Breath Volume	2.492	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.126	03:44
Breath Volume	5.226	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

---

Data Download Date/Time: 01/27/2020 23:32



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA,VICTOR,J  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: RIVERA,VICTOR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17618080A3  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: PILLOT LIND

DATE/TIME OF TEST: 01/10/2020 05:16:09

---

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:06
Subject Sample # 1	0.104	05:07
Breath Volume	1.304	05:07
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.083	05:11
Breath Volume	2.019	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 3	0.094	05:14
Breath Volume	2.093	05:14
Air Blank	0.000	05:15
Control Test	0.077	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

---

Data Download Date/Time: 01/27/2020 23:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAWKINS,DEBORAH,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HAWKINS,DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1791908A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: DEYO

DATE/TIME OF TEST: 01/03/2020 01:17:28

---

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.078	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.195	01:12
Breath Volume		1.605	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.185	01:15
Breath Volume		1.195	01:15
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

---

Data Download Date/Time: 01/27/2020 23:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAWKINS,DEBORAH,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HAWKINS,DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1791908A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: COLON REYES

DATE/TIME OF TEST: 01/12/2020 04:38:42

---

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.078	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.148	04:33
Breath Volume	1.519	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.151	04:37
Breath Volume	2.300	04:37
Air Blank	0.000	04:37
Control Test	0.078	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

---

Data Download Date/Time: 01/27/2020 23:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA,VICTOR,J  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HAWKINS,DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1791908A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DAVILA REYES

DATE/TIME OF TEST: 01/17/2020 02:02:37

---

Diagnostics Check		OK	01:50	
Air Blank		0.000	01:51	
Control Test		0.079	01:51	
Air Blank		0.000	01:52	
Subject Sample	# 1	0.161	01:55	Volume Not Met
Breath Volume		0.066	01:55	
Air Blank		0.000	01:56	
Air Blank		0.000	01:57	
Subject Sample	# 2	0.174	02:00	Volume Not Met
Breath Volume		0.082	02:00	
Air Blank		0.000	02:01	
Control Test		0.079	02:01	
Air Blank		0.000	02:02	
Diagnostics Check		OK	02:02	

---

Data Download Date/Time: 01/27/2020 23:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA,VICTOR,J  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HAWKINS,DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1791908A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DAVILA REYES

DATE/TIME OF TEST: 01/17/2020 02:20:30

---

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:05	
Control Test		0.077	02:06	
Air Blank		0.000	02:06	
Subject Sample	# 1	0.170	02:09	Volume Not Met
Breath Volume		0.125	02:09	
Air Blank		0.000	02:10	
Air Blank		0.000	02:12	
Subject Sample	# 2	0.181	02:14	
Breath Volume		1.652	02:14	
Air Blank		0.000	02:15	
Air Blank		0.000	02:16	
Subject Sample	# 3	0.187	02:18	
Breath Volume		1.425	02:18	
Air Blank		0.000	02:19	
Control Test		0.077	02:19	
Air Blank		0.000	02:20	
Diagnostics Check		OK	02:20	

---

Data Download Date/Time: 01/27/2020 23:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: SCHOPFER

DATE/TIME OF TEST: 01/24/2020 20:03:52

---

Diagnostics Check	OK	19:57
Air Blank	0.000	19:57
Control Test	0.080	19:57
Air Blank	0.000	19:58
Subject Sample # 1	0.195	19:59
Breath Volume	2.324	19:59
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.193	20:02
Breath Volume	2.390	20:02
Air Blank	0.000	20:02
Control Test	0.080	20:03
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

---

Data Download Date/Time: 02/24/2020 09:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 12/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: RAM

DATE/TIME OF TEST: 01/01/2020 01:48:37

---

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.195	01:43
Breath Volume	2.222	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.189	01:46
Breath Volume	2.324	01:46
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

---

Data Download Date/Time: 01/15/2020 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 12/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: BARDALES-OCAMPO

DATE/TIME OF TEST: 01/02/2020 00:41:43

---

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:39	
Control Test	0.081	00:39	
Air Blank	0.000	00:40	
Subject Sample	# 1	00:40	Subject Test Refused
Breath Volume	0.000	00:40	
Air Blank	0.000	00:40	
Control Test	0.081	00:41	
Air Blank	0.000	00:41	
Diagnostics Check	OK	00:41	

---

Data Download Date/Time: 01/15/2020 10:14



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUMMINGS,GARY,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: CUMMINGS,GARY  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 12/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: GOBLE

DATE/TIME OF TEST: 01/03/2020 01:24:13

---

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.080	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.114	01:19
Breath Volume	1.699	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.108	01:22
Breath Volume	2.183	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

---

Data Download Date/Time: 01/15/2020 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRETT,CODY,R  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BARRETT,CODY  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 12/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: ALFARO

DATE/TIME OF TEST: 01/03/2020 21:27:39

---

Diagnostics Check	OK	21:20
Air Blank	0.000	21:21
Control Test	0.080	21:21
Air Blank	0.000	21:21
Subject Sample # 1	0.190	21:22
Breath Volume	3.015	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.197	21:25
Breath Volume	2.699	21:25
Air Blank	0.000	21:26
Control Test	0.079	21:27
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

---

Data Download Date/Time: 01/15/2020 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 12/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 01/11/2020 22:55:28

---

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.081	22:50
Air Blank	0.000	22:50
Subject Sample # 1	0.142	22:50
Breath Volume	2.070	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.139	22:53
Breath Volume	2.125	22:53
Air Blank	0.000	22:54
Control Test	0.080	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

---

Data Download Date/Time: 01/15/2020 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 12/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: PINETE

DATE/TIME OF TEST: 01/13/2020 01:17:42

---

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.081	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.161	01:10
Breath Volume	2.472	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.160	01:15
Breath Volume	1.402	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

---

Data Download Date/Time: 01/15/2020 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRETT,CODY,R  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BARRETT,CODY  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 12/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: GORDILLO RUIZ

DATE/TIME OF TEST: 01/13/2020 01:54:34

---

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.079	01:52	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:53	RFI Detect
Breath Volume	0.000	01:53	
Air Blank	0.000	01:53	
Control Test	0.080	01:53	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

---

Data Download Date/Time: 01/15/2020 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRETT,CODY,R  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BARRETT,CODY  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 12/16/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: GORDILLO RUIZ

DATE/TIME OF TEST: 01/13/2020 02:10:33

---

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:57	
Control Test		0.080	01:57	
Air Blank		0.000	01:57	
Subject Sample	# 1	0.169	02:01	Volume Not Met
Breath Volume		0.972	02:01	
Air Blank		0.000	02:01	
Air Blank		0.000	02:03	
Subject Sample	# 2	0.143	02:04	
Breath Volume		1.273	02:04	
Air Blank		0.000	02:05	
Air Blank		0.000	02:07	
Subject Sample	# 3	0.158	02:08	
Breath Volume		3.453	02:08	
Air Blank		0.000	02:09	
Control Test		0.078	02:09	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:10	

---

Data Download Date/Time: 01/15/2020 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: MONTEMAGNO,CARLO  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: RICE

DATE/TIME OF TEST: 01/18/2020 02:05:31

---

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.081	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.000	02:01
Breath Volume	2.156	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.000	02:04
Breath Volume	3.425	02:04
Air Blank	0.000	02:04
Control Test	0.081	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

---

Data Download Date/Time: 02/15/2020 08:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: MONTEMAGNO,CARLO  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: GAMBA

DATE/TIME OF TEST: 01/19/2020 02:34:01

---

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.080	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:33	
Control Test	0.081	02:33	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:34	

---

Data Download Date/Time: 02/15/2020 08:37



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: RIFFE

DATE/TIME OF TEST: 01/19/2020 20:30:02

---

Diagnostics Check	OK	20:23
Air Blank	0.000	20:24
Control Test	0.081	20:24
Air Blank	0.000	20:24
Subject Sample # 1	0.100	20:25
Breath Volume	2.718	20:25
Air Blank	0.000	20:26
Air Blank	0.000	20:27
Subject Sample # 2	0.100	20:28
Breath Volume	3.089	20:28
Air Blank	0.000	20:29
Control Test	0.080	20:29
Air Blank	0.000	20:29
Diagnostics Check	OK	20:30

---

Data Download Date/Time: 02/15/2020 08:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUICK,BENJAMIN,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: QUICK,BENJAMIN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 01/22/2020 00:01:07

---

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.081	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.272	23:55
Breath Volume	1.542	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.266	23:59
Breath Volume	1.687	23:59
Air Blank	0.000	00:00
Control Test	0.080	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

---

Data Download Date/Time: 02/15/2020 08:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: MONTEMAGNO,CARLO  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/24/2020 00:54:59

---

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.081	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.000	00:50
Breath Volume	3.988	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.000	00:53
Breath Volume	3.933	00:53
Air Blank	0.000	00:54
Control Test	0.081	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

---

Data Download Date/Time: 02/15/2020 08:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: GOMEZ SANCHEZ

DATE/TIME OF TEST: 01/24/2020 20:31:23

---

Diagnostics Check	OK	20:23
Air Blank	0.000	20:24
Control Test	0.080	20:24
Air Blank	0.000	20:24
Subject Sample # 1	0.199	20:26
Breath Volume	1.308	20:26
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.202	20:29
Breath Volume	1.480	20:29
Air Blank	0.000	20:30
Control Test	0.077	20:30
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

---

Data Download Date/Time: 02/15/2020 08:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUMMINGS,GARY,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: CUMMINGS,GARY  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 01/27/2020 03:56:30

---

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:54	
Control Test	0.080	03:54	
Air Blank	0.000	03:54	
Subject Sample	# 1	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Control Test	0.080	03:55	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

---

Data Download Date/Time: 02/15/2020 08:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: MONTEMAGNO,CARLO  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: GULLO

DATE/TIME OF TEST: 01/28/2020 03:04:18

---

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.079	02:57	
Air Blank	0.000	02:58	
Subject Sample	# 1 0.160	03:00	
Breath Volume	1.566	03:00	
Air Blank	0.000	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 2	03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:03	
Control Test	0.079	03:03	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

---

Data Download Date/Time: 02/15/2020 08:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: CHRISTESON

DATE/TIME OF TEST: 01/29/2020 01:15:38

---

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.081	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.000	01:11
Breath Volume	2.699	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.000	01:14
Breath Volume	2.410	01:14
Air Blank	0.000	01:14
Control Test	0.081	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

---

Data Download Date/Time: 02/15/2020 08:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 01/30/2020 00:10:45

---

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.080	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.133	00:06
Breath Volume	3.578	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.127	00:09
Breath Volume	2.535	00:09
Air Blank	0.000	00:09
Control Test	0.079	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

---

Data Download Date/Time: 02/15/2020 08:38



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUMMINGS,GARY,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: CUMMINGS,GARY  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 962587  
DRY GAS EXPIRY DATE: 06/04/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: THURMAN

DATE/TIME OF TEST: 01/30/2020 04:34:53

---

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.194	04:30
Breath Volume	2.105	04:30
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.192	04:33
Breath Volume	1.894	04:33
Air Blank	0.000	04:33
Control Test	0.080	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

---

Data Download Date/Time: 02/15/2020 08:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALFY,RACHEL,M  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SULAIMAN,RAMIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: RODRIQUEZ

DATE/TIME OF TEST: 01/07/2020 02:27:57

---

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.081	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.139	02:23
Breath Volume	1.914	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.139	02:26
Breath Volume	1.839	02:26
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

---

Data Download Date/Time: 01/18/2020 14:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALFY,RACHEL,M  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: PALFY,RACHEL  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: BUEZO

DATE/TIME OF TEST: 01/12/2020 03:58:17

---

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.078	03:51	
Air Blank	0.000	03:51	
Subject Sample # 1	0.187	03:53	
Breath Volume	1.500	03:53	
Air Blank	0.000	03:53	
Air Blank	0.000	03:55	
Subject Sample # 2	0.178	03:56	
Breath Volume	1.679	03:56	
Air Blank	0.000	03:57	
Control Test	0.077	03:57	Interferent Detect
Air Blank	0.000	03:58	

---

Data Download Date/Time: 01/18/2020 14:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANHORN,MICHAEL,T  
BREATH TEST OPERATOR AGENCY: Holmes Beach PD  
ARRESTING OFFICER: VANHORN,MICHAEL  
ARRESTING OFFICER AGENCY: Holmes Beach PD  
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175970  
DRY GAS EXPIRY DATE: 08/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: BOTERO

DATE/TIME OF TEST: 01/25/2020 21:25:48

---

Diagnostics Check	OK	21:19
Air Blank	0.000	21:19
Control Test	0.080	21:20
Air Blank	0.000	21:20
Subject Sample # 1	0.234	21:21
Breath Volume	1.972	21:21
Air Blank	0.000	21:21
Air Blank	0.000	21:23
Subject Sample # 2	0.231	21:24
Breath Volume	1.863	21:24
Air Blank	0.000	21:24
Control Test	0.079	21:25
Air Blank	0.000	21:25
Diagnostics Check	OK	21:25

---

Data Download Date/Time: 02/28/2020 12:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: AGUILAR ESCOBAR

DATE/TIME OF TEST: 01/01/2020 02:51:47

---

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.080	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.193	02:47
Breath Volume	2.195	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:49
Subject Sample # 2	0.173	02:50
Breath Volume	2.417	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

---

Data Download Date/Time: 01/27/2020 10:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MONTES AGAMA

DATE/TIME OF TEST: 01/04/2020 03:22:31

---

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:10	
Control Test		0.079	03:10	
Air Blank		0.000	03:11	
Subject Sample	# 1	0.094	03:14	Volume Not Met
Breath Volume		1.675	03:14	
Air Blank		0.000	03:15	
Air Blank		0.000	03:16	
Subject Sample	# 2	0.188	03:17	
Breath Volume		3.269	03:17	
Air Blank		0.000	03:18	
Air Blank		0.000	03:20	
Subject Sample	# 3	0.184	03:20	
Breath Volume		3.761	03:20	
Air Blank		0.000	03:21	
Control Test		0.077	03:21	
Air Blank		0.000	03:22	
Diagnostics Check		OK	03:22	

---

Data Download Date/Time: 01/27/2020 10:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: GOMEZ GARCIA

DATE/TIME OF TEST: 01/04/2020 19:55:52

---

Diagnostics Check	OK	19:49
Air Blank	0.000	19:49
Control Test	0.079	19:50
Air Blank	0.000	19:50
Subject Sample # 1	0.150	19:51
Breath Volume	1.152	19:51
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.147	19:54
Breath Volume	1.695	19:54
Air Blank	0.000	19:54
Control Test	0.078	19:55
Air Blank	0.000	19:55
Diagnostics Check	OK	19:55

---

Data Download Date/Time: 01/27/2020 10:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: PEREZ SOLIS

DATE/TIME OF TEST: 01/04/2020 23:11:34

---

Diagnostics Check		OK	22:57	
Air Blank		0.000	22:57	
Control Test		0.080	22:58	
Air Blank		0.000	22:58	
Subject Sample	# 1	0.074	23:01	Volume Not Met
Breath Volume		0.902	23:01	
Air Blank		0.000	23:02	
Air Blank		0.000	23:04	
Subject Sample	# 2	0.086	23:05	
Breath Volume		1.480	23:05	
Air Blank		0.000	23:05	
Air Blank		0.000	23:07	
Subject Sample	# 3	0.089	23:09	
Breath Volume		1.265	23:09	
Air Blank		0.000	23:10	
Control Test		0.079	23:10	
Air Blank		0.000	23:11	
Diagnostics Check		OK	23:11	

---

Data Download Date/Time: 01/27/2020 10:31



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUBEL,NEIL,F  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: RUBEL,NEIL  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 01/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 01/22/2020 23:43:14

---

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.081	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.217	23:38
Breath Volume	3.863	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.211	23:41
Breath Volume	2.890	23:41
Air Blank	0.000	23:42
Control Test	0.081	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

---

Data Download Date/Time: 02/12/2020 02:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUBEL,NEIL,F  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: RUBEL,NEIL  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 01/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/26/2020 02:32:41

---

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.080	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.199	02:28
Breath Volume	2.617	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.210	02:31
Breath Volume	2.351	02:31
Air Blank	0.000	02:31
Control Test	0.081	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

---

Data Download Date/Time: 02/12/2020 02:27



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: MELHOUSE

DATE/TIME OF TEST: 01/06/2020 23:32:20

---

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.101	23:27
Breath Volume	2.886	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.102	23:30
Breath Volume	2.785	23:30
Air Blank	0.000	23:31
Control Test	0.079	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

---

Data Download Date/Time: 01/16/2020 20:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: WAYCASTER

DATE/TIME OF TEST: 01/07/2020 02:56:22

---

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:55	
		02:55	RFI Detect
Air Blank	0.000	02:56	

---

Data Download Date/Time: 01/16/2020 20:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: WAYCASTER

DATE/TIME OF TEST: 01/07/2020 03:05:26

---

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.079	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.112	03:00
Breath Volume	5.019	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.109	03:03
Breath Volume	3.781	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

---

Data Download Date/Time: 01/16/2020 20:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: TRISTRAM

DATE/TIME OF TEST: 01/12/2020 02:12:31

---

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.077	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.077	02:07
Breath Volume	1.421	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.085	02:10
Breath Volume	2.386	02:10
Air Blank	0.000	02:11
Control Test	0.078	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

---

Data Download Date/Time: 01/16/2020 20:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: CHARD

DATE/TIME OF TEST: 01/15/2020 22:58:58

---

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.078	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.101	22:54
Breath Volume	2.855	22:54
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.098	22:57
Breath Volume	3.097	22:57
Air Blank	0.000	22:58
Control Test	0.076	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

---

Data Download Date/Time: 01/16/2020 20:09



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GUILLOT

DATE/TIME OF TEST: 01/17/2020 03:46:56

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.079	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.187	03:42
Breath Volume	3.660	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:44
Subject Sample # 2	0.179	03:45
Breath Volume	3.429	03:45
Air Blank	0.000	03:46
Control Test	0.078	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

---

Data Download Date/Time: 02/12/2020 18:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PIZZOFERRATO

DATE/TIME OF TEST: 01/25/2020 03:28:00

---

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.078	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.095	03:22
Breath Volume	1.429	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.091	03:26
Breath Volume	1.242	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:28

---

Data Download Date/Time: 02/12/2020 18:59



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLEY,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: HOOD 2

DATE/TIME OF TEST: 01/05/2020 02:55:37

---

Diagnostics Check		OK	02:49
Air Blank		0.000	02:49
Control Test		0.084	02:49
Air Blank		0.000	02:50
Subject Sample	# 1	0.124	02:50
Breath Volume		2.250	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.127	02:54
Breath Volume		3.437	02:54
Air Blank		0.000	02:54
Control Test		0.083	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

---

Data Download Date/Time: 01/16/2020 17:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLEY,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: GAUBATZ

DATE/TIME OF TEST: 01/24/2020 02:04:39

---

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.082	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.136	01:59
Breath Volume	1.910	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.147	02:03
Breath Volume	2.542	02:03
Air Blank	0.000	02:03
Control Test	0.081	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

---

Data Download Date/Time: 02/12/2020 19:26



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARLEY,JOSEPH,E  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FARLEY,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: CREWS

DATE/TIME OF TEST: 01/26/2020 01:39:27

---

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.082	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.147	01:34
Breath Volume	2.859	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.149	01:37
Breath Volume	3.152	01:37
Air Blank	0.000	01:38
Control Test	0.082	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

---

Data Download Date/Time: 02/12/2020 19:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/01/2020 04:04:52

---

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.077	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.210	03:59
Breath Volume		2.207	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:01
Subject Sample	# 2	0.212	04:03
Breath Volume		2.519	04:03
Air Blank		0.000	04:03
Control Test		0.079	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

---

Data Download Date/Time: 01/24/2020 13:01



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 07:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: RESTREPO BOHORQUEZ

DATE/TIME OF TEST: 01/01/2020 08:36:00

---

Diagnostics Check	OK	08:28
Air Blank	0.000	08:28
Control Test	0.080	08:29
Air Blank	0.000	08:29
Subject Sample # 1	0.182	08:30
Breath Volume	1.835	08:30
Air Blank	0.000	08:31
Air Blank	0.000	08:33
Subject Sample # 2	0.187	08:34
Breath Volume	3.269	08:34
Air Blank	0.000	08:35
Control Test	0.079	08:35
Air Blank	0.000	08:35
Diagnostics Check	OK	08:36

---

Data Download Date/Time: 01/24/2020 13:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: FRANCO,G  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 01/02/2020 01:46:49

---

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.079	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.198	01:42
Breath Volume	1.535	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.194	01:45
Breath Volume	1.210	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

---

Data Download Date/Time: 01/24/2020 13:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MAYORA

DATE/TIME OF TEST: 01/04/2020 05:31:26

---

Diagnostics Check	OK	05:24
Air Blank	0.000	05:25
Control Test	0.078	05:25
Air Blank	0.000	05:25
Subject Sample # 1	0.152	05:26
Breath Volume	3.476	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.157	05:29
Breath Volume	3.542	05:29
Air Blank	0.000	05:30
Control Test	0.078	05:30
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

---

Data Download Date/Time: 01/24/2020 13:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: QUINTANA,J  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: BOADA

DATE/TIME OF TEST: 01/05/2020 02:51:44

---

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.143	02:47
Breath Volume	2.433	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.137	02:50
Breath Volume	1.878	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

---

Data Download Date/Time: 01/24/2020 13:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 01/10/2020 03:41:53

---

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:39	
Control Test	0.079	03:39	
Air Blank	0.000	03:40	
Subject Sample	# 1	03:40	Subject Test Refused
Breath Volume	0.000	03:40	
Air Blank	0.000	03:40	
Control Test	0.077	03:41	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

---

Data Download Date/Time: 01/24/2020 13:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: DANZA

DATE/TIME OF TEST: 01/12/2020 06:24:42

---

Diagnostics Check	OK	06:17
Air Blank	0.000	06:18
Control Test	0.079	06:18
Air Blank	0.000	06:18
Subject Sample # 1	0.089	06:19
Breath Volume	1.324	06:19
Air Blank	0.000	06:20
Air Blank	0.000	06:22
Subject Sample # 2	0.091	06:23
Breath Volume	1.574	06:23
Air Blank	0.000	06:23
Control Test	0.079	06:24
Air Blank	0.000	06:24
Diagnostics Check	OK	06:24

---

Data Download Date/Time: 01/24/2020 13:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 12/27/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: DEL SOL

DATE/TIME OF TEST: 01/17/2020 06:09:45

---

Diagnostics Check		OK	06:01
Air Blank		0.000	06:02
Control Test		0.079	06:02
Air Blank		0.000	06:02
Subject Sample	# 1	0.203	06:05
Breath Volume		1.468	06:05
Air Blank		0.000	06:05
Air Blank		0.000	06:07
Subject Sample	# 2	0.202	06:08
Breath Volume		2.472	06:08
Air Blank		0.000	06:08
Control Test		0.077	06:09
Air Blank		0.000	06:09
Diagnostics Check		OK	06:09

---

Data Download Date/Time: 01/24/2020 13:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: QUINTANA,J  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 01/26/2020 02:03:14

---

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.000	01:58
Breath Volume	2.214	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.000	02:01
Breath Volume	1.957	02:01
Air Blank	0.000	02:02
Control Test	0.079	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

---

Data Download Date/Time: 02/26/2020 15:14



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 01/30/2020 04:13:21

---

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:10	
Control Test	0.079	04:11	
Air Blank	0.000	04:11	
Subject Sample	# 1	04:11	Subject Test Refused
Breath Volume	0.000	04:11	
Air Blank	0.000	04:12	
Control Test	0.078	04:12	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

---

Data Download Date/Time: 02/26/2020 15:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: VARGAS,D  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: PEREZ-POZO

DATE/TIME OF TEST: 01/31/2020 02:40:07

---

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:37	
Control Test	0.079	02:38	
Air Blank	0.000	02:38	
Subject Sample	# 1	02:38	Subject Test Refused
Breath Volume	0.000	02:38	
Air Blank	0.000	02:39	
Control Test	0.078	02:39	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

---

Data Download Date/Time: 02/26/2020 15:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 01/31/2020 07:06:50

---

Diagnostics Check	OK	07:04	
Air Blank	0.000	07:04	
Control Test	0.079	07:04	
Air Blank	0.000	07:05	
Subject Sample	# 1	07:05	Subject Test Refused
Breath Volume	0.000	07:05	
Air Blank	0.000	07:05	
Control Test	0.078	07:06	
Air Blank	0.000	07:06	
Diagnostics Check	OK	07:06	

---

Data Download Date/Time: 02/26/2020 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: STERN,ADAM  
ARRESTING OFFICER AGENCY: Plantation PD  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: VALDEZ GARCIA

DATE/TIME OF TEST: 01/01/2020 03:05:19

---

Diagnostics Check		OK	02:58
Air Blank		0.000	02:59
Control Test		0.080	02:59
Air Blank		0.000	02:59
Subject Sample	# 1	0.275	03:00
Breath Volume		1.976	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.283	03:03
Breath Volume		2.550	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

---

Data Download Date/Time: 01/16/2020 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'DOR,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: O'DOR,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: COMEAUX

DATE/TIME OF TEST: 01/07/2020 07:25:26

---

Diagnostics Check		OK	07:14
Air Blank		0.000	07:14
Control Test		0.080	07:15
Air Blank		0.000	07:15
Subject Sample	# 1	0.193	07:17
Breath Volume		2.140	07:17
Air Blank		0.000	07:17
Air Blank		0.000	07:19
Subject Sample	# 2	0.214	07:20
Breath Volume		2.175	07:20
Air Blank		0.000	07:20
Air Blank		0.000	07:22
Subject Sample	# 3	0.188	07:23
Breath Volume		1.847	07:23
Air Blank		0.000	07:24
Control Test		0.080	07:24
Air Blank		0.000	07:25
Diagnostics Check		OK	07:25

---

Data Download Date/Time: 01/16/2020 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: SAPP,JOSHUA  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: RICHARDS

DATE/TIME OF TEST: 01/12/2020 02:28:32

---

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.242	02:23
Breath Volume	1.394	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.225	02:26
Breath Volume	1.167	02:26
Air Blank	0.000	02:27
Control Test	0.078	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

---

Data Download Date/Time: 01/16/2020 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LOUGHRAN,KEVIN  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: POLK

DATE/TIME OF TEST: 01/12/2020 23:11:27

---

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.079	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.259	23:06
Breath Volume	1.949	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.246	23:09
Breath Volume	1.898	23:09
Air Blank	0.000	23:10
Control Test	0.078	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

---

Data Download Date/Time: 01/16/2020 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: SAPP,JOSHUA  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 01/13/2020 01:53:53

---

Diagnostics Check		OK	01:41	
Air Blank		0.000	01:41	
Control Test		0.079	01:42	
Air Blank		0.000	01:42	
Subject Sample	# 1	0.161	01:45	Volume Not Met
Breath Volume		0.308	01:45	
Air Blank		0.000	01:46	
Air Blank		0.000	01:48	
Subject Sample	# 2	0.185	01:49	
Breath Volume		1.468	01:49	
Air Blank		0.000	01:49	
Air Blank		0.000	01:51	
Subject Sample	# 3	0.192	01:52	
Breath Volume		2.367	01:52	
Air Blank		0.000	01:52	
Control Test		0.077	01:53	
Air Blank		0.000	01:53	
Diagnostics Check		OK	01:53	

---

Data Download Date/Time: 01/16/2020 15:21



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: KEATING,KAYLA  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: VOROZHTSOV

DATE/TIME OF TEST: 01/14/2020 20:54:42

---

Diagnostics Check		OK	20:41	
Air Blank		0.000	20:41	
Control Test		0.079	20:41	
Air Blank		0.000	20:42	
Subject Sample	# 1	0.070	20:45	Volume Not Met
Breath Volume		0.281	20:45	
Air Blank		0.000	20:46	
Air Blank		0.000	20:48	
Subject Sample	# 2	0.078	20:48	
Breath Volume		1.531	20:48	
Air Blank		0.000	20:49	
Air Blank		0.000	20:51	
Subject Sample	# 3	0.071	20:53	
Breath Volume		1.453	20:53	
Air Blank		0.000	20:53	
Control Test		0.079	20:54	
Air Blank		0.000	20:54	
Diagnostics Check		OK	20:54	

---

Data Download Date/Time: 01/16/2020 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BERMAN,MARC  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: FLOREZ RESTREPO

DATE/TIME OF TEST: 01/16/2020 03:09:16

---

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.196	03:04
Breath Volume	1.488	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.200	03:07
Breath Volume	2.085	03:07
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

---

Data Download Date/Time: 01/16/2020 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WASHINGTON, YVETTE, D  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LUNA, KRISTINA  
ARRESTING OFFICER AGENCY: Miramar PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: CAMARGO

DATE/TIME OF TEST: 01/16/2020 18:32:08

---

Diagnostics Check	OK	18:25
Air Blank	0.000	18:26
Control Test	0.079	18:26
Air Blank	0.000	18:27
Subject Sample # 1	0.000	18:27
Breath Volume	3.566	18:27
Air Blank	0.000	18:28
Air Blank	0.000	18:30
Subject Sample # 2	0.000	18:30
Breath Volume	3.257	18:30
Air Blank	0.000	18:31
Control Test	0.079	18:31
Air Blank	0.000	18:32
Diagnostics Check	OK	18:32

---

Data Download Date/Time: 02/21/2020 20:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HAINES,DAVID  
ARRESTING OFFICER AGENCY: Davie PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: PARRISH HOBBS

DATE/TIME OF TEST: 01/16/2020 23:17:18

---

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.080	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.000	23:12
Breath Volume	1.625	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.000	23:15
Breath Volume	1.484	23:15
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

---

Data Download Date/Time: 02/21/2020 20:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ROJAS,RANDY  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 01/17/2020 04:12:48

---

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.130	04:07
Breath Volume	2.074	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.127	04:11
Breath Volume	1.644	04:11
Air Blank	0.000	04:11
Control Test	0.077	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

---

Data Download Date/Time: 02/21/2020 20:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: YASKO,JOE  
ARRESTING OFFICER AGENCY: Lauderhill PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 01/18/2020 16:54:07

---

Diagnostics Check	OK	16:47
Air Blank	0.000	16:48
Control Test	0.079	16:48
Air Blank	0.000	16:48
Subject Sample # 1	0.272	16:49
Breath Volume	2.855	16:49
Air Blank	0.000	16:50
Air Blank	0.000	16:51
Subject Sample # 2	0.265	16:52
Breath Volume	3.027	16:52
Air Blank	0.000	16:53
Control Test	0.078	16:53
Air Blank	0.000	16:54
Diagnostics Check	OK	16:54

---

Data Download Date/Time: 02/21/2020 20:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ARMSTRONG,BRITTANY  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: TELXEIRA

DATE/TIME OF TEST: 01/19/2020 03:12:44

---

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:05
Air Blank	0.000	03:05
Subject Sample # 1	0.119	03:07
Breath Volume	1.554	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.113	03:11
Breath Volume	1.992	03:11
Air Blank	0.000	03:11
Control Test	0.077	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

---

Data Download Date/Time: 02/21/2020 20:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'DOR,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: O'DOR,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: CALLE

DATE/TIME OF TEST: 01/19/2020 04:30:11

---

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.079	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.116	04:25
Breath Volume	3.511	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.116	04:28
Breath Volume	3.011	04:28
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

---

Data Download Date/Time: 02/21/2020 20:51



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: ARAICA

DATE/TIME OF TEST: 01/19/2020 05:00:27

---

Diagnostics Check	OK	04:47	
Air Blank	0.000	04:48	
Control Test	0.078	04:48	
Air Blank	0.000	04:48	
Subject Sample	# 1 0.147	04:49	
Breath Volume	1.656	04:49	
Air Blank	0.000	04:50	
Air Blank	0.000	04:52	
Subject Sample	# 2 0.113	04:55	Volume Not Met
Breath Volume	0.605	04:55	
Air Blank	0.000	04:56	
Air Blank	0.000	04:58	
Subject Sample	# 3 0.140	04:58	
Breath Volume	3.078	04:58	
Air Blank	0.000	04:59	
Control Test	0.076	04:59	
Air Blank	0.000	05:00	
Diagnostics Check	OK	05:00	

---

Data Download Date/Time: 02/21/2020 20:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: PRESUME,JEHUD  
ARRESTING OFFICER AGENCY: Miramar PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: SEALS

DATE/TIME OF TEST: 01/20/2020 20:12:19

---

Diagnostics Check	OK	20:05
Air Blank	0.000	20:05
Control Test	0.079	20:06
Air Blank	0.000	20:06
Subject Sample # 1	0.143	20:07
Breath Volume	1.429	20:07
Air Blank	0.000	20:08
Air Blank	0.000	20:09
Subject Sample # 2	0.147	20:10
Breath Volume	1.859	20:10
Air Blank	0.000	20:11
Control Test	0.079	20:11
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

---

Data Download Date/Time: 02/21/2020 20:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: TERAZON,JOSEPH  
ARRESTING OFFICER AGENCY: Hollywood PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: KORCE II

DATE/TIME OF TEST: 01/21/2020 03:01:25

---

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.158	02:56
Breath Volume	3.171	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.148	02:59
Breath Volume	1.941	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

---

Data Download Date/Time: 02/21/2020 20:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: CHAPPIN DUNCAN,DAVID  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: DELACRUZ HERNANDEZ

DATE/TIME OF TEST: 01/25/2020 20:18:18

---

Diagnostics Check	OK	20:10
Air Blank	0.000	20:10
Control Test	0.080	20:11
Air Blank	0.000	20:11
Subject Sample # 1	0.252	20:13
Breath Volume	3.023	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:16
Subject Sample # 2	0.245	20:16
Breath Volume	2.785	20:16
Air Blank	0.000	20:17
Control Test	0.076	20:17
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

---

Data Download Date/Time: 02/21/2020 20:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS  
ARRESTING OFFICER AGENCY: Miramar PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 01/26/2020 00:47:21

---

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.079	00:41
Air Blank	0.000	00:42
Subject Sample # 1	0.097	00:42
Breath Volume	2.589	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.094	00:45
Breath Volume	1.441	00:45
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

---

Data Download Date/Time: 02/21/2020 20:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: CHAPPIN DUNCAN,DAVID  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 01/26/2020 03:30:16

---

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:19	
Control Test		0.079	03:19	
Air Blank		0.000	03:19	
Subject Sample	# 1	0.119	03:23	Volume Not Met
Breath Volume		0.097	03:23	
Air Blank		0.000	03:23	
Air Blank		0.000	03:25	
Subject Sample	# 2	0.118	03:28	Volume Not Met
Breath Volume		0.351	03:28	
Air Blank		0.000	03:29	
Control Test		0.079	03:29	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:30	

---

Data Download Date/Time: 02/21/2020 20:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: CHAPPIN DUNCAN,DAVID  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 01/26/2020 03:45:05

---

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:34	
Control Test		0.078	03:34	
Air Blank		0.000	03:34	
Subject Sample	# 1	0.120	03:38	Volume Not Met
Breath Volume		0.050	03:38	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 2		03:43	No Sample Provided
Breath Volume		0.000	03:43	
Air Blank		0.000	03:44	
Control Test		0.078	03:44	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

---

Data Download Date/Time: 02/21/2020 20:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: CAMPBELL,DWAYNE  
ARRESTING OFFICER AGENCY: Miramar PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: URIA

DATE/TIME OF TEST: 01/27/2020 05:38:09

---

Diagnostics Check	OK	05:29
Air Blank	0.000	05:30
Control Test	0.080	05:30
Air Blank	0.000	05:31
Subject Sample # 1	0.142	05:32
Breath Volume	2.230	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.130	05:36
Breath Volume	1.191	05:36
Air Blank	0.000	05:37
Control Test	0.078	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

---

Data Download Date/Time: 02/21/2020 20:51



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LEON,JOEL  
ARRESTING OFFICER AGENCY: Hallandale Beach PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: GAVINO

DATE/TIME OF TEST: 01/28/2020 03:18:26

---

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.080	03:16	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:17	
Control Test	0.080	03:17	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

---

Data Download Date/Time: 02/21/2020 20:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: COLLINS,ALBERT  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 01/29/2020 03:37:22

---

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.117	03:32
Breath Volume	1.738	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.122	03:35
Breath Volume	2.359	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

---

Data Download Date/Time: 02/21/2020 20:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: ROSENBLUM

DATE/TIME OF TEST: 01/29/2020 21:50:49

---

Diagnostics Check	OK	21:44
Air Blank	0.000	21:45
Control Test	0.080	21:45
Air Blank	0.000	21:46
Subject Sample # 1	0.000	21:46
Breath Volume	2.238	21:46
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.000	21:49
Breath Volume	2.203	21:49
Air Blank	0.000	21:49
Control Test	0.079	21:50
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

---

Data Download Date/Time: 02/21/2020 20:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARMSTRONG,B,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ARMSTRONG,B  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/30/2020 01:11:08

---

Diagnostics Check		OK	01:04
Air Blank		0.000	01:05
Control Test		0.080	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.238	01:06
Breath Volume		1.605	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.231	01:09
Breath Volume		1.613	01:09
Air Blank		0.000	01:10
Control Test		0.077	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

---

Data Download Date/Time: 02/21/2020 20:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HERNANDEZ,ERNESTO  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: BRAY

DATE/TIME OF TEST: 01/31/2020 15:16:12

---

Diagnostics Check	OK	15:09
Air Blank	0.000	15:10
Control Test	0.080	15:10
Air Blank	0.000	15:10
Subject Sample # 1	0.327	15:11
Breath Volume	1.570	15:11
Air Blank	0.000	15:12
Air Blank	0.000	15:13
Subject Sample # 2	0.336	15:14
Breath Volume	2.085	15:14
Air Blank	0.000	15:15
Control Test	0.077	15:15
Air Blank	0.000	15:16
Diagnostics Check	OK	15:16

---

Data Download Date/Time: 02/21/2020 20:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ARMSTRONG,BRITTANY  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: LOISEAU

DATE/TIME OF TEST: 01/17/2020 00:03:45

---

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:50	
Control Test	0.079	23:51	
Air Blank	0.000	23:51	
Subject Sample	# 2	23:54	No Sample Provided
Breath Volume	0.000	23:54	
Air Blank	0.000	23:55	
Air Blank	0.000	23:57	
Subject Sample	# 3	0.276	23:58
Breath Volume	2.160	23:58	
Air Blank	0.000	23:59	
Air Blank	0.000	00:01	
Subject Sample	# 1	0.296	00:01
Breath Volume	2.085	00:01	
Air Blank	0.000	00:02	
Control Test	0.079	00:03	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

---

Data Download Date/Time: 02/21/2020 21:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: KEATING,KAYLA  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: BOWRA

DATE/TIME OF TEST: 01/17/2020 21:41:19

---

Diagnostics Check		OK	21:26	
Air Blank		0.000	21:26	
Control Test		0.080	21:27	
Air Blank		0.000	21:27	
Subject Sample	# 1	0.230	21:30	Volume Not Met
Breath Volume		0.382	21:30	
Air Blank		0.000	21:31	
Air Blank		0.000	21:33	
Subject Sample	# 2	0.271	21:34	
Breath Volume		1.683	21:34	
Air Blank		0.000	21:34	
Air Blank		0.000	21:36	
Subject Sample	# 3	0.259	21:39	
Breath Volume		1.667	21:39	
Air Blank		0.000	21:40	
Control Test		0.077	21:40	
Air Blank		0.000	21:41	
Diagnostics Check		OK	21:41	

---

Data Download Date/Time: 02/21/2020 21:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ULLOA,ADRIAN  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 01/18/2020 00:52:12

---

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.122	00:47
Breath Volume	1.808	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.122	00:50
Breath Volume	2.242	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

---

Data Download Date/Time: 02/21/2020 21:05



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ROJAS,RANDY  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: MALETTE

DATE/TIME OF TEST: 01/18/2020 06:25:50

---

Diagnostics Check		OK	06:10	
Air Blank		0.000	06:10	
Control Test		0.080	06:11	
Air Blank		0.000	06:11	
Subject Sample	# 1	0.130	06:14	Volume Not Met
Breath Volume		0.421	06:14	
Air Blank		0.000	06:15	
Air Blank		0.000	06:17	
Subject Sample	# 2	0.166	06:19	
Breath Volume		1.683	06:19	
Air Blank		0.000	06:20	
Air Blank		0.000	06:21	
Subject Sample	# 3	0.169	06:24	
Breath Volume		1.156	06:24	
Air Blank		0.000	06:24	
Control Test		0.076	06:25	
Air Blank		0.000	06:25	
Diagnostics Check		OK	06:25	

---

Data Download Date/Time: 02/21/2020 21:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BERMAN,MARC  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: JAIMES

DATE/TIME OF TEST: 01/19/2020 01:36:28

---

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.163	01:31
Breath Volume	1.769	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.166	01:34
Breath Volume	1.386	01:34
Air Blank	0.000	01:35
Control Test	0.078	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

---

Data Download Date/Time: 02/21/2020 21:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: IRELAN,ANDREW  
ARRESTING OFFICER AGENCY: Margate PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: COTO

DATE/TIME OF TEST: 01/19/2020 02:41:32

---

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.079	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.162	02:36
Breath Volume	1.113	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.168	02:39
Breath Volume	1.695	02:39
Air Blank	0.000	02:40
Control Test	0.078	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

---

Data Download Date/Time: 02/21/2020 21:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: GAINZA

DATE/TIME OF TEST: 01/19/2020 03:10:07

---

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.078	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.099	03:03
Breath Volume	1.652	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:05
Subject Sample # 2	0.095	03:08
Breath Volume	1.324	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

---

Data Download Date/Time: 02/21/2020 21:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BARNES,KENNETH  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:34

SUBJECT'S LAST NAME: COMEAUX

DATE/TIME OF TEST: 01/19/2020 14:02:44

---

Diagnostics Check	OK	13:56
Air Blank	0.000	13:56
Control Test	0.079	13:57
Air Blank	0.000	13:57
Subject Sample # 1	0.268	13:58
Breath Volume	2.242	13:58
Air Blank	0.000	13:58
Air Blank	0.000	14:00
Subject Sample # 2	0.257	14:01
Breath Volume	1.316	14:01
Air Blank	0.000	14:01
Control Test	0.080	14:02
Air Blank	0.000	14:02
Diagnostics Check	OK	14:02

---

Data Download Date/Time: 02/21/2020 21:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'DOR,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: O'DOR,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 01/20/2020 00:01:05

---

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.280	23:56
Breath Volume	3.179	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:58
Subject Sample # 2	0.264	23:59
Breath Volume	2.941	23:59
Air Blank	0.000	00:00
Control Test	0.079	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:01

---

Data Download Date/Time: 02/21/2020 21:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ALMERIA,RENDELL  
ARRESTING OFFICER AGENCY: Pembroke Pines PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:17

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/21/2020 06:46:21

---

Diagnostics Check		OK	06:38
Air Blank		0.000	06:39
Control Test		0.080	06:39
Air Blank		0.000	06:40
Subject Sample	# 1	0.224	06:40
Breath Volume		2.078	06:40
Air Blank		0.000	06:41
Air Blank		0.000	06:43
Subject Sample	# 2	0.217	06:44
Breath Volume		1.980	06:44
Air Blank		0.000	06:45
Control Test		0.080	06:45
Air Blank		0.000	06:46
Diagnostics Check		OK	06:46

---

Data Download Date/Time: 02/21/2020 21:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HERNANDEZ,BRIAN  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: DEVASWAMPARAMBIL BAL

DATE/TIME OF TEST: 01/21/2020 16:53:59

---

Diagnostics Check	OK	16:42
Air Blank	0.000	16:43
Control Test	0.081	16:43
Air Blank	0.000	16:44
Subject Sample # 1	0.193	16:46
Breath Volume	1.683	16:46
Air Blank	0.000	16:47
Air Blank	0.000	16:48
Subject Sample # 2	0.201	16:52
Breath Volume	2.828	16:52
Air Blank	0.000	16:52
Control Test	0.078	16:53
Air Blank	0.000	16:53
Diagnostics Check	OK	16:53

---

Data Download Date/Time: 02/21/2020 21:05



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ROJAS,RANDY  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: WOJTCZAK

DATE/TIME OF TEST: 01/24/2020 05:04:44

---

Diagnostics Check		OK	04:58
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Subject Sample	# 1	0.224	05:00
Breath Volume		2.601	05:00
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.231	05:03
Breath Volume		1.867	05:03
Air Blank		0.000	05:03
Control Test		0.078	05:04
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

---

Data Download Date/Time: 02/21/2020 21:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: CHAPPIN DUNCAN,DAVID  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: PEREZ TOBAR

DATE/TIME OF TEST: 01/25/2020 01:11:12

---

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.080	01:04
Air Blank		0.000	01:05
Subject Sample	# 1	0.208	01:06
Breath Volume		2.933	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.211	01:09
Breath Volume		2.832	01:09
Air Blank		0.000	01:10
Control Test		0.078	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

---

Data Download Date/Time: 02/21/2020 21:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BOHM,JONATHAN  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: GUTIERREZ RAMIREZ

DATE/TIME OF TEST: 01/25/2020 02:25:19

---

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.186	02:20
Breath Volume	2.839	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample # 2	0.186	02:23
Breath Volume	2.695	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

---

Data Download Date/Time: 02/21/2020 21:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: FOWLER,SHAUN  
ARRESTING OFFICER AGENCY: Hollywood PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 01/25/2020 14:45:39

---

Diagnostics Check	OK	14:39
Air Blank	0.000	14:39
Control Test	0.080	14:39
Air Blank	0.000	14:40
Subject Sample # 1	0.182	14:41
Breath Volume	2.230	14:41
Air Blank	0.000	14:41
Air Blank	0.000	14:43
Subject Sample # 2	0.183	14:43
Breath Volume	2.996	14:43
Air Blank	0.000	14:44
Control Test	0.080	14:45
Air Blank	0.000	14:45
Diagnostics Check	OK	14:45

---

Data Download Date/Time: 02/21/2020 21:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: PENA,RYAN  
ARRESTING OFFICER AGENCY: Davie PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: PAPP

DATE/TIME OF TEST: 01/25/2020 21:36:26

---

Diagnostics Check	OK	21:27
Air Blank	0.000	21:28
Control Test	0.080	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.159	21:30
Breath Volume	1.410	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample # 2	0.151	21:34
Breath Volume	1.496	21:34
Air Blank	0.000	21:35
Control Test	0.078	21:35
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

---

Data Download Date/Time: 02/21/2020 21:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BASOA,JESSE  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: PARRA

DATE/TIME OF TEST: 01/26/2020 21:21:30

---

Diagnostics Check	OK	21:15
Air Blank	0.000	21:15
Control Test	0.080	21:15
Air Blank	0.000	21:16
Subject Sample # 1	0.168	21:16
Breath Volume	2.289	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.161	21:19
Breath Volume	3.242	21:19
Air Blank	0.000	21:20
Control Test	0.080	21:20
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

---

Data Download Date/Time: 02/21/2020 21:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BAIN,MERVIN  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: GALAN HERNANDEZ

DATE/TIME OF TEST: 01/26/2020 22:13:30

---

Diagnostics Check	OK	22:06
Air Blank	0.000	22:07
Control Test	0.080	22:07
Air Blank	0.000	22:07
Subject Sample # 1	0.151	22:08
Breath Volume	1.726	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.146	22:11
Breath Volume	1.996	22:11
Air Blank	0.000	22:12
Control Test	0.079	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

---

Data Download Date/Time: 02/21/2020 21:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BERMAN,MARC  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: ATEHORTUA

DATE/TIME OF TEST: 01/26/2020 22:58:43

---

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:53
Subject Sample # 1	0.000	22:54
Breath Volume	3.289	22:54
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.000	22:57
Breath Volume	2.277	22:57
Air Blank	0.000	22:57
Control Test	0.080	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

---

Data Download Date/Time: 02/21/2020 21:06



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: STONER,NATHAN  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 01/26/2020 23:49:13

---

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.142	23:44
Breath Volume	2.195	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:46
Subject Sample # 2	0.145	23:47
Breath Volume	2.285	23:47
Air Blank	0.000	23:48
Control Test	0.079	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

---

Data Download Date/Time: 02/21/2020 21:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER,MONTREZ,R  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: WALKER,MARINA  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG923908  
DRY GAS EXPIRY DATE: 08/27/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: OATES

DATE/TIME OF TEST: 01/01/2020 01:24:19

---

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.082	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.176	01:19
Breath Volume	1.273	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.159	01:22
Breath Volume	1.199	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

---

Data Download Date/Time: 01/30/2020 14:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER,MONTREZ,R  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: POTTER,MONTREZ  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG923908  
DRY GAS EXPIRY DATE: 08/27/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: COLVIN

DATE/TIME OF TEST: 01/01/2020 21:05:45

---

Diagnostics Check	OK	21:02	
Air Blank	0.000	21:03	
Control Test	0.082	21:03	
Air Blank	0.000	21:04	
Subject Sample	# 1	21:04	Subject Test Refused
Breath Volume	0.000	21:04	
Air Blank	0.000	21:04	
Control Test	0.081	21:05	
Air Blank	0.000	21:05	
Diagnostics Check	OK	21:05	

---

Data Download Date/Time: 01/30/2020 14:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER,MONTREZ,R  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: FUGATE,KATIE  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG923908  
DRY GAS EXPIRY DATE: 08/27/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 01/02/2020 19:34:57

---

Diagnostics Check	OK	19:32	
Air Blank	0.000	19:32	
Control Test	0.081	19:32	
Air Blank	0.000	19:33	
Subject Sample	# 1	19:33	Subject Test Refused
Breath Volume	0.000	19:33	
Air Blank	0.000	19:34	
Control Test	0.080	19:34	
Air Blank	0.000	19:34	
Diagnostics Check	OK	19:34	

---

Data Download Date/Time: 01/30/2020 14:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER,MONTREZ,R  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: POTTER,MONTREZ  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG923908  
DRY GAS EXPIRY DATE: 08/27/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/02/2020 20:39:24

---

Diagnostics Check	OK	20:31
Air Blank	0.000	20:31
Control Test	0.081	20:32
Air Blank	0.000	20:32
Subject Sample # 1	0.197	20:33
Breath Volume	1.265	20:33
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.199	20:37
Breath Volume	1.476	20:37
Air Blank	0.000	20:38
Control Test	0.079	20:38
Air Blank	0.000	20:39
Diagnostics Check	OK	20:39

---

Data Download Date/Time: 01/30/2020 14:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER,MONTREZ,R  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: HUTCHINSON,JOSEPH  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG923908  
DRY GAS EXPIRY DATE: 08/27/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 01/11/2020 05:14:17

---

Diagnostics Check	OK	05:07
Air Blank	0.000	05:07
Control Test	0.081	05:08
Air Blank	0.000	05:08
Subject Sample # 1	0.269	05:09
Breath Volume	1.808	05:09
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.276	05:12
Breath Volume	1.550	05:12
Air Blank	0.000	05:13
Control Test	0.080	05:13
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

---

Data Download Date/Time: 01/30/2020 14:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER,MONTREZ,R  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: POTTER,MONTREZ  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG923908  
DRY GAS EXPIRY DATE: 08/27/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: GALANFELICIANO

DATE/TIME OF TEST: 01/11/2020 23:07:35

---

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.081	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.155	23:00
Breath Volume	1.773	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:02
Subject Sample # 2	0.140	23:05
Breath Volume	2.023	23:05
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

---

Data Download Date/Time: 01/30/2020 14:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER,MONTREZ,R  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: LAROCHE,THOMAS  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG923908  
DRY GAS EXPIRY DATE: 08/27/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 01/12/2020 05:06:32

---

Diagnostics Check		OK	04:59
Air Blank		0.000	05:00
Control Test		0.080	05:00
Air Blank		0.000	05:01
Subject Sample	# 1	0.156	05:01
Breath Volume		2.515	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.148	05:04
Breath Volume		1.617	05:04
Air Blank		0.000	05:05
Control Test		0.080	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

---

Data Download Date/Time: 01/30/2020 14:59



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER,MONTREZ,R  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: GOBER,JOSHUA  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG923908  
DRY GAS EXPIRY DATE: 08/27/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/12/2020 05:38:22

---

Diagnostics Check	OK	05:30
Air Blank	0.000	05:31
Control Test	0.081	05:31
Air Blank	0.000	05:32
Subject Sample # 1	0.163	05:32
Breath Volume	1.164	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.170	05:36
Breath Volume	1.234	05:36
Air Blank	0.000	05:37
Control Test	0.080	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

---

Data Download Date/Time: 01/30/2020 14:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER,MONTREZ,R  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: CLAYTON,STEVEN  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG923908  
DRY GAS EXPIRY DATE: 08/27/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: GAUL

DATE/TIME OF TEST: 01/17/2020 02:13:52

---

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.082	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.225	02:08
Breath Volume	1.585	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:10
Subject Sample # 2	0.211	02:12
Breath Volume	1.261	02:12
Air Blank	0.000	02:12
Control Test	0.081	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

---

Data Download Date/Time: 01/30/2020 14:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER,MONTREZ,R  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: FUGATE,KATIE  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG923908  
DRY GAS EXPIRY DATE: 08/27/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: BENJUMEA

DATE/TIME OF TEST: 01/20/2020 01:30:04

---

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:27	
Control Test	0.083	01:27	
Air Blank	0.000	01:28	
Subject Sample	# 1	01:28	Subject Test Refused
Breath Volume	0.000	01:28	
Air Blank	0.000	01:29	
Control Test	0.082	01:29	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:30	

---

Data Download Date/Time: 01/30/2020 14:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FITHIAN,JASON,  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: KEATING,K  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:51

SUBJECT'S LAST NAME: TOLENTINO

DATE/TIME OF TEST: 01/04/2020 17:18:09

---

Diagnostics Check	OK	17:11
Air Blank	0.000	17:12
Control Test	0.076	17:12
Air Blank	0.000	17:12
Subject Sample # 1	0.189	17:13
Breath Volume	4.117	17:13
Air Blank	0.000	17:14
Air Blank	0.000	17:15
Subject Sample # 2	0.191	17:16
Breath Volume	3.785	17:16
Air Blank	0.000	17:17
Control Test	0.076	17:17
Air Blank	0.000	17:18
Diagnostics Check	OK	17:18

---

Data Download Date/Time: 01/16/2020 12:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FITHIAN,JASON,P  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: FITHIAN,J  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: MANGUBA

DATE/TIME OF TEST: 01/11/2020 02:30:22

---

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.076	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.293	02:25
Breath Volume	3.175	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:27
Subject Sample # 2	0.289	02:28
Breath Volume	2.273	02:28
Air Blank	0.000	02:29
Control Test	0.077	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

---

Data Download Date/Time: 01/16/2020 12:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FITHIAN,JASON,  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618090A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: RIZZA

DATE/TIME OF TEST: 01/15/2020 00:16:53

---

Diagnostics Check		OK	00:05
Air Blank		0.000	00:06
Control Test		0.077	00:06
Air Blank		0.000	00:07
Subject Sample	# 1	0.247	00:07
Breath Volume		1.773	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.219	00:11
Breath Volume		1.746	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:13
Subject Sample	# 3	0.237	00:15
Breath Volume		2.335	00:15
Air Blank		0.000	00:15
Control Test		0.077	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

---

Data Download Date/Time: 01/16/2020 12:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FITHIAM,JASON,  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: FITHIAN,JASON  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MERCADO

DATE/TIME OF TEST: 01/30/2020 21:48:02

---

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.078	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.167	21:43
Breath Volume	3.402	21:43
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.160	21:46
Breath Volume	2.941	21:46
Air Blank	0.000	21:47
Control Test	0.077	21:47
Air Blank	0.000	21:47
Diagnostics Check	OK	21:48

---

Data Download Date/Time: 02/20/2020 20:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: GARAU

DATE/TIME OF TEST: 01/03/2020 01:11:09

---

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.077	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.291	01:04
Breath Volume	1.644	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:06
Subject Sample # 2	0.287	01:09
Breath Volume	2.488	01:09
Air Blank	0.000	01:10
Control Test	0.077	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

---

Data Download Date/Time: 01/16/2020 15:30



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02618080A3  
DRY GAS EXPIRY DATE: 03/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: LINCOLN

DATE/TIME OF TEST: 01/12/2020 02:03:15

---

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.077	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.106	01:58
Breath Volume	3.781	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.110	02:01
Breath Volume	4.054	02:01
Air Blank	0.000	02:02
Control Test	0.077	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

---

Data Download Date/Time: 01/16/2020 15:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANO,MARDEE,G  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: PATTEN,J  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: BECKFORD

DATE/TIME OF TEST: 01/05/2020 04:14:55

---

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.110	04:10
Breath Volume	1.812	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.110	04:13
Breath Volume	1.605	04:13
Air Blank	0.000	04:13
Control Test	0.081	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

---

Data Download Date/Time: 01/29/2020 00:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JUDD,DEP,B  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: STUAB,D  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: WALTER

DATE/TIME OF TEST: 01/07/2020 18:21:33

---

Diagnostics Check	OK	18:12
Air Blank	0.000	18:13
Control Test	0.080	18:13
Air Blank	0.000	18:14
Subject Sample # 1	0.203	18:16
Breath Volume	1.566	18:16
Air Blank	0.000	18:17
Air Blank	0.000	18:18
Subject Sample # 2	0.199	18:19
Breath Volume	1.703	18:19
Air Blank	0.000	18:20
Control Test	0.081	18:20
Air Blank	0.000	18:21
Diagnostics Check	OK	18:21

---

Data Download Date/Time: 01/29/2020 00:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAFFARAS,CELINA,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: ALLEN,M  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: MARINO

DATE/TIME OF TEST: 01/08/2020 21:28:41

---

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.080	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.179	21:23
Breath Volume	2.632	21:23
Air Blank	0.000	21:24
Air Blank	0.000	21:25
Subject Sample # 2	0.180	21:27
Breath Volume	2.812	21:27
Air Blank	0.000	21:27
Control Test	0.080	21:28
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

---

Data Download Date/Time: 01/29/2020 00:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PROSPER,BENSON,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GRIFFIN,M  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 08:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:55

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/11/2020 09:23:43

---

Diagnostics Check	OK	09:16
Air Blank	0.000	09:17
Control Test	0.080	09:17
Air Blank	0.000	09:18
Subject Sample # 1	0.066	09:19
Breath Volume	1.390	09:19
Air Blank	0.000	09:19
Air Blank	0.000	09:21
Subject Sample # 2	0.059	09:22
Breath Volume	2.050	09:22
Air Blank	0.000	09:22
Control Test	0.080	09:23
Air Blank	0.000	09:23
Diagnostics Check	OK	09:23

---

Data Download Date/Time: 01/29/2020 00:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JUDD,BOBBY,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: LINDSEY,JOSHUA  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 01/12/2020 02:34:01

---

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.080	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.183	02:28
Breath Volume		2.222	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.175	02:32
Breath Volume		2.476	02:32
Air Blank		0.000	02:33
Control Test		0.085	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:34

---

Data Download Date/Time: 01/29/2020 00:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JUDD,DEP,B  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: ERICKSON,CHRISTINA  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MACQUARRIE

DATE/TIME OF TEST: 01/12/2020 03:27:17

---

Diagnostics Check	OK	03:26	
Air Blank		03:27	RFI Detect
Air Blank		03:27	RFI Detect

---

Data Download Date/Time: 01/29/2020 00:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JUDD,DEP,B  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: ERICKSON,CHRISTINA  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MACQUARRIA

DATE/TIME OF TEST: 01/12/2020 03:37:13

---

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.080	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.243	03:32
Breath Volume		2.589	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:34
Subject Sample	# 2	0.242	03:35
Breath Volume		2.570	03:35
Air Blank		0.000	03:36
Control Test		0.079	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

---

Data Download Date/Time: 01/29/2020 00:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EGAN,MELISSA,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: LAROSA,RYAN  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 01/14/2020 03:46:15

---

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.069	03:41
Breath Volume	3.226	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.066	03:44
Breath Volume	3.375	03:44
Air Blank	0.000	03:45
Control Test	0.082	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

---

Data Download Date/Time: 01/29/2020 00:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JUDD,DEP,B  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: LINDSEY,J  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: MERCORELLI

DATE/TIME OF TEST: 01/15/2020 20:55:36

---

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.080	20:49
Air Blank	0.000	20:49
Subject Sample # 1	0.000	20:51
Breath Volume	1.394	20:51
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.000	20:54
Breath Volume	2.082	20:54
Air Blank	0.000	20:54
Control Test	0.080	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

---

Data Download Date/Time: 01/29/2020 00:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,D.,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: WRASSE,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: LARIOS

DATE/TIME OF TEST: 01/16/2020 03:21:25

---

Diagnostics Check		OK	03:13
Air Blank		0.000	03:14
Control Test		0.080	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.188	03:16
Breath Volume		2.421	03:16
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.187	03:19
Breath Volume		2.089	03:19
Air Blank		0.000	03:20
Control Test		0.079	03:20
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

---

Data Download Date/Time: 01/29/2020 00:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EGAN,MELISSA,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: OWENS,N  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: GRANADO VELARDE

DATE/TIME OF TEST: 01/18/2020 01:01:46

---

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.165	00:56
Breath Volume	2.824	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.152	01:00
Breath Volume	2.738	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

---

Data Download Date/Time: 01/29/2020 00:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,D,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: ERICKSON,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: ROY

DATE/TIME OF TEST: 01/22/2020 01:22:18

---

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.081	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.000	01:16
Breath Volume	3.824	01:16
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.000	01:20
Breath Volume	3.800	01:20
Air Blank	0.000	01:21
Control Test	0.081	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

---

Data Download Date/Time: 01/29/2020 00:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,D,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: METROKA,K  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: PLUMMER

DATE/TIME OF TEST: 01/22/2020 02:14:16

---

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:03	
Control Test	0.081	02:03	
Air Blank	0.000	02:04	
Subject Sample # 1	0.133	02:07	
Breath Volume	1.671	02:07	
Air Blank	0.000	02:07	
Air Blank	0.000	02:09	
Subject Sample # 2	0.133	02:12	Volume Not Met
Breath Volume	0.085	02:12	
Air Blank	0.000	02:13	
Air Blank	0.000	02:14	

---

Data Download Date/Time: 01/29/2020 00:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EGAN,MELISSA,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: HARTSMANSTORPHER,CHRIS  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: TALREJA

DATE/TIME OF TEST: 01/22/2020 22:25:27

---

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.080	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.088	22:19
Breath Volume	1.433	22:19
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.095	22:23
Breath Volume	1.347	22:23
Air Blank	0.000	22:24
Control Test	0.082	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

---

Data Download Date/Time: 01/29/2020 00:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JUDD,DEP,B  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: ERICKSON,CHRISTINA  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: BISSONNETTE

DATE/TIME OF TEST: 01/24/2020 22:54:11

---

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.080	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.129	22:49
Breath Volume	2.664	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:51
Subject Sample # 2	0.129	22:52
Breath Volume	2.570	22:52
Air Blank	0.000	22:53
Control Test	0.083	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

---

Data Download Date/Time: 01/29/2020 00:48



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RONGA,DAVID,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: JOSEPH,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 10:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: MATTHEW

DATE/TIME OF TEST: 01/25/2020 11:18:45

---

Diagnostics Check		OK	11:11
Air Blank		0.000	11:12
Control Test		0.080	11:12
Air Blank		0.000	11:13
Subject Sample	# 1	0.301	11:13
Breath Volume		2.886	11:13
Air Blank		0.000	11:14
Air Blank		0.000	11:16
Subject Sample	# 2	0.301	11:17
Breath Volume		3.164	11:17
Air Blank		0.000	11:17
Control Test		0.081	11:18
Air Blank		0.000	11:18
Diagnostics Check		OK	11:18

---

Data Download Date/Time: 01/29/2020 00:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEONARD,ALEYA,D  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: CURTIS,RJ  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: MARCIANO

DATE/TIME OF TEST: 01/25/2020 21:51:38

---

Diagnostics Check	OK	21:44
Air Blank	0.000	21:45
Control Test	0.080	21:45
Air Blank	0.000	21:46
Subject Sample # 1	0.041	21:47
Breath Volume	1.261	21:47
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.036	21:50
Breath Volume	1.523	21:50
Air Blank	0.000	21:50
Control Test	0.080	21:51
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

---

Data Download Date/Time: 01/29/2020 00:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEONARD,ALEYA,D  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: TAYLOR,BOBBY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: STENFELDT

DATE/TIME OF TEST: 01/26/2020 01:27:20

---

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.081	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.194	01:22
Breath Volume	1.824	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.196	01:25
Breath Volume	1.925	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

---

Data Download Date/Time: 01/29/2020 00:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEONARD,ALEYA,D  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: BURGA,FERNANDO  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: COALWELL

DATE/TIME OF TEST: 01/26/2020 03:35:22

---

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.197	03:30
Breath Volume	1.910	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.198	03:33
Breath Volume	2.183	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

---

Data Download Date/Time: 01/29/2020 00:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAFFARAS,CELINA,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: KLOER,C  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 12/04/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SEWELL

DATE/TIME OF TEST: 01/28/2020 02:36:59

---

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.080	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.124	02:32
Breath Volume	2.410	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.124	02:35
Breath Volume	1.992	02:35
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

---

Data Download Date/Time: 01/29/2020 00:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUNN,CDFC,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: TAYLOR,R  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: BUBLITZ

DATE/TIME OF TEST: 01/29/2020 17:22:55

---

Diagnostics Check	OK	17:15
Air Blank	0.000	17:15
Control Test	0.080	17:16
Air Blank	0.000	17:16
Subject Sample # 1	0.255	17:17
Breath Volume	1.761	17:17
Air Blank	0.000	17:18
Air Blank	0.000	17:20
Subject Sample # 2	0.243	17:21
Breath Volume	1.460	17:21
Air Blank	0.000	17:21
Control Test	0.084	17:22
Air Blank	0.000	17:22
Diagnostics Check	OK	17:22

---

Data Download Date/Time: 02/25/2020 01:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JUDD,DEP,B  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: DAVIDSON,MICHAEL  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 01/31/2020 01:08:53

---

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.080	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.177	01:03
Breath Volume	3.531	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.177	01:07
Breath Volume	2.953	01:07
Air Blank	0.000	01:07
Control Test	0.082	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

---

Data Download Date/Time: 02/25/2020 01:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRASER, MICHAEL, C  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: BROWN, BRENDAN  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 01/01/2020 02:27:42

---

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.081	02:22
Air Blank		0.000	02:22
Subject Sample	# 1	0.042	02:23
Breath Volume		4.011	02:23
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.041	02:26
Breath Volume		4.101	02:26
Air Blank		0.000	02:26
Control Test		0.080	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

---

Data Download Date/Time: 01/11/2020 20:42



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEACH,ZACHARY,M  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: BROWN,BRENDAN  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 01/05/2020 04:18:42

---

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.081	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.056	04:14
Breath Volume		2.222	04:14
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.059	04:17
Breath Volume		2.089	04:17
Air Blank		0.000	04:17
Control Test		0.080	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

---

Data Download Date/Time: 01/11/2020 20:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEACH,ZACHARY,M  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: ABDELHADI,AUSTIN  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:21

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/05/2020 17:54:08

---

Diagnostics Check	OK	17:47
Air Blank	0.000	17:48
Control Test	0.081	17:48
Air Blank	0.000	17:49
Subject Sample # 1	0.000	17:49
Breath Volume	2.828	17:49
Air Blank	0.000	17:50
Air Blank	0.000	17:52
Subject Sample # 2	0.000	17:52
Breath Volume	2.335	17:52
Air Blank	0.000	17:53
Control Test	0.079	17:53
Air Blank	0.000	17:54
Diagnostics Check	OK	17:54

---

Data Download Date/Time: 01/11/2020 20:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: HUDSON,JOSHUA  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/09/2020 04:18:51

---

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.080	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.120	04:14
Breath Volume		2.613	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:16
Subject Sample	# 2	0.112	04:17
Breath Volume		1.472	04:17
Air Blank		0.000	04:17
Control Test		0.080	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

---

Data Download Date/Time: 01/11/2020 20:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: HUDSON,JOSHUA  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: SARGENT

DATE/TIME OF TEST: 01/10/2020 01:23:51

---

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.166	01:16
Breath Volume	4.304	01:16
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.141	01:19
Breath Volume	1.621	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 3	0.158	01:22
Breath Volume	2.027	01:22
Air Blank	0.000	01:22
Control Test	0.079	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

---

Data Download Date/Time: 01/11/2020 20:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLLINS,CHASE,M  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: COLLINS,CHASE  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 01/11/2020 04:45:18

---

Diagnostics Check		OK	04:38
Air Blank		0.000	04:38
Control Test		0.079	04:39
Air Blank		0.000	04:39
Subject Sample	# 1	0.254	04:40
Breath Volume		2.597	04:40
Air Blank		0.000	04:41
Air Blank		0.000	04:43
Subject Sample	# 2	0.240	04:43
Breath Volume		2.406	04:43
Air Blank		0.000	04:44
Control Test		0.077	04:44
Air Blank		0.000	04:45
Diagnostics Check		OK	04:45

---

Data Download Date/Time: 01/11/2020 20:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,BRANDON,W  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: WILLIAMS,BRANDON  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 01/11/2020 23:59:50

---

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.089	23:55
Breath Volume	2.105	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.086	23:58
Breath Volume	1.992	23:58
Air Blank	0.000	23:58
Control Test	0.080	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

---

Data Download Date/Time: 02/17/2020 23:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMAHON,KENNETH,J  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: MCMAHON,KENNETH  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: CLINGERMAN

DATE/TIME OF TEST: 01/12/2020 02:37:36

---

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.141	02:32
Breath Volume	2.312	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.126	02:35
Breath Volume	1.863	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

---

Data Download Date/Time: 02/17/2020 23:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARLIER,STEVEN,J  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: WILLIAMS,BRANDON  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: KOON

DATE/TIME OF TEST: 01/13/2020 01:44:48

---

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.081	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.000	01:39
Breath Volume	1.988	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.000	01:43
Breath Volume	3.875	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

---

Data Download Date/Time: 02/17/2020 23:45



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: HUDSON,JOSHUA  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 01/18/2020 00:40:19

---

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.080	00:31
Air Blank		0.000	00:32
Subject Sample	# 1	0.135	00:35
Breath Volume		1.820	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:37
Subject Sample	# 2	0.134	00:38
Breath Volume		1.687	00:38
Air Blank		0.000	00:39
Control Test		0.077	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

---

Data Download Date/Time: 02/17/2020 23:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DANIELS,JAMES,A  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: DANIELS,JAMES  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: BRUNI

DATE/TIME OF TEST: 01/18/2020 01:35:34

---

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.079	01:29
Air Blank		0.000	01:29
Subject Sample	# 1	0.211	01:30
Breath Volume		2.179	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.204	01:33
Breath Volume		3.148	01:33
Air Blank		0.000	01:34
Control Test		0.079	01:34
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

---

Data Download Date/Time: 02/17/2020 23:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMAHON,KENNETH,J  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: BROWN,BRENDAN  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: BREWER

DATE/TIME OF TEST: 01/18/2020 02:06:17

---

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.185	02:01
Breath Volume	1.882	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.184	02:04
Breath Volume	1.964	02:04
Air Blank	0.000	02:05
Control Test	0.078	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

---

Data Download Date/Time: 02/17/2020 23:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMAHON,KENNETH,J  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: BROWN,BRENDAN  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/18/2020 20:46:10

---

Diagnostics Check	OK	20:39
Air Blank	0.000	20:39
Control Test	0.080	20:39
Air Blank	0.000	20:40
Subject Sample # 1	0.207	20:40
Breath Volume	1.761	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.207	20:44
Breath Volume	2.937	20:44
Air Blank	0.000	20:45
Control Test	0.079	20:45
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

---

Data Download Date/Time: 02/17/2020 23:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, MICHAEL, A  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: GARCIA, MICHAEL  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 01/19/2020 15:13:34

---

Diagnostics Check	OK	15:06	
Air Blank	0.000	15:06	
Control Test	0.080	15:06	
Air Blank	0.000	15:07	
Subject Sample	# 1 0.271	15:07	
Breath Volume	1.468	15:07	
Air Blank	0.000	15:08	
Air Blank	0.000	15:10	
Subject Sample	# 2 0.272	15:11	
Breath Volume	1.523	15:11	
Air Blank	0.000	15:12	
Control Test	0.079	15:12	RFI Detect
Air Blank	0.000	15:13	

---

Data Download Date/Time: 02/17/2020 23:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLLINS,CHASE,M  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: COLLINS,CHASE  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: TOBIN

DATE/TIME OF TEST: 01/21/2020 10:37:06

---

Diagnostics Check	OK	10:30
Air Blank	0.000	10:31
Control Test	0.082	10:31
Air Blank	0.000	10:31
Subject Sample # 1	0.164	10:32
Breath Volume	1.937	10:32
Air Blank	0.000	10:33
Air Blank	0.000	10:34
Subject Sample # 2	0.170	10:35
Breath Volume	2.519	10:35
Air Blank	0.000	10:36
Control Test	0.080	10:36
Air Blank	0.000	10:37
Diagnostics Check	OK	10:37

---

Data Download Date/Time: 02/17/2020 23:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, MICHAEL, A  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: ABDELHADI, AUSTIN  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 07:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 01/23/2020 08:16:10

---

Diagnostics Check	OK	08:10
Air Blank	0.000	08:10
Control Test	0.081	08:10
Air Blank	0.000	08:11
Subject Sample # 1	0.000	08:11
Breath Volume	2.449	08:11
Air Blank	0.000	08:12
Air Blank	0.000	08:14
Subject Sample # 2	0.000	08:14
Breath Volume	2.636	08:14
Air Blank	0.000	08:15
Control Test	0.079	08:15
Air Blank	0.000	08:16
Diagnostics Check	OK	08:16

---

Data Download Date/Time: 02/17/2020 23:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLLINS,CHASE,M  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: COLLINS,CHASE  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 10:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: NIESEN

DATE/TIME OF TEST: 01/26/2020 11:26:21

---

Diagnostics Check	OK	11:14	
Air Blank	0.000	11:15	
Control Test	0.081	11:15	
Air Blank	0.000	11:16	
Subject Sample	# 1	11:19	No Sample Provided
Breath Volume	0.000	11:19	
Air Blank	0.000	11:19	
Air Blank	0.000	11:21	
Subject Sample	# 2	11:24	No Sample Provided
Breath Volume	0.000	11:24	
Air Blank	0.000	11:25	
Control Test	0.080	11:25	
Air Blank	0.000	11:26	
Diagnostics Check	OK	11:26	

---

Data Download Date/Time: 02/17/2020 23:46



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NELSON,JAMIE,L  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: NELSON,JAMIE  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: TRUELOCK

DATE/TIME OF TEST: 01/30/2020 19:31:09

---

Diagnostics Check	OK	19:24
Air Blank	0.000	19:25
Control Test	0.080	19:25
Air Blank	0.000	19:26
Subject Sample # 1	0.195	19:26
Breath Volume	2.941	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.197	19:29
Breath Volume	3.277	19:29
Air Blank	0.000	19:30
Control Test	0.079	19:30
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

---

Data Download Date/Time: 02/17/2020 23:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NELSON,JAMIE,L  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: TOMASEGOVICH,KODY  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: OGARA

DATE/TIME OF TEST: 01/31/2020 02:22:43

---

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.155	02:18
Breath Volume	1.554	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.146	02:21
Breath Volume	1.726	02:21
Air Blank	0.000	02:21
Control Test	0.078	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

---

Data Download Date/Time: 02/17/2020 23:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKIPPER,KYLE,A  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: SKIPPER,KYLE  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A1  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 01/31/2020 02:40:28

---

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.079	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.018	02:36
Breath Volume	3.984	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.017	02:38
Breath Volume	4.613	02:38
Air Blank	0.000	02:39
Control Test	0.078	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

---

Data Download Date/Time: 02/17/2020 23:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLT,CATESSA,B  
BREATH TEST OPERATOR AGENCY: Hardee County SO  
ARRESTING OFFICER: WARD,SHANE  
ARRESTING OFFICER AGENCY: Hardee County SO  
LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181632  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/30/2020 22:56:08

---

Diagnostics Check	OK	22:49
Air Blank	0.000	22:50
Control Test	0.079	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.162	22:51
Breath Volume	3.449	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:53
Subject Sample # 2	0.175	22:54
Breath Volume	3.191	22:54
Air Blank	0.000	22:55
Control Test	0.078	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

---

Data Download Date/Time: 02/28/2020 18:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARGUELLO,VICTOR,S  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: ARGUELLO,VICTOR  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: VELAZQUEZ TEJADA

DATE/TIME OF TEST: 01/01/2020 02:31:47

---

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.079	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.158	02:27
Breath Volume	1.332	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.157	02:30
Breath Volume	1.457	02:30
Air Blank	0.000	02:30
Control Test	0.078	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

---

Data Download Date/Time: 01/17/2020 15:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: CHLEBANOWSKI,RYAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: VAILLANCOURT

DATE/TIME OF TEST: 01/01/2020 02:51:24

---

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.078	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.186	02:43
Breath Volume		1.617	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:45
Subject Sample	# 2	0.217	02:46
Breath Volume		1.441	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 3	0.191	02:49
Breath Volume		1.734	02:49
Air Blank		0.000	02:50
Control Test		0.077	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

---

Data Download Date/Time: 01/17/2020 15:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: DONAVON,KATHLEEN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 01/02/2020 19:11:31

---

Diagnostics Check	OK	19:08	
Air Blank	0.000	19:08	
Control Test	0.079	19:09	
Air Blank	0.000	19:09	
Subject Sample	# 1	19:10	Subject Test Refused
Breath Volume	0.000	19:10	
Air Blank	0.000	19:10	
Control Test	0.079	19:10	
Air Blank	0.000	19:11	
Diagnostics Check	OK	19:11	

---

Data Download Date/Time: 01/17/2020 15:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLEN,JOHN,M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: CALDERWOOD,ANDREW  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 10:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:33

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 01/04/2020 11:09:18

---

Diagnostics Check	OK	11:02
Air Blank	0.000	11:03
Control Test	0.079	11:03
Air Blank	0.000	11:04
Subject Sample # 1	0.204	11:04
Breath Volume	3.144	11:04
Air Blank	0.000	11:05
Air Blank	0.000	11:07
Subject Sample # 2	0.192	11:07
Breath Volume	2.351	11:07
Air Blank	0.000	11:08
Control Test	0.077	11:08
Air Blank	0.000	11:09
Diagnostics Check	OK	11:09

---

Data Download Date/Time: 01/17/2020 15:28



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANE,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: LANE,JONATHAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: FREE

DATE/TIME OF TEST: 01/05/2020 01:17:41

---

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.079	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.200	01:13
Breath Volume		1.328	01:13
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.203	01:16
Breath Volume		2.472	01:16
Air Blank		0.000	01:16
Control Test		0.078	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

---

Data Download Date/Time: 01/17/2020 15:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANE,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: WHITE,BRANDON  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: HAZUKA

DATE/TIME OF TEST: 01/05/2020 03:38:33

---

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.079	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.128	03:33
Breath Volume	1.480	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.146	03:36
Breath Volume	1.335	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

---

Data Download Date/Time: 01/17/2020 15:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,ANDREW,J  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: BOZZETTI,AVERY  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HAUTECLOCHE

DATE/TIME OF TEST: 01/06/2020 01:27:25

---

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.126	01:22
Breath Volume	3.125	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.114	01:25
Breath Volume	2.089	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

---

Data Download Date/Time: 01/17/2020 15:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: CHLEBANOWSKI,RYAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: DRAYTON

DATE/TIME OF TEST: 01/12/2020 04:36:11

---

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.080	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.196	04:31
Breath Volume	1.898	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:33
Subject Sample # 2	0.196	04:34
Breath Volume	1.539	04:34
Air Blank	0.000	04:35
Control Test	0.077	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

---

Data Download Date/Time: 01/17/2020 15:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TYLER,BRYAN,D  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: TYLER,BRYAN  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 01/12/2020 05:46:36

---

Diagnostics Check	OK	05:43	
Air Blank	0.000	05:44	
Control Test	0.079	05:44	
Air Blank	0.000	05:45	
Subject Sample	# 1	05:45	Subject Test Refused
Breath Volume	0.000	05:45	
Air Blank	0.000	05:45	
Control Test	0.079	05:45	
Air Blank	0.000	05:46	
Diagnostics Check	OK	05:46	

---

Data Download Date/Time: 01/17/2020 15:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,ANDREW,J  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: ROSCOE,RAYMOND  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: DEMIN

DATE/TIME OF TEST: 01/14/2020 01:07:07

---

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.080	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.183	01:02
Breath Volume		1.976	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.163	01:05
Breath Volume		1.234	01:05
Air Blank		0.000	01:06
Control Test		0.078	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

---

Data Download Date/Time: 01/17/2020 15:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,ANDREW,J  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: DONOVAN,KATHLEEN  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 12/18/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: COUTY

DATE/TIME OF TEST: 01/15/2020 03:54:20

---

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:51	
Control Test	0.080	03:52	
Air Blank	0.000	03:52	
Subject Sample	# 1	03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:53	
Control Test	0.080	03:53	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

---

Data Download Date/Time: 01/17/2020 15:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: CHLEBANOWSKI,RYAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 01/15/2020 21:08:49

---

Diagnostics Check	OK	20:59
Air Blank	0.000	21:00
Control Test	0.079	21:00
Air Blank	0.000	21:01
Subject Sample # 1	0.318	21:04
Breath Volume	1.976	21:04
Air Blank	0.000	21:05
Air Blank	0.000	21:06
Subject Sample # 2	0.329	21:07
Breath Volume	2.246	21:07
Air Blank	0.000	21:07
Control Test	0.075	21:08
Air Blank	0.000	21:08
Diagnostics Check	OK	21:08

---

Data Download Date/Time: 01/17/2020 15:28



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: WAITE,KENNETH  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 01/16/2020 20:32:39

---

Diagnostics Check	OK	20:29	
Air Blank	0.000	20:30	
Control Test	0.080	20:30	
Air Blank	0.000	20:31	
Subject Sample	# 1	20:31	Subject Test Refused
Breath Volume	0.000	20:31	
Air Blank	0.000	20:31	
Control Test	0.080	20:32	
Air Blank	0.000	20:32	
Diagnostics Check	OK	20:32	

---

Data Download Date/Time: 01/17/2020 15:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANE,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: LANE,JONATHAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: BECKER

DATE/TIME OF TEST: 01/19/2020 02:42:43

---

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:39	
Control Test	0.080	02:40	
Air Blank	0.000	02:40	
Subject Sample	# 1	02:41	Subject Test Refused
Breath Volume	0.000	02:41	
Air Blank	0.000	02:41	
Control Test	0.080	02:42	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

---

Data Download Date/Time: 02/03/2020 13:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TYLER,BRYAN,D  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: WAITE,KENNETH  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: WOLOSHIN

DATE/TIME OF TEST: 01/21/2020 04:54:10

---

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:51	
Control Test	0.080	04:52	
Air Blank	0.000	04:52	
Subject Sample	# 1	04:52	Subject Test Refused
Breath Volume	0.000	04:52	
Air Blank	0.000	04:53	
Control Test	0.079	04:53	
Air Blank	0.000	04:54	
Diagnostics Check	OK	04:54	

---

Data Download Date/Time: 02/03/2020 13:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TYLER,BRYAN,D  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: TYLER,BRYAN  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 01/25/2020 04:08:09

---

Diagnostics Check		OK	04:01
Air Blank		0.000	04:02
Control Test		0.079	04:02
Air Blank		0.000	04:03
Subject Sample	# 1	0.250	04:03
Breath Volume		2.449	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.232	04:06
Breath Volume		2.218	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

---

Data Download Date/Time: 02/03/2020 13:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUERRA, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: GUERRA, WILLIAM  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: VILLAVICENCIO'

DATE/TIME OF TEST: 01/26/2020 22:24:53

---

Diagnostics Check	OK	22:24	
Air Blank		22:24	RFI Detect
Air Blank	0.000	22:24	

---

Data Download Date/Time: 02/03/2020 13:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUERRA, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: GUERRA, WILLIAM  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: VILLAVICENCIO

DATE/TIME OF TEST: 01/26/2020 22:29:46

---

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:27	
Control Test	0.080	22:27	
Air Blank	0.000	22:28	
Subject Sample	# 1	22:28	Subject Test Refused
Breath Volume	0.000	22:28	
Air Blank	0.000	22:28	
Control Test	0.080	22:29	
Air Blank	0.000	22:29	
Diagnostics Check	OK	22:29	

---

Data Download Date/Time: 02/03/2020 13:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWOGGER, EDWARD, A  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: SWOGGER, EDWARD  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: MAYORGA

DATE/TIME OF TEST: 01/28/2020 20:39:54

---

Diagnostics Check	OK	20:33
Air Blank	0.000	20:34
Control Test	0.080	20:34
Air Blank	0.000	20:34
Subject Sample # 1	0.232	20:35
Breath Volume	1.785	20:35
Air Blank	0.000	20:36
Air Blank	0.000	20:37
Subject Sample # 2	0.235	20:38
Breath Volume	1.796	20:38
Air Blank	0.000	20:38
Control Test	0.078	20:39
Air Blank	0.000	20:39
Diagnostics Check	OK	20:39

---

Data Download Date/Time: 02/03/2020 13:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONES,SHAUN,G  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: LONES,SHAUN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: CARDENAS BELTRAN

DATE/TIME OF TEST: 01/01/2020 01:11:33

---

Diagnostics Check	OK	01:08	
Air Blank	0.000	01:09	
Control Test	0.079	01:09	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:10	Subject Test Refused
Breath Volume	0.000	01:10	
Air Blank	0.000	01:10	
Control Test	0.079	01:10	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

---

Data Download Date/Time: 01/08/2020 12:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONES,SHAUN,G  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HRADECKY,CORBIN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: PRIDEAUX

DATE/TIME OF TEST: 01/01/2020 02:56:49

---

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.079	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:55	
Control Test	0.078	02:56	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

---

Data Download Date/Time: 01/08/2020 12:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONES,SHAUN,G  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: LONES,SHAUN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:36

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 01/01/2020 18:03:36

---

Diagnostics Check	OK	17:56
Air Blank	0.000	17:56
Control Test	0.080	17:57
Air Blank	0.000	17:57
Subject Sample # 1	0.166	17:59
Breath Volume	2.375	17:59
Air Blank	0.000	17:59
Air Blank	0.000	18:01
Subject Sample # 2	0.168	18:01
Breath Volume	1.644	18:01
Air Blank	0.000	18:02
Control Test	0.079	18:02
Air Blank	0.000	18:03
Diagnostics Check	OK	18:03

---

Data Download Date/Time: 01/08/2020 12:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONES,SHAUN,G  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: LONES,SHAUN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 01/03/2020 17:49:57

---

Diagnostics Check	OK	17:46	
Air Blank	0.000	17:47	
Control Test	0.080	17:47	
Air Blank	0.000	17:48	
Subject Sample	# 1	17:48	Subject Test Refused
Breath Volume	0.000	17:48	
Air Blank	0.000	17:49	
Control Test	0.080	17:49	
Air Blank	0.000	17:49	
Diagnostics Check	OK	17:49	

---

Data Download Date/Time: 01/08/2020 12:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: MARTIN-REYES,JUAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: NOWICKI

DATE/TIME OF TEST: 01/05/2020 18:52:40

---

Diagnostics Check	OK	18:45
Air Blank	0.000	18:46
Control Test	0.080	18:46
Air Blank	0.000	18:47
Subject Sample # 1	0.151	18:47
Breath Volume	1.410	18:47
Air Blank	0.000	18:48
Air Blank	0.000	18:50
Subject Sample # 2	0.160	18:51
Breath Volume	2.046	18:51
Air Blank	0.000	18:51
Control Test	0.079	18:52
Air Blank	0.000	18:52
Diagnostics Check	OK	18:52

---

Data Download Date/Time: 01/08/2020 12:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUINTERO,LISSETTE,  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: SWIRE,BIANDREA  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: FRYER

DATE/TIME OF TEST: 01/06/2020 21:42:15

---

Diagnostics Check	OK	21:39	
Air Blank	0.000	21:39	
Control Test	0.080	21:40	
Air Blank	0.000	21:40	
Subject Sample	# 1	21:40	Subject Test Refused
Breath Volume	0.000	21:40	
Air Blank	0.000	21:41	
Control Test	0.080	21:41	
Air Blank	0.000	21:42	
Diagnostics Check	OK	21:42	

---

Data Download Date/Time: 01/08/2020 12:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUINTERO,LISSETTE,  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: FAISON,MEGAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 12/05/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/07/2020 22:57:49

---

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.138	22:53
Breath Volume	2.683	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:55
Subject Sample # 2	0.140	22:56
Breath Volume	1.835	22:56
Air Blank	0.000	22:56
Control Test	0.078	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

---

Data Download Date/Time: 01/08/2020 12:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HRADECKY,CORBIN,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HRADECKY,CORBIN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: VANERKA

DATE/TIME OF TEST: 01/09/2020 01:38:45

---

Diagnostics Check		OK	01:31
Air Blank		0.000	01:32
Control Test		0.080	01:32
Air Blank		0.000	01:33
Subject Sample	# 1	0.159	01:33
Breath Volume		1.375	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.177	01:37
Breath Volume		1.406	01:37
Air Blank		0.000	01:37
Control Test		0.079	01:38
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

---

Data Download Date/Time: 02/04/2020 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HRADECKY,CORBIN,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: VIDAL,NERELYS  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: YAC XILOJ

DATE/TIME OF TEST: 01/09/2020 21:19:55

---

Diagnostics Check		OK	21:11
Air Blank		0.000	21:12
Control Test		0.079	21:12
Air Blank		0.000	21:13
Subject Sample	# 1	0.217	21:13
Breath Volume		1.183	21:13
Air Blank		0.000	21:14
Air Blank		0.000	21:16
Subject Sample	# 2	0.233	21:18
Breath Volume		1.574	21:18
Air Blank		0.000	21:18
Control Test		0.080	21:19
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

---

Data Download Date/Time: 02/04/2020 12:19



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUINTERO,LISSETTE,  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HARRILL,ANDERSEN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: HRADECKY

DATE/TIME OF TEST: 01/12/2020 04:25:08

---

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:08	
Control Test	0.079	04:09	
Air Blank	0.000	04:09	
Subject Sample	# 1 0.200	04:12	
Breath Volume	1.273	04:12	
Air Blank	0.000	04:13	
Air Blank	0.000	04:15	
Subject Sample	# 2 0.218	04:18	Volume Not Met
Breath Volume	0.144	04:18	
Air Blank	0.000	04:19	
Air Blank	0.000	04:20	
Subject Sample	# 3 0.181	04:23	Volume Not Met
Breath Volume	0.171	04:23	
Air Blank		04:24	Purge Fail
Air Blank	0.000	04:25	

---

Data Download Date/Time: 02/04/2020 12:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUINTERO,LISSETTE,  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: SZYMANSKI,AMIE  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BERG

DATE/TIME OF TEST: 01/13/2020 01:13:24

---

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:10	
Control Test	0.079	01:11	
Air Blank	0.000	01:11	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:12	
Control Test	0.080	01:12	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

---

Data Download Date/Time: 02/04/2020 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JOEL,M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: TORRES,JOEL  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: CERDA

DATE/TIME OF TEST: 01/13/2020 22:50:13

---

Diagnostics Check	OK	22:43
Air Blank	0.000	22:44
Control Test	0.081	22:44
Air Blank	0.000	22:45
Subject Sample # 1	0.071	22:45
Breath Volume	2.023	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.068	22:48
Breath Volume	1.816	22:48
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

---

Data Download Date/Time: 02/04/2020 12:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANE,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: LANE,JONATHAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: BUI  
ARREST TIME: 01:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: HUMPHREY

DATE/TIME OF TEST: 01/14/2020 02:45:15

---

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:42	
Control Test	0.080	02:43	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

---

Data Download Date/Time: 02/04/2020 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUINTERO,LISSETTE,  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HARRILL,ANDERSEN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: RUBIO AMAYA

DATE/TIME OF TEST: 01/16/2020 02:01:45

---

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:59	
Control Test	0.080	01:59	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:00	
Control Test	0.080	02:01	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

---

Data Download Date/Time: 02/04/2020 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONES,SHAUN,G  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: 7C06,SHAUN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 01/17/2020 15:04:20

---

Diagnostics Check	OK	15:00	
Air Blank	0.000	15:01	
Control Test	0.080	15:01	
Air Blank	0.000	15:02	
Subject Sample	# 1	15:02	Subject Test Refused
Breath Volume	0.000	15:02	
Air Blank	0.000	15:03	
Control Test	0.079	15:03	
Air Blank	0.000	15:04	
Diagnostics Check	OK	15:04	

---

Data Download Date/Time: 02/04/2020 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HRADECKY,CORBIN,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HRADECKY,CORBIN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: MILOSZEWSKI

DATE/TIME OF TEST: 01/18/2020 01:36:42

---

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.081	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.118	01:32
Breath Volume	3.542	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.119	01:35
Breath Volume	3.371	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

---

Data Download Date/Time: 02/04/2020 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: VIDAL,NERELYS  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: VIZOSO

DATE/TIME OF TEST: 01/19/2020 00:08:33

---

Diagnostics Check	OK	00:01
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:02
Subject Sample # 1	0.093	00:04
Breath Volume	1.769	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.091	00:06
Breath Volume	1.390	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

---

Data Download Date/Time: 02/04/2020 12:19



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUINTERO,LISSETTE,  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HARRILL,ANDERSEN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 01/21/2020 23:47:18

---

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:44	
Control Test	0.081	23:45	
Air Blank	0.000	23:45	
Subject Sample	# 1	23:45	Subject Test Refused
Breath Volume	0.000	23:45	
Air Blank	0.000	23:46	
Control Test	0.080	23:46	
Air Blank	0.000	23:47	
Diagnostics Check	OK	23:47	

---

Data Download Date/Time: 02/04/2020 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HRADECKY,CORBIN,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HRADECKY,CORBIN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 01/23/2020 23:27:53

---

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.190	23:22
Breath Volume	1.976	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.187	23:26
Breath Volume	2.933	23:26
Air Blank	0.000	23:26
Control Test	0.079	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

---

Data Download Date/Time: 02/04/2020 12:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWOGGER, EDWARD, A  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: SWOGGER, EDWARD  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 183142  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: CROSBY

DATE/TIME OF TEST: 01/24/2020 20:16:52

---

Diagnostics Check	OK	20:13	
Air Blank	0.000	20:14	
Control Test	0.080	20:14	
Air Blank	0.000	20:15	
Subject Sample	# 1	20:15	Subject Test Refused
Breath Volume	0.000	20:15	
Air Blank	0.000	20:15	
Control Test	0.079	20:16	
Air Blank	0.000	20:16	
Diagnostics Check	OK	20:16	

---

Data Download Date/Time: 02/04/2020 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUINTERO,LISSETTE,  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: SZYMANSKI,AMIE  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 01/26/2020 02:36:05

---

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.080	02:33	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:35	
Control Test	0.079	02:35	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

---

Data Download Date/Time: 02/04/2020 12:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUINTERO,LISSETTE,  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: SZYMANSKI,AMIE  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: CAPPS

DATE/TIME OF TEST: 01/27/2020 02:26:32

---

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
Control Test	0.080	02:24	
Air Blank	0.000	02:24	
Subject Sample	# 1	02:25	Subject Test Refused
Breath Volume	0.000	02:25	
Air Blank	0.000	02:25	
Control Test	0.079	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

---

Data Download Date/Time: 02/04/2020 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HRADECKY,CORBIN,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HRADECKY,CORBIN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BETANCOURT

DATE/TIME OF TEST: 01/29/2020 03:28:19

---

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.080	03:26	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:27	
Control Test	0.080	03:27	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

---

Data Download Date/Time: 02/04/2020 12:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUERRA, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: SZYMANSKY, ANN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: BEACH

DATE/TIME OF TEST: 01/30/2020 03:11:12

---

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:08	
Control Test	0.080	03:09	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:10	
Control Test	0.080	03:10	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

---

Data Download Date/Time: 02/04/2020 12:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAYS, MATTHEW, W  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: GOODNOW, TIMOTHY  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 875088  
DRY GAS EXPIRY DATE: 12/12/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: GUERRA

DATE/TIME OF TEST: 01/18/2020 06:55:05

---

Diagnostics Check	OK	06:48
Air Blank	0.000	06:49
Control Test	0.078	06:49
Air Blank	0.000	06:49
Subject Sample # 1	0.000	06:50
Breath Volume	2.460	06:50
Air Blank	0.000	06:51
Air Blank	0.000	06:53
Subject Sample # 2	0.000	06:53
Breath Volume	1.535	06:53
Air Blank	0.000	06:54
Control Test	0.079	06:54
Air Blank	0.000	06:55
Diagnostics Check	OK	06:55

---

Data Download Date/Time: 01/28/2020 15:19



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GURRIERI,JOHN,V  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: GURRIERI,JOHN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 875088  
DRY GAS EXPIRY DATE: 12/12/2020  
VIOLATION CODE: DUI  
ARREST TIME: 12:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: CHIASSON

DATE/TIME OF TEST: 01/24/2020 02:32:59

---

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.077	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.316	02:28
Breath Volume	3.472	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:30
Subject Sample # 2	0.318	02:31
Breath Volume	3.296	02:31
Air Blank	0.000	02:32
Control Test	0.077	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

---

Data Download Date/Time: 01/28/2020 15:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULFORD,G,  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: PULFORD,G  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 875088  
DRY GAS EXPIRY DATE: 12/12/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: GOOLSBY

DATE/TIME OF TEST: 01/26/2020 01:51:55

---

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:40	
Control Test	0.079	01:40	
Air Blank	0.000	01:41	
Subject Sample	# 1 0.191	01:44	Volume Not Met
Breath Volume	1.046	01:44	
Air Blank	0.000	01:45	
Air Blank	0.000	01:46	
Subject Sample	# 2 0.222	01:47	
Breath Volume	2.011	01:47	
Air Blank	0.000	01:48	
Air Blank	0.000	01:49	
Subject Sample	# 3 0.209	01:50	
Breath Volume	1.453	01:50	
Air Blank	0.000	01:50	
Control Test	0.075	01:51	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

---

Data Download Date/Time: 01/28/2020 15:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWDER,MICHAEL,P  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: LOWDER,MICHAEL  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 875088  
DRY GAS EXPIRY DATE: 12/12/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: CLIFFORD

DATE/TIME OF TEST: 01/28/2020 19:54:01

---

Diagnostics Check	OK	19:47
Air Blank	0.000	19:48
Control Test	0.079	19:48
Air Blank	0.000	19:49
Subject Sample # 1	0.000	19:49
Breath Volume	3.816	19:49
Air Blank	0.000	19:50
Air Blank	0.000	19:52
Subject Sample # 2	0.000	19:52
Breath Volume	4.664	19:52
Air Blank	0.000	19:53
Control Test	0.077	19:53
Air Blank	0.000	19:53
Diagnostics Check	OK	19:54

---

Data Download Date/Time: 02/25/2020 14:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,ANTHONY,L  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: ESQUIVEL,ANTHONY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 875088  
DRY GAS EXPIRY DATE: 12/12/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 01/30/2020 03:45:15

---

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.179	03:37
Breath Volume	1.816	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.156	03:40
Breath Volume	1.667	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 3	0.180	03:43
Breath Volume	2.468	03:43
Air Blank	0.000	03:44
Control Test	0.078	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

---

Data Download Date/Time: 02/25/2020 14:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GURRIERI,JOHN,V  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: GURRIERI,JOHN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 875088  
DRY GAS EXPIRY DATE: 12/12/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: INSPRUCKER

DATE/TIME OF TEST: 01/31/2020 19:47:53

---

Diagnostics Check	OK	19:44	
Air Blank	0.000	19:45	
Control Test	0.079	19:45	
Air Blank	0.000	19:46	
Subject Sample	# 1	19:46	Subject Test Refused
Breath Volume	0.000	19:46	
Air Blank	0.000	19:46	
Control Test	0.079	19:47	
Air Blank	0.000	19:47	
Diagnostics Check	OK	19:47	

---

Data Download Date/Time: 02/25/2020 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 01/04/2020 23:23:22

---

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.079	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.145	23:18
Breath Volume	2.734	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:20
Subject Sample # 2	0.151	23:21
Breath Volume	2.789	23:21
Air Blank	0.000	23:22
Control Test	0.079	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

---

Data Download Date/Time: 01/28/2020 15:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: CHARTIER,K  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: GORSKI

DATE/TIME OF TEST: 01/05/2020 02:14:06

---

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.080	02:11	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:13	
Control Test	0.080	02:13	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:14	

---

Data Download Date/Time: 01/28/2020 15:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORCHARD,DUSTIN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: ORCHARD,DUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1031121  
DRY GAS EXPIRY DATE: 10/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 01/07/2020 20:39:06

---

Diagnostics Check	OK	20:30
Air Blank	0.000	20:31
Control Test	0.080	20:31
Air Blank	0.000	20:32
Subject Sample # 1	0.195	20:33
Breath Volume	1.757	20:33
Air Blank	0.000	20:34
Air Blank	0.000	20:36
Subject Sample # 2	0.184	20:37
Breath Volume	1.269	20:37
Air Blank	0.000	20:38
Control Test	0.076	20:38
Air Blank	0.000	20:39
Diagnostics Check	OK	20:39

---

Data Download Date/Time: 01/28/2020 15:09



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CALVET,CHRISTINA,M  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: CALVET,CHRISTINA  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1031121  
DRY GAS EXPIRY DATE: 10/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: LAPAN

DATE/TIME OF TEST: 01/11/2020 03:52:27

---

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.144	03:47
Breath Volume	3.187	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.136	03:50
Breath Volume	3.890	03:50
Air Blank	0.000	03:51
Control Test	0.077	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

---

Data Download Date/Time: 01/28/2020 15:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KARR,LUCAS,B  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: KARR,LUCAS  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1031121  
DRY GAS EXPIRY DATE: 10/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: WOODHOUSE

DATE/TIME OF TEST: 01/12/2020 03:45:23

---

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:34	
Control Test	0.079	03:34	
Air Blank	0.000	03:35	
Subject Sample	# 1	03:38	No Sample Provided
Breath Volume	0.000	03:38	
Air Blank	0.000	03:38	
Air Blank	0.000	03:40	
Subject Sample	# 2	03:44	No Sample Provided
Breath Volume	0.000	03:44	
Air Blank	0.000	03:44	
Control Test	0.080	03:44	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

---

Data Download Date/Time: 01/28/2020 15:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEST,R,  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: WEST,R  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1031121  
DRY GAS EXPIRY DATE: 10/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GRAVATT

DATE/TIME OF TEST: 01/12/2020 04:50:50

---

Diagnostics Check	OK	04:43
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.187	04:45
Breath Volume	2.136	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.184	04:49
Breath Volume	2.160	04:49
Air Blank	0.000	04:49
Control Test	0.077	04:50
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

---

Data Download Date/Time: 01/28/2020 15:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTIGLIONE,SALVATORE,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: CASTIGLIONE,SALVATORE  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1031121  
DRY GAS EXPIRY DATE: 10/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: TREASTER

DATE/TIME OF TEST: 01/13/2020 00:32:43

---

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:30	
Control Test	0.080	00:30	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:31	
Control Test	0.079	00:32	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

---

Data Download Date/Time: 01/28/2020 15:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRINGTON,ALEX,M  
BREATH TEST OPERATOR AGENCY: Belleview PD  
ARRESTING OFFICER: HARRINGTON,A  
ARRESTING OFFICER AGENCY: Belleview PD  
LAST AGENCY INSPECTION: 12/12/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1031121  
DRY GAS EXPIRY DATE: 10/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ANGLE

DATE/TIME OF TEST: 01/13/2020 00:51:26

---

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.153	00:46
Breath Volume	2.449	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.154	00:49
Breath Volume	3.101	00:49
Air Blank	0.000	00:50
Control Test	0.079	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

---

Data Download Date/Time: 01/28/2020 15:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAREY,ANDREW,S  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: VENTRY,WILLIS  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: GAITHER

DATE/TIME OF TEST: 01/01/2020 00:49:44

---

Diagnostics Check		OK	00:37	
Air Blank		0.000	00:37	
Control Test		0.078	00:37	
Air Blank		0.000	00:38	
Subject Sample	# 1		00:41	No Sample Provided
Breath Volume		0.000	00:41	
Air Blank		0.000	00:41	
Air Blank		0.000	00:44	
Subject Sample	# 2	0.196	00:44	No .020 agreement
Breath Volume		2.468	00:44	
Air Blank		0.000	00:45	
Air Blank		0.000	00:47	
Subject Sample	# 3	0.172	00:48	No .020 agreement
Breath Volume		1.503	00:48	
Air Blank		0.000	00:48	
Control Test		0.077	00:49	
Air Blank		0.000	00:49	
Diagnostics Check		OK	00:49	

---

Data Download Date/Time: 01/06/2020 23:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAREY,ANDREW,S  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: VENTRY,WILLIS  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: GAITHER

DATE/TIME OF TEST: 01/01/2020 01:39:33

---

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.157	01:34
Breath Volume	1.726	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.156	01:37
Breath Volume	1.324	01:37
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

---

Data Download Date/Time: 01/06/2020 23:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAREY,ANDREW,S  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: SANDERS,DAVID  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: HAAS

DATE/TIME OF TEST: 01/01/2020 02:26:00

---

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.000	02:21
Breath Volume	1.648	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.000	02:24
Breath Volume	1.679	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:26

---

Data Download Date/Time: 01/06/2020 23:39



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: SCHULTZ,JEREMY  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: MORTON

DATE/TIME OF TEST: 01/04/2020 01:44:40

---

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.079	01:39	
Air Blank	0.000	01:39	
Subject Sample	# 1 0.179	01:43	Volume Not Met
Breath Volume	0.484	01:43	
Air Blank		01:43	Purge Fail
Air Blank	0.000	01:44	

---

Data Download Date/Time: 01/06/2020 23:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: SCHULTZ,JEREMY  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: MORTON

DATE/TIME OF TEST: 01/04/2020 01:47:36

---

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:46	
Control Test	0.073	01:46	Control Outside Tolerance
Air Blank	0.000	01:47	

---

Data Download Date/Time: 01/06/2020 23:39