

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISNEY,WHITNEY,L
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: LUJAN,ARTURO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: RODRIGUEZ-ANTON

DATE/TIME OF TEST: 01/06/2020 23:05:55

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:03	
Control Test	0.082	23:03	
Air Blank	0.000	23:04	
Subject Sample	# 1	23:04	Subject Test Refused
Breath Volume	0.000	23:04	
Air Blank	0.000	23:05	
Control Test	0.081	23:05	
Air Blank	0.000	23:05	
Diagnostics Check	OK	23:05	

Data Download Date/Time: 01/28/2020 12:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,JUSTIN,J
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: LUJAN,ARTURO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: ROBLETO

DATE/TIME OF TEST: 01/07/2020 22:23:28

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:21	
Control Test	0.082	22:21	
Air Blank	0.000	22:21	
Subject Sample	# 1	22:22	Subject Test Refused
Breath Volume	0.000	22:22	
Air Blank	0.000	22:22	
Control Test	0.082	22:22	
Air Blank	0.000	22:23	
Diagnostics Check	OK	22:23	

Data Download Date/Time: 01/28/2020 12:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,RYAN,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: VAZQUEZ,WILLIAM
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: MEDRANO

DATE/TIME OF TEST: 01/11/2020 05:44:21

Diagnostics Check	OK	05:41	
Air Blank	0.000	05:41	
Control Test	0.082	05:42	
Air Blank	0.000	05:42	
Subject Sample	# 1	05:42	Subject Test Refused
Breath Volume	0.000	05:42	
Air Blank	0.000	05:43	
Control Test	0.082	05:43	
Air Blank	0.000	05:44	
Diagnostics Check	OK	05:44	

Data Download Date/Time: 01/28/2020 12:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,RYAN,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: SISNEY,WHITNEY
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 01/14/2020 01:44:45

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:42	
Control Test	0.082	01:42	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:43	
Control Test	0.082	01:44	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 01/28/2020 12:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,ANDREW,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: HERNANDEZ,PETER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: MARLOWE

DATE/TIME OF TEST: 01/16/2020 00:32:39

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.082	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.243	00:28
Breath Volume	3.324	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.234	00:31
Breath Volume	2.039	00:31
Air Blank	0.000	00:31
Control Test	0.081	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 01/28/2020 12:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,RYAN,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: SMITH,RYAN
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: FELL

DATE/TIME OF TEST: 01/25/2020 01:16:30

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:13	
Control Test	0.082	01:14	
Air Blank	0.000	01:14	
Subject Sample	# 1	01:15	Subject Test Refused
Breath Volume	0.000	01:15	
Air Blank	0.000	01:15	
Control Test	0.082	01:15	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 01/28/2020 12:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,CRISTINO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: PEREZ,CRISTINO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 01/28/2020 19:22:35

Diagnostics Check	OK	19:19	
Air Blank	0.000	19:20	
Control Test	0.081	19:20	
Air Blank	0.000	19:20	
Subject Sample	# 1	19:21	Subject Test Refused
Breath Volume	0.000	19:21	
Air Blank	0.000	19:21	
Control Test	0.081	19:22	
Air Blank	0.000	19:22	
Diagnostics Check	OK	19:22	

Data Download Date/Time: 02/27/2020 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILES,STEPHANIE,N
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SILES,STEPHANIE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: CARDOZA

DATE/TIME OF TEST: 01/12/2020 04:44:37

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
Control Test	0.079	04:40	
Air Blank	0.000	04:40	
Subject Sample	# 1	04:43	Subject Test Refused
Breath Volume	0.035	04:43	
Air Blank	0.000	04:43	
Control Test	0.079	04:44	
Air Blank	0.000	04:44	
Diagnostics Check	OK	04:44	

Data Download Date/Time: 01/21/2020 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILES,STEPHANIE,N
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SILES,STEPHANIE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: CARDOZA

DATE/TIME OF TEST: 01/12/2020 05:24:41

Diagnostics Check	OK	05:15
Air Blank	0.000	05:15
Control Test	0.079	05:16
Air Blank	0.000	05:16
Subject Sample # 1	0.140	05:19
Breath Volume	1.308	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.146	05:23
Breath Volume	1.238	05:23
Air Blank	0.000	05:23
Control Test	0.079	05:24
Air Blank	0.000	05:24
Diagnostics Check	OK	05:24

Data Download Date/Time: 01/21/2020 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADDY,STEVEN,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BRADDY,STEVEN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 08:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: SAUCEDO

DATE/TIME OF TEST: 01/19/2020 09:02:40

Diagnostics Check	OK	09:01	
Air Blank	0.000	09:01	
Control Test	0.079	09:01	
Air Blank		09:02	RFI Detect
Air Blank		09:02	RFI Detect

Data Download Date/Time: 01/21/2020 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADDY,STEVEN,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BRADDY,STEVEN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 08:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: SAUCEDO

DATE/TIME OF TEST: 01/19/2020 09:06:55

Diagnostics Check	OK	09:06	
Air Blank		09:06	RFI Detect
Air Blank		09:06	RFI Detect

Data Download Date/Time: 01/21/2020 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADDY,STEVEN,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BRADDY,STEVEN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 08:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: SAUCEDO

DATE/TIME OF TEST: 01/19/2020 09:38:21

Diagnostics Check	OK	09:35	
Air Blank	0.000	09:35	
Control Test	0.079	09:36	
Air Blank	0.000	09:36	
Subject Sample	# 1	09:36	Subject Test Refused
Breath Volume	0.000	09:36	
Air Blank	0.000	09:37	
Control Test	0.079	09:37	
Air Blank	0.000	09:38	
Diagnostics Check	OK	09:38	

Data Download Date/Time: 01/21/2020 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,KEVIN,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,KEVIN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/23/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 01/19/2020 22:04:45

Diagnostics Check		OK	21:52	
Air Blank		0.000	21:53	
Control Test		0.078	21:53	
Air Blank		0.000	21:54	
Subject Sample	# 1	0.141	21:57	Volume Not Met
Breath Volume		0.820	21:57	
Air Blank		0.000	21:58	
Air Blank		0.000	21:59	
Subject Sample	# 2	0.167	22:00	
Breath Volume		1.921	22:00	
Air Blank		0.000	22:01	
Air Blank		0.000	22:02	
Subject Sample	# 3	0.161	22:03	
Breath Volume		1.820	22:03	
Air Blank		0.000	22:03	
Control Test		0.076	22:04	
Air Blank		0.000	22:04	
Diagnostics Check		OK	22:04	

Data Download Date/Time: 01/21/2020 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: MENDEZ SALES

DATE/TIME OF TEST: 01/07/2020 04:00:47

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:58	
Control Test	0.083	03:58	
Air Blank	0.000	03:59	
Subject Sample	# 1	03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	03:59	
Control Test	0.082	04:00	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 01/29/2020 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: HOLLINGSWORTH

DATE/TIME OF TEST: 01/11/2020 20:10:12

Diagnostics Check	OK	20:03
Air Blank	0.000	20:03
Control Test	0.083	20:04
Air Blank	0.000	20:04
Subject Sample # 1	0.129	20:05
Breath Volume	1.441	20:05
Air Blank	0.000	20:06
Air Blank	0.000	20:08
Subject Sample # 2	0.135	20:08
Breath Volume	2.156	20:08
Air Blank	0.000	20:09
Control Test	0.082	20:09
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 01/29/2020 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 01/14/2020 23:59:58

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.082	23:54
Air Blank	0.000	23:54
Subject Sample # 1	0.000	23:55
Breath Volume	2.937	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.000	23:58
Breath Volume	2.429	23:58
Air Blank	0.000	23:59
Control Test	0.082	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 01/29/2020 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILPIN,SCHELISA,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GILPIN,SCHELISA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/18/2020 23:20:58

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.082	23:12	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:15	No Sample Provided
Breath Volume	0.000	23:15	
Air Blank	0.000	23:16	
Air Blank	0.000	23:18	
Subject Sample	# 2	23:19	Subject Test Refused
Breath Volume	0.000	23:19	
Air Blank	0.000	23:20	
Control Test	0.082	23:20	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

Data Download Date/Time: 01/29/2020 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,A
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GILPIN,LISA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: SANCHEZ MENJIEVA

DATE/TIME OF TEST: 01/25/2020 04:38:17

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.082	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.169	04:33
Breath Volume	1.593	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.162	04:36
Breath Volume	2.945	04:36
Air Blank	0.000	04:37
Control Test	0.082	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 01/29/2020 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: VINA

DATE/TIME OF TEST: 01/25/2020 17:56:31

Diagnostics Check	OK	17:43	
Air Blank	0.000	17:43	
Control Test	0.082	17:44	
Air Blank	0.000	17:44	
Subject Sample	# 1	17:47	No Sample Provided
Breath Volume	0.000	17:47	
Air Blank	0.000	17:48	
Air Blank	0.000	17:50	
Subject Sample	# 2	0.194	17:50
Breath Volume	1.355	17:50	
Air Blank	0.000	17:51	
Air Blank	0.000	17:53	
Subject Sample	# 3	0.184	17:54
Breath Volume	1.421	17:54	
Air Blank	0.000	17:55	
Control Test	0.082	17:55	
Air Blank	0.000	17:56	
Diagnostics Check	OK	17:56	

Data Download Date/Time: 01/29/2020 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILPIN,SCHELISA,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: MAYS,D
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 01/26/2020 03:40:42

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.083	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:35	No Sample Provided
Breath Volume	0.000	03:35	
Air Blank	0.000	03:36	
Air Blank	0.000	03:38	
Subject Sample	# 2	03:39	Subject Test Refused
Breath Volume	0.000	03:39	
Air Blank	0.000	03:39	
Control Test	0.082	03:40	
Air Blank	0.000	03:40	
Diagnostics Check	OK	03:40	

Data Download Date/Time: 01/29/2020 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILPIN,SCHELISA,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GILPIN,SCHELISA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: SENTENO

DATE/TIME OF TEST: 01/26/2020 04:50:50

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:42	
Control Test	0.082	04:43	
Air Blank	0.000	04:43	
Subject Sample	# 1	04:46	No Sample Provided
Breath Volume	0.000	04:46	
Air Blank	0.000	04:47	
Air Blank	0.000	04:49	
Subject Sample	# 2	04:49	Subject Test Refused
Breath Volume	0.000	04:49	
Air Blank	0.000	04:49	
Control Test	0.083	04:50	
Air Blank	0.000	04:50	
Diagnostics Check	OK	04:50	

Data Download Date/Time: 01/29/2020 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILPIN,SCHELISA,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GILPIN,SCHELISA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: LOMELI

DATE/TIME OF TEST: 01/27/2020 01:50:40

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:42	
Control Test	0.082	01:42	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:46	No Sample Provided
Breath Volume	0.000	01:46	
Air Blank	0.000	01:47	
Air Blank	0.000	01:49	
Subject Sample	# 2	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:49	
Control Test	0.083	01:50	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 01/29/2020 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,FREDERICK,A
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: RANSOM,CHAKA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 07:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 01/04/2020 08:32:28

Diagnostics Check	OK	08:29	
Air Blank	0.000	08:30	
Control Test	0.082	08:30	
Air Blank	0.000	08:30	
Subject Sample	# 1	08:31	Subject Test Refused
Breath Volume	0.000	08:31	
Air Blank	0.000	08:31	
Control Test	0.082	08:31	
Air Blank	0.000	08:32	
Diagnostics Check	OK	08:32	

Data Download Date/Time: 01/24/2020 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: MELENDEZ,DANIEL
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 01/06/2020 04:31:55

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.083	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.213	04:26
Breath Volume	1.859	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.214	04:30
Breath Volume	3.031	04:30
Air Blank	0.000	04:31
Control Test	0.082	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 01/24/2020 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: MELENDEZ,DANIEL
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: SANZ

DATE/TIME OF TEST: 01/11/2020 02:11:12

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.082	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.204	02:06
Breath Volume		2.562	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:08
Subject Sample	# 2	0.216	02:09
Breath Volume		2.367	02:09
Air Blank		0.000	02:10
Control Test		0.081	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 01/24/2020 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN-
ROSQUETE,STEPHANNE,J
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: BROWN-
ROSQUETE,STEPHANNE
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020

VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male

OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: NAVARRETE

DATE/TIME OF TEST: 01/11/2020 04:00:18

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.082	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.104	03:55
Breath Volume	4.308	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.104	03:58
Breath Volume	3.429	03:58
Air Blank	0.000	03:59
Control Test	0.081	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 01/24/2020 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIND,ABEL,D
BREATH TEST OPERATOR AGENCY: NORTH BAY VILLAGE PD
ARRESTING OFFICER: BRIND,ABEL
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 12/26/2019
Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: BROWNLEE
DATE/TIME OF TEST: 01/19/2020 05:41:07

Diagnostics Check	OK	05:29	
Air Blank	0.000	05:30	
Control Test	0.082	05:30	
Air Blank	0.000	05:31	
Subject Sample	# 1	05:34	No Sample Provided
Breath Volume	0.000	05:34	
Air Blank	0.000	05:34	
Air Blank	0.000	05:36	
Subject Sample	# 2	05:39	No Sample Provided
Breath Volume	0.000	05:39	
Air Blank	0.000	05:40	
Control Test	0.082	05:40	
Air Blank	0.000	05:41	
Diagnostics Check	OK	05:41	

Data Download Date/Time: 01/24/2020 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1175968
BREATH TEST OPERATOR: BRIND,ABEL,D	DRY GAS EXPIRY DATE: 07/30/2022
BREATH TEST OPERATOR AGENCY: NORTH BAY VILLAGE PD	VIOLATION CODE: DUI
ARRESTING OFFICER: BRIND,ABEL	ARREST TIME: 04:07
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD	GENDER: Male
LAST AGENCY INSPECTION: 12/26/2019	OBSERVATION PERIOD BEGAN: 04:45
SUBJECT'S LAST NAME: BROWNLEE	DATE/TIME OF TEST: 01/19/2020 05:59:31

Diagnostics Check	OK	05:56	
Air Blank	0.000	05:57	
Control Test	0.081	05:57	
Air Blank	0.000	05:58	
Subject Sample	# 1	05:58	Subject Test Refused
Breath Volume	0.000	05:58	
Air Blank	0.000	05:58	
Control Test	0.082	05:58	
Air Blank	0.000	05:59	
Diagnostics Check	OK	05:59	

Data Download Date/Time: 01/24/2020 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ,DANIEL,'
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: MELENDEZ,DANIEL
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 06:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:17

SUBJECT'S LAST NAME: VICTORIA

DATE/TIME OF TEST: 01/19/2020 06:44:22

Diagnostics Check	OK	06:38
Air Blank	0.000	06:38
Control Test	0.082	06:38
Air Blank	0.000	06:39
Subject Sample # 1	0.189	06:40
Breath Volume	2.464	06:40
Air Blank	0.000	06:40
Air Blank	0.000	06:42
Subject Sample # 2	0.185	06:42
Breath Volume	1.839	06:42
Air Blank	0.000	06:43
Control Test	0.079	06:43
Air Blank	0.000	06:44
Diagnostics Check	OK	06:44

Data Download Date/Time: 01/24/2020 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAZONOV,VLADISLAV,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: CAMPOS,BRANDON
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:42

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 01/19/2020 15:10:06

Diagnostics Check	OK	15:07	
Air Blank	0.000	15:07	
Control Test	0.080	15:08	
Air Blank	0.000	15:08	
Subject Sample	# 1	15:08	Subject Test Refused
Breath Volume	0.000	15:08	
Air Blank	0.000	15:09	
Control Test	0.081	15:09	
Air Blank	0.000	15:10	
Diagnostics Check	OK	15:10	

Data Download Date/Time: 01/24/2020 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUMPHRY,KERRY,R
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: CALLAN,TONY
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: BONIFACIO SANCHEZ

DATE/TIME OF TEST: 01/21/2020 04:44:08

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.082	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.184	04:39
Breath Volume	3.750	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.186	04:42
Breath Volume	3.210	04:42
Air Blank	0.000	04:43
Control Test	0.081	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 01/24/2020 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUMPHRY,KERRY,R
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: HUMPHRY,KERRY
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 01/25/2020 03:47:35

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:45	
Control Test	0.081	03:45	
Air Blank	0.000	03:46	
Subject Sample	# 1	03:46	Subject Test Refused
Breath Volume	0.000	03:46	
Air Blank	0.000	03:46	
Control Test	0.081	03:47	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 02/27/2020 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONFESSORE,STEVE,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: HACKWORTH,EARL
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SADILIANI

DATE/TIME OF TEST: 01/25/2020 22:28:40

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:26	
Control Test	0.082	22:26	
Air Blank	0.000	22:27	
Subject Sample	# 1	22:27	Subject Test Refused
Breath Volume	0.000	22:27	
Air Blank	0.000	22:27	
Control Test	0.081	22:28	
Air Blank	0.000	22:28	
Diagnostics Check	OK	22:28	

Data Download Date/Time: 02/27/2020 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONFESSORE,STEVE,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: BLACIO,GUSTAVO
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 01/27/2020 22:15:57

Diagnostics Check	OK	22:12	
Air Blank	0.000	22:13	
Control Test	0.082	22:13	
Air Blank	0.000	22:14	
Subject Sample	# 1	22:14	Subject Test Refused
Breath Volume	0.000	22:14	
Air Blank	0.000	22:15	
Control Test	0.082	22:15	
Air Blank	0.000	22:15	
Diagnostics Check	OK	22:15	

Data Download Date/Time: 02/27/2020 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,PEDRO,A
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: GONZALEZ,PABLO
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: GUYTON

DATE/TIME OF TEST: 01/28/2020 17:41:28

Diagnostics Check	OK	17:33
Air Blank	0.000	17:34
Control Test	0.082	17:34
Air Blank	0.000	17:34
Subject Sample # 1	0.180	17:36
Breath Volume	1.250	17:36
Air Blank	0.000	17:36
Air Blank	0.000	17:38
Subject Sample # 2	0.176	17:39
Breath Volume	1.390	17:39
Air Blank	0.000	17:40
Control Test	0.082	17:40
Air Blank	0.000	17:41
Diagnostics Check	OK	17:41

Data Download Date/Time: 02/27/2020 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,PEDRO,A
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: SOCARRAS,PEDRO
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 13:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 01/29/2020 14:43:50

Diagnostics Check	OK	14:41	
Air Blank	0.000	14:41	
Control Test	0.081	14:41	
Air Blank	0.000	14:42	
Subject Sample	# 1	14:42	Subject Test Refused
Breath Volume	0.000	14:42	
Air Blank	0.000	14:42	
Control Test	0.082	14:43	
Air Blank	0.000	14:43	
Diagnostics Check	OK	14:43	

Data Download Date/Time: 02/27/2020 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON,REGINALD,B
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: HORTON,REGINALD
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 01/31/2020 00:23:37

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.179	00:19
Breath Volume	1.988	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.190	00:22
Breath Volume	1.921	00:22
Air Blank	0.000	00:22
Control Test	0.082	00:23
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 02/27/2020 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CELLA,DOMENICO,G
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RONDON,LEORDANIS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: JEAN-LOUIS

DATE/TIME OF TEST: 01/06/2020 00:14:25

Diagnostics Check	OK	00:13	
Air Blank		00:13	RFI Detect
Air Blank	0.000	00:14	

Data Download Date/Time: 01/24/2020 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CELLA,DOMENICO,G
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RONDON,LEORDANIS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: JEAN-LOUIS

DATE/TIME OF TEST: 01/06/2020 00:21:41

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:19	
Control Test	0.079	00:19	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:20	Subject Test Refused
Breath Volume	0.000	00:20	
Air Blank	0.000	00:20	
Control Test	0.079	00:21	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 01/24/2020 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTIS,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: RIVAS,E
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: LUBIN

DATE/TIME OF TEST: 01/11/2020 04:04:03

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.133	03:58
Breath Volume	2.464	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:00
Subject Sample # 2	0.125	04:02
Breath Volume	1.707	04:02
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:04

Data Download Date/Time: 01/24/2020 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTIS,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: RIVAS,E
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: WOSAR

DATE/TIME OF TEST: 01/18/2020 02:05:35

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.079	01:54	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:58	No Sample Provided
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Air Blank	0.000	02:01	
Subject Sample	# 2	02:04	No Sample Provided
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Control Test	0.078	02:04	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 01/24/2020 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGINS,LYLE,P
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: MULET,C
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: PRADO

DATE/TIME OF TEST: 01/27/2020 23:41:40

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:30	
Control Test	0.081	23:31	
Air Blank	0.000	23:31	
Subject Sample	# 1	23:34	No Sample Provided
Breath Volume	0.000	23:34	
Air Blank	0.000	23:35	
Air Blank	0.000	23:37	
Subject Sample	# 2	23:40	No Sample Provided
Breath Volume	0.000	23:40	
Air Blank	0.000	23:40	
Control Test	0.079	23:41	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:41	

Data Download Date/Time: 02/26/2020 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGINS,LYLE,P
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: MULET,C
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: PRADO

DATE/TIME OF TEST: 01/27/2020 23:48:19

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:44	
Control Test	0.080	23:44	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:46	Subject Test Refused
Breath Volume	0.757	23:46	
Air Blank	0.000	23:47	
Control Test	0.080	23:47	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 02/26/2020 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,MIGUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:23

SUBJECT'S LAST NAME: OSCEOLA

DATE/TIME OF TEST: 01/31/2020 08:51:20

Diagnostics Check	OK	08:48	
Air Blank	0.000	08:48	
Control Test	0.079	08:49	
Air Blank	0.000	08:49	
Subject Sample	# 1	08:49	Subject Test Refused
Breath Volume	0.000	08:49	
Air Blank	0.000	08:50	
Control Test	0.079	08:50	
Air Blank	0.000	08:51	
Diagnostics Check	OK	08:51	

Data Download Date/Time: 02/26/2020 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 01/25/2020 03:18:02

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.074	03:13
Breath Volume	3.171	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:15
Subject Sample # 2	0.069	03:16
Breath Volume	2.609	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:18

Data Download Date/Time: 02/26/2020 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATTORNO,ALEXANDER,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GATTORNO,ALEXANDER
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:36

SUBJECT'S LAST NAME: JADOTTE

DATE/TIME OF TEST: 01/27/2020 08:34:41

Diagnostics Check	OK	08:28
Air Blank	0.000	08:28
Control Test	0.080	08:29
Air Blank	0.000	08:29
Subject Sample # 1	0.138	08:30
Breath Volume	2.484	08:30
Air Blank	0.000	08:30
Air Blank	0.000	08:32
Subject Sample # 2	0.144	08:33
Breath Volume	3.808	08:33
Air Blank	0.000	08:33
Control Test	0.080	08:34
Air Blank	0.000	08:34
Diagnostics Check	OK	08:34

Data Download Date/Time: 02/26/2020 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 01/27/2020 23:12:39

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.080	23:07
Air Blank	0.000	23:07
Subject Sample # 1	0.216	23:08
Breath Volume	3.304	23:08
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.215	23:11
Breath Volume	2.218	23:11
Air Blank	0.000	23:11
Control Test	0.079	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 02/26/2020 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGGINS,MICHAEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: HUGGINS,MICHAEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ABUKAMAR

DATE/TIME OF TEST: 01/04/2020 01:42:40

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:39	
Control Test	0.079	01:40	
Air Blank	0.000	01:40	
Subject Sample	# 1	01:40	Subject Test Refused
Breath Volume	0.000	01:40	
Air Blank	0.000	01:41	
Control Test	0.081	01:41	
Air Blank	0.000	01:42	
Diagnostics Check	Fail	01:42	Temperature Regulation Test
Air Blank	0.000	01:42	

Data Download Date/Time: 01/24/2020 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 01/11/2020 04:41:52

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
Control Test	0.081	04:39	
Air Blank	0.000	04:40	
Subject Sample	# 1	04:40	Subject Test Refused
Breath Volume	0.000	04:40	
Air Blank	0.000	04:40	
Control Test	0.081	04:41	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

Data Download Date/Time: 01/24/2020 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLEDO MARQUEZ,MIGUEL,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GONZALEZ,RITA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 01/18/2020 21:45:24

Diagnostics Check	OK	21:42	
Air Blank	0.000	21:43	
Control Test	0.080	21:43	
Air Blank	0.000	21:43	
Subject Sample	# 1	21:44	Subject Test Refused
Breath Volume	0.000	21:44	
Air Blank	0.000	21:44	
Control Test	0.078	21:44	
Air Blank	0.000	21:45	
Diagnostics Check	OK	21:45	

Data Download Date/Time: 01/24/2020 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: JUAREZ

DATE/TIME OF TEST: 01/25/2020 07:01:41

Diagnostics Check	OK	06:47	
Air Blank	0.000	06:48	
Control Test	0.080	06:48	
Air Blank	0.000	06:49	
Subject Sample	# 1 0.200	06:50	
Breath Volume	2.464	06:50	
Air Blank	0.000	06:51	
Air Blank	0.000	06:53	
Subject Sample	# 2 0.171	06:56	Volume Not Met
Breath Volume	0.617	06:56	
Air Blank	0.000	06:57	
Air Blank	0.000	06:58	
Subject Sample	# 3 0.202	07:00	
Breath Volume	2.347	07:00	
Air Blank	0.000	07:00	
Control Test	0.078	07:01	
Air Blank	0.000	07:01	
Diagnostics Check	OK	07:01	

Data Download Date/Time: 02/26/2020 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: ARTEAGA RODRIGUEZ

DATE/TIME OF TEST: 01/04/2020 02:25:55

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.231	02:21
Breath Volume	4.027	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.217	02:24
Breath Volume	2.601	02:24
Air Blank	0.000	02:24
Control Test	0.078	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 01/24/2020 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: QUINTANA

DATE/TIME OF TEST: 01/11/2020 02:54:28

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.079	02:52	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:53	Subject Test Refused
Breath Volume	0.000	02:53	
Air Blank	0.000	02:53	
Control Test	0.080	02:53	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 01/24/2020 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: MARCOS DE LEON

DATE/TIME OF TEST: 01/12/2020 05:07:17

Diagnostics Check	OK	05:05	
Air Blank	0.000	05:06	
Control Test	0.074	05:06	Control Outside Tolerance
Air Blank	0.000	05:07	

Data Download Date/Time: 01/24/2020 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: MARCOS DE LEON

DATE/TIME OF TEST: 01/12/2020 05:18:52

Diagnostics Check		OK	05:09
Air Blank		0.000	05:10
Control Test		0.079	05:10
Air Blank		0.000	05:10
Subject Sample	# 1	0.210	05:12
Breath Volume		1.980	05:12
Air Blank		0.000	05:13
Air Blank		0.000	05:14
Subject Sample	# 2	0.194	05:16
Breath Volume		1.816	05:16
Air Blank		0.000	05:17
Control Test		0.078	05:18
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 01/24/2020 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TAPANES,R
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 01/16/2020 00:42:41

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.079	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.225	00:37
Breath Volume	3.808	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.220	00:40
Breath Volume	2.023	00:40
Air Blank	0.000	00:41
Control Test	0.079	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 01/24/2020 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOSTALY,STEVENS,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: ESPADA,ROME
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: KISWANI

DATE/TIME OF TEST: 01/01/2020 03:33:54

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.078	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.076	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 01/22/2020 21:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORIVAL,PHANEL,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: GONZALEZ,ALEXANDER
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: GROSMAN

DATE/TIME OF TEST: 01/03/2020 06:16:28

Diagnostics Check	OK	06:10
Air Blank	0.000	06:10
Control Test	0.077	06:10
Air Blank	0.000	06:11
Subject Sample # 1	0.169	06:11
Breath Volume	3.046	06:11
Air Blank	0.000	06:12
Air Blank	0.000	06:14
Subject Sample # 2	0.158	06:14
Breath Volume	2.238	06:14
Air Blank	0.000	06:15
Control Test	0.077	06:15
Air Blank	0.000	06:16
Diagnostics Check	OK	06:16

Data Download Date/Time: 01/22/2020 21:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANUEVA, MAURICIO,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: MADERA, KEVIN
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: BABAIN

DATE/TIME OF TEST: 01/06/2020 18:05:02

Diagnostics Check	OK	17:56	
Air Blank	0.000	17:56	
Control Test	0.079	17:57	
Air Blank	0.000	17:57	
Subject Sample	# 1	18:00	No Sample Provided
Breath Volume	0.000	18:00	
Air Blank	0.000	18:01	
Air Blank	0.000	18:03	
Subject Sample	# 2	18:03	Subject Test Refused
Breath Volume	0.000	18:03	
Air Blank	0.000	18:04	
Control Test	0.078	18:04	
Air Blank	0.000	18:04	
Diagnostics Check	OK	18:05	

Data Download Date/Time: 01/22/2020 21:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOSTALY,STEVENS,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: KELNER,DOUGLAS
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 01/14/2020 08:03:01

Diagnostics Check	OK	07:55
Air Blank	0.000	07:55
Control Test	0.078	07:55
Air Blank	0.000	07:56
Subject Sample # 1	0.238	07:57
Breath Volume	1.515	07:57
Air Blank	0.000	07:58
Air Blank	0.000	08:00
Subject Sample # 2	0.235	08:01
Breath Volume	1.406	08:01
Air Blank	0.000	08:02
Control Test	0.077	08:02
Air Blank	0.000	08:02
Diagnostics Check	OK	08:03

Data Download Date/Time: 01/22/2020 21:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IINO,RYUNOSUKE,
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: IINO,RYUNOSUKE
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 12/31/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: BORODINA

DATE/TIME OF TEST: 01/21/2020 22:34:36

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:31	
Control Test	0.078	22:31	
Air Blank	0.000	22:32	
Subject Sample	# 1	22:33	Subject Test Refused
Breath Volume	0.000	22:33	
Air Blank	0.000	22:33	
Control Test	0.078	22:34	
Air Blank	0.000	22:34	
Diagnostics Check	OK	22:34	

Data Download Date/Time: 01/22/2020 21:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYA,JOSE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: MAYA,JOSE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 01/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: BARRERA

DATE/TIME OF TEST: 01/25/2020 01:55:36

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:53	
Control Test	0.077	01:53	
Air Blank	0.000	01:53	
Subject Sample	# 1	01:54	Subject Test Refused
Breath Volume	0.000	01:54	
Air Blank	0.000	01:54	
Control Test	0.077	01:55	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time: 02/15/2020 16:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYA,JOSE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: MAYA,JOSE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 01/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: BARRERA

DATE/TIME OF TEST: 01/25/2020 02:06:30

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:59	
Control Test	0.078	01:59	
Air Blank	0.000	02:00	
Subject Sample	# 1 0.172	02:01	
Breath Volume	1.316	02:01	
Air Blank	0.000	02:02	
Air Blank	0.000	02:04	
Subject Sample	# 2 0.189	02:04	
Breath Volume	1.515	02:04	
Air Blank	0.000	02:05	
Control Test	0.073	02:05	Control Outside Tolerance
Air Blank	0.000	02:06	

Data Download Date/Time: 02/15/2020 16:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULASKI,DAMON,J
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: HAMILTON,BRANDON
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A1
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CONNER

DATE/TIME OF TEST: 01/02/2020 00:08:17

Diagnostics Check		OK	00:02
Air Blank		0.000	00:02
Control Test		0.081	00:02
Air Blank		0.000	00:03
Subject Sample	# 1	0.000	00:03
Breath Volume		2.218	00:03
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.000	00:06
Breath Volume		2.210	00:06
Air Blank		0.000	00:07
Control Test		0.083	00:07
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 01/02/2020 19:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULASKI,DAMON,J
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: MOSCHIANO,DAVID
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A1
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: BUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: SINCLAIR

DATE/TIME OF TEST: 01/16/2020 19:07:27

Diagnostics Check	OK	19:04	
Air Blank	0.000	19:04	
Control Test	0.078	19:05	
Air Blank	0.000	19:05	
Subject Sample	# 1	19:06	Subject Test Refused
Breath Volume	0.000	19:06	
Air Blank	0.000	19:06	
Control Test	0.079	19:06	
Air Blank	0.000	19:07	
Diagnostics Check	OK	19:07	

Data Download Date/Time: 02/11/2020 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVIS,REECE,H
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: MIDOLO,GABRIELLE
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A1
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: BUI
ARREST TIME: 17:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/25/2020 17:50:26

Diagnostics Check	OK	17:44
Air Blank	0.000	17:44
Control Test	0.080	17:45
Air Blank	0.000	17:45
Subject Sample # 1	0.145	17:45
Breath Volume	2.199	17:45
Air Blank	0.000	17:46
Air Blank	0.000	17:48
Subject Sample # 2	0.139	17:48
Breath Volume	2.003	17:48
Air Blank	0.000	17:49
Control Test	0.080	17:49
Air Blank	0.000	17:50
Diagnostics Check	OK	17:50

Data Download Date/Time: 02/11/2020 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVIS,REECE,H
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: RORER,JESSICA
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A1
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: BUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: STICKNEY

DATE/TIME OF TEST: 01/25/2020 20:09:20

Diagnostics Check	OK	20:03
Air Blank	0.000	20:03
Control Test	0.080	20:03
Air Blank	0.000	20:04
Subject Sample # 1	0.179	20:04
Breath Volume	1.675	20:04
Air Blank	0.000	20:05
Air Blank	0.000	20:07
Subject Sample # 2	0.171	20:07
Breath Volume	1.757	20:07
Air Blank	0.000	20:08
Control Test	0.081	20:08
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

Data Download Date/Time: 02/11/2020 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARBOGAST, JOSHUA, S
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: KLEIS, KEVIN
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: BUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 01/12/2020 23:29:00

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.080	23:23
Air Blank	0.000	23:23
Subject Sample # 1	0.160	23:24
Breath Volume	2.933	23:24
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.154	23:27
Breath Volume	2.222	23:27
Air Blank	0.000	23:28
Control Test	0.079	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:29

Data Download Date/Time: 02/05/2020 16:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIS,KEVIN,W
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: ARBOGAST,JOSHUA
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: NIETO GUTIERREZ

DATE/TIME OF TEST: 01/18/2020 16:44:35

Diagnostics Check	OK	16:39	
Air Blank	0.000	16:39	
Control Test	0.081	16:40	
Air Blank	0.000	16:40	
Subject Sample	# 1	16:43	Subject Test Refused
Breath Volume	0.000	16:43	
Air Blank	0.000	16:43	
Control Test	0.082	16:43	
Air Blank	0.000	16:44	
Diagnostics Check	OK	16:44	

Data Download Date/Time: 02/05/2020 16:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: WICKERT

DATE/TIME OF TEST: 01/14/2020 00:00:41

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.078	23:55
Air Blank	0.000	23:55
Subject Sample # 1	0.058	23:56
Breath Volume	2.394	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.058	23:59
Breath Volume	2.351	23:59
Air Blank	0.000	23:59
Control Test	0.078	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 02/24/2020 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 01/14/2020 02:24:15

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.078	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.200	02:16
Breath Volume	2.250	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:18
Subject Sample # 2	0.179	02:19
Breath Volume	1.378	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 3	0.189	02:22
Breath Volume	2.386	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 02/24/2020 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: LOPEZ POJOY

DATE/TIME OF TEST: 01/18/2020 02:35:44

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.187	02:31
Breath Volume	1.433	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.184	02:34
Breath Volume	1.542	02:34
Air Blank	0.000	02:34
Control Test	0.079	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 02/24/2020 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: MARQUEZ MARTINEZ

DATE/TIME OF TEST: 01/19/2020 20:19:56

Diagnostics Check	OK	20:13
Air Blank	0.000	20:14
Control Test	0.079	20:14
Air Blank	0.000	20:14
Subject Sample # 1	0.193	20:15
Breath Volume	1.664	20:15
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.178	20:18
Breath Volume	1.250	20:18
Air Blank	0.000	20:19
Control Test	0.079	20:19
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 02/24/2020 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PLATT

DATE/TIME OF TEST: 01/25/2020 01:59:40

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.078	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.199	01:52
Breath Volume	2.300	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.232	01:55
Breath Volume	3.226	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 3	0.229	01:58
Breath Volume	2.105	01:58
Air Blank	0.000	01:58
Control Test	0.078	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 02/24/2020 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: GUZMAN PEREZ

DATE/TIME OF TEST: 01/28/2020 02:42:35

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.080	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.135	02:36
Breath Volume	1.214	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.129	02:41
Breath Volume	1.652	02:41
Air Blank	0.000	02:41
Control Test	0.080	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 02/24/2020 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICKEY, WILLIAM, Q
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: DICKEY, WILLIAM
ARRESTING OFFICER AGENCY: PORT ST. JOE PD
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: MUNOZ-JIMENEZ

DATE/TIME OF TEST: 01/01/2020 23:14:58

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.079	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.292	23:10
Breath Volume	2.664	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:12
Subject Sample # 2	0.275	23:13
Breath Volume	2.863	23:13
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 01/16/2020 21:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HEERDEN, TY, C
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: VAN HEERDEN, TY
ARRESTING OFFICER AGENCY: PORT ST. JOE PD
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: WILFONG

DATE/TIME OF TEST: 01/04/2020 03:47:31

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:45	
Control Test	0.079	03:45	
Air Blank	0.000	03:45	
Subject Sample	# 1	03:46	Subject Test Refused
Breath Volume	0.000	03:46	
Air Blank	0.000	03:46	
Control Test	0.078	03:46	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 01/16/2020 21:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HEERDEN, TY, C
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: ALDAY, SETH
ARRESTING OFFICER AGENCY: GULF COUNTY SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 01/05/2020 02:20:25

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.078	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.136	02:15
Breath Volume	1.558	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.126	02:18
Breath Volume	2.191	02:18
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 01/16/2020 21:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HEERDEN, TY, C
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: VAN HEERDEN, TY
ARRESTING OFFICER AGENCY: PORT ST. JOE PD
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: THURSBAY

DATE/TIME OF TEST: 01/05/2020 21:32:39

Diagnostics Check	OK	21:26
Air Blank	0.000	21:26
Control Test	0.079	21:27
Air Blank	0.000	21:27
Subject Sample # 1	0.000	21:28
Breath Volume	1.921	21:28
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.000	21:31
Breath Volume	1.394	21:31
Air Blank	0.000	21:31
Control Test	0.080	21:32
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 01/16/2020 21:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HIDEG 17018,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: BORGES LOPEZ

DATE/TIME OF TEST: 01/01/2020 03:53:22

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.080	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.203	03:48
Breath Volume		1.921	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.203	03:51
Breath Volume		1.882	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOFF 19034,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: HAGER

DATE/TIME OF TEST: 01/01/2020 04:33:07

Diagnostics Check		OK	04:21	
Air Blank		0.000	04:21	
Control Test		0.079	04:21	
Air Blank		0.000	04:22	
Subject Sample	# 1	0.064	04:25	Volume Not Met
Breath Volume		1.078	04:25	
Air Blank		0.000	04:26	
Air Blank		0.000	04:28	
Subject Sample	# 2	0.071	04:28	
Breath Volume		2.511	04:28	
Air Blank		0.000	04:29	
Air Blank		0.000	04:31	
Subject Sample	# 3	0.068	04:31	
Breath Volume		2.136	04:31	
Air Blank		0.000	04:32	
Control Test		0.080	04:32	
Air Blank		0.000	04:33	
Diagnostics Check		OK	04:33	

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,A,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: OTERO,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: BARRANTES VARGAS

DATE/TIME OF TEST: 01/01/2020 05:34:25

Diagnostics Check		OK	05:24
Air Blank		0.000	05:24
Control Test		0.080	05:24
Air Blank		0.000	05:25
Subject Sample	# 1	0.194	05:25
Breath Volume		2.453	05:25
Air Blank		0.000	05:26
Air Blank		0.000	05:28
Subject Sample	# 2	0.171	05:29
Breath Volume		1.398	05:29
Air Blank		0.000	05:29
Air Blank		0.000	05:31
Subject Sample	# 3	0.191	05:32
Breath Volume		2.562	05:32
Air Blank		0.000	05:33
Control Test		0.080	05:33
Air Blank		0.000	05:34
Diagnostics Check		OK	05:34

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,A,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: VERBIEST,R
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:02

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 01/01/2020 13:41:04

Diagnostics Check	OK	13:34
Air Blank	0.000	13:34
Control Test	0.080	13:35
Air Blank	0.000	13:35
Subject Sample # 1	0.246	13:36
Breath Volume	1.382	13:36
Air Blank	0.000	13:36
Air Blank	0.000	13:38
Subject Sample # 2	0.236	13:39
Breath Volume	1.660	13:39
Air Blank	0.000	13:40
Control Test	0.080	13:40
Air Blank	0.000	13:40
Diagnostics Check	OK	13:41

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSIAH,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH,JOSIAH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: DIAZ GARCIA

DATE/TIME OF TEST: 01/01/2020 22:40:30

Diagnostics Check	OK	22:32
Air Blank	0.000	22:32
Control Test	0.081	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.197	22:34
Breath Volume	2.550	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.194	22:38
Breath Volume	3.089	22:38
Air Blank	0.000	22:39
Control Test	0.079	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: GRIMMER

DATE/TIME OF TEST: 01/02/2020 18:48:31

Diagnostics Check	OK	18:40
Air Blank	0.000	18:41
Control Test	0.080	18:41
Air Blank	0.000	18:41
Subject Sample # 1	0.174	18:44
Breath Volume	1.785	18:44
Air Blank	0.000	18:44
Air Blank	0.000	18:46
Subject Sample # 2	0.160	18:47
Breath Volume	1.468	18:47
Air Blank	0.000	18:47
Control Test	0.078	18:47
Air Blank	0.000	18:48
Diagnostics Check	OK	18:48

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTER,CHRISTIAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELLISON,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 01/05/2020 05:47:59

Diagnostics Check	OK	05:45	
Air Blank	0.000	05:45	
Control Test	0.080	05:45	
Air Blank	0.000	05:46	
Subject Sample	# 1	05:46	RFI Detect
Breath Volume	0.000	05:46	
Air Blank	0.000	05:47	
Control Test	0.079	05:47	
Air Blank	0.000	05:47	
Diagnostics Check	OK	05:47	

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTER,CHRISTIAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELLISON,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 01/05/2020 05:58:15

Diagnostics Check	OK	05:51
Air Blank	0.000	05:52
Control Test	0.080	05:52
Air Blank	0.000	05:53
Subject Sample # 1	0.172	05:53
Breath Volume	2.382	05:53
Air Blank	0.000	05:54
Air Blank	0.000	05:56
Subject Sample # 2	0.175	05:56
Breath Volume	3.062	05:56
Air Blank	0.000	05:57
Control Test	0.078	05:57
Air Blank	0.000	05:58
Diagnostics Check	OK	05:58

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAEFER,K,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WILLIAMS,T
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: RIVERA ROMAN

DATE/TIME OF TEST: 01/05/2020 07:22:43

Diagnostics Check	OK	07:14
Air Blank	0.000	07:15
Control Test	0.079	07:15
Air Blank	0.000	07:16
Subject Sample # 1	0.129	07:16
Breath Volume	1.843	07:16
Air Blank	0.000	07:17
Air Blank	0.000	07:18
Subject Sample # 2	0.121	07:21
Breath Volume	1.285	07:21
Air Blank	0.000	07:21
Control Test	0.081	07:22
Air Blank	0.000	07:22
Diagnostics Check	OK	07:22

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,KATIE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CALHOUN,KURTIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: GEFFEN

DATE/TIME OF TEST: 01/07/2020 23:02:05

Diagnostics Check	OK	22:54
Air Blank	0.000	22:55
Control Test	0.080	22:55
Air Blank	0.000	22:55
Subject Sample # 1	0.117	22:57
Breath Volume	1.472	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.117	23:00
Breath Volume	1.261	23:00
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,PATRICK,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KIRK,PATRICK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:35

SUBJECT'S LAST NAME: ACKER

DATE/TIME OF TEST: 01/08/2020 10:05:23

Diagnostics Check	OK	09:59
Air Blank	0.000	09:59
Control Test	0.080	09:59
Air Blank	0.000	10:00
Subject Sample # 1	0.121	10:00
Breath Volume	2.472	10:00
Air Blank	0.000	10:01
Air Blank	0.000	10:03
Subject Sample # 2	0.117	10:03
Breath Volume	2.695	10:03
Air Blank	0.000	10:04
Control Test	0.077	10:04
Air Blank	0.000	10:05
Diagnostics Check	OK	10:05

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ORTINO 19064,E
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: SALAVEJUS

DATE/TIME OF TEST: 01/08/2020 22:39:58

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.000	22:35
Breath Volume	2.386	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.000	22:38
Breath Volume	2.390	22:38
Air Blank	0.000	22:39
Control Test	0.080	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: DOTSON

DATE/TIME OF TEST: 01/08/2020 23:10:41

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.080	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.124	23:06
Breath Volume	3.093	23:06
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.124	23:09
Breath Volume	2.796	23:09
Air Blank	0.000	23:09
Control Test	0.079	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LOCK 06117,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 01/09/2020 04:36:47

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.080	04:31
Air Blank	0.000	04:31
Subject Sample # 1	0.000	04:32
Breath Volume	3.957	04:32
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.000	04:35
Breath Volume	3.792	04:35
Air Blank	0.000	04:35
Control Test	0.080	04:36
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CALHOUN 06065,K
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: ANTONIO MARTINEZ

DATE/TIME OF TEST: 01/10/2020 00:41:14

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.192	00:35
Breath Volume	1.847	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.185	00:39
Breath Volume	2.437	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CHAML 19048,A
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: GASPAR ANTONIO

DATE/TIME OF TEST: 01/10/2020 04:24:49

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.079	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.203	04:20
Breath Volume	2.535	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.198	04:23
Breath Volume	3.777	04:23
Air Blank	0.000	04:23
Control Test	0.080	04:24
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEINER,RYAN,T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: RODRIGUEZ,E
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: DREW

DATE/TIME OF TEST: 01/11/2020 03:34:14

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.158	03:28
Breath Volume	2.851	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.163	03:32
Breath Volume	2.816	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BARKLEY,JEFFREY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 01/11/2020 05:23:09

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.080	05:17
Air Blank	0.000	05:17
Subject Sample # 1	0.122	05:18
Breath Volume	1.734	05:18
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 2	0.127	05:21
Breath Volume	1.656	05:21
Air Blank	0.000	05:22
Control Test	0.080	05:22
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 01/11/2020 20:24:29

Diagnostics Check	OK	20:21	
Air Blank	0.000	20:21	
Control Test	0.080	20:22	
Air Blank	0.000	20:22	
Subject Sample	# 1	20:23	Subject Test Refused
Breath Volume	0.000	20:23	
Air Blank	0.000	20:23	
Control Test	0.080	20:23	
Air Blank	0.000	20:24	
Diagnostics Check	OK	20:24	

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS, MATTHEW, K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CICCHINI, PAUL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: RUBICCO

DATE/TIME OF TEST: 01/11/2020 21:47:48

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.155	21:42
Breath Volume	1.437	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.170	21:46
Breath Volume	2.281	21:46
Air Blank	0.000	21:46
Control Test	0.079	21:47
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSIAH,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH,JOSIAH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: FONTALVO ROYERO

DATE/TIME OF TEST: 01/12/2020 01:01:05

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.210	00:56
Breath Volume	3.070	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.218	00:59
Breath Volume	2.871	00:59
Air Blank	0.000	01:00
Control Test	0.078	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINS,KYLE,L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: LUNA,MICHAEL
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: CAMPO

DATE/TIME OF TEST: 01/12/2020 01:24:36

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.186	01:19
Breath Volume	1.527	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.182	01:23
Breath Volume	1.312	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADURO,VICTOR,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MADURO,VICTOR
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 01/12/2020 02:06:06

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.140	02:01
Breath Volume	1.660	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.138	02:04
Breath Volume	2.042	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,KATIE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DEKEYSER,NICKOLAS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: DOBSON

DATE/TIME OF TEST: 01/12/2020 04:21:45

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:07	
Control Test		0.080	04:07	
Air Blank		0.000	04:07	
Subject Sample	# 1	0.145	04:08	
Breath Volume		1.203	04:08	
Air Blank		0.000	04:09	
Air Blank		0.000	04:11	
Subject Sample	# 2	0.087	04:14	Volume Not Met
Breath Volume		0.050	04:14	
Air Blank		0.000	04:15	
Air Blank		0.000	04:17	
Subject Sample	# 3	0.130	04:20	Volume Not Met
Breath Volume		0.089	04:20	
Air Blank		0.000	04:20	
Control Test		0.079	04:21	
Air Blank		0.000	04:21	
Diagnostics Check		OK	04:21	

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREEMAN,MONICA,M
BREATH TEST OPERATOR AGENCY: Lee Port Authority
ARRESTING OFFICER: GALLOWAY,ERIC
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: SUMMERS

DATE/TIME OF TEST: 01/13/2020 01:34:10

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.167	01:29
Breath Volume	3.253	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.161	01:32
Breath Volume	2.617	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY,E,
BREATH TEST OPERATOR AGENCY: Lee Port Authority
ARRESTING OFFICER: BELLO,H
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 01/13/2020 02:32:20

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.080	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.148	02:27
Breath Volume	2.382	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.145	02:30
Breath Volume	2.550	02:30
Air Blank	0.000	02:31
Control Test	0.079	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAWLINS,DUSTIN,P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CONDON,JAMES
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: WILDMAN

DATE/TIME OF TEST: 01/13/2020 08:52:11

Diagnostics Check	OK	08:45
Air Blank	0.000	08:46
Control Test	0.079	08:46
Air Blank	0.000	08:47
Subject Sample # 1	0.127	08:47
Breath Volume	1.824	08:47
Air Blank	0.000	08:48
Air Blank	0.000	08:50
Subject Sample # 2	0.122	08:50
Breath Volume	2.593	08:50
Air Blank	0.000	08:51
Control Test	0.079	08:51
Air Blank	0.000	08:52
Diagnostics Check	OK	08:52

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LA SPIA 16071,G
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 01/13/2020 22:29:28

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.080	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.036	22:25
Breath Volume	2.343	22:25
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.036	22:28
Breath Volume	2.277	22:28
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEFFIN,DAN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUBOIS,MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: GIESSEL

DATE/TIME OF TEST: 01/14/2020 14:39:32

Diagnostics Check		OK	14:26	
Air Blank		0.000	14:27	
Control Test		0.079	14:27	
Air Blank		0.000	14:28	
Subject Sample	# 1	0.114	14:31	Volume Not Met
Breath Volume		0.230	14:31	
Air Blank		0.000	14:31	
Air Blank		0.000	14:33	
Subject Sample	# 2	0.128	14:34	
Breath Volume		1.738	14:34	
Air Blank		0.000	14:35	
Air Blank		0.000	14:37	
Subject Sample	# 3	0.127	14:38	
Breath Volume		1.273	14:38	
Air Blank		0.000	14:38	
Control Test		0.078	14:38	
Air Blank		0.000	14:39	
Diagnostics Check		OK	14:39	

Data Download Date/Time: 01/21/2020 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGOWAN, TYLER, J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MCGOWAN, TYLER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:39

SUBJECT'S LAST NAME: FOUTCH

DATE/TIME OF TEST: 01/08/2020 10:05:34

Diagnostics Check	OK	09:59
Air Blank	0.000	10:00
Control Test	0.080	10:00
Air Blank	0.000	10:01
Subject Sample # 1	0.000	10:01
Breath Volume	2.078	10:01
Air Blank	0.000	10:01
Air Blank	0.000	10:03
Subject Sample # 2	0.000	10:04
Breath Volume	2.972	10:04
Air Blank	0.000	10:04
Control Test	0.079	10:04
Air Blank	0.000	10:05
Diagnostics Check	OK	10:05

Data Download Date/Time: 01/15/2020 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANPELT,J,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: VANPELT,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 01/09/2020 01:17:12

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.223	01:12
Breath Volume		2.277	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:14
Subject Sample	# 2	0.217	01:15
Breath Volume		2.917	01:15
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 01/15/2020 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONSALL, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BONSALL, MATTHEW
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: ARCINIEGA HERNANDEZ

DATE/TIME OF TEST: 01/10/2020 00:52:35

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.081	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.134	00:47
Breath Volume	2.953	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.120	00:51
Breath Volume	1.550	00:51
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 01/15/2020 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: VALCIN,JEFFREY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 01/11/2020 00:32:55

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.082	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.079	00:28
Breath Volume	3.128	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.077	00:31
Breath Volume	2.988	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 01/15/2020 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILLAK, MICHAEL, O
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CARVALHO, ASHLEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: APARICIO-NOYOLA

DATE/TIME OF TEST: 01/12/2020 01:03:40

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.081	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.211	00:58
Breath Volume		1.906	00:58
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.200	01:01
Breath Volume		2.367	01:01
Air Blank		0.000	01:02
Control Test		0.080	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 01/15/2020 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHILKE,CARL,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SCHILKE,CARL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:09

SUBJECT'S LAST NAME: UGLAND

DATE/TIME OF TEST: 01/12/2020 17:42:03

Diagnostics Check	OK	17:33
Air Blank	0.000	17:34
Control Test	0.081	17:34
Air Blank	0.000	17:34
Subject Sample # 1	0.147	17:37
Breath Volume	2.832	17:37
Air Blank	0.000	17:38
Air Blank	0.000	17:39
Subject Sample # 2	0.148	17:40
Breath Volume	2.777	17:40
Air Blank	0.000	17:41
Control Test	0.080	17:41
Air Blank	0.000	17:41
Diagnostics Check	OK	17:42

Data Download Date/Time: 01/15/2020 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONSALL, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BONSALL, M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: SHOLOMITH

DATE/TIME OF TEST: 01/14/2020 20:06:46

Diagnostics Check	OK	20:00
Air Blank	0.000	20:00
Control Test	0.081	20:01
Air Blank	0.000	20:01
Subject Sample # 1	0.290	20:02
Breath Volume	2.367	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:04
Subject Sample # 2	0.283	20:05
Breath Volume	2.105	20:05
Air Blank	0.000	20:05
Control Test	0.080	20:06
Air Blank	0.000	20:06
Diagnostics Check	OK	20:06

Data Download Date/Time: 01/15/2020 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JEAN BAPTISTE,LOUSSANDA
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 01/15/2020 00:46:29

Diagnostics Check		OK	00:39
Air Blank		0.000	00:40
Control Test		0.081	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.227	00:41
Breath Volume		1.316	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.232	00:44
Breath Volume		1.105	00:44
Air Blank		0.000	00:45
Control Test		0.078	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 01/15/2020 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHILKE,CARL,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SCHILKE,CARL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:21

SUBJECT'S LAST NAME: RACIUS

DATE/TIME OF TEST: 01/15/2020 17:50:08

Diagnostics Check	OK	17:42
Air Blank	0.000	17:43
Control Test	0.081	17:43
Air Blank	0.000	17:44
Subject Sample # 1	0.290	17:45
Breath Volume	1.980	17:45
Air Blank	0.000	17:46
Air Blank	0.000	17:48
Subject Sample # 2	0.285	17:48
Breath Volume	1.839	17:48
Air Blank	0.000	17:49
Control Test	0.079	17:49
Air Blank	0.000	17:50
Diagnostics Check	OK	17:50

Data Download Date/Time: 02/05/2020 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MARTINEZ,CARLOS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: FRANCISCO SALVADOR

DATE/TIME OF TEST: 01/16/2020 02:36:29

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.177	02:32
Breath Volume	2.554	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.159	02:34
Breath Volume	1.253	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 02/05/2020 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROBEL, RAYMOND, R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: STROBEL, RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:10

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 01/16/2020 12:42:36

Diagnostics Check	OK	12:36
Air Blank	0.000	12:36
Control Test	0.080	12:36
Air Blank	0.000	12:37
Subject Sample # 1	0.000	12:38
Breath Volume	2.796	12:38
Air Blank	0.000	12:38
Air Blank	0.000	12:40
Subject Sample # 2	0.000	12:41
Breath Volume	3.191	12:41
Air Blank	0.000	12:41
Control Test	0.081	12:41
Air Blank	0.000	12:42
Diagnostics Check	OK	12:42

Data Download Date/Time: 02/05/2020 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DINO,JOSEPH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 01/16/2020 20:53:35

Diagnostics Check	OK	20:47
Air Blank	0.000	20:47
Control Test	0.081	20:47
Air Blank	0.000	20:48
Subject Sample # 1	0.177	20:48
Breath Volume	2.906	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.178	20:52
Breath Volume	2.546	20:52
Air Blank	0.000	20:52
Control Test	0.081	20:52
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 02/05/2020 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISCHER, MICHAEL, A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FISCHER, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 01/21/2020 00:27:31

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.081	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.029	00:23
Breath Volume	3.437	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.021	00:25
Breath Volume	2.457	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 02/05/2020 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 01/21/2020 21:38:12

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.081	21:33
Air Blank	0.000	21:33
Subject Sample # 1	0.000	21:33
Breath Volume	2.203	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.000	21:36
Breath Volume	1.941	21:36
Air Blank	0.000	21:37
Control Test	0.082	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 02/05/2020 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONSALL, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HOYOS, J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: UTUY PASTOR

DATE/TIME OF TEST: 01/23/2020 22:01:00

Diagnostics Check		OK	21:53
Air Blank		0.000	21:54
Control Test		0.081	21:54
Air Blank		0.000	21:55
Subject Sample	# 1	0.254	21:56
Breath Volume		1.613	21:56
Air Blank		0.000	21:57
Air Blank		0.000	21:58
Subject Sample	# 2	0.256	21:59
Breath Volume		1.609	21:59
Air Blank		0.000	22:00
Control Test		0.079	22:00
Air Blank		0.000	22:00
Diagnostics Check		OK	22:01

Data Download Date/Time: 02/05/2020 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CONLIN,EVAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: SOLHEIM

DATE/TIME OF TEST: 01/25/2020 02:50:45

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.086	02:46
Breath Volume	1.683	02:46
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.088	02:49
Breath Volume	1.730	02:49
Air Blank	0.000	02:49
Control Test	0.080	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 02/05/2020 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISCHER, MICHAEL, A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FISCHER, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: EKHOLM

DATE/TIME OF TEST: 01/25/2020 04:41:58

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.080	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.193	04:37
Breath Volume	4.804	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:39
Subject Sample # 2	0.185	04:40
Breath Volume	4.695	04:40
Air Blank	0.000	04:41
Control Test	0.078	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 02/05/2020 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YETTER,ERIC,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BYRD,BRIAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: GURR

DATE/TIME OF TEST: 01/27/2020 20:46:13

Diagnostics Check	OK	20:37
Air Blank	0.000	20:38
Control Test	0.082	20:38
Air Blank	0.000	20:39
Subject Sample # 1	0.007	20:41
Breath Volume	1.175	20:41
Air Blank	0.000	20:42
Air Blank	0.000	20:43
Subject Sample # 2	0.000	20:44
Breath Volume	1.113	20:44
Air Blank	0.000	20:45
Control Test	0.080	20:45
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

Data Download Date/Time: 02/05/2020 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWELL,NATHANAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROWELL,NATHANAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: YOLLES

DATE/TIME OF TEST: 01/01/2020 07:36:09

Diagnostics Check	OK	07:29
Air Blank	0.000	07:30
Control Test	0.079	07:30
Air Blank	0.000	07:30
Subject Sample # 1	0.000	07:31
Breath Volume	2.320	07:31
Air Blank	0.000	07:32
Air Blank	0.000	07:34
Subject Sample # 2	0.000	07:34
Breath Volume	1.621	07:34
Air Blank	0.000	07:35
Control Test	0.079	07:35
Air Blank	0.000	07:36
Diagnostics Check	OK	07:36

Data Download Date/Time: 01/29/2020 08:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAOUNI,K,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: SHAOUNI,K
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MENESES

DATE/TIME OF TEST: 01/12/2020 02:50:06

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.079	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.186	02:45
Breath Volume		1.269	02:45
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.193	02:48
Breath Volume		1.621	02:48
Air Blank		0.000	02:49
Control Test		0.078	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 01/29/2020 08:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER, MARCUS, H
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: VILLACRES, NICHOLAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: CHOPIN

DATE/TIME OF TEST: 01/14/2020 21:27:39

Diagnostics Check		OK	21:20
Air Blank		0.000	21:20
Control Test		0.079	21:21
Air Blank		0.000	21:21
Subject Sample	# 1	0.246	21:23
Breath Volume		1.382	21:23
Air Blank		0.000	21:24
Air Blank		0.000	21:25
Subject Sample	# 2	0.259	21:26
Breath Volume		1.937	21:26
Air Blank		0.000	21:26
Control Test		0.077	21:27
Air Blank		0.000	21:27
Diagnostics Check		OK	21:27

Data Download Date/Time: 01/29/2020 08:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GRAY,CONNER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: PARONTO

DATE/TIME OF TEST: 01/18/2020 01:12:33

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.159	01:08
Breath Volume	1.597	01:08
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.157	01:11
Breath Volume	1.589	01:11
Air Blank	0.000	01:11
Control Test	0.077	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 01/29/2020 08:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GRAY,CONNER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: CHAFFINS

DATE/TIME OF TEST: 01/19/2020 02:28:42

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.211	02:24
Breath Volume	3.632	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:26
Subject Sample # 2	0.212	02:27
Breath Volume	3.644	02:27
Air Blank	0.000	02:27
Control Test	0.078	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 01/29/2020 08:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: VILLACRES,NICHOLAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: CRASSO

DATE/TIME OF TEST: 01/20/2020 01:48:39

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:46	
Control Test	0.079	01:46	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:47	Subject Test Refused
Breath Volume	0.000	01:47	
Air Blank	0.000	01:47	
Control Test	0.079	01:48	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

Data Download Date/Time: 01/29/2020 08:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WINSTEAD,JUSTIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: PIER

DATE/TIME OF TEST: 01/23/2020 23:35:59

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.080	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.103	23:31
Breath Volume	3.148	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:33
Subject Sample # 2	0.098	23:34
Breath Volume	2.902	23:34
Air Blank	0.000	23:35
Control Test	0.078	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 01/29/2020 08:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,NICHOLAS,W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LEE,N
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: BIEGLER

DATE/TIME OF TEST: 01/26/2020 00:30:01

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.165	00:25
Breath Volume	1.687	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.163	00:28
Breath Volume	1.800	00:28
Air Blank	0.000	00:29
Control Test	0.079	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:30

Data Download Date/Time: 01/29/2020 08:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAOUNI,K,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: SHAOUNI,K
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: CORIDDI

DATE/TIME OF TEST: 01/26/2020 23:47:07

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.079	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.007	23:42
Breath Volume	1.863	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.000	23:45
Breath Volume	1.433	23:45
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 01/29/2020 08:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MARCUS,H
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BAKER,MARCUS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 01/28/2020 00:51:24

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.079	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.123	00:46
Breath Volume	3.628	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:48
Subject Sample # 2	0.117	00:49
Breath Volume	2.835	00:49
Air Blank	0.000	00:50
Control Test	0.078	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 01/29/2020 08:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GANGL,B
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 12/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: LAPORTA

DATE/TIME OF TEST: 01/01/2020 01:37:15

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.000	01:33
Breath Volume	1.148	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.000	01:35
Breath Volume	1.457	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 01/06/2020 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ELIAS,DORIN
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 12/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 01/01/2020 03:50:32

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.080	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.121	03:46
Breath Volume	1.589	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.125	03:49
Breath Volume	1.191	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 01/06/2020 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATSON,MITCHELL,J
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BATSON,MITCHELL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 12/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: BAZOALTO SUAREZ

DATE/TIME OF TEST: 01/01/2020 23:42:35

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:37
Air Blank	0.000	23:37
Subject Sample # 1	0.235	23:38
Breath Volume	2.644	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.232	23:41
Breath Volume	2.101	23:41
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 01/06/2020 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,JAMES,P
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: DRISCOLL,JAMES
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 12/02/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 01/02/2020 21:51:07

Diagnostics Check		OK	21:36	
Air Blank		0.000	21:36	
Control Test		0.079	21:37	
Air Blank		0.000	21:37	
Subject Sample	# 1	0.000	21:40	Volume Not Met
Breath Volume		0.015	21:40	
Air Blank		0.000	21:41	
Air Blank		0.000	21:43	
Subject Sample	# 2	0.000	21:44	
Breath Volume		1.105	21:44	
Air Blank		0.000	21:44	
Air Blank		0.000	21:46	
Subject Sample	# 3	0.000	21:49	Volume Not Met
Breath Volume		1.054	21:49	
Air Blank		0.000	21:50	
Control Test		0.081	21:50	
Air Blank		0.000	21:51	
Diagnostics Check		OK	21:51	

Data Download Date/Time: 01/06/2020 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ GAITAN, JUAN CAMILO,
BREATH TEST OPERATOR AGENCY: FHP TROOP F
ARRESTING OFFICER: WADDELL, NATHANIEL
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: GOMEZ GAITAN

DATE/TIME OF TEST: 01/19/2020 16:10:40

Diagnostics Check	OK	16:09	
Air Blank	0.000	16:10	
Control Test	0.079	16:10	
Air Blank		16:10	Improper Sample
Air Blank		16:10	Improper Sample

Data Download Date/Time: 02/03/2020 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: FHP TROOP F
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: GOMEZ GAITAN

DATE/TIME OF TEST: 01/19/2020 16:14:01

Diagnostics Check	OK	16:13	
Air Blank		16:13	Improper Sample
Air Blank		16:14	Improper Sample

Data Download Date/Time: 02/03/2020 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: FHP TROOP F
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: GOMEZ GAITAN

DATE/TIME OF TEST: 01/19/2020 16:17:39

Diagnostics Check	OK	16:16	
Air Blank		16:17	Ambient Fail
Air Blank	0.000	16:17	

Data Download Date/Time: 02/03/2020 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: FHP TROOP F
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: GOMEZ GAITAN

DATE/TIME OF TEST: 01/19/2020 16:20:58

Diagnostics Check	OK	16:20	
Air Blank		16:20	Improper Sample
Air Blank		16:20	Improper Sample

Data Download Date/Time: 02/03/2020 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: HUGHES, DAVID
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 12/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/01/2020 12:00:47

Diagnostics Check	OK	11:54
Air Blank	0.000	11:55
Control Test	0.077	11:55
Air Blank	0.000	11:56
Subject Sample # 1	0.000	11:56
Breath Volume	1.417	11:56
Air Blank	0.000	11:57
Air Blank	0.000	11:59
Subject Sample # 2	0.000	11:59
Breath Volume	1.902	11:59
Air Blank	0.000	11:59
Control Test	0.075	12:00
Air Blank	0.000	12:00
Diagnostics Check	OK	12:00

Data Download Date/Time: 01/23/2020 23:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADRUGO,FELICIA,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CLENDENING,JAMES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: HANCOCK

DATE/TIME OF TEST: 01/13/2020 04:21:23

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:15	
Control Test	0.076	04:15	
Air Blank	0.000	04:16	
Subject Sample # 1	0.171	04:16	
Breath Volume	1.621	04:16	
Air Blank	0.000	04:17	
Air Blank	0.000	04:19	
Subject Sample # 2	0.173	04:19	
Breath Volume	1.863	04:19	
Air Blank	0.000	04:20	
Control Test	0.074	04:20	Control Outside Tolerance
Air Blank	0.000	04:21	

Data Download Date/Time: 01/23/2020 23:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EPPS,QUINTELLA,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CLENDENING,JAMES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 01/16/2020 22:00:14

Diagnostics Check	OK	21:53
Air Blank	0.000	21:54
Control Test	0.078	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.111	21:55
Breath Volume	1.722	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.112	21:58
Breath Volume	1.664	21:58
Air Blank	0.000	21:59
Control Test	0.076	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 01/23/2020 23:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: JEAN, FRANTZ
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 01/20/2020 10:57:56

Diagnostics Check	OK	10:51
Air Blank	0.000	10:52
Control Test	0.078	10:52
Air Blank	0.000	10:52
Subject Sample # 1	0.129	10:53
Breath Volume	1.972	10:53
Air Blank	0.000	10:54
Air Blank	0.000	10:55
Subject Sample # 2	0.132	10:56
Breath Volume	2.218	10:56
Air Blank	0.000	10:57
Control Test	0.076	10:57
Air Blank	0.000	10:57
Diagnostics Check	OK	10:57

Data Download Date/Time: 01/23/2020 23:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: JEAN, FRANTZ
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 01/20/2020 12:27:53

Diagnostics Check	OK	12:20
Air Blank	0.000	12:21
Control Test	0.077	12:21
Air Blank	0.000	12:22
Subject Sample # 1	0.113	12:23
Breath Volume	2.609	12:23
Air Blank	0.000	12:24
Air Blank	0.000	12:25
Subject Sample # 2	0.109	12:26
Breath Volume	2.566	12:26
Air Blank	0.000	12:26
Control Test	0.075	12:27
Air Blank	0.000	12:27
Diagnostics Check	OK	12:27

Data Download Date/Time: 01/23/2020 23:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: JEAN, FRANTZ
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 01/20/2020 15:32:12

Diagnostics Check	OK	15:28	
Air Blank	0.000	15:29	
Control Test	0.078	15:29	
Air Blank	0.000	15:29	
Subject Sample # 1	0.038	15:30	
Breath Volume	1.417	15:30	
Air Blank		15:30	Improper Sample
Air Blank	0.000	15:31	
Control Test	0.078	15:31	
Air Blank	0.000	15:32	
Diagnostics Check	OK	15:32	

Data Download Date/Time: 01/23/2020 23:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: JEAN, FRANTZ
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 01/20/2020 15:40:49

Diagnostics Check	OK	15:34
Air Blank	0.000	15:35
Control Test	0.077	15:35
Air Blank	0.000	15:35
Subject Sample # 1	0.036	15:36
Breath Volume	1.179	15:36
Air Blank	0.000	15:37
Air Blank	0.000	15:38
Subject Sample # 2	0.035	15:39
Breath Volume	1.808	15:39
Air Blank	0.000	15:39
Control Test	0.076	15:40
Air Blank	0.000	15:40
Diagnostics Check	OK	15:40

Data Download Date/Time: 01/23/2020 23:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADRUGO,FELICIA,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CLENDENING,JAMES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 01/25/2020 00:00:20

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:50	
Control Test	0.077	23:51	
Air Blank	0.000	23:51	
Subject Sample	# 1 0.263	23:53	
Breath Volume	1.644	23:53	
Air Blank	0.000	23:54	
Air Blank	0.000	23:55	
Subject Sample	# 2 0.266	23:58	
Breath Volume	1.812	23:58	
Air Blank	0.000	23:59	
Control Test	0.071	23:59	Control Outside Tolerance
Air Blank	0.000	00:00	

Data Download Date/Time: 02/26/2020 03:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CLENDING, JAMES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 01/25/2020 13:34:07

Diagnostics Check	OK	13:28
Air Blank	0.000	13:28
Control Test	0.077	13:28
Air Blank	0.000	13:29
Subject Sample # 1	0.044	13:29
Breath Volume	1.953	13:29
Air Blank	0.000	13:30
Air Blank	0.000	13:32
Subject Sample # 2	0.043	13:32
Breath Volume	1.902	13:32
Air Blank	0.000	13:33
Control Test	0.077	13:33
Air Blank	0.000	13:34
Diagnostics Check	OK	13:34

Data Download Date/Time: 02/26/2020 03:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CLENDING, JAMES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 01/25/2020 15:18:14

Diagnostics Check	OK	15:12
Air Blank	0.000	15:12
Control Test	0.077	15:13
Air Blank	0.000	15:13
Subject Sample # 1	0.010	15:14
Breath Volume	2.007	15:14
Air Blank	0.000	15:14
Air Blank	0.000	15:16
Subject Sample # 2	0.008	15:16
Breath Volume	2.160	15:16
Air Blank	0.000	15:17
Control Test	0.075	15:17
Air Blank	0.000	15:18
Diagnostics Check	OK	15:18

Data Download Date/Time: 02/26/2020 03:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWE,ANGELA,M
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: CORTOPASSI,BARON
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 12/19/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: HOWE

DATE/TIME OF TEST: 01/15/2020 21:37:56

Diagnostics Check	OK	21:26	
Air Blank	0.000	21:27	
Control Test	0.077	21:27	
Air Blank	0.000	21:27	
Subject Sample	# 1	21:30	No Sample Provided
Breath Volume	0.000	21:30	
Air Blank	0.000	21:31	
Air Blank	0.000	21:33	
Subject Sample	# 2	21:36	No Sample Provided
Breath Volume	0.000	21:36	
Air Blank	0.000	21:37	
Control Test	0.076	21:37	
Air Blank	0.000	21:37	
Diagnostics Check	OK	21:37	

Data Download Date/Time: 01/22/2020 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,JOHN,S
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: HAGEDORN,NICK
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/10/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SLAUGHTER

DATE/TIME OF TEST: 01/01/2020 03:31:48

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.190	03:27
Breath Volume	3.136	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.178	03:30
Breath Volume	2.375	03:30
Air Blank	0.000	03:30
Control Test	0.079	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 01/14/2020 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE, RAYMOND, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WHITE, RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: LAVALLE

DATE/TIME OF TEST: 01/06/2020 19:16:51

Diagnostics Check	OK	19:09	
Air Blank	0.000	19:09	
Control Test	0.079	19:10	
Air Blank	0.000	19:10	
Subject Sample # 1	0.092	19:11	
Breath Volume	4.089	19:11	
Air Blank	0.000	19:12	
Air Blank	0.000	19:14	
Subject Sample # 2	0.080	19:15	
Breath Volume	2.734	19:15	
Air Blank	0.000	19:15	
Control Test	0.074	19:16	Control Outside Tolerance
Air Blank	0.000	19:16	

Data Download Date/Time: 01/14/2020 10:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE, RAYMOND, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WHITE, RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: LAVALLE

DATE/TIME OF TEST: 01/06/2020 19:23:25

Diagnostics Check	OK	19:19	
Air Blank	0.000	19:20	
Control Test	0.078	19:20	
Air Blank	0.000	19:20	
Subject Sample # 1	0.086	19:21	
Breath Volume	2.742	19:21	
Air Blank		19:22	RFI Detect
Air Blank	0.000	19:22	
Control Test	0.079	19:22	
Air Blank	0.000	19:23	
Diagnostics Check	OK	19:23	

Data Download Date/Time: 01/14/2020 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE, RAYMOND, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WHITE, RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: LAVALLE

DATE/TIME OF TEST: 01/06/2020 19:31:48

Diagnostics Check	OK	19:24
Air Blank	0.000	19:25
Control Test	0.079	19:25
Air Blank	0.000	19:26
Subject Sample # 1	0.083	19:26
Breath Volume	3.015	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.076	19:30
Breath Volume	2.050	19:30
Air Blank	0.000	19:30
Control Test	0.079	19:31
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 01/14/2020 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEYER, CHARLES, L
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: YOUNG, JESSE
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 01/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:36

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 01/18/2020 09:03:14

Diagnostics Check	OK	08:56
Air Blank	0.000	08:56
Control Test	0.079	08:57
Air Blank	0.000	08:57
Subject Sample # 1	0.299	08:58
Breath Volume	1.558	08:58
Air Blank	0.000	08:59
Air Blank	0.000	09:01
Subject Sample # 2	0.292	09:01
Breath Volume	1.242	09:01
Air Blank	0.000	09:02
Control Test	0.078	09:02
Air Blank	0.000	09:03
Diagnostics Check	OK	09:03

Data Download Date/Time: 02/17/2020 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

INSTRUMENT REGISTERED TO: LIBERTY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROHECKER,KYNDAL,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DEASON,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 01/08/2020 23:08:25

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.102	23:03
Breath Volume	1.378	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.100	23:06
Breath Volume	1.574	23:06
Air Blank	0.000	23:07
Control Test	0.082	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 02/18/2020 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAUFFMAN, MITCHELL, P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: KAUFFMAN, MITCHELL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: LOWERY

DATE/TIME OF TEST: 01/25/2020 02:29:04

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:18	
Control Test	0.079	02:18	
Air Blank	0.000	02:18	
Subject Sample	# 1	02:22	No Sample Provided
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Air Blank	0.000	02:24	
Subject Sample	# 2	02:27	No Sample Provided
Breath Volume	0.000	02:27	
Air Blank	0.000	02:28	
Control Test	0.078	02:28	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 02/03/2020 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRAZEE,COLE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/15/2020 03:37:35

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.082	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.182	03:32
Breath Volume	1.101	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.167	03:36
Breath Volume	1.550	03:36
Air Blank	0.000	03:36
Control Test	0.081	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 01/16/2020 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: STARNES,JACOB
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 01/16/2020 01:39:14

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.082	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.154	01:34
Breath Volume	1.402	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.157	01:37
Breath Volume	1.109	01:37
Air Blank	0.000	01:38
Control Test	0.081	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 01/16/2020 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CROCK,NICOLLE
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 01/19/2020 03:47:59

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.081	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.140	03:43
Breath Volume	1.312	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.142	03:46
Breath Volume	3.214	03:46
Air Blank	0.000	03:47
Control Test	0.081	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 01/23/2020 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS, DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: NUCETTE VIDES

DATE/TIME OF TEST: 01/19/2020 05:18:22

Diagnostics Check	OK	05:11
Air Blank	0.000	05:12
Control Test	0.082	05:12
Air Blank	0.000	05:13
Subject Sample # 1	0.184	05:13
Breath Volume	4.332	05:13
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.185	05:16
Breath Volume	3.167	05:16
Air Blank	0.000	05:17
Control Test	0.081	05:17
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 01/23/2020 15:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GUEVARA, JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PARHAM

DATE/TIME OF TEST: 01/20/2020 02:36:19

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.081	02:34	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:35	
Control Test	0.081	02:35	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 01/23/2020 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORALES,JOSE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: RODEN

DATE/TIME OF TEST: 01/22/2020 19:15:13

Diagnostics Check	OK	19:08
Air Blank	0.000	19:09
Control Test	0.082	19:09
Air Blank	0.000	19:10
Subject Sample # 1	0.237	19:10
Breath Volume	2.097	19:10
Air Blank	0.000	19:11
Air Blank	0.000	19:13
Subject Sample # 2	0.238	19:13
Breath Volume	3.046	19:13
Air Blank	0.000	19:14
Control Test	0.080	19:14
Air Blank	0.000	19:15
Diagnostics Check	OK	19:15

Data Download Date/Time: 01/23/2020 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SLACK,JOHNIE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/25/2020 05:24:47

Diagnostics Check	OK	05:18
Air Blank	0.000	05:18
Control Test	0.082	05:19
Air Blank	0.000	05:19
Subject Sample # 1	0.083	05:20
Breath Volume	2.562	05:20
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.082	05:23
Breath Volume	2.449	05:23
Air Blank	0.000	05:23
Control Test	0.082	05:24
Air Blank	0.000	05:24
Diagnostics Check	OK	05:24

Data Download Date/Time: 01/30/2020 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WHITMORE, AARON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 01/25/2020 08:05:06

Diagnostics Check	OK	07:58
Air Blank	0.000	07:59
Control Test	0.080	07:59
Air Blank	0.000	08:00
Subject Sample # 1	0.175	08:00
Breath Volume	1.582	08:00
Air Blank	0.000	08:01
Air Blank	0.000	08:03
Subject Sample # 2	0.185	08:03
Breath Volume	1.457	08:03
Air Blank	0.000	08:04
Control Test	0.080	08:04
Air Blank	0.000	08:05
Diagnostics Check	OK	08:05

Data Download Date/Time: 01/30/2020 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TUCK,NATHAN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: HARSHMAN

DATE/TIME OF TEST: 01/26/2020 03:25:37

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.082	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.189	03:20
Breath Volume	1.816	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.205	03:24
Breath Volume	3.910	03:24
Air Blank	0.000	03:24
Control Test	0.081	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 01/30/2020 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LANE, KEATON
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 01/26/2020 04:05:09

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.082	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.196	04:00
Breath Volume	2.144	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.191	04:03
Breath Volume	1.835	04:03
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 01/30/2020 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LEE,MIKEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: MARREROVAZQUEZ

DATE/TIME OF TEST: 01/26/2020 05:05:15

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.082	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.176	05:00
Breath Volume	1.246	05:00
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.177	05:03
Breath Volume	1.828	05:03
Air Blank	0.000	05:04
Control Test	0.081	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 01/30/2020 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DE NIJS,SAMIRA
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: RODRIGUEZ VAZQUEZ

DATE/TIME OF TEST: 01/26/2020 22:37:58

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.082	22:32
Air Blank	0.000	22:32
Subject Sample # 1	0.197	22:33
Breath Volume	2.066	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.200	22:36
Breath Volume	1.265	22:36
Air Blank	0.000	22:37
Control Test	0.082	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 01/30/2020 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORRIS,STEVE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 01/29/2020 04:09:31

Diagnostics Check		OK	03:57	
Air Blank		0.000	03:57	
Control Test		0.082	03:58	
Air Blank		0.000	03:58	
Subject Sample	# 1	0.212	04:01	Volume Not Met
Breath Volume		0.132	04:01	
Air Blank		0.000	04:02	
Air Blank		0.000	04:04	
Subject Sample	# 2	0.240	04:04	
Breath Volume		1.183	04:04	
Air Blank		0.000	04:05	
Air Blank		0.000	04:07	
Subject Sample	# 3	0.226	04:08	
Breath Volume		1.355	04:08	
Air Blank		0.000	04:08	
Control Test		0.080	04:08	
Air Blank		0.000	04:09	
Diagnostics Check		OK	04:09	

Data Download Date/Time: 01/30/2020 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BODE,JOSHUA
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: APONTE ROSARIO

DATE/TIME OF TEST: 01/31/2020 02:34:19

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:23	
Control Test	0.082	02:23	
Air Blank	0.000	02:24	
Subject Sample	# 1	02:27	No Sample Provided
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	
Air Blank	0.000	02:29	
Subject Sample	# 2	02:32	No Sample Provided
Breath Volume	0.000	02:32	
Air Blank	0.000	02:33	
Control Test	0.082	02:33	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 02/06/2020 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARWOOD,ARMANDO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: DEL RIO

DATE/TIME OF TEST: 01/11/2020 02:38:31

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.141	02:33
Breath Volume	2.167	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.131	02:37
Breath Volume	1.640	02:37
Air Blank	0.000	02:37
Control Test	0.079	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 01/16/2020 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ERICKSON, STEPHEN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: MADL

DATE/TIME OF TEST: 01/11/2020 04:07:49

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.079	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.057	04:03
Breath Volume	2.378	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:05
Subject Sample # 2	0.054	04:06
Breath Volume	2.097	04:06
Air Blank	0.000	04:06
Control Test	0.078	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 01/16/2020 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HENDRICKS,BROOKE
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MILLARD

DATE/TIME OF TEST: 01/12/2020 04:36:36

Diagnostics Check		OK	04:24	
Air Blank		0.000	04:24	
Control Test		0.079	04:25	
Air Blank		0.000	04:25	
Subject Sample	# 1	0.040	04:28	Volume Not Met
Breath Volume		0.492	04:28	
Air Blank		0.000	04:29	
Air Blank		0.000	04:31	
Subject Sample	# 2	0.055	04:31	
Breath Volume		2.066	04:31	
Air Blank		0.000	04:32	
Air Blank		0.000	04:34	
Subject Sample	# 3	0.053	04:35	
Breath Volume		2.253	04:35	
Air Blank		0.000	04:35	
Control Test		0.079	04:36	
Air Blank		0.000	04:36	
Diagnostics Check		OK	04:36	

Data Download Date/Time: 01/16/2020 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: VAILLANCOURT

DATE/TIME OF TEST: 01/15/2020 22:26:06

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:20	
Control Test	0.079	22:20	
Air Blank	0.000	22:21	
Subject Sample # 1	0.171	22:22	
Breath Volume	2.570	22:22	
Air Blank	0.000	22:23	
Air Blank	0.000	22:25	
Subject Sample # 2	0.160	22:25	
Breath Volume	1.421	22:25	
Air Blank		22:25	Improper Sample
Air Blank	0.000	22:26	

Data Download Date/Time: 01/16/2020 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: VAILLANCOURT

DATE/TIME OF TEST: 01/16/2020 00:07:09

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.079	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.138	00:02
Breath Volume		2.597	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.138	00:05
Breath Volume		2.265	00:05
Air Blank		0.000	00:06
Control Test		0.079	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 01/16/2020 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOMINGUEZ,JAYMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: DRAKE

DATE/TIME OF TEST: 01/16/2020 02:28:45

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.079	02:18	
Air Blank	0.000	02:18	
Subject Sample	# 1	02:21	No Sample Provided
Breath Volume	0.000	02:21	
Air Blank	0.000	02:22	
Air Blank	0.000	02:24	
Subject Sample	# 2	02:27	No Sample Provided
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	
Control Test	0.079	02:28	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 01/16/2020 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOMINGUEZ,JAYMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: ZARATEGONZALEZ

DATE/TIME OF TEST: 01/16/2020 21:36:06

Diagnostics Check	OK	21:30
Air Blank	0.000	21:30
Control Test	0.079	21:30
Air Blank	0.000	21:31
Subject Sample # 1	0.046	21:31
Breath Volume	2.472	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.048	21:34
Breath Volume	3.109	21:34
Air Blank	0.000	21:35
Control Test	0.079	21:35
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 03/12/2020 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,BAYRON,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: GUZMAN,VICTOR
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 12/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 07:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: LINAREZ MARTINEZ

DATE/TIME OF TEST: 01/01/2020 08:14:09

Diagnostics Check	OK	08:06
Air Blank	0.000	08:06
Control Test	0.081	08:07
Air Blank	0.000	08:07
Subject Sample # 1	0.141	08:09
Breath Volume	2.074	08:09
Air Blank	0.000	08:10
Air Blank	0.000	08:12
Subject Sample # 2	0.132	08:12
Breath Volume	1.785	08:12
Air Blank	0.000	08:13
Control Test	0.077	08:13
Air Blank	0.000	08:14
Diagnostics Check	OK	08:14

Data Download Date/Time: 01/03/2020 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUVRES,MICHEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: DOUVRES,MICHAEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 12/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: HOLMSTROM

DATE/TIME OF TEST: 01/01/2020 20:54:56

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.080	20:48
Air Blank	0.000	20:49
Subject Sample # 1	0.242	20:49
Breath Volume	2.250	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.234	20:53
Breath Volume	2.816	20:53
Air Blank	0.000	20:54
Control Test	0.080	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 01/03/2020 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRODTMAN,ERIC,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: STRODTMAN,ERIC
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: JAESCHKE II

DATE/TIME OF TEST: 01/01/2020 23:58:53

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.079	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.162	23:54
Breath Volume	1.480	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:56
Subject Sample # 2	0.154	23:57
Breath Volume	2.046	23:57
Air Blank	0.000	23:57
Control Test	0.078	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 01/03/2020 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HELLER,MICHAEL
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: COSME-FIGUEROA

DATE/TIME OF TEST: 01/03/2020 19:02:29

Diagnostics Check	OK	18:55
Air Blank	0.000	18:55
Control Test	0.080	18:55
Air Blank	0.000	18:56
Subject Sample # 1	0.161	18:56
Breath Volume	1.421	18:56
Air Blank	0.000	18:57
Air Blank	0.000	18:59
Subject Sample # 2	0.157	19:00
Breath Volume	2.050	19:00
Air Blank	0.000	19:01
Control Test	0.078	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

Data Download Date/Time: 01/24/2020 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 01/05/2020 00:35:38

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.081	00:29
Air Blank	0.000	00:30
Subject Sample # 1	0.176	00:30
Breath Volume	2.882	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.167	00:34
Breath Volume	2.523	00:34
Air Blank	0.000	00:34
Control Test	0.078	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 01/24/2020 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: KOFFINAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PALACIO

DATE/TIME OF TEST: 01/05/2020 02:37:05

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.147	02:32
Breath Volume	1.886	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.151	02:35
Breath Volume	1.894	02:35
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:37

Data Download Date/Time: 01/24/2020 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: FERNANDEZ SOTO

DATE/TIME OF TEST: 01/06/2020 01:37:02

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.081	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.166	01:32
Breath Volume	2.648	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:34
Subject Sample # 2	0.163	01:35
Breath Volume	3.093	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:37

Data Download Date/Time: 01/24/2020 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: DEMERY,ANTHONY
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: DECENA

DATE/TIME OF TEST: 01/06/2020 02:39:01

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.131	02:34
Breath Volume	1.859	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample # 2	0.132	02:37
Breath Volume	2.363	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:39

Data Download Date/Time: 01/24/2020 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRODTMAN,ERIC,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SANCHEZ,ROBERTO
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: DOE 3099

DATE/TIME OF TEST: 01/08/2020 01:53:26

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.081	01:50	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:52	
Control Test	0.080	01:52	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 01/24/2020 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 01/11/2020 00:22:03

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.000	00:17
Breath Volume	3.171	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.000	00:20
Breath Volume	2.914	00:20
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:22

Data Download Date/Time: 01/24/2020 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRODTMAN,ERIC,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SHEROTSKI,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: VIDAL-HANE

DATE/TIME OF TEST: 01/11/2020 22:31:52

Diagnostics Check	OK	22:28	
Air Blank	0.000	22:28	
Control Test	0.080	22:29	
Air Blank	0.000	22:29	
Subject Sample	# 1	22:30	Subject Test Refused
Breath Volume	0.000	22:30	
Air Blank	0.000	22:30	
Control Test	0.080	22:31	
Air Blank	0.000	22:31	
Diagnostics Check	OK	22:31	

Data Download Date/Time: 01/24/2020 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAQUERO,DANNY,H
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BAQUERO,DANNY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: CHISHOLM

DATE/TIME OF TEST: 01/12/2020 05:03:22

Diagnostics Check		OK	04:56
Air Blank		0.000	04:57
Control Test		0.080	04:57
Air Blank		0.000	04:58
Subject Sample	# 1	0.188	04:58
Breath Volume		1.960	04:58
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.192	05:01
Breath Volume		2.367	05:01
Air Blank		0.000	05:02
Control Test		0.079	05:02
Air Blank		0.000	05:03
Diagnostics Check		OK	05:03

Data Download Date/Time: 01/24/2020 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ALFRED,N
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WILLIAMS,ALFRED
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:56

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 01/12/2020 07:28:46

Diagnostics Check	OK	07:21
Air Blank	0.000	07:22
Control Test	0.081	07:22
Air Blank	0.000	07:23
Subject Sample # 1	0.000	07:24
Breath Volume	3.199	07:24
Air Blank	0.000	07:24
Air Blank	0.000	07:26
Subject Sample # 2	0.000	07:27
Breath Volume	3.402	07:27
Air Blank	0.000	07:27
Control Test	0.080	07:28
Air Blank	0.000	07:28
Diagnostics Check	OK	07:28

Data Download Date/Time: 01/24/2020 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: PERKOVICH

DATE/TIME OF TEST: 01/12/2020 23:58:00

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.081	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.000	23:53
Breath Volume	1.464	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.000	23:56
Breath Volume	1.753	23:56
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:58

Data Download Date/Time: 01/24/2020 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CASTILLO CONTREARAS

DATE/TIME OF TEST: 01/18/2020 03:46:46

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.080	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.252	03:42
Breath Volume		2.628	03:42
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.239	03:45
Breath Volume		2.226	03:45
Air Blank		0.000	03:45
Control Test		0.079	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 01/24/2020 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: OLKA,MATTHEW
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: SYED

DATE/TIME OF TEST: 01/18/2020 04:55:01

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.079	04:49
Air Blank	0.000	04:50
Subject Sample # 1	0.000	04:50
Breath Volume	3.648	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.000	04:53
Breath Volume	3.070	04:53
Air Blank	0.000	04:54
Control Test	0.080	04:54
Air Blank	0.000	04:54
Diagnostics Check	OK	04:55

Data Download Date/Time: 01/24/2020 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONTRERAS,BRYANT,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CONTRERAS,BRYANT
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 10:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: ALICEA

DATE/TIME OF TEST: 01/19/2020 11:08:39

Diagnostics Check	OK	11:01
Air Blank	0.000	11:02
Control Test	0.081	11:02
Air Blank	0.000	11:02
Subject Sample # 1	0.000	11:03
Breath Volume	4.214	11:03
Air Blank	0.000	11:04
Air Blank	0.000	11:06
Subject Sample # 2	0.000	11:07
Breath Volume	3.367	11:07
Air Blank	0.000	11:07
Control Test	0.079	11:08
Air Blank	0.000	11:08
Diagnostics Check	OK	11:08

Data Download Date/Time: 01/24/2020 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAQUERO,DANNY,H
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BAQUERO,DANNY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/20/2020 01:14:53

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.080	01:08
Air Blank		0.000	01:08
Subject Sample	# 1	0.189	01:10
Breath Volume		1.660	01:10
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.196	01:13
Breath Volume		3.089	01:13
Air Blank		0.000	01:13
Control Test		0.079	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 01/24/2020 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: POULIN

DATE/TIME OF TEST: 01/23/2020 00:22:46

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.149	00:17
Breath Volume	2.308	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.150	00:21
Breath Volume	2.546	00:21
Air Blank	0.000	00:21
Control Test	0.079	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 01/24/2020 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HERNANDEZ,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 01/11/2020 04:06:45

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.078	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.163	04:01
Breath Volume	2.378	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.164	04:05
Breath Volume	1.484	04:05
Air Blank	0.000	04:05
Control Test	0.078	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 01/24/2020 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,EDWIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: COLON,EDWIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: VARGAS CRUZ

DATE/TIME OF TEST: 01/12/2020 05:54:59

Diagnostics Check		OK	05:48
Air Blank		0.000	05:48
Control Test		0.080	05:48
Air Blank		0.000	05:49
Subject Sample	# 1	0.107	05:50
Breath Volume		1.917	05:50
Air Blank		0.000	05:50
Air Blank		0.000	05:52
Subject Sample	# 2	0.110	05:53
Breath Volume		3.566	05:53
Air Blank		0.000	05:54
Control Test		0.079	05:54
Air Blank		0.000	05:54
Diagnostics Check		OK	05:54

Data Download Date/Time: 01/24/2020 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,RIGOBERTO,
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ESCOBAR,RIGOBERTO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 17:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: SCHOFIELD

DATE/TIME OF TEST: 01/31/2020 18:44:53

Diagnostics Check	OK	18:38	
Air Blank	0.000	18:38	
Control Test	0.079	18:38	
Air Blank	0.000	18:39	
Subject Sample # 1	0.117	18:40	
Breath Volume	1.597	18:40	
Air Blank	0.000	18:41	
Air Blank	0.000	18:42	
Subject Sample # 2	0.109	18:43	
Breath Volume	1.687	18:43	
Air Blank	0.000	18:44	
Control Test	0.073	18:44	Control Outside Tolerance
Air Blank	0.000	18:44	

Data Download Date/Time: 02/07/2020 19:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,T,M
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: SMITH,J
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: STRIETER

DATE/TIME OF TEST: 01/01/2020 02:58:26

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.081	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.183	02:54
Breath Volume	3.984	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.176	02:56
Breath Volume	2.632	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 01/29/2020 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTMAN, MATTHEW, R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: GIVANS, GARRETH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: COLBY

DATE/TIME OF TEST: 01/02/2020 22:36:46

Diagnostics Check	OK	22:29
Air Blank	0.000	22:30
Control Test	0.080	22:30
Air Blank	0.000	22:30
Subject Sample # 1	0.295	22:31
Breath Volume	1.503	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.314	22:35
Breath Volume	1.328	22:35
Air Blank	0.000	22:35
Control Test	0.080	22:36
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 01/29/2020 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTMAN, MATTHEW, R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: INGERSOLL, NICHOLAS
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 01/06/2020 03:33:15

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.080	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.134	03:26
Breath Volume	1.886	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.129	03:31
Breath Volume	1.695	03:31
Air Blank	0.000	03:32
Control Test	0.081	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 01/29/2020 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,JASON,M
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: WRIGHT,TYLER
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 12/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 01/25/2020 01:39:54

Diagnostics Check		OK	01:33
Air Blank		0.000	01:34
Control Test		0.082	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.236	01:35
Breath Volume		2.679	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:37
Subject Sample	# 2	0.229	01:38
Breath Volume		1.273	01:38
Air Blank		0.000	01:38
Control Test		0.081	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 02/13/2020 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSON,JASON,A
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: HANSON,JASON
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:03

SUBJECT'S LAST NAME: ARISMENDY

DATE/TIME OF TEST: 01/31/2020 14:58:37

Diagnostics Check		OK	14:45	
Air Blank		0.000	14:45	
Control Test		0.081	14:45	
Air Blank		0.000	14:46	
Subject Sample	# 1	0.236	14:49	Volume Not Met
Breath Volume		0.542	14:49	
Air Blank		0.000	14:50	
Air Blank		0.000	14:51	
Subject Sample	# 2	0.285	14:52	No .020 agreement
Breath Volume		1.902	14:52	
Air Blank		0.000	14:53	
Air Blank		0.000	14:55	
Subject Sample	# 3	0.240	14:56	No .020 agreement
Breath Volume		1.968	14:56	
Air Blank		0.000	14:57	
Control Test		0.079	14:57	
Air Blank		0.000	14:58	
Diagnostics Check		OK	14:58	

Data Download Date/Time: 02/13/2020 16:30