

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITSON,GEOFFREY,B  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: JONES,JONATHAN  
ARRESTING OFFICER AGENCY: Deland PD  
LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: DESFORGES

DATE/TIME OF TEST: 11/01/2020 03:18:07

---

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:15	
Control Test	0.080	03:15	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:16	Subject Test Refused
Breath Volume	0.000	03:16	
Air Blank	0.000	03:17	
Control Test	0.080	03:17	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

---

Data Download Date/Time: 11/16/2020 16:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITSON,GEOFFREY,B  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: PENA,ENMANUEL  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: HERNANDEZ-RIVERA

DATE/TIME OF TEST: 11/01/2020 04:44:36

---

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.079	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.168	04:38
Breath Volume	2.378	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:40
Subject Sample # 2	0.155	04:43
Breath Volume	3.179	04:43
Air Blank	0.000	04:43
Control Test	0.081	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

---

Data Download Date/Time: 11/16/2020 16:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LICKISS,BRANT,C  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: LICKISS,BRANT  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: LAUDA

DATE/TIME OF TEST: 11/08/2020 00:03:32

---

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.080	23:58
Air Blank	0.000	23:58
Subject Sample # 2	0.019	23:59
Breath Volume	2.316	23:59
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.019	00:02
Breath Volume	3.386	00:02
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

---

Data Download Date/Time: 11/16/2020 16:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LICKISS,BRANT,C  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: JOHNSON,DEJHAN  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: GONZALEZ DE JESUS

DATE/TIME OF TEST: 11/26/2020 00:54:21

---

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:43	
Control Test	0.080	00:43	
Air Blank	0.000	00:44	
Subject Sample	# 1	00:47	No Sample Provided
Breath Volume	0.000	00:47	
Air Blank	0.000	00:47	
Air Blank	0.000	00:49	
Subject Sample	# 2	00:52	No Sample Provided
Breath Volume	0.000	00:52	
Air Blank	0.000	00:53	
Control Test	0.080	00:53	
Air Blank	0.000	00:54	
Diagnostics Check	OK	00:54	

---

Data Download Date/Time: 12/07/2020 18:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LICKISS,BRANT,C  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: JOHNSON,DEJHAN  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: GONZALEZ DE JESUS

DATE/TIME OF TEST: 11/26/2020 01:16:00

---

Diagnostics Check		OK	01:07
Air Blank		0.000	01:08
Control Test		0.080	01:08
Air Blank		0.000	01:08
Subject Sample	# 1	0.134	01:09
Breath Volume		1.476	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.127	01:14
Breath Volume		1.558	01:14
Air Blank		0.000	01:15
Control Test		0.080	01:15
Air Blank		0.000	01:15
Diagnostics Check		OK	01:16

---

Data Download Date/Time: 12/07/2020 18:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LICKISS,BRANT,C  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: RIEDINGER,RYAN  
ARRESTING OFFICER AGENCY: Deland PD  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 11/26/2020 03:22:55

---

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:20	
Control Test	0.080	03:20	
Air Blank	0.000	03:21	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:22	
Control Test	0.079	03:22	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

---

Data Download Date/Time: 12/07/2020 18:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITSON,GEOFFREY,B  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: MENDOZA,LUCHAY  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: TISDALE

DATE/TIME OF TEST: 11/28/2020 02:31:24

---

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.067	02:26
Breath Volume	1.839	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.068	02:29
Breath Volume	1.917	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

---

Data Download Date/Time: 12/07/2020 18:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELLIS,ALEXANDER,  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: ELLIS,ALEXANDER  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:16

SUBJECT'S LAST NAME: DEVINE

DATE/TIME OF TEST: 11/04/2020 16:46:04

---

Diagnostics Check	OK	16:39
Air Blank	0.000	16:40
Control Test	0.081	16:40
Air Blank	0.000	16:41
Subject Sample # 1	0.328	16:41
Breath Volume	3.691	16:41
Air Blank	0.000	16:42
Air Blank	0.000	16:44
Subject Sample # 2	0.338	16:44
Breath Volume	3.199	16:44
Air Blank	0.000	16:45
Control Test	0.081	16:45
Air Blank	0.000	16:45
Diagnostics Check	OK	16:46

---

Data Download Date/Time: 11/12/2020 10:03



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRUSCIO,J,J  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: TRUSCIO,J  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 11:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:18

SUBJECT'S LAST NAME: WEGING

DATE/TIME OF TEST: 11/13/2020 12:45:25

---

Diagnostics Check	OK	12:39
Air Blank	0.000	12:39
Control Test	0.081	12:39
Air Blank	0.000	12:40
Subject Sample # 1	0.040	12:40
Breath Volume	2.000	12:40
Air Blank	0.000	12:41
Air Blank	0.000	12:43
Subject Sample # 2	0.040	12:43
Breath Volume	1.781	12:43
Air Blank	0.000	12:44
Control Test	0.081	12:44
Air Blank	0.000	12:45
Diagnostics Check	OK	12:45

---

Data Download Date/Time: 12/07/2020 10:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: MANGOLD,K  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: WILES

DATE/TIME OF TEST: 11/13/2020 17:06:55

---

Diagnostics Check	OK	17:03	
Air Blank	0.000	17:03	
Control Test	0.081	17:04	
Air Blank	0.000	17:04	
Subject Sample	# 1	17:05	Subject Test Refused
Breath Volume	0.000	17:05	
Air Blank	0.000	17:06	
Control Test	0.081	17:06	
Air Blank	0.000	17:06	
Diagnostics Check	OK	17:06	

---

Data Download Date/Time: 12/07/2020 10:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRUSCIO,JESSE,J  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: TRUSCIO,JESSE  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:09

SUBJECT'S LAST NAME: KHAKHAR

DATE/TIME OF TEST: 11/19/2020 14:37:52

---

Diagnostics Check	OK	14:30
Air Blank	0.000	14:31
Control Test	0.082	14:31
Air Blank	0.000	14:31
Subject Sample # 1	0.337	14:32
Breath Volume	2.156	14:32
Air Blank	0.000	14:33
Air Blank	0.000	14:35
Subject Sample # 2	0.338	14:36
Breath Volume	1.906	14:36
Air Blank	0.000	14:36
Control Test	0.081	14:37
Air Blank	0.000	14:37
Diagnostics Check	OK	14:37

---

Data Download Date/Time: 12/07/2020 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELLIS,ALEXANDER,  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: ELLIS,ALEXANDER  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 11/26/2020 00:52:24

---

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.082	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.191	00:47
Breath Volume	1.902	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.200	00:50
Breath Volume	1.765	00:50
Air Blank	0.000	00:51
Control Test	0.082	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

---

Data Download Date/Time: 12/07/2020 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: PRESTON,MIKAYLA  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 11/29/2020 17:59:58

---

Diagnostics Check	OK	17:48	
Air Blank	0.000	17:49	
Control Test	0.082	17:49	
Air Blank	0.000	17:49	
Subject Sample	# 1 0.251	17:52	Volume Not Met
Breath Volume	0.488	17:52	
Air Blank	0.000	17:54	
Air Blank	0.000	17:55	
Subject Sample	# 2	17:58	No Sample Provided
Breath Volume	0.000	17:58	
Air Blank	0.000	17:59	
Control Test	0.078	17:59	
Air Blank	0.000	17:59	
Diagnostics Check	OK	17:59	

---

Data Download Date/Time: 12/07/2020 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER, MARVILLE, H  
BREATH TEST OPERATOR AGENCY: Daytona Beach PD  
ARRESTING OFFICER: TUCKER, MARVILLE  
ARRESTING OFFICER AGENCY: Daytona Beach PD  
LAST AGENCY INSPECTION: 10/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/08/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: O'DONNELL

DATE/TIME OF TEST: 11/11/2020 19:42:44

---

Diagnostics Check	OK	19:36
Air Blank	0.000	19:36
Control Test	0.081	19:37
Air Blank	0.000	19:37
Subject Sample # 1	0.000	19:38
Breath Volume	3.308	19:38
Air Blank	0.000	19:39
Air Blank	0.000	19:41
Subject Sample # 2	0.000	19:41
Breath Volume	2.066	19:41
Air Blank	0.000	19:41
Control Test	0.080	19:42
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

---

Data Download Date/Time: 11/30/2020 12:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,STERLING,R  
BREATH TEST OPERATOR AGENCY: Daytona Beach PD  
ARRESTING OFFICER: DEMNER,AUTUMN  
ARRESTING OFFICER AGENCY: Daytona Beach PD  
LAST AGENCY INSPECTION: 10/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/08/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/15/2020 00:06:01

---

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:52	
Control Test	0.081	23:52	
Air Blank	0.000	23:53	
Subject Sample	# 3 0.160	23:54	
Breath Volume	1.597	23:54	
Air Blank	0.000	23:55	
Air Blank	0.000	23:57	
Subject Sample	# 1 0.159	00:00	Volume Not Met
Breath Volume	0.843	00:00	
Air Blank	0.000	00:01	
Air Blank	0.000	00:02	
Subject Sample	# 2 0.153	00:04	
Breath Volume	2.097	00:04	
Air Blank	0.000	00:05	
Control Test	0.077	00:05	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:06	

---

Data Download Date/Time: 11/30/2020 12:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUNNE,JUSTIN,R  
BREATH TEST OPERATOR AGENCY: Daytona Beach PD  
ARRESTING OFFICER: DUNNE,JUSTIN  
ARRESTING OFFICER AGENCY: Daytona Beach PD  
LAST AGENCY INSPECTION: 10/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/08/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: ORREGO

DATE/TIME OF TEST: 11/20/2020 18:57:11

---

Diagnostics Check	OK	18:50
Air Blank	0.000	18:51
Control Test	0.081	18:51
Air Blank	0.000	18:52
Subject Sample # 1	0.000	18:52
Breath Volume	1.441	18:52
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.000	18:55
Breath Volume	1.425	18:55
Air Blank	0.000	18:56
Control Test	0.080	18:56
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

---

Data Download Date/Time: 11/30/2020 12:33



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MIRINO,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: Deland PD  
ARRESTING OFFICER: JAMES,ROYCE  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: MCCLAUGHRY

DATE/TIME OF TEST: 11/14/2020 23:21:43

---

Diagnostics Check		OK	23:15
Air Blank		0.000	23:16
Control Test		0.079	23:16
Air Blank		0.000	23:16
Subject Sample	# 1	0.249	23:17
Breath Volume		1.664	23:17
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.251	23:20
Breath Volume		1.781	23:20
Air Blank		0.000	23:20
Control Test		0.079	23:21
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

---

Data Download Date/Time: 11/30/2020 19:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALTER,CRAIG,K  
BREATH TEST OPERATOR AGENCY: Deland PD  
ARRESTING OFFICER: JOHNSON,ROMAN  
ARRESTING OFFICER AGENCY: Deland PD  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 11/21/2020 19:20:46

---

Diagnostics Check	OK	19:17	
Air Blank	0.000	19:18	
Control Test	0.079	19:18	
Air Blank	0.000	19:19	
Subject Sample	# 1	19:19	Subject Test Refused
Breath Volume	0.000	19:19	
Air Blank	0.000	19:19	
Control Test	0.080	19:20	
Air Blank	0.000	19:20	
Diagnostics Check	OK	19:20	

---

Data Download Date/Time: 11/30/2020 19:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MIRINO, MICHAEL, A

BREATH TEST OPERATOR AGENCY: Deland PD

ARRESTING OFFICER: DUMONT, TRAVIS

ARRESTING OFFICER AGENCY: Deland PD

LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 970670

DRY GAS EXPIRY DATE: 06/07/2021

VIOLATION CODE: DUI

ARREST TIME: 00:57

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 11/24/2020 01:44:52

---

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:42	
Control Test	0.080	01:42	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:43	
Control Test	0.080	01:44	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

---

Data Download Date/Time: 11/30/2020 19:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MIRINO,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: Deland PD  
ARRESTING OFFICER: JAMES,ROYCE  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: BAIR

DATE/TIME OF TEST: 11/24/2020 21:37:27

---

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.079	21:32
Air Blank	0.000	21:32
Subject Sample # 1	0.243	21:33
Breath Volume	1.562	21:33
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.262	21:35
Breath Volume	2.109	21:35
Air Blank	0.000	21:36
Control Test	0.080	21:36
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

---

Data Download Date/Time: 11/30/2020 19:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MIRINO,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: Deland PD  
ARRESTING OFFICER: JAMES,ROYCE  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 11/28/2020 02:27:07

---

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:24	
Control Test	0.080	02:24	
Air Blank	0.000	02:25	
Subject Sample	# 1	02:25	Subject Test Refused
Breath Volume	0.000	02:25	
Air Blank	0.000	02:26	
Control Test	0.080	02:26	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:27	

---

Data Download Date/Time: 11/30/2020 19:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MIRINO,MICHAEL,A

BREATH TEST OPERATOR AGENCY: Deland PD

ARRESTING OFFICER: JONES,JOHN

ARRESTING OFFICER AGENCY: Deland PD

LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 970670

DRY GAS EXPIRY DATE: 06/07/2021

VIOLATION CODE: DUI

ARREST TIME: 02:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: WATERS

DATE/TIME OF TEST: 11/28/2020 03:41:55

---

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.181	03:37
Breath Volume	2.457	03:37
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.182	03:40
Breath Volume	2.636	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

---

Data Download Date/Time: 11/30/2020 19:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAUN,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: SPINELLI,MIKE  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 10:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:24

SUBJECT'S LAST NAME: WILLARD

DATE/TIME OF TEST: 11/06/2020 11:47:31

---

Diagnostics Check	OK	11:44	
Air Blank	0.000	11:45	
Control Test	0.080	11:45	
Air Blank	0.000	11:45	
Subject Sample	# 1	11:46	Subject Test Refused
Breath Volume	0.000	11:46	
Air Blank	0.000	11:46	
Control Test	0.079	11:46	
Air Blank	0.000	11:47	
Diagnostics Check	OK	11:47	

---

Data Download Date/Time: 11/27/2020 17:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FAIRCLOTH, EDWARD, W  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: GREEN, SETH  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 11/06/2020 18:54:20

---

Diagnostics Check	OK	18:47
Air Blank	0.000	18:48
Control Test	0.080	18:48
Air Blank	0.000	18:49
Subject Sample # 1	0.000	18:49
Breath Volume	2.000	18:49
Air Blank	0.000	18:50
Air Blank	0.000	18:52
Subject Sample # 2	0.000	18:52
Breath Volume	2.925	18:52
Air Blank	0.000	18:53
Control Test	0.079	18:53
Air Blank	0.000	18:54
Diagnostics Check	OK	18:54

---

Data Download Date/Time: 11/27/2020 17:53



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAUN,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: WICHMAN,MICHELLE  
ARRESTING OFFICER AGENCY: Bunnell PD  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 10:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:47

SUBJECT'S LAST NAME: PREZZA

DATE/TIME OF TEST: 11/07/2020 11:14:49

---

Diagnostics Check		OK	11:08
Air Blank		0.000	11:08
Control Test		0.080	11:09
Air Blank		0.000	11:09
Subject Sample	# 1	0.269	11:10
Breath Volume		3.707	11:10
Air Blank		0.000	11:11
Air Blank		0.000	11:12
Subject Sample	# 2	0.277	11:13
Breath Volume		2.253	11:13
Air Blank		0.000	11:13
Control Test		0.079	11:14
Air Blank		0.000	11:14
Diagnostics Check		OK	11:14

---

Data Download Date/Time: 11/27/2020 17:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TWING,BRANDON,R  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: MORGANTE,T  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 11/08/2020 01:40:15

---

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.156	01:35
Breath Volume	1.769	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:37
Subject Sample # 2	0.165	01:38
Breath Volume	1.269	01:38
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

---

Data Download Date/Time: 11/27/2020 17:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCARTHY, EVAN, V  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: BARNETT, JOSEPH  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 11/10/2020 17:10:39

---

Diagnostics Check		OK	17:02
Air Blank		0.000	17:03
Control Test		0.080	17:03
Air Blank		0.000	17:04
Subject Sample	# 1	0.255	17:05
Breath Volume		2.535	17:05
Air Blank		0.000	17:05
Air Blank		0.000	17:07
Subject Sample	# 2	0.265	17:09
Breath Volume		2.382	17:09
Air Blank		0.000	17:09
Control Test		0.079	17:10
Air Blank		0.000	17:10
Diagnostics Check		OK	17:10

---

Data Download Date/Time: 11/27/2020 17:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEYER,JEFFREY,D  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: BIFANO,A  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/13/2020 00:59:00

---

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:51	
Control Test	0.080	00:51	
Air Blank	0.000	00:51	
Subject Sample	# 1	00:55	No Sample Provided
Breath Volume	0.000	00:55	
Air Blank	0.000	00:55	
Air Blank	0.000	00:57	
Subject Sample	# 2	00:57	Subject Test Refused
Breath Volume	0.000	00:57	
Air Blank	0.000	00:58	
Control Test	0.080	00:58	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:59	

---

Data Download Date/Time: 11/27/2020 17:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NICOLAS,MITA,  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: SCHERR,S  
ARRESTING OFFICER AGENCY: Flagler Beach PD  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: BROMAGEN

DATE/TIME OF TEST: 11/14/2020 19:14:57

---

Diagnostics Check		OK	19:07
Air Blank		0.000	19:07
Control Test		0.080	19:08
Air Blank		0.000	19:08
Subject Sample	# 1	0.198	19:09
Breath Volume		1.277	19:09
Air Blank		0.000	19:10
Air Blank		0.000	19:11
Subject Sample	# 2	0.192	19:13
Breath Volume		1.382	19:13
Air Blank		0.000	19:14
Control Test		0.080	19:14
Air Blank		0.000	19:14
Diagnostics Check		OK	19:14

---

Data Download Date/Time: 11/27/2020 17:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCARTHY, EVAN, V  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: SCHERR, EVAN  
ARRESTING OFFICER AGENCY: Flagler Beach PD  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: CZARNECKI

DATE/TIME OF TEST: 11/15/2020 17:57:27

---

Diagnostics Check	OK	17:53	
Air Blank	0.000	17:53	
Control Test	0.080	17:53	
Air Blank	0.000	17:54	
Subject Sample	# 1	17:56	Subject Test Refused
Breath Volume	0.000	17:56	
Air Blank	0.000	17:56	
Control Test	0.079	17:56	
Air Blank	0.000	17:57	
Diagnostics Check	OK	17:57	

---

Data Download Date/Time: 11/27/2020 17:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AUGONE,JOSEPH,A  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: BEANE,B  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: AUGONE

DATE/TIME OF TEST: 11/19/2020 14:03:01

---

Diagnostics Check	OK	13:55
Air Blank	0.000	13:55
Control Test	0.079	13:55
Air Blank	0.000	13:56
Subject Sample # 1	0.000	13:58
Breath Volume	2.078	13:58
Air Blank	0.000	13:58
Air Blank	0.000	14:00
Subject Sample # 2	0.000	14:01
Breath Volume	1.382	14:01
Air Blank	0.000	14:02
Control Test	0.079	14:02
Air Blank	0.000	14:02
Diagnostics Check	OK	14:03

---

Data Download Date/Time: 11/27/2020 17:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON,SHANICE,J  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: SCHERR,E  
ARRESTING OFFICER AGENCY: Flagler Beach PD  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 11/19/2020 15:11:00

---

Diagnostics Check	OK	15:04
Air Blank	0.000	15:04
Control Test	0.080	15:05
Air Blank	0.000	15:05
Subject Sample # 1	0.081	15:06
Breath Volume	2.972	15:06
Air Blank	0.000	15:06
Air Blank	0.000	15:08
Subject Sample # 2	0.077	15:09
Breath Volume	2.664	15:09
Air Blank	0.000	15:10
Control Test	0.079	15:10
Air Blank	0.000	15:10
Diagnostics Check	OK	15:11

---

Data Download Date/Time: 11/27/2020 17:53



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: SCHERR,EVAN  
ARRESTING OFFICER AGENCY: Flagler Beach PD  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: BOYLSTON

DATE/TIME OF TEST: 11/23/2020 22:05:39

---

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:03	
Control Test	0.080	22:03	
Air Blank	0.000	22:04	
Subject Sample	# 1	22:04	Subject Test Refused
Breath Volume	0.000	22:04	
Air Blank	0.000	22:04	
Control Test	0.079	22:05	
Air Blank	0.000	22:05	
Diagnostics Check	OK	22:05	

---

Data Download Date/Time: 11/27/2020 17:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NICOLAS,MITA,  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: LOGSDON,D  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: COMBS

DATE/TIME OF TEST: 11/24/2020 02:36:10

---

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.080	02:34	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:35	
Control Test	0.080	02:35	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

---

Data Download Date/Time: 11/27/2020 17:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NICOLAS,MITA,  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: SCHERR,E  
ARRESTING OFFICER AGENCY: Flagler Beach PD  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 11/25/2020 02:47:25

---

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.186	02:42
Breath Volume	1.511	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.199	02:45
Breath Volume	1.898	02:45
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

---

Data Download Date/Time: 11/27/2020 17:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STOGDON,BRAD,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: STOGDON,BRAD  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 11/26/2020 05:06:23

---

Diagnostics Check	OK	05:03	
Air Blank	0.000	05:04	
Control Test	0.080	05:04	
Air Blank	0.000	05:04	
Subject Sample	# 1	05:05	Subject Test Refused
Breath Volume	0.000	05:05	
Air Blank	0.000	05:05	
Control Test	0.080	05:05	
Air Blank	0.000	05:06	
Diagnostics Check	OK	05:06	

---

Data Download Date/Time: 11/27/2020 17:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: MANKA,KYLE  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 11/27/2020 02:09:35

---

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:07	
Control Test	0.080	02:07	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:08	Subject Test Refused
Breath Volume	0.000	02:08	
Air Blank	0.000	02:08	
Control Test	0.080	02:08	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

---

Data Download Date/Time: 11/27/2020 17:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001149

INSTRUMENT REGISTERED TO: LAKE HELEN P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SCOTT,F  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: RIDEOUT,TIMOTHY  
ARRESTING OFFICER AGENCY: Deland PD  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: KANIA

DATE/TIME OF TEST: 11/21/2020 03:40:04

---

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.185	03:35
Breath Volume	3.039	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.193	03:38
Breath Volume	2.906	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:40

---

Data Download Date/Time: 12/02/2020 11:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO  
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: HEATH

DATE/TIME OF TEST: 11/02/2020 00:48:05

---

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:45	
Control Test	0.080	00:45	
Air Blank	0.000	00:46	
Subject Sample	# 1	00:46	Subject Test Refused
Breath Volume	0.000	00:46	
Air Blank	0.000	00:47	
Control Test	0.079	00:47	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:48	

---

Data Download Date/Time: 11/23/2020 10:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO  
ARRESTING OFFICER: TERRERO,FRAICOR  
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO  
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: REYNA

DATE/TIME OF TEST: 11/02/2020 03:01:50

---

Diagnostics Check	OK	03:01	
Air Blank		03:01	RFI Detect
Air Blank	0.000	03:01	

---

Data Download Date/Time: 11/23/2020 10:10



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO  
ARRESTING OFFICER: TERRERO,FRAICOR  
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO  
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: REYNA

DATE/TIME OF TEST: 11/02/2020 03:10:39

---

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.081	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.167	03:06
Breath Volume	3.835	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.169	03:08
Breath Volume	3.535	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

---

Data Download Date/Time: 11/23/2020 10:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO  
ARRESTING OFFICER: TERRERO,FRAICOR  
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO  
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: MATUTE SALAZAR

DATE/TIME OF TEST: 11/13/2020 23:48:15

---

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.079	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.138	23:43
Breath Volume	3.968	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.141	23:46
Breath Volume	2.347	23:46
Air Blank	0.000	23:47
Control Test	0.079	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

---

Data Download Date/Time: 11/23/2020 10:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLINE, GLENN,  
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO  
ARRESTING OFFICER: TEST, CLINE  
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 13:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:25

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 11/24/2020 13:56:32

---

Diagnostics Check	OK	13:50
Air Blank	0.000	13:51
Control Test	0.081	13:51
Air Blank	0.000	13:52
Subject Sample # 1	0.000	13:52
Breath Volume	1.527	13:52
Air Blank	0.000	13:52
Air Blank	0.000	13:54
Subject Sample # 2	0.000	13:55
Breath Volume	1.609	13:55
Air Blank	0.000	13:55
Control Test	0.081	13:55
Air Blank	0.000	13:56
Diagnostics Check	OK	13:56

---

Data Download Date/Time: 12/15/2020 07:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP TROOP F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP TROOP F  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: FERNANDEZ ESPINO

DATE/TIME OF TEST: 11/28/2020 02:14:09

---

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.075	02:09
Breath Volume	3.605	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.072	02:12
Breath Volume	2.820	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

---

Data Download Date/Time: 12/15/2020 07:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP TROOP F  
ARRESTING OFFICER: KOPE,MICHAEL  
ARRESTING OFFICER AGENCY: FHP TROOP F  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CIONI

DATE/TIME OF TEST: 11/29/2020 02:00:22

---

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:54
Subject Sample # 1	0.182	01:55
Breath Volume	3.957	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.169	01:58
Breath Volume	2.191	01:58
Air Blank	0.000	01:59
Control Test	0.078	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

---

Data Download Date/Time: 12/15/2020 07:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP TROOP F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP TROOP F  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: DIETER

DATE/TIME OF TEST: 11/29/2020 23:31:53

---

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.026	23:26
Breath Volume	1.351	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.029	23:30
Breath Volume	2.253	23:30
Air Blank	0.000	23:30
Control Test	0.079	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

---

Data Download Date/Time: 12/15/2020 07:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: BRUNER,SEAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 10/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 10:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: RAEDER

DATE/TIME OF TEST: 11/02/2020 11:17:26

---

Diagnostics Check		OK	11:11
Air Blank		0.000	11:11
Control Test		0.082	11:12
Air Blank		0.000	11:12
Subject Sample	# 1	0.000	11:13
Breath Volume		2.210	11:13
Air Blank		0.000	11:13
Air Blank		0.000	11:15
Subject Sample	# 2	0.000	11:16
Breath Volume		2.289	11:16
Air Blank		0.000	11:16
Control Test		0.081	11:16
Air Blank		0.000	11:17
Diagnostics Check		OK	11:17

---

Data Download Date/Time: 11/22/2020 03:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: BRUNER,SEAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 10/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 14:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 11/02/2020 15:19:04

---

Diagnostics Check	OK	15:13
Air Blank	0.000	15:13
Control Test	0.082	15:13
Air Blank	0.000	15:14
Subject Sample # 1	0.000	15:14
Breath Volume	1.628	15:14
Air Blank	0.000	15:15
Air Blank	0.000	15:17
Subject Sample # 2	0.000	15:17
Breath Volume	2.019	15:17
Air Blank	0.000	15:18
Control Test	0.082	15:18
Air Blank	0.000	15:18
Diagnostics Check	OK	15:19

---

Data Download Date/Time: 11/22/2020 03:20



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: BRUNER,SEAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 10/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 10:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: JACINTHE

DATE/TIME OF TEST: 11/03/2020 10:38:13

---

Diagnostics Check	OK	10:32
Air Blank	0.000	10:32
Control Test	0.082	10:33
Air Blank	0.000	10:33
Subject Sample # 1	0.000	10:33
Breath Volume	1.414	10:33
Air Blank	0.000	10:34
Air Blank	0.000	10:36
Subject Sample # 2	0.000	10:36
Breath Volume	1.441	10:36
Air Blank	0.000	10:37
Control Test	0.082	10:37
Air Blank	0.000	10:38
Diagnostics Check	OK	10:38

---

Data Download Date/Time: 11/22/2020 03:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: BRUNER,SEAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 10/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: ZIEGLER

DATE/TIME OF TEST: 11/15/2020 03:30:38

---

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.081	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.212	03:26
Breath Volume	2.589	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.203	03:29
Breath Volume	4.402	03:29
Air Blank	0.000	03:29
Control Test	0.082	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

---

Data Download Date/Time: 11/22/2020 03:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WEBER, RAY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 10/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 08:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 11/16/2020 08:54:56

---

Diagnostics Check	OK	08:49
Air Blank	0.000	08:49
Control Test	0.082	08:49
Air Blank	0.000	08:50
Subject Sample # 1	0.000	08:50
Breath Volume	2.375	08:50
Air Blank	0.000	08:51
Air Blank	0.000	08:53
Subject Sample # 2	0.000	08:53
Breath Volume	2.433	08:53
Air Blank	0.000	08:53
Control Test	0.081	08:54
Air Blank	0.000	08:54
Diagnostics Check	OK	08:54

---

Data Download Date/Time: 11/22/2020 03:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: JACKSON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 08:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: MANZARI

DATE/TIME OF TEST: 11/17/2020 09:13:04

---

Diagnostics Check	OK	09:06
Air Blank	0.000	09:07
Control Test	0.082	09:07
Air Blank	0.000	09:08
Subject Sample # 1	0.000	09:08
Breath Volume	2.812	09:08
Air Blank	0.000	09:09
Air Blank	0.000	09:11
Subject Sample # 2	0.000	09:11
Breath Volume	2.828	09:11
Air Blank	0.000	09:12
Control Test	0.082	09:12
Air Blank	0.000	09:12
Diagnostics Check	OK	09:13

---

Data Download Date/Time: 11/22/2020 03:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: JACKSON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: BUCCI

DATE/TIME OF TEST: 11/18/2020 17:27:22

---

Diagnostics Check	OK	17:21
Air Blank	0.000	17:21
Control Test	0.083	17:21
Air Blank	0.000	17:22
Subject Sample # 1	0.117	17:22
Breath Volume	2.003	17:22
Air Blank	0.000	17:23
Air Blank	0.000	17:25
Subject Sample # 2	0.107	17:25
Breath Volume	1.671	17:25
Air Blank	0.000	17:26
Control Test	0.082	17:26
Air Blank	0.000	17:27
Diagnostics Check	OK	17:27

---

Data Download Date/Time: 11/22/2020 03:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,STEVEN,J  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: EVANS,STEVEN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 11/22/2020 03:28:13

---

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.082	03:26	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:27	
Control Test	0.082	03:27	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

---

Data Download Date/Time: 12/21/2020 14:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: JACKSON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 11:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: HAYNES

DATE/TIME OF TEST: 11/23/2020 11:33:51

---

Diagnostics Check	OK	11:27
Air Blank	0.000	11:28
Control Test	0.082	11:28
Air Blank	0.000	11:29
Subject Sample # 1	0.000	11:29
Breath Volume	1.316	11:29
Air Blank	0.000	11:30
Air Blank	0.000	11:32
Subject Sample # 2	0.000	11:32
Breath Volume	1.605	11:32
Air Blank	0.000	11:32
Control Test	0.082	11:33
Air Blank	0.000	11:33
Diagnostics Check	OK	11:33

---

Data Download Date/Time: 12/21/2020 14:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WEBER, RAY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 16:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/23/2020 16:50:10

---

Diagnostics Check	OK	16:44
Air Blank	0.000	16:44
Control Test	0.082	16:45
Air Blank	0.000	16:45
Subject Sample # 1	0.000	16:45
Breath Volume	1.234	16:45
Air Blank	0.000	16:46
Air Blank	0.000	16:48
Subject Sample # 2	0.000	16:48
Breath Volume	1.363	16:48
Air Blank	0.000	16:49
Control Test	0.082	16:49
Air Blank	0.000	16:50
Diagnostics Check	OK	16:50

---

Data Download Date/Time: 12/21/2020 14:18



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WEBER, RAY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 08:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 11/24/2020 08:39:15

---

Diagnostics Check	OK	08:33
Air Blank	0.000	08:33
Control Test	0.082	08:34
Air Blank	0.000	08:34
Subject Sample # 1	0.000	08:35
Breath Volume	3.292	08:35
Air Blank	0.000	08:35
Air Blank	0.000	08:37
Subject Sample # 2	0.000	08:37
Breath Volume	3.828	08:37
Air Blank	0.000	08:38
Control Test	0.081	08:38
Air Blank	0.000	08:39
Diagnostics Check	OK	08:39

---

Data Download Date/Time: 12/21/2020 14:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WEBER, RAY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 09:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: DUKE

DATE/TIME OF TEST: 11/24/2020 09:37:51

---

Diagnostics Check		OK	09:32
Air Blank		0.000	09:32
Control Test		0.082	09:32
Air Blank		0.000	09:33
Subject Sample	# 1	0.000	09:33
Breath Volume		2.492	09:33
Air Blank		0.000	09:34
Air Blank		0.000	09:36
Subject Sample	# 2	0.000	09:36
Breath Volume		2.601	09:36
Air Blank		0.000	09:36
Control Test		0.080	09:37
Air Blank		0.000	09:37
Diagnostics Check		OK	09:37

---

Data Download Date/Time: 12/21/2020 14:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WEBER, RAY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 09:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: MILLAN

DATE/TIME OF TEST: 11/24/2020 09:53:51

---

Diagnostics Check	OK	09:47
Air Blank	0.000	09:48
Control Test	0.081	09:48
Air Blank	0.000	09:49
Subject Sample # 1	0.000	09:49
Breath Volume	1.906	09:49
Air Blank	0.000	09:50
Air Blank	0.000	09:52
Subject Sample # 2	0.000	09:52
Breath Volume	1.937	09:52
Air Blank	0.000	09:52
Control Test	0.081	09:53
Air Blank	0.000	09:53
Diagnostics Check	OK	09:53

---

Data Download Date/Time: 12/21/2020 14:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WEBER, RAY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 10:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: DICKS

DATE/TIME OF TEST: 11/24/2020 10:41:49

---

Diagnostics Check	OK	10:35
Air Blank	0.000	10:36
Control Test	0.082	10:36
Air Blank	0.000	10:37
Subject Sample # 1	0.000	10:37
Breath Volume	1.570	10:37
Air Blank	0.000	10:38
Air Blank	0.000	10:40
Subject Sample # 2	0.000	10:40
Breath Volume	1.617	10:40
Air Blank	0.000	10:40
Control Test	0.081	10:41
Air Blank	0.000	10:41
Diagnostics Check	OK	10:41

---

Data Download Date/Time: 12/21/2020 14:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WEBER, RAY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 11:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: HAMLIN

DATE/TIME OF TEST: 11/24/2020 11:40:11

---

Diagnostics Check	OK	11:34
Air Blank	0.000	11:34
Control Test	0.081	11:35
Air Blank	0.000	11:35
Subject Sample # 1	0.000	11:35
Breath Volume	2.246	11:35
Air Blank	0.000	11:36
Air Blank	0.000	11:38
Subject Sample # 2	0.000	11:38
Breath Volume	2.945	11:38
Air Blank	0.000	11:39
Control Test	0.081	11:39
Air Blank	0.000	11:40
Diagnostics Check	OK	11:40

---

Data Download Date/Time: 12/21/2020 14:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WEBER, RAY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 13:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: JACINTHE

DATE/TIME OF TEST: 11/24/2020 13:35:11

---

Diagnostics Check	OK	13:29
Air Blank	0.000	13:29
Control Test	0.081	13:30
Air Blank	0.000	13:30
Subject Sample # 1	0.000	13:30
Breath Volume	1.558	13:30
Air Blank	0.000	13:31
Air Blank	0.000	13:33
Subject Sample # 2	0.000	13:33
Breath Volume	1.589	13:33
Air Blank	0.000	13:34
Control Test	0.081	13:34
Air Blank	0.000	13:35
Diagnostics Check	OK	13:35

---

Data Download Date/Time: 12/21/2020 14:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WEBER, RAY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 13:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: GARZA

DATE/TIME OF TEST: 11/24/2020 13:50:10

---

Diagnostics Check	OK	13:44
Air Blank	0.000	13:44
Control Test	0.082	13:45
Air Blank	0.000	13:45
Subject Sample # 1	0.000	13:45
Breath Volume	3.074	13:45
Air Blank	0.000	13:46
Air Blank	0.000	13:48
Subject Sample # 2	0.000	13:48
Breath Volume	3.558	13:48
Air Blank	0.000	13:49
Control Test	0.080	13:49
Air Blank	0.000	13:50
Diagnostics Check	OK	13:50

---

Data Download Date/Time: 12/21/2020 14:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WEBER, RAY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 13:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 11/24/2020 14:10:18

---

Diagnostics Check	OK	14:04
Air Blank	0.000	14:04
Control Test	0.081	14:05
Air Blank	0.000	14:05
Subject Sample # 1	0.000	14:06
Breath Volume	1.558	14:06
Air Blank	0.000	14:06
Air Blank	0.000	14:08
Subject Sample # 2	0.000	14:08
Breath Volume	1.464	14:08
Air Blank	0.000	14:09
Control Test	0.081	14:09
Air Blank	0.000	14:10
Diagnostics Check	OK	14:10

---

Data Download Date/Time: 12/21/2020 14:19



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WEBER, RAY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 14:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 11/24/2020 15:08:42

---

Diagnostics Check	OK	15:06	
Air Blank	0.000	15:07	
Control Test	0.081	15:07	
Air Blank		15:08	Purge Fail
Air Blank	0.000	15:08	

---

Data Download Date/Time: 12/21/2020 14:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WEBER, RAY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 14:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 11/24/2020 15:16:51

---

Diagnostics Check	OK	15:11
Air Blank	0.000	15:11
Control Test	0.081	15:11
Air Blank	0.000	15:12
Subject Sample # 1	0.000	15:12
Breath Volume	1.894	15:12
Air Blank	0.000	15:13
Air Blank	0.000	15:15
Subject Sample # 2	0.000	15:15
Breath Volume	2.472	15:15
Air Blank	0.000	15:15
Control Test	0.080	15:16
Air Blank	0.000	15:16
Diagnostics Check	OK	15:16

---

Data Download Date/Time: 12/21/2020 14:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WEBER, RAY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 15:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 11/24/2020 15:38:25

---

Diagnostics Check	OK	15:32
Air Blank	0.000	15:33
Control Test	0.082	15:33
Air Blank	0.000	15:33
Subject Sample # 1	0.000	15:34
Breath Volume	1.636	15:34
Air Blank	0.000	15:34
Air Blank	0.000	15:36
Subject Sample # 2	0.000	15:37
Breath Volume	1.480	15:37
Air Blank	0.000	15:37
Control Test	0.081	15:37
Air Blank	0.000	15:38
Diagnostics Check	OK	15:38

---

Data Download Date/Time: 12/21/2020 14:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WEBER, RAY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 15:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: HOKE

DATE/TIME OF TEST: 11/24/2020 15:53:28

---

Diagnostics Check	OK	15:47
Air Blank	0.000	15:47
Control Test	0.082	15:48
Air Blank	0.000	15:48
Subject Sample # 1	0.000	15:49
Breath Volume	1.867	15:49
Air Blank	0.000	15:49
Air Blank	0.000	15:51
Subject Sample # 2	0.000	15:52
Breath Volume	1.886	15:52
Air Blank	0.000	15:52
Control Test	0.081	15:52
Air Blank	0.000	15:53
Diagnostics Check	OK	15:53

---

Data Download Date/Time: 12/21/2020 14:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WEBER, RAY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 15:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: RAEDER

DATE/TIME OF TEST: 11/24/2020 16:10:20

---

Diagnostics Check	OK	16:04
Air Blank	0.000	16:04
Control Test	0.083	16:05
Air Blank	0.000	16:05
Subject Sample # 1	0.000	16:06
Breath Volume	2.175	16:06
Air Blank	0.000	16:06
Air Blank	0.000	16:08
Subject Sample # 2	0.000	16:08
Breath Volume	2.175	16:08
Air Blank	0.000	16:09
Control Test	0.081	16:09
Air Blank	0.000	16:10
Diagnostics Check	OK	16:10

---

Data Download Date/Time: 12/21/2020 14:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: BRUNER,SEAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 11/25/2020 01:29:09

---

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.082	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.095	01:24
Breath Volume	1.539	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:26
Subject Sample # 2	0.091	01:27
Breath Volume	1.347	01:27
Air Blank	0.000	01:28
Control Test	0.082	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

---

Data Download Date/Time: 12/21/2020 14:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: JACKSON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 07:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/25/2020 08:15:14

---

Diagnostics Check	OK	08:09
Air Blank	0.000	08:09
Control Test	0.082	08:09
Air Blank	0.000	08:10
Subject Sample # 1	0.000	08:10
Breath Volume	1.804	08:10
Air Blank	0.000	08:11
Air Blank	0.000	08:13
Subject Sample # 2	0.000	08:13
Breath Volume	1.843	08:13
Air Blank	0.000	08:14
Control Test	0.080	08:14
Air Blank	0.000	08:15
Diagnostics Check	OK	08:15

---

Data Download Date/Time: 12/21/2020 14:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: JACKSON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 08:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 11/25/2020 09:06:00

---

Diagnostics Check	OK	08:59
Air Blank	0.000	09:00
Control Test	0.082	09:00
Air Blank	0.000	09:01
Subject Sample # 1	0.000	09:01
Breath Volume	1.691	09:01
Air Blank	0.000	09:02
Air Blank	0.000	09:04
Subject Sample # 2	0.000	09:04
Breath Volume	1.121	09:04
Air Blank	0.000	09:05
Control Test	0.081	09:05
Air Blank	0.000	09:05
Diagnostics Check	OK	09:06

---

Data Download Date/Time: 12/21/2020 14:19



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: BRUNER,SEAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/29/2020 01:07:28

---

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.082	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.189	01:02
Breath Volume		1.871	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.180	01:05
Breath Volume		1.316	01:05
Air Blank		0.000	01:06
Control Test		0.082	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

---

Data Download Date/Time: 12/21/2020 14:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: SAFIEDDINE,JAMAL  
ARRESTING OFFICER AGENCY: Davenport PD  
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: FERRER MELENDEZ

DATE/TIME OF TEST: 11/29/2020 22:32:48

---

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.083	22:27
Air Blank	0.000	22:27
Subject Sample # 1	0.058	22:28
Breath Volume	1.558	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:30
Subject Sample # 2	0.058	22:31
Breath Volume	1.589	22:31
Air Blank	0.000	22:31
Control Test	0.081	22:32
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

---

Data Download Date/Time: 12/21/2020 14:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENSON,DREW,M  
BREATH TEST OPERATOR AGENCY: Eglin AFB PD  
ARRESTING OFFICER: HARVEY,JEREMIAH  
ARRESTING OFFICER AGENCY: Eglin AFB PD  
LAST AGENCY INSPECTION: 10/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1283438  
DRY GAS EXPIRY DATE: 02/25/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: JAMIESON

DATE/TIME OF TEST: 11/01/2020 05:07:02

---

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:59	
Control Test	0.082	04:59	
Air Blank	0.000	04:59	
Subject Sample	# 1 0.222	05:01	
Breath Volume	1.535	05:01	
Air Blank	0.000	05:02	
Air Blank	0.000	05:03	
Subject Sample	# 2 0.230	05:05	
Breath Volume	1.917	05:05	
Air Blank	0.000	05:06	
Control Test	0.082	05:06	
Air Blank		05:06	RFI Detect
Air Blank	0.000	05:07	

---

Data Download Date/Time: 12/01/2020 06:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HENSON,DREW,M

BREATH TEST OPERATOR AGENCY: Eglin AFB PD

ARRESTING OFFICER: HARVEY,JEREMIAH

ARRESTING OFFICER AGENCY: Eglin AFB PD

LAST AGENCY INSPECTION: 10/05/2020

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 1283438

DRY GAS EXPIRY DATE: 02/25/2023

VIOLATION CODE: DUI

ARREST TIME: 03:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: JAMIESON

DATE/TIME OF TEST: 11/01/2020 05:30:43

---

Diagnostics Check	OK	05:29	
Air Blank		05:30	RFI Detect
Air Blank	0.000	05:30	

---

Data Download Date/Time: 12/01/2020 06:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HENSON,DREW,M

BREATH TEST OPERATOR AGENCY: Eglin AFB PD

ARRESTING OFFICER: HARVEY,JEREMIAH

ARRESTING OFFICER AGENCY: Eglin AFB PD

LAST AGENCY INSPECTION: 10/05/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1283438

DRY GAS EXPIRY DATE: 02/25/2023

VIOLATION CODE: DUI

ARREST TIME: 03:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: JAMIESON

DATE/TIME OF TEST: 11/01/2020 05:33:36

---

Diagnostics Check	OK	05:33	
Air Blank		05:33	RFI Detect
Air Blank		05:33	RFI Detect

---

Data Download Date/Time: 12/01/2020 06:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENSON,DREW,M  
BREATH TEST OPERATOR AGENCY: Eglin AFB PD  
ARRESTING OFFICER: HARVEY,JEREMIAH  
ARRESTING OFFICER AGENCY: Eglin AFB PD  
LAST AGENCY INSPECTION: 10/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1283438  
DRY GAS EXPIRY DATE: 02/25/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: JAMIESON

DATE/TIME OF TEST: 11/01/2020 05:45:19

---

Diagnostics Check	OK	05:38
Air Blank	0.000	05:39
Control Test	0.082	05:39
Air Blank	0.000	05:39
Subject Sample # 1	0.218	05:40
Breath Volume	1.550	05:40
Air Blank	0.000	05:41
Air Blank	0.000	05:42
Subject Sample # 2	0.215	05:43
Breath Volume	1.332	05:43
Air Blank	0.000	05:44
Control Test	0.082	05:44
Air Blank	0.000	05:45
Diagnostics Check	OK	05:45

---

Data Download Date/Time: 12/01/2020 06:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENSON,DREW,M  
BREATH TEST OPERATOR AGENCY: Eglin AFB PD  
ARRESTING OFFICER: BYLICKI,DANIEL  
ARRESTING OFFICER AGENCY: Eglin AFB PD  
LAST AGENCY INSPECTION: 10/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1283438  
DRY GAS EXPIRY DATE: 02/25/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 11/01/2020 21:36:07

---

Diagnostics Check	OK	21:31	
Air Blank	0.000	21:32	
Control Test	0.083	21:32	
Air Blank	0.000	21:32	
Subject Sample	# 1	21:34	Subject Test Refused
Breath Volume	0.000	21:34	
Air Blank	0.000	21:35	
Control Test	0.083	21:35	
Air Blank	0.000	21:36	
Diagnostics Check	OK	21:36	

---

Data Download Date/Time: 12/01/2020 06:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MILLER,RAYMOND,S

BREATH TEST OPERATOR AGENCY: Eglin AFB PD

ARRESTING OFFICER: BYLICKI,DANIEL

ARRESTING OFFICER AGENCY: Eglin AFB PD

LAST AGENCY INSPECTION: 10/05/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1283438

DRY GAS EXPIRY DATE: 02/25/2023

VIOLATION CODE: DUI

ARREST TIME: 19:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/04/2020 20:33:59

---

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:30	
Control Test	0.083	20:30	
Air Blank	0.000	20:31	
Subject Sample	# 1	20:32	Subject Test Refused
Breath Volume	0.000	20:32	
Air Blank	0.000	20:33	
Control Test	0.082	20:33	
Air Blank	0.000	20:33	
Diagnostics Check	OK	20:33	

---

Data Download Date/Time: 12/01/2020 06:03



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OGLE,SHAWN,T  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: KAVANAUGH,ANTHONY  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: LESSLEY

DATE/TIME OF TEST: 11/13/2020 19:23:38

---

Diagnostics Check	OK	19:16
Air Blank	0.000	19:17
Control Test	0.080	19:17
Air Blank	0.000	19:18
Subject Sample # 1	0.220	19:19
Breath Volume	1.398	19:19
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 2	0.226	19:22
Breath Volume	1.855	19:22
Air Blank	0.000	19:22
Control Test	0.079	19:23
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

---

Data Download Date/Time: 12/13/2020 11:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OGLE,SHAWN,T  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: MYERS,KEVIN  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 11/14/2020 00:06:59

---

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.081	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.039	00:02
Breath Volume	1.273	00:02
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.039	00:05
Breath Volume	1.441	00:05
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

---

Data Download Date/Time: 12/13/2020 11:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OGLE,SHAWN,T  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: OGLE,SHAWN  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:43

SUBJECT'S LAST NAME: BENSINK

DATE/TIME OF TEST: 11/14/2020 07:13:02

---

Diagnostics Check		OK	07:06
Air Blank		0.000	07:06
Control Test		0.080	07:06
Air Blank		0.000	07:07
Subject Sample	# 1	0.153	07:08
Breath Volume		2.226	07:08
Air Blank		0.000	07:08
Air Blank		0.000	07:10
Subject Sample	# 2	0.154	07:11
Breath Volume		2.984	07:11
Air Blank		0.000	07:12
Control Test		0.081	07:12
Air Blank		0.000	07:12
Diagnostics Check		OK	07:13

---

Data Download Date/Time: 12/13/2020 11:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEARSON,ROBERT,B  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: SCOTT,KARI  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:51

SUBJECT'S LAST NAME: TUTONI

DATE/TIME OF TEST: 11/22/2020 07:20:10

---

Diagnostics Check	OK	07:17	
Air Blank	0.000	07:17	
Control Test	0.081	07:18	
Air Blank	0.000	07:18	
Subject Sample	# 1	07:18	Subject Test Refused
Breath Volume	0.000	07:18	
Air Blank	0.000	07:19	
Control Test	0.081	07:19	
Air Blank	0.000	07:20	
Diagnostics Check	OK	07:20	

---

Data Download Date/Time: 12/13/2020 11:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,RICHARD,L  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: OGLE,SHAWN  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: WHIPPY

DATE/TIME OF TEST: 11/23/2020 20:55:48

---

Diagnostics Check	OK	20:53	
Air Blank	0.000	20:53	
Control Test	0.081	20:53	
Air Blank	0.000	20:54	
Subject Sample	# 1	20:54	Subject Test Refused
Breath Volume	0.000	20:54	
Air Blank	0.000	20:54	
Control Test	0.082	20:55	
Air Blank	0.000	20:55	
Diagnostics Check	OK	20:55	

---

Data Download Date/Time: 12/13/2020 11:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,RICHARD,L  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: OGLE,SHAWN  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: TOUPS

DATE/TIME OF TEST: 11/27/2020 23:33:42

---

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:31	
Control Test	0.081	23:31	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:32	Subject Test Refused
Breath Volume	0.000	23:32	
Air Blank	0.000	23:32	
Control Test	0.080	23:33	
Air Blank	0.000	23:33	
Diagnostics Check	OK	23:33	

---

Data Download Date/Time: 12/13/2020 11:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OGLE,SHAWN,T  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: BAKAYSA,MICHAEL  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: MOSELEY

DATE/TIME OF TEST: 11/30/2020 05:44:51

---

Diagnostics Check	OK	05:38
Air Blank	0.000	05:39
Control Test	0.080	05:39
Air Blank	0.000	05:39
Subject Sample # 1	0.000	05:40
Breath Volume	3.437	05:40
Air Blank	0.000	05:41
Air Blank	0.000	05:43
Subject Sample # 2	0.000	05:43
Breath Volume	2.968	05:43
Air Blank	0.000	05:43
Control Test	0.081	05:44
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

---

Data Download Date/Time: 12/13/2020 11:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,JUSTIN,A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: ADAMS,JUSTIN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 11/09/2020 19:29:37

---

Diagnostics Check		OK	19:20
Air Blank		0.000	19:21
Control Test		0.082	19:21
Air Blank		0.000	19:22
Subject Sample	# 1	0.209	19:24
Breath Volume		1.753	19:24
Air Blank		0.000	19:25
Air Blank		0.000	19:26
Subject Sample	# 2	0.213	19:28
Breath Volume		1.886	19:28
Air Blank		0.000	19:28
Control Test		0.080	19:28
Air Blank		0.000	19:29
Diagnostics Check		OK	19:29

---

Data Download Date/Time: 11/30/2020 20:22



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,JUSTIN,A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: ADAMS,JUSTIN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: CULLER

DATE/TIME OF TEST: 11/28/2020 00:03:04

---

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.081	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.119	23:58
Breath Volume	4.402	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.117	00:01
Breath Volume	3.703	00:01
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:03

---

Data Download Date/Time: 11/30/2020 20:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,JUSTIN,A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: ADAMS,JUSTIN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: PYLE

DATE/TIME OF TEST: 11/28/2020 02:43:54

---

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.175	02:38
Breath Volume	1.945	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.184	02:42
Breath Volume	2.300	02:42
Air Blank	0.000	02:42
Control Test	0.081	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

---

Data Download Date/Time: 11/30/2020 20:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,JUSTIN,A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: ADAMS,JUSTIN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: WALTZ

DATE/TIME OF TEST: 11/29/2020 00:40:51

---

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.306	00:36
Breath Volume	1.871	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.300	00:39
Breath Volume	1.410	00:39
Air Blank	0.000	00:39
Control Test	0.080	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

---

Data Download Date/Time: 11/30/2020 20:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D  
BREATH TEST OPERATOR AGENCY: Clermont PD  
ARRESTING OFFICER: REBANDO, BRIAN  
ARRESTING OFFICER AGENCY: Clermont PD  
LAST AGENCY INSPECTION: 10/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1149031  
DRY GAS EXPIRY DATE: 05/30/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: JIMENEZ MENDEZ

DATE/TIME OF TEST: 11/01/2020 02:53:02

---

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.081	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.096	02:48
Breath Volume	2.062	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.096	02:51
Breath Volume	2.050	02:51
Air Blank	0.000	02:52
Control Test	0.081	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:53

---

Data Download Date/Time: 11/19/2020 12:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DUBI,UZIEL,Y

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: DUBI,UZIEL

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 10/17/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1149031

DRY GAS EXPIRY DATE: 05/30/2022

VIOLATION CODE: DUI

ARREST TIME: 23:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BEVERSTOCK

DATE/TIME OF TEST: 11/07/2020 00:49:42

---

Diagnostics Check		OK	00:42
Air Blank		0.000	00:43
Control Test		0.081	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.172	00:44
Breath Volume		1.324	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.189	00:48
Breath Volume		1.542	00:48
Air Blank		0.000	00:48
Control Test		0.080	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

---

Data Download Date/Time: 11/19/2020 12:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DUBI,UZIEL,Y

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: GUY,KEVIN

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 10/17/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1149031

DRY GAS EXPIRY DATE: 05/30/2022

VIOLATION CODE: DUI

ARREST TIME: 22:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: GILLEY

DATE/TIME OF TEST: 11/07/2020 23:56:18

---

Diagnostics Check	OK	23:49
Air Blank	0.000	23:50
Control Test	0.080	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.202	23:51
Breath Volume	1.949	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.193	23:54
Breath Volume	2.242	23:54
Air Blank	0.000	23:55
Control Test	0.081	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

---

Data Download Date/Time: 11/19/2020 12:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARPER,J,L

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: PEREZ,C

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 10/17/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1149031

DRY GAS EXPIRY DATE: 05/30/2022

VIOLATION CODE: DUI

ARREST TIME: 17:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 11/10/2020 18:31:09

---

Diagnostics Check	OK	18:20
Air Blank	0.000	18:21
Control Test	0.081	18:21
Air Blank	0.000	18:22
Subject Sample # 1	0.227	18:23
Breath Volume	3.207	18:23
Air Blank	0.000	18:24
Air Blank	0.000	18:25
Subject Sample # 2	0.195	18:26
Breath Volume	1.359	18:26
Air Blank	0.000	18:27
Air Blank	0.000	18:28
Subject Sample # 3	0.218	18:29
Breath Volume	2.402	18:29
Air Blank	0.000	18:30
Control Test	0.082	18:30
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

---

Data Download Date/Time: 11/19/2020 12:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARPER,J,L

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: REBANDO,B

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 10/17/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1149031

DRY GAS EXPIRY DATE: 05/30/2022

VIOLATION CODE: DUI

ARREST TIME: 01:46

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 11/16/2020 02:23:44

---

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:21	
Control Test	0.081	02:21	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Control Test	0.081	02:23	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

---

Data Download Date/Time: 11/19/2020 12:47



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,JOSE,G  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: DELGADO,JOSE  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: HESEMEYER

DATE/TIME OF TEST: 11/01/2020 01:39:37

---

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.082	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.212	01:35
Breath Volume		2.257	01:35
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.210	01:38
Breath Volume		1.562	01:38
Air Blank		0.000	01:38
Control Test		0.081	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

---

Data Download Date/Time: 11/16/2020 22:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: ANGERVIL,CLAUDY  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: BAECHER

DATE/TIME OF TEST: 11/01/2020 22:10:07

---

Diagnostics Check	OK	22:03
Air Blank	0.000	22:03
Control Test	0.081	22:03
Air Blank	0.000	22:04
Subject Sample # 1	0.241	22:04
Breath Volume	1.671	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.234	22:08
Breath Volume	1.375	22:08
Air Blank	0.000	22:09
Control Test	0.080	22:09
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

---

Data Download Date/Time: 11/16/2020 22:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PECCI,Z,A  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: PECCI,Z  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: HAYDAK

DATE/TIME OF TEST: 11/03/2020 16:07:19

---

Diagnostics Check		OK	16:00
Air Blank		0.000	16:01
Control Test		0.082	16:01
Air Blank		0.000	16:02
Subject Sample	# 1	0.229	16:02
Breath Volume		2.628	16:02
Air Blank		0.000	16:03
Air Blank		0.000	16:05
Subject Sample	# 2	0.233	16:05
Breath Volume		2.519	16:05
Air Blank		0.000	16:06
Control Test		0.081	16:06
Air Blank		0.000	16:07
Diagnostics Check		OK	16:07

---

Data Download Date/Time: 11/16/2020 22:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: POLIZZI,CHRISTINE  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: PARADISO

DATE/TIME OF TEST: 11/04/2020 00:50:27

---

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.081	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.207	00:45
Breath Volume		1.945	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.206	00:48
Breath Volume		1.671	00:48
Air Blank		0.000	00:49
Control Test		0.081	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

---

Data Download Date/Time: 11/16/2020 22:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: GITLIN,CORY  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ECHEANDIA

DATE/TIME OF TEST: 11/04/2020 02:38:42

---

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.081	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.211	02:33
Breath Volume	3.296	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.228	02:37
Breath Volume	3.000	02:37
Air Blank	0.000	02:37
Control Test	0.081	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

---

Data Download Date/Time: 11/16/2020 22:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: THOMAS,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CHARRI

DATE/TIME OF TEST: 11/07/2020 02:00:28

---

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.082	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.000	01:55
Breath Volume	1.792	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.000	01:59
Breath Volume	3.390	01:59
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

---

Data Download Date/Time: 11/16/2020 22:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,RICKY,  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: PEREZ,RICKY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 11/08/2020 02:26:11

---

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.078	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.157	02:21
Breath Volume	1.539	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.148	02:24
Breath Volume	2.000	02:24
Air Blank	0.000	02:25
Control Test	0.081	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

---

Data Download Date/Time: 11/16/2020 22:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,RICKY,  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: PEREZ,RICKY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 11/14/2020 02:15:17

---

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:07	
Control Test	0.081	02:07	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:10	No Sample Provided
Breath Volume	0.000	02:10	
Air Blank	0.000	02:11	
Air Blank	0.000	02:13	
Subject Sample	# 2	02:13	Subject Test Refused
Breath Volume	0.000	02:13	
Air Blank	0.000	02:14	
Control Test	0.082	02:14	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:15	

---

Data Download Date/Time: 11/16/2020 22:21



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: ANGERVIL,CLAUDY  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: CARRIKER

DATE/TIME OF TEST: 11/14/2020 21:46:39

---

Diagnostics Check	OK	21:40
Air Blank	0.000	21:40
Control Test	0.082	21:40
Air Blank	0.000	21:41
Subject Sample # 1	0.000	21:42
Breath Volume	1.359	21:42
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.000	21:45
Breath Volume	1.292	21:45
Air Blank	0.000	21:45
Control Test	0.081	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

---

Data Download Date/Time: 11/16/2020 22:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUBOIS,BROOKS,E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: DUBOIS,BROOKS  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: HORNIK

DATE/TIME OF TEST: 11/15/2020 20:31:27

---

Diagnostics Check	OK	20:25
Air Blank	0.000	20:25
Control Test	0.082	20:25
Air Blank	0.000	20:26
Subject Sample # 1	0.217	20:26
Breath Volume	2.105	20:26
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.212	20:29
Breath Volume	2.304	20:29
Air Blank	0.000	20:30
Control Test	0.082	20:30
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

---

Data Download Date/Time: 11/16/2020 22:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAY,ERICK,J  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: DAY,ERICK  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 11/17/2020 15:17:09

---

Diagnostics Check	OK	15:11
Air Blank	0.000	15:11
Control Test	0.082	15:11
Air Blank	0.000	15:12
Subject Sample # 1	0.143	15:12
Breath Volume	1.570	15:12
Air Blank	0.000	15:13
Air Blank	0.000	15:15
Subject Sample # 2	0.137	15:15
Breath Volume	1.359	15:15
Air Blank	0.000	15:16
Control Test	0.081	15:16
Air Blank	0.000	15:17
Diagnostics Check	OK	15:17

---

Data Download Date/Time: 12/03/2020 23:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKES,JOHN,D  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: WILKES,JOHN  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: MENDEZ MARTIN

DATE/TIME OF TEST: 11/18/2020 00:37:24

---

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.082	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.107	00:32
Breath Volume	3.628	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.107	00:35
Breath Volume	4.066	00:35
Air Blank	0.000	00:36
Control Test	0.082	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

---

Data Download Date/Time: 12/03/2020 23:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUEDIGER,CHRISTOPHER,L  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: RUEDIGER,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:19

SUBJECT'S LAST NAME: FOLEY

DATE/TIME OF TEST: 11/18/2020 11:51:17

---

Diagnostics Check	OK	11:44
Air Blank	0.000	11:44
Control Test	0.080	11:45
Air Blank	0.000	11:45
Subject Sample # 1	0.276	11:46
Breath Volume	2.828	11:46
Air Blank	0.000	11:46
Air Blank	0.000	11:48
Subject Sample # 2	0.270	11:49
Breath Volume	3.335	11:49
Air Blank	0.000	11:50
Control Test	0.080	11:50
Air Blank	0.000	11:51
Diagnostics Check	OK	11:51

---

Data Download Date/Time: 12/03/2020 23:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: POLIZZI,CHRISTINE  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: DIAZ-HERNANDEZ

DATE/TIME OF TEST: 11/22/2020 00:44:27

---

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.083	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.194	00:39
Breath Volume	2.804	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.194	00:42
Breath Volume	2.632	00:42
Air Blank	0.000	00:43
Control Test	0.082	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

---

Data Download Date/Time: 12/03/2020 23:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA  
MOCTEZUMA,JONATHAN,J  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: SPUGNARDI,TIMOTHY  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021

VIOLATION CODE: DUI  
ARREST TIME: 04:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: RIVERA MOCTEZUMA

DATE/TIME OF TEST: 11/26/2020 05:33:35

---

Diagnostics Check	OK	05:26
Air Blank	0.000	05:27
Control Test	0.081	05:27
Air Blank	0.000	05:28
Subject Sample # 1	0.237	05:29
Breath Volume	4.222	05:29
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample # 2	0.227	05:31
Breath Volume	1.148	05:31
Air Blank	0.000	05:32
Control Test	0.079	05:32
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

---

Data Download Date/Time: 12/03/2020 23:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: POLIZZI,CHRISTINE  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: VALDERRAMA

DATE/TIME OF TEST: 11/27/2020 01:27:28

---

Diagnostics Check		OK	01:19
Air Blank		0.000	01:20
Control Test		0.081	01:20
Air Blank		0.000	01:21
Subject Sample	# 1	0.242	01:22
Breath Volume		2.066	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.253	01:25
Breath Volume		1.925	01:25
Air Blank		0.000	01:26
Control Test		0.079	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

---

Data Download Date/Time: 12/03/2020 23:13



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRADY,REBECCA,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BRADY,REBECCA  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 11:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 11/28/2020 12:52:44

---

Diagnostics Check	OK	12:46
Air Blank	0.000	12:46
Control Test	0.080	12:47
Air Blank	0.000	12:47
Subject Sample # 1	0.263	12:48
Breath Volume	3.250	12:48
Air Blank	0.000	12:48
Air Blank	0.000	12:50
Subject Sample # 2	0.263	12:51
Breath Volume	3.511	12:51
Air Blank	0.000	12:51
Control Test	0.078	12:52
Air Blank	0.000	12:52
Diagnostics Check	OK	12:52

---

Data Download Date/Time: 12/03/2020 23:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANDERS,JAMES,R  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: MCKEON,AMANDA  
ARRESTING OFFICER AGENCY: Jupiter Island  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: KILLGORE

DATE/TIME OF TEST: 11/29/2020 10:40:08

---

Diagnostics Check	OK	10:33	
Air Blank	0.000	10:33	
Control Test	0.082	10:33	
Air Blank	0.000	10:34	
Subject Sample # 1	0.170	10:36	
Breath Volume	3.324	10:36	
Air Blank	0.000	10:36	
Air Blank	0.000	10:38	
		10:38	Improper Sample
Air Blank	0.000	10:39	
Control Test	0.081	10:39	
Air Blank	0.000	10:40	
Diagnostics Check	OK	10:40	

---

Data Download Date/Time: 12/03/2020 23:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANDERS,JAMES,R  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: MCKEON,A  
ARRESTING OFFICER AGENCY: Jupiter Island  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: KILLGORE

DATE/TIME OF TEST: 11/29/2020 11:26:13

---

Diagnostics Check	OK	11:18
Air Blank	0.000	11:18
Control Test	0.081	11:19
Air Blank	0.000	11:19
Subject Sample # 1	0.139	11:20
Breath Volume	1.582	11:20
Air Blank	0.000	11:21
Air Blank	0.000	11:23
Subject Sample # 2	0.142	11:24
Breath Volume	2.476	11:24
Air Blank	0.000	11:25
Control Test	0.080	11:25
Air Blank	0.000	11:26
Diagnostics Check	OK	11:26

---

Data Download Date/Time: 12/03/2020 23:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARAFA,YUSUF,A  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: ARAFA,YUSUF  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 11/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:39

SUBJECT'S LAST NAME: PRESSLEY

DATE/TIME OF TEST: 11/01/2020 18:07:12

---

Diagnostics Check	OK	17:59
Air Blank	0.000	17:59
Control Test	0.080	18:00
Air Blank	0.000	18:00
Subject Sample # 1	0.180	18:01
Breath Volume	1.496	18:01
Air Blank	0.000	18:02
Air Blank	0.000	18:04
Subject Sample # 2	0.184	18:05
Breath Volume	1.929	18:05
Air Blank	0.000	18:06
Control Test	0.080	18:06
Air Blank	0.000	18:07
Diagnostics Check	OK	18:07

---

Data Download Date/Time: 11/03/2020 09:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSS,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 11/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: MAZILE

DATE/TIME OF TEST: 11/04/2020 04:42:00

---

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.082	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.000	04:37
Breath Volume	1.503	04:37
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.000	04:40
Breath Volume	1.312	04:40
Air Blank	0.000	04:41
Control Test	0.082	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:42

---

Data Download Date/Time: 12/01/2020 12:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARAFA,YUSUF,A  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MELCHER,ANDREW  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 11/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: BOADWINE

DATE/TIME OF TEST: 11/05/2020 17:19:57

---

Diagnostics Check		OK	17:10
Air Blank		0.000	17:11
Control Test		0.081	17:11
Air Blank		0.000	17:11
Subject Sample	# 1	0.285	17:12
Breath Volume		2.152	17:12
Air Blank		0.000	17:12
Air Blank		0.000	17:14
Subject Sample	# 2	0.254	17:15
Breath Volume		1.644	17:15
Air Blank		0.000	17:15
Air Blank		0.000	17:17
Subject Sample	# 3	0.265	17:18
Breath Volume		2.605	17:18
Air Blank		0.000	17:19
Control Test		0.079	17:19
Air Blank		0.000	17:19
Diagnostics Check		OK	17:19

---

Data Download Date/Time: 12/01/2020 12:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: WEBER,JEFFREY  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 11/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 11/15/2020 16:57:19

---

Diagnostics Check	OK	16:50
Air Blank	0.000	16:51
Control Test	0.081	16:51
Air Blank	0.000	16:52
Subject Sample # 1	0.180	16:52
Breath Volume	1.972	16:52
Air Blank	0.000	16:53
Air Blank	0.000	16:55
Subject Sample # 2	0.178	16:55
Breath Volume	1.960	16:55
Air Blank	0.000	16:56
Control Test	0.080	16:56
Air Blank	0.000	16:57
Diagnostics Check	OK	16:57

---

Data Download Date/Time: 12/01/2020 12:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WATSON,KEVIN,V  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: WATSON,KEVIN  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 11/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 11/16/2020 00:44:11

---

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:41	
Control Test	0.081	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:42	Subject Test Refused
Breath Volume	0.000	00:42	
Air Blank	0.000	00:43	
Control Test	0.081	00:43	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

---

Data Download Date/Time: 12/01/2020 12:23



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSS,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 11/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: DEVANEY

DATE/TIME OF TEST: 11/22/2020 04:38:06

---

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.080	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.168	04:33
Breath Volume	2.527	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:35
Subject Sample # 2	0.175	04:36
Breath Volume	2.445	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

---

Data Download Date/Time: 12/01/2020 12:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,DARYL,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: TURNER,WILLIAM  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 11/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: KENYON

DATE/TIME OF TEST: 11/22/2020 22:26:48

---

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.081	22:21
Air Blank	0.000	22:21
Subject Sample # 1	0.129	22:22
Breath Volume	3.039	22:22
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.130	22:25
Breath Volume	2.980	22:25
Air Blank	0.000	22:25
Control Test	0.081	22:26
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

---

Data Download Date/Time: 12/01/2020 12:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSS,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 11/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: COUPLAND

DATE/TIME OF TEST: 11/27/2020 00:12:55

---

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.081	00:07	
Air Blank	0.000	00:08	
Subject Sample # 1	0.136	00:09	
Breath Volume	1.957	00:09	
Air Blank	0.000	00:09	
Air Blank	0.000	00:11	
Subject Sample # 2	0.124	00:11	
Breath Volume	1.230	00:11	
Air Blank		00:12	RFI Detect
Air Blank	0.000	00:12	

---

Data Download Date/Time: 12/01/2020 12:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARAFA,YUSUF,A  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: ARAFA,YUSUF  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 11/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: HENRY III

DATE/TIME OF TEST: 11/29/2020 02:24:34

---

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.081	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.205	02:19
Breath Volume	3.644	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.199	02:23
Breath Volume	3.089	02:23
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

---

Data Download Date/Time: 12/01/2020 12:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ABRAHAM,ERIC,D  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: JOLLY,JAMES  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HERRON

DATE/TIME OF TEST: 11/04/2020 03:53:01

---

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.084	03:47
Breath Volume	1.757	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.074	03:51
Breath Volume	1.347	03:51
Air Blank	0.000	03:52
Control Test	0.082	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:53

---

Data Download Date/Time: 11/05/2020 18:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ABRAHAM,ERIC,D  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: ABRAHAM,ERIC  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 17:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 11/06/2020 19:05:11

---

Diagnostics Check	OK	18:56
Air Blank	0.000	18:56
Control Test	0.080	18:57
Air Blank	0.000	18:57
Subject Sample # 1	0.223	18:58
Breath Volume	3.351	18:58
Air Blank	0.000	18:59
Air Blank	0.000	19:01
Subject Sample # 2	0.210	19:03
Breath Volume	4.320	19:03
Air Blank	0.000	19:04
Control Test	0.078	19:04
Air Blank	0.000	19:05
Diagnostics Check	OK	19:05

---

Data Download Date/Time: 11/29/2020 15:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAND,CURTIS,L  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: BLAND,CURTIS  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: BARNARD

DATE/TIME OF TEST: 11/13/2020 05:37:30

---

Diagnostics Check		OK	05:30
Air Blank		0.000	05:31
Control Test		0.078	05:31
Air Blank		0.000	05:32
Subject Sample	# 1	0.145	05:32
Breath Volume		4.128	05:32
Air Blank		0.000	05:33
Air Blank		0.000	05:35
Subject Sample	# 2	0.129	05:36
Breath Volume		2.578	05:36
Air Blank		0.000	05:36
Control Test		0.079	05:36
Air Blank		0.000	05:37
Diagnostics Check		OK	05:37

---

Data Download Date/Time: 11/29/2020 15:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSE,STEVEN,D  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: HOUSE,STEVEN  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 16:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: NICKERSON

DATE/TIME OF TEST: 11/13/2020 17:03:40

---

Diagnostics Check	OK	16:57
Air Blank	0.000	16:57
Control Test	0.080	16:57
Air Blank	0.000	16:58
Subject Sample # 1	0.354	16:58
Breath Volume	1.105	16:58
Air Blank	0.000	16:59
Air Blank	0.000	17:01
Subject Sample # 2	0.349	17:02
Breath Volume	1.195	17:02
Air Blank	0.000	17:02
Control Test	0.080	17:03
Air Blank	0.000	17:03
Diagnostics Check	OK	17:03

---

Data Download Date/Time: 11/29/2020 15:43



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TURNER,JONATHAN,R  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: HALE,CALEB  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 11/14/2020 03:41:27

---

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.081	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.197	03:35
Breath Volume	2.417	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.180	03:39
Breath Volume	1.148	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

---

Data Download Date/Time: 11/29/2020 15:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ABRAHAM,ERIC,D  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: ABRAHAM,ERIC  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: BAUMANN

DATE/TIME OF TEST: 11/14/2020 22:44:47

---

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.080	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.282	22:40
Breath Volume	1.304	22:40
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.299	22:43
Breath Volume	1.695	22:43
Air Blank	0.000	22:43
Control Test	0.079	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

---

Data Download Date/Time: 11/29/2020 15:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAND,CURTIS,L  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: BLAND,CURTIS  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 11/15/2020 20:51:01

---

Diagnostics Check		OK	20:44
Air Blank		0.000	20:45
Control Test		0.080	20:45
Air Blank		0.000	20:46
Subject Sample	# 1	0.248	20:46
Breath Volume		2.601	20:46
Air Blank		0.000	20:47
Air Blank		0.000	20:49
Subject Sample	# 2	0.233	20:49
Breath Volume		2.722	20:49
Air Blank		0.000	20:50
Control Test		0.079	20:50
Air Blank		0.000	20:50
Diagnostics Check		OK	20:51

---

Data Download Date/Time: 11/29/2020 15:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ABRAHAM,ERIC,D  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: ABRAHAM,ERIC  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: BURNETT

DATE/TIME OF TEST: 11/15/2020 23:14:03

---

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:09	
Control Test	0.081	23:10	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:12	Subject Test Refused
Breath Volume	0.000	23:12	
Air Blank	0.000	23:13	
Control Test	0.081	23:13	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:14	

---

Data Download Date/Time: 11/29/2020 15:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THORE,RICHARD,A  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: MCNIEL,RAYMOND  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 11/19/2020 22:11:33

---

Diagnostics Check	OK	22:03
Air Blank	0.000	22:04
Control Test	0.081	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.217	22:06
Breath Volume	3.679	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.211	22:10
Breath Volume	1.835	22:10
Air Blank	0.000	22:10
Control Test	0.080	22:10
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

---

Data Download Date/Time: 11/29/2020 15:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAND,CURTIS,L  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: BLAND,CURTIS  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: ASHBY

DATE/TIME OF TEST: 11/22/2020 00:53:51

---

Diagnostics Check	OK	00:49	
Air Blank	0.000	00:50	
Control Test	0.081	00:50	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:52	Subject Test Refused
Breath Volume	0.242	00:52	
Air Blank	0.000	00:52	
Control Test	0.081	00:53	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

---

Data Download Date/Time: 11/29/2020 15:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAND,CURTIS,L  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: BLAND,CURTIS  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: CHESTNUT

DATE/TIME OF TEST: 11/22/2020 04:00:50

---

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:58	
Control Test	0.081	03:58	
Air Blank	0.000	03:59	
Subject Sample	# 1	03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	03:59	
Control Test	0.080	04:00	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

---

Data Download Date/Time: 11/29/2020 15:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSE,STEVEN,D  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: BAKER,KENTON  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 10:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: BROE

DATE/TIME OF TEST: 11/22/2020 10:42:26

---

Diagnostics Check	OK	10:35
Air Blank	0.000	10:36
Control Test	0.081	10:36
Air Blank	0.000	10:37
Subject Sample # 1	0.000	10:38
Breath Volume	2.015	10:38
Air Blank	0.000	10:38
Air Blank	0.000	10:40
Subject Sample # 2	0.000	10:41
Breath Volume	1.921	10:41
Air Blank	0.000	10:41
Control Test	0.080	10:41
Air Blank	0.000	10:42
Diagnostics Check	OK	10:42

---

Data Download Date/Time: 11/29/2020 15:43



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ABRAHAM,ERIC,D  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: JOHNSON,ROBERT  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 11/24/2020 23:55:27

---

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.081	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.216	23:50
Breath Volume	2.742	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:52
Subject Sample # 2	0.210	23:53
Breath Volume	2.785	23:53
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

---

Data Download Date/Time: 11/29/2020 15:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VASQUEZ,EVEN,  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: DREW,LIAM  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 10/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 11/01/2020 04:29:45

---

Diagnostics Check		OK	04:18	
Air Blank		0.000	04:18	
Control Test		0.081	04:18	
Air Blank		0.000	04:19	
Subject Sample	# 1	0.160	04:22	Volume Not Met
Breath Volume		0.109	04:22	
Air Blank		0.000	04:23	
Air Blank		0.000	04:24	
Subject Sample	# 2	0.194	04:28	Volume Not Met
Breath Volume		0.085	04:28	
Air Blank		0.000	04:28	
Control Test		0.080	04:29	
Air Blank		0.000	04:29	
Diagnostics Check		OK	04:29	

---

Data Download Date/Time: 11/26/2020 13:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COREY,GAVIN,T  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: MILLER,JONATHON  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 10/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: EADDY

DATE/TIME OF TEST: 11/08/2020 04:22:08

---

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.081	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.163	04:17
Breath Volume	2.003	04:17
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.165	04:20
Breath Volume	1.847	04:20
Air Blank	0.000	04:21
Control Test	0.081	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

---

Data Download Date/Time: 11/26/2020 13:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLON, RAYMOND, M  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: BUSTAMANTE, SERGIO  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 10/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CORONILLA

DATE/TIME OF TEST: 11/13/2020 01:56:07

---

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.080	01:41	
Air Blank	0.000	01:42	
Subject Sample # 1	0.000	01:45	
Breath Volume	1.972	01:45	
Air Blank	0.000	01:45	
Air Blank	0.000	01:47	
Subject Sample # 2	0.000	01:51	Volume Not Met
Breath Volume	1.437	01:51	
Air Blank	0.000	01:51	
Air Blank	0.000	01:53	
Subject Sample # 3	0.000	01:54	
Breath Volume	1.222	01:54	
Air Blank	0.000	01:55	
Control Test	0.080	01:55	
Air Blank	0.000	01:56	
Diagnostics Check	OK	01:56	

---

Data Download Date/Time: 11/26/2020 13:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COREY,GAVIN,T  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: COREY,GAVIN  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 10/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 11/14/2020 23:50:20

---

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:47	
Control Test	0.080	23:48	
Air Blank	0.000	23:48	
Subject Sample	# 1	23:48	Subject Test Refused
Breath Volume	0.000	23:48	
Air Blank	0.000	23:49	
Control Test	0.080	23:49	
Air Blank	0.000	23:50	
Diagnostics Check	OK	23:50	

---

Data Download Date/Time: 11/26/2020 13:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VASQUEZ,EVEN,  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: SCELFO,ZACHARY  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 10/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 11/20/2020 03:10:56

---

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.079	03:08	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:10	
Control Test	0.080	03:10	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

---

Data Download Date/Time: 11/26/2020 13:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMBY,TAYLOR,N  
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD  
ARRESTING OFFICER: HARTUNG,NICHOLAS  
ARRESTING OFFICER AGENCY: Tyndall AFB PD  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A4  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: ALDANA

DATE/TIME OF TEST: 11/15/2020 01:03:29

---

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.236	00:59
Breath Volume		2.175	00:59
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.235	01:01
Breath Volume		2.082	01:01
Air Blank		0.000	01:02
Control Test		0.080	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

---

Data Download Date/Time: 12/11/2020 11:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENDERSON,ELIZABETH,D  
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD  
ARRESTING OFFICER: MURRAY,DAKOTA  
ARRESTING OFFICER AGENCY: Tyndall AFB PD  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A4  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: MCCATHY

DATE/TIME OF TEST: 11/30/2020 20:30:45

---

Diagnostics Check	OK	20:29	
Air Blank	0.000	20:29	
Control Test	0.081	20:30	
Air Blank		20:30	RFI Detect
Air Blank	0.000	20:30	

---

Data Download Date/Time: 12/11/2020 11:32



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENDERSON,ELIZABETH,D  
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD  
ARRESTING OFFICER: MURRAY,DAKOTA  
ARRESTING OFFICER AGENCY: Tyndall AFB PD  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A4  
DRY GAS EXPIRY DATE: 09/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 11/30/2020 20:44:54

---

Diagnostics Check	OK	20:33	
Air Blank	0.000	20:33	
Control Test	0.080	20:34	
Air Blank	0.000	20:34	
Subject Sample	# 1 0.180	20:37	Volume Not Met
Breath Volume	0.000	20:37	
Air Blank	0.000	20:38	
Air Blank	0.000	20:40	
Subject Sample	# 2	20:43	No Sample Provided
Breath Volume	0.000	20:43	
Air Blank	0.000	20:43	
Control Test	0.080	20:44	
Air Blank	0.000	20:44	
Diagnostics Check	OK	20:44	

---

Data Download Date/Time: 12/11/2020 11:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEYMOUR,KEVIN,R  
BREATH TEST OPERATOR AGENCY: Auburndale PD  
ARRESTING OFFICER: SEYMOUR,KEVIN  
ARRESTING OFFICER AGENCY: Auburndale PD  
LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 11/04/2020 22:44:44

---

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.079	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.234	22:38
Breath Volume	1.945	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:40
Subject Sample # 2	0.238	22:43
Breath Volume	1.867	22:43
Air Blank	0.000	22:43
Control Test	0.079	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

---

Data Download Date/Time: 11/24/2020 16:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWAN,DANNY,A  
BREATH TEST OPERATOR AGENCY: Auburndale PD  
ARRESTING OFFICER: SWAN,DANNY  
ARRESTING OFFICER AGENCY: Auburndale PD  
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 11/27/2020 22:10:13

---

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.079	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.000	22:05
Breath Volume	3.472	22:05
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.000	22:08
Breath Volume	3.328	22:08
Air Blank	0.000	22:09
Control Test	0.079	22:09
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

---

Data Download Date/Time: 12/15/2020 17:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWAN,DANNY,A  
BREATH TEST OPERATOR AGENCY: Auburndale PD  
ARRESTING OFFICER: HAMILTON,RICHARD  
ARRESTING OFFICER AGENCY: Auburndale PD  
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: HALVERSON

DATE/TIME OF TEST: 11/27/2020 23:46:08

---

Diagnostics Check	OK	23:36	
Air Blank	0.000	23:37	
Control Test	0.080	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 1	23:41	No Sample Provided
Breath Volume	0.000	23:41	
Air Blank	0.000	23:41	
Air Blank	0.000	23:43	
Subject Sample	# 2	23:44	Subject Test Refused
Breath Volume	0.000	23:44	
Air Blank	0.000	23:45	
Control Test	0.081	23:45	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:46	

---

Data Download Date/Time: 12/15/2020 17:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDDERS,NORMAN,R  
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores  
ARRESTING OFFICER: RAYNOR,JASON  
ARRESTING OFFICER AGENCY: Daytona Beach PD  
LAST AGENCY INSPECTION: 10/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1305520  
DRY GAS EXPIRY DATE: 04/01/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: HOOPER

DATE/TIME OF TEST: 11/08/2020 04:01:51

---

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:52	
Control Test		0.080	03:52	
Air Blank		0.000	03:53	
Subject Sample	# 1	0.208	03:54	No .020 agreement
Breath Volume		1.195	03:54	
Air Blank		0.000	03:54	
Air Blank		0.000	03:56	
Subject Sample	# 2	0.238	03:57	No .020 agreement
Breath Volume		1.839	03:57	
Air Blank		0.000	03:57	
Air Blank		0.000	03:59	
Subject Sample	# 3	0.259	04:00	No .020 agreement
Breath Volume		2.675	04:00	
Air Blank		0.000	04:00	
Control Test		0.079	04:01	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

---

Data Download Date/Time: 11/23/2020 02:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J  
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores  
ARRESTING OFFICER: CASTELLANO,VINCENT  
ARRESTING OFFICER AGENCY: Daytona Beach Shores  
LAST AGENCY INSPECTION: 10/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1305520  
DRY GAS EXPIRY DATE: 04/01/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: DI POLITO

DATE/TIME OF TEST: 11/13/2020 23:27:44

---

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.080	23:22
Air Blank	0.000	23:22
Subject Sample # 1	0.082	23:23
Breath Volume	2.128	23:23
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.083	23:26
Breath Volume	1.460	23:26
Air Blank	0.000	23:26
Control Test	0.078	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

---

Data Download Date/Time: 11/23/2020 02:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J  
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores  
ARRESTING OFFICER: MOWEN,RYAN  
ARRESTING OFFICER AGENCY: Daytona Beach Shores  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1305520  
DRY GAS EXPIRY DATE: 04/01/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: ALAS CASTILLO

DATE/TIME OF TEST: 11/28/2020 05:15:38

---

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.080	05:10
Air Blank	0.000	05:10
Subject Sample # 1	0.177	05:11
Breath Volume	2.347	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.176	05:14
Breath Volume	2.199	05:14
Air Blank	0.000	05:14
Control Test	0.079	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

---

Data Download Date/Time: 12/17/2020 00:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRESENZI,G,  
BREATH TEST OPERATOR AGENCY: Edgewater PD  
ARRESTING OFFICER: CRESENZI,G  
ARRESTING OFFICER AGENCY: Edgewater PD  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1217042  
DRY GAS EXPIRY DATE: 10/10/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: PINNEY

DATE/TIME OF TEST: 11/17/2020 06:05:15

---

Diagnostics Check	OK	05:58
Air Blank	0.000	05:59
Control Test	0.079	05:59
Air Blank	0.000	06:00
Subject Sample # 1	0.224	06:00
Breath Volume	3.128	06:00
Air Blank	0.000	06:01
Air Blank	0.000	06:03
Subject Sample # 2	0.221	06:03
Breath Volume	1.382	06:03
Air Blank	0.000	06:04
Control Test	0.079	06:04
Air Blank	0.000	06:05
Diagnostics Check	OK	06:05

---

Data Download Date/Time: 11/29/2020 14:37



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FEGER,JOSHUA,D  
BREATH TEST OPERATOR AGENCY: Edgewater PD  
ARRESTING OFFICER: FEGER,JOSHUA  
ARRESTING OFFICER AGENCY: Edgewater PD  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1217042  
DRY GAS EXPIRY DATE: 10/10/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 11/20/2020 18:44:50

---

Diagnostics Check	OK	18:38
Air Blank	0.000	18:38
Control Test	0.078	18:38
Air Blank	0.000	18:39
Subject Sample # 1	0.179	18:40
Breath Volume	1.867	18:40
Air Blank	0.000	18:41
Air Blank	0.000	18:42
Subject Sample # 2	0.184	18:43
Breath Volume	1.941	18:43
Air Blank	0.000	18:43
Control Test	0.077	18:44
Air Blank	0.000	18:44
Diagnostics Check	OK	18:44

---

Data Download Date/Time: 11/29/2020 14:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINZ,ANTHONY,E  
BREATH TEST OPERATOR AGENCY: Edgewater PD  
ARRESTING OFFICER: FEGER,JOSHUA  
ARRESTING OFFICER AGENCY: Edgewater PD  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1217042  
DRY GAS EXPIRY DATE: 10/10/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:18

SUBJECT'S LAST NAME: AMMONS

DATE/TIME OF TEST: 11/24/2020 14:01:23

---

Diagnostics Check	OK	13:52
Air Blank	0.000	13:52
Control Test	0.078	13:52
Air Blank	0.000	13:53
Subject Sample # 1	0.314	13:54
Breath Volume	1.144	13:54
Air Blank	0.000	13:55
Air Blank	0.000	13:57
Subject Sample # 2	0.304	13:59
Breath Volume	1.250	13:59
Air Blank	0.000	14:00
Control Test	0.077	14:00
Air Blank	0.000	14:01
Diagnostics Check	OK	14:01

---

Data Download Date/Time: 11/29/2020 14:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRESENZI,G,  
BREATH TEST OPERATOR AGENCY: Edgewater PD  
ARRESTING OFFICER: JENKINS,A  
ARRESTING OFFICER AGENCY: New Smyrna Beach PD  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1217042  
DRY GAS EXPIRY DATE: 10/10/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: ERHARD

DATE/TIME OF TEST: 11/26/2020 01:29:38

---

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.079	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.226	01:24
Breath Volume		1.152	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.212	01:28
Breath Volume		1.617	01:28
Air Blank		0.000	01:28
Control Test		0.079	01:29
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

---

Data Download Date/Time: 11/29/2020 14:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVILA,NICOLE,M  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: RETTE,CODY  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 10/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1185592  
DRY GAS EXPIRY DATE: 08/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 11/09/2020 16:49:39

---

Diagnostics Check	OK	16:43
Air Blank	0.000	16:43
Control Test	0.082	16:43
Air Blank	0.000	16:44
Subject Sample # 1	0.282	16:44
Breath Volume	2.257	16:44
Air Blank	0.000	16:45
Air Blank	0.000	16:47
Subject Sample # 2	0.266	16:48
Breath Volume	2.003	16:48
Air Blank	0.000	16:48
Control Test	0.082	16:49
Air Blank	0.000	16:49
Diagnostics Check	OK	16:49

---

Data Download Date/Time: 12/13/2020 04:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMBSH,CARLYTON,R  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: HAMBSH,CARLYTON  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 10/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1185592  
DRY GAS EXPIRY DATE: 08/13/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: KENNEDY PROCTOR

DATE/TIME OF TEST: 11/12/2020 19:44:14

---

Diagnostics Check	OK	19:38
Air Blank	0.000	19:38
Control Test	0.082	19:38
Air Blank	0.000	19:39
Subject Sample # 1	0.000	19:39
Breath Volume	2.472	19:39
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.000	19:42
Breath Volume	2.472	19:42
Air Blank	0.000	19:43
Control Test	0.081	19:43
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

---

Data Download Date/Time: 12/13/2020 04:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STARLING,DARREN,J  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: VELEZ,EDWARD  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 10/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1185592  
DRY GAS EXPIRY DATE: 08/13/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: LARSON

DATE/TIME OF TEST: 11/21/2020 02:00:05

---

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.082	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.209	01:55
Breath Volume	1.886	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.219	01:58
Breath Volume	1.417	01:58
Air Blank	0.000	01:59
Control Test	0.082	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

---

Data Download Date/Time: 12/13/2020 04:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVILA,NICOLE,M  
BREATH TEST OPERATOR AGENCY: Port Orange PD  
ARRESTING OFFICER: ALEVISATOS,JARED  
ARRESTING OFFICER AGENCY: Port Orange PD  
LAST AGENCY INSPECTION: 10/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1185592  
DRY GAS EXPIRY DATE: 08/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 11/28/2020 22:38:50

---

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:27	
Control Test	0.081	22:27	
Air Blank	0.000	22:28	
Subject Sample	# 1 0.152	22:28	
Breath Volume	1.757	22:28	
Air Blank	0.000	22:29	
Air Blank	0.000	22:31	
Subject Sample	# 2 0.143	22:34	Volume Not Met
Breath Volume	0.015	22:34	
Air Blank	0.000	22:34	
Air Blank	0.000	22:36	
Subject Sample	# 3 0.150	22:37	
Breath Volume	1.906	22:37	
Air Blank	0.000	22:37	
Control Test	0.080	22:38	
Air Blank	0.000	22:38	
Diagnostics Check	OK	22:38	

---

Data Download Date/Time: 12/13/2020 04:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MAGUIRE,MICHAEL  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 11/01/2020 05:52:36

---

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.082	05:45
Air Blank	0.000	05:45
Subject Sample # 1	0.069	05:47
Breath Volume	1.816	05:47
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample # 2	0.062	05:51
Breath Volume	1.406	05:51
Air Blank	0.000	05:51
Control Test	0.081	05:52
Air Blank	0.000	05:52
Diagnostics Check	OK	05:52

---

Data Download Date/Time: 11/05/2020 11:41



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: AZEVEDO,FABIO  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 11/01/2020 07:21:05

---

Diagnostics Check		OK	07:14
Air Blank		0.000	07:14
Control Test		0.081	07:15
Air Blank		0.000	07:15
Subject Sample	# 1	0.126	07:16
Breath Volume		2.210	07:16
Air Blank		0.000	07:16
Air Blank		0.000	07:18
Subject Sample	# 2	0.127	07:19
Breath Volume		2.394	07:19
Air Blank		0.000	07:20
Control Test		0.082	07:20
Air Blank		0.000	07:20
Diagnostics Check		OK	07:21

---

Data Download Date/Time: 11/05/2020 11:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ROCCHILD,MIGUEL  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/04/2020 02:31:44

---

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.082	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.363	02:27
Breath Volume	2.281	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.349	02:30
Breath Volume	1.738	02:30
Air Blank	0.000	02:30
Control Test	0.081	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

---

Data Download Date/Time: 11/05/2020 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: INABNET,JUSTIN  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: BUTTENBENDER

DATE/TIME OF TEST: 11/04/2020 05:39:25

---

Diagnostics Check	OK	05:32
Air Blank	0.000	05:33
Control Test	0.081	05:33
Air Blank	0.000	05:33
Subject Sample # 1	0.181	05:34
Breath Volume	2.566	05:34
Air Blank	0.000	05:35
Air Blank	0.000	05:36
Subject Sample # 2	0.175	05:37
Breath Volume	1.203	05:37
Air Blank	0.000	05:38
Control Test	0.082	05:38
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

---

Data Download Date/Time: 11/05/2020 11:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DOMINGUEZ,JAYMES  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: ORTIZ DIAZ

DATE/TIME OF TEST: 11/05/2020 03:18:40

---

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.082	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.076	03:14
Breath Volume	2.777	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.076	03:17
Breath Volume	3.125	03:17
Air Blank	0.000	03:17
Control Test	0.081	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

---

Data Download Date/Time: 11/05/2020 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SANCHEZ,KEVIN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: NEVILUS

DATE/TIME OF TEST: 11/05/2020 03:53:55

---

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:42	
Control Test		0.081	03:43	
Air Blank		0.000	03:43	
Subject Sample	# 1	0.106	03:46	Volume Not Met
Breath Volume		0.066	03:46	
Air Blank		0.000	03:47	
Air Blank		0.000	03:49	
Subject Sample	# 2	0.000	03:52	Volume Not Met
Breath Volume		0.015	03:52	
Air Blank		0.000	03:53	
Control Test		0.081	03:53	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

---

Data Download Date/Time: 11/05/2020 11:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WATKINS,ROMAN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 11/15/2020 05:02:05

---

Diagnostics Check		OK	04:54
Air Blank		0.000	04:55
Control Test		0.082	04:55
Air Blank		0.000	04:56
Subject Sample	# 1	0.128	04:57
Breath Volume		1.269	04:57
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.134	05:00
Breath Volume		1.421	05:00
Air Blank		0.000	05:01
Control Test		0.080	05:01
Air Blank		0.000	05:01
Diagnostics Check		OK	05:02

---

Data Download Date/Time: 12/07/2020 14:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MENDOZA,ANGEL  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: MCKAY

DATE/TIME OF TEST: 11/21/2020 01:46:23

---

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.082	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.157	01:40
Breath Volume	1.535	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.159	01:44
Breath Volume	1.507	01:44
Air Blank	0.000	01:45
Control Test	0.081	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

---

Data Download Date/Time: 12/07/2020 14:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DEL CASTILLO,BRET  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: HAJ ALI AKBARI

DATE/TIME OF TEST: 11/21/2020 03:51:09

---

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.082	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.233	03:46
Breath Volume	1.519	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.235	03:49
Breath Volume	2.691	03:49
Air Blank	0.000	03:50
Control Test	0.081	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

---

Data Download Date/Time: 12/07/2020 14:54



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GARARD,WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: BLACKSHEAR

DATE/TIME OF TEST: 11/23/2020 21:43:39

---

Diagnostics Check	OK	21:36
Air Blank	0.000	21:37
Control Test	0.082	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.135	21:39
Breath Volume	1.312	21:39
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.135	21:42
Breath Volume	1.609	21:42
Air Blank	0.000	21:42
Control Test	0.081	21:43
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

---

Data Download Date/Time: 12/07/2020 14:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLBROOK,NICHOLAS,B  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: MYNY,ERIC  
ARRESTING OFFICER AGENCY: Cocoa PD  
LAST AGENCY INSPECTION: 10/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A4  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: LAUFFER

DATE/TIME OF TEST: 11/06/2020 01:19:50

---

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.078	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.188	01:15
Breath Volume	1.742	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:17
Subject Sample # 2	0.182	01:18
Breath Volume	1.722	01:18
Air Blank	0.000	01:18
Control Test	0.076	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

---

Data Download Date/Time: 11/25/2020 19:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLBROOK,NICHOLAS,B  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: SCHALLON,CHRISTIAN  
ARRESTING OFFICER AGENCY: Rockledge PD  
LAST AGENCY INSPECTION: 10/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175968  
DRY GAS EXPIRY DATE: 07/22/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: BRANHAM

DATE/TIME OF TEST: 11/09/2020 20:25:56

---

Diagnostics Check	OK	20:19
Air Blank	0.000	20:20
Control Test	0.079	20:20
Air Blank	0.000	20:20
Subject Sample # 1	0.153	20:21
Breath Volume	2.093	20:21
Air Blank	0.000	20:22
Air Blank	0.000	20:24
Subject Sample # 2	0.152	20:24
Breath Volume	1.648	20:24
Air Blank	0.000	20:25
Control Test	0.077	20:25
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

---

Data Download Date/Time: 12/09/2020 09:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,JAMES,P  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: CLARK,JAMES  
ARRESTING OFFICER AGENCY: Rockledge PD  
LAST AGENCY INSPECTION: 10/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175968  
DRY GAS EXPIRY DATE: 07/22/2022  
VIOLATION CODE: DUI  
ARREST TIME: 14:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 11/13/2020 15:22:37

---

Diagnostics Check	OK	15:16
Air Blank	0.000	15:16
Control Test	0.078	15:16
Air Blank	0.000	15:17
Subject Sample # 1	0.000	15:18
Breath Volume	3.609	15:18
Air Blank	0.000	15:18
Air Blank	0.000	15:20
Subject Sample # 2	0.000	15:21
Breath Volume	2.406	15:21
Air Blank	0.000	15:21
Control Test	0.077	15:22
Air Blank	0.000	15:22
Diagnostics Check	OK	15:22

---

Data Download Date/Time: 12/09/2020 09:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARNER, TYRELL, J  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: BLASSINGAME, PAUL  
ARRESTING OFFICER AGENCY: Rockledge PD  
LAST AGENCY INSPECTION: 10/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175968  
DRY GAS EXPIRY DATE: 07/22/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: SLEY

DATE/TIME OF TEST: 11/21/2020 01:37:05

---

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:34	
Control Test	0.079	01:35	
Air Blank	0.000	01:35	
Subject Sample	# 1	01:35	Subject Test Refused
Breath Volume	0.000	01:35	
Air Blank	0.000	01:36	
Control Test	0.078	01:36	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:37	

---

Data Download Date/Time: 12/09/2020 09:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLBROOK,NICHOLAS,B  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: CRONIN,ERIC  
ARRESTING OFFICER AGENCY: Cocoa PD  
LAST AGENCY INSPECTION: 10/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11759668  
DRY GAS EXPIRY DATE: 07/22/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SHUFRO

DATE/TIME OF TEST: 11/27/2020 22:11:23

---

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.078	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.000	22:06
Breath Volume	1.437	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.000	22:10
Breath Volume	1.316	22:10
Air Blank	0.000	22:10
Control Test	0.078	22:10
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

---

Data Download Date/Time: 12/09/2020 09:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLBROOK,NICHOLAS,B  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: SCHALLON,CHRISTIAN  
ARRESTING OFFICER AGENCY: Rockledge PD  
LAST AGENCY INSPECTION: 10/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175968  
DRY GAS EXPIRY DATE: 07/22/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 11/28/2020 01:41:02

---

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:38	
Control Test	0.079	01:39	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:40	
Control Test	0.078	01:40	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:41	

---

Data Download Date/Time: 12/09/2020 09:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUBRANO,OLIVIA,C  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: LABOY,LONITA  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 11/05/2020 02:30:11

---

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.097	02:25
Breath Volume	2.421	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.077	02:28
Breath Volume	2.210	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

---

Data Download Date/Time: 11/29/2020 07:19



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: VAZQUEZ,A  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: FALLACE

DATE/TIME OF TEST: 11/06/2020 22:46:34

---

Diagnostics Check	OK	22:39
Air Blank	0.000	22:39
Control Test	0.081	22:39
Air Blank	0.000	22:40
Subject Sample # 1	0.152	22:41
Breath Volume	2.824	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.148	22:44
Breath Volume	2.644	22:44
Air Blank	0.000	22:45
Control Test	0.078	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

---

Data Download Date/Time: 11/29/2020 07:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIVINS,HEATHER,C  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: BLAZEWICZ,ALISON  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: CRUZ CRUZ

DATE/TIME OF TEST: 11/07/2020 03:06:37

---

Diagnostics Check		OK	02:59
Air Blank		0.000	03:00
Control Test		0.081	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.210	03:02
Breath Volume		1.425	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.200	03:05
Breath Volume		1.121	03:05
Air Blank		0.000	03:05
Control Test		0.079	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

---

Data Download Date/Time: 11/29/2020 07:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: VAZQUEZ,A  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SPINNING

DATE/TIME OF TEST: 11/07/2020 23:47:35

---

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.081	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.000	23:43
Breath Volume	3.082	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.000	23:46
Breath Volume	2.816	23:46
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

---

Data Download Date/Time: 11/29/2020 07:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: BLZEWICZ,A  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: NAZAR

DATE/TIME OF TEST: 11/08/2020 19:34:06

---

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.081	19:26
Air Blank	0.000	19:26
Subject Sample # 1	0.191	19:28
Breath Volume	2.035	19:28
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.201	19:32
Breath Volume	3.597	19:32
Air Blank	0.000	19:33
Control Test	0.080	19:33
Air Blank	0.000	19:34
Diagnostics Check	OK	19:34

---

Data Download Date/Time: 11/29/2020 07:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: VAZQUEZ,A  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HARNEY

DATE/TIME OF TEST: 11/12/2020 23:02:16

---

Diagnostics Check	OK	23:01	
Air Blank		23:01	RFI Detect
Air Blank	0.000	23:02	

---

Data Download Date/Time: 11/29/2020 07:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: VAZQUEZ,A  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HARNEY

DATE/TIME OF TEST: 11/12/2020 23:11:21

---

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.081	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.145	23:06
Breath Volume	1.957	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:08
Subject Sample # 2	0.142	23:09
Breath Volume	1.906	23:09
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

---

Data Download Date/Time: 11/29/2020 07:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: PEARSON,M  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 11/13/2020 00:05:14

---

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.081	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.000	00:00
Breath Volume	1.625	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.000	00:03
Breath Volume	1.425	00:03
Air Blank	0.000	00:04
Control Test	0.080	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

---

Data Download Date/Time: 11/29/2020 07:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITE,T,  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: DURGIN,S  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: TOPP

DATE/TIME OF TEST: 11/14/2020 00:27:19

---

Diagnostics Check		OK	00:20
Air Blank		0.000	00:21
Control Test		0.081	00:21
Air Blank		0.000	00:22
Subject Sample	# 1	0.000	00:23
Breath Volume		2.667	00:23
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.000	00:25
Breath Volume		1.300	00:25
Air Blank		0.000	00:26
Control Test		0.080	00:26
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

---

Data Download Date/Time: 11/29/2020 07:19



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRISON,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: IERVASL,M  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 11/14/2020 04:55:18

---

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.080	04:49
Air Blank	0.000	04:50
Subject Sample # 1	0.185	04:50
Breath Volume	2.675	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.180	04:53
Breath Volume	2.476	04:53
Air Blank	0.000	04:54
Control Test	0.079	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

---

Data Download Date/Time: 11/29/2020 07:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRISON,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: VAZQUEZ,A  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: CORDERO CASTANEDA

DATE/TIME OF TEST: 11/15/2020 00:56:26

---

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.079	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.038	00:51
Breath Volume	3.722	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.036	00:54
Breath Volume	3.359	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

---

Data Download Date/Time: 11/29/2020 07:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUBRANO,OLIVIA,C  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: DOLCI,PETER  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: MATIAS

DATE/TIME OF TEST: 11/15/2020 04:33:39

---

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.107	04:27
Breath Volume	2.082	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.099	04:31
Breath Volume	1.687	04:31
Air Blank	0.000	04:32
Control Test	0.080	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

---

Data Download Date/Time: 11/29/2020 07:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: BLAZEWICZ,A  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 11/17/2020 22:03:05

---

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.081	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.146	21:57
Breath Volume	1.882	21:57
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.135	22:01
Breath Volume	1.777	22:01
Air Blank	0.000	22:02
Control Test	0.081	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

---

Data Download Date/Time: 11/29/2020 07:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRISON,A,  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: KEY,J  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 11/20/2020 00:34:30

---

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:28	
Control Test	0.080	00:29	
Air Blank	0.000	00:29	
Subject Sample	# 1 0.260	00:29	
Breath Volume	3.128	00:29	
Air Blank	0.000	00:30	
Air Blank	0.000	00:32	
Subject Sample	# 2 0.240	00:32	
Breath Volume	3.472	00:32	
Air Blank	0.000	00:33	
Control Test	0.066	00:33	Control Outside Tolerance
Air Blank	0.000	00:34	

---

Data Download Date/Time: 11/29/2020 07:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRISON,A,  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: KEY,J  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 11/20/2020 00:55:05

---

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:46	
Control Test	0.076	00:46	
Air Blank	0.000	00:47	
Subject Sample	# 1 0.255	00:47	
Breath Volume	3.394	00:47	
Air Blank	0.000	00:48	
Air Blank	0.000	00:50	
Subject Sample	# 2 0.231	00:50	
Breath Volume	3.601	00:50	
Air Blank	0.000	00:51	
Air Blank	0.000	00:53	
Subject Sample	# 3 0.246	00:53	
Breath Volume	3.410	00:53	
Air Blank	0.000	00:54	
Control Test	0.070	00:54	Control Outside Tolerance
Air Blank	0.000	00:55	

---

Data Download Date/Time: 11/29/2020 07:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEILL, MICHAEL, A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: BLAZAWICZ, ALISON  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 11/21/2020 00:33:08

---

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.081	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.000	00:28
Breath Volume	1.984	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.000	00:31
Breath Volume	2.644	00:31
Air Blank	0.000	00:32
Control Test	0.081	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

---

Data Download Date/Time: 11/29/2020 07:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,B,  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: FANNING,A  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 12:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:52

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 11/21/2020 13:21:26

---

Diagnostics Check	OK	13:15
Air Blank	0.000	13:15
Control Test	0.080	13:15
Air Blank	0.000	13:16
Subject Sample # 1	0.207	13:16
Breath Volume	4.492	13:16
Air Blank	0.000	13:17
Air Blank	0.000	13:19
Subject Sample # 2	0.198	13:19
Breath Volume	3.250	13:19
Air Blank	0.000	13:20
Control Test	0.079	13:20
Air Blank	0.000	13:21
Diagnostics Check	OK	13:21

---

Data Download Date/Time: 11/29/2020 07:19



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEAL,ADAM,M  
BREATH TEST OPERATOR AGENCY: Indialantic PD  
ARRESTING OFFICER: BLAZEWICZ,ALISON  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 11/22/2020 01:30:50

---

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.000	01:26
Breath Volume	2.750	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.000	01:29
Breath Volume	2.242	01:29
Air Blank	0.000	01:29
Control Test	0.080	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

---

Data Download Date/Time: 11/29/2020 07:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HESSION,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: WILLINGHAM,JARRETT  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 15:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:36

SUBJECT'S LAST NAME: CHERRY

DATE/TIME OF TEST: 11/22/2020 16:04:09

---

Diagnostics Check	OK	15:57
Air Blank	0.000	15:58
Control Test	0.081	15:58
Air Blank	0.000	15:58
Subject Sample # 1	0.109	15:59
Breath Volume	1.238	15:59
Air Blank	0.000	16:00
Air Blank	0.000	16:01
Subject Sample # 2	0.101	16:02
Breath Volume	1.542	16:02
Air Blank	0.000	16:03
Control Test	0.081	16:03
Air Blank	0.000	16:04
Diagnostics Check	OK	16:04

---

Data Download Date/Time: 11/29/2020 07:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HICKS,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: West Melbourne PD  
ARRESTING OFFICER: WILSON,JORDAN  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 10/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: GAYLE

DATE/TIME OF TEST: 11/22/2020 23:52:28

---

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.081	23:46
Air Blank	0.000	23:47
Subject Sample # 1	0.044	23:48
Breath Volume	1.265	23:48
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.037	23:50
Breath Volume	1.105	23:50
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

---

Data Download Date/Time: 11/29/2020 07:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: BLAZEWICZ,A  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ALGAHTANI

DATE/TIME OF TEST: 11/27/2020 02:36:21

---

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.199	02:31
Breath Volume	2.804	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.204	02:34
Breath Volume	1.835	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

---

Data Download Date/Time: 11/29/2020 07:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITE,T,  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: DURGIN,S  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 11/28/2020 21:40:11

---

Diagnostics Check		OK	21:25	
Air Blank		0.000	21:26	
Control Test		0.081	21:26	
Air Blank		0.000	21:26	
Subject Sample	# 1	0.259	21:30	Volume Not Met
Breath Volume		0.121	21:30	
Air Blank		0.000	21:30	
Air Blank		0.000	21:32	
Subject Sample	# 2	0.259	21:34	
Breath Volume		1.417	21:34	
Air Blank		0.000	21:35	
Air Blank		0.000	21:36	
Subject Sample	# 3	0.256	21:38	
Breath Volume		1.613	21:38	
Air Blank		0.000	21:39	
Control Test		0.078	21:39	
Air Blank		0.000	21:40	
Diagnostics Check		OK	21:40	

---

Data Download Date/Time: 11/29/2020 07:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRISON,ALEXIS,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: HARRISON,A  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 11/29/2020 21:18:06

---

Diagnostics Check	OK	21:11
Air Blank	0.000	21:11
Control Test	0.081	21:12
Air Blank	0.000	21:12
Subject Sample # 1	0.000	21:13
Breath Volume	2.382	21:13
Air Blank	0.000	21:14
Air Blank	0.000	21:16
Subject Sample # 2	0.000	21:16
Breath Volume	3.003	21:16
Air Blank	0.000	21:17
Control Test	0.080	21:17
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

---

Data Download Date/Time: 12/31/2020 12:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 11/28/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22620080A2

DRY GAS EXPIRY DATE: 10/05/2022

VIOLATION CODE: DUI

ARREST TIME: 22:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 11/28/2020 22:22:30

---

Diagnostics Check		OK	22:16
Air Blank		0.000	22:16
Control Test		0.079	22:16
Air Blank		0.000	22:17
Subject Sample	# 1	0.154	22:18
Breath Volume		2.761	22:18
Air Blank		0.000	22:18
Air Blank		0.000	22:20
Subject Sample	# 2	0.155	22:20
Breath Volume		2.480	22:20
Air Blank		0.000	22:21
Control Test		0.080	22:21
Air Blank		0.000	22:22
Diagnostics Check		OK	22:22

---

Data Download Date/Time: 12/13/2020 18:13





# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22620080A2

DRY GAS EXPIRY DATE: 10/05/2022

VIOLATION CODE: DUI

ARREST TIME: 19:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 11/12/2020 19:17:38

---

Diagnostics Check	OK	19:13	
Air Blank	0.000	19:14	
Control Test	0.079	19:14	
Air Blank	0.000	19:15	
Subject Sample	# 1	19:16	Subject Test Refused
Breath Volume	0.000	19:16	
Air Blank	0.000	19:16	
Control Test	0.078	19:17	
Air Blank	0.000	19:17	
Diagnostics Check	OK	19:17	

---

Data Download Date/Time: 12/03/2020 14:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22620080A2

DRY GAS EXPIRY DATE: 10/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:20

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: GANNON

DATE/TIME OF TEST: 11/19/2020 02:37:37

---

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.080	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.145	02:33
Breath Volume	1.414	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.139	02:36
Breath Volume	1.796	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

---

Data Download Date/Time: 12/03/2020 14:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22620080A2

DRY GAS EXPIRY DATE: 10/05/2022

VIOLATION CODE: DUI

ARREST TIME: 23:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: DELVA

DATE/TIME OF TEST: 11/22/2020 23:32:22

---

Diagnostics Check		OK	23:19	
Air Blank		0.000	23:19	
Control Test		0.079	23:19	
Air Blank		0.000	23:20	
Subject Sample	# 1	0.141	23:23	Volume Not Met
Breath Volume		0.128	23:23	
Air Blank		0.000	23:24	
Air Blank		0.000	23:25	
Subject Sample	# 2	0.162	23:27	
Breath Volume		1.285	23:27	
Air Blank		0.000	23:27	
Air Blank		0.000	23:29	
Subject Sample	# 3	0.165	23:30	
Breath Volume		1.222	23:30	
Air Blank		0.000	23:31	
Control Test		0.078	23:31	
Air Blank		0.000	23:32	
Diagnostics Check		OK	23:32	

---

Data Download Date/Time: 12/03/2020 14:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: NEWTON,ZACHARY  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: MAISONET

DATE/TIME OF TEST: 11/01/2020 01:54:35

---

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.080	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.092	01:50
Breath Volume	2.464	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.091	01:53
Breath Volume	2.378	01:53
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

---

Data Download Date/Time: 11/04/2020 22:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: KINKOPF,TIMOTHY  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 11/01/2020 02:08:43

---

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.080	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.208	02:04
Breath Volume		2.699	02:04
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.202	02:07
Breath Volume		2.691	02:07
Air Blank		0.000	02:07
Control Test		0.080	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

---

Data Download Date/Time: 11/04/2020 22:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MASSIP,LEONARDO  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: BENOIT

DATE/TIME OF TEST: 11/01/2020 05:36:14

---

Diagnostics Check	OK	05:30
Air Blank	0.000	05:30
Control Test	0.080	05:31
Air Blank	0.000	05:31
Subject Sample # 1	0.161	05:31
Breath Volume	2.144	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.161	05:34
Breath Volume	1.917	05:34
Air Blank	0.000	05:35
Control Test	0.081	05:35
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

---

Data Download Date/Time: 11/04/2020 22:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINKOPF,TIM,E  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: HIDEROA,KEVIN  
ARRESTING OFFICER AGENCY: Altamonte Springs PD  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 11/04/2020 01:43:24

---

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.082	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.105	01:38
Breath Volume	3.042	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.100	01:41
Breath Volume	3.265	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

---

Data Download Date/Time: 11/04/2020 22:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MCDONNELL,BRANDON  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: FYE

DATE/TIME OF TEST: 11/07/2020 05:11:58

---

Diagnostics Check		OK	05:04
Air Blank		0.000	05:04
Control Test		0.080	05:05
Air Blank		0.000	05:05
Subject Sample	# 1	0.188	05:07
Breath Volume		2.281	05:07
Air Blank		0.000	05:07
Air Blank		0.000	05:09
Subject Sample	# 2	0.203	05:10
Breath Volume		2.062	05:10
Air Blank		0.000	05:11
Control Test		0.080	05:11
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

---

Data Download Date/Time: 12/31/2020 22:17



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: UCELO,BRYAN  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: HEALY

DATE/TIME OF TEST: 11/08/2020 03:55:23

---

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.138	03:49
Breath Volume	1.582	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:51
Subject Sample # 2	0.131	03:53
Breath Volume	1.597	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

---

Data Download Date/Time: 12/31/2020 22:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: HIDEROA, KEVIN  
ARRESTING OFFICER AGENCY: Altamonte Springs PD  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: FIVEK

DATE/TIME OF TEST: 11/12/2020 03:55:30

---

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.137	03:50
Breath Volume	1.242	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:52
Subject Sample # 2	0.150	03:53
Breath Volume	2.257	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

---

Data Download Date/Time: 12/31/2020 22:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: KAISER,DAVID  
ARRESTING OFFICER AGENCY: Sanford PD  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: DOXTATER

DATE/TIME OF TEST: 11/13/2020 03:25:54

---

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:14	
Control Test	0.081	03:15	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:18	No Sample Provided
Breath Volume	0.000	03:18	
Air Blank	0.000	03:19	
Air Blank	0.000	03:21	
Subject Sample	# 2	03:24	No Sample Provided
Breath Volume	0.000	03:24	
Air Blank	0.000	03:25	
Control Test	0.081	03:25	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

---

Data Download Date/Time: 12/31/2020 22:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: KAISER,DAVID  
ARRESTING OFFICER AGENCY: Sanford PD  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: DOXTATER

DATE/TIME OF TEST: 11/13/2020 03:37:37

---

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.177	03:31
Breath Volume	1.859	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.177	03:35
Breath Volume	2.242	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

---

Data Download Date/Time: 12/31/2020 22:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: KINKOPF,TIMOTHY  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 11/15/2020 02:50:29

---

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.211	02:45
Breath Volume	2.273	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.205	02:48
Breath Volume	2.152	02:48
Air Blank	0.000	02:49
Control Test	0.077	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

---

Data Download Date/Time: 12/31/2020 22:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANCION,CALIX,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: ANCION,CALIX  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: SWENSON

DATE/TIME OF TEST: 11/15/2020 16:32:10

---

Diagnostics Check	OK	16:25
Air Blank	0.000	16:26
Control Test	0.080	16:26
Air Blank	0.000	16:26
Subject Sample # 1	0.000	16:27
Breath Volume	1.457	16:27
Air Blank	0.000	16:28
Air Blank	0.000	16:30
Subject Sample # 2	0.000	16:30
Breath Volume	1.222	16:30
Air Blank	0.000	16:31
Control Test	0.081	16:31
Air Blank	0.000	16:32
Diagnostics Check	OK	16:32

---

Data Download Date/Time: 12/31/2020 22:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: POZADA, ANDREW  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 11/17/2020 02:07:15

---

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.081	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.000	02:03
Breath Volume	1.714	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.000	02:05
Breath Volume	1.269	02:05
Air Blank	0.000	02:06
Control Test	0.081	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

---

Data Download Date/Time: 12/31/2020 22:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: PFANNERSTILL, JACOB  
ARRESTING OFFICER AGENCY: Altamonte Springs PD  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 11/18/2020 01:54:45

---

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.081	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.280	01:50
Breath Volume		2.515	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:52
Subject Sample	# 2	0.278	01:53
Breath Volume		2.679	01:53
Air Blank		0.000	01:53
Control Test		0.079	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

---

Data Download Date/Time: 12/31/2020 22:17



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MASSIP,LEONARDO  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: CARRILLE

DATE/TIME OF TEST: 11/20/2020 01:20:40

---

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.081	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.247	01:16
Breath Volume		1.703	01:16
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.245	01:19
Breath Volume		1.484	01:19
Air Blank		0.000	01:19
Control Test		0.081	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

---

Data Download Date/Time: 12/31/2020 22:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: RODRIGUEZ,EMMANUEL  
ARRESTING OFFICER AGENCY: Altamonte Springs PD  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: RODRIGUEZ-FRANQUI

DATE/TIME OF TEST: 11/20/2020 04:43:57

---

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.082	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.185	04:39
Breath Volume	1.824	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.184	04:42
Breath Volume	3.796	04:42
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

---

Data Download Date/Time: 12/31/2020 22:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: FUENTES,PEREYRA  
ARRESTING OFFICER AGENCY: Sanford PD  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MENDEZ BARMARA

DATE/TIME OF TEST: 11/20/2020 22:00:16

---

Diagnostics Check	OK	21:53
Air Blank	0.000	21:54
Control Test	0.081	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.173	21:55
Breath Volume	2.082	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.183	21:58
Breath Volume	1.746	21:58
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

---

Data Download Date/Time: 12/31/2020 22:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: SMITH,BLAKE  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: BOJARA

DATE/TIME OF TEST: 11/21/2020 01:55:55

---

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:45	
Control Test	0.081	01:45	
Air Blank	0.000	01:45	
Subject Sample	# 1	01:48	No Sample Provided
Breath Volume	0.000	01:48	
Air Blank	0.000	01:49	
Air Blank	0.000	01:51	
Subject Sample	# 2	01:54	No Sample Provided
Breath Volume	0.000	01:54	
Air Blank	0.000	01:55	
Control Test	0.081	01:55	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

---

Data Download Date/Time: 12/31/2020 22:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: SAAVERDRA,JUAN  
ARRESTING OFFICER AGENCY: Casselberry PD  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: DUARTE

DATE/TIME OF TEST: 11/21/2020 03:22:54

---

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.080	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.212	03:18
Breath Volume		3.277	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:20
Subject Sample	# 2	0.210	03:21
Breath Volume		3.296	03:21
Air Blank		0.000	03:22
Control Test		0.080	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

---

Data Download Date/Time: 12/31/2020 22:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: PFANNERSTILL,JACOB  
ARRESTING OFFICER AGENCY: Altamonte Springs PD  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: BRIDDELL

DATE/TIME OF TEST: 11/23/2020 04:06:02

---

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.082	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.159	04:01
Breath Volume	2.421	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:03
Subject Sample # 2	0.146	04:04
Breath Volume	2.359	04:04
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:06

---

Data Download Date/Time: 12/31/2020 22:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MACGILLIVARY, SEAN  
ARRESTING OFFICER AGENCY: Lake Mary PD  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: MERCEDES JR

DATE/TIME OF TEST: 11/26/2020 02:25:42

---

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.161	02:21
Breath Volume	1.976	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.162	02:24
Breath Volume	2.070	02:24
Air Blank	0.000	02:24
Control Test	0.081	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

---

Data Download Date/Time: 12/31/2020 22:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: QUINTANA,ABEL  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: LABORDE

DATE/TIME OF TEST: 11/28/2020 01:51:55

---

Diagnostics Check		OK	01:45
Air Blank		0.000	01:45
Control Test		0.081	01:46
Air Blank		0.000	01:46
Subject Sample	# 1	0.174	01:47
Breath Volume		4.230	01:47
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.157	01:50
Breath Volume		4.277	01:50
Air Blank		0.000	01:51
Control Test		0.081	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

---

Data Download Date/Time: 12/31/2020 22:17



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: TENNANT, CHRIS  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: DAVILA-GALAN

DATE/TIME OF TEST: 11/30/2020 02:01:10

---

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.082	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.092	01:56
Breath Volume	2.304	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.095	01:59
Breath Volume	2.062	01:59
Air Blank	0.000	02:00
Control Test	0.081	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

---

Data Download Date/Time: 12/31/2020 22:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916501  
DRY GAS EXPIRY DATE: 06/14/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 11/28/2020 22:14:59

---

Diagnostics Check	OK	22:07	
Air Blank	0.000	22:08	
Control Test	0.080	22:08	
Air Blank	0.000	22:09	
Subject Sample # 1	0.032	22:10	
Breath Volume	1.425	22:10	
Air Blank	0.000	22:11	
Air Blank	0.000	22:12	
Subject Sample # 2	0.029	22:13	
Breath Volume	1.718	22:13	
Air Blank	0.000	22:14	
Control Test	0.077	22:14	
Air Blank		22:14	RFI Detect
Air Blank		22:14	RFI Detect

---

Data Download Date/Time: 12/05/2020 04:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: HIRSHMAN,BRIAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916501  
DRY GAS EXPIRY DATE: 06/14/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 11/28/2020 23:06:08

---

Diagnostics Check	OK	22:59
Air Blank	0.000	23:00
Control Test	0.079	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.151	23:01
Breath Volume	1.925	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:03
Subject Sample # 2	0.147	23:04
Breath Volume	1.496	23:04
Air Blank	0.000	23:05
Control Test	0.078	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

---

Data Download Date/Time: 12/05/2020 04:19