

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,CRISTINO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: SMITH,DANIEL
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:52

SUBJECT'S LAST NAME: MONFISTON

DATE/TIME OF TEST: 11/06/2020 08:28:16

Diagnostics Check	OK	08:25	
Air Blank	0.000	08:25	
Control Test	0.079	08:26	
Air Blank	0.000	08:26	
Subject Sample	# 1	08:26	Subject Test Refused
Breath Volume	0.000	08:26	
Air Blank	0.000	08:27	
Control Test	0.079	08:27	
Air Blank	0.000	08:28	
Diagnostics Check	OK	08:28	

Data Download Date/Time: 11/10/2020 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKER,ALEXA,C
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: BECKER,ALEXA
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 11/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 11/16/2020 23:13:28

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:11	
Control Test	0.078	23:11	
Air Blank	0.000	23:11	
Subject Sample	# 1	23:12	Subject Test Refused
Breath Volume	0.000	23:12	
Air Blank	0.000	23:12	
Control Test	0.079	23:12	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:13	

Data Download Date/Time: 12/30/2020 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,R,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: RUBIDO,E
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 11/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: CARRANZA

DATE/TIME OF TEST: 11/17/2020 05:09:12

Diagnostics Check	OK	05:06	
Air Blank	0.000	05:06	
Control Test	0.079	05:07	
Air Blank	0.000	05:07	
Subject Sample	# 1	05:07	Subject Test Refused
Breath Volume	0.000	05:07	
Air Blank	0.000	05:08	
Control Test	0.079	05:08	
Air Blank	0.000	05:09	
Diagnostics Check	OK	05:09	

Data Download Date/Time: 12/30/2020 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,ANDREW,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: SISNEY,WHITNEY
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 11/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 11/19/2020 21:33:22

Diagnostics Check	OK	21:32	
Air Blank		21:32	RFI Detect
Air Blank	0.000	21:33	

Data Download Date/Time: 12/30/2020 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,ANDREW,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: SISNEY,WHITNEY
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 11/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: HARRIS JR

DATE/TIME OF TEST: 11/19/2020 21:39:09

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:36	
Control Test	0.080	21:36	
Air Blank	0.000	21:37	
Subject Sample	# 1	21:37	Subject Test Refused
Breath Volume	0.000	21:37	
Air Blank	0.000	21:38	
Control Test	0.080	21:38	
Air Blank	0.000	21:39	
Diagnostics Check	OK	21:39	

Data Download Date/Time: 12/30/2020 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEXANDER,E
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: CRESPO,CHRISTIAN
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 11/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: DELGADO ACOSTA

DATE/TIME OF TEST: 11/21/2020 05:22:14

Diagnostics Check		OK	05:14
Air Blank		0.000	05:15
Control Test		0.079	05:15
Air Blank		0.000	05:16
Subject Sample	# 1	0.270	05:17
Breath Volume		2.421	05:17
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 2	0.270	05:20
Breath Volume		3.339	05:20
Air Blank		0.000	05:21
Control Test		0.078	05:21
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 12/30/2020 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEXANDER,E
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: SANCHEZ,JENNY
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 11/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 11/29/2020 07:09:39

Diagnostics Check	OK	07:06	
Air Blank	0.000	07:07	
Control Test	0.079	07:07	
Air Blank	0.000	07:08	
Subject Sample	# 1	07:08	Subject Test Refused
Breath Volume	0.000	07:08	
Air Blank	0.000	07:08	
Control Test	0.079	07:09	
Air Blank	0.000	07:09	
Diagnostics Check	OK	07:09	

Data Download Date/Time: 12/30/2020 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEXANDER,E
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: RODRIGUEZ,ALEXANDER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 11/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GALDAMEZ

DATE/TIME OF TEST: 11/30/2020 01:53:54

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.081	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:52	
Control Test	0.081	01:53	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 12/30/2020 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEETZ,JOHN,W
BREATH TEST OPERATOR AGENCY: Calhoun County SO
ARRESTING OFFICER: SCHEETZ,JOHN
ARRESTING OFFICER AGENCY: Calhoun County SO
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: FERRELL

DATE/TIME OF TEST: 11/13/2020 15:36:03

Diagnostics Check	OK	15:30
Air Blank	0.000	15:30
Control Test	0.081	15:31
Air Blank	0.000	15:31
Subject Sample # 1	0.000	15:31
Breath Volume	2.757	15:31
Air Blank	0.000	15:32
Air Blank	0.000	15:34
Subject Sample # 2	0.000	15:34
Breath Volume	2.125	15:34
Air Blank	0.000	15:35
Control Test	0.081	15:35
Air Blank	0.000	15:35
Diagnostics Check	OK	15:36

Data Download Date/Time: 12/10/2020 17:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEETZ,JOHN,W
BREATH TEST OPERATOR AGENCY: Calhoun County SO
ARRESTING OFFICER: SCHEETZ,JOHN
ARRESTING OFFICER AGENCY: Calhoun County SO
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: PROB/PAROL
ARREST TIME: 07:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: FERRELL

DATE/TIME OF TEST: 11/30/2020 08:20:21

Diagnostics Check	OK	08:14
Air Blank	0.000	08:14
Control Test	0.080	08:15
Air Blank	0.000	08:15
Subject Sample # 1	0.000	08:16
Breath Volume	2.167	08:16
Air Blank	0.000	08:16
Air Blank	0.000	08:18
Subject Sample # 2	0.000	08:18
Breath Volume	2.476	08:18
Air Blank	0.000	08:19
Control Test	0.080	08:19
Air Blank	0.000	08:20
Diagnostics Check	OK	08:20

Data Download Date/Time: 12/10/2020 17:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 10/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/01/2020 03:19:25

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:16	
Control Test	0.080	03:16	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Control Test	0.077	03:18	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 11/25/2020 12:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: COSNER,STEVEN
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 10/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: GERMAIN

DATE/TIME OF TEST: 11/01/2020 05:26:57

Diagnostics Check	OK	05:20
Air Blank	0.000	05:20
Control Test	0.080	05:21
Air Blank	0.000	05:21
Subject Sample # 1	0.093	05:22
Breath Volume	1.917	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.092	05:25
Breath Volume	2.578	05:25
Air Blank	0.000	05:26
Control Test	0.080	05:26
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 11/25/2020 12:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: MELENDEZ,DANIEL
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 10/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 11/02/2020 02:00:23

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.160	01:52
Breath Volume	2.890	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.130	01:56
Breath Volume	1.839	01:56
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 3	0.146	01:58
Breath Volume	3.117	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 11/25/2020 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: MORENO,DANIELLA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 10/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 11/08/2020 02:33:15

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:29	
Control Test	0.080	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:32	
Control Test	0.080	02:32	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

Data Download Date/Time: 11/25/2020 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: MELENDEZ,DANIEL
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 10/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: CONCHA

DATE/TIME OF TEST: 11/10/2020 04:19:41

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.190	04:14
Breath Volume	1.304	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.188	04:18
Breath Volume	1.480	04:18
Air Blank	0.000	04:18
Control Test	0.079	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 11/25/2020 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGODA,JOSHUA,S
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: VEGODA,JOSHUA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 10/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 14:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:52

SUBJECT'S LAST NAME: TAPIA

DATE/TIME OF TEST: 11/14/2020 16:03:16

Diagnostics Check	OK	16:00	
Air Blank	0.000	16:00	
Control Test	0.080	16:00	
Air Blank	0.000	16:01	
Subject Sample	# 1	16:01	Subject Test Refused
Breath Volume	0.000	16:01	
Air Blank	0.000	16:02	
Control Test	0.079	16:02	
Air Blank	0.000	16:03	
Diagnostics Check	OK	16:03	

Data Download Date/Time: 11/25/2020 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: CARBALLO,PETER
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 10/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: BETANCOURT

DATE/TIME OF TEST: 11/15/2020 06:07:24

Diagnostics Check	OK	06:00
Air Blank	0.000	06:00
Control Test	0.080	06:01
Air Blank	0.000	06:01
Subject Sample # 1	0.114	06:02
Breath Volume	1.355	06:02
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 2	0.113	06:05
Breath Volume	1.726	06:05
Air Blank	0.000	06:06
Control Test	0.080	06:06
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

Data Download Date/Time: 11/25/2020 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN-
ROSQUETE,STEPHANNE,J
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: BROWN-
ROSQUETE,STEPHANNE
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 10/29/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022

VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male

OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: JOSHUA

DATE/TIME OF TEST: 11/16/2020 00:40:39

Diagnostics Check		OK	00:28	
Air Blank		0.000	00:29	
Control Test		0.080	00:29	
Air Blank		0.000	00:30	
Subject Sample	# 1	0.233	00:33	Volume Not Met
Breath Volume		0.632	00:33	
Air Blank		0.000	00:34	
Air Blank		0.000	00:35	
Subject Sample	# 2	0.279	00:36	
Breath Volume		2.410	00:36	
Air Blank		0.000	00:36	
Air Blank		0.000	00:38	
Subject Sample	# 3	0.284	00:39	
Breath Volume		2.339	00:39	
Air Blank		0.000	00:39	
Control Test		0.079	00:40	
Air Blank		0.000	00:40	
Diagnostics Check		OK	00:40	

Data Download Date/Time: 11/25/2020 12:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,PEDRO,A
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: BLANDON,ALLAN
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 12:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 11/30/2020 13:41:04

Diagnostics Check	OK	13:31
Air Blank	0.000	13:31
Control Test	0.080	13:32
Air Blank	0.000	13:32
Subject Sample # 1	0.173	13:33
Breath Volume	1.242	13:33
Air Blank	0.000	13:33
Air Blank	0.000	13:35
Subject Sample # 2	0.196	13:36
Breath Volume	2.218	13:36
Air Blank	0.000	13:37
Air Blank	0.000	13:38
Subject Sample # 3	0.187	13:39
Breath Volume	1.699	13:39
Air Blank	0.000	13:40
Control Test	0.080	13:40
Air Blank	0.000	13:40
Diagnostics Check	OK	13:41

Data Download Date/Time: 12/21/2020 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANUEVA, MAURICIO,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: GOMEZ, RICHARD
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GIBERT

DATE/TIME OF TEST: 11/27/2020 03:00:37

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.151	02:56
Breath Volume	2.175	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.154	02:59
Breath Volume	1.921	02:59
Air Blank	0.000	02:59
Control Test	0.079	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 12/19/2020 02:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOSTALY, STEVENS,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: GOMEZ, RICHARD
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: KOSTADINOV

DATE/TIME OF TEST: 11/28/2020 02:03:31

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.146	01:59
Breath Volume	2.238	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.139	02:02
Breath Volume	1.414	02:02
Air Blank	0.000	02:02
Control Test	0.078	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 12/19/2020 02:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/05/2020 23:21:39

Diagnostics Check	OK	23:18	
Air Blank	0.000	23:19	
Control Test	0.078	23:19	
Air Blank	0.000	23:19	
Subject Sample	# 1	23:20	Subject Test Refused
Breath Volume	0.000	23:20	
Air Blank	0.000	23:20	
Control Test	0.079	23:21	
Air Blank	0.000	23:21	
Diagnostics Check	OK	23:21	

Data Download Date/Time: 11/25/2020 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPPE,ZACHARY,D
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: JACOBS,GARRETT
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 11/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 17:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:47

SUBJECT'S LAST NAME: RUNDGREN

DATE/TIME OF TEST: 11/29/2020 18:19:13

Diagnostics Check	OK	18:10
Air Blank	0.000	18:10
Control Test	0.081	18:11
Air Blank	0.000	18:11
Subject Sample # 1	0.141	18:14
Breath Volume	1.191	18:14
Air Blank	0.000	18:15
Air Blank	0.000	18:17
Subject Sample # 2	0.140	18:17
Breath Volume	1.773	18:17
Air Blank	0.000	18:18
Control Test	0.081	18:18
Air Blank	0.000	18:19
Diagnostics Check	OK	18:19

Data Download Date/Time: 12/29/2020 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 11/10/2020 00:12:52

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:06	
Control Test	0.077	00:07	
Air Blank	0.000	00:07	
Subject Sample	# 1 0.221	00:08	
Breath Volume	1.359	00:08	
Air Blank	0.000	00:08	
Air Blank	0.000	00:10	
Subject Sample	# 2	00:11	Subject Test Refused
Breath Volume	0.000	00:11	
Air Blank	0.000	00:11	
Control Test	0.079	00:12	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 11/23/2020 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 11/19/2020 22:20:08

Diagnostics Check	OK	22:13
Air Blank	0.000	22:14
Control Test	0.081	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.084	22:15
Breath Volume	2.566	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.089	22:18
Breath Volume	2.632	22:18
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 11/23/2020 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TEST,CLINE
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: SYS CHECK
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 11/24/2020 14:01:33

Diagnostics Check	OK	13:57	
Air Blank	0.000	13:57	
Control Test	0.079	13:57	
Air Blank	0.000	13:58	
Subject Sample # 1	0.000	13:58	
Breath Volume	1.445	13:58	
Air Blank	0.000	13:59	
Air Blank		14:00	RFI Detect
Air Blank		14:01	RFI Detect
Air Blank	0.000	14:01	

Data Download Date/Time: 12/17/2020 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: BOPP

DATE/TIME OF TEST: 11/28/2020 20:46:07

Diagnostics Check	OK	20:40
Air Blank	0.000	20:40
Control Test	0.079	20:40
Air Blank	0.000	20:41
Subject Sample # 1	0.000	20:41
Breath Volume	2.687	20:41
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample # 2	0.000	20:44
Breath Volume	2.609	20:44
Air Blank	0.000	20:45
Control Test	0.078	20:45
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

Data Download Date/Time: 12/17/2020 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BACON,DEDRICK
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: HERNANDEZ-LOPEZ

DATE/TIME OF TEST: 11/29/2020 00:29:26

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.079	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.192	00:25
Breath Volume	2.679	00:25
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.194	00:27
Breath Volume	2.652	00:27
Air Blank	0.000	00:28
Control Test	0.078	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 12/17/2020 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: DELGADO CASTRO

DATE/TIME OF TEST: 11/29/2020 02:03:03

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.154	01:57
Breath Volume	1.222	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.166	02:01
Breath Volume	1.382	02:01
Air Blank	0.000	02:02
Control Test	0.077	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:03

Data Download Date/Time: 12/17/2020 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BAUTISTA

DATE/TIME OF TEST: 11/29/2020 02:44:08

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.234	02:39
Breath Volume	2.214	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:41
Subject Sample # 2	0.245	02:42
Breath Volume	1.980	02:42
Air Blank	0.000	02:43
Control Test	0.077	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 12/17/2020 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICKEY, WILLIAM, Q
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: VANHEERDEN, TY
ARRESTING OFFICER AGENCY: PORT ST. JOE PD
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: WEBSTER

DATE/TIME OF TEST: 11/28/2020 02:41:11

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.078	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.197	02:36
Breath Volume	2.421	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.203	02:39
Breath Volume	2.246	02:39
Air Blank	0.000	02:40
Control Test	0.078	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 12/30/2020 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBIN,GASTON,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: SANTIAGO,JAFET
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: SHUKLA

DATE/TIME OF TEST: 11/05/2020 19:29:39

Diagnostics Check		OK	19:15
Air Blank		0.000	19:15
Control Test		0.080	19:16
Air Blank		0.000	19:16
Subject Sample	# 1	0.325	19:19
Breath Volume		1.597	19:19
Air Blank		0.000	19:19
Air Blank		0.000	19:21
Subject Sample	# 2	0.294	19:23
Breath Volume		1.292	19:23
Air Blank		0.000	19:24
Air Blank		0.000	19:26
Subject Sample	# 3	0.315	19:28
Breath Volume		1.652	19:28
Air Blank		0.000	19:28
Control Test		0.079	19:29
Air Blank		0.000	19:29
Diagnostics Check		OK	19:29

Data Download Date/Time: 11/30/2020 18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WINSTEAD,JUSTIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 11/07/2020 02:47:24

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:39	
Control Test	0.081	02:39	
Air Blank	0.000	02:40	
Subject Sample	# 1	02:43	No Sample Provided
Breath Volume	0.000	02:43	
Air Blank	0.000	02:43	
Air Blank	0.000	02:45	
Subject Sample	# 2	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:46	
Control Test	0.081	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 11/30/2020 18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MARK,E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HANIGH,M
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:27

SUBJECT'S LAST NAME: NAIL

DATE/TIME OF TEST: 11/07/2020 14:57:17

Diagnostics Check	OK	14:50
Air Blank	0.000	14:51
Control Test	0.080	14:51
Air Blank	0.000	14:52
Subject Sample # 1	0.000	14:53
Breath Volume	1.769	14:53
Air Blank	0.000	14:53
Air Blank	0.000	14:55
Subject Sample # 2	0.000	14:55
Breath Volume	1.886	14:55
Air Blank	0.000	14:56
Control Test	0.079	14:56
Air Blank	0.000	14:57
Diagnostics Check	OK	14:57

Data Download Date/Time: 11/30/2020 18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIDNER,CODY,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LEME,SANTIAGO
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: O'BRYAN

DATE/TIME OF TEST: 11/09/2020 05:39:49

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.081	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.000	05:34
Breath Volume	1.445	05:34
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.000	05:38
Breath Volume	3.417	05:38
Air Blank	0.000	05:38
Control Test	0.080	05:39
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

Data Download Date/Time: 11/30/2020 18:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIDNER,CODY,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WEIDNER,CODY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/17/2020 20:49:50

Diagnostics Check	OK	20:42
Air Blank	0.000	20:43
Control Test	0.082	20:43
Air Blank	0.000	20:43
Subject Sample # 1	0.000	20:44
Breath Volume	1.632	20:44
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.000	20:48
Breath Volume	1.316	20:48
Air Blank	0.000	20:48
Control Test	0.081	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 11/30/2020 18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELIONE, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCLAUGHLIN, BRADLEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: COLLETTE

DATE/TIME OF TEST: 11/18/2020 01:55:53

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.082	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.197	01:51
Breath Volume	3.539	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:53
Subject Sample # 2	0.193	01:54
Breath Volume	3.894	01:54
Air Blank	0.000	01:54
Control Test	0.080	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 11/30/2020 18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBIN,GASTON,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROJAS,E
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: EATON

DATE/TIME OF TEST: 11/19/2020 23:50:25

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.081	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.000	23:45
Breath Volume	2.132	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.000	23:49
Breath Volume	1.972	23:49
Air Blank	0.000	23:49
Control Test	0.081	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 11/30/2020 18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIDNER,CODY,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCLAUGHLIN,BRADLEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 11/21/2020 05:06:35

Diagnostics Check		OK	04:59
Air Blank		0.000	05:00
Control Test		0.081	05:00
Air Blank		0.000	05:01
Subject Sample	# 1	0.178	05:01
Breath Volume		1.605	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.179	05:05
Breath Volume		1.632	05:05
Air Blank		0.000	05:05
Control Test		0.080	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 11/30/2020 18:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ESPINOSA,ANGEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: MADERA

DATE/TIME OF TEST: 11/22/2020 06:17:11

Diagnostics Check		OK	06:06
Air Blank		0.000	06:07
Control Test		0.080	06:07
Air Blank		0.000	06:08
Subject Sample	# 1	0.185	06:11
Breath Volume		1.792	06:11
Air Blank		0.000	06:11
Air Blank		0.000	06:13
Subject Sample	# 2	0.165	06:15
Breath Volume		1.453	06:15
Air Blank		0.000	06:16
Control Test		0.078	06:16
Air Blank		0.000	06:17
Diagnostics Check		OK	06:17

Data Download Date/Time: 11/30/2020 18:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBIN,GASTON,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BARRERA,C
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: RAZZINO

DATE/TIME OF TEST: 11/23/2020 21:53:43

Diagnostics Check	OK	21:48	
Air Blank	0.000	21:49	
Control Test	0.081	21:49	
Air Blank	0.000	21:49	
Subject Sample	# 1	21:52	Subject Test Refused
Breath Volume	0.000	21:52	
Air Blank	0.000	21:52	
Control Test	0.081	21:53	
Air Blank	0.000	21:53	
Diagnostics Check	OK	21:53	

Data Download Date/Time: 11/30/2020 18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WAGNER,ROBERT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: SELLARDS

DATE/TIME OF TEST: 11/26/2020 17:20:07

Diagnostics Check	OK	17:13
Air Blank	0.000	17:13
Control Test	0.081	17:14
Air Blank	0.000	17:14
Subject Sample # 1	0.169	17:15
Breath Volume	2.242	17:15
Air Blank	0.000	17:16
Air Blank	0.000	17:17
Subject Sample # 2	0.165	17:18
Breath Volume	2.527	17:18
Air Blank	0.000	17:19
Control Test	0.079	17:19
Air Blank	0.000	17:20
Diagnostics Check	OK	17:20

Data Download Date/Time: 11/30/2020 18:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBIN,GASTON,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROBIN,GASTON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: HENSON

DATE/TIME OF TEST: 11/28/2020 05:49:47

Diagnostics Check	OK	05:40
Air Blank	0.000	05:40
Control Test	0.080	05:40
Air Blank	0.000	05:41
Subject Sample # 1	0.169	05:41
Breath Volume	2.230	05:41
Air Blank	0.000	05:42
Air Blank	0.000	05:44
Subject Sample # 2	0.196	05:45
Breath Volume	2.105	05:45
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 3	0.188	05:48
Breath Volume	3.281	05:48
Air Blank	0.000	05:48
Control Test	0.080	05:49
Air Blank	0.000	05:49
Diagnostics Check	OK	05:49

Data Download Date/Time: 11/30/2020 18:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBIN,GASTON,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROBIN,GASTON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: BISGARD

DATE/TIME OF TEST: 11/30/2020 00:14:17

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.080	00:08
Air Blank	0.000	00:08
Subject Sample # 1	0.183	00:09
Breath Volume	1.687	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:11
Subject Sample # 2	0.173	00:12
Breath Volume	1.601	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 11/30/2020 18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HERNANDEZ TORRES

DATE/TIME OF TEST: 11/01/2020 03:21:30

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.078	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.124	03:17
Breath Volume		2.292	03:17
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.125	03:20
Breath Volume		2.199	03:20
Air Blank		0.000	03:20
Control Test		0.078	03:20
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 11/10/2020 15:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EMANS,ALEX,R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GROSSKLAS,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: SWISHER

DATE/TIME OF TEST: 11/01/2020 03:51:26

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.078	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.000	03:47
Breath Volume	1.484	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.000	03:50
Breath Volume	1.382	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 11/10/2020 15:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,S,
BREATH TEST OPERATOR AGENCY: FHP TROOP F
ARRESTING OFFICER: TERUZ,G
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: TRUONG

DATE/TIME OF TEST: 11/02/2020 01:50:13

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.078	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.067	01:45
Breath Volume	1.167	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.067	01:48
Breath Volume	1.359	01:48
Air Blank	0.000	01:49
Control Test	0.078	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 11/10/2020 15:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 11/02/2020 02:03:12

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:53	
Control Test		0.078	01:53	
Air Blank		0.000	01:54	
Subject Sample	# 1	0.142	01:57	Volume Not Met
Breath Volume		0.167	01:57	
Air Blank		0.000	01:57	
Air Blank		0.000	01:59	
Subject Sample	# 2		02:01	Subject Test Refused
Breath Volume		0.000	02:01	
Air Blank		0.000	02:02	
Control Test		0.077	02:02	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 11/10/2020 15:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOEL,Y,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BIERMAN,AMANDA
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: FRIEDLANDER

DATE/TIME OF TEST: 11/04/2020 03:33:17

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.078	03:21	
Air Blank	0.000	03:22	
Subject Sample	# 1 0.230	03:23	
Breath Volume	2.023	03:23	
Air Blank	0.000	03:24	
Air Blank	0.000	03:26	
Subject Sample	# 2 0.189	03:29	Volume Not Met
Breath Volume	0.539	03:29	
Air Blank	0.000	03:29	
Air Blank	0.000	03:31	
Subject Sample	# 3	03:31	Subject Test Refused
Breath Volume	0.000	03:31	
Air Blank	0.000	03:32	
Control Test	0.078	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 11/10/2020 15:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,ANDREW,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WOODS,HUNTER
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ANDRES

DATE/TIME OF TEST: 11/01/2020 02:43:29

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.209	02:38
Breath Volume	1.285	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.225	02:41
Breath Volume	2.550	02:41
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 11/17/2020 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,THOMAS,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ENDELMANN,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 11/07/2020 01:11:41

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.081	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.156	01:07
Breath Volume	4.054	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.157	01:10
Breath Volume	3.550	01:10
Air Blank	0.000	01:10
Control Test	0.080	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 12/08/2020 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NOVA,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 11/07/2020 02:54:51

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.161	02:50
Breath Volume	1.140	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.147	02:53
Breath Volume	1.320	02:53
Air Blank	0.000	02:53
Control Test	0.080	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 12/08/2020 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MENDOZA,URIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: HERNANEZ AGUILAR

DATE/TIME OF TEST: 11/07/2020 20:10:22

Diagnostics Check		OK	19:58	
Air Blank		0.000	19:58	
Control Test		0.080	19:59	
Air Blank		0.000	19:59	
Subject Sample	# 1	0.122	20:02	Volume Not Met
Breath Volume		1.000	20:02	
Air Blank		0.000	20:03	
Air Blank		0.000	20:05	
Subject Sample	# 2	0.135	20:05	
Breath Volume		1.496	20:05	
Air Blank		0.000	20:06	
Air Blank		0.000	20:08	
Subject Sample	# 3	0.137	20:08	
Breath Volume		1.390	20:08	
Air Blank		0.000	20:09	
Control Test		0.079	20:09	
Air Blank		0.000	20:10	
Diagnostics Check		OK	20:10	

Data Download Date/Time: 12/08/2020 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,N
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MORRIS,H
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 11/08/2020 00:35:08

Diagnostics Check		OK	00:19	
Air Blank		0.000	00:20	
Control Test		0.080	00:20	
Air Blank		0.000	00:21	
Subject Sample	# 1	0.124	00:22	
Breath Volume		1.242	00:22	
Air Blank		0.000	00:22	
Air Blank		0.000	00:24	
Subject Sample	# 2	0.121	00:27	Volume Not Met
Breath Volume		0.777	00:27	
Air Blank		0.000	00:28	
Air Blank		0.000	00:30	
Subject Sample	# 3	0.106	00:33	Volume Not Met
Breath Volume		0.894	00:33	
Air Blank		0.000	00:34	
Control Test		0.079	00:34	
Air Blank		0.000	00:35	
Diagnostics Check		OK	00:35	

Data Download Date/Time: 12/08/2020 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,N
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MORRIS,H
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 11/08/2020 00:50:27

Diagnostics Check		OK	00:38	
Air Blank		0.000	00:39	
Control Test		0.079	00:39	
Air Blank		0.000	00:40	
Subject Sample	# 1	0.087	00:43	Volume Not Met
Breath Volume		0.121	00:43	
Air Blank		0.000	00:43	
Air Blank		0.000	00:45	
Subject Sample	# 2	0.105	00:48	Volume Not Met
Breath Volume		0.078	00:48	
Air Blank		0.000	00:49	
Control Test		0.080	00:49	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 12/08/2020 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,S,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MULRY,T
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: BRONCATTI

DATE/TIME OF TEST: 11/08/2020 04:45:06

Diagnostics Check		OK	04:38
Air Blank		0.000	04:39
Control Test		0.080	04:39
Air Blank		0.000	04:40
Subject Sample	# 1	0.262	04:40
Breath Volume		1.781	04:40
Air Blank		0.000	04:41
Air Blank		0.000	04:43
Subject Sample	# 2	0.263	04:43
Breath Volume		1.984	04:43
Air Blank		0.000	04:44
Control Test		0.079	04:44
Air Blank		0.000	04:45
Diagnostics Check		OK	04:45

Data Download Date/Time: 12/08/2020 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STEARNS,M
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: TERIAN

DATE/TIME OF TEST: 11/11/2020 23:40:36

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.108	23:35
Breath Volume	1.910	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.116	23:39
Breath Volume	1.605	23:39
Air Blank	0.000	23:39
Control Test	0.079	23:40
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 12/08/2020 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEAR,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: AK

DATE/TIME OF TEST: 11/13/2020 02:35:18

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.087	02:30
Breath Volume	2.875	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.081	02:33
Breath Volume	2.132	02:33
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 12/08/2020 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARBOGAST, JOSHUA, S
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: KNUTSON, PETER
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: DISNEY

DATE/TIME OF TEST: 11/13/2020 20:31:33

Diagnostics Check	OK	20:24
Air Blank	0.000	20:25
Control Test	0.080	20:25
Air Blank	0.000	20:26
Subject Sample # 1	0.132	20:26
Breath Volume	1.675	20:26
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.134	20:29
Breath Volume	1.617	20:29
Air Blank	0.000	20:30
Control Test	0.080	20:30
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

Data Download Date/Time: 12/08/2020 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THROESCH,PATRICK,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SOLIS,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: OBERMAN

DATE/TIME OF TEST: 11/14/2020 03:05:26

Diagnostics Check		OK	02:58
Air Blank		0.000	02:58
Control Test		0.080	02:59
Air Blank		0.000	02:59
Subject Sample	# 1	0.225	03:00
Breath Volume		4.968	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.219	03:03
Breath Volume		4.183	03:03
Air Blank		0.000	03:04
Control Test		0.080	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 12/08/2020 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,E,W
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KELLY,M
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: MCDONOUGH

DATE/TIME OF TEST: 11/14/2020 03:28:03

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.125	03:23
Breath Volume	2.871	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.122	03:26
Breath Volume	2.722	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:28

Data Download Date/Time: 12/08/2020 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,S,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TERUZ,G
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: RAMOS CARRILLO

DATE/TIME OF TEST: 11/15/2020 01:19:23

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.096	01:14
Breath Volume	1.449	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.092	01:17
Breath Volume	1.382	01:17
Air Blank	0.000	01:18
Control Test	0.078	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 12/08/2020 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLMSTEAD,J,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STARN,S,N
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: ROSAIRO

DATE/TIME OF TEST: 11/16/2020 00:03:13

Diagnostics Check	OK	23:56
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.142	23:58
Breath Volume	1.394	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.136	00:01
Breath Volume	1.269	00:01
Air Blank	0.000	00:02
Control Test	0.079	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 12/08/2020 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GASH, C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: SOLORIO

DATE/TIME OF TEST: 11/16/2020 01:31:08

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.080	01:25
Air Blank		0.000	01:25
Subject Sample	# 1	0.242	01:26
Breath Volume		2.675	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.235	01:29
Breath Volume		2.074	01:29
Air Blank		0.000	01:30
Control Test		0.079	01:30
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 12/08/2020 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBBONS,S,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SOLIS,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: VOLPE

DATE/TIME OF TEST: 11/16/2020 01:59:24

Diagnostics Check		OK	01:47	
Air Blank		0.000	01:47	
Control Test		0.079	01:48	
Air Blank		0.000	01:48	
Subject Sample	# 1	0.177	01:51	Volume Not Met
Breath Volume		1.109	01:51	
Air Blank		0.000	01:52	
Air Blank		0.000	01:54	
Subject Sample	# 2	0.195	01:54	
Breath Volume		1.531	01:54	
Air Blank		0.000	01:55	
Air Blank		0.000	01:57	
Subject Sample	# 3	0.176	01:57	
Breath Volume		1.722	01:57	
Air Blank		0.000	01:58	
Control Test		0.079	01:58	
Air Blank		0.000	01:59	
Diagnostics Check		OK	01:59	

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINT-FORT,P,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PSCHIGADA,W
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:35

SUBJECT'S LAST NAME: MARISSETTE

DATE/TIME OF TEST: 11/17/2020 14:41:16

Diagnostics Check	OK	14:27	
Air Blank	0.000	14:27	
Control Test	0.081	14:28	
Air Blank	0.000	14:28	
Subject Sample # 1	0.000	14:29	
Breath Volume	1.429	14:29	
Air Blank	0.000	14:29	
Air Blank	0.000	14:31	
Subject Sample # 2	0.000	14:35	Volume Not Met
Breath Volume	0.066	14:35	
Air Blank	0.000	14:35	
Air Blank	0.000	14:37	
Subject Sample # 3	0.000	14:39	
Breath Volume	1.707	14:39	
Air Blank	0.000	14:40	
Control Test	0.080	14:40	
Air Blank	0.000	14:41	
Diagnostics Check	OK	14:41	

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINT-FORT,P,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PSCHIGADA,W
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:35

SUBJECT'S LAST NAME: MARISSETTE

DATE/TIME OF TEST: 11/17/2020 14:59:36

Diagnostics Check	OK	14:48	
Air Blank	0.000	14:48	
Control Test	0.080	14:48	
Air Blank	0.000	14:49	
Subject Sample	# 1	14:52	No Sample Provided
Breath Volume	0.000	14:52	
Air Blank	0.000	14:53	
Air Blank	0.000	14:55	
Subject Sample	# 2	14:58	No Sample Provided
Breath Volume	0.000	14:58	
Air Blank	0.000	14:58	
Control Test	0.080	14:59	
Air Blank	0.000	14:59	
Diagnostics Check	OK	14:59	

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOEL,Y,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SILL,N
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: ELUSME

DATE/TIME OF TEST: 11/18/2020 02:05:55

Diagnostics Check		OK	01:58
Air Blank		0.000	01:59
Control Test		0.081	01:59
Air Blank		0.000	02:00
Subject Sample	# 1	0.203	02:01
Breath Volume		2.554	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:03
Subject Sample	# 2	0.204	02:04
Breath Volume		2.421	02:04
Air Blank		0.000	02:04
Control Test		0.080	02:05
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,E,W
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCCOY,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 11/19/2020 20:38:10

Diagnostics Check	OK	20:35	
Air Blank	0.000	20:35	
Control Test	0.081	20:35	
Air Blank	0.000	20:36	
Subject Sample	# 1	20:36	Subject Test Refused
Breath Volume	0.000	20:36	
Air Blank	0.000	20:37	
Control Test	0.081	20:37	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,S,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SHERAN,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: TORIBIO JACOBO

DATE/TIME OF TEST: 11/19/2020 23:41:17

Diagnostics Check	OK	23:34
Air Blank	0.000	23:35
Control Test	0.081	23:35
Air Blank	0.000	23:35
Subject Sample # 1	0.206	23:36
Breath Volume	1.355	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.220	23:39
Breath Volume	2.273	23:39
Air Blank	0.000	23:40
Control Test	0.080	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THROESCH,PATRICK,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SOLIS,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: HADDAD-JACOBS

DATE/TIME OF TEST: 11/20/2020 00:08:47

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.081	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.164	00:04
Breath Volume	1.527	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.176	00:07
Breath Volume	1.691	00:07
Air Blank	0.000	00:07
Control Test	0.081	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 11/20/2020 01:44:41

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.082	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.145	01:40
Breath Volume	1.402	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.142	01:43
Breath Volume	1.507	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LEONI,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 11/21/2020 00:52:09

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.081	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.172	00:47
Breath Volume	1.460	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:49
Subject Sample # 2	0.184	00:50
Breath Volume	2.097	00:50
Air Blank	0.000	00:51
Control Test	0.081	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,S,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MOLCAN,M
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: DROTOVICK

DATE/TIME OF TEST: 11/21/2020 23:08:51

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.081	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.241	23:04
Breath Volume	1.589	23:04
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.240	23:07
Breath Volume	2.210	23:07
Air Blank	0.000	23:07
Control Test	0.080	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVAS,W,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: APONTE,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: MENDEZPEREZ

DATE/TIME OF TEST: 11/22/2020 04:46:23

Diagnostics Check		OK	04:40
Air Blank		0.000	04:40
Control Test		0.082	04:41
Air Blank		0.000	04:41
Subject Sample	# 1	0.166	04:42
Breath Volume		1.792	04:42
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.158	04:44
Breath Volume		1.789	04:44
Air Blank		0.000	04:45
Control Test		0.081	04:45
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LEONI, S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: HURD

DATE/TIME OF TEST: 11/23/2020 20:04:43

Diagnostics Check	OK	19:58
Air Blank	0.000	19:58
Control Test	0.081	19:59
Air Blank	0.000	19:59
Subject Sample # 1	0.175	20:00
Breath Volume	1.570	20:00
Air Blank	0.000	20:00
Air Blank	0.000	20:02
Subject Sample # 2	0.177	20:03
Breath Volume	1.207	20:03
Air Blank	0.000	20:03
Control Test	0.080	20:04
Air Blank	0.000	20:04
Diagnostics Check	OK	20:04

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,E,W
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SACKMANN,A
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: HIMICK

DATE/TIME OF TEST: 11/25/2020 06:30:34

Diagnostics Check	OK	06:23
Air Blank	0.000	06:24
Control Test	0.081	06:24
Air Blank	0.000	06:25
Subject Sample # 1	0.161	06:25
Breath Volume	1.535	06:25
Air Blank	0.000	06:26
Air Blank	0.000	06:28
Subject Sample # 2	0.167	06:28
Breath Volume	1.300	06:28
Air Blank	0.000	06:29
Control Test	0.081	06:29
Air Blank	0.000	06:30
Diagnostics Check	OK	06:30

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER JR,S,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: LENNON

DATE/TIME OF TEST: 11/25/2020 22:02:38

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.081	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.310	21:57
Breath Volume	2.113	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.305	22:00
Breath Volume	1.718	22:00
Air Blank	0.000	22:01
Control Test	0.080	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBBONS,S,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LEONI,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: PLACE

DATE/TIME OF TEST: 11/28/2020 00:46:08

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.082	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.133	00:39
Breath Volume	2.125	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.130	00:44
Breath Volume	2.277	00:44
Air Blank	0.000	00:45
Control Test	0.081	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VAUGHAN,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: NEGOSHIAN

DATE/TIME OF TEST: 11/28/2020 05:13:34

Diagnostics Check	OK	05:07
Air Blank	0.000	05:07
Control Test	0.081	05:08
Air Blank	0.000	05:08
Subject Sample # 1	0.154	05:09
Breath Volume	1.507	05:09
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.154	05:12
Breath Volume	1.816	05:12
Air Blank	0.000	05:12
Control Test	0.081	05:12
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRATT,NICHOLAS,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PRATT,NICHOLAS
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: GOLWALA

DATE/TIME OF TEST: 11/28/2020 10:08:36

Diagnostics Check	OK	10:06	
Air Blank	0.000	10:06	
Control Test	0.081	10:06	
Air Blank	0.000	10:07	
Subject Sample # 1	0.000	10:07	
Breath Volume	2.214	10:07	
Air Blank		10:08	RFI Detect
Air Blank		10:08	RFI Detect
Air Blank	0.000	10:08	

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRATT,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PRATT,NICHOLAS
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: GOLWALA

DATE/TIME OF TEST: 11/28/2020 10:16:32

Diagnostics Check	OK	10:10	
Air Blank	0.000	10:11	
Control Test	0.081	10:11	
Air Blank	0.000	10:11	
Subject Sample # 1	0.000	10:12	
Breath Volume	2.050	10:12	
Air Blank	0.000	10:12	
Air Blank	0.000	10:14	
Subject Sample # 2		10:15	RFI Detect
Breath Volume	0.000	10:15	
Air Blank	0.000	10:15	
Control Test	0.080	10:15	
Air Blank	0.000	10:16	
Diagnostics Check	OK	10:16	

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRATT,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PRATT,NICHOLAS
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: GOLWALA

DATE/TIME OF TEST: 11/28/2020 10:41:16

Diagnostics Check	OK	10:35
Air Blank	0.000	10:35
Control Test	0.081	10:35
Air Blank	0.000	10:36
Subject Sample # 1	0.000	10:36
Breath Volume	2.773	10:36
Air Blank	0.000	10:37
Air Blank	0.000	10:39
Subject Sample # 2	0.000	10:39
Breath Volume	2.414	10:39
Air Blank	0.000	10:40
Control Test	0.081	10:40
Air Blank	0.000	10:41
Diagnostics Check	OK	10:41

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIS,KEVIN,W
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: RODRIGUEZ PITRE,VANESSA
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/28/2020 23:00:57

Diagnostics Check	OK	22:54
Air Blank	0.000	22:55
Control Test	0.081	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.000	22:56
Breath Volume	2.777	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.000	22:59
Breath Volume	2.789	22:59
Air Blank	0.000	23:00
Control Test	0.081	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIS,KEVIN,W
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: JACKSON,TAJAHS
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: BUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: LESSON

DATE/TIME OF TEST: 11/29/2020 19:26:00

Diagnostics Check	OK	19:19
Air Blank	0.000	19:20
Control Test	0.081	19:20
Air Blank	0.000	19:21
Subject Sample # 1	0.007	19:21
Breath Volume	2.960	19:21
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.007	19:24
Breath Volume	3.570	19:24
Air Blank	0.000	19:25
Control Test	0.081	19:25
Air Blank	0.000	19:25
Diagnostics Check	OK	19:26

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THROESCH,PATRICK,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KORPOLINSKI,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 11/30/2020 02:33:59

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.142	02:29
Breath Volume	3.386	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample # 2	0.136	02:32
Breath Volume	3.250	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 12/08/2020 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEPITONE, MICHAEL, A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CLENDENING, JAMES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 11/14/2020 13:47:50

Diagnostics Check	OK	13:44	
Air Blank	0.000	13:44	
Control Test	0.078	13:45	
Air Blank	0.000	13:45	
Subject Sample	# 1	13:46	RFI Detect
Breath Volume	0.000	13:46	
Air Blank	0.000	13:46	
Control Test	0.080	13:47	
Air Blank	0.000	13:47	
Diagnostics Check	OK	13:47	

Data Download Date/Time: 11/28/2020 15:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEPITONE, MICHAEL, A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CLENDENING, JAMES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 11/14/2020 14:00:12

Diagnostics Check	OK	13:50	
Air Blank	0.000	13:51	
Control Test	0.079	13:51	
Air Blank	0.000	13:52	
Subject Sample # 1	0.223	13:54	
Breath Volume	1.511	13:54	
Air Blank	0.000	13:56	
Air Blank	0.000	13:57	
Subject Sample # 2	0.222	13:58	
Breath Volume	1.773	13:58	
Air Blank		13:59	Purge Fail
Air Blank	0.000	14:00	

Data Download Date/Time: 12/02/2020 07:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MIDDLEBROOKS, ALLAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 11/14/2020 00:17:16

Diagnostics Check		OK	00:10
Air Blank		0.000	00:11
Control Test		0.079	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.000	00:12
Breath Volume		2.671	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.000	00:15
Breath Volume		2.742	00:15
Air Blank		0.000	00:16
Control Test		0.080	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 11/18/2020 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,GENE,A
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: DAVIS,TIMMY
ARRESTING OFFICER AGENCY: Apalachicola PD
LAST AGENCY INSPECTION: 11/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: CRUM

DATE/TIME OF TEST: 11/20/2020 19:41:28

Diagnostics Check	OK	19:34
Air Blank	0.000	19:35
Control Test	0.080	19:35
Air Blank	0.000	19:36
Subject Sample # 1	0.165	19:36
Breath Volume	1.945	19:36
Air Blank	0.000	19:37
Air Blank	0.000	19:39
Subject Sample # 2	0.169	19:39
Breath Volume	1.933	19:39
Air Blank	0.000	19:40
Control Test	0.078	19:40
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 12/17/2020 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTOPASSI,BARON,H
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: PETERSON,AMBER
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 11/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: K532538982550

DATE/TIME OF TEST: 11/21/2020 16:52:07

Diagnostics Check	OK	16:40	
Air Blank	0.000	16:41	
Control Test	0.079	16:41	
Air Blank	0.000	16:42	
Subject Sample	# 1	16:45	No Sample Provided
Breath Volume	0.000	16:45	
Air Blank	0.000	16:45	
Air Blank	0.000	16:47	
Subject Sample	# 2	16:50	No Sample Provided
Breath Volume	0.000	16:50	
Air Blank	0.000	16:51	
Control Test	0.080	16:51	
Air Blank	0.000	16:52	
Diagnostics Check	OK	16:52	

Data Download Date/Time: 12/17/2020 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,FAUSTO,A
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: ROSARIO,FAUSTO
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: R9087543321

DATE/TIME OF TEST: 11/03/2020 12:14:46

Diagnostics Check	OK	12:09
Air Blank	0.000	12:09
Control Test	0.078	12:09
Air Blank	0.000	12:10
Subject Sample # 1	0.000	12:10
Breath Volume	1.902	12:10
Air Blank	0.000	12:11
Air Blank	0.000	12:13
Subject Sample # 2	0.000	12:13
Breath Volume	1.781	12:13
Air Blank	0.000	12:13
Control Test	0.078	12:14
Air Blank	0.000	12:14
Diagnostics Check	OK	12:14

Data Download Date/Time: 12/09/2020 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARDO,ADALBERTO,
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: PARDO,ADALBERTO
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 11/04/2020 00:13:30

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:11	
Control Test	0.078	00:11	
Air Blank	0.000	00:11	
Subject Sample	# 1	00:12	Subject Test Refused
Breath Volume	0.000	00:12	
Air Blank	0.000	00:12	
Control Test	0.078	00:12	
Air Blank	0.000	00:13	
Diagnostics Check	OK	00:13	

Data Download Date/Time: 12/09/2020 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BRIAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ADAMCZYK,ANDREW
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 10/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: MAYNE

DATE/TIME OF TEST: 11/08/2020 02:14:40

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.079	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.009	02:10
Breath Volume	2.199	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.008	02:13
Breath Volume	2.152	02:13
Air Blank	0.000	02:13
Control Test	0.079	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 11/09/2020 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BRIAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WARREN,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 11/22/2020 02:18:47

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.079	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.124	02:14
Breath Volume	2.808	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.124	02:17
Breath Volume	2.867	02:17
Air Blank	0.000	02:17
Control Test	0.079	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 12/09/2020 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 10/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 11/01/2020 03:30:25

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.125	03:26
Breath Volume	2.753	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.122	03:28
Breath Volume	2.187	03:28
Air Blank	0.000	03:29
Control Test	0.083	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 11/05/2020 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUCHOCKI,RAYMOND,J
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: JOHNSON,TREVOR
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: RENATO

DATE/TIME OF TEST: 11/14/2020 02:30:43

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:27	
Control Test	0.082	02:28	
Air Blank	0.000	02:28	
Subject Sample # 1	0.223	02:29	
Breath Volume	1.503	02:29	
Air Blank		02:29	Purge Fail
Air Blank	0.000	02:30	

Data Download Date/Time: 12/08/2020 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUCHOCKI,RAYMOND,J
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: JOHNSON,TREVOR
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: RENATO

DATE/TIME OF TEST: 11/14/2020 02:37:04

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.079	02:34	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.246	02:35	
Air Blank	0.000	02:36	
Control Test	0.081	02:36	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 12/08/2020 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: GLOVER,ELIZABETH
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 11/14/2020 03:28:33

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.081	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1 0.239	03:27	
Breath Volume	2.292	03:27	
Air Blank		03:27	Purge Fail
Air Blank	0.000	03:28	

Data Download Date/Time: 12/08/2020 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: GLOVER,ELIZABETH
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 11/14/2020 03:36:36

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.229	03:32
Breath Volume	1.183	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.235	03:35
Breath Volume	2.613	03:35
Air Blank	0.000	03:35
Control Test	0.081	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 12/08/2020 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MARTINA,RYAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: BAGIAS

DATE/TIME OF TEST: 11/28/2020 00:33:49

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.081	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.159	00:29
Breath Volume	3.792	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:31
Subject Sample # 2	0.149	00:32
Breath Volume	3.375	00:32
Air Blank	0.000	00:32
Control Test	0.082	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 12/08/2020 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SODERHOLM,CHELSEA,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SODERHOLM,CHELSEA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:04

SUBJECT'S LAST NAME: DONALDSON

DATE/TIME OF TEST: 11/28/2020 15:56:08

Diagnostics Check		OK	15:46
Air Blank		0.000	15:46
Control Test		0.081	15:47
Air Blank		0.000	15:47
Subject Sample	# 1	0.347	15:48
Breath Volume		1.992	15:48
Air Blank		0.000	15:48
Air Blank		0.000	15:50
Subject Sample	# 2	0.378	15:50
Breath Volume		1.308	15:50
Air Blank		0.000	15:51
Air Blank		0.000	15:53
Subject Sample	# 3	0.350	15:54
Breath Volume		1.601	15:54
Air Blank		0.000	15:55
Control Test		0.081	15:55
Air Blank		0.000	15:56
Diagnostics Check		OK	15:56

Data Download Date/Time: 12/08/2020 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 11/06/2020 00:08:02

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.080	00:02
Air Blank	0.000	00:02
Subject Sample # 1	0.124	00:03
Breath Volume	4.734	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:05
Subject Sample # 2	0.122	00:06
Breath Volume	4.921	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:08

Data Download Date/Time: 12/08/2020 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: SHIVER

DATE/TIME OF TEST: 11/08/2020 03:39:23

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.100	03:34
Breath Volume	3.082	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.101	03:37
Breath Volume	3.160	03:37
Air Blank	0.000	03:38
Control Test	0.081	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 12/08/2020 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MARTINA, RYAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: HORACE

DATE/TIME OF TEST: 11/08/2020 05:02:16

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.080	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.093	04:57
Breath Volume	2.148	04:57
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.094	05:00
Breath Volume	2.246	05:00
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 12/08/2020 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD,LEE,M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WILLIAMS,GARRETT
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: CRISP

DATE/TIME OF TEST: 11/09/2020 15:42:17

Diagnostics Check	OK	15:35
Air Blank	0.000	15:36
Control Test	0.080	15:36
Air Blank	0.000	15:37
Subject Sample # 1	0.000	15:37
Breath Volume	1.910	15:37
Air Blank	0.000	15:38
Air Blank	0.000	15:40
Subject Sample # 2	0.000	15:40
Breath Volume	1.148	15:40
Air Blank	0.000	15:41
Control Test	0.081	15:41
Air Blank	0.000	15:42
Diagnostics Check	OK	15:42

Data Download Date/Time: 12/08/2020 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: STUTZMAN

DATE/TIME OF TEST: 11/10/2020 23:08:11

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.080	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.154	23:03
Breath Volume	3.078	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.148	23:06
Breath Volume	3.250	23:06
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 12/08/2020 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITAKER,HUNTER,H
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: DEMING,W
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ETHRIDGE

DATE/TIME OF TEST: 11/12/2020 01:56:41

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.079	01:50
Air Blank		0.000	01:51
Subject Sample	# 1	0.213	01:52
Breath Volume		1.703	01:52
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.204	01:55
Breath Volume		1.546	01:55
Air Blank		0.000	01:55
Control Test		0.079	01:56
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 12/08/2020 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITAKER,HUNTER,H
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SHECKLES,M
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 11/20/2020 05:42:10

Diagnostics Check	OK	05:35
Air Blank	0.000	05:36
Control Test	0.080	05:36
Air Blank	0.000	05:37
Subject Sample # 1	0.226	05:37
Breath Volume	2.363	05:37
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample # 2	0.228	05:40
Breath Volume	3.144	05:40
Air Blank	0.000	05:41
Control Test	0.080	05:41
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

Data Download Date/Time: 12/08/2020 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,R,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: STEELE,R
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: CARRIDGE

DATE/TIME OF TEST: 11/26/2020 02:13:27

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.142	02:09
Breath Volume	1.976	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.137	02:11
Breath Volume	2.085	02:11
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 12/08/2020 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 11/02/2020 18:35:36

Diagnostics Check	OK	18:28
Air Blank	0.000	18:28
Control Test	0.081	18:28
Air Blank	0.000	18:29
Subject Sample # 1	0.342	18:30
Breath Volume	1.726	18:30
Air Blank	0.000	18:31
Air Blank	0.000	18:33
Subject Sample # 2	0.352	18:34
Breath Volume	2.097	18:34
Air Blank	0.000	18:34
Control Test	0.080	18:34
Air Blank	0.000	18:35
Diagnostics Check	OK	18:35

Data Download Date/Time: 11/24/2020 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: MAPPWILLIAMS

DATE/TIME OF TEST: 11/05/2020 02:39:36

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.079	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.112	02:35
Breath Volume	2.839	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.115	02:38
Breath Volume	2.949	02:38
Air Blank	0.000	02:38
Control Test	0.079	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 11/05/2020 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARPENTER, BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 11/06/2020 01:37:02

Diagnostics Check		OK	01:29
Air Blank		0.000	01:30
Control Test		0.079	01:30
Air Blank		0.000	01:31
Subject Sample	# 1	0.236	01:32
Breath Volume		1.593	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.243	01:35
Breath Volume		1.585	01:35
Air Blank		0.000	01:36
Control Test		0.077	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:37

Data Download Date/Time: 12/07/2020 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHAOUNI,KALEY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: SPARKMAN

DATE/TIME OF TEST: 11/06/2020 02:34:42

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.078	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.250	02:30
Breath Volume	2.820	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.249	02:33
Breath Volume	2.535	02:33
Air Blank	0.000	02:33
Control Test	0.078	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 12/07/2020 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOREL,ABIGAIL
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: KIRBY

DATE/TIME OF TEST: 11/06/2020 03:49:42

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.067	03:45
Breath Volume	1.195	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.065	03:48
Breath Volume	1.117	03:48
Air Blank	0.000	03:48
Control Test	0.079	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 12/07/2020 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FAGUNDES,FELIPE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: CLEARWATERS

DATE/TIME OF TEST: 11/07/2020 03:32:11

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.079	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.211	03:27
Breath Volume		2.730	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.212	03:30
Breath Volume		2.917	03:30
Air Blank		0.000	03:31
Control Test		0.080	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 12/07/2020 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ,YAISEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: LOUIDOR

DATE/TIME OF TEST: 11/07/2020 05:54:28

Diagnostics Check		OK	05:48
Air Blank		0.000	05:48
Control Test		0.079	05:49
Air Blank		0.000	05:49
Subject Sample	# 1	0.184	05:50
Breath Volume		2.812	05:50
Air Blank		0.000	05:50
Air Blank		0.000	05:52
Subject Sample	# 2	0.184	05:53
Breath Volume		1.234	05:53
Air Blank		0.000	05:53
Control Test		0.080	05:53
Air Blank		0.000	05:54
Diagnostics Check		OK	05:54

Data Download Date/Time: 12/07/2020 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOREL,ABIGAIL
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: LODATO

DATE/TIME OF TEST: 11/08/2020 01:06:39

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.079	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.145	01:02
Breath Volume		1.875	01:02
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.141	01:05
Breath Volume		1.523	01:05
Air Blank		0.000	01:05
Control Test		0.079	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 12/07/2020 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDERSON,JAMES
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: ALBA

DATE/TIME OF TEST: 11/08/2020 02:58:31

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.097	02:54
Breath Volume	1.683	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.098	02:57
Breath Volume	1.207	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 12/07/2020 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORSO,ANTHONY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: COLLERAN

DATE/TIME OF TEST: 11/08/2020 03:41:16

Diagnostics Check		OK	03:34
Air Blank		0.000	03:35
Control Test		0.079	03:35
Air Blank		0.000	03:36
Subject Sample	# 1	0.277	03:36
Breath Volume		1.671	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.281	03:39
Breath Volume		3.093	03:39
Air Blank		0.000	03:40
Control Test		0.079	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 12/07/2020 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: STOLK,BRUCE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 11/09/2020 05:13:08

Diagnostics Check		OK	05:05
Air Blank		0.000	05:05
Control Test		0.080	05:05
Air Blank		0.000	05:06
Subject Sample	# 1	0.232	05:08
Breath Volume		2.003	05:08
Air Blank		0.000	05:09
Air Blank		0.000	05:11
Subject Sample	# 2	0.222	05:11
Breath Volume		2.320	05:11
Air Blank		0.000	05:12
Control Test		0.079	05:12
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

Data Download Date/Time: 12/07/2020 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRYANT,ALFONZIA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: AVILES

DATE/TIME OF TEST: 11/10/2020 04:06:43

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.079	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.144	04:02
Breath Volume	1.125	04:02
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.157	04:05
Breath Volume	2.035	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 12/07/2020 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BELFORD,DAVID
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:46

SUBJECT'S LAST NAME: CALEROALVAREZ

DATE/TIME OF TEST: 11/11/2020 14:17:59

Diagnostics Check	OK	14:11
Air Blank	0.000	14:12
Control Test	0.079	14:12
Air Blank	0.000	14:12
Subject Sample # 1	0.000	14:13
Breath Volume	2.035	14:13
Air Blank	0.000	14:14
Air Blank	0.000	14:16
Subject Sample # 2	0.000	14:16
Breath Volume	1.730	14:16
Air Blank	0.000	14:17
Control Test	0.080	14:17
Air Blank	0.000	14:17
Diagnostics Check	OK	14:17

Data Download Date/Time: 12/07/2020 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: EDWARDS,ANTHONY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: BAMACABAMACA

DATE/TIME OF TEST: 11/13/2020 04:27:00

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.188	04:22
Breath Volume		2.929	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.189	04:25
Breath Volume		3.386	04:25
Air Blank		0.000	04:26
Control Test		0.079	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:27

Data Download Date/Time: 12/07/2020 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIDEAUX,ADAM
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/13/2020 19:50:41

Diagnostics Check	OK	19:41
Air Blank	0.000	19:41
Control Test	0.079	19:41
Air Blank	0.000	19:42
Subject Sample # 1	0.254	19:42
Breath Volume	1.585	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.275	19:46
Breath Volume	2.242	19:46
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 3	0.268	19:49
Breath Volume	1.343	19:49
Air Blank	0.000	19:49
Control Test	0.079	19:50
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

Data Download Date/Time: 12/07/2020 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LEWIS,RYAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: JIMENEZVILLAHERMOSA

DATE/TIME OF TEST: 11/14/2020 01:24:56

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.182	01:20
Breath Volume	2.589	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:22
Subject Sample # 2	0.180	01:23
Breath Volume	2.535	01:23
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 12/07/2020 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ,STEPHANIE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 11/14/2020 20:11:23

Diagnostics Check	OK	20:05
Air Blank	0.000	20:05
Control Test	0.080	20:05
Air Blank	0.000	20:06
Subject Sample # 1	0.162	20:06
Breath Volume	1.910	20:06
Air Blank	0.000	20:07
Air Blank	0.000	20:09
Subject Sample # 2	0.155	20:09
Breath Volume	3.402	20:09
Air Blank	0.000	20:10
Control Test	0.080	20:10
Air Blank	0.000	20:11
Diagnostics Check	OK	20:11

Data Download Date/Time: 12/07/2020 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 11/15/2020 04:36:32

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.079	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.151	04:31
Breath Volume	1.144	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.149	04:35
Breath Volume	1.683	04:35
Air Blank	0.000	04:35
Control Test	0.079	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARPENTER, BLAKE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 11/15/2020 05:00:33

Diagnostics Check	OK	04:54
Air Blank	0.000	04:54
Control Test	0.079	04:54
Air Blank	0.000	04:55
Subject Sample # 1	0.120	04:56
Breath Volume	2.953	04:56
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.118	04:59
Breath Volume	2.609	04:59
Air Blank	0.000	04:59
Control Test	0.079	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MAYFIELD,JOHN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: MATEO

DATE/TIME OF TEST: 11/17/2020 10:56:11

Diagnostics Check	OK	10:50
Air Blank	0.000	10:50
Control Test	0.080	10:51
Air Blank	0.000	10:51
Subject Sample # 1	0.000	10:51
Breath Volume	1.503	10:51
Air Blank	0.000	10:52
Air Blank	0.000	10:54
Subject Sample # 2	0.000	10:54
Breath Volume	1.972	10:54
Air Blank	0.000	10:55
Control Test	0.080	10:55
Air Blank	0.000	10:56
Diagnostics Check	OK	10:56

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CERVELLERA, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 11/18/2020 21:58:36

Diagnostics Check	OK	21:50
Air Blank	0.000	21:51
Control Test	0.078	21:51
Air Blank	0.000	21:52
Subject Sample # 1	0.308	21:54
Breath Volume	2.062	21:54
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.314	21:57
Breath Volume	1.648	21:57
Air Blank	0.000	21:57
Control Test	0.079	21:58
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCHOEN,ELLIS
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: CHOC LOPEZ

DATE/TIME OF TEST: 11/18/2020 22:59:33

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.079	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.203	22:55
Breath Volume	2.574	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.198	22:58
Breath Volume	1.191	22:58
Air Blank	0.000	22:58
Control Test	0.078	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,BRIAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: HUBBS

DATE/TIME OF TEST: 11/21/2020 01:25:11

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.079	01:21	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:23	Subject Test Refused
Breath Volume	0.000	01:23	
Air Blank	0.000	01:24	
Control Test	0.081	01:24	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VENTURA, CHRISTOPHER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: AZCANIOMANTILLA

DATE/TIME OF TEST: 11/21/2020 02:05:50

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.080	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.167	02:01
Breath Volume	1.234	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.162	02:04
Breath Volume	1.320	02:04
Air Blank	0.000	02:04
Control Test	0.080	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 11/21/2020 02:47:54

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.039	02:43
Breath Volume	2.132	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:45
Subject Sample # 2	0.039	02:46
Breath Volume	2.523	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOMINGUEZ,JAYMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SASTOQUEALEJO

DATE/TIME OF TEST: 11/21/2020 03:41:59

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.170	03:37
Breath Volume	1.402	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:39
Subject Sample # 2	0.155	03:40
Breath Volume	1.222	03:40
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARBONE,ANDRES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 11/21/2020 04:37:18

Diagnostics Check	OK	04:34	
Air Blank	0.000	04:34	
Control Test	0.080	04:35	
Air Blank	0.000	04:35	
Subject Sample	# 1	04:35	Subject Test Refused
Breath Volume	0.000	04:35	
Air Blank	0.000	04:36	
Control Test	0.080	04:36	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SINGSON, ROLAND
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: CETOLENE

DATE/TIME OF TEST: 11/21/2020 18:21:22

Diagnostics Check	OK	18:14
Air Blank	0.000	18:14
Control Test	0.079	18:15
Air Blank	0.000	18:15
Subject Sample # 1	0.125	18:17
Breath Volume	2.640	18:17
Air Blank	0.000	18:17
Air Blank	0.000	18:19
Subject Sample # 2	0.118	18:19
Breath Volume	1.910	18:19
Air Blank	0.000	18:20
Control Test	0.080	18:20
Air Blank	0.000	18:21
Diagnostics Check	OK	18:21

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ, YAISEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: RICKERT

DATE/TIME OF TEST: 11/22/2020 03:06:30

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.079	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.172	03:02
Breath Volume		1.316	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.168	03:05
Breath Volume		1.496	03:05
Air Blank		0.000	03:05
Control Test		0.080	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRAZEE,COLE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 11/22/2020 23:12:22

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.152	23:07
Breath Volume	3.515	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.156	23:10
Breath Volume	2.562	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HEBEL,GRIFFIN
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: BARTHELMASS

DATE/TIME OF TEST: 11/23/2020 04:20:24

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.181	04:15
Breath Volume	1.777	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.183	04:18
Breath Volume	1.148	04:18
Air Blank	0.000	04:19
Control Test	0.078	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPARKS,ELIZABETH
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: GALINDEZ

DATE/TIME OF TEST: 11/24/2020 04:29:48

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.079	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.106	04:25
Breath Volume	1.203	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:27
Subject Sample # 2	0.114	04:28
Breath Volume	1.695	04:28
Air Blank	0.000	04:28
Control Test	0.079	04:29
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOLMES,MARIAH
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 11/24/2020 23:01:28

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.079	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.189	22:57
Breath Volume	1.527	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.185	23:00
Breath Volume	1.113	23:00
Air Blank	0.000	23:00
Control Test	0.079	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WISECUP,AUSTIN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 11/25/2020 00:42:36

Diagnostics Check		OK	00:29	
Air Blank		0.000	00:30	
Control Test		0.079	00:30	
Air Blank		0.000	00:30	
Subject Sample	# 1	0.143	00:34	Volume Not Met
Breath Volume		0.132	00:34	
Air Blank		0.000	00:34	
Air Blank		0.000	00:36	
Subject Sample	# 2	0.127	00:37	
Breath Volume		1.210	00:37	
Air Blank		0.000	00:38	
Air Blank		0.000	00:40	
Subject Sample	# 3	0.134	00:41	
Breath Volume		1.222	00:41	
Air Blank		0.000	00:41	
Control Test		0.078	00:42	
Air Blank		0.000	00:42	
Diagnostics Check		OK	00:42	

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 11/26/2020 00:00:43

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.079	23:55
Air Blank	0.000	23:55
Subject Sample # 1	0.160	23:56
Breath Volume	1.261	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.161	23:59
Breath Volume	1.210	23:59
Air Blank	0.000	23:59
Control Test	0.080	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DEL CASTILLO,BRET
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: CHEESMAN

DATE/TIME OF TEST: 11/26/2020 02:09:15

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.080	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.247	02:04
Breath Volume		1.667	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:06
Subject Sample	# 2	0.231	02:07
Breath Volume		1.199	02:07
Air Blank		0.000	02:08
Control Test		0.079	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HEBEL,GRIFFIN
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: LUCENA MIRANDA

DATE/TIME OF TEST: 11/26/2020 05:26:55

Diagnostics Check		OK	05:20
Air Blank		0.000	05:21
Control Test		0.080	05:21
Air Blank		0.000	05:22
Subject Sample	# 1	0.148	05:22
Breath Volume		1.488	05:22
Air Blank		0.000	05:22
Air Blank		0.000	05:24
Subject Sample	# 2	0.150	05:25
Breath Volume		2.250	05:25
Air Blank		0.000	05:26
Control Test		0.080	05:26
Air Blank		0.000	05:26
Diagnostics Check		OK	05:26

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WATKINS,ROMAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: HILLMAN

DATE/TIME OF TEST: 11/27/2020 21:12:31

Diagnostics Check	OK	21:04
Air Blank	0.000	21:05
Control Test	0.079	21:05
Air Blank	0.000	21:05
Subject Sample # 1	0.168	21:07
Breath Volume	1.683	21:07
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 2	0.170	21:11
Breath Volume	1.578	21:11
Air Blank	0.000	21:11
Control Test	0.079	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: STARNES,JACOB
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 11/28/2020 04:41:02

Diagnostics Check		OK	04:29	
Air Blank		0.000	04:29	
Control Test		0.080	04:30	
Air Blank		0.000	04:30	
Subject Sample	# 1	0.143	04:33	Volume Not Met
Breath Volume		0.066	04:33	
Air Blank		0.000	04:34	
Air Blank		0.000	04:36	
Subject Sample	# 2	0.137	04:39	Volume Not Met
Breath Volume		0.382	04:39	
Air Blank		0.000	04:40	
Control Test		0.079	04:40	
Air Blank		0.000	04:40	
Diagnostics Check		OK	04:41	

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: STARNES,JACOB
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 11/28/2020 04:55:16

Diagnostics Check		OK	04:43	
Air Blank		0.000	04:44	
Control Test		0.079	04:44	
Air Blank		0.000	04:45	
Subject Sample	# 1	0.114	04:48	Volume Not Met
Breath Volume		0.738	04:48	
Air Blank		0.000	04:48	
Air Blank		0.000	04:50	
Subject Sample	# 2	0.122	04:53	Volume Not Met
Breath Volume		0.035	04:53	
Air Blank		0.000	04:54	
Control Test		0.079	04:54	
Air Blank		0.000	04:55	
Diagnostics Check		OK	04:55	

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RUSSO,ANDREW
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MEJIAARGUETA

DATE/TIME OF TEST: 11/29/2020 03:07:19

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.080	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.097	03:02
Breath Volume	1.593	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.091	03:05
Breath Volume	1.562	03:05
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LEE,MIKEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: ROCHE

DATE/TIME OF TEST: 11/30/2020 02:41:49

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:30	
Control Test		0.079	02:31	
Air Blank		0.000	02:31	
Subject Sample	# 1	0.170	02:34	Volume Not Met
Breath Volume		0.472	02:34	
Air Blank		0.000	02:35	
Air Blank		0.000	02:37	
Subject Sample	# 2	0.194	02:40	Volume Not Met
Breath Volume		0.359	02:40	
Air Blank		0.000	02:40	
Control Test		0.079	02:41	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 12/07/2020 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 11/15/2020 17:01:13

Diagnostics Check	OK	16:54
Air Blank	0.000	16:54
Control Test	0.079	16:55
Air Blank	0.000	16:55
Subject Sample # 1	0.260	16:56
Breath Volume	2.152	16:56
Air Blank	0.000	16:57
Air Blank	0.000	16:58
Subject Sample # 2	0.240	16:59
Breath Volume	1.265	16:59
Air Blank	0.000	17:00
Control Test	0.079	17:00
Air Blank	0.000	17:01
Diagnostics Check	OK	17:01

Data Download Date/Time: 12/16/2020 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 11/21/2020 01:49:11

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.079	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.056	01:44
Breath Volume	4.367	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:46
Subject Sample # 2	0.053	01:47
Breath Volume	3.839	01:47
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 12/16/2020 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 11/22/2020 05:30:42

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.079	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.163	05:24
Breath Volume	2.195	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.153	05:29
Breath Volume	1.968	05:29
Air Blank	0.000	05:29
Control Test	0.080	05:30
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 12/16/2020 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUCHNIR,EUGENE,V
BREATH TEST OPERATOR AGENCY: Quincy PD
ARRESTING OFFICER: KOUCHNIR,EUGENE
ARRESTING OFFICER AGENCY: Quincy PD
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 11/22/2020 22:49:41

Diagnostics Check	OK	22:43
Air Blank	0.000	22:44
Control Test	0.078	22:44
Air Blank	0.000	22:44
Subject Sample # 1	0.201	22:45
Breath Volume	1.656	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:47
Subject Sample # 2	0.211	22:48
Breath Volume	1.382	22:48
Air Blank	0.000	22:48
Control Test	0.078	22:49
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 12/16/2020 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JOSHUA,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DIXON,MARCUS
ARRESTING OFFICER AGENCY: Chattahoochee PD
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: VALLE-LANDAVERDE

DATE/TIME OF TEST: 11/24/2020 00:28:17

Diagnostics Check		OK	00:17
Air Blank		0.000	00:18
Control Test		0.079	00:18
Air Blank		0.000	00:18
Subject Sample	# 1	0.368	00:19
Breath Volume		2.605	00:19
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.345	00:23
Breath Volume		2.429	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 3	0.342	00:26
Breath Volume		2.277	00:26
Air Blank		0.000	00:27
Control Test		0.078	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 12/16/2020 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/25/2020 21:51:41

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:42	
Control Test	0.079	21:42	
Air Blank	0.000	21:42	
Subject Sample	# 1 0.096	21:46	Volume Not Met
Breath Volume	0.734	21:46	
Air Blank	0.000	21:46	
Air Blank	0.000	21:48	
Subject Sample	# 2	21:50	Subject Test Refused
Breath Volume	0.277	21:50	
Air Blank	0.000	21:50	
Control Test	0.080	21:51	
Air Blank	0.000	21:51	
Diagnostics Check	OK	21:51	

Data Download Date/Time: 12/16/2020 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/25/2020 22:00:31

Diagnostics Check	OK	21:54	
Air Blank	0.000	21:55	
Control Test	0.079	21:55	
Air Blank	0.000	21:56	
Subject Sample	# 1	21:58	Subject Test Refused
Breath Volume	0.812	21:58	
Air Blank	0.000	21:59	
Control Test	0.081	21:59	
Air Blank	0.000	22:00	
Diagnostics Check	OK	22:00	

Data Download Date/Time: 12/16/2020 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRODTMAN,ERIC,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: TANNER,BENJAMIN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: ZELAYA-PIZARRO

DATE/TIME OF TEST: 11/05/2020 04:07:44

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.130	04:01
Breath Volume	1.312	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.117	04:05
Breath Volume	1.222	04:05
Air Blank	0.000	04:06
Control Test	0.078	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 11/23/2020 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUVRES,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: JOHNSON,ANDREW
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 10/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: GARCIA ROSADO

DATE/TIME OF TEST: 11/07/2020 04:44:28

Diagnostics Check	OK	04:38
Air Blank	0.000	04:38
Control Test	0.080	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.083	04:39
Breath Volume	1.531	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.088	04:42
Breath Volume	1.417	04:42
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 11/23/2020 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUVRES,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: JOHNSON,ANDREW
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 10/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/08/2020 00:11:47

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.080	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.104	00:07
Breath Volume	3.105	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.103	00:10
Breath Volume	3.210	00:10
Air Blank	0.000	00:10
Control Test	0.080	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 11/23/2020 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUVRES,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: DOUVRES,MICHAEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 10/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: PICON COLON

DATE/TIME OF TEST: 11/08/2020 03:31:09

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:16	
Control Test		0.080	03:16	
Air Blank		0.000	03:17	
Subject Sample	# 1	0.034	03:18	
Breath Volume		1.316	03:18	
Air Blank		0.000	03:18	
Air Blank		0.000	03:20	
Subject Sample	# 2	0.031	03:23	Volume Not Met
Breath Volume		1.015	03:23	
Air Blank		0.000	03:24	
Air Blank		0.000	03:26	
Subject Sample	# 3	0.030	03:29	Volume Not Met
Breath Volume		0.222	03:29	
Air Blank		0.000	03:30	
Control Test		0.080	03:30	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 11/23/2020 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUVRES,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: JOHNSON,ANDREW
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 10/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: FAIRLEY

DATE/TIME OF TEST: 11/12/2020 21:29:00

Diagnostics Check	OK	21:21
Air Blank	0.000	21:22
Control Test	0.080	21:22
Air Blank	0.000	21:23
Subject Sample # 1	0.241	21:24
Breath Volume	2.496	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:26
Subject Sample # 2	0.244	21:27
Breath Volume	2.562	21:27
Air Blank	0.000	21:28
Control Test	0.079	21:28
Air Blank	0.000	21:28
Diagnostics Check	OK	21:29

Data Download Date/Time: 11/23/2020 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,BRIAN,G
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: LANZO,MIGUEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 10/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 11/14/2020 05:15:21

Diagnostics Check		OK	05:03	
Air Blank		0.000	05:04	
Control Test		0.078	05:04	
Air Blank		0.000	05:05	
Subject Sample	# 1	0.181	05:08	Volume Not Met
Breath Volume		0.144	05:08	
Air Blank		0.000	05:08	
Air Blank		0.000	05:10	
Subject Sample	# 2	0.175	05:13	Volume Not Met
Breath Volume		0.968	05:13	
Air Blank		0.000	05:14	
Control Test		0.077	05:14	
Air Blank		0.000	05:15	
Diagnostics Check		OK	05:15	

Data Download Date/Time: 11/23/2020 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHOBREY,SPENCER,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHOBREY,SPENCER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 11/19/2020 23:30:59

Diagnostics Check	OK	23:22
Air Blank	0.000	23:23
Control Test	0.076	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.127	23:26
Breath Volume	1.667	23:26
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.120	23:29
Breath Volume	1.582	23:29
Air Blank	0.000	23:30
Control Test	0.077	23:30
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 11/23/2020 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ENFINGER

DATE/TIME OF TEST: 11/20/2020 23:42:18

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.200	23:37
Breath Volume	4.429	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.194	23:40
Breath Volume	3.136	23:40
Air Blank	0.000	23:41
Control Test	0.078	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 11/23/2020 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES, RONNIE,
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: BARRETT, ASHLEY
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 10/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 11/23/2020 00:37:34

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.080	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.000	00:32
Breath Volume	2.144	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.000	00:36
Breath Volume	2.316	00:36
Air Blank	0.000	00:36
Control Test	0.081	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 11/23/2020 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUVRES,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: DOUVRES,MICHAEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: REAVES

DATE/TIME OF TEST: 11/27/2020 01:49:05

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.079	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.251	01:44
Breath Volume	2.214	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:46
Subject Sample # 2	0.247	01:47
Breath Volume	2.132	01:47
Air Blank	0.000	01:48
Control Test	0.078	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:49

Data Download Date/Time: 12/15/2020 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORD, GLENN, W
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: FORD, GLENN
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: FEBO ORTIZ

DATE/TIME OF TEST: 11/30/2020 18:38:48

Diagnostics Check	OK	18:30
Air Blank	0.000	18:31
Control Test	0.081	18:31
Air Blank	0.000	18:32
Subject Sample # 1	0.127	18:32
Breath Volume	1.320	18:32
Air Blank	0.000	18:33
Air Blank	0.000	18:35
Subject Sample # 2	0.124	18:37
Breath Volume	1.253	18:37
Air Blank	0.000	18:37
Control Test	0.079	18:38
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

Data Download Date/Time: 12/15/2020 09:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,ISAAC,N
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: REYES,ISAAC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 07:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:22

SUBJECT'S LAST NAME: GUILLAUME

DATE/TIME OF TEST: 11/07/2020 09:51:41

Diagnostics Check		OK	09:42
Air Blank		0.000	09:42
Control Test		0.081	09:43
Air Blank		0.000	09:43
Subject Sample	# 1	0.188	09:44
Breath Volume		2.667	09:44
Air Blank		0.000	09:44
Air Blank		0.000	09:46
Subject Sample	# 2	0.167	09:47
Breath Volume		1.562	09:47
Air Blank		0.000	09:47
Air Blank		0.000	09:49
Subject Sample	# 3	0.166	09:50
Breath Volume		2.257	09:50
Air Blank		0.000	09:50
Control Test		0.080	09:51
Air Blank		0.000	09:51
Diagnostics Check		OK	09:51

Data Download Date/Time: 11/30/2020 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLEWELLYN,A.,
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: STANLEY,W.
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 10/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: COOKSEY

DATE/TIME OF TEST: 11/09/2020 20:01:31

Diagnostics Check	OK	19:55
Air Blank	0.000	19:55
Control Test	0.079	19:55
Air Blank	0.000	19:56
Subject Sample # 1	0.196	19:56
Breath Volume	3.347	19:56
Air Blank	0.000	19:57
Air Blank	0.000	19:59
Subject Sample # 2	0.191	19:59
Breath Volume	3.187	19:59
Air Blank	0.000	20:00
Control Test	0.080	20:00
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

Data Download Date/Time: 11/26/2020 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERDUE,JOSHUA,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: PERDUE,JOSHUA
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 10/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: TARLETON

DATE/TIME OF TEST: 11/22/2020 03:53:20

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.144	03:48
Breath Volume	1.488	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.142	03:51
Breath Volume	1.656	03:51
Air Blank	0.000	03:52
Control Test	0.081	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 11/26/2020 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,DEMETRIUS,L
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: BROWN,DEMETRIUS
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 11/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:59

SUBJECT'S LAST NAME: PEACOCK

DATE/TIME OF TEST: 11/26/2020 15:27:37

Diagnostics Check	OK	15:20
Air Blank	0.000	15:21
Control Test	0.080	15:21
Air Blank	0.000	15:21
Subject Sample # 1	0.291	15:22
Breath Volume	3.433	15:22
Air Blank	0.000	15:23
Air Blank	0.000	15:25
Subject Sample # 2	0.287	15:26
Breath Volume	2.761	15:26
Air Blank	0.000	15:26
Control Test	0.079	15:27
Air Blank	0.000	15:27
Diagnostics Check	OK	15:27

Data Download Date/Time: 12/28/2020 13:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: NELSON,BRYAN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: BARGER

DATE/TIME OF TEST: 11/03/2020 23:54:22

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.081	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.165	23:49
Breath Volume	2.222	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.163	23:52
Breath Volume	1.964	23:52
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 11/06/2020 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: AFFELD,MATHEW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/04/2020 02:26:12

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.081	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.139	02:21
Breath Volume	1.785	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:23
Subject Sample # 2	0.129	02:24
Breath Volume	1.203	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 11/06/2020 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: WYATT,WILLIAM
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 11/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: NATAREZ CLAROS

DATE/TIME OF TEST: 11/22/2020 03:19:10

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.081	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.157	03:14
Breath Volume	1.941	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.158	03:17
Breath Volume	2.761	03:17
Air Blank	0.000	03:18
Control Test	0.081	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 11/23/2020 08:47